

# Aberration-corrected high-NA open-top selective-plane illumination microscopy for biological imaging

RYAN MCGORTY,<sup>1</sup> DAN XIE,<sup>2</sup> AND BO HUANG<sup>2,3,\*</sup>

<sup>1</sup> Dept. of Physics and Biophysics, University of San Diego, CA 92110, USA

<sup>2</sup> Dept. of Pharmaceutical Chemistry, University of California, San Francisco, CA 94158, USA

<sup>3</sup> Dept. of Biochemistry and Biophysics, University of California, San Francisco, CA 94158, USA

\*Bo.Huang@ucsf.edu

**Abstract:** Selective-plane illumination microscopy (SPIM) provides unparalleled advantages for volumetric imaging of living organisms over extended times. However, the spatial configuration of a SPIM system often limits its compatibility with many widely used biological sample holders such as multi-well chambers and plates. To solve this problem, we developed a high numerical aperture (NA) open-top configuration that places both the excitation and detection objectives on the opposite of the sample coverglass. We carried out a theoretical calculation to analyze the structure of the system-induced aberrations. We then experimentally showed aberration-correction using adaptive optics combined with static optical components, demonstrating near-diffraction-limited performance in imaging fluorescently labeled cells.

© 2017 Optical Society of America

**OCIS codes:** (080.080) Geometric Optics; (110.0110) Imaging systems; (110.0180) Microscopy.

## References and links

1. Jan Huiskens, Jim Swoger, Filippo Del Bene, Joachim Wittbrodt, and Ernst H. K. Stelzer. Optical Sectioning Deep Inside Live Embryos by Selective Plane Illumination Microscopy. *Science*, 305(5686):1007–1009, August 2004.
2. Loïc A. Royer, William C. Lemon, Raghav K. Chhetri, Yanan Wan, Michael Coleman, Eugene W. Myers, and Philipp J. Keller. Adaptive light-sheet microscopy for long-term, high-resolution imaging in living organisms. *Nature Biotechnology*, advance online publication, October 2016.
3. Sébastien Wolf, Willy Supatto, Georges Debrégeas, Pierre Mahou, Sergei G. Kruglik, Jean-Marc Sintes, Emmanuel Beaurepaire, and Raphaël Candelier. Whole-brain functional imaging with two-photon light-sheet microscopy. *Nature Methods*, 12(5):379–380, May 2015.
4. Yicong Wu, Alireza Ghitani, Ryan Christensen, Anthony Santella, Zhuo Du, Gary Rondeau, Zhirong Bao, Daniel Colón-Ramos, and Hari Shroff. Inverted selective plane illumination microscopy (iSPIM) enables coupled cell identity lineaging and neurodevelopmental imaging in *Caenorhabditis elegans*. *Proceedings of the National Academy of Sciences of the United States of America*, 108(43):17708–17713, October 2011.
5. Bi-Chang Chen, Wesley R. Legant, Kai Wang, Lin Shao, Daniel E. Milkie, Michael W. Davidson, Chris Janetopoulos, Xufeng S. Wu, John A. Hammer, Zhe Liu, Brian P. English, Yuko Mimori-Kiyosue, Daniel P. Romero, Alex T. Ritter, Jennifer Lippincott-Schwartz, Lillian Fritz-Laylin, R. Dyche Mullins, Diana M. Mitchell, Joshua N. Bembenek, Anne-Cecile Reymann, Ralph Böhme, Stephan W. Grill, Jennifer T. Wang, Geraldine Seydoux, U. Serdar Tulu, Daniel P. Kiehart, and Eric Betzig. Lattice light-sheet microscopy: Imaging molecules to embryos at high spatiotemporal resolution. *Science*, 346(6208):1257998, October 2014.
6. Matthew B. Bouchard, Venkatakaushik Voleti, César S. Mendes, Clay Lacefield, Wesley B. Grueber, Richard S. Mann, Randy M. Bruno, and Elizabeth M. C. Hillman. Swept confocally-aligned planar excitation (SCAPE) microscopy for high-speed volumetric imaging of behaving organisms. *Nature Photonics*, 9(2):113–119, February 2015.
7. Petr Strnad, Stefan Gunther, Judith Reichmann, Uros Krzic, Balint Balazs, Gustavo de Medeiros, Nils Norlin, Takashi Hiiragi, Lars Hufnagel, and Jan Ellenberg. Inverted light-sheet microscope for imaging mouse pre-implantation development. *Nature Methods*, 13(2):139–142, February 2016.
8. Ryan McGorty, Harrison Liu, Daichi Kamiyama, Zhiqiang Dong, Su Guo, and Bo Huang. Open-top selective plane illumination microscope for conventionally mounted specimens. *Optics Express*, 23(12):16142, June 2015.
9. R. H. Freeman and J. E. Pearson. Deformable mirrors for all seasons and reasons. *Applied Optics*, 21(4):580–588, February 1982.
10. A. M. Weiner. Femtosecond pulse shaping using spatial light modulators. *Review of Scientific Instruments*, 71(5):1929–1960, May 2000.
11. Martin J. Booth. Adaptive optics in microscopy. *Philosophical Transactions of the Royal Society of London A: Mathematical, Physical and Engineering Sciences*, 365(1861):2829–2843, December 2007.

12. Martin J. Booth, Mark A. A. Neil, Rimantas Juškaitis, and Tony Wilson. Adaptive aberration correction in a confocal microscope. *Proceedings of the National Academy of Sciences*, 99(9):5788–5792, April 2002.
13. D. Débarre, T. Vieille, and E. Beaufort. Simple characterisation of a deformable mirror inside a high numerical aperture microscope using phase diversity. *Journal of Microscopy*, 244(2):136–143, November 2011.
14. P. Kner, J.W. Sedat, D.A. Agard, and Z. Kam. High-resolution wide-field microscopy with adaptive optics for spherical aberration correction and motionless focusing. *Journal of microscopy*, 237(2):136–147, February 2010.
15. Na Ji, Daniel E. Milkie, and Eric Betzig. Adaptive optics via pupil segmentation for high-resolution imaging in biological tissues. *Nature Methods*, 7(2):141–147, February 2010.
16. Delphine Débarre, Edward J. Botcherby, Tomoko Watanabe, Shankar Srinivas, Martin J. Booth, and Tony Wilson. Image-based adaptive optics for two-photon microscopy. *Optics Letters*, 34(16):2495, August 2009.
17. Ignacio Izeddin, Mohamed El Beheiry, Jordi Andilla, Daniel Ciepiewski, Xavier Darzacq, and Maxime Dahan. PSF shaping using adaptive optics for three-dimensional single-molecule super-resolution imaging and tracking. *Optics Express*, 20(5):4957, February 2012.
18. Daniel Burke, Brian Patton, Fang Huang, Joerg Bewersdorf, and Martin J. Booth. Adaptive optics correction of specimen-induced aberrations in single-molecule switching microscopy. *Optica*, 2(2):177–185, February 2015.
19. Cyril Bourgenot, Christopher D. Saunter, Jonathan M. Taylor, John M. Girkin, and Gordon D. Love. 3d adaptive optics in a light sheet microscope. *Optics Express*, 20(12):13252–13261, June 2012.
20. Diwakar Turaga and Timothy E. Holy. Aberrations and their correction in light-sheet microscopy: a low-dimensional parametrization. *Biomedical Optics Express*, 4(9):1654–1661, August 2013.
21. Bridget M. Hanser, Mats G. L. Gustafsson, David A. Agard, and John W. Sedat. Phase retrieval for high-numerical-aperture optical systems. *Optics Letters*, 28(10):801–803, May 2003.
22. B. M. Hanser, M. G. L. Gustafsson, D. A. Agard, and J. W. Sedat. Phase-retrieved pupil functions in wide-field fluorescence microscopy. *Journal of Microscopy*, 216(Pt 1):32–48, October 2004.
23. Raju Tomer, Khaled Khairy, Fernando Amat, and Philipp J. Keller. Quantitative high-speed imaging of entire developing embryos with simultaneous multiview light-sheet microscopy. *Nature Methods*, 9(7):755–763, July 2012.

---

## 1. Introduction

For over a decade selective-plane illumination microscopy (SPIM), or light-sheet microscopy, has been successful in imaging the development of whole organisms [1, 2], observing individual neurons fire across large volumes [3], and observing the dynamics of single cells [4, 5]. SPIM systems often require special sample loading approaches, in which small tubes or cylinders of agarose gel hold the sample in the typically tight space between two objectives. This configuration makes it difficult for SPIM systems to use conventional samples holders for cell microscopy, particularly for high-content imaging where microfluidic devices or multi-well plates would be the ideal. Efforts have been devoted to resolve this incompatibility between SPIM systems and high-throughput, high-resolution imaging, however, the published methods are either optically demanding [6] or require a specially manufactured sample holder wherein the specimen is supported by a thin membrane [7]. A more user-friendly design which resembles typical inverted microscopes and suites existing popular sample holders has yet to be developed.

Previously, we demonstrated an open-top SPIM configuration to work with a variety of sample holders [8]. The two objectives, one for creating the excitation light sheet and the other for imaging, are installed underneath the sample holder (Figure 1). Mounting of the imaging sample in the open-top SPIM is similar to that in an inverted microscope, where the sample is carried by a flat glass surface such as a coverslip or the glass bottom of a multi-well plate. However, this convenience of sample loading comes at the expense of image quality. Because a piece of glass tilted with respect to the optical axis is between the objective and the sample, severe optical aberrations reduce the image quality. If the open-top SPIM uses a low numerical aperture (NA) imaging objective (e.g., NA = 0.3), the geometry-based aberration would primarily consist of spherical aberration and astigmatism, both of which are analytically simple. We employed two elements to help counter these aberrations. The first was a water prism to replace the tilted air-glass-water interface with a tilted water-glass-water interface. The second measure was a cylindrical lens placed in the imaging path. With these correction methods, a resolution of about 1 Åm was achieved using a 0.3 NA objective in our open-top SPIM. While micron-level SPIM resolution is adequate in many applications, it falls short of revealing details at the subcellular

level. However, using a higher NA objective in our open-top SPIM system would exacerbate the optical aberrations and the water prism and cylindrical lens would no longer suffice for correction.

Fortunately, adaptive optics (AOs), such as deformable mirrors [9] and spatial light modulators [10], have provided a promising approach for complex aberration correction in biomedical imaging [11]. An AO device can add a prescribed phase distortion to actively compensate a distorted wavefront. In biological imaging, it is widely recognized that thick specimens can cause substantial aberrations due to spatial variations of refractive index. In the past two decades, AO devices have been implemented to address this problem of aberration in a variety of microscopy methods, including confocal microscopy [12, 13], wide-field microscopy [14], two-photon microscopy [15, 16], single-molecule imaging [17, 18] and SPIM [19, 20]. However, most existing works have focused on specimen-induced aberrations; system-induced aberrations, which typically reflect alignment imperfection rather than inherent features of imaging systems, remains a less discussed aspect.

Here, we extend the open-top SPIM concept to the case of a high-NA imaging objective with the assistance of AO techniques. Since the majority of the overall aberration is inherent in the system, we carry out a theoretical analysis to estimate its magnitude, spatial structure and major components. Then, using a phase-retrieval method, we experimentally characterize the optical aberration in the system and use a deformable mirror for correction. To test the imaging capability of the system on biological samples, we imaged fluorescently labeled cells with the preconfigured counter-aberration pattern added on the deformable mirror. Even without sample-specific corrections, considerable improvement on the resolution was achieved.

## 2. An estimation of coverslip-induced optical aberration

### 2.1. Theoretical analysis

In this section we evaluate optical aberrations induced by the presence of a tilted coverslip between a point source and a water-dipping objective. Figure 1 illustrates the geometry of the imaging system, wherein the optical axis of the objective is aligned with the  $z$ -direction. We assume that the point light source is located at the center of the field of view (FOV)<sup>1</sup>, and its emission is isotropic. The maximum cone of light that can enter the imaging system is defined by the numerical aperture (N.A.) of the objective:

$$\text{N.A.} = n_1 \sin \alpha, \quad (1)$$

where  $n_1$  is the refractive index of the immersion medium. Assume that the coverslip has thickness  $d$ , refractive index  $n_2$ , and orientation  $\hat{\mathbf{n}}$ . Without loss of generality, we can let  $\hat{\mathbf{n}}$  be contained in the  $x$ - $z$  plane, i.e.,

$$\hat{\mathbf{n}} = \sin \gamma \hat{\mathbf{e}}_x + \cos \gamma \hat{\mathbf{e}}_z, \quad (2)$$

where  $\gamma$  is the tilt angle of the coverslip with respect to the  $x$ - $y$  plane. Within the collection cone of the objective, the direction of a light ray,  $\hat{\mathbf{k}}$ , can be specified by the azimuth angle  $\theta$  and the polar angle  $\varphi$ :

$$\hat{\mathbf{k}} = \sin \theta \cos \varphi \hat{\mathbf{e}}_x + \sin \theta \sin \varphi \hat{\mathbf{e}}_y + \cos \theta \hat{\mathbf{e}}_z, \quad (3)$$

and  $\theta$  satisfies the condition  $\alpha + \pi/2 \leq \theta \leq \pi$ . On the upper water-glass interface, the angle of incidence,  $\phi$ , is determined by

$$\sin \phi = |\hat{\mathbf{k}} \times \hat{\mathbf{n}}|, \quad (4)$$

$$\cos \phi = |\hat{\mathbf{k}} \cdot \hat{\mathbf{n}}|. \quad (5)$$

<sup>1</sup>The phase change on the pupil plane caused by a displacement of the point source from the center of FOV can be represented by the first 4 Zernike modes, and thus does not introduce additional optical aberrations.

In addition, the  $s$ - and  $p$ - directions with respect to the plane of incidence is determined by

$$\hat{\mathbf{s}} = (\hat{\mathbf{k}} \times \hat{\mathbf{n}}) / |\hat{\mathbf{k}} \times \hat{\mathbf{n}}| = \csc \phi (\hat{\mathbf{k}} \times \hat{\mathbf{n}}), \quad (6)$$

$$\hat{\mathbf{p}} = \hat{\mathbf{s}} \times \hat{\mathbf{k}}. \quad (7)$$

According to Snell's law, the angle of refraction,  $\phi'$ , is associated with  $\phi$  by

$$n_1 \sin \phi = n_2 \sin \phi', \quad (8)$$

and  $\phi'$  can be calculated by substituting Eq. (5) into Eq. (8). Let  $\beta = \phi - \phi'$ , then in the  $\hat{\mathbf{s}} - \hat{\mathbf{k}} - \hat{\mathbf{p}}$  coordinates, the direction vector of the refracted ray,  $\hat{\mathbf{k}}'$ , can be denoted by

$$\hat{\mathbf{k}}' = \cos \beta \hat{\mathbf{k}} - \sin \beta \hat{\mathbf{p}}. \quad (9)$$

The upper water-coverslip interface can be described by the equation of plane

$$\mathbf{r} - \mathbf{r}_0 \cdot \hat{\mathbf{n}} = 0, \quad (10)$$

where  $\mathbf{r}_0$  is an arbitrary point in the plane. If we set the origin at the point of incidence, then  $\mathbf{r}_0$  can be set as  $\mathbf{0}$ , and Eq. (10) is simplified as

$$x \sin \gamma + z \cos \gamma = 0. \quad (11)$$

The lower coverslip-water interface, which is translated from the upper interface in the direction of  $\hat{\mathbf{n}}$  by  $-d$ , can be described by the equation of plane

$$\mathbf{r} - \mathbf{r}_d \cdot \hat{\mathbf{n}} = 0, \quad (12)$$

where  $\mathbf{r}_d = -d\hat{\mathbf{n}} + \mathbf{r}_0$ . Substituting Eq. (11) into Eq. (12), the latter can be simplified as

$$\mathbf{r} \cdot \hat{\mathbf{n}} = -d. \quad (13)$$

Furthermore, the refracted ray can be described by the equations of a line

$$\begin{aligned} \frac{x - x_0}{k'_x} &= \frac{y - y_0}{k'_y}, \\ \frac{y - y_0}{k'_y} &= \frac{z - z_0}{k'_z}, \end{aligned} \quad (14)$$

where  $\mathbf{r}_0$  is the point of incidence, and  $k'_{x,y,z}$  are the  $x, y, z$  components of  $\hat{\mathbf{k}}'$  obtained in Eq. (9). Combining Eq. (13) and Eq. (14), the intersection between the lower interface and the refracted ray, i.e., the exit site of the refracted ray,  $\mathbf{r}_e = [x_e, y_e, z_e]^T$ , can be determined by solving the linear system

$$\mathbf{M} \cdot \mathbf{r}_e = \mathbf{b}, \quad (15)$$

where

$$\mathbf{M} = \begin{bmatrix} k'_z & 0 & -k'_x \\ 0 & k'_z & -k'_y \\ \sin \gamma & 0 & \cos \gamma \end{bmatrix},$$

and  $\mathbf{b} = 0, 0, -d^T$ . The optical path length (OPL) of the refracted ray inside the coverslip is

$$\text{OPL}_c = n_2 |\mathbf{r}_e - \mathbf{r}_0| = n_2 |\mathbf{r}_e|. \quad (16)$$




Fig. 1. (a) The actual layout of the excitation and imaging paths. The sample sits on top of the cover-glass bottomed holder, which is carried by a motorized stage. (b) The geometry of the imaging system. The ray of incidence (green line) crosses the upper surface of a coverslip; the normal vector  $\hat{n}$  of the coverslip plane is tilted by angle  $\gamma$  from z-axis in the x-z plane. The ray is refracted at the interface, travels inside the coverslip (red line), exits from the bottom interface and propagates in the original direction. Without the coverslip, the ray is supposed to propagate along the dashed line.

As depicted by Figure 1, below  $z_e$ , i.e., the height of the intersection  $\mathbf{r}_e$ , the refracted ray and the virtual original ray (the dashed line) will travel in parallel over the same distance in the immersion medium before they reach the front lens of the objective. This equal distance does not introduce extra aberration, however, the lateral displacement of the ray is equivalent to a displacement of the point source on the sample plane, and thus increases the optical path by

$$\Delta l = n_1 \sin \theta |\mathbf{r}_e - (z_e - z_0) \hat{\mathbf{k}} / k_z|. \quad (17)$$

Adding the OPL contributions in Eq. (16) and Eq. (17), the total OPL of the refracted ray is

$$\text{OPL}_r = \text{OPL}_c + \Delta l, \quad (18)$$

while from  $z = 0$  to  $z = z_e$ , the aberration-free OPL should be

$$\text{OPL}_i = n_1 |(z_e - z_0) \hat{\mathbf{k}} / k_z|. \quad (19)$$

Therefore, the overall phase shift measured by the wavelength is

$$\Delta \Phi = \frac{\text{OPL}_r - \text{OPL}_i}{\lambda}. \quad (20)$$

## 2.2. Simulation

Based on the derivations in 2.1, we simulate optical aberrations on the pupil plane with different coverslip tilt angles. The NA of the objective is set at 0.8 or 1.0, and  $n_1, n_2$  are set at 1.33, 1.52, respectively. Figure 2 shows the absolute and relative aberration patterns on the pupil plane at  $0^\circ, 15^\circ, 35^\circ$  (maximum practical tilt for an 1.0 NA objective because of geometric constraints), and  $45^\circ$  (maximum practical tilt for an 0.8 NA objective). In each case, the phase shift as a function of  $\theta$  and  $\varphi$  is calculated using Eq. (20); the tilt and piston components are extracted from the x cross section and subtracted from the raw pupil function. In addition to




Fig. 2. (a) Optical aberrations on the pupil plane of a NA=1.0 objective under 0°, 15°, 35°, 45° coverslip tilt. The linear non-aberrational components (piston, tip/tilt) are subtracted from the phase difference calculated with Eq. (20). The white dashed circle represents 0.8 NA. At 45° coverslip tilt, the pupil function is cropped to NA = 0.8 because of objective geometric constraint. The left column shows the absolute aberration in each case, and the right column shows the relative aberration with respect to that in the zero-tilt case. Color bar unit: wavelengths. (b) Cross sections of the absolute (left column) and relative (right column) aberrations. Unit: wavelength.

the coverglass-induced spherical aberrations already present at 0° tilt, calculation results show an increasing amount of phase aberrations with higher tilt angles. At both 35° (for 1.0 NA objective) and 45° (for 0.8 NA objective) coverslip tilt, the steepest change in phase aberrations occurs at the rightmost edge of the pupil plane, where light rays hit the coverglass at the highest incident angles. Other than this edge, the phase aberrations are rather smooth and well suited for correction with a deformable mirror (DM).

Besides phase aberrations discussed above, light intensity is diminished due to reflections at the water-coverslip interface. The power loss is negligible at small angles of incidence, but can be significant at large angles of incidence; the latter would occur for a large coverslip tilt angle. At each interface, the reflectances for *s*-polarized light and *p*-polarized light,  $R_s$  and  $R_p$ , can be determined by the Fresnel equations:

$$R_s = \frac{|n_1 \cos \phi - n_2 \cos \phi'|^2}{|n_1 \cos \phi + n_2 \cos \phi'|^2}, \quad (21)$$

$$R_p = \frac{|n_1 \cos \phi' - n_2 \cos \phi|^2}{|n_1 \cos \phi' + n_2 \cos \phi|^2}, \quad (22)$$

where  $\phi$ ,  $\phi'$  are the angles of incidence and refraction, respectively. Suppose that the incident light is unpolarized, i.e.,  $P_s = P_p = P_0/2$ , then the transmitted power after crossing the two




Fig. 3. Total transmittance of rays within the collection cone at  $\gamma = 0^\circ, 15^\circ, 35^\circ, 45^\circ$  degrees, calculated for NA = 0.8 (white, dashed circle) and 1.0 (full circle). The cross sections along the x-direction are shown in the bottom row.

interfaces is

$$P_t \phi = P_s (1 - R_s)^2 + P_p (1 - R_p)^2 = P_0 \left[ 1 - (R_s + R_p) + \frac{R_s^2 + R_p^2}{2} \right], \quad (23)$$

so the total transmittance through the coverslip is

$$T_c = 1 - (R_s + R_p) + \frac{R_s^2 + R_p^2}{2}. \quad (24)$$

Figure 3 shows  $T_c(\theta, \varphi)$  on the pupil plane of an N.A.= 0.8/1.0 objective for the same four tilt angles in Figure 2. It can be seen that while the transmittance within most of the pupil plane was close to 1 in all cases, at  $35^\circ$  (for 1.0 NA objective) and  $45^\circ$  (for 0.8 NA objective) coverslip tilt, the amplitude loss on the rightmost edge of the pupil plane is significant. This effect caused an effectively small clipping of the pupil plane and hence a small loss of NA. On the other hand, this reduction of transmittance exactly coincide with the part of pupil plane containing the sharpest variation of phase aberrations. Consequently, imperfect phase aberration correction at this rightmost edge of the pupil plane should not have a large effect of the imaging performance.

### 3. Experiment

#### 3.1. Construction of the AO-SPIM

To test the high-NA objective performance when imaging across a tilted coverslip, we layout our SPIM system horizontally on an optical table in order to more easily test different high-NA objectives and tilt angles of the sample (Fig. 4). A long working-distance  $20\times 0.42$  NA air objective (Mitutoyo) in combination with an  $f = 200$  mm cylindrical lens (Thorlabs) are used for excitation; a  $40\times 0.8$  NA infinity-corrected, water-dipping objective (MRD07420, Nikon), is used for imaging. With this objective combination, the 0.8 NA imaging objective can operate at a coverslip tilt angle of  $40^\circ$ . The collimated laser beam forms a light-sheet after passing the excitation optics. By co-focusing the two objectives, the light-sheet can be confined to the focal plane of the high-NA objective.

To accommodate the water-dipping objective, we constructed a small water-containing cube (dimensions  $3.5 \text{ cm} \times 3.5 \text{ cm} \times 3.5 \text{ cm}$ ). The imaging objective dips into the chamber from one side and is sealed with an O-ring. The excitation light enters through a glass window on a perpendicular side. The top of the chamber is open to allow the sample to be inserted from above. The sample is either attached to a coverslip or held between a glass slide and coverslip

and held perpendicular to the optical table. By using a manual rotation stage we can tilt the slide or coverslip at a known angle with respect to the optical axis. A motorized translation stage (MT1-Z8, Thorlabs) allows us to move the sample along the optical axis of the imaging objective. The excitation light is provided by a 473 nm diode-pumped solid state laser (MBL-FN-473-50mW, Changchun New Industries Optoelectronics, China). Images are magnified by a pair of relay lenses and captured by a scientific CMOS camera (Flash 4.0, Hamamatsu).




Fig. 4. (a) Experimental setup. (b) Results of simulating our optical design with Zemax. We simulate a point-source situated behind a coverglass tilted around  $40^\circ$  to the optical axis. The first few optical elements of a  $40\times$  water dipping objective are shown. Light rays are shown emerging from the point source and being refracted at the water-glass and glass-water interfaces. The simulated point spread function is shown. We incorporated a cylindrical lens downstream from the objective in this simulation to counter some of the aberrations.

In our previous low NA open-top SPIM system, the system-induced optical aberration is partially corrected by a 10 m focal length cylindrical lens (SCX-50.8-5000.0, Melles-Griot) placed before the second relay lens. A Zemax simulation of the point-spread function (PSF) at the focus shows that the optical aberration using the 0.8 NA objective, even in the presence of the counter-aberrating cylindrical lens, is severe when the tilt angle is  $40^\circ$ . To correct the rest of the aberration, we place a 140-actuator deformable mirror (Multi-DM, Boston Micromachines) at the conjugate plane of the objective back pupil. Each actuator has a maximum stroke of  $3.5 \text{ \AA}$ , which is equivalent to 7 wavelengths for fluorescent signals around 500 nm. The conjugate pupil diameter is reduced to 4 mm by the telescope formed by the tube lens (TL,  $f = 200 \text{ mm}$ ) and the first relay lens (RL1,  $f = 100 \text{ mm}$ ) to match the active area (4.4 mm diameter) of the deformable mirror.

### 3.2. Aberration measurement and correction

In order to experimentally characterize the aberration in the present setup, we record the PSFs of sub-diffraction limit sized fluorescent beads to numerically retrieve the pupil function [21]. We use 200-nm diameter fluorescent beads (FBs) that are either attached to a glass coverslip or embedded in 1% agarose gel sandwiched between a coverslip and a glass slide. The sample is then stepped in the z-direction as depicted in Fig. 4. Unless otherwise noted, the PSFs are acquired in the epi-illumination mode where the excitation and emitted light both pass through the same objective. Typically, we record 21 images separated by  $\Delta z = 0.5 \text{ \mu m}$  of a single bead to find the pupil function using the iterative phase-retrieval method previously described [21, 22].

An example of a pupil function retrieved for a coverslip tilt angle around  $40^\circ$  is shown in Fig. 5(a). Phase retrieval procedure is carried out over 7 iterations, through which a cumulative pupil pattern is produced. These iterations are needed as the phase-retrieval process does not always, especially in the presence of large aberrations with multiple phase wraps, produce error-free pupil functions. Additionally, we iterate the process to mitigate the effects of uncertainty




Fig. 5. (a) The accumulated pupil function from iterations 1—7. In each iteration, the retrieved pupil function is reversed and discretized into  $12 \times 12$  segments, then scaled up and applied to the 140-segment deformable mirror. The overall correction pattern added on the deformable mirror is shown in (b). (c) With the accumulated correction in (b), a majority of the system aberration is eliminated except at the rightmost edge of the pupil plane. (d) The coefficients of 5–22 zernike components in the retrieved system aberration (a).

in the relationship between the voltage applied to the deformable mirror's actuators and their displacement. The retrieved pupil function is reversed and discretized into  $12 \times 12$  segments as shown in Fig. 5(b), then scaled to a voltage to be applied to the deformable mirror for aberration correction. The associated Zernike coefficients of modes 5–22 are shown in Fig. 5(d). At the end of the 7 iterations, the overall range of voltages applied on the deformable mirror segments takes 69% of the mirror stroke, which is equivalent to 4.7 wavelengths. The retrieved pupil function has little aberration except on the rightmost edge (Fig. 5(c)), and the RMS error over the pupil plane is 0.21 wavelength.

Fig. 6(a)-(c) shows orthogonal slices through the three-dimensional point-spread function when imaging across an approximately  $40^\circ$ -tilted coverglass. The cylindrical lens is in place in the detection path to initially correct as much of the aberrations as possible. Fitting a Gaussian function to a line profile through this PSF yields a full-width at half maximum (FWHM) of  $4.6 \mu\text{m}$  along the long axis of the PSF,  $0.5 \mu\text{m}$  in  $x$ , and  $0.6 \mu\text{m}$  in  $y$  directions (see Fig. 6 (j)). After six iterations of phase-retrieval and correction with the deformable mirror the PSF quality is significantly improved, as shown in Fig. 6 (d)-(f). With this correction, the FWHM of the Gaussian-fitted PSF is  $2.6 \mu\text{m}$  in  $z$  and  $0.4 \mu\text{m}$  in both  $x$  and  $y$  (Fig. 6(k)). We noticed a slight tilt of the PSF in the  $xz$  cross section, possibly resulted from the clipping of the pupil plane at its rightmost edge due to reduced transmission. In addition to the proof-of-principle setup using epi-illumination, we also measured the PSF using light-sheet illumination, i.e., illuminating the sample via the path depicted in Fig. 4 that is orthogonal to the imaging objective. As shown in Fig. 6(g)-(i), due to the reduced illumination depth of the light sheet, the axial extent of the PSF is further reduced. The axial FWHM of the Gaussian-fitted PSF with the light-sheet illumination is  $1.9 \mu\text{m}$  (figure not shown).

Although the air-water interface in the excitation path introduces aberrations into the excitation




Fig. 6. The PSF of our microscope without and with aberration corrections from the deformable mirror, all displayed with log scaling. All scale bars are  $1 \mu\text{m}$ . In (a)-(c), only the cylindrical lens placed in the imaging path corrects some of the aberrations. The deformable mirror is flat. In (d)-(f), a pattern is placed on the deformable mirror found through iteratively finding the pupil function and adjusting the deformable mirror. In (g)-(i), a sheet of excitation light coming from an orthogonally placed objective is used rather than using wide-field illumination as in the previous cases. (a), (d), and (g) display slices of the PSF in the x-y plane. (b), (e), and (h) display slices in the z-y plane. (c), (f), and (i) display slices in the z-x plane. (j), (k): Cross sections of the PSFs in x, y, and z directions before and after the aberration correction.

path and degrades the quality of the light sheet, we have not yet made attempts to correct these aberrations. As the excitation light sheet is formed with a much lower NA objective (0.42 NA in our case) the aberrations will be much less severe. However, additional efforts on the excitation side would be necessary to achieve a diffraction-limited thickness of the light sheet.

We compared PSFs measured for FB samples prepared in the two approaches, i.e., FBs are directly attached to the backside of a coverslip and embedded in agarose gel, respectively. Over a volume extending several tens of microns in all dimensions, we do not observe any significant difference between the PSFs in the two cases. However, when one images farther from the point where the PSF is acquired to retrieve the patterns added on the deformable mirror, we expect the performance of the microscope to degrade. We plan to determine the size of the isoplanatic patch in the future.

### 3.3. Imaging biological samples

To demonstrate the biological applications of our microscope, we imaged human embryonic kidney 293 (HEK) cells with fluorescently labeled nuclear lamina. Like the FB samples, the

cells were placed between a glass slide and a coverslip, with the latter facing the imaging objective. The excitation was implemented in the light-sheet mode, and image stacks were acquired by moving the sample in  $1\ \mu\text{m}$  steps along the optical axis of the imaging objective. Immediately prior to imaging the cells, the pupil function under the specific tilt angle of the sample had been characterized with  $200\ \text{nm}$  FBs, and the pattern was converted to voltages and applied to the deformable mirror accordingly. Fig. 7 shows images of a few cells. It can be seen that with the correction applied, the image exhibits a sharper contrast than the uncorrected image (Fig. 7(c)), even though the correction only compensates system-induced aberration while leaving the sample-induced aberrations unaddressed.




Fig. 7. HEK cells with fluorescently labeled nuclear lamina were imaged with our SPIM using a cylindrical lens and a deformable mirror to reduce aberrations. (a) Upper left: a frame of the acquired stack. Upper right and bottom:  $y$ - $z$  and  $x$ - $z$  view of the slices along the dashed lines. The displayed volume covers  $22\ \mu\text{m}$  along the  $z$ -axis. (b) The zoomed image of a cell. The signal intensities along the dashed line before and after the aberration correction are illustrated in (c).

We note that switching out the FB sample used for aberration characterization for the sample of labeled cells invariably alters, however slightly, the tilt angle of the coverglass, the coverglass thickness and the distance of the sample beyond the coverglass. However, we found that the slight variations in these parameters do not significantly change the necessary pattern on the deformable mirror to obtain high-resolution images. This robustness suggests that the major component of the overall aberration is system-induced and remains static given the alignment is preserved. Thus, for samples which do not vary largely in composition over the whole volume, e.g., single layers of fixed cells, tissues preserved in buffer, the correction pattern calculated from a point source can be directly applied to the deformable mirror. This “once and forever” mode would no longer suffice for samples that change drastically over time and space, e.g., developing fly embryos [23]. Faster computational routines would be required to dynamically shape the correction pattern, and multiple AO devices may be needed in the optical path for complementary optical corrections [2].

#### **4. Conclusion**

We have demonstrated the implementation of a high-resolution open-top SPIM system, which allows one to use specimen prepared in conventional sample holders. Central to the new system is a high-NA objective which collects signals through a tilted coverslip. Although this configuration introduces considerable aberration into the images, we have shown that such aberration can be reliably retrieved by measuring emissions of a point source, and then be largely reversed by adding an opposite aberration pattern on a deformable mirror. Simulations of the imaging system and experimental results both suggest that uncorrected aberrations introduced by the tilted coverslip between the objective and the sample primarily exists at one edge of the exit pupil. This design provides an optically simple and versatile approach to image high-content samples at sub-micron resolution on a SPIM platform.

#### **Acknowledgments**

The HEK cells used for imaging were kindly provided by Sayaka Sekine. This work is supported by NIH (R33EB019784) and the UCSF Program for Breakthrough Biomedical Research.