Clinical Validation of a Non-Invasive Prenatal Test for Genome-Wide Detection of

Fetal Copy Number Variants

Authorship:

Roy B. Lefkowitz[1], John A. Tynan[1], Tong Liu[1], Yijin Wu[1], Amin R. Mazloom[1], Eyad

Almasri[1], Grant Hogg[1], Vach Angkachatchai[1],Chen Zhao[3], Daniel S. Grosu[2], Graham

McLennan[2], Mathias Ehrich[2†]

Acknowledgements:

The authors would like to acknowledge the efforts of those who aided completion of this
study: Nathan Faulkner and Ashley Diaz for sample procurement; Sarah Ahn,
Christine Safanayong, Haiping Lu, Joshua White, Lisa Chamberlain for sample
processing and DNA extraction; David Wong, Jana Schroth, Marc Wycoco, Jesse Fox
for library preparation and sequencing, Pat Liu for bioinformatics pipeline processing.
We would also like to acknowledge the contribution of samples by Ronald Wapner at
Columbia University.

Current Affiliations:

[1]Sequenom Laboratories, 3595 John Hopkins Ct, San Diego, CA, USA

[2]Sequenom Inc., 3595 John Hopkins Ct, San Diego, CA, USA

[3]Illumina, San Diego, CA, USA

[†] To whom correspondence should be addressed:

Dr. Mathias Ehrich, MD

Sequenom Laboratories

3595 John Hopkins Court, San Diego, CA, 92121

Phone: 858-202-9068

Email: mehrich@sequenom.com

Running Head: Clinical Validation of Genome-Wide Cell-Free DNA Testing

Keywords: non-invasive prenatal testing, genome-wide, cell-free DNA, chromosomal CNV, sub-chromosomal CNV, microdeletions

Abbreviations: bootstrap confidence level (BCL); chorionic villus sampling (CVS); chromosomal abnormality decision tree (CADET); cell-free DNA (cfDNA); circular binary segmentation (CBS); copy number variants (CNVs); non-invasive prenatal testing (NIPT); sex chromosomal aneuploidies (SCAs); massively parallel sequencing (MPS), triomsy 21 (T21); trisomy 18 (T18); trisomy 13 (T13)

1. ABSTRACT

Background: Current cell-free DNA (cfDNA) assessment of fetal chromosomes does not analyze and report on all chromosomes. Hence, a significant proportion of fetal chromosomal abnormalities are not detectable by current non-invasive methods. Here we report the clinical validation of a novel NIPT designed to detect genome-wide gains and losses of chromosomal material ≥7 Mb and losses associated with specific deletions <7 Mb.

Objective: The objective of this study is to provide a clinical validation of the sensitivity and specificity of a novel NIPT for detection of genome-wide abnormalities.

Study Design: This retrospective, blinded study included maternal plasma collected from 1222 study subjects with pregnancies at increased risk for fetal chromosomal abnormalities that were assessed for trisomy 21 (T21), trisomy 18 (T18), trisomy 13 (T13), sex chromosome aneuploidies (SCAs), fetal sex, genome-wide copy number variants (CNVs) 7 Mb and larger,

and select deletions smaller than 7 Mb. Performance was assessed by comparing test results

with findings from G-band karyotyping, microarray data, or high coverage sequencing.

Results: Clinical sensitivity within this study was determined to be 100% for T21, T18, T13, and

SCAs, and 97.7% for genome-wide CNVs. Clinical specificity within this study was determined

to be 100% for T21, T18, and T13, and 99.9% for SCAs and CNVs. Fetal sex classification had

an accuracy of 99.6%.

Conclusion: This study has demonstrated that genome-wide non-invasive prenatal testing

(NIPT) for fetal chromosomal abnormalities can provide high resolution, sensitive, and specific

detection of a wide range of sub-chromosomal and whole chromosomal abnormalities that were

previously only detectable by invasive karyotype analysis. In some instances, this NIPT also

provided additional clarification about the origin of genetic material that had not been identified

by invasive karyotype analysis.

2. INTRODUCTION

Since its introduction in 2011, NIPT has had a significant impact on prenatal care. In only four

years, NIPT has evolved into a standard option for high-risk pregnancies [1]. Content has also

evolved from exclusive T21 testing to include T18, T13, SCAs, and select microdeletions. This

'standard' content can be expected to detect 80-83% of chromosomal abnormalities detected by

karyotyping in a general screening population [2-4], however this leaves a gap of approximately

17-20% of alternative chromosomal/sub-chromosomal abnormalities not detected.

Consequently, obtaining comprehensive information about the genetic makeup of the fetus

requires an invasive procedure. To overcome these limitations, NIPT should be extended to

cover the entire genome. However, it is challenging to maintain a very high specificity and

positive predictive value when interrogating all accessible regions in the genome [5]. In previous

reports, we have overcome these technical hurdles [6]. Furthermore, a recent study by Yin et al.

demonstrated feasibility for non-invasive genome-wide detection of sub-chromosomal

abnormalities [7]. In this report, we have improved the assay and the statistical methods to

enable comprehensive genome-wide detection of CNVs ≥7 Mb [8]. We present the results of a

large blinded clinical study of more than 1200 samples including more than 100 samples with

common aneuploidies detectable by traditional NIPT and over 30 samples affected by sub-

chromosomal CNVs.

3. METHODS

STUDY DESIGN

This blinded, retrospective clinical study included samples from women considered at increased

risk for fetal aneuploidy based on advanced maternal age ≥35, a positive serum screen, an

abnormal ultrasound finding, and/or a history of aneuploidy. Archived samples were selected for

inclusion in the study by an unblinded internal third party according to the requirements

documented in the study plan. The samples were then blind-coded to all operators and the

analysts who processed the samples. After sequencing, an automated bioinformatics analysis

was performed to detect whole chromosome aneuploidies and sub-chromosomal CNVs. The

results were compiled electronically and were reviewed by a subject matter expert who

assigned the final classification. This manual review mimics the process in the clinical

laboratory, where cases are reviewed by a laboratory director before a result is signed out. The

complete list of classification results was provided to the internal third party for determination of

concordance. Analyzed samples had confirmation of positive or negative events by either G-

band karyotype or microarray findings from samples collected through either CVS or

amniocentesis. Circulating cell-free "fetal" DNA is believed to originate largely from placental

trophoblasts. Genetic differences between the fetus and the placenta can occur (e.g. confined

placental mosaicism), leading to discordance between NIPT results and cytogenetic studies on

amniocytes or postnatally obtained samples [9]. Results from CVS by chromosomal microarray

were thus considered the most accurate ground truth. Therefore, discordant results originating

from amniocytes (karyotype or microarray) were resolved by sequencing at high coverage (an

average of 226 million reads per sample). Sequencing depth has been shown as the limiting

factor in NIPT methods, with increased depth allowing improved detection of events in samples

with lower fetal fractions or improved detection of smaller events [10]. High coverage

sequencing has been used in multiple studies to unambiguously identify sub-chromosomal

events ([6,11-13]) and was used here as a reference for performance evaluation in discrepant

amniocentesis samples.

Details of the sample demographics are described in Table 1. Indications for invasive testing are

described in Table 2.

SAMPLE COLLECTION

In total, 1222 maternal plasma samples had been previously collected using Investigational

Review Board (IRB) approved protocols (protocol numbers: Compass IRB #00508, WIRB

#20120148, Compass IRB #00351, Columbia University IRB #AAAN9002) with a small subset

(9 samples) comprising remnant plasma samples collected from previously consented patients

in accordance with the FDA Guidance on Informed Consent for *In Vitro* Diagnostic Devices

Using Leftover Human Specimens that are Not Individually Identifiable (25 April 2006). All

subjects provided written informed consent prior to undergoing any study related procedures.

LIBRARY PREPARATION, SEQUENCING, AND ANALYTICAL METHODS

Libraries were prepared and quantified as described by Tynan et al.[14]. To reduce noise and

increase signal, sequencing depth for this analysis was increased to target 32M reads per

sample. Sequencing reads were aligned to hg19 using Bowtie 2 [15]. The genome was then

partitioned into 50-kbp non-overlapping segments and the total number of reads per segment

was determined, by counting the number of reads with 5' ends overlapping with a segment.

Segments with high read-count variability or low map-ability were excluded. The 50 kbp read-

counts were then normalized to remove coverage and GC biases, and other higher-order

artifacts using the methods previously described in Zhao et al [6].

The presence of fetal DNA was quantified using the regional counts of whole genome single-

end sequencing data as described by Kim et al. [16].

GENOME-WIDE DETECTION OF ABNORMALITIES

Circular binary segmentation (CBS) was used to identify CNVs throughout the entire genome by

segmenting each chromosome into contiguous regions of equal copy number [17]. A segment-

merging algorithm was then used to compensate for over segmentation by CBS when the

signal-to-noise ratio (SNR) was low [18]. Z-scores were calculated for both CBS-identified CNVs

and whole chromosome variants by comparing the signal amplitude with a reference set of

samples in the same region. The measured Z-scores form part of an enhanced version of

Chromosomal Aberration DEcision Tree (CADET) previously described in detail in Zhao et al

[6]. CADET incorporates the Z statistics for a CBS-detected CNV to assess the its statistical

significance and a log odds ratio (LOR) to provide a measure of the likelihood of an event being

real, based on an observed fraction of fetal DNA across the genome (see supplemental

methods for [6]).

To further improve specificity of CNV detection, bootstrap analysis was performed as an

additional measure for the confidence of the candidate CNVs. The within sample read count

variability was compared to a normal population (represented by 371 euploid samples) and

quantified by bootstrap confidence level (BCL). In order to assess within sample variability, the

bootstrap resampling described below was applied to every candidate CNV.

For each identified segment within the CNV, the median shift of segment fraction from the

normal level across the chromosome was calculated. This median shift was then corrected to

create a read count baseline for bootstrapping. Next, a bootstrapped segment of the same

segment length as the candidate CNV was randomly sampled with replacement from the

baseline read counts. The median shift was then applied to this bootstrapped fragment. The

segment fraction of the bootstrapped fragment was calculated as follows:

$$\textbf{segment fraction} = \frac{\sum read\ counts\ within\ segment}{\sum read\ counts\ across\ the\ autosome}$$

This process was repeated 1000 times to generate a bootstrap distribution of segment fractions

for an affected population. A normal reference distribution was created based on the segment

fraction of the same location as the candidate CNV in 371 euploid samples. A threshold was

then calculated as the segment fraction that was at least 3.95 median absolute deviations away

from the median segment fraction of the reference distribution. Lastly, the BCL was calculated

as the proportion of bootstrap segments whose fractions had absolute z-statistics above the

significance threshold.

A whole chromosome or sub-chromosomal abnormality is detected as follows:

1. A chromosome is classified to have a trisomy or monosomy if:

161              $|z_{CHR}| \geq C$, $LOR_{CHR} > 0$, BCL ≥ 0.99, and $|z_{CHR}| \geq \alpha\ |z_{CBS}|$;

162   2. A chromosome is classified to have a sub-chromosomal abnormality if:

163              $|z_{CBS}| \geq C$, $LOR_{CBS} > 0$, BCL ≥ 0.99, and $|z_{CHR}| < \alpha\ |z_{CBS}|$

where C is a predefined z-score cutoff that controls the tradeoff between sensitivity and

specificity. For this study C was 3 for chromosome 21 trisomy and monosomy, and 3.95 for all

other CNVs. The comparison $|z_{CHR}| \geq \alpha\ |z_{CBS}|$ is used to distinguish a whole chromosome event

from a sub-chromosomal event. α denotes the type 1 error for misclassification of abnormalities

as an aneuploidy. Simulations (see supplemental methods for [6]) showed that α = 0.8 resulted

in a misclassification of abnormalities at close to 0%; as such, this value was used for this study.

4. RESULTS

The study comprised a total of 1222 maternal plasma samples. After unblinding, eleven

samples were excluded because they had no or insufficient karyotype or microarray information

(see Figure 1) and three samples were excluded because of confirmed mosaicism.

Forty-two (42) of the remaining 1208 samples were flagged as non-reportable using quality

criteria that had been established prior to analysis, leaving 1166 reportable samples for

analyses. Technical failure criteria included but were not limited to: low library concentration, low

raw autosomal counts, high GC bias, poor normalization, and high bin variability (see Figure 1).

Biological failure criteria included low fetal fraction (less than 4%) and large maternal CNV

events. The most common reason for failure was low fetal fraction (n= 11). During review of the

data, one sample was signed out as T18 (and was included in the analyzed cohort), even

though it did not meet the autosomal count minimum. This sample had sufficient counts for the

determination of standard aneuploidies, but not sufficient counts for the detection of sub-

chromosomal CNVs. The 42 non-reportable samples showed no evidence of enrichment for

whole-chromosome/sub-chromosomal positive samples (5 positive of 42 non-reportable

samples vs. 204 positive of 1166 reportable samples). Concordant with previous studies, the

overall reportable rate on first aliquots of maternal plasma was 96.5%. The overall no call rate

per patient is expected to be approximately 1% when a second aliquot is available, based on

published clinical laboratory experience [19].

T21, T18, and T13 Detection

Among the 1166 reportable samples, there were 85 T21 samples, 27 T18 samples, and 15 T13 samples (see Figure 1). All euploid, T21, T18, and T13 samples (determined by invasive diagnostic procedures) were classified correctly by NIPT. One sample was classified by NIPT as T21 but had a normal (46, XX) karyotype by amniocentesis (see Supplemental Table 1). This discrepancy was adjudicated through high coverage 'uniplex' sequencing (typically with >180 million reads) according to our study plan (see Supplemental Figure 1D). The sample showed 16.3% fetal fraction, but the gain of genetic material from chromosome 21 was concordant with 6.5% fetal fraction. This is suggestive of confined placental mosaicism, a relatively common cause of discordant results between amniocyte results and CVS or placentally-derived cell free DNA analysis [9]. Table 3 summarizes the performance for T21, T18, and T13. For T21, the sensitivity was 100% (95% CI: 94.6%-100%) and the specificity was 100% (95% CI: 99.6%-100%). For T18, the sensitivity was 100% (95% CI: 84.4%-100%) and the specificity was 100% (95% CI: 99.6%-100%). For T13, the sensitivity was 100% (95% CI: 74.7%-100%) and the specificity was 100% (95% CI: 99.6%-100%).

SCA Detection

In the 1166 samples reportable for all chromosomal abnormalities, there were 21 samples that were flagged as non-reportable specifically for SCA classification based on thresholds for the chromosome X z-score and chromosome Y z-score as described by Mazloom et al. [20]. There was also one additional sample with an apparent maternal XXX that was flagged as non-reportable for SCA because the maternal event distorted the sex chromosomal representation to a degree that fetal events could not be classified. Among the remaining 1144 samples reportable for SCAs, there were 7 discordant positives that were classified as normal (46, XX) by karyotype and as XO by sequencing at 6-plex (see Supplemental Table 1). In all discordant cases, the karyotype had been obtained from amniocyte samples. This phenomenon is well described and may be attributed to varying levels of placental or maternal mosaicism [21].

215 Uniplex sequencing confirmed 6 of the 7 XO samples. The 7[th] sample had a non-reportable

216 result at uniplex coverage, hence the existing amniocentesis result was used as truth resulting

217 in one false positive assignment. Overall, the sensitivity for SCA was 100% (95% CI: 84.0%-

218 100%) with a specificity of 99.9% (95% CI: 99.4%-100%) (see Table 3).

Genome Wide Detection of CNVs

The test was also designed to detect CNVs equal to or larger than 7 Mb (including whole

chromosome abnormalities other than T13, T18, T21, and SCAs); as well as select

microdeletions smaller than 7 Mb [5]. Among the 1166 samples reportable for sub-chromosomal

abnormalities, there were 43 samples that had positive results for a variety of CNV aberrations:

Wolf-Hirschhorn syndrome deletions, DiGeorge syndrome deletions, Prader-Willi/Angelman

syndrome deletions, Cri du Chat syndrome deletions, and a variety of ≥7 Mb CNVs including

both sub-chromosomal CNVs and whole chromosome trisomies. Several examples of NIPT

detected CNVs confirmed by microarray or karyotype are shown in Figure 2.

Overall, the sensitivity for detection of whole chromosome and sub-chromosomal abnormalities

other than T13, T18, T21, and SCAs was 97.7% (95% CI: 86.2%-99.9%) and the specificity was

99.9% (95% CI: 99.4%-100%) (see Table 3). One case was clearly mosaic for T22 by both

standard coverage and uniplex sequencing (see Supplemental Table 1 and Supplement Figure

1E), but was classified as normal by microarray analysis. Because the invasive diagnosis came

from microarray analysis of cells derived from CVS, our study design considered this as the gold

standard, and this outcome was considered a discordant positive. Another case showed no gain

or loss of genetic material with both standard and uniplex sequencing for a sample that had a

46, XX, der(12)t(12;19)(p13.1;q13.1) karyotype (see Supplemental Figure 2B). This outcome

was considered as a discordant negative given that the invasive procedure was CVS. A set of 7

samples were classified as full chromosomal trisomies, and because the karyotype or

microarray results were derived from amniocytes in these cases, uniplex sequencing was

performed for adjudication per the study design. The trisomy finding was confirmed in each case

by uniplex sequencing, and these findings were considered as concordant positives

(Supplemental Figure 1).

In addition to the high accuracy demonstrated in the results, in some cases NIPT could also

provide clarification about the origin of extra genetic material when the G-banding pattern was

not sufficiently clear. In one case, the amniocyte karyotype finding, 46, XX, der(5)t(5;?)(p15.3;?),

indicated a deletion on chromosome 5 and a duplication of unknown origin (see Figure 3). NIPT

identified an 18.6 Mb duplication representing the entire short arm of chromosome 17, indicating

this fetal tissue possessed a trisomy of chromosome 17p that was not clarified by standard

karyotype. Trisomy 17p is associated with developmental delay, growth retardation, hypotonia,

digital abnormalities, congenital heart defects, and distinctive facial features [22]. Additional

cases are shown in Supplemental Figures 3.

253 5. DISCUSSION

Couples and families who are seeking prenatal information about the genetic health of their

baby have multiple options today – ranging from non-invasive tests that screen for select

chromosomal abnormalities to invasive procedures including karyotype and microarray testing

that can deliver the most comprehensive genomic assessment. Current NIPT methods provide

information about a limited set of conditions, that typically include T21, T18, T13 and SCAs. In

this study we expand testing to the entire genome. The results demonstrate that high sensitivity

and specificity were maintained while adding genome-wide clinically relevant content.

Most NIPT tests suffer from the problem of multiple hypothesis testing; when multiple regions of

the genome are tested independently, false positive rates become additive [23]. This can result

in unacceptably high false positive rates and lead to maternal anxiety and additional

unnecessary clinical testing. Methods that target individual regions of the genome cannot

overcome these limitations because they are part of the design of the test [24]. We have shown

that these limitations can be overcome by using a genome-wide approach, as described here

and previously [6]. In this study we demonstrate that the test can provide excellent performance

for CNVs ≥7Mb and select microdeletions <7Mb, across the genome.

This study has two main limitations. Although we analyze 2- to 6-fold more samples with sub-

chromosomal CNVs than previous studies [6,12,24], the total number of affected samples is still

relatively small compared to studies validating performance for detection common whole-

chromosome aneuploidies [25]. It remains challenging to statistically validate detection of rare

chromosomal variations. A second challenge commonly observed in novel technologies with

increased performance is the compilation of samples with appropriate ground truth. In our study

we face the technical challenge of comparing sequencing versus G-band karyotype or

microarray, as well as the biological aspect of cfDNA versus amniocyte or CVS. We have used

a method that prioritizes CVS over amniocentesis because of the placental origin of cell free

DNA [9]. Factors that can lead to discordant results include confined placental mosaicism,

maternal mosaicism, vanishing twins, undetected tumors, and technical errors.

An interesting observation in this study was the sometimes subjective nature of karyotyping.

When G-banding is used as an analytical method, the optical resolution is occasionally not

sufficient to determine the origin of additional genetic material, making a clinical interpretation

more difficult. These ambiguities are eliminated when using NIPT because the test requires

sequencing alignment to the genome before DNA gains and losses are determined. However,

G-banding will, for the foreseeable future, be the superior methodology to determine copy

number neutral structural changes.

In pregnancies that can benefit from additional information, this test provides more clinically

relevant results than previous NIPT options. However, its role as a follow-up test to abnormal

ultrasound findings or as a general population screen will likely be debated for the foreseeable

future. It has been argued that the incidence of clinically relevant sub-chromosomal CNVs could

be as high as 1.7% in a general population [26], and early microarray data [27] indicate that as

many as 30% of sub-chromosomal CNVs could be >7Mb. In consequence, the incidence for

these large CNVs may be around 0.5% in the general population, substantially higher than the

incidence of T21.

In summary, this clinical study provides validation for an approach that extends the clinical

validity of cfDNA testing, now providing detection of T21, T18, T13, SCAs, fetal sex, and

genome-wide detection of sub-chromosomal and whole chromosomal abnormalities. Overall,

sub-chromosomal abnormalities and aneuploidies other than T13, T18, T21, and SCAs were

detected with a combined clinical sensitivity and specificity of 97.7% and 99.9%, respectively.

This enables comprehensive non-invasive chromosomal assessment that was previously

available only by karyotype, and in some cases, may clarify cryptic findings otherwise

identifiable only by microarray.

6. REFERENCES

1. **ACOG (2015) Cell-free DNA screening for fetal aneuploidy. Committee Opinion No. 640. Obstetrics & Gynecology 126: e31-37.**

2. **Mary Norton RW, Miriam Kuppermann, Laura Jelliffe-Pawlowski, Robert Currier. Cell free DNA analysis vs sequential screening as primary testing considering all fetal chromosomal abnormalities; 2015 02 February 2015 - 07 February 2015; San Diego. pp. S2.**

3. **Norton ME, Jelliffe-Pawlowski LL, Currier RJ (2014) Chromosome abnormalities detected by current prenatal screening and noninvasive prenatal testing. Obstetrics & Gynecology 124: 979-986.**

4. **Wellesley D, Dolk H, Boyd PA, Greenlees R, Haeusler M, et al. (2012) Rare chromosome abnormalities, prevalence and prenatal diagnosis rates from population-based congenital anomaly registers in Europe. European Journal of Human Genetics 20: 521-526.**

5. **Helgeson J, Wardrop J, Boomer T, Almasri E, Paxton W, et al. (2015) Clinical outcome of subchromosomal events detected by whole - genome noninvasive prenatal testing. Prenatal diagnosis 35: 999-1004.**

6. **Zhao C, Tynan J, Ehrich M, Hannum G, McCullough R, et al. (2015) Detection of Fetal Subchromosomal Abnormalities by Sequencing Circulating Cell-Free DNA from Maternal Plasma. Clinical chemistry 61: 608-616.**

7. **Yin A-h, Peng C-f, Zhao X, Caughey BA, Yang J-x, et al. (2015) Noninvasive detection of fetal subchromosomal abnormalities by semiconductor sequencing of maternal plasma DNA. Proceedings of the National Academy of Sciences: 201518151.**

8. **Lefkowitz RB, Liu T, Wu Y, Hogg G, Tim RC, et al. (2015) Noninvasive Whole Genome Karyotype Assay for Subchromosomal Abnormalities In Preparation.**

9. **Grati FR, Malvestiti F, Ferreira JC, Bajaj K, Gaetani E, et al. (2014) Fetoplacental mosaicism: potential implications for false-positive and false-negative noninvasive prenatal screening results. Genetics in Medicine 16: 620-624.**

10. **Fan HC, Quake SR (2010) Sensitivity of noninvasive prenatal detection of fetal aneuploidy from maternal plasma using shotgun sequencing is limited only by counting statistics. PLoS One 5: e10439.**

11. **Jensen TJ, Kim SK, van den Boom D, Deciu C, Ehrich M (2014) Noninvasive Detection of a Balanced Fetal Translocation from Maternal Plasma. Clinical chemistry 60: 1298-1305.**

12. **Srinivasan A, Bianchi DW, Huang H, Sehnert AJ, Rava RP (2013) Noninvasive detection of fetal subchromosome abnormalities via deep sequencing of maternal plasma. The American Journal of Human Genetics 92: 167-176.**

13. **Peters D, Chu T, Yatsenko SA, Hendrix N, Hogge WA, et al. (2011) Noninvasive prenatal diagnosis of a fetal microdeletion syndrome. New England Journal of Medicine 365: 1847-1848.**

14. Tynan J, Kim S, Mazloom A, Zhao C, McLennan G, et al. (2015) Application of risk score analysis to
low ‐ coverage whole genome sequencing data for the noninvasive detection of trisomy 21,
trisomy 18, and trisomy 13. Prenatal Diagnosis.

15. Langmead B, Salzberg SL (2012) Fast gapped-read alignment with Bowtie 2. Nature methods 9:
357-359.

16. Kim SK, Hannum G, Geis J, Tynan J, Hogg G, et al. (2015) Determination of fetal DNA fraction from
the plasma of pregnant women using sequence read counts. Prenatal Diagnosis 35: 810-815.

17. Olshen AB, Venkatraman E, Lucito R, Wigler M (2004) Circular binary segmentation for the analysis
of array ‐ based DNA copy number data. Biostatistics 5: 557-572.

18. Willenbrock H, Fridlyand J (2005) A comparison study: applying segmentation to array CGH data
for downstream analyses. Bioinformatics 21: 4084-4091.

19. McCullough RM, Almasri EA, Guan X, Geis JA, Hicks SC, et al. (2014) Non-invasive prenatal
chromosomal aneuploidy testing-clinical experience: 100,000 clinical samples.

20. Mazloom AR, Džakula Ž, Oeth P, Wang H, Jensen T, et al. (2013) Noninvasive prenatal detection of
sex chromosomal aneuploidies by sequencing circulating cell ‐ free DNA from maternal
plasma. Prenatal diagnosis 33: 591-597.

21. Wang Y, Chen Y, Tian F, Zhang J, Song Z, et al. (2014) Maternal mosaicism is a significant
contributor to discordant sex chromosomal aneuploidies associated with noninvasive prenatal
testing. Clinical chemistry 60: 251-259.

22. Jinno Y, Matsuda I, Kajii T. Trisomy 17p due to A t (5; 17)(p15; p11) pat translocation; 1981. pp.
123-125.

23. Miller RG (1981) Simultaneous statistical inference. New York: Springer-Verlag.

24. Wapner RJ, Babiarz JE, Levy B, Stosic M, Zimmermann B, et al. (2015) Expanding the scope of
noninvasive prenatal testing: detection of fetal microdeletion syndromes. American journal of
obstetrics and gynecology 212: 332.e331-339.

25. Palomaki GE, Kloza EM, Lambert-Messerlian GM, Haddow JE, Neveux LM, et al. (2011) DNA
sequencing of maternal plasma to detect Down syndrome: an international clinical validation
study. Genetics in medicine 13: 913-920.

26. Wapner RJ, Martin CL, Levy B, Ballif BC, Eng CM, et al. (2012) Chromosomal microarray versus
karyotyping for prenatal diagnosis. New England Journal of Medicine 367: 2175-2184.

27. Shaffer LG, Dabell MP, Fisher AJ, Coppinger J, Bandholz AM, et al. (2012) Experience with
microarray ‐ based comparative genomic hybridization for prenatal diagnosis in over 5000
pregnancies. Prenatal diagnosis 32: 976-985.

372
373

TABLES

Demographic	Median	Range
Maternal Age (Yrs) (Ntotal = 1177)	36.0	17.8-47
Gestational Age (Wks) (Ntotal =1183)	17	8-38
Maternal Weight (Lbs) (Ntotal =1168)	150	93-366
Procedure	**Percent**	**(Naffected/Ntotal)**
CVS	14.5	(175/1203)
Amniocentesis	85.2	(1025/1203)
Both	0.2	(3/1203)
Confirmation	**Percent**	**(Naffected/Ntotal)**
Karyotype	90.4	(1089/1205)
Microarray	5.8	(70/1205)
Both	3.8	(46/1205)

Table 1. Demographic and pregnancy-related data. Choice of invasive procedure, choice of diagnostic test, and demographic data were not available for all 1208 samples included in the study. Ntotal refers to the number of samples where that data was available.

Indication for Testing	Euploid (n=1009)	T21 (n=87)	T18 (n=29)	T13 (n=15)	SCA (n=26)	CNV (n=43)	Total (n=1209)
Positive Serum Screening	404(40%)	36(41.4%)	13(44.8%)	4(26.7%)	3(11.5%)	16(37.2%)	476(39.4%)
Maternal Age > 35 Yrs	418(41.4%)	28(32.2%)	10(34.5%)	4(26.7%)	2(7.7%)	10(23.3%)	472(39%)
Ultrasound Abnormality	152(15.1%)	34(39.1%)	6(20.7%)	10(66.7%)	16(61.5%)	21(48.8%)	239(19.8%)
Positive Family History	33(3.3%)	4(4.6%)	1(3.4%)	0(0%)	1(3.8%)	1(2.3%)	40(3.3%)
Not Specified	85(8.4%)	11(12.6%)	5(17.2%)	2(13.3%)	4(15.4%)	2(4.7%)	109(9%)

Table 2. Indication for testing. Each cell lists the count and percentage of different indications
for testing for different types of chromosomal abnormalities. Samples may have multiple
indications for testing. This table includes both reportable (n=1166) and non-reportable samples
(total n=42), for a total of 1208 samples. One sample had no indication for testing, leaving a
total of 1207 samples for the above analysis. As 2 of the trisomy 18 samples also had an SCA,
the total number of abnormalities and euploid samples sum up to 1209.

Abnormality	Concordant Positive	Discordant Positive	Concordant Negative	Discordant Negative	Sensitivity	Specificity
T21	85	0	1081	0	100% (95% CI: 94.6%-100%)	100% (95% CI: 99.6%-100%)
T18	27	0	1139	0	100% (95% CI: 84.4%-100%)	100% (95% CI: 99.6%-100%)
T13	15	0	1151	0	100% (95% CI: 74.7%-100%)	100% (95% CI: 99.6%-100%)
SCA	26	1	1117	0	100% (95% CI: 84.0%-100%)	99.9% (95% CI: 99.4%-100%)
CNVs*	42	1	1122	1	97.7% (95% CI: 86.2%-99.9%)	99.9% (95% CI: 99.4%-100%)

*- Genome copy number variants including 8 samples with detected whole chromosome trisomies, and 35 samples with sub-chromosomal CNVs

Abnormality	Concordant Male	Discordant Male	Concordant Female	Discordant Female	Accuracy
Fetal Sex	583	4	578	1	99.6% (95% CI: 98.9%-99.8%)

Table 3. Clinical performance. Concordance was established using confirmation from a combination of microarray, G-band karyotype, and high coverage sequencing data. There were 42 samples that were non-reportable for all chromosomal abnormalities based on the breakdown in Figure 1. However, this table also takes into account additional reportability criteria for SCA based on chromosome X and Y Z-scores.

FIGURE LEGENDS

Figure 1. Flow chart showing sample exclusions and reasons for non-reportable samples. One of the trisomy 18 samples was also an XXY sample and one of the other trisomy 18 samples was also an X sample. For samples that had discordant results for sequencing versus karyotype or microarray derived from testing of amniocytes, uniplex sequencing was used for confirmation.

Figure 2. Examples of karyotype- and microarray-detected CNVs detected by sequencing. A) Chromosome 22 ideogram showing sequencing-based detection of a 22q11.21 deletion (DiGeorge syndrome) that was confirmed by karyotype and microarray. B) Chromosome 18 ideogram for a sample with a karyotype 46, XY, i(18)(q10). C) Chromosome 7 and 13 ideograms for a sample with a karyotype 46, XX, der(7)t(7;13)(q36;q22).

Figure 3. Clarification of karyotype findings by sequencing of maternal plasma. Normal regions are colored blue while deletions and duplications that are detected with high bootstrap confidence levels (>0.99) are colored red. Low bootstrap confidence events are colored yellow. Chromosome 5 & 17 ideograms are shown for a sample with a karyotype of 46, XX, der(5)t(5;?)(p15.3;?). The unidentified duplication on the karyotype is from chromosome 17.

Figure 1

Figure 2

A.

Microarray:
2.6Mb deletion of 22Q11.21

Sequencing:
22q11.21 deletion

Chr 22

B.

Karyotype:
46, XY, i(18)(q10)

Sequencing:
18p11.32-p11.21 deletion
18p11.21-q23 duplication

Chr 18

C.

Karyotype:
46, XX, der(7)t(7;13)(q36;q22)

Sequencing:
7q36.2-q36.3 deletion
13q31.1-q34 duplication

Chr 7

Chr 13

Figure 3

483