1
2 **A large field of view two-photon mesoscope with subcellular resolution for *in vivo***
3 **imaging**
4

5 Sofroniew, N. J. [1, *], Flickinger, D. [1, *], King, J. [2], Svoboda, K. [1]
6
7 [1] Janelia Research Campus, Ashburn VA 20147, USA
8 [2] Vidrio Technologies, Ashburn VA 20147, USA
9
10 * These authors contributed equally to this work

11
12 **Abstract**
13

Imaging is used to map activity across populations of neurons. Microscopes with
cellular resolution have small (< 1 millimeter) fields of view and cannot
simultaneously image activity distributed across multiple brain areas. Typical large
field of view microscopes do not resolve single cells, especially in the axial
dimension. We developed a 2-photon random access mesoscope (2p-RAM) that
allows high-resolution imaging anywhere within a volume spanning multiple brain
areas ($\varnothing$ 5 mm x 1 mm cylinder). 2p-RAM resolution is near diffraction limited
(lateral, 0.66 μm, axial 4.09 μm at the center; excitation wavelength = 970 nm;
numerical aperture = 0.6) over a large range of excitation wavelengths. A fast three-
dimensional scanning system allows efficient sampling of neural activity in arbitrary
regions of interest across the entire imaging volume. We illustrate the use of the 2p-
RAM by imaging neural activity in multiple, non-contiguous brain areas in
transgenic mice expressing protein calcium sensors.

33   **Introduction**
34
Over the last ten years cellular calcium imaging has become widely used to image
activity in neuronal populations (Grienberger and Konnerth 2012, Peron et al. 2015a).
Calcium imaging samples activity of all labeled neurons in an imaging volume and can
readily be combined with visualization of cell type markers to analyze activity in specific
cell types, the nodes of neural circuits (Kerlin et al. 2010, Fu et al. 2014, Peron et al.
2015a).
41
The development of sensitive protein sensors for neuronal function has been a major
driver for cellular activity imaging *in vivo* (Nagai et al. 2004, Tian et al. 2009, Horikawa
et al. 2010, Zhao et al. 2011, Chen et al. 2013, Dana et al. 2016). More sensitive
indicators allow measurement from larger numbers of neurons at fixed signal-to-noise
ratios (SNR) (Wilt et al. 2013, Peron et al. 2015a). Future improvements in protein
sensors are expected to further increase the numbers of neurons that can be probed
simultaneously.
49
The vast majority of cellular imaging *in vivo* has been done with 2-photon microscopy
(Denk and Svoboda 1997). 2-photon microscopy confines excitation in scattering tissue,
which in turn underlies three-dimensional contrast with subcellular resolution (Denk et al.
1994). 2-photon microscopy demands that individual resolution elements (on the order of
$1\ \mu m^3$) are sampled sequentially, limiting its speed. Spurred by the development of fast
solid state cameras, wide-field methods such as wide-field microscopy (Ziv et al. 2013),
light-field microscopy (Prevedel et al. 2014), and light-sheet microscopy (Holekamp et
al. 2008, Ahrens et al. 2013) have been used to measure neural activity. Scattering and
out-of-focus fluorescence rapidly degrade signal and contrast when imaging in scattering
tissue with wide-field methods compared to 2-photon microscopy (Peron et al. 2015a).
As a consequence, at any point in the image, signals from multiple neurons are mixed
together. Unmixing these signals to extract neural activity of individual neurons is a
challenging computational problem, which remains unsolved in general. In contrast, the
high resolution and image contrast provided by 2-photon microscopy allows extraction of
fluorescence signals corresponding to the activity of individual neurons and subcellular
compartments (Stosiek et al. 2003, Sato et al. 2007, Akerboom et al. 2012, Chen et al.
2013) (Petreanu et al. 2012).
67
Together with calcium imaging, 2-photon microscopy is now routinely used to examine
behavior-related activity in populations of neurons (Harvey et al. 2012, Huber et al. 2012,
Peron et al. 2015b). The majority of studies image only dozens to hundreds of neurons in
one brain region at a time. The advent of sensitive protein indicators for calcium (Nagai
et al. 2004, Tian et al. 2009, Horikawa et al. 2010, Zhao et al. 2011, Chen et al. 2013) and
transgenic animals expressing these indicators (Dana et al. 2014, Madisen and al 2015)
promises much higher throughput. 2-photon microscopy has already been used together
with GCaMP6 (Chen et al. 2013) to measure activity in approximately 1800 pyramidal
neurons at 7 Hz using 50 mW of power at the sample (Peron et al. 2015b). Optimization
of the delivered laser power and 2-photon excitation efficiency will allow tracking of

more than 10,000 pyramidal neurons at similar signal-to-noise ratios and sampling
frequencies. This opens up the possibility of tracking representative neuronal populations
across multiple brain regions.
81
Even relatively simple animal behaviors involve multiple brain regions, which are often
non-contiguous (Hernandez et al. 2010, Guo et al. 2014). Simultaneous tracking of
activity in these brain areas is not possible using standard microscopes, because
mechanical movement of microscopes or specimens is slow. High-resolution microscopes
have fields of view (FOV) that are smaller than most brain areas, whereas large FOV
microscopes have small numerical apertures and cannot resolve individual neurons. To
bridge the gap between single cells and brain regions we developed a mesoscale 2-photon
microscope capable of imaging a FOV of 5 mm, with subcellular resolution. We use the
term 'mesoscale microscopy' to refer to imaging with subcellular resolution and fields of
view spanning multiple brain areas (several millimeters).
92
We implemented fast lateral and axial scanning to both maximize the number of neurons
that can be sampled nearly simultaneously and the flexibility in choosing the imaged
neurons. The fluorescence signal collected per unit time is ultimately limited by the
average power that can be delivered to the specimen, which in turn is limited by heating.
In the mouse brain the maximum allowed power is approximately $P_{max} = 200$ mW
(unpublished observations). The signal rate for one scanning beam is $S_1 = a\,P_{max}^2$, where
a is a parameter that depends on the illumination, the experimental preparation (e.g. the
type and concentration of the fluorescent probe), and fluorescence detection. In another
popular scheme the excitation beam is partitioned into n beams for multiplexed excitation
(Lecoq et al. 2014, Stirman et al. 2015). Dividing the power into n laser foci for
multiplexed imaging, the signal is $S_n = a\,n\,(P_{max}/n)^2$. This division corresponds to
reduced overall signal levels (by $1/n$) and a corresponding reduction in the number of
neurons sampled, at fixed signal-to-noise ratio. Increasing the power of each individual
beam beyond a total (summed) power of $P_{max} = 200$ mW would cause heating and
associated damage. Given the relatively slow dynamics of calcium-related fluorescence
(Peron et al. 2015b), serial scanning with a single laser focus maximizes the number of
neurons that can be probed.
110
Several 2-photon microscopes have been designed for large field of view *in vivo* imaging.
Tsai *et al* designed an elegant scanning system based on off-the-shelf components,
producing a scan field of 10 mm (Tsai et al. 2015). Stirman *et al* used a custom objective
(Stirman et al. 2014) with a field of view of 3.5 mm. Neither system provides subcellular
resolution in the axial dimension (14 μm and 12 μm, respectively), implying that signals
from multiple neurons and neuropil may be conflated. In current implementations these
instruments use standard galvo mirrors, which are too slow for some types of functional
imaging. These microscopes also do not implement rapid axial scanning, which is critical
for targeting specific neurons in different parts of the FOV. In addition, the microscopes
are corrected for aberrations at only one excitation wavelength (800 nm and 910 nm
respectively), limiting their use to a subset of fluorescent probes. A large FOV (5 mm)
mesolens for confocal microscopy, with numerical aperture 0.47, has also been reported
(McConnell et al. 2016).

124
We have developed a 2-photon random access mesoscope (2p-RAM**)**, optimized for
imaging populations of neurons in multiple brain regions during behavior. 2p-RAM
provides diffraction-limited resolution in a cylindrical volume ($\varnothing$ 5 mm x 1 mm). It is
fully corrected for excitation light in the 900-1070 nm range, allowing 2-photon imaging
of most widely used fluorescent proteins. 2p-RAM is based on a fast three-dimensional
scanning system that allows random access to hundreds of thousands of neurons for
targeted interrogation.
132
133 **<u>Results</u>**
134
135 We developed a 2-photon random access mesoscope (2p-RAM**)** with subcellular
136 resolution. We first outline the key specifications, which correspond to specific
137 engineering challenges. We then explain the optical path of the mesoscope, followed by a
138 brief description of the design process. Finally we illustrate the mesoscope performance
139 in calibration experiments and imaging neural activity *in vivo*.
140
141 **Key specifications**
142
First, the diameter of the FOV should be five millimeters. This size allows sampling of
most cortical regions in the mouse brain that are nearly coplanar. For example, it would
be possible to image the primary somatosensory cortex, the primary motor cortex and
parts of secondary motor cortex nearly simultaneously (Figure 1A). Second, the
mesoscope should have better than cellular resolution, implying axial resolutions
substantially smaller than the typical size of a soma (diameter, 10 μm) (Peters and
Kaiserman-Abramof 1970). We specified a numerical aperture (NA) of 0.6, which
translates to approximately 0.61 μm lateral resolution and 4.25 μm axial resolution (for λ
= 970 nm; Figure 1B) (Zipfel et al. 2003). Third, the mesoscope should produce
diffraction-limited performance and high two-photon excitation efficiency over the range
of λ = 900 – 1070 nm. This spectral range corresponds to fluorescent protein sensors
based on GFP and various red fluorescence proteins (Akerboom et al. 2013). Fourth, the
mesoscope should maximize the collection of fluorescence signal, ideally with a large
collection NA. We specified a collection NA of 1.0 (specifications listed Figure 1C).
157
Fifth, a key challenge for mesoscale 2-photon microscopy is efficient laser scanning.
Faster scanning provides better time resolution for measurements of neural activity,
reduced photodamage, and tracking larger numbers of neurons. Rapid scanning methods
rely on resonant scanning mirrors (Fan et al. 1999) or acousto optic deflectors (AODs)
(Reddy and Saggau 2005), both of which have scan angles on the order of a few degrees,
corresponding to relatively small (several hundred μm) scans in the specimen plane. To
access the entire scan field rapidly we designed a lateral scan unit with multiple scanners
in series. A fast resonant scan is moved over the specimen in a flexible manner using a
galvanometer scanner, allowing rapid sampling of activity in widely dispersed brain
regions (Figure 1D).
168
The sixth and final challenge is scanning along the optical axis of the microscope. In

most scanning microscopes, focusing is achieved by moving the objective. Because of the
mechanical inertia of the bulky objective, focusing is more than two orders of magnitude
slower than lateral scanning (Gobel et al. 2007). This limit is unacceptable for mesoscale
imaging because the structures of interest to be sampled in different brain regions are
typically in focal planes that differ by up to several hundred micrometers (Figure 1E). We
addressed this problem by implementing a rapid remote focusing unit in which a light-
weight mirror is moved instead of the objective (Botcherby et al. 2012). Rapid remote
focusing also allowed us to relax flatness of field specifications, because rapid focusing
can partially correct for curved scan fields online.
179
180   **Implementation**
181
182   *Overview*
183
The 2p-RAM was assembled on a vertically mounted breadboard (Figure 2A, 2C, 2E).
The refractive optics of the microscope are predicted to introduce 25,000 fs^2 group-delay
dispersion (GDD) (at λ=1000 nm). GDD will cause the light arriving at the focus to be
spread out in time, resulting in reduced efficiency for two-photon excitation (Denk et al.
1995). To reduce GDD at the sample, the Ti:Sapphire laser beam first passed through a
prism-based GDD compensation unit (Akturk et al. 2006) (Figure 2B). With GDD
compensation enabled, we measured a pulse width of 106 fs at the sample using an
autocorrelator (Carpe; APE) (see Figure 2 Supplement 2), comparable to the pulse width
emitted by our laser (Mai Tai HP; Spectra Physics).
193
The microscope is motorized to allow flexible access to the specimen: it moves in x, y
(> 50 mm travel); it rotates around an axis that passes through the specimen plane and is
parallel to the long axis of the microscope (±20 degrees travel); it moves up and down
along the beam axis for coarse focusing (travel +38 mm -13 mm). Since the microscope
moves with respect to the laser, the beam was threaded into the microscope through a
multi-stage periscope.
200
201   *Remote focus unit*
202
Within the core of the microscope, the beam first enters the remote focus (RF) unit
through a polarizing beam splitting cube and a quarter wave plate. The RF unit consists
of a custom RF objective, and a small mirror (mass, 170 mg; PF03-03-P01, Thorlabs)
mounted on a voice coil (LFA 2010 with a sca814 controller; Equipment Solutions). The
beam passes through the RF objective and is reflected by the mirror back through the RF
objective. The quarter wave plate and polarizing beam splitting cube together then direct
the beam into the lateral scan unit. The coatings of the RF objective were optimized for
high transmission in the IR range. These coatings and the small number of elements (5)
combined to give high transmission through the RF objective (97%).
212
The entrance aperture of the RF objective is conjugated to the entrance aperture of the
imaging objective. The RF mirror is approximately conjugate to the focus in the
specimen plane. Axial movement of the RF mirror changes the axial location of the focus

in the specimen (scale factor, 1.23; i.e. a 100 µm move of the RF mirror causes a 123 µm
move of the focus). Depending on the axial position of the RF mirror, the wavefront
entering the imaging objective will be converging, parallel, or diverging. For standard
microscope objectives, which obey the sine condition, only parallel rays will converge to
a diffraction-limited point in the specimen (Born and Wolf 1980). Spherical wavefronts
in the objective pupil will produce spherical aberration at the focus. Rays entering the
objective back aperture at its periphery will focus to a different point along the axis,
thereby producing a larger focal volume with lower peak intensity. These spherical
aberrations degrade resolution and contrast. The RF objective was designed to counteract
the spherical aberrations produced by the imaging objective, so that diffraction-limited
imaging can be achieved throughout the microscope's imaging volume (Botcherby et al.
2012). The imaging objective also has axial chromatic aberration that varies with RF
depth, which is compensated by the RF objective. The RF unit allows rapid axial
movement of the focal plane, since the RF mirror can be moved much faster than the
objective or the specimen.
231
232 *Scanning system*
233
We designed a lateral scanning unit to efficiently scan arbitrary regions of interest in the
FOV of the microscope. A resonant scanner produces a fast line scan (24 kHz line rate)
(CRS 12K, Cambridge Technology). Because of the limited angular range of the scanner
(10 degrees, optical peak-to-peak) this results in a scan range of 0.6 mm or less in the
specimen plane. This scan line is moved over the specimen using a pair of galvo
scanners.
240
Microscopes using remote focusing have to deal with non-planar wavefronts in their
pupil locations. Maintaining conjugation between all scan mirrors and optic pupils is a
critical issue for these microscopes. Otherwise, the movement of any scan mirror that is
not conjugated to a subsequent optic pupil, such as the entrance pupil of the objective,
will cause a lateral shift of the wavefront at that pupil. Wavefront non-planarity and
lateral shift can cause significant optical aberrations.
247
The beam is directed through a pupil relay (PR1; magnification, 0.44) into the resonant
scanner (open aperture, 4.8 mm), and then through a second identical pupil relay in
reverse (PR2; magnification, 2.3) into a virtually conjugated galvo pair (VCGP). The
VCGP unit conjugates the x and y scan galvo pair without a large-aperture relay made
from refractive optics. Alternatively conjugation could have been accomplished using a
custom pupil relay using refractive optics. However, due to its large size, this relay would
be the most complex and expensive part of the microscope. The VCGP provides a
compact and cost-effective solution.
256
The VCGP uses three galvo mirrors (rather than two) to ensure that the pupils of the x
and y scanning galvos are coincident. The first two mirrors (open aperture, 20 mm;
Cambridge Technology 6M2420X40B025S4 mirror on 6240HM40B galvo) are moved in
the same plane so that the beam rotates about a pupil centered on the third mirror (open
aperture, 14 mm, Cambridge 6M2314X44B025S4 mirror on 6231HM44C galvo), which

rotates in the orthogonal direction. In this manner the pupil remains substantially
stationary over a large scan range. The 20 mm, virtually conjugated galvo mirrors each
have a scan range of 40 deg (optical peak-to-peak, OPTP). However, the virtual
conjugate arrangement of these mirrors results in a total scan range of 32.2 deg OPTP at
the following pupil. The 14 mm galvo mirror has a scan range of 44 deg OPTP. The
20 mm mirrors can perform a step across their entire range with step-and-settle time of
3.4 ms. Smaller steps with these mirrors, and all steps with the 14 mm mirror, are faster.
269
270 *Objective*
271
The exit pupil of the VCGP is imaged by another pupil relay (PR3; magnification, 2.3)
onto the entrance pupil (diameter, 25.6 mm) of the imaging objective (NA, 0.6, focal
length 21 mm) (Figure 2D). The excitation beam is routed into the imaging objective
using a shortpass dichroic (DM1; size 70 x 99 mm; 750 nm shortpass custom coating
design from Alluxa). The front lens has a tip diameter of 15 mm. The distance between
the edge of the front lens and the top of the specimen is 2.7 mm. Because remote
focusing moves the focal plane into the specimen without movement of the objective or
specimen, the effective working distance of the objective is larger than 3 mm. *In vivo*
imaging is usually performed through a microscope coverglass (Trachtenberg et al. 2002,
Holtmaat et al. 2009). The optical system is corrected to operate in a diffraction-limited
manner through a # 4 coverglass (450 μm of BK7 glass; equivalent to a stack of three # 1
coverglasses). Power transmission through the entire system is 13% (Figure 2B).
284
285 *Fluorescence collection path*
286
287 Fluorescence light passes through the primary dichroic (DM1) into the detector arm of
288 the microscope. An optional second dichroic mirror (DM2) allows for an ancillary optical
289 path for one-photon imaging and photostimulation. A third dichroic mirror (DM3)
290 divides the fluorescence path into two signals, each directed into one of two GaAsP
291 photomultiplier tubes (PMTs) (Hamamatsu H11706-40).
292
In a 2-photon microscope, the etendue of fluorescence light collection by the objective
lens must be matched or exceeded by the etendue of the light detector. This condition is
critical for optimal detection of precious signal photons. For most microscopes this
condition is not an issue, as PMTs have photocathodes that are much larger than the
objective FOV, as well as sizable collection angles (Tsai et al. 2002). For the 2p-RAM
the imaging objective collects emitted light with an NA of 1.0, and a collection FOV of 6
mm (Oheim et al. 2001). The diameter of the imaging objective pupil relevant for
fluorescence detection is 42 mm. The etendue of this collected emission light exceeds the
etendue of the current state-of-the-art PMTs which have a photocathode diameter of 5
mm and a usable NA of approximately 0.9 (unpublished measurements). This makes it
impossible for all of the fluorescence light collected by the objective to be directed onto
the PMT using conventional imaging schemes (Mainen et al. 1999, Tsai et al. 2002). To
overcome this problem we used oil to directly couple a custom-made condenser lens to
the glass enveloping the PMT. This allowed the collection NA of the PMT to reach 1.4.
In this manner all of the fluorescence that is collected by the imaging objective is

projected onto the PMT for detection (not counting reflection losses).
309
Our collection path does not image the back pupil of the objective onto the PMT (Denk et
al. 1995, Mainen et al. 1999), as that design would require significantly more complex
collection optics. As a result, the spatial distribution of the signal on the PMT
photocathode varies with location of the signal source within the sample. The spot size of
the projected light from a point source in the sample on the photocathode varies from 1.5
mm to 2.7 mm, depending on the location of the source point. The microscope image is
therefore sensitive to inhomogeneities in the photosensitivity of the PMT photocathode.
For the most critical applications it may be necessary to select PMTs with homogeneous
photocathodes or employ computational methods to correct for the inhomogeneity.
319
320 **Optical design**
321
The excitation path, including the RF objective, pupil relays 1-3, and imaging objective,
was modeled in silico (Zemax). We searched for configurations of optical elements with
diffraction-limited performance and high two-photon excitation efficiency across the
⌀5x1 mm imaging volume, for excitation wavelengths over the range 900-1070 nm.
Performance was assessed using the Strehl ratio (SR), which is a measure of the quality
of optical image formation. Without aberrations an optical system has SR = 1; aberrations
drive SR toward zero. We optimized for SR > 0.8 across the imaging volume, a
commonly used criterion for diffraction-limited performance.
330
331 Two-photon excitation efficiency depends on the propagation time delay difference
332 (PTD) for different rays across the pupil of the system (Estrada-Silva et al. 2009). For
333 PTD not much less than the pulse duration, the efficiency of two-photon excitation would
334 be severely reduced. PTD was kept to less than 50 fs (rms). As far as we know no other
335 microscopy system has been designed with optimization of the PTD as a design criterion.
336
Other design parameters included: The number, size and complexity of the optical
elements; focusing range of the RF unit; nature and cost of the optical glasses. Field
curvature was limited to less than 160 μm over the 5 mm FOV. Finally, tolerancing was
performed to minimize sensitivities and guide manufacturing (Code V). No stock lenses
were used for these modules, as doing so in most cases would not allow sufficient control
over the tolerances to meet performance specifications. The element count for each
module was: RF objective, 5; pupil relay 1/2 (identical), 4; pupil relay 3, 6; imaging
objective, 6. (The modules were manufactured by Jenoptik. Jupiter, Fl; part numbers RF
objective: 14163200, pupil relay 1/2: 14506000, pupil relay 3: 14163100, imaging
objective: 14163000; for a table of critical distances between these elements necessary
for assembly see Figure 2 Supplement 1). A detailed description of the custom optics and
the design process will be published elsewhere.
349
350 **Calibration experiments**
351
352 *Brightness across the field of view*
353

We measured the signal brightness across the FOV. In a homogeneous sample, such as a
solution of fluorophore, image brightness is determined by the two-photon excitation
efficiency and the collection efficiency for fluorescence, both of which depend on
position in the FOV. We acquired an image of a 1 mm thick bath of fluorescein ($\lambda = 970$
nm). At an imaging depth of 500 um, brightness was constant up to approximately
2.5 millimeters from the center, beyond which image brightness collapsed (Figure 3). The
image showed inhomogeneities across the resonant scan (608 μm), apparent as shading in
the vertical stripes in Figure 3A. The shading is produced by wave front error introduced
by pupil relay 2 during rapid resonant scanning. Smaller resonant scans would reduce this
shading. Inhomogeneities in brightness within the specified FOV were on the order of
30 % (Figure 3C), including a small dip in brightness in the center of the field (this dip is
caused by spatial inhomogeneity in the PMT sensitivity across its photocathode; see
below).
367
368   *Resolution*
369
We next measured the resolution of the microscope. Fluorescent microbeads (0.5 μm;
Fluoresbrite Calibration Grade YG microspheres CAT# 18859, Polysciences Inc.) were
embedded in agar (1.5 %; Agarose, Type III-A Sigma A9793) at a dilution of 1:10,000.
Three-dimensional images stacks containing images of beads were collected at various
positions in the imaging volume ($\lambda = 970$ nm). Resolution was estimated based on the
lateral and axial full-width-at-half-max (FWHM) of the bead images (Figure 4). The
lateral FWHM of the bead image was 0.66 ± 0.003 μm (SEM) and the axial FWHM was
4.09 ± 0.06 μm (SEM). Approximately 2 millimeters from the center of the FOV, the
resolution started to degrade gradually (Figure 4I, J, K). At the edge of the FOV (2.5 mm
from the center) the lateral FWHM was 0.89 ± 0.006 μm (SEM), and the axial FWHM
6.88 ± 0.07 μm (SEM). At the center of the FOV, performance was diffraction-limited up
to 1 mm deep into the sample (Figure 4 Supplement 1). This performance was similar at
different positions in the field and different imaging depths up to 1 mm into the sample
(Figure 4 Supplement 2). Resolution was consistent across wavelengths from 900 nm to
1040 nm (the limit of the tuning range of our Ti:Sapphire laser) (Figure 4 Supplement 3).
Resolution also remained essentially unchanged for small (up to $\pm 2.5°$) tilts of the
coverglass with respect to the beam axis (Figure 4 Supplement 4). We conclude that the
2p-RAM is capable of subcellular resolution over the entire imaging volume (⌀5x1 mm
cylinder) and is robust to moderate misalignments.
389
The brightness of the imaged beads dropped off towards the edge of the FOV (up to a
factor of four 2.5 mm from the center of the FOV; Figure 4K). This is consistent with the
observed increase in the size of the PSF (Figure 4J). In 2-photon microscopy, the
brightness of objects that are on the order of, or smaller than the diffraction limit are
highly sensitive to degradation in the PSF. Decreased fluorescence collection efficiency
at the edges of the FOV could also contribute to dimmer images at the edge of the FOV.
Conversely, resolution was consistent across the center of the field, (Figure 4I, J)
suggesting that the dip in brightness (Figure 3) at the center of the field is likely due to a
dip in PMT sensitivity in the center of the FOV (see below).
399

400   *Collection efficiency*
401
Biological 2-photon microscopy is typically limited by photodamage produced by
excitation light. Better detection of fluorescence signal allows reduced excitation for
fixed SNR. Optimizing fluorescence detection is therefore a critical aspect of designing a
2-photon microscope for *in vivo* imaging. In scattering samples the objective numerical
aperture and its FOV contribute to fluorescence detection (Oheim et al. 2001). We
therefore characterized the signal collection efficiency of our system as a function of
position of the signal source and numerical aperture (Figure 5).
409
410   We separately varied the position of a light source and its NA while measuring the signal
411   detected by the PMT (Figure 5A). A pinhole was mounted in front of the light source
412   (LED) and imaged using an imaging system into the specimen plane. The NA of the light
413   cone was controlled using an aperture in an intervening aperture plane. The microscope
414   was moved on a 0.5 millimeter grid around this assembly.
415
416   The PMT signal was remarkably constant across the FOV and beyond (Figure 5B). The
417   signal showed a modest dip in the center of the FOV (approximately 20% reduced
418   sensitivity). This dip was tracked down to inhomogeneous sensitivity on the PMT
419   photocathode (data not shown). This dip also likely accounts for the lower brightness at
420   the center of the field (Figure 3).
421
We next changed the NA of the light source. The total light entering the objective was
roughly proportional to the second power of the NA (Figure 5C). The signal detected by
the PMT was proportional to the signal entering the objective up to NA 1.0, the specified
NA for fluorescence detection of the IO (Figure 5C, 5D). We conclude that our
fluorescence collection system performs nearly optimally across the specified FOV
(⌀ 5 mm). The overall transmission of the collection path, including reflection and
transmission losses, was 71 %.
429
430   *Field curvature and its correction using remote focusing*
431
432   To characterize the field curvature of the microscope we imaged a thin and flat
433   fluorescent sample (Figure 6A). If the scan transects the sample the fluorescent image is a
434   circle (Figure 6B). The radius of the circle increases monotonically as the RF is adjusted
435   downward. In this way we mapped the field curvature of the microscope. The focal
436   position changed by up to 158 μm at 2.5 mm from the center of the FOV.
437
We implemented a field curvature correction by the RF unit in our scan control software.
Following a 300 μm step, the RF mirror settles to better than 2.5 μm in 6.77 ± 0.02 ms
(SEM) (Figure 6, Supplement 1). The correction was done on a line-by-line basis, using
the average position of the resonant mirror during the line as the position to correct. This
scheme corrects the field curvature associated with movements of the slower galvo
mirrors (i.e. along the vertical image stripes). Field curvature along the resonant scan
could not be corrected in this manner because the RF unit does not have the necessary

bandwidth. Not correcting the field curvature along the resonant axis (Figure 6C, bottom)
causes striping artifacts present in the images (Figure 7).
447
448 ***In vivo* imaging**
449
We next imaged neurons in transgenic mice expressing fluorescent proteins. Mice were
prepared with a stainless-steel head plate and a D-shaped cranial window using standard
surgical procedures (Peron et al. 2015b). The D-shaped window was 450 μm thick and
made from three layers of custom cut coverglass (Potomac Photonics). The window
spanned from the lambda suture to 3 mm anterior of Bregma and from the midline 5 mm
laterally, allowing optical access to a large fraction of the cortical surface on the left
hemisphere.
457
Low magnification, high-resolution images ($\lambda = 970$ nm) were acquired in anesthetized
mice expressing GCaMP6f in a subset of pyramidal neurons (GP5.17 mice) or GCaMP6s
in the vast majority of pyramidal neurons (GP4.3) (Dana et al. 2014). Superficial images
showed the vascular pattern on the surface of the brain (GP5.17, Figure 7A, B; GP4.3,
Figure, 7C, Movie 1). Higher magnification views revealed single neurons and their
dendrites (Figure 7B). Resolution and contrast degrade with imaging depth *in vivo*,
caused by aberration and scattering (Ji et al. 2012). However, images of neurons appeared
as ring-like structures up to 600 μm deep in the tissue (Movie 1), reflecting GCaMP
exclusion from the nucleus (Tian et al. 2009) and demonstrating subcellular resolution of
the 2p-RAM.
468
469 Deeper imaging, approximately 450 μm below the surface of the brain, revealed the
470 outlines of the layer 4 barrels of the whisker representation area of the somatosensory
471 cortex, and the lower limb and upper limb somatosensory areas (Figure 7E, 7F, 7
472 Supplement 1A; GP4.3). Layer 4 has decreased fluorescence because stellate cells do not
473 express fluorescent proteins in these mice (Dana et al.).
474
We next imaged GFP-labeled nuclei of excitatory neuron in emx1-Cre x lsl h2b-GFP
mice (Madisen and al 2015). We imaged the full FOV in 2 μm z-steps (up to 600 μm
deep) and 1.2 x 1.3 μm^2 pixels (Figure 7 Supplement 2A). We automatically detected
nuclei in three dimensions using a peak finding algorithm (peak_local_max, scikit-image)
(Figure 7 Supplement 2B). We detected a total of 207,359 nuclei in the volume (Figure 7
Supplement 2C). This is a lower bound on the number of neurons accessible for
functional imaging in the 2p-RAM.
482
483 In mice expressing YFP in a sparse subset of pyramidal neurons (YFP-H) (Feng et al.
484 2000) we were able to resolve individual spines (Figure 7 Supplement 3).
485
486 We next imaged neural activity in behaving mice running on a treadmill in tactile virtual
487 reality (Sofroniew et al. 2014, Sofroniew et al. 2015). We imaged a 4.4 mm x 4.2 mm
488 FOV at 1.9 Hz in layer 2/3 (Figure 8 supplement 1A, B, Movie 2; GP4.3 mice). A simple
489 correlation analysis revealed over 3000 active neurons (Figure 8 supplement 1C, D).
490

The 2p-RAM can link activity at the level of brain regions and single neurons. For
example, low magnification movies showed activity across multiple brain regions
including somatosensory, parietal, and motor areas (Figure 8A; GP5.17 mice). We
correlated the average activity from an area (region 1) in somatosensory cortex with
activity in the rest of brain (Vanni and Murphy 2014)(Figure 8B). We then imaged region
1 and three additional, widely separated regions (each 600x600 μm^2) at higher speed (9.6
Hz) (Figure 8C, Movie 3). Individual neurons showed fluorescence transients
corresponding to trains of action potentials (Chen et al. 2013). We manually segmented
individual neurons in each of the regions and computed the change in fluorescence
($\Delta F/F$) over time (Figure 8E). We measured average activity in region 1 by computing
the maximum of the $\Delta F/F$ traces across all neurons in that region at each time point. We
then correlated this trace against all the $\Delta F/F$ traces of individual neurons (Figure 8D,
8F). Some neurons in the different areas showed high correlation with this average trace,
whereas others showed uncorrelated activity (Okun et al. 2015). In this manner, the 2p-
RAM can be used to relate activity across nearly the entire dorsal cortex with the activity
of individual neurons in specific areas during behavior.
507
508   **<u>Discussion</u>**
509
510   We designed a 2-photon random access mesoscope (2p-RAM) that allows rapid
511   microscopic investigation of neural tissue over large imaging volumes. The 2p-RAM has
512   a numerical aperture of 0.6 and an imaging volume corresponding to a ⌀5x1 mm
513   cylinder. The imaging volume is approximately 100-fold larger compared to other
514   microscopes with comparable resolution.
515
516   The 2p-RAM is optimized for high-resolution *in vivo* imaging. Lateral scanning is
517   performed using a resonant scanner in series with a xy galvo scanner. Rapid scanning is
518   achieved by moving the fast resonant scan over the FOV using the galvo mirrors. Even
519   faster and more flexible scanning could potentially be achieved using acousto-optic
520   deflectors (Duemani Reddy et al. 2008) instead of resonant scanners.
521
Rapid axial scanning is critical for mesoscale functional imaging in the intact brain. This
is because neurons of interest will generally be at different imaging depths in different
brain regions. It is also often of interest to interrogate the same brain region at different
imaging depths. In the 2p-RAM, axial scanning is based on aberration-free focusing
using a remote mirror (Botcherby et al. 2012). The mirror is moved in a plane nearly
conjugate to the specimen plane to move the focus. Aberrations are corrected by passing
the beam through a RF objective that produces complimentary aberrations to the imaging
objective. In published optical arrangements, the beam is scanned laterally first, and is
then sent into the RF system (Botcherby et al. 2012). This design puts constraints on the
RF objective. In our system the beam is passed through the RF system before lateral
scanning. This configuration allowed the etendue, and thus size, complexity, and cost, of
the RF objective to be much lower than that of the imaging objective. We also optimized
the coatings and the element count of the RF objective to minimize power losses that are
normally high in remote focusing systems. The optics subsequent to the RF objective
could be made such that the remote focusing is invariant to lateral scanning, and thus

diffraction-limited focusing can be achieved over a large volume with this arrangement.
Alternative axial scanning techniques, such as rapidly moving the objective (Gobel et al.
2007), are impractical due to the large size of the 2p-RAM objective. Electrotunable
lenses introduce spherical and chromatic aberrations (Grewe et al. 2011).
541
The 2p-RAM has relatively large field curvature (Figure 6). Rapid remote focusing can
help to compensate for field curvature in real time during scanning. Online correction of
field curvature can only be achieved for the movements produced by the slower galvo
scanners. The movements of the resonant scanner are too fast to be fully compensated,
because RF systems implemented with either galvos (Botcherby et al. 2012), voice coils
(Figure 6 Supplement 1), or piezos (unpublished observations) cannot match the resonant
scanning bandwidth. Fast resonant scan lines are by necessity tilted, which is more
pronounced close to the edge of the FOV. For many *in vivo* applications the residual field
curvature is expected to be only a small inconvenience, because neurons of interest in
different brain areas are typically in different focal planes.
552
The serial scan system allows for both low-magnification imaging across a large area and
high-resolution imaging across smaller, but widely separated, areas. Low-magnification
imaging can be combined with online analysis to identify brain areas with particular
activity patterns of interest. Once these areas are found, high-resolution imaging can
interrogate populations of individual neurons within those areas. The low-magnification
image can be used to optimally place these smaller imaging regions around the
vasculature.
560
We performed GDD compensation to maximize peak power delivered for a given
average power. Larger peak power then can be used to maintain the same signal to noise
ratio during faster scanning with smaller pixel dwell times. Optimal placement of
imaging regions and faster scanning combined with smaller jump times between regions
will allow for simultaneous tracking of tens of thousands neurons across the cortex at
10 Hz.
567
568 Microscopes with yet larger fields of view are possible. However, given fixed NA, larger
569 fields of view correspond to larger objectives. Larger and PMTs will be required to
570 efficiently collect the fluorescence transmitted by these giant objectives.
571
There is considerable interest in more natural, unconstrained, and therefore more variable
animal behaviors. Relating the underlying neural population activity to behavior will
require single-trial analysis, which is limited by the number of neurons that can be
recorded in multiple brain regions (Churchland et al. 2007). Mesoscale recording
systems, such as 2p-RAM, allow recordings from thousands of neurons across multiple
brain areas. These large numbers are critical to achieve single-trial decoding of neural
activity patterns during behavior.
579
580
581

Figure Captions

Figure 1. Specifications for the 2p-RAM.

A) Schematic of the dorsal surface of the mouse cortex (from Allen Brain Atlas, Brain

Explorer 2). The overlaid circle corresponds to the specified FOV (∅ 5 mm). MOp:

primary somatomotor, MOs: secondary somatomotor, SSp-ul: primary somatosensory

upper limb, SSp-ll: primary somatosensory lower limb, SSp-un: primary somatosensory

unassigned, SSp-tr: primary somatosensory trunk, SSp-bfd: primary somatosensory barrel

field, PTLp: posterior parietal association, RSP: retrosplenial, VISam: anteromedial

visual, VISpm: posteromedial visual, VISp: primary visual.

B) Schematic of a neuron (approximate soma diameter, 10 μm), and a diffraction limited

point spread function (PSF) corresponding to NA 0.6 and 970 nm. Subcellular resolution

implies that the PSF must be contained within a single neuron.

C) Key specifications for the mesoscope.

D) Imaging parameters used in the experiments. In all cases a fast resonant scan (time per

scan line, 42 μs; 24 kHz line rate) is made over patches 608 μm wide. The time per patch

is 42 μs x number of lines. The entire FOV can be covered in nine stripes. A slow (0.7

Hz), high-resolution scan with 3500 lines per stripe was used to cover a large portion the

FOV (i). A fast (4.3 Hz), low-resolution scan across the entire scan range has up to 500

lines per stripe (iv). More typical scans will involve sampling multiple smaller patches at

high resolution and frame rates (v). See implementation for discussion of the scanning

patterns.

E) Schematic showing a typical use-case. The entire imaging volume is a cylinder with a

5 mm diameter and 1 mm depth. The fast scan is made over patches in different parts of

the volume.

Figure 2. 2p-RAM optics and optomechanics.

A) Optical layout. Several routing mirrors were omitted for clarity. See text for detailed

description.

B) Power transmission measurements at 970 nm.

C) CAD model of the microscope. The microscope was assembled on a vertically

mounted breadboard measuring 1.0 m x 0.5 m. The objective is 63 mm wide, 120 mm

long, and weighs 1.43 kg.

D) Photograph of microscope objective. (D, E, photocredit; Matt Staley).

E) Photograph of vertically mounted breadboard.

Figure 2. Supplement 1. Distances between the critical 2p-RAM components.

A) CAD model of the microscope with critical distances for assembling the 2p-RAM.

PR: pupil relay, RFO: remote focus objective, IO: immersion objective, res.: resonant

scan mirror.

Figure 2. Supplement 2. Pulse width measurements.

We measured the pulse width (full-width-at-half-max) using an autocorrelator. The

autocorrelator was placed after the first telescope and EOM. For a Gaussian laser pulse,

the width of the autocorrelation is 1.41 times the pulse width of the individual laser pulse.

A) Autocorrelation at the autocorrelator without GDD compensation. B) Autocorrelation

at the autocorrelator with GDD compensation. The pulse is now very broad due to the
pre-chirp. C) The autocorrelation at the specimen. As a sample we used a uniform
fluorescent solution (Sigma F6377, 100 μM, 1 mm thick) and recorded fluorescence
using the PMT.
632
Figure 3. Image brightness.
A) Image of a uniform fluorescent sample (fluorescein solution, Sigma F6377, 100 μM, 1
mm thick under 450 μm of coverglass, average of 30 frames). The image was acquired in
9 stripes, each approximately 600 μm wide (see Figure 1D, iv for details). The intensity
variations in the image are caused by changes both in excitation and collection efficiency
across the field.
B) Y profile (along slow scan axis) through slice at the central RF position normalized to
the peak signal, which was located off the main two axes.
C) X profile (along fast scan axis) profile through central slice. The dips in signal at the
edge of the stripes are caused by aberrations at large resonant scan angles.
643
Figure 4. Point spread functions.
Point spread functions were measured at 13 different positions in the field at the middle
depth of the imaging volume. The sample consisted of fluorescent microbeads (diameter,
0.5 μm, embedded in 1 mm thick 1.5% agar, under 450 μm of coverglass) imaged at
970 nm. The excitation intensity was 8.75 mW, below excitation saturation of the beads.
At each position, a 61 μm x 61 μm x 35 μm volume (average of 30 acquisitions) was
acquired with 8.4 x 8.4 x 2 pixels per μm. Bead that were separated by at least 2 μm x 2
μm x 12 μm from the edges of the volume and other beads were analyzed. 2-13 PSFs per
volume were measured.
A) Example measurement at the center of the FOV.
B) Mean intensity projections through the XY, XZ, and YZ slices of the bead.
C) Gaussian fit through the lateral slice through the bead image.
D) Same as C), axial slice.
E-H) Same as A-D) for a bead at the edge of the FOV (2.4 mm from the center on the
top).
I) Average lateral full-width-at-half-max (FWHM) of the central XY slice of a
bead, averaged across the X and Y directions.
J) Average axial FWHM.
K) Average normalized brightness of the beads. Error bars, standard deviation.
663
Figure 4 Supplement 1. Point spread functions across imaging depth
Point spread functions were measured at five different depths at the center of the FOV.
Same sample and imaging conditions as Figure 4.
A) Example measurement at the center of the FOV close to the cover glass at the top of
the sample.
B) Mean intensity projections through the XY, XZ, and YZ slices of the bead.
C) Gaussian fit through the lateral slice through the bead image.
D) Same as C), axial slice.
E-H) Same as A-D) for a bead at the bottom of the imaging volume (RF set to 1 mm
deep).

I) Average lateral full-width-at-half-max (FWHM) of the central XY slice of a bead,
averaged across the X and Y directions.
J) Average axial FWHM.
K) Average normalized brightness of the beads. Error bars, standard deviation.
678
Figure 4 Supplement 2. Point spread functions across the imaging volume
Point spread functions were measured at 15 different locations across the imaging
volume. Three planes (depths; 0 μm, blue; 300 μm, grey; 800 μm, red) were sampled
using the RF system. In each plane, five regions of the FOV were sampled, including the
center and four extrema (2.4 mm from the center) in all quadrants (see legend). Same
sample and imaging conditions as Figure 4.
A) Average lateral full-width-at-half-max (FWHM) of the central XY slice of a bead,
averaged across the X and Y directions.
B) Average axial FWHM.
C) Average normalized brightness of the beads. Error bars, standard deviation.
689
Figure 4 Supplement 3. Point spread functions at different excitation wavelengths.
Point spread functions were measured at the center of the FOV for different excitation
wavelengths. For each wavelength dispersion compensation was adjusted to give
maximum signal from a fluorescein bath and same average power was used (conditions
otherwise as in Figure 4).
A) Average lateral full-width-at-half-max (FWHM).
B) Average axial FWHM.
697
Figure 4 Supplement 4. Point spread functions as a function of coverglass angle
A) Schematic illustrating how PSF size was measured at different coverglass angles.
B) Mean intensity projections through the XZ slices of a bead at different coverglass
angles.
C) Average lateral full-width-at-half-max (FWHM) .
D) Average axial FWHM.
704
Figure 5. Measurement of collection efficiency.
A) Apparatus for measuring signal collection across the FOV and determining the NA of
the collection optics. The light source consisted of a LED (530 nm) and adjacent pinhole,
with a miniature optical relay that projected the pinhole onto the top of a microscope
slide. The projected spot had a 460 μm diameter. An aperture in the relay could be
adjusted to change the NA of the light source from 0 to 1.
B) Normalized signal intensity measured with the PMT (not shown in A) as a function of
position of the light source. The objective was moved in a 0.5 mm grid relative to the
light source. Right, Y slice through the center of the image. Left, X slice through the
center of the image. The peak signal was located off the main two axes.
C) Signal intensity measured by the PMT as a function of the numerical aperture of the
light source (blue) overlaid on the power entering the objective (gray). As the NA of the
light source increases the power entering the objective scales approximately as the square
of the NA, as expected. The power entering the objective was measured with a high NA

accepting power meter (Thor; S170C). Measurement was taken with the 460 μm spot in
the center of the field.
D) Peak normalized ratio of the intensity measured by the PMT to the power entering the
objective as a function of the NA of the light source (axis sampled non-uniformly). The
signal intensity measured through the collection optics scales with the power entering the
objective until it drops off around NA 1.0, significantly larger than the 0.6 excitation NA
of the objective.
726
Figure 6. Characterization and correction of field curvature
A) Field curvature was measured using a thin fluorescein sample (11.2 μm, sandwiched
between a microscope slide and 450 μm of coverglass, imaging conditions as in Figure
3A). At a particular RF position the image of the sample was a fluorescent ring. We
measured the diameter of the ring as a function of RF position.
B) Image of the sample without any correction for field curvature. The sample appears as
a thin ring.
C) Measured field curvature (gray: before field curvature correction, blue: after field
curvature correction).
D) The RF mirror was programmed to compensate for the field curvature. The
compensation is done on a line-by-line basis using the average position of the resonant
mirror during the line as the point to correct. This compensation is able to correct the
field curvature along each stripe well (top), but is unable to correct the field curvature
within a resonant scan as the resonant mirror moves too fast (bottom).
741
Figure 6 Supplement 1. Characterization of remote focus mirror positioner (voice
coil).
A) Motion profile (blue) during of the RF mirror during a 300 μm move of the focal
position. The command signal (grey) contained a 3.3 ms ramp to prevent oscillations of
the actuator. Inset corresponds to a zoom during the settle period.
B) The step and settle times (to < 2.5 μm) as a function of step size.
748
Figure 7. Imaging fluorescent proteins in anesthetized transgenic mice
A) Low magnification image (one section, 120 μm below the dura). The mouse expressed
GCaMP6f under the thy-1 promoter (GP 5.17 line). Scanning parameters as in Fig. 1Di.
Field curvature correction was enabled. Average of 30 scans; power 120 mW.
B) Higher magnification image (dashed box in A). Scanning parameters as in Fig. 1Dii.
C) Low magnification image (one section, 200 μm below the dura). The mouse expressed
GCaMP6s under the thy-1 promoter (GP 4.3 line). Scanning parameters as in Fig. 1Di.
Field curvature correction was enabled. Average of 30 scans; power 120 mW. Data
corresponds to Movie 1.
D) Higher magnification image (dashed box in C). Single neurons with nuclear excluded
GCaMP are clearly visible.
E) Low magnification image at a deeper focal plane (one section, 450 μm below the
dura). Same mouse and scanning parameters as C); power 175 mW.
F) Higher magnification image (dashed box in E) showing major somatosensory areas in
layer 4. These areas appear as regions with reduced fluorescence since GCaMP6s
expression is minimal in layer 4 stellate cells.

765
Figure 7 Supplement 1. Overlay of a typical field of view with the reference atlas.
A) Overlay of the image from Figure 7E with the mouse cortex (from Allen Brain Atlas,
Brain Explorer 2). MOp: primary somatomotor, MOs: secondary somatomotor, SSp-ul:
primary somatosensory upper limb, SSp-ll: primary somatosensory lower limb, SSp-un:
primary somatosensory unassigned, SSp-tr: primary somatosensory trunk, SSp-bfd:
primary somatosensory barrel field, PTLp: posterior parietal association, RSP:
retrosplenial, VISam: anteromedial visual, VISpm: posteromedial visual, VISp: primary
visual.
774
Figure 7 Supplement 2. Counting nuclei
A) Max projection, 100 μm to 200 μm deep, imaged in a mouse expressing GFP in the
nuclei of excitatory neurons (emx1-cre x lsl-h2b-GFP). The image was locally contrast
enhanced with a contrast-limited adaptive histogram equalization (CLAHE, scikit-
image).
B) Blow-up of a single section showing automatically detected nuclei in red. Nuclei were
detected in 3D using a local peak finding algorithm (peak_local_max, scikit-image) after
median filtering.
C) Number of detected nuclei per 2 μm thick slice, smoothed with a length 3 median
filter.
785
786 **Figure 7 Supplement 3. Dendritic images**
787 A) One slice through a mouse expressing YFP in a subset of neurons.
788 B) Zoom in over a dendrite with clearly visible spines.
789
Figure 8. *In vivo* imaging of neural activity
A) Low magnification image from a mouse expressing GCaMP6f under the thy-1
promoter (GP 5.17 line) (parameters described in Figure 1Div; 120 mW average laser
power). Sampling rate, 4.3 Hz.
B) Cross-correlation map of the mean activity in region 1 with the activity in all pixels
after smoothing with a 5x5 pixel boxcar filter. A highly correlated spot of activity is
visible around region 1.
C) Four fields of view (corresponding to boxes in A) that were acquired at high
resolution and frame rate (scan parameters describe in Figure 1Dv; power, 120 mW;
sampling rate, 9.6 Hz). Regions of interest (orange) were manually drawn around
individual neurons.
D) Each neuron from the four fields of view colored according to its correlation value
with the average activity in region 1. The average activity in region 1 was computed by
taking the maximum values across all the $\Delta F / F$ traces in that region.
E) $\Delta F / F$ traces for 16 neurons extracted from the four separate regions (four per
region) colored according to their correlation coefficient shown in D. Sampling rate,
9.6 Hz. Data corresponds to Movie 3.
F) Histograms of the correlation coefficients for the neurons in D in the four separate
regions.
809
Figure 8 Supplement 1. Neural activity in a large field of view

A) Image from a mouse expressing GCaMP6s under the thy-1 promoter (GP 4.3 line)
(parameters described in Figure 1Diii, FOV 4.4 mm x 4.2 mm; 120 mW average laser
power; sampling rate 1.9 Hz). Image corresponds to Movie 3.
B) Local correlation image, showing the result of cross-correlating each pixel with a 4-
pixel neighborhood around it.
C) 3179 regions of interest were manually drawn around active neurons, using the local
correlation image as a reference.
D) $\Delta F / F$ traces for 100 neurons. The neurons were selected based on having high
skewness during that time period.
820
821   **Movie 1. Multi-resolution information in a 2p-RAM image stack.**
822   Same data as in Figure 7C.
823
824   **Movie 2. Functional imaging of a large field of view.**
825   Same data as in Figure 8, Supplement 1. Movie data was smoothed and downsampled by
826   a factor of two in time.
827
828   **Movie 3. Low resolution imaging at low speed, followed by multi-area imaging at**
829   **higher speeds.**
830   Same data as in Figure 8A, C. Low resolution movie data smoothed and downsampled by
831   a factor of three in time. High resolution movie data median filtered in space with filter
832   size two, and smoothed and down sampled by a factor of five in time.
833

Author contributions

Funding

Howard Hughes Medical Institute (HHMI)

NJS, DF, KS

Author contributions

NJS, Conception and design, experiments, analysis and interpretation of data, drafting the

article; DF, Conception and design, optical and mechanical engineering, experiments,

analysis and interpretation of data, revising the article; JK, Conception and design,

software development, revising the article; KS, Conception and design, experiments,

analysis and interpretation of data, drafting the article.

Acknowledgements

We thank Courtney Davis and Simon Peron for help with cranial window surgeries;

David Tracy and Julie Bentley for help with optical design; Jon Arnold for mechanical

engineering support; Jay Kumler, Dan Sykora, and Matt Falanga (Jenoptik) for help with

optical design and testing; Jeff Brooker and Shane Patton (Thorlabs) for custom

mechanical engineering; Aaron Kerlin and Vasily Goncharov for help with the remote

focus system; Jeremy Freeman for help with analysis; Kaspar Podgorski, Arseny

Finkelstein, Na Ji, Simon Peron, and Aaron Kerlin for comments on the manuscript.

Code, data, and CAD models at https://github.com/sofroniewn/2pRAM-paper.

856
857   References
858
Ahrens, M. B., M. B. Orger, D. N. Robson, J. M. Li and P. J. Keller (2013). "Whole-brain
functional imaging at cellular resolution using light-sheet microscopy." Nature
Methods **10**(5): 413-420.
Akerboom, J., N. Carreras Calderon, L. Tian, S. Wabnig, M. Prigge, J. Tolo, A. Gordus,
M. B. Orger, K. E. Severi, J. J. Macklin, R. Patel, S. R. Pulver, T. J. Wardill, E. Fischer, C.
Schuler, T. W. Chen, K. S. Sarkisyan, J. S. Marvin, C. I. Bargmann, D. S. Kim, S. Kugler,
L. Lagnado, P. Hegemann, A. Gottschalk, E. R. Schreiter and L. L. Looger (2013).
"Genetically encoded calcium indicators for multi-color neural activity imaging and
combination with optogenetics." Frontiers in molecular neuroscience **6**: 2.
Akerboom, J., T. W. Chen, T. J. Wardill, L. Tian, J. S. Marvin, S. Mutlu, N. C. Calderon, F.
Esposti, B. G. Borghuis, X. R. Sun, A. Gordus, M. B. Orger, R. Portugues, F. Engert, J. J.
Macklin, A. Filosa, A. Aggarwal, R. A. Kerr, R. Takagi, S. Kracun, E. Shigetomi, B. S.
Khakh, H. Baier, L. Lagnado, S. S. Wang, C. I. Bargmann, B. E. Kimmel, V. Jayaraman,
K. Svoboda, D. S. Kim, E. R. Schreiter and L. L. Looger (2012). "Optimization of a
GCaMP Calcium Indicator for Neural Activity Imaging." The Journal of neuroscience :
the official journal of the Society for Neuroscience **32**(40): 13819-13840.
Akturk, S., X. Gu, M. Kimmel and R. Trebino (2006). "Extremely simple single-prism
ultrashort- pulse compressor." Optics express **14**(21): 10101-10108.
Born, M. and E. Wolf (1980). Principles of Optics. Oxford, Pergamon.
Botcherby, E. J., C. W. Smith, M. M. Kohl, D. Debarre, M. J. Booth, R. Juskaitis, O.
Paulsen and T. Wilson (2012). "Aberration-free three-dimensional multiphoton
imaging of neuronal activity at kHz rates." Proc Natl Acad Sci U S A **109**(8): 2919-
2924.
Chen, T. W., T. J. Wardill, Y. Sun, S. R. Pulver, S. L. Renninger, A. Baohan, E. R.
Schreiter, R. A. Kerr, M. B. Orger, V. Jayaraman, L. L. Looger, K. Svoboda and D. S. Kim
(2013). "Ultrasensitive fluorescent proteins for imaging neuronal activity." Nature
499(7458): 295-300.
Churchland, M. M., B. M. Yu, M. Sahani and K. V. Shenoy (2007). "Techniques for
extracting single-trial activity patterns from large-scale neural recordings." Curr
Opin Neurobiol **17**(5): 609-618.
Dana, H., T.-W. Chen, A. Hu, B. C. Shields, C. Guo, L. Looger, D. Kim and K. Svoboda
(2014). "Thy1-GCaMP6 Transgenic Mice for Neuronal Population Imaging In Vivo."
PLoS One **9**(9): e108697.
Dana, H., B. Mohar, Y. Sun, S. Narayan, A. Gordus, J. P. Hasseman, G. Tsegaye, G. T.
Holt, A. Hu, D. Walpita, R. Patel, J. J. Macklin, C. I. Bargmann, M. B. Ahrens, E. R.
Schreiter, V. Jayaraman, L. L. Looger, K. Svoboda and D. S. Kim (2016). "Sensitive red
protein calcium indicators for imaging neural activity." eLife **5**.
Denk, W., K. R. Delaney, A. Gelperin, D. Kleinfeld, B. W. Strowbridge, D. W. Tank and
R. Yuste (1994). "Anatomical and functional imaging of neurons using 2-photon
laser scanning microscopy." J. Neurosci. Meth. **54**: 151-162.

Denk, W., D. W. Piston and W. W. Webb (1995). Two-photon molecular excitation in
laser-scanning microscopy. <u>Handbook of Biological Confocal Microscopy</u>. J. B.
Pawley. New York, Plenum Press**:** 445-458.
Denk, W. and K. Svoboda (1997). "Photon upmanship: why multiphoton imaging is
more than a gimmick." <u>Neuron</u> **18**: 351-357.
Duemani Reddy, G., K. Kelleher, R. Fink and P. Saggau (2008). "Three-dimensional
random access multiphoton microscopy for functional imaging of neuronal activity."
<u>Nat Neurosci</u> **11**(6): 713-720.
Estrada-Silva, F. C., J. Garduno-Mejia, M. Rosete-Aguilar, C. J. Roman-Moreno and R.
Ortega-Martinez (2009). "Aberration effects on femtosecond pulses generated by
nonideal achromatic doublets." <u>Appl Opt</u> **48**(24): 4723-4734.
Fan, G. Y., H. Fujisaki, A. Miyawaki, R. K. Tsay, R. Y. Tsien and M. H. Ellisman (1999).
"Video-rate scanning two-photon excitation fluorescence microscopy and ratio
imaging with cameleons." <u>Biophysical journal</u> **76**(5): 2412-2420.
Feng, G., R. H. Mellor, M. Bernstein, C. Keller-Peck, Q. T. Nguyen, M. Wallace, J. M.
Nerbonne, J. W. Lichtman and J. R. Sanes (2000). "Imaging neuronal subsets in
transgenic mice expressing multiple spectral variants of GFP." <u>Neuron</u> **28**(1): 41-51.
Fu, Y., J. M. Tucciarone, J. S. Espinosa, N. Sheng, D. P. Darcy, R. A. Nicoll, Z. J. Huang
and M. P. Stryker (2014). "A cortical circuit for gain control by behavioral state." <u>Cell</u>
156(6): 1139-1152.
Gobel, W., B. M. Kampa and F. Helmchen (2007). "Imaging cellular network
dynamics in three dimensions using fast 3D laser scanning." <u>Nat Methods</u> **4**(1): 73-
79.
Grewe, B. F., F. F. Voigt, M. van 't Hoff and F. Helmchen (2011). "Fast two-layer two-
photon imaging of neuronal cell populations using an electrically tunable lens."
<u>Biomed Opt Express</u> **2**(7): 2035-2046.
Grienberger, C. and A. Konnerth (2012). "Imaging calcium in neurons." <u>Neuron</u>
73(5): 862-885.
Guo, Z. V., N. Li, D. Huber, E. Ophir, D. A. Gutnisky , J. T. Ting, G. Feng and K. Svoboda
(2014). "Flow of cortical activity underlying a tactile decision in mice." <u>Neuron</u>
81(1): 179-194.
Harvey, C. D., P. Coen and D. W. Tank (2012). "Choice-specific sequences in parietal
cortex during a virtual-navigation decision task." <u>Nature</u> **484**(7392): 62-68.
Hernandez, A., V. Nacher, R. Luna, A. Zainos, L. Lemus, M. Alvarez, Y. Vazquez, L.
Camarillo and R. Romo (2010). "Decoding a perceptual decision process across
cortex." <u>Neuron</u> **66**(2): 300-314.
Holekamp, T. F., D. Turaga and T. E. Holy (2008). "Fast three-dimensional
fluorescence imaging of activity in neural populations by objective-coupled planar
illumination microscopy." <u>Neuron</u> **57**(5): 661-672.
Holtmaat, A., T. Bonhoeffer, D. K. Chow, J. Chuckowree, V. De Paola, S. B. Hofer, M.
Hubener, T. Keck, G. Knott, W. C. Lee, R. Mostany, T. D. Mrsic-Flogel, E. Nedivi, C.
Portera-Cailliau, K. Svoboda, J. T. Trachtenberg and L. Wilbrecht (2009). "Long-term,
high-resolution imaging in the mouse neocortex through a chronic cranial window."
<u>Nat Protoc</u> **4**(8): 1128-1144.
Horikawa, K., Y. Yamada, T. Matsuda, K. Kobayashi, M. Hashimoto, T. Matsu-ura, A.
Miyawaki, T. Michikawa, K. Mikoshiba and T. Nagai (2010). "Spontaneous network

activity visualized by ultrasensitive Ca(2+) indicators, yellow Cameleon-Nano."
Nature Methods **7**(9): 729-732.
Huber, D., D. A. Gutnisky, S. Peron, D. H. O'Connor, J. S. Wiegert, L. Tian, T. G. Oertner,
L. L. Looger and K. Svoboda (2012). "Multiple dynamic representations in the motor
cortex during sensorimotor learning." Nature **484**(7395): 473-478.
Ji, N., T. R. Sato and E. Betzig (2012). "Characterization and adaptive optical
correction of aberrations during in vivo imaging in the mouse cortex." Proceedings
of the National Academy of Sciences of the United States of America **109**(1): 22-27.
Kerlin, A. M., M. L. Andermann, V. K. Berezovskii and R. C. Reid (2010). "Broadly
tuned response properties of diverse inhibitory neuron subtypes in mouse visual
cortex." Neuron **67**(5): 858-871.
Lecoq, J., J. Savall, D. Vucinic, B. F. Grewe, H. Kim, J. Z. Li, L. J. Kitch and M. J. Schnitzer
(2014). "Visualizing mammalian brain area interactions by dual-axis two-photon
calcium imaging." Nature Neuroscience **17**(12): 1825-1829.
Madisen, L. and e. al (2015). "Transgenic mice for intersectional targeting of neural
sensors and effectors with high specificity
and performance." Neuron **in press**.
Mainen, Z. F., M. Maletic-Savatic, S. H. Shi, Y. Hayashi, R. Malinow and K. Svoboda
(1999). "Two-photon imaging in living brain slices." Methods **18**: 231-239.
McConnell, G., J. Tragardh, R. Amor, J. Dempster, E. Reid and W. B. Amos (2016). "The
mesolens project at the University of Strathclyde." F1000Res.
Nagai, T., S. Yamada, T. Tominaga, M. Ichikawa and A. Miyawaki (2004). "Expanded
dynamic range of fluorescent indicators for Ca2+ by circularly permuted yellow
fluorescent proteins." Proc Natl Acad Sci U S A **101**(29): 10554-10559.
Oheim, M., E. Beaurepaire, E. Chaigneau, J. Mertz and S. Charpak (2001). "Two-
photon microscopy in brain tissJue: parameters influencing the imaging depth." J
Neurosci Methods **111**(1): 29-37.
Okun, M., N. A. Steinmetz, L. Cossell, M. F. Iacaruso, H. Ko, P. Bartho, T. Moore, S. B.
Hofer, T. D. Mrsic-Flogel, M. Carandini and K. D. Harris (2015). "Diverse coupling of
neurons to populations in sensory cortex." Nature **521**(7553): 511-515.
Peron, S., T. W. Chen and K. Svoboda (2015a). "Comprehensive imaging of cortical
networks." Curr Opin Neurobiol **32**: 115-123.
Peron, S. P., J. Freeman, V. Iyer, C. Guo and K. Svoboda (2015b). "A Cellular
Resolution Map of Barrel Cortex Activity during Tactile Behavior." Neuron **86**(3):
783-799.
Peters, A. and I. R. Kaiserman-Abramof (1970). "The small pyramidal neuron of the
rat cerebral cortex. The perikaryon, dendrites and spines." Am J Anat **127**(4): 321-
355.
Petreanu, L., D. A. Gutnisky, D. Huber, N. L. Xu, D. H. O'Connor, L. Tian, L. Looger and
K. Svoboda (2012). "Activity in motor-sensory projections reveals distributed
coding in somatosensation." Nature **489**(7415): 299-303.
Prevedel, R., Y. G. Yoon, M. Hoffmann, N. Pak, G. Wetzstein, S. Kato, T. Schrodel, R.
Raskar, M. Zimmer, E. S. Boyden and A. Vaziri (2014). "Simultaneous whole-animal
3D imaging of neuronal activity using light-field microscopy." Nature Methods
11(7): 727-730.

Reddy, G. D. and P. Saggau (2005). "Fast three-dimensional laser scanning scheme
using acousto-optic deflectors." Journal of biomedical optics **10**(6): 64038.
Sato, T. R., N. W. Gray, Z. F. Mainen and K. Svoboda (2007). "The Functional
Microarchitecture of the Mouse Barrel Cortex." PLoS Biol **5**(7): e189.
Sofroniew, N. J., J. D. Cohen, A. K. Lee and K. Svoboda (2014). "Natural whisker-
guided behavior by head-fixed mice in tactile virtual reality." The Journal of
neuroscience : the official journal of the Society for Neuroscience **34**(29): 9537-
9550.
Sofroniew, N. J., Y. A. Vlasov, S. Andrew Hires, J. Freeman and K. Svoboda (2015).
"Neural coding in barrel cortex during whisker-guided locomotion." eLife **4**.
Stirman, J. N., I. T. Smith , M. W. Kudenov and S. L. Smith (2014). "Wide field-of-view,
twin-region two-photon imaging across extended cortical networks." bioRxiv.
Stirman, J. N., I. T. Smith , M. W. Kudenov and S. L. Smith (2015). "Wide field-of-view,
twin-region two-photon imaging across extended cortical networks." bioRxiv.
Stosiek, C., O. Garaschuk, K. Holthoff and A. Konnerth (2003). "In vivo two-photon
calcium imaging of neuronal networks." Proc Natl Acad Sci U S A **100**(12): 7319-
7324.
Tian, L., S. A. Hires, T. Mao, D. Huber, M. E. Chiappe, S. H. Chalasani, L. Petreanu, J.
Akerboom, S. A. McKinney, E. R. Schreiter, C. I. Bargmann, V. Jayaraman, K. Svoboda
and L. L. Looger (2009). "Imaging neural activity in worms, flies and mice with
improved GCaMP calcium indicators." Nat Methods **6**(12): 875-881.
Trachtenberg, J. T., B. E. Chen, G. W. Knott, G. Feng, J. R. Sanes, E. Welker and K.
Svoboda (2002). "Long-term in vivo imaging of experience-dependent synaptic
plasticity in adult cortex." Nature **420**(6917): 788-794.
Tsai, P. S., C. Mateo, J. J. Field, C. B. Schaffer, M. E. Anderson and D. Kleinfeld (2015).
"Ultra-large field-of-view two-photon microscopy." Optics express **23**(11): 13833-
13847.
Tsai, P. S., N. Nishimura, E. J. Yoder, A. White, E. Dolnick and D. Kleinfeld (2002).
Principles, design and construction of a two photon scanning microscope for in vitro
and in vivo studies. Methods for In Vivo Optical Imaging. R. Frostig. Boca Raton, CRC
Press**:** 113-171.
Vanni, M. P. and T. H. Murphy (2014). "Mesoscale transcranial spontaneous activity
mapping in GCaMP3 transgenic mice reveals extensive reciprocal connections
between areas of somatomotor cortex." J Neurosci **34**(48): 15931-15946.
Wilt, B. A., J. E. Fitzgerald and M. J. Schnitzer (2013). "Photon shot noise limits on
optical detection of neuronal spikes and estimation of spike timing." Biophysical
journal **104**(1): 51-62.
Zhao, Y., S. Araki, J. Wu, T. Teramoto, Y. F. Chang, M. Nakano, A. S. Abdelfattah, M.
Fujiwara, T. Ishihara, T. Nagai and R. E. Campbell (2011). "An expanded palette of
genetically encoded Ca(2) indicators." Science **333**(6051): 1888-1891.
Zipfel, W. R., R. M. Williams and W. W. Webb (2003). "Nonlinear magic: multiphoton
microscopy in the biosciences." Nature biotechnology **21**(11): 1369-1377.
Ziv, Y., L. D. Burns, E. D. Cocker, E. O. Hamel, K. K. Ghosh, L. J. Kitch, A. El Gamal and
M. J. Schnitzer (2013). "Long-term dynamics of CA1 hippocampal place codes." Nat
Neurosci **16**(3): 264-266.
1035

A)

1 mm

B) Diffraction limited PSF
at NA 0.6, 970nm :
4.25 µm axial FWHM
0.61 µm lateral FWHM

PSF

5 µm

C) Key specifications:
Field of view: 5.0 mm
Depth range: 1 mm
Excitation NA: 0.6
Collection NA: 1.0
Excitation wavelenth range: 900 nm - 1070 nm

D)

Example scan patterns

	i)	ii)	iii)	iv)	v)
	1 mm	100 µm	1 mm	1 mm	300 µm
# stripes	9	1	7	9	4
# lines / stripe	3500	1024	1750	500	512
Area (mm²)	4.4 x 5.4	0.6 x 0.6	4.4 x 4.2	6.3 x 5.4	4 x .6 x .6
Pixel size (µm²)	1.2 x 1.3	0.6 x 0.6	2.4 x 2.6	12 x 13	1.2 x 1.2
Frame rate (Hz)	0.7	21	1.9	4.3	9.5
Figure panel	3A, C, E	3B	3-1	4A, 5A 8B, 8D	4C

E) Random access scanning

1 mm

5.0 mm

Brain area of interest

Region of interest scan
512 x 512 pixels
600 x 600 µm²
22 ms

Jump ~ 6 ms

A)

EOM: electro-optic modulator
GDD: group delay dispersion
QWP: quarter waveplate
PBS: polarizing beam splitter
RFO: remote focus objective
VC: voice coil
Res.: 12kHz resonant scanner
PR: pupil relay
DM: dichroic mirror
PMT: photomultiplier tube
IO: imaging objective

B)

Transmission

*1: 91% after EOM

*2: 55% after prism based GDD comp

*3: 35% before remote focus objective

*4: 17% in between galvo pair

*5: 13% out of objective

C)

D)

5 cm

E)

20 cm

A)

1: RFO mounting flange to RF mirror: 32.50 mm
2: PR1 mounting flange to RFO mounting flange: 341.43 mm
3: PR1 mounting flange to res. mirror: 57.95 mm
4: PR2 mounting flange to res. mirror: 46.01 mm
5: PR2 mounting flange to single xy point: 399.68 mm
6: PR3 inside mounting flange to single xy mirror point: 423.88 mm
7: PR3 inside mounting flange to top surface of IO flange: 224.05 mm

A) After EOM
FWHM 202 fs   Pulse width 142 fs

B) After compensation
FWHM 1426 fs   Pulse width 1008 fs

C) At sample
FWHM 150 fs   Pulse width 106 fs

Figure 2
Supplement 2

A)

RF 500 µm

1 mm

B)

Position (mm)

0.0 0.2 0.4 0.6 0.8 1.0
Normalized intensity

C)

1.0
0.8
0.6
0.4
0.2
0.0

Normalized intensity

-3   -2   -1   0   1   2   3
Position (mm)

Figure 4

Figure 4
Supplement 1

A) Coverglass angle

Bead sample

B) 1 um

-5.5° -2.5° 0° 2.5° 5.5°

C) Lateral FWHM (μm)

D) Axial FWHM (μm)

Coverglass angle (°)

Figure 4
Supplement 4

A)
Primary dichroic mirror
Imaging objective
Slide
Aperture
LED

B)
1 mm

Normalized intensity

Position (mm)
-4 -3 -2 -1 0 1 2 3 4

Position (mm)
Normalized intensity

C)
Normalized power
Entering objective
Collected
Numerical aperture

D)
Relative collection
Numerical aperture

A) Thin fluorescein sample / Field curvature

B) Uncorrected / 1 mm

C) uncorrected / corrected / vertical slice / horizontal slice / Depth (µm) / Position (mm)

D) Corrected / 1 mm

Figure 7

A) 120 um deep - GP5.17

Medial
Anterior

1 mm

B)

100 µm

C) 200 um deep - GP4.3

1 mm

D)

100 µm

E) 450 um deep - GP4.3

1 mm

F)

Hindlimb S1

Barrel cortex

C1

500 µm

A)

1 mm

A) emx1-cre x lsl-h2b-GFP
100 - 200 μm max projection

B)

C)

A)

B)

A)

B)

C)

correlation 0.7 / 0.2

1 mm

1 mm

1 mm

D)

5 ΔF/F

Time (s)