Light Sheet Theta Microscopy for High-resolution Quantitative Imaging of Large Biological Systems

Bianca Migliori[1,4,#], Malika S. Datta[1,#], Mehmet C. Apak[1] & Raju Tomer[1,2,3,*]

[1]Department of Biological Sciences

[2]Neurotechnology Center

[3]Data Science Institute

Columbia University, New York, NY 10027, USA.

[4]Department of Neuroscience, Karolinska Institut, Stockholm, Sweden

[#]Equal contributions

*Correspondence: raju.tomer@columbia.edu

Abstract

Recent advances in tissue clearing methods are enabling unprecedented optical access to large intact biological systems. These advances fuel the need for high-speed microscopy approaches to image large transparent samples quantitatively and at high resolution. While Light Sheet Microscopy (LSM), with its high planar imaging speed and low photo-bleaching, can be effective, scaling up to larger imaging volumes has been hindered by the use of orthogonal light-sheet illumination. To address this fundamental limitation, we have developed an approach - Light Sheet Theta Microscopy (LSTM) – that uniformly illuminates samples from the same side as the detection objective, eliminating limits on lateral dimensions without sacrificing imaging depth and speed. We present detailed characterization of LSTM, and show that this approach achieves rapid high-resolution imaging of large intact samples with superior uniform image quality than LSM. LSTM is a significant step in high-resolution quantitative mapping of structure and function of large intact biological systems.

Introduction

The emergence of various tissue clearing and molecular labelling methods over the last decade is enabling unprecedented optical access to the structure and function of intact biological systems [1-14]. Most of these methods employ a cocktail of chemicals for membrane lipid dissolution and/or refractive index smoothening to render the tissue transparent [15]. Together with parallel advances in high-speed microscopy methods, these approaches have already proven to be highly effective in mapping of organs as large as the intact adult mouse brain [1,4,16,17]. By providing a highly-detailed 3D view of the architecture of normal and abnormal intact tissues, these methods can accelerate our understanding of the structure and function of brains, a key goal of high profile BRAIN initiative, as well as provide mechanistic insights into the pathophysiology of the microarchitecture of diseased tissues. However, scaling up these approaches, while maintaining uniform high imaging quality, faces the challenges of clearing and labelling of large samples combined with high-resolution quantitative 3D imaging. Here we address some of these challenges by developing a conceptually distinct microscopy framework: Light Sheet Theta Microscopy (LSTM). Building upon the principles of light sheet microscopy (LSM)[18,19], LSTM allows high-speed quantitative imaging of large intact tissues at high uniform resolution.

Light sheet microscopy was first introduced more than a century ago [18,19]; it uses a thin sheet of light for planar illumination of a sample and an orthogonally arranged wide-field detection arm for simultaneously capturing the emitted signal with a high speed CCD or sCMOS camera[16]. Compared to other commonly used 3D imaging modalities, confocal and 2-photon microscopy, LSM places the minimum possible energy load on the sample and provides orders of magnitude faster imaging. The samples larger than the field-of-view (FOV) of a microscope are imaged by sequential acquisition of overlapping image stacks[1], which are then computationally stitched to

result in the final image volumes. LSM has been highly successful for experimentations in developmental biology [20-24], cell biology [25-27], high-resolution whole brain neuroanatomy [1,17,28] and neural activity mapping experiments [28-33].

The sizes of samples that can be imaged with LSM is restricted along two dimensions, the detection and the illumination axes (**Figure 1**). Sample illumination is therefore restricted to few millimeters deep and wide, without a limit on length. Although LSM has been used for rapid high-resolution imaging of samples as large as mouse brains [1], image quality is progressively reduced towards the sample's center because of illumination light scattering (even with two-sided illumination). The progressive reduction of image quality is even more severe for larger rat brain tissues[34]. These fundamental limitations have precluded the use of LSM for high-resolution quantitative imaging of large samples such as rodent brain tissues, and for imaging of a sample with laterally extended geometries such as thick slices of human brain or physically expanded tissues (e.g. using Expansion Microscopy[35] and MAP[36]). Recently, alternative optical configurations of LSM - including the rotation of the illumination and the detection axes by 45° relative to the sample surface normal as done in OCPI, iSPIM and diSPIM implementations [23,29,37,38] and illumination light sheets through the detection objectives itself[31] have partially addressed these limitations. While these implementations have been highly successful for rapid imaging of small samples (e.g. *C. elegans* embryos) or small volumes of mouse brain cortex, they suffer from other geometry and image quality constraints (**Figure 1a**) when applied to larger samples.

We have developed Light Sheet Theta Microscopy (LSTM) to address some of these limitations. LSTM achieves planar imaging by employing obliquely arranged illumination light sheets from the same side of the sample as the detection objective. This configuration alleviates limitations on the lateral dimensions of the sample, while providing similar or better imaging depth, uniform

high-resolution, low photo-bleaching and high imaging speed for larger samples. Here we characterize the LSTM approach using examples that include mouse and rat brain, as well as human brain slices. Through high-speed quantitative imaging of larger samples, LSTM could facilitate mapping of an entire post-mortem human brain (slab-by-slab) in a practical time-frame.

Results

Light Sheet Theta Microscopy (LSTM)

LSTM includes a standard wide-field detection arm and two symmetrically arranged non-orthogonal ($\theta<90°$, relative to the sample surface normal) illumination arms for the generation of thin sheets of light that intersect at the detection focal plane (**Figure 1,2**). This approach results in a thin line illumination profile which is then scanned, in synchrony with the line-by-line rolling shutter detection of an sCMOS camera (virtual slit effect, Tomer et al 2014), to achieve thin optical sectioning (**Figure 1**). In contrast to LSM, the non-orthogonal optical configuration of LSTM does not place any restrictions on the lateral dimensions of the imaging volume, while still allowing access to the complete working distance of the detection objective, provides high imaging speeds (20 milliseconds per image acquisition, limited only by sCMOS camera speed[1]) and low photo-bleaching (**Figure 3d**). To achieve planar illumination, we designed two modes of line scanning: (1) 1-axis scanning (1-AS), which involves translation of light sheets perpendicular to their propagation direction (**Figure 1c-left**). (2) simultaneous 2-axes scanning (2-AS): concurrent translation of light sheet perpendicular to and along the propagation direction so that the thinnest part of the light sheet intersects with the detection plane (**Figure 1c-right**). The LSTM 1-AS approach provides a simpler implementation, although at the cost of non-uniformity in the planar

illumination and low axial resolution (because of the need to use low numerical aperture illumination). The 2-AS approach allows for uniform planar illumination and detection to enable high-resolution quantitative imaging. To characterize the two LSTM modes and to compare them with LSM, we imaged micron-sized fluorescent beads and CLARITY-cleared[1,7] human brain tissue stained with nuclear marker DAPI (**Figure 1d, Supplementary Figure 1, Supplementary Video 1**). The resulting image volumes reveals that LSTM 2-AS indeed allows for uniform high axial resolution across the entire field-of-view, whereas both LSTM 1-AS and LSM imaging produce reduced image quality on the periphery of the field-of-view (**Figure 1d, Supplementary Figure 1 and Supplementary Video 1)**. Simultaneous two-sided illumination (**Figure 1b-right**), from two symmetrically arranged illumination arms, provides higher signal and may reduce the illumination artifacts caused by opaque objects in the illumination path thus improving the uniform planar illumination and detection pre-requisite for achieving quantitatively accurate imaging of large transparent samples.

LSTM implementation

A standard LSM system consists of a fixed orthogonal arrangement of illumination and detection optical arms that are typically positioned using independently mounted optical components. For example, CLARITY-Optimized Light-sheet Microscopy (COLM, Tomer et al 2014) uses a horizontally spread out configuration in which each component is free to move independently. The need to maintain degrees of freedom on translation as well as rotation of the entire illumination assemblies as a whole poses challenges, particularly during fine adjustments of optical alignment. To implement the illumination arms as a rigid monolithic assembly that can be easily rotated and translated, we built the first prototype on a vertically mounted breadboard (**Figure 2, Supplementary Figure 2**), using a caging system to connect all the optics to rigid frames. The

entire assemblies were then connected to the breadboard via x-y manual translation stages to allow for finer positioning adjustments. We also designed an open top sample mounting strategy by employing a 3D printed chamber (**Supplementary Figure 2b**) attached to a high accuracy x-y-z motorized stage assembly. Biological samples were mounted in a quartz cuvette of the appropriate size, tightly connected to the bottom of the sample chamber (**Supplementary Figure 2b**). We also developed adapters to mount a prism mirror for the optical alignments. The entire sample chamber assembly can be translated in 3 dimensions to acquire the image volumes. This approach allows for full exploration of various parameters of the system (such as the angular separation between the illumination and detection arms) and acquiring data from large samples by providing rigid monolithic illumination and detection units with translational and rotational degrees of freedoms.

The final overall LSTM illumination configuration includes a LASER source, collimators (~10 mm output beam diameter), ETL, cylindrical lens, galvo scanner, scan lens, tube lens and illumination objective (**Figure 2**). In addition, we incorporated an iris, after the collimator, to remove the peripheral spread of Gaussian beams, a one dimensional slit, before cylindrical lens, to control the effective numerical aperture of illumination and a second iris at the conjugate plane, between scan lens and tube lens, to control the light sheet height. The detection arm is composed of a detection objective, emission filter, tube lens and an sCMOS camera.

Since LSTM involves scanning of a line illumination-detection profile generated by the intersection of the light sheet and the detection plane, we used static sheets (generated by the use of a cylindrical lens and the associated optics), instead of a dynamic sheet (generated by rapid scanning of a pencil beam) to maximize imaging speeds. The cropping of peripheral parts of the large input diameter beam with an iris ensured a relatively uniform intensity distribution profile across the static light sheet. We used a galvo scanner to achieve rapid translation of light sheets

perpendicular to their propagation direction. Finally, for 2-AS mode, we also needed rapid translation of the thinnest part of the sheet along the propagation direction. Possible approaches here include fast piezo motors to translate the illumination objectives, using holographic spatial light modulators or an electrically tunable lens (ETL) driving induce divergence and convergence of a collimated beam. The use of piezo motors for rapid scanning of objective often results in vibrations and require additional settling time[28,30], and the spatial light modulators are limited in modulation speed because of slower refresh rates. ETLs, on the other hand, can achieve high frequency modulation of focal point position without the need for moving optics of significant mass[39]. We thus tested an ETL based approach and found it to be highly effective for achieving uniform simultaneous 2-axes scanning (**Figure 1d, Supplementary Video 1**).

The LSTM assembly was optically aligned by placing a prism mirror (with fine scratches in the center, see **Supplementary Figure 2b** for mounting arrangements) in the focal plane of detection optics, to visualize the location and cross-section of the light sheet relative to the detection focal plane. The light sheet positioning parameters were optimized such that the thinnest part was in alignment with the center of the field-of-view of the detection plane. Next, the mirror was replaced with a high concentration (>2%) agarose gel containing fluorescent beads (Note, the high concentration of agarose was used to ensure that only the surface plane of the agarose gel was visible during the alignment optimization.). By examining the extent and quality of the illuminated beads located on the surface, optimal galvo scanner and ETL parameters for achieving uniform planar illumination across the entire field-of-view were identified.

LSTM characterization

A series of calculations were performed to assess and compare various properties of LSTM (summarized in **Figure 3**). First, we devised a method to calculate the physical geometric constraints of arranging a given set of detection and illumination objectives in a non-orthogonal configuration (**Figure 3A, Supplementary Figure 3**). The main physical parameters used in the calculations are the working distances and the diameters of both the illumination and detection objectives. We calculated the range of physically-allowable, relative angular arrangements that enable light sheets to intersect the detection focal plane at their thinnest parts, while also ensuring that illumination objectives remain above the physical extent of the detection objective (**Figure 3A, Supplementary Figure 3**). For instance, only angular configuration of 43-62 degrees for Macro 4x/0.28NA/29.5mmWD (Olympus) as the illumination objective and 10x/0.6NA/8mmWD (Olympus) as the detection objective are possible (**Figure 3, Supplementary Figure 3**). Note that the working distance of this illumination objective is given for use in air, and therefore we calculated the approximate effective working distance as shown in **Supplementary Figure 3a**. Next, determined the influence of angular separation of illumination and detection arms on the resulting image volumes. We first calculated and compared the illumination path length in LSTM and LSM; shorter the illumination path length the better the image quality. In LSM, the illumination light sheet needs to penetrate the entire width of the sample for complete coverage, whereas in LSTM the effective illumination path length depends on the angular arrangement and the tissue thickness(t): $t/\cos(\theta)$. As shown in **Figure 3b**, LSTM is expected to outperform LSM for high quality imaging of large samples.

In LSTM, from an illumination path length stand point, minimizing the angular separation will increase the imaging quality. However, when the effect of θ on the effective light sheet thickness (approximated as $b/\sin(\theta)$, **Figure 3c**), which determines the axial resolution, is measured, an

inverse relationship is found: the more the θ the better the axial resolution. Because illumination is provided via a relatively low NA objective (0.28) for which the light scattering has much smaller effect on the illumination side, we decided to maximize the angular separation (~60°) to achieve higher axial resolution. All the experiments were performed using this configuration.

LSTM allows rapid quantitative imaging of large samples with high uniform resolution.

We first tested the use of lower NA illumination (hence larger field-of-view and thicker light sheets) in LSTM 1-AS configuration in a large sample, a CLARITY-cleared thick coronal section of *Thy1-eYFP* transgenic mouse brain (**Figure 4**). While, the LSTM 1-AS mode allowed for high-quality imaging of the section, image quality was reduced (marked with dotted-rectangles) for peripheral most portions of the field-of-view, even for the low NA illumination configuration. This result is similar to the imaging performance of a LSM system employing Gaussian beams for illumination.

By allowing use of high NA objectives for illumination (hence thinner sheet), LSTM 2-AS mode enables high-resolution imaging with uniform quality. To assess the quantitative imaging performance of LSTM, we performed imaging of cleared intact mouse central nervous system of *Thy1-eYFP* transgenic mouse. As demonstrated in **Figure 5a** and **Supplementary Video 3**, LSTM enables rapid high-resolution quantitative imaging of these large samples without any reduction in the image quality across the sample dimensions. We further imaged a large (~9.6mm x 13.5 mm x 5.34 mm) coronal slice of CLARITY-cleared *Thy1-eYFP* transgenic mouse brain, with 10x/0.6NA/8mmWD (**Figure 5b, Supplementary Video 4**) and 25x/1.0NA/8mm (**Figure 5c**) objectives, and larger input beam diameter (to employ the full available NA of 0.28) of the illumination objective. Note that this sample was expanded ~1.5-2 fold by incubation in glycerol

solution[28] to result in ~1.5-2 folds expansion. As demonstrated by zoom-in views of various locations of the samples, LSTM indeed provides high uniform quality across the entire samples.

Finally, we demonstrate that LSTM outperforms LSM for uniform high-resolution imaging of large samples (**Figure 6**). For this we cleared a large slice of rat brain and stained it with a relatively uniform label to visualize all the blood vessels. Previous attempts of using LSM to image rat brain resulted in poor image quality apart from the most peripheral parts of the tissue[34], as also expected because of illumination scattering. We challenged the LSTM approach with this very large sample (~2 centimeters wide and >5 mm deep) for a direct comparison with LSM imaging performance. Note that we decided to choose a highly transparent tissue (**Figure 6a inset)** and a broad label for a fair comparison. As shown in **Figure 6 and Supplementary video 5**, indeed LSTM allowed rapid uniform high-resolution imaging of the entire tissue, whereas LSM resulted in progressively reduced image quality towards the middle of the sample, similar to the previous report[34]. To complement these observations, we also performed imaging of a large piece of human brain tissue labelled with a uniform nuclear label DAPI. Human brain tissue, being dense, scatters the illumination light heavily and thus has been proven to challenging to be imaged by LSM approach. As shown in **Supplementary Figure 1b**, LSTM indeed enabled uniform high-resolution imaging of large piece of human brain tissue.

Thus, LSTM allows high-resolution quantitative imaging of large intact biological systems with no limitations on the lateral dimensions and high uniform quality superior to LSM. The sample thickness that can be imaged remains limited by the working distance of the detection objective and also by level of tissue transparency and the penetration of labelling reagents.

Discussion

Understanding the function of complex biological systems – especially highly complex mammalian brains - requires access to the intricate details of underlying structure and molecular composition, along with the functional dynamics. Over the last decades, a number of methods for clearing tissues to facilitate the interrogation of the structure and molecular architecture of large intact tissues have been developed. Methods such as CLARITY (Chung et al 2013, Tomer et al 2014) can render opaque tissues transparent while at the same time preserving structural and molecular content. Together with advances in high-speed imaging methods, these approaches provide new possibilities for understanding the functioning of tissues in health, and in diseased states such as malignant tumors and neurological damage. Methods that allow acquisition of high-resolution information on a practical time-frame such as LSM (Stelzer 2015, Migliori et al 2016) are effective for imaging of these transparent tissues because of their inherent low photo-bleaching and the high imaging speeds.

The LSM approach of illuminating a sample with a thin sheet of light, and the detecting the emitted signal with an orthogonally arranged detection arm provides two main advantages: minimal energy load and high imaging speed. We developed a highly-optimized implementation of LSM, called COLM[1,28] which allowed high-resolution imaging of entire intact mouse brains in matter of hours. However, LSM as a general approach has been restricted in the lateral dimension of image volumes because of the use of illumination light sheets from orthogonal side of the samples, putting a hard-physical limit on the imaging volume dimensions. Meanwhile, physical tissue expansion approaches [35,36] for achieving high imaging resolution are producing even larger samples. To address these challenges, we developed a conceptually distinct imaging framework, called Light Sheet Theta Microscopy (LSTM). As for LSM, LSTM is based on planar illumination, but

achieves this goal by using non-orthogonally arranged illumination objectives to produce light sheets that intersect the detection plane in a line profile, which is then synchronously (along with line-by-line detection of sCMOS camera) scanned along the detection plane. An immediate advantage of such a configuration is that it alleviates the restrictions on lateral sample dimensions, while providing uniform image quality for achieving true quantitative imaging. We found that a strategy of simultaneous 2-axes illumination (i.e. along and perpendicular to light sheet propagation) provided the best imaging performance. In comparison to LSM, LSTM allows imaging of larger samples such as a ~2 cm wide and ~5mm thick rat brain slice - with high uniform resolution across the entire sample.

LSTM is expected to allow uniform high-resolution imaging of large samples including thick slabs of cleared and labelled post-mortem healthy and diseased human brains as well as imaging of large animal intact brains, including rat and primate brains. Moreover, LSTM will facilitate *in situ* detection of thousands of transcripts directly in tissue samples. Future work will include integration of super-resolution approaches (such as structured illumination) and simultaneous multi-view imaging.

Methods

LSTM implementation. The optical layout and physical implementation details are presented in **Figure 2, Supplementary Figure 2**. A complete parts list and their sources are listed in the **Supplementary Table 1**. In the LSTM implementation, two thin light sheets are generated by using two illumination optical arms each containing LASER source, cylindrical lens, vertical slit, iris, electrically tunable lens, galvo scanner, scan lens, tube lens, mirror, and the illumination

objective (Olympus Macro 4x/0.28 NA). Since the illumination objectives were air objectives, we used a 3D printed cap (using Ultimaker 2+extended), containing a 1 inch diameter quartz coverslip, to seal the objective for oil immersion use. The emitted signal is detected by a detection arm which is composed of a detection objective, tube lens, and sCMOS camera (Hamamatsu Orca Flash 4.0 V3). The illumination arms were mounted at an approximately 60° angle relative to the detection arm on a vertically mounted breadboard fixed on an optic table with pillar posts. To facilitate the optical alignment of the system, all the three optical arms were mounted on two manual translation stages which were attached to the breadboard. We used a 3D printed (using Ultimaker 2+extended) open top sample chamber that was filled with an immersion oil of Refractive Index 1.454 (Cargille labs). The sample was mounted in a quartz cuvette (FireFly Scientific) which was then attached to the base of the sample chamber (**Supplementary Figure 2**). The microscope was controlled with a custom LabView software (available on request). The 3D model of LSTM microscope was made with Autodesk Inventor 2017.

LSTM geometric constraint calculations. The physical geometric constraints on the arrangement of illumination objectives were calculated by analyzing the two opposite bounds (**Figure 3a, Supplementary Figure 3**): the illumination objective not to touch the detection objective (**Figure 3a top**) and the illumination objective not going below the physical extent of the detection objective (**Figure 3a middle**). The range of allowable angular positions was calculated by taking the effective working distances and the objective diameters into account as shown in the schematics of **Supplementary Figure 3b**, resulting in the following relationships:

$$W2 * \sin(\theta i) = \frac{D1}{2} + D2 * \frac{\cos(\theta i)}{2}$$

$$W2 * \cos(\theta f) = W1 + D2 * \frac{\sin(\theta f)}{2}$$

where W1 and W2 are the working distances of the detection and illumination objectives respectively, D1 and D2 are the diameters of the detection and illumination objectives respectively, and θi and θf are the boundary angular positions, as shown in **Figure 3a, Supplementary Figure 3b**. We used Macro 4x/0.28NA/29.5WD (Olympus) as the illumination objectives. These objectives are designed to be used in air, therefore we used Snell's law to calculate the approximate effective working distance in a refractive index liquid of 1.454, as shown in **Supplementary Figure 3a**, resulting in ~44 mm, and the objective diameter (lowest part of the tapered ending) was measured to be ~28 mm. For most of the experiments we used 10x/0.6NA/8mmWD objective (Olympus) in the detection arm with values of W1 = 8, and D1 = 40. For this combination, we found the allowable angular range to be 43.3° to 62.3° degrees. This calculated range was our initial guideline for identifying the maximum possible angular positioning during system alignment. We used ~60° as the final angle separation. Calculations were performed in Matlab.

Illumination depth calculations. We used geometric calculations (**Figure 3b**) to estimate the maximum illumination path lengths of LSTM as $t / \cos(\theta)$, where t is the sample thickness to be imaged and θ is the angle between the illumination propagation direction and the detection axis. The maximum illumination path length in LSM would be the same as the sample width (w). We calculated the ratio of these illumination path lengths which was converted into a binary representation by thresholding at 1, and plotted as a heat map, shown in **Figure 3b**.

LSTM effective light sheet thickness calculations. Due to the non-orthogonal incidence of the light sheet on the detection plane, the effective light sheet thickness can be approximated as the projection of the original thickness on to the detection direction, resulting in $b / \sin(\theta)$, where b is

the original light sheet thickness at the most focused position, and θ is the angle of incidence relative the detection axis. The relationship was plotted in the graph shown in **Figure 3c**.

Imaging experiments. We used the passive CLARITY method (as described previosuly[1]) for all the tissue clarification experiments. The hydrogel monomer (HM) solution recipe consisted of 1-4% (wt/vol) acrylamide, 0.05% (wt/vol) bisacrylamide, 4% paraformaldehyde, 1x Phosphate Buffered Saline, deionized water, and 0.25% thermal initiation VA-044(Wako Chemicals, NC0632395). For whole brain clearing, trans-cardial perfusion was performed with 20 ml HM solution, followed by overnight incubation at 4°C. The rate brain was perfused with 4% paraformaldehyde (PFA), post-fixed for 16 hours and then frozen in isopentane for storage. The frozen brain was thawed at room temperature in PBS buffer, sliced and incubated in HM solution overnight at 4°C. The human brain tissue was incubated in 4% PFA for ~2 days, followed by incubation in HM solution up to days at 4°C. All the perfused tissues were de-gassed and transferred to 37°C for 3-4 hours for hydrogel polymerization. The tissues were cleared by incubating (with shaking) in clearing buffer (4% (wt/vol) SDS, 0.2 M boric acid, pH 8.5) at 37°C until clear (2-3 weeks). The rat brain and human brain slices were incubated in the HM solution overnight at 4°C followed by degassing to replace the oxygen with nitrogen, and incubation at 37°C for 3-4 hours for polymerization. After washing off any remaining HM solution, the tissues were incubated in buffered clearing solution (4% (wt/vol) SDS, 0.2 M boric acid, pH 8.5) at 37°C with shaking until the tissue became clear (1-4 weeks). Afterwards, the tissue was washed with 0.2 M boric acid buffer (pH 8.5) with 0.1% Triton X-100 for up to 24 hours. The cleared tissue was labelled with DAPI (1 μg/mL final concentration) and/or blood vessels marker tomato lectin (Vector Labs, FL-1171), by incubating in the labelling solution for 3-4 days. After washing with the buffered solution (0.2 M boric acid buffer, pH 7.5, 0.1% Triton X-100), the tissue was

transferred into 85-87% glycerol solution in graded fashion (i.e. 25%, 50%, 65% and finally 87%) for final clearing and imaging. All the image volumes were acquired with a 2 or 5 microns step-size.

Image analyses. TeraStitcher [40] based pipeline[1] was used for stitching of acquired image stack tiles of all the datasets. Maximum intensity projections, and other linear image contrast adjustments were performed using Fiji [41,42] and Matlab. All volume renderings were performed using Amira (FEI). All the fluorescent beads image analysis was performed using Fiji. To calculate the axial Full Width at Half Maximum (FWHM), x-z projections of beads image stacks were used. For individual beads a line intensity profile was calculated along the central position, followed by manual calculations of full width at half maximum intensity values.

Data availability statement. All the datasets reported in this paper, ranging in tens of terabytes will be made available on request. Complete CAD model of LSM and other related resources will be made available with the manuscript and on a dedicated resources webpage. Complete parts list is included as a supplementary table.

Materials & Correspondence requests to Raju Tomer (raju.tomer@columbia.edu).

AUTHOR CONTRIBUTIONS

R.T. conceived the project and designed the microscopes. B.M., M.S.D and R.T. built the microscopes. M.S.D., B.M. and R.T. performed the imaging experiments. M.C.A. assisted with all the experiments. R.T. and B.M. analyzed the data. R.T., B.M. and M.S.D. wrote the paper. R.T. supervised all aspects of the work.

ACKNOWLEDGEMENTS

We thank Weijian Yang and Rafael Yuste for discussions and advice on the ETL use. We are grateful to Rafael Yuste, Darcy Kelley and Oliver Hobert for general advice and reading of the manuscript draft. This work is supported by Columbia University Arts & Sciences startup grant to R.T. We thank Ola Hermanson for supporting B.M. and helpful discussions. We would like to thank Serge Przedborski for providing rodent brain tissues and Peter Canoll for the human brain tissue.

Figure 1

Figure 1. **Light Sheet Theta Microscopy (LSTM) for high-resolution quantitative imaging of large intact samples**. (a) Schematic comparison of Light Sheet microscopy (LSM), inverted LSM and Light Sheet Theta Microscopy (LSTM). LSM involves the illumination and the detection optics to be orthogonally arranged, thus putting limitations on the sample size along two of the sample dimensions i.e. along the detection and the illumination axis. Alternative implementations such as OCPI, iSPIM (shown) and SCAPE are partially effective to address this for smaller samples, however at the cost of reduction in the effective detection working distance (magenta arrowheads) and image quality needed for quantitative imaging of larger samples. The LSTM configuration, which involves illumination through non-orthogonal axes (<90°), can effectively address these limitations, while providing high imaging speed, uniform resolution across large field-of-view, and low photo-bleaching. (b) Schematic representation of line detection strategy for optical sectioning. One or two thin sheets of light intersect with the detection plane in a line, which is then synchronously scanned (along with the rolling shutter of sCMOS camera) in the focal plane, resulting in uniform planar imaging. (c) Schematic summary of two modes of LSTM line scanning. The 1-Axis scanning (1-AS) involves the translation of the light sheet perpendicular to their propagation direction, whereas the simultaneous 2-Axes scan (2-AS) involves the translation perpendicularly as well as along the illumination direction such that the thinnest part of the sheet intersects with the detection plane. The latter mode results in uniform axial resolution across the entire field-of-view. (d) Comparison of optical sectioning in LSTM 1-AS, LSTM 2-AS and LSM configurations. Left shows x-z projection of 1 micron fluorescent beads imaged with LSTM 1-AS, LSTM 2-AS and LSM using the same detection objective (Olympus 10X/0.6NA/8mmWD). The illumination and detection objectives in LSTM configuration were arranged at ~60° angular separation (as discussed in **Figure 3**). Axial FWHMs (Full Width at Half Maximum) were

calculated from beads positioned at different positions along the field-of-view and plotted in the bottom graph with corresponding colors (blue: LSTM in 2-AS mode, Green: LSTM in 1-AS mode, red: LSM). LSTM 2-AS achieves uniform optical section over the entire field-of-view whereas both LSTM in 1-AS mode and LSM results in similarly reduced axial resolution on the peripheries. Right shows the x-z projections (20 microns thick) of an image stack acquired from a DAPI stained CLARITY-cleared Human brain slice, using LSTM 1-Axis and LSTM 2-Axes scan modes with 10X/0.6NA/8mmWD (Olympus) detection objective (**Supplementary Figure 1** compares different exposure settings, **Supplementary Video 1** provides detailed 3D volume rendering.). This example demonstrates that LSTM enables the use of very thin light sheets for achieving good axial resolution over a large field-of-view. All scale bars are 100 microns.

Figure 2

Figure 2. **LSTM microscopy implementation**. (a) Schematic summary of the optical path of LSTM microscopy. Two light sheets are generated by using a cylindrical lens (CL), scan lens (SL), tube lens (TL) and illumination objectives (Olympus Macro 4x). The galvo scanner is used to translate the light sheets perpendicular to their propagation direction, and the electrically tunable lens (ETL) is used to translate the thinnest part of the sheet along the propagation direction. Custom made collimators were used to generate an input beam of ~10 mm diameter, which is then trimmed through an iris. A slit is placed after the ETL to reduce the beam diameter in one direction in order to control the effective numerical aperture of the illumination. An additional Iris is placed at the conjugate plane between the SL and TL to control the light sheet height for minimizing any unwanted illumination. The illumination axes are arranged at an approximate angle of 60° with the detection axis. A custom 3D-printed cap with a quartz coverslip was attached to the illumination objective (designed for imaging in air). The detection arm consists of the detection objective (Olympus 10x/0.6NA/8mmWD or 25x/1.0NA/8mmWD), the tube lens and a sCMOS camera (Hamamatsu Orca Flash V3.0). (b) 3D model of the entire LSTM microscope. A vertical breadboard was used to mount the optical components, via x-y translation stages to allow for degree of freedom during the optical alignment. A sample chamber was mounted on a 3 axes (x,y,z) motorized stage. Samples were mounted in a quartz cuvette of appropriate size, which was then attached at the bottom of the sample chamber. Imaging of the entire sample was performed by moving the sample chamber. **Figure 3, Supplementary Figures 2-3, Supplementary Video 2 and Supplementary Table 1** provide further details.

Figure 3

Figure 3.   LSTM design parameters. (a) Physical geometric constraints for arranging the detection and illumination objectives. The schematics summarize the analyses for a particular set of objectives that were commonly used for LSTM imaging: Olympus Macro 4x/0.28NA/29.5WD for illumination, and Olympus 10x/0.6NA/8mmWD for detection. Two extreme angular positions are shown. Note that the specified working distance of the illumination objective (in air) was used to calculate the elongated working distance (owing to the high refractive immersion oil), as shown in **Supplementary Figure 3**. For all the imaging experiments, the angular separation between the illumination and detection optics was set to ~60°. The bottom image shows an example of the

physical implementation of generating illuminating thin sheets. (b) Comparison of the illumination path length in the LSTM and LSM. In LSM, the illumination light sheet will need to penetrate the entire width of the sample, whereas in LSTM it is dependent on the angular arrangement and tissue thickness (t): $t/\cos(\theta)$. The left graph plots the dependence of illumination depth of LSTM on θ. Each color curve represents a different tissue thickness(t), as marked by the arrow in the direction of increase in thickness. The right graph compares the illumination depth required to image a sample of given width and thickness. The ratio of the LSTM and LSM illumination path was converted into a binary representation to summarize the parameters ranges where LSTM (magenta) and LSM (cyan) need smaller illumination depth, and hence will provide better image quality. (c) Effective planar illumination thickness can be approximated as $b/\sin(\theta)$, where b is the light sheet thickness. The right graph plots the effective light sheet thickness as a function of θ for different values of b (the arrow points in the direction of increasing b value). (d) Comparison of energy load on the sample while imaging with LSM and LSTM. The two rows compare the time-accumulated energy load in LSM and LSTM for imaging of a single plane and a single image stack tile. The energy load in LSTM is dependent on the sample thickness for a given angular arrangement, and in LSM on the sample width. For larger samples LSTM energy load is similar to that in LSM. Note the schematics only show one sided illumination configuration for both LSTM and LSM.

Figure 4

Figure 4. LSTM imaging quality in 1-AS mode. A thick slice of CLARITY-cleared *Thy1-eYFP* transgenic mouse brain was imaged with LSTM using 1-Axis scanning. Low numerical aperture illumination was used to generate a large field-of-view (and hence thicker) light sheets. The bounding box is 8.3mm x 7.3 mm x 2 mm. X-Y and X-Z maximum intensity projections of a stack at marked positions are shown on the right side. Similar to standard LSM, the image quality partially degrades on the periphery of the stack (outlined by the yellow dotted rectangles). The scale bar is 100 microns.

Figure 5

Figure 5. Rapid uniform high-resolution imaging of intact mouse central nervous system. (a) A CLARITY-cleared intact *Thy1-eYFP* transgenic mouse brain, with attached spinal cord, was imaged with LSTM microscopy (10x/0.6NA objective). High-resolution 3D rendering, using 2x2 fold down-sampled data, is shown for the entire tissue and for a sub-volume marked by magenta rectangles. Two orthogonal views, and zoomed-in images are shown as marked by corresponding colors. The bounding boxes are 11.8mm x 27.6 mm x 5.2 mm for the whole sample, and 5.1mm x 3.1mm x 3.5mm for the sub-volume. A detailed volume rendering is shown in **Supplementary Video 3**. (b) A large coronal slice of a *Thy1-eYFP* transgenic mouse brain was imaged with LSTM to demonstrate the uniform high quality imaging of LSTM. The volume rendering was performed using 4x4 fold down-sampled data. Zoomed-in images are shown for the marked colored squares, demonstrating image quality across different locations of the image volume. The bounding box is 9.6mm x 13.5 mm x 5.34 mm. **Supplementary Video 4** show detailed volumetric rendering. (c) The same sample was imaged with a high-NA 25x/1.0NA objective. The volume rendering was performed from 2x2 fold down-sampled images, and the zoomed-in images are shown at different locations across the sample, as marked by the colored squares. The bounding box is 6mm x 9.6 mm x 0.5 mm.

Figure 6

Figure 6. LSTM enables uniform high-quality imaging of large samples. (a) For an unbiased comparison of the imaging performance of LSTM and LSM, we used a highly cleared large tissue sample (~2 cm wide and ~5 mm, inset in the top image) from rat brain. The sample was stained for visualizing the vasculature distribution. The images shown are maximum intensity z-projections, and were acquired using the same 10x/0.6NA objective. The bottom graph profiles the mean intensity (in the vertical direction) across the specified region of interest (ROI) marked with dotted rectangles. In LSM, the intensity signal is progressively degraded towards the interior

of the sample, whereas LSTM allows uniform quality across the entire sample. The scale bars are 1 millimeter. (b) An example image stack from the LSTM dataset. Optical slices (50 microns thick) are shown at three different depths as identified (orange) in the stack. The bounding box of the stack is 1mm x 1mm x 5mm. The scale bars are 100 microns. A detailed volume rendering is shown in **Supplementary Video 5**.

REFERENCES

Tomer, R., Ye, L., Hsueh, B. & Deisseroth, K. Advanced CLARITY for rapid and high-resolution imaging of intact tissues. *Nature protocols* **9**, 1682-1697, doi:10.1038/nprot.2014.123 (2014).

Renier, N. *et al.* iDISCO: a simple, rapid method to immunolabel large tissue samples for volume imaging. *Cell* **159**, 896-910, doi:10.1016/j.cell.2014.10.010 (2014).

Pan, C. *et al.* Shrinkage-mediated imaging of entire organs and organisms using uDISCO. *Nat Methods* **13**, 859-867, doi:10.1038/nmeth.3964 (2016).

Romanov, R. A. *et al.* Molecular interrogation of hypothalamic organization reveals distinct dopamine neuronal subtypes. *Nat Neurosci* **20**, 176-188, doi:10.1038/nn.4462 (2017).

Dodt, H. U. *et al.* Ultramicroscopy: three-dimensional visualization of neuronal networks in the whole mouse brain. *Nat Methods* **4**, 331-336, doi:10.1038/nmeth1036 (2007).

Susaki, E. A. *et al.* Whole-brain imaging with single-cell resolution using chemical cocktails and computational analysis. *Cell* **157**, 726-739, doi:10.1016/j.cell.2014.03.042 (2014).

Chung, K. *et al.* Structural and molecular interrogation of intact biological systems. *Nature* **497**, 332-337, doi:10.1038/nature12107 (2013).

Yang, B. *et al.* Single-cell phenotyping within transparent intact tissue through whole-body clearing. *Cell* **158**, 945-958, doi:10.1016/j.cell.2014.07.017 (2014).

Hama, H. *et al.* Scale: a chemical approach for fluorescence imaging and reconstruction of transparent mouse brain. *Nat Neurosci* **14**, 1481-1488, doi:10.1038/nn.2928 (2011).

Erturk, A. *et al.* Three-dimensional imaging of solvent-cleared organs using 3DISCO. *Nature protocols* **7**, 1983-1995, doi:10.1038/nprot.2012.119 (2012).

Ke, M. T., Fujimoto, S. & Imai, T. SeeDB: a simple and morphology-preserving optical clearing agent for neuronal circuit reconstruction. *Nat Neurosci* **16**, 1154-1161, doi:10.1038/nn.3447 (2013).

Becker, K., Jahrling, N., Saghafi, S., Weiler, R. & Dodt, H. U. Chemical clearing and dehydration of GFP expressing mouse brains. *PLoS One* **7**, e33916, doi:10.1371/journal.pone.0033916 (2012).

Murray, E. *et al.* Simple, Scalable Proteomic Imaging for High-Dimensional Profiling of Intact Systems. *Cell* **163**, 1500-1514, doi:10.1016/j.cell.2015.11.025 (2015).

Kuwajima, T. *et al.* ClearT: a detergent- and solvent-free clearing method for neuronal and non-neuronal tissue. *Development* **140**, 1364-1368, doi:10.1242/dev.091844 (2013).

Richardson, D. S. & Lichtman, J. W. Clarifying Tissue Clearing. *Cell* **162**, 246-257, doi:10.1016/j.cell.2015.06.067 (2015).

Migliori, B., Datta, M. S. & Tomer, R. Advanced light microscopy enables rapid mapping of brain structure and function in high resolution. *Laser Focus World* **52**, 47-52 (2016).

Lerner, T. N. *et al.* Intact-Brain Analyses Reveal Distinct Information Carried by SNc Dopamine Subcircuits. *Cell* **162**, 635-647, doi:10.1016/j.cell.2015.07.014 (2015).

Siedentopf, H. & Zsigmondy, R. Uber Sichtbarmachung und Grössenbestimmung ultramikroskopischer Teilchen, mit besonderer Anwendung auf Goldrubingläser. *Annalen der Physik* **10**, 1-39 (1903).

Stelzer, E. H. Light-sheet fluorescence microscopy for quantitative biology. *Nat Methods* **12**, 23-26, doi:10.1038/nmeth.3219 (2015).

Huisken, J., Swoger, J., Del Bene, F., Wittbrodt, J. & Stelzer, E. H. Optical sectioning deep inside live embryos by selective plane illumination microscopy. *Science* **305**, 1007-1009, doi:10.1126/science.1100035 (2004).

Keller, P. J., Schmidt, A. D., Wittbrodt, J. & Stelzer, E. H. Reconstruction of zebrafish early embryonic development by scanned light sheet microscopy. *Science* **322**, 1065-1069, doi:10.1126/science.1162493 (2008).

Reynaud, E. G., Peychl, J., Huisken, J. & Tomancak, P. Guide to light-sheet microscopy for adventurous biologists. *Nat Methods* **12**, 30-34, doi:10.1038/nmeth.3222 (2015).

Wu, Y. *et al.* Spatially isotropic four-dimensional imaging with dual-view plane illumination microscopy. *Nat Biotechnol* **31**, 1032-1038, doi:10.1038/nbt.2713 (2013).

Preibisch, S., Saalfeld, S., Schindelin, J. & Tomancak, P. Software for bead-based registration of selective plane illumination microscopy data. *Nat Methods* **7**, 418-419, doi:10.1038/nmeth0610-418 (2010).

Planchon, T. A. *et al.* Rapid three-dimensional isotropic imaging of living cells using Bessel beam plane illumination. *Nat Methods* **8**, 417-423, doi:10.1038/nmeth.1586 (2011).

Gao, L. *et al.* Noninvasive imaging beyond the diffraction limit of 3D dynamics in thickly fluorescent specimens. *Cell* **151**, 1370-1385, doi:10.1016/j.cell.2012.10.008 (2012).

Chen, B. C. *et al.* Lattice light-sheet microscopy: imaging molecules to embryos at high spatiotemporal resolution. *Science* **346**, 1257998, doi:10.1126/science.1257998 (2014).

Tomer, R. *et al.* SPED Light Sheet Microscopy: Fast Mapping of Biological System Structure and Function. *Cell* **163**, 1796-1806, doi:10.1016/j.cell.2015.11.061 (2015).

Holekamp, T. F., Turaga, D. & Holy, T. E. Fast three-dimensional fluorescence imaging of activity in neural populations by objective-coupled planar illumination microscopy. *Neuron* **57**, 661-672, doi:10.1016/j.neuron.2008.01.011 (2008).

Ahrens, M. B., Orger, M. B., Robson, D. N., Li, J. M. & Keller, P. J. Whole-brain functional imaging at cellular resolution using light-sheet microscopy. *Nat Methods* **10**, 413-420, doi:10.1038/nmeth.2434 (2013).

Bouchard, M. B. *et al.* Swept confocally-aligned planar excitation (SCAPE) microscopy for high speed volumetric imaging of behaving organisms. *Nat Photonics* **9**, 113-119, doi:10.1038/nphoton.2014.323 (2015).

Vladimirov, N. *et al.* Light-sheet functional imaging in fictively behaving zebrafish. *Nat Methods* **11**, 883-884, doi:10.1038/nmeth.3040 (2014).

Chhetri, R. K. *et al.* Whole-animal functional and developmental imaging with isotropic spatial resolution. *Nat Methods* **12**, 1171-1178, doi:10.1038/nmeth.3632 (2015).

Stefaniuk, M. *et al.* Light-sheet microscopy imaging of a whole cleared rat brain with Thy1-GFP transgene. *Sci Rep* **6**, 28209, doi:10.1038/srep28209 (2016).

Chen, F., Tillberg, P. W. & Boyden, E. S. Optical imaging. Expansion microscopy. *Science* **347**, 543-548, doi:10.1126/science.1260088 (2015).

Ku, T. *et al.* Multiplexed and scalable super-resolution imaging of three-dimensional protein localization in size-adjustable tissues. *Nat Biotechnol* **34**, 973-981, doi:10.1038/nbt.3641 (2016).

Wu, Y. *et al.* Inverted selective plane illumination microscopy (iSPIM) enables coupled cell identity lineaging and neurodevelopmental imaging in Caenorhabditis elegans. *Proc Natl Acad Sci U S A* **108**, 17708-17713, doi:10.1073/pnas.1108494108 (2011).

Strnad, P. *et al.* Inverted light-sheet microscope for imaging mouse pre-implantation development. *Nat Methods* **13**, 139-142, doi:10.1038/nmeth.3690 (2016).

Fahrbach, F. O., Voigt, F. F., Schmid, B., Helmchen, F. & Huisken, J. Rapid 3D light-sheet microscopy with a tunable lens. *Opt Express* **21**, 21010-21026, doi:10.1364/OE.21.021010 (2013).

Bria, A. & Iannello, G. TeraStitcher - a tool for fast automatic 3D-stitching of teravoxel-sized microscopy images. *BMC Bioinformatics* **13**, 316, doi:10.1186/1471-2105-13-316 (2012).

Schindelin, J. *et al.* Fiji: an open-source platform for biological-image analysis. *Nat Methods* **9**, 676-682, doi:10.1038/nmeth.2019 (2012).

Schneider, C. A., Rasband, W. S. & Eliceiri, K. W. NIH Image to ImageJ: 25 years of image analysis. *Nat Methods* **9**, 671-675 (2012).