PHENOS: a high-throughput and flexible tool for microorganism growth phenotyping on solid media

David B. H. Barton[1], Danae Georghiou[1], Neelam Dave[1], Majed Alghamdi[1], Thomas Walsh[1], Edward J. Louis[1±], Steven S. Foster[1*]

*Corresponding author: ssf5@le.ac.uk

±Co-corresponding author: ejl21@le.ac.uk

[1]Department of Genetics & Genome Biology, University of Leicester, Leicester LE1 7RH, UK.

Other email addresses: D.B.H.B.: dbh8@le.ac.uk; D.G.: dg217@le.ac.uk; N.D.: njd25@le.ac.uk; M.A.: mafa2@le.ac.uk; T.W.: tw164@protonmail.com

Abstract

BACKGROUND Microbial arrays, with a large number of different strains on a single plate printed with robotic precision, underpin an increasing number of genetic and genomic approaches. These include Synthetic Genetic Array analysis, high-throughput Quantitative Trait Loci (QTL) analysis and 2-hybrid techniques. Measuring the growth of individual colonies within these arrays is an essential part of many of these techniques but is useful for any work with arrays. Measurement is typically done using intermittent imagery fed into complex image analysis software, which is not especially accurate and is challenging to use effectively. We have developed a simple and fast alternative technique that uses a pinning robot and a commonplace microplate reader to continuously measure the thickness of colonies growing on solid agar, complemented by a technique for normalizing the amount of cells initially printed to each spot of the array in the first place. We have developed software to automate the process of combining multiple sets of readings, subtracting agar absorbance, and visualizing colony thickness changes in a number of informative ways. **RESULTS** The "PHENOS" pipeline (<u>phen</u>otyping <u>on</u> <u>s</u>olid media), optimized for *Saccharomyces* yeasts, produces highly reproducible growth curves and is particularly sensitive to low-level growth. We have empirically determined a formula to estimate colony cell count from an absorbance measurement, and shown this to be comparable with estimates from measurements in liquid. We have also validated the technique by reproducing the results of an earlier QTL study done with conventional liquid phenotyping, and found PHENOS to be considerably more sensitive. **CONCLUSIONS** "PHENOS" is a cost effective and reliable high-throughput technique for quantifying growth of yeast arrays, and is likely to be equally very useful for a range of other types of microbial arrays. A detailed guide to the pipeline and software is provided with the installation files at https://github.com/gact/phenos.

Keywords

Arrays; Yeast; Phenotyping; QTL; SGA; Microbes; Growth

Background

Around the turn of the new millennium, following developments in DNA microarray technology, microbial geneticists began to work with large regularly ordered arrays of colonies of genetically distinct microorganisms. For example, *Saccharomyces cerevisiae* deletion mutant arrays were announced in 1999 [1] and made it possible to rapidly screen the whole yeast genome for synthetic lethal interactions in Synthetic Genetic Array (SGA) analyses [2, 3]. That same technique was soon expanded to other species like the fission yeast *Schizosaccharomyces pombe* [4], and the bacterium *Escherichia coli* [5]. Other labs were also putting gene expression libraries into *E. coli* or in *S. cerevisiae* arrays for high-throughput screening [6, 7].

As these approaches gained ground, robotic colony manipulation systems became commercially viable tools to help create and duplicate these microbial arrays. For example, the VersArray™ Colony Picker and Arrayer Systems from BioRad Laboratories, the QPix 400 Series Microbial Colony Pickers from Molecular Devices, and the easy-to-use ROTOR HDA Colony Manipulation Robot (Singer Instruments in Somerset, UK) which was specifically conceived for SGA analysis.

In just the last few years, the ROTOR HDA system alone has been used for purposes as diverse as screening *S. cerevisiae* for thermotolerance [8], working with *S. pombe* auxotroph deletion libraries [9], phenotyping the wild yeast *Lachancea quebecensis* [10], screening mutagenized strains of the unicellular alga *Chlamydomonas reinhardtii* [11], and generating *E. coli* synthetic genetic arrays [12]. The ROTOR HDA uses disposable plastic 'repads' of pins to print cells from one array to another in a single action. Long-pin repads can be used to transfer material to or from microtitre plates containing liquid cultures and to or from solid agar lawns in rectangular dishes (such as Singer's own proprietory 'PlusPlates™'). 96 spot arrays can easily be combined into 384 arrays, or 384 into 1536, by superimposing each source array with an offset. Larger arrays can easily be broken down into multiple smaller arrays by reversing that process.

Microbial growth can be quantified in liquid media in microtitre plates, and analysed using packages

such as GrowthRates [13] or GATHODE [14]. It can most accurately be measured by measuring

turbidity using specialist machines such the Bioscreen C (Oy Growth Curves Ab Ltd) [15, 16] which

holds two 100× 600 µl sample plates with a proprietary honeycomb format meant to create more

even heat distribution through the samples. However, it is challenging to simultaneously inoculate

these hexagonal-grid wells from the rectangular-grid high-density ROTOR arrays, and to find a level

of agitation sufficient to prevent flocculation that does not introduce bubbles or risk cross-

contamination between wells. Another solution, the oCelloScope (Philips), uses standard microtitre

plates but can only read 96 wells at once. The Omnilog (Biolog) can handle 50x 96-well plates

simultaneously, but at a considerable cost and footprint.

It is becoming more common and is certainly more convenient to measure the growth of whole

colonies growing on solid agar lawns in rectangular dishes. This is usually achieved by feeding digital

snapshots, taken intermittently, into sophisticated image-analysis software, such as the Screenmill

Colony Measurement Engine [17] (as open-source macro for the Java-based ImageJ software),

Colonyzer [18] (a command-line app written in Python 2.7), YeastXtract [19] or the Colony-Live

imaging system [20]. Specialist instruments have now become commercially available which perform

this imaging and analysis in an automated fashion, for example the Phenobooth (Singer Instruments,

Somerset, UK). Such an instrument greatly simplifies the process, and incorporating incubation and

imaging functions into the same device overcomes limitations in the number of timepoints at which

measurements are taken. However, image-based solutions remain vulnerable to confusion by specks

and bubbles, particularly when measuring small colonies.

To overcome problems with these existing approaches and save time and money, we developed an

alternative technique which uses a commonplace microplate reader (BMG Labtech's FLUOstar

Omega, which can incubate between 25 and 45°C and accommodate 96, 384 and 1536 plates) to

measure the thickness of colonies printed with the ROTOR at 20 minute intervals. We have

demonstrated that this measure is a reliable proxy for overall colony size. We developed the

PHENOS software, with a graphical user interface (GUI), to run on Windows PCs alongside the

microplate reader's own software. This software combines multiple output files from the microplate

readers and automatically generates a variety of data visualizations, and output files containing

various growth curve summary values. The software is written in Python 2.7 and so could be

adapted to other operating systems without great difficulty. It was designed to make it relatively

easy to add new parser classes to recognise and read new formats of input files that might be

generated by different microplate readers. Alternatively, users may prefer to write their own code

for analyzing the data, or may find their microplate reader's native software sufficient for their

purposes.

We validated the PHENOS pipeline by comparing its measurements to two additional low-

throughput means of measuring colony population: use of a cell counting chamber under a

microscope, and absorbance in liquid media using a spectrophotometer. Comparing both

absorbance methods to the manual cell count, we found PHENOS to be of comparable accuracy to

the spectrophotometer. We also used this data to empirically derive a formula to estimate cell count

from PHENOS readings. Because PHENOS measures absorbance through the thickness of a colony (a

roughly cylindrical volume) as a proxy for overall colony size (a roughly hemispherical volume), there

is not a straightforward linear relationship between the absorbance readings and cell count.

For additional validation and to demonstrate one of its potential uses, we used PHENOS

measurements to reproduce a Quantitative Trait Loci (QTL) analysis previously undertaken using

conventional liquid media growth measurements. QTL analysis is a powerful approach to

understanding complex genetic architecture by correlating phenotypes with genotypes [21]. It can

be substantially automated through the use of large arrays of genetically diverse microorganisms.

With sufficient genetic markers determined (e.g. by sequencing) for each strain in the array, growth

phenotypes and genotypes can be passed into the statistical analysis package R/qtl [22]. Previously

our lab described a set of 576 genetically-diverse haploid F1 *S. cerevisiae* strains derived from

crosses between four geographically-diverse parent strains, chosen because they encapsulate much

of the genetic diversity discovered in a genomic survey of *S. cerevisiae* [23]. The parents were

YPS128 (herein designated 'A' for American), DBVPG6765 ('E' for European), Y12 ('S' for Sake) and

DBVPG6044 ('W' for West African), each made genetically tractable with knockouts of the *URA3*

uracil-metabolism gene and *HO* gene (thus preventing mating type switching and keeping strains

haploid) [24]. From each of the six pairwise crosses between these parental strains, 24 tetrads were

dissected, yielding 96 haploid F1s. These progeny are genetically and phenotypically extremely

diverse. They were genotyped at over 200 markers using high-resolution melt quantitative PCR [25]

to distinguish parental allele types, and then phenotyped for up to 23 different stress conditions in

liquid media. 82 significant QTLs were described in total, including 4 distinct regions for resistance to

the herbicide paraquat [26]. We phenotyped the same panel on solid YPD media containing 500

µg/µl paraquat and used the PHENOS pipeline to automatically generate data for entry into R/qtl

[25]. We found the PHENOS method reproduced the QTLs from the earlier study but showed greater

sensitivity as 18 additional regions were identified.

We consider a couple of the distinctive characteristics of PHENOS growth curves: an initial drop in

absorbance is commonplace, and as colonies grow and absorbance measurements increase,

variance in those measurements also increases. We investigate these properties experimentally and

report their probable causes in our discussion and conclusions. Neither affects the utility of PHENOS,

which produces growth curves that are both highly reproducible and accurate enough for

statistically sensitive analyses such as QTL analysis.

QTL analysis on genetically diverse populations, by revealing genetic causes of phenotypic variation,

is a valuable tool for determining the mode of action of drugs and stresses. But PHENOS will also be

useful for measuring population trait variability in general, synthetic genetic interactions, assessing

the quality of printed arrays, or for any of the other whole colony array techniques mentioned

above.

Methods

We will first describe, in general terms, the methods used in the PHENOS pipeline itself, including

the punch-in technique for normalizing the amount of cells printed to each spot. For much greater

detail, users can consult the guide which is available with the source code for the PHENOS software

at https://github.com/gact/phenos. As part of this walkthrough, we also show examples of the

visualizations that the PHENOS software produces. We then describe two methods used to validate

this approach to phenotyping. One of these, using QTL analysis, provides a demonstration of how

the PHENOS pipeline can be used in practice. We also describe a simple experiment to investigate

properties of the PHENOS growth curves.

Overview

An overview of the experimental side of the PHENOS pipeline is given in Figure 1. The experimental

plate is put into the microplate reader before any cells have been printed to it, in order to measure

and later subtract out absorbance due to agar. As part of the printing process, starting cell masses

can be partially normalized using the 'punch-in' technique described below. The microplate reader is

set up to store each set of continuous readings as a different text file, and it is possible to generate

growth curves even from intermittent readings. As summarised in Figure 2, the PHENOS software

then gathers information from the user about the experiment (for instance the treatment and the

name of the array being used, which will be used to look up information about individual strains in

the layout, which is stored in a separate file), combines multiple files taken from a single

experimental plate, and provides visualizations and analyses. The PHENOS software is written for

Windows (so that it can operate alongside the microplate reader software) in Python 2.7, and has a

simple GUI-based interface which can be launched by double-clicking on a desktop icon. It stores

readings and other experimental data in HDF5 databases to allow complex queries and comparisons

between multiple experiments, for example to measure final growth levels as a treatment/control

ratio for use as a phenotype in QTL analysis.

For our experiments, we typically compare normal growth (usually on standard rich YPD media) to

growth under a phenotype-discriminating treatment (e.g. the same media containing a genotoxin).

Our control experiments are normally arrayed at 384 density, with 2-4 replicates of each strain, on

40ml of YPD agar media: 10g/l Yeast Extract, 20g/l Bacto Peptone, 20g/l D-Glucose, 1% adenine

solution (0.5% adenine in 0.05M HCl), adjusted to pH 6.3 using 1M HCl, then autoclaved with 20g/l

Bacto agar. Treatment experiments are grown on the same media but with a treatment added (e.g. a

toxic compound at a discriminating concentration, or a different incubation temperature).

Absorbance readings (optical density, or OD) are taken at 600nm (on a FLUOstar Optima microplate

reader, a 595nm filter can be used), every 20 minutes. Control experiments are run for

approximately 65 hours (over a weekend) so they can be compared to as many treatment

experiments as possible, as some treatment conditions do not produce discriminating growth

phenotypes until much later. For many growth phenotypes, however, an overnight program run for

~16 hours is sufficient to discriminate between phenotypes based on final growth levels. Measuring

full growth curves gives experimenters the clearest overview from which to judge the best summary

values and time window to discriminate between different growth phenotypes. Once this judgement

has been made, it is possible to increase productivity by taking only intermittent 'snapshot'

measurements of each plate at the appropriate points. Our snapshot protocols take readings of a

plate four times in rapid succession (taking under 20 minutes for a 384 array), in order to average

over the measurement variance observed at high absorbance values, a phenomenon discussed in

more detail below.

Examples of the visualizations produced by the PHENOS software are given in Figure 3. Additionally,

the PHENOS software includes the option to generate time-lapse mp4 movies of colony thickness

measurements, which are an effective means for calling attention to any location-related biases in

growth. For each experiment (combining any number of sets of measurements of one experimental

plate), PHENOS also generates a text file summary of the combined experimental data, a text file

containing summary values of the curve for each colony and, if genotype data has been provided, a

text file containing both summary values and genotype data suitable for analysis using the R/qtl

package.

The PHENOS software is written with a flexible architecture that allows users to configure which

visualizations and summary values are included in the output by editing a text configuration file

called 'config.txt'. Each summary value is handled by a specific 'phenotype calculator' subroutine,

and new ones can be written into the software with relative ease. These calculator routines are

named in 'camel case', following programming convention. The currently available calculators are

shown in Figure 4 and listed in 'config.txt'. The calculators used by default can be changed by editing

this file, or particular calculators can be specified for particular treatments. The PHENOS software

can be used to pair a treatment experiment with a control experiment, creating a 'controlled

experiment' database entry that allows for more complex phenotype calculators that compare the

two curves (e.g. 'TreatmentRatioCalc', which is the most commonly used for our QTL analysis).

Normalization by the punch-in technique

For best results, there should be as little variation as possible in the amount of cells printed to each

position in the plate to be phenotyped. We have developed a 'punch-in' technique to normalize

these printed masses, which is summarised in Figure 5.

We create special single-use 'punch-in' plates with double the usual thickness of agar medium (8mm

rather than 4mm), and allow the colonies to grow on it overnight. We use long-pin repads and turn

the ROTOR pressure and overshoot settings to maximum, in order to drive the pins into the agar and

push excess cell mass up the sides of the pin, leaving a more uniform amount on the tip, which is

then printed to the target (experimental) plate as normal. When working with 384 arrays and

repads, the ROTOR cannot deliver sufficient pressure to punch a 384 repad into normal agar.

Therefore, a soft agar punch-in plate, with 25% of the usual agar, must be prepared (and used within

24 hours). The yeast array must be printed to this plate, using low pressure and overshoot settings

so as to not punch in prematurely. Only 384 repads should be used to print to soft agar plates

because 96 repads will punch in to them immediately even at the lowest possible settings, which will

prevent proper punch-in normalization when they are later used as a source plate. Printed punch-in

plates should be incubated for ~16 hours before use, and then used as a source plate almost

immediately. Pinning settings for this source plate must be set to maximum pressure and overshoot,

and sufficient repeat pinnings to drive the pins all the way to the bottom of the soft agar precisely

once. The detailed PHENOS guide provides extra tips for mastering this technique, and it should be

practised before being used for actual experiments. We have observed differences in normalization

efficiency are most often due to different makes of agar or inconsistent media pH producing

variation in agar hardness, or due to punch-in plates not being thick enough (we recommend 8mm).

Cell count validation

We created multiple 96-density arrays on solid media of a single haploid strain with weak flocculation (EW01fc403), and induced a range of colony sizes by incubating the plates from one edge and for different lengths of time. Each plate was measured using PHENOS, with a 'snapshot' program to take four successive readings, and immediately afterwards colonies of different sizes were extracted along with a plug of agar by using a 1 ml pipette tip with 15-20 mm cut off. Colony and agar were deposited into 1 ml deflocculation buffer (50 mM KAc, 5 mM EDTA buffer, pH 4.5), and cells rinsed from the agar by thorough pipetting. The solution was then removed to a cuvette, leaving agar behind, and blanked readings taken in a Helios Gamma spectrophotometer (Thermo Scientific). Thereafter, 5 µl was transferred to a Neubauer improved cell counting chamber for counting at 40x magnification. Number of cells = $10000 \times C \times (25/S)$ where C=cells counted, S=subsquares considered. In some cases, freshly printed colonies were extracted into only 100 µl deflocculation buffer for cell counting without also being read in the spectrophotometer. Data was tabulated in Microsoft Excel, and microplate reader measurements were adjusted to remove absorbance due to agar, and to average readings from the snapshot program. For spectrophotometer readings with $OD_{600}>1$, 1:10 dilutions were used instead and results multiplied by 10 to account for known inaccuracies at high optical densities.

QTL validation

The 576 F1 *S. cerevisiae* strains previously described [26], comprising six crosses which we redesignated as 'AE', 'AS', 'AW', 'ES', 'EW' and 'SW', were re-arrayed as three 384-density arrays, each containing two of the crosses, with two biological replicates of each. All were incubated at 30°C for at least 16 hours, both with and without 500 µg/µl paraquat mixed into the YPD agar media (after autoclaving), and measured every 20 minutes. For each replicate, we determined the average OD_{600} (with agar absorbance subtracted) at ~16 hours as a fraction of the control growth (without paraquat) using the TreatmentRatio phenotype calculator in the PHENOS software. Biological

replicates were kept separate rather than averaged together in order to improve statistical power.

Taking these biological replicates as input, R/qtl was used to calculate genotype probabilities at

intervals of 1 centiMorgan, perform single-QTL analysis on the input data, run single-QTL analyses on

1000 permuted datasets, and estimate a LOD threshold at 5% significance level based on those

permutation analyses. Before being input to R/qtl, each permuted dataset was generated by

stratified permutation within tetrads, ensuring that no sample was permuted with its biological

replicate. QTL peaks were identified at loci with LOD values exceeding the threshold, and 1.5-LOD

support intervals were estimated by extension from the QTL peaks in either direction until the LOD

value fell by more than 1.5 LOD units below the peak LOD value.

Analysis of measurement variance

In order to assess the impact of material uniformity on variance in successive measurements, we

took multiple successive readings of a range of organic and non-organic substances with different

textures and levels of uniformity or granularity: yeast colonies now in stationary phase, artificial

spots made from woodglue and yeast extract mixed with biro ink, and layers of printer paper,

opaque plastic sheeting (cut from disposable weighing boats), and low-density 2mm foam packing

material. The standard deviation of multiple successive measurements was calculated for each

material type.

Results

Early dips in PHENOS growth curves

A common peculiarity of PHENOS curves is a slight dip that is often observed over the first hour of

incubation (clearly visible in Figure 4). We attribute this to microscopic bubbles which we have

observed in freshly printed spots (Figure 6), which behave like lenses and diffract the microplate

reader's light source until they have evaporated away. While curves can be normalized against each

other according to their minimum values, we observe in replicate plots that curves are usually well

aligned over most of the time course if the shrinkage is simply ignored.

Cell count validation

A range of differently sized colonies were measured in multiple ways, as described in the

corresponding methods section. The measurements and calculations are in Additional File 1, and are

displayed in Figure 7. To fit curves to each set of comparisons – spectrophotometer vs cell count,

and PHENOS vs log_{10} (cell count) – Excel's LINEST function was used to determine parameters that

minimize the sum of least squares, giving best curve fits relating A (average measurement with agar

subtracted) to P (predicted cell count). With parameters rounded, the best fit for PHENOS is

$P=77900e^{(1.76A)}$, while for liquid spectrophotometry the relationship is $P=8000(725A+7)$. For average

predicted cell counts below 6 million, corresponding to a microplate reader measurement (without

agar) of ~2.5, the predicted cell counts for the two methods show linear agreement with $R^2=0.77$.

Following the statistical approach of Bland & Altman for comparing two methods of measurement

[27], it can be seen in Figure 8 that at higher absorbance measurements both methods show

comparable degrees of inaccuracy, although the PHENOS formula tends to underestimate the size of

large colonies.

Variance in measurements at high absorbance values

We observe that higher absorbance readings, above 2.0 (typically reached in 10-20 hours on YPD at

30°C) exhibit greater variation between successive measurements, and often in a manner that is

synchronised across most colonies on a plate (as can be seen in Figure 3E), but with no discernible

periodicity. This variance is equally noticeable with more frequent measurements, with and without

incubation, and whether or not the reader is covered up to limit any light leakage or heat flow. We

tested the hypothesis that this might be due to microscopic differences in the starting position of

each pass of readings, and therefore would be more pronounced for colonies with more elaborate

morphologies or, in general, with the granularity of the material being measured. We took multiple

successive measurements of a range of organic and non-organic substances with different

granularity, but observed no difference between them (Figure 9, Additional File 2).

These results indicate that reading variation at high absorbance values is not attributable to

positioning error and granularity. We conclude that this is simply an unavoidable hardware

limitation, though easily overcome by averaging readings within a time window, as in our snapshot

protocols.

QTL validation

PHENOS was used to derive treatment ratios for the 576 F1 haploids growing with and without 500

µg/µl paraquat. Sample curves are shown in Figure 10.

QTL analysis with R/qtl yielded 61 QTLs overall, covering 22 distinct regions, including all four found

in the previous study [26], with comparable LOD scores, as illustrated in Figure 11. The original study

in liquid media, and this reproduction of it, both had low resolution due to large haplotype blocks

and low density genetic markers. Much greater resolution, down to individual SNPs, can be achieved

by repeating a process of crossing and sporulation to create advanced intercross lines with a high

level of genomic admixture, and generating much larger numbers of markers through whole genome

sequencing [16].

Discussion

Despite the reading variation at higher absorbance values described above (and equally problematic

in liquid spectrophotometry) PHENOS can provide accurate scores of colony size across a very wide

range of sizes, including very small ones that might not be picked up at all by imaging-based

approaches. Analyses of inter- and intra-plate variability (Figures 9 and 10 respectively) demonstrate

that this is a robust and highly reproducible assay.

We've already described and investigated a couple of quirks of PHENOS growth curves. These make

it difficult to reliably fit the curves to conventional logistic growth curve models [28]. Additionally,

we've observed some non-standard yeast growth patterns in response to various genotoxins. Strong

doses can produce curves that that show growth initially but then trend downwards as cell death

and dispersal thins the colony. Some compounds with lower stability, such as methyl

methanesulfonate, can produce curves which plateau and maybe fall before then entering a second

phase of growth after the survivors begin to proliferate. These are interesting responses and worthy

of study even though they cannot be parameterized as conventional logistic growth models. For the

purposes of QTL analysis, the most powerful phenotype measurement is the summary value that

shows the greatest variation between sensitive and resistant strains, and is least influenced by

experimental factors such as normalization efficiency, or least prone to miscomputation. This may

vary on a case by case basis, depending on the growth patterns concerned, so it is best to view the

full growth curves before deciding on the best summary values (and time windows) to use.

As shown in Figure 4, the PHENOS software allows for a range of summary values to be generated

from a PHENOS curve, including ones which are analogous to the efficiency (final growth), rate

(maximum slope), and lag (growth delay) parameters commonly derived from logistic growth curves

in liquid media, [26, 28]. The 'MaxSlopeCalc' value, analogous to growth rate, can be miscomputed

when there are multiple growth phases, so replicate plots should be consulted to confirm that

inflection points are being correctly identified. The 'LagCalc' value, analogous to growth lag, often

correlates with the initial absorbance of cells right after printing (the 'printed mass'). For this reason,

by default, PHENOS generates a scatterplot visualization comparing the two values for each curve so

that any correlation becomes evident. For the key purpose of discriminating between strains, the

growth level in a hand-picked time window ('AverageWithoutAgarCalc'), and its cross-experiment

derivative ('TreatmentRatioCalc') are usually the most direct and reliably computed values.

So-called 'edge effects' are a common concern of working with microbial arrays. The colonies around

the edge of an array might grow differently due to differences in heat-dispersal (in liquids) or

nutrient availability (on solid media). In the solid media arrays used for PHENOS, we only observe

significant edge effects after about 40 hours in arrays growing on normal YPD media. At this point

peripheral colonies continue growing for longer due to greater availability of nutrients at the edge of

the array, but we see no evidence of significant differences in earlier growth, or on plates where

growth is limited by treatments instead of nutrient availability.

In some circumstances PHENOS may not be the most suitable strategy for growth phenotyping. If a

treatment compound is expensive it will be inefficiently deployed by being dispersed in 40 ml of

solid media. Reducing the amount of media to 20ml per plate may mitigate this, but can introduce

artefacts due to the agar slightly contracting from the edges of the plate during incubation.

Alternatively, it might be possible to dispense hot agar into the wells of a microtitre plate but care

would need to be taken to avoid bubbles and unevenness (possibly with the use of additives such as

surfactants) that would compromise printing quality, and this would be harder in 384-well plates

than in 96.

Any treatment compound must be soluble of course, and sufficiently stable in solid media, and if it

affects the pH too strongly it will also make the agar too soft. If the treatment compound affects the

agar colour too much then it may increase the base agar absorbance readings so much that growth

curves appear to plateau prematurely, because the maximum possible overall reading, for FLUOstar

microplate readers at least, is 3.5. A typical *S. cerevisiae* colony at maximum size in a 384-array

growing on normal YPD agar will register an OD_{600} without agar of up to 2.5, so the agar absorbance

must be less than 1 when working with this organism. Complex phenotypes such as colony

morphology characteristics cannot easily be measured by PHENOS in its current state. Furthermore,

we expect that growth on solid media will not always reflect growth under liquid conditions. The

latter may be more relevant to applications in brewing and wine-making, although solid media

growth is likely to be much more similar to how yeast grow in the wild.

We have found 384 arrays strike the best balance between scale and reliability but the ROTOR is

capable of printing 1536 arrays and, with a minor vendor adjustment, the FLUOstar Omega is

capable of reading them. 1536 arrays are challenging to duplicate, especially if they contain diverse

growth phenotypes, as slow-growing colonies are easily lost, often due to larger surrounding

colonies keeping a repad pin from making full contact. Only short-pin repads are available for 1536

arrays, and their pins have a very small cross section which necessitates a much finer alignment

between printing and reading coordinates than the machines can usually achieve. This problem may

be overcome by using 4 long-pin 384 repads, each normalized from a different punch-in plate, but

this reduces the convenience and speed of the method. Furthermore, the greater proximity of

colonies in 1536 arrays substantially reduces growth speed and peak colony size, and brings forward

the time point at which edge effects become prominent. Also, mating pheromone inhibition or

similar effects may complicate growth in mixed arrays.

Conclusions

In successfully normalized experiments, PHENOS produces proxy growth curves with a high degree

of reproducibility between replicates (Figures 3F, 10), and with greater ease and automation than

conventional methods. We've reported results that use *S. cerevisiae*, but within our lab the exact

same approach has also been used to measure the growth of other yeast strains such as *S.*

eubayanus and *S. uvarum.* Nor need PHENOS be limited to working with yeast. We have briefly

experimented with using PHENOS to measure *E. coli* growth (strain DH5a) on Luria Broth agar, and

obtained growth curves with a maximum change in OD_{600} of 0.62 after 16 hours (Additional File 4).

While less optically absorbent than yeast, this should still be perfectly sufficient to discriminate

growth phenotypes. Provided that an organism can grow on a solid medium in the first place,

PHENOS may prove a viable approach to growth phenotyping.

We have also briefly experimented with measuring whole-colony fluorescence in tandem with

optical absorbance and have encouraging preliminary results, although extra software will be

needed to handle the data files produced by this procedure.

For labs in possession of a colony arraying robot, colony thickness as measured by an incubating

microplate reader is a very convenient and sensitive alternative to more expensive specialist

solutions for quantifying growth. It sidesteps problems with imaging-based solutions such as poor

contrast or bubbles or specks which can confuse image analysis, as well as the general difficulty of

setting up a high-quality imaging and image analysis pipeline in the first place. It avoids problems

with liquid phenotyping such as format conversion, unequal heat distribution, evaporation, and the

technical challenge of fine-tuning agitation levels to prevent flocculation without introducing

bubbles or cross-well contamination. Furthermore, there are clear safety benefits to having toxic

treatment compounds contained within solid agar, and the same will be true if working with more

dangerous organisms than yeast

List of abbreviations

426   OD: optical density; PHENOS: phenotyping on solid media; QTL: quantitative trait loci; SGA: synthetic

427   genetic array; YPD: yeast extract peptone dextrose growth media.

Figure legends

Figure 1 Overview of the PHENOS experimental workflow. We distinguish between three types of

plates: stock plates for array maintenance, punch-in plates for normalization, and experimental

plates for measurement.

Figure 2 Overview of the PHENOS analysis workflow. The PHENOS software combines files and

visualizes and compares experiments.

Figure 3 Examples of visualizations produced by the PHENOS software. The colour bars and some

of the labels have been enlarged for this publication. **A)** Agar thickness of an unprinted plate,

showing a low-quality plate with a bulge due to inadequately mixed media. For a good quality plate,

all squares would be the same colour (the colour scale shows absorbance at each location in the

array—darker colours signifying thicker agar—and is comparable between different visualizations).

There should be no pattern to the size of squares either: the largest and smallest squares are the

highest and lowest readings on that plate respectively. Squares are also labelled with the array

position, although those labels are not legible in these small reproductions. **B)** Printing quality, based

on initial readings immediately after printing, with agar subtracted. As with A, the colour scale is

absolute and comparable between different experiments, but the spot scaling shows where the

largest and smallest printed masses on the plate are located. Each spot is labelled with the strain

name, which is legible on full size images. **C)** Final measurements of a plate as measured by the

microplate reader (radius being proportional to OD_{600} with agar absorbance subtracted), next to **D)**

an actual image of the same plate. In this case, there are four replicates of each strain. **E)** Control

growth curves for all 384 colonies on a YPD plate, coloured by the initial printed mass as described

for B. In a successful treatment experiment such as is shown in C & D, there would be much greater

diversity in these curves. **F)** Replicate plots, showing growth curves of all four replicates of a single

strain for which growth has been equally suppressed (compare to E) by hydroxyurea. Dots on these

curves indicate where the PHENOS software has calculated the inflection point of each curve to lie,

i.e. the point of maximum growth. Curves are also labelled with the array position of the colony.

Figure 4 'Phenotype calculators' within the PHENOS software. Each generates a different summary

value. Some, indicated with asterisks, e.g. AverageWithoutAgarCalc, require a particular time

window to be specified by the user: for a given curve, all timepoints within that window are then

averaged, to account for the variance usually observed at higher absorbance readings. Distinctive

properties of PHENOS curves, such as initial reductions in absorbance, are discussed within the text

and Figure 6.

Figure 5 Punch-in technique for normalizing printed cell masses.

Figure 6 Bubbles in printed cell patches. These evaporate over the first hour after printing, leading

to initial dips in PHENOS growth curves.

Figure 7 Accuracy of PHENOS colony measurements. A) Microplate reader measurements plotted

against cell counts. **B)** Liquid spectrophotometry measurements plotted against cell counts; for initial

readings >1, 1:10 dilutions were read and their results multiplied by 10 to empirically adjust for non-

linearity at high absorbances. Together, these measurements were used to empirically determine

formulae for predicting cell counts from measured values.

Figure 8 Comparison of predicted cell counts by two methods: microplate absorbance and

conventional liquid spectrophotometry, both at OD_{600}.

Figure 9 Variation in multiple successive readings of a variety of materials. This variation is a

function of overall average absorbance, but barely affected by different degrees of granularity.

Figure 10 Sample curves from two strains showing extremely different growth responses. Growth

on solid media containing 500 µg/µl paraquat (red curves) compared to normal YPD (blue curves).

Curves are labelled with experiment IDs and plate positions, and dots indicate PHENOS-calculated

inflection points.

Figure 11 QTL validation. *S. cerevisiae* QTLs for paraquat resistance, determined using PHENOS,

compared to QTLs found in the same set of strains grown in liquid media. In both experiments, 576

strains were phenotyped, each an F1 progeny from one of six different crosses of four geographically

diverse parental strains (A= American, E= European, S= Saki, W=West African). Different crosses

produce different QTLs, though one QTL on chromosome XII is common to all the crosses. The

results show that PHENOS reproduces the liquid media QTLs and is more sensitive.

Additional Files

Additional File 1 Validation by cell counting (xlsx). Raw data and calculations.

Additional File 2 Testing with multiple materials (xlsx). Raw data and calculations.

Additional File 3A Validation by paraquat QTLs (xlsx). Full list of QTL intervals and their details.

Additional File 3B Validation by paraquat QTLs measurements (zip). Initial DAT and csv files.

Additional File 4 E coli results (xlsx). Raw data and calculations.

Declarations

Ethics approval and consent to participate

Not applicable.

Consent for publication

Not applicable.

Availability of data and materials

The datasets supporting the conclusions of this article are included within the article (and its additional files). The PHENOS software and test files are available from https://github.com/gact/phenos/, under an MIT license. This software is written for Windows, and requires Python 2.7, along with the modules "scipy", "numpy", "tables", "matplotlib", "xlrd", "brewer2mpl" and their dependencies. The yeast strains used in QTL validation are available on request.

Competing interests

The authors declare that they have no competing interests.

Funding

The work was funded by the University of Leicester under a Burgess Centre grant, with additional funding for consumables and genotyping provided by BBSRC grant BB/L022508/1 to EJL. DG was funded by the University of Leicester under a Pritchard PhD studentship award.

Author's contributions

DBHB invented the technique, wrote the PHENOS software and drafted the manuscript. ND tested early versions and provided feedback. MA performed the paraquat phenotyping. TW provided help with R/qtl and Github. DG tested the software in a variety of contexts and drove improvements. EJL and SSF jointly supported the project and helped draft the manuscript. All authors read and approved the final manuscript.

Acknowledgements

We wish to thank Maria Viskaduraki for statistical advice, and Tuğba Şan and Noor Bafana for testing aspects of the pipeline in its early stages.

References

References

1. Winzeler EA, Shoemaker DD, Astromoff A, Liang H, Anderson K, Andre B, et al. Functional

characterization of the S. cerevisiae genome by gene deletion and parallel analysis. Science.

1999;285:901–6.

2. Tong AHY, Evangelista M, Parsons AB, Xu H, Bader GD, Pagé N, et al. Systematic Genetic Analysis

with Ordered Arrays of Yeast Deletion Mutants. Science (80-.). 2001;294.

3. Scherens B, Goffeau A. The uses of genome-wide yeast mutant collections. Genome Biol. BioMed

Central; 2004;5:229.

4. Roguev A, Wiren M, Weissman JS, Krogan NJ. High-throughput genetic interaction mapping in the

fission yeast Schizosaccharomyces pombe. Nat. Methods. 2007;4:861–6.

5. Typas A, Nichols RJ, Siegele DA, Shales M, Collins SR, Lim B, et al. High-throughput, quantitative

analyses of genetic interactions in E. coli. Nat. Methods. 2008;5:781–7.

6. Holz C, Lueking A, Bovekamp L, Gutjahr C, Bolotina N, Lehrach H, et al. A human cDNA expression

library in yeast enriched for open reading frames. Genome Res. Cold Spring Harbor Laboratory Press;

2001;11:1730–5.

7. Harrington JJ, Sherf B, Rundlett S, Jackson PD, Perry R, Cain S, et al. Creation of genome-wide

protein expression libraries using random activation of gene expression. Nat. Biotechnol. Nature

Publishing Group; 2001;19:440–5.

8. Meersman E, Steensels J, Struyf N, Paulus T, Saels V, Mathawan M, et al. Tuning Chocolate Flavor

through Development of Thermotolerant Saccharomyces cerevisiae Starter Cultures with Increased

Acetate Ester Production. Appl. Environ. Microbiol. American Society for Microbiology; 2015;82:732–

46.

9. Malecki M, Bähler J. Identifying genes required for respiratory growth of fission yeast. Wellcome

open Res. Europe PMC Funders; 2016;1:12.

10. Freel KC, Friedrich A, Sarilar V, Devillers H, Neuvéglise C, Schacherer J. Whole-Genome

Sequencing and Intraspecific Analysis of the Yeast Species *Lachancea quebecensis*. Genome Biol.

Evol. Oxford University Press; 2016;8:733–41.

11. Zhang R, Patena W, Armbruster U, Gang SS, Blum SR, Jonikas MC. High-Throughput Genotyping

of Green Algal Mutants Reveals Random Distribution of Mutagenic Insertion Sites and

Endonucleolytic Cleavage of Transforming DNA. Plant Cell. 2014;26:1398–409.

12. Côté J-P, French S, Gehrke SS, MacNair CR, Mangat CS, Bharat A, et al. The Genome-Wide

Interaction Network of Nutrient Stress Genes in Escherichia coli. MBio. American Society for

Microbiology; 2016;7:e01714-16.

13. Hall BG, Acar H, Nandipati A, Barlow M. Growth Rates Made Easy. Mol. Biol. Evol. 2014;31:232–8.

14. Jung PP, Christian N, Kay DP, Skupin A, Linster CL. Protocols and Programs for High-Throughput

Growth and Aging Phenotyping in Yeast. Schacherer J, editor. PLoS One. Public Library of Science;

2015;10:e0119807.

15. Warringer J, Ericson E, Fernandez L, Nerman O, Blomberg A. High-resolution yeast phenomics

resolves different physiological features in the saline response. Proc. Natl. Acad. Sci. U. S. A.

2003;100:15724–9.

16. Cubillos FA, Parts L, Salinas F, Bergström A, Scovacricchi E, Zia A, et al. High Resolution Mapping

of Complex Traits with a Four-Parent Advanced Intercross Yeast Population. Genetics. 2013;

17. Dittmar JC, Reid RJD, Rothstein R. ScreenMill: A freely available software suite for growth

measurement , analysis and visualization of high-. BMC Bioinformatics. 2010;11:353.

18. Lawless C, Wilkinson DJ, Young A, Addinall SG, Lydall DA. Colonyzer: automated quantification of

micro-organism growth characteristics on solid agar. BMC Bioinformatics. 2010;11:287.

19. Shah NA, Laws RJ, Wardman B, Zhao LP, Hartman JL. Accurate, precise modeling of cell

proliferation kinetics from time-lapse imaging and automated image analysis of agar yeast culture

arrays. BMC Syst. Biol. 2007;1:3.

20. Takeuchi R, Tamura T, Nakayashiki T, Tanaka Y, Muto A, Wanner BL, et al. Colony-live —a high-

throughput method for measuring microbial colony growth kinetics— reveals diverse growth effects

of gene knockouts in Escherichia coli. BMC Microbiol. 2014;14:171.

21. Wilkening S, Lin G, Fritsch ES, Tekkedil MM, Anders S, Kuehn R, et al. An Evaluation of High-

Throughput Approaches to QTL Mapping in Saccharomyces cerevisiae. Genetics. 2013;

22. Broman KW, Sen S. A Guide to QTL Mapping with R/qtl. New York, NY: Springer New York; 2009.

23. Liti G, Carter DM, Moses AM, Warringer J, Parts L, James SA, et al. Population genomics of

domestic and wild yeasts. Nature. 2009;458:337–41.

24. Cubillos FA, Louis EJ, Liti G. Generation of a large set of genetically tractable haploid and diploid

Saccharomyces strains. FEMS Yeast Res. 2009;9:1217–25.

25. Wittwer CT, Reed GH, Gundry CN, Vandersteen JG, Pryor RJ. High-Resolution Genotyping by

Amplicon Melting Analysis Using LCGreen. Clin. Chem. 2003;49.

26. Cubillos FA, Billi E, Zörgö E, Parts L, Fargier P, Omholt S, et al. Assessing the complex architecture

of polygenic traits in diverged yeast populations. Mol. Ecol. 2011;20:1401–13.

27. Martin Bland J, Altman D. Statistical methods for assessing agreement between two methods of

clinical measurement. Lancet. 1986;327:307–10.

28. Richards OW. The Growth of the Yeast Saccharomyces cerevisiae. I. The Growth Curve, its

Mathematical Analysis, and the Effect of Temperature on the Yeast Growth. Ann. Bot. Oxford

University Press; 42:271–83.

SINGER ROTOR HDA

MICROPLATE READER

Keep maintained

Normal pressure pin

1) STOCK PLATE

Low pressure print

2) PUNCH-IN PLATE*

*8mm soft agar for long pin 384 repads

Incubate ~16hrs

*High pressure **punch-in***

2) PUNCH-IN PLATE*

Dispose

Normal pressure print

3) EXPERIMENTAL PLATE

3) EXPERIMENTAL PLATE

Check in PHENOS results that plate is flat

FILE 'a'

Incubate while measuring...

3) EXPERIMENTAL PLATE

FILE 'b'

Additional measurements as needed

FILE 'c' etc

R u n   P H E N O S

Normal pressure and overshoot

Low pressure and overshoot

Well grown colony after ~16hrs

Maximum pressure, maximum overshoot, repeat pinnings until pin goes to bottom

Excess pushed up pin

Uniform amount left on tip

Normal pressure and overshoot

Normalized printed mass

Stock plate

Punch-in plate (double thickness, soft agar needed for 384 arrays)

Experimental plate