Adaption of a Conventional ELISA to a 96-well ELISA-Array for Measuring the Antibody Responses to Influenza virus proteins, viruses and vaccines

Eric Waltari[1*], Esteban Carabajal[1], Mrinmoy Sanyal[2], Natalia Friedland[2], Krista M. M[c]Cutcheon[1]

[1]Chan Zuckerberg Biohub, San Francisco, CA, USA
[2]Stanford ChEM-H and Department of Biochemistry, Stanford University School of Medicine, Stanford, CA, USA

* Correspondence:
Eric Waltari
Eric.Waltari@czbiohub.org

Keywords: ELISA-Array, infectious disease, protein array, titer

Abstract

We describe an adaptation of conventional ELISA methods to an ELISA-Array format using non-contact Piezo printing of up to 30 spots of purified recombinant viral fusion proteins, vaccine and virus on 96 well high-protein binding plates. Antigens were printed in 1 nanoliter volumes of protein stabilizing buffer using as little as 0.25 nanograms of protein, 2000-fold less than conventional ELISA. The performance of the ELISA-Array was demonstrated by serially diluting n=8 human post-flu vaccination plasma samples starting at a 1/1000 dilution and measuring binding to the array of Influenza antigens. Plasma polyclonal antibody levels were detected using a cocktail of biotinylated anti-human kappa and lambda light chain antibodies, followed by a Streptavidin-horseradish peroxidase conjugate and the dose-dependent signal was developed with a precipitable TMB substrate. Intra- and inter-assay precision of absorbance units among the eight donor samples showed mean CVs of 4.8% and 10.8%, respectively. The plasma could be differentiated by donor and antigen with titer sensitivities ranging from 1×10^3 to 4×10^6, IC_{50} values from 1×10^4 to 9×10^6, with monoclonal antibody sensitivities in the ng/mL range. Equivalent sensitivities of ELISA versus ELISA-Array, compared using plasma and an H1N1 HA trimer, were achieved on the ELISA-Array printed at 0.25ng per 200um spot and 1000ng per ELISA 96-well. Vacuum-sealed array plates were shown to be stable when stored for at least 2 days at ambient temperature and up to 1 month at 4-8°C. By the use of any set of

printed antigens and analyte matrices the methods of this multiplexed ELISA-Array format can

be broadly applied in translational research.

INTRODUCTION

The activity of humoral antibodies provide the best correlation to long-term immune memory

and protection (Antia et al. 2018). During the first two weeks of exposure to a pathogen, the

majority of antibodies found in the serum derive from plasmablasts, either rapidly re-activated

from memory B cell pools or expanded from newly stimulated, somatically hypermutated and

differentiated B cells upon contact with antigen in lymph tissue (De Silva and Klein 2015).

During recovery, some plasmablasts will home to the bone marrow where they terminally

differentiate into long-lived plasma cells stably secreting antibodies that circulate in serum for

many years (Abbas, Lichtman, and Pillai 2014; Yoshida et al. 2010). Cellular and molecular

events leading to antigen-specific B cell expansion, differentiation, homing and fate are complex

and not predictable in outcome. In lieu, serum can be used to measure the binding kinetics,

magnitude, specificity and cross-reactivity of the secreted antibodies in response to infection or

vaccination. Serological testing can help evaluate an individual's susceptibility, exposure or

protection from past, existing and future pathogens. It is also possible to make positive or

negative correlations of binding characteristics to serum neutralization activity or antibody

enhanced disease (Katzelnick et al. 2017). Analytical methods characterizing antibodies ideally

have the ability to measure the robustness, specificity and genetic breadth of activity to

pathogens. Humoral responses are typically quantified by titer in naïve, acute, convalescent and

recovery sera in the context of natural infection or pre- and post- vaccination and correlated to *in*

vitro activity assays and clinical signs of immune protection (e.g. Antia et al., 2018; Lowell et

al., 2017; Madore et al., 2010).

The enzyme-linked-immunosorbent-assay (ELISA) first described by Engvall and

Perlmann (1972), is commonly used to measure specific antibody-antigen binding. Variants and

derivatives of the ELISA have become assay workhorses of immunology laboratories and a host

of compatible reagents, consumables, plate washers, multichannel pipettes, robotic liquid

handlers, and assay formats have been developed and are available from multiple vendors. A

conventional antigen ELISA single plex format passively coats antigens on a 96-well high

capacity protein binding surface (e.g. Nunc Maxisorp™, ThermoFisher Scientific, Waltham,

MA) and indirectly titers primary antibody binding by secondary binding of polyclonal

antibodies conjugated to horseradish peroxidase (HRP), which turns over the colorimetric

3,3',5,5'-tetramethylbenzidine (TMB) substrate for assay readout. Secondary antibodies are

typically directed against a constant region of the heavy or light chain of the primary antibody,

such as polyclonal anti-Fc directed to IgG, IgM, IgA or IgE, anti-kappa or anti-lambda light

chains. A common variation to boost sensitivity includes using a biotinylated secondary antibody

with a Streptavidin-horseradish peroxidase (HRP) conjugate. Although fluorescent reporters

have the advantage of allowing for multiplexed detection using different dyes, the use of HRP

has been shown to be more sensitive because the enzymatic turnover of colorimetric or

chemiluminescent substrates amplifies the signal (Gogalic et al. 2018).

The principles of ELISA have been adapted using advances in the protein array field to

increase the throughput, efficiency and scope of data in immunoassays (reviewed in Kingsmore

2006). Printing proteins can be carried out by passive adsorption without requiring modification

or chemical coupling to nanoparticles or other surfaces. This advantage and advancements in

nozzle technology allow for flexibility and precision in spotting picolitre volumes of purified,

crude, or complex proteinaceous substrates (Barbulovic-Nad et al. 2006). Furthermore, a

superior level of sensitivity can be achieved in miniaturized ligand-binding assays, as shown by

Ekins' ambient analyte assay theory (Ekins 1989). Obtaining higher sensitivity in a system that

uses smaller amounts of capture molecules and smaller amounts of sample can be explained by

two main features. First, the binding reaction occurs at a high target concentration; and second,

the capture-molecule–target complex is found only in the small area of the spot, resulting in a

high local signal (Templin et al. 2002). The most published format for protein array printing in

the infectious disease research setting is onto glass slides functionalized with nitrocellulose,

perhaps because both of the technical ease and that high density arrays are made possible by

printing onto this high protein binding flat surface (Davies et al. 2005; Desbien et al. 2013;

Koopmans et al. 2012; Nakajima et al. 2018; Price et al. 2013; te Beest et al. 2014). An

alternative format amenable for use in research labs is an ELISA-based microarray printed

directly onto the bottom of a 96-well plate (Mendoza et al. 1999). This method has been

validated against single plex assays (Liew et al. 2007) and has been adopted for biomarker

discovery in research labs (W. Huang et al. 2018; Y. Huang and Zhu 2017) and commercial

assays (e.g. PBL Assay Science, Piscataway, NJ; Quantarix, Billerica, MA; BioVendor LLC,

Asheville, NC; RayBiotech Inc., Peachtree Corners, GA). However, to date the 96-well format

has been infrequently applied to infectious disease antigens (Kang et al. 2012; Wang et al. 2015),

warranting more published examples and methods of applied research in this area.

For our multiplexed infectious disease research, we decided to capitalize on the

resources, familiarity and knowledge readily available for the conventional 96-well plate ELISA

and adapt the workflow directly to an in-house 96-well plate ELISA-Array format. The only

changes in the assay format were at the first and final steps. Using Maxisorp™ 96-well plates in

the first step, in lieu of coating a single antigen per well, we printed 1 nanoliter volumes of 8

viral antigens, in triplicate, per well. In the final step, a precipitating form of TMB substrate and

an array plate reader were needed for the ELISA-Array instead of the soluble TMB form and

general lab plate reader. The remainder of the workflow and reagents were identical in both

formats. The development and testing of the ELISA-Array was carried out using healthy human

donor plasma sampled post-vaccination (from the 2018 FluLaval vaccine), and assayed on

printed vaccine, recombinant hemagglutinin trimers, and purified FluB virus. Here we provide

data characterizing the ELISA-Array methods, advantages, precision, robustness, sensitivity,

stability and utility in infectious disease research.

RESULTS

Initial optimization of printing conditions

Although many operating conditions for printing followed the standard recommendations of the

manufacturer of the sciFLEXARRAYER S12 instrument, several specific parameters were

optimized for this ELISA-Array application. We tested variations in printing protein

concentration, drop volume and formulation buffer using goat anti-human Fc polyclonal

antibody (Jackson ImmunoResearch, West Grove, PA) as a probe with commercially available

human reference serum (Bethyl Laboratories, Montgomery, TX) as an analyte. The probe was

varied by diluting a PBS stock in a 1:1 volume of each of three sciSPOT protein formulation

buffers D1, D11 and D12 (Scienion AG, Berlin, Germany). Probe was dispensed in 1, 2 or 4

drops from a 384-well source plate at 25, 100 and 400 ug/mL final. PBS in formulation buffers

without the anti-human Fc protein was used for a background control. The probes of the printed

arrays bound to the Fc region of IgG within the human plasma, then are detected with HRP

conjugate antibodies specific to the kappa light chain of the IgG antibody in a traditional

sandwich ELISA format.

The signal intensity increased with increasing protein printed, and 400 ug/mL provided

the highest signal. The spot size increased with drop number, but the sensitivity was similar

between 2 and 4 drops. The protein stabilizing D12 buffer offered the highest sensitivity among

formulation buffers to approximately 4 ng/ml concentrations of IgG detected from human sera.

These data are shown in Supplementary Figure S1. The final printing parameters used in this

report for Influenza antigens are described in the methods section. Twelve 96-well plates were

printed in one batch with the Influenza antigens listed in Table 1 and using the array pattern

illustrated in Figure 1.

Assay miniaturization gain of sensitivity in ELISA-Array

According to the ambient assay theory (Ekins 1989), miniaturizing the ELISA to an array print

of 0.25ng of protein in a 200um spot (with a surface area of $15.6mm^2$, or $0.02ng/mm^2$) should

yield higher sensitivity than coating in the same proportion over an entire 96-well (with a surface

area of $320mm^2$). We tested this by comparing the signal sensitivity obtained using human

immune reference plasma binding purified H1N1 HA trimer, either printed in 0.25ng spots in

triplicate or coated in a 96-well at 1000, 100, 10 or 1 ng in duplicate. Following the same assay

methods with the exception of the final TMB substrate (soluble for the ELISA and precipitating

for the ELISA-Array), equivalent IC_{50} values were obtained only when the 96-well was coated

with 2000-fold more total protein, or >150-fold more/mm^2 (Figure 2).

Data analysis

After calculating median intensity in absorbance units (AU) of each triplicate set of antigen

spots, we fit standard 4-parameter logistic (4P) curves of intensity against plasma or mAb

concentration with PRISM (GraphPad, San Diego, CA). In each plate we tested three negative

controls to calculate the lower limit of detection, and on average the LOD value was less than 5

AU. Because the variance in readings at values less than 10 AU was high (data not shown), we

set a lower limit of detection (LLOQ) at 10 AU. From the 4P curves fit to each sample we

calculated both the titer at which the curve passed the LLOQ and the IC_{50} values. Across our

assays we observed that the upper intensity range was never greater than 180 AU and thus set

limits of 0-200 AU in 4P curve fitting. Because at high concentrations the hook effect can lead to

reduced intensity readings (Tighe et al. 2015), we disregard any decreased values at high analyte

concentrations.

ELISA-Array Sensitivity and Specificity

In each ELISA-Array assay, polyclonal human immune reference plasma and monoclonal

antibodies of known binding activity were used to control for assay performance and determine

sensitivity. The dose-dependent binding of each of these controls over three independent assays

(Figure 3), IC_{50} values and LLOQ are reported in Table 2. mAb A is known to be a neutralizing

antibody recognizing a conformationally dependent epitope on the stalk region of HA trimers

(Kallewaard et al. 2016) and was detected against the array of Influenza antigens from 10-

120ng/mL, well below the quantitative ug/mL range of relevant protective antibody levels *in vivo*

(Crum☐Cianflone et al. 2012; Semenova et al. 2004). Reference plasma showed Influenza

antigen binding IC_{50} values of 1.4×10^5 to 9×10^6 and titers of 6.4×10^4 to 1×10^6 (Figure 3 and

Table 2). These values reflect a polyclonal mixture of antibodies of any isotype since detection

was not limited to IgG (a cocktail of anti-kappa and anti-lambda light chain secondary antibodies

was used). The correlation of binding titers to protection varies by disease and for Influenza has

not been shown to be predictive (Madore et al. 2010). However, since binding is a pre-requisite

for neutralization activity, plasma titer can demonstrate the variation, breadth and magnitude of

viral antigen specificity between individuals.

The specificity of the assay was tested by measuring cross-reactivity at 200nM

concentrations of two irrelevant mAbs to any of the printed proteins: one directed to the

envelope protein of Dengue virus and the other to the RSV fusion protein. No signal was

observed in the assay with these mAbs. Specificity was also tested by printing a GFP-foldon-

Avitag-6His trimer in each well as a negative control protein, at the same concentration as the

Influenza A HA trimers. This control showed no binding to donor plasma or to control antibodies

mAb A and mAb B.

The lack of binding of control mAb A to the HA trimer of A/Shanghai/02/2013 H7N9

was not expected based on publications of this mAb binding to other H7 strains of HA, albeit at

lower affinities than other HA subtypes (Kallewaard et al. 2016). Reference plasma and other

donor plasma did bind the H7 trimer (Figures 3 and 4). A repeat test print of the H7 HA trimer at

0.5, 0.375 and 0.25ng/spot did not change the binding results, nor did testing on a regular ELISA

format (data not shown). Further optimization of this antigen, and comparisons to other strains

are needed in order to draw conclusions on cross-reactive antibody binding to H7.

ELISA-Array assay performance

The ELISA-Array assay was qualified using a selected in-house human reference plasma and

eight individual human plasma samples from day 28 post-vaccination with the 2018 FluLaval

quadrivalent vaccine (GlaxoSmithKline, Research Triangle Park, NC). All array plates used for

performance testing were from one print batch, stored at 4ºC. To increase accuracy, we avoided

making large dilutions by preparing stock solutions of 10x reference plasma, 10x control mAbs,

100x secondary biotinylated antibody mixture, and 100x streptavidin conjugate in assay diluent.

These were aliquoted and stored at -80ºC. Each plasma donor was also aliquoted undiluted and

stored at -80ºC. Although not done in these assays, it would be optimal in the future to briefly

spin down donor plasma before assaying to clear the sample of lipid and other aggregates.

Aliquots were freshly thawed for each assay, and the same lot of assay diluent and TMB

substrate were also used throughout all assays. The final concentrations of assay materials are

described in the methods section. Intra-assay precision was determined by running n=3 plate

assays on the first day after printing the arrays. Inter-assay precision was determined by running

an additional two plates one and two weeks later. Precision was calculated by the variance

between plates of titer and IC_{50} values for reference plasma and each of the eight donors for all

array antigens. Intra- and Inter-assay precision data is shown for the reference plasma and two

donors in Figure 4, and Tables 3 and 4, respectively, and for all eight plasma samples in

Supplementary Tables S1 and S2. The precision of absorbance units among the reference plasma

and two donors showed mean CV of 4.8% intra-assay and 10.8% inter-assay, and 6.0% intra-

assay and 12.5% inter-assay among the eight donor samples. There were a few examples of high

variance inter-assay, in samples of low dilutions. This may be due to weak binding or

interference from the serum matrix. The plasma titers could be differentiated by donor and

antigen with sensitivities ranging from 1×10^3 to 4×10^6 and IC_{50} values from 1×10^4 to 9×10^6

(Figure 4, Table 4, and Table S2). For example, we measured a robust titer in the reference

plasma donor to all array antigens (Figure 4). In contrast, robust titers in donor 1 plasma were

216 measured only to the vaccine itself, to Influenza B viruses and to the individual antigens of

217 H1N1 and H3N2 HA trimers matching the strains used in the vaccine (Table 1 and Figure 4).

218 There was only weak binding to HA trimers not in the vaccine (i.e. H2, H5, H7), indicating

219 insufficient cross-reactive antibodies were elicited in donor 1. A plot of IC_{50} values in Figure 5

220 for three donors shows the overall tight standard deviations between three assays performed over

221 three weeks, and visually quantifies differences between antibody binding for each of the array

222 antigens and donors. We cannot differentiate pre-existing antibody immunity from vaccine

223 responses in these samples, but the quantitative nature of the data would allow for this to be done

224 using titer and IC_{50} comparisons with pre-vaccine plasma, not included in this study. Three

225 rounds of freeze thaws of the reference plasma from -80ºC showed no change in IC_{50} values or

226 titers (data not shown). One plate assay was also run by a second operator to evaluate the

227 robustness of the method, which was determined to be equivalent to intra-assay precision (Plate

228 4 in Figure 6 and Table S3).

230 **Stability testing of printed plates**

231 Printed array plates were covered with a foil plate seal and vacuum-sealed immediately after the

232 overnight curing step and stored at 4ºC. They were found to be stable stored in this manner for

233 up to one month (Figure 6). At 8 weeks post-print, plates stored at 4ºC showed about a 2-fold

234 drop in IC_{50} and the titer shifted to one higher dilution in the ¼ titration series (i.e. a change from

235 $1x10^6$ to $2.6 x10^5$). Significant losses in activity were also measured for vacuum-sealed plates

236 stored for 1 week at either ambient temperature or 37 ºC (Figure 6). All of the stability assay data

237 including variability for each antigen and plasma sample are provided in Supplementary Tables

238 S4 and S5. Using a different print lot, we tested the plates for 2-days at ambient temperature and

239 found them to be stable (data not shown).

241 **TABLES**

242 Table 1. Antigens included in the protein microarrays.

Influenza Strain	Subtype
A/Michigan/45/2015[*]	H1N1
A/Japan/305+/1957	H2N2
A/Singapore/INFIMH-16-0019/2016[*]	H3N2
A/Viet Nam/1194/2004	H5N1
A/Shanghai/02/2013	H7N9
B/Phuket/3073/2013-like virus[*] (B/Yamagata/16/88 lineage)	Influenza B
B/Colorado/06/2017-like virus[*] (B/Victoria/2/87 lineage)	Influenza B
Flulaval Vaccine 2018-2019	[*]H1N1, [*]H3N2, [*]Influenza B -1, [*]Influenza B -2

243 [*]components of the WHO recommended seasonal flu vaccine for 2018-19

245 Table 2. Sensitivity of reference mAbs and reference plasma pAb on ELISA-array antigens.

Antigen	IC$_{50}$ (ng/mL)	LLOQ (ng/mL)	Plasma titer IC$_{50}$	Plasma LLOQ
mAb A			Reference plasma	
Vaccine	1900	470	1.4e5	2.6e5
H1N1	110	10	8.5e6	1.0e6
H2N2	380	120	1.3e5	2.6e5
H3N2	170	30	4.0e5	2.6e5
H5N1	40	10	3.0e5	2.6e5
H7N9	>30	>30	2.9e5	6.4e4
mAb B				
Vaccine	7.3	2		
Influenza B1	238	125	9.0e6	1.0e6
Influenza B2	25	2	4.6e6	1.0e6

248   Table 3. Intra-assay precision from three donor plasma samples on ELISA-array antigens.[1,2,3]

	Reference				Donor 1				Donor 6			
	1	2	3	%CV	1	2	3	%CV	1	2	3	%CV
Flulaval seasonal vaccine 2018-2019												
1e3	99.5	114.4	118.6	9.1	133.0	136.6	135.8	1.4	150.4	144.4	146.3	2.1
4e3	152.5	146.5	144.6	2.8	100.6	104.6	112.7	5.8	114.4	108.7	101.9	5.8
1.6e4	135.4	137.2	132.9	1.6	106.3	108.7	100.6	3.9	58.1	52.7	49.8	7.9
6.4e4	89.0	87.7	90.0	1.3	53.4	53.5	49.5	4.4	**17.7**	**14.4**	**15.5**	10.7
2.6e5	**36.4**	**36.8**	**32.8**	6.3	**15.5**	**15.1**	**14.1**	4.8	3.5	2.6	2.6	OOR
1e6	9.9	9.6	9.6	OOR	4.2	4.4	3.0	OOR	2.2	0.4	0.2	OOR
4.1e6	1.6	0.0	1.9	OOR	0.5	0.2	1.1	OOR	0.0	0.1	0.2	OOR
1.6e7	0.4	0.4	1.0	OOR	0.1	0.1	0.2	OOR	0.5	0.0	0.0	OOR
H1N1 HA trimer A/Michigan/45/2015												
1e3	100.2	103.7	118.4	9.0	149.3	145.3	153.3	2.7	138.7	149.6	140.3	4.1
4e3	148.0	142.6	150.9	2.9	154.7	153.5	158.8	1.8	125.9	123.5	121.0	2.0
1.6e4	143.5	132.3	139.4	4.1	118.1	119.7	119.2	0.7	70.7	67.9	67.5	2.5
6.4e4	100.1	101.1	103.7	1.8	58.5	61.8	62.4	3.5	**21.4**	**21.0**	**22.6**	3.8
2.6e5	45.3	48.4	43.8	5.1	**15.8**	**16.8**	**18.2**	7.2	4.7	5.4	5.7	OOR
1e6	**11.9**	**13.2**	**13.5**	6.6	3.1	4.1	3.2	OOR	0.8	0.0	0.7	OOR
4.1e6	3.1	0.9	3.5	OOR	0.0	1.0	0.0	OOR	0.3	0.1	0.2	OOR
1.6e7	0.1	0.1	1.2	OOR	0.0	0.1	0.2	OOR	0.2	0.0	0.0	OOR
H2N2 HA trimer A/Japan/305+/1957												
1e3	109.6	129.4	141.9	12.9	58.7	65.8	66.8	6.9	113.0	118.6	101.3	7.9
4e3	115.1	145.6	137.9	11.9	**24.7**	**25.4**	**26.1**	2.8	88.3	99.7	96.8	6.2
1.6e4	111.8	111.7	104.4	3.9	5.9	5.6	5.8	OOR	46.1	50.6	49.3	4.8
6.4e4	80.9	80.0	80.9	0.7	1.1	0.7	0.8	OOR	**12.7**	**12.8**	**14.3**	6.8
2.6e5	**28.5**	**32.7**	**30.5**	6.9	0.1	0.4	0.2	OOR	2.5	2.7	0.5	OOR
1e6	7.6	8.0	8.6	OOR	0.1	0.1	0.0	OOR	0.6	0.1	0.3	OOR
4.1e6	0.8	2.0	0.0	OOR	0.1	0.0	0.0	OOR	0.2	0.0	0.1	OOR
1.6e7	0.2	0.1	0.8	OOR	0.1	0.1	0.0	OOR	0.1	0.0	0.0	OOR
H3N2 trimer A/Singapore/INFIMH-16-0019/2016												
1e3	103.0	114.5	111.2	5.4	147.4	157.5	158.8	4.1	104.9	114.4	107.8	4.5
4e3	127.6	124.7	132.6	3.1	122.5	128.0	125.9	2.2	61.4	61.3	62.4	1.0
1.6e4	89.5	89.1	88.9	0.3	66.4	68.8	68.1	1.9	**18.0**	**21.0**	**22.7**	11.5
6.4e4	43.2	45.4	45.0	2.6	**20.6**	**22.2**	**21.0**	3.9	4.9	4.8	5.7	OOR
2.6e5	**12.7**	**12.4**	**11.7**	3.9	5.8	5.1	5.3	OOR	0.4	0.8	0.5	OOR
1.0e6	3.2	3.1	2.4	OOR	0.5	0.7	0.5	OOR	0.5	0.0	0.2	OOR
4.1e6	0.5	0.0	1.2	OOR	0.4	0.3	0.7	OOR	0.1	0.3	0.4	OOR
1.6e7	0.2	0.1	0.9	OOR	0.1	0.0	0.2	OOR	0.0	0.0	0.0	OOR
H5N1 A/Viet Nam/1194/2004												
1e3	130.6	120.0	124.8	4.2	43.6	61.7	72.9	24.9	115.3	120.5	113.7	3.0
4e3	112.7	129.5	130.9	8.1	**25.3**	**33.1**	**29.7**	13.4	71.7	71.7	73.3	1.3
1.6e4	71.3	82.7	88.2	10.7	6.3	8.6	8.8	OOR	**26.0**	**26.7**	**28.2**	4.2
6.4e4	49.2	54.5	52.1	5.0	0.5	2.2	1.1	OOR	5.4	6.4	6.1	OOR
2.6e5	**14.3**	**14.5**	**13.7**	3.0	0.2	0.2	0.7	OOR	0.4	0.9	0.3	OOR
1.0e6	2.8	3.1	2.9	OOR	0.4	0.1	0.0	OOR	1.0	0.1	0.2	OOR
4.1e6	0.3	0.0	0.8	OOR	0.2	0.0	0.6	OOR	0.1	0.1	0.2	OOR
1.6e7	0.0	0.0	1.1	OOR	0.0	0.0	0.0	OOR	0.1	0.0	0.0	OOR

	Reference				Donor 1				Donor 6			
	1	2	3	%CV	1	2	3	%CV	1	2	3	%CV
H7N1 HA trimer A/Shanghai/02/2013												
1e3	81.6	69.0	80.9	9.2	**23.5**	**27.5**	**17.3**	22.5	78.4	102.1	94.6	13.2
4e3	96.7	95.5	104.1	4.7	0.3	2.8	4.7	OOR	42.7	51.1	50.3	9.7
1.6e4	81.4	73.7	78.4	5.0	0.0	0.5	3.4	OOR	**16.5**	**16.3**	**15.7**	2.7
6.4e4	**35.8**	**37.1**	**36.7**	1.9	0.2	0.2	0.0	OOR	3.6	2.8	3.8	OOR
2.6e5	10.3	7.7	9.5	OOR	0.1	0.0	0.4	OOR	0.2	0.5	0.8	OOR
1.0e6	2.2	2.0	2.0	OOR	0.0	0.0	0.0	OOR	0.3	0.1	0.1	OOR
4.1e6	0.7	0.0	0.8	OOR	0.1	0.2	0.3	OOR	0.1	0.0	0.1	OOR
Influenza B/Phuket/3073/2013-like virus (B/Yamagata/16/88 lineage)												
1e3	71.9	118.8	129.4	28.7	154.7	159.6	158.8	1.7	109.4	139.5	138.5	13.3
4e3	38.5	96.1	105.4	45.3	160.4	162.7	165.6	1.6	145.0	144.3	150.9	2.5
1.6e4	152.0	160.6	146.6	4.6	142.7	148.3	147.4	2.0	108.5	112.7	111.0	1.9
6.4e4	130.1	150.9	151.3	8.4	86.2	86.9	84.5	1.4	50.3	54.8	56.6	6.0
2.6e5	87.6	90.9	86.4	2.7	**31.7**	**30.9**	**35.0**	6.6	**15.1**	**15.0**	**15.3**	1.0
1.0e6	**33.4**	33.9	33.8	0.8	6.9	8.1	7.9	OOR	3.1	2.8	4.7	OOR
4.1e6	7.8	**10.0**	**10.0**	13.6	1.2	1.7	2.1	OOR	0.2	0.9	0.4	OOR
Influenza B/Colorado/06/2017-like virus (B/Victoria/2/87 lineage)												
1e3	112.9	98.2	152.2	23.1	166.6	168.4	170.1	1.0	117.7	132.9	143.1	9.7
4e3	118.8	123.7	145.9	11.1	167.9	163.1	161.7	2.0	150.1	152.8	151.4	0.9
1.6e4	150.4	165.4	158.3	4.8	125.1	123.2	119.6	2.3	99.0	100.2	100.8	0.9
6.4e4	136.6	138.0	138.7	0.8	62.0	63.1	59.4	3.1	40.9	41.4	42.8	2.4
2.6e5	72.9	72.7	68.4	3.5	**19.8**	**21.6**	**21.9**	5.4	**10.6**	**10.0**	**11.1**	5.0
1.0e6	**25.0**	**26.1**	**24.4**	3.4	4.7	6.2	5.5	OOR	3.0	1.8	2.9	OOR
4.1e6	6.5	6.5	8.2	OOR	1.0	0.4	1.7	OOR	0.4	0.5	0.1	OOR

249   [1]Hook effect on low dilution points of titrations and all CV values are shown in grey.

250   [2]OOR indicates values below the lower level of quantitation signal of 10 (LLOQ).

251   [3]Values in bold are the titers where signal is greater or equal than the LLOQ of 10.

254   Table 4. Inter-assay precision from three donor plasma samples on ELISA-array antigens.[1,2,3]

	Reference				Donor 1				Donor 6			
	1	2	3	%CV	1	2	3	%CV	1	2	3	%CV
Flulaval seasonal vaccine 2018-2019												
1e3	105.2	97.3	99.4	4.1	138.3	110.5	132.3	11.5	148.4	138.3	142.4	3.6
4e3	146.8	150.0	102.8	19.8	108.7	110.7	106.6	1.9	116.2	95.9	128.4	14.5
1.6e4	128.4	109.1	120.0	8.1	97.2	112.4	89.8	11.6	133.0	138.7	134.7	2.1
6.4e4	77.5	84.2	75.8	5.6	46.1	60.5	40.3	21.3	105.6	94.9	92.3	7.2
2.6e5	**31.3**	**29.1**	**32.0**	5.1	**13.1**	**15.6**	**11.1**	17.0	51.4	41.9	42.2	11.9
1e6	7.4	9.1	8.0	OOR	4.3	5.5	3.2	OOR	**15.3**	**12.5**	**10.9**	17.1
4.1e6	1.0	1.1	0.6	OOR	0.5	0.4	0.4	OOR	3.0	3.5	3.9	OOR
1.6e7	0.1	0.4	0.2	OOR	0.0	0.4	0.5	OOR	0.6	0.7	0.4	OOR
H1N1 HA trimer A/Michigan/45/2015												
1e3	124.5	119.8	102.6	10.0	155.7	131.1	143.3	8.6	132.8	144.2	129.0	5.9
4e3	156.2	139.4	136.6	7.3	153.9	151.9	143.0	3.9	121.2	129.2	107.6	9.2
1.6e4	141.6	129.7	137.5	4.5	114.2	119.0	106.8	5.5	71.8	74.1	57.6	13.2
6.4e4	97.3	95.3	98.4	1.6	58.1	64.5	52.4	10.4	**25.6**	**28.8**	**18.7**	21.1
2.6e5	46.4	42.4	49.2	7.5	**17.4**	**18.7**	**14.2**	14.0	5.2	6.4	4.2	OOR
1e6	**11.5**	**11.3**	**12.3**	4.6	4.2	5.1	3.0	OOR	0.2	0.8	0.6	OOR
4.1e6	1.7	2.5	3.9	OOR	0.0	0.0	0.1	OOR	0.1	0.0	0.2	OOR
1.6e7	0.4	0.3	0.2	OOR	0.6	0.2	0.0	OOR	0.0	0.1	0.0	OOR
H2N2 HA trimer A/Japan/305+/1957												
1e3	110.7	123.7	80.5	21.1	67.4	41.8	57.1	23.2	107.8	122.4	99.5	10.5
4e3	142.3	116.1	104.6	16.0	**24.7**	**28.5**	**25.9**	7.3	83.0	91.6	98.6	8.6
1.6e4	120.8	96.6	127.5	14.1	5.2	6.9	5.7	OOR	53.8	54.1	45.7	9.3
6.4e4	78.9	73.8	85.1	7.1	1.1	0.6	0.7	OOR	**17.3**	**19.7**	**12.7**	21.6
2.6e5	**30.2**	**23.7**	**34.6**	18.4	0.1	0.0	0.5	OOR	2.3	3.6	2.0	OOR
1e6	6.0	6.4	8.8	OOR	0.3	0.0	0.3	OOR	0.5	0.6	0.4	OOR
4.1e6	0.8	0.2	1.8	OOR	0.2	0.0	0.0	OOR	0.1	0.2	0.1	OOR
1.6e7	0.1	0.7	0.4	OOR	0.2	0.1	0.4	OOR	0.0	0.4	0.0	OOR
H3N2 trimer A/Singapore/INFIMH-16-0019/2016												
1e3	97.8	115.4	108.1	8.2	157.4	136.2	150.3	7.3	112.1	100.1	85.1	13.7
4e3	140.2	114.0	119.5	11.1	128.6	133.8	119.6	5.6	63.4	69.3	54.3	12.1
1.6e4	95.3	71.9	88.5	14.1	65.7	76.6	61.9	11.2	**22.9**	**23.2**	**16.0**	19.7
6.4e4	42.1	**35.2**	39.2	8.9	**19.8**	**25.4**	**19.6**	15.2	5.7	5.2	4.2	OOR
2.6e5	**11.6**	9.4	**11.6**	11.5	5.2	5.9	4.5	OOR	1.3	0.9	0.3	OOR
1.0e6	3.5	1.8	2.8	OOR	4.9	0.7	0.3	OOR	0.2	0.2	0.2	OOR
4.1e6	0.5	0.9	0.3	OOR	0.3	0.5	0.1	OOR	0.3	0.2	0.2	OOR
1.6e7	0.4	0.0	0.2	OOR	0.5	0.2	0.2	OOR	0.1	0.1	0.0	OOR
H5N1 A/Viet Nam/1194/2004												
1e3	115.0	96.6	86.7	14.4	64.4	21.4	36.8	53.3	109.3	65.7	68.2	30.2
4e3	142.5	94.1	97.9	24.1	**32.3**	**19.9**	**21.1**	28.0	72.8	66.2	66.0	5.6
1.6e4	79.1	48.2	83.2	27.2	8.8	5.8	6.2	OOR	**27.7**	**30.4**	**23.0**	13.7
6.4e4	49.0	42.7	53.0	10.8	1.3	0.7	1.1	OOR	7.2	7.0	4.9	OOR
2.6e5	**14.7**	**12.0**	**16.2**	14.8	0.2	0.2	0.1	OOR	1.1	1.1	0.5	OOR
1.0e6	2.7	3.3	3.5	OOR	0.1	0.2	0.0	OOR	0.2	0.2	0.2	OOR
4.1e6	0.5	0.6	0.5	OOR	0.1	0.0	0.1	OOR	0.0	0.1	0.2	OOR
1.6e7	0.4	0.2	0.1	OOR	0.0	0.2	0.1	OOR	0.0	0.0	0.1	OOR

	Reference				Donor 1				Donor 6			
	1	2	3	%CV	1	2	3	%CV	1	2	3	%CV
H7N1 HA trimer A/Shanghai/02/2013												
1e3	95.1	115.2	98.9	10.3	**24.9**	8.8	**21.2**	11.5	92.8	83.6	85.7	5.5
4e3	113.7	84.2	97.7	15.0	1.5	6.1	5.7	OOR	57.3	66.4	54.4	10.6
1.6e4	86.7	75.7	82.9	6.8	0.7	7.3	0.6	OOR	**19.2**	**23.1**	**17.7**	14.0
6.4e4	**34.9**	**33.6**	42.9	2.7	0.2	0.1	0.0	OOR	4.7	4.6	3.3	OOR
2.6e5	9.9	9.9	**13.0**	OOR	0.0	0.0	0.1	OOR	3.0	2.2	3.0	OOR
1.0e6	2.4	2.1	3.1	OOR	0.2	0.3	0.3	OOR	0.9	0.2	0.8	OOR
4.1e6	0.3	0.3	0.5	OOR	0.2	0.1	0.0	OOR	0.1	0.3	0.1	OOR
Influenza B/Phuket/3073/2013-like virus (B/Yamagata/16/88 lineage)												
1e3	111.6	86.4	101.8	12.7	140.5	130.5	153.5	8.2	138.1	91.6	104.7	21.5
4e3	103.1	127.5	94.7	15.7	157.7	158.3	158.2	0.2	152.9	144.4	141.4	4.1
1.6e4	150.0	153.2	145.7	2.5	142.6	148.5	142.6	2.4	116.6	123.5	107.7	6.8
6.4e4	142.4	135.6	146.5	3.9	84.2	90.8	81.9	5.4	61.4	63.2	51.2	11.1
2.6e5	87.5	86.5	94.1	4.6	**30.7**	**35.1**	**28.8**	10.3	**17.2**	**19.9**	**12.8**	21.5
1.0e6	**33.5**	**35.4**	**38.8**	7.6	7.3	8.9	6.4	OOR	4.1	5.4	3.7	OOR
4.1e6	9.6	8.1	9.9	OOR	0.6	0.9	1.3	OOR	0.7	0.9	0.7	OOR
Influenza B/Colorado/06/2017-like virus (B/Victoria/2/87 lineage)												
1e3	103.4	112.5	99.1	6.6	168.5	164.0	164.4	1.5	137.8	147.5	110.3	14.6
4e3	142.9	133.5	133.8	3.9	161.3	162.6	158.3	1.4	149.8	162.4	146.3	5.6
1.6e4	161.3	159.7	158.8	0.8	119.9	122.3	121.6	1.0	102.4	111.7	95.6	7.8
6.4e4	130.2	127.5	125.6	1.8	59.8	66.9	62.6	5.7	47.9	50.8	**37.8**	15.0
2.6e5	69.2	70.3	74.4	3.9	**19.1**	**22.9**	**19.3**	10.5	**12.8**	**14.0**	9.3	20.2
1.0e6	**26.0**	**24.0**	**27.6**	6.8	4.8	6.4	5.0	OOR	3.0	3.8	2.5	OOR
4.1e6	6.4	6.8	6.8	OOR	0.2	0.6	0.9	OOR	0.5	0.6	1.0	OOR

255   [1]Hook effect on low dilution points of titrations and all CV values are shown in grey.

256   [2]OOR indicates values below the lower level of quantitation of 10.

257   [3]Values in bold are the titers where signal is greater or equal than the LLOQ of 10.

259 Table 5. Intra-assay and Inter-assay precision of plasma IC_{50} values on ELISA-Array antigens.

	Reference				Donor 1				Donor 6			
	1	2	3	%CV	1	2	3	%CV	1	2	3	%CV
Flulaval seasonal vaccine 2018-2019												
Intra	1.2e5	1.1e5	1.1e5	6.0	2.2e5	2.2e5	2.6e5	10.1	1.1e4	1.2e4	1.6e4	20.3
Inter	1.5e5	1.4e5	1.4e5	4.8	3.0e5	1.4e5	3.4e5	40.5	1.3e4	7.1e5	1.2e4	28.0
H1N1 HA trimer A/Michigan/45/2015												
Intra	8.6e6	7.3e6	8.6e6	8.8	2.24e5	2.0e5	2.2e5	6.3	6.4e5	8.1e5	7.1e5	11.8
Inter	1.0e5	8.4e6	7.1e6	17.6	2.6e5	2.4e5	2.6e5	4.1	5.6e5	6.2e5	8.1e5	20.0
H2N2 HA trimer A/Japan/305+/1957												
Intra	8.8e6	1.0e5	1.2e5	16.7	5.1e4	5.6e4	5.5e4	5.3	1.1e4	8.3e5	6.4e5	24.8
Inter	1.4e5	1.3e5	1.2e5	9.2	6.0e4	1.7e4	3.8e4	55.3	6.6e5	1.1e4	6.6e5	29.9
H3N2 trimer A/Singapore/INFIMH-16-0019/2016												
Intra	4.0e5	3.6e5	4.6e5	13.5	8.1e5	8.7e5	9.4e5	7.3	2.6e4	3.7e4	3.0e4	18.7
Inter	4.6e5	4.1e5	3.4e5	14.9	9.1e5	5.4e5	9.7e5	29.1	3.3e4	1.7e4	2.0e4	36.3
H5N1 A/Viet Nam/1194/2004												
Intra	5.2e5	5.2e5	4.2e5	12.0	2.4e4	3.0e4	7.6e4	66.0	2.3e4	2.7e4	2.0e4	14.4
Inter	5.6e5	1.9e5	1.4e5	75.8	3.6e4	9.3e5	2.6e4	56.2	1.9e4	9.8e5	7.9e5	48.4
H7N1 HA trimer A/Shanghai/02/2013												
Intra	2.4e5	2.6e5	3.0e5	13.0	>1e3	>1e3	>1e3	OOR	3.9e4	4.2e4	3.5e4	8.2
Inter	3.2e5	3.7e5	1.9e5	30.6	>1e3	>1e3	>1e3	OOR	2.3e4	1.2e4	2.1e4	31.6
Influenza B/Phuket/3073/2013-like virus (B/Yamagata/16/88 lineage)												
Intra	3.3e6	3.3e6	3.2e6	2.3	1.4e5	1.4e5	1.6e5	7.1	3.2e5	2.2e5	3.2e5	20.1
Inter	3.1e6	3.3e6	2.1e5	112.0	1.4e5	1.3e5	2.5e5	39.1	2.7e5	2.0e5	2.0e4	124.0
Influenza B/Colorado/06/2017-like virus (B/Victoria/2/87 lineage)												
Intra	4.2e6	5.2e6	4.5e6	11.4	2.6e5	2.7e5	3.0e5	7.7	2.3e5	2.8e5	3.2e5	15.7
Inter	4.5e6	5.7e6	3.8e6	21.3	2.9e5	2.3e5	2.5e5	10.0	2.7e5	3.8e5	5.1e5	30.6

DISCUSSION

An ability to print microarrays in a format for a 96-well ELISA-Array was first published by

Mendoza et al. (1999), and its utility for infectious disease testing has been demonstrated with

antibody arrays to encephalitis viruses (Kang et al. 2012) and viral antigen arrays to Flaviviridae

(Wang et al. 2015), using a non-contact piezo Bio-Dot Printing System (Biodot, Irvine, CA). As

with these two prior studies, we printed using non-contact piezo nozzles, but in smaller volumes

using a Scienion S12 instrument. We compared binding data in arrays using a variety of

Influenza antigen types including recombinant viral protein HA trimers, vaccine and viruses. The

parallel identification of viral antigen binding was carried out in a quantitative manner by

performing full dose-response curves of human plasma with an analysis of the precision of titer

and IC_{50} values. These data allowed for a comparison of the abundance and context of antibodies

from natural exposure or vaccination within a single individual and between individuals. In

future arrays it would be of interest to include the Influenza neuramidase protein, a second viral

surface protein that can be targeted by neutralizing antibodies (Memoli et al. 2016). It is

important to note that the quality and relevance of protein array data is only as good as the

proteins printed. The use of reference monoclonal antibodies or plasma with known activity is

helpful to characterize the integrity of the protein reagents and printing conditions. Such

reference reagents can also serve to bridge data between different array print lots and stability as

well as between data from different operators or labs. Stability testing of the ELISA-Array plates

supports the ability to ship plates on cold packs to a collaborating research lab, with a tolerance

of up to 48 hours at ambient temperature. The collaborating lab would only incur the lesser cost

and training required for the array reader.

There is great potential to use what is learned from protein arrays in research labs

towards the design and testing of vaccines and for the development of simplified rapid point-of-

care testing (POCT) of infectious diseases. To date, POCT efforts have taken the form of printed

arrays on lateral flow test strips, a promising technology needing further development to be

useful in endemic, low resource settings (Kim, Chung, and Kang 2019; Urusov, Zherdev, and

Dzantiev 2019). Protein array data can also be a valuable tool for the identification of individuals

most likely to have acquired broadly cross-reactive and/or potent neutralizing antibodies to

infectious disease. Passive immunization with monoclonal antibodies can be highly effective in

controlling viral pathogenesis (Salazar et al. 2017), and an ability to screen the serum of human

donors suspected to be clinically protected from disease in a multiplexed and quantitative format

can help identify the best donor for antibody discovery. In this application it is valuable to have

functional neutralization assay data on the same sera, to correlate to binding.

Overall, we have provided new methods and qualification data to support applications

ELISA-Array assay format for infectious disease research. The key advantages we observed with

this technology included the passive coating in a protein stabilizing buffer, the low protein

reagent consumption with nanoliter printing, and the ability to perform quantitative analyses

using nearly the same workflow, reagents and lab equipment as used in the conventional ELISA.

METHODS

Chemicals

Phosphate buffered saline (Gibco DPBS, calcium and magnesium free, pH 7.2) and EDTA

(0.5M Ambion) were obtained from Thermofisher (Waltham, MA). PBS with 0.05%

Polysorbate-20 was purchased as a 20x stock from Teknova Inc. (Hollister, CA). Fraction V

bovine serum albumin (BSA), gamma-Globulins from bovine blood (BGG), ProClin 300 and

CHAPS were purchased from Sigma-Aldrich (St. Louis, MO). Fetal Bovine Serum (FBS), US

source, triple 0.1um filtered, was sourced from Omega Scientific (Tarzana, CA).

General instrument and experimental parameters

The Scienion sciFLEXARRAYER S12 instrument (Scienion AG, Berlin, Germany) has been

optimized for non-contact, piezo-acoustic dispensing of ultra-low volumes from an inert coated

glass capillary in a climate-controlled (temperature, dewpoint and humidity) environment, with

precise XYZ axis control and on-board camera and software for QC of each spot in the array.

315   We printed our arrays with the PDC70 type 3 nozzle (Scienion) due to its reduced

dispense volume and the specific hydrophobic coating optimized to improve the dispense

stability of protein solutions. The system liquid is Milli-Q Water filtered through the Milli-Q

Ultrapure Water System (Millipore Sigma, Burlington, MA) that is subsequently degassed for at

least 30 minutes in a sonicating water bath. Proteins were spotted on low dust, black, clear

bottom, high protein binding Fluotrac™ 600, Greiner Bio-One 96-well plates (Fisher Scientific,

Waltham, MA), positioned on a ceramic platform under vacuum. The printing was carried out at

60% humidity and ambient temperature, with drop volumes of 330-350 pL. The 384-well source

plate (Scienion) was kept at dewpoint. Drop stability and array quality were assessed for quality

for each run. Prior to dispensing into the plates, autodrop detection was used to assess drop

stability by quantifying the velocity, deviation and drop volume for each protein spotted. In

addition, all plates were imaged with the on-board head camera after the completion of spotting

to ensure correct alignment and spot diameters. Printed arrays were incubated overnight at 75%

relative humidity and ambient room temperature to allow adsorption of the proteins to the

binding surface of the plate. Plates were then vacuum packaged and stored at 4°C until ready for

use.

The sciREADER CL2 (Scienion) is used for the colorimetric reading of the final assay of arrays.

After images are taken of each well, the software analysis program aligns the spot pattern to the

imaged spots and calculates a median intensity in absorbance units (AU).

Initial ELISA-Array assay used for optimization of parameters

Initial 96-well printed arrays were printed according to the general instrument parameters

described above. All assay steps were performed at ambient temperature. Blocking solution

(sciBLOCK Protein D1M solution, Scienion) was added at 200 µL/well with a multichannel

pipet and allowed to incubate without agitation for 1 hour. The block solution was manually

removed and the plate washed 1x by adding 300 µL/well of sciWASH Protein D1 agitating at

350rpm for 5 minutes on a Bioshake iQ thermomixer. Dilutions of human reference serum (with

IgG quantified at 4.4 mg/ml; Bethyl Laboratories) were made in blocking buffer (PBS, 0.05%

Tween-20, and 0.5% BSA), and 100 µL/well was added to the plate incubated for 1 hour with

gentle agitation (250 rpm). The arrays were manually washed 3 times with 2-5 min agitations in

between washes. A secondary polyclonal goat antibody (Southern Biotech, Birmingham, AL)

directed to the human kappa light chain and conjugated with horseradish peroxidase (HRP) was

used at 1/5000 in blocking buffer for 1 hour. After a second round of 3 manual washes the signal

was developed with sciCOLOR T2, a precipitating TMB reagent (Scienion), for 15 minutes.

Influenza hemagglutinin proteins

Recombinant hemagglutinin (HA) ectodomain constructs were made using gene blocks (IDT,

Newark, NJ) cloned into a pADD2 plasmid using EcoRI and XhoI restriction sites. The cloning

was performed with In-Fusion HD Cloning kit (Takara Bio, Mountain View, CA). Each

construct consisted of the native HA signal sequence, HA ectodomain, a trimeric foldon

domain of T4 fibronectin, an Avi tag sequence (GLNDIFEAQKIEWHE), and a hexa-histidine

affinity tag (Whittle J. et al., 2014). A negative control construct was made with the tags fused to

the GFP protein. Plasmids were purified with NucleoBond® Xtra Maxi kit (Macherey Nagel,

Düren, Germany) and transfected into Expi293 cells grown in a 1:2 mixture of Expi293 and

Freestyle (Gibco, Thermofisher Scientific) media. For transfection, 50 µg of plasmid was pre-

incubated with 1.3 ml of FectoPRO transfection reagent (Polyplus, New York, USA) and added

to 1L of media. At day 4, the media was clarified by centrifugation (7500xg, 15 min), filtered,

and diluted 2-fold with PBS. The media was then batch incubated with HisPur Ni-NTA resin

(Thermofisher Scientific) for 2 hours at 4°C and loaded on a gravity flow column. The resin was

washed with 20 column volumes of PBS with 5 mM imidazole and eluted in 4 ml of PBS with

250 mM imidazole. The eluted protein was concentrated and loaded on a Superdex 200 16/60

column (GE Healthcare, Chicago, IL) pre-equilibrated with PBS. The elution fractions

corresponding to the trimeric HA proteins were pooled, concentrated, and stored in 10% glycerol

in PBS at -20°C.

Human plasma and reference antibodies

Post-vaccination (28 days) blood samples were collected from nine healthy volunteers who

received the 2018/2019 seasonal influenza vaccine (FluLaval quadrivalent vaccine,

GlaxoSmithKline, Research Triangle Park, NC) in the fall of 2018 following a protocol approved

by Stanford University (IRB protocol 48130). Plasma was separated from heparinized blood

samples by centrifugation at 500g for 15 minutes and stored at -20°C. One donor was identified

with the highest plasma titer to all Influenza antigens and was used as a reference in all assays.

Two specific positive control reference antibodies with characterized low nanomolar affinities to

HA trimers were cloned and made recombinantly as human IgG1 in Expi293 cells with methods

described previously (Durham et al. 2019): MEDI8852 for InfA group 1 and 2 HA (Kallewaard

et al), and TF19 for InfB HA (unpublished in-house reagent). Two recombinant mAbs were used

as cross-reactivity controls, J9 directed to the Dengue viral envelope (Durham et al. 2019) and

3D3 to the RSV fusion protein (Collarini et al. 2009).

Printing of the Influenza protein array

Protein stocks of recombinant proteins at 0.5 mg/mL in PBS or vaccine stocks were diluted 1:1

in D12 buffer, mixed by pipetting, transferred to a 384 well polypropylene plate

(sciSOURCEPLATE, Scienion), and centrifuged for 2 min at 1800xg ambient temperature to

eliminate debris or air bubbles. The pattern printed was a 6x6 spot array on each well, and each

protein or vaccine along with positive and negative controls was printed in triplicate, 3 spots per

well, with 3 drops printed per spot. A single lot of twelve 96-well plates were printed in one day,

and after overnight curing were either subject to the ELISA-Array assay the next day (plates 1-

4), subject to temperature variations for one week (plates 5-7), or subject to varying incubation

times at 4°C (plates 8-10).

Influenza ELISA-Array assay

Each 96-well printed array was printed according to the general instrument parameters described

above. All assay steps were performed at ambient temperature, incubations except the blocking

step were done with low agitation on a Titer Plate Shaker (Lab-line Instruments, Melrose, IL)

and washes were done using a BioTek ELx405 plate washer (Bio-Tek Instruments, Winooski,

VT). High agitation is avoided as it leads to comets around the spots, which interferes with

accurate spot definition during reading of absorbance intensity. Array plates were washed 1x 300

µL/well before immediately adding 200 µL/well of assay diluent (PBS, 0.5% BSA, 2% filtered

FBS, 0.2% BGG, 0.25% CHAPS, 5mM EDTA, 0.05% Polysorbate-20 and 0.05% ProClin 300,

pH 7.2) down the sides of the wells with a multichannel pipette. Plates were allowed to block for

1 hour. Human plasma was diluted 1/1000 in assay diluent and serially diluted ¼ for n=8 points.

After manual removal of blocking solution with a multichannel pipette, plasma titrations were

added with one donor per column of the plate and incubated for 2 hours. The arrays were washed

3x 300 µL/well on the plate washer before adding 100 µL/well of a 1/5000 cocktail of

biotinylated secondary polyclonal goat antibodies (Southern Biotech) directed to the human

kappa and lambda light chains in assay diluent. After 1 hour of incubation and 3x 300 µL/well

washing, a SA-HRP high sensitivity conjugate (Pierce) was added for 1 hour. After a final 3x

300 µL/well wash, residual buffer was manually removed with a pipette and 50 µL/well of

sciCOLOR T2 added for 20 minutes.

Conventional ELISA

The same high protein binding plates as used in the ELISA-Array were coated overnight at 4°C

with 100 µL/well of the H1N1 A/Michigan/45/2015 strain of HA trimer in PBS, pH 7.2. The

ELISA assay was performed in the same manner as the ELISA-Array with the exception of the

development step. Plates were developed for 10 minutes with 50 µL/well soluble TMB substrate

(KPL Sure Blue 1-component, SeraCare Life Sciences, Milford, MA) and stopped with 50

µL/well of TMB Stop solution (KPL).

FIGURE LEGENDS

Figure 1. The ELISA-Array print pattern. The 6x6 array was printed in each well of a 96-well
plate. The outer edges contain fiducial spots of biotinylated anti-kappa antibody for orientation
(#11). All other spots are printed in triplicate, using anti-IgG Fc antibody as a positive control
(#1) and GFP-foldon as a negative control (#10). In the top half of the pattern are Influenza A
HA group I proteins (#2-4) and the 2018 vaccine (#5), and in the bottom half are Influenza A HA
group II proteins (#6-7) and Influenza B HA proteins (#8-9). At right is an image of one array
well developed after binding reference plasma on H1N1 A/Michigan/45/2015 HA trimer.

Figure 2. A comparison of the amount of protein needed in ELISA versus ELISA-Array.
Two sets of data are shown from a titration of a reference plasma on the H1N1
A/Michigan/45/2015 HA trimer. On the left, a conventional ELISA is shown using 4 different
antigen coat amounts. On the right, ELISA-array data is shown for the same antigen printed at
0.25ng where an equivalent IC_{50} is obtained when compared to the highest (1ug) antigen coat on
the conventional ELISA. The dashed line in ELISA-array plot indicates the lower limit of
quantification (10 AU).

Figure 3. The sensitivity of the ELISA-Array. Three sets of data are shown with the titration
of two monoclonal anti-Influenza antibodies and a reference plasma on the Influenza array
antigens. Panels show titrations of an anti-Influenza A mAb (left), an anti-Influenza B mAb
(center) and a reference plasma (right) against the array of Influenza A antigens (left & right),
Influenza B antigens (center & right), and the FluLaval 2018-2019 vaccine (all). Data is from 3
inter-assay plates, with mean and SD shown.

Figure 4. Dose-response curves of 3 donors on the Influenza antigen array. The dose-
dependent binding curves of a reference plasma (red), donor 1 plasma (blue) and donor 6 plasma
(green) to the Influenza antigen array are shown. The top left panel shows plasma titrations
against Influenza A group I HA trimers, the top right panel shows titrations against Influenza A
group II HA trimers, the bottom left panel shows titrations against Influenza B viruses, and the
bottom right panel shows titrations against the FluLaval 2018-2019 vaccine. Data is from 3 inter-
assay plates, with mean and SD shown.

Figure 5. IC_{50} comparisons of 3 donors on the Influenza antigen array. IC_{50} values
determined from 4-parameter fits of array antigen binding curves using a reference plasma (red),
donor 1 plasma (blue) and donor 6 plasma (green) are plotted. The top left panel shows IC_{50}
values against Influenza A group I HA trimers, the top right panel shows IC_{50} values against
Influenza A group II HA trimers, the bottom left panel shows IC_{50} values against Influenza B
viruses, and the bottom right panel shows IC_{50} values against the FluLaval 2018-2019 vaccine.
Data is from 3 inter-assay plates, with mean and SD shown. The IC_{50} value for donor 1 plasma
against H7 is not shown because the titer was greater than the minimum 1/1000 dilution.

Figure 6. Intra-assay, Inter-assay and Stability Performance of the ELISA-Array. Three
dose-dependent binding curves are shown using a reference plasma and the H1N1
A/Michigan/45/2015 HA trimer. The left panel shows both intra-assay error (plates #1-3) as well
as inter-operator robustness (plates #1-3 vs. #4). The center panel shows inter-assay error with
plates tested after increasing time periods stored at 4 ℃. The right panel shows stability after 1
week at ambient temperature , at 37 ℃ and up to 8 weeks at 4 ℃.

SUPPLEMENTARY MATERIAL

Supplementary Figure S1. Optimization of printing parameters. Initial tests of signal
intensity using different Scienion formulation buffers (D01 vs. D11 vs. D12), protein stock
concentrations (25 vs 100 vs 400 µg/mL), and number of drops (1 vs. 2 vs. 3). All spots are of
goat anti-human Fc, made in triplicate, and detected using anti-kappa HRP. BLK indicates buffer
background controls printed in the same array. Values in signal intensity are of human reference
serum dilutions from dark to light green of 1/1,000, 1/5,000, 1/10,000, 1/100,000, and
1/1,000,000.
480
481 **Supplementary Table S1.** Intra-assay precision (n=3 plates) from all eight donor plasma
482 samples, reference plasma, and 2 anti-Influenza mAbs on ELISA-array antigens.
483
484 **Supplementary Table S2.** Inter-assay precision (n=3 plates) from all eight donor plasma
485 samples, reference plasma, and 2 anti-Influenza mAbs on ELISA-array antigens.
486
487 **Supplementary Table S3.** Inter-operator robustness (n=2 plates) from all eight donor plasma
488 samples, reference plasma, and 2 anti-Influenza mAbs on ELISA-array antigens.
489
490 **Supplementary Table S4.** Stability of ELISA-Array plates at varying temperatures for 1 week
491 (n=3 plates) from all eight donor plasma samples, reference plasma, and 2 anti-Influenza mAbs
492 on ELISA-array antigens.
493
494 **Supplementary Table S5.** Stability of ELISA-Array plates under longer-term 4 ℃ storage (n=5
495 plates) from all eight donor plasma samples, reference plasma, and 2 anti-Influenza mAbs on
496 ELISA-array antigens.
497

AUTHOR CONTRIBUTIONS

KM and MS conceived and designed the study. EW, EC and KM carried out the experiments and

performed data analysis. NF produced the recombinant viral antigens. MS coordinated and

collected the donor plasma and provided the vaccine and virus materials. EW and KM wrote the

manuscript. All the authors read and approved the final manuscript.

ACKNOWLEDGEMENTS

The authors would like to thank Joshua Cantlon and Robert Kardish for their expert help with

establishing the initial parameters and training for ELISA-array printing.

FUNDING

This work was supported by the Chan Zuckerberg Biohub.

REFERENCES

Abbas, Abul K., Andrew H. Lichtman, and Shiv Pillai. 2014. *Cellular and Molecular*
Immunology. Philadelphia, PA: Elsevier Health Sciences.
Antia, Alice, Hasan Ahmed, Andreas Handel, Nichole E. Carlson, Ian J. Amanna, Rustom Antia,
and Mark Slifka. 2018. "Heterogeneity and Longevity of Antibody Memory to Viruses
and Vaccines." *PLoS Biology* 16 (8): e2006601.
https://doi.org/10.1371/journal.pbio.2006601.
Barbulovic-Nad, Irena, Michael Lucente, Yu Sun, Mingjun Zhang, Aaron R. Wheeler, and
Markus Bussmann. 2006. "Bio-Microarray Fabrication Techniques—a Review." *Critical*
Reviews in Biotechnology 26 (4): 237–59. https://doi.org/10.1080/07388550600978358.
Beest, Dennis te, Erwin de Bruin, Sandra Imholz, Jacco Wallinga, Peter Teunis, Marion
Koopmans, and Michiel van Boven. 2014. "Discrimination of Influenza Infection
(A/2009 H1N1) from Prior Exposure by Antibody Protein Microarray Analysis." *PloS*
One 9 (11): e113021. https://doi.org/10.1371/journal.pone.0113021.
Collarini, Ellen J., F. Eun-Hyung Lee, Orit Foord, Minha Park, Gizette Sperinde, Hai Wu,
William D. Harriman, Stephen F. Carroll, Stote L. Ellsworth, and Larry J. Anderson.
2009. "Potent High-Affinity Antibodies for Treatment and Prophylaxis of Respiratory
Syncytial Virus Derived from B Cells of Infected Patients." *The Journal of Immunology*
183 (10): 6338–45. https://doi.org/10.4049/jimmunol.0901373.
Crum-Cianflone, N. F., G. Collins, G. Defang, E. Iverson, Lynn E. Eberly, C. Duplessis, J.
Maguire, A. Ganesan, B. K. Agan, and T. Lalani. 2012. "Immunoglobulin G Subclass
Levels and Antibody Responses to the 2009 Influenza A (H1N1) Monovalent Vaccine
among Human Immunodeficiency Virus (HIV)-infected and HIV-uninfected Adults."
Clinical & Experimental Immunology 168 (1): 135–41. https://doi.org/10.1111/j.1365-
2249.2011.04550.x.
Davies, D. Huw, Xiaowu Liang, Jenny E. Hernandez, Arlo Randall, Siddiqua Hirst, Yunxiang
Mu, Kimberly M. Romero, Toai T. Nguyen, Mina Kalantari-Dehaghi, and Shane Crotty.
2005. "Profiling the Humoral Immune Response to Infection by Using Proteome
Microarrays: High-Throughput Vaccine and Diagnostic Antigen Discovery."
Proceedings of the National Academy of Sciences 102 (3): 547–52.
https://doi.org/10.1073/pnas.0408782102.
De Silva, Nilushi S., and Ulf Klein. 2015. "Dynamics of B Cells in Germinal Centres." *Nature*
Reviews Immunology 15 (3): 137–48. https://doi.org/10.1038/nri3804.
Desbien, Anthony L., Neal Van Hoeven, Steven J. Reed, Allen C. Casey, John D. Laurance,
Susan L. Baldwin, Malcolm S. Duthie, Steven G. Reed, and Darrick Carter. 2013.
"Development of a High Density Hemagglutinin Protein Microarray to Determine the
Breadth of Influenza Antibody Responses." *Biotechniques* 54 (6): 345–48.
https://doi.org/10.2144/000114041.
Durham, Natasha D., Aditi Agrawal, Eric Waltari, Derek Croote, Fabio Zanini, Mallorie Fouch,
Edgar Davidson, Olivia Smith, Esteban Carabajal, and John E. Pak. 2019. "Broadly
Neutralizing Human Antibodies against Dengue Virus Identified by Single B Cell
Transcriptomics." *ELife* 8: e52384. https://doi.org/10.7554/eLife.52384.

Ekins, Roger P. 1989. "Multi-Analyte Immunoassay." *Journal of Pharmaceutical and*
Biomedical Analysis 7 (2): 155–68. https://doi.org/10.1016/0731-7085(89)80079-2.
Engvall, Eva, and Peter Perlmann. 1972. "Enzyme-Linked Immunosorbent Assay, ELISA: III.
Quantitation of Specific Antibodies by Enzyme-Labeled Anti-Immunoglobulin in
Antigen-Coated Tubes." *The Journal of Immunology* 109 (1): 129–35.
Gogalic, Selma, Ursula Sauer, Sara Doppler, and Claudia Preininger. 2018. "Investigating
Colorimetric Protein Array Assay Schemes for Detection of Recurrence of Bladder
Cancer." *Biosensors* 8 (1): 10. https://doi.org/10.3390/bios8010010.
Huang, Wei, Kelly Whittaker, Huihua Zhang, Jian Wu, Si-Wei Zhu, and Ruo-Pan Huang. 2018.
"Integration of Antibody Array Technology into Drug Discovery and Development."
Assay and Drug Development Technologies 16 (2): 74–95.
https://doi.org/10.1089/adt.2017.808.
Huang, Yi, and Heng Zhu. 2017. "Protein Array-Based Approaches for Biomarker Discovery in
Cancer." *Genomics, Proteomics & Bioinformatics* 15 (2): 73–81.
https://doi.org/10.1016/j.gpb.2017.03.001.
Kallewaard, Nicole L., Davide Corti, Patrick J. Collins, Ursula Neu, Josephine M. McAuliffe,
Ebony Benjamin, Leslie Wachter-Rosati, Frances J. Palmer-Hill, Andy Q. Yuan, and
Philip A. Walker. 2016. "Structure and Function Analysis of an Antibody Recognizing
All Influenza A Subtypes." *Cell* 166 (3): 596–608.
https://doi.org/10.1016/j.cell.2016.05.073.
Kang, Xiaoping, Yuchang Li, Li Fan, Fang Lin, Jingjing Wei, Xiaolei Zhu, Yi Hu, Jing Li,
Guohui Chang, and Qingyu Zhu. 2012. "Development of an ELISA-Array for
Simultaneous Detection of Five Encephalitis Viruses." *Virology Journal* 9 (1): 56.
https://doi.org/10.1186/1743-422X-9-56.
Katzelnick, Leah C., Lionel Gresh, M. Elizabeth Halloran, Juan Carlos Mercado, Guillermina
Kuan, Aubree Gordon, Angel Balmaseda, and Eva Harris. 2017. "Antibody-Dependent
Enhancement of Severe Dengue Disease in Humans." *Science* 358 (6365): 929–32.
https://doi.org/10.1126/science.aan6836.
Kim, Hanbi, Doo-Ryeon Chung, and Minhee Kang. 2019. "A New Point-of-Care Test for the
Diagnosis of Infectious Diseases Based on Multiplex Lateral Flow Immunoassays."
Analyst 144 (8): 2460–66. https://doi.org/10.1039/C8AN02295J.
Kingsmore, Stephen F. 2006. "Multiplexed Protein Measurement: Technologies and
Applications of Protein and Antibody Arrays." *Nature Reviews Drug Discovery* 5 (4):
310. https://doi.org/10.1038/nrd2006.
Koopmans, M., Erwin de Bruin, G.-J. Godeke, I. Friesema, R. Van Gageldonk, Maarten
Schipper, Adam Meijer, Rob van Binnendijk, G. F. Rimmelzwaan, and M. D. De Jong.
2012. "Profiling of Humoral Immune Responses to Influenza Viruses by Using Protein
Microarray." *Clinical Microbiology and Infection* 18 (8): 797–807.
https://doi.org/10.1111/j.1469-0691.2011.03701.x.
Liew, Michael, Matthew C. Groll, James E. Thompson, Sara L. Call, Joann E. Moser, Justin D.
Hoopes, Karl Voelkerding, Carl Wittwer, and Rex S. Spendlove. 2007. "Validating a
Custom Multiplex ELISA against Individual Commercial Immunoassays Using Clinical
Samples." *Biotechniques* 42 (3): 327–33. https://doi.org/10.2144/000112332.
Lowell, George H., Sagit Ziv, Svetlana Bruzil, Ron Babecoff, and Tamar Ben-Yedidia. 2017.
"Back to the Future: Immunization with M-001 Prior to Trivalent Influenza Vaccine in

598 2011/12 Enhanced Protective Immune Responses against 2014/15 Epidemic Strain."
599 *Vaccine* 35 (5): 713–15. https://doi.org/10.1016/j.vaccine.2016.12.063.
600 Madore, Dace V., Bruce D. Meade, Fran Rubin, Carolyn Deal, and Freyja Lynn. 2010.
601 "Utilization of Serologic Assays to Support Efficacy of Vaccines in Nonclinical and
602 Clinical Trials: Meeting at the Crossroads." *Vaccine* 28 (29): 4539–47.
603 https://doi.org/10.1016/j.vaccine.2010.04.094.
604 Memoli, Matthew J., Pamela A. Shaw, Alison Han, Lindsay Czajkowski, Susan Reed, Rani
605 Athota, Tyler Bristol, Sarah Fargis, Kyle Risos, and John H. Powers. 2016. "Evaluation
606 of Antihemagglutinin and Antineuraminidase Antibodies as Correlates of Protection in an
607 Influenza A/H1N1 Virus Healthy Human Challenge Model." *MBio* 7 (2): e00417-16.
608 https://doi.org/10.1128/mBio.00417-16.
609 Mendoza, Leo G., P. McQuary, A. Mongan, Rajeswari Gangadharan, Stafford Brignac, and
610 Maren Eggers. 1999. "High-Throughput Microarray-Based Enzyme-Linked
611 Immunosorbent Assay (ELISA)." *Biotechniques* 27 (4): 778–88.
612 https://doi.org/10.2144/99274rr01.
613 Nakajima, Rie, Medalyn Supnet, Algis Jasinskas, Aarti Jain, Omid Taghavian, Joshua Obiero,
614 Donald K. Milton, Wilbur H. Chen, Michael Grantham, and Richard Webby. 2018.
615 "Protein Microarray Analysis of the Specificity and Cross-Reactivity of Influenza Virus
616 Hemagglutinin-Specific Antibodies." *MSphere* 3 (6): e00592-18.
617 https://doi.org/10.1128/mSphere.00592-18.
618 Price, Jordan V., Justin A. Jarrell, David Furman, Nicole H. Kattah, Evan Newell, Cornelia L.
619 Dekker, Mark M. Davis, and Paul J. Utz. 2013. "Characterization of Influenza Vaccine
620 Immunogenicity Using Influenza Antigen Microarrays." *PloS One* 8 (5): e64555.
621 https://doi.org/10.1371/journal.pone.0064555.
622 Salazar, Georgina, Ningyan Zhang, Tong-Ming Fu, and Zhiqiang An. 2017. "Antibody
623 Therapies for the Prevention and Treatment of Viral Infections." *Npj Vaccines* 2: 19.
624 https://doi.org/10.1038/s41541-017-0019-3.
625 Semenova, V. A., E. Steward-Clark, K. L. Stamey, T. H. Taylor, D. S. Schmidt, S. K. Martin, N.
626 Marano, and C. P. Quinn. 2004. "Mass Value Assignment of Total and Subclass
627 Immunoglobulin G in a Human Standard Anthrax Reference Serum." *Clin. Diagn. Lab.*
628 *Immunol.* 11 (5): 919–23. https://doi.org/10.1128/CDLI.11.5.919-923.2004.
629 Templin, Markus F., Dieter Stoll, Monika Schrenk, Petra C. Traub, Christian F. Vöhringer, and
630 Thomas O. Joos. 2002. "Protein Microarray Technology." *Drug Discovery Today* 7 (15):
631 815–22. https://doi.org/10.1016/S1359-6446(00)01910-2.
632 Tighe, Patrick J., Richard R. Ryder, Ian Todd, and Lucy C. Fairclough. 2015. "ELISA in the
633 Multiplex Era: Potentials and Pitfalls." *PROTEOMICS–Clinical Applications* 9 (3–4):
634 406–22. https://doi.org/10.1002/prca.201400130.
635 Urusov, Alexandr E., Anatoly V. Zherdev, and Boris B. Dzantiev. 2019. "Towards Lateral Flow
636 Quantitative Assays: Detection Approaches." *Biosensors* 9 (3): 89.
637 https://doi.org/10.3390/bios9030089.
638 Wang, D., Y. Zheng, X. Kang, X. Zhang, H. Hao, W. Chen, L. Liu, X. Li, L. Li, and Q. Yuan.
639 2015. "A Multiplex ELISA-Based Protein Array for Screening Diagnostic Antigens and
640 Diagnosis of Flaviviridae Infection." *European Journal of Clinical Microbiology &*
641 *Infectious Diseases* 34 (7): 1327–36. https://doi.org/10.1007/s10096-015-2353-6.
642 Yoshida, Taketoshi, Henrik Mei, Thomas Dörner, Falk Hiepe, Andreas Radbruch, Simon
643 Fillatreau, and Bimba F. Hoyer. 2010. "Memory B and Memory Plasma Cells."

644   *Immunological Reviews* 237 (1): 117–39. https://doi.org/10.1111/j.1600-
645   065X.2010.00938.
646

Figure 1. The ELISA-Array print pattern.

1: xIgG Fc (positive control)
2: HA H1 (group I)
3: HA H2 (group I)
4: HA H5 (group I)
5: FluILaval vaccine (2018)
6: HA H3 (group II)
7: HA H7 (group II)
8: HA FluB (B/Yamagata)
9: HA FluB (B/Victoria)
10: GFP foldon (negative control)
11: xkappa biotin (fiducials)

11	2	3	4	5	11
11	2	3	4	5	10
1	2	3	4	5	10
1	6	7	8	9	10
1	6	7	8	9	
11	6	7	8	9	11

Figure 2. A comparison of the amount of protein needed in ELISA *versus* ELISA-Array.

$IC_{50}= 1.1 \times 10^5$

$IC_{50}= 1.3 \times 10^5$

Figure 3. The sensitivity of the ELISA-Array.

Figure 4. Dose-response curves of 3 donors on the Influenza antigen array.

Figure 5. IC$_{50}$ comparisons of 3 donors on the Influenza antigen array.

Figure 6. Intra-assay, Inter-assay and Stability Performance of the ELISA-Array.