**A streamlined workflow for automated cryo focused ion beam milling**

**Sebastian Tacke[1,4], Philipp Erdmann[2,4], Zhexin Wang[1], Sven Klumpe[2], Michael Grange[1],**
**Radim Kříž[3], Matej Dolník[3], John Mitchels[3], Jürgen Plitzko[2,*], Stefan Raunser[1,*]**

[1]Department of Structural Biochemistry, Max Planck Institute of Molecular Physiology, Otto-
Hahn-Str. 11, 44227 Dortmund, Germany

[2]Department Molecular Structural Biology, Max Planck Institute of Biochemistry, Am
Klopferspitz 18, 82152 Martinsried, Germany

[3]Thermo Fisher Scientific, Materials & Structural Analysis, Vlastimila Pecha 1282/12, 62700
Brno, Czech Republic.

[4]These authors contributed equally

[*]Correspondence: stefan.raunser@mpi-dortmund.mpg.de, plitzko@biochem.mpg.de

# Abstract

Cryo-electron tomography (cryo-ET) is an emerging technique to study the cellular
architecture and the structure of proteins at high resolution *in situ*. Most biological specimens
are too thick to be directly investigated and are therefore thinned by milling with a focused
ion beam under cryogenic conditions (cryo-FIB). This procedure is unautomated and prone to
contaminations, which makes it a tedious and time-consuming process, often leading to
suboptimal results. Here, we present new hardware and software tools that overcome the
current limitations. We developed a new glove box and a high vacuum cryo transfer system
and installed a stage heater, a cryo-shield and a cryo-shutter in the FIB milling microscope.
This reduces the ice contamination during the transfer and milling process and simplifies the
handling of the sample. In addition, we introduce a new software application that automates
the key milling steps. Together, these improvements allow for high-quality, high-throughput
cryo-FIB milling. This paves the way for new types of experiments, which have been previously
considered infeasible.

## Introduction

Electron tomography at cryogenic temperatures (cryo-ET) offers the unique possibility to structurally analyse biological macromolecules in their native cellular environment [Grange2017, Bartesaghi2012]. Using this method, it is even possible to determine protein structures at near-atomic resolution by applying sub-volume averaging techniques [Schur2016].

One prerequisite for electron cryo microscopy is that the sample is thinner than the inelastic mean-free-path length of electrons (approximately 350 nm for 300 keV electrons [Vulovic2013]. Thicker samples result in an increased number of inelastically scattered electrons, which decreases the quality of the image. The majority of biological samples (e.g. eukaryotic cells or tissues) are too thick for cryo-ET and need to be thinned.

Cryo-ultramicrotomy is one approach that can be used to reduce specimen thickness. Here, the sample is cut into ribbons of vitreous sections that are thin enough to be imaged by cryo-ET [Al-Amoudi2004]. However, this method is technically demanding and prone to sample-altering artefacts such as knife marks, compression or crevasses [Al-Amoudi2005]. Another approach uses a focused ion beam under cryogenic conditions (cryo-FIB) in a scanning electron microscope (SEM) to thin samples [Marko2007, Rigort2010]. In recent years, numerous studies have demonstrated the suitability of this technique for preparing biological samples for cryo-ET without the artefacts known from cryo-ultramicrotomy [Hagen2015, Mahamid2016].

Current best-practice protocols for cryo-FIB milling include: i) vitrification of the sample, ii) identification of regions-of-interest, iii) rough ablation of the surrounding sample material and iv) polishing of the target region, leaving a thin lamella ready for cryo-ET [Medeiros2018]. However, the entire procedure involves multiple handling, transfer, milling and imaging steps. During these steps, samples need to be handled with forceps and it is difficult even for highly experienced practitioners not to destroy the thin and fragile lamellae. Moreover, since the vitrified sample needs to be cooled constantly to avoid de-vitrification, it is especially prone to any kind of ice contamination. Contamination of polished lamellae is particularly disruptive to cryo-ET as ice crystals often obscure the view on features otherwise visible in the lamellae. It is even possible that ice crystals conceal the entire lamella, which prevents any meaningful data acquisition (Figure 1a). Besides the transfer, which bears a high risk of exposing the sample to water, a potential source of (amorphous) ice contamination is

the cryo-FIB/SEM itself. Despite the high vacuum inside the SEM, water molecules from the

residual gas inside the microscope chamber can still deposit on the sample that is kept well

below -140 °C to avoid devitrification (Figure 1b). Depending on the partial pressure of water

inside the chamber, this effect can be severe and ultimately limits the number of lamellae that

can be prepared per session.

During the milling process, most of the steps to date need to be adjusted manually,

resulting in relatively long operator times at the cryo-FIB/SEM. Because of the lack of

automation, unattended overnight runs are currently not possible, limiting the total number

of preparations per week. Taken together, the highest throughput reported so far is limited

to 16 lamellae per session (10 hours) [Medeiros2018].

To obtain optimal lamellae for cryo-ET and to increase the overall throughput of cryo-

FIB milling, three major hurdles need to be overcome: First, all types of contamination need

to be reduced to a minimum, and anhydrous environments need to be established for the

handling and transfer of the sample. At the same time, surface contamination inside the

microscope needs to be decreased. Second, the handling of the lamellae must be improved

to avoid lamella cracking. Third, all steps for the production of the lamellae on the system

need to be automated in order to reduce heterogeneity and increase throughput.

Here, we present an integrated workflow that addresses and overcomes all three

bottlenecks. In a first step, we implemented a specially designed glove box and a high vacuum

cryo transfer system into the current workflow to reduce ice contamination during handling

and transfer. Secondly, we tremendously reduced the contamination rates inside the cryo-

FIB/SEM by the installation of a stage heater, a cryo-shield and a cryo-shutter. Thirdly, we

demonstrate that most of the milling steps can be automated by a new software application.

We show that in combination the new hardware and software increase the overall throughput

of lamella production, enabling new types of experiments, which are considered infeasible

with current setups.

## Results

### Sample preparation and handling

97    After vitrification by plunge or high-pressure freezing, the EM grid with the sample is placed

98    into the Autogrid, then put into the Aquilos Autogrid shuttle and transferred to the cryo-

99    FIB/SEM. During these steps the sample needs to be kept at temperatures below -140˚C to

avoid hexagonal ice formation. Many cryo-EM laboratories have implemented air-dehumidifier rooms. While those help to reduce the overall humidity, and thereby ice contamination, they cannot completely prevent it.

To eliminate ice contamination during handling and simplify implementation of the technique in laboratories unable to control room humidity, we developed a glove box system (Figure 2a) that allows the grid preparation steps to be performed in a tightly regulated atmosphere. During operation, the glove box is constantly purged with dry nitrogen gas to keep the humidity below the detection limit (1 %) of the humidity sensors used. Beyond functionality, the glove box system was designed to provide maximal convenience during grid preparation. To guarantee that the sample can be brought into the box without disturbing the anhydrous environment inside, we equipped the glove box with a load lock and a port for a high vacuum cryo transfer system. The direct supply of liquid nitrogen is provided by a liquid nitrogen reservoir, which is located inside the glove box. Furthermore, we installed a set of heated tool holders in the box that allow to heat and dry different tools, such as forceps, clipping tools, cassette gripper, etc. directly after usage.

To increase the ease of handling grids inside the box, we designed and built a dedicated universal preparation station (Figure 2b), again maximising the convenience for the user. The station contains an isolated liquid nitrogen bath with three different preparation modules for each step of grid preparation, namely autoloader grid assembly, cryo-FIB shuttle loading and autoloader cassette loading. The users can arrange the needed modules according to their working routine.

To avoid contamination and devitrification during sample transfer from the plunge freezer to the glove box and from there to the cryo-FIB/SEM, we implemented a previously developed high vacuum cryo transfer system [Tacke2016] (Figure 2c), which consists of a transfer rod, a liquid nitrogen reservoir, a vacuum chamber, a valve and a plug & play adapter for the glove box or cryo-FIB/SEM. To avoid devitrification during transfer, the cryo-FIB shuttle is actively cooled by the liquid nitrogen reservoir as soon as the transfer rod is retracted in its parking position. Moreover, because the cooled surfaces of the nitrogen reservoir act as a cryo-pump, the vacuum level is kept constant as long as the reservoir is filled [Tacke2016].

With the introduced devices, we are able to minimise the artefacts, currently associated with existing workflows streamlining the entire grid preparation whilst improving user convenience, efficiency and flexibility. Especially during sample screening, where many

samples have to be prepared and transferred over a longer timeframe, the benefit of the devices is apparent. Due to the reduced humidity, the preparation station can stay cooled for several hours without ice contamination and the high vacuum cryo transfer system allows the necessary quick and contamination-free transfer.

**Hardware modification of the cryo-FIB**

Although cryo-FIB/SEMs are typically operated at high vacuum conditions, contamination of the sample inside the microscope cannot be completely avoided, due to various sources of contamination in the FIB chamber. One of these is the ablated sample material which re-deposits on the specimen. This kind of contamination is greatest at the beginning of lamellae preparation when the sample material is milled in bulk and decreases to a minimum towards the final polishing step. It is therefore not a major problem for the production of lamellae if the milling is performed in two stages: first all sites are roughly milled to a thickness of approximately 500 - 800 nm and only then the final thinning is performed.

The second source of contamination is residual water inside the vacuum chamber. Its deposition/sublimation rate depends on two factors: i) the fraction of water in the residual gas mixture (partial pressure) and ii) the surface temperature of the sample material (Supplementary Figure 1a). We installed a large cryo-shield and a cryo-shutter to reduce the partial pressure of water in the system to less than $3 \times 10^{-9}$ mbar and heated the sample to max. -165 ˚C using a newly installed stage heater (Figure 3). Combined, these actions moved the ratio between deposition and sublimation closer towards an equilibrium, i.e. minimal deposition rates, which is optimal for cryo-FIB milling (Supplementary Figure 1a) (see also [Umrath1983], [Echlin1992]).

Due to the increased pumping speed, the partial pressure decreases from 2-3 x $10^{-7}$ mbar to approximately $3 \times 10^{-9}$ mbar when the cryo-shield is cooled with liquid nitrogen. After 30 min, the system reaches the lowest partial pressure of water (Supplementary Figure 1b). The cryo-shutter in the direct vicinity of the sample (Figure 3) reduces the partial pressure in this position even further. Under these conditions the contamination rate is $5.6 \pm 1.0$ nm $h^{-1}$ in the milling position compared to $85 \pm 7$ nm $h^{-1}$ without these modifications (Supplementary Figure 2, Supplementary Table 1). It should be mentioned, that the initial contamination rate is different for each microscope. Typically, the contamination rate for the used cryo-FIB/SEM

is specified to be smaller than 50 nm h$^{-1}$. In combination, the improved vacuum conditions
allow for milling over a longer time course and pave the way for supervised automated milling
strategies.

**Automated lamella production pipeline**

The lamella production workflow can be separated in four major steps: i) definition of
region(s)-of-interest (ROI), ii) site preparation, iii) rough milling and iv) thinning
(Supplementary Figure 3). In this study, we used the software application MAPS (Thermo
Fisher Scientific) to visually identify ROIs directly in the cryo-FIB/SEM (Supplementary Figure
3a). If this is not possible or rare biological events, such as for instance cell-cell fusions, are
investigated, the correlation of light cryo microscopy and cryo-FIB/SEM is needed. After
defining the milling position, the eucentric height and the milling angle are set for each lamella
site. We then used the new software application AutoTEM Cryo (Thermo Fisher Scientific) for
the next steps. AutoTEM Cryo executes site preparation, rough milling and thinning in a
supervised or fully automated manner (Supplementary Figure 3b).

First, AutoTEM Cryo imports the list of lamella sites from the MAPS software and all
parameters relevant for milling, such as ion beam currents, dimensions of the initial milling
pattern, desired lamellae dimensions etc., are defined. Once set, the program saves all
relevant parameters in template files. This allows to mill future batches exactly under the
same conditions and thereby helps to improve the reproducibility of the process.
Furthermore, with the help of the defined templates, the milling process can be easily
optimized for different types of specimen.

After defining all relevant parameters, the automated milling process starts (Figure 4).
If two different materials are used during EM grid production (e.g. copper and carbon), the
stability of the fragile carbon layer might suffer during freezing due to the different expansion
coefficients of the two materials. During site preparation, the ROI is centered and thereafter
relief cuts are milled, to release the stress of the support layer (Figure 4a) [Wolff2019]. The
subsequent rough milling is divided into four steps (Figure 4b). First, the bulk material is
removed with relatively high ion beam currents (0.5 – 1 nA) until a lamella thickness of
approximately 2-3 µm is reached (Figure 4c). In the following three milling steps, the ion
currents are successively decreased from approximately 0.5 nA to 0.1 nA and the lamella
thickness is gradually reduced to approximately 150 - 250 nm (Figure 4c). To produce a

uniform thickness, the lamella is tilted by a few degrees prior to milling to compensate for the
ion beam divergence [Schaffer2017]. Finally, the lamella is polished with relatively low ion
beam currents (30 -50 pA) to obtain a thickness of 100 - 200 nm (Figure 4d).
199

**Efficiency estimation and quality assessment**

To estimate the efficiency of our lamellae preparation pipeline, that is the production
of thin and intact lamellae, we separated the handling steps into two parts: i) the pre-
processing and ii) the post-processing. The pre-processing includes the assembly of the
autoloader grid, its loading to the cryo-FIB shuttle and the transfer of the shuttle from the
glove box to the cryo-FIB. The post-processing step comprises the transport from cryo-
FIB/SEM to the glove box and the transfer of the autoloader grid from the cryo-FIB shuttle to
the autoloader cassette.

In recently published preparation protocols, the success rate of obtaining suitable
specimens for proficient users was estimated to be 80% for the pre-processing and 90% for
the post-processing step, which results in a total efficiency of 72% [Medeiros2018]. In reality,
the efficiency is probably even lower because some lamellae will not be usable for cryo-ET
analysis.

Our improvements of the pre-processing and post-processing steps resulted in a
considerably higher efficiency of above 95% for the entire handling (pre- and post-processing).
In addition, all intact lamellae were almost contamination free (Figure 5a) and only in rare
cases, ice crystals could be found in the vicinity of the lamellae. Importantly, since the
contamination inside the cryo-FIB/SEM was reduced to a minimum, almost no contamination
layer could be identified in the final tomograms (Figure 5b) and approximately 90% of all intact
lamellae were in principle useable for cryo-ET investigations. Taken together, the high
efficiency and low contamination rate reflect the impact of the newly implemented hardware
components.

To assess the quality of the automated lamellae preparation, we applied the same
approach as described above. We investigated different types of specimens and quantified
the success rates for the rough milling and polishing. We considered a milling task as
successfully terminated, if the following criteria were fulfilled: i) the lamella was intact, i.e. not
broken, ii) the desired thickness was reached, iii) tracking of the lamella site was successful,
i.e. the right position of the sample was milled. We determined an average success rate of

95% and 88% for the rough milling and polishing, respectively (Table 1). With the estimated efficiency of 95% for the pre- and post-processing and the calculated efficiency of 84% for the milling process, the efficiency of the total workflow can be estimated to be around 80%. Although the success rate was much higher in the case of manual milling (above 95%), the total throughput is much higher since overnight runs are possible at this point. Moreover, the operator time is significantly reduced.

To demonstrate that lamellae, prepared by automated cryo-FIB milling using AutoTEM cryo, are suitable for subtomogram averaging, a set of ribosomes were identified on several lamellae (Figure 6a,b). With only 15,000 particles, a structure of 14 Å (0.143 FSC criteria) was calculated (Figure 6c, e). Since the orientations and locations can be precisely determined, we were able to identify polysomes (Figure 6d).

Taken together, these results clearly demonstrate that the streamlined workflow not only increases the overall throughput of the lamellae production but also generates results that are suitable for in-situ high-resolution investigations.

## Discussion

Contamination of the samples, inconvenient sample handling and the lack of automation are the three dominant bottlenecks of current cryo-FIB workflows, which can decrease the quality of the final lamellae and limit the overall throughput per milling session. In the present study we overcame the aforementioned limitations by the introduction of new hardware components (Table 2) and a new software to automate milling procedures.

To reduce sample contamination throughout the entire preparation process, we introduced a glove box and a cryo high vacuum transfer system. Moreover, we significantly reduced the contamination rates inside the vacuum system, mainly because of the installation of an additional cryo-shield and a cryo-shutter. Now, the entire handling and transfer steps as well as the cryo-FIB milling can be performed in an almost anhydrous environment. In order to streamline the handling of the grids carrying fragile lamellae, we introduced a new preparation station, which aims at avoiding all types of artefacts related to sample handling with maximal convenience for the user.

The contamination rate of 88 nm h$^{-1}$ in the initial cryo-FIB/SEM setup limited the total number of prepared lamellae per session. Due to the improved and new hardware, we could

decrease the contamination rate to 5.6 nm $h^{-1}$ in the milling position. Therefore, the lamella

production is not restricted to a specific time frame anymore and, in principle, an 24/7

operation is possible now.

Since this is difficult to realize by manual operation, we introduced the new software

application AutoTEM Cryo, which successively processes the queue of lamella sites in a

supervised to automated manner, allowing for the automatic batch milling of many lamellae.

With the help of AutoTEM Cryo, the time spent on the microscope is now reduced to

approximately 5 minutes per lamella site at the beginning of the milling process. Because of

the automation of Aquilos cryo-FIB milling, we can now fully exploit the benefits of the

hardware improvements and can indeed process up to 28 lamellae per session. In addition,

the supervised automated milling process allows users to explore the influence of different

milling parameters and strategies in a systematic manner. Due to the higher throughput,

quality assessment routines can be developed and once these routines are established, the

influence of different milling parameters can more readily be investigated, for example the

influence of high ion currents (2-3 nA) on the quality of the cryo-lamella, which is not yet fully

understood.

Taken together, the combination of new hard- and software increases the overall

throughput of lamellae production, improves the quality of the lamellae, and simplifies the

process of cryo-FIB milling and sample transfer. In addition, the methods described in this

work enable new types of experiments, which were previously difficult to realize.

## Acknowledgements

282  We thank Roger Wepf for many fruitful discussions and the exchange of ideas. We also thank

283  Sagar Khavnekar and Florian Beck for help with subtomogram averaging. This work was

284  supported by funding from the Max Planck Society (to S.R. and J.P.). M.G. is supported by an

285  EMBO Long-Term Fellowship.

## Material and Methods

**Glove box, preparation station and cryo high vacuum transfer**

The glove box has been manufactured by Microscopy Solutions Py. Ltd and modified by us

accordingly. Before sample preparation, the glove box is purged overnight with dry nitrogen.

Before the sample is brought into the glove box through the load lock, the liquid nitrogen

reservoir and the preparation station are filled with liquid nitrogen. Then, EM grids are clipped

to the Autogrids, if not already done. Thereafter, the Autogrids are loaded into the cryo-FIB

shuttle and transferred to the Aquilos microscope via the cryo high vacuum transfer system.

After the milling session has started, both the glove box as well as the preparation station are

baked out at 40-50 °C with the help of the installed heaters. After milling, the cryo-FIB shuttle

with the sample is transferred back to the glove box and the autoloader grids are unloaded,

rotated by 90° and mounted to the autoloader cassette. Finally, the cassette is loaded into the

NanoCab (Thermo Fisher Scientific) and transferred to the electron cryo microscope (Titan

Krios, Thermo Fisher Scientific).

**Cryo-FIB/SEM**

The developments described in this study were carried out on an Aquilos cryo-FIB/SEM

(Thermo Fisher Scientific), equipped with an additional ion getter pump (Agilent), a larger

cryo-shield (manufactured by CryoVac GmbH & Co KG) and stage heater (in-house

production). The instrument is also equipped with a gas injection system (GIS) and a platinum

sputter coater. After insertion, the grids are coated with an organometallic platinum layer to

protect the sample while milling. The sputter coating is utilized to avoid charging effects

during imaging in the SEM and can either be done before or after GIS deposition. For scanning

electron imaging, an acceleration voltage of 2-5 kV with an electron current of max. 28 pA was

used. The ion optics were operated at 30 kV acceleration voltage with different ion currents

(50 – 700 pA). All ion currents are specified by the user and should be tuned for every new

sample material. The alignment of the ion optics is crucial and was performed at least once a

week to avoid massive image shifts while changing the ion currents. For scanning electron

imaging, the installed Everhart-Thonley detector was utilized.

**Automated milling via AutoTEM Cryo**

Although the relevant milling parameter as ion beam currents, milling patterns and milling

time might vary for different types of sample, a brief overview is given:

After the identification of ROIs, there are two possible workflows: 1) The eucentric

height and the milling angle are set for each lamella site using the MAPS application (Thermo

Fisher Scientific). Typically, this preparation step takes approximately 5 min for each lamella

site. Alternatively, 2), After setting the ROI positions in MAPS and then transferring them to

AutoTEM Cryo (Thermo Fisher Scientific), the eucentric position, tilts and beam shift values can be set in AutoTEM using the preparation step. This can take around 2-3 minutes per site and has the benefit of storing the beam shifts.

In the first milling step, the exact lamella positions were defined and relevant milling parameters were set. The lamella width is mostly defined by the type of sample material and the condition of the specific ROI. Lamella widths of 8 – 10 µm were typically used. The desired lamella thickness was project dependent: in some cases, the project required a lamella thickness below 150 nm and in other cases, a thickness of 250 nm was still within the required limits. For the first estimation of milling times, a total sample thickness of 15 µm was assumed.

In the first milling step, the site preparation, relief cuts were milled to release the stress from the support film of the grid, which might occur during plunge freezing. Here, a width of 1-3 µm and a height of 10-15 µm were typically used. The relief cuts were milled at relatively high ion beam currents (0.5–1 nA) for approximately 5 min.

After site preparation, most of the bulk material was removed during rough milling. This task is separated into four different milling steps. During the first milling step, most of the sample material is ablated with relatively high ion beam currents (0.5–1 nA) until a lamella thickness of approximately 2-3 µm is reached. In the following milling steps, the ion currents were consecutively reduced from 0.5 to 0.1 nA. After these milling steps, the lamellae thickness was already reduced to 300-600 nm. For the initial rough milling the front and rear pattern height were set to 5-7 µm. All later milling pattern were calculated accordingly. Since most of the sample material was removed during rough milling, most of the time was spent during this step (20-30min).

The polishing step was executed in up to five successive milling steps with an ion beam current of 30-50 pA. Since most of the sample material was already ablated, the last thinning process took only approximately 5-10 min.

**Cryo-electron tomography and tomogram reconstruction**

All data sets were collected on a Titan Krios (Thermo Fisher), operated at an acceleration voltage of 300kV. The microscope is equipped with a field emission gun, a BioQuantum imaging filter (Gatan, Munich, Germany) with a set energy width of 20eV and a K3 direct electron detector (Gatan, Munich, Germany). The software package SerialEM was used for automated data acquisition [Mastronade2005]. For reconstruction of the tomograms, the

software package IMOD was used [Kremer1996]. Imaging conditions are listed in

Supplementary Table 2.

**Template Matching, Polysome Detection and Visualization**

Tomograms were binned (IMOD bin 8x, 14.32 Å pix$^{-1}$) and template matching was performed

using PyTom [Hrabe2012]. First a reference was constructed from approximately 300

manually picked ribosomes, which were aligned in PyTom using the fast-rotational matching

(FRM) algorithm. The reference obtained from this was then truncated at the small subunit to

allow later removal of false positives. For each tomogram, the 800 highest-scoring cross-

correlation peaks were extracted, and the subtomograms aligned and classified using a

maximum likelihood approach in Relion 3 [Zivanov2018] to remove false positives. This was

judged by the absent small subunit in the false positive class averages. After initial processing

at bin 8x, further classification and refinement of the 2x binned (3.58 Å pix$^{-1}$) subtomogram

averages was performed in Relion, including normalization and CTF estimation with CTFFIND

4.1.5 [Rohou2015]. The final list of particles (approximately 15000) yielded a ribosome

structure at 14 Å resolution (0.143 FSC criterion), which was used in all animations. Polysome

detection was performed analogously to published using an inhouse Matlab script

[Brandt2010].

**Figures**

**Figure 1: Possible artefacts arising during handling, transfer and processing.** a) This overview shows a rare case, were all possible artifacts were found on a single lamella. If the sample material is transferred in a hydrous environment, ice crystals of different sizes (highlighted by green dashed lines) can contaminate the lamella. Moreover, the fragile lamella can be easily damaged during handling, and in worst case, cracks across the lamella can appear (orange solid lines). Obviously, data acquisition at the region-of-interest (orange triangles) is impossible in such cases. Scale bar, 1 µm. b) Even if these artefacts can be avoided, intact lamellae can be corrupted by contaminations occurring during milling. The water fraction of the residual gas inside the vacuum chamber deposits by time onto the lamella (green dashed line). As indicated (white arrows), the thickness of the polished lamella was originally 180 nm but due to the water deposition inside the cryo-FIB/SEM, it increased to approximately 280 nm, reducing the quality of the lamella. The water deposition can be identified unambiguously since the lamella was coated with platinum (green triangle) in the end of the milling procedure. Scale bar, 100 nm.

**a**

- IN₂ reservoir → $IN_2$ reservoir
- Port for cryo high-vacuum transfer
- Gloves (not shown)
- Heated tool holder
- Load lock
- Heated tool holder
- Preparation station

**b**

- Shuttle loading module
- Tool holders
- Adapter NanoCab
- Transfer Module
- Cassette loading module
- Rail system
- Adapter cryo high-vacuum transfer
- Cryo-FIB shuttle
- Preparation modules

**c**

- Transfer rod
- Parking position
- Vacuum window
- IN₂ reservoir → $IN_2$ reservoir
- Valve
- Vacuum chamber
- Adapter for cryo-FIB/SEM & Glove box

**Figure 2: Schematic of the glove box, the preparation station, and the high vacuum cryo transfer system.** a) The glove box was designed to enable convenient sample handling in an anhydrous environment. To reduce the humidity to a minimum, the glove box is purged during operation with nitrogen and a cold trap is installed in the back of the $lN_2$ reservoir. $lN_2$ can be taken from this reservoir for applications inside the glove box. For convenient handling, heated tool holders are installed. A load lock enables the insertion of tools without breaking the humidity barrier. b) To guarantee a comfortable and secure handling of the sample material, a new preparation station was developed. The preparation station consists of four parts: the liquid nitrogen bath, the preparation modules, the adapter for the transfer systems and the adapter for the NanoCab. All modules can be re-arranged inside the liquid nitrogen bath according to the preferences of the operator and are designed for left- and right-handed persons. Clipped autoloader C-clip rings can be loaded onto the cryo-FIB shuttle or to the autoloader cassette using the corresponding modules. The preparation station is mounted onto a rail system to ensure highest flexibility inside the glove box. c) The high vacuum cryo transfer system consists of a transfer rod for picking up the cryo-FIB shuttle, a vacuum chamber, liquid nitrogen reservoir, a valve and the adapter for the cryo-FIB/SEM and glove box. The sample can be picked up by the transfer rod and is inserted in the vacuum chamber. Inside the vacuum chamber the cryo-FIB shuttle is placed in a parking position. Here, the shuttle is actively cooled by the $lN_2$ reservoir. The liquid nitrogen acts also as a cryo-pump and maintains the vacuum level, while the high vacuum cryo transfer is disconnected from the pumps. The $lN_2$ volume holds for approximately 30 min. After the shuttle is positioned in the parking position, the valve towards the glove box is closed and the high vacuum cryo transfer is evacuated. Thereafter, the valve at the transfer system is closed and the sample can be transferred to the cryo-FIB/SEM.

**Figure 3: Schematic of the cryo-shield and cryo-shutter**. The total surface area of the cryo-
shield is approximately 880 cm$^2$. Assuming a pumping speed of 15 l cm$^{-2}$ s$^{-1}$ [Bello2017], the
total pumping speed can be estimated to be 13,000 l s$^{-1}$. To reduce the partial pressure in the
vicinity of the sample even further, a cryo-shutter can be inserted during milling (blue arrow).
The cryo-shutter covers most of the grid and in combination with the installed stage heater
(not shown), the contamination rate decreases to 5.7 nm h$^{-1}$. With these improvements, the
sample can be prepared and stored in the milling position (orange arrows) until imaging
without major water deposition.

**Figure 4: AutoTEM pipeline for automated lamellae preparation.** a) After the lamella site is centered, micro-expansion joints (blue frame) are milled next to the lamella site to release possible stress of the support layer. Scale bar, 10 µm. b) Thereafter, rough milling (blue frame) is performed in four consecutive steps. First, the bulk material is ablated with relatively high currents (0.5 – 1 nA). In the subsequent steps, the lamella is thinned to a thickness between 250 – 500 nm with an ion current of 0.5 to 0.1 nA. Scale bar, 10 µm. c) In the last step, the lamella is polished (blue frame) with low ion beam currents (30 – 50 pA). Scale bar, 10 µm. In order to correct for possible stage drifts, specific regions of the ROI (green frames in a), b) and c)) are imaged frequently and compared by a newly developed, cross-correlation based algorithm. d) The final lamella thickness is optimally about 100 to 200 nm. Scale bar, 10 µm. e) Overview of the different lamella sites, prepared during the same session (dashed orange circles). Scale bar, 50 µm (Inset scale bar: 10 µm).

**Figure 5: Quality assessment of the pre- and post-processing step.** a) Overview of a lamella. The lamella is only covered with small ice crystals (green circles) which do not prevent data acquisition in these areas. Areas of the residual platinum layer can be identified (blue dashed line). Although these areas should not be used for data acquisition, the useable area was large enough to take six tomograms. In some areas, non-vitrified regions can be identified (green dashed eclipse) but within the region of interest (orange triangles), no crystalline ice can be found. Scale bar: 1 µm. b) Orthogonal slice through the tomogram shown in a). The average thickness is approximately 280 nm (orange dashed line). The lamella was stored approximately 30 min in the cryo-FIB/SEM prior to imaging. The green dashed line indicates the theoretical contamination layer, which would have deposit on the lamella without the introduced hardware modifications. As indicated by the orange circle (see also orange circle in a)), the ice contamination is in direct proximity to the lamella. This shows that the contamination rate is reduced. The slice was Gauss filtered and averaged (6 slices). Scale bar: 300 nm.

455
**Figure 6: Suitability of automatically milled lamellae for subtomogram averaging.** a)
Representative overview of an automatically milled lamella. The lamella was sputter coated

to reduce charging artefacts. Scale bar: 2.5 μm. b) Macromolecular complexes such as
ribosomes are readily detectable. Scale bar: 100 nm. c) Subtomograms can be averaged to
high resolution (14 Å at 0.143 FSC criterion). d) Additionally, orientations and locations can be
precisely determined, which enabled the detection of polysomes up to a pentamer (insert and
boxed region P in a)).

| Sample | Number of ROI (lamellae sites) | Number of successfully roughly milled lamellae | Number of successfully polished lamellae |
|---|---|---|---|
| Myofibrils | 21 | 20 | 10 |
| A9 cells | 7 | 6 | 6 |
| *S. cerevisiae* | 94 | 89 | 85 |
| *E. coli* | 23 | 23 | 21 |
| Neurons | 3 | 3 | 2 |
| HeLa cells | 6 | 6 | 6 |
| Sum | 154 | 147 | 130 |
| **Success rate** | | **95%** | **88%** |

466  **Table 1: Milling statistics for different types of sample material using the automated**

467  **lamella production pipeline.**

| Step in workflow | Possible contamination | Implication | Current technology/ limitations | New technology/ improvements |
|---|---|---|---|---|
| **Sample Vitrification** | Ice particles | Lack of visibility of sample surface | Air-dehumidified rooms | Not addressed |
| **Cryo-FIB/SEM Loading** | Ice particles and condensation | Lack of visibility and curtaining | Aquilos loading station:<br><br>No regulated environment<br>Limited runtime (~ 1 h) | Glove Box, preparation station, high vacuum cryo transfer:<br>Anhydrous environment<br>Extended runtime (> 3 h) |
| **Cryo-FIB/SEM Milling** | Condensation | Increased thickness of lamella | Cryo-shielding:<br><br>Contamination rate ~ 50 nm h$^{-1}$<br>Limit of throughput | Extra shielding, cryo-shutter, stage heater:<br>Contamination rate ~ 6 nm h$^{-1}$<br>High throughput |
| **Cryo-FIB/SEM Unloading** | Ice particles and condensation | Increased thickness of lamella | Aquilos loading station:<br><br>No regulated environment<br>Limited runtime (~ 1 h)<br>Cassette loading not possible | Glove Box, preparation station, high vacuum cryo transfer:<br>Anhydrous environment<br>Extended runtime (> 3 h)<br>Cassette loading possible |
| **Sample Storage** | Ice particles | Loss of viewable areas | Grid box and Dewar | Not addressed |

**Table 2: Overview of the improvements of the cryo-FIB/SEM workflow**

**References**

[Grange2017] Grange, M, Vasishtan, D., & Grünewald, K. Cellular electron cryo tomography

and in situ sub-volume averaging reveal the context of microtubule-based processes. *J. Struct.*

*Biol.* **197**, 181-190 (2017).

[Bartesaghi2012] Bartesaghi A. et al. Secondary Structure Determination by Constrained

Single-Particle Cryo-Electron Tomography. *Structure* **20**, 2003-2013 (2012).

[Schur2016] Schur, F. K. M. et al. An atomic model of HIV-1 capsid-SP1 reveals structures

regulating assembly and maturation. *Science* **353**, 506-508 (2016).

[Vulovic2013] Vulovic, M. et al. Image formation modelling in cryo-electron microscopy. *J.*

*Struct. Biol.* **183**, 19-32 (2013).

[Al-Amoudi2004] Al-Alamoudi, A., Norlen, L.P.O., & Dubochet J. Cryo-electron microscopy of

vitreous sections of native biological cells and tissues. *J. Struct. Biol.* **148**, 131-135 (2004)

[Al-Amoudi2005] Al-Alamoudi, A., Studer, D. & Dubochet, J. Cutting artefacts and cutting

process in vitreous sections for cryo-electron microscopy. *J. Struct. Biol.* **150**, 109-121 (2005).

[Marko2007] Marko, M. et al. Focused-ion-beam thinning of frozen-hydrated biological

specimens for cryo-electron microscopy. *Nat. Methods* **4**, 215-217 (2007).

[Rigort2010] Rigort, A. et al. Micromachining tools and correlative approaches for cellular

cryo-electron tomography. *J. Struct. Biol.* **172**, 169-179 (2010).

[Hagen2015] Hagen C. et al. Structural basis of vesicle formation at the inner nuclear

membrane. *Cell* **163** 1692-1701 (2015).

[Mahamid2016] Mahamid J. et al Visualizing the molecular sociology at the HeLa cell nuclear

periphery. *Science* **351**, 969–972 (2016).

[Medeiros2018] Medeiros, J.M. et al. Robust workflow and instrumentation for cryo-focused

ion beam milling of samples for electron cryotomography. *Ultramicroscopy* **190**, 1-11 (2018).

[Tacke016] Tacke, S. et al. A Versatile High-Vacuum Cryo-transfer System for Cryo-microscopy

and Analytics. *Biophysical Journal* **110** 4, 758-765 (2016).

[Echlin1992] Echling, P. Low-Temperature Microscopy and Analysis. Plenum Publishing

Corporation, New York (1992)

[Umrath1983] Umrath, W.W. Berechnung von Gefriertrocknungszeiten für die

elektronenmikroskopische Präparation. *Mikroskopie* **40**, 9-34 (1983).

[Wolff2019] Wolff, G. Mind the gap: Micro-expansion joints drastically decrease the bending

of FIB-milled cryo-lamellae. *J. Struct. Biol.* In Press (2019).

[Schaffer2017] Schaffer, M. et al. Optimized cryo-focused ion beam sample preparation aimed

at in situ structural studies of membrane proteins. *J. Struct. Biol.* **197**, 73-82 (2017).

[Mastronade2005] Mastronarde D.N. Automated electron microscope tomography using

robust prediction of specimen movements. *J. Struct. Biol.* 152, 36–51 (2005).

[Hrabe2012] Hrabe, T. et al. PyTom: a python-based toolbox for localization of

macromolecules in cryo-electron tomograms and subtomogram analysis. *J. Struct. Biol.* **178**,

177-188 (2012).

[Zivanov2018] Zivanov, J. et al. New tools for automated high-resolution cryo-EM structure

determination in RELION-3. *eLife* **7**, e42166 (2018).

[Rohou2015] Rohou, A. and N. Grigorieff, CTFFIND4: Fast and accurate defocus estimation

from electron micrographs. *J. Struct. Biol.* **192**, 216-221 (2015).

[Brandt2010] Brandt, F. et al. The Three-Dimensional Organization of Polyribosomes in Intact

Human Cells. *Mol. Cell* **39**, 560-569 (2010).

[Kremer1996] Kremer, J.R., Mastronarde, D.N. & McIntosh, J.R. Computer visualization of

three-dimensional image data using IMOD. *J. Struct. Biol.* **116**, 1-6 (1996).

[Bello2017] Bello, I. Vacuum and Ultravacuum: Physics and Technology. CRC Press (2017)

**Supplementary Figure 1: Phase diagram of water.** a) Phase diagram of water and the associated mass flux (deposition/sublimation). The dashed line indicates the devitrification temperature of water. The black asterisk marks the partial pressure value of the cryo-FIB/SEM before the hardware modifications. The arrows indicate the used strategy to minimize the ice deposition onto the sample. By the installation of the cryo-shield, the partial pressure of water dropped to $3 \times 10^{-9}$ mbar (black arrow 1) and the contamination rate decreased to (72 +/- 4) nm h$^{-1}$. If the cryo-shutter is inserted, the partial pressure in the vicinity of the sample decreases by approximately one order of magnitude (black arrow 2), which reduces the contamination rate to (12.8 +/- 1.6) nm h$^{-1}$. To reduce the contamination rate even more, the sample is heated to approximately -150°C (black arrow 3). Herewith, the contamination rate is reduced to (5.6 +/- 1.0) nm h$^{-1}$ (blue asterisk). b) Partial pressure of water inside the cryo-FIB/SEM chamber measured during operation (orange graph). As soon as the cryo-shield is cooled (black arrow), water molecules from the residual gas get trapped and the partial pressure decreases. After 30 min, the system reaches its final pressure value (solid green arrow). The impact of the cryo-shield is indicated by the green dashed arrow. At this time point, the cryo-stage reached already -185°C and the cryo-shield temperature was still above -100°C, where the pumping speed of the cryo-shield is negligible. As indicated by the blue area, this value matches with the partial pressure regime of standard cryo-FIB/SEM instruments.

**Supplementary Figure 2: Contamination rate measurement in the milling position.** a) Before
imaging, the stage, cryo-shield and the cryo-shutter were cooled. After reaching the final
temperature of approximately -185°C, the stage was heated to -165°C. Since there might be a
temperature offset of 15°C between stage and shuttle, we assume that the sample has a
temperature of -150°C. Whereas the cryo-shield reached a temperature of -185°C, the cryo-
shutter could only be cooled to -177°C. To determine the contamination rate in the milling
position, a holey gold EM grid (UltraAuFoil holey gold grid) was mounted to the cryo-FIB/SEM
shuttle and loaded into the instrument. For the baseline measurements (t$_1$) the thickness of
the holey gold film was determined (blue frame). To avoid ice ablation during focusing, beam
tuning was done next to the region-of-interest (black frame). b) After 5.4 hours (t$_2$), another
series of thickness measurement was performed on the same holes. c) Baseline
measurements: Line profiles were averaged horizontally (blue arrows in a)). The upper and
the lower edge of the profile (dotted black line) were fitted with Gaussian function. If fitting
was not possible, edge positions were determined manually. The centers of the edges were
used to calculate the thickness. d) Thickness determination after 5.4 h: The line profile was
averaged and the thickness was determined as described in c). Finally, both values were
corrected regarding the tilting angle of the grid and the total increase in thickness was
determined. Herewith, the contamination rate can be calculated. Scale bars, 2 µm.

a

b

**Supplementary Figure 3: Software applications.** a) Overview of the MAPS software application. Either ROIs are identified by light cryo microscopy and uploaded in the MAPS software or identified directly in the overview image as shown here. b) Overview of the AutoTEM Cryo software application. The software pipeline is separated into three different tasks (orange frame): i) Site preparation, ii) rough milling iii) thinning. Each step consists of several sub-tasks (green frame) which can be tuned to guarantee an optimal performance for each lamella site (blue frame).

| Stage temperature [°C] | Shield temperature [°C] | Status shutter | Contamination rate [nm h$^{-1}$] (n) |
|---|---|---|---|
| -188 | RT | Not inserted | 85 ± 7 (15) |
| -188 | -177 | Not inserted | 72 ± 4 (8) |
| -188 | -177 | Inserted | 12.8 ± 1.6 (15) |
| -165 | -177 | Inserted | 5.6 ± 1.0 (15) |

**Supplementary Table 1: Influence of the different hardware components on the contamination rate inside the cryo-FIB/SEM.** The number of holes, which were used for thickness determination (sample size), is given by the number n. For each condition, two to three positions, containing at least four holes, were selected. All measurements of one condition were performed on the same day. The uncertainty is estimated by the standard deviation of the different conditions.

| Sample | Defocus [µm] | Tilting range [°] | Increments [°] | Pixel size [Å] | Total Dose [e$^-$/Å] |
|---|---|---|---|---|---|
| **Myofibrils overview** | 20 | -29 to 39 dose symmetric | 2 | 11.3 | 17 |
| *E. coli* **overview** | 50 | 0 | 0 | 22.8 | 10 |
| *E. coli* **tomograms** | 1.5-3.5 | ± 60 dose symmetric | 3 | 1.79 | 120 |

**Supplementary Table 2: Imaging conditions for the presented data.** The imaging conditions for the two tomogram slices (shown in Figure 5a,b and Figure 6b) and the TEM image (shown in Figure 6a) are listed.