Scalable method for micro-CT analysis enables large scale quantitative characterization of brain lesions and implants

David B. Kastner[1,2*], Viktor Kharazia[2*], Rhino Nevers[2], Daniela A. Astudillo-Maya[2], Greer M. Williams[1], Zhounan Yang[1], Clay Smyth[2], Luca Della Santina[3], Dilworth Y. Parkinson[4], Loren M. Frank[2,5]

[1]*Department of Psychiatry, University of California, San Francisco, CA 94143, USA*

[2]*Kavli Institute for Fundamental Neuroscience and Department of Physiology, University of California, San Francisco, CA 94158, USA*

[3]*Deparment of Ophthalmology, University of California, San Francisco, CA 94143, USA*

[4]*Advanced Light Sources Division, Lawrence Berkeley National Labs, Berkeley, CA 94720, USA*

[5]*Howard Hughes Medical Institute*

[*]*These authors contributed equally to this work*

Corresponding author: David B. Kastner, david.kastner2@ucsf.edu

Abstract

Anatomic evaluation is an important aspect of many studies in neuroscience; however, it often lacks information about the three-dimensional structure of the brain. Micro-CT imaging provides an excellent, nondestructive, method for the evaluation of brain structure, but current applications to neurophysiological or lesions studies require removal of the skull and hazardous chemicals, dehydration or embedding, limiting their scalability and utility. Here we present a protocol using eosin in combination with bone decalcification to enhance contrast in the tissue and then employ monochromatic and propagation phase-contrast micro-CT imaging to enable the imaging of brain structure with the preservation of the surrounding skull. Instead of relying on descriptive, time-consuming, or subjective methods, we develop simple quantitative analyses to map the locations of recording electrodes and to characterize the presence and extent of brain lesions.

Introduction

Evaluating brain structure is a critical confirmatory step for much of systems neuroscience. This often occurs through histological sectioning and light-microscopic analysis. Although such a technique provides a high-resolution view of brain structures, it is time-consuming, difficult to scale for high-throughput analyses, and, perhaps most problematic, it destroys the three-dimensional structure of the brain. Additionally, histological processing is often accompanied by descriptive analyses and not quantitative evaluation. Newer tissue clearing methods[1,2] maintain the three-dimensional structure of the brain for high resolution imaging, but remain expensive, labor intensive, and not well suited to evaluate lesions or electrode localization, mainstays of systems neuroscientific inquiry. Micro-CT presents an excellent middle ground for evaluating large-scale three-dimensional brain anatomy while maintaining the original structure[3-9], providing an excellent substrate for subsequent quantitative analyses[10-15].

The majority of studies that visualize neuroanatomy using micro-CT have not evaluated the technique for quantifying brain lesions and electrode placement, instead focusing on tumors[4,5,16], cavernous malformations[10,17], disease pathology[15,18-20] and cellular and subcellular visualization[11-13,21,22]. A recent protocol was developed for micro-CT imaging of lesions and electrode placement[8] providing an elegant proof of principle of the use of micro-CT for systems neuroscience. However, that protocol, although clear and well described, contains many involved steps, some of which require expensive and hazardous elements, including osmium. It also removed the skull prior to imaging, leaving out valuable information and removing a structural support for neural hardware. Furthermore, in using osmium, it becomes far more challenging to process the brain for subsequent standard histologic techniques, requiring a decision to be made between either imaging the full three-dimensional volume, or using standard histologic processing.

Phase contrast micro-CT has proven capable of solving all of the above-mentioned problems. Utilizing phase contrast techniques to process micro-CT images generates sufficient contrast in the brain[3,6,7,9,15,23] without, heavy metals or complex protocols. In addition, brain structures can be resolved with the overlying skull intact[9], histology can be performed on the same samples following the CT[5,11,14-16,20,21,24], and laborious embedding of the tissue is not necessary[5]. However, phase contrast processing, to the best of our knowledge, has not been applied to brains with recording electrodes in place, nor has it been used to evaluate brains

lesioned for behavioral studies. Moreover, quantitative approaches to electrode localization or
lesion quantification using phase contrast processing have not been presented.

Here we present a scalable and inexpensive protocol utilizing monochromatic X-rays and
propagation phase contrast to image and quantify brain lesions and electrode locations within the
brain. We adapt decalcification techniques[25,26] to image the brain with critical parts of the skull
intact. Furthermore, we avoid any dehydration or embedding of the tissue, and enhance contrast
to the post-mortem brain using eosin as an additional contrast agent[27,28]. The resulting protocol
allows for evaluation of brain structure relative to the coordinates determined by skull sutures
and permits further processing with standard histologic techniques.

68 Material and Methods

Animals: All experiments were conducted in accordance with University of California, San
Francisco Institutional Animal Care and Use Committee and US National Institutes of Health
guidelines. Datasets were collected from male Long-Evans rats that were fed standard rat chow
(LabDiet 5001). Long-Evans rats used for the lesioning were either of a wild-type genotype or of
a *Fmr1* KO genotype[29] and were bred at Medical College of Wisconsin, the remaining rats were
purchased from Charles River. We did not observe any gross anatomical differences between the
genotypes; therefore, we combined all brains for the analysis of the lesions.

Surgery: Anesthesia was induced using ketamine, xylazine, atropine, and isoflurane. For the
animals that underwent surgery, they were either implanted with a multielectrode microdrive, or
underwent lesioning of the dorsal and intermediate hippocampus. For the microdrive implant,
craniotomies and accompanying durotomies were placed at +2.6 ML and -4.0 AP for the dorsal
hippocampal canula and +4.7 ML and -5.7 AP for the intermediate hippocampal canula. Values
are relative to bregma. Screws were imbedded into the skull to support the acrylic casing that
secured the microdrive onto the skull of the rat. For the lesion rats, craniotomies were drilled
bilaterally to allow for access to all 24 injection sites (Table 1) and the injection needle (33
gauge) was inserted through the dura. For each lesion site, the needle was lowered 0.1 mm below
the target coordinates and then raised back to the target prior to injection. 120 nl of NMDA (10
mg/ml) was injected at each site at a rate of 0.1 µl/min. The needle was left at the site for 2
minutes after finishing the injection. For the rats that underwent control surgeries, everything
was the same as the lesion rats, except no NMDA was injected, and the needle did not dwell at
the hypothetical injection site. The animals that underwent the lesion surgery (either with or
without injection of the NMDA) received Diazepam 10 mg/kg intraperitoneally post-operatively.

Electrophysiological recording: Tetrodes were individually moved toward the hippocampus
daily. Physiological data, digitized at 30 kHz, and was recorded using the SpikeGadgets data
acquisition system and the Trodes software package (SpikeGadgets).

Processing for micro-CT: Prior to perfusion, animals were anesthetized with isoflurane, and
then euthanized with pentobarbital. Animals were perfused intracardially with 100 mL of
phosphate buffer solution (PBS), followed by 250 mL of chilled 4% paraformaldehyde in PBS.
The skull was dissected out with the brain intact, with most of the facial bones trimmed off, and
all soft tissues removed. A 5-mm-wide opening was made using a rongeur on the ventral side
along the midline (optional), for improved diffusion of fixative. The sample was then placed in
50mL of the same fixative for 48 hours at 4°C.

Subsequently, the sample was rinsed in PBS with 0.05% sodium azide (PBS-azide) and
decalcified by placing it in a container with 250 mL of 8% EDTA (Sigma-Aldrich) for 6-7 days
at room temperature on an orbital shaker. The solution was replaced once midway through the
decalcification.

After rinsing again in PBS-azide, the decalcified skull was further trimmed to a final size
on both anterior and posterior sides using a razor blade. Care was taken to avoid compressing or
otherwise distorting the brain. The trimmed sample was placed in 50mL of 5% Eosin Y (Sigma-
Aldrich) for 6-7 days at 4°C on an orbital shaker. The solution was replaced once after 48 hours.
The sample was then rinsed in PBS and stored in 50mL PBS-azide at 4°C prior to micro-CT
scanning.

For Nissl staining, the brain (with skull removed) was cut using a Vibratome 3000 (Leica
Biosystems); 50-80 um thick coronal sections were collected into a 24-well dish with PBS-azide
and stored at 4°C. Sections were mounted on adhesive slides, air-dried, and stained using a
traditional cresyl violet staining protocol.

We confirmed the benefits of the eosin staining by processing and imaging a brain
without eosin and found that the eosin provided substantial improvements in contrast (data not
shown).

Micro-CT scanning and image reconstruction: Scanning was done at beamline 8.3.2 (hard X-
ray micro-tomography) at the Advanced Light Source (ALS) at Lawrence Berkeley National
Laboratory (Fig S1). All scans were carried out at room temperature.

For the brain with the implanted microdrive, the brain anterior and posterior to the
hippocampus were removed to decrease the width of the brain to minimize the field of view for
the imaging. Furthermore, as many as possible of the skull screws were removed to minimize
interference. As the microdrive was still attached to the skull via acrylic and remaining skull
screws, it provided the support for the brain. Consistent with a previous protocol[5] the brain was
out of solution for the scanning. X-rays passed through the sample along the A-P/M-L axes as
the sample rotated.

For the rest of the brains, the sample was placed snuggly in a conical tube and was not in
solution (Fig S1), consistent with a prior protocol[5]. X-rays passed through the sample along the
M-L/D-V axes as the sample rotated. This is the narrowest dimension of the brain, thereby
allowing for the minimal field of view and finest pixel size (see Table 2), as limited by the
number of pixels on the available detector at the beamline that can accommodate the full field of
view of the brain.
Monochromatic X-rays were generated using a multilayer mirror monochromator. Energy
of 25 keV was used. The flux is approximately $\sim 10^5$ hv/sec/μm^2. The source size is $\sim$220 x 30
μm full width at half maximum, and the maximum beam size at the sample is $\sim$40mm x 4.6 mm.
The sample was placed 54 cm from the scintillator (Fig S1) to enable the maximal usage of
propagation phase contrast imaging available at the beamline (because of its small hutch, this is
the maximal sample to scintillator distance for this beamline). For each scan, a total of 1313
projection angles with an exposure time of 100 – 125 ms each were used (Table 2). Detection
was accomplished with a 0.5 mm thick YAG:Ce scintillator, a Nikon zoom lens, and a
PCO.Edge sCMOS detector. The zoom was set to resolve 7 μm lines and spaces (14 μm period)
on a test pattern in 2 dimensions. We did not measure precisely the 3-dimensional resolution as it
was not necessary for the conclusions of this study, but our results are consistent with an

approximately 14 μm isotropic resolution. Radiation dose was not calculated for this study as we
did not detect any detrimental effects of the scanning at the resolution used (data not shown).
147   The images were reconstructed using tomopy using a script designed at the ALS
(https://bitbucket.org/berkeleylab/als-microct-python/src/master/tomopy832/tomopy832.py). See
Table 2 for the parameters used for the reconstruction of the various brains. Multiple tiles were
manually stitched together to create the final imaged volume (Table 2). It is the stitching together
of the multiple tiles that causes the striped patterning see in (Fig 1 & Vid 1). Phase retrieval was
accomplished using the standard Paganin method[30], as implemented in tomopy.

Image visualization and analysis: Fiji[31-33] was used to visualize virtual slices of the brain and to
create a mask of the hippocampal region for the lesion analysis using the Segmentation Editor
tools. Imaris (Oxford Instruments) was used for three-dimensional reconstruction of the brain.
Slicer-3D 4.10.2 (https://www.slicer.org/)[34,35] was used to register and align the brains
(http://hdl.handle.net/1926/1291) by finding the linear transformation (rotation and translation)
to place the entirety of the imaged brains into a shared coordinate system. The alignment of the
brains was not restricted to align any specific region of the brain (such as just the hippocampus)
and rather worked to best align the whole brain and skull.

161   For the lesion analysis we were interested in understanding the similarity of the
hippocampal formation between brains as that was the target of the lesioning. Therefore, we
created a mask for the left and right hippocampal region. The mask was intentionally generous to
ensure that the entire hippocampal formation was included for all brains. For each virtual slice
we calculated the similarity between all pairs of brains using the structural similarity metric
(SSIM)[36,37]. For SSIM we used a gaussian smoothing filter for the images with a sigma of 50
pixels. This allowed for better comparison between the brains due to inherent differences in the
exact positioning of the brain structures. A dendrogram that grouped lesioned and control brains
was calculated using single-linkage hierarchical clustering over the average SSIM for all pairs of
brains. The similarity matrix was subtracted from 1 to create a dissimilarity matrix as the input to
the hierarchical clustering.

Theta phase offset calculation: We also calculated the phase of the endogenous ~8 Hz theta
rhythm at each electrode so that differences in phase across electrodes could be related to
electrode locations. The local field potential (LFP) was defined as the voltage from each tetrode
after being low-pass filtered below 400 Hz and down-sampled to 1.5 kHz. To extract theta
oscillations, the LFP data were band-pass filtered with a Parks-McClellan optimal equiripple FIR
filter (pass band between 4 and 10 Hz, stop bands below 3 Hz and above 11 Hz) applied in both
the forward and reverse directions to prevent phase distortion[38]. Instantaneous theta phase was
calculated as the angle of the Hilbert-transformed theta oscillation.

180   Analysis of theta phase was limited to time periods when the rat was running, which was
defined as a body speed greater than 10 cm/s[38]. Instantaneous phase differences between pairs of
tetrodes were measured as the angle between the complex phase vectors for each tetrode. For
each running period, the average pairwise phase difference was calculated as the circular mean
over time of the angle between phase vectors, and the average phase offset over the entire
behavior session was calculated as the circular mean of individual average phase differences for
each running period.

Results

Brain structure with intact skull

189
We sought to develop a protocol that does not require osmication, dehydration of the
tissue, nor embedding in resin or paraffin, but could still provide sufficient contrast to image
brain structures in the presence of electrodes or lesions. We wanted to avoid osmication and
other heavy metals as they can add laborious steps to the processing of the brain and can often
add significant expense. We wanted to avoid dehydration of the tissue and embedding in resin or
paraffin as these techniques often lead to substantial shrinkage of tissue[39,40]. We therefore
adapted protocols using eosin[27,28] to provide enhanced contrast within the brain, and imaged the
brain using monochromatic X-rays at the Advanced Light Source at Lawrence Berkeley National
Lab.

Typically, the skull is removed to image the brain, as it provides a barrier for X-ray
transmission. However, the skull provides the reference frame for the stereotactic coordinates
utilized to target specific brain structures while the animal is alive. Removing the dorsal skull
therefore discards a piece of information critical for evaluating and refining anatomical targeting.
Furthermore, the skull also provides a rigid structure to prevent alterations to the brain
dimensions due to histological processing. Therefore, in addition to processing the brain with
eosin, we adapted decalcification techniques to leave most of the skull in place but allow for
more transmitted X-rays to better resolve brain anatomy. Leaving the skull intact also allowed
for a direct confirmation that the brain maintained its integrity through the processing and
imaging (see Anderson & Maga[41] for a counterexample showing brain shrinkage relative to the
skull).

With our method we are able to image the whole rat brain with the overlying skull (Fig 1
& Vid 1). The eosin combined with phase contrast processing provides sufficient contrast to the
brain to allow for the visualization of many different brain regions. Given that the imaging
provides a three-dimensional volume, we can virtually section the brain to any given region of
interest along any plane of interest (Fig 1B). Having the skull intact allows for setting the frame
of reference of the brain as one would in surgery by leveling bregma and lambda.

An advantage of using eosin as a contrast agent is that eosin is commonly used stain for
histology, and therefore still allows for classic histologic imaging following the micro-CT.
Therefore, using the same brain, following the micro-CT, we sectioned the brain and stained it
with Cresyl Violet to demonstrate the Nissl substance in the neurons. Nissl-stained brain sections
exhibited well preserved morphology and can be used as anatomic reference to corresponding
micro-CT virtual slices (Fig 2). The Nissl-stained brains also confirmed the preservation of
relevant brain structure through the preparation and imaging steps.

Lesion quantification

223
Lesions provide a useful perturbation for evaluating the necessity of brain regions for
behavior. However, validating a lesion using sectioning, staining and imaging presents a
challenge for quantification, often leading to more descriptive analyses. Therefore, we utilized
our protocol to quantitatively evaluate the presence and extent of hippocampal lesions.

The CT protocol enables clear visualization of the lesioned area of the brain (Fig 3A).
Quantification of the lesion could be done by image segmentation, as was done with previous
micro-CT processing[8]; however, image segmentation is laborious and prone to subjectivity.

Therefore, we sought a more automatable and objective approach for quantifying the presence
and extent of lesions.

We processed and collected micro-CT images of 25 rat brains—12 rats underwent control
surgery and 13 rats underwent hippocampal lesion surgery (see Material and Methods). We
aligned all brains to a common reference and created a mask to extract the three-dimensional
region of the hippocampus across all brains (Fig 3B). We calculated the similarity of all pairs of
brains for each virtual slice using SSIM (Fig 3C), a similarity metric designed to capture the
perceptual similarity between images[36,37]. We clustered the brains using hierarchical clustering
based on the average pairwise similarity matrix across the bilateral dorsal hippocampal formation
and bilateral dorsal-most portion of the intermediate hippocampal formation (the portion of the
hippocampal formation that was lesioned (Table 1)). The lesion and control brains separate into
different groups by the clustering (Fig 3D).

To visualize the extent of the lesion, for all brains and virtual slices we found the most
similar control brain and then plotted those values for the group of control and lesion brains
across the dorsal and intermediate/ventral hippocampus for both the left and right sides of the
brain (Fig 3E). Across the entire extent of the dorsal hippocampus, and the initial part of the
intermediate hippocampus the two groups of animals were clearly separate. Towards the later
part of the intermediate hippocampus and into ventral hippocampus the brains were
indistinguishable, both being similar to control brains. This is consistent with the intended
location of the lesions, since far intermediate and ventral hippocampus were not targeted (Table
1).

Electrode localization

Beyond just providing potentially useful landmarks, the skull also provides a stable
structure to support a microelectrode drive, thereby enabling minimal disruption of electrodes
during processing and scanning. Therefore, we evaluated if we could image functional electrodes
implanted in the brain and assess their location relative to brain structures. Independently
moveable tetrodes were implanted above dorsal and intermediate hippocampus and recordings
were taken as animals explored a spatial environment. Subsequently the animal was sacrificed,
and the skull and brain prepared for imaging with the microdrive remaining on the skull to
minimally disrupt the electrodes. The brain, with skull and microdrive intact, underwent imaging
with micro-CT.

The electrodes were easily visualized, and the resolution was sufficiently fine to
distinguish the four individual electrodes that comprised each tetrode wire (Fig S2A). Although
the hardware used to stabilize the microdrive on the skull degraded the imaging quality
somewhat, there remained sufficient signal to resolve the pyramidal cell layer of the
hippocampus (Fig 4A). All that was required to identify individual electrodes was a visual
comparison of the layout of the canula of the drive and the arrangement of the electrodes from
the micro-CT (Fig 4B).

As with the lesions, here, with the electrodes, the goal is not just to visualize, but to also
extract relevant quantitative information. Again, it would be possible to perform image
segmentation for the hippocampal cell layer, but that would be quite laborious and subjective at
the most critical parts due to the large signal produced by the electrodes. Instead we sought to
extract coordinates that could form the basis of a systematic two-dimensional space on which we
could place the locations of the electrodes. For this analysis, we were most interested in a two-

dimensional space for the anterior-posterior and medial-lateral axis, since the electrodes were
moved over the course of the recording, thereby making the dorsal-ventral coordinates of the CT
not representative of the recording location.

We took advantage of the control brains that we scanned for the lesion cohort (n=12) and
utilized those brains to define a consistent two-dimensional coordinate system. As the origin of
the space we utilize the center of most anterior portion of the hippocampal CA1 cell layer (Fig
S2B). And then as point 1,1 we utilized the lateral extent of the hippocampal (CA1/2/3) cell layer
at the point where the dorsal and ventral portions of the hippocampus fuse to form intermediate
hippocampus (Fig S2B). We chose these points as they were easily visible in all brains and
would be likely visible in an electrode implanted brain as well. Within this space, we then
plotted, for each brain, sample locations of the lateral aspect of the hippocampal cell layer and
the medial aspect of the dentate gyrus cell layer from anterior to posterior (Fig 4C&S2B). Across
all brains the locations were very similar.

We then aligned the brain with the implanted electrodes to the same reference brain as
the lesion cohort and normalized the space by locating the equivalent points. In this normalized
space, the lateral aspect the hippocampal cell layer and the medial aspect of the dentate gyrus
cell layer fell within the range of the control brains, confirming the uniformity of the space. This
then allows for the placement of the anterior-posterior and medial-lateral location of the
electrodes (Fig 4C).

As the goal of localizing the electrodes is to inform the physiology recorded on those
electrodes, we took advantage of the finding of theta phase offset in the local field potential as
the hippocampus evolves along its septo-temporal axis[38,42]. Within this two-dimensional space
we confirmed a phase offset of the theta rhythm as the hippocampus progress from dorsal to
intermediate (Fig 4C&D) (see Materials and Methods). Furthermore, this technique provides a
straightforward reference to measure and discover further axes along which such things could
vary.

Discussion

We have presented a cost-effective and scalable protocol for utilizing micro-CT to image
postmortem brain structure with the overlying skull (Fig 1 & Vid 1). The protocol allows for the
quantification of lesions (Fig 3) and localization of electrodes (Fig 4) within the brain. As
contrast for the micro-CT is provided by eosin and phase contrast processing, the processing for
the presented protocol does not preclude further histological sectioning and imaging of the brain
(Fig 2).

This work builds upon the prior protocol, developed for systems neuroscientific
questions, that uses osmium as the contrast agent for micro-CT[8], which allows for use with
commercially available micro-CT machines. Our protocol offers three key advantages over that
previous protocol for systems neuroscience. First, our protocol does not require multiple
hazardous processing steps such as osmication, dehydration and resin embedding—as described
in Masís et al.[8]—replacing it with a single step of 5% eosin in water. Second, it allows for brain
imaging with a partially intact skull, permitting the direct measurement of skull landmarks
relative to brain structures, and if necessary, subsequent micro-CT of the eosin-stained brain
without a skull (data not shown). Third, our protocol greatly simplifies the subsequent processing

for histology as the tissue can be cut and stained immediately after micro-CT with a range of
standard histological stains. Our microscopic analysis confirmed well-preserved neural
morphology after incubation in 5% eosin (Fig 2). Our protocol does, however, benefit from
access to a synchrotron source to take full advantage of phase contrast processing. Although,
there are a limited number of synchrotrons, they are available to all. Recent work has shown that
phase contrast processing could be possible on commercially available micro-CTs[13], allowing for
future studies to potentially relax the need for a synchrotron source.

In developing the presented method, we sought to keep the protocol simple. Given the
rapid advancement of micro-CT technology and processing, it is entirely possible that even
further simplifications to the protocol could occur. Our goal was to use the method as a means
towards neuroanatomic evaluation for systems neuroscience experiments. Given our ability to
evaluate lesions and electrode placement, this method has achieved that goal. As such we have
not compared our method to other plausible options, such as those using different staining
techniques. Further studies can explore such comparisons.

Our protocol, while applied to the rat brain with a focus on the hippocampus, is
compatible with other species and other brain regions. With the expansion of organisms used for
genetic manipulations, neural recordings, and behavior this protocol can be very useful for all
stages of the process. Our protocol also presents an excellent way to quickly generate a
stereotactic atlas for guiding further experimentation. We note however, that micro-CT has
limitations on the maximal field of view available, making this methodology less appropriate for
larger brains.

There have been tremendous advances in recording technology to extract precise
quantitative physiological information from many regions and cells throughout the brain[43-45].
However, the way in which we evaluate the precise locations at which those devices recorded
often rests on descriptive and qualitative metrics. Given the ever-expanding corpus of precise
anatomical information, there is a need to improve the precision and quantitation of electrode
locations, and our protocol provides such an approach. It also allows for rapid and more
objective quantification of lesion extent and other features that one might wish to compare across
brains.

Acknowledgements: We thank J Masis and A Joshi for helpful discussions. We thank EL
Denovellis, H Barnard and A MacDowell for technical assistance. This work was supported by
grants from the Jane Coffin Childs Memorial Fund for Medical Research (DBK), the UCSF
Physician Scientist Scholars Program (DBK), an NIH R25 (R25MH060482) (DBK), an NIH
Core Grant for Vision Research (NIH-NEI EY002162) (LDS), Research to Prevent Blindness
(RPB Unrestricted Grant to the UCSF Department of Ophthalmology) (LDS) the Simons
Foundation for Autism Research grant (291584) (LMF) and the Howard Hughes Medical
Institute (LMF). This research used resources of the Advanced Light Source, a DOE Office of
Science User Facility under contract no. DE-AC02-05CH11231.

Author contribution: DBK, VK and DYP conceived of the study. DBK, VK and DYP
developed the protocol. DBK and DYP collected the data. DBK, RN, ZY and CS performed the
surgeries. DBK, RN, LDS and LMF developed the analyses. DBK, RN, DAM, ZY, and GMW
analyzed the data. DBK, VK, DYP and LMF wrote the paper.

Conflict of interest: VK is employed (part-time) by Neuralink Corp. The other authors declare
no conflicts of interest.

Figure Legends

Figure 1: micro-CT image of the intact brain with overlying skull. A) Three-dimensional
representation of the overlying skull. B) The same brain and skull from A, virtually sliced along
the sagittal plane (top), horizontal plane (bottom right) and coronal plane (bottom left).

Figure 2: Comparison of virtual slices to standard histological sections following Nissl (Cresyl
Violet) stain. The same brain was utilized for both imaging modalities. A-E show different
coronal slices along the anterior-posterior axis. Within each slice, sample brain regions and
anatomic landmarks that can be visualized both in the CT and Nissl stain are indicated using the
brain atlas by Paxinos and Watson[46]. Scale bar in A reflects scaling for all panels.

Figure 3: Visualization and quantification of hippocampal lesions. A) Virtual coronal slice
through example brains that underwent control surgery (left) and hippocampal lesion surgery
(right). B) Example virtual slices of the hippocampus following aligning all brains and masking
the hippocampus. Dorsal hippocampus is shown along the anterior-posterior axis and
intermediate hippocampus is shown along the dorsal-ventral axis. Jagged edges seen in the
dorsal-ventral slices reflect the fact that the mask was created and interpolated along the anterior-
posterior axis. Scale bar applies to all images within this panel. C) Pairwise structural similarity
metric (SSIM) between all pairs of brains. Each pixel represents the value of the SSIM averaged
across all virtual slices bilaterally throughout the dorsal hippocampal formation and the most
dorsal part of the intermediate hippocampal formation. Matrix is ordered based on the
hierarchical cluster from panel D as indicated by the group labels at top. Dotted blue lines
demarcate the major groups from panel D. Left axis shows which are the control and lesion
brains. D) Dendrogram showing the results of the hierarchical clustering of all brain based on the
similarity matrix from panel C. Brains labeled as either control (C, black lines) or lesion (L,
green lines). E) Median (+- interquartile range) similarity of control (black) and lesion (green)
brains to nearest intact brain as a function of anatomical location. Top left and right plots show
the similarity metric calculated along anterior-posterior axis of the right and left dorsal
hippocampus. Bottom left and far right plots show the similarity metric calculated along the
dorsal-ventral axis of the right and left intermediate/ventral hippocampus, respectively.

Figure 4: Visualization and quantification of electrode localization. A) Virtual coronal slice
through a brain with 22 tetrodes implanted, two of which can be visualized in the slice. B)
Virtual horizontal slice through the tetrodes (left). This image was processed to highlight the
tetrodes (see Table 2 for parameters). Image of the bottom of the canula of the drive implanted
into the brain (right) with the locations of the different tetrodes numbered. Xs indicate locations
that did not have tetrodes. C) Two-dimension coordinate system defined by the most anterior
part of the hippocampal cell layer (Fig S2B) and the lateral part of hippocampal cell layer at the
point of intermediate hippocampal fusion (Fig S2B). Left-most gray lines show the locations of
the medial part of dentate gyrus cell layer for 12 different brains. Rightmost gray lines show the
locations of the lateral part of the hippocampal cell layer. Diamonds demarcate the same points
for the electrode implanted brain. Circles represent the location of the electrode implanted into
the brain (one tetrode was utilized as reference and was therefore not included in this plot).
Colors within the circles show the mean phase offset of the theta rhythm in the local field

potential relative to the electrode denoted by the white circle. D) Example local field potential
recordings from electrodes from 4 different tetrodes with traces colored by the phase offset.
Black, top, trace is the reference electrode. Vertical dotted lines indicate the peak of the theta
rhythm in the reference (top) trace.

Video 1: Video showing the entire three-dimensional rendering of the same brain from Fig 1 and
virtually slicing the brain along the sagittal, coronal and horizontal axes.

Supplementary Figure 1: Layout of micro-CT at the ALS. 1 indicated the location of the x-ray
source. 2 indicated the location of the sample, which rotates. 3 indication the location of the
collector optics.

Supplementary Figure 2: A) Fine scale tetrode structure can be visualized with the micro-CT.
Left shows a virtual horizontal slice, where the four individual electrodes that make up a single
tetrode can be seen. Right show a virtual coronal slice, where the individual electrode and the
braiding of the tetrode can be seen. B) Example brain showing the location of the coordinates the
make up the two-dimension space from Fig 4C. Virtual coronal slices of an example brain, with
sample points of the two-dimensional space demarcated as black dots. At AP=0 the dot shows
the origin point. At AP=1 the rightmost dot shows the point (1,1). For the rest of the slices the
rightmost dot shows the medial part of the dentate cell layer and the leftmost dot shows the
lateral aspect of the hippocampal cell layer. Scale bar applies to all images within the panel.

AP (mm)	ML (mm)	DV (mm)
-2.8	±1.4	-3.0
-3.3	±2.4	-3.0
-4.1	±1.8	-2.8
-4.1	±3.4	-2.8
-4.8	±2.0	-2.8
-4.8	±4.2	-3.1
-5.5	±2.6	-3.0
-5.5	±3.6	-2.9
-5.5	±5.0	-3.5
-5.5	±5.0	-5.5
-6.2	±4.0	-3.4
-6.2	±5.4	-4.4

424 **Table 1:** Stereotaxic coordinates of NMDA infusions to produce lesions within the hippocampal
425 formation. Adapted from Kim and Frank[47]. The coordinates are given for a Long Evans rat skull
426 which is leveled so that bregma and lambda line in the same horizontal plane. AP:
427 anteroposterior, ML: mediolateral, DV: dorsoventral. The AP and ML coordinates are measured
428 from bregma. DV coordinates are measured from the dural surface above the brain.

	full brain (Fig 1&2)	lesion brains (Fig 3)	implant brain: electrodes (Fig 4)	implant brain: anatomy (Fig 4)
energy	25 keV	25 keV	25 keV	25 keV
angles	1313	1313	1313	1313
exposure	125 ms	100 ms	125 ms	125 ms
pixel size	7.205 μm/pixel	7.84 μm/pixel	7.72 μm/pixel	7.72 μm/pixel
scintillator distance	540 mm	540 mm	540 mm	540 mm
# tiles	32	8	4	4
~scan time	150 min	32 min	25 min	25 min
outlier removal in 1d	true	true	False	true
ring removal	polar ring	polar ring	Fourier-wavelet	polar ring
max width for ring removal	100	100	100	200
min value of image to filter	0.5	0.6	-3000	0.6
max portion of image to filter	1.2	1.2	3000	1.2
manual centering	false	false	true	true
phase retrieval	true	true	false	true
alpha regularization	0.0005	0.0005	—	0.0005

Table 2: Scanning parameters and reconstruction parameters for Tomopy for all of the brains. The lesion brains include all of the brains described in Fig 3, including both the brains that underwent control and lesion surgery. The implanted brain was processed in two ways, one two resolve the brain anatomy and the other to maximally resolve the electrodes. The x-ray beam is only so wide; therefore, to scan an entire volume multiple tiles in the vertical dimension were taken and then the tiles were manually stitched together. The lesion brains were imaged from the striatum to the cerebellum to focus on the brain regions of interest and to save time.

1.   Chung, K. *et al.* Structural and molecular interrogation of intact biological systems. *Nature* **497,** 332–337 (2013).
2.   Renier, N. *et al.* iDISCO: a simple, rapid method to immunolabel large tissue samples for volume imaging. *Cell* **159,** 896–910 (2014).

442   3.   Marinescu, M. *et al.* Synchrotron radiation X-ray phase micro-computed tomography as a
443   new method to detect iron oxide nanoparticles in the brain. *Mol Imaging Biol* **15,** 552–
444   559 (2013).
445   4.   Pfeiffer, F. *et al.* High-resolution brain tumor visualization using three-dimensional x-ray
446   phase contrast tomography. *Phys Med Biol* **52,** 6923–6930 (2007).
447   5.   de Crespigny, A. *et al.* 3D micro-CT imaging of the postmortem brain. *J Neurosci Methods*
448   **171,** 207–213 (2008).
449   6.   Zhu, P. *et al.* Low-dose, simple, and fast grating-based X-ray phase-contrast imaging. *Proc*
450   *Natl Acad Sci USA* **107,** 13576–13581 (2010).
451   7.   Stampanoni, M. *et al.* in *Developments in X-Ray Tomography VII* (ed. Stock, S. R.) **7804,**
452   78040L (International Society for Optics and Photonics, 2010).
453   8.   Masís, J. *et al.* A micro-CT-based method for quantitative brain lesion characterization
454   and electrode localization. *Sci Rep* **8,** 5184 (2018).
455   9.   Croton, L. C. P. *et al.* In situ phase contrast X-ray brain CT. *Sci Rep* **8,** –12 (2018).
456   10.   Choi, J. P. *et al.* Micro-CT Imaging Reveals Mekk3 Heterozygosity Prevents Cerebral
457   Cavernous Malformations in Ccm2-Deficient Mice. *PLoS One* **11,** e0160833 (2016).
458   11.   Hieber, S. E. *et al.* Tomographic brain imaging with nucleolar detail and automatic cell
459   counting. *Sci Rep* **6,** –11 (2016).
460   12.   Dyer, E. L. *et al.* Quantifying Mesoscale Neuroanatomy Using X-Ray Microtomography.
461   *eNeuro* **4,** (2017).
462   13.   Toepperwien, M., van der Meer, F., Stadelmann, C. & Salditt, T. Three-dimensional virtual
463   histology of human cerebellum by X-ray phase-contrast tomography. *Proc Natl Acad Sci*
464   *USA* **115,** 6940–6945 (2018).
465   14.   Strotton, M. C. *et al.* Optimising complementary soft tissue synchrotron X-ray
466   microtomography for reversibly-stained central nervous system samples. *Sci Rep* **8,** –18
467   (2018).
468   15.   Pinzer, B. R. *et al.* Imaging brain amyloid deposition using grating-based differential
469   phase contrast tomography. *Neuroimage* **61,** 1336–1346 (2012).
470   16.   Huang, S. *et al.* In-line phase-contrast and grating-based phase-contrast synchrotron
471   imaging study of brain micrometastasis of breast cancer. *Sci Rep* **5,** 9418 (2015).
472   17.   Zhou, Z. *et al.* Cerebral cavernous malformations arise from endothelial gain of MEKK3-
473   KLF2/4 signalling. *Nature* **532,** 122–126 (2016).
474   18.   Cedola, A. *et al.* X-Ray Phase Contrast Tomography Reveals Early Vascular Alterations and
475   Neuronal Loss in a Multiple Sclerosis Model. *Sci Rep* **7,** 5890–11 (2017).
476   19.   Massimi, L. *et al.* Exploring Alzheimer's disease mouse brain through X-ray phase contrast
477   tomography: From the cell to the organ. *Neuroimage* **184,** 490–495 (2019).
478   20.   Toepperwien, M., van der Meer, F., Stadelmann, C. & Salditt, T. Correlative x-ray phase-
479   contrast tomography and histology of human brain tissue affected by Alzheimer's
480   disease. *Neuroimage* **210,** (2020).
481   21.   Khimchenko, A. *et al.* Hard X-Ray Nanoholotomography: Large-Scale, Label-Free, 3D
482   Neuroimaging beyond Optical Limit. *Adv Sci (Weinh)* **5,** 1700694 (2018).
483   22.   de Castro Fonseca, M. *et al.* High-resolution synchrotron-based X-ray microtomography
484   as a tool to unveil the three-dimensional neuronal architecture of the brain. *Sci Rep* **8,** –
485   13 (2018).

486   23.   Schulz, G. *et al.* High-resolution tomographic imaging of a human cerebellum:
487   comparison of absorption and grating-based phase contrast. *J R Soc Interface* **7,** 1665–
488   1676 (2010).
489   24.   Beltran, M. A. *et al.* Interface-specific x-ray phase retrieval tomography of complex
490   biological organs. *Phys Med Biol* **56,** 7353–7369 (2011).
491   25.   Hatta, H. *et al.* A simple and rapid decalcification procedure of skeletal tissues for
492   pathology using an ultrasonic cleaner with D-mannitol and formic acid. *Acta Histochem.*
493   **116,** 753–757 (2014).
494   26.   Savi, F. M., Brierly, G. I., Baldwin, J., Theodoropoulos, C. & Woodruff, M. A. Comparison
495   of Different Decalcification Methods Using Rat Mandibles as a Model. *J. Histochem.*
496   *Cytochem.* **65,** 705–722 (2017).
497   27.   Busse, M. *et al.* Three-dimensional virtual histology enabled through cytoplasm-specific
498   X-ray stain for microscopic and nanoscopic computed tomography. *Proc Natl Acad Sci*
499   *USA* **115,** 2293–2298 (2018).
500   28.   Müller, M. *et al.* Nucleus-specific X-ray stain for 3D virtual histology. *Sci Rep* **8,** 17855
501   (2018).
502   29.   Miller, E. A. *et al.* Robust and replicable measurement for prepulse inhibition of the
503   acoustic startle response. *Mol. Psychiatry* (2020).
504   30.   Paganin, D., Mayo, S. C., Gureyev, T. E., Miller, P. R. & Wilkins, S. W. Simultaneous phase
505   and amplitude extraction from a single defocused image of a homogeneous object. *J*
506   *Microsc* **206,** 33–40 (2002).
507   31.   Schindelin, J. *et al.* Fiji: an open-source platform for biological-image analysis. *Nat*
508   *Methods* **9,** 676–682 (2012).
509   32.   Schindelin, J., Rueden, C. T., Hiner, M. C. & Eliceiri, K. W. The ImageJ ecosystem: An open
510   platform for biomedical image analysis. *Mol. Reprod. Dev.* **82,** 518–529 (2015).
511   33.   Schneider, C. A., Rasband, W. S. & Eliceiri, K. W. NIH Image to ImageJ: 25 years of image
512   analysis. *Nat Methods* **9,** 671–675 (2012).
513   34.   Kikinis, R., Pieper, S. D. & Vosburgh, K. G. in *Intraoperative Imaging and Image-Guided*
514   *Therapy* **13,** 277–289 (Springer, New York, NY, 2014).
515   35.   Fedorov, A. *et al.* 3D Slicer as an image computing platform for the Quantitative Imaging
516   Network. *Magn Reson Imaging* **30,** 1323–1341 (2012).
517   36.   Wang, Z., Bovik, A. C., Sheikh, H. R. & Simoncelli, E. P. Image quality assessment: from
518   error visibility to structural similarity. *IEEE Trans Image Process* **13,** 600–612 (2004).
519   37.   Wang, Z. & Bovik, A. C. Mean Squared Error: Love It or Leave It? A new look at signal
520   fidelity measures. *Ieee Signal Processing Magazine* **26,** 98–117 (2009).
521   38.   Lubenov, E. V. & Siapas, A. G. Hippocampal theta oscillations are travelling waves. *Nature*
522   **459,** 534–539 (2009).
523   39.   Richardson, D. S. & Lichtman, J. W. Clarifying Tissue Clearing. *Cell* **162,** 246–257 (2015).
524   40.   Rosenberg, H. S. & Marcontell, J. Whole-mount, paraffin imbedding as a method for
525   preservation of congenitally malformed hearts. *American Heart Journal* **67,** 379–382
526   (1964).
527   41.   Anderson, R. & Maga, A. M. A Novel Procedure for Rapid Imaging of Adult Mouse Brains
528   with MicroCT Using Iodine-Based Contrast. *PLoS One* **10,** e0142974 (2015).

529   42.   Patel, J., Fujisawa, S., Berényi, A., Royer, S. & Buzsáki, G. Traveling theta waves along the
530   entire septotemporal axis of the hippocampus. *Neuron* **75,** 410–417 (2012).

531   43.   Chung, J. E. *et al.* High-Density, Long-Lasting, and Multi-region Electrophysiological
532   Recordings Using Polymer Electrode Arrays. *Neuron* **101,** 21–31.e5 (2018).

533   44.   Jun, J. J. *et al.* Fully integrated silicon probes for high-density recording of neural activity.
534   *Nature* **551,** 232–236 (2017).

535   45.   Voigts, J., Newman, J. P., Wilson, M. A. & Harnett, M. T. An easy-to-assemble, robust, and
536   lightweight drive implant for chronic tetrode recordings in freely moving animals. *bioRxiv*
537   **16,** 11–23 (2020).

538   46.   Paxinos, G. & Watson, C. *The Rat Brain in Stereotaxic Coordinates*. (Academic Press,
539   2013).

540   47.   Kim, S. M. & Frank, L. M. Hippocampal lesions impair rapid learning of a continuous
541   spatial alternation task. *PLoS One* **4,** e5494 (2009).

Figure 1

A

B
sagittal

2 mm

horizontal

2 mm

coronal

2 mm

Figure 2

aca: anterior commissure, ant   DG: dentate gyrus
CA1/2/3: hippocampal cell layer   DMPAG: dorsomedial periaqueductl grey
Cb: cerrebellar lobules   fmi: forceps minor corpus callosum   opt: optic tract
cc: corpus callosum   fmj: forceps major corpus callosum   PFl: paraflocculus
cp: cerebral peduncle   ic: internal capsule   Pir: piriform cortex
CPu: caudate putamen (striatum)   lo: lateral olfactory tract   rf: rhinal fissure

Figure 3

Figure 4

A

0.1 mm   0.1 mm

B

AP=1.6

AP=1

AP=0.55

AP=0.14

AP=0

2 mm