Scalable method for micro-CT analysis enables large scale quantitative
characterization of brain lesions and implants

David B. Kastner[1,2*], Viktor Kharazia[2*], Rhino Nevers[2], Daniela Astudillo Maya[2], Greer M.
Williams[1], Zhounan Yang[1], Clay Smyth[2], Luca Della Santina[3], Dilworth Y. Parkinson[4], Loren M.
Frank[2,5]

[1]*Department of Psychiatry, University of California, San Francisco, CA 94143, USA*

[2]*Kavli Institute for Fundamental Neuroscience and Department of Physiology, University of*
California, San Francisco, CA 94158, USA

[3]*Deparment of Ophthalmology, University of California, San Francisco, CA 94143, USA*

[4]*Advanced Light Sources Division, Lawrence Berkeley National Labs, Berkeley, CA 94720, USA*

[5]*Howard Hughes Medical Institute*

[*]*These authors contributed equally to this work*

Corresponding author: David B. Kastner, david.kastner2@ucsf.edu

Abstract

Anatomic evaluation is an important aspect of many studies in neuroscience; however, it often lacks information about the three-dimensional structure of the brain. With appropriate contrast, micro-CT imaging provides an excellent, nondestructive, method for the evaluation of brain structure, but currently available protocols are labor intensive, use hazardous compounds, complicate standard histological analysis, and require the removal of the skull. Here we present a simple protocol that solves those problems. We use eosin in combination with bone decalcification to create contrast in the tissue and then employ monochromatic and propagation phase-contrast micro-CT imaging to enable the imaging of brain structure with the preservation of the surrounding skull. Instead of relying on descriptive, time-consuming, or subjective methods, we develop simple quantitative analyses to map the locations of recording electrodes and to characterize the presence and extent of brain lesions.

Introduction

Evaluating brain structure presents a critical confirmatory step for much of systems neuroscience. This often occurs through histological sectioning and light-microscopic analysis. Although such a technique provides a high-resolution view of brain structures, it is time-consuming, difficult to scale for high-throughput analyses, and, perhaps most problematic, it destroys the three-dimensional structure of the brain. Additionally, histological processing is often accompanied by description analyses and not quantitative evaluation. Newer, tissue clearing, methods (Chung et al., 2013; Renier et al., 2014) maintain the three-dimensional structure of the brain for high resolution imaging, but remain expensive, labor intensive, and not well equipped to evaluate lesions or electrode localization, mainstays of systems neuroscientific inquiry. Micro-CT presents an excellent middle ground for evaluating brain anatomy while maintaining three-dimensional structure (Ding et al., 2019; Dyer et al., 2017; Masís et al., 2018).

A previous, elegant, protocol was developed for micro-CT imaging of the post-mortem rodent brain (Masís et al., 2018); however, that protocol, although clear and well described, contains many involved steps, some of which require expensive and hazardous elements, including osmium. There is no evidence that that protocol could work if the skull was left intact, and simple quantitative methods for the analysis of the imaged volume were not established. Furthermore, in using osmium, it becomes far more challenging to process the brain for subsequent standard histologic techniques, requiring a decision to be made between either imaging the full three-dimensional volume, or using standard histologic processing.

Drawing upon the soft-organ literature (Busse et al., 2018; Müller et al., 2018) we present an easy, scalable and inexpensive protocol for providing contrast to the post-mortem brain using eosin as the contrast agent. Furthermore, we adapt decalcification techniques (Hatta et al., 2014; Savi et al., 2017) to image the brain with critical parts of the skull intact. To extract maximal contrast from the brain, we scan the brain with monochromatic X-rays and propagation phase contrast imaging, both of which benefit from utilizing a synchrotron source for the X-rays. The resulting protocol is simple and allows for evaluation of brain structure

relative to the coordinates determined by skull sutures. It does not preclude further processing with standard histologic techniques. In addition, we present quantitative methods to localize electrodes within the brain and describe the extent of lesions within the brain.

Material and Methods

Animals: All experiments were conducted in accordance with University of California, San Francisco Institutional Animal Care and Use Committee and US National Institutes of Health guidelines. Datasets were collected from male Long-Evans rats that were fed standard rat chow (LabDiet 5001).

Surgery: Anesthesia was induced using ketamine, xylazine, atropine, and isoflurane. For the animals that underwent surgery, they were either implanted with a multielectrode microdrive, or underwent lesioning of the dorsal and intermediate hippocampus. For the microdrive implant, craniotomies and accompanying durotomies were placed at +2.6 ML and -4.0 AP for the dorsal hippocampal canula and +4.7 ML and -5.7 AP for the intermediate hippocampal canula. Values are relative to bregma. Screws were imbedded into the skull to support the acrylic casing that secured the microdrive onto the skull of the rat. For the lesion rats, craniotomies were drilled bilaterally to allow for access to all 24 injection sites (Table 1) and the injection needle (33 gauge) was inserted through the dura. For each lesion site, the needle was lowered 0.1 mm below the target coordinates and then raised back to the target prior to injection. 120 nl of NMDA (10 mg/ml) was injected at each site at a rate of 0.1 μl/min. The needle was left at the site for 2 minutes after finishing the injection. For the rats that underwent control surgeries, everything was the same as the lesion rats, except no NMDA was injected, and the needle did not dwell at the hypothetical injection site. The animals that underwent the lesion surgery (either with or without injection of the NMDA) received Diazepam 10 mg/kg intraperitoneally post-operatively.

Electrophysiological recording: Tetrodes were individually moved toward the hippocampus daily. Physiological data, digitized at 30 kHz, and was recorded using the SpikeGadgets data acquisition system and the Trodes software package (SpikeGadgets).

Processing for micro-CT: Prior to perfusion, animals were anesthetized with isoflurane, and then euthanized with pentobarbital. Animals were perfused intracardially with 100 mL of phosphate buffer solution (PBS), followed by 250 mL of chilled 4% paraformaldehyde in PBS. The skull was dissected out with the brain intact, with most of facial bones trimmed off, and all soft tissues removed. A 5-mm-wide opening was made using a rongeur on the ventral side along the midline (optional), for improved diffusion of fixative. The sample was then placed in 50mL of the same fixative for 48 hours at 4°C.

Subsequently, the sample was rinsed in PBS with 0.05% sodium azide (PBS-azide) and decalcified by placing it in a container with 250 mL of 8% EDTA (Sigma-Aldrich) for 6-7 days at room temperature on an orbital shaker. The solution was replaced once midway through the decalcification.

After rinsing again in PBS-azide, the decalcified skull was further trimmed to a final size
on both anterior and posterior sides using a razor blade. Care was taken to avoid compressing
or otherwise distorting the brain. The trimmed sample was placed in 50mL of 5% Eosin Y
(Sigma-Aldrich) for 6-7 days at 4°C on an orbital shaker. The solution was replaced once after 48
hours. The sample was then rinsed in PBS and stored in 50mL PBS-azide at 4°C prior to micro-CT
scanning.

For Nissl staining, the brain (with skull removed) was cut using a Vibratome 3000 (Leica
Biosystems); 50-80 um thick coronal sections were collected into a 24-well dish with PBS-azide
and stored at 4°C. Sections were mounted on adhesive slides, air-dried, and stained using a
traditional cresyl violet staining protocol.

Micro-CT scanning and image reconstruction: Scanning was done at beamline 8.3.2
(tomography) at the Advanced Light Source (ALS) at Lawrence Berkeley National Laboratory (Fig
S1). All scans were carried out at room temperature.

For the brain with the implanted microdrive, the brain anterior and posterior to the
hippocampus were removed to decrease the width of the brain to maximize resolution for the
imaging. Furthermore, as many as possible of the skull screws were removed to minimize
interference. As the microdrive was still attached to the skull via acrylic and remaining skull
screws, it provided the support for the brain. The brain was out of solution for the scanning. X-
rays passed through the sample along the A-P/M-L axes as the sample rotated.

For the rest of the brains, the sample was placed snuggly in a conical tube and was not
in solution (Fig S1). X-rays passed through the sample along the M-L/D-V axes as the sample
rotated. This is the narrowest dimension of the brain, thereby allowing for the finest resolution
~7 μm/pixel (see Table 2).
Monochromatic X-rays were generated using a multilayer mirror monochromator.
Energy of 25 keV was used, and the sample was placed 54 cm (maximal distance for beamline)
from the collector (Fig S1) to enable maximal usage of propagation phase contrast imaging. A
total of 1313 angles with an exposure time of 100 – 125 ms were used to collect the image
(Table 2).
The images were reconstructed using tomopy using a script designed at the ALS
(https://bitbucket.org/berkeleylab/als-microct-python/src/master/tomopy832/tomopy832.py).
See Table 2 for the parameters used for the reconstruction of the various brains. Multiple tiles
were manually stitched together to create the final imaged volume (Table 2).

Image visualization and analysis: Fiji (Schindelin et al., 2012; 2015; Schneider et al., 2012) was
used to visualize virtual slices of the brain and to create a mask of the hippocampal region for
the lesion analysis using the Segmentation Editor tools. Imaris (Oxford Instruments) was used
for three-dimensional reconstruction of the brain. Slicer-3D 4.10.2 (https://www.slicer.org/)
(Fedorov et al., 2012; Kikinis et al., 2014) was used to register and align the brains
(http://hdl.handle.net/1926/1291) by finding the linear transformation (rotation and
translation) to place the entirety of the imaged brains into a shared coordinate system. The
alignment of the brains was not restricted to align any specific region of the brain (such as just
the hippocampus) and rather worked to best align the whole brain and skull.

For the lesion analysis we were interested in understanding the similarity of the
hippocampal formation between brains as that was the region of lesioning. Therefore, we
created a mask for the left and right hippocampal region. The mask was intentionally generous
to ensure that the entire hippocampal formation was included for all brains. For each virtual
slice we found the similarity between all pairs of brains using the structural similarity metric
(SSIM) (Wang and Bovik, 2009; Wang et al., 2004). For SSIM we used a gaussian smoothing filter
for the images with a sigma of 50 pixels. A dendrogram that grouped lesioned and control
brains was calculated using single-linkage hierarchical clustering over the average SSIM for all
pairs of brains. The similarity matrix was subtracted from 1 to create a dissimilarity matrix as
the input to the hierarchical clustering.

Theta phase offset calculation: The local field potential (LFP) was defined as the voltage from
each tetrode after being low-pass filtered below 400 Hz and down-sampled to 1.5 kHz. To
extract theta oscillations, the LFP data were band-pass filtered with a Parks-McClellan optimal
equiripple FIR filter (pass band between 4 and 10 Hz, stop bands below 3 Hz and above 11 Hz)
applied in both the forward and reverse directions to prevent phase distortion (Lubenov and
Siapas, 2009). Instantaneous theta phase was calculated as the angle of the Hilbert-
transformed theta oscillation.

Analysis of theta phase was limited to time periods when the rat was running, which
was defined as a body speed greater than 10 cm/s (Lubenov and Siapas, 2009). Instantaneous
phase differences between pairs of tetrodes were measured as the angle between the complex
phase vectors for each tetrode. For each running period, the average pairwise phase difference
was calculated as the circular mean over time of the angle between phase vectors, and the
average phase offset over the entire behavior session was calculated as the circular mean of
individual average phase differences for each running period.

Results:

Brain structure with intact skull

To utilize micro-CT for structural imaging of the brain it is critical to add a substrate to provide
contrast to resolve different brain structures. In developing our protocol, our first goal was to
create as simple a procedure as possible, and one that does not require osmication,
dehydration of the tissue, nor imbedding in resin (Masís et al., 2018). We therefore developed a
protocol wherein we use eosin (Busse et al., 2018) to provide contrast within the brain, and
imaged the brain using monochromatic X-rays at the Advanced Light Source at Lawrence
Berkeley National Lab. The monochromatic, lower energy, X-ray source is far less harsh than
would be polychromatic white light, obviating the need for any further protective processing to
the brain, even with the greater flux of X-rays provided by a synchrotron.

Typically, the skull would be removed to image the brain, as it provides a barrier for light
transmission. However, the skull provides the reference frame for the stereotactic coordinates
utilized to target specific brain structures while the animal is alive. Removing the dorsal skull
therefore discards a piece of information critical for evaluating and refining anatomical
targeting. Furthermore, the skull also provides a rigid structure to prevent alterations to the

brain dimensions due to histological processing. Therefore, in addition to processing the brain
with eosin, we also adapted decalcification techniques to leave most of the skull in place but
allow for enough transmitted X-rays to image the underlying brain.

With our method we are able to image the whole rat brain with the overlying skull (Fig 1
& Vid 1). The eosin provides sufficient contrast to the brain to allow for the visualization of
many different brain regions. Given that the imaging provides a three-dimensional volume, we
can virtually section the brain to any given region of interest along any plane of interest (Fig
1B). Having the skull intact allows for setting the frame of reference of the brain as one would
in surgery by leveling bregma and lambda.

An advantage of using eosin to provide contrast to the brain is that it is commonly used
stain for histology, and therefore still allows for classic histologic imaging following the micro-
CT. Therefore, using the same brain, following the micro-CT, we sectioned the brain and stained
it with Cresyl Violet to demonstrate the Nissl substance in the neurons. Nissl-stained brain
sections exhibited well preserved morphology and can be used as anatomic reference to
corresponding micro-CT virtual slices (Fig 2).

Lesion quantification

Lesions provide a useful perturbation for evaluating the necessity of brain regions for behavior.
However, validating a lesion using sectioning, staining and imaging presents a challenge for
quantification, often leading to more descriptive analyses. Therefore, we utilized our protocol
to quantitatively evaluate the presence and extent of hippocampal lesions.

The CT protocol enables clear visualization of the lesioned area of the brain (Fig 3A).
Quantification of the lesion could be done by image segmentation, as was done with previous
micro-CT processing (Masís et al., 2018); however, image segmentation is laborious and prone
to subjectivity. Therefore, we sought a more automatable and objective approach for
quantifying the presence and extent of lesions.

We processed and collected micro-CT images of 25 rat brains—12 rats underwent
control surgery and 13 rats underwent hippocampal lesion surgery (see Material and Methods).
We aligned all brains to a common reference and created a mask to extract the three-
dimensional region of the hippocampus across all brains (Fig 3B). We calculated the similarity of
all pairs of brains for each virtual slice using SSIM (Fig 3C), a similarity metric designed to
capture the perceptual similarity between images (Wang et al., 2004; Wang and Bovik, 2009).
We clustered the brains using hierarchical clustering based on the average pairwise similarity
matrix across the bilateral dorsal hippocampal formation and bilateral dorsal-most portion of
the intermediate hippocampal formation (the portion of the hippocampal formation that was
lesioned (Table 1)). The lesion and control brains separate into different groups by the
clustering (Fig 3D).

To visualize the extent of the lesion, for all brains and virtual slices we found the most
similar control brain and then plotted those values for the group of control and lesion brains
across the dorsal and intermediate/ventral hippocampus for both the left and right sides of the
brain (Fig 3E). Across the entire extent of the dorsal hippocampus, and the initial part of the
intermediate hippocampus the two groups of animals were clearly separate. Towards the later

part of the intermediate hippocampus and into ventral hippocampus the brains were
indistinguishable, both being similar to control brains. This is consistent with the intended
location of the lesions, since far intermediate and ventral hippocampus were not targeted
(Table 1).

Electrode localization

Beyond just providing potentially useful landmarks, the skull also provides a stable structure to
support a microelectrode drive, thereby enabling minimal disruption of electrodes during
processing and scanning. Therefore, we evaluated if we could image functional electrodes
implanted in the brain and assess their location relative to brain structures. Independently
moveable tetrodes were implanted above dorsal and intermediate hippocampus. The
microdrive remained on the skull to minimally disrupt the electrodes. The brain, with skull and
microdrive intact, underwent imaging with micro-CT.

The electrodes were easily visualized, and the resolution was sufficiently fine to
distinguish the four individual electrodes that comprised each tetrode wire (Fig S2A). Although
the hardware used to stabilize the microdrive on the skull degraded the imaging quality
somewhat, there remained sufficient signal to resolve the pyramidal cell layer of the
hippocampus (Fig 4A). All that was required to identify individual electrodes was a visual
comparison of the layout of the canula of the drive and the arrangement of the electrodes from
the micro-CT (Fig 4B).

As with the lesions, here, with the electrodes, the goal is not just to visualize, but to also
extract relevant quantitative information. Again, it would be possible to perform image
segmentation for the hippocampal cell layer, but that would be quite laborious and subjective
at the most critical parts due to the large signal produced by the electrodes. Instead we sought
to extract coordinates that could form the basis of a systematic two-dimensional space on
which we could place the locations of the electrodes. For this analysis, we were most interested
in a two-dimensional space for the anterior-posterior and medial-lateral axis, since the
electrodes were moved over the course of the recording, thereby making the dorsal-ventral
coordinates of the CT not representative of the recording location.

We took advantage of the control brains that we scanned for the lesion cohort (n=12)
and utilized those brains to define a consistent two-dimensional coordinate system. As the
origin of the space we utilized the center of most anterior portion of the hippocampal CA1 cell
layer (Fig S2B). And then as point 1,1 we utilized the lateral extent of the hippocampal
(CA1/2/3) cell layer at the point where the dorsal and ventral portions of the hippocampus fuse
to form intermediate hippocampus (Fig S2B). We chose these points as they were easily visible
in all brains and would be likely visible in an electrode implanted brain as well. Within this
space, we then plotted, for each brain, sample locations of the lateral aspect of the
hippocampal cell layer and the medial aspect of the dentate gyrus cell layer from anterior to
posterior (Fig 4C&S2B). Across all brains the locations were very similar.

We then aligned the brain with the implanted electrodes to the same reference brain as
the lesion cohort and normalized the space by locating the equivalent points. In this
normalized space, the lateral aspect the hippocampal cell layer and the medial aspect of the

dentate gyrus cell layer fell within the range of the control brains, confirming the uniformity of
the space. This then allows for the placement of the anterior-posterior and medial-lateral
location of the electrodes (Fig 4C).

As the goal of localizing the electrodes is to inform the physiology recorded on those
electrodes, we took advantage of the finding of theta phase offset in the local field potential as
the hippocampus evolves along its septo-temporal axis (Lubenov and Siapas, 2009; Patel et al.,
2012). Within this two-dimensional space we confirmed a phase offset of the theta rhythm as
the hippocampus progress from dorsal to intermediate (Fig 4C&D) (see Materials and
Methods). Furthermore, this technique provides a straightforward reference to measure and
discover further axes along which such things could vary.

Discussion

We have presented a simple, cost-effective and scalable protocol for utilizing micro-CT to image
postmortem brain structure with the overlying skull (Fig 1 & Vid 1). The protocol allows for the
quantification of lesions (Fig 3) and localization of electrodes (Fig 4) within the brain. As
contrast for the micro-CT is provided by eosin, the processing for the presented protocol does
not preclude further histological sectioning and imaging of the brain (Fig 2).

This work builds upon the prior protocol that uses osmium as the contrast agent for
micro-CT (Masís et al., 2018), which allows for use with commercially available micro-CT
machines. Our protocol offers three key advantages over that previous protocol. First, our
protocol does not require multiple hazardous processing steps such as osmication, dehydration
and resin embedding—as described in Masís et al. (2018)—replacing it with a single step of 5%
eosin in water. Second, it allows for brain imaging with a partially intact skull, permitting the
direct measurement of skull landmarks relative to brain structures. If necessary, micro-CT of
the eosin-stained brain can also be done without a skull (data not shown). Third, our protocol
greatly simplifies the subsequent processing for histology as the tissue can be cut and stained
immediately after micro-CT with a range of standard histological stains. Our microscopic
analysis confirmed well-preserved neural morphology after incubation in 5% eosin (Fig 2). Our
protocol does, however, benefit from access to a synchrotron source to enable enough flux of
X-rays to allow for monochromatic and propagation phase contrast imaging. Although, there
are a limited number of synchrotrons, they are available to all.

Our protocol, while applied to the rat brain with a focus on the hippocampus, is
compatible with other species and other brain regions. With the expansion of organisms used
for genetic manipulations, neural recordings, and behavior this protocol can be very useful for
all stages of the process. Our protocol also presents an excellent way to quickly generate a
stereotactic atlas for guiding further experimentation. We note however, that micro-CT has
limitations on the maximal field of view available, making this methodology less appropriate for
larger brains.

There have been tremendous advances in recording technology to extract precise
quantitative physiological information from many regions and cells throughout the brain (J. E.
Chung et al., 2018; Jun et al., 2017; Voigts et al., 2020). However, the way in which we evaluate

the precise locations at which those devices recorded often rests on descriptive and qualitative
metrics. Given the ever-expanding corpus of precise anatomical information, there is a need to
improve the precision and quantitation of electrode locations, and our protocol provides such
an approach. It also allows for rapid and more objective quantification of lesion extent and
other features that one might wish to compare across brains.

Acknowledgements: We thank J Masis and A Joshi for helpful discussions. We thank H Barnard
and A MacDowell for technical assistance. This work was supported by grants from the Jane
Coffin Childs Memorial Fund for Medical Research (DBK), the UCSF Physician Scientist Scholars
Program (DBK), an NIH R25 (R25MH060482) (DBK), an NIH Core Grant for Vision Research (NIH-
NEI EY002162) (LDS), Research to Prevent Blindness (RPB Unrestricted Grant to the UCSF
Department of Ophthalmology) (LDS) the Simons Foundation for Autism Research grant
(291584) (LMF) and the Howard Hughes Medical Institute (LMF). This research used resources
of the Advanced Light Source, a DOE Office of Science User Facility under contract no. DE-
AC02-05CH11231.

Author contribution: DBK, VK and DYP conceived of the study. DBK, VK and DYP developed the
protocol. DBK and DYP collected the data. DBK, RN, ZY and CS performed the surgeries. DBK,
RN, LDS and LMF developed the analyses. DBK, RN, DAM, ZY, and GMW analyzed the data. DBK,
VK, DYP and LMF wrote the paper.

Conflict of interest: VK is employed (part-time) by Neuralink Corp. The other authors declare no
conflicts of interest.

Figure Legends:

Figure 1: micro-CT image of the intact brain with overlying skull. A) Three-dimensional
representation of the overlying skull. B) The same brain and skull from A, virtually sliced along
the sagittal plane (top), horizontal plane (bottom right) and coronal plane (bottom left).

Figure 2: Comparison of virtual slices to standard histological sections following Nissl (Cresyl
Violet) stain. The same brain was utilized for both imaging modalities. A-E show different
coronal slices along the anterior-posterior axis. Within each slice sample brain regions and
anatomic landmarks that can be visualized both in the CT and Nissl stain are indicated using the
brain atlas by Paxinos and Watson (2013).

Figure 3: Visualization and quantification of hippocampal lesions. A) Virtual coronal slice
through example brains that underwent control surgery (left) and hippocampal lesion surgery
(right). B) Example virtual slices of the hippocampus following aligning all brains and masking
the hippocampus. Dorsal hippocampus is shown along the anterior-posterior axis and
intermediate hippocampus is shown along the dorsal-ventral axis. Jagged edges seen in the
dorsal-ventral slices reflect the fact that the mask was created and interpolated along the
anterior-posterior axis. Scale bar applies to all images within this panel. C) Pairwise structural
similarity metric (SSIM) between all pairs of brains. Each pixel represents the value of the SSIM
averaged across all virtual slices bilaterally throughout the dorsal hippocampal formation and

the most dorsal part of the intermediate hippocampal formation. Matrix is ordered based on
the hierarchical cluster from panel D as indicated by the group labels at top. Dotted blue lines
demarcate the major groups from panel D. Left axis shows which are the control and lesion
brains. D) Dendrogram showing the results of the hierarchical clustering of all brain based on
the similarity matrix from panel C. Brains labeled as either control (C, black lines) or lesion (L,
green lines). E) Median (+- interquartile range) similarity of control (black) and lesion (green)
brains to intact brains as a function of anatomical location. Far left and far right plots show the
similarity metric calculated along the dorsal-ventral axis of the right and left
intermediate/ventral hippocampus, respectively. Middle left and middle right plots show the
similarity metric calculated along anterior-posterior axis of the right and left dorsal
hippocampus, respectively.

Figure 4: Visualization and quantification of electrode localization. A) Virtual coronal slice
through a brain with 22 tetrodes implanted, two of which can be visualized in the slice. B)
Virtual horizontal slice through the tetrodes (left). This image was processed to highlight the
tetrodes (see Table 2 for parameters). Image of the bottom of the canula of the drive implanted
into the brain (right) with the locations of the different tetrodes numbered. Xs indicate
locations that did not have tetrodes. C) Two-dimension coordinate system defined by the most
anterior part of the hippocampal cell layer (Fig S2B) and the lateral part of hippocampal cell
layer at the point of intermediate hippocampal fusion (Fig S2B). Left-most gray lines show the
locations of the medial part of dentate gyrus cell layer for 12 different brains. Rightmost gray
lines show the locations of the lateral part of the hippocampal cell layer. Diamonds demarcate
the same points for the electrode implanted brain. Circles represent the location of the
electrode implanted into the brain (one tetrode was utilized as reference and was therefore not
included in this plot). Colors within the circles show the mean phase offset of the theta rhythm
in the local field potential relative to the electrode denoted by the white circle. D) Example local
field potential recordings from electrodes from 4 different tetrodes with traces colored by the
phase offset. Black, top, trace is the reference electrode. Vertical dotted lines indicate the peak
of the theta rhythm in the reference (top) trace.

Supplementary Figure 1: Layout of micro-CT at the ALS. 1 indicated the location of the x-ray
source. 2 indicated the location of the sample, which rotates. 3 indication the location of the
collector optics.

Supplementary Figure 2: A) Fine scale tetrode structure can be visualized with the micro-CT.
Left shows a virtual horizontal slice, where the four individual electrodes that make up a single
tetrode can be seen. Right show a virtual coronal slice, where the individual electrode and the
braiding of the tetrode can be seen. B) Example brain showing the location of the coordinates
the make up the two-dimension space from Fig 4C. Virtual coronal slices of an example brain,
with sample points of the two-dimensional space demarcated as black dots. At AP=0 the dot
shows the origin point. At AP=1 the rightmost dot shows the point (1,1). For the rest of the
slices the rightmost dot shows the medial part of the dentate cell layer and the leftmost dot
shows the lateral aspect of the hippocampal cell layer. Scale bar applies to all images within the
panel.

AP (mm)	ML (mm)	DV (mm)
-2.8	±1.4	-3.0
-3.3	±2.4	-3.0
-4.1	±1.8	-2.8
-4.1	±3.4	-2.8
-4.8	±2.0	-2.8
-4.8	±4.2	-3.1
-5.5	±2.6	-3.0
-5.5	±3.6	-2.9
-5.5	±5.0	-3.5
-5.5	±5.0	-5.5
-6.2	±4.0	-3.4
-6.2	±5.4	-4.4

381 **Table 1:** Stereotaxic coordinates of NMDA infusions to produce lesions within the hippocampal
382 formation. Adapted from Kim and Frank (2009). The coordinates are given for a Long Evans rat
383 skull which is leveled so that bregma and lambda line in the same horizontal plane. AP:
384 anteroposterior, ML: mediolateral, DV: dorsoventral. The AP and ML coordinates are measured
385 from bregma. DV coordinates are measured from the dural surface above the brain.

	full brain (Fig 1&2)	lesion brains (Fig 3)	implant brain: electrodes (Fig 4)	implant brain: anatomy (Fig 4)
energy	25 keV	25 keV	25 keV	25 keV
angles	1313	1313	1313	1313
exposure	125 ms	100 ms	125 ms	125 ms
resolution	7.205 µm/pixel	7.84 µm/pixel	7.72 µm/pixel	7.72 µm/pixel
scintillator distance	540 mm	540 mm	540 mm	540 mm
# tiles	32	8	4	4
~scan time	150 min	32 min	25 min	25 min
outlier removal in 1d	true	true	False	true
ring removal	polar ring	polar ring	Fourier-wavelet	polar ring

max width for ring removal	100	100	100	200
min value of image to filter	0.5	0.6	-3000	0.6
max portion of image to filter	1.2	1.2	3000	1.2
manual centering	false	false	true	true
phase retrieval	true	true	false	true
alpha regularization	0.0005	0.0005	—	0.0005

Table 2: Scanning parameters and reconstruction parameters for Tomopy for all of the brains. The lesion brains include all of the brains described in Fig 3, including both the brains that underwent control and lesion surgery. The implanted brain was processed in two ways, one two resolve the brain anatomy and the other to maximally resolve the electrodes. The x-ray beam is only so wide; therefore, to scan an entire volume multiple tiles in the vertical dimension were taken and then the tiles were manually stitched together. The lesion brains were imaged from the striatum to the cerebellum to focus on the brain regions of interest and to save time.

Busse, M., Müller, M., Kimm, M.A., Ferstl, S., Allner, S., Achterhold, K., Herzen, J., Pfeiffer, F., 2018. Three-dimensional virtual histology enabled through cytoplasm-specific X-ray stain for microscopic and nanoscopic computed tomography. Proc Natl Acad Sci USA 115, 2293–2298. doi:10.1073/pnas.1720862115

Chung, K., Wallace, J., Kim, S.-Y., Kalyanasundaram, S., Andalman, A.S., Davidson, T.J., Mirzabekov, J.J., Zalocusky, K.A., Mattis, J., Denisin, A.K., Pak, S., Bernstein, H., Ramakrishnan, C., Grosenick, L., Gradinaru, V., Deisseroth, K., 2013. Structural and molecular interrogation of intact biological systems. Nature 497, 332–337. doi:10.1038/nature12107

Ding, Y., Vanselow, D.J., Yakovlevi, M.A., Katz, S.R., Lin, A.Y., Clark, D.P., Vargas, P., Xin, X., Copper, J.E., Canfield, V.A., Ang, K.C., Wang, Y., Xiao, X., De Carlos, F., van Rossum, D.B., La Riviere, P., Cheng, K.C., 2019. Computational 3D histological phenotyping of whole zebrafish by X-ray histotomograhpy. Elife 8. doi:10.7554/eLife.44898

Dyer, E.L., Gray Roncal, W., Prasad, J.A., Fernandes, H.L., Gürsoy, D., De Andrade, V., Fezzaa, K., Xiao, X., Vogelstein, J.T., Jacobsen, C., Kording, K.P., Kasthuri, N., 2017. Quantifying Mesoscale Neuroanatomy Using X-Ray Microtomography. eNeuro 4. doi:10.1523/ENEURO.0195-17.2017

Fedorov, A., Beichel, R., Kalpathy-Cramer, J., Finet, J., Fillion-Robin, J.-C., Pujol, S., Bauer, C., Jennings, D., Fennessy, F., Sonka, M., Buatti, J., Aylward, S., Miller, J.V., Pieper, S., Kikinis, R.,

2012. 3D Slicer as an image computing platform for the Quantitative Imaging Network.
Magn Reson Imaging 30, 1323–1341. doi:10.1016/j.mri.2012.05.001
Hatta, H., Tsuneyama, K., Nomoto, K., Hayashi, S., Miwa, S., Nakajima, T., Nishida, T., Nakanishi,
Y., Imura, J., 2014. A simple and rapid decalcification procedure of skeletal tissues for
pathology using an ultrasonic cleaner with D-mannitol and formic acid. Acta Histochem.
116, 753–757. doi:10.1016/j.acthis.2014.01.006
Kikinis, R., Pieper, S.D., Vosburgh, K.G., 2014. 3D Slicer: A Platform for Subject-Specific Image
Analysis, Visualization, and Clinical Support, in: Intraoperative Imaging and Image-Guided
Therapy. Springer, New York, NY, New York, NY, pp. 277–289. doi:10.1007/978-1-4614-
7657-3_19
Kim, S.M., Frank, L.M., 2009. Hippocampal lesions impair rapid learning of a continuous spatial
alternation task. PLoS One 4, e5494. doi:10.1371/journal.pone.0005494.t001
Lubenov, E.V., Siapas, A.G., 2009. Hippocampal theta oscillations are travelling waves. Nature
459, 534–539. doi:10.1038/nature08010
Masís, J., Mankus, D., Wolff, S.B.E., Guitchounts, G., Joesch, M., Cox, D.D., 2018. A micro-CT-
based method for quantitative brain lesion characterization and electrode localization. Sci
Rep 8, 5184. doi:10.1038/s41598-018-23247-z
Müller, M., Kimm, M.A., Ferstl, S., Allner, S., Achterhold, K., Herzen, J., Pfeiffer, F., Busse, M.,
2018. Nucleus-specific X-ray stain for 3D virtual histology. Sci Rep 8, 17855.
doi:10.1038/s41598-018-36067-y
Patel, J., Fujisawa, S., Berényi, A., Royer, S., Buzsáki, G., 2012. Traveling theta waves along the
entire septotemporal axis of the hippocampus. Neuron 75, 410–417.
doi:10.1016/j.neuron.2012.07.015
Paxinos, G., Watson, C., 2013. The Rat Brain in Stereotaxic Coordinates. Academic Press.
Renier, N., Wu, Z., Simon, D.J., Yang, J., Ariel, P., Tessier-Lavigne, M., 2014. iDISCO: a simple,
rapid method to immunolabel large tissue samples for volume imaging. Cell 159, 896–910.
doi:10.1016/j.cell.2014.10.010
Savi, F.M., Brierly, G.I., Baldwin, J., Theodoropoulos, C., Woodruff, M.A., 2017. Comparison of
Different Decalcification Methods Using Rat Mandibles as a Model. J. Histochem.
Cytochem. 65, 705–722. doi:10.1369/0022155417733708
Schindelin, J., Arganda-Carreras, I., Frise, E., Kaynig, V., Longair, M., Pietzsch, T., Preibisch, S.,
Rueden, C., Saalfeld, S., Schmid, B., Tinevez, J.-Y., White, D.J., Hartenstein, V., Eliceiri, K.,
Tomancak, P., Cardona, A., 2012. Fiji: an open-source platform for biological-image analysis.
Nat Methods 9, 676–682. doi:10.1038/nmeth.2019
Schindelin, J., Rueden, C.T., Hiner, M.C., Eliceiri, K.W., 2015. The ImageJ ecosystem: An open
platform for biomedical image analysis. Mol. Reprod. Dev. 82, 518–529.
doi:10.1002/mrd.22489
Schneider, C.A., Rasband, W.S., Eliceiri, K.W., 2012. NIH Image to ImageJ: 25 years of image
analysis. Nat Methods 9, 671–675. doi:10.1038/nmeth.2089
Wang, Z., Bovik, A.C., 2009. Mean Squared Error: Love It or Leave It? A new look at signal
fidelity measures. Ieee Signal Processing Magazine 26, 98–117.
doi:10.1109/MSP.2008.930649

Wang, Z., Bovik, A.C., Sheikh, H.R., Simoncelli, E.P., 2004. Image quality assessment: from error
visibility to structural similarity. IEEE Trans Image Process 13, 600–612.
doi:10.1109/tip.2003.819861
460

Figure 1

A

B sagittal

2 mm

horizontal   2 mm

coronal

2 mm

Figure 2

aca: anterior commissure, ant
CA1/2/3: hippocampal cell layer
Cb: cerebellar lobules
cc: corpus callosum
cp: cerebral peduncle
CPu: caudate putamen (striatum)

DG: dentate gyrus
DMPAG: dorsomedial periaqueductl grey
fmi: forceps minor corpus callosum
fmj: forceps major corpus callosum
ic: internal capsule
lo: lateral olfactory tract

opt: optic tract
PFl: paraflocculus
Pir: piriform cortex
rf: rhinal fissure

Figure 4

A

0.1 mm   0.1 mm

B

AP=1.6

AP=1

AP=0.55

AP=0.14

AP=0

2 mm