Exploiting prior knowledge about biological macromolecules in cryo-EM structure determination

Dari Kimanius[1], Gustav Zickert[2], Takanori Nakane[1], Jonas Adler[2], Sebastian Lunz[4], Carola-Bibiane Schönlieb[4]*, Ozan Öktem[2]*, and Sjors H.W. Scheres[1]*

[1]MRC Laboratory of Molecular Biology, Cambridge, UK
[2]Department of Mathematics, Royal Institute of Technology (KTH), Sweden
[4]Department of Applied Mathematics and Theoretical Physics, University of Cambridge, UK

March 25, 2020

∗ **For correspondence:** cbs31@cam.ac.uk, ozan@kth.se, scheres@mrc-lmb.cam.ac.uk

Abstract

Three-dimensional reconstruction of the electron scattering potential of biological macromolecules from electron cryo-microscopy (cryo-EM) projection images is an ill-posed problem. The most popular cryo-EM software solutions to date rely on a regularisation approach that is based on the prior assumption that the scattering potential varies smoothly over three-dimensional space. Although this approach has been hugely successful in recent years, the amount of prior knowledge it exploits compares unfavourably to the knowledge about biological structures that has been accumulated over decades of research in Structural Biology. Here, we present a regularisation framework for cryo-EM structure determination that exploits prior knowledge about biological structures through a convolutional neural network that is trained on known macromolecular structures. We insert this neural network into the iterative cryo-EM structure determination process through an approach that is inspired by Regularisation by Denoising. We show that the new regularisation approach yields better reconstructions than the current state-of-the-art for simulated data and discuss options to extend this work for application to experimental cryo-EM data.

1   Introduction

In cryo-EM single particle analysis, the three-dimensional structure of biological macromolecules is reconstructed from two-dimensional projection images of multiple copies of these molecules in different relative orientations. The requirement to image under low-dose conditions, in order to limit damage to the radiation-sensitive samples, leads to high levels of noise in cryo-EM images and the need to average over many images. Moreover, individual macromolecules, or particles, adopt unknown orientations in the sample and may adopt multiple conformations. The resulting large numbers of unknown variables render the optimimisation problem ill-posed, which combined with the high noise level in the data lead to a challenging reconstruction problem.

The most widely used approach to cryo-EM structure determination to date is based on the maximum-a-posteriori (MAP) estimator. This approach was introduced to the cryo-EM field through its implementation in the Relion program [1, 2]. Other software packages, like Cryosparc [3] and Thunder [4], have since built on the same approach. The MAP approach differs from common alternative approaches in two aspects. Firstly, it marginalises over the unknown orientation (and conformational class) assignments of the particles; and secondly, it uses an explicit penalisation term as a regulariser.

Whereas previously popular software packages typically aimed to find the best orientation and class for each particle, the marginalisation approach rather uses a weighted average of all possible orientations. Marginalisation was introduced to the cryo-EM field by Sigworth [5]. Using simulated data, he showed that marginalisation over the orientations for 2D alignment against a single reference leads to reduced sensitivity to the choice of the starting reference and the ability to align images with very low signal-to-noise ratios. Marginalisation over class assignments was later found to be particularly useful for 2D [6] and 3D [7] classification of experimental cryo-EM images. Since its original conception in Xmipp [6], marginalisation over class assignments has been adopted in many software packages, including Relion [2], Thunder [4], Frealign/Cistem [8], Sphire [9], and Cryosparc [3].

In general, regularisation can be used to prevent overfitting when solving an ill-posed problem. Before the introduction of the MAP approach, cryo-EM software tools would prevent overfitting mostly through heuristic low-pass approaches and Wiener filter approaches [10, 11]. The MAP approach pioneered the optimisation of an explicitly regularised target function by expressing the reconstruction problem in an empirical Bayesian framework. The prior information that underlies the regularisation in this approach is the assumption that rapidly changing density values in real space cryo-EM reconstructions, or high powers at higher spatial frequencies in Fourier space, are unlikely. In other words, one assumes that cryo-EM reconstructions are smooth. By expressing both the (marginal) likelihood function and the prior in Fourier space, a standard L_2-Tikhonov regularisation approach is formulated [12]. Its MAP solution can be obtained through expectation-maximisation [13], where an analytical solution to the maximisation step exists. The resulting algorithm yields a 3D Wiener filter that removes high-resolution noise from the reconstruction, while estimating all necessary parameters from the data [2]. The ability to obtain efficiently filtered reconstructions without the need for user-tunable parameters probably played an important role in the rapid uptake of MAP optimisation in the field [14].

Nevertheless, the smoothness prior seems an information-poor choice when compared to the knowledge about biological macromolecules that has been acquired through decades of structural biology research. For example, we know that the density for the macromolecules in cryo-EM reconstructions is surrounded by flat solvent density and that proteins are made up of amino-acid chains that fold into well-defined secondary structure elements like α-helices and β-strands. Although a wealth of structural information exists, it has in practice been difficult to hand-craft a formulation of a prior function that can be incorporated into the optimisation algorithm that underlies cryo-EM structure determination.

Machine learning based on deep neural networks, or deep learning in short, can capture this prior information and has recently seen tremendous success in a wide range of computer vision tasks [15, 16, 17]. Convolutional neural networks have been shown to perform equally well or better than conventional state-of-the-art methods in inverse problems for several imaging modalities [18], including de-noising [19, 20, 21], computed tomography [22, 23] and magnetic resonance imaging [24, 25]. Generally, this performance is attributed to the deep neural networks' ability to efficiently learn statistical correlations about both the noise and the signal through the examples in the training data set.

Regularisation by denoising (RED), is an image-recovery framework that enables the plug-in of various denoisers into a MAP optimisation protocol [26]. The RED framework was originally centered around an explicit expression for a prior that incorporates the output of the denoiser. Commonly used optimisation algorithms require the objective function to be differentiable. To achieve this for the proposed prior expression, several conditions must hold for the denoiser function: local homogeneity and Jacobian symmetry. However, most state-of-the-art denoisers, including denoising convolutional neural networks, fail to exhibit these properties, in particular Jacobian symmetry. Despite this, RED has been shown to perform well also with these denoiser functions. Hence, the explicit prior approach cannot explain the behaviour of this framework. In response to these issues, Reehorst and Schniter proposed a new framework, score-matching by denoising (SMD), which showed that RED achieves its performance by approximating the "score" or the gradient of the prior distribution. This approach circumvents the requirement of an explicit prior expression and further does away with the above mentioned conditions on the denoiser function [27].

Here, we propose a cryo-EM structure determination procedure that is inspired by the RED protocol to integrate a convolutional neural network to express prior information. Cryo-EM reconstruction typically uses a vector-valued regularisation parameter to model variations in signal-to-noise ratios across different spatial frequencies. So-called "gold-standard" recommendations in the field are to estimate these through the Fourier shell correlation (FSC) between independently obtained reconstructions from halves of the data set [28]. This estimate typically varies greatly between data sets and throughout different refinement steps. Therefore, we expand RED to account for a vector-valued regularisation parameter that varies throughout the reconstruction and present a theoretical framework, inspired by SMD, to show that this approach is valid in the case of more realistic Gaussian noise. Because we train our convolutional neural network on tens of thousands of cryo-EM reconstructions from simulated data and test it using data that is simulated in the same manner, this work does, not yet, represent a finalised solution to improve cryo-EM reconstruction from experimental data. Rather, it serves as a general proof of principle that learned priors can improve cryo-EM reconstruction. Future research directions, like the exploration of alternative optimisation algorithms and different strategies to design and train the neural networks, are discussed.

2 Theory

2.1 MAP with a Gaussian Prior

The approach taken here is based on computing the MAP estimator in the Bayesian formulation of the structure recovery problem. Data is 2D images $\boldsymbol{y}_1, ..., \boldsymbol{y}_m \in \mathbb{C}^N$ that are digitised Fourier transforms of the 2D projection

images generated by the individual particles. All particles represent the same macro-molecule up to a 3D orientation (rotation and translation) and $\boldsymbol{x} \in \mathbb{C}^M$ is the Fourier transform of its 3D structure. The task is to recover the model $\boldsymbol{x}$ (Fourier transform of 3D structure) from 2D Fourier transform images $\boldsymbol{y}_1, ..., \boldsymbol{y}_m$ (data).

Disregarding noise and measurement errors, the data generated by a model $\boldsymbol{x}$ with 3D orientation ϕ is expressible as $\boldsymbol{H}^\phi \boldsymbol{x} \in \mathbb{C}^N$. Here, $\boldsymbol{H}^\phi \colon \mathbb{C}^M \to \mathbb{C}^N$ is the $(N \times M)$-matrix that is the frequency space model for phase contrast transmission electron microscope (TEM) image formation (forward model) of a particle (a model in $\mathbb{C}^M$ with 3D orientation ϕ) under the weak phase object approximation. The forward model also accounts for the TEM optics by multiplying the electron exit wave in frequency space with a known contrast transfer function (CTF).

The Bayesian setting is based on formulating structure recovery as statistical estimation. We therefore begin by introducing $\mathbb{C}^N$-valued random variables $\mathrm{y}_1, ..., \mathrm{y}_m$ that generate data and the $\mathbb{C}^M$-valued random variable x that generates models. Finally, Φ is the random variable that generates particle orientations ϕ. Assuming independent uncorrelated Gaussian noise on the Fourier components, the conditional probability density of $(\mathrm{y}_i \mid \mathrm{x}, \Phi)$, which is the probability density of the data $\boldsymbol{y}_i \in \mathbb{C}^N$ given the model $\boldsymbol{x} \in \mathbb{C}^M$ and orientation ϕ, is

$$P(\boldsymbol{y}_i | \boldsymbol{x}, \phi) := C(\boldsymbol{\sigma}_i^2) \exp\left(-\frac{1}{2} \left| \boldsymbol{\sigma}_i^{-1} \left(\boldsymbol{y}_i - \boldsymbol{H}_i^\phi \boldsymbol{x} \right) \right|^2 \right). \tag{1}$$

Here, $\boldsymbol{H}_i^\phi \colon \mathbb{C}^M \to \mathbb{C}^N$ is the aforementioned TEM forward model for a particle with orientation ϕ and with particle dependent CTF, the vector $\boldsymbol{\sigma}_i \in \mathbb{R}_{\geqslant 0}^N$ contains the frequency-dependent power of noise for each component of the i:th image and C is a normalisation factor. Assuming independence between individual images and marginalising over the hidden variables, ϕ, the likelihood for the entire data set $\boldsymbol{y} = (\boldsymbol{y}_1, ..., \boldsymbol{y}_m)$ becomes

$$P(\boldsymbol{y}|\boldsymbol{x}) = \prod_i \int_\phi P(\boldsymbol{y}_i|\boldsymbol{x}, \phi) P(\phi) d\phi, \tag{2}$$

where $\Phi \sim P(\phi)$ is a probability density for the hidden variables, which will be ignored in what follows.

In MAP we aim to optimize the posterior probability, or for computational feasibility, the equivalent log-posterior probability, here defined as $\mathcal{L}(\boldsymbol{x}|\boldsymbol{y})$. Given a prior probability density $P(\boldsymbol{x})$ for the model $\boldsymbol{x}$, the log-posterior is (up to an additive constant) given by Bayes' theorem as

$$\mathcal{L}(\boldsymbol{x}|\boldsymbol{y}) := \log P(\boldsymbol{x}|\boldsymbol{y}) = \log P(\boldsymbol{y}|\boldsymbol{x}) + \log P(\boldsymbol{x}). \tag{3}$$

Now, inserting eq. (2) and taking the gradient with respect to $\boldsymbol{x}$ yields

$$\nabla \mathcal{L}(\boldsymbol{x}, \boldsymbol{y}) = \sum_i \mathbb{E}_\Phi \left[\boldsymbol{\sigma}_i^{-2} (\boldsymbol{H}_i^\Phi)^* (\mathrm{y}_i - \boldsymbol{H}_i^\Phi \mathrm{x}) \mid \mathrm{x} = \boldsymbol{x}, \mathrm{y}_i = \boldsymbol{y}_i \right] + \nabla \log P(\boldsymbol{x}), \tag{4}$$

where $(\cdot)^*$ is the adjoint operator. In most implementations of MAP-based cryo-EM refinement algorithms the prior density is assumed to be a Gaussian

$$P(\boldsymbol{x}) := C(\boldsymbol{\tau}^2) \exp\left(-\frac{1}{2} \left| \boldsymbol{\tau}^{-1} \boldsymbol{x} \right|^2 \right). \tag{5}$$

Here, $\boldsymbol{\tau} \in \mathbb{R}_{\geqslant 0}^M$ contains the frequency-dependent estimate of the power of the signal in the model. This distribution arises through the central limit theorem, by assuming random distribution of atoms in a periodic real space box, and can thus be intuitively regarded as a minimal assumption criteria about the density [29]. More intuitively, this prior biases the refinement to a smoother real-space representation by suppressing large Fourier components. There is no analytical solution of the equation $\nabla \mathcal{L}(\boldsymbol{x}|\boldsymbol{y}) = \boldsymbol{0}$, but the Expectation Maximisation algorithm [13] applied to $\mathcal{L}$ results in the iterative scheme:

$$\boldsymbol{x}^{(n+1)} = \frac{\boldsymbol{B}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y} \right)}{\boldsymbol{K}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y} \right) + \boldsymbol{\tau}^{-2}}, \tag{6}$$

where

$$\boldsymbol{B}(\boldsymbol{x}, \boldsymbol{y}) := \sum_i \mathbb{E}_\Phi \left[\boldsymbol{\sigma}_i^{-2} (\boldsymbol{H}_i^\Phi)^* \mathrm{y}_i \mid \mathrm{x} = \boldsymbol{x}, \mathrm{y}_i = \boldsymbol{y}_i \right]$$

$$\boldsymbol{K}(\boldsymbol{x}, \boldsymbol{y}) := \sum_i \mathbb{E}_\Phi \left[\boldsymbol{\sigma}_i^{-2} (\boldsymbol{H}_i^\Phi)^* \boldsymbol{H}_i^\Phi \mid \mathrm{x} = \boldsymbol{x}, \mathrm{y}_i = \boldsymbol{y}_i \right]. \tag{7}$$

In practice, $\boldsymbol{\sigma}$ and $\boldsymbol{\tau}$ are both samples of random variables and estimates of these are updated at every iteration n, but for simplicity these updates are ignored in the notation used.

The vector-valued regularisation parameter $\boldsymbol{\tau}$ is commonly based on the radially averaged FSC calculated from reconstructions of two independent halves of the data set [2]. For each reconstruction, $k = \{1, 2\}$, $\boldsymbol{\tau}_k$ is calculated through

$$\boldsymbol{\tau}_k^2 = \frac{1}{\boldsymbol{K}(\boldsymbol{x}_k, \boldsymbol{y}_{i \in k})} \left(\frac{\text{FSC}(\boldsymbol{x}_1, \boldsymbol{x}_2)}{1 - \text{FSC}(\boldsymbol{x}_1, \boldsymbol{x}_2)} \right), \tag{8}$$

where $\boldsymbol{y}_{i \in k}$ is the k:th-subset of the measured data $\boldsymbol{y}$ and $\boldsymbol{x}_k$ is the reconstructed density from each such subset. Intuitively, as more signal is accumulated into each of the two reconstructions, the FSC increases, which in effect reduces the regularisation applied in eq. (6).

2.2 MAP with non-Gaussian prior

Applying the expectation-maximisation method to find the MAP estimator is convenient when working with Gaussian priors since the M-step has a closed form solution. For a general prior, in the M-step we need to solve the non-linear equation

$$\boldsymbol{K}(\boldsymbol{x}^{(n)}, \boldsymbol{y})\boldsymbol{x}^{(n+1)} - \nabla \log P(\boldsymbol{x}^{(n+1)}) = \boldsymbol{B}(\boldsymbol{x}^{(n)}, \boldsymbol{y}) \tag{9}$$

We can approximately solve this equation by adding and subtracting a Gaussian prior, which gives

$$\nabla \log P(\boldsymbol{x}) = -\boldsymbol{\tau}^{-2}\boldsymbol{x} + \left(\nabla \log P(\boldsymbol{x}) + \boldsymbol{\tau}^{-2}\boldsymbol{x} \right). \tag{10}$$

Assuming that the second term varies slowly with $\boldsymbol{x}$, which would be the case when the prior is approximately Gaussian, we get an approximate MAP estimator for arbitrary differentiable priors

$$\boldsymbol{x}^{(n+1)} = \frac{\boldsymbol{B}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y}\right) + \left(\nabla \log P(\boldsymbol{x}^{(n)}) + \boldsymbol{\tau}^{-2}\boldsymbol{x}^{(n)} \right)}{\boldsymbol{K}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y}\right) + \boldsymbol{\tau}^{-2}} \tag{11}$$

While the Gaussian component of the prior can be readily dealth with using e.g. the FSC as above, the non-linear part is hard to specify mathematically. The regularization by denoising (RED) method allows us to approximate the gradient of the log-prior using a neural network trained to denoise volumes. The RED prior formula has been shown to accurately estimate the gradient of the prior through score matching by denoising (SMD) [27].

When dealing with independent additive Gaussian noise, regularization by denoising (RED) can be used to integrate an external denoiser into an iterative image restoration protocol [26]. We show in the Appendix that the MMSE estimator $\boldsymbol{f} : \mathbb{C}^M \to \mathbb{C}^M$ under Gaussian noise with co-variance $\boldsymbol{\tau}^{-2}$ approximates the gradient of the log-prior according to $\nabla \log P(\boldsymbol{x}) \approx \boldsymbol{\tau}^{-2} \left[\boldsymbol{f}(\boldsymbol{x}) - \boldsymbol{x} \right]$, which in turn implies

$$\nabla \log P(\boldsymbol{x}) + \boldsymbol{\tau}^{-2}\boldsymbol{x} \approx \boldsymbol{\tau}^{-2}\boldsymbol{f}(\boldsymbol{x}). \tag{12}$$

Inserting (11) into (12) gives an approximate expectation-maximisation estimator for priors using a learned denoiser.

$$\boldsymbol{x}^{(n+1)} = \frac{\boldsymbol{B}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y}\right) + \lambda \boldsymbol{f}_{\hat{\boldsymbol{\theta}}}\left(\tilde{\boldsymbol{x}}^{(n)}\right) \boldsymbol{\tau}^{-2}}{\boldsymbol{K}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y}\right) + \boldsymbol{\tau}^{-2}} \quad \text{with} \quad \tilde{\boldsymbol{x}}^{(n)} = \frac{\boldsymbol{B}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y}\right)}{\boldsymbol{K}\left(\boldsymbol{x}^{(n)}, \boldsymbol{y}\right)}. \tag{13}$$

In the above, $\tilde{\boldsymbol{x}}$ is the unregularised $\boldsymbol{x}$ and $0 \leqslant \lambda \leqslant 1$ is an empirical parameter. Note that (13) is equivalent to (6) when $\lambda = 0$, so λ balances the influence of denoiser vs. conventional Gaussian prior.

In the above RED-inspired protocol, the denoiser acts on the unregularised map that contains information from the most recent alignment of the experimental data. Further adaptation can be obtained by making use of the fact that RELION is used for the refinement. To do this, we consider an MMSE estimator $\hat{f} : \mathbb{C}^M \to \mathbb{C}^M$ that is trained to "denoise" RELION refinement volumes. More precisely, we consider the following supervised statistical learning problem:

$$\hat{f} \in \underset{f : \mathbb{C}^M \to \mathbb{C}^M}{\arg\min} \; \mathbb{E}_{(\hat{\mathbb{x}}, \tilde{\mathbb{x}})}\left\{ \left\| \hat{\mathbb{x}} - f(\tilde{\mathbb{x}}) \right\|^2 \right\} \quad \Longrightarrow \quad \hat{f}(\tilde{\boldsymbol{x}}) = \mathbb{E}\left[\hat{\mathbb{x}} \mid \tilde{\mathbb{x}} = \tilde{\boldsymbol{x}} \right]. \tag{14}$$

In the above, $\hat{\mathbb{x}}$ is the random variable that generates appropriately coarsened (low-pass filtered) 3D structures of macro-molecular assemblies from the PDB and $\tilde{\mathbb{x}}$ is the random variable that generates RELION refinement structures.

Since we lack knowledge about the joint probability distribution of $(\hat{\mathbb{x}}, \tilde{\mathbb{x}})$, the conditional expectation in the left-hand-side of (14) is replaced by its empirical counterpart given by supervised training data $(\hat{\boldsymbol{x}}_i, \tilde{\boldsymbol{x}}_i) \in \mathbb{C}^M \times \mathbb{C}^M$ for $i = 1, \ldots, m$ that are random draws from $(\hat{\mathbb{x}}, \tilde{\mathbb{x}})$, i.e., $\hat{\boldsymbol{x}}_i$ is the true 3D structure derived from PDB (after appropriate coarsening) and $\tilde{\boldsymbol{x}}_i$ is a corresponding RELION refinement structure. Furthermore, a deep neural network is used to parameterise $\mathbb{C}^M$-valued mappings on $\mathbb{C}^M$, thereby replacing the infinite dimensional minimisation over such functions with a finite dimensional minimisation over the deep neural

network parameters. The resulting counterpart of (14) can then be written as the following empirical risk minimisation problem:

$$\hat{\theta} \in \arg\min_{\theta} \sum_t \left\| \hat{\boldsymbol{x}}_t - f_\theta(\tilde{\boldsymbol{x}}_t) \right\|^2. \tag{15}$$

If the network has sufficient model capacity and there is enough supervised training data, then the above yields an approximation to the conditional mean estimator, i.e., $\boldsymbol{f}_{\hat{\theta}} \approx \mathbb{E}\big[\hat{\mathbb{x}} \mid \tilde{\mathbb{x}} = \tilde{\boldsymbol{x}}\big]$.

3 Experimental Design

3.1 Convolutional Neural Network

Synthetic training data was generated from 543 atomic structures that were downloaded from the Protein Data Bank (PDB) [cite]. All downloaded structures were solved using X-ray crystallography to at least 4 Å resolution; had molecular weights between 40 and 100 kDa; and consisted of a single protein chain. 110 structures also contained nucleic acids. Atomic models were converted to volumetric density maps in the MRC format [cite] using the pdb2map-tool in EMAN2 [30]. In order to obtain simulated projection data sets at different signal-to-noise ratios (SNRs), the intensity values of these maps were multiplied with three different constants: 0.014, 0.012 and 0.010. Next, the density maps were projected onto images of (96×96) pixels, with a pixel-size of 1.5Å, and independent Gaussian noise of unity variance was added. The resulting projection data sets had SNRs in the range 0.0003 to 0.0100 and on average 0.0020, which was calculated as the average per-pixel variance of the projection images divided by the variance of the added noise. No attempts were made to model the typical CTF of an electron microscope. Individual projection data sets of 10,000 projections each were generated for the different structures and SNRs using random projection directions that uniformly sampled the orientation space. Each of these projection data sets was then used for standard, unmasked "3D auto-refinement" in Relion [2], which was started from a 30Å low-pass filtered version of the ground truth map. New external-reconstruct functionality in Relion-3.1 (also see below) was used to obtain unregularised half-maps at each of the intermediate iterations of all refinements. In total, this process yielded over 25,000 unmasked 3D density maps at different stages of refinement, which were further augmented with size-preserving rotations.

To verify the validity of the underlying assumptions about Gaussian distribution of noise for the derivations in the theory section we also trained a network using pairs of noise free ground truth and pure Gaussian noise distorted maps. For this purpose we used the estimated frequency space distribution of noise calculated from the Relion intermediate reconstructions. During training, random Gaussian noise was generated in frequency space matching this profile. The noise was then convoluted with a real space mask that dampened the noise levels in the solvent region. However, unless specified the presented results are for the a denoiser trained on Relion intermediate reconstructions.

For this work we chose a U-Net [31] for the denoiser and trained the network using Tensorflow [32] in Python. The network was trained with residual learning [19] via ADAM [33] with a mini-batch size of 10 maps for 27 epochs. The U-net consisted of five down- and up-sampling stages that were implemented through max-pooling and transposed convolutional layers [34], respectively. Each stage was made up of two consecutive blocks repeating BN+ReLU+Conv. The input-output pairs for training consisted of the unregularised and unmasked reconstructions from intermediate Relion iterations and their respective noise-free ground truth maps. Here, the coarsening (low-pass filtering) applied to the ground truth (see eq. (15)) was achieved by low-pass filtering it to match the resolution of the corresponding noisy input map. This is done by multiplying the Fourier transform of the ground truth with the FSC of each noisy input map. The FSC is estimated for each map by Relion during the reconstruction of the training data set from the half-maps.

3.2 Assessment of the RED approach

Four PDB structures were selected to test the performance of the RED approach (Table 1). The test structures were excluded from the training data set and had a minimum of 8Å RMSD difference for at least 1,000 aligned atom pairs with any other structure in the training data set. The four test structures represent different levels of difficulty for the reconstruction, mostly arising from differences in their molecular weight (MW) and their overall shapes (as projection images of near-spherical structures are harder to align than those with more anisotropic shapes). In order of increasing difficulty the test structures are listed in Table 1, from the elongated, larger structure 4AIL to the smaller, near-spherically shaped structure 4M82. Projection data sets for testing were made in the same way as the projection data sets that were used to generate the training set for the convolutional neural network. Test data sets were generated at four different SNRs by multiplying the original maps from the pdb2map tool by 0.014, 0.012, 0.01 and 0.008, which yields the average SNRs (i) 0.0038, (ii) 0.0028, (iii) 0.0019 and (iv) 0.0012, respectively.

Table 1: Characteristics of structures in the test data set. Compactness is expressed as the ratio between the smallest and the largest diameter of the molecule. Relative image SNR is expressed as the per-pixel average variance of signal over noise relative to that of the structure with the maximum SNR (4AIL).

	4AIL	4BB9	4BTF	4M82
Molecular weight (kDa)	96	70	54	46
Relative image SNR	1.00	0.52	0.27	0.26
Compactness	0.42	0.64	0.62	0.74
α-helical content (%)	38	49	46	39
β-strand content (%)	21	10	10	12
Loop content (%)	36	41	44	49
Nucleotide content (%)	5	0	0	0

The single-pass performance of the denoiser can be examined by applying it once to unregularized maps and evaluating the ratio between the L^p-difference to ground truth from the denoised map and the unregularized map. We evaluated the average of this ratio (using both L^1 and L^2) for the entire training data set, as a function of the estimated nominal resolution of each map.

Standard, unmasked Relion 3D auto-refinements, with updates based on eq. (6) were compared to refinements with updates based on eq. (13). Again, all refinements were started from initial models that were obtained by 30 Å low-pass filtering of the ground truth map and used the same refinement parameters. The results of both types of refinements were compared based on the reported half-map FSC and the FSC against the ground truth and the angular error relative to the true projection angles.

3.3   Implementation details

Instead of implementing the update formula in eq. (13) directly in the C++ code of Relion, we created an external reconstruction functionality in its refinement program. When using the `--external_reconstruct` command-line option, the `relion_refine` program will write out MRC files containing $\boldsymbol{B}(\boldsymbol{x}, \boldsymbol{y})$ and $\boldsymbol{K}(\boldsymbol{x}, \boldsymbol{y})$ for both half-maps at each iteration step, together with the current estimate for FSC and $\boldsymbol{\tau}^2$. The program will then wait until reconstructions with a specified filename for both half-maps exist on the file system, read those maps back in, and proceed with the next iteration of refinement. This functionality was coupled to a script in Python to perform the update formula in eq. (13) using Tensorflow [32]. The Python script for external reconstruction using the proposed RED protocol, together with the architecture and the weights of the convolutional neural network that was used for the presented results may be downloaded from github (`https://github.com/3dem/externprior`).

4   Results

4.1   Performance of the denoiser

We first tested the effects of applying the denoiser on individual, intermediate, unregularised reconstructions from a standard Relion refinement (Figure 1 A & B). After the first iteration of the refinement of the 4AIL projection data at SNR (i), the unregularised half-reconstruction contains only lower resolution components and exhibits notable rotational smearing. The latter is often observed in low-resolution refinement intermediates and is caused by uncertainty in the angular assignments. Application of the denoiser reduces the rotational smearing and brings the map closer to the true signal, as confirmed by visual comparison and the FSC with the ground truth map.

At iteration 16 of the same refinement, the unregularised map contains higher resolution components, but noise components are clearly visible in the volume surrounding the protein, which should be devoid of features. The denoiser efficiently removes the latter and, again, both visual comparison as well as FSC calculation with the ground truth map confirm that the denoised map is closer to the ground truth signal than the unregularised map.

The average ratio between the difference to ground truth from the denoised map and the unregularized map was calculated for L^1 and L^2, at each nominal resolution (Figure 1 C). When the resolution of the input map is worse than 10Å the denoiser fails to produce a significant improvement on average. As the resolution improves beyond 10Å the performance gradually improves and eventually plateaus beyond 4.5Å.

Figure 1: Single-pass denoising performance results, at low (**A**) and high (**B**) resolution, showing central slices of unregularised input map, output from the denoiser and ground truth for reconstructions of a structure from the test data set (PDB-ID: 4AIL). Color-scale spans between minimum and maximum density values in each slice. FSC curves are showing correlation with ground truth before (black) and after (red) denoising. The resolution at $FSC = 0.5$ is shown for each curve. **C** shows the average ratio between the L^p-difference from ground truth to the denoised map and the unregularized map, for the entire training data set, as a function of the nominal resolution.

Figure 2: FSCs of reconstructions at four different SNRs (rows) of four structures (columns). Comparing regular Relion (black) with RED, using a denoiser trained on Gaussian noise (green) and on Relion intermediate reconstructions (purple). Dotted lines show the half-map FSC and solid lines show FSC between the regularised map and ground truth. All maps are first multiplied with a solvent mask with a smooth edge before comparison to ground truth. Upper shaded area shows the FSC when model is reconstructed with zero angular error.

Figure 3: FSCs of reconstructions at four different SNRs (rows) of four structures (columns). Comparing regular Relion (black) with RED (purple) and confidence weighted RED (red), both using denoiser trained on Relion intermediate reconstructions. Dotted lines show the half-map FSC and solid lines show FSC between the regularised map and ground truth. All maps are first multiplied with a solvent mask with a smooth edge before comparison to ground truth. Upper shaded area shows the FSC when model is reconstructed with zero angular error. Note that purple lines are showing the same results as the purple lines in figure 2.

Figure 4: Central slices of reconstruction results of 4M82 at SNR (*iv*) and their respective ground truth. Regular Relion (**A**), RED (**B**), certainty weighted RED (**C**). Each ground truth map has been low-pass filtered to match the estimated resolution of the reconstructed maps by multiplying their Fourier transform with the half-map FSC. The result for RED is showing high-contrast features that do not correlate well with ground truth. This issue is significantly alleviated by certainty weighting.

Figure 5: The distribution of angular error of reconstructions at four different SNRs (rows) of four structures (columns). Comparing regular Relion (black) with confidence weighted RED (red).

Figure 6: Central slices of the reconstructed maps in the test data set. Each pair comparing regular Relion (left) with confidence weighted RED (right).

Figure 7: Iso-surface visualisations of the reconstruction results with regular Relion, certainty weighted RED together with ground truth for 4AIL at SNR (*iii*) (left) and 4BB9 at SNR (*i*) (right).

4.2 Performance of the RED approach

Next, we tested the performance of the RED approach on projection data sets of the four test structures at four different SNRs. First we run the RED with a denoiser trained on Gaussian noise and compare the results with that for a denoiser trained on Relion intermediate reconstructions (Figures 2). The RED based approache outperforms the standard approach in all refinements performed, as measured by FSCs between the resulting reconstructions and the ground truth maps (solid green and purple lines in Figure 2). This suggests that Gaussian noise can partially explain the distortions. However, the denoiser trained on Relion intermediate reconstructions performs better, confirming that Gaussian noise is an incomplete model of the distortions observed in Relion intermediate reconstructions.

Both the RED approach and standard Relion auto-refinement produced a reconstruction with a resolution close to the Nyquist frequency for the easiest test structure, 4AIL, at the highest SNR. Conversely, both approaches yielded a low-resolution solution that was devoid of recognisable protein features for the most difficult test structure, 4M82, at the lowest SNR. Therefore, the range of structures and SNRs of our test data set represents a wide range of scenarios, from easy to hard refinements. The gains in performance tend to be higher at higher SNRs; they are small for the most difficult refinements.

Also measures of reconstruction quality that are reported by Relion without having access to the ground truth, i.e. the half-map FSCs, improve when using the RED approach (dotted green and purple lines in Figure 2). However, for the refinement of 4M82 at the three lowest SNRs, we noticed a severe overestimation of resolution based on the half-map FSCs using the RED approach. A visual examination of the maps (figure 4 B) reveals high-contrast features that are similar in appearance to those of biological macromolecules, but do not correspond to ground truth. This points to the denoiser introducing false high-resolution features into the reconstructions that are shared among the two half-maps. This then leads to an overestimation of the resolution, which is based on the correlation between the half-maps. Since the half-map FSC is the only reliable estimate of resolution when ground truth is missing this issue poses a major problem for the validation of the reconstruction results.

4.3 Certainty Weighting

The λ-parameter in eq. (13) can be used to fall-back on the Gaussian prior in a scenario where confidence in the performance of the denoiser is lacking. Inspired by the single-pass performance of the denoiser (Figure 1 C) we tested a simple tuning scheme for λ to address the issue of resolution overestimation. We will refer to this modification as *certainty weighted RED*.

At nominal resolutions worse than 10Å, as estimated by Relion's half-map FSC, we assign a full fall-back onto the Gaussian prior, by setting $\lambda = 0$. This is where the single-pass performance indicates that the denoiser results in little improvement in the reconstruction (see Figure 1C). To avoid impairing the performance at high-resolutions, however, we set $\lambda = 1$ at nominal resolutions better than 4.5Å, which is approximately where the single-pass performance of the network begins to peak. For any nominal resolution between 4.5Å and 10Å we use a linear interpolation between zero and one.

Using this approach we see that the overestimation of the half-map FSC is significantly alleviated (dotted red lines in Figure 3) without affecting the overall quality of the reconstructions as measured by the FSC with ground truth (solid red lines in figure 3); the errors in the angular assignments (Figure 5); and the visual

appearance of the resulting structures (Figure 6 and 7). A visual inspection of the 4M82 map resulting from the refinement with the lowest SNR also suggests that the reconstruction no longer accumulates false structural features with poor correlation to ground truth (Figure 4).

5   Conclusions & Discussions

Convolutional neural networks for denoising are typically trained on pairs of noiseless and noisy representations of the signal. It is thus crucial to have access to accurate noiseless ground truth signal, which makes it challenging to apply these networks in areas where such ground truth is impossible or expensive to acquire, such as medical imaging [35]. To circumvent this problem, we used synthetic ground truth cryo-EM reconstructions and used a simplistic physical forward model to generate simulated projection images. In doing so, we explicitly did not aim to obtain a network to denoise reconstructions from experimental cryo-EM data. Rather, we used test data sets that were generated using the same forward model to provide a proof-of-principle that learned priors can improve the reconstruction process. Our results convincingly make this case: the RED approach outperformed standard auto-refinement in Relion in all tests performed.

The standard approach in Relion uses an L2-Tikhonov regularisation on the Fourier components of the reconstruction [2]. In practice, this purely Fourier-based regularisation term is often complemented with an *ad hoc* regularisation in real-space by setting all densities to a constant value in the volume outside a user-specified three-dimensional mask around the reconstruction. Often, such masks are generated after an initial refinement has yielded a preliminary reconstruction. In the tests performed here, no solvent-mask was provided. Thereby, the improvements observed by the RED approach reflect the difference between a purely Fourier-based L2-Thikonov regularisation and learned prior in real-space. The observed differences with masked refinements would be smaller.

A major identified pitfall of this method is the risk of injecting similar features that resemble those of biological macromolecules into both half-maps, which then lead an over-estimation of resolution based on the half-map FSC. Without access to ground truth, the half-map FSC is the single most important measure of quality for the reconstruction, which make this issue a major concern for the usefulness of the method. However, we were able to resolve this issue with an empirical approach that we call confidence-weighted RED. By falling back onto the less informative Gaussian prior at nominal resolutions where the denoiser is known to produce little improvement in the reconstruction, we managed to sufficiently alleviate the problem, while retaining equal overall performance. This approach is based on performance statistics gathered by applying the denoiser to a representative data set, e.g. the training or test data set, and requires no significant additional effort compared to regular RED. It is noteworthy that the denoiser performance likely varies over different resolution shells. Therefore, a more detailed investigation into the confidence weighting scheme could lead to further improvements.

We envision multiple other avenues for improving the approach set out in this paper. Firstly, different network architectures may be devised and networks may be trained through alternative loss functions to improve their performance, foremost in the low resolution range. A large variety of successful network architectures have been developed in other areas of image processing, e.g. FFDNet [36] or RCAN [37]. Furthermore, the denoiser may be trained on smaller patches of the reconstruction, rather than on the entire map as done here, and the patches may be combined using a sliding window. This approach is appealing due to the inherent flexibility in terms of particle size and training time. Moreover, the use of smaller patches may be intrinsically better suited to deal with the inherent variability in map quality that is caused by different extents of order that exist in many biological macromolecules. However, as networks trained on smaller patches no longer see the entire box, where a particle at the center is surrounded by solvent, this approach may be less powerful in flattening the solvent. One solution could be the combination of multiple denoisers that are trained in different resolution domains. Confidence weighting may then be employed to each of these denoisers according to their performance in each domain. For instance, a patch-based denoiser dedicated to the high-resolution domain could be combined with a denoiser with a global view of the map at low resolution.

Secondly, the networks should be optimised for the denoising of reconstructions from experimental cryo-EM data. Experimental cryo-EM data has different characteristics in both the signal and the noise than the synthetic data used here. Experimental noise is not independent in the pixels, but falls off with higher spatial frequencies. Both the signal and part of the noise are affected by the CTF of the microscope, and in particular the lowest spatial frequencies in experimental cryo-EM reconstructions are poorly modelled by the simplistic forward model of protein atoms in vaccuum used in this paper. Several options exist to generate more realistic pairs of maps for training the denoiser. Refinements with experimental cryo-EM data downloaded from the EMPIAR data base [38] may be used to generate training data. In such a scenario, generating image pairs with ground truth will be difficult, but one could generate noisy intermediate reconstructions from relatively small subsets of the data and provide the high-resolution maps from the complete data sets as substitutes for the ground truth. This is particularly feasible when the networks will only be trained on intermediate-resolution reconstructions

(also see below). Alternatively, one could train a convolutional neural network to learn a data-driven forward model from pairs of atomic models and their corresponding cryo-EM reconstructions, to generate more realistic ground truth maps, and then use generative adversarial networks (GANs) to generate the corresponding noisy data. For this purpose, conditional GANs (cGAN) are a particularly suiting candidate, since pairs of ground truth and noisy data are available [39]. Similarly, cycle-consistent adversarial networks (CycleGAN) [40] or style transfer GANs [41] may be used to relate the two data domains in cases where pairs are lacking, e.g. for low resolution reconstructions without a matching ground truth or for generalizing the data set to new experimental conditions with few existing examples [42].

Thirdly, the approach described here may be adapted to use optimisation algorithms other than expectation maximisation. A gradient-driven approach with mini-batches would for instance be based on an update formula derived from eq. (4) and (29) to inject prior knowledge into the reconstruction more often, which has the potential of improving convergence speed. For this class of algorithms, Adversarial Regularisers might be a more natural candidate compared to RED, since this method models the prior more directly and thus enables for better control of the properties of the generated gradient [43]. Alternatively, algorithms based on ADMM and plug-and-play denoisers are of potential interest for investigation [44, 45].

Finally, an important consideration when employing prior information in solving any inverse problem is that the prior information used in solving the problem can no longer be used for external validation of the results. This is relevant when injecting prior information about biological macromolecules into the cryo-EM reconstruction process. In current cryo-EM approaches such information is not used at all, and the appearance of protein and nucleic-acid-like features in the cryo-EM reconstructions is often implicitly used to confer confidence in the correctness of the maps. However, one could imagine a more informative prior generating reconstructions in which such features are incorrectly "invented" from noise in the maps. Commonly used procedures to prevent overfitting, most notably splitting the data set into two independent halves, do not necessarily protect against interpretation of such false features. Therefore, new validation tools may need to be devised to guard against the interpretation of false features. One option could be to only use information up to a given intermediate resolution when training the denoiser. In that way, the presence of protein or nucleic acid-like features beyond that resolution could still be used to provide confidence in the result. This approach is particularly appealing because the high resolution signal is typically not contributing much to the alignment of the individual noisy projection images anyway [46].

We expect that, with further improvements along the lines discussed above, the approach presented in this paper will result in a tool for structure determination that will outperform the current state-of-the-art, also for experimental cryo-EM data. This has the exciting potential to improve any reconstruction from existing data sets and to expand the applicability of the technique to samples that are currently beyond its scope.

6   Acknowledgements

We thank Jake Grimmett and Toby Darling for help with high-performance computing. C.B.S. acknowledges support from the Leverhulme Trust project on 'Breaking the non-convexity barrier', the Engineering and Physical Sciences Research Council (EPSRC: EP/M00483X/1, EP/S026045/1 and Centre EP/N014588/1), the RISE projects CHiPS and NoMADS, the Cantab Capital Institute for the Mathematics of Information and the Alan Turing Institute; O.Ö acknowledges support by the Swedish Foundation for Strategic Research grants (AM13-0049 and ID14-0055); S.H.W.S. acknowledges support by the UK Medical Research Council (MC_UP_A025_1013) and Wave 1 of the UK Research and Innovation (UKRI) Strategic Priorities Fund under the EPSRC (EP/T001569/1), particularly the "AI for Science" theme within that grant and The Alan Turing Institute.

References

[1] S. H. Scheres, A Bayesian view on cryo-EM structure determination, Journal of molecular biology 415 (2) (2012) 406–418.

[2] S. H. Scheres, RELION: implementation of a bayesian approach to cryo-EM structure determination, Journal of structural biology 180 (3) (2012) 519–530.

[3] A. Punjani, J. L. Rubinstein, D. J. Fleet, M. A. Brubaker, cryosparc: algorithms for rapid unsupervised cryo-em structure determination, Nature methods 14 (3) (2017) 290.

[4] M. Hu, H. Yu, K. Gu, Z. Wang, H. Ruan, K. Wang, S. Ren, B. Li, L. Gan, S. Xu, et al., A particle-filter framework for robust cryo-EM 3D reconstruction, Nature methods 15 (12) (2018) 1083–1089.

[5] F. J. Sigworth, A maximum-likelihood approach to single-particle image refinement, Journal of structural biology 122 (3) (1998) 328–339.

[6] S. H. Scheres, M. Valle, R. Nuñez, C. O. Sorzano, R. Marabini, G. T. Herman, J.-M. Carazo, Maximum-likelihood multi-reference refinement for electron microscopy images, Journal of molecular biology 348 (1) (2005) 139–149.

[7] S. H. Scheres, H. Gao, M. Valle, G. T. Herman, P. P. Eggermont, J. Frank, J.-M. Carazo, Disentangling conformational states of macromolecules in 3D-EM through likelihood optimization, Nature methods 4 (1) (2007) 27–29.

[8] T. Grant, A. Rohou, N. Grigorieff, cistem, user-friendly software for single-particle image processing, Elife 7 (2018) e35383.

[9] T. Moriya, M. Saur, M. Stabrin, F. Merino, H. Voicu, Z. Huang, P. A. Penczek, S. Raunser, C. Gatsogiannis, High-resolution single particle analysis from electron cryo-microscopy images using sphire, JoVE (Journal of Visualized Experiments) 123 (2017) e55448.

[10] P. A. Penczek, Fundamentals of three-dimensional reconstruction from projections, in: Methods in enzymology, Vol. 482, Elsevier, 2010, pp. 1–33.

[11] J. Frank, Electron tomography: methods for three-dimensional visualization of structures in the cell, Springer Science & Business Media, 2008.

[12] H. W. Engl, M. Hanke, A. Neubauer, Regularization of inverse problems, Vol. 375, Springer Science & Business Media, 1996.

[13] A. P. Dempster, N. M. Laird, D. B. Rubin, Maximum likelihood from incomplete data via the em algorithm, Journal of the Royal Statistical Society: Series B (Methodological) 39 (1) (1977) 1–22.

[14] R. Fernandez-Leiro, S. H. Scheres, Unravelling biological macromolecules with cryo-electron microscopy, Nature 537 (7620) (2016) 339–346.

[15] D. A. Forsyth, J. Ponce, Computer vision: a modern approach, Prentice Hall Professional Technical Reference, 2002.

[16] W. Rawat, Z. Wang, Deep convolutional neural networks for image classification: A comprehensive review, Neural computation 29 (9) (2017) 2352–2449.

[17] M. Arjovsky, S. Chintala, L. Bottou, Wasserstein gan, arXiv preprint arXiv:1701.07875.

[18] S. Arridge, P. Maass, O. Öktem, C.-B. Schönlieb, Solving inverse problems using data-driven models, Acta Numerica 28 (2019) 1–174.

[19] K. Zhang, W. Zuo, Y. Chen, D. Meng, L. Zhang, Beyond a gaussian denoiser: Residual learning of deep cnn for image denoising, IEEE Transactions on Image Processing 26 (7) (2017) 3142–3155.

[20] W. Jifara, F. Jiang, S. Rho, M. Cheng, S. Liu, Medical image denoising using convolutional neural network: a residual learning approach, The Journal of Supercomputing 75 (2) (2019) 704–718.

[21] L. Gondara, Medical image denoising using convolutional denoising autoencoders, in: 2016 IEEE 16th International Conference on Data Mining Workshops (ICDMW), IEEE, 2016, pp. 241–246.

[22] J. Adler, O. Öktem, Solving ill-posed inverse problems using iterative deep neural networks, Inverse Problems 33 (12) (2017) 124007.

[23] J. Adler, O. Öktem, Learned primal-dual reconstruction, IEEE transactions on medical imaging 37 (6) (2018) 1322–1332.

[24] K. Hammernik, T. Klatzer, E. Kobler, M. P. Recht, D. K. Sodickson, T. Pock, F. Knoll, Learning a variational network for reconstruction of accelerated mri data, Magnetic resonance in medicine 79 (6) (2018) 3055–3071.

[25] M. Mardani, E. Gong, J. Y. Cheng, S. S. Vasanawala, G. Zaharchuk, L. Xing, J. M. Pauly, Deep generative adversarial neural networks for compressive sensing mri, IEEE transactions on medical imaging 38 (1) (2018) 167–179.

[26] Y. Romano, M. Elad, P. Milanfar, The little engine that could: Regularization by denoising (red), SIAM Journal on Imaging Sciences 10 (4) (2017) 1804–1844.

[27] E. T. Reehorst, P. Schniter, Regularization by denoising: Clarifications and new interpretations, IEEE Transactions on Computational Imaging 5 (1) (2018) 52–67.

[28] R. Henderson, A. Sali, M. L. Baker, B. Carragher, B. Devkota, K. H. Downing, E. H. Egelman, Z. Feng, J. Frank, N. Grigorieff, et al., Outcome of the first electron microscopy validation task force meeting, Structure 20 (2) (2012) 205–214.

[29] A. Wilson, The probability distribution of x-ray intensities, Acta Crystallographica 2 (5) (1949) 318–321.

[30] G. Tang, L. Peng, P. R. Baldwin, D. S. Mann, W. Jiang, I. Rees, S. J. Ludtke, Eman2: an extensible image processing suite for electron microscopy, Journal of structural biology 157 (1) (2007) 38–46.

[31] O. Ronneberger, P. Fischer, T. Brox, U-net: Convolutional networks for biomedical image segmentation, in: International Conference on Medical image computing and computer-assisted intervention, Springer, 2015, pp. 234–241.

[32] M. Abadi, A. Agarwal, P. Barham, E. Brevdo, Z. Chen, C. Citro, G. S. Corrado, A. Davis, J. Dean, M. Devin, S. Ghemawat, I. Goodfellow, A. Harp, G. Irving, M. Isard, Y. Jia, R. Jozefowicz, L. Kaiser, M. Kudlur, J. Levenberg, D. Mané, R. Monga, S. Moore, D. Murray, C. Olah, M. Schuster, J. Shlens, B. Steiner, I. Sutskever, K. Talwar, P. Tucker, V. Vanhoucke, V. Vasudevan, F. Viégas, O. Vinyals, P. Warden, M. Wattenberg, M. Wicke, Y. Yu, X. Zheng, TensorFlow: Large-scale machine learning on heterogeneous systems, software available from tensorflow.org (2015).
URL https://www.tensorflow.org/

[33] D. P. Kingma, J. Ba, Adam: A method for stochastic optimization, arXiv preprint arXiv:1412.6980.

[34] V. Dumoulin, F. Visin, A guide to convolution arithmetic for deep learning, arXiv preprint arXiv:1603.07285.

[35] Y. Huo, Z. Xu, S. Bao, A. Assad, R. G. Abramson, B. A. Landman, Adversarial synthesis learning enables segmentation without target modality ground truth, in: 2018 IEEE 15th International Symposium on Biomedical Imaging (ISBI 2018), IEEE, 2018, pp. 1217–1220.

[36] K. Zhang, W. Zuo, L. Zhang, FFDNet: Toward a fast and flexible solution for cnn-based image denoising, IEEE Transactions on Image Processing 27 (9) (2018) 4608–4622.

[37] Y. Zhang, K. Li, K. Li, L. Wang, B. Zhong, Y. Fu, Image super-resolution using very deep residual channel attention networks, in: Proceedings of the European Conference on Computer Vision (ECCV), 2018, pp. 286–301.

[38] A. Iudin, P. K. Korir, J. Salavert-Torres, G. J. Kleywegt, A. Patwardhan, Empiar: a public archive for raw electron microscopy image data, Nature methods 13 (5) (2016) 387.

[39] P. Isola, J.-Y. Zhu, T. Zhou, A. A. Efros, Image-to-image translation with conditional adversarial networks, in: Proceedings of the IEEE conference on computer vision and pattern recognition, 2017, pp. 1125–1134.

[40] X. Zhu, Y. Liu, Z. Qin, J. Li, Data augmentation in emotion classification using generative adversarial networks, arXiv preprint arXiv:1711.00648.

[41] L. A. Gatys, A. S. Ecker, M. Bethge, Image style transfer using convolutional neural networks, in: Proceedings of the IEEE conference on computer vision and pattern recognition, 2016, pp. 2414–2423.

[42] M. Januszewski, V. Jain, Segmentation-enhanced CycleGAN, bioRxiv (2019) 548081.

[43] S. Lunz, O. Öktem, C.-B. Schönlieb, Adversarial regularizers in inverse problems, in: Advances in Neural Information Processing Systems, 2018, pp. 8507–8516.

[44] S. V. Venkatakrishnan, C. A. Bouman, B. Wohlberg, Plug-and-play priors for model based reconstruction, in: 2013 IEEE Global Conference on Signal and Information Processing, IEEE, 2013, pp. 945–948.

[45] S. Bigdeli, D. Honzátko, S. Süsstrunk, L. A. Dunbar, Image restoration using plug-and-play cnn map denoisers, arXiv preprint arXiv:1912.09299.

[46] R. Henderson, S. Chen, J. Z. Chen, N. Grigorieff, L. A. Passmore, L. Ciccarelli, J. L. Rubinstein, R. A. Crowther, P. L. Stewart, P. B. Rosenthal, Tilt-pair analysis of images from a range of different specimens in single-particle electron cryomicroscopy, Journal of molecular biology 413 (5) (2011) 1028–1046.

[47] E. Parzen, On estimation of a probability density function and mode, The annals of mathematical statistics 33 (3) (1962) 1065–1076.

7   Appendix

SMD describes the results of RED using a white Gaussian noise model with a fixed variance as the underlying assumption for the image distortion. In this appendix we present an approach for handling data with varying levels of distortion. In practice, this leads to a dynamically adjusted regularization parameter that is to be estimated during reconstruction, which is described in eq. (12). We show how this also leads to a training protocol that minimizes the model mapping to a coarsened ground truth.

We define the random vector $\mathrm{x} \sim P_{\text{true}}$, where P_{true} is the distribution of undistorted ground truth data. Here, we assume a Gaussian noise model formalized as $\tilde{\mathrm{x}} = \mathrm{x} + \boldsymbol{\epsilon}(\mathrm{x})$, where $\boldsymbol{\epsilon}(\mathrm{x}) \sim \mathcal{N}(\mathbf{0}, \boldsymbol{\nu}(\mathrm{x}))$ and $\boldsymbol{\nu}(\mathrm{x})$ is a $M \times M$ co-variance matrix that depends on x to describe the shape of the Gaussian. For the application in this paper this dependency does not have to be explicitly defined, rather an empirical approach is sufficient where the training data set is extended to also include observations of $\boldsymbol{\nu}$. In section 3.1 we specify how this was done in practice. One can device an empirical prior distribution that approximates P_{true} through Kernel Density Estimation (KDE) [47], given a large representative corpus of training data, $\{\boldsymbol{x}\}_{t=1}^T$, drawn from P_{true}, and the corresponding $\{\boldsymbol{\nu}_t\}_{t=1}^T$. This becomes

$$P_{\text{KDE}}(\boldsymbol{x}) := \frac{1}{T} \sum_t \mathcal{N}(\boldsymbol{x}|\boldsymbol{x}_t, \boldsymbol{\nu}_t). \tag{16}$$

We define an MMSE function as

$$\hat{\boldsymbol{x}}_{\text{MMSE}} = \arg\min_{\boldsymbol{f}} \mathbb{E}\left\{|\hat{\mathrm{x}} - \boldsymbol{f}(\tilde{\mathrm{x}})|^2\right\} \quad \text{with} \quad \hat{\mathrm{x}} = (\boldsymbol{\nu}(\mathrm{x}))^{-1}\mathrm{x} \tag{17}$$

In practice, due to the higher relative error of high-frequency components, $\hat{\mathrm{x}}$ corresponds to a coarsened (low-pass filtered) version of x. We also define

$$\boldsymbol{\psi} = \arg\min_{\boldsymbol{f}} \mathbb{E}\left\{\left|(\boldsymbol{\nu}(\mathrm{x}))^{-1} - \boldsymbol{f}(\tilde{\mathrm{x}})\right|^2\right\}. \tag{18}$$

Inspired by the proof for Tweedie's formula we then get

$$\nabla \log P_{\text{KDE}}(\boldsymbol{x}) = \frac{\nabla P_{\text{KDE}}(\boldsymbol{x})}{P_{\text{KDE}}(\boldsymbol{x})} \tag{19}$$

$$= \frac{\sum_t \boldsymbol{\nu}_t^{-1}(\boldsymbol{x}_t - \boldsymbol{x})\mathcal{N}(\boldsymbol{x}|\boldsymbol{x}_t, \boldsymbol{\nu}_{\boldsymbol{x}_t})}{\sum_t \mathcal{N}(\boldsymbol{x}|\boldsymbol{x}_t, \boldsymbol{\nu}_{\boldsymbol{x}_t})} \tag{20}$$

$$= \frac{\sum_t \boldsymbol{\nu}_t^{-1}\boldsymbol{x}_t\mathcal{N}(\boldsymbol{x}|\boldsymbol{x}_t, \boldsymbol{\nu}_t)}{\sum_t \mathcal{N}(\boldsymbol{x}|\boldsymbol{x}_t, \boldsymbol{\nu}_t)} - \frac{\sum_t \boldsymbol{\nu}_t^{-1}\mathcal{N}(\boldsymbol{x}|\boldsymbol{x}_t, \boldsymbol{\nu}_t)}{\sum_t \mathcal{N}(\boldsymbol{x}|\boldsymbol{x}_t, \boldsymbol{\nu}_t)}\boldsymbol{x} \tag{21}$$

$$= \hat{\boldsymbol{x}}_{\text{MMSE}}(\boldsymbol{x}) - \boldsymbol{\psi}(\boldsymbol{x})\boldsymbol{x}, \tag{22}$$

Now, it follows that

$$\hat{\boldsymbol{\theta}} = \arg\min_{\boldsymbol{\theta}} \mathbb{E}\left\{|\nabla \log P_{\text{KDE}}(\tilde{\mathrm{x}}) - \boldsymbol{\psi}(\tilde{\mathrm{x}})\left[\boldsymbol{f}_{\boldsymbol{\theta}}(\tilde{\mathrm{x}}) - \tilde{\mathrm{x}}\right]|^2\right\} \tag{23}$$

$$= \arg\min_{\boldsymbol{\theta}} \mathbb{E}\left\{|\nabla \log P_{\text{KDE}}(\tilde{\mathrm{x}}) + \boldsymbol{\psi}(\tilde{\mathrm{x}})\tilde{\mathrm{x}} - \boldsymbol{\psi}(\tilde{\mathrm{x}})\boldsymbol{f}_{\boldsymbol{\theta}}(\tilde{\mathrm{x}})|^2\right\} \tag{24}$$

$$= \arg\min_{\boldsymbol{\theta}} \mathbb{E}\left\{|\hat{\boldsymbol{x}}_{\text{MMSE}}(\tilde{\mathrm{x}}) - \boldsymbol{\psi}(\tilde{\mathrm{x}})\boldsymbol{f}_{\boldsymbol{\theta}}(\tilde{\mathrm{x}})|^2\right\} \tag{25}$$

$$= \arg\min_{\boldsymbol{\theta}} \mathbb{E}\left\{|\hat{\mathrm{x}} - \hat{\boldsymbol{x}}_{\text{MMSE}}(\tilde{\mathrm{x}})|^2 + |\hat{\boldsymbol{x}}_{\text{MMSE}}(\tilde{\mathrm{x}}) - \boldsymbol{\psi}(\tilde{\mathrm{x}})\boldsymbol{f}_{\boldsymbol{\theta}}(\tilde{\mathrm{x}})|^2\right\} \tag{26}$$

$$= \arg\min_{\boldsymbol{\theta}} \mathbb{E}\left\{|\hat{\mathrm{x}} - \boldsymbol{\psi}(\tilde{\mathrm{x}})\boldsymbol{f}_{\boldsymbol{\theta}}(\tilde{\mathrm{x}})|^2\right\}. \tag{27}$$

In the above, eq. (25) follows from eq. (19) - (22) and eq. (27) uses the orthogonality principle. Hence optimizing the weights, $\boldsymbol{\theta}$ through

$$\hat{\boldsymbol{\theta}} = \arg\min_{\boldsymbol{\theta}} \mathbb{E}\left\{|\hat{\mathrm{x}} - \boldsymbol{\psi}(\tilde{\mathrm{x}})\boldsymbol{f}_{\boldsymbol{\theta}}(\tilde{\mathrm{x}})|^2\right\} \tag{28}$$

equates matching the gradient of the log-KDE-prior through the modified RED formula:

$$\nabla \log P_{\text{KDE}}(\tilde{\mathrm{x}}) \approx \boldsymbol{\psi}(\mathrm{x})\left[\boldsymbol{f}_{\hat{\boldsymbol{\theta}}}(\tilde{\mathrm{x}}) - \tilde{\mathrm{x}}\right]. \tag{29}$$

Here, we propose to train the denoiser to learn $\boldsymbol{\psi}$, while still retaining $\boldsymbol{\psi}$ in eq. (29) during reconstruction. In effect this yields a denoiser that applies a low-pass filter in addition to denoising the map, which is a more

conservative approach that prevents uncertain high-frequency information to be injected into the refinement. Hence the optimization of the weights, $\hat{\boldsymbol{\theta}}$, is instead carried out through

$$\hat{\boldsymbol{\theta}} = \arg\min_{\boldsymbol{\theta}} \mathbb{E}\left\{|\hat{\mathbb{x}} - \boldsymbol{f}_{\boldsymbol{\theta}}(\tilde{\mathbb{x}})|^2\right\}. \tag{30}$$

However, we are left with estimating $\boldsymbol{\psi}$ in eq. (29), which is applied during refinement. Ignoring the off-diagonal components of $\boldsymbol{\psi}$ and estimating only the diagonal components through eq. (8), i.e. $\boldsymbol{\psi} \approx \boldsymbol{\tau}^{-2}$, yields eq. (29) equivalent to eq. (12), excluding the empirical parameter, λ.