Title

Multi-modal Nonlinear Optical and Thermal Imaging Platform for Label-Free Characterization of Biological Tissue

Authors

Wilson R Adams,[1] Brian Mehl,[2] Eric Lieser,[2] Manqing Wang,[3] Shane Patton,[2] Graham A Throckmorton,[3] J Logan Jenkins,[3] Jeremy B Ford,[3] Rekha Gautam,[3] Jeff Brooker,[2] E. Duco Jansen,[1,4] Anita Mahadevan-Jansen[1,4,*]

Affiliations

[1]Vanderbilt University, Dept of Biomedical Engineering, Nashville, TN, 37235, USA.

[2]Thorlabs Imaging Research, Sterling, VA, USA.

[3]Chongqing University, College of Bioengineering, Chongqing, China.

[4]Vanderbilt University Medical Center, Dept of Neurosurgery, Nashville, TN, 37232, USA

*anita.mahadevan-jansen@vanderbilt.edu

Abstract

The ability to characterize the combined structural, functional, and thermal properties of biophysically dynamic samples is needed to address critical questions related to tissue structure, physiological dynamics, and disease progression. Towards this, we have developed an imaging platform that enables multiple nonlinear imaging modalities to be combined with thermal imaging on a common sample. Here we demonstrate label-free multimodal imaging of live cells, excised tissues, and live rodent brain models. While potential applications of this technology are wide-ranging, we expect it to be especially useful in addressing biomedical research questions aimed at the biomolecular and biophysical properties of tissue and their physiology.

Main Text – [6468 words]

With increased focus on high-speed biological processes at molecular, chemical, and biophysical levels, there has come a critical need to develop tools that can keep pace with the number of active research areas in biomedicine. Technical developments in nonlinear optical microscopy have revolutionized our ability to study biophysical and biochemical properties of tissues, both in terms of the wide range of samples, improved contrast modalities, and unprecedented speeds at which imaging is now possible. However, combining direct spatial temperature measurements alongside multiphoton modes of contrast have not been widely adopted. Multiphoton imaging approaches have enabled deep-tissue imaging across numerous model systems, gleaning important insights on the structure and function of cells and tissues (*1, 2*) while thermal imaging can yield complementary information to relate to study metabolism, circulation, and immune response. Temperature plays a crucial role in biology yet has yet to be extensively demonstrated in any substantial capacity towards biological microscopy applications. Thus, instrumentation to explore the structural, functional, and thermal properties of tissue would prove useful in studying biophysical dynamics in tissue physiology.

Multiphoton fluorescence (MPF) and second harmonic generation (SHG), are particularly useful in highlighting tissue structure and function (*3, 4*). Calcium, voltage, and molecule-sensitive reporters continue to push the boundaries of multiphoton applications in biomedical research. Vibrational spectroscopic contrast using nonlinear Raman imaging (NRI), namely coherent anti-Stokes Raman scattering (CARS) and stimulated Raman scattering (SRS), has shown particularly wide-ranging applicability. These include rapid identification of tumor margins (*5*), high-dimensional structural labeling (*6*), other biological processes such as drug delivery (*7*), neurotransmitter release (*8*), and electrophysiological dynamics (*9*). Such imaging can highlight the molecular organization and biochemical composition of tissue over time–

offering a glimpse into sub- and inter-molecular dynamics without using exogeneous labels.
Multiplexing functional, biochemical, and biophysical (e.g. temperature) contrast in cells and
tissues can offer a unique tool in evaluating and imaging biological and physiological processes
that may not be possible with the same degree of exploratory power in separate instruments (*1*).
Nonlinear optical microscopy is inherently suitable for multimodal imaging applications.
Integrating the different modalities that fall under nonlinear microscopy can help alleviate
unnecessary system complexity while providing complementary information about tissue
structure and function. Notably, CARS, SRS, MPF, and SHG are all optical processes that can
occur and be detected simultaneously or sequentially by selecting the appropriate filters and
detectors for the techniques used. Several groups have demonstrated selective combinations of
nonlinear imaging strategies to address a range of biological questions, *in vitro* and *in vivo (10)*,
particularly sensing fast biological processes such as neuronal action potentials and calcium
activity with combined SRS and calcium-sensitive MPF (*9*). Integration of SRS with optical
coherence tomography (OCT) has been shown to augment nonspecific scattering-based contrast
with vibrational specificity to image lipid distributions in excised human adipose tissue (*11*).
More recently, researchers have employed four or more nonlinear imaging modalities, including
CARS, MPF, SHG, and third harmonic generation for live tissue and intravital imaging towards
wound healing and cancer metastasis (*12, 13*). While most of these demonstrated approaches
have been applied towards observing biological processes over the span of multiple hours at the
molecular and biochemical levels, they are limited in their ability to observe snapshots of sub-
second functional, biochemical, and thermal processes.
The influence of temperature in physiological and biochemical processes – particularly
surrounding sample damage and physiological modulation - has recently become an area of
prominence across biomedical research disciplines (*14-16*). As optical imaging and perturbation
technologies continue to advance biomedical research, it becomes critical to consider
thermodynamic effects particularly on live specimens albeit practically difficult. A flexible
imaging system that incorporates molecular, biochemical, and biophysical information from cells
to tissues, *in vitro* to *in vivo,* would enable the study of rapid and dynamic biophysical processes
from multiple perspectives.
Measuring temperatures in biological samples at high resolutions spatially and temporally
has proven difficult and continues to be an active area of research. Groups have multiplexed
blackbody thermal imaging and fluorescence microscopy at cellular resolutions *in vitro* (*17, 18*),
however water absorption generally obscures the ability to visualize cell morphology with
thermal imaging alone. Moreover, optical components that suffice for use between 0.4 and 14
μm wavelengths to encompass visible optical and thermal infrared imaging are not readily
available. Finite element heat transfer modeling and Monte Carlo simulations of photon transport
in scattering media is often regarded as the benchmark approach for temperature estimation of
dynamic photothermal processes in biological samples with high water concentrations (*19*).
Several indirect approaches have been demonstrated including temperature-dependent changes in
fluorophore emission, and probe beam deflection microscopy (*16, 19, 20*). Several indirect
approaches have been demonstrated including temperature-dependent changes in fluorophore
emission, and probe beam deflection microscopy (*16, 20*). While point-based methods (utilizing
thermocouples, customized miniature sensors, or intrinsic fluorescence of rare-earth doped glass
waveguides) are more direct, they sacrifice spatial information obtained with temperature
mapping (*21*). Previous work has utilized thermal imaging with laser speckle imaging to study
cerebral blood flow changes during different methods of anesthesia (*22*). However, despite the

potential utility of combining thermal imaging with multimodal nonlinear microscopy, such a

combination has not been previously reported. A system that can combine multimodal nonlinear

microscopy with thermal imaging would offer unprecedented flexibility to study biological

processes in real time.

The goal of this project is to integrate multiple nonlinear imaging methods with thermal

imaging so that biophysical, biochemical, and molecular information from dynamic biological

processes may be imaged with micron level spatial resolution and millisecond level temporal

resolution. Towards this effort, we present an imaging platform that integrates single-band

CARS, SRS, MPF, SHG, and wide field thermal microscopy (ThM) for characterizing tissues at

varying time scales. Applications of our imaging system are demonstrated imaging both *in vitro*

with neural cell cultures and tissue specimen as well as *in vivo* with acute craniotomy rat

preparations. Our platform offers a novel approach enabling imaging techniques with otherwise

incompatible optical paths (due to physical limitations of hardware) to be applied on a common

sample. We demonstrate that this platform provides a robust tool for researchers to probe the

physiology of dynamic biological systems, through the unique integration of vibrational

spectroscopic, functional fluorescence, and thermal contrast.

System Design

The framework of the Multimodal Advanced Nonlinear and Thermal Imaging System, or

MANTIS, is a three-armed imaging turret (Customized Bergamo II Platform, Thorlabs Imaging

Research, Sterling, VA, USA) that respectively images a sample with either nonlinear imaging,

thermal imaging, or widefield white light reflectance imaging which is reconfigurable to adapt

additional measurement approaches (Fig. 1A). The mechanical arrangement of distinct imaging

arms allows for modalities with incompatible optical instrumentation (e.g. ultrafast near infrared

imaging with endogenous shortwave infrared measurements) to be performed sequentially

without disturbing or repositioning the sample. A FLIR SC8300-series high-speed indium-

antimonide CCD camera equipped with a 4X germanium imaging objective (FLIR Systems Inc.,

Nashua, NH, USA | Fig. 1B) is attached to the thermal imaging arm of the microscope. The

thermal camera relies on the endogenous black/grey body emission of a sample between 3-5μm

wavelength under the assumption of homogenous emissivity to infer sample temperatures.

The nonlinear imaging section of MANTIS, depicted in Fig. 1C with a simplified

schematic, is coupled to two nonlinear laser sources supplying femtosecond (Insight DS+,

Spectra Physics, Fremont, CA, USA) and picosecond (picoEmerald S, A.P.E, Berlin, DE) laser

lines. Higher spectral bandwidth (~15nm) femtosecond laser pulses provides high peak powers

necessary for optimal multiphoton and higher harmonic generation *in vivo (1)*. The picosecond

laser has narrower bandwidth (~0.5nm, or 10cm^{-1}), which is critical for maintaining the spectral

resolution necessary for single-band nonlinear Raman imaging while minimizing power at the

sample for safe *in vivo* applications (<20mW average power) *(23)*. Both laser systems operate

with a pulse repetition rate of 80MHz and output two beams necessary for pump and stokes

excitation of NRI contrast processes. The shorter wavelength tunable laser line undergoes a

tunable path length delay while the longer wavelength laser line is intensity-modulated at

20MHz via an electro-optical modulator (Thorlabs, Newton, NJ, USA) to facilitate SRS. Each

laser's output is spatially and temporally co-linear and directed with a series of mirrors into the

scan head of the nonlinear imaging arm of MANTIS.

The imaging optics of the nonlinear imaging arm are based on conventional upright laser-

scanning microscopy. The scanning optics consist of a pair of galvanometric mirrors (Thorlabs,

Newton, NJ, USA), which are imaged onto the back focal plane of a commercial objective lens

(Olympus XLUMPLFLN, 20X 1.0NA | Nikon CFI Apochromat NIR 60X, 1.0NA) via a 4f
optical relay. This scan relay performs a fixed 4X magnification of the laser beam diameter to
accommodate the back-aperture pupil size of the largest objective lenses we use relative to the
entrance beam diameter (SL50-2P2 & TL200-2P2, Thorlabs, Newton, NJ, USA). Two epi-
detection ports with a photomultiplier tube (CARS/MPF/SHG | GaAsP Amplified PMT,
Thorlabs, USA) and a large area 50-V reverse-biased photodiode (SRS | A.P.E. GmbH, Berlin,
DE) are used for imaging epi-detected contrast. Two additional detachable forward detection
ports were also built to accommodate coherent imaging modalities in transparent samples in
transmission mode. Stimulated Raman loss for SRS was demodulated via a commercial lock-in-
amplifier (A.P.E. Gmbh, Berlin, DE). All emission filters and excitation wavelengths included
with the system are summarized in Table S1 (Semrock, Brattleboro, VT, USA). Scanning and
detection hardware for imaging is controlled through ThorImageLS version 2.1, 3.0, or 3.2
(Thorlabs Imaging Research, Sterling, VA, USA). All images shown are raw with linear
intensity rescaling, with all analysis being performed in FIJI (*24*).
The third imaging arm of MANTIS initially features a color complementary metal oxide
semiconductor (CMOS) camera (Thorlabs Inc., Newton, NJ, USA) with a variable focal length
lens (Navitar Inc., Rochester, NY, USA) for widefield white-light reflectance imaging of
samples (Fig. 1A). However, this arm was built into the imaging system to readily enable
integration of additional contrast modalities that may be further incompatible with multimodal
nonlinear microscopy – such as interferometric contrast or stereoscopic surgical guidance. Each
arm is locked into position by a custom-designed spring-loaded locking system against the
rotating turret base (Fig. 1A). The maximum fields of view of each imaging arm is ~800µm-x-
800µm for nonlinear microscopy with a 20X objective, 3mm-x-4mm for thermal microscopy
with a 4X objective, and 50mm-x-50mm for widefield white light reflectance imaging.
System Performance
To achieve high-speed imaging at subcellular resolution, MANTIS was designed to resolve at
least 1-µm lateral resolution in the nonlinear imaging arm. Performance was verified on a series
of standards, (Fig. **2**), with measured axial and lateral resolutions for each modality summarized
in Table S2 (data summarized in Table S1). Images of 0.5-µm diameter polystyrene (PS, latex,
Polysciences Inc. Warrington, PA, USA) beads with CARS and SRS were obtained to generate
a point spread function (PSF), which is reported here as the full-width half-maximum intensity
cross-section of beads fit to a Gaussian curve. Fluorescent bead calibration standards with a 250-
nm diameter (Polysciences Inc. Warrington, PA, USA) were used to measure the PSF of MPF.
All PSF calculations were performed using the MetroloJ plugin in FIJI (*25*). Measured
resolutions, reported in Table S2, are on par with results published literature when imaging with
numerical apertures approaching 1.0 (*3, 23, 26*). Spectral separation of polymethyl-methacrylate
(PMMA, acrylic, Polysciences Inc. Warrington, PA, USA)) and PS beads with single-band SRS
are shown in Fig. **2**B. Acrylic (PMMA) beads are depicted using the 2927-cm^{-1} band – an
asymmetric CH_3 resonance- (yellow), and latex (PS) beads are depicted spectrally separate with
the 3053-cm^{-1} band– an asymmetric CH_2 resonance- in blue. The signal-to-noise ratio (SNR),
measured at the edge of a vegetable oil meniscus at the 2927-cm^{-1} band with SRS (Fig. S2), was
calculated to be 34.6 (*27*).
Resolution standards to verify SHG imaging resolution are not commercially available. In
place of a controlled standard, a collagen-rich biological sample (a porcine mitral valve) was
imaged at a high spatial sampling density (250-nm/px) and the finest resolvable fibrillar
structures were measured as a proxy for lateral resolution. Sample images of mitral valve

collagen are shown in Fig. **2**E and Fig. S3. Example calculations of fibril diameter are show in
Fig. S1. Fibrillar collagen bundles and quaternary structure are clearly visualized. While fibrillar
structures vary substantially in diameter (*3*), the resolving power of SHG on MANTIS (Table 2)
is well within our targeted resolution goals being able to resolve sub-micron spatial features.
The FLIR thermal camera resolution was estimated by identifying the smallest resolvable
group on a 1951 United States Air Force resolution target. Resolving group 6, element 4 equates
to a lateral resolution of 6.9-μm. The depth of focus of the thermal camera is about 40-μm, based
on an f/4.0 aperture stop of the objective lens in the thermal camera. Since the depth of focus of
the thermal camera is substantially larger than the depth of focus of the nonlinear imaging
system, registration in the axial dimension straightforward. The factory measured temperature
resolution is specified to be 0.1°C. The nonlinear and thermal microscopy field view were
independently adjusted to centrally registered the corner of a 10-um grid target. Switching
between thermal and nonlinear microscopy fields of view was found to repeatable within 1-μm
radial to imaging turret rotation and 10-μm tangent to imaging turret rotation (Fig. 1A, yellow
arrows). Since field of view of the thermal camera is nearly four times that of the nonlinear
microscopy field of view – making the coincidence of each modality's field of view with each
other straight-forward Figure S5. Once aligned, the fields of view were observed to remain well
registered for more than a week following initial alignment. It takes about 90-180 seconds to
switch between nonlinear and thermal imaging arms over the same sample. The sub-millimeter
repeatability of positioning multiple imaging arms readily enables multiplexing of thermal and
nonlinear microscopy on the same field of view without needing to reposition the sample.
***In vitro* Imaging**
To verify the capability of MANTIS to perform *in vitro* imaging, cultured NG108 cells – a
spiking neuron-astrocytoma hybridoma cell line - were imaged with CARS and MPF, as seen in
Fig. **3**. Briefly, cells were plated on poly-D-lysine coated cover glass for 48hr prior to imaging
and maintained in Dulbecco's Modified Eagle Medium (DMEM) supplemented with 10%v/v
fetal bovine serum and 5mM L-glutamine. 12-hours prior to imaging, medium was replaced with
DMEM containing 3%v/v FBS to morphologically differentiate adherent cells. Performing
CARS imaging at 2830-cm^{-1} resonance highlights a lipid-dominant CH_2 symmetric stretch mode,
which relates to cellular projections and lipid-rich intracellular contents such as organelles, lipid
droplets, and vesicles (Fig. **3**A). Autofluorescence from NADH and FAD can be used to
extrapolate metabolic information relating to relative aerobic metabolism dynamics and
metabolic cofactor distribution throughout the cell, Together, combined CARS and MPF
demonstrate the structural and functional capacities of multimodal nonlinear imaging *in vitro*
(Fig. **3**B and C). Blackbody thermal imaging of cells in media in an upright configuration was
not feasible due to the strong absorption of water in the wavelength range measured by the
thermal camera (3 to 5-μm). Imaging with cellular resolution with thermal microscopy is
possible in an upright configuration (Figure S4A), with the ability to resolve cellular temperature
distribution. However, once medium is introduced to the imaging field of view (Figure S4B),
cellular morphology is occluded due to strong water absorption of the sensed wavelength range.
This has been addressed by others by utilizing inverted imaging configurations (*17, 28, 29*) ,
however resolvable cellular morphology were not demonstrated.
Combined SHG and MPF was demonstrated on fresh *ex vivo* porcine mitral valve
samples (Figure S5). Tissue samples were harvested from porcine cardiac tissue postmortem and
stored in phosphate buffered saline for 72hrs prior to imaging. Samples were mounted on a
standard microscopy slide and imaged through leveled and supported cover glass in contact with

the tissue surface. Simultaneous autofluorescence and SHG under 930-nm illumination with the
femtosecond laser source yielded high resolution microstructure of endogenous elastin and
collagen, free of exogenous labeling (*30*). Changes in collagen and elastin can dangerously
disrupt the structure and function of heart valves (*31*). Having methods to study new
interventions to improve heart valve function could be valuable to cardiac biomechanics
researchers. Combining SHG with SRS at 2880-cm^{-1} to broadly highlight lipids allows for the
visualization of tissue architecture in unfixed frozen sections of murine cervix (Figure S6).
Distinct chemical and structural differences are notable between the epithelium and stroma of the
cervix. The stroma is dense with collagen and low in cellular content, while the epithelium is rich
in lipids and proteins due to cellularization. The lamina propria of blood vessels are also
discernable with SHG, appearing as fine tubular structures with strong SHG signal, providing an
indication of tissue vascularization. Visualizing the changes in vasculature, cellular, extracellular
matrix protein distribution throughout cervix shows potential for studying parturition processes
and cancer progression in the cervix without the need for exogenous labelling.
245   Fig. **4**, Figure S6, and Figure S7 demonstrate multimodal nonlinear imaging with CARS,
SRS, SHG, and MPF in an ex vivo rat sciatic nerve preparation. Nerves were harvested from
Sprague-Dawley rats and imaged immediately postmortem. Contrast from CARS and SRS
primarily highlights myelin (Figure S7B&C). The CARS and SRS images correspond well with
the MPF images of FluoroMyelin green (ThermoFisher Inc | Figure S7E, Figure S8C and D) but
are less prone to heterogeneous dye uptake. While myelin plays a major functional role in nerve
signal conduction, collagen is a major structural component of the sciatic nerve that provides a
protective exterior sheath and additional mechanical stability within the nerve. Collagen contrast
from SHG can be used to visualize the epineurium and fascicle-residing collagen (Figure S9).
These four modalities together therefore offer key structural insight to the sciatic nerve and can
be used to study nerve injury, regeneration, and disease. Furthermore, the ability to flexibly
apply these techniques to different sample types, between live cells and excised tissue specimen,
highlights a key design goal of the imaging platform.
258   The key advancement allowed by the MANTIS platform is the ability to multiplex
optically incompatible imaging techniques such as thermal imaging with nonlinear optical
microscopy. Thermal imaging can be used to evaluate fast photothermal processes, such as
infrared neural stimulation (*32, 33*), with coregistered contrast from nonlinear imaging
modalities to study real time biophysical dynamics. To assess the multimodal capabilities of
MANTIS, readout of multimodal nonlinear contrast was directly correlated with high-speed
temperature changes during infrared neural stimulation in a fresh *ex vivo* sciatic nerve
preparation (Fig. **4**). Point-scanned line repeats were employed to obtain high framerate
nonlinear signals during IR stimulation. Intriguingly, our results show a notable decrease in
nonlinear signals that tracks closely with temperature increases over the period of one second.
Notably, SHG signals did not exhibit a discernable decrease in signal, which is attributable to the
lack of collagen present through where line scans were acquired (center of the field of view, Fig.
4). However, decreases in SHG signal during IR stimulation have been verified in other FOVs,
suggesting this signal decrease is a more general physical phenomenon. Decreases in nonlinear
signal were found to be caused by a defocusing artifact induced by the spatial thermal gradient
inherent to infrared neural stimulation in the microscope's field of view. Resolving photothermal
effects of IR heating in biological tissues at high speed, such as during INS, will provide
invaluable information in studying *in vivo* applications of INS.

In vivo Imaging

An acute craniotomy preparation in an anesthetized rat model was utilized to validate the *in vivo* capabilities of MANTIS. All animal use protocols were approved by Vanderbilt University Institute for Animal Care and Use Committee (VU-IACUC). Sprague-Dawley rats were anesthetized with intraperitoneal injections of ketamine (40-80mg/kg) and xylazine (10mg/kg) for 40 minutes prior to surgery. Anesthetic depth was monitored every 10 minutes and maintained with follow-up half-dose anesthetic injections every 90-120 minutes as needed. Animals were mounted in a stereotaxic frame under the microscope objective while a 2mm-x-2mm section of skull and dura mater overlying the somatosensory cortex was carefully removed. Tissue hydration was maintained with sterile saline throughout imaging. Animals were sacrificed by anesthetic overdose and cervical dislocation following experiments. During imaging experiments, pixel dwell times needed to be increased to 8-µs to account for signal loss in moving the SRS detection path into an epi-detection configuration. Epi-SRS detection was implemented by replacing the dichroic mirror immediately preceding to the objective in the incident light path with a polarizing beam splitter cube and quarter wave plate (*23*). Images were initially acquired at a high spatial sampling density (0.5-µm/px, 800x800 px) with a 0.2Hz framerate. However, reducing the spatial sampling density and overall image sizes (3µm/px, 256x256 px) over a comparable field of view, we were able to achieve 2Hz framerates. Subsequently, repeated point-based line scanning could be implemented with a 488Hz line rate to observe faster biophysical dynamics. This flexibility emphasizes the trade-offs between spatial and temporal sampling in nonlinear microscopy approaches *in vivo*. Acquiring images using CARS, MPF, or SHG needed to be performed sequentially, due to the availability of only one PMT detector at the time of experiments. However, any one of the three PMT-requiring modalities can be multiplexed with SRS contrast, since SRS uses a separate detector in an adjacent detection path. Excitation and detection wavelengths are summarized in Table S1.

Examples of *in vivo* images can be seen in Fig. **5**. A lipid-rich CARS resonance at 2850cm^{-1} (Fig. **5**C) and a protein-dominant SRS resonance at 2930cm^{-1} (Fig. **5**B) appears to highlight superficial cellular features as well as vascular perfusion within blood vessels. Autofluorescence contrast from blue and green detection channels are most commonly used to measure endogenous NADH (Fig. **5**G, Table S1) and FAD (Fig. **5**E, Table S1). This diffuse NADH signal is likely arising from the neuropil at the pial surface of the cortex, while FAD tends to present more sparsely superficially in the outer cortex *in vivo*; this weak fluorescence is often spectrally indistinct from other biological autofluorescence (*34, 35*). Contrast from SHG (**Fig**. 5 revealed unknown morphologies, which may be due to cytoskeletal structures across the pial surface, but is more likely to be residual collagen from the dura mater after surgical preparation. Thermal imaging (**Fig**. 5H) of the brain surface appears to reveal mostly topographical contrast - vasculature is clearly visible. The lack of optical penetration depth of measured wavelengths of the thermal camera are likely responsible for the brain surface topographical contrast. Temperature fluctuations across the surface of the brain appear to be minimal, even with breathing, heart rate, and motion artifacts. Larger vessels appear slightly warmer than smaller vessels. Focusing artifacts appeared to have some effect on the accurate approximation of the brain surface temperature within 1°C – particularly noticeable at the edges of the field of view, which likely arise from out of focus signal from underlying bulk tissue due to the high degree of surface curvature in the sample (**Fig**. 5H). Cellular morphologies were not apparent in thermal images of the brain surface. Thermal imaging of the brain surface was capable of acquisition speeds approaching 180Hz framerates with a full field of view. This

imaging rate can be substantially increased to 500Hz or more by cropping the image sensor readout without loss in spatial resolution or pixel binning without reduction in field of view. The vessel morphologies visible between thermal and CARS/SRS contrast made fine registration of thermal and nonlinear imaging fields straightforward with rigid registration of manually labelled image features.

Combined Multimodal Nonlinear and Thermal Microscopy

Infrared neural stimulation (INS) utilizes pulsed short-wave infrared laser light to transiently invoke a thermal gradient in neural tissue resulting in activation of neural cells. One of the major concerns about neuromodulation with rapid targeted thermal gradients is that the change in temperature could cause cell damage. Cells are naturally prone to damage at elevated temperatures; however, the role of temperature-time history is often underappreciated in the context of biological thermodynamics. Integrating fast nonlinear and thermal microscopy offers a particularly useful platform for studying the physical and functional impacts of INS on neural cells.

The ability to correlate high resolution temporal and spatial thermal information (Fig. 6E&F) with the functional information afforded by nonlinear imaging (Fig. 6A-D), available in the MANTIS platform offers a unique opportunity to visualize biochemical, biophysical, and biomolecular dynamics of cultured NG108 cells during INS. An 8ms pulse of 1875nm infrared light was delivered to the cells via a 400μm diameter optical fiber at average radiant exposures spanning 0.5 (stimulating) and 3J/cm^2 (damaging) while imaging at 10Hz framerates. Multiphoton fluorescence of cells imaged in saline containing a damage indicator, propidium iodide (PI, 1uM concentration, Fig. 6C, Table S2), differentiated healthy and damaged cells due to IR exposure. Increases in relative fluorescence greater than 3% were presumed to be indicative of cell damage. Simultaneously, SRS imaging at a CH$_3$ resonance (2930cm^{-1}, lipid/protein) provides endogenous biochemical information of cells during INS. In this study, SRS images were primarily used to segment cell morphologies to extract viability status (MPF) and interpolate thermodynamics (thermal imaging).

Thermal images during INS were acquired at 34Hz framerates after all nonlinear imaging and stimulation experiments were completed. Temperature information is not simultaneously available with nonlinear observations due to physical limitations. Since the thermal properties of cells mounted in imaging saline behave thermodynamically like water upon INS, post-hoc registration of thermal information with *a priori* knowledge of thermal and nonlinear fields of view was used. Multimodal image registration was performed with an iterative closest point algorithm implemented in MATLAB to register manually selected features of a fluorescent target (Chroma, Rochester, NY, USA) observable in both nonlinear and thermal imaging modes. Cell centroids were calculated from cell morphologies identified with SRS (Fig. 6A) in FIJI utilizing a seeded watershed segmentation algorithm (*24, 36*). Spatial transformations from aforementioned field of view registration were applied to the cell centroid positions to obtain their positions in the thermal camera field of view. Cell positions in the thermal camera field of view were used to interpolate temperature-time and temperature-space thermal data on a per-cell basis utilizing built-in 2D interpolation functions in MATLAB.

Since the stimulating IR light (1875nm) is strongly absorbed by water, the immersion medium required for nonlinear imaging had a substantial impact on the heating of cells, leading to potentially large discrepancies in thermal imaging measurements. To minimize optical absorption of 1875nm stimulation light by the aqueous immersion medium necessary for the nonlinear microscopy objective (Olympus XLUMPLNFL 20X, 1.0NA), spectroscopic-grade

deuterated water (Sigma Aldrich, St. Louis, MO, USA) was used as an immersion medium (*33*).
Cells remained immersed in normal imaging media and were separated from heavy water by the
coverslip on which the cells were mounted. The reduction in IR absorption in the objective lens
immersion media was enough to allow for representative spatial thermal measurements of INS
on aqueous cellular samples with the thermal camera. Functional observations of cell death
aligned well with peak spatial temperature maps (Fig. 6E, inset). It is expected that absolute
temperature measurements at the sample are likely to differ with and without the presence of
deuterated immersion medium during INS, however the spatial and temporal dynamics of IR-
induced cellular thermal gradients are expected to remain consistent and comparable across all
exposure conditions. With this assumption, physiological comparisons can be effectively drawn
with the understanding that the observed absolute temperatures are likely slightly higher than
during nonlinear image experiments where immersion medium may impact INS.
Spatially registered observations of cell functional, chemical, and temperature dynamics
is only made possible by a combined and accurately coregistered nonlinear and thermal imaging
platform such as the MANTIS platform. From these multimodal imaging experiments, we
observed that cells that experience more rapid changes in temperatures as a function of time were
more likely to be damaged (Fig. 6E, F). Damaged cells are more likely to be located near low
spatial thermal gradient values, which correspond to local maxima in spatial heating profiles
(Fig. 6E, inset). However, cells outside the fiber illumination would be expected to present
similar spatial thermal gradients as cells at peak levels of heating - introducing some
heterogeneity in the spatial thermal gradient information. Considering spatial and temporal
thermal gradient information together clarifies the thermodynamic difference between heated and
unheated cells. Numerous cells appear to survive rapid temperature changes while others do not,
however damaged cells are far more likely to undergo rapid heating. This observation
demonstrates the variability in cell physiological responses due to INS. Distinct increases in
mean multiphoton fluorescence timeseries indicate cell damage within 30 seconds of IR
exposure. Cell damage appears to coincide with increases in mean SRS CH_3 signal following IR
stimulation. The basis for SRS signal increase is currently unclear, though we speculate
attribution to endogenous cellular damage responses (e.g. lipid vacuolization, organelle damage,
and increased chaperone protein expression such as heat shock proteins). There remains to be
extensive validation of changes in endogenous SRS signal in the context of live cell imaging and
damage beyond lipid storage and cell membrane dynamics. Experimentally, SRS images were
primarily used in Fig. 6 to identify, segment, and locate cells for temperature and viability data
calculations. However, this platform readily enables such explorations into the molecular basis of
SRS signal changes simultaneously with more established live cell fluorescence imaging probes.
Discussion
We demonstrate the first imaging system that combines CARS, SRS, MPF, SHG, and thermal
imaging into a single microscope for biomedical imaging applications. The similarities in
illumination and detection instrumentation needed for nonlinear excitation has been exploited in
the past to design multimodal imaging platforms. But as demonstrated here, our novel system
design includes a movable turret that allows for overlaying these different modalities to achieve a
more complete picture of biological processes to be imaged. This logically suggests that another
imaging modality that typically could not be integrated into a nonlinear optical microscopy path
may be similarly incorporated. Such flexibility of optical microscopy design provides the
opportunity to study physiology in unique and dynamic ways.

The proposed imaging system, MANTIS, was built with modularity and expandability in
mind. On the nonlinear imaging arm, the current configuration has four optional detection ports -
two each in the epi- and forward detection configurations. Additional channels can easily be
added to expand simultaneous imaging capabilities based on research needs. Lock-in detection
arms can be refitted with the appropriate optical filters for transient absorption imaging. Sum-
frequency generation imaging can also be readily integrated for studying ordered molecular
orientation and interfacial phenomena in biological samples. The widefield reflectance imaging
arm may be useful for integrating laser speckle, diffuse reflectance, optical coherence
tomography and microscopy, or spatial frequency domain imaging with the correct illumination
optics to map tissue blood flow, oxygenation, and optical properties. Furthermore, the widefield
reflectance and thermal imaging arm were built on detachable 96-mm optical rails so that other
imaging methods may be integrated based on future research needs. The multi-armed imaging
concept can also yield more utility out of a condensed instrument footprint, which may be useful
in places where laboratory floorspace is limited.
The scanning optics used in nonlinear imaging arm can achieve 1kHz scan rates. A
galvo-scanning pair was chosen over a faster galvo-resonant scanning pair, which typically offer
an order of magnitude increase in scanning rates. By doing so, improved control over image
sampling densities can be achieved to readily span subcellular and multicellular scales. This is
exemplified by the difference in sampling density observed between Fig. **3** and Fig. **5**. As
demonstrated in Fig. **4**, our imaging system can perform nonlinear imaging with line scans
approaching 0.5kHz with detectable amounts of signal. This is enough to resolve high-speed
biophysical dynamics, such as during neural modulation or optogenetic stimulation, from a
functional, chemical, and physical standpoint. Kilohertz bandwidth scanning rates are sufficient
to achieve visualization of neuronal action potentials as demonstrated by Lee *et al.* with balanced
detection SRS and calcium fluorescence microscopy (*9*).
Both femtosecond and picosecond lasers were included in the design of MANTIS for
signal and spectral resolution considerations. The bandwidth of the picosecond laser source
(around 0.5nm) allows for higher spectral resolutions (around $10cm^{-1}$) which is critical for
nonlinear Raman imaging (*37*). However, the peak powers of the picosecond source are orders of
magnitude lower than that of the femtosecond laser, yielding less signal during multiphoton
fluorescence and higher harmonic imaging. At the time of construction, the use of broadband
nonlinear Raman techniques, such as pulse shaping and spectral focusing had yet to be realized
for fast imaging *in vivo* (*38*). As such, design considerations to optimize narrow band spectral
resolution and *in vivo* imaging speed were a priority. A number of broadband spectral NRI
techniques have since been described (*39, 40*), with a handful demonstrated *in vivo* (*10*). The
flexibility of the existing instrumentation on MANTIS readily allows for the integration such
broadband approaches.
Using blackbody thermal microscopy to characterize sample temperatures in conjunction
with nonlinear imaging, or any laser-scanning microscopy approach, has yet to be previously
published. Estimating temperature with a thermal camera provides a more direct measurement
than other approaches, such as fluorescence-based methods. The ability to measure thermal
information alongside the functional and structural information offered by nonlinear microscopy
provides a unique instrument to explore new questions in biophysics, particularly *ex vivo* and *in*
vivo. Practically, thermally microscopy presents some advantages and drawbacks depending on
the model systems being imaged. We have found that high resolution fluorescence microscopy
can be difficult to interpret in highly thermally dynamic systems with high numerical apertures at

high imaging speeds, which became apparent in applying infrared neural stimulation in sciatic
nerve (Fig. **4**) as well as in cells. Having thermal information to contrast and systematically
compensate for thermally induced effects is something that our imaging platform lends itself to
accomplish. Since imaging depth with thermal microscopy is limited due to the optical
penetration depth of water in the 3-5µm wavelength spectral regime, this means that SWIR based
imaging methods are ideally suited for measuring surface temperatures in water-dominant
samples such as biological tissues. This is even more apparent *in vitro*, where imaging through
any amount of aqueous medium occludes cellular morphology (Figure S4). While others have
demonstrated blackbody thermal imaging of cells via inverted microscopy through cover glass
(*17, 18*), visualizing cell morphology with thermal imaging has not been previously
demonstrated. Thermal microscopy on MANTIS can be performed readily and reliably on
tissues *ex vivo* (Fig. **4**F) and *in vivo* (Fig. **5**H). Topographical features such as blood vessels
tissue surfaces are clearly visible. The pronounced topography of these anatomical features is
helpful when performing fine registration nonlinear and thermal imaging fields. Temperature at
aqueous interfaces are useful in approximating spatial temperature distributions *in vitro*.
Combining surface temperatures with finite element heat transfer modeling and precise control
over sample thermodynamics can be used to render volumetric estimates of temperature *in vitro*
(*18*). In combination with fluorescence or other nonlinear imaging modalities, imaging the
thermodynamics of adherent cells *in vitro* or in tissues with thermal microscopy could be quite
useful. Consequently, multimodal methods almost become an essential requirement to correlate
temperature and real-time cellular observations. Our microscopy platform readily allows such
observation to be made and flexibly conFig.d to address a wide range of biological and
physiological questions.
As expected, cell viability at elevated temporal thermal gradients values is more likely to
result in cell death than at lower gradient values (Fig. 6F, axes histograms). However, as evident
by the overlap in cell viability as a function of thermodynamics in Fig. 6F, peak thermal
gradients are not absolutely predictive (Fig. 6F, scatter plot). While cells heated quickly are more
likely to become damaged, rapid heating is not necessarily a death sentence for cells. The
variability in thermally evoked cell damage illustrates the need for tools to study underlying
functional and biomolecular dynamics evoked by IR light on a cell-to-cell basis. Exposure time
and time after exposure becomes a crucial dimension of cellular physiology to explore in the
context of infrared neural stimulation. Endogenous lipid/protein signals (Fig. 6A&B) and a
fluorescence marker for cell damage (Fig. 6C&D) on average appear to increase following strong
levels of INS. Registration of thermal and nonlinear imaging fields within a couple of microns is
key to enable direct correlation of temperature-time dynamics with cell functionality since cells
are not readily visible with thermal imaging.
To practically extend this imaging platform, cell viability markers can easily be
substituted for calcium sensors, voltage probes, FRET constructs, or gene transcription assays,
alongside hyperspectral or bio-orthogonal SRS approaches, to systematically study the functional
impact of INS on other aspects of cellular physiology. SRS often provides relatively nonspecific
biomolecular information relative to fluorescence imaging strategies; combining functional
fluorescence and SRS imaging strategies during INS can expand the current understanding of the
impact of cellular physiological dynamics on SRS signals. Similar approaches can be used to
understand the effects of rapid temperature changes on more specific aspects of cellular
physiology to improve our general understanding of the impact of thermodynamics on cell
physiology.

Beyond INS, similar approaches as demonstrated in Fig. 6 may readily be adapted to
study laser tissue interactions associated with optogenetic or nanoplasmonic neuromodulation,
laser preconditioning of immunological response, and photodynamic therapy of infectious or
cancerous model systems where temperature changes may have a substantial impact. These types
of experiments are readily enabled only by an imaging platform that integrates nonlinear and
thermal microscopy with high enough spatial and temporal resolution capabilities to study the
processes at hand. While neuroscientific questions provide a valuable benchmark in terms of
evaluating imaging speed in the context of biology, this platform is just as applicable to other
areas of biomedical research.
Undoubtedly, an imaging platform that simultaneously integrates thermal and nonlinear
imaging would be practically important in answering many questions in the afore mentioned
disciplines. It becomes technically difficult to accommodate the disparity in optical detection
wavelengths for use in a broad range of cellular, tissue, and live animal preparations. Tradeoffs
between sample flexibility and optical access were major considerations in the imaging system's
three-armed design. Inherent hardware limitations prevent the design a truly simultaneous
multimodal nonlinear and thermal imaging system. Nonetheless, the MANTIS platform offers a
novel approach to creatively answer fundamental and translationally relevant questions about
biology and thermodynamics which are applicable broadly to biomedical research.

Conclusion

We present a novel and flexible multimodal optical imaging system that combines nonlinear and
thermal microscopies for biomedical imaging applications. Our platform can be used to overlay
functional, structural, and biochemical information from a single specimen tracked over time at
high speed and spatial resolution. Such approaches can be applied to study dynamic biophysical
processes in cells and tissues, *in vitro* and *in vivo*. In conjunction with continually expanding
molecular biology tools, this instrument and related imaging methods will aid in studying
physiological and biochemical processes from multiple perspectives in many fields including
neuroscience, cancer biology, metabolic disease, tissue biomechanics, and much more.

References

532
1.　S. Yue, M. N. Slipchenko, J. X. Cheng, Multimodal Nonlinear Optical Microscopy. *Laser Photon Rev* **5**, (2011).
2.　W. R. Zipfel, R. M. Williams, W. W. Webb, Nonlinear magic: multiphoton microscopy in the biosciences. *Nat Biotechnol* **21**, 1369-1377 (2003).
3.　P. J. Campagnola *et al.*, Three-dimensional high-resolution second-harmonic generation imaging of endogenous structural proteins in biological tissues. *Biophysical journal* **82**, 493-508 (2002).
4.　F. Helmchen, W. Denk, Deep tissue two-photon microscopy. *Nature Methods* **2**, 932-940 (2005).
5.　M. Ji *et al.*, Detection of human brain tumor infiltration with quantitative stimulated Raman scattering microscopy. *Science Translational Medicine* **7**, 309ra163-309ra163 (2015).
6.　L. Wei *et al.*, Super-multiplex vibrational imaging. *Nature*, (2017).
7.　M. Lee, W. J. Tipping, A. Serrels, A. N. Hulme, V. G. Brunton. (2017), vol. 10069, pp. 100690T-100690T-100698.
8.　D. Fu, W. Yang, X. S. Xie, Label-free Imaging of Neurotransmitter Acetylcholine at Neuromuscular Junctions with Stimulated Raman Scattering. *J Am Chem Soc*, (2016).
9.　H. J. Lee *et al.*, Label-Free Vibrational Spectroscopic Imaging of Neuronal Membrane Potential. *The journal of physical chemistry letters*, 1932-1936 (2017).
10.　D. Zhang, P. Wang, M. N. Slipchenko, J. X. Cheng, Fast vibrational imaging of single cells and tissues by stimulated Raman scattering microscopy. *Accounts of chemical research* **47**, 2282-2290 (2014).
11.　F. E. Robles, K. C. Zhou, M. C. Fischer, W. S. Warren, Stimulated Raman scattering spectroscopic optical coherence tomography. *Optica* **4**, 243 (2017).
12.　H. Tu *et al.*, Concurrence of extracellular vesicle enrichment and metabolic switch visualized label-free in the tumor microenvironment. *Science Advances* **3**, (2017).
13.　S. A. Boppart *et al.*, Label-free optical imaging technologies for rapid translation and use during intraoperative surgical and tumor margin assessment. *Journal of biomedical optics* **23**, 021104 (2017).
14.　S. F. Owen, M. H. Liu, A. C. Kreitzer, Thermal constraints on in vivo optogenetic manipulations. *Nat. Neurosci.* **22**, 1061-1065 (2019).
15.　A. Picot *et al.*, Temperature Rise under Two-Photon Optogenetic Brain Stimulation. *Cell Rep* **24**, 1243-1253 e1245 (2018).
16.　K. Podgorski, G. Ranganathan, Brain heating induced by near-infrared lasers during multiphoton microscopy. *J Neurophysiol* **116**, 1012-1023 (2016).
17.　M. L. Denton *et al.*, Effect of ambient temperature and intracellular pigmentation on photothermal damage rate kinetics. *Journal of biomedical optics* **24**, 1-15, 15 (2019).
18.　A. J. Walsh, G. P. Tolstykh, S. Martens, B. L. Ibey, H. T. Beier, Action potential block in neurons by infrared light. *Neurophotonics* **3**, 040501 (2016).
19.　B. Choi, a. J. Welch, Analysis of thermal relaxation during laser irradiation of tissue. *Lasers in surgery and medicine* **29**, 351-359 (2001).
20.　D. Moreau, C. Lefort, R. Burke, P. Leveque, R. P. O'Connor, Rhodamine B as an optical thermometer in cells focally exposed to infrared laser light or nanosecond pulsed electric fields. *Biomedical optics express* **6**, 4105-4117 (2015).
21.　J. B. Ford *et al.*, Identifying the Role of Block Length in Neural Heat Block to Reduce Temperatures During Infrared Neural Inhibition. *Lasers in surgery and medicine* **0**, (2019).
22.　T. Suzuki, N. Oishi, H. Fukuyama, Simultaneous infrared thermal imaging and laser speckle imaging of brain temperature and cerebral blood flow in rats. *Journal of biomedical optics* **24**, 031014 (2018).

23. C. W. Freudiger *et al.*, Label-Free Biomedical Imaging with High Sensitivity by Stimulated Raman
Scattering Microscopy. *Science* **322**, 1857-1861 (2008).
24. J. Schindelin *et al.*, Fiji: an open-source platform for biological-image analysis. *Nat Methods* **9**,
676-682 (2012).
25. C. C. Matthews, Fabrice P. . (2006).
26. C. Xu, W. Zipfel, B. S. Jason, R. M. Williams, W. W. Webb, Multiphoton Fluorescence Excitation:
New Spectral Windows for Biological Nonlinear Microscopy. *Proceedings of the National*
Academy of Sciences of the United States of America **93**, 10763-10768 (1996).
27. D. Zhang, M. N. Slipchenko, J. X. Cheng, Highly Sensitive Vibrational Imaging by Femtosecond
Pulse Stimulated Raman Loss. *The journal of physical chemistry letters* **2**, 1248-1253 (2011).
28. T. Y. Choi *et al.*, Thermal evaluation of laser exposures in an in vitro retinal model by
microthermal sensing. *Journal of biomedical optics* **19**, 97003 (2014).
29. M. L. Denton *et al.*, Spatially correlated microthermography maps threshold temperature in
laser-induced damage. *Journal of biomedical optics* **16**, 1-12, 12 (2011).
30. K. Tilbury *et al.*, Second harmonic generation microscopy analysis of extracellular matrix changes
in human idiopathic pulmonary fibrosis. *Journal of biomedical optics* **19**, 086014 (2014).
31. M. Prunotto, P. P. Caimmi, M. Bongiovanni, Cellular pathology of mitral valve prolapse.
Cardiovasc Pathol **19**, e113-117 (2010).
32. J. Wells *et al.*, Biophysical mechanisms of transient optical stimulation of peripheral nerve.
Biophysical journal **93**, 2567-2580 (2007).
33. M. G. Shapiro, K. Homma, S. Villarreal, C. P. Richter, F. Bezanilla, Infrared light excites cells by
changing their electrical capacitance. *Nat Commun* **3**, 736 (2012).
34. E. Stuntz *et al.*, Endogenous Two-Photon Excited Fluorescence Imaging Characterizes Neuron
and Astrocyte Metabolic Responses to Manganese Toxicity. *Sci Rep* **7**, 1041 (2017).
35. M. A. Yaseen *et al.*, Fluorescence lifetime microscopy of NADH distinguishes alterations in
cerebral metabolism in vivo. *Biomedical optics express* **8**, 2368-2385 (2017).
36. D. Legland, I. Arganda-Carreras, P. Andrey, MorphoLibJ: integrated library and plugins for
mathematical morphology with ImageJ. *Bioinformatics* **32**, 3532-3534 (2016).
37. C. Y. Lin *et al.*, Picosecond spectral coherent anti-Stokes Raman scattering imaging with principal
component analysis of meibomian glands. *Journal of biomedical optics* **16**, 021104 (2011).
38. D. Fu, G. Holtom, C. Freudiger, X. Zhang, X. S. Xie, Hyperspectral imaging with stimulated Raman
scattering by chirped femtosecond lasers. *The journal of physical chemistry. B* **117**, 4634-4640
(2013).
39. C. E. Dedic, T. R. Meyer, J. B. Michael, Single-shot ultrafast coherent anti-Stokes Raman
scattering of vibrational/rotational nonequilibrium. *Optica* **4**, 563 (2017).
40. C. H. Camp, Jr. *et al.*, High-Speed Coherent Raman Fingerprint Imaging of Biological Tissues. *Nat*
Photonics **8**, 627-634 (2014).
616

Funding
Funding for this work was provided by the following grants: AFOSR DURIP FA9550-15-1-
0328, AFOSR FA9550-14-1-0303, AFOSR FA 9550-17-1-0374. WRA was supported through
the ASEE NDSEG Fellowship.
Acknowledgements
The authors would also like to thank Dr. Bryan Millis, Dr. Christine O'Brien, and Dr. Andrea
Locke for their feedback and guidance on the manuscript. The authors would also like to thank
Dr. Matthew Bersi, Joshua Bender, and Dr. W David Merryman for providing the porcine mitral
valve samples.

Author Contributions

626
AMJ, EDJ, and JB conceived the idea for the manuscript. WRA, BM, EL, SP, JB and AMJ
designed, and built the imaging system. AMJ and EDJ secured funding support for the published
work. MW performed *in vivo* surgical preparations and assisted in imaging. RG, GT, JLJ, and
JBF provided samples and experimental guidance imaging tissue samples. WRA assisted in all
sample preparations, performed all imaging experiments, image processing, data analysis, and
prepared the manuscript. All authors contributed to editing manuscript.

Data Access

634 Any raw data and processing scripts are readily available from the corresponding authors upon
635 request.

Conflicts of Interest Statement

637 BM, EL, SP, and JB are employees of Thorlabs Imaging Research, LLC. The remaining authors
638 report no additional conflicts of interest.

Supplemental Materials

Table S1: Summary of emission filters and excitation wavelengths used for multimodal imaging. All filters obtained from Semrock (Brattleboro, VT, USA).

Table S2: Summary of measured system resolutions across multiple imaging modalities.

Table S3: Detailed summary of point spread function calculations for nonlinear imaging modalities.

Figure S1: Sample Lateral Resolution Characterization for SHG imaging in a porcine mitral valve sample. Gaussian fitting performed in FIJI.

Figure S2: Signal to Noise calculation with a Vegetable Oil meniscus in FIJI. SNR was calculated to be 34.6.

Figure S3: To-scale representation of the overlap of multimodal imaging fields of view for the 3 imaging arms of MANTIS.

Figure S4: Thermal images of cultured 3T3 Fibroblasts. A) Cells imaged without any aqueous medium. B) Image of aqueous medium front advancing over cells. The absorption of water in the short-wave infrared is high, making cell culture medium difficult to image through with blackbody thermal contrast.

Figure S5: Porcine Mitral Valve imaging with endogenous autofluorescence from elastin (left) and collagen SHG (right). Image width is 520-µm.

Figure S6: Simultaneous SHG (cyan) and SRS (2880-cm-1, lipid dominant resonance, orange) imaging of an unstained murine cervix unfixed frozen section. The stroma (dense in collagen) and epithelium are easily discernable based on relative concentrations on ratio of SHG and SRS signals. Blood vessels lamina propria are also visible with SHG contrast.

Figure S7: Composite multimode images of an *ex vivo* rat sciatic nerve. (A) CARS signal at 2927cm^{-1} (B, green), SRS signal at 2927cm^{-1} (C, red), SHG signal (D, cyan), and multiphoton fluorescence of FluoroMyelin Green (E, Grey). All scale bars are 100um.

Figure S8: Ex vivo rat sciatic nerve samples imaged with multimodal nonlinear imaging from Supp. Fig. S7. Two different modalities can be combined in different ways to visualize tissue structure.

Figure S9: Supplementary Figure 9: Ex vivo rat sciatic nerve samples imaged with SRS (red, myelin at 2927cm-1) and SHG (cyan, collagen). Rescaled SHG images are shown to highlight intrafasicular collagen, which is in lower abundance and overall signal then epineurial collagen.

Figures

Fig. 1: MANTIS System Design. A) The layout of MANTIS. Capable of performing multimodal
nonlinear and thermal imaging over the same sample. Additionally, white light imaging on a third
imaging arm is available for further expansion of the system. B) Schematic of wide-field thermal
microscopy, which infers the temperature of a sample based on the blackbody emission observed between
wavelengths of 3 and 5μm. C) General optical schematic of a multimodal nonlinear imaging system
capable of integrating CARS, SRS, MPF and SHG imaging. DM – dichroic mirror, BP – bandpass filter,
PMT – photomultiplier tube, EOM – electro-optical modulator, InSb CCD – indium antimonide charge
coupled device.

Fig. 2: MANTIS System Performance. A) Resolution of CARS (Green), SRS (Red) with a 1μm polystyrene bead, and Multiphoton Fluorescence (Magenta) with a 500-nm yellow-fluorescent polystyrene bead. Measured resolutions summarized in Table S1. B) Preparation of mixed beads containing 2μm polystyrene beads (blue, 2927cm^{-1}) with 1-10μm PMMA beads (yellow, 3053cm^{-1}). Peaks specific to each bead type C) CARS (Green) and SRS (Red) of signal-to-noise profile measured at a vegetable oil – air meniscus at 2927cm^{-1}. D) Thermal Imaging of a 1951 U.S. Air Force Target. E) SHG imaging of a porcine mitral valve ex vivo.

691 **Fig. 3: *in vitro* imaging capabilities of MANTIS.** NG108 cells observed with CARS (A) at 2830cm^{-1}, a
692 protein-dominant peak, NADH autofluorescence (B), FAD autofluorescence (C), and as a composite
693 overlay of the three channels (D). Images were acquired separately and coregistered together with a rigid
694 transformation.

Fig. 4: Ex vivo rat sciatic nerve samples imaged with high-speed multimodal nonlinear and thermal microscopy. A) Combined image of B-E. Images and corresponding high-speed line scans through the centers of the fields of view with B) CARS at 2880-cm^{-1}. C) SRS at 2880cm^{-1}, D) FluoroMyelin Green, E) SHG. Red lines on graphs indicate the exposure of the tissue to 1875nm light from an optical fiber. Light exposure causes a local temperature increase over 1 second (thermal images, F& G) yielding a corresponding decrease in nonlinear signals.

703 **Fig. 5: Multimodal images from a rat brain in vivo.** A) Composite image of B (SRS, 2930cm⁻¹) and C
704 (CARS, 2850cm⁻¹) highlighting lipid and protein-dominated signals, respectively. D) Composite image of
705 the same field of view in A of E (Green autofluorescence), F (SHG signal), and G (Blue autofluorescence,
706 likely NADH). (H) Thermal images of the brain surface, temperatures shown in °C.

Fig. 6: Fast Multimodal Nonlinear and Thermal Microscopy of NG108 cells during in infrared neural stimulation – Examples of an A) SRS average intensity projection image of cells with indicated centroids and (B) pooled average SRS timeseries of live (magenta) and damaged (green) cells during IR stimulation across all observed cells. (C) Example of a normalized sum intensity projection of cell viability image timeseries used to indicate cell viability via uptake and fluorescence of propidium iodide. (D) Average pooled MPF timeseries of all observed live and damaged cells during INS. Increases in MPF signal after INS above 3% dF/F over 30 seconds were considered damaged/dead. (E) Thermal images of NG108 sample preparation during peak temperatures of INS. Registration of thermal and nonlinear fields of view (E, inset) allows for precise spatial temperature mapping. Cellular details are not observable due to strong absorption by the aqueous imaging medium. Coupled with high speed thermal imaging, (F) high resolution temporal and spatial thermal observations can be directly correlated with functional cellular outcomes such as cell viability. Data shows results of 10 different experiments including 3 different INS intensity conditions, n = 1144 cells.

Supplemental Materials

Table S1: Summary of emission filters and excitation wavelengths used for multimodal imaging. All filters obtained from Semrock (Brattleboro, VT, USA).

Table S2: Summary of measured system resolutions across multiple imaging modalities.

Table S3: Detailed summary of point spread function calculations for nonlinear imaging modalities.

Figure S1: Sample Lateral Resolution Characterization for SHG imaging in a porcine mitral valve sample. Gaussian fitting performed in FIJI.

Figure S2: Signal to Noise calculation with a Vegetable Oil meniscus in FIJI. SNR was calculated to be 34.6.

Figure S3: To-scale representation of the overlap of multimodal imaging fields of view for the 3 imaging arms of MANTIS.

Figure S4: Thermal images of cultured 3T3 Fibroblasts. A) Cells imaged without any aqueous medium. B) Image of aqueous medium front advancing over cells. The absorption of water in the short-wave infrared is high, making cell culture medium difficult to image through with blackbody thermal contrast.

Figure S5: Porcine Mitral Valve imaging with endogenous autofluorescence from elastin (left) and collagen SHG (right). Image width is 520-µm.

Figure S6: Simultaneous SHG (cyan) and SRS (2880-cm-1, lipid dominant resonance, orange) imaging of an unstained murine cervix unfixed frozen section. The stroma (dense in collagen) and epithelium are easily discernable based on relative concentrations on ratio of SHG and SRS signals. Blood vessels lamina propria are also visible with SHG contrast.

Figure S7: Composite multimode images of an *ex vivo* rat sciatic nerve. (A) CARS signal at 2927cm^{-1} (B, green), SRS signal at 2927cm^{-1} (C, red), SHG signal (D, cyan), and multiphoton fluorescence of FluoroMyelin Green (E, Grey). All scale bars are 100um.

Figure S8: Ex vivo rat sciatic nerve samples imaged with multimodal nonlinear imaging from Supp. Fig. S7. Two different modalities can be combined in different ways to visualize tissue structure.

Figure S9: Supplementary Figure 9: Ex vivo rat sciatic nerve samples imaged with SRS (red, myelin at 2927cm-1) and SHG (cyan, collagen). Rescaled SHG images are shown to highlight intrafasicular collagen, which is in lower abundance and overall signal then epineurial collagen.

753   **Supplementary Tables**

Table S1: Summary of emission filters and excitation wavelengths used for multimodal imaging. All filters obtained from Semrock (Brattleboro, VT, USA).

Contrast	Filter (Center/Passband FWHM)	Excitation Wavelengths
CH-band CARS	625nm/90nm	792-803nm & 1040nm
SRS	890nm/310nm	792-803nm & 1040nm
NADH/Blue Fluorescence	460nm/30nm	780nm
FAD/Green Fluorescence	525nm/30nm	934nm
SHG	460nm/30nm	900nm
Propidium Iodide / Orange Fluorescence	625nm/90nm	1040nm

Table S2: Summary of measured system resolutions across multiple imaging modalities.

Modality	Lateral Resolution (μm)	Axial Resolution (μm)
CARS	0.632	3.009
SRS	0.833	3.136
MPF	0.359	1.502
SHG	0.388	X
Thermal	6.9	X

755
756

757 **Table S3: Detailed summary of point spread function calculations for nonlinear imaging modalities**.

						Mean	STD
CARS							
x	0.646	0.615	0.636	x		**0.6323**	0.0158
y	0.773	0.69	0.772	y		**0.7450**	0.0476
z	2.758	3.282	2.986	z		**3.0087**	0.2627
SRS							
x	0.849	0.743	0.906	x		**0.8327**	0.0827
y	0.896	0.793	0.86	y		**0.8497**	0.0523
z	3.36	2.813	3.235	z		**3.1360**	0.2866
MPF							
x	0.379	0.272	0.426	x		**0.3590**	0.0789
y	0.286	0.413	0.377	y		**0.3587**	0.0655
z	1.473	1.432	1.601	z		**1.5020**	0.0882
SHG							
L	0.388	0.383	0.433	L		**0.4013**	0.0275

758
759

 Supplementary Figures:

761
762   Figure S1: Sample Lateral Resolution Characterization for SHG imaging in a porcine mitral valve sample.
763   Gaussian fitting performed in FIJI.

765   Figure S2: Signal to Noise calculation with a Vegetable Oil meniscus in FIJI. SNR was calculated to be
766   34.6.

Figure S3: To-scale representation of the overlap of multimodal imaging fields of view for the 3 imaging arms of MANTIS.

771 Figure S4: Thermal images of cultured 3T3 Fibroblasts. A) Cells imaged without any aqueous medium.
772 B) Image of aqueous medium front advancing over cells. The absorption of water in the short-wave
773 infrared is high, making cell culture medium difficult to image through with blackbody thermal contrast.

775 Figure S5: Porcine Mitral Valve imaging with endogenous autofluorescence from elastin (left) and
776 collagen SHG (right). Image width is 520-µm.

Combined

SHG

SRS ~ 2880cm⁻¹

Figure S6: Simultaneous SHG (cyan) and SRS (2880-cm-1, lipid dominant resonance, orange) imaging of an unstained murine cervix unfixed frozen section. The stroma (dense in collagen) and epithelium are easily discernable based on relative concentrations on ratio of SHG and SRS signals. Blood vessels lamina propria are also visible with SHG contrast.

783 Figure S7: Composite multimode image of an *ex vivo* rat sciatic nerve. (A) CARS signal at 2927cm-1 (B,
784 green), SRS signal at 2927cm-1 (C, red), SHG signal (D, cyan), and multiphoton fluorescence of
785 FluoroMyelin Green (E, Grey). All scale bars are 100um.

A

B

CARS (green) | SRS at 2927-cm⁻¹ (red)

SHG (cyan) | Multiphoton Fluorescence
of FluoroMyelin Green (grey)

C

D

CARS at 2927-cm⁻¹ (green) | Multiphoton
Fluorescence of FluoroMyelin Green (grey)

SRS at 2927-cm⁻¹ (red) | Multiphoton
Fluorescence of FluoroMyelin Green (grey)

E

F

CARS at 2927-cm⁻¹ (greed) | .SHG (cyan)

SRS at 2927-cm⁻¹ (red) | SHG (cyan)

787   Figure S8: Ex vivo rat sciatic nerve samples imaged with multimodal nonlinear imaging from Supp. Fig.
788   7. Two different modalities can be combined in different ways to visualize tissue structure.

SRS at 2927cm⁻¹ vs. SHG

SHG contrast adjusted to
show intrafasicular collagen
structures

790   Figure S9: Ex vivo rat sciatic nerve samples imaged with SRS (red, myelin at 2927cm-1) and SHG (cyan,
791   collagen). Rescaled SHG images are shown to highlight intrafasicular collagen, which is in lower
792   abundance and overall signal then epineurial collagen.