# An approach for normalization and quality control for NanoString RNA expression data

Arjun Bhattacharya[†1], Alina M. Hamilton[†2], Helena Furberg[3], Eugene Pietzak[4], Mark P. Purdue[5], Melissa A. Troester[2,6], Katherine A. Hoadley[*7,8], Michael I. Love[*1,8]

1. Department of Biostatistics, University of North Carolina at Chapel Hill, Chapel Hill, NC, 27514, USA

2. Department of Pathology and Laboratory Medicine, University of North Carolina at Chapel Hill, Chapel Hill, NC, 27514, USA

3. Department of Epidemiology and Biostatistics, Memorial Sloan Kettering Cancer Center, New York, NY, 10065, USA

4. Urology Service, Department of Surgery, Memorial Sloan Kettering Cancer Center, New York, New York, United States

5. Division of Cancer Epidemiology and Genetics, National Cancer Institute, Rockville, MD, 20850, USA

6. Department of Epidemiology, University of North Carolina at Chapel Hill, Chapel Hill, NC, 27514, USA

7. Lineberger Comprehensive Cancer Center, University of North Carolina at Chapel Hill, Chapel Hill, NC, 27514, USAss

8. Department of Genetics, University of North Carolina at Chapel Hill, Chapel Hill, NC, 27514, USA

*To whom correspondence should be addressed.

Email: K.A.H. hoadley@med.unc.edu, M.I. milove@email.unc.edu

[†]The authors wish it to be known that, in their opinion, the first 2 authors should be regarded as joint First Authors.

*The authors wish it to be known that, in their opinion, the last 2 authors should be regard as joint Last Authors.

**ABSTRACT**

The NanoString RNA counting assay for formalin-fixed paraffin embedded samples is unique in its sensitivity, technical reproducibility, and robustness for analysis of clinical and archival samples. While commercial normalization methods are provided by NanoString, they are not optimal for all settings, particularly when samples exhibit strong technical or biological variation or where housekeeping genes have variable performance across the cohort. Here, we develop and evaluate a more comprehensive normalization procedure for NanoString data with steps for quality control, selection of housekeeping targets, normalization, and iterative data visualization and biological validation. The approach was evaluated using a large cohort ($N = 1,649$) from the Carolina Breast Cancer Study, two cohorts of moderate sample size ($N = 359$ and $130$), and a small published dataset ($N = 12$). The iterative process developed here eliminates technical variation (e.g. from different study phases or sites) more reliably than the three other methods, including NanoString's commercial package, without diminishing biological variation, especially in long-term longitudinal multi-phase or multi-site cohorts. We also find that probe sets validated for nCounter, such as the PAM50 gene signature, are impervious to batch issues. This work emphasizes that systematic quality control, normalization, and visualization of NanoString nCounter data is an imperative component of study design that influences results in downstream analyses.

**Keywords**: NanoString nCounter expression; gene expression normalization; quality control; data visualization

**INTRODUCTION**

The NanoString nCounter platform offers a targeted strategy for gene expression quantification using a panel of up to 800 genes without requiring cDNA synthesis or amplification steps [1]. The technology offers advantages in sensitivity, technical reproducibility, and strong robustness for profiling formalin-fixed, paraffin-embedded (FFPE) samples [2]. Given these advantages, nCounter is increasingly used for longitudinal studies involving FFPE samples carried out over several years [3] and diagnostic assays in clinical settings [4,5].

Proper normalization and quality control of gene expression is necessary prior to statistical analysis to

reduce unwanted variation that may be associated with technical batches or RNA degradation from

sample fixation [6,7]. While some sources of variation can be enumerated a priori (e.g. different research

centers, batches over time, or RNA preservation methods), not all can be captured. In all cases, it is

advisable to define a quality control and normalization pipeline to detect and account for technical

variation in downstream statistical modeling. All normalization methods deal with a trade-off between bias

that needs correction and bias or variance that may be introduced in normalization [8].

Many approaches have been developed to normalize nCounter data. NanoString provides two forms of

normalization in its commonly-used nSolver Analysis Software [9]: (A) a graphical user interface with

optional background correction and positive-control and housekeeping gene normalization and (B) the

Advanced Analysis tool, which draws on the NormqPCR R package [10,11] to select co-expressed

housekeeping genes prior to normalization. The NanoStringNorm package implements the nSolver

algorithms in R [12]. The R packages NanoStringDiff and RCRnorm use hierarchical modeling methods

that incorporate information from the positive, negative, and housekeeping controls for normalization

[13,14]. The NACHO R package proposes a simple quality control and visualization pipeline that

precedes normalization using either NanoStringNorm or NanostringDiff [15], though, without post-

normalization visualization to assess normalization quality. When technical replicates are available,  a

method from Molania et al, Remove Unwanted Variation-III (RUV-III), can be used along with an iterative

normalization process where several parameters (i.e. number of housekeeping genes, number of

detected outliers, number of dimensions of technical noise) are tuned with relevant visual and biological

checks [7]. RUV-III normalization frequently outperformed nSolver normalization by more efficiently

removing technical sources of variation while preserving biological variation [7]. Since many cohorts do

not have technical replicates, we extend Molania et al's iterative framework using RUVSeq [6–8], a

precursor of RUV-III.

Here, we provide a framework for the quality control and normalization of mRNA expression count data

from the NanoString nCounter platform, using a large dataset ($N = 1{,}649$) of breast tumor expression

from the Carolina Breast Cancer Study (CBCS) and three other cohorts of differing sample size ($N =$

$12, 130,$ and $359$). We illustrate some of the pitfalls in the nSolver method of background correction and

positive control normalization, provide an alternative approach that uses RUVSeq [6,8], and benchmark

our framework against other normalization methods [9,13,14]. We find that, especially in longitudinal,

multi-phase or multi-site cohorts, RUVSeq outperforms nSolver in removing differences across technical

sources of variation. Lastly, we provide quality checks for normalization and outline the impact of proper

normalization on inference for biological associations and expression-based disease subtyping.

**MATERIAL AND METHODS**

***Data collection***

We used four cohorts with nCounter gene expression data to evaluate differences between normalization

procedures. Cohort details and the normalization parameters for each cohort are given below and

summarized in **Supplemental Table 1**.

*CBCS gene expression data*

The Carolina Breast Cancer Study (CBCS) is a multi-phase cohort of women with breast cancer in North

Carolina. Samples were collected during three study phases: Phase 1 (1993-1996), Phase 2 (1996-

2001), and Phase 3 (2008-2013). Paraffin-embedded tumor blocks were reviewed and assayed for gene

expression using the NanoString nCounter system as discussed previously [3,16,17]. Study phase gives

the relative age of the tumor block. In total, 1,649 samples from patients with invasive breast cancer from

CBCS, across all three study phases, were analyzed on a custom panel of 417 genes. All assays were

performed in the Translational Genomics Laboratory (TGL) at the University of North Carolina at Chapel

Hill (UNC). After quality control and normalization, 1,264 samples remained in the nSolver-normalized

data, and 1,219 samples remained in the RUVSeq-normalized data. This dataset was used to benchmark

against NanoStringDiff [13] and RCRnorm [14], using the same 1,264 samples in the nSolver-normalized

set.

*Bladder tumor gene expression data*

FFPE Biospecimens from 42 samples of NMIBC from UNC (Chapel Hill, NC) and 88 samples from a

study conducted by the Memorial Sloan Kettering Cancer Center (New York, NY) with non-muscle

invasive bladder cancer (NMIBC) were analyzed. RNA was isolated using the RNeasy FFPE Kit (Qiagen)

at UNC and NanoString assays were performed at the TGL at UNC using a custom codeset consisting of

440 endogenous and 6 housekeeping genes. After quality control and normalization, 86 samples

remained in both the nSolver-normalized and RUVSeq-normalized datasets.

*Kidney tumor gene expression data*

This study includes 359 samples from patients with clear cell renal cell carcinoma (CCRCC) with fresh-

frozen tissue collected as part of a large case-control study of kidney cancer conducted in central and

eastern Europe [18] . Slides for each case were reviewed by a pathologist to assess tumor stage and

grade [19]. Manual microdissection was performed to remove non-tumor tissue. Frozen sections were

placed directly in Trizol reagent (Invitrogen, Carlsbad, CA), homogenized for 2 minutes on ice, and RNA

was isolated using the manufacturer's protocol. NanoString assays were performed at UNC TGL using a

custom codeset consisting of 62 endogenous and 6 housekeeping genes commonly studied in kidney

cancer. After quality control and normalization, 331 samples remained in both the nSolver- and RUVSeq-

normalized data.

*Sabry et al gene expression data*

We downloaded raw RCC files from Sabry et al [20] from the NCBI Gene Expression Omnibus (GEO)

with accession number GSE130286 and imported them using functions in NanoStringQCPro [21]. This

dataset comprised of 12 samples, all of which remained after normalization with both procedures. The

dataset measured 706 endogenous genes with 40 housekeeping genes from the NanoString nCounter

Human Myeloid Innate Immunity Panel [20].

**Quality control and normalization**

The full quality control and normalization process using nSolver and RUVSeq is summarized in **Figure 1**,

starting with familiarization of the raw data (**Figure 1.1),** technical quality control (**Figure 1.2**)**,** pre-

normalization assessment of housekeeping genes (**Figure 1.3**) and data visualization to detect

problematic samples and assess whether flagged samples should be removed (**Figure 1.4**).

Normalization is performed with either nSolver or RUVSeq (**Figure 1.5**), and the processed expression

data is assessed for validity through relevant visualization and biological checks (**Figure 1.6**). If validation

is unsatisfactory and technical variation is still present, this process is iterated.

*Technical quality controls flags*

nSolver provides quality control (QC) flags to assess the quality of the data for imaging, binding density,

linearity of the positive controls, and limit of detection. The definition and implementation of this QC is

summarized in nSolver [9] and NanoStringNorm [12] documentation. We mark any sample that is flagged

in at least one of these four QC assessments as technical quality control. We use these QC flags in both

nSolver normalization and RUVSeq normalization.

*Below limit of detection quality control*

We use high proportions of both endogenous and housekeeping genes below the limit of detection (LOD)

as a QC flag to assess reduced assay or sample quality. The per-sample LOD is defined as the mean of

the counts of negative control probes for a given sample. We assessed the percent of counts below the

LOD in the housekeeping genes per sample to flag both poor quality samples and housekeeping genes

with problems in their measurement. We used samples with all housekeeping genes above the LOD as a

reference group to determine the regular distribution of genes below the LOD. Samples were flagged if

(1) they had more than one housekeeping gene below the LOD and (2) the percent of endogenous genes

below the LOD was greater than the top quartile of the distribution of percent below LOD in the reference

group.

*Housekeeping gene assessment*

Housekeeping genes serve two purposes: 1) for QC purposes to remove samples with overall poor

quality and 2) for assessing the amount of technical variation present in the normalization procedure.

NanoString documentation suggests that ideal housekeeping genes are highly expressed, have similar

coefficients of variation, and have expression values that correlate well with other housekeeping genes

across all samples [9,12]. Because of these definitions, these targets will ideally vary only due to the level

of technical variation present. RUVSeq relies on housekeeping genes, i.e. genes not influenced by the

condition of interest (e.g. cancer subtype), with no assumptions on co-expression of all housekeeping

genes. To assess the potential for housekeeping correction to introduce bias, housekeeping genes were

assessed for differential expression across a primary biological covariate of interest (estrogen receptor

status in CBCS, tumor stage in the kidney and bladder cancer data, and treatment groups in Sabry et al

[20]) using negative binomial regression on the raw counts from the MASS package [22].

***nSolver normalization***

*Background correction*

NanoString guidelines suggest background correction [9,12] by either subtraction or thresholding for an

estimated background noise level for experiments in which low expressing targets are common, or when

the presence of a transcript has an important research implication [7,12]. Data from all four cohorts

considered do not necessarily fall under this criterion, and accordingly, we did not background correct by

either method. To demonstrate the effect of background correction, we tested nSolver-normalized gene

expression with and without background thresholding in CBCS using relative log expression (RLE) plots.

*Positive control and housekeeping gene-based normalization*

The arithmetic mean of the geometric means of the positive controls for each lane was computed and

then divided by the geometric mean of each lane to generate a lane-specific positive control normalization

factor [9,12]. The counts for every gene were multiplied by their lane-specific normalization factor. To

account for any noise introduced into the nCounter assay by positive normalization, the housekeeping

genes were used similarly as the positive control genes to compute housekeeping normalization factors

to scale the expression values [9,12]. NanoString flagged samples with large housekeeping gene scaling

factors (we call this a housekeeping QC flag) and large positive control scaling factors (positive QC flag)

but note that samples with these flags simply indicate that a sample is divergent from other samples in

the dataset and do not necessarily require removal. Pre-normalization visualization (**Figure 1.4)** is

important for confirming the inclusion or removal of these samples.

***RUVSeq normalization pipeline***

*Normalization*

The RUVSeq-based normalization process (**Figure 1.5**)**,** an alternative approach to nSolver

normalization, proceeds following quality control and housekeeping assessment. Distributional

differences were scaled between lanes using upper-quartile normalization [23]. Unwanted technical

factors were estimated in the resulting gene expression data with the RUVg function from RUVSeq [8].

Unwanted variation was estimated using the final set of endogenous housekeeping genes on the

NanoString gene expression panel [24,25]. In general, the number of dimensions of unwanted variation to

remove was chosen by iteratively normalizing the data for a given number of dimensions and checking for

the removal of known technical factors already identified in the raw expression data (e.g. study phase),

and presence of key biological variation (e.g. bimodality of ESR1 expression in the CBCS breast cancer

data where estrogen receptor status is a known predominant feature). Further details about choosing this

dimension are given by Gagnon-Bartsch et al and Risso et al [6,8]. DESeq2 was used to compute a

variance stabilizing transformation of the original count data [25], and estimated unwanted variation was

removed using the removeBatchEffects function from limma [26]. Ultimately, we removed 1, 1, 3, and 1

dimensions of unwanted variation from CBCS, kidney cancer, bladder cancer, and the Sabry et al

datasets, respectively. RLE plots, principal component analysis and heatmaps were used to detect any

potential outliers before and after normalization.

***Alternative normalization methods for benchmarking***

Using CBCS data, we compared the normalized datasets from nSolver, RUVSeq, NanoStringDiff [13],

and RCRnorm [14] with the raw data through visualization methods outlined above (**Figure 1.1 to 1.4**,

RLE plots and scatter plots of principal components over important technical and biological sources of

variation). Details about these methods are provided in **Supplemental Table 2**.

***Downstream analyses***

We used several data visualization or benchmarking methods for each cohort.

*Silhouette width analysis in CBCS*

Silhouette width, a measure used to assess how similar a sample is to its own group (i.e. study phase) as

compared to other groups, was used to determine the impact of the two normalization procedures on

technical and biological variation [27]. Many samples with large silhouettes can be interpreted as

indicating that the different study phases are distinct and that a batch effect is still present in the data.

*eQTL analysis in CBCS*

We assessed the additive relationship between the gene expression values and germline genotypes with

linear regression analysis using MatrixEQTL [28], applying the same linear model as detailed in previous

work [29]. Briefly, for each gene and SNP in our data, we constructed a simple linear regression, where

the dependent variable is the scaled expression of the gene with zero mean and unit variance, the

predictor of interest is the dosage of the alternative allele of the SNP, and the adjusting covariates are the

top five principal components of the genotype matrix. We considered both cis- (SNP is less than 0.5 Mb

from the gene) and trans-eQTLs in our analysis. We adjusted for multiple testing via the Benjamini-

Hochberg procedure [30].

*PAM50 subtyping in CBCS*

We classified each subject into PAM50 subtypes using the procedure summarized by Parker et al [31,32].

Briefly, for each sample, we computed the Euclidean distance of the log-scale expression values for the

50 PAM50 genes to the PAM50 centroids for each of the molecular subtypes. Each sample was classified

to the subtype with the minimal distance [31]. The PAM50 genes were clustered hierarchically for both

samples and genes and visualized in heatmaps. Subtype concordance was assessed between

normalization methods excluding normal-like cases.

*RNA-seq normalization and distance correlation analysis in CBCS*

We obtained a separate set of samples (not included in the analysis described above) from CBCS with

both RNA-seq and nCounter expression (on a different codeset of 166 genes). We followed a standard

RNA-seq normalization process with DESeq2 [25], using the median of ratios method to estimate scaling

factors [24]. We calculated the distance correlation and conducted a multivariate permutation test of

independence between the RNA-seq data set (subset to the overlapping genes on the NanoString

codeset) with each of the nSolver-normalized and RUVSeq-normalized nCounter data using the energy

package [33]. The distance correlation and associated permutation test allow for detection of non-

independence across multivariate datasets of different distribution.

*Differential expression analysis with Sabry et al. dataset [20]*

We conducted differential expression analysis to compare both normalization methods in the Sabry et al.

dataset [20] using DESeq2 [25], and adjusting for multiple testing with the Benjamini-Hochberg [30]

procedure. We compared differential expression across IL-2–primed NK cells vs. NK cells alone and

CTV-1-primed NK cells for 6 hours vs. NK cells alone.

**RESULTS**

We evaluated the ability of normalization methods to remove technical variation while retaining

biologically meaningful variation across four cohorts of differing sample size and varying sources of

technical bias (**Supplemental Table 1**). Known sources of technical variation included age of sample

(study phase) and different study sites. The cohorts varied in preservation methods; two cohorts used

fresh-frozen specimens, while two used archival FFPE specimens.  The number of genes measured for

both endogenous genes and housekeeping genes also varied by study. In addition, some studies used

validated and optimized code sets for specific gene signatures versus a more general code set.

In cohorts with large technical biases, RUVSeq provided superior normalization with more robust removal

of technical variation and provided stronger biological associations compared to other normalization

methods. In two of the datasets, we found that downstream analyses performed on data normalized with

nSolver and RUVSeq detected substantially different biological associations. However, when few strong

technical biases were present or if a validated and optimized code set (e.g. PAM50 genes) was used,

nSolver and RUVSeq performed comparably.

***Case study: Carolina Breast Cancer Study***

*Evaluation of background correction*

Background thresholding led to increased per-sample variance while per-sample medians remained

relatively similar (**Supplemental Figure 1A**). The distributions of per-sample median expression values

were more right-skewed (greater mean than median) when using background thresholding prior to

normalization compared to not using background thresholding (**Supplemental Figure 1B**). Based on this

analysis, we did not perform background correction prior to normalization for all cohorts analyzed.

*Quality assessment of expression levels using LOD of housekeeping genes*

We used the housekeeping genes to assess if the lack of expression of endogenous genes was due to

biology or due to technical failures. We compared the level of missing endogenous genes in samples with

all housekeeping genes present to those with increasing number of housekeeping genes below LOD.

There was a strong positive correlation for increasing proportions of genes below the LOD in both the

endogenous and housekeeping genes (**Figure 2A** and **Supplemental Figure 2)**. Samples with higher

numbers of genes below the LOD were from earlier phases of CBCS (i.e. Phase 1 from 1993-1996 and

Phase 2 from 1996-2001), and thus associated with sample age (**Figure 2A** and **Supplemental Figure**

**3**). Samples with a higher proportion of endogenous genes below the LOD had increased numbers of QC

flags as well (**Supplemental Figure 2**).

*Evaluation of normalization methods*

We benchmarked RUVSeq and nSolver with two other normalization methods, NanoStringDiff [13] and

RCRnorm [14].  We observed differences across the four normalization strategies (described in

**Supplemental Table 2**), namely greater remaining technical variation using nSolver and NanoStringDiff

than RCRnorm and RUVSeq **(Figure 2B-D)**. A large portion of the variation in the raw expression could

be attributed to study phase **(Supplemental Figure 4A).** While all methods reduced study phase

associated variation compared to the raw data, there were considerable differences in the deviations from

the median log-expressions in the nSolver- and NanoStringDiff-normalized expression that are not

present in the RUVSeq- and RCRnorm-normalized data **(Figure 2B)**. The nSolver and NanoStringDiff

methods retained technical variation, either not fully corrected or re-introduced during the nSolver

normalization process.

We examined the ability of each normalization method to retain biological variation. Estrogen Receptor

(ER) status is one of the most important clinical and biological features in breast cancer and is used for

determining course of treatment [34,35]. ER status drives many of the molecular classifications [36–38]

and even drives separate classification of breast tumors in TCGA's pan-cancer analysis of 10,000 tumors

[39].   In the raw expression, variation due to ER status was captured in PC2 rather than PC1 (study age);

however, after RUVSeq-normalization, ER status was reflected predominantly in PC1 (**Figure 2C**).  In the

nSolver-, NanoStringDiff-, and RCRnorm-normalized data, ER status was shared between PC1 and PC2,

suggesting that unresolved technical variation was still present. RUVSeq demonstrated

effective removal of technical variation and boosting of the true biological signal. The PAM50 molecular

subtypes [31], which are also linked with ER status, were also clearly separated by PC1 for RUVSeq-

normalized data, but this was not the case for nSolver-, NanoStringDiff-, or RCRnorm-normalization

(**Supplemental Figure 4B**).  These results suggest that RUVSeq-normalization best balances the

removal of technical variation with the retention of important axes of biological variation, with RCRnorm

showing better performance than nSolver and NanoStringDiff, but not superior to RUVSeq. A significant

disadvantage of RCRnorm is its computational cost: RCRnorm was unable to run on the CBCS dataset

($N = 1278$ after QC) on a 64-bit operating system with 8 GB of installed RAM, requiring RCRnorm-

normalization to be performed on a high-performance cluster. We summarize the maximum memory used

by method in CBCS in **Supplemental Table 2**.

We used silhouette width to assess extent of unwanted technical variation from study phase remaining by

the normalization methods. Larger positive silhouette values indicate within-group similarity (i.e. samples

clustering by study phase). Per-sample silhouettes across the alternatively normalized datasets showed

that RUVSeq best addressed the largest source of technical variation identified in the raw data (**Figure**

**2D**; **Supplemental Figure 5A**) while also not removing a significant portion of biological variation

(**Supplemental Figure 5B**). NanoStringDiff also demonstrated less similarity of samples across study

phase similar to RUVSeq but removed biologically relevant similarity of samples grouped by ER status.

Due to the performance of NanoStringDiff and computational limitations of RCRnorm, for subsequent

analyses and datasets, we only illustrate differences between nSolver- and RUVSeq-normalized data.

*Genomic analyses and expression profiles across normalization methods*

We evaluated the impact of normalization choice on downstream analyses including eQTLs, PAM50

molecular subtyping, known expression patterns, and similarity to RNA-seq data. In a full cis-trans eQTL

analysis accounting for race and genetic-based ancestry, we found considerably more eQTLs using

nSolver as opposed to RUVSeq, thresholding at nominal $P < 10^{-3}$ (2,050 vs. 1,143). We identified strong

cis-eQTL signals in both normalized datasets; however, stronger FDR values were identified with

RUVSeq (**Figure 3A,** densely populated around the 45-degree line**)**. We observed considerably more

trans-eQTLs using nSolver, including a higher proportion of trans-eQTLs across various FDR-adjusted

significance levels (**Figure 3B; Supplemental Figure 7**). We suspected that spurious trans-eQTLs may

have resulted from residual technical variation in expression data that was confounded with study phase,

subsequently being identified as a QTL due to ancestry differences across study phase. In cross-

chromosomal trans-eQTL analysis, distributions of absolute differences in minor allele frequency (MAF)

for trans-eSNPs across women of African and European ancestry were wide for both methods

(**Supplemental Figure 7)**. However, we observed substantially more trans-eSNPs with moderate

absolute MAF differences across study phase with nSolver, compared to RUVSeq. This provides some

evidence for the presence of residual confounding technical variation in the nSolver-normalized

expression data leading to spurious trans-eQTL results (with a directed acyclic graph for this hypothesis

in **Supplemental Figure 8**), though we cannot confirm this with eQTL analysis alone.

We compared each normalization method for the ability to classify breast cancer samples into PAM50

intrinsic molecular subtype using the classification scheme outlined by Parker et al [31]. Our PAM50

subtyping calls were robust across normalization methods with 91% agreement and a Kappa of 0.87

(95% CI (0.85, 0.90)) (**Supplemental Table 3**). Among discordant calls, approximately half had low

confidence values from the subtyping algorithm, and half had differences in correlations to centroids less

than 0.1 between the discordant calls (data not shown).  Most of these discordant calls were among

HER2-enriched, luminal B and luminal A subtypes, which are molecularly similar [40].

We observed noticeable differences between the RUVSeq- and nSolver-normalized gene expression

when visualized after hierarchical clustering via heatmaps, similar to the principal component analysis.

Using this method, we identified 14 additional samples with strong technical errors in the nSolver-

normalized data not previously marked by QC flags (**Supplemental Figure 9**)**,** emphasizing the need for

post-normalization data visualization. In early breast cancer clustering papers, the first major division was

by ER status separating basal-like and HER2-enriched molecular subtypes (predominantly ER-negative)

from luminal A and B molecular subtypes (predominantly ER-positive) [31]. This pattern was observed in

RUVSeq-data but only partially preserved with nSolver normalization (**Supplemental Figure 9**)**.** Rather,

nSolver data clustering was driven by a combination of ER status and study phase. Study phase

dominated two of the groups and were formed by Phase 1 and Phase 3 samples, respectively—samples

with a 10+ year difference in age.

Lastly, we compared normalization choices for NanoString data to RNA-seq data performed on the same

samples. CBCS collected RNA-seq measurements for 70 samples that have data on a different nCounter

codeset (162 genes instead of 417) and RNA-seq normalized using standard procedures. A permutation-

based test of independence using the distance correlation [33,41] revealed that the distance correlation

between the RNA-seq and nSolver data was small and near 0 (distance correlation = 0.051, $P = 0.24$)

while the distance correlation between the RNA-seq and RUVSeq- data was larger (distance correlation =

0.36, $P = 0.02$). The permutation-based test rejected the null hypothesis of independence (distance

correlation of zero for unrelated datasets) between RUVSeq-normalized nCounter data and RNA-seq

data but fails to reject the null hypothesis for nSolver-normalization nCounter and RNA-seq data. We

conclude that RUVSeq produced normalized data with closer relation to the RNA-seq, in terms of

distance correlation and test of independence, compared to nSolver.

***Case study: differential expression analysis in natural killer cells***

We looked at the impact of the two normalization methods in a small cohort ($N = 12$) on DE analysis

across natural killer (NK) cells primed for tumor-specific cells and cytokines from Sabry et al [20]. RLE

plots before and after normalization showed minor differences between the two normalization methods

(**Supplemental Figure 9**).

Using DESeq2 [25], we identified genes differentially expressed in NK cells primed by CTV-1 or IL-2

cytokines compared to unprimed NK cells at FDR-adjusted $P < 0.05$. The two normalization methods led

to a different number of differentially expressed genes with a limited overlap of significant genes by both

methods (**Figure 4A).** The raw $P$-value histograms from differential expression analysis using nSolver-

normalized expression exhibited a slope toward 0 for $P$-values under 0.3, which can indicate issues with

unaccounted-for correlations among samples [42], such as residual technical variation. The distributions

of $P$-values using the RUVSeq-normalized data were closer to uniform throughout the range [0,1] for most

genes (**Figure 4B**). While the $\log_2$-fold changes were correlated between the two normalization

procedures, the genes found to be differentially expressed only with nSolver-normalized data tended to

have large standard errors with RUVSeq-normalized data and therefore not statistically significant using

RUVSeq (**Figure 4C**). These differences in DE results emphasize the importance of properly validating

normalization prior to downstream genomic analyses.

***Case study: bladder cancer gene expression***

RUVSeq reduced technical variation (study site) while maintaining the biological variation (tumor grade).

RUVSeq data showed the most homogeneity in per-sample median deviation of log-expressions

compared to raw and nSolver data (**Figure 5A**). The first principal component of nSolver data had

significant differences by study sites, which was not present in RUVSeq data (**Figure 5B**). In addition,

there was a stronger biological association with tumor grade in the first principal component of expression

using RUVSeq data (**Figure 5C**).

*Case study: kidney cancer gene expression*

We only found subtle differences in the deviations from the median expression between the normalization

procedures for the kidney cancer dataset (**Figure 6A**). This cohort did not have the same known technical

variables observed in the other cohorts such as study site or sample age, and the RNA came from fresh-

frozen material (**Supplemental Table 1**). We evaluated normalization methods on a source of technical

variation, DV300, the proportion of RNA fragments detected at greater than 300 base pairs as a source of

technical variation, and tumor stage as a biological variable of interest. The first two principal components

colored by level of DV300 (**Figure 6B**) and tumor stage (**Figure 6C**) showed little difference across the

two normalization methods. When there were limited sources of technical variation and a robust, high

quality dataset, we found both normalization methods performed equally well.

**DISCUSSION**

Proper normalization is imperative in performing correct statistical inference from complex gene

expression data. Here, we outline a sequential framework for NanoString nCounter RNA expression data

that provides both quality control checks, considerations for choosing housekeeping genes, and iterative

normalization with biological validation using both NanoString's nSolver software [9,12] and RUVSeq

[6,8]. We show that RUVSeq provided a superior normalization to nSolver on three out of four datasets by

more efficiently removing sources of technical variation, while retaining robust biological associations.

We also benchmark RUVSeq-normalization with two other normalization methods implemented in R and

show that RUVSeq outperformed all methods in reducing technical variation.

We observed that normalization methods were sensitive to the quality and the set of housekeeping

genes. Several genes thought to behave exclusively in a "housekeeping" fashion in fact associate with

biological variables under certain conditions [43] or across different tissue types [44]. A careful validation

of housekeeping gene stability on a case-by-case basis and separately for new studies, considering both

technical and biological sources of variation in each dataset, is therefore imperative for an optimized

normalization procedure.

We developed a quality metric to assess sample quality: samples with high proportions of genes detected

below the LOD in both endogenous genes and housekeepers were indicative of either low-quality

samples or reduced assay efficiency. Sample age was correlated with higher proportions of genes below

the LOD in both endogenous and housekeeping genes, which was likely due to RNA degradation over

time. We stress that missing counts in endogenous genes alone does not suggest poor sample quality in

the absence of additional QC flags but could represent genes not expressed and therefore not detected

under certain biological conditions or cell types. An example includes using an immuno-oncology gene

panel in a tumor sample with little to no immune cell infiltration. Conversely, many samples with counts

below the LOD in both endogenous genes and housekeepers had additional quality control flags including

those derived from nSolver's assessment of data quality. We excluded these samples for analysis in both

the nSolver- and RUVSeq-based procedures.

nSolver-normalized data was prone to residual unwanted technical variation when there were known

technical biases, such as in CBCS and the bladder example. We checked for known biological

associations that are intrinsic to the sample, as in eQTL analysis, to judge the performance of the

normalization process [45,46]. A full cis-trans eQTL analysis using nSolver- and RUVSeq-normalized data

showed a strong cis-eQTL signal in data from both normalization methods. We found significantly more

trans-eQTLs with the nSolver-normalized data (**Figure 3**). However, many of the trans-eSNPs for the loci

found with nSolver-normalized data tended to have moderate MAF differences across phase, leading us

to suspect they were spurious associations driven by residual technical variation in gene expression

(**Supplemental Figure 8**). Such spurious associations from population stratification have been described

in many previous studies of eQTL analysis [47–50].

The choice of normalization procedure is less of a concern in cohorts with minimal sources of technical

variation or in nCounter targeted gene panels that have been optimized for robust measurement across

preservation methods. In the CBCS breast cancer cohort, we identified significant differences in gene

expression between normalization methods across the entire gene set (417 total genes). However,

PAM50 subtyping was robust across the two normalization procedures. The genes in the PAM50

classifier were selected due to their consistent measurement in both FFPE and fresh frozen breast

tissues [31], suggesting that robustly measured genes may be less affected by different normalization

procedures. Furthermore, we see minimal differences in residual technical variation in the kidney cancer

dataset and the Sabry et al dataset, both of which were measured on either robustly validated genes or

nCounter panels. The kidney cancer example had newer, fresh-frozen specimens that were profiled using

a small and well-validated set of genes important in that cancer type.  This dataset gives an opportunity to

stress the importance of the general principles of normalization: as Gagnon-Bartsch et al and Molania et

al recommend [6,7], normalization should be a part of scientific process and should be approached

iteratively with visual inspection and biological validation to tune the process. One normalization

procedure is not necessarily applicable to all datasets and must be re-evaluated on each dataset.

In conclusion, we outline a systematic and iterative framework for the normalization of NanoString

nCounter expression data. Even without background correction, a technique which has been shown to

impair normalization of microarray expression data [51,52], we believe that relying solely on positive

control and housekeeping gene-based normalization may result in residual technical variation after

normalization. Here, we show the merits of a comprehensive procedure that includes sample quality

control checks including the addition of new checks, assessments of housekeeping genes, normalization

with RUVSeq [6,8] and data analysis with popular count-based R/Bioconductor packages, as well as

iterative data visualization and biological validation to assess normalization. Researchers must pay close

attention to the normalization process and systematically assess pipelines that best suit each dataset.

**KEY POINTS**

• The NanoString nCounter RNA counting assay, an attractive option in archived samples, has

sub-optimal quality control and normalization pipelines.

- We provide an iterative framework for nCounter data with steps for quality control, normalization, and visualization/validation using RUVSeq.
- Using four real datasets, we show that our framework eliminates technical variation more reliably than other methods, including NanoString's provided software nSolver, without diminishing biological variation.
- We stress that quality control and normalization must be emphasized in study design and evaluated using proper visualization and other checks, or else results in downstream analyses may be biased.

**AVAILABILITY**

Relevant R code for these analyses are freely available at https://github.com/bhattacharya-a-bt/CBCS_normalization [53]. Summary statistics for eQTL analysis are available at https://github.com/bhattacharya-a-bt/CBCS_TWAS_Paper [54], as a part of Bhattacharya et al [29].

CBCS genotype datasets analyzed in this study are not publicly available as many CBCS patients are still being followed and accordingly CBCS data is considered sensitive; the data is available from M.A.T upon reasonable request. Raw and normalized expression data from CBCS will be available on GEO upon publication. For replication or review prior to publication, this data can be accessed from GEO through a reviewer token or requested from M.A.T. Data from the bladder and kidney cancer datasets may be provided by the authors upon reasonable request.

**ACCESSION NUMBERS**

Raw RCC files for nCounter expression from Sabry et al [20] are available NCBI Gene Expression Omnibus (GEO) with the accession numbers GSE130286. Raw and normalized expression data from CBCS will be available on GEO upon publication. For replication prior to publication, this data can be requested from the authors.

**SUPPLEMENTARY DATA**

Document S1: Supplemental Tables and Figures

**ACKNOWLEDGEMENT**

We thank the Carolina Breast Cancer Study participants and volunteers. We thank Halei Benefield, Xiaohua Gao, Erin Kirk, Linnea Olsson, and Jessica Tse for their invaluable support during the research process.

**FUNDING**

Susan G. Komen® provided financial support for CBCS study infrastructure.  Funding was provided by the National Institutes of Health, National Cancer Institute P01-CA151135, P50-CA05822, and U01-CA179715 to M.A.T.  M.I.L. is supported by P01-CA142538 and P30-ES010126. KAH is supported by a Komen Career Catalyst Grant (CCR16376756). AMH is supported by 1T32GM12274. The Translational Genomics Laboratory is supported in part by grants from the National Cancer Institute (3P30CA016086) and the University of North Carolina at Chapel Hill University Cancer Research Fund. The kidney cancer study and gene expression analysis were supported by the Intramural Research Program of the National Institutes of Health and the National Cancer Institute.

The funders had no role in the design of the study, the collection, analysis, or interpretation of the data, the writing of the manuscript, or the decision to submit the manuscript for publication.

**CONFLICT OF INTEREST**

The authors have no conflicts of interest to disclose.

**REFERENCES**

1. Geiss GK, Bumgarner RE, Birditt B, et al. Direct multiplexed measurement of gene expression with color-coded probe pairs. Nat. Biotechnol. 2008; 26:317–325

2. Veldman-Jones MH, Brant R, Rooney C, et al. Evaluating Robustness and Sensitivity of the NanoString Technologies nCounter Platform to Enable Multiplexed Gene Expression Analysis of Clinical Samples. Cancer Res. 2015; 75:2587–2593

3. Troester MA, Sun X, Allott EH, et al. Racial Differences in PAM50 Subtypes in the Carolina Breast Cancer Study. JNCI J. Natl. Cancer Inst. 2018; 110:176–182

4. Wallden B, Storhoff J, Nielsen T, et al. Development and verification of the PAM50-based Prosigna breast cancer gene signature assay. BMC Med. Genomics 2015; 8:54

5. Vieira AF, Schmitt F. An Update on Breast Cancer Multigene Prognostic Tests-Emergent Clinical Biomarkers. Front. Med. 2018; 5:248

6. Gagnon-Bartsch JA, Speed TP. Using control genes to correct for unwanted variation in microarray data. Biostatistics 2012; 13:539–552

7. Molania R, Gagnon-Bartsch JA, Dobrovic A, et al. A new normalization for Nanostring nCounter gene expression data. Nucleic Acids Res. 2019; 47:6073–6083

8. Risso D, Ngai J, Speed TP, et al. Normalization of RNA-seq data using factor analysis of control genes or samples. Nat. Biotechnol. 2014; 32:896–902

9. . nSolver™ 4.0 Analysis Software User Manual. 2018;

10. Vandesompele J, De Preter K, Pattyn F, et al. Accurate normalization of real-time quantitative RT-PCR data by geometric averaging of multiple internal control genes. Genome Biol. 2002; 3:research0034.1

11. Perkins JR, Dawes JM, McMahon SB, et al. ReadqPCR and NormqPCR: R packages for the reading, quality checking and normalisation of RT-qPCR quantification cycle (Cq) data. BMC Genomics 2012; 13:296

12. Waggott D, Chu K, Yin S, et al. Gene expression NanoStringNorm: an extensible R package for the pre-processing of NanoString mRNA and miRNA data. Bioinforma. Appl. NOTE 2012; 28:1546–1548

13. Wang H, Horbinski C, Wu H, et al. NanoStringDiff: a novel statistical method for differential expression

analysis based on NanoString nCounter data. Nucleic Acids Res. 2016; 44:gkw677

14. Jia G, Wang X, Li Q, et al. Rcrnorm: An integrated system of random-coefficient hierarchical regression models for normalizing nanostring ncounter data. Ann. Appl. Stat. 2019; 13:1617–1647

15. Canouil ML, Bouland GA, Lie Bonnefond A, et al. NACHO: an R package for quality control of NanoString nCounter data. Bioinformatics 2020; 36:970–971

16. D'Arcy M, Fleming J, Robinson WR, et al. Race-associated biological differences among Luminal A breast tumors. Breast Cancer Res. Treat. 2015; 152:437–448

17. Hall IJ, Moorman PG, Millikan RC, et al. Comparative Analysis of Breast Cancer Risk Factors among African-American Women and White Women. Am. J. Epidemiol. 2005; 161:40–51

18. Brennan P, Van Der Hel O, Moore LE, et al. Tobacco smoking, body mass index, hypertension, and kidney cancer risk in central and eastern Europe. Br. J. Cancer 2008; 99:1912–1915

19. Moore LE, Nickerson ML, Brennan P, et al. Von Hippel-Lindau (VHL) inactivation in sporadic clear cell renal cancer: Associations with germline VHL polymorphisms and etiologic risk factors. PLoS Genet. 2011; 7:

20. Sabry M, Zubiak A, Hood SP, et al. Tumor- and cytokine-primed human natural killer cells exhibit distinct phenotypic and transcriptional signatures. PLoS One 2019; 14:e0218674

21. Nickles D, Sandmann T, Ziman R, et al. NacoStringQCPro.

22. Venables WN, Ripley BD. Modern Applied Statistics with S. 2002;

23. Bullard JH, Purdom E, Hansen KD, et al. Evaluation of statistical methods for normalization and differential expression in mRNA-Seq experiments. BMC Bioinformatics 2010; 11:94

24. Anders S, Huber W. Differential expression analysis for sequence count data. Genome Biol. 2010; 11:R106

25. Love MI, Huber W, Anders S. Moderated estimation of fold change and dispersion for RNA-seq data with DESeq2. Genome Biol. 2014; 15:550

26. Ritchie ME, Phipson B, Wu D, et al. limma powers differential expression analyses for RNA-sequencing and microarray studies. Nucleic Acids Res. 2015; 43:e47–e47

27. Rousseeuw PJ. Silhouettes: A graphical aid to the interpretation and validation of cluster analysis. J. Comput. Appl. Math. 1987; 20:53–65

28. Shabalin AA. Gene expression Matrix eQTL: ultra fast eQTL analysis via large matrix operations. Bioinformatics 2012; 28:1353–1358

29. Bhattacharya A, García-Closas M, Olshan AF, et al. A Framework for Transcriptome-Wide Association Studies in Breast Cancer in Diverse Study Populations. bioRxiv 2019; 769570

30. Benjamini Y, Hochberg Y. Controlling the False Discovery Rate: A Practical and Powerful Approach to Multiple. Source J. R. Stat. Soc. Ser. B 1995; 57:

31. Parker JS, Mullins M, Cheang MCU, et al. Supervised risk predictor of breast cancer based on intrinsic subtypes. J. Clin. Oncol. 2009; 27:1160–1167

32. Gendoo DMA, Ratanasirigulchai N, Schröder M, et al. genefu: a package for breast cancer gene expression analysis. 2018;

33. Székely GJ, Rizzo ML. The Energy of Data. Annu. Rev. Stat. Its Appl. 2017; 4:447–479

34. Dai X, Xiang L, Li T, et al. Cancer hallmarks, biomarkers and breast cancer molecular subtypes. J. Cancer 2016; 7:1281–1294

35. Elizabeth M, Hammond H, Hayes DF, et al. American Society of Clinical Oncology/College of American Pathologists Guideline Recommendations for Immunohistochemical Testing of Estrogen and Progesterone Receptors in Breast Cancer. J. Clin. Oncol. 2010; 28:2784–2795

36. Curtis C, Shah SP, Chin SF, et al. The genomic and transcriptomic architecture of 2,000 breast tumours reveals novel subgroups. Nature 2012; 486:346–352

37. Perou CM, Sørile T, Eisen MB, et al. Molecular portraits of human breast tumours. Nature 2000; 406:747–752

38. Sørlie T, Tibshirani R, Parker J, et al. Repeated observation of breast tumor subtypes in independent gene expression data sets. Proc. Natl. Acad. Sci. U. S. A. 2003; 100:8418–8423

39. Hoadley KA, Yau C, Hinoue T, et al. Cell-of-Origin Patterns Dominate the Molecular Classification of 10,000 Tumors from 33 Types of Cancer. Cell 2018; 173:291-304.e6

40. Picornell AC, Echavarria I, Alvarez E, et al. Breast cancer PAM50 signature: Correlation and concordance between RNA-Seq and digital multiplexed gene expression technologies in a triple negative breast cancer series. BMC Genomics 2019; 20:452

41. Mantel N. The detection of disease clustering and a generalized regression approach. Cancer Res.

1967; 27:209–220

42. Breheny P, Stromberg A, Lambert J. P-Value histograms: Inference and diagnostics. High-Throughput 2018; 7:

43. Sikand K, Singh J, Ebron JS, et al. Housekeeping gene selection advisory: glyceraldehyde-3-phosphate dehydrogenase (GAPDH) and β-actin are targets of miR-644a. PLoS One 2012; 7:e47510

44. Barber RD, Harmer DW, Coleman RA, et al. GAPDH as a housekeeping gene: analysis of GAPDH mRNA expression in a panel of 72 human tissues. Physiol. Genomics 2005; 21:389–395

45. Raulerson CK, Ko A, Kidd JC, et al. Adipose Tissue Gene Expression Associations Reveal Hundreds of Candidate Genes for Cardiometabolic Traits. 2019;

46. Aguet F, Brown AA, Castel SE, et al. Genetic effects on gene expression across human tissues. Nature 2017; 550:204–213

47. Lee C. Genome-wide expression quantitative trait loci analysis using mixed models. Front. Genet. 2018; 9:

48. Jiang N, Wang M, Jia T, et al. A robust statistical method for association-based eQTL analysis. PLoS One 2011; 6:

49. Hyun MK, Ye C, Eskin E. Accurate discovery of expression quantitative trait loci under confounding from spurious and genuine regulatory hotspots. Genetics 2008; 180:1909–1925

50. Mao W, Hausler R, Chikina M. DataRemix: a universal data transformation for optimal inference from gene expression datasets.

51. Irizarry RA, Hobbs B, Collin F, et al. Exploration, normalization, and summaries of high density oligonucleotide array probe level data. Biostatistics 2003; 4:249–264

52. Freytag S, Gagnon-Bartsch J, Speed TP, et al. Systematic noise degrades gene co-expression signals but can be corrected. BMC Bioinformatics 2015; 16:309

53. Bhattacharya A, Hamilton AM, Troester MA, et al. Code and summary results for 'An approach for normalization and quality control for NanoString RNA expression data'. 2020;

54. Bhattacharya A, Garcia-Closas M, Olshan AF, et al. Code, models, and results for CBCS TWAS Paper. Github 2019;

**FIGURE CAPTIONS**

Figure 1: **Graphical summary of both nSolver and RUVSeq normalization pipelines.** The quality control and normalization process starts with familiarization with the data (**Step 1**) and technical quality control to flag samples with potentially poor quality (**Step 2**). After a set of housekeeping genes are selected (**Step 3**), important unwanted technical variables are also investigated through visualization techniques (**Step 4**). Problematic samples (e.g. those that are flagged multiple times in technical quality control checks) are excluded. Next, the data is normalized using upper quartile normalization and RUVSeq (**Step 5**), and the normalized data is visualized to assess the removal of unwanted technical variation and retention of important biological variation (**Step 6**). Steps 3—6 are iterated until technical variation is satisfactorily removed, changing the set of housekeeping genes or the number of dimensions of unwanted technical variation ($k$) estimated using RUVSeq. This data can then be used for downstream analysis (**Step 7**).

Figure 2: **Quality control and normalization validation in CBCS. (A)** Boxplot of percent of endogenous genes below the limit of detection (LOD) ($Y$-axis) over varying numbers of the 11 housekeeping genes below LOD ($X$-axis), colored by CBCS study phase. Note that the $X$-axis scale is decreasing. **(B)** Kernel density plots of deviations from median per-sample $\log_2$-expression from the raw, nSolver-, RUVSeq-, NanoStringDiff-, and RCRnorm-normalized expression matrices, colored by CBCS study phase. **(C)** Plots of the first principal component ($X$-axis) vs. second principal component ($Y$-axis) colored by estrogen receptor subtype of the raw, nSolver-, RUVSeq-, NanoStringDiff-, and RCRnorm-normalized expression data. **(D)** Violin plots of the distribution of per-sample silhouette values, as calculated to study phase, using raw, nSolver-, RUVSeq-, NanoStringDiff-, and RCRnorm-normalized expression. The boxplot shows the 25% quartile, median, and 75% quartile of the distribution, and the plotted triangle shows the mean of the distribution.

Figure 3: **eQTL analysis in CBCS. (A)** Cis-trans plots of eQTL results from nSolver-normalized (left) and RUVSeq-normalized data with chromosomal position of eSNP on the $X$-axis and the transcription start site of associated gene in the eQTL (eGene) on the $Y$-axis. Points for eQTLs are colored by FDR-

adjusted $P$-value of the association. The dotted line provides a 45-degree reference line for cis-eQTLs. **(B)** Number of cis- (left) and trans-eQTLs (right) across various FDR-adjusted significance levels. The number of eQTLs identified in nSolver-normalized data is shown in red and the number of eQTLs identified in RUVSeq-normalized data is shown in blue.

Figure 4: **Differential expression analysis from Sabry et al** [20]**. (A)** Venn diagram of the number of differentially expressed genes using nSolver-normalized (blue) and RUVSeq-normalized data (red) across comparisons for IL-2-primed (top) and CTV-1-primed NK cells (bottom). **(B)** Raw $P$-value histograms for differential expression analysis using nSolver-normalized (blue) and RUVSeq-normalized (red) data across the two comparisons. **(C)** Scatterplots of $\log_2$-fold changes from differential expression analysis using RUVSeq-normalized data ($X$-axis) and nSolver-normalized data ($Y$-axis) for any gene identified as differentially expressed in either one of the two datasets. Points are colored by the datasets in which that given gene was classified as differentially expressed. The size of point reflects the standard error of the effect size as estimated in the RUVSeq-normalized data. $X = 0, Y = 0$, and the 45-degree lines are provided for reference.

Figure 5: **Normalization differences in bladder cancer dataset. (A)** RLE plot from bladder cancer dataset, colored by assay month. **(B)** Boxplot of first principal component of expression by tumor collection site (location) across nSolver- (top) and RUVSeq-normalized (bottom) data. **(C)** Boxplot of first principal component of expression by tumor grade across nSolver- (top) and RUVSeq-normalized (bottom) data.

Figure 6: **Equal performance of normalization procedures in kidney cancer dataset. (A)** RLE plot of per-sample deviations from the median for raw, nSolver-, and RUVSeq-normalized data. **(B)** Scatter plot of the first and second principal component of nSolver- (left) and RUVSeq-normalized (right) expression, colored by high and low DV300. **(C)** Scatter plot of the first and second principal component of nSolver- (left) and RUVSeq-normalized (right) expression, colored by tumor stage.

## 1. Data familiarization

- *Determine limit of detection*
- *Determine raw median expression per sample*

## 2. Technical quality control

- ***Using nSolver Functions:*** *Flag samples with Imaging, Binding Density, Positive Control Linearity, and Limit of Detection QC flags*
- ***Using Endogenous Genes:*** *Flag samples with high proportions of endogenous genes below the limit of detection (LOD)*
- ***Using Housekeeping Genes:*** *Flag samples with high proportions of housekeeping genes below the LOD*

## 3. Identify housekeeping genes for normalization

- *Assess expression of housekeeping genes across biological variables*
- *Flag housekeeping genes frequently detected below the LOD*

## 4. Pre-normalization data visualization

- *Create RLE plots/principal component plots to visually inspect flagged samples and identify outliers indicative of sample/assay-level failure*
- *Assess variation across technical and experimental variables*

## 4a. Exclude problematic samples

## 5. RUVSeq normalization

- *Perform upper quartile normalization (Bullard 2010)*
- *Perform normalization with RUVg (Risso 2014)*

**Iterate over *k***

**Unsatisfactory**

## 6a. Visualization

- *Create RLE plots/principle component plots*
- *Assess variation across technical variables*

## 6b. Biological checks

- *Assess known intrinsic biological associations/patterns*

## 7. Downstream analysis