Multi-centre, multi-vendor reproducibility of 7T QSM and R$_2$* in the human brain: results from the UK7T study

Catarina Rua[a*], William T Clarke[b], Ian D Driver[c], Olivier Mougin[d], Andrew T. Morgan[e], Stuart Clare[b], Susan Francis[d], Keith Muir[e], David Porter[e], Richard Wise[c], Adrian Carpenter[a], Guy Williams[a], James B Rowe[f,g], Richard Bowtell[d], Christopher T Rodgers[a]

a) Wolfson Brain Imaging Centre, Department of Clinical Neurosciences, University of Cambridge, Cambridge, United Kingdom.
   (The Wolfson Brain Imaging Centre, Box 65, Cambridge Biomedical Campus, Cambridge, UK, CB2 0QQ)
b) Wellcome Centre for Integrative Neuroimaging, FMRIB, Nuffield Department of Clinical Neurosciences, University of Oxford, Oxford, United Kingdom.
   (Wellcome Centre for Integrative Neuroimaging, FMRIB, Level 0, John Radcliffe Hospital, Oxford, United Kingdom, OX3 9DU)
c) Cardiff University Brain Research Imaging Centre, School of Psychology, Cardiff University, Cardiff, United Kingdom.
   (Cardiff University Brain Research Imaging Centre, Cardiff University, Maindy Road, Cardiff, CF24 4HQ)
d) Sir Peter Mansfield Imaging Centre, School of Physics and Astronomy, University of Nottingham, Nottingham, United Kingdom.
   (Sir Peter Mansfield Imaging Centre, University of Nottingham, University Park, Nottingham, NG7 2RD)
e) Imaging Centre of Excellence, University of Glasgow, Glasgow, United Kingdom
   (Imaging Centre of Excellence, Queen Elizabeth University Hospital, Langlands Dr, Glasgow, United Kingdom, G51 4LB)
f) Department of Clinical Neurosciences and Cambridge University Hospitals NHS Trust, University of Cambridge, Cambridge, United Kingdom
   (Department of Clinical Neurosciences, Herchel Smith Building, Cambridge Biomedical Campus, Cambridge CB2 0SZ)
g) Medical Research Council Cognition and Brain Sciences Unit, University of Cambridge, Cambridge, United Kingdom.
   (MRC Cognition and Brain Sciences Unit, University of Cambridge, 15 Chaucer Road, Cambridge, CB27EF)

* To whom correspondence should be addressed

Abstract

We present the reliability of ultra-high field T_2^* MRI at 7T, as part of the UK7T Network's "Travelling Heads" study. T_2^*-weighted MRI images can be processed to produce quantitative susceptibility maps (QSM) and R_2^* maps. These reflect iron and myelin concentrations, which are altered in many pathophysiological processes. The relaxation parameters of human brain tissue are such that R_2^* mapping and QSM show particularly strong gains in contrast-to-noise ratio at ultra-high field (7T) *vs* clinical field strengths (1.5 - 3T). We aimed to determine the inter-subject and inter-site reproducibility of QSM and R_2^* mapping at 7T, in readiness for future multi-site clinical studies.

Methods: Ten healthy volunteers were scanned with harmonised single- and multi-echo T_2^*-weighted gradient echo pulse sequences. Participants were scanned five times at each "home" site and once at each of four other sites. The five sites had 1x Philips, 2x Siemens Magnetom, and 2x Siemens Terra scanners. QSM and R_2^* maps were computed with the Multi-Scale Dipole Inversion (MSDI) algorithm (https://github.com/fil-physics/Publication-Code). Results were assessed in relevant subcortical and cortical regions of interest (ROIs) defined manually or by the MNI152 standard space.

Results and Discussion: Mean susceptibility (χ) and R_2^* values agreed broadly with literature values in all ROIs. The inter-site within-subject standard deviation was 0.001 – 0.005 ppm (χ) and 0.0005 – 0.001 ms^{-1} (R_2^*). For χ this is 21-95% better than 3T reports, and 15-124% better for R_2^*. The median ICC from within- and cross-site R_2^* data was 0.98 and 0.91, respectively. Multi-echo QSM had greater variability *vs* single-echo QSM especially in areas with large B_0 inhomogeneity such as the inferior frontal cortex. Across sites, R_2^* values were more consistent than QSM in subcortical structures due to differences in B_0-shimming. On a between-subject level, our measured χ and R_2^* cross-site variance is comparable to within-site variance in the literature, suggesting that it is reasonable to pool data across sites using our harmonised protocol.

Conclusion: The harmonized UK7T protocol and pipeline delivers over a 2-fold improvement in the coefficient of reproducibility for QSM and R_2^* at 7T compared to previous reports of multi-site reproducibility at 3T. These protocols are ready for use in multi-site clinical studies at 7T.

Keywords

7 tesla; MRI; Quantitative Susceptibility Mapping; R_2^* mapping; Multi-centre; Reproducibility.

1. Introduction

Neurodegenerative diseases are a significant global health burden. In many instances, neurodegeneration is associated with the deposition of iron in the brain. Understanding the patterns of deposition and their association with other risk factors is a key area of clinical research, but progress has been limited by the need to scale over multi-centre trials.

A popular approach to estimating iron concentration in the human brain uses gradient-echo (GE) magnetic resonance imaging (MRI). In grey matter, iron is mainly found in the protein ferritin, where it exists in a paramagnetic state (Langkammer et al., 2012). This paramagnetic iron interacts with the MRI scanner's static magnetic field (B_0) causing local dipolar field perturbations. These accentuate the rate of transverse signal decay causing T_2^* relaxation in surrounding tissue, which is visible as decreasing signal amplitude with increasing echo time in a series of GE images. This effect causes an increase in the *rate* of transverse relaxation, R_2^*, which correlates well with non-heme iron concentrations in grey matter (Gelman et al., 1999; Langkammer et al., 2010), and has been used to investigate the distribution of iron in the healthy brain and in disease (Haacke et al., 2005; Yao et al., 2009; Li et al., 2019).

The local presence of iron (and to a lesser extent myelin and calcium) also affects the signal phase of GE images because of the effect of the field perturbation on the local Larmor frequency (House et al., 2007; He et al., 2009; Lee et al., 2012). Quantitative Susceptibility Mapping (QSM) methods attempt to deconvolve these dipole phase patterns to identify the sources of the magnetic field inhomogeneity. In other words, QSM estimates quantitative maps of tissue magnetic susceptibility χ from GE phase data (Li and Leigh, 2004; Reichenbach, 2012; Wang and Liu, 2015). This approach has shown sensitivity to several neurological conditions (Lotfipour et al., 2012; Acosta-Cabronero et al., 2013; Blazejewska et al., 2015; Acosta-Cabronero et al., 2016) and offers advantages over magnitude R_2^* such as having reduced blooming artifacts or being able to distinguish between paramagnetic and diamagnetic substances (Eskreis-Winkler et al., 2017).

R_2^* imaging and QSM have been shown to provide reproducible results in single-site
and cross-site studies at 1.5T and 3T (Hinoda et al., 2015; Cobzas et al., 2015; Deh et
al., 2015; Lin et al., 2015; Santin et al., 2017; Feng et al., 2018; Spincemaille et al.,
2019).

The dipole-inversion problem at the heart of QSM methods benefits from the
increased sensitivity to magnetic susceptibility variation and spatial resolution at ultra-
high fields ($B_0 \geq 7$ T) (Yacoub et al., 2001; Reichenbach et al., 2001; Tie-Qiang et al.,
2006; Duyn et al., 2007; Wharton and Bowtel, 2010). At 7T, close attention must be
paid to B_0 shimming and gradient linearity to achieve accurate QSM and R_2^* mapping
(Yang et al., 2010). Head position is also an important factor that affects the
susceptibility anisotropy (Lancione et al., 2017; Li et al., 2017).

In this study, we introduce single-echo and multi-echo GE imaging protocols for QSM
and R_2^* mapping at 7T which were standardised on three different 7T MRI scanner
platforms, from two different vendors. We applied this standardised protocol in the
UK7T Network's "Travelling Heads" study on 10 subjects scanned at 5 sites. We report
reproducibility for derived R_2^* and QSM maps and make recommendations for the
design of future multi-centre studies.

# Site	Vendor	Scanner Model	Gradient Performance	Installation Date (Month-Year)	Software Version
1 Wellcome Centre for Integrative Neuroimaging (FMRIB), University of Oxford	Siemens	Magnetom 7T	70 mT m^{-1} 200 mT m^{-1} ms^{-1}	Dec-2011	VB17A
2 Cardiff University Brain Research Imaging Centre, Cardiff University	Siemens	Magnetom 7T	70 mT m^{-1} 200 mT m^{-1} ms^{-1}	Dec-2015	VB17A
3 Sir Peter Mansfield Imaging Centre, University of Nottingham	Philips	Achieva 7T	40 mT m^{-1} 200 mT m^{-1} ms^{-1}	Sep-2005	R5.1.7.0
4 Wolfson Brain Imaging Centre, University of Cambridge	Siemens	Magnetom Terra	80 mT m^{-1} 200 mT m^{-1} ms^{-1}	Dec-2016	VE11U
5 Imaging Centre of Excellence, University of Glasgow	Siemens	Magnetom Terra	80 mT m^{-1} 200 mT m^{-1} ms^{-1}	Mar-2017	VE11U

120 **Table 1:** Details of the scanners and hardware used for the UK7T Network's Travelling
121 Heads study.
122

2. Methods

2.1. Measurement setup

Ten healthy volunteers (3 female, 7 male; age 32.0±5.9 years) were recruited: comprising two subjects from each of the five 7T imaging sites in the UK7T Network (described in Table 1). Each subject was scanned five times at their "home" site, and once at the other sites, under local ethics approval for multi-site studies obtained at Site-4 (HBREC.2017.08). Scans for each subject were completed within a period of between 83 and 258 days.

In every scan session, B_0 shimming was performed using the vendors' default second-order (or third-order for Site-4 and Site-5) B_0-shimming routines. B_1^+-calibration was performed initially using the vendor's default adjustment scans. A 3D DREAM sequence (Nehrke et al., 2012; Ehses et al., 2019) was subsequently acquired and the transmit voltage (or power attenuation) was then adjusted for all subsequent imaging based on the mean flip-angle from the brain in an anatomically-specified axial slice of the 3D DREAM flip angle map as described in Clarke et al. (2019). Single-echo (SE) 0.7mm isotropic resolution T_2^*-weighted GE data were then acquired with: TE/TR=20/31ms; FA=15°; bandwidth=70Hz/px; in-plane acceleration-factor=4 (Sites-1/2/4/5) or 2x2 (Site-3); FOV=224x224x157mm^3; scan-time=~9min. Multi-echo (ME) 1.4mm isotropic resolution T_2^*-weighted GE data were acquired with: TE_1/TR=4/43ms; 8 echoes with monopolar gradient readouts; echo-spacing=5ms; FA=15°; bandwidth=260Hz/px; acceleration-factor=4 (Sites-1/2/4/5) or 2x1.5 (Site-3); FOV=269x218x157mm^3; scan-time ~6min (Sites-1/2/4/5) and ~4min (Site-3). For Siemens data, coil combination was performed using a custom implementation of Roemer's algorithm, as previously described (Clarke et al., 2019). Subject 6's SE scan failed to reconstruct using Roemer's method on data from the 1[st] visit at Site-5 so a sum-of-squares (SoS) algorithm was used for coil combination for that scan instead. A 0.7mm isotropic MP2RAGE scan was used for within- and cross-site registration as previously described (Mougin et al., 2019).

2.2. QSM and R_2* data processing

QSM maps were generated from both the SE and ME T_2*-weighted datasets using the Multi-Scale Dipole Inversion (MSDI) algorithm, as implemented in QSMbox v2.0 (Acosta-Cabronero et al., 2018). Briefly: first the local field was estimated by phase unwrapping (Abdul-Rahman et al., 2005) and weighted least squares phase echo fitting was performed on the ME data. Then, for both SE and ME data, background field was removed using the Laplacian Boundary Value (LBV) method followed by the variable Spherical Mean Value (vSMV) algorithm with an initial kernel radius of 40mm (Zhou et al., 2014; Acosta-Cabronero et al., 2018). MSDI inversion was estimated with two scales: the self-optimised lambda method was used on the first scale with filtering performed using a kernel with 1mm radius, and on the second scale the regularization term was set to $\lambda=10^{2.7}$ (the optimal value for *in-vivo* 7T datasets found in (Acosta-Cabronero et al., 2018)) and filtering was done with a kernel radius set to 5mm. Brain masks used in the analysis were obtained with FSL's Brain Extraction Tool (BET) with fractional intensity threshold=0.2 for SE data (Smith, 2002). These were then mapped to ME data space.

On the ME data, QSM was reconstructed seven more times: with the shortest echo (TE1=4 ms), with the two shortest echoes (i.e. TE_1/TE_2 = 4/9 ms), with the three shortest echoes (i.e. $TE_1/TE_2/TE_3$ = 4/9/14 ms), and so forth.

On the ME dataset, voxel-wise quantitative maps of R_2* were obtained using the Auto-Regression on Linear Operations (ARLO) algorithm for fast monoexponential fitting (Pei et al., 2015).

2.3. Data Registration

The neck was cropped from the magnitude data with FSL's "robustfov" command (https://fsl.fmrib.ox.ac.uk/fsl/), applied to the SE data and the 4[th] echo of the ME data. High-resolution SE and ME templates were made from this cropped data for each subject with antsMultivariateTemplateConstruction2.sh from the Advanced Normalization Tools (ANTs, http://stnava.github.io/ANTs/). Two approaches were compared: transformations using rigid registration with mutual information similarity metric (denoted as "Rigid" below) or using symmetric diffeomorphic image registration

with cross-correlation similarity metric (denoted "SyN" below). Other settings were

kept the same for both approaches: 4 steps with 0.1 gradient step size, maximum

iterations per step 1000, 500, 250 and 100, smoothing factors per step of 4, 3, 2, and 1

voxels, and shrink factors per step of 12x, 8x, 4x, and 2x. The resulting registrations

were then applied to the QSM and R_2* maps which were averaged to create SE and ME

QSM and R_2* templates for each subject.

2.4. Selection of Regions of Interest (ROIs)

Five regions of interest (Substantia Nigra, Red Nucleus, Caudate Nucleus, Putamen and

Globus Pallidus) were manually segmented based on the subject-specific QSM

templates of the SE data registered with the "SyN" approach. In order to minimize the

amount of segmentation variability, these ROIs were then mapped to the SE "Rigid",

and ME "SyN" and ME "Rigid" spaces with nearest neighbour interpolation and via

non-linear registrations obtained with the default settings in the

antsRegistrationSyN.sh command in ANTs.

Magnitude data were first registered to the T_1-weighted MP2RAGE scans (Rigid

transformations; MI similarity metric) and later to the standard T_1 "MNI152 brain"

(Montreal Neurological Institute 152) (using settings in antsRegistrationSyN.sh) applied

to the SE data and to the 1^{st} echo of the ME data. These registrations were then used

to map the 48 probabilistic cortical ROIs, "cortical ROIs", from the Harvard-Oxford

Cortical Atlas and the 21 probabilistic subcortical ROIs, "subcortical ROIs", from the

Harvard Oxford Subcortical Atlas to the QSM and R_2* template spaces.

The T_1-weighted MP2RAGE data was bias-field corrected, brain extracted, and

segmented into five tissues using SPM (https://www.fil.ion.ucl.ac.uk/spm/): the grey

matter (GM), white matter (WM) and cerebral-spinal fluid (CSF) volumes were mapped

into each subject-specific QSM template space. Then, using "fslmaths" from FSL

(https://fsl.fmrib.ox.ac.uk/fsl/), the mapped cortical ROIs were thresholded at 10% of

the "robust range" of non-zero voxels and multiplied by the GM tissue map in order to

obtain GM-specific cortical ROIs. The mapped subcortical ROIs were thresholded at

50% of the "robust range" of non-zero voxels. From these, any CSF voxels were

excluded from the left and right Caudate Nucleus, Putamen and Globus Pallidus, and
the voxel sets from the left and right counterparts were merged together.

From the SE and ME data, average χ and R_2* values were extracted from the manual
and Atlas-based ROIs for all volunteers and sessions in template space (values given in
Supplementary Material 1).

In order to estimate where the magnetic field is spatially more variable, field-maps
were first estimated from the ME datasets. ΔB_0 was then calculated per-voxel as the
average difference between the field in a voxel and its immediate nearest neighbors.
The average ΔB_0 was extracted for each of the cortical ROIs and averaged across all
subjects and sessions. Then the cortical ROIs were divided into two groups based on
the ΔB_0 values: wherever $|\Delta B_0| > 0.005\ Hz$ the ROI was grouped into "high ΔB_0"
regions, otherwise it was grouped into "low ΔB_0" regions.

We explored three possible susceptibility reference regions for QSM processing. The
average QSM signal was extracted from:

1. A whole brain mask, "wb";

2. A whole-brain CSF mask eroded in two steps, "csf";

3. A manually placed cylindrical ROI in the right ventricle, "cyl" (across all subjects
231   the ROI volume was 104±11 mm^3).

233   2.5. Statistical Analysis

234   Statistical analysis was performed with R 3.5.3 (R Core Team, 2013). Cross-site analysis
235   used only the 1st scan at the "home" site along with the scans at the other four sites.
236   To obtain the within subject average, AV_w, the χ and R_2* values were averaged within
237   the same site and across the sites and then averaged across subjects:

$$AV_w = \frac{\sum_{i=1}^{m} \left(\sum_{j=1}^{n} x_{ij}/n\right)}{m} \qquad [1]$$

238   where n is the number of sessions ($n = 5$ for within-site and cross-site) and m the
239   number of subjects. Relative reliability was measured using the intra-class correlation
240   coefficient (ICC) from within and cross-site data independently for each ROI (Weir,
241   2005):

$$ICC = \frac{MS_b - MS_w}{MS_b + MS_w(n-1)} \qquad [2]$$

242   where MS_b and MS_w are the between-subjects and within-subjects mean square from

243   a random-effects, one-way analysis of variance (ANOVA) model. Intra-subject absolute

244   variability is assessed by measuring the within-subject standard-deviation (SD$_w$)

245   calculated as (Santin et al., 2017):

246
$$SD_w = \sqrt{\frac{\sum_{i=1}^{m}\sigma_i{}^2}{m}} \text{ with } \sigma_i = \frac{\sqrt{\sum_{j=1}^{n}(x_{ij}-\bar{x}_i)^2}}{n-1} \qquad [3]$$

247   where $\bar{x}_i = \sum_{j=1}^{n} x_{ij}/n$ is the replicate average for each subject. SD$_w$ was computed

248   using within-site data and cross-site data independently. Similarly, cross-subject

249   variability was calculated by measuring the between-subject standard-deviation (SD$_b$):

$$SD_b = \sqrt{\frac{\sum_{i=1}^{m}\sum_{j=1}^{n}(x_{ij}-x_{avg})^2}{n \times m - 1}} \qquad [4]$$

where $x_{avg} = \sum_{i=1}^{m}\sum_{j=1}^{n} x_{ij}/(n \times m)$ is the measurement average across subjects and

sessions. Note that SD$_b$ is computed using data from all sites.

Statistical testing on AV$_w$, SD$_w$ and ICC values extracted from manual and template-

based ROIs was done by first fitting the data with normal, log-normal, gamma and

logistic distributions. The goodness-of-fit statistics for the parametric distributions

were calculated and the distribution which showed the lowest Akaikes Information

Criterion was then used on a general linear model fitting. All models included as fixed

main effects ROI number and data type (within- and cross-site). When evaluating the

data registration type, the model also included registration type ("Rigid" and "SyN") as

a fixed main effect. When testing for QSM reference, the model also included

reference region ("wb", "csf", and "cyl") as a fixed main effect. On ME QSM data, a

model was fitted which also included the number of echoes processed as a fixed main

effect. When comparing the manual and subcortical ROIs, the ROI type (manual vs.

atlas-based) was also included as a fixed main effect. Finally, on the data from the

cortical ROIs, ROI number was replaced with "high ΔB_0" and "low ΔB_0" ROI type as

covariate. A p-value less than 0.05 was considered significant.

266
267 **Figure 1:** Representative slices of SE χ (A) ME χ (B) and R_2* maps (C) from an example
268 subject. The right Caudate Nucleus (a), Putamen (b) and Globus Pallidus (c) are shown
269 in green.
270

3. Results

Figure 1 shows QSM and R_2* maps for one example subject. Basal ganglia structures,

including Caudate Nucleus, Putamen and Globus Pallidus are clearly visible consistent

with previous findings (Langkammer et al., 2010; Wang et al., 2015; Betts et al., 2016;

Acosta-Cabronero et al., 2016). Supplementary Material 2 Figure 1 highlights the

difference in QSM data quality when using our chosen Roemer coil combination

method *vs* using sum-of-squares coil combination.

Figure 2: Mean and standard deviation literature values of QSM (A) and R₂* (B). The
mean and standard deviation results from this study are also plotted. For data with the
symbol '§' the standard error of the mean was originally reported and has been
rescaled by reported N. Shaded regions correspond to literature data.

3.1. QSM and R₂* results and literature

Figure 2 compares average χ and R₂* values calculated in this study in the five manual

ROIs and three corresponding atlas-based subcortical ROIs against literature ranges.

The SE χ-values and ME χ-values from this study are consistent with literature values at

1.5T, 3T and 7T. R₂* values from this study also agree closely with 7T literature values.

Figure 3. Boxplots from data obtained on the manual ROIs of within- and cross-site AV_w (A), SD_w (B) and ICC (C) of SE and ME QSM, and R_2^*. Data from each ROI is shown with a different marker for each boxplot. Legend: SN=Substantia Nigra; RN: Red Nucleus; CN: Caudate Nucleus; Pu: Putamen; GP: Globus Pallidus.

3.2. Reproducibility of QSM and R_2^*

Figure 3 shows boxplots over ROIs of the within- and cross-site AV_w (A), SD_w (B) and ICC (C) values for the manual ROIs on the χ and R_2^* maps. The AV_w from R_2^* maps measured on the same site is systematically higher compared to the AV_w measured across sites (p < 0.0001; e.g., on the Putamen ROI, $AV_{w\_within\text{-}site}$ = 0.0493 ms^{-1} vs $AV_{w\_cross\text{-}site}$ = 0.0489 ms^{-1}). On this comparison, QSM data did not show significant

differences between within-site and cross-site groups for either SE data (p = 0.053) or

ME data (p = 0.65).

From all the data in the manual ROIs, the median SD_w of SE χ-values was approximately

29% lower than for ME χ-values (p = 0.0010). There was a significantly larger SD_w cross-

site compared to within-site on SE χ data (p < 0.0001; e.g., on the PN ROI, $SD_{w\_within-site}$ =

0.00088 ppm vs $SD_{w\_cross-site}$ = 0.0014 ppm), ME χ (p = 0.033) and on R_2* data (p <

0.0001).

The ICC values for within- and cross-site R_2* data (median ICC was 0.98 and 0.91,

respectively) were found to be significantly higher than values for SE χ (median ICC was

0.89 and 0.64, respectively) or for ME χ (median was ICC 0.76 and 0.38, respectively) (p

= 0.00011). For all measurements, the ICC for cross-site data was significantly lower

than for within-site data (SE QSM: p < 0.0001; ME QSM: p = 0.017; R_2*: p < 0.0001).

Similar statistics were obtained for AV_w, SD_w and ICC measurements in the altas-based

cortical ROIs (Table 2, Supplementary Material 2).

Figure 4. Voxel-wise within- and cross-site standard deviation of an example subject
from SE and ME QSM and R_2* data with data registered with "Rigid" (A) and "SyN" (B)
transformations. Arrows point to regions where the SD_w decreased with the "SyN"
transformations (green) are compared to "Rigid" (white). The right Caudate Nucleus
(a), Putamen (b) and Globus Pallidus (c) are outlined in white.
323

3.3 Registration

The within- and cross-site standard deviations for one axial slice from one example

subject using "Rigid" and "SyN" registration approaches are shown in Figure 4.

Generally, with both registration methods, within-site and cross-site SD_w increases in

veins, in the orbitofrontal regions and at the cortical surface (white and green arrows,

Figure 4). These are areas associated with large B_0 inhomogeneities and gradient non-

linearity. However, there is a decrease in the cross-site standard deviation in the

orbitofrontal region and close to the edges of the cortex when using the "SyN"

compared to the "Rigid" method (green arrows, Figure 4).

On the manual ROIs increased variability was observed for R_2* on "Rigid" registered

data compared to "SyN" (SD_w: $p < 0.0001$; ICC: $p < 0.013$) but not for SE or ME χ: for

example, the median cross-site R_2* SD_w from all ROIs was 0.00066 ms^{-1} using "SyN"

method and 0.00086 ms^{-1} using the "Rigid" registration method. On the atlas-based

cortical ROIs, the same significant trend was observed for R_2* and SE χ data (Table 2,

Supplementary Material 2).

340
341 **Figure 5:** Boxplots from data obtained on the manual ROIs of within- and cross-site SD_w
342 (red and green, respectively) of SE QSM (A) and ME QSM (B) with a whole-brain
343 reference (wb), with a csf reference (csf), and with a cylinder reference (cyl). Data from

each ROI is shown with a different marker for each boxplot. Legend: SN=Substantia Nigra; RN: Red Nucleus; CN: Caudate Nucleus; Pu: Putamen; GP: Globus Pallidus.

3.4 QSM referencing

To assess the optimal QSM susceptibility referencing, Figure 5 shows boxplots of the SD_w for SE and ME χ using different referencing methods on the manual ROIs. On SE χ data, compared to "wb" correction (chosen correction for this study), the "csf" reference did not increase significantly the SD_w (p = 0.93) but with "cyl" the median SD_w increased by approximately 14% (p < 0.0001).

ME χ data showed an increase in the median SD_w of, respectively, 11% (p = 0.00096) and 8% (p = 0.00064) when using "csf" and "cyl" methods for correction. The effect of varying the referencing of QSM data was similar in within-site and cross-site data, for all methods tested.

Figure 6. Boxplots from data obtained on the manual ROIs of within- and cross-site SD_w for ME QSM calculated with different number of echoes. Increasing trend on the median SD_w observed with increasing number of echoes (dotted lines). Legend: SN=Substantia Nigra; RN: Red Nucleus; CN: Caudate Nucleus; Pu: Putamen; GP: Globus Pallidus.

3.5 ME QSM

On average across all the manual ROIs and compared to single echo data, multi-echo data (using two or more echoes) showed a significant 14% increase of the SD_w (Figure 6) and 3% of the ICC (Table 1, Supplementary Material 2). This supports the SE and ME χ comparison in Section 3.2. Similar behaviour was observed on the atlas-based cortical ROIs (Table 2, Supplementary Material 2). In the atlas-based cortical ROIs, long

echo times (i.e. using 6 or more echoes) showed an average increase of 15.7% in SD_w (p

< 0.0001) compared to using 2 to 5 echoes and a decrease of 1.75% in ICC (p < 0.0001)

(Table 2, Supplementary Material 2).

3.6 ROI selection

There is a small but significant higher average χ from manually drawn ROIs compared

to the atlas-based subcortical ROIs in SE QSM data (p < 0.0001; e.g. 0.042±0.009 ppm

vs 0.033±0.010 ppm in the caudate nucleus) and in ME QSM data (p < 0.0001; e.g.

0.048±0.010 ppm vs 0.038±0.011 ppm in the caudate nucleus) (Figure 2). Similarly, for

R_2^* (e.g. 0.041±0.004 ms^{-1} vs 0.039±0.006 ms^{-1} in the caudate nucleus) this difference

was significant (p < 0.0001). In addition, the SD_w was, on average, approximately two

times higher and the ICC lower in the atlas-based subcortical ROIs compared to the

manual ROIs in all datasets (SD_w: SE QSM p < 0.0001, ME QSM p < 0.0001, R_2^* p <

0.0001; ICC: SE QSM p = 0.00021, ME QSM p = 0.0023, R_2^* p = 0.012). So, ROI selection

should be done consistently in a study.

3.7 Spatial distribution of the magnetic field

On the altas-based cortical ROIs the SD_w increased by approximately 28% and 88% on

"high ΔB_0" regions compared to "low ΔB_0" regions on ME χ and R_2^* data, respectively

(p = 0.0011 and p < 0.0001) (Table 2, Supplementary Material 2). Similarly, ICC values

decreased significantly for SE and ME χ and R_2^* values.

4. Discussion

In this paper, the reproducibility of QSM χ and R_2^* measurements in cortical and

subcortical regions of the brain was assessed for the first time in a multi-site study at

7T for two different protocols (a single-echo 0.7mm isotropic T_2^*-weighted scan and a

1.5mm isotropic multi-echo T_2^*-weighted scan), using three different scanner

platforms provided by two different vendors.

Previous studies at 1.5T and 3T have shown good reproducibility for χ and R_2^* data

acquired on the same scanner or across sites (1.5T and 3T) (Hinoda et al., 2015; Cobzas

et al., 2015; Deh et al., 2015; Lin et al., 2015; Santin et al., 2017; Feng et al., 2018;

Spincemaille et al., 2019). In terms of QSM and depending on the subcortical region,
intra-scanner 3T repeatability studies report an SD_w of 0.002-0.005 ppm (Feng et al.,
2018) and 0.004-0.006 ppm (Santin et al., 2017), and the cross-site 3T study by Lin et
al. (2015) reported an average SD_W of 0.006-0.010 ppm. We observed a within-site SD_w
range of 0.0009-0.004 ppm and cross-site SD_w range of 0.001-0.005 ppm at 7T. The
latter is therefore 21-95% better than within sites studies at 3T.

The range of within-site SD_w values for R_2* was averaged 0.0003-0.001 ms^{-1} in our
study and the cross-site SD_w range was 0.0005-0.001 ms^{-1}. The cross-site values are
comparable to the *same site* reported at 3T: 0.0005-0.0009 ms^{-1} (Feng et al., 2018),
0.0006-0.002 ms^{-1} (Santin et al., 2017). Compared to the latter, our cross-site results
show a reduction of 15-124% in R_2* variability.

The higher values of cross-site SD_w compared to the within-site values in our study may
be attributed to the different gradient systems and automatic distortion corrections
used in the different scanner platforms and to the different approaches to shimming,
which lead to different geometrical distortions and dropout regions (Yang et al., 2010).
We showed that the use of a non-linear registration method (here, "SyN" in ANTs)
significantly reduced the inter-scanner variability of cortical QSM compared to rigid-
body registration, indicating that differences in geometric distortion across scanners
were present. The R_2* results for both cortical and subcortical structures also show
significantly lower inter-scanner variability when a non-linear registration was used.

In this study, the reproducibility of QSM using single-echo (SE), high-resolution (0.7
mm isotropic resolution; TE=20ms) and multi-echo standard-resolution (ME) standard-
resolution (1.4 mm isotropic resolution; TE=4, 9, 14, 19, 24, 29, 34 and 39 ms)
protocols were compared, and the results show that the ME QSM data has a
significantly higher variability than SE QSM. Although ME QSM data has been
combined with a magnitude-weighted least squares regression of phase to echo time,
it may carry incorrect phase from late echoes of the echo train that suffered multiple
phase wraps. This has also been verified with an analysis on multi-echo QSM data
reconstructed with different numbers of echoes: long echo times increase significantly
the test-retest variability.

R_2^* values show significantly lower variability, reflected in the higher ICC within and

across-sites compared to corresponding values for χ in subcortical areas. This may be

because the χ estimation is globally more sensitive to background field inhomogeneity

compared to magnitude data. However, in orbitofrontal and lower temporal regions

large through-plane field variations from tissue-air interfaces dominate the field

changes and produce dropouts in the signal magnitude and increase the background

phase, affecting both QSM and R_2^* maps by increasing variability and decreasing ICC.

QSM can only determine relative susceptibility differences (Cheng et al., 2009) and

most approaches to calculation of susceptibility from measured phase yield maps in

which the average value of susceptibility is zero over the masked imaging volume.

Issues related to referencing of QSM data have been investigated (Feng et al., 2018;

Straub et al., 2017), with aim of finding a reference region or tissue to which all

susceptibility values are referred that produces well-defined and reproducible values

of susceptibility. Here we investigated how the choice of reference affects the within-

site and cross-site variability of measured susceptibility at ultra-high-field. We tested

three accepted reference regions: total whole brain signal, "wb", whole brain CSF

eroded in order to exclude any pial or skull surfaces, "csf", and a manually selected

cylindrical ROI in the right ventricle, "cyl". We found that the "cyl" referencing

generally increased the variability of the cross-site and within-site susceptibility

measurements in cortical and subcortical ROIs compared to "wb" referencing. In the

case of ME acquisition the "csf" referencing also increased the variability relative to

"wb" data. This may be because of imprecision in systematically obtaining average

QSM signal from CSF regions. Referencing using a small ROI in the ventricles might be

prone to subjectivity given the natural variation in ventricle size in healthy subjects and

in disease. Furthermore, the ventricles do not contain pure CSF: they are traversed by

blood vessels with a different χ (Sullivan et al., 2002). This makes whole-brain

referencing attractive in many situations. Yet, in patient cohorts where there is

substantial iron load in subcortical structures (Snyder and Connor, 2009), whole brain

referencing might not be an appropriate approach. In this case, the more appropriate

approach will be to choose a small reference region which shows no changes in the

particular disease to be "zero" susceptibility at a cost of a slight increase in SD.

To eliminate operator-dependent bias in segmentation when determining brain
structures, we have analysed data using both manual and atlas-based segmentation.
From our results, manual ROIs showed significantly lower variability compared to atlas-
based methods. This happens because of imprecision in registration between MNI and
subject space as well as the empirical thresholding that was chosen to obtain the
subcortical ROIs. However, traditional manual drawing of ROIs for cohort studies is
difficult, time consuming and potentially unsuitable as it biases results towards
particular cohorts (Collins et al., 2003) so it may not always be the most appropriate
approach.

In this study, harmonized protocols were produced for all five scanners without any
significant sequence alterations, as a product 3D gradient echo (GE) sequence was
readily available on all systems (the product 'gre' sequence from Siemens and the
product 'ffe' from Philips). The protocols and an example dataset are provided in
(Clarke, 2018). Generally, we also relied on the vendors' reconstruction. However, at
the end of the reconstruction pipeline of the Siemens systems we adopted a different
coil combination approach based on Roemer et al. (1990) and Walsh et al. (2000), to
match the SENSE approach implemented on Philips scanners (Pruessmann et al., 1999;
Robinson et al., 2017). This was required due to artifacts appearing on phase images in
Siemens data reconstructed with the vendor's pipeline, such as open-ended fringe
lines or singularities (Chavez et al., 2002) (Figure 1, Supplementary Material 2). These
reduce the consistency of the QSM results (Santin et al., 2017). However, other coil
combination methods such as a selective channel combination approach (Vegh et al.,
2016) or the COMPOSER (COMbining Phase data using a Short Echo-time Reference
scan) method (Bollmann et al., 2018) have also been shown to reduce open-ended
fringe lines and noise in the signal phase. For future investigations, the raw k-space
data collected from all sites in this study has been stored and is available from the
authors upon request.

On the QSM reconstruction, an imperfect background field filtering can influence the
reproducibility of QSM data. For this reason, we performed background removal in

two steps as implemented in QSMbox v2.0 and as described in (Acosta-Cabronero et
al., 2018): first with the LBV approach and then followed by the vSMV method.
Regularized field-to-susceptibility inversion strategies have been proposed to
overcome the ill-posed problem in QSM with data acquired at a single head orientation
(de Rochefort et al., 2010). We opted to use the MSDI implementation in QSMbox v2.0
(Acosta-Cabronero et al., 2018), as it ranked top-10 in all metrics of the 2016 QSM
Reconstruction Challenge (Langkammer et al., 2018), and also now includes a new self-
optimized local scale, which results in a better preservation of phase noise texture and
low susceptibility contrast features. On the second step, the regularization factor, λ,
used for this study was set to $10^{2.7}$, as recommended by Acosta-Cabronero et al. (2018)
based on an L-curve analysis (Hansen et al., 1993) with high-resolution 7T data.
505

506
Figure 7. Illustration of the feasibility of a 7T QSM clinical study. χ (A) and R_2^* (B) for
four ROIs (Substantia Nigra, SN; Caudate Nucleus, CN; Putamen, Pu; Globus Pallidus,
GP) from healthy volunteer (HV) and synthetic "patient" (PT) data for which AV_{lit} and
SD_{lit} were obtained from Langkammer et al. (2016) and SD_b were calculated from data
of the current study. AV_{lit} values for R_2^* were linearly scaled to 7T according to Yao et
al. (2007). Blue bars show the $AV_{lit} \pm SD_{lit}$ and green bars the $AV_{lit} \pm SD_b$. Statistical

differences between HV and PT obtained from Langkammer et al. (2016) are also shown. For each ROI, the sample size that would have been needed to give a significant effect was calculated from the group means, AV_{lit}, and the SD_b per ROI and is shown in circles.

To minimise confounding effects of age or pathology, we assessed test-retest reliability and cross-site variability with ten healthy young subjects. The cross-site, between-subject standard-deviation, SD_b, measured in this study was evaluated together with healthy and Parkinson's disease data from (Langkammer et al., 2016). A power analysis revealed a sample size that would have been required for a multi-site clinical study in each ROI as shown in Figure 7. For all the significant ROIs the number of subjects that would have been required per group was less or equal to 44. Since this is lower than the sample size we have used in this study (90 healthy volunteer scans) and the numbers in the Langkammer study (66 patients and 58 control subjects), it gives strong confidence of feasibility for future 7T QSM clinical studies.

5. Conclusion

We investigated test-retest reliability and reproducibility of T_2*-weighted imaging protocols at ultra-high field MRI. Considering the increase in susceptibility effects at 7T, we found that variability of measurements of QSM χ and R_2* in the basal ganglia are reduced compared to reports from lower field strengths, 1.5T and 3T. Scanner hardware differences give more modest improvements for cortical measurements of QSM χ and R_2*. Multi-echo protocols do not benefit from long echo times as these increase the imprecision in the estimation of QSM. We suggest that 7T MRI is suitable for multicentre quantitative analyses of brain iron, in health and disease.

6. Acknowledgements

The UK7T Network and this work was funded by the UK's Medical Research Council (MRC) [MR/N008537/1]. We thank Julio Acosta-Cabronero for making the QSMbox publically available.

7. Centre funding

The Wellcome Centre for Integrative Neuroimaging is supported by core funding from the Wellcome Trust (203139/Z/16/Z).

Cardiff University Brain Research Imaging Centre is supported by the UK Medical Research Council (MR/M008932/1) and the Wellcome Trust (WT104943).

This research was co-funded by the NIHR Cambridge Biomedical Research Centre. The views expressed are those of the author(s) and not necessarily those of the NHS, the NIHR or the Department of Health and Social Care. The Cambridge 7T MRI facility is co-funded by the University of Cambridge and the Medical Research Council (MR/M008983/1).

8. Individual funding

CTR is funded by a Sir Henry Dale Fellowship from the Wellcome Trust and the Royal Society [098436/Z/12/B]. JBR is supported by the Wellcome Trust (WT103838).

9. References

Abdul-Rahman, H., Gdeisat, M., Burton, D., Michael, L., 2005. Fast three-dimensional phase-unwrapping algorithm based on sorting by reliability following a non-continuous path. App. Optic. 46, 6623-6635. https://doi.org/10.1117/12.611415

Acosta-Cabronero, J., Milovic, C., Mattern, H., Tejos, C., Speck, O., Callaghan, M.F., 2018. A robust multi-scale approach to quantitative susceptibility mapping. NeuroImage 183, 7-24. https://doi.org/10.1016/j.neuroimage.2018.07.065

Acosta-Cabronero, J., Betts, M.J., Cardenas-Blanco, A., Yang, S., Nestor, P.J., 2016. In Vivo MRI mapping of brain iron deposition across the adult lifespan. J. Neurosci. 36, 364–374. https://doi.org/10.1523/JNEUROSCI.1907-15.2016

Acosta-Cabronero, J., Williams, G.B., Cardenas-Blanco, A., Arnold, R.J., Lupson, V., Nestor, P.J., 2013. In vivo quantitative susceptibility mapping (QSM) in Alzheimer's disease. PLoS One 8, e81093. https://doi.org/10.1371/journal.pone.0081093

Barbosa, J.H.O., Santos, A.C., Salmon, C.E.G., 2015. Susceptibility weighted imaging: differentiating between calcification and hemosiderin. Radiologia brasileira 48(2), 93-100. https://doi.org/10.1590/0100-3984.2014.0010

Betts, M.J., Acosta-Cabronero, J., Cardenas-Blanco, A., Nestor, P.J. and Düzel, E., 2016. High-resolution characterisation of the aging brain using simultaneous quantitative susceptibility

mapping (QSM) and R2* measurements at 7 T. Neuroimage 138, pp.43-63.
https://doi.org/10.1016/j.neuroimage.2016.05.024

Bilgic, B., Pfefferbaum, A., Rohlfing, T., Sullivan, E.V., Adalsteinsson, E., 2012. MRI estimates of
brain iron concentration in normal aging using quantitative susceptibility
mapping. Neuroimage 59(3), 2625-2635. https://doi.org/10.1016/j.neuroimage.2011.08.077

Blazejewska, A.I., Al-Radaideh, A.M., Wharton, S., Lim, S.Y., Bowtell, R.W., Constantinescu, C.S.,
Gowland, P.A., 2015. Increase in the iron content of the substantia nigra and red nucleus in
multiple sclerosis and clinically isolated syndrome: a 7 Tesla MRI study. J. Magn. Reson.
Imaging 41, 1065-1070. https://doi.org/10.1002/jmri.24644

Bollmann, S., Robinson, S.D., O'Brien, K., Vegh, V., Janke, A., Marstaller, L., Reutens, D., Barth,
M., 2018. The challenge of bias-free coil combination for quantitative susceptibility mapping at
ultra-high field. Magn. Reson. Med. 79(1), 97-107. https://doi.org/10.1002/mrm.26644

Chavez, S., Xiang, Q., Li, A., 2002. Understanding phase maps in MRI: a new cutline phase
unwrapping method. IEEE Trans Med Imaging 21, 966–977.
https://doi.org/10.1109/TMI.2002.803106

Cheng, Y.-C.N., Neelavalli, J., Haacke, E.M., 2009. Limitations of calculating field distributions
and magnetic susceptibilities in MRI using a Fourier based method. Phys. Med. Biol. 54(5),
1169–1189. https://doi.org/10.1088/0031-9155/54/5/005

Choi, S., Li, X., Harrison, D.M., 2019. The impact of coregitration of gradient recalled echo
images on quantitative susceptibility and R2* mapping at 7T. bioRxiv.
https://doi.org/10.1101/529891

Clarke, W.T., 2018. UK7T Network harmonized neuroimaging protocols.
https://ora.ox.ac.uk/objects/uuid:55ca977f-62df-4cbf-b300-2dc893e36647.

Clarke, W.T., Mougin, O., Driver, I.D., Rua, C., Morgan, A., Asghar, M., Clare, S., Francis, S.,
Wise, R., Rodgers, C.T., Carpenter, T.A., Muir, K., Bowtell, R., 2019. Multi-site harmonization of
7 Tesla MRI neuroimaging protocols. NeuroImage 206, 116335.
https://doi.org/10.1016/j.neuroimage.2019.116335

Cobzas, D., Sun, H.F., Walsh, A.J., Lebel, R.M., Blevins, G., Wilman, A.H., 2015. Subcortical gray
matter segmentation and voxel-based analysis using transverse relaxation and quantitative
susceptibility mapping with application to multiple sclerosis. J. Magn. Reson. Imaging 42(6),
1601–1610. https://doi.org/10.1002/jmri.24951

Collins, D.L., Zijdenbos, A.P., Paus, T., Evans, A.C., 2003. Use of registration for cohort studies.
Medical image registration.

de Rochefort, L., Liu, T., Kressler, B., Liu, J., Spincemaille, P., Lebon, V., Wu, J., Wang, Y., 2010.
Quantitative susceptibility map reconstruction from MR phase data using Bayesian
regularization: validation and application to brain imaging. Magn. Reson. Med. 63, 194–206.
https://doi.org/10.1002/mrm.22187

Deh, K., Nguyen, T.D., Eskreis-Winkler, S., Prince, M.R., Spincemaille, P., Gauthier, S.,
Kovanlikaya, I., Zhang, Y., Wang, Y., 2015. Reproducibility of quantitative susceptibility
mapping in the brain at two field strengths from two vendors. J. Magn. Reson. Imaging 42,
1592–1600. https://doi.org/10.1002/jmri.24943

Deistung, A., Schäfer, A., Schweser, F., Biedermann, U., Güllmar, D., Trampel, R., Turner, R.,
Reichenbach, J.R., 2013. High-resolution MR imaging of the human brainstem in vivo at 7
Tesla. Frontiers in human neuroscience 7, 710. https://doi.org/10.3389/fnhum.2013.00710

Duyn, J.H., van Gelderen, P., Li, T.Q., de Zwart, J.A., Koretsky, A.P. and Fukunaga, M., 2007.
High-field MRI of brain cortical substructure based on signal phase. Proceedings of the
National Academy of Sciences, 104(28), pp.11796-11801.
https://doi.org/10.1073/pnas.0610821104

Ehses, P., Brenner, D., Stirnberg, R., Pracht, E.D., Stöcker, T., 2019. Whole-brain B1-mapping
using three-dimensional DREAM. Magn. Reson. Med. 82(3), 924-934.
https://doi.org/10.1002/mrm.27773

Eskreis-Winkler, S., Zhang, Y., Zhang, J., Liu, Z., Dimov, A., Gupta, A. and Wang, Y., 2017. The
clinical utility of QSM: disease diagnosis, medical management, and surgical planning. NMR in
Biomedicine 30(4), p.e3668. https://doi.org/10.1002/nbm.3668

Feng, X., Deistung, A., Reichenbach, J.R., 2018. Quantitative susceptibility mapping (QSM) and
R2* in the human brain at 3 T: Evaluation of intra-scanner repeatability. Z. Med. Phys. 28, 36–
48. https://doi.org/10.1016/j.zemedi.2017.05.003

Gelman, N., Gorell, J.M., Barker, P.B., Savage, R.M., Spickler, E.M., Windham, J.P., Knight, R.A.,
1999. MR imaging of human brain at 3.0 T: preliminary report on transverse re-laxation rates
and relation to estimated iron content. Radiology 210, 759–767.
https://doi.org/10.1148/radiology.210.3.r99fe41759

Haacke, E.M., Cheng, N., House, M.J., Liu, Q., Neelavalli, J., Ogg, R.J., Khan, A., Ayaz, M., Kirsch,
W., Obenaus, A., 2005. Imaging iron stores in the brain using magnetic resonance imaging.
Magn. Reson. Imag. 23, 1-25. https://doi.org/10.1016/j.mri.2004.10.001

Hansen, P.C., O'Leary, D.P., 1993. The use of the l-curve in the regularization of discrete ill-
posed problems. SIAM J Sci Comput 14(6), 1487-1503. https://doi.org/10.1137/0914086

He, X., Yablonskiy, D.A., 2009. Biophysical mechanisms of phase contrast in gradient echo MRI.
Proc. Natl. Acad. Sci. U.S.A. 106, 13558–13563. https://doi.org/10.1073/pnas.0904899106

Hinoda, T., Fushimi, Y., Okada, T., Fujimoto, K., Liu, C., Yamamoto, A., Okada, T., Kido, A.,
Togashi, K., 2015. Quantitative susceptibility mapping at 3 T and 1.5 T: evaluation of
consistency and reproducibility. Invest. Radiol. 50, 522-530.
https://doi.org/10.1097/RLI.0000000000000159

House, M.J., Pierre, T.G.S., Kowdley, K.V., Montine, T., Connor, J., Beard, J., Berger, J., Siddaiah,
N., Shankland, E., Jin, L.W., 2007. Correlation of proton transverse relaxation rates (R2) with
iron concentrations in postmortem brain tissue from Alzheimer's disease patients. Magn.
Reson. Med. 57, 172–180. https://doi.org/10.1002/mrm.21118

Keuken, M.C., Bazin, P.L., Backhouse, K., Beekhuizen, S., Himmer, L., Kandola, A., Lafeber, J.J.,
Prochazkova, L., Trutti, A., Schäfer, A., Turner, R., 2017. Effects of aging on T1, T2*, and QSM
MRI values in the subcortex. Brain Structure and Function 222(6), 2487-2505.
https://doi.org/10.1007/s00429-016-1352-4

Lancione, M., Tosetti, M., Donatelli, G., Cosottini, M., Costagli, M., 2017. The impact of white
matter fiber orientation in single-acquisition quantitative susceptibility mapping. NMR Biomed.
30, e3798. https://doi.org/10.1002/nbm.3798

Langkammer, C., Schweser, F., Shmueli, K., Kames, C., Li, X., Guo, L., Milovic, C., Kim, J., Wei, H.,
Bredies, K., Buch, S., 2018. Quantitative susceptibility mapping: report from the 2016
reconstruction challenge. Magn. Reson. Med. 79(3), 1661-73.
https://doi.org/10.1002/mrm.26830

Langkammer, C., Pirpamer, L., Seiler, S., Deistung, A., Schweser, F., Franthal, S., Homayoon, N.,
Katschnig-Winter, P., Koegl-Wallner, M., Pendl, T., Stoegerer, E.M., 2016. Quantitative
susceptibility mapping in Parkinson's disease. PLoS One, 11(9), e0162460.
https://doi.org/10.1371/journal.pone.0162460

Langkammer, C., Schweser, F., Krebs, N., Deistung, A., Goessler, W., Scheurer, E., Sommer, K.,
Reishofer, G., Yen, K., Fazekas, F., Ropele, S., Reichenbach, J.R., 2012. Quantitative
susceptibility mapping (QSM) as a means to measure brain iron? A post mortem validation
study. Neuroimage 62(3), 1593–1599. https://doi.org/10.1016/j.neuroimage.2012.05.049

Langkammer, C., Krebs, N., Goessler, W., Scheurer, E., Ebner, F., Yen, K., Fazekas, F., Ropele, S.,
2010. Quantitative MR imaging of brain iron: a postmortem validation study. Radiology 257(2),
455-462. https://doi.org/10.1148/radiol.10100495

Lee, J., Shmueli, K., Kang, B.T., Yao, B., Fukunaga, M., van Gelderen, P., Palumbo, S., Bosetti, F.,
Silva, A.C., Duyn, J.H., 2012. The contribution of myelin to magnetic susceptibility-weighted
contrasts in high-field MRI of the brain. Neuroimage 59, 3967–3975.
https://doi.org/10.1016/j.neuroimage.2011.10.076

Li, G., Zhai, G., Zhao, X., An, H., Spincemaille, P., Gillen, K.M., Ku, Y., Wang, Y., Huang, D., Li, J.,
2019. 3D texture analysis within substantia nigra of Parkinson's disease patients on
quantitative susceptibility maps and R2* maps. NeuroImage 188, 465-472.
https://doi.org/10.1016/j.neuroimage.2018.12.041

Li, L., Leigh, J.S., 2004. Quantifying arbitrary magnetic susceptibility distributions with MR.
Magn. Reson. Med. 51, 1077–1082. https://doi.org/10.1002/mrm.20054

Li, W., Liu, C., Duong, T.Q., van Zijl, P.C., Li, X., 2017. Susceptibility tensor imaging (STI) of the
brain. NMR Biomed. 30(4), p.e3540. https://doi.org/10.1002/nbm.3540

Lin, P.Y., Chao, T.C., Wu, M.L., 2015. Quantitative susceptibility mapping of human brain at 3T:
a multisite reproducibility study. AJNR Am J. Neuroradiol. 36, 467-474.
https://doi.org/10.3174/ajnr.A4137

Lotfipour, A.K., Wharton, S., Schwarz, S.T., Gontu, V., Schaefer, A., Peters, A.M., Bowtell, R.W.,
Auer, D.P., Gowland, P.A., Bajaj, P.S., 2012. High resolution magnetic susceptibility mapping of
the substantia nigra in Parkinson's disease. J. Magn. Reson. Imag. 35, 48-55.
https://doi.org/10.1002/jmri.22752

Mougin, O., Clarke, W., Driver, I., Rua, C., Morgan, A.T., Francis, S., Muir, K., Carpenter, A.,
Rodgers, C., Wise, R., Porter, D., Clare, S., Bowtell, R., 2019. Robustness of PSIR segmentation
and R1 mapping at 7T: a travelling head study. Proc. Intr. Soc. Mag. Reson. Med. 27, 237.

Nehrke, K., Bornert, P., 2012. DREAM--a novel approach for robust, ultrafast, multislice B(1)
mapping. Magn. Reson. Med. 68(5), 1517-1526. https://doi.org/10.1002/mrm.24158

Pei, M., Nguyen, T.D., Thimmappa, N.D., Salustri, C., Dong, F., Cooper, M.A., Li, J., Prince, M.R.,
Wang, Y., 2015. Algorithm for fast monoexponential fitting based on auto-regression on linear
operations (ARLO) of data. Magn. Reson. Med. 73, 843-850.
https://doi.org/10.1002/mrm.25137

Pruessmann, K.P., Weiger, M., Scheidegger, M.B., Boesiger, P., 1999. SENSE: sensitivity
encoding for fast MRI. Magn. Reson. Med. 42(5), 952-962. https://doi.org/10.1002/(SICI)1522-
2594(199911)42:5<952::AID-MRM16>3.0.CO;2-S

R Core team, 2013. R: A language and environment for statistical computing. R Foundation for
Statistical Computing, Vienna, Austria. http://www.R-project.org/.

Reichenbach, J.R., 2012. The future of susceptibility contrast for assessment of anatomy and
function. Neuroimage 62, 1311-1315. https://doi.org/10.1016/j.neuroimage.2012.01.004

Reichenbach, J.R., Jonetz-Mentzel, L., Fitzek, C., Haacke, E.M., Kido, D.K., Lee, B.C., Kaiser,
W.A., 2001. High-resolution blood oxygen-level dependent MR venography (HRBV): a new
technique. Neuroradiology 43, 364–369. https://doi.org/10.1007/s002340000503

Robinson, S.D., Bredies, K., Khabiova, D., Dymerska, B., Marques, J.P., Schweser, F., 2017. An
illustrated comparison of processing methods for MR phase imaging and QSM: combining
array coil signals and phase unwrapping. NMR Biomed. 30, e3601.
https://doi.org/10.1002/nbm.3601

Roemer, P.B., Edelstein, W.A., Hayes, C.E., Souza, S.P., Mueller, O.M., 1990. The NMR phased
array. Magn. Reson. Med. 16(2), 192-225. https://doi.org/10.1002/mrm.1910160203

Santin, M.D., Didier, M., Valabregue, R., Yahia Cherif, L., García-Lorenzo, D., Loureiro de Sousa,
P., Bardinet, E., Lehéricy, S., 2017. Reproducibility of R2* and quantitative susceptibility
mapping (QSM) reconstruction methods in the basal ganglia of healthy subjects. NMR Biomed.
30(4), e3491. https://doi.org/10.1002/nbm.3491

Schweser, F., Deistung, A., Lehr, B.W., Reichenbach, J.R., 2011. Quantitative imaging of intrinsic
magnetic tissue properties using MRI signal phase: an approach to in vivo brain iron
metabolism?. Neuroimage 54(4), 2789-2807.
https://doi.org/10.1016/j.neuroimage.2010.10.070

Smith, S.M., 2002. Fast robust automated brain extraction. Human Brain Mapping 17(3), 143-
155. https://doi.org/10.1002/hbm.10062

Snyder A.M., Connor J.R., 2009. Iron, the substantia nigra and related neurological disorders.
Biochimica et Biophysics Acta 1790, 606-614. https://doi.org/10.1016/j.bbagen.2008.08.005

Spincemaille, P., Liu, Z., Zhang, S., Kovanlikaya, I., Ippoliti, M., Makowski, M., Watts, R., de
Rochefort, L., Venkatraman, V., Desmond, P. and Santin, M.D., 2019. Clinical integration of
automated processing for brain quantitative susceptibility mapping: multi-site reproducibility
and single-site robustness. Journal of Neuroimaging, 29(6), pp.689-698.
https://doi.org/10.1111/jon.12658

Straub, S., Schneider, T.M., Emmerich, J., Freitag, M.T., Ziener, C.H., Schlemmer, H.P., Ladd,
M.E., Laun, F.B., 2017. Suitable reference tissues for quantitative susceptibility mapping of the
brain. Magn. Reson. Med. 78, 204-214. https://doi.org/10.1002/mrm.26369

Sullivan, E.V., Pfefferbaum, A., Adalsteinsson, E., Swan, G.E., Carmelli, D., 2002. Differential
rates of regional brain change in callosal and ventricular size: a 4-year longitudinal MRI study
of elderly men. Cereb. Cortex 12(4), 438-45. https://doi.org/10.1093/cercor/12.4.438

Tie-Qiang, T., Gelderen, P., Merkle, H., Talagala, L., Koretsky, A.P., Duyn, J., 2006. Extensive
heterogeneity in white matter intensity in high-resolution T2*-weighted MRI of the human
brain at 7.0 T. NeuroImage 32, 1032-1040. https://doi.org/10.1016/j.neuroimage.2006.05.053

Vegh, V., O'Brien, K., Barth, M. and Reutens, D.C., 2016. Selective channel combination of MRI
signal phase. Magn. Reson. Med. 76(5), 1469-1477. https://doi.org/10.1002/mrm.26057

Walsh, D.O., Gmitro, A.F., Marcellin, M.W., 2000. Adaptive reconstruction of phased array MR
imagery. Magn. Reson. Med. 43(5), 682-690. https://doi.org/10.1002/(SICI)1522-
2594(200005)43:5<682::AID-MRM10>3.0.CO;2-G

Wang, Y., Liu, T., 2015. Quantitative susceptibility mapping (QSM): decoding MRI data for a
tissue magnetic biomarker. Magn. Reson. Med. 7, 82-101. https://doi.org/10.1002/mrm.25358

Weir, J. P., 2005. Quantifying test-retest reliability using the intraclass correlation coefficient
and the SEM. J. Strength Cond. Res. 19(1), 231-240.

Wharton, S., Bowtell, R., 2010. Whole-brain susceptibility mapping at high field: a comparison
of multiple-and single-orientation methods. Neuroimage 53(2), 515-525.
https://doi.org/10.1016/j.neuroimage.2010.06.070

Yacoub, E., Shmuel, A., Pfeuffer, J., Van de Moortele, P.F., Adriany, G., Andersen, P., Vaughan,
J.T., Merkle, H., Ugurbil, K., Hu, X., 2001. Imaging brain function in humans at 7 Tesla. Magn.
Reson. Med. 45, 588-594. https://doi.org/10.1002/mrm.1080

Yang, X., Sammet, S., Schmalbrock, P., Knopp, M. V., 2010. Postprocessing correction for
distortions in T2* decay caused by quadratic cross-slice b0 inhomogeneity. Magn. Reson. Med.
63(1): 1258-1268. https://doi.org/10.1002/mrm.22316

Yao, B., Li, T., van Gelderen, P., Shmueli, K., De Zwart, J.A., Duyn, J.H., 2009. Neuro image
susceptibility contrast in high field MRI of human brain as a function of tissue iron content.
Neuroimage 44(4), 1259-66. https://doi.org/10.1016/j.neuroimage.2008.10.029

Yao, B., van Gelderen, P., de Zwart, J.A., Duyn, J.H., 2007. Brain iron in MR imaging: R2* and
phase shift at different field strengths. Proc. Intr. Soc. Mag. Reson. Med. 15, 2165.

Zhou, D., Liu, T., Spincemaille, P., Wang, Y., 2014. Background field removal by solving the
Laplacian boundary value problem. NMR Biomed. 27(3), 312-319.
https://doi.org/10.1002/nbm.3064