Two-photon fluorescence imaging of live neurons using a multimode optical fiber

Raphaël Turcotte[1,2], Carla C. Schmidt[1], Martin J. Booth[2]*, and Nigel J. Emptage[1]*.

1. Department of Pharmacology, University of Oxford, Mansfield Road, Oxford OX1 3QT, United Kingdom

2. Department of Engineering Science, University of Oxford, Parks Road, Oxford OX1 3PJ, United Kingdom

***Corresponding authors**: nigel.emptage@pharm.ox.ac.uk and martin.booth@eng.ox.ac.uk

Abstract

An ability to optically visualize neurons and their subcellular components across brain regions is integral to our understanding of neuronal function[1]. Multimode optical fibers (MMFs), combined with wavefront control methods, have achieved minimally-invasive *in vivo* imaging of neurons in deep-brain regions with diffraction-limited spatial resolution[2-5]. To date MMF systems have been based on linear fluorescence and are therefore only able to perform two-dimensional imaging due to an absence of confocal filtering[6]. Here we report a method to achieve optical sectioning in live brain tissue by taking advantage of the nonlinearity of two-photon excited fluorescence (TPEF). In addition, we provide a new approach for the determination of transmission matrix for TPEF, a necessary step for the light shaping requirement of MMF imaging. Central to these advances is the use of a laser source able to generate both continuous wave (CW) light and femtosecond pulses. This configuration allows the system to be compact with minimal distal optics, enhancing suitability for neuroscience applications. With our system we demonstrate TPEF imaging of neurons and their dendrites in live brain slices through a single 60 μm-core MMF with a numerical aperture of 0.37. Taken together these developments represent an important step towards three-dimensional (3D) deep-brain imaging *in vivo*.

Main text

Step-index MMFs have emerged as a central component of a minimally-invasive micro-endoscope technology owing to their small size relative to other devices[2]. Indeed, their diameter, between 60-150 µm, enables direct insertion several centimeters into the living mammalian brain while causing minimal tissue damage and, hence, is likely to be of minimal consequence to the neuronal network[2-5]. An essential component of MMF technology is wavefront control, required to circumvent the random phase delays and mode coupling that occurs inside a fiber[7,8]. By shaping the wavefront entering a MMF using a spatial light modulator (SLM), a diffraction-limited focused point of light can be formed at the distal end of the fiber, suitable for the excitation of fluorophores in a sample. Digital point-scanning can be realized by updating the SLM pattern and thus translating the focal volume. Although the axial location of the focal plane can be precisely controlled, imaging is not 3D when illuminating with one-photon excitation; out-of-focus background excitation is present, restricting the use of the technology to sparsely labelled samples[2]. Confocality has been demonstrated using reflectance contrast using a digital pinhole[6], however, the broad emission spectra of fluorophores mean it is challenging to translate this approach to fluorescence imaging. The importance of multiphoton processes for intrinsic 3D imaging is well recognized and femtosecond laser pulses have been spatiotemporally focused[9] through MMFs for TPEF[10] and coherent anti-Stokes Raman scattering[11] imaging as well as for microfabrication[12]. Unfortunately, these optical systems require substantial distal optics and have not been shown to be compatible with biological imaging. We sought to address this limitation and have developed a MMF system capable of multiphoton imaging in living tissue. The device is optimized for translation into neuroscience research, although we anticipate that it will also be suitable for other biological applications.

The central concept for the work is the use of a dual mode laser source, able to switch between CW and pulsed operation (Mira Optima 900-F, Coherent, 830 nm). The light source was integrated in an optical system that has been optimized for live animal imaging through compactness[13], stability, and mobility[2]. Indeed, the modular design of the system is particularly well suited for *in vivo*

applications (Fig. 1); a detailed description is provided in the Supplementary Methods. Light from the source module was directed toward an imaging module using free-space optics and to a calibration module using a single mode optical fiber (SMF). The fixed imaging module comprised a liquid-crystal SLM (LC-SLM) for modulating the wavefront, a pinhole for filtering undesired diffraction, and optics to relay and couple the light from the LC-SLM into the MMF[2,13-15]. The movable calibration module was used in the evaluation of the transmission matrix, which describes the complex field propagation through the MMF, by imaging a selected plane some distance away from the MMF distal facet while it interfered with the SMF reference beam[16]. A transmission matrix, evaluated with the laser in CW mode, was used to calculate the patterns required on the LC-SLM for shaping the wavefront such that the foci for digital point-scanning could be generated in the biological sample at the plane of calibration[3]. Importantly, this configuration enabled us to evaluate the transmission matrix with minimal distal optics, eliminating the need to position a focused beam with free-space optics at the distal end of the fiber, and also to easily remove the distal optics in order to position the specimen for imaging. These changes afford a substantial improvement in the technology for neuroscience and life science applications.

The performance of our system was first assessed by generating foci at the MMF distal end for multiphoton fluorescence excitation. For this purpose, illumination foci were sequentially captured using the calibration module for both laser operating modes (Fig. 2). With CW light, a near diffraction-limited light distribution was achieved in the center of the fiber (full width at half maximum (FWHM): 2.6 ± 0.1 µm; $1/e^2$ width: 4.3 ± 0.2 µm). A maximum intensity projection of illumination foci recorded at different lateral locations revealed that the performance was maintained across the field-of-view (Fig. 2a), that is, uniform illumination was achieved. With the calibration performed in CW mode, it was then necessary to switch to pulsed mode to generate multiphoton fluorescence. The change in the laser temporal mode had minimal impact on the illumination foci despite the high sensitivity of the transmission matrix to perturbations (Fig. 2b)[17]. With the laser in pulse mode the width of the illumination foci increased (FWHM: 2.9 ± 0.1 µm) and tapered off more

slowly at the edges than for the CW mode (Fig. 2e; $1/e^2$ width: 5.7 ± 0.6 μm). Some deterioration in

the focus quality was expected because calibration is wavelength dependent and spectral broadening

would occur on transitioning from CW to pulsed mode (Fig. 2d; FWHM CW: 1.9 ± 0.3 nm; FWHM

pulse: 9.8 ± 0.3 nm). However, it is important to recognize that as the intensity of TPEF is dependent

on the square of the photon distribution, then the square of the illumination foci (Fig. 2b) should be

evaluated to characterize the TPEF excitation profile (Fig. 2c). When assessed in this way, a similar

point quality is achieved in pulsed mode as for the CW mode, with a FWHM of 2.5 ± 0.1 μm with

almost entirely suppressed extended edges (Fig. 2e; $1/e^2$ width: 4.0 ± 0.1 μm). Thus, we have

successfully implemented spatial focusing of femtosecond laser pulses through the MMF.

Having established a procedure for generating the desired light distribution in the sample, we

tested the TPEF imaging capability of the MMF system using fluorescent beads. Figure 3a shows an

image of 2-μm red fluorescent beads acquired at a distance of 40 μm from the distal facet of a MMF

25 mm in length with a core diameter of 60 μm and a numerical aperture of 0.37. An excellent signal-

to-noise ratio was achieved (SNR = 6.0 ± 0.6) with a total incident power at the sample of ~10 mW.

The fluorescence background from the focal plane has been a limiting factor for attaining high

dynamic range imaging. While the enhancement in peak intensity at the focus can easily reach values

between 100 and 350 over the average background illumination[3,4], the background fluorescence from

bright objects in the focal plane but outside the desired focus can easily overwhelm the signal from a

dim object at the focus, especially considering that the fluorescence background from all pixels is

integrated. Further suppression of the in-focus background from the nonlinearity of TPEF would

therefore be beneficial for increasing the dynamic range of the imaging system (Supplementary Fig.

S1).

As our design is optimized for live-tissue imaging we assessed its performance by imaging

Alexa Fluor 594 labelled neurons in live brain slices from the rat hippocampus. Fluorescent structures

were readily located and could be identified as neuronal cell soma (Fig. 3b). Pyramidal cell apical

dendrites were also clearly visible projecting from the soma. In an attempt to visualize smaller

features, we followed an apical dendrite away from the cell body and were able to detect secondary dendritic branching from the primary apical dendrite (Fig. 3c and Supplementary Fig. S1). The ability to visualize healthy cells within living tissue represents an important step for multiphoton MMF applications as previous demonstrations have all been confined to artificial samples[10,11]. Two particular challenges are faced when imaging live-tissue imaging: the first, achieving an adequate signal from the living structure, and the second, minimising phototoxic damage. Our successful imaging illustrates that these issues can be surmounted. Beyond the 3D imaging capability of the multiphoton approach, the intrinsic optical sectioning of nonlinear optical processes also reduces the out-of-focus fluorescence background, previously a considerable impediment for routine *in vivo* MMF imaging with linear fluorescence.

In summary, we have achieved multiphoton imaging of neurons and their dendrites in living brain tissue using wavefront control through a MMF. This is the first demonstration of this type and was accomplished by integrating a dual CW/pulsed laser source into a micro-endoscopic system that enabled spatial focusing of femtosecond pulses with minimal distal optics. We expect that this method will enable 3D imaging of densely labelled mammalian brains *in vivo* once implemented at optimized wavelengths and not just 830 nm. In addition, with enhancement in speed and computation, it is anticipated that it will be possible to improve focusing by maximizing point-by-point the nonlinear fluorescent signal obtained with the current calibration procedure. Increased computing capacity will also support MMFs with higher numerical aperture, e.g. from 0.37 to 0.66, offering control over a larger number of fiber modes, and thus substantial increase signal generation and collection for TPEF. Finally, we observed that the non-linearity of the multiphoton process exacerbated non-uniformities in the illumination. However, these non-uniformities are independent of the calibration method as they arise intrinsically from the fiber geometry and could be alleviated by performing spatially-variant deconvolution in order to make more regular the spatial response[18].

Acknowledgements

RT, CCS, MJB and NE acknowledge support from the Biotechnology and Biological Sciences Research Council (BBSRC) (BB/P0273OX/1). RT and MJB acknowledge support from the European Research Council (ERC) (H2020, 695140).

Conflict of Interest

The authors declare no conflict of interest.

References

1.   Ji, N. The practical and fundamental limits of optical imaging in mammalian brains. *Neuron* **83,** 1242–1245 (2014).

2.   Vasquez-Lopez, S. A. *et al.* Subcellular spatial resolution achieved for deep-brain imaging in vivo using a minimally invasive multimode fiber. *Light Sci. Appl.* **7,** 110 (2018).

3.   Turcotte, R., Schmidt, C. C., Emptage, N. J. & Booth, M. J. Focusing light in biological tissue through a multimode optical fiber: refractive index matching. *Opt. Lett.* **44,** 2386–2389 (2019).

4.   Ohayon, S., Caravaca-Aguirre, A., Piestun, R. & DiCarlo, J. J. Minimally invasive multimode optical fiber microendoscope for deep brain fluorescence imaging. *Biomed. Opt. Express* **9,** 1492–1509 (2018).

5.   Turtaev, S. *et al.* High-fidelity multimode fibre-based endoscopy for deep brain in vivo imaging. *Light Sci. Appl.* **7,** 92 (2018).

6.   Loterie, D. *et al.* Digital confocal microscopy through a multimode fiber. *Opt. Express* **23,** 23845 (2015).

7.   Dholakia, K. & Čižmár, T. Shaping the future of manipulation. *Nat. Photonics* **5,** 335–342 (2011).

8. Di Leonardo, R. & Bianchi, S. Hologram transmission through multi-mode optical fibers. *Opt.*

Express **19,** 247–254 (2011).

9. Morales-Delgado, E. E., Farahi, S., Papadopoulos, I. N., Psaltis, D. & Moser, C. Delivery of

focused short pulses through a multimode fiber. *Opt. Express* **23,** 9109–9120 (2015).

10. Morales-Delgado, E. E., Psaltis, D. & Moser, C. Two-photon imaging through a multimode

fiber. *Opt. Express* **23,** 32158–32170 (2015).

11. Trägårdh, J. *et al.* Label-free CARS microscopy through a multimode fiber endoscope. *Opt.*

Express **27,** 30055–30066 (2019).

12. Morales-Delgado, E. E. *et al.* Three-dimensional microfabrication through a multimode optical

fiber. *Opt. Express* **25,** 7031 (2017).

13. Plöschner, M. & Čižmár, T. Compact multimode fiber beam-shaping system based on GPU

accelerated digital holography. *Opt. Lett.* **40,** 197–200 (2015).

14. Neil, M. A., Booth, M. J. & Wilson, T. Dynamic wave-front generation for the characterization

and testing of optical systems. *Opt. Lett.* **23,** 1849–1851 (1998).

15. Čižmár, T. & Dholakia, K. Shaping the light transmission through a multimode optical fibre:

complex transformation analysis and applications in biophotonics. *Opt. Express* **19,** 18871–

18884 (2011).

16. Popoff, S. M. *et al.* Measuring the transmission matrix in optics: an approach to the study and

control of light propagation in disordered media. *Phys. Rev. Lett.* **104,** 100601 (2010).

17. Loterie, D., Psaltis, D. & Moser, C. Bend translation in multimode fiber imaging. *Opt. Express*

25, 6263–6273 (2017).

18. Lauer, T. R. Deconvolution with a spatially-variant PSF. in *Astronomical Data Analysis II.*

*Proc. of SPIE***,** 167–173 (2002).

Figure 1. A dual mode laser enables multiphoton imaging with a MMF system optimized for minimally invasive deep-brain micro-endoscopy. Schematic of the experimental system. (**a**) The source module contains a laser able to generate CW and pulse outputs. (**b**) The imaging module includes an SLM for wavefront control during imaging and calibration, as well as a PMT for detection of the epi-propagating signal. (**c**) The calibration module is used for recording illumination focus data from the imaging module and, with the reference beam from the source module, for evaluating the transmission matrix. *CW*: continuous wave; *PBS*: polarizing beam splitter; *SMF*: single-mode fiber; *SLM*: spatial light modulator; *DM*: dichroic mirror; *MMF*: multimode fiber; *PMT*: photomultiplier tube; *OL*: objective lens; *NPBS*: non-polarizing beam splitter; *CCD*: charge-couple device camera.

Figure 2. TPEF nonlinearity facilitates wavefront control for focusing light through a MMF. (**a,b**) Illumination foci achieved with wavefront control through a MMF and recorded sequentially at different positions (maximum intensity projection) when the laser is operating in (**a**) CW mode and (**b**) pulsed mode. (**c**) Squared of the intensity distribution shown in (b) to illustrate the TPEF spots. Scale bar: 15 μm; inset width: 7.6 μm. (**d**) Spectra of the laser source in CW and pulsed mode. (**e**) Intensity profile for the inset foci shown in a-c.

Figure 3. MMF multiphoton imaging of live brain tissue reveals neuronal bodies and dendrites. (**a**) Multiphoton image of red fluorescent beads (2 μm in diameter) acquired through a MMF. (**b,c**) Multiphoton images in live brain tissue from the rat hippocampus acquired through a MMF showed clearly identifiable (**b**) neuronal cell bodies and apical dendrites as well as (**c**) smaller dendrites branching off from an apical dendrite, after the labelling of 2 to 3 neurons with AlexaFluor594. Scale bar: 15 μm.

(a)

(b)

(c)

Laser CW/Pulse

PBS

PMT

DM

Iris

SLM

MMF

SMF

OL

NPBS

CCD

Figure 1

(a) Continuous wave

(b) Pulsed

(c) Pulsed squared

(d) CW / Pulsed — Intensity [1] vs Wavelength [nm]

(e) CW / Pulsed / Pulsed² — Intensity [1] vs Position [µm]

Figure 2

Figure 3