Field-deployable molecular diagnostic platform for arbovirus and *Wolbachia* detection in

Aedes aegypti

Natalie N. Rutkowski, Yuemei Dong, and George Dimopoulos*

W. Harry Feinstone Department of Molecular Microbiology and Immunology, Bloomberg

School of Public Health, Johns Hopkins University, Baltimore, Maryland, United States of

America

*Correspondence should be addressed to GD
E-mail: gdimopo1@jhu.edu

Abstract

Background: Surveillance of mosquito infection status is critical for planning and deployment

of proper mosquito control initiatives. Concurrently, *Wolbachia* is being widely used as a control

method for arboviral transmission. Point-of-care (POC) detection assays are necessary for

monitoring the infection prevalence and geographic range of viruses as well as *Wolbachia* in

mosquito vector populations. We therefore assessed the novel qPCR bCUBE molecular

diagnostic system as a tool for virus and *Wolbachia* detection.

Results: We developed a reliable, specific, and sensitive diagnostic assay for detecting Zika

virus and dengue virus serotype 2 using the real-time qPCR platform bCUBE. With bCUBE-

based qRT-PCR, both *Wolbachia* bacterium and virus RNA could be reliably detected in

individually infected *Ae. aegypti* mosquitoes and in pools of 5, 10, or 15 mosquitoes.

Conclusions: The portable qPCR bCUBE diagnostic platform is capable of detecting Zika and

dengue virus as well as *Wolbachia* in mosquitoes and therefore has potential as a practical field-

deployable diagnostic test for vector-borne disease surveillance programs.

48 **Keywords:** *Aedes aegypti*, Zika, dengue, diagnostics, qPCR

Introduction

Arthropod-borne diseases threaten over two-thirds of the global population and are exhibiting an ongoing expansion of their geographic range and prevalence as a result of climate change, urbanization, and globalization (1-4). According to the World Health Organization, dengue virus infection has increased 30-fold over the past 50 years and now affects nearly 100 million people, primarily in the Americas and Asia (5, 6). Recent outbreaks of Zika virus have also led to widespread concern because of the virus's ability to cause newborn malformations (7, 8).

Dengue and Zika viruses are members of the genus *Flavivirus* that are primarily transmitted by the *Aedes* mosquito (7, 9, 10). Other members of the *Flavivirus* genus are also recognized as vector-borne pathogens of public health significance, including West Nile virus (WNV), yellow fever virus (YFV), Japanese encephalitis virus (JEV), and Chikungunya virus (CHIKV). These viruses result in similar flu-like symptoms with the potential to progress to neuroinvasive outcomes. *Aedes* mosquitoes, with their aggressive blood-feeding behavior, have allowed for efficient human-mosquito transmission of these arboviruses (3, 11). The geographical presence of this mosquito vector has dramatically increased in the last few decades, leading to an expanded transmission of these arboviruses (12-14). The lack of vaccines and treatment against these arboviruses highlights the importance of mosquito control and surveillance strategies (15). Current and future mosquito-targeted control strategies are, and will be, having a significant epidemiological impact but also require mosquito and pathogen surveillance (15). Surveillance of geographical distribution of the vector mosquitoes and the pathogens they carry is an essential component of disease prevention and control.

Wolbachia pipientis is increasingly being used as a method for limiting arboviral transmission in

dengue and Zika virus-endemic countries. This endosymbiotic bacterium is passed from mother

to offspring and has been shown to suppress dengue and Zika virus transmission in *Aedes*

mosquitoes (16, 17). Field-release trials require continual monitoring of *Wolbachia*-infected *Ae.*

aegypti in mosquito populations (18-20). Real-time qPCR has been used as the primary method

for detecting *Wolbachia* genetic material in mosquitoes (21).

Accurate, rapid, and cost-effective mosquito and pathogen surveillance is critical for monitoring

infection prevalence and thereby mitigating transmission risk. Methods currently used for

arboviral detection include viral culture, antibody detection, antigen detection, and RNA

detection using quantitative real-time RT-PCR (qPCR). Currently, the qPCR method is the gold

standard because of its use of specific molecular markers (22, 23). Many of these assays are

laborious and laboratory-based and require expensive and bulky instruments, making them

incompatible with low-resource regions (24). Recently, several novel, advanced point-of-care

(POC) diagnostic measures have been developed for detecting mosquito-borne viruses in the

field, including honey-baited nucleic acid preservation cards (25, 26), loop-mediated isothermal

amplification (LAMP) (27), biosensors (28), and adaptations of near-infrared spectrometry

techniques (29). However, these techniques have documented limitations, including cross-

reactivity with other flaviviruses or requirements for training on specialized equipment, making

the adoption of these new diagnostic tools difficult (30, 31). Although the global burden of

emerging outbreaks of Zika and dengue is clearly recognized, there is a gap in resources for

endemic countries and regions, which are consistently plagued by a lack of equipment and

adequate resources to consistently monitor the prevalence and range of infected mosquitoes (5,

32, 33).

Disease surveillance and integrated vector control are essential for curbing disease transmission.

Nucleic acid-based testing to detect viral RNA allows for specific and sensitive virus monitoring

in mosquito surveillance programs (24). Several conventional and more recent real-time PCR-

based assays have been established for mosquitoes and their vectored pathogens, and the ready

availability of genome sequences for both vectors and pathogens can support the identification of

additional PCR-compatible molecular markers (34-36).

In the present study, we evaluated a novel and portable real-time PCR platform, bCUBE (Hyris,

Ltd), as a PCR-based arboviral detection method with potential for field-deployability. bCUBE

makes possible the genetic testing of biological samples in any setting, at any time, with real-

time access to results on its dedicated cloud-based software platform. This technology is a

portable device, similar in size to a Bluetooth speaker, which is used for biological analysis in

several fields, including agricultural pest control. The device can be operated from a laptop,

tablet, or cellular device through an easy-to-use gateway and generates centralized data analysis

immediately after a reaction. The device is capable of performing thermocycling reactions such

as real-time PCR, as well as loop-mediated isothermal amplification (LAMP). The system can be

calibrated to distinguish between positive and negative samples in a single reaction with

predetermined conditions that have been established ahead of time in the laboratory. This

feature allows the bCUBE to be operated by individuals lacking in depth training in qPCR assays

and data analysis skills. We have now explored the use of bCUBE technology for detection of

both dengue and Zika virus in *Aedes aegypti,* optimizing and standardizing the sample

preparation method to be used with a commercially available one-step qRT-PCR assay kit.

Finally, we have developed a bCUBE-compatible qPCR diagnostic assay for the surveillance of

arboviral pathogens and *Wolbachia* in *Aedes* mosquitoes.

Methods

Ethical Statement. This study was carried out in strict accordance with the recommendations in

the Guide for the Care and Use of Laboratory Animals of the National Institutes of Health (NIH).

Mice were used according to an animal protocol (permit # MO15H144) approved by the Animal

Care and Use Committee of the Johns Hopkins University and were used as the blood source for

the maintenance of the mosquito colonies. Commercial anonymous human blood (InterState

Blood Bank) was used for Zika and dengue virus infection assays in mosquitoes, and informed

consent was therefore not applicable.

Mosquito rearing and mosquito infections. *Ae. aegypti* Liverpool strain LVPB12 mosquitoes

were maintained on 10% sucrose solution under standard insectary conditions at 27±0.5°C and

75-80% humidity with a 12-h light/dark cycle. Mosquitoes were reared using a standard rearing

protocol, and colonies were maintained on Swiss Webster mice (Charles River Laboratories).

DENV2 and ZIKV infection of mosquitoes was carried out as previously described (37). Virus

supplemented blood was fed to mosquitoes using an artificial membrane glass feeder, and

infected mosquitoes were double-caged and incubated in a reach-in incubator under conditions

similar to those in the standard insectary chamber described above. *Wolbachia w*AlbB-infected

Ae. aegypti (WB1) mosquitoes were reared and maintained under the standard insectary

conditions described above and in (38).

Cell culture and virus propagation. *Aedes albopictus* C6/36 cells (ATCC CRL-1660) were

cultured, and viral stocks were prepared as previously described in (39). In brief, C6/36 cells

were cultured in MEM medium (Gibco) supplemented with 10% heat-inactivated fetal bovine

serum (FBS), 1% penicillin-streptomycin, and 1% non-essential amino acids and maintained in a

tissue culture incubator at 32°C and 5% CO_2. Baby hamster kidney strain 21 (BHK-21, ATCC

CCL-10) cells were maintained at 37°C and 5% CO_2 in the DMEM medium (Gibco)

supplemented with 10% fetal bovine serum (FBS), 1% penicillin-streptomycin, and 5µg/ml

Plasmocin^{TM}. DENV serotype 2 New Guinea C strain (DENV2) and ZIKV strain FSS 13025

(ZIKV) were used in the indicated experiments. For viral stock preparation, C6/36 cells grown to

80% confluence were infected with ZIKV and DENV2 at a multiplicity of infection (MOI) of 10

and incubated at 32°C and 5% CO_2 for 5 days (DENV2) and 6 days (ZIKV). Virus was

harvested by three freeze-thaw cycles using dry ice and a water bath (37°C), then centrifuged at

2,000 rpm for 10 min at 4°C. The supernatant from this cell lysis was mixed with the original

cell culture supernatant to yield the final viral stock. Viral stocks were aliquotted and stored at -

80°C for long-term storage. Viral stock titration was done by plaque assay as described below.

Primer design for real-time quantitative RT-PCR (qPCR). The ZIKV envelope (E) protein

was chosen as the target for ZIKV detection, and primers were developed based on previously

established sequences (40). Dengue virus serotype 2 (DENV2) qPCR was performed using

previously developed protocols that employ primers detecting serotype 2 in the 3'UTR region

(41) (Table 1). The primers were modified and optimized for bCUBE qPCR (Table 1). Primer

sequences were assessed with NCBI Blast to search for potential cross-reactivity against other

viruses, bacteria, and vectors. Previously established primers for *w*AlbB detection were used for

Wolbachia-infected *Ae. aegypti* (42). The sequences of the primers are summarized in Table 1.

Primer	Sequence 5'-3'	Nucleotide position	Amplicon size (bp)	Accession Number	Reference
ZIKV-forward	AGCAACATGGCGGAGGTAAG	1,128-1,147	145	FSS13025	(34)*
ZIKV-reverse	CTGTCCACTAACGTTCTTTTGCAGA	1,249-1,273			
DENV-2-forward	TCCCTTCCAAATCGCAGCAACAATG	10,517-10,541	168	NC_001474.2	(35)*
DENV-2-reverse	CGTTCTGTGCCTGGAATGATG	10,665-10,685			
wAlbB-forward	CCTTACCTCCTGCACAACAA	213,522-213,541	110	CP031221.1	(36)
wAlbB-reverse	GGATTGTCCAGTGGCCTTA	213,394-213,412			
WNV-forward	TTGTGTTGGCTCTCTTGGCGTTCTT	233-257	408	AF196835	(38)
WNV-reverse	CAGCCGACAGCACTGGACATTCATA	640-616			
JEV-forward	GGCAGAAAGCAAAACAAAAGA	390-410	367	AF080251	(39)
JEV-reverse	CGGATCTCCTGCTTCGCTTGG	736-756			
YFV-forward	CACGGCATGGTTCCTTCCA	5,656-5,674	71	MN708497	(37)
YFV-reverse	ACTCTTTCCAGCCTTACGCAAA	5,707-5,728			
CHIKV-forward	TACAGGGCTCATACCGCATC	10,357-10,376	154	NC_004162	(40)
CHIKV-reverse	AAAGGTGTCCAGGCTGAAGA	10,492-10,511			

Table 1. **Primer pairs for viral detection and cross-reactivity panel for real-time qPCR**
using the bCUBE.
*ZIKV and DENV-2 primers were modified and optimized for qRT-PCR. The cross-reactivity
panel primer sequences were included to confirm viral RNA obtained from BEI Resources. Four
viruses were included in this study: West Nile virus (WNV), Japanese encephalitis virus (JEV),
yellow fever virus (YFV), and Chikungunya virus (CHIKV). Primer sequence, nucleotide
position, and amplicon size are listed.
174

Cross-reactivity panel of mosquito-borne viruses. Viral nucleic acid sequences included in the

cross-reactivity panel were CHIKV (H20235 ST MARTIN 2013), JEV (India R53567), WNV

(CO 1862), and YFV (17D), obtained from BEI Resources. A NanoDrop Spectrophotometer

(ThermoFisher Scientific) was used to measure the RNA concentration of these viral samples.

Previously established primers were used, as listed in Table 1, for qPCR to confirm the presence

of corresponding viral RNA in the samples (43-46).

Total RNA preparation for DENV2- and ZIKV-infected whole mosquitoes or tissues. Total

RNA extraction was performed using a squash buffer (10 mM Tris base, 1 mM EDTA, 50 mM

NaCl) supplemented with 1:8 part Proteinase K (Qiagen, 15mg/ml). Mosquito abdomens with

the midguts and heads with thoraces were individually collected in 50 µl of squash buffer at 7

and 14 days, respectively, after an infectious blood meal and stored at -80°C until extracted.

Proteinase K was added at 1/8 volume to each sample to give a final concentration of 15 mg/ml

and homogenized with a cordless Pellet Pestle Motor (Kontes) for 40-60 s. Samples were

incubated at 57°C for 5 min, followed by 95°C for 5 min for enzyme deactivation. The

supernatant from this crude RNA extraction was used immediately for qPCR or stored at -80°C

until use.

gDNA preparation of *Wolbachia*-infected mosquitoes. Total gDNA from *Wolbachia*-infected

total mosquitoes or tissues was prepared as described above (47). These crude gDNA extractions

were immediately used for qPCR or stored at -80°C until use.

Laboratory standard real-time PCR with Applied Biosystems equipment. A StepOnePlus

Real-Time PCR system (Applied Biosystems) was used for comparative analysis. The mosquito

housekeeping gene ribosomal protein S17 (*rps17)* was amplified in *Aedes aegypti* using the two

different PCR machines (Applied Biosystems and bCUBE) with the same real-time PCR

settings. SYBR Green PCR Master Mix (Applied Biosystems) was used according to

manufacturer's instructions. Mosquitoes were collected individually in 50 µl of squash buffer,

and crude gDNA was extracted. Identical samples were applied to both the bCUBE- and

StepOnePlus-based qPCR systems for comparative evaluation.

Hyris bCUBE real-time RT-PCR. The portable qPCR machine Hyris bCUBE 2.0 thermocycler

(Hyris, London, UK) and the 16-well cartridges were kindly provided by Hyris Inc. The Hyris

data analysis platform was used for this study. GoTaq 1-Step RT-qPCR kit (Promega) was used

at a final volume of 20 µl with: 0.4 µl of kit reverse transcriptase enzyme, 10 µl of RT-PCR

buffer, 0.3 µl of 10 µM of forward and reverse primer, 9 µl of DNA/RNA free water, and 1 µl of

crude RNA or gDNA sample. Each sample was performed in technical duplicate. Each cartridge

run included one negative and one positive control. The following thermocycling settings were

used: reverse transcription for 15 min at 37°C, heat-inactivation of reverse transcriptase at 95°C

for 10 min, followed by 30 PCR cycles at 95°C for 10 s, 60°C for 30 s, and 72°C for 30 s.

Melting curve analysis was performed at the end by cooling to 60°C, followed by heating to

95°C at 0.05°C/s. Automatic data analysis was generated with the Hyris data analysis platform.

Two technical replicates were performed on the platform for each biological replicate. Three

biological replicates were included for each assay.

Absolute quantification of viral copy numbers through qRT-PCR. *In vitro*-transcribed RNA

of the E gene of ZIKV and the 3'-UTR region of DENV2 were used for absolute quantification

and cloned into a TOPO-TA vector. Total RNA was extracted using TRIzol reagent

(ThermoFisher Scientific). RNA (2 µg) was used for cDNA synthesis with M-MLV Reverse

Transcriptase (Promega). Conventional PCR was used with the primer (Table 1) to amplify a

DNA fragment of 168-bp for DENV2 and 145-bp for ZIKV. PCR was followed by PCR

purification using a QIAquick PCR Purification kit and protocol (Qiagen). About 50 ng of

cleaned PCR product of either DENV2 or ZIKV was separately cloned into a TOPO-TA PCR

cloning vector (ThermoFisher Scientific) according to the manufacturer's instructions. The

ligation product was transformed with TOP10-competent cells (ThermoFisher Scientific), and

positive clones grown on LB selection agar plates were screened through colony-PCR, followed

by plasmid mini-prep (Qiagen) and sequencing confirmation. The plasmids of the final

confirmed-positive clones were purified using the Maxiprep kit (Qiagen) according to the

manufacturer's instructions. Plasmid DNA was used to prepare absolute standards. A HiScribe

T7 High Yield RNA Synthesis Kit (NEB) was used for RNA synthesis, and the concentration of

the RNA was checked with a NanoDrop Spectrophotometer (ThermoFisher Scientific).

Molecular weights were converted to copy numbers using the New England BioLabs Calculator

(https://nebiocalculator.neb.com/#!/ssrnaamt). The concentration of RNA was adjusted to 10^{10}

copies/µl and serially diluted 10 times for the standard curve on the bCUBE. The standard curve

was generated using GoTaq 1-Step RT-qPCR (Promega).

Viral titration by plaque assay. The titers of ZIKV and DENV2 in the original viral stocks and

in the infected mosquitoes were determined by plaque assays in BHK-21 cells. Each experiment

was performed with at least three biological replicates. Whole mosquito or mosquito tissue

samples were collected at 7 and 14 days post-infectious blood meal in 150 µl of complete

DMEM medium with glass beads. Tissue samples were homogenized with a Bullet Blender

(Next Advance) and serially diluted with DMEM complete medium. One or two days before

plaque assay, the BHK-21 cells were split at a 1:10 dilution and grown on 24-well plates to 80%

confluency. Serially diluted mosquito or viral samples (100 µl each) were added to the BHK-21

cells, followed by incubation at room temperature for 15 min on a rocking shaker (VWR) and

subsequent incubation at 37°C with 5% CO_2 in a cell incubator (ThermoScientific) for another

45 min. The 24-well plates with infected BHK-21 cells were overlaid with 1 ml of 0.8%

methylcellulose in complete DMEM medium with 2% FBS and incubated for 5 to 6 days in the

cell culture incubator (37°C and 5% CO_2). Plaques were fixed and developed with staining

reagent (1% crystal violet in 1:1 methanol/acetone solution) at room temperature for

approximately 30 min. Plates were rinsed with DI water and air-dried, and plaques were counted

and multiplied by the dilution factor to calculate the plaque forming units (PFUs) per sample.

DENV2 and ZIKV mosquito pools. Preliminary pooled sample experiments involved a total of

300 *Aedes aegypti*: 276 uninfected, 12 infected with DENV2 and 12 infected with ZIKV. First,

12 mosquitoes were infected with ZIKV and 12 with DENV2. Following confirmation of

infection by RT-qPCR of individual infected mosquitoes, each was placed into a pool of

uninfected mosquitoes. Four different pools were used to measure the sensitivity of the infection

detection for pools of 5, 10, 15, and 20. The squash buffer volumes used were 250, 450, 700, and

950 µl, respectively, followed by the addition of 1:8 part Proteinase K.

Statistical Analysis. Graphs were generated using GraphPad Prism Software version 8. The

estimation of several diagnostic parameters for leg samples, including sensitivity, specificity,

accuracy, and positive and negative predictive values was calculated using the web-based

software MedCalc Diagnostic Test Evaluation Calculator.

Results

The reliability of the portable bCUBE qPCR machine is equal to that of a laboratory-standard qPCR system

The reliability of the novel bCUBE real-time PCR system was validated by comparing the system to a standard qPCR instrument (Applied Biosystems StepOnePlus, ABI). Our proof-of-principle studies used mosquito tissue homogenates (abdomen with the midgut, and head with the thorax) in squash buffer that were tested with the mosquito housekeeping gene (ribosomal protein S17, *rps*17). *Aedes aegypti* were dissected into crude compartments, including the abdomen and head plus thorax. Identical samples were processed using both qPCR machines, identical settings, and identical SYBR Green Master Mix and instructions (Applied Biosystems). As shown in Fig. 1, the housekeeping *Rps*17 gene was detected in all 24 samples by both machines. The cycle threshold (Ct) values for the *Rps*17 gene in *Ae. aegypti* were detected significantly earlier in the bCUBE system than in the laboratory standard procedure, with a mean difference of 2.40 Ct value (paired t-test, p-value < 0.0001). This result suggests that the bCUBE qPCR is reliable in detecting a mosquito gene from a single mosquito tissue using a crude sample-preparation method.

Fig. 1. Comparison of *Ae. aegypti Rps*17 gene detection in abdomen (ABD) and head plus thorax (HT) samples using bCUBE and Applied Biosystems real-time PCR. *Ae. aegypti* abdomens and head plus thoraces were collected, and gDNA was extracted using the squash buffer method. The housekeeping gene *Rps*17 was amplified using the bCUBE and Applied Biosystems real-time PCR systems. Cycle threshold (Ct) values are plotted. Statistical significance was determined by Mann-Whitney test (****, $p < 0.0001$; Mann-Whitney).

bCUBE-based one-step qRT-PCR is sufficiently sensitive for DENV2 and ZIKV detection

We performed sensitivity studies on the bCUBE-based qPCR system based on absolute

quantification standard curve analysis. Cloned fragments of DENV2 and ZIKV were quantitated

for RNA copy number and serially diluted to establish a qPCR standard curve. End-point

limitations were set to < 30 for Zika virus, whereas dengue end-point limitations were set to < 29

based on standard curve (Supplementary Fig. 1). Correlation coefficient (R^2) values were 0.99

for DENV2 and 0.98 for ZIKV. Based on the standard curve analyses, the limit of detection

range was 10 viral RNA copy numbers for DENV2 and ZIKV. Amplification above these values

was attributed to non-specific amplification.

bCUBE-based qRT-PCR is specific for DENV2 and ZIKV detection

The specificity of the ZIKV and DENV2 assays was then determined using a cross-reactivity

panel. To address challenges involving primer cross-reactivity with other arboviruses (46, 48) as

well as the tendency toward false-positive amplification in negative samples, we tested the

specificity of the assay against a panel of *Flaviviruses* and one *Alphavirus*. The frequently co-

circulating arboviral RNA samples, obtained from BEI resources (Table 2) for this study, were:

CHIKV (H20235 ST MARTIN 2013), JEV (India R53567), WNV (CO 1862), and YFV (17D).

The ZIKV and DENV2 primer pairs showed a high degree of specificity for amplification of

their respective virus RNAs (Table 2), since no amplification of other virus RNAs occurred.

Family	Genus	Species	Strain	BEI No.	Result for ZIKV qPCR Assay	Result for DENV-2 qPCR Assay
Flaviviridae	*Flavivirus*	Zika virus			+	-
		Dengue virus serotype 2			-	+
		Japanese Encephalitis virus	India	NR-9592	-	-
		West Nile virus	CO 1862	NR-50434	-	-
		Yellow Fever virus	17D	NR-2869	-	-
Togaviridae	*Alphavirus*	Chikungunya virus	St Martin 2013	NR-50130	-	-

Table 2. Evaluation of the specificity of bCUBE-based qRT-PCR for DENV2 and ZIKV detection
A cross-reactivity panel of frequently co-circulating viruses was used to evaluate the specificity of bCUBE-based qRT-PCR for DENV2 and ZIKV. Arboviral RNA samples were obtained from BEI Resources. The presence of all cross-reactivity panel RNA samples was quantified on NanoDrop for molecular weight and amplified on qRT-PCR with their own primer sequences.

bCUBE-based qRT-PCR is as accurate as plaque assay in detecting infectious viral RNA

PCR detection of non-infectious viral RNA is possible when viral RNA is not actively

replicating (49). Therefore, the viral RNA copy number and viral infectious RNA load can differ

in mosquitoes. We thus sought to evaluate the infection prevalence using bCUBE qRT-PCR and

plaque assays as opposed to titer comparisons. Three biological replicates of *Ae. aegypti* were

infected with ZIKV and DENV-2. Each replicate of infected mosquitoes was collected and

separated into two halves to compare infectious prevalence as determined by plaque assay and

bCUBE qPCR analysis. In each of the three replicates, no-significant differences were seen in

infection prevalence values for both ZIKV and DENV2 between the two methods. These results

suggest that bCUBE qRT-PCR analysis can accurately detect infectious ZIKV AND DENV2

(Fig. 2).

Fig 2. ZIKV AND DENV2 infection prevalence in *Ae. aegypti* as detected in the bCUBE versus plaque assay. *Ae. aegypti* were infected with Zika (A) and dengue (B) virus via an artificial blood meal. Each biological replicate was split into two groups and analyzed using the bCUBE assay (black bars), and the other half was used for the plaque assay (grey bars). No significant difference was detected between the plaque assay and bCUBE qPCR in terms of infectious prevalence (ns: not significant; Fisher's exact test).

ZIKV and DENV-2 RNA can be detected and quantified in individual mosquito samples by bCUBE assay

Ae. aegypti mosquitoes were experimentally infected with ZIKV (n=112) and DENV2 (n=112) through an artificial blood feeder containing anonymous human blood supplemented with virus, and samples from the mosquitoes were evaluated for infection status with the bCUBE platform at 7 and 14 days post-blood meal. Samples with Ct values < 30 were considered positive for ZIKV and < 29 for DENV2 infection. Each sample was run in duplicate to confirm infection. Both ZIKV and DENV2 viral RNA were detected from an individually infected mosquito using the squash buffer extraction method and the bCUBE platform. Viral RNA could be detected and

351   quantitated in samples containing single abdomen with midguts and head with thoraces of

352   infected mosquitoes (Fig. 3).

Fig 3. Viral RNA concentrations of individual *Ae. aegypti* tissues collected at various time points.

Individual *Ae. aegypti* that were infected with Zika (A) and dengue virus (B) were collected at 7 or 14 days post-infectious blood meal. Abdomens with midgut samples (ABD) were collected on day 7 post-infectious blood meal, and head with thorax samples (HT) were collected on day 14. Corresponding leg (L) samples were also collected on day 14. Viral RNA extraction was performed using the squash method, followed by amplification using 1-step qPCR on the bCUBE. Uninfected mosquitoes were not included.

Next we investigated detection of virus in leg samples of *Ae. aegypti* (Fig. 3). Forty-six leg

samples were collected from mosquitoes that were also analyzed for infection in the heads and

thoraces to confirm infectious status. We found significant differences between the Ct values for

the head with thorax samples versus the leg samples. For the leg samples, six false negative

sample results were obtained for DENV2, and two false negative results were obtained for

ZIKV. One false positive result was detected for ZIKV (Fig. 4). These results indicate that

analysis of infection using leg samples is not reliable. No amplification occurred in uninfected

Aedes aegypti used as a negative control. Because of the false negative results, the DENV2

sensitivity of leg samples was 76% (95% CI, 54.87% to 90.64%) and the diagnostic specificity of

leg samples was 100% (95% CI, 83.89% to 100.00%). The sensitivity for ZIKV of leg samples

was 94.59% (95% CI, 81.81% to 99.34%), and the specificity was 88.89% (95% CI, 51.75% to

99.72%). The overall accuracy of the bCUBE qPCR platform for detecting leg samples was

93.48% for ZIKV (95% CI, 82.10 to 98.63%) and 86.96% for DENV2 (95% CI, 73.75% to

95.06%) (Fig. 4).

Statistic for ZIKV	Value	95% CI
Sensitivity	94.59%	81.81%-99.34%
Specificity	88.89%	51.75%-99.72%
Positive Likelihood Ratio	8.51	1.34-54.12
Negative Likelihood Ratio	0.06	0.02-0.24
Disease prevalence	80.43%	66.09%-90.64%
Positive Predictive Value	97.22%	84.63%-99.55%
Negative Predictive Value	80.00%	50.47%-94.01%
Accuracy	93.48%	82.10%-98.63%

Statistic for DENV2	Value	95% CI
Sensitivity	76.00%	54.87%-90.64%
Specificity	100.00%	83.89%-100.00%
Positive Likelihood Ratio	-	-
Negative Likelihood Ratio	0.24	0.12-0.48
Disease prevalence	54.35%	39.01%-69.10%
Positive Predictive Value	100.00%	-
Negative Predictiv Value	77.78%	63.53%-87.55%
Accuracy	86.96%	73.74%-95.06%

378 **Fig. 4.** Infected *Ae. aegypti* (n=46) were collected, and identical samples of head with thorax and
379 legs were amplified for Zika virus (A) and DENV2 (B) using bCUBE-based qRT-PCR. Infection
380 status was compared (****, p<0.0001; paired two-tailed t-test).

bCUBE-based qRT-PCR detects ZIKV and DENV-2 viral RNA in pooled *Ae. aegypti*

samples

Surveillance programs monitoring ZIKV and DENV2 transmission will typically assay the

presence of virus in pooled samples ranging from 5 to 50 mosquitoes (50). We pooled individual

ZIKV- or DENV2-infected mosquitoes with 5, 10, 15, or 20 uninfected *Aedes* in order to assess

the assay's sensitivity with pooled samples. The infection status of individual infected

mosquitoes was confirmed by viral RNA extraction using squash buffer followed by bCUBE-

based qRT-PCR. Following confirmation of infection, positive mosquitoes were placed into four

different pools for further viral detection. ZIKV could be detected in all pooled samples. DENV2

could be detected in pools of 5, 10, and 15 mosquitoes, but DENV2 viral RNA could not be

detected in one of the three biological replicates of the 20-pooled mosquitoes. We found

significant differences between the individual infected samples and pooled samples; however,

the detection in pooled samples was still within the range for a positive identification

(Supplementary Table 1).

bCUBE-based qRT-PCR detects *Wolbachia* in individual *Ae. aegypti*

*Wolbachia w*AlbB was amplified from individual mosquitoes with qRT-PCR using the bCUBE

platform (n=35) (Supplementary Table 2). Uninfected *Ae. aegypti* (n=10) were used as negative

controls. All *Wolbachia* -infected mosquitoes were positive for *Wolbachia* infection, and the

negative controls did not amplify. These data suggest that bCUBE qPCR can be used for

Wolbachia detection employing previously established primers.

<u>Discussion</u>

Arboviral pathogens represent an expanding threat across the world. Monitoring the vector and

its infectious cargo is essential for planning adequate mosquito control to reduce transmission.

Endemic countries frequently lack surveillance resources for monitoring pathogen and vector

distribution. Therefore, cost-effective field-deployable pathogen detection methods can greatly

facilitate surveillance and control efforts. We sought to develop a method for detecting viral

pathogens in their *Aedes aegypti* vector with the potential for field deployment. Currently,

pathogen detection in the field is limited by expensive lab-based equipment and reactions and the

necessity for highly trained personnel to process samples. Real-time qPCR has long been

considered the gold standard for viral detection; however, inadequate access to sample

preparation methods and the requirement for bulky equipment render this method incompatible

with fieldwork.

417
As an alternative to laboratory PCR, we have now explored the use of bCUBE, a novel qPCR

platform, with a simple nucleic acid extraction method for detection of arboviruses and

Wolbachia in mosquito vectors. Actions were taken to limit the cost needed for analysis. We

used DNA dye binding qPCR commercial kits as an alternative to hydrolysis probe-based qPCR

kits. Furthermore, we extracted DNA/RNA using the squash buffer methodology as opposed to

expensive extraction kits. It is important to note that current bCUBE cartridges are restricted to

16-wells thereby limiting the number of samples run during each experiment. Hyris bCUBE

software can be programmed to display results as positive or negative result as opposed to

interpreting cycle threshold and melting curve values. This minimizes training for remote

personnel.

We confirmed the reliability of the bCUBE by comparing its performance against laboratory

standards. Therefore, bCUBE is capable of performing real-time qPCR while overcoming the

barriers presented by the need for bulky equipment. We then employed a crude extraction

method using squash buffer to extract viral DENV2 and ZIKV RNA from mosquitoes for use in

the bCUBE system. To address challenges involving primer cross-reactivity with other

arboviruses (46, 48) as well as the tendency toward false-positive amplification in negative

samples, we tested the specificity of the assay against a panel of frequently co-circulating

arboviruses. The results confirmed the assay's specificity for DENV2 and ZIKV RNA, with the

potential to apply these primer pairs for field conditions. To negate any non-specific

amplification of the host's genetic material, we used a standard curve to determine the limit of

detection and subsequently, the cut off values for detecting ZIKV and DENV2. This allowed for

positive identification of viral RNA from infected mosquito samples while limiting the

amplification of false positive samples as a result of background noise. Following primer

optimization, we compared the bCUBE assay to plaque assays for determining infection

prevalence. Because qPCR is generally capable of detecting low RNA copy numbers in

individual samples, it is important to evaluate the assay's potential for detecting infectious viral

RNA rather than viral RNA that is not replicating (49, 51). No significant differences were seen

between the bCUBE and plaque assay for three biological replicates, suggesting that the bCUBE

assay is able to detect infectious viral RNA.

We evaluated the use of *Ae. aegypti* leg samples for viral detection. Some infected leg samples

turned out to be false negatives suggesting this sample type is unreliable for assaying virus.

However, mosquito surveillance agencies typically evaluate arboviral infection status in pooled

samples as opposed to assaying individual mosquitoes. ZIKV RNA was detected in pools of 5,

10, 15, and 20 with uninfected mosquitoes in three biological replicates. DENV RNA was not

detected in one of three biological replicates of 20-mosquito pools, suggesting possible

limitations for DENV detection. This highlights the bCUBE platform's potential for pool testing

in the field.

Given RNA's inherent instability and frequency of mutation, it is likely for sequence variability

to occur under field conditions. Therefore, it is important to vigilantly monitor the viral strains

that are co-circulating in a region to accurately and specifically detect viral RNA in mosquitoes

when applying this methodology in the field. It is critical to perform absolute quantification

analyses and determine end-point limitation values during optimization to avoid amplification of

false positive samples as a result of qPCR background noise. Potential future studies involve

bringing the portable platform into the field to analyze its feasibility for performing qPCR under

true field conditions. The versatility of the assay can be applied to future studies involving

Plasmodium falciparum detection in *Anopheles* species and monitoring of various species of the

Anopheles vector as well as detecting insecticide resistance in field-caught mosquitoes.

Conclusion

We have developed a highly sensitive qPCR method for detection and quantitation of DENV2

and ZIKV in single mosquitoes using the portable qPCR Hyris bCUBE platform. Our assay

allowed for sensitive and reliable DENV2 and ZIKV detection that make it a potential tool for

mosquito surveillance programs in endemic countries facing arboviral outbreaks. The detection

of *Wolbachia* in *Ae. aegypti* further demonstrates the platform's ability to be used as a multi-

detection assay. By overcoming the challenges of costs associated with reagents and equipment,

the bCUBE qPCR platform provides a potential field-deployable laboratory resources for

mosquito surveillance agencies in remote regions.

Acknowledgments

The authors thank Hyris, Inc. for providing the bCUBE and Mr Stefano Lo Priore and Mr

Lorenzo Colombo for technical advices. We also thank Dr. Zhiyong Xi from Michigan State

University for kindly providing *Wolbachia w*AlbB-infected *Ae. aegypti* strain WB1 and Dr.

Deborah McClellan for editorial assistance. We thank the Johns Hopkins Malaria Research

Institute Insectary core facility. This work has been supported by the National Institutes of

Health grant R21AI136456, and the Bloomberg Philanthropies.

References

1. Lee H, Halverson S, Ezinwa N. Mosquito-Borne Diseases. Prim Care. 2018;45(3):393-407.
2. Yun SI, Lee YM. Zika virus: An emerging flavivirus. J Microbiol. 2017;55(3):204-19.
3. Epelboin Y, Talaga S, Epelboin L, Dusfour I. Zika virus: An updated review of competent or
naturally infected mosquitoes. PLoS Negl Trop Dis. 2017;11(11):e0005933.
4. Esser HJ, Mögling R, Cleton NB, van der Jeugd H, Sprong H, Stroo A, et al. Risk factors
associated with sustained circulation of six zoonotic arboviruses: a systematic review for
selection of surveillance sites in non-endemic areas. Parasit Vectors. 2019;12(1):265.
5. Bhatt S, Gething PW, Brady OJ, Messina JP, Farlow AW, Moyes CL, et al. The global
distribution and burden of dengue. Nature. 2013;496(7446):504-7.
6. Guo C, Zhou Z, Wen Z, Liu Y, Zeng C, Xiao D, et al. Global Epidemiology of Dengue
Outbreaks in 1990-2015: A Systematic Review and Meta-Analysis. Front Cell Infect
Microbiol. 2017;7:317.
7. Petersen LR, Jamieson DJ, Powers AM, Honein MA. Zika Virus. N Engl J Med.
2016;374(16):1552-63.
8. Rasmussen SA, Jamieson DJ, Honein MA, Petersen LR. Zika Virus and Birth Defects--
Reviewing the Evidence for Causality. N Engl J Med. 2016;374(20):1981-7.
9. Ayres CF. Identification of Zika virus vectors and implications for control. Lancet Infect Dis.
2016;16(3):278-9.
10. Patterson J, Sammon M, Garg M. Dengue, Zika and Chikungunya: Emerging Arboviruses in
the New World. West J Emerg Med. 2016;17(6):671-9.

11. Weaver SC, Reisen WK. Present and future arboviral threats. Antiviral Res. 2010;85(2):328-
45.
12. Benelli G, Mehlhorn H. Declining malaria, rising of dengue and Zika virus: insights for
mosquito vector control. Parasitol Res. 2016;115(5):1747-54.
13. Kraemer MUG, Reiner RC, Jr., Brady OJ, Messina JP, Gilbert M, Pigott DM, et al. Past and
future spread of the arbovirus vectors Aedes aegypti and Aedes albopictus. Nat Microbiol.
2019;4(5):854-63.
14. Lequime S, Richard V, Cao-Lormeau VM, Lambrechts L. Full-genome dengue virus
sequencing in mosquito saliva shows lack of convergent positive selection during
transmission by Aedes aegypti. Virus Evol. 2017;3(2):vex031.
15. Ayllon T, Campos RM, Brasil P, Morone FC, Camara DCP, Meira GLS, et al. Early
Evidence for Zika Virus Circulation among Aedes aegypti Mosquitoes, Rio de Janeiro,
Brazil. Emerg Infect Dis. 2017;23(8):1411-2.
16. Walker T, Johnson PH, Moreira LA, Iturbe-Ormaetxe I, Frentiu FD, McMeniman CJ, et al.
The wMel Wolbachia strain blocks dengue and invades caged Aedes aegypti populations.
Nature. 2011;476(7361):450-3.
17. Dutra HL, Rocha MN, Dias FB, Mansur SB, Caragata EP, Moreira LA. Wolbachia Blocks
Currently Circulating Zika Virus Isolates in Brazilian Aedes aegypti Mosquitoes. Cell Host
Microbe. 2016;19(6):771-4.
18. Kamtchum-Tatuene J, Makepeace BL, Benjamin L, Baylis M, Solomon T. The potential role
of Wolbachia in controlling the transmission of emerging human arboviral infections. Curr
Opin Infect Dis. 2017;30(1):108-16.
19. Garcia GA, Sylvestre G, Aguiar R, da Costa GB, Martins AJ, Lima JBP, et al. Matching the
genetics of released and local Aedes aegypti populations is critical to assure Wolbachia
invasion. PLoS Negl Trop Dis. 2019;13(1):e0007023.
20. Flores HA, O'Neill SL. Controlling vector-borne diseases by releasing modified mosquitoes.
Nat Rev Microbiol. 2018;16(8):508-18.
21. Joubert DA, O'Neill SL. Comparison of Stable and Transient Wolbachia Infection Models in
Aedes aegypti to Block Dengue and West Nile Viruses. PLoS Negl Trop Dis.
2017;11(1):e0005275.
22. Waggoner JJ, Pinsky BA. Zika Virus: Diagnostics for an Emerging Pandemic Threat. J Clin
Microbiol. 2016;54(4):860-7.
23. Nasci RS, Gottfried KL, Burkhalter KL, Kulasekera VL, Lambert AJ, Lanciotti RS, et al.
Comparison of vero cell plaque assay, TaqMan reverse transcriptase polymerase chain
reaction RNA assay, and VecTest antigen assay for detection of West Nile virus in field-
collected mosquitoes. J Am Mosq Control Assoc. 2002;18(4):294-300.
24. Ramírez AL, van den Hurk AF, Meyer DB, Ritchie SA. Searching for the proverbial needle
in a haystack: advances in mosquito-borne arbovirus surveillance. Parasit Vectors.
2018;11(1):320.
25. Burkhalter KL, Wiggins K, Burkett-Cadena N, Alto BW. Laboratory Evaluation of
Commercially Available Platforms to Detect West Nile and Zika Viruses From Honey Cards.
J Med Entomol. 2018;55(3):717-22.
26. Girod R, Guidez A, Carinci R, Issaly J, Gaborit P, Ferrero E, et al. Detection of Chikungunya
Virus Circulation Using Sugar-Baited Traps during a Major Outbreak in French Guiana.
PLoS Negl Trop Dis. 2016;10(9):e0004876.

27. Yaren O, Alto BW, Gangodkar PV, Ranade SR, Patil KN, Bradley KM, et al. Point of
sampling detection of Zika virus within a multiplexed kit capable of detecting dengue and
chikungunya. BMC Infect Dis. 2017;17(1):293.
28. Wasik D, Mulchandani A, Yates MV. Point-of-Use Nanobiosensor for Detection of Dengue
Virus NS1 Antigen in Adult Aedes aegypti: A Potential Tool for Improved Dengue
Surveillance. Analytical Chemistry. 2018;90(1):679-84.
29. Fernandes JN, dos Santos LMB, Chouin-Carneiro T, Pavan MG, Garcia GA, David MR, et
al. Rapid, noninvasive detection of Zika virus in Aedes aegypti mosquitoes by
near-infrared spectroscopy. Science Advances. 2018;4(5):eaat0496.
30. Singh RK, Dhama K, Karthik K, Tiwari R, Khandia R, Munjal A, et al. Advances in
Diagnosis, Surveillance, and Monitoring of Zika Virus: An Update. Front Microbiol.
2017;8:2677.
31. Dejnirattisai W, Supasa P, Wongwiwat W, Rouvinski A, Barba-Spaeth G, Duangchinda T, et
al. Dengue virus sero-cross-reactivity drives antibody-dependent enhancement of infection
with zika virus. Nat Immunol. 2016;17(9):1102-8.
32. Benelli G. Managing mosquitoes and ticks in a rapidly changing world - Facts and trends.
Saudi J Biol Sci. 2019;26(5):921-9.
33. Molyneux DH, Savioli L, Engels D. Neglected tropical diseases: progress towards addressing
the chronic pandemic. Lancet. 2017;389(10066):312-25.
34. Tien WP, Lim G, Yeo G, Chiang SN, Chong CS, Ng LC, et al. SYBR green-based one step
quantitative real-time polymerase chain reaction assay for the detection of Zika virus in field-
caught mosquitoes. Parasit Vectors. 2017;10(1):427.
35. Xu MY, Liu SQ, Deng CL, Zhang QY, Zhang B. Detection of Zika virus by SYBR green
one-step real-time RT-PCR. J Virol Methods. 2016;236:93-7.
36. Charrel RN, Leparc-Goffart I, Pas S, de Lamballerie X, Koopmans M, Reusken C.
Background review for diagnostic test development for Zika virus infection. Bull World
Health Organ. 2016;94(8):574-84d.
37. Das S, Garver L, Ramirez JR, Xi Z, Dimopoulos G. Protocol for dengue infections in
mosquitoes (A. aegypti) and infection phenotype determination. J Vis Exp. 2007(5):220.
38. Xi Z, Khoo CC, Dobson SL. Wolbachia establishment and invasion in an Aedes aegypti
laboratory population. Science. 2005;310(5746):326-8.
39. Sim S, Ramirez JL, Dimopoulos G. Dengue virus infection of the Aedes aegypti salivary
gland and chemosensory apparatus induces genes that modulate infection and blood-feeding
behavior. PLoS Pathog. 2012;8(3):e1002631.
40. Lanciotti RS, Kosoy OL, Laven JJ, Velez JO, Lambert AJ, Johnson AJ, et al. Genetic and
serologic properties of Zika virus associated with an epidemic, Yap State, Micronesia, 2007.
Emerg Infect Dis. 2008;14(8):1232-9.
41. Frentiu FD, Zakir T, Walker T, Popovici J, Pyke AT, van den Hurk A, et al. Limited dengue
virus replication in field-collected Aedes aegypti mosquitoes infected with Wolbachia. PLoS
Negl Trop Dis. 2014;8(2):e2688.
42. Axford JK, Ross PA, Yeap HL, Callahan AG, Hoffmann AA. Fitness of wAlbB Wolbachia
Infection in Aedes aegypti: Parameter Estimates in an Outcrossed Background and Potential
for Population Invasion. Am J Trop Med Hyg. 2016;94(3):507-16.
43. Danet L, Beauclair G, Berthet M, Moratorio G, Gracias S, Tangy F, et al. Midgut barriers
prevent the replication and dissemination of the yellow fever vaccine in Aedes aegypti. PLoS
Negl Trop Dis. 2019;13(8):e0007299.

44. Lanciotti RS, Kerst AJ, Nasci RS, Godsey MS, Mitchell CJ, Savage HM, et al. Rapid
detection of west nile virus from human clinical specimens, field-collected mosquitoes, and
avian samples by a TaqMan reverse transcriptase-PCR assay. J Clin Microbiol.
2000;38(11):4066-71.
45. Khan SA, Borah J, Chowdhury P, Dutta P, Mahanta J. Characterization of Japanese
encephalitis virus (JEV) genotype III clinical isolates in northeast India. Trans R Soc Trop
Med Hyg. 2015;109(8):522-8.
46. Wu W, Wang J, Yu N, Yan J, Zhuo Z, Chen M, et al. Development of multiplex real-time
reverse-transcriptase polymerase chain reaction assay for simultaneous detection of Zika,
dengue, yellow fever, and chikungunya viruses in a single tube. J Med Virol.
2018;90(11):1681-6.
47. Goncalves DDS, Hooker DJ, Dong Y, Baran N, Kyrylos P, Iturbe-Ormaetxe I, et al.
Detecting wMel Wolbachia in field-collected Aedes aegypti mosquitoes using loop-mediated
isothermal amplification (LAMP). Parasit Vectors. 2019;12(1):404.
48. Carrillo-Hernandez MY, Ruiz-Saenz J, Villamizar LJ, Gomez-Rangel SY, Martinez-
Gutierrez M. Co-circulation and simultaneous co-infection of dengue, chikungunya, and zika
viruses in patients with febrile syndrome at the Colombian-Venezuelan border. BMC Infect
Dis. 2018;18(1):61.
49. Aubry M, Richard V, Green J, Broult J, Musso D. Inactivation of Zika virus in plasma with
amotosalen and ultraviolet A illumination. Transfusion. 2016;56(1):33-40.
50. Gu W, Lampman R, Novak RJ. Assessment of arbovirus vector infection rates using variable
size pooling. Med Vet Entomol. 2004;18(2):200-4.
51. Musso D, Richard V, Broult J, Cao-Lormeau VM. Inactivation of dengue virus in plasma
with amotosalen and ultraviolet A illumination. Transfusion. 2014;54(11):2924-30.
623