1 **Validation of a single-step, single-tube reverse transcription-loop-mediated**

2 **isothermal amplification assay for rapid detection of SARS-CoV-2 RNA**

1.1 Author names

Jean Y. H. Lee[1,2], Nickala Best[3], Julie McAuley[1], Jessica L. Porter[1], Torsten Seemann[1,4], Mark B.

Schultz[4], Michelle Sait[4], Nicole Orlando[4], Karolina Mercoulia[4], Susan A. Ballard[4], Julian Druce[5],

Thomas Tran[5], Mike G. Catton[5], Melinda J. Pryor[6], Huanhuan L. Cui[6], Angela Luttick[6], Sean

McDonald[3], Arran Greenhalgh[3], Jason C. Kwong[1,7], Norelle L. Sherry[4,7], Maryza Graham[8], Tuyet

Hoang[1,4], Marion Herisse[1], Sacha J. Pidot[1], Deborah A. Williamson[1,4,9], Benjamin P. Howden[1,4,7], Ian

R. Monk[1,#], and Timothy P. Stinear[1,*,#]

1.2 Affiliation

1. Department of Microbiology and Immunology, University of Melbourne at the Peter Doherty

Institute for Infection and Immunity, Melbourne, Victoria, Australia.

2. Department of Infectious Diseases, Monash Health, Clayton, Victoria, Australia.

3. GeneWorks Pty Ltd, Thebarton, South Australia, Australia.

4. Microbiological Diagnostic Unit Public Health Laboratory, University of Melbourne at the Peter

Doherty Institute for Infection and Immunity, Melbourne, Victoria, Australia.

5. Victorian Infectious Diseases Reference Laboratory, Melbourne Health at the Peter Doherty

Institute for Infection and Immunity, Melbourne, Victoria, Australia.

6. 360Biolabs, Melbourne, Victoria, Australia

7. Department of Infectious Diseases, Austin Health, Heidelberg, Victoria, Australia

8. Department of Microbiology, Monash Health, Clayton, Victoria, Australia

9. Melbourne Health, Melbourne, Victoria, Australia

1.3   Corresponding author

* Corresponding author, tstinear@unimelb.edu.au

Joint senior authors

1.4   Keyword

SARS-CoV-2, RT-LAMP, universal transport media, nasopharyngeal swabs

2. Abstract

Introduction: The SARS-CoV-2 pandemic of 2020 has resulted in unparalleled requirements for

RNA extraction kits and enzymes required for virus detection, leading to global shortages. This has

necessitated the exploration of alternative diagnostic options to alleviate supply chain issues.

Aim: To establish and validate a reverse transcription loop-mediated isothermal amplification (RT-

LAMP) assay for the detection of SARS-CoV-2 from nasopharyngeal swabs.

Methodology: We used a commercial RT-LAMP mastermix from OptiGene Ltd in combination with

a primer set designed to detect the CDC N1 region of the SARS-CoV-2 nucleocapsid (N) gene. A

single-tube, single-step fluorescence assay was implemented whereby as little as 1 µL of universal

transport medium (UTM) directly from a nasopharyngeal swab could be used as template, bypassing

the requirement for RNA purification. Amplification and detection could be conducted in any thermocycler capable of holding 65°C for 30 minutes and measure fluorescence in the FAM channel at one-minute intervals.

Results: Assay evaluation by assessment of 157 clinical specimens previously screened by E-gene RT-qPCR revealed assay sensitivity and specificity of 87% and 100%, respectively. Results were fast, with an average time-to-positive (Tp) for 93 clinical samples of 14 minutes (SD $\pm$7 minutes). Using dilutions of SARS-CoV-2 virus spiked into UTM, we also evaluated assay performance against FDA guidelines for implementation of emergency-use diagnostics and established a limit-of-detection of 54 Tissue Culture Infectious Dose 50 per ml ($TCID_{50}$ mL^{-1}), with satisfactory assay sensitivity and specificity. A comparison of 20 clinical specimens between four laboratories showed excellent interlaboratory concordance; performing equally well on three different, commonly used thermocyclers, pointing to the robustness of the assay.

Conclusion: With a simplified workflow, N1-STOP-LAMP is a powerful, scalable option for specific and rapid detection of SARS-CoV-2 and an additional resource in the diagnostic armamentarium against COVID-19.

3. Data summary

The authors confirm all supporting data, code and protocols have been provided within the article or through supplementary data files.

4. Introduction

The current SARS-CoV-2 pandemic has created an unprecedented global demand for rapid diagnostic testing. Most countries are employing real-time quantitative PCR (RT-qPCR) for confirmation of infection [1-7]. Consequently, a global shortage of RNA extraction kits as well as RT-qPCR assay kits and their associated reagents has ensued [8-10]. Therefore, alternative diagnostics not dependent on these commonly used materials are required. First described 20 years ago by Notomi *et al.*, the loop-mediated isothermal amplification (LAMP) assay is robust, rapid and straightforward, yet retains high sensitivity and specificity [11]. These features have seen the LAMP assay and the inclusion of a reverse transcriptase (RT-LAMP) implemented for a broad range of molecular diagnostic applications extending from infectious diseases, including detection of the original SARS-CoV-1 virus [12], bacteria and parasites [13-15] to cancer [16]. The advantages of RT-LAMP include; using different reagents than RT-qPCR, the potential for direct processing of samples without the need for prior RNA extraction and an extremely rapid turn-around time. Several groups have now described different RT-LAMP assays for detection of SARS-CoV-2 RNA [17-24].

In this study, we developed a RT-LAMP assay that targets the CDC N1 region of the SARS-CoV-2 nucleocapsid gene (N-gene) [3] and used a commercial mastermix from OptiGene Ltd. This mix contains a proprietary reverse transcriptase for cDNA synthesis and the thermophilic GspSSD strand-displacing polymerase/reverse transcriptase for DNA amplification (www.optigene.co.uk) with a dsDNA intercalating fluorescent dye. Detection was achieved by measuring the increase in fluorescence as amplification products accumulate. The **N1** gene **S**ingle **T**ube **Op**tigene **LAMP** assay (hereafter called N1-STOP-LAMP) was assessed for the direct detection of SARS-CoV-2 RNA, following *FDA Policy for Diagnostic Tests for Coronavirus Disease-2019 during the Public Health Emergency* against the four parameters and acceptance criteria summarised in Table 1 [5]. Validation

87 samples were human upper respiratory tract specimens, collected using nasopharyngeal flocked swabs

88 stored in universal transport media (UTM).

5. Methods

Specimen collection and handling

Nasopharyngeal swabs were collected by qualified healthcare professionals from patients meeting the

epidemiological and clinical criteria as specified by the Victorian Department of Health and Human

Services at the time of swab collection [1] between the 23rd of March 2020 and 4th of April 2020.

Copan flocked swabs collected and directly inoculated on site in either 1 mL or 3 mL of UTM®

(Catalogue N^{o}s., 330C and 350C respectively) were used. Samples were collected at metropolitan

hospitals in Melbourne, Victoria, Australia, and transported to the Doherty Institute Public Health

Laboratories for further testing as per World Health Organization recommendations [6]. All swabs

were processed in a class II biological safety cabinet.

Cell culture and SARS-CoV-2 RNA extraction

Vero cells (within 30 passages from the original American Type Culture Collection [ATCC] stock)

were maintained in Minimal Essential Media (MEM) supplemented with 10% heat-inactivated foetal

bovine serum (FBS), 10 μM HEPES, 2 mM glutamine and antibiotics. Cell cultures were maintained

at 37^{o}C in a 5% CO_2 incubator. All virus infection cultures were conducted within the High

Containment Facilities in a PC3 laboratory at the Doherty Institute. To generate stocks of SARS-CoV-

2, confluent Vero cell monolayers were washed once with MEM without FBS (infection media) then

infected with a known amount SARS-CoV-2 virus originally isolated from a patient [25]. After 1h

incubation at 37°C in a 5% CO2 incubator to enable virus binding, infection media containing 1 mg

mL^{-1} TPCK-trypsin was added, and flasks returned to the incubator. After 3d incubation and

microscopic confirmation of widespread cytopathic effect (CPE), the supernatants were harvested,

and filter sterilised through a 0.45 µm syringe filter. To assess infectious SARS-CoV-2 viral titres,

both Tissue Culture Infectious Dose 50 (TCID$_{50}$) and plaque assays were performed. Briefly, serial

dilutions of the stock virus were added to washed monolayers of Vero cells. After 1h incubation to

allow virus to adhere, for the TCID$_{50}$ assay, infection media containing 1 µg mL^{-1} TPCK-Trypsin was

added, while for the plaque assay the infected cell monolayer was overlaid with Leibovitz-15 (L15)

media supplemented with 0.9% agarose (DNA grade, Sigma), antibiotics and 2 µg mL^{-1} TPCK

Trypsin. After 3d incubation, the dilution of stock required to cause CPE in at least 50% of wells

(TCID$_{50}$) was determined via back calculation of microscopic confirmation of CPE in wells for a

given dilution. For quantitation via plaque assay, plaques present in the monolayer were

macroscopically visualised, individually counted and back calculated for each given dilution to

determine the number of Plaque Forming Units (PFU) per mL in the original stock. Stocks of SARS-

CoV-2 used by this study had a TCID$_{50}$ of 5.4×10^5 mL^{-1} and plaque assay gave 9.78×10^5 PFU mL^{-1}.

To heat inactivate the virus, 200 µL of neat stock was heated to 60°C for 30 min, then cooled.

Inactivation was confirmed via complete lack of CPE and plaque formation using both TCID$_{50}$ and

plaque assays. To prepare SARS-CoV-2 RNA from stocks, 500 µL aliquots were thawed and RNA

extracted using the RNeasy mini Kit (Qiagen) according to the manufacturer's specifications, with an

elution volume of 50 µL. Based on RNA concentrations, the total virus harvested from Vero cell

culture was 3.1×10^9 copies mL^{-1}, suggesting a high number of non-infectious virus particles in the

virus stocks.

RNA extraction from UTM for RT-qPCR

A 200 µL aliquot of Copan UTM® was processed through the QIAsymphony DSP Virus/Pathogen

Mini Kit (Qiagen, Cat# 937036) following manufacturer's instructions on the QIAsymphony SP

instrument. The RNA was eluted in 60 µL of recommended buffer.

RNA extraction from UTM SPRI beads for N1-STOP-LAMP

Solid-phase reversible immobilisation (SPRI) on carboxylated paramagnetic beads (Sera-Mag™

Magnetic SpeedBeads™, from GE Healthcare) were prepared for RNA binding as described

(https://openwetware.org/wiki/SPRI_bead_mix). RNA purification was performed in 96-well plates

with initial lysis by the addition of 25 µL of 6GTD lysis buffer (7.08 g 10 mL^{-1} guanidine thiocyanate

(6M), made up to 8.2 mL with water, 1 mL Tris HCL [pH 8.0] and 800 µL of 1M dithiothreitol),

mixed by pipetting ten times and incubation at room temperature for 1 min [26]. To this, 75 µL of

100% ethyl alcohol and 20 µL of prepared SPRI beads were added, mixed by pipetting ten times and

incubated at room temperature for 5 min. The RNA-bead complex was then immobilised by placing

the 96-well plate on a magnetic rack and incubated again at room temperature for 5 min. The

supernatant was then discarded, and the beads washed twice in 200 µL of freshly prepared 80% ethyl

alcohol (v/v) with 30 sec room temperature incubation between each wash. The beads were air-dried

for 2 min at room temperature before RNA was eluted by the addition of 20 µL of nuclease-free

water.

E-gene RT-qPCR

A one-step RT-qPCR was conducted with the primers as described by Corman *et al*., targeting the

viral envelope E-gene of SARS-CoV-2 E_Sarbeco_F: 5'-

ACAGGTACGTTAATAGTTAATAGCGT-3', 5'-E_Sarbeco_R:

ATATTGCAGCAGTACGCACACA-3', E_Sarbeco_P: 5'-FAM-

ACACTAGCCATCCTTACTGCGCTTCG-BHQ1-3') [27]. A 20 µL reaction was assembled

consisting of 1x qScript XLT One-Step RT-qPCR ToughMix Low ROX (2x) (QuantaBio), 400 nM

E_Sarbeco_F, 400 nM E_Sarbeco_R 200 nM E_Sarbeco_P (Probe) and 5 µL of purified RNA. The

following program was conducted on an ABI 7500 Fast instrument: 55°C for 10 min for reverse

transcription, 1 cycle of 95°C for 3 min and then 45 cycles of 95°C for 15 sec and 58°C for 30 sec.

N1-STOP-LAMP

A 50-reaction bottle of dried Reverse Transcriptase Isothermal Mastermix (Optigene, ISO-DR004-

RT) was rehydrated with 750 µL of resuspension buffer and vortexed gently to mix. The mix contains

both a proprietary reverse transcriptase and the GspSSD LF DNA polymerase that has both reverse

transcriptase and strand-displacing DNA polymerase activity. The N1-STOP-LAMP assay uses six

standard LAMP oligonucleotide primers that target the sequence spanning the CDC N1 region of the

SARS-CoV-2 nucleocapsid gene. Sequences of the primers are available upon request (GeneWorks

Ltd). Each N1-STOP-LAMP reaction contained: 15 µL of mastermix, 5 µL of 5x primer stock and 4

µL of water. Mastermix was reconstituted in a separate biological safety cabinet to that used for

template addition. The source of RNA template for the reaction consisted of either 1 µL of purified

SARS-CoV-2 RNA, 1 µL of SARS-CoV-2 purified virus or 1 µL of UTM from a nasopharyngeal

swab. A no-template control (1 µL water) was included in all runs. Reactions were assembled in

either 8-tube Genie strips (OP-00008, OptiGene Ltd) or 96-MicroAmp-Fast-Optical reaction plate

(Applied Biosystems). Strip tubes reactions were capped, or for 96-MicroAmp-Fast-Optical reaction

plates, sealed with MicroAmp Optical adhesive film (Applied Biosystems) and tapped to remove

bubbles. Strip tubes were loaded onto the Genie-II or Genie-III (OptiGene Ltd) and 96-well plates run

on a QuantStudio 7 thermocycler (ThermoFisher). Reactions were incubated at 65°C for 30 min with

fluorescence acquisition every 30 sec (Genie instruments) or 1 min (QuantStudio 7). A positive result

was indicated by an increase in fluorescence at an emission wavelength of 540 nm (FAM channel)

above a defined threshold, recorded as time-to-positive (Tp) expressed in min:sec.

Dilution series of purified SARS-CoV-2

A virus stock with of 5.4×10^5 $TCID_{50}$ mL^{-1} was serially diluted to 10^{-6} in a generic UTM (comprising

per liter: 8 g NaCl, 0.2 g KCl, 1.15 g Na_2HPO_4, 0.2 g KH_2PO_4, 4 mL 0.5% phenol red 0.5%, 5 g

gelatin, 950 mL tissue culture water, 1 mL Fungizone [5 mg mL^{-1}], 20 mL penicillin/streptomycin) in

biological triplicates, with the 10^{-1} through to the 10^{-6} dilutions tested by N1-STOP-LAMP (1 μL).

RNA was extracted from a 200 μL aliquot of each dilution as described above and 5 μL of the

purified RNA used as template in E-gene RT-qPCR or N1-STOP-LAMP assay.

Interlaboratory comparison of N1-STOP-LAMP

A panel of 20 blinded clinical samples (13 positive and 7 negative) with cycle threshold (Ct) values

previously established by E-gene RT-qPCR were aliquoted and distributed to three different

laboratories for independent testing by N1-STOP-LAMP assay (Table S2).

Biological specificity

To test for cross reactivity of the N1-STOP-LAMP assay, a control panel of respiratory pathogens

(NATRPC2-BIO, ZeptoMetrix) was screened (Table S3). A 1 μL aliquot of NATrol RP1 or RP2, or

samples spiked with SARS-CoV-2 virus were used as template for N1-STOP-LAMP.

In silico nucleotide sequence comparisons of the N1 region of SARS-CoV-2

To assess the inclusivity and exclusivity of the N1 region targeted by the LAMP assay 2755 publicly available SARS-CoV-2 genomes were downloaded and filtered to remove entries of less than 29 kb using *Seqtk seq* (v1.3-r106). Sequence gaps were replaced with 'N' using *Seqkit* (v0.12.0). After this quality control step, homology searches were then conducted using *NCBI BLAST+ blastn* using *Genepuller.pl* (https://github.com/tseemann/bioinfo-scripts/blob/master/bin/gene-puller.pl) to find the region in each of the remaining 2738 genomes matching the 5' region of the N1 sequence. The resulting sequence coordinates of those hits were then used to extract the 240 bp region from all 2738 genomes using *Genepuller.pl* and aligned with Clustalo (v1.2.4). The alignment was visualised with *Mesquite* (v3.61).

Statistical analysis

Data analysis was managed using GraphPad Prism (v8.4.1). Sensitivity and specificity testing were performed using the Wilson/Brown hybrid method as deployed in GraphPad Prism.

6. Results

Assessing detection sensitivity of N1-STOP-LAMP using purified RNA.

We began by assessing the limit of detection (LoD) of N1-STOP-LAMP under optimum conditions using a 10-fold dilution series of purified RNA, prepared from a titred SARS-CoV-2 virus stock. Although LAMP assays are not strictly quantitative, a positive correlation between Tp and RNA

concentration was evident. This was indicated by an absolute detection threshold between 50 and 500

viral genome copies per reaction. We achieved a reliable detection of 5/5 replicates at 500 viral

genome copies and detection of 4/5 replicates at 50 viral genome copies per reaction (Fig. 1A). Of

note, this threshold is likely an underestimate of the true detection limit, as we assumed all RNA

yielded from the viral stock generated from the Vero cell supernatant was of viral origin. Using

$TCID_{50}$, the absolute LoD of N1-STOP-LAMP was between 0.001 and 0.01 $TCID_{50}$ per reaction

(equivalent to 1-10 $TCID_{50}$ mL^{-1}).

Direct detection of SARS-CoV-2 RNA in clinical specimens and specimen inactivation

The nasopharyngeal specimens received by our public health laboratories were collected using Copan

flocked swabs in either 1 mL or 3 mL of UTM®. To assess the impact of sample matrix on the LAMP

assay, we performed pilot experiments with UTM from swabs taken from SARS-CoV-2 negative

specimens, adding increasing amounts of UTM to N1-STOP-LAMP. Sterile UTM had no impact on

N1-STOP-LAMP (data not shown). However, the nasopharyngeal secretions in patient samples were

observed to be inhibitory. Most pronounced when 5 µL of patient sample was used as the direct RNA

template (Fig. 1B), but relieved at a 1/5 dilution of the patient sample, thus 5 µL of a 1/5 dilution of

the patient sample or a 1 µL of neat patient sample was selected as the optimum template volume for

N1-STOP-LAMP (Fig. 2A). Experiments were also conducted using dry swabs eluted in PBS. We

tested elution in 0.5 mL, 1.0 mL and 1.5 mL and found that an elution volume of 1.5 mL of PBS was

a good compromise between unnecessary dilution of potential virus in the sample and sufficient

dilution to alleviate assay inhibition (data not shown). To reduce the risk associated with handling

clinical specimens containing infectious SARS-CoV-2, we also assessed the impact of a heat

inactivation step on detection sensitivity. Pre-treatment at 60°C for 30 min led to a >5-$\log_{10}$

inactivation of the virus accessed by PFU and $TCID_{50}$ determination (data not shown). No impact on

N1-STOP-LAMP detection sensitivity was observed, with equivalent Tp between each treatment

(Table 2, Fig. 1C).

Comparative detection sensitivity of N1-STOP-LAMP versus E-gene RT-qPCR.

The N1-STOP-LAMP assay was then directly compared with E-gene RT-qPCR, using a dilution of

titred SARS-CoV-2 virus stock in UTM across a 6-$\log_{10}$ dilution series. A RT-qPCR calibration curve

was established from triplicate extraction experiments for each of the six dilutions (Fig. 2A). The

assays were prepared with 5 µL of purified viral RNA as template for both the N1-STOP-LAMP and

RT-qPCR assays. A 1 µL and 5 µL aliquot of each virus dilution in UTM was also tested directly in

the N1-STOP-LAMP assay (*i.e.* without RNA extraction). The side-by-side comparisons showed E-

gene RT-qPCR was up to 1-$\log_{10}$ more sensitive than N1-STOP-LAMP, with RT-qPCR detecting 2/3

replicates at the lowest dilution of 0.54 $TCID_{50}$ mL^{-1} and N1-STOP-LAMP detecting no viral RNA at

this concentration (Table 3, Fig. 2). At 54 $TCID_{50}$ mL^{-1}, the RT-qPCR and N1-STOP-LAMP detected

3/3 replicates. There was no difference in N1-STOP-LAMP sensitivity using either 5 µL UTM added

directly to the test or purified RNA, but the assay lost another 1-$\log_{10}$ sensitivity where 1 µL of neat

UTM was added directly to the N1-STOP-LAMP reaction (Fig. 2B).

Establishing N1-STOP-LAMP limit of detection (LoD)

The FDA guidelines for implementation of emergency-use diagnostics defines LoD as the lowest

concentration at which 19/20 replicates are positive (Table 1). Informed by the previous experiment

(Fig. 2B), we selected a SARS-CoV-2 concentration of 54 $TCID_{50}$ mL^{-1} (in UTM) and tested 20x 1 µL

aliquots by N1-STOP-LAMP. We observed 20/20 positive reactions with an average Tp of 15.8 mins,

inter-quartile range 12-19 mins and coefficient of variation of 31.18% (Fig. 3A).

Evaluation of N1-STOP-LAMP against FDA criteria.

The FDA guidelines for implementation of emergency-use diagnostics also require an assessment of assay performance using at least 30 contrived positive and 30 negative clinical specimens (authentic or contrived), with at least 20 of the positives at a concentration of 1-2x the LoD (Table 1). We obtained 30 nasopharyngeal dry swabs from healthy, anonymous volunteers. Swabs were eluted in 1.5 mL of PBS and 1 µL aliquots were tested directly by N1-STOP-LAMP. All eluates were negative. We then spiked the eluates with SARS-CoV-2 virus at 1x and 2x the LoD (approx. 50 and 100 $TCID_{50}$ mL^{-1}, respectively). Results showed that N1-STOP-LAMP detected 20/20 samples spiked with 1x the assay LoD and 10/10 samples at 2x LoD (Fig. 3B, Table 4). N1-STOP-LAMP thus meets the FDA Clinical Evaluation criteria (Table 1).

Clinical specimen evaluation of N1-STOP-LAMP against E-gene RT-qPCR gold standard

We then sought to evaluate N1-STOP-LAMP performance against E-gene RT-qPCR, setting the latter assay as the current 'gold standard'. We directly screened 50 negative clinical specimens and 107 positive nasopharyngeal clinical specimens as established by E-gene RT-qPCR (Table S1). Results showed N1-STOP-LAMP assay sensitivity of 87%, specificity of 100%, positive predictive value of 100% and negative predictive value of 78% (Table 5, Fig. 3). As noted during initial sensitivity experiments (Table 3), a positive correlation was observed between N1-STOP-LAMP Tp and E-gene RT-qPCR Ct (Fig. 4). The average Tp among the 93 N1-LAMP-STOP positive specimens was 14 min (SD $\pm$ 7 min). For the false-negative samples with sufficient clinical material remaining (5 out of 14), RNA extraction using a simple, rapid magnetic bead purification protocol (takes 20 min) was performed, then N1-STOP-LAMP repeated. Four of the 5 samples returned a positive result with Tp

range from 10:00 – 25:30 min:sec (Table S1). Indicating that the reduced sensitivity of N1-STOP-

LAMP compared to RT-qPCR can be improved upon by using samples with a higher RNA template

concentration and purity with only a small increase in time to results.

Confirmation of N1-STOP-LAMP inclusivity and exclusivity

The FDA guidelines for implementation of emergency-use diagnostics require an assessment of

inclusivity and exclusivity of the oligonucleotide primers used in an assay. To assess the former, we

performed *in silico* screening of the region spanned by the six primers used for the N1-STOP-LAMP

assay against all available SARS-CoV-2 genome sequences (Supplementary data file 1). This 240 bp

sequence spans the CDC N1 region of the SARS-CoV-2 N-gene [3]. Alignment of this region against

2738 genomes revealed 100% conservation amongst all entries, showing that this a conserved target

sequence and that the primers used will perform equally across all identified lineages of the virus.

To assess the exclusivity criterion, we used a nucleotide BLAST search of the N1 region against the

NCBI Genbank *nt* database and observed no non-SARS-CoV-2 sequence matches above 80%

nucleotide identity, in-line with FDA cross-reactivity requirement. Further to this analysis, we also

performed *in vitro* testing with the NATRPC2-BIO (ZeptoMetrix) specificity panel, a control panel

consisting of 22 common respiratory pathogens. None of these pathogens were detected by N1-STOP-

LAMP (Table S3). Thus, both *in silico* and *in vitro* testing confirmed the specificity of N1-STOP-

LAMP for SARS-CoV-2.

Interlaboratory comparisons

In order to explore the ability of other laboratories to readily deploy N1-STOP-LAMP, we prepared

an external-quality-assessment (EQA) panel of 20 clinical specimens (Table S2). These were previous

positive and negative clinical samples as assessed by E-gene RT-qPCR that had been heat-inactivated

for safety. These specimens covered a range of virus titres (Ct values 16.3 – 30.6). The panel was

tested simultaneously in four different laboratories (Labs A – D). The results showed excellent

correspondence between all laboratories for ten positive samples, the majority with E-gene RT-qPCR

Ct values <26 (Fig. 5, Table S2). Above this Ct (*i.e.* those samples with lower virus titres), the

laboratories returned variably positive results (Samples 14, 16 & 18), reflecting that these specimens

had virus concentrations at or beyond the N1-STOP-LAMP LoD (Table S2). Concordance among the

seven negative results was overall very good, however one laboratory (Lab C) returned a false

positive result (Sample 2), highlighting the issue of contamination with sensitive molecular tests

(Table S2). Of note, the laboratories used a range of different instruments for amplification/detection

including: OptiGene Genie II and III, BioRad CFX and ThermoFisher Quantstudio 7. This trial

suggests that N1-STOP-LAMP is a robust and transferrable assay format.

7. Discussion

A critical component of an effective COVID-19 pandemic response is the rapid and robust detection

of positive clinical samples [28]. This has required massively upscaled SARS-CoV-2 diagnostic

capabilities. Particularly, a worldwide focus on developing improved technologies [29]. There have

been more than 20 molecular tests recently receiving FDA Emergency Use Authorisations (EUAs)

[5]. In this current study, we demonstrate the potential of N1-STOP-LAMP. We focused efforts on the

LAMP assay because it has been previously employed for the rapid and robust detection of numerous

RNA viruses including Zika, Chikungunya, Influenza and SARS-CoV-1 [12, 30-33] and was of great

assistance during these outbreaks, particularly in resource constrained settings. Advantages of LAMP

testing include rapid turnaround time, ease of implementation, non-standard reagent use and potential

utility at point of care [19-21]. To date, there has been limited detail on the performance

characteristics of emerging RT-LAMP based tests in head-to-head comparisons with gold standard

assays. Such information is critical to ensuring the safe deployment of new tests, given the clinical

and public health consequences of erroneous test results.

Here we have shown the specific detection of SARS-CoV-2 RNA directly from clinical swabs

without the need for RNA purification, using a LAMP assay targeting the N1 region of the SARS-

CoV-2 N-gene in conjunction with the reverse transcriptase/GspSSD LF DNA polymerase mastermix

from OptiGene Ltd. With a detection limit of 54 $TCID_{50}$ mL^{-1} the N1-STOP-LAMP assay met each of

the four FDA criteria for EUA. Based on reported data, as opposed to head-to-head comparisons, the

N1-STOP-LAMP detection limit, is higher than several recently reported FDA-EUA COVID-19 tests.

Expressed as virus copies per mL the N1-STOP-LAMP LoD is estimated at 54,000 virus copies per

mL (assumes our $TCID_{50}$ underestimates virus copy number by 1000-fold, see methods), compared to

Cepheid GeneXpert Xpress 250 virus copies per mL or Abbott ID NOW COVID-19 test 125 virus

copies per mL. Other EUA assays reporting LoDs similar to N1-STOP-LAMP include the Luminex

ARIES SARS-CoV-2 assay (75,000 copies per mL) and the GenMark ePlex SARS-CoV-2 test

(100,000 copies per mL). Given these performance metrics, and the test's high positive predictive

value (100%), we see N1-STOP-LAMP as well suited for widespread screening of large populations

when COVID-19 prevalence is low. N1-STOP-LAMP could be used in large-scale, national testing

programs to support aggressive COVID-19 contact tracing and disease suppression or elimination

activities. The test could also play a role in near-point-of-care settings, such as outbreak investigation

in hospitals and nursing homes, providing rapid-turnaround-time of results for health-care workers

and highly vulnerable patients. The relatively simple assay format of N1-STOP-LAMP is also suitable

16737403-file00.docx

for deployment in lower and middle-income countries, where access to sophisticated laboratory

infrastructure is limited.

During this evaluation we noted opportunities for improvement of N1-STOP-LAMP. We observed

that although 1 µL of UTM was tolerated in the OptiGene RT-LAMP formulation, larger quantities of

sample matrix (UTM or PBS containing additional human material from the nasopharyngeal swab)

were inhibitory to the RT-LAMP reaction. If more than 1 µL of spiked (purified virus or RNA as

template) sample matrix was used in the assay, no amplification was observed (Fig. 1B). However,

under ideal conditions (*i.e.*, no sample inhibition from human components), the impact of using an

increased template volume was shown, where 5 µL of UTM added directly to the N1-STOP-LAMP

reaction was equal to the sensitivity of using purified RNA (Fig. 2B). Furthermore, addition of a

simple RNA purification step, that does not require commercial kits, improved the detection

sensitivity of N1-STOP-LAMP to rival RT-qPCR. Our follow-up of five false-negative results (Table

5, Fig. 3C), demonstrated the feasibility of including a simple, scalable and rapid magnetic-bead RNA

purification step [26], with which we were able to detect SARS-CoV-2 RNA with N1-STOP-LAMP

in 4/5 UTM specimens that had previously tested negative by the assay (Table S1). Other

opportunities for improvement include the use of specific fluorescent probes to detect amplification

products, thus facilitating multiplex reactions and the inclusion of an internal amplification control

within each test [24,34].

As mentioned, recognised strengths of the LAMP assay include the rapid time to test result, with the

majority of the positive clinical samples yielding a result in under 15 min, and the robust test format,

testing directly from the swab eluate. Our preferred specimen type for N1-STOP-LAMP is a dry

nasopharyngeal swab. Detailed examination of optimum swabs for the detection of influenza and

other RNA viruses has shown dry swabs offer superior performance [35]. In the current study we have

shown that N1-STOP-LAMP has satisfactory performance using dry swabs eluted in 1.5 mL of PBS

(Fig. 3B); a format that potentially doubles the effective concentration of virus in the sample

compared to 3 mL of UTM in the Copan format. Another strength of N1-STOP-LAMP is that the

assay readily scales from small sample numbers (e.g. 8-well with portable detection unit) to high

throughput (run with sample robotics and 96-well thermocyclers in a centralised laboratory).

While establishing the assay, we noted susceptibility to contamination that was easily addressed by

the standard precautions used to prevent template carryover for any nucleic acid amplification test,

such as: physical separation of mastermix preparation from sample inoculation; not opening post-

amplification reaction tubes or plates; and inclusion of negative extraction controls. Addition of

uracil-N-glycosylase has been reported as a strategy to limit the risk of amplicon contamination,

however this is associated with a loss of detection sensitivity [36].

In this report we have shown N1-STOP-LAMP is robust diagnostic test for the specific and rapid

detection of SARS-CoV-2. It is an alternative molecular test for SARS-CoV-2 that can be readily and

seamlessly deployed, particularly when access to standard RT-qPCR based approaches are limited.

8. Author statements

8.1 Authors and contributors

Jean Lee:

Conceptualisation

Data curation

Formal analysis

Funding acquisition

Investigation

Methodology

Project administration

Resources

Software

Supervision

Validation

Visualisation

Writing - original draft

Writing - review & editing

Nickala Best:

Conceptualisation

Data curation

Formal analysis

Investigation

Methodology

Resources

Validation

Visualisation

Writing - original draft

Writing - review & editing

Julie McAuley:

Conceptualisation

Data curation

Formal analysis

Investigation

Methodology

Resources

Validation

Writing - review & editing

Jessica Porter:

Data curation

Formal analysis

Investigation

Methodology

Writing - review & editing

Torsten Seemann:

Conceptualisation

Investigation

Methodology

Software

Mark Schultz:

Conceptualisation

Investigation

Methodology

Software

Michelle Sait:

Formal analysis

Investigation

Methodology

Software

Nicole Orlando:

Formal analysis

Investigation

Methodology

Karolina Mercoulia:

Data curation

Investigation

Methodology

Susan Ballard:

Conceptualisation

Data curation

Formal analysis

Investigation

Methodology

Julian Druce:

Conceptualisation

Data curation

Formal analysis

Investigation

Methodology

Thomas Tran:

Investigation

Methodology

Mike Catton:

Conceptualisation

Investigation

Supervision

Writing - review & editing

Melinda Pryor:

Conceptualisation

Investigation

Methodology

Supervision

Writing - review & editing

Huanhuan Cui:

Investigation

Methodology

Writing - review & editing

Angela Luttick:

Investigation

Methodology

Writing - review & editing

Sean McDonald:

Conceptualisation

Formal analysis

Investigation

Methodology

Writing - review & editing

Arran Greenhalgh:

Conceptualisation

Formal analysis

Funding acquisition

Investigation

Methodology

Writing - review & editing

Jason Kwong:

Conceptualisation

Formal analysis

Funding acquisition

Methodology

Writing - original draft

Writing - review & editing

Norelle Sherry:

Conceptualisation

Formal analysis

Investigation

Methodology

Validation

Writing - original draft

Writing - review & editing

Maryza Graham:

Conceptualisation

Formal analysis

Investigation

Methodology

Validation

Writing - review & editing

Tuyet Hoang:

Conceptualisation

Funding acquisition

Investigation

Methodology

Validation

Writing - review & editing

Marion Herisse:

Conceptualisation

Funding acquisition

Investigation

Methodology

Resources

Validation

Writing - review & editing

Sacha Pidot:

Investigation

Methodology

Resources

Writing - review & editing

Deborah Williamson:

Conceptualisation

Formal analysis

Funding acquisition

Investigation

Methodology

Writing - review & editing

Benjamin Howden:

Conceptualisation

Formal analysis

Funding acquisition

Investigation

Methodology

Project administration

Resources

Writing - original draft

Writing - review & editing

Ian Monk:

Conceptualisation

Data curation

Formal analysis

Funding acquisition

Investigation

Methodology

Project administration

Resources

Validation

Writing - original draft

Writing - review & editing

Timothy Stinear:

Conceptualisation

Data curation

Formal analysis

Funding acquisition

Investigation

Methodology

Project administration

Resources

Software

Supervision

Validation

Visualisation

Writing - original draft

Writing - review & editing

8.2   Conflicts of interest

The authors declare the following conflict of interest: Nickala Best, Sean McDonald, and Arran Greenhalgh are employees of GeneWorks, a commercial entity that distributes OptiGene reagents in Australia.

8.3   Funding information

DAW is supported by an Emerging Leadership Investigator Grant from the National Health and Medical Research Council (NHMRC) of Australia (APP1174555). TPS is supported by NHMRC Research Fellowship (APP1105525). BPH is supported by NHMRC Practitioner Fellowship (APP1105905).

8.4   Ethical approval

This study was conducted in accordance with the *National Health and Medical Research Council of Australia National Statement for Ethical Conduct in Human Research* 2007 (Updated 2018). The study was exempt from requiring specific approvals, as it involved the use of existing collections of data or records that contained non-identifiable data about human beings [37].

16737403-file00.docx

9. References

1. Victorian Department Health and Human Services 2020. Coronavirus disease 2019 (COVID-19) Guidelines for health services and general practitioners - Version 17, 5/04/2020. State Government of Victoria, Australia. https://www.dhhs.vic.gov.au/health-services-and-general-practitioners-coronavirus-disease-covid-19 (accessed 7 April 2020)

2. UK National Health Service 2020. Guidance and standard operating procedure: COVID-19 virus testing in NHS laboratories, 16/03/2020. NHS, London. https://www.england.nhs.uk/coronavirus/publication/guidance-and-standard-operating-procedure-covid-19-virus-testing-in-nhs-laboratories/ (accessed 7 April 2020)

3. Center for Disease Control. 2020. A 2019-Novel Coronavirus (2019-nCoV) real-time RT-PCR panel primers and probes, 24/01/2020. US Department of Health & Human Services, CDC, Atlanta. https://www.cdc.gov/coronavirus/2019-ncov/downloads/rt-pcr-panel-primer-probes.pdf (accessed 7 April 2020)

4. Centre for Disease Control. 2020. Real-time RT-PCR panel for detection 2019-Novel Coronavirus, 24/01/2020. US Department of Health & Human Services, CDC, Atlanta. https://www.cdc.gov/coronavirus/2019-ncov/downloads/rt-pcr-panel-for-detection- instructions.pdf (accessed 7 April 2020).

5. Food and Drug Administration. 2020. Policy for diagnostic tests for Coronavirus disease – 2019 during the public health emergency – Immediately in effect guidance for clinical laboratories, commercial manufacturers, and food and Drug administration staff, 16/03/2020. US Department of Health & Human Services, Food and Drug

Administration. https://www.fda.gov/regulatory-information/search-fda-guidance-
documents/policy-diagnostic-tests-coronavirus-disease-2019-during-public-health-
emergency (accessed 7 April 2020).

6. World Health Organization. 2020. Laboratory testing for coronavirus disease 2019
(COVID-19) in suspected human cases, 2/03/2020. World Health Organisation,
Geneva. https://apps.who.int/iris/bitstream/handle/10665/331329/WHO-COVID-19-
laboratory-2020.4-eng.pdf?sequence=1&isAllowed=y (accessed 7 April 2020).

7. World Health Organization. 2020. Coronavirus disease (COVID-19) technical
guidance: Laboratory testing for 2019-nCoV in humans. 3. Molecular assays to
diagnose COVID-19, 03/2020. World Health Organisation, Geneva.
https://www.who.int/emergencies/diseases/novel-coronavirus-2019/technical-
guidance/laboratory-guidance (accessed 7 April 2020)

8. Kelly P. 2020. Deputy Chief Medical Officer's press conference about COVID-19 on 16
March. Australian Government. https://www.health.gov.au/news/deputy-chief-medical-
officers-press-conference-about-covid-19-on-16-march (accessed 29 April 2020).

9. Arnold C. COVID-19: Biomedical research in a world under social-distancing
measures. *Nat Med* 2020. doi:10.1038/d41591-020-00005-1.

10. Anonymous. Open for outbreaks. *Nat Biotechnol* 2020; 38:377. doi:10.1038/s41587-020-
0499-y

11. Notomi T, Okayama H, Masubuchi H, Yonekawa T, Watanabe K *et al*. Loop-mediated
isothermal amplification of DNA. *Nucleic Acids Res* 2000; 28:E63. doi:10.1093/nar/28.12.e63

12. Hong TC, Mai QL, Cuong DV, Parida M, Minekawa H *et al*. Development and evaluation of
a novel loop-mediated isothermal amplification method for rapid detection of severe acute

respiratory syndrome coronavirus. *J Clin Microbiol* 2004; 42:1956-61. doi:
10.1128/jcm.42.5.1956-1961.2004

13. Iwamoto T, Sonobe T, Hayashi K. Loop-mediated isothermal amplification for direct
detection of *Mycobacterium tuberculosis* complex, *M. avium*, and *M. intracellulare* in sputum
samples. *J Clin Microbiol* 2003; 41:2616-22. doi: 10.1128/JCM.41.6.2616-2622.2003

14. Mori Y, Kanda H, Notomi T. Loop-mediated isothermal amplification (LAMP): recent
progress in research and development. *J Infect Chemother* 2013; 19:404-11. doi:
10.1007/s10156-013-0590-0

15. Poon LL, Wong BW, Ma EH, Chan KH, Chow LM *et al*. Sensitive and inexpensive
molecular test for falciparum malaria: detecting *Plasmodium falciparu*m DNA directly from
heat-treated blood by loop-mediated isothermal amplification. *Clin Chem* 2006; 52:303-6.
doi: 10.1373/clinchem.2005.057901

16. Tsujimoto M, Nakabayashi K, Yoshidome K, Kaneko T, Iwase T *et al*. One-step nucleic acid
amplification for intraoperative detection of lymph node metastasis in breast cancer patients.
Clin Cancer Res 2007; 13:4807-16. doi: 10.1158/1078-0432.CCR-06-2512

17. Zhang Y, Odiwuor N, Xiong J, Sun L, Nyaruaba RO et al. Rapid Molecular Detection
of SARS-CoV-2 (COVID-19) Virus RNA Using Colorimetric LAMP. *medRxiv* 2020.
doi: 10.1101/2020.02.26.20028373.

18. Yu L, Wu S, Hao X, Li X, Liu X *et al*. Rapid colorimetric detection of COVID-19
coronavirus using a reverse tran-scriptional loop-mediated isothermal amplification
(RT-LAMP) diagnostic plat-form: iLACO. *MedRxiv* 2020.
doi:10.1101/2020.02.20.20025874.

19.   Yan C, Cui J, Huang L, Du B, Chen L *et al*. Rapid and visual detection of 2019 novel

715   coronavirus (SARS-CoV-2) by a reverse transcription loop-mediated isothermal

716   amplification assay. *Clin Microbiol Infect* 2020. doi:10.1016/j.cmi.2020.04.001.

717   20.   Park G-S, Ku K, Baek S-H, Kim S-J, Kim S-I *et* al. Development of Reverse

718   Transcription Loop-mediated Isothermal Amplification (RT-LAMP) Assays

719   Targeting SARS-CoV-2. *J Mol Diagnostics* 2020. doi: 10.1016/j.jmoldx.2020.03.006.

720   21.   Pang J, Wang MX, Ang IYH, Tan SHX, Lewis RF *et al*.Potential Rapid Diagnostics, Vaccine

721   and Therapeutics for 2019 Novel Coronavirus (2019-nCoV): A Systematic Review. *J Clin

722   Med* 2020; *9:*623. doi: 10.3390/jcm9030623

723   22.   Osterdahl MF, Lee KA, Lochlainn ML, Wilson S, Douthwaite S *et al*. Detecting

724   SARS-CoV-2 at point of care: Preliminary data comparing Loop-mediated isothermal

725   amplification (LAMP) to PCR. *MedRxiv*. 2020. doi: 10.1101/2020.04.01.20047357.

726   23.   Lu R, Wu X, Wan Z, Li Y, Zuo L *et al*. Development of a Novel Reverse Transcription Loop-

727   Mediated Isothermal Amplification Method for Rapid Detection of SARS-CoV-2. Virol Sin

728   2020. doi: 10.1007/s12250-020-00218-1

729   24.   Broughton JP, Deng X, Yu G, Fasching CL, Servellita V *et al*. CRISPR-Cas12-based

730   detection of SARS-CoV-2. Nat Biotechnol 2020. doi:10.1038/s41587-020-0513-4.

731   25.   Caly L, Druce J, Roberts J, Bond K, Tran T *et al*. Isolation and rapid sharing of the

732   2019 novel coronavirus (SARS-CoV-2) from the first patient diagnosed with COVID-

733   19 in Australia. *Med J Aust* 2020. doi:10.5694/mja2.50569.

734   26.   He H, Li R, Chen Y, Pan P, Tong W *et al*. Integrated DNA and RNA extraction using

735   magnetic beads from viral pathogens causing acute respiratory infections. *Sci Rep* 2017;

736   7:45199. doi: 10.1038/srep45199

737   27.   Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A *et al*. Detection of 2019 novel

738   coronavirus (2019-nCoV) by real-time RT-PCR. Euro Surveill 25. doi: 10.2807/1560-

739   7917.ES.2020.25.3.2000045

740   28.   Baek YH, Um J, Antigua KJC, Park JH, Kim Y *et* al. Development of a reverse

741   transcription-loop-mediated isothermal amplification as a rapid early-detection

742   method   for   novel   SARS-CoV-2.   2020.   Emerg   Microbes   Infect

743   doi:10.1080/22221751.2020.1756698:1-31.

744   29.   Nguyen T, Duong Bang D, Wolff A. 2019 Novel Coronavirus Disease (COVID-19): Paving

745   the Road for Rapid Detection and Point-of-Care Diagnostics. Micromachines (Basel) 2020;

746   11(3), 306. doi:10.3390/mi11030306.

747   30.   FIND. 2020. SARS-CoV-2 diagnostic pipeline, 2020 https://www.finddx.org/covid-

748   19/pipeline/ (accessed 29 April 2020).

749   31.   Jayawardena S, Cheung CY, Barr I, Chan KH, Chen H *et al*. Loop-mediated isothermal

750   amplification for influenza A (H5N1) virus. *Emerg Infect Dis* 2007; 13:899-901. doi:

751   10.3201/eid1306.061572

752   32.   Lopez-Jimena B, Wehner S, Harold G, Bakheit M, Frischmann S *et al*. Development of a

753   single-tube one-step RT-LAMP assay to detect the Chikungunya virus genome. *PLoS Negl*

754   *Trop Dis* 2018; 12:e0006448. doi: 10.1371/journal.pntd.0006448

755   33.   Silva S, Paiva MHS, Guedes DRD, Krokovsky L, Melo FL *et al*. Development and Validation

756   of Reverse Transcription Loop-Mediated Isothermal Amplification (RT-LAMP) for Rapid

757   Detection of ZIKV in Mosquito Samples from Brazil. *Sci Rep* 2019; 9:4494. doi:

758   10.1038/s41598-019-40960-5

34.   Bhadra S, Riedel TE, Lakhotia S, Tran ND, Ellington AD. High-surety isothermal

amplification and detection of SARS-CoV-2, including with crude enzymes. *BioRxiv*

2020. doi: https://doi.org/10.1101/2020.04.13.039941.

35.   Moore C, Corden S, Sinha J, Jones R. Dry cotton or flocked respiratory swabs as a simple

collection technique for the molecular detection of respiratory viruses using real-time

NASBA. *J Virol Meth* 2008; 153:84-9. doi: 10.1016/j.jviromet.2008.08.001

36.   Kim EM, Jeon HS, Kim JJ, Shin YK, Lee YJ *et al*. Uracil-DNA glycosylase-treated reverse

transcription loop-mediated isothermal amplification for rapid detection of avian influenza

virus preventing carry-over contamination. J Vet Sci 2016; 17:421-5. doi:

10.4142/jvs.2016.17.3.421

37.   National Health and Medical Research Council. 2018. National Statement on Ethical

Conduct in Human Research 2007 (Updated 2018). The National Health and Medical

Research Council, the Australian Research Council and Universities Australia.

Commonwealth   of   Australia,   Canberra.   https://nhmrc.gov.au/about-

us/publications/national-statement-ethical-conduct-human-research-2007-updated-

2018.

10. Figures and tables

Fig. 1. Limit of detection of N1 STOP-LAMP and establishing optimal sample template volume and

inactivation conditions. (A) Plot showing performance of the N1 LAMP assay across 5-log$_{10}$ dilution

of purified viral RNA. Y-axis is time-to-positive (Tp) and x-axis is estimate of viral genomes/reaction

based on starting RNA concentration. The number of replicates per dilution (n) and number of

positive replicates per dilution is indicated. (B) Example N1-STOP-LAMP amplification plots for

SARS-CoV-2 positive clinical sample N^{o}. 52, showing inhibitory impact of 5 μL of a neat sample

matrix on LAMP and the effect of diluting the UTM in water. (C) N1-STOP-LAMP amplification

plots for three SARS-CoV-2 positive clinical samples, showing no loss in detection sensitivity after

specimen heat-treatment of 60°C for 30 min to inactivate virus.

Fig. 2: Comparison of E-gene RT-qPCR with N1-STOP-LAMP and limit of detection. Titred virus

was serially diluted 10-fold in UTM in triplicate and RNA was extracted from each replicate dilution.

A 1 µL aliquot of each UTM dilution was also tested directly by N1-STOP-LAMP. (A) Shows

calibration curve for the E-gene RT-qPCR. A curve was interpolated using linear regression. (B)

792   Comparative performance of N1-STOP-LAMP with 5 µL purified RNA versus 1 µL or 5 µL of UTM

793   directly added to the reaction.

Fig. 3: Clinical evaluation of N1-STOP-LAMP. (A) Establishing limit of detection according to FDA Emergency Use Authorisations (EUA) guidelines. Twenty, 1 µL replicates of SARS-CoV-2 virus in Universal Transport Medium (UTM) at a concentration of 54 $TCID_{50}$ mL^{-1} were tested by direct N1-STOP-LAMP. All data points plotted, with average and 95% CI indicated. (B) Assessment of assay against FDA EUA clinical performance criteria. Thirty healthy volunteer dry nasopharyngeal swabs eluted in 1.5 mL of PBS were screened by N1-STOP-LAMP. Shown are time-to-positives (Tp). Mean and 95% CI are indicated. (C). Plot showing correspondence between 107 E-gene RT-qPCR positive and N1-STOP-LAMP, where 1 µL of clinical specimen in UTM was added directly to the RT-LAMP reaction. Y-axis is LAMP Tp and x-axis is RT-qPCR cycle threshold (Ct). The dotted line indicates Tp 30 mins. Results plotted above this line (encircled) were RT-qPCR positive but N1-STOP-LAMP negative. (D) Results of 4-way interlaboratory comparison of N1-STOP-LAMP. Ten clinical samples

806   previously tested positive by E-gene RT-qPCR were used to create a test panel for comparing assay

807   performance between laboratories.

808 **Table 1.** FDA validation study recommendations for SARS-CoV-2 molecular diagnostics [5].

Method Validation Parameter	Validation description and acceptance criteria
1. Limit of Detection (LoD)	Test a dilution series of three replicates per concentration. LoD defined as lowest concentration at which 19/20 replicates positive.
2. Clinical evaluation	Test at least 30 contrived reactive specimens and 30 non-reactive specimens. At least 20 specimens should have virus/RNA concentration of 1-2x LoD, remainder spanning the assay range. Acceptance defined as 95% agreement 1-2x LoD, and 100% agreement at other concentrations and for negative specimens.
3. Inclusivity	*In silico* analysis indicating percent identity matches of target region against publicly available SARS-CoV-2 sequences. Acceptance defined as 100% match against all entries with the selected primers/probes.
4. Cross-reactivity	*In silico* analysis of the assay primers/probes compared to respiratory microbiota and other viral pathogens. Acceptance defined as <80% homology between all primers/probes and non-target sequences present in public databases.

810 **Table 2.** Assessment of heat treatment of universal transport medium specimens on detection

811 sensitivity.

Sample N^{o}	E-gene RT-qPCR Ct	Treatment	
		No heat Tp (min:sec)	Heat Tp (min:sec)
50	21.76	10:15	10:45
51	24.40	17:45	15:45
52	15.05	07:15	07:45

812   Ct = cycle threshold; Tp = time-to-positive

814 **Table 3.** Comparison of N1-STOP-LAMP and RT-qPCR performance using titred virus diluted in

815 universal transport medium.

SARS-CoV-2 (TCID$_{50}$ mL^{-1})	N	N1-LAMP (1 µL direct)		N1-LAMP (5 µL direct)		N1-LAMP (5 µL of 60 µL)		RT-qPCR (5 µL of 60 µL)	
		Tp (Avg)	CV (%)	Tp (Avg)	CV (%)	Tp (Avg)	CV (%)	Ct (Avg)	CV (%)
54000	3	06.33	02.28	06.41	02.25	05.33	02.71	18.41	0.33
5400	3	07.42	01.95	07.58	05.04	06.33	06.03	21.40	3.75
540	3	08.92	11.33	09.25	02.70	07.50	03.33	24.77	1.03
54	3	15.75	12.60	17.67	48.49	10.67	02.71	27.94	1.13
5.4	3	nd	-	16.25	08.70[#]	16.88	07.33[#]	31.10	1.17
0.54	3	nd	-	nd	nd	nd	-	36.96	2.79[#]

816 Tp = time-to-positive; CV = Coefficient of variation; nd = not detected; [#]2/3 replicates positive at this dilution.

818 **Table 4.** Summary of clinical evaluation comparisons against FDA criteria.

	Negative	1-2x LoD Tp (min:sec)	2x LoD Tp (min:sec)
N1-STOP-LAMP +	0	20 (07:30 – 22:00)[#]	10 (08:30 – 21:00)[#]
N1-STOP-LAMP -	30	0	0

819 LoD = Limit of detection; Tp = time-to-positive; [#]range indicated within parentheses.

821   **Table 5.** Comparison of N1-STOP-LAMP versus E-gene RT-qPCR using clinical specimens.

	RT-qPCR +	RT-qPCR -	Total
N1-STOP-LAMP +	93	0	93
N1-STOP-LAMP -	14	50	64
Total	107	50	157

Supplementary data

Table S1. Clinical data positivity summary (107 E-gene RT-qPCR positives vs N1-STOP-LAMP Tp).

Sample	E-gene RT-qPCR Ct	N1-STOP-LAMP Tp (min:sec)	N1-STOP-LAMP retest of RNA
1	23.0	17:45	
2	21.0	08:05	
3	23.0	09:45	
4	26.0	28:00	
5	nd	nd	
6	21.2	11:00	
7	17.8	10:00	
8	16.3	08:15	
9	19.3	08:45	
10	18.6	08:30	
11	25.1	20:00	
12	22.0	13:45	
13	28.1	14:00	
14	23.2	11:00	
15	24.3	20:15	
16	28.3	29:15	
17	23.1	17:30	
18	26.1	28:15	
19	16.4	07:45	
20	20.9	11:45	

21	27.2	16:45		
22	20.1	10:15		
23	20.4	11:00		
24	20.0	08:45		
25	18.9	09:00		
26	21.6	11:45		
27	27.1	27:00		
28	21.9	16:00		
29	17.0	09:45		
30	15.4	09:00		
31	19.3	17:15		
32	20.6	11:15		
33	15.1	09:30		
34	27.6	20:30		
35	28.5	21:45		
36	16.3	09:15		
37	16.5	09:45		
38	23.1	22:30		
39	31.5	nd		
40	30.5	nd		
41	25.9	23:15		
42	21.3	13:30		
43	30.6	29:15		
44	25.4	15:45		
45	22.8	13:15		
46	27.2	29:15		
47	21.7	08:30		

48	16.5	07:15
49	20.4	08:30
50	21.8	10:15
51	24.4	12:00
52	15.1	08:00
53	28.8	27:00
54	18.2	07:45
55	17.2	06:00
56	26.7	09:00
57	23.9	11:30
58	16.2	06:30
59	17.8	07:15
60	27.8	24:15
61	27.2	11:45
62	35.8	nd
63	22.8	11:45
64	17.0	08:00
65	20.0	16:45
66	19.5	08:45
67	24.8	12:15
68	29.2	28:30
69	32.3	nd
70	26.6	20:30
71	33.1	28:00
72	20.0	07:30
73	18.6	08:30
74	17.5	07:50

16737403-file00.docx

75	30.7	25:30	
76	20.9	13:00	
77	19.0	07:15	
78	26.3	11:45	
79	22.6	08:30	
80	29.3	nd	
81	20.9	08:30	
82	21.5	14:15	
83	20.0	07:45	
84	26.6	nd	
85	28.5	nd	
86	28.2	26:15	
87	30.3	nd	
88	27.6	13:15	
89	20.0	09:15	
90	27.4	27:45	
91	18.7	07:15	
92	30.0	20.30	
93	28.0	nd	10:00
94	33.2	nd	nd
95	19.7	08:45	
96	28.2	nd	17:45
97	34.7	09:00	
98	25.8	nd	25:30
99	21.3	12:15	
100	16.5	07:45	
101	33.4	25:15	

102	25.1	19:00	
103	20.3	09:45	
104	26.8	nd	12:25
105	25.1	16:00	
106	22.4	12:30	
107	21.4	08:25	

825 Ct = cycle threshold; Tp = time-to-positive

827   **Supplementary Table S2.** Results of N1-STOP-LAMP external quality assessment among four

828   laboratories.

Sample	E-gene RT-qPCR Ct	N1-STOP-LAMP Tp				Agreement with RT-qPCR
		Lab A	Lab B	Lab C	Lab D	
1	16.3	06.31	08:00	05:12	04:48	4/4
2	nd	nd	nd	19:54	nd	3/4
3	16.5	06.54	04:26	06:06	04:54	4/4
4	nd	nd	nd	nd	nd	4/4
5	20.4	08.51	08:44	08:12	07:24	4/4
6	nd	nd	nd	nd	nd	4/4
7	21.7	09.30	05:58	07:12	06:54	4/4
8	nd	nd	nd	nd	nd	4/4
9	22.1	12.3	12:06	11:36	07:00	4/4
10	nd	nd	nd	nd	nd	4/4
11	22.8	08.45	05:18	07:24	07:54	4/4
12	25.4	16.45	16:16	09:30	07:06	4/4
13	25.9	24.30	26:21	20:48	25:12	4/4
14	27.2	nd	24:32	nd	nd	1/4
15	27.6	29.15	25:18	23:18	28:06	4/4
16	28.5	29.15	nd	nd	nd	1/4
17	29	nd	nd	nd	nd	0/4
18	30.6	29.15	23:04	nd	24:00	3/4
19	nd	nd	nd	nd	nd	4/4
20	nd	nd	nd	nd	nd	4/4

16737403-file00.docx

829   Ct = cycle threshold; Tp = time-to-positive

830   **Supplementary Table S3.** N1-STOP-LAMP screen of NATtrol™ Respiratory Panel 2 (RP2)

831   Controls.

Virus/Organism	Strain	RP2 Control 1	RP2 Control 2	N1-STOP-LAMP
Adenovirus Type 1	N/A	Positive	Negative	Negative
Adenovirus Type 3	N/A	Positive	Negative	Negative
Adenovirus Type 31	N/A	Positive	Negative	Negative
C. pneumoniae	CWL-029	Positive	Negative	Negative
Influenza A 2009 H1N1pdm	A/NY/02/2009	Positive	Negative	Negative
Influenza A 2009 H3N2	A/Brisbane/10/07	Positive	Negative	Negative
Human Metapneumovirus Type 8	Peru6-2003	Positive	Negative	Negative
M. pneumoniae	M129	Positive	Negative	Negative
Parainfluenza Type 1	N/A	Positive	Negative	Negative
Parainfluenza Type 4	N/A	Positive	Negative	Negative
Rhinovirus Type 1A	N/A	Positive	Negative	Negative
B. parapertussis	A747	Negative	Positive	Negative
B. pertussis	A639	Negative	Positive	Negative
Coronavirus 229E	N/A	Negative	Positive	Negative
Coronavirus HKU-1	Recombinant	Negative	Positive	Negative
Coronavirus NL63	N/A	Negative	Positive	Negative
Coronavirus OC43	N/A	Negative	Positive	Negative
Influenza A H1N1	A/New Cal/20/99	Negative	Positive	Negative
Influenza B	B/Florida/02/06	Negative	Positive	Negative

Parainfluenza Type 2	N/A	Negative	Positive	Negative
Parainfluenza Type 3	N/A	Negative	Positive	Negative
RSV Type A	2006 isolate	Negative	Positive	Negative

833   **Supplementary data file 1.** Global initiative on sharing all influenza data (GISAID) submitters table.