Full-length 16S rRNA gene amplicon analysis of human gut microbiota using

MinION™ nanopore sequencing confers species-level resolution

Yoshiyuki Matsuo[1*], Shinnosuke Komiya[2,3], Yoshiaki Yasumizu[4,5], Yuki Yasuoka[5],

Katsura Mizushima[6], Tomohisa Takagi[6], Kirill Kryukov[7,8], Aisaku Fukuda[9], Yoshiharu

Morimoto[2], Yuji Naito[6], Hidetaka Okada[3], Hidemasa Bono[10,11], So Nakagawa[7] and

Kiichi Hirota[1]

[1] Department of Human Stress Response Science, Institute of Biomedical Science,

Kansai Medical University, Hirakata, Japan.

[2] HORAC Grand Front Osaka Clinic, Osaka, Japan.

[3] Obstetrics and Gynecology, Kansai Medical University Graduate School of Medicine,

Hirakata, Japan.

[4] Department of Experimental Immunology, Immunology Frontier Research Center,

Osaka University, Osaka, Japan.

[5] Faculty of Medicine, Osaka University, Osaka, Japan.

[6] Molecular Gastroenterology and Hepatology, Kyoto Prefectural University of

Medicine, Kyoto, Japan.

[7] Department of Molecular Life Science, Tokai University School of Medicine, Isehara,

Japan.

[8] Department of Genomics and Evolutionary Biology, National Institute of Genetics,

Mishima, Japan.

[9] IVF Osaka Clinic, Osaka, Japan.

[10] Database Center for Life Science (DBCLS), Research Organization of Information

and Systems, Mishima, Japan.

[11] Program of Biomedical Science, Graduate School of Integrated Sciences for Life,

Hiroshima University, Higashi-Hiroshima, Japan.

* Correspondence:

Yoshiyuki Matsuo

ysmatsuo-kyt@umin.ac.jp

[1]Department of Human Stress Response Science, Institute of Biomedical Science,

Kansai Medical University, 2-5-1 Shin-machi, Hirakata, Osaka 573-1010, Japan.

Abstract

Background: Species-level genetic characterization of complex bacterial communities has important clinical applications in both diagnosis and treatment. Amplicon sequencing of the 16S ribosomal RNA (rRNA) gene has proven to be a powerful strategy for the taxonomic classification of bacteria. This study aims to improve the method for full-length 16S rRNA gene analysis using the nanopore long-read sequencer MinION™. We compared it to the conventional short-read sequencing method in both a mock bacterial community and human fecal samples.

Results: We modified our existing protocol for full-length 16S amplicon sequencing by MinION™. A new strategy for library construction with an optimized primer set overcame PCR-associated bias and enabled taxonomic classification across a broad range of bacterial species. We compared the performance of full-length and short-read 16S amplicon sequencing for the characterization of human gut microbiota with a complex bacterial composition. The relative abundance of dominant bacterial genera was highly similar between full-length and short-read sequencing. At the species level, MinION™ long-read sequencing had better resolution for discriminating between members of particular taxa such as *Bifidobacterium*, allowing an accurate representation of the sample bacterial composition.

Conclusions: Our present microbiome study, comparing the discriminatory power of full-length and short-read sequencing, clearly illustrated the analytical advantage of sequencing the full-length 16S rRNA gene, which provided the requisite species-level resolution and accuracy in clinical settings.

Keywords:

16S rRNA, gut microbiota, MinION™, nanopore sequencing

Background

Recent advances in DNA sequencing technology have had a revolutionary impact on clinical microbiology [1]. Next-generation sequencing (NGS) technology enables parallel sequencing of DNA on a massive scale to generate vast quantities of accurate data. NGS platforms are now increasingly used in the field of clinical research [2]. Metagenomic sequencing offers numerous advantages over traditional culture-based techniques that have long been the standard test for detecting pathogenic bacteria. This method is particularly useful for characterizing uncultured bacteria and novel pathogens [3].

Among sequence-based bacterial analyses, amplicon sequencing of the 16S ribosomal RNA (rRNA) gene has proven to be a reliable and efficient option for taxonomic classification [4, 5]. The bacterial 16S rRNA gene contains nine variable regions (V1 to V9) that are separated by highly conserved sequences across different taxa. For bacterial identification, the 16S rRNA gene is first amplified by polymerase chain reaction (PCR) with primers annealing to conserved regions and then sequenced. The sequencing data are subjected to bioinformatic analysis in which the variable regions are used to discriminate between bacterial taxa [6].

Since the conventional parallel-type short-read sequencer cannot yield reads covering the full length of the 16S rRNA gene [7], several regions of it have been targeted for sequencing, which often causes ambiguity in taxonomic classification [8]. New sequencing platforms have overcome these technical restrictions, particularly those affecting read length. A prime example is the MinION™ sequencer from Oxford Nanopore Technologies, which is capable of producing long sequences with no theoretical read length limit [9-11]. MinION™ sequencing targets the entire 16S rRNA gene, allowing the identification of bacteria with more accuracy and sensitivity [12, 13]. Furthermore, MinION™ produces sequencing data in real time, which reduces turnaround time for data processing [14, 15].

Given these features of MinION™ sequencing, we had previously conducted full-length

16S amplicon sequencing analyses using the MinION™ platform coupled to a

bioinformatics pipeline, which allowed us to identify bacterial pathogens with a total

analysis time of under two hours [16]. However, we also found that the approach of

using the commercial 16S Barcoding Kit (SQK-RAB204) available from Oxford

Nanopore Technologies has a limited ability to detect particular taxa such as

Bifidobacterium [16]. This is probably due to sequence mismatches in the primer used

for 16S gene amplification [17]. Deviations or aberrancies in the *Bifidobacterium*

composition in the human gut have been reported in several diseases including obesity,

allergy, and inflammatory disorders [18]. Based on their putative health-promoting

effects, several strains of *Bifidobacterium* have been utilized as probiotics [19]. Within

these contexts, the species-level characterization of *Bifidobacterium* diversity in human

gut microbiota is potentially important in clinical practice.

Our 16S rRNA gene sequence analysis using MinION™ has been tested only with pre-

characterized mock bacterial DNA and a limited number of pathogenic bacteria from a

patient-derived sample [20]. Its applicability to highly complex bacterial communities

has not yet been thoroughly investigated. Therefore, in this study we modified our

existing protocol for 16S amplicon sequencing by MinION™ and applied it to human

gut microbiota with a complex bacterial composition [21], including *Bifidobacterium*, to

determine whether full-length 16S rRNA gene sequencing with MinION™ is an

effective characterization tool.

Results

Classification of the mock bacterial community

The 16S rRNA gene sequence of *Bifidobacterium* has three base mismatches with the

27F forward primer provided in the commercial sequencing kit (16S Barcoding Kit,

SQK-RAB204, Oxford Nanopore Technologies; Additional File 2: Supplementary Fig.

S1a), which biases amplification toward underrepresentation of *Bifidobacterium* species

(Additional File 2: Supplementary Fig S2, Additional File 3: Supplementary Table S1-

S3). To overcome this drawback, we introduced three degenerate bases to the 16S rRNA

gene-specific sequences of the primer (Additional File 2: Supplementary Fig. S1b). The

competence of the modified primer set was then evaluated by 16S rRNA gene sequence

analysis of a ten-species mock community. The V1-V9 region of the 16S rRNA gene

was amplified by the four-primer PCR method with the rapid adapter attachment

chemistry and sequenced (Fig. 1a). MinION™ sequencing generated 8651 pass reads

(Table 1). Following adapter trimming and size selection, 6972 reads (80.6% of pass

reads with an average lead length of 1473 bp) were retained for bacterial identification.

Full-length 16S amplicon sequencing with the modified primer set led to the successful

identification of all the ten bacterial genera, including *Bifidobacterium* (Fig. 1b,

Additional File 3: Supplementary Table S4). At the species level, 92.5% of analyzed

reads were correctly assigned to each bacterial taxon included in the mock community,

demonstrating the excellent discriminatory power of this full-length sequencing method

for species identification (Fig. 1c). The only exception was *Bacillus cereus*.

Discrimination of *Bacillus cereus* from the closely related species such as *Bacillus*

anthracis and *Bacillus thuringiensis* was not fully achieved with the 16S gene analysis

using MinION™ (Additional File 4).

We compared the resolution of full-length and short-read 16S amplicon sequencing for

the taxonomic classification of bacteria. The V3-V4 region was amplified by four-

primer PCR from the ten-species mock community DNA, and the samples were

sequenced on MinION™. After removing the adapter/barcode sequences and filtering

reads by length, 96189 reads with an average length of 454.9 bp for downstream

analysis were yielded (Table 1). In contrast to full-length sequencing with the highest

resolution, a significant number of V3-V4 reads could not be classified down to the

genus level, but could only be assigned to a higher taxonomic rank (Fig. 1b, 1c). Most

reads derived from *Enterococcus faecalis* and *Escherichia coli* were not assigned down

to each taxon, as more than two species produced the same similarity score for the V3-

V4 sequence read queries (Additional File 5). Such reads were classified to the parent

taxon, as more specific classification was impossible (Additional File 3: Supplementary

Table S5). The classifications were not affected by increasing the number of analyzed

reads to 10000 (Additional File 2: Supplementary Fig. S3, Additional File 3:

Supplementary Table S6).

For eight of the ten bacterial strains constituting the mock community, each subset of

V1-V9 sequencing reads classified to the specific genus were assigned with almost

complete accuracy (98.2-100%) to the corresponding species (Fig. 2). V3-V4 short-read

sequencing showed a discriminatory power comparable to that of V1-V9 full-length

sequencing in the classification of three genera (*Deinococcus*, *Rhodobacter*, and

Streptococcus) with more than 98% of reads correctly assigned. However, the V3-V4

region was not suitable for species-level identification of other taxa, such as *Clostridium*

and *Staphylococcus*. Only 0.2% of the V3-V4 reads belonging to the genus *Clostridium*

were assigned to *Clostridium beijerinckii*, a component of the mock community. In

contrast, 92.3% of *Clostridium* reads obtained from V1-V9 full-length sequencing were

correctly classified at the species level. These results suggest a lower resolution of the

V3-V4 region for species-level classification, emphasizing the advantage of long-read

sequencing for obtaining an accurate representation of the sample bacterial composition.

Classification of human fecal bacteria

We assessed the performance of our full-length 16S amplicon sequencing approach in

the context of a highly complex bacterial community. The V1-V9 region was amplified

by four-primer PCR from six human fecal samples (F1-F6) and analyzed by MinION™

sequencing. (Table 2). In Fig. 3, the numbers of species detected are plotted against the

numbers of reads analyzed. The curve started to plateau at around 20000 reads. There

was a highly significant correlation between the read numbers 20000 and 30000

(Pearson's correlation coefficient $r > 0.999$, Additional File 6: Supplementary Table S7).

Based on these observations, randomly sampled 20000 reads were used in further

analysis to determine the bacterial composition of the human gut.

For comparison, amplicon sequencing of the V3-V4 region was also conducted using

the MinION™ (Table 2) and the Illumina MiSeq™ platform (Table 3). The processed

reads from each data set were allocated to the reference bacterial genome using our

bioinformatics pipeline to determine the bacterial compositions (Additional File 7 for

V1-V9 MinION™ sequencing, Additional File 8 for V3-V4 MinION™ sequencing, and

Additional File 9 for V3-V4 MiSeq™ sequencing). From MiSeq™ sequencing data, the

bacterial composition was also analyzed by the operational taxonomic unit (OTU)-

based approach using the QIIME 2 (ver. 2019.7) pipeline (Additional File 2:

Supplementary Fig. S4, Additional File 10) [22, 23]. Although *Bacteroides* was

underrepresented in the OTU-based analysis, the two analytical methods (our

bioinformatics pipeline and OTU-based method) produced similar taxonomic profiles in

the dominant phylotypes for the MiSeq™ data. This result confirmed the validity of our

method for the taxonomic classification of the bacterial community.

The three sequencing methods (V1-V9 MinION™ sequencing, V3-V4 MinION™

sequencing, and V3-V4 MiSeq™ sequencing) revealed similar profiles for the six fecal

samples at the genus level (Fig. 4). Statistically significant similarities have been found

in the relative genus abundances across these sequencing methods. Thus, at the genus

level, V1-V9 full-length MinION™ sequencing exhibited a discriminatory power

comparable to that of high-quality short-read sequencing with MiSeq™ technology.

The species-level taxonomic resolution achieved by full-length sequencing of the

16S rRNA gene using MinION™

While genus classification using long versus short reads was relatively comparable, we

observed considerable differences across amplified regions in the species-level profiling

of human gut microbiota. As shown in Fig. 5, the number of ambiguous reads that were

not assigned to species but could be classified at a higher level was significantly greater

in the V3-V4 data set in comparison than in the V1-V9 data set.

When species compositions of the dominant taxa (*Bifidobacterium*, *Blautia*, and

Bacteroides) were analyzed, the V1-V9 and V3-V4 sequencing produced comparable

results for *Blautia* (Additional File 2: Supplementary Fig. S5, Additional File 6:

Supplementary Table S8) and *Bacteroides* genus (Additional File 2: Supplementary Fig.

S6, Additional File 6: Supplementary Table S9) in most of the fecal samples. For

Bifidobacterium, there appeared to be considerable deviations in the relative abundances

of some species depending on the sequencing method used (Fig. 6, Additional File 6:

Supplementary Table S10). Notably, most of the *Bifidobacterium* reads generated by

V1-V9 MinION™ sequencing were classified into the *Bifidobacterium* species that

were isolated from human sources [18, 24]. A significant number of the V3-V4 reads,

however, were assigned to *Bifidobacterium* species of non-human origin (Additional

File 2: Supplementary Fig. S7). From V3-V4 MiSeq™ sequencing data, the OTU-based

classification analysis using the QIIME 2 pipeline also revealed a lower resolution of

short-read sequencing for taxonomic separation of *Bifidobacterium* genus. Except for

Bifidobacterium longum, *Bifidobacterium* species could not be reliably identified by the

V3-V4 sequencing strategy (Additional File 2: Supplementary Fig. S8). These results

suggest that species classification of *Bifidobacterium* based on V3-V4 sequencing can

potentially lead to misidentification and biased community profiling that lacks

biological significance.

Discussion

16S rRNA amplicon sequencing is a powerful strategy for taxonomic classification of

bacteria and has been extensively employed for analyzing samples from environmental

and clinical sources [5, 25, 26]. We assessed the performance of MinION™ sequencing

by comparing the resolution of the V1-V9 and V3-V4 reads for the taxonomic

classification of bacteria. Due to the error-prone nature of MinION™ sequencing, the

existing OTU-based approach, requiring at least 97% sequence identity threshold, has

been considered unsuitable for taxonomic classification of MinION™ reads [27, 28].

Instead, the reads were analyzed by the direct read mapping method that assigns

sequences to taxonomic bins based on the similarity to a reference database [14, 15].

Long-read MinION™ sequencing with the optimized primer set successfully identified

Bifidobacterium species leading to a better representation of the species composition of

the mock community. For improving the classification results, the reads were filtered by

length to eliminate those outside the expected size range. Typically, extremely short

reads possess only one primer-binding site, suggesting that they are derived from

incomplete sequencing. There also exist unexpectedly longer reads with a continuous

sequence structure in which two 16S amplicons are linked end-to-end. Because these

reads can potentially result in unclassified reads or misclassification, they were

eliminated before alignment to the reference sequences of the bacterial genome.

We also modified library construction for MinION™ sequencing with a four-primer

PCR strategy, which enabled ligase-free adapter attachment to occur in a single-step

reaction. The four-primer PCR generates amplicons with particular chemical

modifications at the 5' ends to which adapter molecules can be attached non-

enzymatically. Unlike the ligation-based approach, the PCR products amplified by the

four-primer method are subjected directly to the adapter attachment reaction without

repairing their 5' ends, substantially reducing the time required for sample preparation.

Furthermore, because the protocol is free of Nanopore's transposase-based technology

(e.g. Rapid Sequencing Kit, SQK-RAD004) that cleaves DNA molecules to produce

chemically modified ends for library construction, the PCR products are kept intact,

enabling sequencing of the entire amplified region. Thus, the four-primer PCR-based

method allowed us to perform amplicon sequencing on the MinION™ platform with

user-defined arbitrary primer pairs, taking advantage of the rapid adapter attachment

chemistry. This method can be applied to a wide range of sequence-based analyses,

including detection of functional genetic markers like antimicrobial resistance genes

and identification of genetic variations in targeted loci [11, 29, 30].

Our present microbiome study, comparing the discriminatory power of the V1-V9 and

V3-V4 reads sequenced on the MinION™ platform, clearly illustrated the advantage of

sequencing the entire 16S rRNA gene. The full-length 16S gene sequencing provided

better resolution than short-read sequencing for discriminating between members of

certain bacterial taxa, including *Clostridium*, *Enterococcus*, *Escherichia*, and

Staphylococcus. Consistently, comprehensive *in silico* experiments using sequencing a

sequencing data set consisting of different regions of the 16S rRNA gene have shown

that the choice of the regions to be sequenced substantially affects the classification

results [6, 31]. As shown here and in previous publications, short-read sequencing of the

16S rRNA gene may be a reasonable option for providing a rough estimation of

bacterial diversity. However, it was not suited for analysis requiring species-level

resolution and accuracy, which was afforded by sequencing the entire 16S rRNA gene.

It is important to note that even full-length 16S gene analysis fails to discriminate some

closely related species such as members of *Bacillus cereus* group, indicating the

limitations of the 16S amplicon sequencing alone in species allocation. Long read

sequencing targeting other phylogenetic markers may be an alternative to 16S rRNA

amplicon sequencing and provide better resolution for bacterial identification.

In the analysis of the human fecal samples, we used the taxonomic resolution of the V3-

V4 region sequenced with MiSeq™, which generates highly accurate reads, as a

benchmark for the taxonomic resolution of the full-length 16S gene sequenced with

MinION™. The relative abundance of dominant bacterial taxa was highly similar at the

genus level between full-length MinION™ and short-read MiSeq™ sequencing.

Despite the lower read quality, the full-length sequencing by MinION™ enabled

reliable identification of bacterial genera with an accuracy comparable to MiSeq™

technology. The MiSeq™ platform enables 16S rRNA gene sequencing on a massive

scale with reduced cost (approximately 20 USD per sample). Considering a low

equipment price (1000 USD) and an affordable per-run cost (approximately 50 USD per

sample), the MinION™ sequencer could be a viable option for practical applications in

clinical microbiology.

At the species level, MinION™ long-read sequencing had better resolution for accurate

identification of the composition of human gut microbiota. Composition profiles of

Bifidobacterium, one of the dominant genera present in the human gut [32], appeared to

differ considerably between the two sequencing platforms. While most MinION™ V1-

V9 reads were assigned to *Bifidobacterium* species of human origin, a significant

number of the MiSeq™ V3-V4 reads were assigned to non-human *Bifidobacterium*

species [24]. Such improbable errors in species classification may be attributed to the

lower resolution provided by the V3-V4 region, but in fact, 16S rRNA gene sequence

analysis of mouse gut microbiota has revealed that V3-V4 reads sequenced on MiSeq™

are not well-suited for classifying *Bifidobacterium* species, consistent with our findings

[33]. Since the taxonomic assignment is a critical step for analyzing the bacterial

diversity and community composition, the reference database quality and the alignment

algorithm must be further evaluated for each sequencing data set.

Conclusions

Our modified protocol for 16S amplicon sequencing overcame known limitations, such

as the primer-associated bias toward the underrepresentation of *Bifidobacterium*, and

enabled taxonomic classification across a broad range of bacterial species.

Benchmarking with MiSeq™ sequencing technology demonstrated the analytical

advantage of sequencing the full-length 16S gene with MinION™, which provided the

requisite species-level resolution and accuracy. With the recent progress in nanopore

sequencing chemistry and base-calling algorithms, sequencing accuracy is continuously

improving [34, 35]. This will soon enable us to exploit the full potential of MinION™

long-read sequencing technology. High-quality long sequences will allow better

discrimination between closely related species, and even bacterial strains, in sequence-

based bacterial analyses.

Methods

Mock bacterial community DNA

A mixture of bacterial DNA (10 Strain Even Mix Genomic Material, MSA-1000) was

obtained from the American Type Culture Collection (ATCC, Manassas, VA, USA),

comprising genomic DNA prepared from the following ten bacterial strains: *Bacillus*

cereus (ATCC 10987), *Bifidobacterium adolescentis* (ATCC 15703), *Clostridium*

beijerinckii (ATCC 35702), *Deinococcus radiodurans* (ATCC BAA-816), *Enterococcus*

faecalis (ATCC 47077), *Escherichia coli* (ATCC 700926), *Lactobacillus gasseri* (ATCC

33323), *Rhodobacter sphaeroides* (ATCC 17029), *Staphylococcus epidermidis* (ATCC

12228), and *Streptococcus mutans* (ATCC 700610).

Fecal DNA

DNA was extracted from six human fecal samples using the NucleoSpin® Microbial

DNA Kit (Macherey-Nagel, Düren, Germany), as described previously [36]. Briefly,

human feces stored using the Feces Collection Kit (Techno Suruga Lab, Shizuoka,

Japan) were subjected to mechanical disruption by bead-beating, and DNA was isolated

using silica membrane spin columns. Extracted DNA was purified with the Agencourt

AMPure® XP (Beckman Coulter, Brea, CA, USA).

16S rRNA gene sequencing on the MinION™ platform

Four-primer PCR with rapid adapter attachment chemistry generated 16S gene

amplicons with modified 5′ ends for simplified post-PCR adapter attachment following

the manufacturer's instructions with slight modifications. For amplification of the V1-

V9 region of the 16S rRNA gene, the following inner primers were used, with 16S

rRNA gene-specific sequences underlined: forward primer (S-D-Bact-0008-c-S-20 [37])

with anchor sequence 5′-

TTTCTGTTGGTGCTGATATTGC<u>AGRGTTYGATYMTGGCTCAG</u>-3′ and reverse

primer (1492R) with anchor sequence 5′-

ACTTGCCTGTCGCTCTATCTTCCGGYTACCTTGTTACGACTT-3′. For

amplification of the V3-V4 region, the following inner primers were used, with 16S

rRNA gene-specific sequences underlined: 341F with anchor sequence 5′-

TTTCTGTTGGTGCTGATATTGCCCTACGGGNGGCWGCAG-3′ and 806R with

anchor sequence 5′-

ACTTGCCTGTCGCTCTATCTTCGGACTACHVGGGTWTCTAAT-3′. PCR

amplification of 16S rRNA genes was conducted using the KAPA2G™ Robust HotStart

ReadyMix PCR Kit (Kapa Biosystems, Wilmington, MA, USA) in a total volume of 25

µl containing inner primer pairs (50 nM each) and the barcoded outer primer mixture

(3%) from the PCR Barcoding Kit (SQK-PBK004; Oxford Nanopore Technologies,

Oxford, UK). Amplification was performed with the following PCR conditions: initial

denaturation at 95 °C for 3 min, 5 cycles of 95 °C for 15 sec, 55 °C for 15 sec, and

72 °C for 30 sec, 30 cycles of 95 °C for 15 sec, 62 °C for 15 sec, and 72 °C for 30 sec,

followed by a final extension at 72 °C for 1 min. Amplified DNA was purified using

AMPure® XP (Beckman Coulter) and quantified by a NanoDrop® 1000 (Thermo

Fischer Scientific, Waltham, MA, USA). A total of 100 ng of DNA was incubated with

1 µl of Rapid Adapter at room temperature for 5 min. The prepared DNA library (11 µl)

was mixed with 34 µl of Sequencing Buffer, 25.5 µl of Loading Beads, and 4.5 µl of

water, loaded onto the R9.4 flow cell (FLO-MIN106; Oxford Nanopore Technologies),

and sequenced on the MinION™ Mk1B. MINKNOW software ver. 1.11.5 (Oxford

Nanopore Technologies) was used for data acquisition.

16S rRNA gene sequencing on the MiSeq™ platform

Sequencing libraries were constructed as described previously [36]. Briefly, the V3-V4

regions of the 16S rRNA gene were amplified using a 16S (V3–V4) Metagenomic

Library Construction Kit for NGS (Takara Bio Inc, Kusatsu, Japan). The following

primers were used (16S rRNA gene-specific sequences are underlined): 341F with

overhang adapter 5′-

TCGTCGGCAGCGTCAGATGTGTATAAGAGACAG<u>CCTACGGGNGGCWGCAG</u>-

3′ and 806R with overhang adapter 5′-

GTCTCGTGGGCTCGGAGATGTGTATAAGAGACAG<u>GGACTACHVGGGTWTCT</u>

<u>AAT</u>-3′. The second PCR was performed using the Nextera® XT Index Kit (Illumina,

San Diego, CA, USA) for sample multiplexing with index adapters. The libraries were

sequenced on the MiSeq™ platform using the MiSeq™ Reagent Kit v3 (2 × 250 bp;

Illumina).

Bioinformatics analysis

Albacore software ver. 2.3.4 (Oxford Nanopore Technologies) was used for basecalling

the MinION™ sequencing data (FAST5 files) to generate pass reads (FASTQ format)

with a mean quality score > 7. The adapter and barcode sequences were trimmed using

the EPI2ME Fastq Barcoding workflow ver. 3.10.4 (Oxford Nanopore Technologies).

The reads were filtered by size using SeqKit software ver. 0.10.0 [38], retaining 1300-

1950 bp sequences for the V1-V9 region and 350-600 bp sequences for the V3-V4

region, based on the size distribution of 16S rRNA gene sequences in the SILVA

database ver. 132 [39, 40]. The average Phred quality score was assessed using

NanoPlot ver. 1.27.0 [41]. The processed reads from each set were analyzed using our

bioinformatics pipeline [42], as described previously [14, 15]. Briefly, FASTQ files

were converted to FASTA files. Simple repetitive sequences were masked using the

TANTAN program ver. 18 with default parameters [43]. To remove reads derived from

human DNA, we searched each read against the human genome (GRCh38) using

minimap2 ver. 2.14 with a map-ont-option [44]. Then, unmatched reads were regarded

as reads derived from bacteria. For each read, a minimap2 search with 5850

representative bacterial genome sequences stored in the GenomeSync database

(Additional File 1) [45] was performed. For each read, the species showing the highest

minimap2 score were assigned to the query sequence. When more than two species

showed the same similarity score, the reads were classified at any higher taxonomic

rank covering all the identified species. Taxa were determined based on the NCBI

taxonomy database [46]. Low-abundance taxa with less than 0.01% of total reads were

discarded from the analysis.

Statistical analyses

Differences between groups were evaluated by one-way analysis of variance (ANOVA)

followed by Dunnett's test for multiple comparisons. The Pearson correlation

coefficient was computed to compare the bacterial compositions analyzed by different

sequencing methods. Statistical significance was defined by a P-value < 0.05. Statistical

analyses were performed with Prism8 (GraphPad Software, Inc. La Jolla, CA, USA).

Statement of ethics

The Sunkaky Institutional Review Board approved this study (No. 2017-27). All

participants provided written informed consent.

Abbreviations

NGS: next-generation sequencing; OTU: operational taxonomic unit; PCR: polymerase

chain reaction; rRNA: ribosomal RNA

Additional files

Additional File 1: Representative bacterial genomes stored in the GenomeSync

database.

Additional File 2: Fig. S1. Sequence heterogeneities of the 27F primer-annealing site

in 16S rRNA genes. **Fig. S2.** Evaluation of 16S rRNA PCR primers for identification of

bacterial species. **Fig. S3.** Effect of read number on taxonomic classification. **Fig. S4.**

Rarefaction curves of observed OTUs in 16S V3-V4 amplicon sequencing of human

fecal samples using the MiSeq™ platform. **Fig. S5.** Species composition of *Blautia* in

human fecal samples. **Fig. S6.** Species composition of *Bacteroides* in human fecal

samples. **Fig. S7.** Deviations in the relative abundances of *Bifidobacterium* species in

human fecal samples. **Fig. S8.** Comparison of species composition of fecal

Bifidobacterium between classification methods.

Additional File 3: Tables S1-S6. Taxonomic assignment of the mock community

analyzed by MinION™ sequencing.

Additional File 4: Alignment search results for V1-V9 amplicon sequencing of the

mock community.

Additional File 5: Alignment search results for V3-V4 amplicon sequencing of the

mock community.

Additional File 6: Table S7. Correlations between numbers of reads and numbers of

detected species in 16S rRNA gene sequencing of human fecal samples. **Table S8.**

Comparison of species composition of fecal *Blautia* between sequencing methods.

Table S9. Comparison of species composition of fecal *Bacteroides* between sequencing

methods. **Table S10.** Comparison of species composition of fecal *Bifidobacterium*

between sequencing methods.

Additional File 7: Taxonomic profile of human fecal samples from MinION™

sequencing (amplicons: V1-V9).

Additional File 8: Taxonomic profile of human fecal samples from MinION™

sequencing (amplicons: V3-V4).

Additional File 9: Taxonomic profile of human fecal samples from MiSeq™

sequencing (amplicons: V3-V4).

Additional File 10: Taxonomic profiles of human fecal samples from MiSeq™

sequencing (amplicons: V3-V4, taxonomic classification by OTU-based analysis using

the QIIME 2 pipeline).

Declarations

Ethics approval and consent to participate

This study was approved by the Sunkaky Institutional Review Board (No. 2017-27).

Consent for publication

All participants provided written informed consent.

Availability of data and materials

The sequence datasets supporting the conclusions of this article are available in the

DDBJ DRA database (https://www.ddbj.nig.ac.jp/dra/index-e.html) under accession

numbers DRR225043 to DRR225065.

Competing interests

The authors declare that they have no competing interests.

Funding

This work was supported by Japan Society for the Promotion of Science KAKENHI

Grant Number JP19K09339 to YMa and the branding program as a world-leading

research university on intractable immune and allergic diseases supported by the

Ministry of Education, Culture, Sports, Science and Technology of Japan.

Authors' contributions

YMa, KK, AF, YMo, YN, SN and KH designed and supervised the study. SK, AF, YMo,

and HO contributed to sample collection. YMa, KM, and TT conducted the

experiments. YMa, YYasumizu, YYasuoka, and SN analyzed the data. YMa wrote the

manuscript. YYasumizu, HB, SN, and KH contributed to editing the manuscript. All

authors read and approved the final manuscript.

Acknowledgements

We are grateful to Tadashi Imanishi (Tokai University School of Medicine) and Shino

Matsukawa (Kyoto University Hospital) for helpful comments and discussion. We

would like to thank Editage (www.editage.com) for English language editing.

References

1. Chiu CY, Miller SA. Clinical metagenomics. Nat Rev Genet. 2019;20(6):341-55.

2. Loman NJ, Misra RV, Dallman TJ, Constantinidou C, Gharbia SE, Wain J, et al. Performance comparison of benchtop high-throughput sequencing platforms. Nat Biotechnol. 2012;30(5):434-9.

3. Didelot X, Bowden R, Wilson DJ, Peto TEA, Crook DW. Transforming clinical microbiology with bacterial genome sequencing. Nat Rev Genet. 2012;13(9):601-12.

4. Clarridge JE, 3rd. Impact of 16S rRNA gene sequence analysis for identification of bacteria on clinical microbiology and infectious diseases. Clin Microbiol Rev. 2004;17(4):840-62, table of contents.

5. Langille MG, Zaneveld J, Caporaso JG, McDonald D, Knights D, Reyes JA, et al. Predictive functional profiling of microbial communities using 16S rRNA marker gene sequences. Nat Biotechnol. 2013;31(9):814-21.

6. Johnson JS, Spakowicz DJ, Hong BY, Petersen LM, Demkowicz P, Chen L, et al. Evaluation of 16S rRNA gene sequencing for species and strain-level microbiome analysis. Nat Commun. 2019;10(1):5029.

7. Ravi RK, Walton K, Khosroheidari M. MiSeq: A Next Generation Sequencing Platform for Genomic Analysis. Methods Mol Biol. 2018;1706:223-32.

8. Kuczynski J, Lauber CL, Walters WA, Parfrey LW, Clemente JC, Gevers D, et al. Experimental and analytical tools for studying the human microbiome. Nat Rev Genet. 2011;13(1):47-58.

9. Leggett RM, Clark MD. A world of opportunities with nanopore sequencing. J Exp Bot. 2017;68(20):5419-29.

10. Quick J, Ashton P, Calus S, Chatt C, Gossain S, Hawker J, et al. Rapid draft sequencing and real-time nanopore sequencing in a hospital outbreak of

Salmonella. Genome Biol. 2015;16:114.

11. Leggett RM, Alcon-Giner C, Heavens D, Caim S, Brook TC, Kujawska M, et al.

Rapid MinION profiling of preterm microbiota and antimicrobial-resistant

pathogens. Nat Microbiol. 2020;5(3):430-42.

12. Benitez-Paez A, Sanz Y. Multi-locus and long amplicon sequencing approach to

study microbial diversity at species level using the MinION portable nanopore

sequencer. Gigascience. 2017;6(7):1-12.

13. Shin H, Lee E, Shin J, Ko SR, Oh HS, Ahn CY, et al. Elucidation of the bacterial

communities associated with the harmful microalgae Alexandrium tamarense

and Cochlodinium polykrikoides using nanopore sequencing. Sci Rep.

2018;8(1):5323.

14. Mitsuhashi S, Kryukov K, Nakagawa S, Takeuchi JS, Shiraishi Y, Asano K, et

al. A portable system for rapid bacterial composition analysis using a nanopore-

based sequencer and laptop computer. Sci Rep. 2017;7(1):5657.

15. Nakagawa S, Inoue S, Kryukov K, Yamagishi J, Ohno A, Hayashida K, et al.

Rapid sequencing-based diagnosis of infectious bacterial species from

meningitis patients in Zambia. Clin Transl Immunology. 2019;8(11):e01087.

16. Kai S, Matsuo Y, Nakagawa S, Kryukov K, Matsukawa S, Tanaka H, et al.

Rapid bacterial identification by direct PCR amplification of 16S rRNA genes

using the MinION nanopore sequencer. FEBS Open Bio. 2019;9(3):548-57.

17. Kim SW, Suda W, Kim S, Oshima K, Fukuda S, Ohno H, et al. Robustness of

gut microbiota of healthy adults in response to probiotic intervention revealed by

high-throughput pyrosequencing. DNA Res. 2013;20(3):241-53.

18. Arboleya S, Watkins C, Stanton C, Ross RP. Gut Bifidobacteria Populations in

Human Health and Aging. Front Microbiol. 2016;7:1204.

19. Backhed F, Ley RE, Sonnenburg JL, Peterson DA, Gordon JI. Host-bacterial

mutualism in the human intestine. Science. 2005;307(5717):1915-20.

20.   Tanaka H, Matsuo Y, Nakagawa S, Nishi K, Okamoto A, Kai S, et al. Real-time

538   diagnostic analysis of MinION-based metagenomic sequencing in clinical

539   microbiology evaluation: a case report. JA Clin Rep. 2019;5(1):24.

540   21.   Arumugam M, Raes J, Pelletier E, Le Paslier D, Yamada T, Mende DR, et al.

541   Enterotypes of the human gut microbiome. Nature. 2011;473(7346):174-80.

542   22.   Caporaso JG, Kuczynski J, Stombaugh J, Bittinger K, Bushman FD, Costello

543   EK, et al. QIIME allows analysis of high-throughput community sequencing

544   data. Nat Methods. 2010;7(5):335-6.

545   23.   Bolyen E, Rideout JR, Dillon MR, Bokulich NA, Abnet CC, Al-Ghalith GA, et

546   al. Reproducible, interactive, scalable and extensible microbiome data science

547   using QIIME 2. Nat Biotechnol. 2019;37(8):852-7.

548   24.   Milani C, Lugli GA, Duranti S, Turroni F, Bottacini F, Mangifesta M, et al.

549   Genomic encyclopedia of type strains of the genus Bifidobacterium. Appl

550   Environ Microbiol. 2014;80(20):6290-302.

551   25.   Human Microbiome Project C. A framework for human microbiome research.

552   Nature. 2012;486(7402):215-21.

553   26.   Srinivasan R, Karaoz U, Volegova M, MacKichan J, Kato-Maeda M, Miller S, et

554   al. Use of 16S rRNA gene for identification of a broad range of clinically

555   relevant bacterial pathogens. PLoS One. 2015;10(2):e0117617.

556   27.   Santos A, van Aerle R, Barrientos L, Martinez-Urtaza J. Computational methods

557   for 16S metabarcoding studies using Nanopore sequencing data. Comput Struct

558   Biotechnol J. 2020;18:296-305.

559   28.   Ma X, Stachler E, Bibby K. Evaluation of Oxford Nanopore MinION™

560   Sequencing for 16S rRNA Microbiome Characterization. bioRxiv. 2017

561   https://doi.org/10.1101/099960

562   29.   Quick J, Grubaugh ND, Pullan ST, Claro IM, Smith AD, Gangavarapu K, et al.

563   Multiplex PCR method for MinION and Illumina sequencing of Zika and other

virus genomes directly from clinical samples. Nat Protoc. 2017;12(6):1261-76.

30. Cornelis S, Gansemans Y, Deleye L, Deforce D, Van Nieuwerburgh F. Forensic

SNP Genotyping using Nanopore MinION Sequencing. Sci Rep. 2017;7:41759.

31. Bukin YS, Galachyants YP, Morozov IV, Bukin SV, Zakharenko AS, Zemskaya

TI. The effect of 16S rRNA region choice on bacterial community

metabarcoding results. Sci Data. 2019;6:190007.

32. Ventura M, Turroni F, Motherway MO, MacSharry J, van Sinderen D. Host-

microbe interactions that facilitate gut colonization by commensal

bifidobacteria. Trends Microbiol. 2012;20(10):467-76.

33. Shin J, Lee S, Go MJ, Lee SY, Kim SC, Lee CH, et al. Analysis of the mouse gut

microbiome using full-length 16S rRNA amplicon sequencing. Sci Rep.

2016;6:29681.

34. Magi A, Semeraro R, Mingrino A, Giusti B, D'Aurizio R. Nanopore sequencing

data analysis: state of the art, applications and challenges. Brief Bioinform.

2018;19(6):1256-72.

35. Rang FJ, Kloosterman WP, de Ridder J. From squiggle to basepair:

computational approaches for improving nanopore sequencing read accuracy.

Genome Biol. 2018;19(1):90.

36. Takagi T, Naito Y, Inoue R, Kashiwagi S, Uchiyama K, Mizushima K, et al.

Differences in gut microbiota associated with age, sex, and stool consistency in

healthy Japanese subjects. J Gastroenterol. 2019;54(1):53-63.

37. Klindworth A, Pruesse E, Schweer T, Peplies J, Quast C, Horn M, et al.

Evaluation of general 16S ribosomal RNA gene PCR primers for classical and

next-generation sequencing-based diversity studies. Nucleic Acids Res.

2013;41(1):e1.

38. Shen W, Le S, Li Y, Hu F. SeqKit: A Cross-Platform and Ultrafast Toolkit for

FASTA/Q File Manipulation. PLoS One. 2016;11(10):e0163962.

39.   Quast C, Pruesse E, Yilmaz P, Gerken J, Schweer T, Yarza P, et al. The SILVA

592   ribosomal RNA gene database project: improved data processing and web-based

593   tools. Nucleic Acids Res. 2013;41(Database issue):D590-6.

594   40.   Silva reference files. https://mothur.org/wiki/silva_reference_files/

595   41.   De Coster W, D'Hert S, Schultz DT, Cruts M, Van Broeckhoven C. NanoPack:

596   visualizing and processing long-read sequencing data. Bioinformatics.

597   2018;34(15):2666-9.

598   42.   Genome Search Toolkit. http://kirill-kryukov.com/study/tools/gstk/

599   43.   Frith MC. A new repeat-masking method enables specific detection of

600   homologous sequences. Nucleic Acids Res. 2011;39(4):e23.

601   44.   Li H. Minimap2: pairwise alignment for nucleotide sequences. Bioinformatics.

602   2018;34(18):3094-100.

603   45.   GenomeSync. http://genomesync.org

604   46.   Federhen S. The NCBI Taxonomy database. Nucleic Acids Res.

605   2012;40(Database issue):D136-43.

Figure legends

Fig. 1 16S rRNA gene sequence analysis using the MinION™ nanopore sequencer. **a**

Workflow of 16S rRNA amplicon sequencing on the MinION™ platform. Sequencing

libraries are generated by the four-primer PCR-based strategy, enabling simplified post-

PCR adapter attachment. At the initial stage of PCR, the 16S rRNA gene is amplified

with the inner primer pairs. The resulting PCR products are targeted for amplification

with the outer primers to introduce the barcode and tag sequences at both ends, to which

adapter molecules can be attached in a single-step reaction. **b, c** Taxonomic assignments

of a mock community analyzed by MinION™ sequencing. The V1-V9 or V3-V4 region

of the 16S rRNA gene was amplified from a pre-characterized mock community sample

comprising ten bacterial species and sequenced on the MinION™ platform. Three

thousand reads were randomly selected from the processed data set and aligned directly

to the reference genome database of 5850 representative bacterial species. The pie

charts represent taxonomic profiles at the (b) genus and (c) species levels. Slices

corresponding to misclassified (assigned to bacteria not present in the mock

community) or unclassified (not classified at the species level but placed in a higher

taxonomic rank) reads are exploded. The relative abundance (%) of each taxon is

shown.

Fig. 2 Accurate taxonomic assignment afforded by full-length MinION™ sequencing of

the 16S rRNA gene. Classification accuracy compared between full-length (V1-V9) and

partial (V3-V4) 16S sequencing data obtained from composition profiling of the ten-

species mock community. The donut charts show the proportions of reads correctly

assigned to the species constituting the mock community. The percentage of correctly

classified reads is shown in the center hole. NA: not assigned (no reads were classified

in *Escherichia* genus).

Fig. 3 16S rRNA gene sequence analysis of human gut microbiota. Six human fecal

samples (F1-F6) were subjected to full-length 16S rRNA amplicon sequencing via

MinION™. Numbers of detected species are plotted against numbers of reads used for

taxonomic classification.

Fig. 4 Comparison of taxonomic profiles of human gut microbiota between sequencing

methodologies. Six fecal samples (F1-F6) were analyzed by sequencing the entire 16S

rRNA gene using MinION™ (N_V1-V9). For comparison, the V3-V4 region was

sequenced on MinION™ (N_V3-V4) or MiSeq™ platforms (I_V1-V9). Randomly

sampled 20000 reads from each data set were allocated to the reference genome

database of 5850 representative bacterial species. A heat map shows the relative genus

abundance (%) of classified reads. The 15 most abundant taxa are shown. The Pearson

correlation coefficient (*r*) between sequencing methods was computed. Asterisks

indicate significant correlations at $P < 0.05$.

Fig. 5 Comparison of taxonomic resolution. The percentages of ambiguous reads not

assigned to the species level are plotted for six fecal samples analyzed by MinION™

(N_V1-V9 and N_V3-V4) or MiSeq™ (I_V3-V4). Horizontal bars represent mean

values. * $P < 0.05$ (statistically significant).

Fig. 6 Species composition of *Bifidobacterium* in six fecal samples. MinION™ V1-V9

sequencing confers species-level resolution for bacterial composition profiling. Results

obtained by the three sequencing methods are shown. The legends show the 14 most

abundant *Bifidobacterium* species.

Table 1 MinION™ sequencing statistics for the mock community sample

Primers	Pass reads				Trimmed reads		Filtered reads		
	No. of reads	Min (bp)	Avg (bp)	Max (bp)	No. of reads	Avg (bp)	No. of reads	Avg (bp)	Q score
V1-V9	8651	237	1497	3292	8455	1367.1	6972 (80.6%)	1473	9.0
V3-V4	101372	180	585.7	1977	99937	451.8	96189 (94.9%)	454.9	9.2

Min: minimum read length, Avg: average read length, Max: maximum read length, Q score: average Phred quality score. The percentage of reads retained after size filtering is shown in parentheses.

Table 2 Statistics of MinION™ sequencing data for human fecal samples

Sample	Pass reads				Trimmed reads		Filtered reads		
	No. of reads	Min (bp)	Avg (bp)	Max (bp)	No. of reads	Avg (bp)	No. of reads	Avg (bp)	Q score
F1/N V1-V9	104895	186	1521.1	4549	103100	1386.4	89752 (85.6%)	1463.7	9.4
F2/N V1-V9	84065	169	1393.8	4253	82458	1259.8	60326 (71.8%)	1461.4	9.4
F3/N V1-V9	76968	168	1474.3	4829	74479	1343.3	60,713 (78.9%)	1465.5	9.4
F4/N V1-V9	114060	168	1541.7	4836	111436	1410.6	100569 (88.2%)	1469.9	9.4
F5/N V1-V9	85912	177	1536.0	4877	83038	1405.4	74168 (86.3%)	1474.2	9.4
F6/N V1-V9	108938	213	1525.1	4866	106857	1393.5	93146 (85.5%)	1467.4	9.4
F1/N V3-V4	52864	160	568.8	2759	52283	435.2	48494 (91.7%)	442.5	9.1
F2/N V3-V4	92816	174	583.4	2886	91989	442.8	89016 (95.9%)	444.7	9.2
F3/N V3-V4	60200	163	568.5	2062	59435	434.6	55706 (92.5%)	441.1	9.2
F4/N V3-V4	83021	202	578.0	2050	81734	446.1	77995 (93.9%)	450.0	9.2
F5/N V3-V4	78409	167	578.4	1796	76135	447.8	72526 (92.5%)	453.1	9.2
F6/N V3-V4	74931	114	580.3	2246	73946	446.1	71330 (95.2%)	449.1	9.2

N: Oxford Nanopore MinION™, Min: minimum read length, Avg: average read length, Max: maximum read length, Q score: average Phred quality score. The percentage of reads retained after size filtering is shown in parentheses.

Table 3 Statistics of MiSeq™ sequencing data for human fecal samples

Sample	Paired reads No. of reads	Merged reads No. of reads	Avg (bp)	Filtered reads No. of reads	Avg (bp)	Q Score
F1/I V3-V4	66242	63821	449.3	63778 (96.3%)	449.5	31.4
F2/I V3-V4	68824	66640	447.6	66490 (96.6%)	448.3	32.3
F3/I V3-V4	132057	128095	446.9	127999 (96.9%)	447.1	31.6
F4/I V3-V4	103532	100945	451.4	100853 (97.4%)	451.7	32.8
F5/I V3-V4	72136	70521	451	70459 (97.7%)	451.3	33.2
F6/I V3-V4	52182	50907	449	50841 (97.4%)	449.5	33.2

I: Illumina MiSeq™, Avg: average read length, Q score: average Phred quality score. The percentage of reads retained after size filtering is shown in parentheses.

Fig. 1

a

Bacterial DNA

16S gene-specific inner primers
Barcoded outer primers

Amplification of 16S rRNA gene

Ligase-free adapter attachment

MinION™ sequencing

Anchor sequence
Barcode sequence
tag for adapter attachment
Rapid sequencing adapter

b

V1-V9

V3-V4

3,000 reads

3,000 reads

Genus	V1-V9	V3-V4
Bacillus	5.07%	9.30%
Bifidobacterium	5.63%	6.60%
Clostridium	16.97%	19.70%
Deinococcus	10.23%	16.10%
Enterococcus	7.30%	0.13%
Escherichia	4.50%	0.00%
Lactobacillus	11.63%	5.77%
Rhodobacter	12.97%	9.17%
Staphylococcus	12.73%	5.97%
Streptococcus	11.87%	8.30%
mis- or unclassified	1.10%	18.97%

c

V1-V9

V3-V4

3,000 reads

3,000 reads

Species	V1-V9	V3-V4
B. cereus	0.40%	0.17%
B. adolescentis	5.63%	5.53%
C. beijerinckii	15.67%	0.03%
D. radiodurans	10.23%	16.07%
E. faecalis	7.30%	0.00%
E. coli	4.50%	0.00%
L. gasseri	11.47%	4.57%
R. sphaeroides	12.97%	9.07%
S. epidermidis	12.50%	1.20%
S. mutans	11.87%	8.20%
mis- or unclassified	7.47%	55.17%

Fig. 2

Fig. 3

Fig. 4

Fig. 5

Fig. 6

a

Bifidobacterium species of human origin

- 🟢 *B. adolescentis*
- 🟡 *B. angulatum*
- 🟣 *B. animalis subsp. lactis*
- 🟧 *B. breve*
- 🔵 *B. dentium*
- 🟩 *B. gallicum*
- 🔵 *B. longum*
- 🟢 *B. scardovii*

Bifidobacterium species of non-human origin

- 🔴 *B. callitrichos*
- 🟤 *B. gallinarum*
- 🟪 *B. lemurum*
- 🟣 *B. minimum*
- 🔵 *B. saguini*
- 🟢 *B. thermophilum*

- ⬜ Unclassified *Bifidobacterium* (not assigned to a species)