# Generating Artificial Sensations with Spinal Cord Stimulation in Primates and Rodents

Amol P. Yadav[1-4], Shuangyan Li[3,11,12], Max O. Krucoff[1,8], Mikhail A. Lebedev[2,9,10], Muhammad M. Abd-El-Barr[1], Miguel A.L. Nicolelis[1-7]

[1]Department of Neurosurgery, [2]Center for Neuroengineering, [3]Department of Neurobiology, [4]Department of Biomedical Engineering, [5]Department of Psychology and Neuroscience, [6]Department of Neurology, Duke University, Durham, NC 27710, [7]Edmond and Lily Safra International Institute of Neuroscience, Natal, Brazil 59066060, [8]Department of Neurosurgery, Medical College of Wisconsin, Milwaukee, WI 53226, [9]Center for Bioelectric Interfaces of the Institute for Cognitive Neuroscience, National Research University Higher School of Economics, Moscow, Russia 101000, [10]Department of Information and Internet Technologies of Digital Health Institute, I.M. Sechenov First Moscow State Medical University, Moscow, Russia 119146, [11]State Key Laboratory of Reliability and Intelligence of Electrical Equipment, School of Electrical Engineering, [12]Tianjin Key Laboratory Bioelectromagnetic Technology and Intelligent Health, Hebei University of Technology, Tianjin 300130, P.R.China

Corresponding Author:

Amol P. Yadav, Ph.D.

Department of Neurosurgery

Duke University,

Box 103905, Durham, NC 27710

amol.yadav@duke.edu

**Abstract**

For patients who have lost sensory function due to a neurological injury such as spinal cord injury (SCI), stroke, or amputation, spinal cord stimulation (SCS) may provide a mechanism for restoring somatic sensations via an intuitive, non-visual pathway. Inspired by this vision, here we trained rhesus monkeys and rats to detect and discriminate patterns of epidural SCS. Thereafter, we constructed psychometric curves describing the relationship between different SCS parameters and the animal's ability to detect SCS and/or changes in its characteristics. We found that the stimulus detection threshold decreased with higher frequency, longer pulse-width, and increasing duration of SCS. Moreover, we found that monkeys were able to discriminate temporally- and spatially-varying patterns (i.e. variations in frequency and location) of SCS delivered through multiple electrodes. Additionally, sensory discrimination of SCS-induced sensations in rats obeyed Weber's law of just noticeable differences. These findings suggest that by varying SCS intensity, temporal pattern, and location different sensory experiences can be evoked. As such, we posit that SCS can provide intuitive sensory feedback in neuroprosthetic devices.

**Main**

Lack of sensory feedback from a brain-controlled actuator or prosthetic device is a major

hindrance to successful integration of the neuroprosthesis in activities of daily life and

rehabilitative protocols (1-3). The somatosensory cortex (S1) and thalamus have been proposed

as potential targets for neurostimulation that could produce naturalistic somatosensory percepts

(4-10). However, stimulating these brain areas requires surgical implantation of deep intracranial

electrodes – a procedure associated with significant risks. While peripheral nerve stimulation

provides a less invasive alternative, sensations evoked with this method are highly localized, and

thus limited in their applicability as a general-purpose sensory input pathway to the brain (11-13).

Previously, our group demonstrated that electrical stimulation of the dorsal surface of the spinal

cord can be used to transmit sensory information to the brain or between multiple brains in rodents

(14). Building on this previous work, here we explored whether rats and nonhuman primates can

learn to detect and discriminate artificial sensations generated with dorsal thoracic epidural spinal

cord stimulation (SCS). Understanding the psychophysical relationship between SCS parameters

and the ability to detect sensations is critical for the development of novel neuroprosthetic devices

that restore function to people with sensory disabilities. We examined how sensory discrimination

changes when SCS parameters are varied in both rodent and primate models, and we asked

whether animals can learn to discriminate sensations generated by SCS patterns that vary in

frequency and spatial location. After training the animals to discriminate SCS patterns, we

determined whether artificial sensations evoked by SCS of variable frequency follow Weber's law

of just noticeable differences (JND) - a critical property defining sensory discrimination.

**Results**

We implanted three rhesus monkeys with percutaneous epidural SCS electrodes at the dorsal

thoracic spinal level and trained them to perform a two-alternative forced choice task (2AFC) using

a joystick-controlled cursor (Figure 1a, Supplementary Figures 1b, 1c, and 1d). In a typical

**Figure 1**

Figure 1: Surgery and experimental task setup. a) We implanted three non-human primates (rhesus monkeys) with SCS percutaneous leads over the T6-T10 dorsal epidural surface of the spinal cord. Leads were externalized from the lower back area and secured inside a custom plastic housing. Leads were manually accessed by the experimenter for daily training and connected to a custom pulse stimulator. b) Monkeys were seated in a primate chair in front of a computer monitor with access to a hand-controlled joystick. They participated in a two-alternative forced choice task (2AFC) by moving the joystick controlling a cursor on the screen in order to receive a juice reward. c) On each trial, monkeys had to hold the cursor inside the center circle for 1 sec. After that, targets appeared on the left and right side of the center. Monkeys were presented with 'SCS-ON' (biphasic, 100 Hz, 200 µs, 1 sec) cue or 'SCS-OFF' cue when the targets appeared. After a brief, variable hold period (100-1000 ms), the center circle disappeared which indicated them to move the joystick. Monkeys had to move the cursor inside the left target on 'SCS-ON' trials and inside the right target on 'SCS-OFF' trials. Correct response resulted in juice reward. In the SCS discrimination task, stimulation was delivered at 200 µs for 1 sec. Monkeys had to select left target for 100 Hz stimulus and right target for 200 Hz stimulus in the frequency discrimination task. For spatial discrimination task, monkeys had to choose left target when stimulation was delivered at electrode pair 1 and right target for stimulation at electrode pair 2.

experimental session, a monkey was seated in a chair in front of a monitor that displayed task-

related cues. The animals moved a hand-held joystick to control a cursor on a screen (Figure 1b).

A typical trial consisted of a brief (300 - 500 ms) center hold period after which two targets

appeared. After a brief preparatory period (250 – 1000 ms) during which a trial cue was presented,

the monkeys had to move the cursor into one of the targets to obtain a juice reward. Monkeys

were initially trained to identify the correct target using a visual cue; however, during the

experimental sessions, no visual cues were presented, and they selected a target by interpreting

SCS cues alone. In the detection task, monkeys had to select the left target if SCS was delivered

during the preparatory period and right target if no SCS was delivered (Figure 1c). In the

discrimination task, monkeys had to select the left target for the 100 Hz stimuli and right target for

the 200 Hz stimuli (frequency discrimination) and the left target for electrode pair 1 and right target

for electrode pair 2 (spatial discrimination).

*Monkeys learned to detect SCS stimuli*

Monkeys M, O, and K learned to detect SCS-induced sensory percepts evoked using

percutaneous dorsal thoracic epidural electrodes (T7 for monkey M, T5-T6 for monkey O, T5-T6

for monkey K). Performance of all monkeys started below chance levels of 50% and reached

above 90% after learning (Figure 2a). Monkey M started detection performance at 49% and

reached a maximum of 93% in 16 days; monkey O started at 49% and reached a maximum of

97% in 10 days; and monkey K started at 41% and reached a maximum of 90% in 8 days.

*Electrode thresholds and electrode mapping*

Once the monkeys learned to detect SCS sensations, we used psychometric analysis to

determine the detection thresholds for different electrode combinations (Figure 2b). We observed

that the detection thresholds varied from 315.6 $\mu$A to 340 $\mu$A for monkey O and 197 $\mu$A to 748 $\mu$A

for monkey K for different cathode-anode electrode pairs (Figure 2c, right). Once electrode

 **Figure 2**

Figure 2: Monkeys learned to detect SCS stimuli. a) Learning curves (sigmoidal fits) for monkeys M, O, and

K showing behavioral performance (fraction correct trials) as a function of training days. b) Psychometric

function showing fraction trials detected as a function of stimulation amplitude in monkey O. Detection

threshold is defined as amplitude at which monkeys achieved 75% performance on detection task. c)

Mapping of bipolar electrode pairs on monkey K's body where stimulation on right-side electrode at

suprathreshold amplitude elicited minor muscle twitches or skin flutter (color coded by electrode pairs and

corresponding sensory thresholds shown on right).  Monkey body shape is adapted from (15).

thresholds were determined, we mapped the bipolar electrode pairs to locations on the monkey's

body by stimulating at suprathreshold amplitudes and observing stimulation-induced minor

muscle twitches or skin flutter. We observed that muscle twitches/skin flutter were elicited in the

trunk and abdomen area only at suprathreshold values but not at sensory threshold values (Figure

2c). We also noted that in both monkeys K and O experimentally determined sensory thresholds

were always lower than the observed motor thresholds for each cathode-anode electrode pair

(Supplementary Figure 3b).

*Sensitivity to detection of sensory percepts in primates*

Thereafter, we investigated the psychophysical relationship between stimulation parameters and

detection of sensory percepts by varying stimulation amplitudes along with stimulation frequency,

pulse-width, or duration of stimulation while keeping the other two parameters constant.

We varied amplitude from 50 $\mu$A to 800 $\mu$A for pulse-widths of 50 $\mu$s, 100 $\mu$s, 200 $\mu$s, and 400 $\mu$s

for monkey K, and pulse-widths of 100 $\mu$s, 200 $\mu$s, and 400 $\mu$s for monkey O. Frequency and

duration of stimulation were held constant (Figures 3a, 3e, and Supplementary Figure 2a). We

observed that stimulus detection threshold significantly decreased with increasing stimulation

pulse-width ($p < 0.05$, repeated measures one-way ANOVA) for both animals (Figure 3h).

We varied amplitude from 50 $\mu$A to 800 $\mu$A for frequencies of 10 Hz, 20 Hz, 50 Hz, 100 Hz, 200

Hz, and 500 Hz for monkey K, and frequencies of 20 Hz, 50 Hz, 100 Hz, 200 Hz, and 500 Hz for

monkey O while keeping pulse-width and duration of stimulation constant (Figures 3b, 3f, and

Supplementary Figure 2b). We observed that stimulation detection threshold significantly

decreased with increasing stimulation frequency ($p < 0.05$, repeated measures one-way ANOVA)

for both animals (Figure 3i).

We varied amplitude from 50 $\mu$A to 600 $\mu$A for duration of 50 ms, 100 ms, 500 ms, and 1000 ms

for monkey K, and amplitude between 50 $\mu$A to 700 $\mu$A for duration of 50 ms, 100 ms, and 500

ms for monkey O, while keeping pulse-width and frequency of stimulation constant (Figures 3c,

3g, and Supplementary Figure 2c). We observed that stimulation detection threshold significantly

**Figure 3**

Figure 3: Psychophysical evaluation of SCS sensory detection with varying stimulation parameters in

primates. Once monkeys learned to detect SCS stimuli at the standard parameters (frequency: 100 Hz,

pulse width: 200 µs, duration: 1 sec), we allocated different blocks of sessions where (pulse width &

amplitude; panels 'a' and 'e'); (frequency & amplitude; panels 'b' and 'f'); and (duration & amplitude; panels

'c' and 'g') were varied while keeping other parameters constant. In monkey K, in a separate block,

frequency and duration was varied with other parameters constant (pulse width: 200 µsec and amplitude:

325 µA). Psychometric curves in a-g are sigmoidal fits. Panels a-d represent psychometric curves for

monkey K, while panels e-g represent psychometric curves fitted to data averaged across monkeys K and

O (circles and error bars are mean ± sem). Panels h, i, j, indicate normalized detection thresholds

(normalized by maximum amplitude used in the experiment block of each individual monkey) averaged

across both monkeys (mean ± sem). Detection threshold were calculated as 75% fraction detected at the

stimulation parameters shown in panels a-c (monkey K) and Supplementary Figure 2 a-c (monkey O). P-

values were calculated using repeated measures one-way ANOVA. Panel k shows threshold duration

obtained as 75% detection from curves in panel d as a function of frequency.

decreased with increasing stimulation duration ($p<0.05$, repeated measures one-way ANOVA) for

both animals (Figure 3j).

In monkey K, we varied both frequency and duration of stimulation while keeping amplitude and

pulse-width of stimulation constant. We observed that as the frequency of stimulation increased,

the duration of stimulation to reach detection threshold decreased (Figures 3d and 3k). Monkey

K was able to detect a sensory percept generated by merely two stimulation pulses delivered at

1000 Hz.

*Sensitivity to detection of sensory percepts in rats*

We have previously shown that rats learn to detect sensations generated by epidural SCS

delivered at the T2 spinal level (14). In order to study the psychophysical performance of rats

 **Figure 4**

Figure 4: Behavioral task setup and rats learned to detect SCS stimuli. a) The experiment task consisted

of a closed behavioral chamber with two reward ports on one side of the chamber. The ports were covered

with vertical sliding doors. Five rats were implanted with bipolar stimulation electrodes on the dorsal surface

of the spinal cord at the T3 spinal level. b) Task consisted of a house light turning 'on', followed by sensory

cues for 1 second. During the cue period SCS was either delivered or not delivered. After the cue period

both reward doors opened, and the rat had to make a nose poke response in either of the ports to receive

water reward. If SCS was delivered rats had to poke inside the left port and if not delivered, then they had

to poke inside the right reward port. Poking in the correct reward poke initiated a water reward, while

incorrect pokes were not rewarded. c) Rats learned to detect SCS stimuli over a period of 15 – 25 days as

indicated – learning curve showing task performance indicated by fraction trials detected as a function of

training sessions. Circles and error bars indicate mean ± sem.

**Figure 5**

Figure 5: Psychophysical evaluation of SCS sensory detection with varying stimulation parameters in rats.

Once rats learned to detect SCS stimuli at the standard parameters (frequency: 100 Hz, pulse width: 200

μs, duration: 1 sec), we allocated different blocks of sessions where (pulse width & amplitude; panel 'a');

(frequency & amplitude; panel 'b'); and (duration & amplitude; panel 'c') were varied while keeping other

parameters constant. a, b, c) Psychometric curves from representative rats showing a leftward shift of

curves as the pulse-width (rat 4), frequency (rat 2), and duration (rat 5) of stimulation increased. d, e, f)

Psychometric curves with averaged data across five rats indicate leftward shift as pulse-width, frequency,

and duration are increased. X-axis represents relative amplitude values (for each rat raw amplitude values

were subtracted by minimum amplitude and the difference was divided by amplitude step size). Circles and

error bars are mean ± sem across five rats. Curves in panels 'a-f' are sigmoidal fits. Panels 'g', 'h', and 'i'

indicate normalized detection thresholds. Thresholds were calculated as 75% fraction detected at different

stimulation parameters consistent with panels 'a-f' and then normalized by maximum amplitude used in the

experimental block of each rat. Circles and error bars are mean ± sem across five rats. P-values were

calculated using repeated measures one-way ANOVA.

pertaining to sensory detection, initially we trained rats to detect SCS stimuli using a 2-AFC task

in a slightly modified behavioral chamber (Figure 4a and 4b). We trained five rats to detect SCS

stimuli delivered at the T3 spinal level (frequency: 100 Hz, pulse-width: 200 μs, duration: 1 sec,

biphasic pulses at 243.7±57.9 μA, Figure 4c).

Once rats learned the basic detection task, we varied stimulation parameters such as: pulse width

(50 – 400 μs); frequency (10 – 500 Hz); and duration (50 – 1000 ms) independently with

stimulation amplitude (Figures 5a, 5d – pulse width; 5b, 5e – frequency; and 5c, 5f – duration).

Similar to the results in monkeys, we observed that stimulation detection thresholds significantly

decreased with increasing pulse-width (Figure 5g, $p < 0.0001$, repeated measures one-way

ANOVA), frequency (Figure 5h, $p < 0.0001$, repeated measures one-way ANOVA) and duration

(Figure 5i, $p < 0.05$, repeated measures one-way ANOVA) for all rats.

**Figure 6**

<u>Figure 6: Monkeys learned to discriminate SCS stimuli varying in frequency and spatial location of</u>

<u>electrodes.</u> a,b) Monkeys K and O learned to discriminate SCS stimuli delivered at the same electrode

location but varying in frequency (100 Hz vs 200 Hz). Monkey O was stimulated at same amplitude while

monkey K was stimulated at the respective threshold amplitude for 100 Hz and 200 Hz. c,d) Monkeys K

and O learned to discriminate stimulation delivered at electrode pair 1 (T6 - T7 spinal level) vs electrode

pair 2 (T7-T8 spinal level). Curves in panels b and d indicate sigmoidal fits to fraction correct trials displayed

as a function of training days.

*Sensory discrimination in primates*

We then trained monkeys K and O to discriminate SCS that varied in frequency as well as location

of stimulation. On frequency discrimination (100 Hz vs 200 Hz, Figure 6a), monkey O's

performance improved from 29% on day 1 to 96% on day 17 of training, while monkey K's

performance improved from 74% on day 1 to 81% on day 11 (Figure 6b). Spatial discrimination

was achieved by stimulating electrode pair 1 versus electrode pair 2 (Figure 6c), where monkey

O's performance improved from 46% on day 1 to 97% on day 11 and monkey K's performance

improved from 74% on day 1 to 86% on day 7 (Figure 6d).

*Sensory discrimination and Weber's law in rats*

We had previously shown that rats can learn to discriminate up to four different time-varying

patterns of stimulation (14). In our current work, we trained rats to discriminate SCS stimuli with

different frequencies using the same behavioral setup that was used for the detection task

(Supplementary Figure 4a). In the basic training, rats learned to discriminate between 10 Hz and

100 Hz of stimulation delivered at pulse-width of 200 $\mu$sec, and duration of 1 sec (Supplementary

Figure 4b). Thereafter, we studied whether discrimination of sensations induced by different

stimulation frequencies follows the rules of Weber's law (16), which states that Just-Noticeable

Differences (JNDs) between a standard frequency and comparison frequency should linearly

increase with the standard frequency of stimulation. To this end, we determined JNDs at different

standard frequencies (10 – 100 Hz) where the comparison frequency was higher than the

standard (Figure 7a). We observed that JNDs had a significant linearly increasing relationship

with the standard frequency of stimulation (Figure 7b, p< 0.0001, linear regression test). After

that, we kept the standard as a higher frequency value (100 – 400 Hz) and decreased the

comparison frequency randomly from that value (Figure 7c). In this case also, we observed that

the JNDs for lower frequency comparison significantly increased linearly as the standard

frequency increased (Figure 7d, p< 0.0001, linear regression test). These results suggest that the

JND rule defined by Weber's law holds true for sensory discrimination of SCS frequencies.

**Figure 7**

Figure 7: Discrimination of higher and lower comparison frequency obeys Weber's Law. a) Fraction of trials

correctly discriminated from standard frequency of 10 Hz, 20 Hz, 40 Hz, 60 Hz, 80 Hz, and 100 Hz, when

compared to a range of higher frequency stimuli. c) Fraction of trials correctly discriminated from standard

frequency of 110 Hz, 220 Hz, 310 Hz, 380 Hz, and 400 Hz, when compared to a range of lower frequency

stimuli. Circles and error bars in 'a' and 'c' indicate mean ± sem. Curves indicate sigmoid fits to the data.

Just noticeable difference (JND) is considered as the comparison frequency value that achieves 75 %

discrimination ability. b, d) Just noticeable difference as a function of standard frequency is indicated by

black circles and error bars. Black line is linear regression fit to the data and pink bounds indicate 95%

confidence bound of the regression line. JNDs associated with higher frequency and lower frequency

comparison were significantly linearly related with standard frequency. P-values were calculated using

linear regression test.

**Discussion**

In this study, we found that rhesus monkeys and rats can learn to detect and discriminate artificial

sensations induced by SCS following several days of exposure. The threshold for detecting SCS

decreases with increasing pulse width, frequency, and duration of the stimulus. We also

documented the ability of monkeys to discriminate sensations that are generated by stimulation

pulses with varying frequency and spatial location. In rats, we showed that the just-noticeable

differences (JNDs) from a perceivable stimulus frequency were linearly related to the perceivable

frequency when it was compared with a stimulus that had either higher or lower frequency. These

results demonstrated the unique ability of SCS as a novel transmission channel to the brain to

encode highly contextual sensory information.

Our results on behavioral sensitivity to detection of sensation in rodents and primates were

comparable to those observed with Intracortical Microstimulation (ICMS) of S1 (5, 17). Notably,

sensitivity to stimulus amplitude increased with increasing pulse width, frequency, and duration

of stimulation. ICMS amplitude using currently FDA-approved UTAH arrays is usually restricted

below ~100 μA due to the possibility of brain tissue damage, which limits the amplitude range for

neuroprosthetic applications between detection thresholds of 20 - 50 μA and maximum allowable

safe amplitude of ~100 μA. In our SCS study, detection thresholds had a wider range from 200 –

800 μA at different stimulation settings and all detection thresholds were consistently on the non-

damaging side of the boundary between damaging and non-damaging stimulation delineated by

Shannon equation [$\log(D) = k – \log(Q)$, with $k = 1.85$] on log charge density versus log charge

per phase plot (Supplementary Figure 3a) (18). Assuming a commercially accepted maximum

charge density of 30 $\mu C/cm^2$ and minimum pulse-width of 50 μs (19), it could be estimated that

maximum SCS current of ~80 mA in monkeys and ~3 mA in rats could be delivered using

electrode contacts (monkeys: 0.1319 $cm^2$; rats: 0.005 $cm^2$) reported in our study without causing

tissue damage.

Frequency modulation has been historically considered a promising method for providing sensory

feedback with several studies showing that animals are capable of discriminating ICMS

frequencies and that frequency modulation obeys Weber's law. (4, 20-22), While ICMS amplitude

modulation in monkeys failed to follow Weber's law (5), experiments in rats showed that

modulation of perceived intensity by amplitude and pulse-width modulation followed Weber's law

(17). In our study, we investigated whether rats and monkeys could learn to discriminate SCS

frequencies. Although monkeys O and K learned to discriminate 100 Hz SCS from 200 Hz, after

taking a closer look at their learning curves it was evident that they displayed different learning

behaviors in the frequency discrimination task (Figures 6b and 6d). Monkey O started at lower

discrimination performance at earlier training sessions but reached higher level of performance

toward the end of training, whereas monkey K's performance started higher than chance and

improved marginally as the training progressed. These differences in learning behavior could be

attributed to the fact that monkey O was stimulated at constant amplitude for both frequency

values (100 Hz and 200 Hz) while monkey K was stimulated at each frequency's threshold
amplitude. It is quite possible that monkey O was discriminating differences in perceived
magnitudes of the sensory percepts while monkey K was discriminating differences in the
qualitative nature of the percepts induced. Rats are capable of discriminating temporal patterns
of SCS when the number of pulses are kept constant but the frequency of stimulus is varied (14).
In addition to that, our current results indicate that JNDs associated with SCS frequency
discrimination in rats clearly follow Weber's law because JNDs linearly increased with standard
stimulation frequency (Figure 7b and 7d). There were at least three discriminable percepts
between 10 Hz and 200 Hz. It can be argued that applying frequency modulation simultaneously
with amplitude and pulse-width modulation would potentially increase the number of distinct
discriminable percepts that are possible within the amplitude range allowed on current SCS
electrodes. Therefore, further experiments exploring the relationship between sensory
discrimination and frequency, pulse-width, amplitude, and duration of stimulation are necessary
to understand how these parameters relate to perceived intensity and quality of sensation evoked
by SCS.

A major advantage of SCS is its ability to target multiple dermatomes simultaneously with a single
electrode array. In particular, a single, commercially available SCS lead with multiple contacts
can evoke sensations in multiple dermatomes simultaneously due to the bilateral sensory
representation of the entire lower body in the ascending dorsal column fibers. It is quite apparent
from our results that monkeys can learn to discriminate the spatial location of the sensations
evoked by SCS (Figure 6d). This suggests that we can take full advantage of the medio-
lateral/rostro-caudal somatotopy represented in the dorsal column fibers in combination with
spatiotemporal stimulation patterns to electrically induce targeted tactile or proprioceptive
sensations in the body. This view is also supported by evidence from computational studies which
indicate that epidural SCS activates dorsal column fibers up to a depth of 0.2-0.25 mm from the

dorsal surface (23-27). However, additional work on miniaturizing electrode contacts and accurate

mapping of SCS-induced sensations needs to be performed to be able to elicit precise sensations.

In addition, the ability of SCS to modulate neuronal activity in supraspinal brain structures is quite

desirable from a neuroprosthetic as well as a therapeutic application standpoint (14, 28-31). A

major limitation of our study is the short experimental time (Supplementary Figure 1a) we had

available for primates – maximum of 5 months – due to the risk of infection associated with

externalized SCS leads. A fully implantable stimulation system, like the one implanted in chronic

pain patients could potentially extend our study indefinitely and allow us to perform longer

experiments in monkeys. Nevertheless, in rodents we were able to perform longer post-implant

experiments because the electrodes and their wires were fully enclosed inside the body.

In conclusion, we have successfully demonstrated that SCS can be used to encode sensory

information in both rats and monkeys and all together our results demonstrate that SCS-induced

sensations obey a robust encoding scheme in the brain. Additionally, our behavioral experiments

serve as a test bed for future animal studies which could elucidate the neural mechanisms

underlying SCS-based sensory detection and discrimination. We envision that SCS can be

developed as an artificial sensory feedback channel for delivering targeted tactile and

proprioceptive information to the brain.

**Methods**

All animal procedures were approved by the Duke University Institutional Animal Care and Use

Committee and in accordance with National Institute of Health Guide for the Care and Use of

Laboratory Animals. Three adult rhesus macaque monkeys (Macaca mulatta), monkeys 'M', 'O',

and 'K' and Long Evans rats (300 - 350 g) participated in the experiments.

*Monkey spinal implant surgery*

Monkeys M, K, and O were implanted with 8-contact cylindrical percutaneous leads (Model 3186,

diameter 1.4mm, contact length 3 mm, spacing 4 mm, Abbott Laboratories) bilaterally to the spinal

mid-line in the dorsal epidural space at T6 – T8 spinal level. Monkey M had two leads with 8

stimulation contacts each, monkey O had two leads, one with 8 contacts (right side) and one with

4 contacts (left side, Model 3146, similar electrode contact dimensions as Model 3186), and

monkey K had two leads with 8 stimulation contacts each. Experimental procedures for monkeys

M, O, and K lasted approximately 45, 135, and 150 days post-implant after which the leads were

explanted (Supplementary Figure 1a).

Implant surgery was performed under general anesthesia using standard procedures typical of

human implantation (for details see supplementary information and Supplementary Figures 1b

and 1c). Once leads were implanted in the epidural space, a small hole was created in the skin

off-midline to externalize the distal end of the lead. Externalized leads were enclosed in a custom

plastic cap which was sutured to the skin (Supplementary Figure 1d). The plastic cap allowed for

access to the leads by the researcher but protection from the animal. The animal wore a protective

vest after surgery and throughout the experimental period which prevented its access to the

plastic cap sutures to its back.

*Monkey SCS detection task*

Monkeys were trained to perform a two-alternative forced choice task where they were seated in

a chair facing a computer monitor which indicated trial progression (Figure 1b). On each trial a

center target appeared, and the monkey had to move a cursor which was joystick-controlled inside

the center target (Figure 1c). After a brief hold period of 500 milliseconds (ms) inside the center

target, two targets appeared on either side of the center target. Each monkey had to hold the

cursor inside the central target for a brief period of 500 ms – 1 second. During this hold period

movement cues were presented. If SCS was presented (charge-balanced, cathode-first, 200 µs

biphasic square pulses at 100 Hz for 1 second using custom microstimulator (32)), then the

monkey had to move the cursor to the left target to obtain a juice reward. If SCS was absent, then

the monkey had to move the cursor inside the right target. At the end of the hold period, the center

target disappeared, thus cuing the monkey to initiate cursor movement toward the reward.

Incorrect target reaches were not rewarded. Initiation of movement prior to the end of the hold

period terminated the trial without reward and a blank screen was displayed 3 seconds before

starting next trial. The learning performance of monkeys was studied using percentage correct

(PC) trials.

*Monkey psychometric evaluation of detection thresholds*

Once monkeys were trained on the detection task, detection thresholds were determined for

different electrode combinations using psychometric testing. Particularly, during 'Stimulation ON'

i.e. 'left target rewarded' trials, the stimulation amplitude was randomly varied from 50 $\mu$A to a

manually determined upper limit which was below the motor threshold. Only left target trials were

analyzed and a percentage correct (PC) performance at each stimulation amplitude value was

determined. A sigmoid curve was fit to the PC values and 75% was considered as detection

threshold. This was repeated for several electrode combinations.

*Monkey electrode mapping*

In monkey K, once sensory detection thresholds were determined, we mapped the location of

electrode pairs to location on the monkey's body by sedating the monkey and stimulating those

electrodes above the sensory threshold values. Areas on the body surface that elicited minor

muscle twitches or skin flutter were marked (Figure 2c, left) and the motor thresholds were noted.

In monkey O, these observations were not made under sedation but while it was seated in the

primate chair.

*Monkey detection thresholds as stimulation parameters vary*

During sets of consecutive sessions, we varied amplitude and frequency or amplitude and pulse-

width or amplitude and duration of stimulation while keeping other parameters constant (for

stimulation parameter ranges, see Supplementary Table 1). The standard parameters that

remained constant while others were varied were frequency: 100 Hz, pulse width: 200 $\mu$sec, and

duration: 1 sec. In monkey K, we varied frequency (10 – 1000 Hz) and duration (1 – 2000 ms) of

stimulation simultaneously while keeping pulse width and amplitude constant at 200 $\mu$s and 325

$\mu$A respectively. Detection thresholds for each monkey were normalized by maximum amplitude

used in the experiment block of that monkey before statistical analysis.

*Monkey sensory discrimination*

In the frequency discrimination task, each monkey was instructed to move the cursor inside the

left target for 100 Hz and right target for 200 Hz respectively (Figure 6a). In the spatial

discrimination task, the monkey was instructed to move the cursor inside the left target for

electrode pair 1 and inside right target for electrode pair 2 respectively (Figure 6c).

*Rat pre-training and SCS electrode implantation*

Moderately water deprived rats were placed inside the behavioral chamber for 2 days to

acclimatize to the behavioral environment. The behavioral chamber had two doors on one side of

the walls which enclosed water reward ports (Figure 4a), slightly modified from the one previously

described (33). Rats were gradually trained to receive water reward from the ports. Initially, both

reward doors were kept open and rats learned to receive water by licking at the water dispensing

spout. Later, the doors were kept closed and would open a few seconds after the house light

turned on. In subsequent sessions, left and right doors would open on alternate trials and rats

learned to obtain reward from each port alternatively. The pre-surgical training period lasted

approximately 8-10 days.

Thereafter custom designed SCS electrodes (1 mm X 0.5 mm contacts arranged transversely in

a bipolar configuration with 0.25 mm spacing using a 0.025mm thickness platinum foil, Goodfellow

Cambridge Limited, England) were implanted into the epidural space under T3 vertebra as

described in our previous article (29). After the rats recovered from the spinal surgical procedures,

cathode leading stimulation pulse trains were delivered at the SCS electrodes using a multi-

channel constant current stimulator (Master-8, A.M.P.I, Jerusalem, Israel) at stimulation settings

which were determined depending on the behavioral task under consideration (Figure 4a).

*Rat sensory detection task*

After recovery from surgery, rats were introduced to a two-alternative forced choice task (2AFC)

to learn detection of SCS stimuli in the chamber (Figure 4b). At the beginning of each trial, a light

in the chamber was turned on for 1 second as a reminder for the animals to pay attention. After

the light turned off, the rats received a sensory cue for 1 second. The sensory cue either consisted

of cathode-leading bipolar square pulse trains (pulse width: 200 μs, Frequency: 100 Hz, duration:

1 s) or no stimulation pulses (interval: 1 s) at each trial. After a brief delay of 0.5 seconds both

reward doors opened, and rats had to respond by choosing the left door for 'SCS ON' trials and

right door for 'SCS OFF' trials. Incorrect responses were not rewarded. During the learning of this

basic detection task, the intensity of the delivered current was determined before each session

and set using procedures described before (14, 29, 31) (mean±std, intensity at 100 Hz was 243.7

± 57.9 μA).

*Rat sensory detection psychophysics*

Once rats learned the basic sensory detection task and their performance was above 80%,

stimulation parameters were varied in a systematic manner during the SCS-ON trials. In different

experimental sessions, stimulation parameters such as frequency and amplitude, pulse-width and

amplitude, and pulse train duration and amplitude were varied while other parameters were kept

constant (standard parameters: pulse width: 200 µs; Frequency: 100 Hz; duration: 1 s), and

sensory detection threshold amplitude was determined (for stimulation parameter ranges see

Supplementary Table 1). For all the conditions, the detectable level of the amplitude was defined

as 75% accuracy of behavioral performance. Detection thresholds for each rat were normalized

by maximum amplitude used in the experiment block of that rat before statistical analysis.

*Rat sensory discrimination task*

In a 2AFC task, rats were presented with either a low frequency stimulus or a high frequency

stimulus during the sensory cue period in the behavioral chamber. For either frequency, the

stimulus was delivered at the same amplitude (determined at each session), pulse width (200 µs),

and duration (1 sec). After a brief delay period following sensory cue presentation, rats had to

choose the left door for higher frequency stimulus and right door for lower frequency stimulus to

obtain reward (Supplementary Figure 4a). Initially, rats were trained to discriminate between 10

Hz and 100 Hz frequency. Incorrect trails were not rewarded.

*Weber's law and sensory discrimination*

Once rats learned to discriminate 10 Hz stimulus from 100 Hz stimulus, demonstrated by

consistent discrimination performance above 80%, the lower frequency (standard frequency) was

kept constant during right door trials while the higher frequency (comparison frequency) was

randomized between 20 Hz and 110 Hz during left door trials. Sensory discrimination performance

between a standard frequency and comparison frequency was determined as the fraction of trials

successfully discriminated for that particular standard and comparison pair. Thus, fraction trials

discriminated was plotted as a function of comparison frequency to obtain Just-Noticeable

Differences (JNDs) determined as 75 % value on the curve. The same experiment was repeated

for different standard frequency values such as 20 Hz vs (40 – 220 Hz), 40 Hz vs (60 – 240 Hz),

60 Hz vs (85 – 310 Hz), 80 Hz vs (110 – 380 Hz),  and 100 Hz vs (130 – 400 Hz) to obtain JNDs

for each standard frequency value and to test for Weber's law.

Similarly, the experiment was repeated for discrimination of a low-frequency comparison stimulus

from a high-frequency standard stimulus. The standard frequency values were 110, 220, 310,

380, and 400 Hz, while the comparison frequency was randomized from (100 – 10 Hz), (200 – 20

Hz), (285 – 60 Hz), (350 – 80 Hz), and (370 – 100 Hz) respectively for each of the standard

frequency values. JNDs were calculated for each of the standard frequency values as explained

before to test for Weber's law.

*Statistical Analysis*

Repeated measures one-way ANOVA was used to test the significance of relationship between

stimulation parameters and sensory detection thresholds in both rats and monkeys (Figures 3h,

3i, 3j, 5g, 5h, and 5i). To test whether JNDs were significantly linearly related to standard

frequency in the sensory discrimination task, the linear regression test was used (Figures 7b and

7d).

**References**

1.     Lebedev MA, Nicolelis MA. Brain-Machine Interfaces: From Basic Science to Neuroprostheses and

Neurorehabilitation. Physiol Rev. 2017;97(2):767-837. doi: 10.1152/physrev.00027.2016. PubMed PMID:

28275048.

2.     Krucoff MO, Rahimpour S, Slutzky MW, Edgerton VR, Turner DA. Enhancing Nervous System

Recovery through Neurobiologics, Neural Interface Training, and Neurorehabilitation. Frontiers in

neuroscience. 2016;10. doi: ARTN 584

10.3389/fnins.2016.00584. PubMed PMID: WOS:000390598100001.

3.     Bensmaia SJ, Miller LE. Restoring sensorimotor function through intracortical interfaces: progress

and looming challenges. Nat Rev Neurosci. 2014;15(5):313-25. doi: 10.1038/nrn3724. PubMed PMID:

24739786.

4.     O'Doherty JE, Lebedev MA, Ifft PJ, Zhuang KZ, Shokur S, Bleuler H, Nicolelis MA. Active tactile

exploration using a brain-machine-brain interface. Nature. 2011;479(7372):228-31. doi:

10.1038/nature10489. PubMed PMID: 21976021; PMCID: 3236080.

5.     Kim S, Callier T, Tabot GA, Gaunt RA, Tenore FV, Bensmaia SJ. Behavioral assessment of sensitivity

to intracortical microstimulation of primate somatosensory cortex. Proceedings of the National Academy

of Sciences of the United States of America. 2015;112(49):15202-7. doi: 10.1073/pnas.1509265112.

PubMed PMID: 26504211; PMCID: 4679002.

6.     Tabot GA, Dammann JF, Berg JA, Tenore FV, Boback JL, Vogelstein RJ, Bensmaia SJ. Restoring the

sense of touch with a prosthetic hand through a brain interface. P Natl Acad Sci USA. 2013;110(45):18279-

84. doi: 10.1073/pnas.1221113110. PubMed PMID: WOS:000326550800062.

7.      Flesher SN, Collinger JL, Foldes ST, Weiss JM, Downey JE, Tyler-Kabara EC, Bensmaia SJ, Schwartz

AB, Boninger ML, Gaunt RA. Intracortical microstimulation of human somatosensory cortex. Science

translational medicine. 2016;8(361):361ra141. Epub 2016/11/01. doi: 10.1126/scitranslmed.aaf8083.

PubMed PMID: 27738096.

8.      Heming E, Sanden A, Kiss ZH. Designing a somatosensory neural prosthesis: percepts evoked by

different patterns of thalamic stimulation. Journal of neural engineering. 2010;7(6):064001. doi:

10.1088/1741-2560/7/6/064001. PubMed PMID: 21084731.

9.      Swan BD, Gasperson LB, Krucoff MO, Grill WM, Turner DA. Sensory percepts induced by microwire

array and DBS microstimulation in human sensory thalamus. Brain Stimul. 2018;11(2):416-22. doi:

10.1016/j.brs.2017.10.017. PubMed PMID: 29126946; PMCID: PMC5803348.

10.     Dadarlat MC, O'Doherty JE, Sabes PN. A learning-based approach to artificial sensory feedback

leads to optimal integration. Nature neuroscience. 2015;18(1):138-44. Epub 2014/11/25. doi:

10.1038/nn.3883. PubMed PMID: 25420067; PMCID: PMC4282864.

11.     Strauss I, Valle G, Artoni F, D'Anna E, Granata G, Di Iorio R, Guiraud D, Stieglitz T, Rossini PM,

Raspopovic S, Petrini FM, Micera S. Characterization of multi-channel intraneural stimulation in

transradial amputees. Sci Rep. 2019;9(1):19258. Epub 2019/12/19. doi: 10.1038/s41598-019-55591-z.

PubMed PMID: 31848384; PMCID: PMC6917705.

12.     Petrini FM, Valle G, Bumbasirevic M, Barberi F, Bortolotti D, Cvancara P, Hiairrassary A, Mijovic P,

Sverrisson AO, Pedrocchi A, Divoux JL, Popovic I, Lechler K, Mijovic B, Guiraud D, Stieglitz T, Alexandersson

A, Micera S, Lesic A, Raspopovic S. Enhancing functional abilities and cognitive integration of the lower

limb prosthesis. Science translational medicine. 2019;11(512). Epub 2019/10/04. doi:

10.1126/scitranslmed.aav8939. PubMed PMID: 31578244.

13.    Petrini FM, Bumbasirevic M, Valle G, Ilic V, Mijovic P, Cvancara P, Barberi F, Katic N, Bortolotti D,

Andreu D, Lechler K, Lesic A, Mazic S, Mijovic B, Guiraud D, Stieglitz T, Alexandersson A, Micera S,

Raspopovic S. Sensory feedback restoration in leg amputees improves walking speed, metabolic cost and

phantom pain. Nat Med. 2019;25(9):1356-63. Epub 2019/09/11. doi: 10.1038/s41591-019-0567-3.

PubMed PMID: 31501600.

14.    Yadav AP, Li D, Nicolelis MAL. A Brain to Spine Interface for Transferring Artificial Sensory

Information. Sci Rep. 2020;10(1):900. Epub 2020/01/23. doi: 10.1038/s41598-020-57617-3. PubMed

PMID: 31964948.

15.    Bellanca RU, Lee GH, Vogel K, Ahrens J, Kroeker R, Thom JP, Worlein JM. A simple alopecia scoring

system for use in colony management of laboratory-housed primates. J Med Primatol. 2014;43(3):153-61.

Epub 2014/02/28. doi: 10.1111/jmp.12107. PubMed PMID: 24571509; PMCID: PMC4438708.

16.    Ekman G. Weber Law and Related Functions. Journal of Psychology. 1959;47(2):343-52. doi: Doi

10.1080/00223980.1959.9916336. PubMed PMID: WOS:A1959CAK1200020.

17.    Bjanes DA, Moritz CT. A Robust Encoding Scheme for Delivering Artificial Sensory Information via

Direct Brain Stimulation. IEEE Trans Neural Syst Rehabil Eng. 2019;27(10):1994-2004. Epub 2019/08/24.

doi: 10.1109/TNSRE.2019.2936739. PubMed PMID: 31443035.

18.    Merrill DR, Bikson M, Jefferys JG. Electrical stimulation of excitable tissue: design of efficacious

and    safe    protocols.    J    Neurosci    Methods.    2005;141(2):171-98.    Epub    2005/01/22.    doi:

10.1016/j.jneumeth.2004.10.020. PubMed PMID: 15661300.

19.    Cogan SF, Ludwig KA, Welle CG, Takmakov P. Tissue damage thresholds during therapeutic

electrical stimulation. Journal of neural engineering. 2016;13(2):021001. Epub 2016/01/23. doi:

10.1088/1741-2560/13/2/021001. PubMed PMID: 26792176; PMCID: PMC5386002.

20.    Romo R, Hernandez A, Zainos A, Brody CD, Lemus L. Sensing without touching: psychophysical

performance based on cortical microstimulation. Neuron. 2000;26(1):273-8. PubMed PMID: 10798410.

21.    O'Doherty JE, Shokur S, Medina LE, Lebedev MA, Nicolelis MAL. Creating a neuroprosthesis for

active tactile exploration of textures. Proceedings of the National Academy of Sciences of the United

States of America. 2019;116(43):21821-7. Epub 2019/10/09. doi: 10.1073/pnas.1908008116. PubMed

PMID: 31591224; PMCID: PMC6815176.

22.    Callier T, Brantly NW, Caravelli A, Bensmaia SJ. The frequency of cortical microstimulation shapes

artificial touch. Proceedings of the National Academy of Sciences of the United States of America.

2020;117(2):1191-200. Epub 2019/12/28. doi: 10.1073/pnas.1916453117. PubMed PMID: 31879342;

PMCID: PMC6969512.

23.    Holsheimer J, Barolat G. Spinal geometry and paresthesia coverage in spinal cord stimulation.

Neuromodulation : journal of the International Neuromodulation Society. 1998;1(3):129-36. doi:

10.1111/j.1525-1403.1998.tb00006.x. PubMed PMID: 22150980.

24.    Barolat G, Massaro F, He J, Zeme S, Ketcik B. Mapping of sensory responses to epidural stimulation

of the intraspinal neural structures in man. Journal of neurosurgery. 1993;78(2):233-9. Epub 1993/02/01.

doi: 10.3171/jns.1993.78.2.0233. PubMed PMID: 8421206.

25.    Holsheimer J, Buitenweg JR. Review: Bioelectrical mechanisms in spinal cord stimulation.

Neuromodulation : journal of the International Neuromodulation Society. 2015;18(3):161-70; discussion

70. doi: 10.1111/ner.12279. PubMed PMID: 25832787.

26.    North RB, Sieracki JM, Fowler KR, Alvarez B, Cutchis PN. Patient-interactive, microprocessor-

controlled neurological stimulation system. Neuromodulation : journal of the International

Neuromodulation    Society.    1998;1(4):185-93.    Epub    1998/10/01.    doi:    10.1111/j.1525-

1403.1998.tb00015.x. PubMed PMID: 22151030.

27.    Holsheimer J. Which Neuronal Elements are Activated Directly by Spinal Cord Stimulation.

Neuromodulation : journal of the International Neuromodulation Society. 2002;5(1):25-31. Epub

2002/01/01. doi: 10.1046/j.1525-1403.2002._2005.x. PubMed PMID: 22151778.

28.    Yadav AP, Nicolelis MAL. Electrical stimulation of the dorsal columns of the spinal cord for

Parkinson's disease. Mov Disord. 2017;32(6):820-32. doi: 10.1002/mds.27033. PubMed PMID: 28497877.

29.    Yadav AP, Fuentes R, Zhang H, Vinholo T, Wang CH, Freire MA, Nicolelis MA. Chronic spinal cord

electrical    stimulation    protects    against    6-hydroxydopamine    lesions.    Sci    Rep.    2014;4:3839.    doi:

10.1038/srep03839. PubMed PMID: 24452435; PMCID: 3899601.

30.    Yadav AP, Borda E, Nicolelis MA. Closed loop spinal cord stimulation restores locomotion and

desynchronizes corticostriatal beta oscillations [abstract]. Mov Disord. 2018;33.

31.    Pais-Vieira M, Yadav AP, Moreira D, Guggenmos D, Santos A, Lebedev M, Nicolelis MA. A Closed

Loop Brain-machine Interface for Epilepsy Control Using Dorsal Column Electrical Stimulation. Sci Rep.

2016;6:32814. doi: 10.1038/srep32814. PubMed PMID: 27605389; PMCID: PMC5015048.

32.    Hanson TL, Omarsson B, O'Doherty JE, Peikon ID, Lebedev MA, Nicolelis MA. High-side digitally

current controlled biphasic bipolar microstimulator. IEEE Trans Neural Syst Rehabil Eng. 2012;20(3):331-

40. Epub 2012/02/14. doi: 10.1109/TNSRE.2012.2187219. PubMed PMID: 22328184; PMCID:

PMC3502026.

33.    Pais-Vieira M, Chiuffa G, Lebedev M, Yadav A, Nicolelis MA. Building an organic computing device

with multiple interconnected brains. Sci Rep. 2015;5:11869. doi: 10.1038/srep11869. PubMed PMID:

26158615; PMCID: 4497302.

34.    Krucoff MO, Zhuang K, MacLeod D, Yin A, Byun YW, Manson RJ, Turner DA, Oliveira L, Lebedev

MA. A novel paraplegia model in awake behaving macaques. Journal of neurophysiology.

2017;118(3):1800-8. Epub 2017/07/14. doi: 10.1152/jn.00327.2017. PubMed PMID: 28701540; PMCID:

PMC5596146.

**Acknowledgements**

We thank Gary Lehew for assistance with experimental setup, Tamara Phillips for assistance with monkey handling, Paul Thompson for assistance with primate task software, Laura Oliveira and Susan Halkiotis for technical assistance, and Joseph O'Doherty for comments on previous version of manuscript. This research was supported by Duke Institute for Brain Sciences Germinator Award and Duke Neurosurgery Research Support offered to Amol Yadav, NIH R25 awarded to Max Krucoff, and Hartwell Foundation grant awarded to Miguel Nicolelis.

**Author Contributions**

A.P.Y., S.L., M.O.K., and M.A.L, designed experiments, A.P.Y. and S.L. performed rodent experiments and surgeries, M.O.K and M.M.A performed primate surgeries, A.P.Y performed primate experiments, A.P.Y. and S.L. analyzed data, A.P.Y. drafted manuscript, all authors edited manuscript, M.A.L.N. supervised work.

**Competing financial interests**

The authors declare no competing financial or non-financial interests

**Supplementary Information and Figures**

Primate implant surgery and timeline

Two weeks prior to surgery, the animal was acclimated to a protective vest. One day before surgery, the animal's back was shaved with an electrical clipper under mild sedation. The day of surgery, the animal underwent general anesthesia and endotracheal tube placement by staff veterinarians. The animal was then positioned prone on a radiolucent fluoroscopy table. The spine was placed into gentle flexion to open the interspinous spaces by placing the knees and pillows under the abdomen of the monkey. Pressure points were carefully padded. The surgical area was cleaned and prepped with 2% chlorhexidine gluconate and 70% isopropyl alcohol solution (ChloraPrepTM) and allowed to dry for 3 minutes. The surgical area was then draped in a sterile fashion. A dose of antibiotics was given within one hour prior to incision.

An anterior-posterior (AP) C-arm fluoroscopic x-ray was then taken to mark the level of the lowest rib, and the radiographic shadows of the spinous processes were positioned in the middle of the pedicles to ensure minimal rotation. After injecting a local anesthetic (1.5% lidocaine with 1:100,000 epinephrine), a 2cm incision was made at this location down to the fascia, and a self-retaining retractor was placed to hold the skin open. A trocar was then introduced through the fascia just lateral to midline between the spinous processes and advanced so the tip just entered the epidural space. Epidural location was confirmed using a loss of resistance technique, fluoroscopic x-ray, and lack of cerebrospinal fluid (CSF) (34). Next, an 8-contact cylindrical percutaneous lead was introduced through the trocar and steered through dorsal epidural space to a target location in the mid-thoracic spine, or approximately across T8 (Supplementary Figure 1b). Once in a satisfactory position, the stylet and trocar were removed. The lead was then secured to the fascia using a 2-0 silk suture, and then a subcutaneous suprafascial pocket was created off midline to secure lead excess. A small hole was then created in the skin off midline to externalize the distal end of the lead. The same procedure was then repeated on the opposite

site, ultimately positioning the electrodes next to one another in the midline dorsal epidural

thoracic space with a slight cranio-caudal offset (Supplementary Figure 1c).

Once both electrodes were externalized, the incision was closed with inverted dermal vicryl

sutures and a running 4-0 nylon suture. The incision was then covered with a thin layer of

bacitracin ointment, and piece of telfa was then placed over the incision and stapled to the skin.

Next, a custom plastic cap was sutured to the animal's skin with 3-0 nylon sutures to cover the

externalized leads (Supplementary Figure 1d). Such a cap allowed for access to the leads by the

researcher but protection from the animal. The telfa was removed 3 days after surgery. Surgical

sutures were removed 7-10 days after surgery. The sutures in the cap needed to be replaced

approximately monthly under sedation.

Monkey M's right lead stopped working a few days post-implant while the left lead stopped

working approximately 45 days post-implant. Hence the leads were explanted thereafter.  Monkey

O's left lead stopped working a few days post-implant while the right lead worked for 135 days

after which they were extracted due to a skin infection at the externalization site. Monkey K's both

left and right leads were working until they were explanted approximately 150 days after surgery

due to a skin erosion at externalization site.

# Supplementary Figure 1

<u>Supplementary Figure 1: Timeline of primate experiments and primate surgery procedures.</u>

a) Monkeys were trained on the joystick task prior to spinal implant surgery. After they recovered from the

surgery behavioral training began for detection of SCS-based sensations. After that sensory detection

psychophysical experiments were performed followed by training on discrimination of SCS-based

sensations. Monkey M learned the SCS detection task, however, 45 days after surgery its implanted spinal

electrodes stopped working and hence were explanted. Spinal implants of Monkey O and K lasted until

they both were trained on the SCS sensory discrimination task. However, 135 and 150 days after surgery

for both monkeys the implants stopped working and were subsequently explanted. b) Anterior-posterior

(AP) fluoroscopic x-ray showing the introduction of the 8-contact cylindrical lead through the trocar into the

mid-thoracic dorsal epidural space. Blue arrow: trocar, Green arrow: 8-contact electrode array, Orange

arrow: spinous process. c) Final AP (left) and lateral (right) x-rays showing placement of dorsal mid-thoracic

epidural electrodes with slight cranio-caudal offset. d) Externalized electrodes with custom plastic cap

sutured to the skin, and incision temporarily covered with telfa stapled to the skin. Cap open (left), cap

closed (right).

**Supplementary Figure 2:**

<u>Supplementary Figure 2: Psychophysical evaluation of SCS sensory detection in monkey 'O' with varying</u>

<u>stimulation parameters.</u> Once monkey O learned to detect SCS stimuli at the standard parameters

(frequency: 100 Hz, pulse width: 200 µs, duration: 1 sec), we allocated different blocks of sessions where

(pulse width & amplitude; panel 'a' ); (frequency & amplitude; panel 'b'); and (duration & amplitude; panel

'c') were varied while keeping other parameters constant. Psychometric curves in a-c are sigmoidal fits to

fraction of trials that were successfully detected as a function of amplitude.

**Supplementary Figure 3**

Supplementary Figure 3: Charge versus charge density and sensory versus motor thresholds.    a) Sensory

thresholds for monkeys (blue circles) and rats (red circles) determined by psychophysical evaluation at

multiple stimulation parameters of pulse-width, frequency, and duration were below tissue damaging levels.

Black dotted line indicates boundary between damaging and non-damaging stimulation defined by k = 1.85

in the Shannon equation [log(D) = k – log (Q)] on a log charge density (D) versus log charge per phase (Q)

plot. b) Sensory thresholds were considerably lower than the amplitude levels where muscle twitches or

skin fluctuations were observed on the animals' body. Purple and turquoise circles represent multiple

electrode pairs tested in monkeys O and K respectively.

 **Supplementary Figure 4**

Supplementary Figure 4: Rats learned to discriminate SCS stimuli varying in frequency. a) Behavioral setup for frequency discrimination task. Rats had to choose left reward port for higher frequency of SCS and right reward port for lower frequency of stimulation. Stimulation was delivered at a constant amplitude at the standard parameters (pulse width: 200 µs, duration: 1 sec) and rats had to choose the correct reward port associated with each frequency to obtain water reward. b) Learning curves for five rats that learned to discriminate two frequencies (10 Hz vs 100 Hz). Curves indicate sigmoidal fits to fraction of trials that were successfully discriminated as a function of session number.

757     **Supplementary Table 1**

| Animal | Amplitude | Pulse Width | Frequency | Duration |
|---|---|---|---|---|
| Monkey O | 50 – 800 µA, 50 µA step | 100 – 400 µs | 20 – 500 Hz | 50 – 500 ms |
| Monkey K | 50 – 800 µA, 50 µA step | 50 – 400 µs | 10 – 500 Hz | 50 – 1000 ms |
| Rat 1 | 50 – 325 µA, 25 µA step | 50 – 400 µs | 10 – 500 Hz | 50 – 1000 ms |
| Rat 2 | 50 – 325 µA, 25 µA step | 50 – 400 µs | 10 – 500 Hz | 50 – 1000 ms |
| Rat 3 | 50 – 325 µA, 25 µA step | 50 – 400 µs | 10 – 500 Hz | 50 – 1000 ms |
| Rat 4 | 50 – 187.5 µA, 12.5 µA step | 50 – 400 µs | 10 – 500 Hz | 50 – 1000 ms |
| Rat 5 | 50 – 175 µA, 12.5 µA step | 50 – 400 µs | 10 – 500 Hz | 50 – 1000 ms |

Supplementary Table 1: Stimulation parameter ranges tested at psychophysical evaluation of

sensory detection. Animals were initially trained to detect SCS stimuli at standard stimulation

parameters – frequency: 100 Hz, pulse-width: 200 µs, and duration: 1000 ms) as shown in Figures

2a and 4c. After that sensory thresholds were determined by using stimulation settings shown

above. In blocks of consecutive sessions, amplitude and frequency or amplitude and pulse-width

or amplitude and duration of stimulation were varied while keeping other parameters constant. In

additional block of sessions in monkey K, frequency (10 – 1000 Hz) and duration (1 – 2000 ms)

of stimulation were simultaneously varied while keeping pulse width and amplitude constant at

200 µs and 325 µA respectively.