Immunomagnetic Sequential Ultrafiltration (iSUF) Platform for Enrichment and Purification of Extracellular Vesicles from Biofluids

Jingjing Zhang,[1#] Luong TH Nguyen,[1#] Richard Hickey,[1] Nicole Walters,[1] Andre F. Palmer,[1] Eduardo Reátegui[1,2*]

[1] William G. Lowrie Department of Chemical and Biomolecular Engineering, The Ohio State University, Columbus OH 43210

[2] Comprehensive Cancer Center, The Ohio State University, Columbus, OH 43210

*corresponding author: reategui.8@osu.edu

[#] equally contributed

Abstract

Extracellular vesicles (EVs) derived from tumor cells have the potential to provide a much-needed source of non-invasive molecular biomarkers for liquid biopsies. However, current methods for EV isolation have limited specificity towards tumor-derived EVs that limit their clinical use. Here, we present an approach called immunomagnetic sequential ultrafiltration (iSUF) that consists of sequential stages of purification and enrichment of EVs (nonspecifically and specifically) in 2h. In iSUF, EVs present in different volumes of biofluids (0.1 mL to 100 mL) can be significantly enriched (up to 1000 times), with 99.9 % removal of contaminating proteins (e.g., albumin). The yields of cell culture media (CCM), serum, and urine EVs corresponded to 98% ± 3.6%, 94% ± 2.0% and 95% ± 2.0%, respectively (p > 0.05). The final step of iSUF enables the separation of tumor-specific EVs by incorporating immunomagnetic beads specific to a target subpopulation of EVs. Serum from a small cohort of clinical samples from metastatic breast cancer patients and healthy donors were processed by the iSUF platform and the isolated EVs from patients showed significantly higher expression levels of breast cancer biomarkers (i.e., HER2, CD24, and miR21).

Introduction

Extracellular vesicles (EVs) are increasingly recognized as relevant diagnostic and therapeutic entities that are present in different biofluids[1]. EVs are lipid particles with sizes that vary from 50 nm to a few microns[2]. EVs are endogenously shed from the surface of cells through distinct mechanisms, leading to different types of vesicles[3]. Multivesicular bodies that contain smaller vesicles can fuse with the plasma membrane to release their internal vesicles (i.e., exosomes, 40 nm to 200 nm)[4]. Larger lipid vesicles can directly bud from the plasma membrane as microvesicles (200 nm to 1 μm)[5]. EVs carry a wide variety of biological cargo, including proteins, RNA, and DNA fragments, giving EVs unique roles in regulating cell-cell communication[6]. Moreover, it has been shown that tumor EVs (tEVs) can tune cellular microenvironments at distant sites to promote angiogenesis, invasiveness, immunosuppression, and metastasis[7–10].

Different proof of concept studies have used tEVs to develop liquid biopsy assays to diagnose and monitor cancer at different stages[11,12]. EVs are more abundant than other circulating biomarkers (e.g., circulating tumor cells), and they are structurally more robust[13]. However, tEVs present in biofluids are surrounded by massive amounts of normal EVs (nEVs; secreted by healthy cells), and other biomolecules (e.g., albumin, lipoproteins, globulins)[14], thus novel purification methods are required to isolate tEVs[15]. A recent survey on the methods used for isolation and characterization of EVs from research laboratories around the world reveals that more than 80% of researchers use ultracentrifugation (UC) for the isolation of EVs and western blotting for protein characterization[16]. Although UC and density gradient methods can be used to process different biofluids, they are labor-intensive, produce protein aggregate contaminants, and

are nonspecific towards EV type (derived from a tumor or normal cells)[17–19]. Moreover, in some cases, additional washing steps (further UC rounds) are required to purify the EVs from contaminant proteins and aggregates[20]. Other EV isolation methods, including polymeric or salt precipitation kits[21], size exclusion chromatography (SEC) columns (e.g., qEVs)[22], and nano/microdevices have limitations[23]. Precipitation kits have low EV recovery rates, lack specificity, and have low purity[24]. qEVs can separate EVs into different size fractions with high purity and low protein contamination, but have low EV recovery rates and are nonspecific for EV subpopulation[25]. Recently, immunoaffinity methods that were developed for cell separation have been adapted for specific EV isolation[26]. Microfluidic and plasmonic devices have been functionalized with antibodies to target different EV populations[27,28]. However, the majority of these approaches target tetraspanins and annexins, which are ubiquitous proteins present in all EVs[29]. Other attempts used epithelial cell adhesion molecule (EpCAM); however, this antigen is also expressed on normal epithelial EVs[30]. Recently, we demonstrated the use of nanostructured polymeric brushes conjugated with epidermal growth factor receptor (EGFR) and integrated into a microfluidic channel to enhance specificity towards tumor-derived EVs isolated from glioblastoma (GBM) patients[31]. Although this approach can achieve a remarkable 94% specificity towards tEVs, the limited amount of biofluid processed (1 to 1.5 mL of serum or plasma) and the retention of albumin, significantly limits its use for proteomics[32].

Compromises have to be made when using a particular technology/methodology for the isolation of EVs[33]. Currently, there is a trade-off between sample volume and specificity in EV isolation technologies that limits quantitative molecular analysis of EV

contents, ultimately impacting the utility of EVs in cancer diagnostics[34]. Here, we present a novel approach termed immunomagnetic sequential ultrafiltration (iSUF) that overcomes current limitations for EV enrichment and purification. iSUF combines three stages of ultrafiltration and immunoaffinity separation: a tangential flow filtration (TFF) step, a standard centrifugal concentration step, and a magnetic-bead antibody-based EV capture step. Using iSUF, we demonstrate that small or large volumes of biofluid can be processed (~ 100 µL or > 100 mL) while concomitantly removing 99.9 % of contaminating proteins (e.g., albumin, lipoproteins, globulins). We have demonstrated the use of iSUF for the enrichment of EVs present in three different types of biofluids: cell culture media (CCM), serum, and urine for which the sample processing time was under 2 h. Another feature of iSUF is that it can enrich EVs up to 1000 times with an EV recovery efficiency higher than 95 %, which overcomes the limitations of other commercially available methods. To further validate the clinical utility of iSUF, we have processed serum samples from 10 metastatic breast cancer patients and demonstrated the presence of HER2, CD24 and miR21 biomarkers at significantly higher levels compared to healthy controls ($p < 0.05$).

Materials and Methods

Materials

Hollow fiber cartridges (molecular weight cut off, MWCO: 500 kDa, material: polysulfone, surface area: 28 cm^2) were supplied by Repligen (Rancho Dominguez, CA). Amicon® ultra-15 centrifugal filter units (MWCO: 10, 30 kDa) were purchased from MilliporeSigma (Burlington, MA). Streptavidin-coated magnetic particles (3.0-3.9 µm) were obtained from Spherotech (Lake Forest, IL). For capturing EVs, Cetuximab was

purchased from ImClone LLC (Branchburg, NJ), EpCAM, CD63, HER2 were obtained from R&D Systems (Minneapolis, MN). Antibodies were biotinylated using an EZ-Link™ micro Sulfo-NHS-biotinylation kit (ThermoFisher Scientific, Waltham, MA). For surface marker EV detection, CD24 was obtained from Novus Biological LLC (Centennial, CO), and HER2 was obtained from Cell Signaling Technology (Danvers, MA).

Cell culture and supernatant collection

U-251 glioblastoma (GBM), MCF-7 breast, and A375 melanoma cancer cell lines were supplied by American Type Culture Collection (ATCC, Manassas, VA). Cell lines were cultured in their recommended culture medium[35] containing 10% FBS and 1% penicillin-streptomycin at 37°C in a 5% CO_2 incubator. For isolation of EVs from CCM, U251, MCF7, and A375 were grown in T75 flasks to 90% cell confluence, followed by washing the cells twice with PBS. Culture medium with 10 % EV-depleted FBS was added to cells for 24 h. CCM was centrifuged at 1, 000 x g for 5 min at room temperature (RT) to discard cell debris before further processing. EV-depleted FBS was prepared by using the permeate of FBS filtered by tangential flow filtration (TFF) (MWCO: 300 kDa). The concentration of EVs present in the purified FBS was not detectable when a tunable resistive pulse sensing method was used as described below.

Healthy donor serum collection

10 mL of whole blood from healthy donors was collected into BD SST serum tubes (Thermo Fisher Scientific, Waltham, MA). Tubes were rocked 10 times and then gently placed upright to coagulate for 60 min. Then, the tubes were centrifuged at RT at 1,100 $\times$ g for 10 min. The serum was subsequently aspirated carefully and stored in 1 mL aliquots at -80 ºC. Prior collection of biofluids from donors, informed consent was

obtained. All blood samples were collected under an approved Institutional Review Board at The Ohio State University (IRB# 2018H0268). All methods were carried out in accordance with relevant guidelines and regulations.

Healthy donor urine collection

Urine was also collected from the same healthy donors above, by either a first-morning or second-morning standard collection protocol[36]. The urine volume collected was 10 to 100 mL. Urine was collected in sterilized 50-mL centrifuge tubes containing 4.2 mL protease inhibitor - a mixture of 1.67 mL 100 mM sodium azide (NaN_3), 2.5 mL phenylmethylsulfonyl fluoride (PMSF) and 50 µl Leupeptin (Millipore Sigma)[37]. After collection, urine samples were frozen at -80 °C until processing time. Prior collection of biofluids from donors, informed consent was obtained. All urine samples were collected under an approved Institutional Review Board at The Ohio State University (IRB# 2019C0189). All methods were carried out in accordance with relevant guidelines and regulations.

Cancer patient samples

1 mL of serum was collected from 10 metastatic breast cancer patients. Samples were stored at -80 ºC until use. All patient samples were collected from the biospecimens biobank through the Total Cancer Care (TCC) Program at the James Comprehensive Cancer Center at The Ohio State University.

Processing biofluids using the iSUF platform

The schematic workflow of the iSUF platform is shown in Fig. 1. In stage 1, TFF was used to concentrate and diafiltrate the biofluid. During sample concentration, fluid

from the sample feed reservoir was removed as filtrate/permeate from the TFF filter. Diafiltration is a fractionation process that removes smaller molecules (filtrate/permeate) through the filter, and leaves larger molecules in the reservoir by adding a diafiltration solution into the reservoir at the same rate as the filtrate is generated. Briefly, a TFF pump circulates the biofluid through a hollow fiber filter cartridge at a controlled flow rate. Sample fractionation depends on the hollow fiber membrane pore size (MWCO), which should be large enough to permeate proteins and free nucleic acids while small enough to retain EVs. During the concentration step, freely permeable molecules are partially removed. To remove the remaining contaminants, a diafiltration step with PBS is necessary. The diafiltration processing time is proportional to the biofluid volume in the system[38], so diafiltration started with a total biofluid volume of 7 mL, which was the sum of the dead volume of the KrosFlo® KR2i TFF system (2 mL in the product container) and 5 mL remaining in the tubing that can be subject to further processing later. The liquid remaining in the tubing was necessary to protect the EVs from drying out and to enable constant volume diafiltration. Hence, CCM and urine were pre-concentrated to a total volume of 7 mL. For the processing of serum, 0.5 mL of sample was diluted to a total volume of 7 mL in PBS and then processed with diafiltration. The input flow rate was kept at 35 mL/min using a peristaltic pump (Cole-Parmer, Vernon Hills, IL). The sample volume after stage 1 was approximately 2 mL (dead volume of the KrosFlo® KR2i system). At stage 2, ultra-centrifugal units with 10 kDa MWCO were used to further concentrate the samples to 100 µL at 3,000 × g for 20 min. For specific isolation of subpopulations of EVs, stage 3 of iSUF, Streptavidin-coated magnetic beads were functionalized with biotinylated capture antibodies (e.g., EpCAM, HER2, EGFR) overnight at 4 ℃ to target tEVs in spiked

samples as well as patient samples. 100 µL of the processed sample (after stage 2) was incubated with the antibody-coated beads for 1 h at RT. The efficiency of the iSUF platform for isolating tEVs was evaluated using flow cytometry and fluorescence microscopy. EVs were also characterized for their size, concentration, morphology, and molecular content (e.g., protein, RNA).

Processing biofluids using ultracentrifugation (UC)

CCM, serum, and urine samples were filtered using a syringe filter (pore size: 1.0 µm) and transferred to ultracentrifuge tubes (Beckman Coulter, Brea, CA) gently using a syringe and blunt needle (Becton, Dickinson and Company, Franklin Lakes, NJ). Ultracentrifuge tubes were sealed with a cordless tube topper (Beckman Coulter) after balancing, then were placed in a Type 55.2 Ti rotor (Beckman Coulter) and centrifuged in the Optima L-80 XP ultracentrifuge (Beckman Coulter) for 90 min at 4 ºC at 100,000 × g for 2.5 h. The supernatants were discarded carefully after UC, and pellets were re-suspended in 100 µL of PBS.

Processing biofluids using commercially available EV isolation kits

Using the Total Exosome Isolation Reagent (TEIR, Invitrogen, Carlsbad, CA), EVs were isolated from 0.5 mL serum according to the manufacturer's instructions. Briefly, 0.5 mL of serum was mixed with a proprietary reagent provided in a kit and incubated for 30 min at 4 ºC. After mixing, the sample was centrifuged at 10,000 × g for 10 min at RT. EVs pellets were resuspended in 100 µL PBS. For size-exclusion chromatography, 0.5 mL serum was loaded into a qEV column (Izon Science, Medford, MA), and flushed with PBS, fractions 7-12 were collected and pooled in 3 mL according to the manufacturer's instructions.

Immunofluorescence staining

After EVs were captured on the functionalized magnetic beads using a cocktail of a recombinant chimeric EGFR monoclonal antibody (Cetuximab, Erbitux®, ImClone LLC, Branchburg, NJ), a goat EpCAM/TROP-1 polyclonal antibody (#AF960, R&D Systems, Minneapolis, MN) and a goat ErbB2/Her2 polyclonal antibody (#AF1129, R&D Systems), they were blocked with 3% (w/v) BSA and 0.05% (v/v) Tween® 20 in PBS for 1 hr at RT, and then stained with either a rabbit HER2/ErbB2 monoclonal antibody – PE conjugate (#98710S, Cell Signaling Technology, Danvers, MA) or a mouse CD24 monoclonal antibody - Alexa Fluor™ 594 conjugate (#NB10077903AF594, Novus Biologicals™) for 1 hr at RT.

Molecular beacon design and quantification

Molecular beacon (MB) (listed 5′–3′) targeting miR-21 used in this study was T+CA A+CA /iCy3/ +TCA +GT+C T+GA TAA GCT AAC TTA TCA GAC TGA /3BHQ_2/. Locked nucleic acid (LNA) nucleotides (positive sign (+) bases) were incorporated into oligonucleotide strands to improve the thermal stability and nuclease resistance of MBs for incubation at 37 °C. The designed MBs were custom synthesized and purified by Sigma-Aldrich. An aqueous solution of MBs in PBS was vigorously mixed with a lipid formulation of dioleoyl-3-trimethylammonium propane (DOTAP), cholesterol, phosphatidylcholine (POPC) and 1, 2-distearoyl-sn-glycero-3-phosphoethanolamine-poly(ethylene glycol) (DSPE- PEG) in 200 proof ethanol, and then sonicated for 5 min using an ultrasonic bath. The MB/lipid mixture was subsequently injected into PBS, vortexed, and sonicated for 5 min. Finally, it was dialyzed with a 20 kDa MWCO dialysis

bag to remove free MBs. After EVs were captured on magnetic beads, they were incubated with the prepared MBs for 2 hr at 37 °C before imaging.

Flow cytometry and fluorescence microscopy.

tEVs from U251 GBM cells were stained with a lipophilic fluorescent dye, SP-DiOC18(3) (ThermoFisher Scientific) for 20 min, the excess dye was washed out. 100 µL fluorescent EVs (~10^{11} particles/mL) were then spiked into 500 µL of serum (~10^{12} particles/mL) and 100 mL of urine (~10^9 particles/mL) from healthy donors. Each of the samples was processed by iSUF and recovered in 100 µL of PBS (stage 1 to 3). Non-spiked fluorescent tEVs were also captured on functionalized beads as a positive control. After washing with PBS, the captured EVs were analyzed by imaging flow cytometry (Amnis, ImageStream^X Mark II Imaging Flow Cytometer, Luminexcorp, Austin, TX); also images were taken using a fluorescence microscope (Nikon Eclipse Ti Inverted Microscope System) with a 100✕ oil immersion lens. For comparison, samples were also processed by UC and resuspended in 100 µL PBS. For total RNA quantification, captured EVs were lysed, and RNA was extracted and quantified using the same procedures as mentioned above.

Statistical Analysis.

Data are expressed as the mean ± STD. A significant test between different mean values was evaluated using one-way ANOVA in JMP Pro 16 software provided by The Ohio State University. Differences between samples were considered statistically significant for $p < 0.05$.

Results

Optimization of the iSUF platform

To overcome current limitations of the enrichment and purification of EVs and on-demand EV subpopulation characterization, we developed the iSUF platform (**Fig. 1 A**) which includes three stages: (1) tangential flow filtration (TFF) for the concentration and purification of EVs, (2) centrifugation for volume reduction and further enrichment of EVs, and (3) immunomagnetic affinity selection for desired EV subpopulation isolation. We used the iSUF platform to process various volumes (0.1-100 mL) of different types of biofluid (CCM, serum, and urine).

To design stage 1, many parameters of TFF processing required optimization, including the selection of membrane pore size (MWCO), sample processing temperature, sample flow rate, pressure, and sample protein concentration. We tested membrane filters with two MWCO sizes (300 and 500 kDa) to determine the optimal MWCO that maximizes the removal of free proteins and nucleic acids while reducing processing time. Our experiments showed that 500 kDa filter membranes were able to remove 99.9 % of free proteins, with over 99 % of EV recovery. When a 300 kDa membrane filter was used, only 80 % of free proteins were removed. Moreover, a 500 kDa membrane filter was chosen since it processed samples 2~3 times faster than a 300 kDa membrane filter (**Supplementary Table 1**). The TFF stage 1 of iSUF was run at 4°C to minimize EV degradation[39]. We further tried to optimize sample processing time, which was highly dependent on the flow rate. The flow rate was linearly associated with the shear rate generated by the filter based on the manufacturer's protocol (Repligen), which exerted a shear force on the EVs. Then, we used a flow rate of 35 mL/min to maintain a shear rate below 5000/s[40]. High flow rates increased the system pressure, mainly when the protein concentration of the sample was high (> 15 mg/mL). We kept the pressure of the system

below 10 psig to avoid leakage and maximize the lifespan of the 500 kDa filter. We used dilutions of fetal bovine serum (FBS) to test the effect of protein concentration on system pressure at 35 mL/min. Our results showed that protein concentration must be equal or lower than 15 mg/mL to maintain the pressure of the system below 10 psig to protect the filter (**Supplementary Fig. 1**).

At stage 2 of iSUF, EVs were loaded into centrifugal filter units and were spun down at 3,000 x *g*. We compared the recovery rates of EVs and processing time for different filter pore sizes. A 3 kDa filter unit obtained over 99% recovery rate and took 60 min to spin down, while a 10 kDa unit obtained a 95% recovery rate in 20 min, and a 30 kDa filter obtained only a 70% recovery rate and took 15 min to spin down (**Supplementary Fig. 2**). We selected the 10 kDa filter unit to maintain a high recovery rate while reducing sample processing time. To enrich tEVs (stage 3), we tested incubation of 3 µm magnetic beads with different concentrations of biotinylated antibodies (10, 20, 100 µg/mL) for 1 h and 2 h at RT, and overnight at 4 °C. Overnight incubation with 100 µg/mL of antibody exceeded the bead-antibody binding efficiency, while 20 µg/mL of antibody was able to yield a high bead-antibody binding efficiency (>90%). Then, we examined the ratio of beads to EVs at 5/100, 20/100, and 80/100 µL during the bead/tEVs incubation step. The 20/100 µL ratio achieved the highest bead-tEVs capture efficiency (>90%).

iSUF platform for processing biofluids

A flow rate of 35 mL/min for stage 1 of iSUF was applied since the protein concentrations of the different biofluids were below 15 mg/mL (**Supplementary Table 2**). We tested the ability of our platform to purify and concentrate EVs from three different

biofluids (i.e., CCM, serum, and urine; **Fig. 1B**). 50 mL of CCM, and 100 mL of urine were concentrated to 7 mL achieving 76 ± 5% and 78 ± 6% of free protein and nucleic acid removal. Subsequently, 150 mL and 200 mL of PBS diafiltration buffer were used to remove the remaining protein contaminants from CCM and urine in 80 min and 100 min, respectively (> 99.9%). For serum, the initial high concentration of proteins (> 80 mg/mL) required an initial dilution of 0.5 mL of the sample in 7 mL of PBS. Subsequently, 300 mL of PBS diafiltration buffer was used to remove the remaining protein contaminants from serum in 120 min (> 99.9%). We first tested iSUF with a 10% BSA solution for which an SDS-PAGE gel showed extensive removal of albumin (**Supplementary Fig. 3**). Moreover, analysis of the purified samples by an SDS-PAGE gel indicated that iSUF removed BSA from CCM, human serum albumin (HSA) and globulins from serum, and Tamm-Horsfall glycoprotein (THF) and HSA from urine. (**Supplementary Fig. 4, 5, 6**).

Concentration, size distribution and microscopy characterization of EVs

The enrichment factor (EF) was determined as the ratio of the concentration of EVs pre-processing (original biofluid) to the concentration of EVs post-processing by iSUF (in 100 µL). For 50 mL of CCM, 0.5 mL of serum, and 100 mL of urine, the EFs were 489 ± 18, 4.8 ± 0.1, and 942 ± 19, respectively (n = 5; **Fig. 2A**). Accordingly, the efficiency of EV recovery rate (the ratio of the total number of EVs post-processing by iSUF to the total number of EVs pre-processing) was 98% ± 3.6%, 95% ± 2.0% and 94% ± 2.0% for CCM, serum, and urine, respectively. Considering that EVs are heterogeneous in size, we also tested the EF across a wide size range of EVs (40 nm - 1µm) in CCM with similar results (n = 5; p > 0.05; **Fig. 2B**). Moreover, the EV concentrations obtained by iSUF was greater than the concentrations obtained by UC. This difference was consistent across

the 40 nm to 1 µm size range (n = 5; $p < 0.05$; **Fig. 2C**), with iSUF enriching EVs at two to three orders of magnitude higher than UC (approximately 10^{11} EVs concentrated by iSUF and 10^9 for UC). Similar results were obtained with comparisons for the enrichment of EVs present in serum and urine. With iSUF, EVs from serum and urine were enriched almost at the same level (10^{12} EVs/mL; **Fig. 2D**). We confirmed the presence of EVs by using atomic force and electron microscopy on the different biofluids processed. The majority of isolated EVs exhibited a round morphology with heterogeneous size distribution. Cryo transmission electron microscopy images (TEM) images of isolated EVs showed the presence of a double-layered lipid membrane, a representative characteristic of EVs. (**Fig. 2E & 2F**).

To further test our iSUF platform, we performed comparative studies with different commercially available EV isolation methods: qEV, TEI, and UC. 0.5 mL of serum from healthy donors were processed with different EV isolation platforms. EVs demonstrated similar smaller and larger size distribution for all platforms, but iSUF obtained a higher EV concentration compared to the other methods within the 40 nm - 1 µm size range (**Fig. 3A**). We also compared the mean size of subpopulations of EVs based on their physical characteristics using different methods[41]. **Fig. 3B** shows the mean size of small EVs (sEVs) and medium/large EVs (m/lEVs). For both size ranges, EVs processed by iSUF were significantly smaller than UC (n = 5; $p < 0.05$). For l/m EVs, EVs purified by UC were larger than all other techniques (n = 5; $p < 0.05$). Then, we compared the total concentration and purity of isolated EVs using different methods. iSUF enriched EVs significantly more efficiently than all the tested methods (n = 5; $p < 0.05$; **Fig. 3C**). The concentration of EVs enriched by iSUF was 51 ± 23, 7.0 ± 2.0, and 56 ± 20 times higher

than qEV, TEI, and UC, respectively. The purity of isolated EVs was normalized and evaluated in terms of the ratio between EV concentration and remaining contaminating protein concentration present in samples after purification using the different methods. The purity of the isolated EVs by iSUF was up to 100-10000 times higher than the different methods tested (**Fig. 3C**).

Molecular content quantification and characterization of EVs

We quantified the total amount of protein and RNA present in EVs isolated by iSUF for different biofluids. For CCM, the quantity of RNA obtained was 11 ± 8.2 ng/mL, thus giving a 9-fold higher concentration of RNA when compared to UC (**Fig. 4A**). For protein analysis, the protein concentration was 6.2 ± 1.8 µg/mL, almost 8-fold more protein obtained than when the same biofluid was processed by UC (**Fig. 4B**). The difference in protein concentration was also demonstrated using CD63 and CD 9 western blot analysis (**Fig. 4C**). Total RNA and protein quantification were carried out for serum and urine from 5 healthy donors. 0.5 mL of serum and 100 mL of urine processed by iSUF produced 54 ± 40 ng/mL and 0.3 ± 0.1 ng/mL of RNA, respectively (**Fig. 4D**). The protein concentration was 1250 ± 480 µg/mL for serum and 3.1 ± 3.0 µg/mL for urine (**Fig. 4E**). We verified the presence of EVs using western blotting for CD63 and CD9 biomarkers (**Fig. 4F**). Moreover, we compared the RNA content obtained from EVs enriched from 0.5 mL of serum using different commercially available methods. The total RNA content obtained using iSUF was 27 ± 20 ng, which was significantly higher than other methods that only obtained 10 ± 9.8 ng (**Fig. 4G**).

Next, we processed different volumes of serum and urine with iSUF to identify the equivalent volumes of biofluid that will produce comparable concentrations of EVs, total

RNA, and total protein. We started with 0.5 mL of serum and different volumes of urine (i.e., 100, 75, 65,10 mL). For different volumes of urine and 0.5 mL of serum, the enriched concentrations of EVs were comparable ($\sim 10^{12}$, **Supplementary Fig. 7**). The total RNA content obtained from EVs from serum and urine also exhibited comparable values of 60 ± 35 ng and 75 ± 40 ng, respectively. However, analysis of protein content of the isolated EVs from both biofluids showed that the protein content of EVs isolated from urine was 10-fold lower than the protein content of EVs in serum, which was 0.4 ± 0.1 mg and 40 ± 10 mg, respectively.

Immunomagnetic affinity selection of the iSUF platform

tEVs are surrounded by a large amount of normal EVs (nEVs) that require removal to perform accurate molecular analysis[42]. One way to isolate tEVs is to exploit the presence of specific surface markers[43]. Our iSUF platform enables the separation of subpopulations of EVs by capturing them on magnetic beads through immunoaffinity. To demonstrate our approach, we used a model system that consisted of spiking fluorescently labeled tEVs from a cancer cell line (i.e., U251 EVs) in 0.5 mL of serum and 100 mL of urine from healthy donors (n = 3). 3-μm magnetic beads were functionalized with EGFR as the capture antibody. All samples were processed through all stages of iSUF, and the captured subpopulation of EVs was characterized and quantified using TIRF and imaging flow cytometry (**Fig. 5A**). A comparable fluorescent signal was obtained between the positive control (tEVs directly from U251 cell supernatant) and spiked samples processed in serum and urine, which verified the high recovery efficiency of tEVs (88 ± 1 for serum, and 93 ± 2 for urine), and their purity. The biofluids spiked with tEVs were also processed by UC, which yielded significantly lower fluorescent signal (3.1

± 0.1 times lower, **Fig. 5A**). Furthermore, to determine tEV RNA isolation efficiency by iSUF, lysed U251-EVs (positive control) and bead-isolated U251-EVs were extracted. iSUF showed over 90% RNA isolation efficiency for the EGFR+ tEVs (**Fig. 5B**). Moreover, the RNA content of bead-extracted tEVs from serum using the iSUF platform was 7-fold higher than UC, which was mainly attributed to the high yield of total EVs recovered in iSUF (**Fig. 5C**). The high RNA content of tEVs from serum isolated using iSUF also confirmed the ability of iSUF to enrich and retrieve tumor targets from mixed populations of EVs.

Detection of proteins and miRNA in EVs from clinical samples

The emergence of targeted therapies require a precise characterization of the molecular subtypes of a patient's tumor[44]. For breast cancer (BC) patients (e.g., luminal A/B, triple-negative), the molecular subtype is typically determined by testing a tissue biopsy for the presence or absence of three essential proteins: estrogen receptor (ER), progesterone receptor (PR), and human epidermal growth factor receptor 2 (HER2)[45]. However, a tissue-based tumor profile is subject to sampling bias, provides only a snapshot of tumor heterogeneity, and cannot be obtained repeatedly[46]. These limitations restrict early cancer detection and significantly contribute to the overdiagnosis and overtreatment of breast cancer patients[47]. The analysis of EVs present in biofluids of cancer patients provides a minimally invasive method to quantify different biomarkers that would enable precise diagnosis or response to treatment[48]. To test the potential use of the iSUF platform in screening biomarkers for BC, we processed 0.5 mL of serum from 10 metastatic BC patients and quantified the expression levels of HER2, CD24, and miR21 on patient isolated EVs[49–51]. After stage 3 of iSUF, magnetic beads were isolated

using a magnet and incubated with detection antibodies and MBs. Then, the magnetic beads were processed with imaging flow cytometry and TIRF microscopy. Through imaging analysis (**Fig. 6A**), we obtained representative histograms that quantify the number of beads and their corresponding fluorescent intensity (**Fig. 6A**). Similar experiments were carried out with healthy controls for which the total number of positive beads (i.e., nonspecific binding) was significantly lower ($p < 0.05$). To compare both cohorts, we defined a total fluorescence intensity quantity as the area under the curve for the histograms obtained for BC patients and healthy controls. Both protein and miRNA show significantly higher expression in BC patient samples compared to healthy controls ($p < 0.05$; **Fig. 6B**), which suggests that iSUF can differentiate tumor biomarkers between BC patients with healthy controls (**Supplementary Table 4)**.

Discussion

Recently, EVs have been explored for diagnostic and therapeutic applications, including liquid biopsy assays for cancer diagnostics, and nanocarriers for drugs and nucleic acids[52,53]. The large number of tEVs compared to other rare biomarkers (i.e., circulating tumor cells, CTCs) makes them more statistical reliable[54]. Innovative methods developed for the enrichment and purification of EVs should remove all contaminants (e.g., free proteins), have a high yield, work amongst different biofluids, and maintain the integrity of the EVs. We have engineered a new platform that includes a TFF enrichment and purification stage, a centrifugal-unit enrichment stage, and a magnetic separation stage for specific isolation of subpopulations of EVs (e.g., tEVs). In stages 1 and 2, iSUF performs enrichment and purification of all EVs (e.g., tEVs and nEVs). In stage 3, tEVs are isolated based on their tumor-specific surface markers. 300 kDa and 500 kDa TFF

filters were initially selected for this study because the molecular weights of the major free proteins in CCM (e.g., BSA, 65 kDa), serum (e.g., HSA, 65 kDa) and urine (e.g., THP, 98 kDa) were below 300 kDa[55–57]. After testing, a 500 kDa TFF filter was finally chosen for subsequent experiments because of its minimal EV loss and faster processing time (~ 2 h).

Different biofluids required different TFF diafiltration buffer volumes to remove the majority of contaminants. CCM and urine contain relatively low concentrations of protein contaminants[58,59], so a small total diafiltration buffer volume (150-200 mL) was required to remove the majority of free proteins. Unlike CCM and urine, serum has a significant amount of free proteins[60]. Therefore, a minimum of 300 mL of diafiltration buffer was necessary to remove 99.9% of the free protein in serum.

Although TFF can concentrate and purify samples, the final product volume is mostly dependent on the dead volume of the specific TFF system (i.e., KrosFlo® KR2i TFF system, 2 mL). Therefore, further enrichment is necessary to concentrate the samples (e.g., spin down to - 100 μL) for applications such as the diagnosis of rare biomarkers[61]. Centrifugal units with different MWCOs (3, 10, and 30 kDa) were evaluated in terms of processing time and yield in our study. A larger MWCO size required shorter centrifugation time, but more EVs were lost because prolonged high-speed centrifugation elongated EVs into an oval shape which made them squeeze through the filter membrane[62,63]. The centrifugal unit with the 10 kDa MWCO was found to be optimal with the shortest processing time and highest yield.

We also compared EVs purified from serum and urine samples of healthy donors. EVs from urine and serum can serve as prognosis biomarkers for clinical analysis[64,65].

Serum is the most commonly used biofluid in the clinical setting with the highest EV concentration[66]. Compared to serum, urine collection is minimally invasive and can be obtained in larger volumes[67], but it usually suffers from much lower EV concentration[68]. Interestingly, using our iSUF platform, the EV RNA concentration in the final products of urine samples (originating from ~ 100 mL of collected sample volume) were comparable to those of serum samples (originating from 0.5 mL of collected sample volume). This suggests that clinical diagnosis by urinary EVs is possible. However, there is still a concern towards urinary EV collection because of large variabilities in the urine volume and its EV concentration. More efforts are needed to come up with a gold standard protocol for urine collection and processing. For protein in urinary EVs, we found that the amount of protein was lower than the amount of protein obtained from serum EVs, which might be explained by the degradation of EV membrane proteins by urine proteases[69].

In this study, we compared the performance of iSUF with three different EV isolation techniques (TEI, qEV, and UC) to determine EV concentration, purity, and quantity of RNA recovered. As other authors have discussed[70,71], we found that UC, a traditional method for EV isolation, has raised concerns about the integrity, yield, and purity of EVs after purification. Interestingly, we found UC-isolated EVs were larger than other platforms, one possible explanation is the presence of extensive levels of protein aggregates that cause bias. Although qEV obtained a relatively pure product, its recovery rate was low; TEI obtained a relatively higher EV number, while retaining massive amounts of free protein. Therefore, low EV concentration and purity will impact the accuracy of molecular analysis of EVs[72].

We are interested in the enrichment and isolation of tEVs to characterize tumor-related proteins and RNAs. Like other immunoaffinity methods[73–75], iSUF captured and isolated EVs using specific tumor surface proteins. iSUF differentiated metastatic BC patients from healthy donors by detecting significantly higher expression levels of proteins and RNA biomarkers present in EVs (e.g., HER2, CD24, and miR21). Based on previous reports, HER2 and miR21 are cancer-associated protein and microRNA species, and are known to be overexpressed in metastatic BCs[76,77]. Compared to miR21 and HER2, CD24 is relatively less investigated in BC but was previously identified as being released from BC stem cells[78]. Furthermore, a recent study indicated that serum CD24 is elevated among BC patients[79]. Moreover, it is important to note that one of these biomarkers may not be a reliable predictor of BC alone. However, the combination of several biomarkers can serve as a tool for BC risk assessment.

In conclusion, iSUF was proposed for rapid, efficient, and specific isolation of EVs from different biofluids. The EV recovery efficiency of iSUF was above 95%, with 90% specific isolation of tEVs and with negligible concentrations of free proteins and nucleic acids. Although iSUF does not process a sample in the shortest time (80 min ~ 120 min), its versatility working with different biofluids, sample volumes, and high purity after sample processing constitute unmatched advantages over current methods used in the field. Overall, we found that the iSUF platform isolated and enriched EVs from a scaled-up sample volume with high purity and yield in a sterile and quick manner simultaneously, and isolated tEVs with high specificity, while other current methods could not guarantee all of those conditions at the same time. Furthermore, we recognize that the iSUF platform

has potentially broad clinical applications beyond liquid biopsies for cancer diagnosis or monitoring.

Conflicts of interest

Acknowledgments

We acknowledge all the patients and healthy volunteers who participated in this study. This work was supported by the U.S. National Institutes of Health (NIH) grants UG3TR002884 (ER). and R01HL126945, R01HL138116, and R01EB021926 (AFP). Additional support for E.R. was provided by the William G. Lowrie Department of Chemical and Biomolecular Engineering and the Comprehensive Cancer Center at The Ohio State University.

Contributions

E.R and A.F.P. conceived the idea. E. R. design the study. J.Z. and L.T.H.N. designed and performed experiments and drafted the original manuscript. J.Z. performed data analysis, prepared all tables and figures with inputs from all authors. R. H. performed equipment installation and calibration. N.W. performed sample collection and supported with editing the manuscript. A.F.P. supervised the TFF experiments and edited the manuscript. E.R. supervised the whole study and edited the manuscript. All authors provided critical feedback and helped to shape the research, analysis and manuscript.

References

1. Raposo, G. & Stoorvogel, W. Extracellular Vesicles: Exosomes, Microvesicles, and Friends. *J. Cell Biol.* **200**, 373 (2013).

2. Minciacchi, V. R., Freeman, M. R. & Di, D. Seminars in Cell & Developmental Biology Extracellular Vesicles in Cancer : Exosomes , Microvesicles and the Emerging Role of Large Oncosomes. *Semin. Cell Dev. Biol.* **40**, 41–51 (2015).

3. Yáñez-Mó, M. *et al.* Biological Properties of Extracellular Vesicles and Their Physiological Functions. *J. Extracell. Vesicles* **4**, 27066 (2015).

4. Théry, C., Zitvogel, L. & Amigorena, S. Exosomes: Composition, Biogenesis and Function. *Nat. Rev. Immunol.* **2**, 569 (2002).

5. Cocucci, E., Racchetti, G. & Meldolesi, J. Shedding microvesicles: artefacts no more. *Trends in Cell Biology* (2009). doi:10.1016/j.tcb.2008.11.003

6. Yáñez-mó, M. *et al.* Biological properties of extracellular vesicles and their physiological functions. **3078**, (2015).

7.   Filipazzi, P., Bürdek, M., Villa, A., Rivoltini, L. & Huber, V. Recent advances on the role of tumor exosomes in immunosuppression and disease progression. *Seminars in Cancer Biology* (2012). doi:10.1016/j.semcancer.2012.02.005

8.   Kahlert, C. & Kalluri, R. Exosomes in tumor microenvironment influence cancer progression and metastasis. *Journal of Molecular Medicine* (2013). doi:10.1007/s00109-013-1020-6

9.   Li, K., Chen, Y., Li, A., Tan, C. & Liu, X. Exosomes play roles in sequential processes of tumor metastasis. *International Journal of Cancer* (2019). doi:10.1002/ijc.31774

10.   Wang, J. *et al.* Multiple myeloma exosomes establish a favourable bone marrow microenvironment with enhanced angiogenesis and immunosuppression. *J. Pathol.* (2016). doi:10.1002/path.4712

11.   Brock, G., Castellanos-rizaldos, E., Hu, L., Coticchia, C. & Skog, J. Liquid biopsy for cancer screening , patient stratification and monitoring. **4**, 280–290 (2015).

12.   Li, A., Zhang, T., Zheng, M., Liu, Y. & Chen, Z. Exosomal proteins as potential markers of tumor diagnosis. *Journal of Hematology and Oncology* (2017). doi:10.1186/s13045-017-0542-8

13.   Zhang, W. *et al.* Liquid Biopsy for Cancer: Circulating Tumor Cells, Circulating Free DNA or Exosomes? *Cell. Physiol. Biochem.* (2017). doi:10.1159/000458736

14. Théry, C., Aled, C., Sebastian, A. & Graça, R. Isolation and Characterization of Exosomes from Cell Culture Supernatants. *Curr. Protoc. Cell Biol.* **3.22**, 1–29 (2006).

15. Li, P., Kaslan, M., Lee, S. H., Yao, J. & Gao, Z. Progress in exosome isolation techniques. *Theranostics* (2017). doi:10.7150/thno.18133

16. Momen-Heravi, F. Isolation of Extracellular Vesicles by Ultracentrifugation. *Methods Mol. Biol.* (2017). doi:10.1007/978-1-4939-7253-1_3

17. Mol, E. A., Goumans, M. J., Doevendans, P. A., Sluijter, J. P. G. & Vader, P. Higher functionality of extracellular vesicles isolated using size-exclusion chromatography compared to ultracentrifugation. *Nanomedicine Nanotechnology, Biol. Med.* (2017). doi:10.1016/j.nano.2017.03.011

18. Yuana, Y., Levels, J., Grootemaat, A. & Sturk, A. Co-isolation of extracellular vesicles and high- density lipoproteins using density gradient ultracentrifugation. **3078**, (2014).

19. Koh, Y. Q., Almughlliq, F. B., Vaswani, K., Peiris, H. N. & Murray, D. Exosome enrichment by ultracentrifugation and size exclusion chromatography by ultracentrifugation (thrombin treated plasma + UC + SEC). 865–874 (2018).

20. Greening, D. W., Xu, R., Ji, H., Tauro, B. J. & Simpson, R. J. A protocol for exosome isolation and characterization: Evaluation of ultracentrifugation, density-gradient separation, and immunoaffinity capture methods. in *Methods in Molecular Biology* (2015). doi:10.1007/978-1-4939-2550-6_15

21. Furi, I., Momen-heravi, F. & Szabo, G. Extracellular vesicle isolation : present and future. **5**, 3–5 (2017).

22. Gámez-Valero, A. *et al.* Size-Exclusion Chromatography-Based Isolation Minimally Alters Extracellular Vesicles' Characteristics Compared to Precipitating Agents. *Sci. Rep.* **6**, 33641 (2016).

23. Kang, Y. *et al.* High-purity capture and release of circulating exosomes using an exosome-specific dual-patterned immunofiltration (ExoDIF) device. 13495–13505 (2017). doi:10.1039/c7nr04557c

24. Rekker, K. *et al.* Comparison of Serum Exosome Isolation Methods for Microrna Profiling. *Clin. Biochem.* **47**, 135 (2014).

25. Stranska, R. *et al.* Comparison of membrane affinity - based method with size - exclusion chromatography for isolation of exosome - like vesicles from human plasma. *J. Transl. Med.* 1–9 (2018). doi:10.1186/s12967-017-1374-6

26. Yang, F., Liao, X., Tian, Y. & Li, G. Exosome separation using microfluidic systems: size-based, immunoaffinity-based and dynamic methodologies. *Biotechnology Journal* (2017). doi:10.1002/biot.201600699

27. Liga, A., Vliegenthart, A. D. B., Oosthuyzen, W., Dear, J. W. & Kersaudy-Kerhoas, M. Exosome isolation: a microfluidic road-map. *Lab Chip* **15**, 2388 (2015).

28. Im, H. *et al.* Label-Free Detection and Molecular Profiling of Exosomes With a Nano-Plasmonic Sensor. *Nat. Biotechnol.* **32**, 490 (2014).

29. Lin, J. *et al.* Exosomes : Novel Biomarkers for Clinical Diagnosis. **2015**, (2015).

30.   Jiang, L. *et al.* EpCAM-dependent extracellular vesicles from intestinal epithelial cells maintain intestinal tract immune balance. *Nat. Commun.* (2016). doi:10.1038/ncomms13045

31.   Reátegui, E. *et al.* Engineered nanointerfaces for microfluidic isolation and molecular profiling of tumor-specific extracellular vesicles. *Nat. Commun.* (2018). doi:10.1038/s41467-017-02261-1

32.   Abramowicz, A., Widlak, P. & Pietrowska, M. Proteomic analysis of exosomal cargo: The challenge of high purity vesicle isolation. *Mol. Biosyst.* (2016). doi:10.1039/c6mb00082g

33.   Ziaei, P., Berkman, C. E. & Norton, M. G. Review: Isolation and Detection of Tumor-Derived Extracellular Vesicles. *ACS Appl. Nano Mater.* **1**, 2004–2020 (2018).

34.   Gurunathan, S., Kang, M.-H., Jeyaraj, M., Qasim, M. & Kim, J.-H. Review of the Isolation, Characterization, Biological Function, and Multifarious Therapeutic Approaches of Exosomes. *Cells* (2019). doi:10.3390/cells8040307

35.   ATCC. Cell Culture Guides. *ATCC online* (2016).

36.   Zhou, H. *et al.* Collection , storage , preservation , and normalization of human urinary exosomes for biomarker discovery. *Kidney Int.* **69**, 1471–1476 (2006).

37.   Pisitkun, T., Shen, R.-F. & Knepper, M. A. Identification and proteomic profiling of exosomes in human urine. *Proc. Natl. Acad. Sci. U. S. A.* **101**, 13368 (2004).

38.   Schwartz, L. & Seeley, K. Introduction to Tangential Flow Filtration for Laboratory and Process Development Applications. *PALL Corp.* (2014).

39. Lee, M., Ban, J., Im, W. & Kim, M. Influence of Storage Condition on Exosome Recovery. **304**, 299–304 (2016).

40. Laboratories, S. MicroKros and MidiKros Hollow Fiber membranes for Tangential Laboratory Separations.

41. Théry, C. *et al.* Minimal information for studies of extracellular vesicles 2018 (MISEV2018): a position statement of the International Society for Extracellular Vesicles and update of the MISEV2014 guidelines. *J. Extracell. Vesicles* (2018). doi:10.1080/20013078.2018.1535750

42. Soung, Y. H., Ford, S., Zhang, V. & Chung, J. Exosomes in cancer diagnostics. *Cancers* (2017). doi:10.3390/cancers9010008

43. Xu, R., Greening, D. W., Zhu, H. J., Takahashi, N. & Simpson, R. J. Extracellular vesicle isolation and characterization: Toward clinical application. *Journal of Clinical Investigation* (2016). doi:10.1172/JCI81129

44. Guo, L. & Guo, N. Exosomes: Potent regulators of tumor malignancy and potential bio-tools in clinical application. *Critical Reviews in Oncology/Hematology* (2015). doi:10.1016/j.critrevonc.2015.04.002

45. Onitilo, A. A., Engel, J. M., Greenlee, R. T. & Mukesh, B. N. Breast Cancer Subtypes Based on ER/PR and Her2 Expression: Comparison of Clinicopathologic Features and Survival. *Clin. Med. Res.* (2009). doi:10.3121/cmr.2009.825

46. Halvaei, S. *et al.* Exosomes in Cancer Liquid Biopsy : A Focus on Breast Cancer. *Mol. Ther. Nucleic Acid* **10**, 131–141 (2018).

47. Shimomura, A. *et al.* Novel combination of serum microRNA for detecting breast cancer in the early stage. *Cancer Sci.* (2016). doi:10.1111/cas.12880

48. Lässer, C. Exosomes in diagnostic and therapeutic applications: Biomarker, vaccine and RNA interference delivery vehicle. *Expert Opinion on Biological Therapy* (2015). doi:10.1517/14712598.2015.977250

49. Krishnamurti, U. & Silverman, J. F. HER2 in breast cancer: A review and update. *Advances in Anatomic Pathology* (2014). doi:10.1097/PAP.0000000000000015

50. Li, S., Yang, X., Yang, J., Zhen, J. & Zhang, D. Serum microRNA-21 as a potential diagnostic biomarker for breast cancer: a systematic review and meta-analysis. *Clin. Exp. Med.* (2016). doi:10.1007/s10238-014-0332-3

51. Sims, A. H., Howell, A., Howell, S. J. & Clarke, R. B. Origins of breast cancer subtypes and therapeutic implications. *Nature Clinical Practice Oncology* (2007). doi:10.1038/ncponc0908

52. Kalluri, R. The biology and function of exosomes in cancer. *Journal of Clinical Investigation* (2016). doi:10.1172/JCI81135

53. De Toro, J., Herschlik, L., Waldner, C. & Mongini, C. Emerging roles of exosomes in normal and pathological conditions: New insights for diagnosis and therapeutic applications. *Frontiers in Immunology* (2015). doi:10.3389/fimmu.2015.00203

54. Sheridan, C. Exosome cancer diagnostic reaches market. *Nature biotechnology* (2016). doi:10.1038/nbt0416-359

55. Wong, C. S. F. *et al.* Optimized exosome isolation protocol for cell culture supernatant and human plasma. *J. Extracell. Vesicles* **4**, 27031 (2015).

56. Grigor'Eva, A. E. *et al.* Contamination of exosome preparations, isolated from biological fluids. *Biomeditsinskaya Khimiya* (2017). doi:10.18097/PBMC20176301091

57. Cheruvanky, A. *et al.* Rapid isolation of urinary exosomal biomarkers using a nanomembrane ultrafiltration concentrator. *Am. J. Physiol. Physiol.* **292**, F1657–F1661 (2007).

58. Langdon, S. P. *Cancer Cell Culture. Cancer Cell Culture* (2003). doi:10.1385/1592594069

59. Yang, C. Y. *et al.* Diagnostic accuracy of urine protein/creatinine ratio is influenced by urine concentration. *PLoS One* (2015). doi:10.1371/journal.pone.0137460

60. Busher, J. T. *Serum Albumin and Globulin. Clinical Methods: The History, Physical, and Laboratory Examinations* (1990).

61. Alix-Panabières, C. & Pantel, K. Clinical prospects of liquid biopsies. *Nat. Biomed. Eng.* **1**, 0065 (2017).

62. Chernyshev, V. S. *et al.* Size and shape characterization of hydrated and desiccated exosomes. *Anal. Bioanal. Chem.* (2015). doi:10.1007/s00216-015-8535-3

63. Skliar, M. & Chernyshev, V. S. Imaging of Extracellular Vesicles by Atomic Force Microscopy. *J. Vis. Exp.* (2019). doi:10.3791/59254

64. Smith, E. R., Zurakowski, D., Saad, A., Scott, R. M. & Moses, M. A. Urinary biomarkers predict brain tumor presence and response to therapy. *Clin. Cancer Res.* **14**, 2378–2386 (2008).

65. Su, Y. Y. *et al.* Upregulated expression of serum exosomal miR-375 and miR-1307 enhance the diagnostic power of CA125 for ovarian cancer. 1–9 (2019).

66. Macías, M., Rebmann, V., Mateos, B., Varo, N. & Perez-gracia, J. L. Comparison of six commercial serum exosome isolation methods suitable for clinical laboratories . Effect in cytokine analysis. (2019).

67. Dear, J. W., Street, J. M. & Bailey, M. A. Urinary exosomes: A reservoir for biomarker discovery and potential mediators of intrarenal signalling. *Proteomics* (2013). doi:10.1002/pmic.201200285

68. Gu, C. *et al.* Clinical significance of urine prostatic exosomal protein in the diagnosis of prostate cancer. **9**, 1074–1078 (2019).

69. Musante, L. *et al.* Proteases and protease inhibitors of urinary extracellular vesicles in diabetic nephropathy. *J. Diabetes Res.* (2015). doi:10.1155/2015/289734

70. Buschmann, D. *et al.* Evaluation of serum extracellular vesicle isolation methods for profiling miRNAs by next-generation sequencing. *J. Extracell. Vesicles* **7**, (2018).

71. Helwa, I. *et al.* A comparative study of serum exosome isolation using differential ultracentrifugation and three commercial reagents. *PLoS One* (2017). doi:10.1371/journal.pone.0170628

72. Syn, N. L., Wang, L., Chow, E. K. H., Lim, C. T. & Goh, B. C. Exosomes in Cancer Nanomedicine and Immunotherapy: Prospects and Challenges. *Trends in Biotechnology* (2017). doi:10.1016/j.tibtech.2017.03.004

73. Yoo, C. E. *et al.* A direct extraction method for microRNAs from exosomes captured by immunoaffinity beads. *Anal. Biochem.* (2012). doi:10.1016/j.ab.2012.09.008

74. Sharma, P. *et al.* Immunoaffinity-based isolation of melanoma cell-derived exosomes from plasma of patients with melanoma. *J. Extracell. Vesicles* (2018). doi:10.1080/20013078.2018.1435138

75. Simpson, R. J., Lim, J. W. E. & Moritz, R. L. Exosomes : proteomic insights and diagnostic potential. **6**, 267–283 (2009).

76. Schwarzenbach, H. Clinical relevance of circulating, cell-free and Exosomal microRNAs in plasma and serum of breast cancer patients. *Oncology Research and Treatment* (2017). doi:10.1159/000478019

77. Hannafon, B. N. *et al.* Plasma exosome microRNAs are indicative of breast cancer. *Breast Cancer Res.* (2016). doi:10.1186/s13058-016-0753-x

78. Ricardo, S. *et al.* Breast cancer stem cell markers CD44, CD24 and ALDH1: Expression distribution within intrinsic molecular subtype. *J. Clin. Pathol.* (2011). doi:10.1136/jcp.2011.090456

79. Rupp, A. K. *et al.* Loss of EpCAM expression in breast cancer derived serum exosomes: Role of proteolytic cleavage. *Gynecol. Oncol.* (2011). doi:10.1016/j.ygyno.2011.04.035

Figure Captions

Figure 1. The Immunomagnetic sequential ultrafiltration (iSUF) platform. A) Schematic representation of iSUF stages 1, 2, 3, including tangential flow filtration (TFF)

purification, centrifugal unit enrichment, and tEV immunoaffinity isolation, respectively. In stage 1, biofluids were processed using a 500 kDa TFF filter, EVs were retained in the retentate and enriched in 7 mL while free proteins and nucleic acids permeated through the TFF filter. Then the PBS valve was opened to start TFF diafiltration until removing 99.9 % of free proteins and nucleic acids. Finally, EVs were recovered in 2 mL of PBS after flushing the TFF system with air. In stage 2, EVs were centrifuged using a 10 kDa centrifugal unit at 3,000 x g and enriched in 100 µL. In stage 3, tEVs were captured using antibodies immobilized to streptavidin-coated magnetic beads (i.e., EGFR) and subsequently pulled out with a magnet. B) iSUF stage 1 purification performance for different biofluids. Cell supernatant (CCM), urine, and serum took 80, 100, and 120 min to obtain 99.9 % efficiency of free protein and nucleic acid removal.

Figure 2. iSUF enrichment performance and characterization of EVs separated from CCM, serum and urine using iSUF. A) Enrichment factor (EF) for EVs present in CCM, serum, and urine after iSUF (n = 5 for each biofluid; p>0.05) (mean ± STD). Enrichment factors (EFs) were calculated as the ratio of EV concentration in biofluids present before and after iSUF processing. B) EFs for different size ranges of EVs for CCM (n = 3; p > 0.05) (mean ± STD). C) qNano measurements of the size distribution and concentration of EVs in the original CCM, after iSUF processing, and after UC processing. The EV concentration after iSUF was significantly higher than in the original CCM and after UC processing (n = 3; p < 0.05) (mean ± STD). D) qNano measurements of the size distribution and concentration of EVs in serum and urine after iSUF processing. E) AFM and SEM images of EVs from U251 CCM after iSUF processing and UC processing. Images were also obtained for EVs in serum and urine after iSUF processing. F) Transmission-EM images of EVs present in CCM after iSUF processing.

Figure 3. Comparison of size distribution and purity of EVs isolated with different platforms. A) Size distribution of EVs isolated from serum using different platforms. iSUF demonstrated the highest EV concentration within the 40 nm – 1 μm size range. B) Left. Mean size boxplot for small EVs (sEVs). sEVs isolated using qEV and TEI demonstrated no differences with iSUF or UC, UC-sEVs showed larger mean size than iSUF-sEVs (n = 5; p < 0.05) (mean ± STD). Right. Mean size boxplot for medium/large EVs (m/l EVs). UC-m/l EVs were larger than iSUF, qEV and TEI m/l EVs (n = 5; p < 0.05) (mean ± STD) C) Concentration and purity of isolated EVs using different platforms. Black boxplots were the absolute concentration of EVs isolated from 0.5 mL of serum; red dots corresponding to the right y-axis were purities defined as the EV concentration divided by the remaining free protein concentration. iSUF isolated EVs from serum most efficiently with high purity

(n = 5; p < 0.05) (mean ± STD), qEV yielded pure EVs, but at a lower concentration, TEI isolated more EVs but at relatively low purity, and UC recovered EVs with the lowest concentration and purity.

Figure 4. Characterization of protein and RNA content of EVs isolated from different biofluids. A) Quantity of total RNA extracted from CCM-EVs using UC and iSUF (n = 6; p < 0.05). B) Comparison of proteins extracted from CCM-EVs using UC and iSUF (n = 6; p < 0.05). C) Western blot analysis of CD63 and CD9 expression in CCM-EVs using iSUF and UC. The expression level of CD63 and CD9 was calculated using an equal mass of protein lysates from samples. D) Quantification of total RNA extracted from serum and urine EVs isolated by iSUF. E) Quantification of EV proteins isolated from

serum and urine EVs isolated by iSUF. F) Western blot analysis of CD63 and CD9 expression in serum and urine EVs isolated by iSUF. G). Total RNA of EVs isolated using different EV isolation methods. Black boxplots were the absolute amount of RNA isolated from 0.5 mL of serum. iSUF isolated the most RNA contents from serum EVs. (n = 5; $p < 0.05$) (mean ± STD).

Figure 5. Characterization of stained tEVs isolated from U251-CCM and spiked in different biofluids. A) Upper. Fluorescence images of tEVs on magnetic beads using TIRF microscopy. Lower. Comparison of fluorescence intensities quantified using imaging flow cytometry. Stained U251-EVs isolated by iSUF was a positive control. tEVs isolated from serum and urine samples using iSUF demonstrated no significant difference (n = 3; p > 0.05), while UC isolated tEVs differed with iSUF and the positive control (n = 3; p < 0.05) (mean ± STD). B) Quantification of total RNA isolated from U251-EVs before and after bead extraction showed no differences (n = 3; p > 0.05). C) Total RNA extracted from serum samples using iSUF was significantly higher than using UC (n = 3; p < 0.05) (mean ± STD).

Figure 6. Characterization of tEVs isolated breast cancer (BC) patient serum using iSUF. A) Left. Characteristic fluorescence images of HER2, CD24 and miR21 expression on isolated EVs immobilized on magnetic beads from BC patients and healthy donors. Right. Fluorescence intensity histograms of HER2, CD24, and miR21 of isolated EVs for BC patients and healthy donors (n = 10; p < 0.05). B) Total fluorescence intensity quantification of HER2, CD24 and miR21 expression level on isolated EVs from BC patients and healthy donors, patients demonstrated higher expression level than healthy donors (n = 10; p < 0.05). P: patient; H: healthy donor.

Supplementary Information

Size distribution and concentration of EVs

A tunable resistive pulse sensing (TRPS) method (Izon Sciences, Boston, MA) was employed for quantification of the size and concentration of EVs. 45 μL of biofluid was pipetted into nanopore membranes (NP150, NP300, NP600, NP800, and NP1000), and then pressure (10 mbar) and voltage (0.38V, 0.32V, 0.26V, 0.18V and 0.12V) were applied. Every single EV causes a resistive pulse that can be used to calculate EV size and concentration. Polystyrene nanoparticles of different known sizes and concentrations were used for calibration.

Atomic force microscopy (AFM)

A clean mica substrate was vapor-phase coated with 3-aminopropyltriethoxysilane (APTES, Millipore Sigma, Burlington, MA) in a vacuum chamber and then dried overnight at 65 °C. Subsequently, 10 μL of purified EVs were incubated on the surface for 30 min at RT. Unbound EVs were extensively rinsed with PBS and then with DI water. The samples were air-dried again before imaging using an AFM (Asylum Research MFP-3D-BIO AFM, Oxford Instruments, Abingdon, United Kingdom).

Scanning electron microscopy (SEM)

Clean coverslips were soaked in 0.25 mg/mL Zetag solution (BASF, Southfield, MI, USA) for 30 min, followed by overnight air drying at RT. Purified EVs were attached to the coated coverslip for 30 min at RT by physisorption. EVs were fixed in 2% glutaraldehyde (MilliporeSigma, Burlington, MA) and 0.1 M sodium cacodylate solution (Electron Microscopy Sciences, Hatfield, PA, USA) for 3h. After washing with 0.1 M sodium cacodylate solution, EVs were incubated in 1% osmium tetraoxide (Electron

Microscopy Sciences) and 0.1 M sodium cacodylate for 2h. The sample was subsequently rinsed with 0.1 M sodium cacodylate solution before dehydration in increasing concentrations of ethanol (50, 70, 85, 95, and 100%, ThermoFisher Scientific) for 30 min each. Next, the samples were transferred to a CO_2 critical point dryer (tousimis, Rockville, MD, USA). Finally, the samples were coated with ∼ 2 nm of gold using a sputtering machine (Leica EM ACE 600, Buffalo Grove, IL) and imaged using SEM (Apreo ii, FEI, Thermo Fisher Scientific).

Transmission electron microscopy (TEM)

3 µL EVs purified from CCM using iSUF platform were applied to a glow discharged lacey carbon coated copper grid (400 mesh, Pacific Grid-Tech, San Francisco, CA) and flash-frozen in liquid ethane using an automated vitrification device (FEI Vitrobot Mark IV, FEI, Hillsboro, OR). The sample was then visualized in a Glacios™ Cryo-TEM (ThermoFisher Scientific).

Protein extraction and quantification

EV samples were lysed in RIPA buffer (Abcam) with the addition of Thermo Scientific™ Halt™ Protease and Phosphatase Inhibitor Cocktails for 15 min on ice. EV samples (with/without lysis) were then pipetted to a 96-well plate and their protein concentrations were quantified using a Pierce™ Rapid Gold BCA Protein Assay kit (ThermoFisher Scientific). EV protein concentration was determined by subtracting the amount of free protein (without lysis) from the total (with lysis) in the purified EV sample.

Sodium dodecyl sulfate-polyacrylamide gel electrophoresis (SDS-PAGE)

Proteins in the final product were denatured and reduced in the presence of NuPAGE Reducing Agent in NuPAGE LDS Sample Buffer at 95 ºC for 10 min. The

proteins were then separated in mini gel tank (ThermoFisher Scientific) using NuPAGE 4-12% Bis-Tris Protein Gel in NuPAGE MOPS SDS Running Buffer for 50 min at 200 V. After separation, the proteins were stained with Coomassie Brilliant Blue G-250 Dye.

Western blotting

After separation by SDS-PAGE, proteins were transferred onto a polyvinylidene fluoride (PVDF) membrane (ThermoFisher Scientific) and then blocked with 3% bovine serum albumin (BSA) and 0.05% Tween in PBS for 1 hr at RT. Primary antibodies against tetraspanin surface markers such as CD63, CD81, and CD9 were incubated with the EVs overnight at 4 °C (Santa Cruz Biotechnology, Inc Dallas, TX, USA). The next day, the PVDF membrane was incubated with an HRP conjugated secondary antibody for 1h at RT. Finally, the sample was incubated with SuperSignal™ West Femto Maximum Sensitivity Substrate for 5 min at RT before imaging using a C-digit blot scanner (LI-COR, Lincoln, NE, USA).

RNA quantification

Total RNA was extracted with QIAzol Lysis Reagent and then purified using a miRNeasy Mini kit according to the manufacturer's protocol (Qiagen, Germantown, MD, USA). After purification, the RNA concentration was quantified using a Qubit microRNA Assay Kit at excitation/emission wavelengths of 500/525 nm.

Supplementary Table 1. Comparison of 300 and 500 kDa TFF filter performance on CCM, urine and serum.

TFF filter MWCO	CCM (50 mL)		Urine (100 mL)		Serum (0.5 mL)	
	Protein removal efficiency	Processing time	Protein removal efficiency	Processing time	Protein removal efficiency	Processing time
300 kDa	>78%	150 min	>80%	200 min	>75%	300 min
500 kDa	>99%	80 min	>99%	100 min	>99%	120 min

Supplementary Figure 1. Pressure in the TFF stage of iSUF processing when processing bovine serum albumin (BSA) solutions at different concentrations at a fixed flow rate of 35 mL/min. 15 mg/mL was the maximum protein concentration in the TFF system to maintain the system pressure below 10 psig.

Supplementary Figure 2. EV recovery rate and processing time of centrifugal units (iSUF, stage 2) with different MWCOs. 3 kDa, 10 kDa, and 30 kDa MWCO centrifugal units obtained over 99%, 95%, and 70% recovery rate, respectively. They took 60 min (3 kDa MWCO), 20 min (10 kDa MWCO) and 15 min (30 kDa MWCO) to spin down the sample to a final volume of 100 µL.

Supplementary Table 2. Protein concentration and removal efficiency of the TFF concentration step and diafiltration step.

	U251 Supernatant	Urine	Serum
Initial protein concentration (mg/mL)	1.20 ± 0.13 (in 50 mL)	1.40 ± 0.50 (in 10~100 mL)	10.0 ± 2.0 (dilute 0.5 mL of serum to total volume of 7mL in PBS)
Concentration step Protein removal efficiency (%)	76 ± 5	78 ± 6	Not applied
Concentration step protein concentration (mg/mL)	2.30 ± 0.38	1.90 ± 0.31	Not applied
Diafiltration step protein concentration (mg/mL)	0	0	0.23 ± 0.06
Total protein removal efficiency (%)	>99.9	>99.9	99.6 ± 1.0

Supplementary Figure 3. SDS-PAGE of 10% BSA solution before and after iSUF processing. BSA was extensively removed after iSUF processing.

Supplementary Figure 4. SDS-PAGE of CCM before and after iSUF processing. BSA was extensively removed after iSUF processing.

Supplementary Figure 5. SDS-PAGE of serum samples before and after iSUF processing. HSA was extensively removed after iSUF processing.

Supplementary Figure 6. SDS-PAGE of urine samples before and after iSUF processing. HSA was extensively removed after iSUF processing.

Supplementary Table 3. EV concentration, purity and miRNA content comparison among different EV purification platforms.

	Concentration (particles/mL)	Free protein (μg/mL)	Purity (particles/μg)	miRNA (ng)
qEV	8.19E10±2.62E10	15.39±4.06	5.94E09±3.08E09	8.29±7.58
UC	7.46E10±2.25E10	11841.0±2375.6	6.77E06±3.30E06	5.52±5.02
TEI	5.70E11±2.79E11	67.2±29.8	1.32E10±1.43E10	15.23±8.24
iSUF	4.15E12±4.66E11	6.32±4.28	1.32E12±1.26E12	53.72±41.02

Supplementary Figure 7. Comparison of EV concentration, protein, and RNA content in iSUF-urine and iSUF-serum samples from seven healthy donors. The EV concentration and RNA content showed comparable values between urine and serum (n = 10; p > 0.05), while protein content in urine was 10 times lower than in serum.

Supplementary Table 4.

Breast cancer patient and healthy donor information

	P1	P2	P3	P4	P5
Concentration (EVs/ mL)	1.06E11	9.67E11	1.77E10	4.81E10	2.54E11
Mean size (nm)	124	130	127	139	120
Age	66	55	81	60	78
Sex	F	F	F	F	F
Histology	Ductal carcinoma in site	Infiltrating dust carcinoma, NOS	Lobular carcinoma, NOS	Lobular carcinoma, NOS	Infiltrating dust carcinoma, NOS
	P6	P7	P8	P9	P10
Concentration (EVs/mL)	5.78E10	1.27E11	3.19E10	9.13E10	3.02E9
Mean size (nm)	135	135	140	120	128
Age	63	68	32	62	50
Sex	F	F	F	F	F
Histology	Carcinoma	Carcinoma	Infiltrating dust carcinoma, NOS	Adenocarcinoma	Adenocarcinoma

Healthy donor information

	H1	H2	H3	H4
Concentration (EVs/mL)	3.64E12	1.57E12	7.93E11	4.69E11
Mean size (nm)	128	134	145	119
Age	25	24	21	24
Sex	F	F	F	F
	H5	H6	H7	H8
Concentration (EVs/mL)	1.27E12	2.49E12	3.85E11	4.67E12
Mean size (nm)	110	128	134	119
Age	27	28	24	25
Sex	F	F	F	F