Upscaling of *In Vivo* HR-pQCT Images Enables Accurate Simulations of Human Microstructural Bone Adaptation

Authors: Nicholas Ohs (1), Duncan C. Tourolle né Betts (1), Penny R. Atkins (1), Stephanie Sebastian (1), Bert van Rietbergen (2), Michael Blauth (3), Patrik Christen (1,4), Ralph Müller (1)

Affiliations:

1. Institute for Biomechanics, ETH Zurich, Zurich, Switzerland
2. Orthopaedic Biomechanics, Department of Biomedical Engineering, Eindhoven University of Technology, Eindhoven, The Netherlands
3. Department for Trauma Surgery, Innsbruck University Hospital, Innsbruck, Austria
4. Institute for Information Systems, FHNW, Brugg, Switzerland

Abstract

In silico trials of treatments in a virtual physiological human (VPH) would revolutionize research in the biomedical field. Hallmarks of bone disease and treatments can already be simulated in pre-clinical models and in *ex vivo* data of humans using microstructural bone adaptation simulations. The increasing availability of *in vivo* high resolution peripheral quantitative computed tomography (HR-pQCT) images provides novel opportunities to validate and ultimately utilize microstructural bone adaptation simulations to improve our understanding of bone diseases and move towards *in silico* VPH decision support systems for clinicians.

In the present study, we investigated if microstructural bone adaptation simulations of *in vivo* human HR-pQCT images yielded accurate results. Since high-resolution ground truth images cannot be obtained *in vivo,* we applied an *ex vivo* approach to study resolution dependence and the effect of upscaling on morphometric accuracy. To address simulation initialisation issues, we developed an input regularisation approach to reduce initialisation shocks observed in microstructural bone adaptation simulations and evaluated upscaling as a way to improve the accuracy of model inputs. Finally, we compared our *ex vivo* results to simulations run on *in vivo* images to investigate whether *in vivo* image artefacts further affect simulation outcomes.

Keywords

Upscaling, HR-pQCT, Simulation, Mechanoregulation, Microstructure, Bone Adaptation

Introduction

Simulations are considered the third pillar of modern science next to models and experiments. In the biomedical field the creation of a virtual physiological human (VPH) is seen as one of the most important goals (Fenner et al. 2008). The vision of the VPH is to provide researchers with a model that allows rapid hypothesis testing via *in silico* trials and provides doctors with a virtual patient as a decision-support system for their daily work (Viceconti and Hunter 2016). The role of bone, both structurally and physiologically, indicates that a validated model for microstructural bone adaptation and (re)modelling is a significant component to any VPH model. Previous studies have shown that various aspects of bone diseases and their treatments can be simulated in pre-clinical models (Müller 2005; Ruimerman et al. 2005; Schulte et al. 2013; Levchuk et al. 2014). Importantly, the advection based model by Adachi et al. (2001) can produce results comparable to population data when *ex vivo* images are used as an input (Badilatti et al. 2016). However, the translation of microstructural simulations to clinical image data has largely been constrained by the availability of high quality images and validation data.

With the introduction and increased use of HR-pQCT, large amounts of clinically relevant data have been gathered which provide the basis to validate and parameterise *in silico* models of bone (Boutroy et al. 2005, 2008; Sornay-Rendu et al. 2007, 2017; Kirmani et al. 2009; Nicks et al. 2012; Nishiyama et al. 2013; Nishiyama and Shane 2013; Yu et al. 2014; Zhu et al. 2014). However, combining current microstructural bone adaptation simulations with HR-pQCT is non-trivial. Existing simulations either utilize synthetic images (Ruimerman et al. 2005) or high-resolution micro-CT images which cannot be obtained clinically (Müller 2005; Schulte et al. 2013; Levchuk et al. 2014; Badilatti et al. 2016). Furthermore, HR-pQCT images tend to have more noise (Rajapakse et al. 2009) and other potential imaging artefacts, such as those due to movement (Pialat et al. 2012).

The reduction in resolution is a known obstacle for the translation of computational techniques from the lab into the clinical setting (Tjong et al. 2012; Manske et al. 2015; Christen et al. 2016).

Alsayednoor et al. (2018), for example, showed that applying a single threshold to HR-pQCT images

cannot yield both correct morphometric indices and mechanical properties on HR-pQCT images.

However, *in silico* microstructural bone adaptations rely on having correct digital representations of

morphometrics *and* mechanics as the simulation couples these two properties. To study the effects

of image resolution in detail, convergence studies have been performed which showed a clear

dependence on image resolution for various mechanical and morphometric parameters (Müller et al.

1996; Christen et al. 2016). In order to counteract this algorithmic dependence on image resolution,

upscaling low-resolution images to desktop micro-CT images on which the algorithms have been

validated can help to produce accurate results (Rajapakse et al. 2009; Schulte et al. 2019). For

example, upscaling of magnetic resonance imaging (MRI) data has been shown to produce micro-FE

results in good agreement with those from micro-CT images of a higher resolution (Rajapakse et al.

2009). While it is clear that upscaling does not yield the same effects as scanning at a higher

resolution (Cooper et al. 2007), as upscaling cannot compensate for information missing in the

image, techniques, like mesh refinement, are widely used in numerical applications to improve

simulation accuracy by providing a better digital representation of the information contained in the

images.

In previous studies using the algorithm by Adachi et al. (2001), initial iterations, which showed

aberrant results, were regarded as part of the model initialization and excluded from analysis

(Schulte et al. 2013; Badilatti et al. 2016). However, the exclusion of the initial iterations would lead

to an inherent divergence between the clinical *in vivo* and *in silico* baseline models, which thus

precludes direct comparison of simulation outcomes against clinically observed changes in bone

microstructure, and hinders validation of the *in silico* model. The aberrant results of the initial

iterations are caused by an initialisation shock, which is common when modelling coupled systems,

like that of advection and finite-element methods by Adachi et al. (2001), and are related to

mismatches between experimental input data and simulation parameters (Balmaseda and Anderson

2009; Mulholland et al. 2015). In the context of microstructural bone adaptation simulations, the

results of these mismatches can be observed in the large sudden changes in parameters, such as the

total bone volume or the overall structural stiffness. Reducing this shock behaviour allows for the

inclusion of all simulation iterations, such that identical baseline models can be used for the clinical

and *in silico* models and results can be directly compared.

The goal of the present study was to determine if microstructural simulations of *in vivo* human HR-

pQCT images based on the approach by Adachi et al. (2001) yield accurate results, such that they

could be used as part of a VPH. Since high-resolution ground truth micro-CT images cannot be

obtained *in vivo,* we applied an *ex vivo* approach to study the resolution dependence and the effects

of upscaling on morphometric accuracy over the course of a microstructural bone adaptation

simulation. However, we first had to address two initialisation issues: initialisation shocks and

disagreement of mechanics and morphometrics in model inputs at HR-pQCT resolution. Thus, we

developed an input regularisation approach to reduce initialisation shocks observed in

microstructural bone adaptation simulations and studied upscaling as a method to improve the

accuracy of model inputs with respect to both mechanics and morphometry. Finally, we compared

our *in silico* results to simulations run on *in vivo* images to investigate whether additional *in vivo*

image artefacts affect simulation outcomes.

Materials

High Resolution *Ex Vivo* Micro-CT Images

Five distal radii were obtained from female cadavers at the Amsterdam Medical Center as part of a

previous study (De Jong et al. 2016). The donors' ages varied between 58 and 95 years and the bone

volume fraction (BV/TV) of the samples varied from 7 to 20% where BV/TV was inversely related with

age. The medical history of the cadaveric specimens was unknown. High resolution CT images were

obtained at an isotropic voxel-size of 25 μm with a vivaCT 80 (70 kV, 114 μA, 300 ms integration

time), a micro-CT device by Scanco Medical AG (Switzerland). Images were Gauss-filtered (sigma =

1.2, support = 1) and the trabecular region was hand-contoured by a trained operator for each scan

using the software of the scanner manufacturer.

In Vivo HR-pQCT Images

Five patients (four female, one male) were recruited at Innsbruck Medical University as part of a

radius fracture study. Patients provided informed consent and participated in a study approved by

the ethics committee of the Medical University of Innsbruck. The age of the patients ranged from 26

to 80 years and their BV/TV from 21 to 7%. In this study, images of the unfractured, contralateral

radius were used. Scans were performed with an isotropic voxel-size of 61 µm using an XtremeCT II

(68 kV, 1470 µA, 43 ms integration time), a clinical HR-pQCT device by Scanco Medical AG

(Switzerland). Images were Gauss-filtered (sigma = 1.2, support = 1) and the trabecular region was

hand-contoured by a trained operator using the software of the scanner manufacturer.

Generation of Low-Resolution Images

The high-resolution *ex vivo* micro-CT grey-scale images were downscaled to resolutions of 40, 61, and

82 µm. These three resolutions will be referred to as low-resolutions in this paper. Currently, 61 and

82 µm are the highest resolutions available for clinical CT scanners (XtremeCT I and II, Scanco

Medical AG, Switzerland), these resolutions will be referred to as the clinically relevant resolutions.

Resizing was performed using the scikit-image (Van Der Walt et al. 2014) rescale function in Python

(Python Software Foundation 2020) with third-order interpolation and anti-aliasing enabled. The

binary hand-contoured masks for the high-resolution images were converted to a floating-point data-

type and resized like the micro-CT images. Finally, a threshold of 50% of the maximum image value

was applied to obtain binary masks for the low-resolution images.

Generation of Upscaled Images

Upscaled images were created from the low-resolution images by applying the scikit-image resize

Python function, again with third-order interpolation and anti-aliasing enabled, to a resolution of 25

µm. When creating the upscaled images, the conversion of image dimensions between the different

resolutions was not unique (i.e. due to rounding of the integer image dimensions after scaling with a

floating point number, differences in size of one voxel could occur). To ensure that the upscaled

images have the exact same dimensions as the original images, the scikit image resize function was

used. The resize function is identical to the rescale function with the exception that it resizes images

to a target image dimension instead of resizing using a target scaling factor. For the upscaled images,

the same hand-contoured masks for the trabecular region were used as for the high-resolution *ex*

vivo micro-CT grey-scale images to avoid influences of differences in masks on our results. Herein,

the images upscaled from 40, 61, and 82 μm to 25 μm are referred to as u40 μm, u61 μm, and u82

μm, respectively.

The 61 μm clinical *in vivo* HR-pQCT images were re-sampled using the scikit-image rescale function,

as was used for the *ex vivo* micro-CT images, to generate datasets at resolutions of 25, 40, and 82

μm. The masks of the HR-pQCT images were similarly rescaled.

Methods

Remodelling Simulations

Micro-FE Analysis

For the micro-FE analysis, we used the parallel octree solver parOsol (Flaig 2012) on the

supercomputer Piz Daint at the Swiss National Supercomputing Centre (CSCS, Lugano, Switzerland).

Output parameters of strain energy density (SED) and the apparent compressive stiffness along the

longitudinal axis were evaluated. Boundary conditions were determined using a load estimation

algorithm developed by Christen et al. (2013). This algorithm tries to linearly combine three different

load cases to achieve the most homogeneous SED distribution possible across the given bone

structure. The target mean SED value was 0.02 MPa, as has been used previously (Christen et al.

2016). Furthermore, soft pads were added to the distal and proximal ends of the images with a pad-

thickness of 246 μm and a Young's modulus of 15 MPa, which has previously been found to improve

the load estimation (Christen et al. 2013). For all experiments, we computed the load-estimation

using the high-resolution files and applied the same loading conditions to the low-resolution and

upscaled files. This method of load-estimation removes the voxel-size dependency of the algorithm

as a confounding factor.

The micro-FE simulations for images with resolutions higher than 50 µm were run on a 50 µm

hexahedral mesh since the mechanical signalling implemented in the microstructural bone

adaptation simulation is roughly equivalent to a blurring with a sigma of 100 µm. Therefore, the

additional resolution in the SED would not yield differing results. The use of the 50 µm hexahedral

mesh also reduced the computational resources required to run simulations (i.e. for the 25 µm

images, this reduction was an order of magnitude).

Remodelling Algorithm

The strain-adaptive *in silico* microstructural bone adaptation simulation by Adachi et al. (2001) was

re-implemented in Python using NumPy (Van Der Walt et al. 2011) and pybind11 (Jakob et al. 2017).

In short, this algorithm is iterative; for each step, SED (a result of the micro-FE simulation) is

translated, via a mechanostat, into the velocity field of an adapted advection equation. Within this

advection equation, the mass transfer is constrained within a proximity of the bone surfaces, which

results in changes to the bone microstructure. Due to the implementation, all changes are limited to

the trabecular region of the simulated structure.

Binary Model Generation from Micro-CT Data

181   The high-resolution images were segmented using a threshold of 450 mg HA/cm^3, which has also

182   been chosen in previous studies on this dataset (Müller et al. 1996; Christen et al. 2016), and the

183   bone volume over total volume (BV/TV) for the trabecular regions was computed for reference.

184   Finally, voxels identified as bone were set to 750 mg HA / cm^3 and background voxels were set to 0

185   mg HA / cm^3.

Regularized Model Generation

To ensure that the remodelling simulation operates only on the bone surface, the input to the algorithm is required to be binary except for surface voxels that can be represented with intermediate values. To compare the effects of using a conventional binary input or using an input allowing partially filled voxels at the surface layer, we implemented a regularization method that preserves information of the grey-scale image at the surface of bone structures (Figure 1, left). First, a regularization threshold was applied to each high-resolution grey-scale image. Then, surface voxels (empty voxels in direct face-to-face contact with full voxels) of the intermediate binary structure were identified using a Von Neumann neighbourhood. For each surface voxel, grey-scale values from the original grey scale image were converted to a value in the range of zero to one relative to the regularization threshold. The regularization threshold was chosen such that the grey-scale BV/TV of the resulting structure was identical to the one computed for the respective conventional binary structure. Finally, the entire structure was multiplied by the same density value as the conventional binary input (750 mg HA / cm^3).

Parameters for Microstructural Bone Adaptation Simulation

Simulation parameters were chosen such that both formation and resorption were observed in the simulations and were used for all samples and resolutions. For this reason, a narrow lazy zone (0.0196 MPa to 0.0204 MPa) was chosen. The maximum velocity of the mechanostat was set to an arbitrary value of 12 µm/month. The slopes of the mechanostat were set to 8000 µm/year/MPa. The chosen value for slope resulted in generally high velocities and greater changes per time unit, due to the very narrow linear regime of the mechanostat. The choice of simulation parameters allowed for large differences between the different resolutions.

To ensure that the choice of time step between consecutive micro-FE calls did not alter the results, a time step of approximately 1.9 months was chosen. The simulated time period was set to 5 years, resulting in a shorter final iteration step.

Study Design

The validity of using *in vivo* HR-pQCT data as an input for advection based microstructural bone adaptation simulations was investigated using four virtual experiments. Experiment A addresses the issue of initialisation shocks and compares the current approach of generating simulation input models with a novel regularization method. The regularized approach retains grey-scale information, allowing the simulation to initialise with a structure closer to the original one. The goal was to compare the behaviour of the two approaches during the initial iteration steps to identify the approach that exhibits the least amount of initialization shock. Experiment B compared the mechanical and morphological properties of regularized input models generated from high-resolution images, which had been downscaled or down- and then upscaled, that were bone volume fraction matched to the original high-resolution image. The aim of this experiment was to quantify differences between regularized input models with respect to mechanics and morphometrics that may confound simulations. In experiment C, microstructural bone adaptation simulations were run on models of all three *ex vivo* image sets (high-resolution, downscaled, and down- then upscaled) to assess if observed differences in the simulations were due to a lack of fidelity in the input data or the numerical grid. Finally, in experiment D, the *in vivo* images were rescaled to the same resolutions used in experiment C and the convergence was quantified with respect to resolution. The results were compared to those from experiment C to assess what effect differences in image quality and factors other than resolution have on the outcome of the simulation.

Experiment A: Effect of Regularized Input Models on Initialisation Shock

Two simulation input model generation approaches were compared (Figure 1, right), the current state of the art through binary representation and a regularized model with partially filled voxels at the surface. Both models were generated from the high resolution *ex vivo* micro-CT image data set. The partially filled voxels approximate a bone surface with sub-voxel precision. Microstructural bone adaptation simulations were run and the discontinuity in BV/TV and compressive stiffness for the initial simulation steps were quantified for both methods.

Experiment B: Effect of Upscaling on Mechanical and Morphometric Accuracy

For experiments B, regularized input models were produced for all three micro-CT image sets: high-resolution, low-resolution, and upscaled low-resolution images. In the following, these are called reference, low-resolution, and upscaled regularized input models. For all images, the reference BV/TV was always the one obtained from the respective binary high-resolution image.

To quantify the agreement in mechanical properties between reference, low-resolution, and upscaled regularized input models of the same bone structure, micro-FE analyses were performed on the data set at each of these resolutions, respectively (Figure 2). SED distributions, mean SED, mean static parameters (BV/TV, trabecular number (Tb.N), trabecular thickness (Tb.Th), trabecular spacing (Tb.Sp), and structural model index (SMI)), and standard deviations for all mean values were computed for comparison between the three regularized input model types. Additionally, the Kolmogorov-Smirnov statistic was computed between the SED of each sample of the low-resolution and upscaled regularized input models and the corresponding reference regularized input model to quantify mechanical agreement. Finally, the adequacy of the chosen threshold was assessed through comparison of the SED distribution from FE analysis of the low-resolution regularized input models for a range of thresholds (575 - 775 mg HA/cm^3). The magnitudes of the peak SED were compared to the reference SED distribution.

Experiment C: Effect of Upscaling on Morphometric Accuracy Throughout a Microstructural Bone Adaptation Simulation

Microstructural bone adaptation simulations were run on the input models generated for experiment B. The simulations run on the reference regularized input models were considered the best approximations of the *in vivo* remodelling process and were used as reference to quantify errors. These simulations are referred to as the reference simulations (Figure 3). Static parameters (BV/TV, Tb.N, Tb.Th, Tb.Sp, and SMI) and formed and resorbed volume over time were computed.

Experiment D: Effect of *In Vivo* Image Artefacts on Convergence of Upscaled HR-pQCT Simulation

Microstructural bone adaptation simulations were run on regularized input models of the *in vivo* HR-pQCT images and their rescaled version (Figure 3) using the same simulation parameters and calculating the same static parameters as in experiment C. The reference BV/TV for the regularized input model generation for all resolutions was based on the respective original HR-pQCT resolution binary structure.

Evaluation and Statistics

All simulations and evaluations were performed within the trabecular mask. SED was evaluated for non-empty voxels. SED distributions were represented using the SciPy Gaussian kernel density approximation (Virtanen et al. 2019). NumPy (Van Der Walt et al. 2011) was used to compute the Kolmogorov-Smirnov (KS) statistic, mean and standard deviations, as well as BV/TV, which was computed through integration of the model density within the trabecular mask. All other static parameters (Tb.N, Tb.Th, Tb.Sp, SMI) were computed using the scanner manufacturer's image processing language (IPL) (Hildebrand et al. 1999). Before calling the IPL functions, models were up-scaled to 25 µm to remove the voxel-size dependency of IPL functions as a confounding factor. Finally, formed and resorbed volume over time was computed by integration of positive and negative density changes using NumPy.

Comparisons for experiment A were done using a paired Student t-test. For experiments B and C, to determine significance, two-way analysis of variance (two-way ANOVA) was performed for each measured parameter as an omnibus test with the two categorical groups: resolution and upscaling. If heteroscedasticity was detected using a Levene test, heteroscedasticity consistent covariance matrices of type HC3 were used. Post-hoc group comparisons were done using paired Student t-tests and p-values were corrected using the Holm-Bonferroni correction for multiple comparisons. For experiment D, paired Student t-tests were performed and p-values were corrected using the Holm-Bonferroni correction for multiple comparisons. The level of significance was set to 0.05. For the

287   Student t-tests and Levene tests, scipy 1.3.1 was used. The ANOVA was done using statsmodels

288   (Seabold and Perktold 2010) 0.10.2.

Results

Experiment A: Effect of Regularized Input Models on Initialisation Shock

The usage of regularized input models removed initialization shocks found in simulations run on

conventional input models. For the conventional binary input models, the change in apparent

compressive stiffness after the first iteration was a factor of 5.9±0.8 larger than the maximum of all

other iteration steps (Figure 4). For BV/TV, a factor of 2.3±1.6 increase was observed in the first

iteration (Figure 4). Visually, we observe that for some samples a clear shock in BV/TV was present.

In contrast, for the first iteration of the regularized input model approach, the change in apparent

compressive stiffness is indistinguishable from the rest of the simulation with a computed factor of

0.5±0.5 increase, which is significantly lower than for the conventional method (p<0.001). For BV/TV,

we obtained a factor of 2.1±0.6 increase (Figure 4) which is not significantly different from the

threshold method. Since the regularized input model approach removed the initialization shock in

apparent stiffness across all samples, we performed all other experiments using this approach.

The average regularization threshold for the high-resolution images was 563.6±6.6 mg HA / cm^3. For

the low-resolution images, the average regularization thresholds were 601.8±10.0, 605.2±14.7, and

590.6±18.9 mg HA / cm^3 for 40, 61, and 82 μm, respectively.

Experiment B: Effect of Upscaling on Mechanical and Morphometric Accuracy

For all tested thresholds, the SED distributions for the low-resolution regularized input models did

not visually match the SED distribution of the reference regularized input models (Figure 5a).

However, for BV/TV matched thresholds, the peaks of the SED distributions aligned well.

Qualitatively, the difference in SED distribution from the reference model increased with voxel-size

(Figure 5b and 5c). The SED distributions for upscaled regularized input models were almost an order

of magnitude closer to the reference than for the low-resolution regularized input models (p<0.05 for

all; Table 1). Specifically, deviations in mean SED for upscaled resolutions were less than 5% while

deviations were up to 42% for low-resolutions and the KS statistic was below 0.03 for all upscaled

resolutions compared to up to 0.22 for low-resolutions (Table 1).

Resolution, applied upscaling, and the interaction between resolution and upscaling had a significant

effect on all measured static morphometric parameters, except for the interaction of the effects for

SMI and the effect of upscaling on Tb.Th (Table 1). Without upscaling, significant differences in the

mechanical and static parameters (except SMI) were observed for all lower resolutions (40, 61, and

82 μm) (Table 1). No differences in BV/TV were observed, as BV/TV was matched. With upscaling,

deviations in the static parameters were significantly lower for each of the lower resolutions (u40,

u61, and u82 μm, respectively) (Table 1).

The average regularization threshold for the upscaled images were 561.7±6.8, 530.9±9.6, and

496.6±13.7 mg HA / cm^3 for the three upscaled resolutions (u40, u61, and u82 μm), respectively.

Experiment C: Effect of Upscaling on Morphometric Accuracy Throughout a Microstructural Bone Adaptation Simulation

For the reference simulations, BV/TV was initially reduced in the range of 1.8% to 14.8%, followed by

an increase in BV/TV in the range of 3.8% to 38.8% (Figure S1). Tb.N decreased in the range of 5.2%

to 22.8%. Trabecular thickness increased in the range of 29.1% to 42.4%, except for one sample for

which it decreased by 8.8%. SMI was dependent on the specific sample, with differences in the range

of -15.0% to 18.5%; two samples experienced an increase and three samples a decrease in SMI.

Resolution, upscaling, and their interaction had a significant effect ($p<0.001$) on all measured static

and dynamic parameters for the maximum deviations observed during the simulation.

Low-Resolution Simulations

Differences in bone-structure between the reference simulation and the low-resolution simulation

were visible and increased over the course of the simulation (Figure 6). Deviations in static

parameters significantly increased over the course of the simulation compared to the initial models

(**Error! Reference source not found.**, Figure 7). Compared to the reference simulations, BV/TV and

Tb.N were underestimated for all low-resolution simulations (p<0.01), SMI and Tb.Sp were

overestimated (p<0.05), and Tb.Th did not follow a clear trend (Figure 77).

Bone formation and resorption rates were significantly larger for low-resolution simulations

compared to the reference simulations (Table 1). Visually, the formation rate for the low-resolution

simulations peaked at a later time point and had lower peak values compared to the reference

simulation (Figure 8). For the resorption rate, peak delay and widening was also observed for the

low-resolution simulations (Figure 8). However, the magnitude of the resorption rate peaks increased

with voxel-size and the initial resorption rate decayed slower compared to formation rates, for which

the opposite was observed.

Upscaled Simulations

Visually, differences in bone structure over the course of the simulation compared to the reference

simulations were drastically reduced for the upscaled simulations (Figure 6). Accuracy in BV/TV was

improved by an order of magnitude for the highest available clinical resolution (s61 µm) compared to

the low-resolution simulations. The accuracy of all other static parameters was also significantly

improved (Table 1, Figure 77). The maximum deviations in static parameters were not significantly

greater than those of the input models for all upscaled resolutions.

For the dynamic parameters, accuracy of bone formation and resorption per time unit was

significantly increased for the upscaled images (Table 1). The formation and resorption rates peaked

at the same time-point across all resolutions, within the temporal resolution of the simulation. The

peaks of both rates were of similar magnitude across all resolutions. Visually, the overall shapes of

the formation and resorption curves were similar across all resolutions for the duration of the

simulation period, with no noticeable widening or shift of peaks (Figure 88).

Experiment D: Effect of *In Vivo* Image Artefacts on Convergence of Upscaled

HR-pQCT Simulations

After an initial drop in BV/TV in the range of 7.2% to 24.2%, the simulations on the *in vivo* HR-pQCT

data upscaled to 25μm showed varying behaviour. Three samples showed an increase in BV/TV, one

sample had a close to stable BV/TV over time, and one sample experienced a further reduction in

BV/TV before BV/TV began increasing after half of the simulation time. Tb.N. decreased for all

samples over time in the range of 12.0% to 42.9%, whereas Tb.Th. increased over time in the range

of 13.8% to 65.1%. Tb.Sp also increased in the range of 13.6% to 78.0%. SMI decreased for all

samples in the range of 7.8% to 21.8%. The spread of SMI values across all samples decreased over

the course of the simulation by 47.8%.

Comparing the convergence of the different static parameters with respect to resolution between

the low-resolution simulations from experiment C and the simulations from experiment D, no

significant differences could be found except for Tb.Th at 61 and 82 μm resolutions (p<0.001) (Figure

99). Hence, the effects of noise and other additional imaging artefacts from *in vivo* HR-pQCT were

smaller than the effects of model resolution, which dominated the convergence errors observed in

the static parameters (Figure 99).

Discussion

The objectives of this study were to investigate whether microstructural simulations of *in vivo* human

HR-pQCT images yielded accurate results and were a viable tool as part of a VPH.

Experiment A: Effect of Regularized Input Models on Initialisation Shock

Since one of the ideas of the VPH is to provide doctors with a decision support system (Viceconti and

Hunter 2016), the goal of every microstructural bone adaptation simulation must be to achieve parity

between the simulated structure and the structure observed *in vivo.* We observed that when using

conventional model inputs, the apparent compressive stiffness showed an initialization shock

behaviour (Figure 4), with a change approximately six times larger than any other change in stiffness

over the course of the simulation. While initialisation shocks have not been studied in the context of

microstructural bone adaptation simulations, in the context of ocean climate models, Mulholland et

al. (Mulholland et al. 2015) identified that the removal of certain model components can result in

abrupt changes in the dynamics of the system. The analogy for microstructural bone adaptation

simulations are the mismatch of applied boundary conditions and the true, but unknown, *in vivo*

boundary conditions. This mismatch can also be interpreted as the removal of certain boundary

condition forces at the beginning of the *in silico* adaptation. We tackled this challenge by employing

the load estimation algorithm by Christen et al. (2013) which tries to estimate the *in vivo* applied

loads more closely than the uniaxial compression boundary conditions typically used with HR-pQCT

radius data (Burghardt et al. 2011).

Another potential cause for initialisation shocks could be the abrupt change of surface geometry

from *in vivo* microstructural bone adaptation to *in silico* bone adaptation simulations. Using a

regularized input, we used information in the grey-scale image that is normally cut off, improving the

input model generation to reduce the initialisation shock and the associated effect on the results. For

all simulations, the developed regularized input model approach removed the shock behaviour in

apparent compressive stiffness (Figure 4). While the magnitude of the change in BV/TV for the initial

iteration step did not change with the new regularized input model, change in BV/TV looks smoother

using this approach (Figure 4). Importantly, larger changes in BV/TV are expected for this type of

simulation, as the structure adapts to the applied boundary conditions over the course of the

simulation, yielding less changes as the structure reaches a shaped optimized to the applied

boundary conditions. We conclude that the first iteration step does not have to be excluded if this

new approach is used, which allows direct comparisons of morphometrics and mechanics to *in vivo*

measurements. Furthermore, the fact that the initialisation shock was removed might indicate that

the regularized input model is a more accurate mechanical representation of the *in vivo* bone

structure than the conventional binary version.

Experiment B: Effect of Upscaling on Mechanical and Morphometric Accuracy

Another obstacle to overcome when running microstructural bone adaptation simulations on *in vivo* HR-pQCT images was finding an accurate digital representation of a bone captured *in vivo* with HR-pQCT with respect to mechanics and morphometry (Alsayednoor et al. 2018), which is an obvious requirement of bone adaptation simulations. We found that with the use of upscaling, the choice of a single threshold provided regularized input models that agreed well for BV/TV, mechanical properties, and other tested morphometric parameters (i.e. Tb.N, Tb.Sp, and SMI). Furthermore, the linear trend of the thresholds for the different upscaled resolutions indicated that even in the absence of a high resolution ground truth, an appropriate threshold for accurate morphometrics and mechanics can be chosen. In contrast, our results for images that were not upscaled agreed with previous research (Alsayednoor et al. 2018), which showed no agreement between mechanics and morphometrics for various thresholds (Figure 5). This lack of agreement also holds true for thresholds optimized to match BV/TV (Figure 5), a method used in a previous study by Christen et al. to investigate the voxel size dependence of a micro-FE based load estimation algorithm (Christen et al. 2016). The culmination of these results indicates that upscaling may also be useful for other applications using images of HR-pQCT resolution.

Experiment C: Effect of Upscaling on Morphometric Accuracy Throughout a Microstructural Bone Adaptation Simulation

To study the accuracy of microstructural bone adaptation simulations for images with HR-pQCT resolutions, we used high-resolution micro-CT images as ground truth, as this method has been previously validated to simulate realistic bone structures over time (Badilatti et al. 2016). The use of different low-resolution voxel-sizes (40, 61, and 82 μm) resulted in deviations in morphometric parameters of more than 30% in comparison to the reference simulations, however deviations as small as 15% may indicate disease, such as osteoporosis (Zhang et al. 2010). Importantly, observed differences in parameters were not consistent, thus could not be corrected or recovered via calibration curves, as has been possible in previous studies (Müller et al. 1996). In comparison to

static morphologic parameters, dynamic parameters were more dependent on image resolution. This

can be understood by realizing that static parameters are averaged over the bone structure, such

that local differences in morphometrics, e.g. trabecular number, may be concealed. On the other

hand, dynamic parameters are additive and sum all changes over the given volume. Thus, even minor

differences in structure result in larger deviations in dynamic parameters. Importantly, standard

deviations of dynamic parameters from a previous study (Schulte et al. 2013) were lower by up to an

order of magnitude compared to the deviations we observed for the highest clinically available

resolution (80% vs 38% for formation, 187% vs 18% for resorption) (Table 1). Therefore, the accuracy

we observed would be insufficient given the magnitude of natural variation previously observed for

these dynamic parameters. Overall, our results show that microstructural bone adaptation

simulations run on native clinical scanner resolutions suffer from poor accuracy in the assessment of

static morphometric and dynamic parameters; thus, limiting future use as a model for human bone

adaptation.

Running the same simulations on the upscaled images (u40, u61, and u82 μm) resulted in a drastic

reduction in static parameter deviations to less than 10%. For BV/TV, these deviations were near 1%

(Table 1), which is similar to the reproducibility limit of BV/TV for the clinical setting (0.84-1.14%)

(MacNeil and Boyd 2008; Mueller et al. 2009). The improved accuracy in assessment of static

parameters from the upscaled images is likely due to the need for higher resolution to correctly

represent the reference model in terms of mechanics and morphology (Figure 6, Table 1). Any

deviations of the regularized input model led to error accumulation throughout the simulation, such

as missing thin individual trabecular structures that grow in thickness in the reference simulations

(Figure 6). These errors ultimately led to larger morphometric deviations of the final structure (Figure

77). Only the dynamic parameters obtained from the upscaled simulations correctly captured the

overall curve profile of the high-resolution simulations, matching the position, width, and height of

peak formation and resorption rates. The large deviations of the downscaled resolutions indicate

that there may have also been an intrinsic voxel-size dependence of the algorithm, independent of

the initial model. The observed maximum deviations in dynamic parameters are also at least a factor

of two smaller than the natural variation observed in inbred mice indicating sufficient accuracy of the

method for human *in vivo* applications (Schulte et al. 2013).

Experiment D: Effect of *In Vivo* Image Artefacts on Convergence of Upscaled HR-pQCT Simulations

Finally, we investigated the effects of clinically observed image artefacts, e.g. higher noise levels, by

using *in vivo* HR-pQCT images. Comparing the simulations run on these images to those run on

upscaled versions of the same images, no significant difference was observed in the convergence of

the different static parameters (Figure 99), except for Tb.Th. Since the *ex vivo* and *in vivo* dataset are

not from the same study participants, it is possible that this difference in Tb.Th is due to unknown

physiological differences between the *in vivo* subject group and the *ex vivo* samples. Furthermore,

with respect to the overall deviation observed in Tb.Th, the observed significant difference is still

small, especially given the fact that Tb.Th is known to be difficult to capture with HR-pQCT resolution

(MacNeil and Boyd 2007; Manske et al. 2015) and are therefore likely not clinically relevant.

Limitations

This study is, however, not without limitations. One limitation of this study was the lack of a high

resolution ground truth scan of the patient radii used in Experiment D. However, the artefacts most

commonly associated with *in vivo* HR-pQCT images, such as motion artefact, are difficult to recreate

with cadaveric specimen, while micro-CT images cannot be obtained from patients due to the

radiation dosage and imaging volume. Thus, we utilized both high-resolution cadaveric images and

clinically acquired *in vivo* HR-pQCT images of patients to assess these factors independently.

An additional limitation of this study is the sample size (n=5 for both *ex vivo* and *in vivo* experiments).

However, the small spread in deviations across subjects observed from the results of the upscaled

simulations indicates that a larger sample size may not be warranted. Importantly, the inclusion of

additional samples would have required an excess of computational resources due to the high

resolution of the simulations.

Conclusions

In conclusion, we found model resolution to be the dominating image property which drove

convergence errors in microstructural bone adaptation simulations. Importantly, upscaling drastically

reduced errors in mechanical analyses, static morphological parameters, and dynamic parameters,

resulting in simulation outcomes that, even for clinically available resolutions, were similar to those

from high-resolution images. Initialisation errors were avoided with the use of upscaling and the

proposed regularization method, which generated model input that closely represented the true

bone structure with respect to both mechanics and morphometry. With these results, we conclude

that microstructural bone adaptation simulations can be run on *in vivo* HR-pQCT images and yield

realistic results, given a validated set of parameters. Hence, these simulations provide a powerful

tool to study disease related bone microstructure changes in patients, as part of the VPH vision.

Acknowledgements

500 Funding for the DACH Fx Project from the Swiss National Science Foundation (Lead Agency,

501 320030L_170205), German Research Foundation (IG 18/19-1, SI 2196/2-1), and Austrian Science

502 Fund (I 3258-B27) is gratefully acknowledged. This work was supported by a grant from the Swiss

503 National Supercomputing Centre (CSCS) under project ID s841.

References

Adachi T, Tsubota KI, Tomita Y, Scott JH (2001) Trabecular surface remodeling simulation for

cancellous bone using microstructural voxel finite element models. J Biomech Eng 123:403–409.

https://doi.org/10.1115/1.1392315

Alsayednoor J, Metcalf L, Rochester J, et al (2018) Comparison of HR-pQCT- and microCT-based finite

element models for the estimation of the mechanical properties of the calcaneus trabecular

bone. Biomech Model Mechanobiol 17:1715–1730. https://doi.org/10.1007/s10237-018-1051-

6

Badilatti SD, Christen P, Parkinson I, Müller R (2016) Load-adaptive bone remodeling simulations

reveal osteoporotic microstructural and mechanical changes in whole human vertebrae. J

Biomech 49:3770–3779. https://doi.org/10.1016/j.jbiomech.2016.10.002

Balmaseda M, Anderson D (2009) Impact of initialization strategies and observations on seasonal

forecast skill. Geophys Res Lett 36:1–5. https://doi.org/10.1029/2008GL035561

Boutroy S, Bouxsein ML, Munoz F, Delmas PD (2005) In vivo assessment of trabecular bone

microarchitecture by high-resolution peripheral quantitative computed tomography. J Clin

Endocrinol Metab 90:6508–6515. https://doi.org/10.1210/jc.2005-1258

Boutroy S, Van Rietbergen B, Sornay-Rendu E, et al (2008) Finite element analysis based on in vivo

HR-pQCT images of the distal radius is associated with wrist fracture in postmenopausal

women. J Bone Miner Res 23:392–399. https://doi.org/10.1359/jbmr.071108

Burghardt AJ, Link TM, Majumdar S (2011) High-resolution computed tomography for clinical imaging

of bone microarchitecture. Clin Orthop Relat Res 469:2179–2193.

https://doi.org/10.1007/s11999-010-1766-x

Christen P, Ito K, Knippels I, et al (2013) Subject-specific bone loading estimation in the human distal

radius. J Biomech 46:759–766. https://doi.org/10.1016/j.jbiomech.2012.11.016

Christen P, Schulte FA, Zwahlen A, et al (2016) Voxel size dependency, reproducibility and sensitivity

of an in vivo bone loading estimation algorithm. J R Soc Interface 13:20150991.

https://doi.org/10.1098/rsif.2015.0991

Cooper D, Turinsky A, Sensen C, Hallgrimsson B (2007) Effect of voxel size on 3D micro-CT analysis of

cortical bone porosity. Calcif Tissue Int 80:211–219. https://doi.org/10.1007/s00223-005-0274-

6

De Jong JJA, Arts JJ, Meyer U, et al (2016) Effect of a cast on short-term reproducibility and bone

parameters obtained from HR-pQCT measurements at the distal end of the radius. J Bone Jt

Surg - Am Vol 98:356–362. https://doi.org/10.2106/JBJS.O.00127

Fenner JW, Brook B, Clapworthy G, et al (2008) The EuroPhysiome, STEP and a roadmap for the

virtual physiological human. Philos Trans R Soc A Math Phys Eng Sci 366:2979–2999.

https://doi.org/10.1098/rsta.2008.0089

Flaig C (2012) A highly scalable memory efficient multigrid solver for µ-finite element analyses. ETH

Zurich

Hildebrand T, Laib A, Müller R, et al (1999) Direct Three-Dimensional Morphometric Analysis of

Human Cancellous Bone: Microstructural Data from Spine, Femur, Iliac Crest, and Calcaneus. J

Bone Miner Res 14:1167–1174. https://doi.org/10.1359/jbmr.1999.14.7.1167

Jakob W, Rhinelander J, Moldovan D (2017) pybind11 - Seamless operability between C++11 and

Python

Kirmani S, Christen D, Van Lenthe GH, et al (2009) Bone structure at the distal radius during

adolescent growth. J Bone Miner Res 24:1033–1042. https://doi.org/10.1359/jbmr.081255

Levchuk A, Zwahlen A, Weigt C, et al (2014) The Clinical Biomechanics Award 2012 - Presented by the

European Society of Biomechanics: Large scale simulations of trabecular bone adaptation to

loading and treatment. Clin Biomech 29:355–362.

https://doi.org/10.1016/j.clinbiomech.2013.12.019

MacNeil JA, Boyd SK (2008) Improved reproducibility of high-resolution peripheral quantitative

computed tomography for measurement of bone quality. Med Eng Phys 30:792–799.

https://doi.org/10.1016/j.medengphy.2007.11.003

MacNeil JA, Boyd SK (2007) Accuracy of high-resolution peripheral quantitative computed

tomography for measurement of bone quality. Med Eng Phys 29:1096–1105.

https://doi.org/10.1016/j.medengphy.2006.11.002

Manske SL, Zhu Y, Sandino C, Boyd SK (2015) Human trabecular bone microarchitecture can be

assessed independently of density with second generation HR-pQCT. Bone 79:213–221.

https://doi.org/10.1016/j.bone.2015.06.006

Mueller TL, Stauber M, Kohler T, et al (2009) Non-invasive bone competence analysis by high-

resolution pQCT: An in vitro reproducibility study on structural and mechanical properties at the

human radius. Bone 44:364–371. https://doi.org/10.1016/j.bone.2008.10.045

Mulholland DP, Laloyaux P, Haines K, Balmaseda MA (2015) Origin and impact of initialization shocks

in coupled atmosphere-ocean forecasts. Mon Weather Rev 143:4631–4644.

https://doi.org/10.1175/MWR-D-15-0076.1

Müller R (2005) Long-term prediction of three-dimensional bone architecture in simulations of pre-,

peri- and post-menopausal microstructural bone remodeling. Osteoporos Int 16:25–35.

https://doi.org/10.1007/s00198-004-1701-7

Müller R, Koller B, Hildebrand T, et al (1996) Resolution dependency of microstructural properties of

cancellous bone based on three-dimensional mu-tomography. Technol Health Care 4:113–119

Nicks KM, Amin S, Atkinson EJ, et al (2012) Relationship of age to bone microstructure independent

of areal bone mineral density. J Bone Miner Res 27:637–644.

https://doi.org/10.1002/jbmr.1468

Nishiyama KK, Macdonald HM, Hanley DA, Boyd SK (2013) Women with previous fragility fractures

can be classified based on bone microarchitecture and finite element analysis measured with

HR-pQCT. Osteoporos Int 24:1733–1740. https://doi.org/10.1007/s00198-012-2160-1

Nishiyama KK, Shane E (2013) Clinical imaging of bone microarchitecture with HR-pQCT. Curr.

Osteoporos. Rep. 11:147–155

Pialat JB, Burghardt AJ, Sode M, et al (2012) Visual grading of motion induced image degradation in

high resolution peripheral computed tomography: Impact of image quality on measures of bone

density and micro-architecture. Bone 50:111–118. https://doi.org/10.1016/j.bone.2011.10.003

Python Software Foundation (2020) Python Language Reference, Version 3.7.

https://www.python.org

Rajapakse CS, Magland J, Zhang H, et al (2009) Implications of noise and resolution on mechanical

properties of trabecular bone estimated by image-based finite-element analysis. J Orthop Res

27:1263–1271. https://doi.org/10.1002/jor.20877

Ruimerman R, Hilbers P, Van Rietbergen B, Huiskes R (2005) A theoretical framework for strain-

related trabecular bone maintenance and adaptation. J Biomech 38:931–941.

https://doi.org/10.1016/j.jbiomech.2004.03.037

Schulte FA, Christen P, Badilatti SD, et al (2019) Virtual supersampling as post-processing step

preserves the trabecular bone morphometry in human peripheral quantitative computed

tomography scans. PLoS One 14:1–11. https://doi.org/10.1371/journal.pone.0212280

Schulte FA, Zwahlen A, Lambers FM, et al (2013) Strain-adaptive in silico modeling of bone

adaptation - A computer simulation validated by in vivo micro-computed tomography data.

Bone 52:485–492. https://doi.org/10.1016/j.bone.2012.09.008

Seabold S, Perktold J (2010) statsmodels: Econometric and statistical modeling with python. In: 9th

Python in Science Conference

Sornay-Rendu E, Boutroy S, Duboeuf F, Chapurlat RD (2017) Bone Microarchitecture Assessed by HR-

pQCT as Predictor of Fracture Risk in Postmenopausal Women: The OFELY Study. J Bone Miner

Res 32:1243–1251. https://doi.org/10.1002/jbmr.3105

Sornay-Rendu E, Boutroy S, Munoz F, Delmas PD (2007) Alterations of cortical and trabecular

architecture are associated with fractures in postmenopausal women, partially independent of

decreased BMD measured by DXA: The OFELY study. J Bone Miner Res 22:425–433.

https://doi.org/10.1359/jbmr.061206

Tjong W, Kazakia GJ, Burghardt AJ, Majumdar S (2012) The effect of voxel size on high-resolution

peripheral computed tomography measurements of trabecular and cortical bone

microstructure. Med Phys 39:1893–1903. https://doi.org/10.1118/1.3689813

Van Der Walt S, Colbert SC, Varoquaux G (2011) The NumPy array: A structure for efficient numerical

computation. Comput Sci Eng 13:22–30. https://doi.org/10.1109/MCSE.2011.37

Van Der Walt S, Schönberger JL, Nunez-Iglesias J, et al (2014) Scikit-image: Image processing in

python. PeerJ 2014:1–18. https://doi.org/10.7717/peerj.453

Viceconti M, Hunter P (2016) The Virtual Physiological Human: Ten Years After. Annu Rev Biomed Eng

18:103–123. https://doi.org/10.1146/annurev-bioeng-110915-114742

Virtanen P, Gommers R, Oliphant TE, et al (2019) SciPy 1.0--Fundamental Algorithms for Scientific

Computing in Python. 1–22

Yu WS, Chan KY, Yu FWP, et al (2014) Bone structural and mechanical indices in Adolescent Idiopathic

Scoliosis evaluated by high-resolution peripheral quantitative computed tomography (HR-

pQCT). Bone 61:109–115. https://doi.org/10.1016/j.bone.2013.12.033

Zhang ZM, Li ZC, Jiang LS, et al (2010) Micro-CT and mechanical evaluation of subchondral trabecular

bone structure between postmenopausal women with osteoarthritis and osteoporosis.

Osteoporos Int 21:1383–1390. https://doi.org/10.1007/s00198-009-1071-2

Zhu TY, Griffith JF, Qin L, et al (2014) Alterations of bone density, microstructure, and strength of the

distal radius in male patients with rheumatoid arthritis: A case-control study with HR-pQCT. J

Bone Miner Res 29:2118–2129. https://doi.org/10.1002/jbmr.2221

Tables

629   Table 1 Comparison of microstructural bone adaptation simulation outcomes for regularized model

630   inputs of micro-CT images that were downscaled from low-resolution and then upscaled back to

631   high-resolution. Parameters were compared against the high-resolution micro-CT image reference

632   simulation and relative deviations in percent are shown

			Downscaled from 25 µm			Upscaled to 25 µm		
			40 µm	61 µm	82 µm	u40 µm	u61 µm	u82 µm
Regularized Input Model Deviation	Mechanics	Mean SED	4.1 ± 1.3% *	17.2 ± 2.0% ***, ###	42.2 ± 7.1% **, ††	0.4 ± 0.6% ‡	2.4 ± 1.6% ‡‡	5.1 ± 3.4% ‡‡
		KS	0.022 ± 0.007 *	0.083 ± 0.023 *, #	0.257 ± 0.083 *, †	0.003 ± 0.004 ‡	0.012 ± 0.009 ‡	0.026 ± 0.021 ‡
	Static Parameters	Tb.N	-7.9 ± 2.7% **	-19.8 ± 4.5% **, ##	-30.1 ± 5.5% **, †††	-0.5 ± 0.1% **, ‡‡	-6.6 ± 1.3% **, ‡‡, ##	-13.4 ± 2.7% **, ‡‡, ††
		Tb.Sp	9.8 ± 2.6% **	26.8 ± 6.0% **, ##	46.5 ± 9.8% **, ††	0.5 ± 0.1% **, ‡‡	7.1 ± 1.4% **, ‡‡, ##	15.9 ± 3.4% **, ‡‡, ††
		Tb.Th	-3.0 ± 1.7% **	6.5 ± 3.8% **, ##	16.8 ± 5.8% **, ††	0.5 ± 0.1%	5.8 ± 1.2% ##	12.4 ± 2.7% *, ††
		SMI	5.5 ± 4.5%	13.9 ± 11.4%	20.5 ± 17.0%	0.03 ± 0.2%	1.8 ± 2.1%	6.6 ± 5.0%
	Static Parameters	BV/TV	-6.7 ± 2.4% **, ¥¥	-21.0 ± 5.2% **, ¥¥, ##	-45.2 ± 6.2% ***, ¥¥¥, ††	-0.3 ± 0.4% ‡‡	-1.4 ± 1.7% ‡‡	-3.2 ± 3.7% ‡‡‡
		Tb.N	-11.1 ± 3.8% **, ¥	-33.8 ± 8.5% **, ¥¥, ##	-57.2 ± 7.4% ***, ¥¥¥, ††	-0.6 ± 0.2% **, ‡‡	-6.8 ± 1.8% **, ‡‡, ##	-14.9 ± 4.2% **, ‡‡‡, ††
		Tb.Sp	14.1 ± 4.2% *, ¥	57.9 ± 16.4% *, ¥, #	154.4 ± 33.3% **, ¥¥, ††	0.6 ± 0.3% *, ‡	7.5 ± 2.1% *, ‡, #	18.2 ± 5.8% *, ‡‡, †
		Tb.Th	-6.3 ± 0.8% ***, ¥	10.3 ± 4.0% *, ¥, ##	18.6 ± 6.6% *, †	0.5 ± 0.1% **, ‡‡‡	5.9 ± 1.2% **, ##	12.9 ± 3.0% **, ††
		SMI	12.6 ± 3.8% *, ¥	31.4 ± 8.9% *, ¥, #	48.4 ± 13.5% *, ¥¥, #, †	0.2 ± 0.2% ‡	2.8 ± 0.9% *, ‡, #	8.2 ± 2.3% *, ‡, †
	Dynamic Parameters	Bone Formation	70.9 ± 1.9% ***	80.3 ± 2.1% ***, ###	86.2 ± 2.7% ***, †	1.5 ± 0.7% ***, ‡‡‡	14.7 ± 2.1% ***, ‡‡‡, ###	29.0 ± 3.6% ***, ‡‡‡, †††
		Bone Resorption	73.0 ± 5.4% ***	186.7 ± 23.4% ***, ##	365.6 ± 61.3% **, ††	0.8 ± 0.9% ‡‡‡	4.7 ± 2.3% *, ‡‡‡, #	11.5 ± 7.2% ‡‡

SED: strain energy density; KS: Kolmogorov Smirnov statistic; Tb.N: Trabecular number, Tb.Sp: Trabecular spacing; Tb.Th: Trabecular thickness;

SMI: Structure model index; s40, s61, and s82 µm: 40, 61, and 82 µm upscaled to 25 µm, respectively

Statistical significance is indicated as follows: Difference from zero (the reference simulation) (*), differences between the same resolution with

and without upscaling (‡), difference between 40 and 61 µm (#), difference between 61 and 82 µm (†), and difference between regularized input

model and maximum deviations (¥)

Figures

Figure 1

Fig 1 Experiment A (Effect of Regularized Input Models on Initialisation Shock) evaluated the

reduction of initialisation shock behaviour in bone volume fraction and compressive stiffness due to

a novel input regularization approach compared to the conventional input approach. Left: Illustration

of the conventional threshold approach used to convert a CT image into a valid input for the load

adaptation simulation. Right: Regularization approach

643   Figure 2

Fig 2 Experiment B (Effect of Upscaling on Mechanical and Morphometric Accuracy) evaluated the

mechanical agreement between the different resolutions for matched BV/TV. For each micro-CT

image, high-, low-resolution, and upscaled images were created (left column). Low-resolution: 40

μm, 61 μm, and 82 μm and upscaled images: Same downscaled resolutions as the low-resolution

image but upscaled back to 25 μm. Micro-finite-element (micro-FE) models were generated using the

regularization model generation approach of experiment A (Figure 1) and matching bone volume

fraction (BV/TV) for each image to the reference (25 μm). A micro-FE analysis was run and strain

energy density (SED) (shown in the jet colour-map) and static parameters were computed for each

model. SED distributions between low and upscaled resolution images were compared using the

Kolmogorov-Smirnov statistic to see which one more closely matches the reference high-resolution

regularized input model mechanically. Mean static parameters were compared to see which one

more closely matches the reference model morphologically

Figure 3

Fig 3 Overview of experiments C (Effect of Upscaling on Morphometric Accuracy throughout a

Microstructural Bone Adaptation Simulation) and D (Effect of *In Vivo* Image Artefacts on

Convergence of Upscaled HR-pQCT Simulation). Top: Regularized input models from experiment B

were taken and microstructural bone adaptation simulations were run for all input models. Bottom:

A separate HR-pQCT dataset was also converted to regularized input models and simulations were

run for all input models. For experiment C, static parameters and dynamic parameters were

computed. The goal of experiment C was to compare how accurate low-resolution versus upscaled

resolution simulations were relative to the reference simulations. For experiment D, static

parameters were computed as well. The goal of experiment D was to compare the voxel size

dependency of down-scaled micro-CT and rescaled HR-pQCT images to determine if additional

artefacts introduced by the HR-pQCT images had a strong influence on the outcome of the bone-

adaptation simulation

Figure 4

Fig 4 Initialisation shocks in bone volume fraction (BV/TV) and relative apparent stiffness are reduced

for regularized input models in comparison to conventional input models, shown for a representative

sample of experiment A (Effect of Regularized Input Models on Initialisation Shock). The initialization

shock visible in BV/TV and stiffness (left) for the conventional model generation was not present

when using the regularization approach (right)

Figure 5

Fig 5 Visual agreement in strain energy density (SED) distributions are improved for upscaled bone

volume fraction (BV/TV) matched regularized input models for a representative sample of

experiment B (Effect of Upscaling on Mechanical and Morphometric Accuracy). (a) The strain energy

density (SED) distributions for the reference resolution (25 μm) (blue) was flatted compared to the

down-scaled 61 μm model (green), which used a threshold chosen to match the bone volume

fraction (BV/TV) of the reference resolution model. Thresholds corresponding to BV/TV values

between 7 and 23% are shown in different shades of grey with a worse agreement with the

reference SED distribution. (b)Comparing across all resolutions, the low-resolution models did not

capture the SED distribution of the reference model. (c) However, upscaled resolutions clearly

captured the distribution of the reference model across all resolutions, with deviations being an

order of magnitude smaller compared to models generated without upscaling

Figure 6

Fig 6 Visual difference after 0, 33, 66, and 100% of the simulated time are reduced for upscaled

regularized input models for a representative sample of experiment C (Effect of Upscaling on

Morphometric Accuracy Throughout a Microstructural Bone Adaptation Simulation). Top:

comparison of reference to low-resolution simulation. Bottom: comparison of reference to upscaled

simulation. For the upscaled simulation, very small structures were still lost, due to very thin

trabeculae that cannot be captured in a 61 μm image, but the major part of the bone structure

remodelled identical to the reference simulation for upscaled images

Figure 7

Fig 7 Deviations from the reference simulation in static parameters are reduced for upscaled

regularized input models across all resolutions as found in experiment C (Effect of Upscaling on

Morphometric Accuracy Throughout a Microstructural Bone Adaptation Simulation). Deviations in

static parameters from the reference simulation (25 µm) over the course of the simulation for low-

resolution (left) and upscaled regularized input models (right) are shown. Computed static

parameters are: Bone volume fraction (BV/TV), trabecular number (Tb.N), trabecular spacing (Tb.Sp),

trabecular thickness (Tb.Th), and structure model index (SMI). BV/TV was matched for the initial

model, resulting in perfect agreement between the resolutions. Generally, for the different samples,

upscaling improves the agreement with the reference (25 µm) simulations. Deviations were more

predictable after upscaling of the image

Figure 8

Fig 8 Deviations from the reference simulation in dynamic parameters are reduced for upscaled

regularized input models across all resolutions as found in experiment C (Effect of Upscaling on

Morphometric Accuracy Throughout a Microstructural Bone Adaptation Simulation). Dynamic

parameter results of simulation run on low-resolution (left) and upscaled regularized input models

(right) are shown. Bone formation and resorption over time (top), and deviations of these

parameters from the reference simulation (25 µm) (bottom). Reference simulation results were

aligned with the use of upscaled regularized input models with respect to the amount and time-point

of formation and resorption events. Deviations were an order of magnitude smaller for the

simulations run on upscaled regularized input models

Fig 9 Agreement in convergence between regularized input models from downscaled micro-CT and

rescaled HR-pQCT as found in experiment D (Effect of *In Vivo* Image Artefacts on Convergence of

Upscaled HR-pQCT Simulation). Convergence behaviour of low-resolution simulations from

experiment C (Effect of Upscaling on Morphometric Accuracy Throughout a Microstructural Bone

Adaptation Simulation) and from rescaled HR-pQCT image simulations (top left). Maximum mean

deviations from the reference (25 μm) simulations and corresponding standard deviations are shown

for all computed static parameters. Significant differences were observed for trabecular thickness

739   (Tb.Th); the mean maximum deviations of all other parameters are not significantly different,

740   indicating that voxel-size was the dominating factor on the simulation outcomes

Supplementary

Figure S1

Fig S1 Visual deviations from the reference simulation in static parameters are reduced for upscaled

regularized input models across all resolutions as found in experiment C (Effect of Upscaling on

Morphometric Accuracy Throughout a Microstructural Bone Adaptation Simulation). Static

parameters over the course of the simulation for low-resolution (left) and upscaled regularized input

models (right) are shown. Computed static parameters are: Bone volume fraction (BV/TV), trabecular

number (Tb.N), trabecular spacing (Tb.Sp), trabecular thickness (Tb.Th), and structure model index

(SMI). BV/TV was matched for the initial model, which is why initially we got perfect agreement

between the resolutions. Overall we see for the different samples that upscaling improves the

agreement with the reference (25 μm) simulations