1   **Validation and performance of a quantitative IgG assay for the screening of**

2   **SARS-CoV-2 antibodies**

3
4
5
6

7   Ana M. Espino[1*#], Petraleigh Pantoja[1,2*] and Carlos A. Sariol[1,2,3*#]
8
9

1 Department of Microbiology and Medical Zoology, University of Puerto Rico-Medical Sciences

Campus, San Juan, PR, USA. 2 Unit of Comparative Medicine, Caribbean Primate Research

Center and Animal Resources Center, University of Puerto Rico-Medical Sciences Campus,

San Juan, PR, USA, 3 Department of Internal Medicine, University of Puerto Rico-Medical

Sciences Campus, San Juan 00936 PR, USA.

* These authors contributed equally to the development of the assay

Corresponding authors: carlos.sariol1@upr.edu and ana.espino1@upr.edu

Running Title: UPR-MSC CovIgG-Assay

Abstract

The current COVID-19 epidemic imposed an unpreceded challenge to the scientific community in terms of treatment, epidemiology, diagnosis, social interaction, fiscal policies and many other areas. The development of accurate and reliable diagnostic tools (high specificity and sensitivity) is crucial in the current period, the near future and in the long term. These assays should provide guidance to identify immune presumptive protected persons, potential plasma, and/or B cell donors and vaccine development among others. Also, such assays will be contributory in supporting prospective and retrospective studies to identify the prevalence and incidence of COVID-19 and to characterize the dynamics of the immune response. As of today, only thirteen serological assays have received the Emergency Use Authorization (EUA) by the U.S. Federal Drug Administration (FDA). In this work we describe the development and validation of a quantitative IgG enzyme-linked immunoassay (ELISA) using the recombinant SARS-CoV-2 Spike Protein S1 domain, containing the receptor-binding domain (RBD), showing 98% sensitivity and 98.9% specificity. The assay showed to be useful to test for SARS-CoV-2 IgG antibodies in plasma samples from COVID-19-recovered subjects as potential donors for plasmapheresis. This assay is currently under review by the Federal Drug Administration for an Emergency Use Authorization request (Submission Number EUA201115).

Introduction

The current severe acute respiratory syndrome coronavirus 2 (SARS CoV-2) pandemic and the resulting unprecedented outbreak of coronavirus disease 2019 (COVID-19) have shifted the paradigm for viral research, epidemiology and diagnostic. Both molecular and serological methods have been developed at an extraordinary speed. As of April 2, 2020 only four months after the virus was detected for first time in Wuhan region, 28 companies obtained Emergency Use Authorization (EUA) approvals from US Federal Drug Administration (FDA) for their commercial Reverse Transcription-Polymerase Chain Reaction (RT-PCR) diagnostics. Those assays are intended to detect the virus during the acute phase of the infection, providing no information regarding the immunological status of these patients. By the same time, from the more than 25 rapid serological tests available only one had the EUA granted. These rapid tests are relatively simple to perform and interpret and therefore require limited test operator training. The main drawback of these rapid tests is that the specificity and particularly the sensitivity are lower than the standard Enzyme-linked Immunosorbent Assays (ELISA). As of June 1, 2020 FDA had received more than 198 notifications from manufacturers confirming they have validated and intend to distribute their tests in the market. However only 13 of those tests have indeed the EUA from FDA. Moreover, in May 2020, FDA removed 28 SARS-CoV-2 serological tests from the notification list of tests offered during the COVID-10 emergency for not having an EUA request. Choosing an appropriate test to screen for the presence of humoral immune response to SARS-CoV-2 is critical. Such serologic tests are expected to play a key role in the fight against COVID-19 by helping to identify individuals who had developed an adaptive immune response and may be at lower risk of infection. Also, validated serological tests are needed to confirm which subjects, being confirmed positive for COVID-19, truly developed a substantial humoral immune response and may be considered as plasma donors (1). Different antigens have been used to detect antibodies against other novel coronavirus such as SARS-CoV and MERS-CoV (2-5). From these previous works it can be concluded that spike-derived (S) antigens are more sensitive, specific

and accurate than nucleocapsid protein-derived (NP) antigens. Also results from assays using S

antigens correlated much better with the neutralizing titers than those using NP antigens (6). A

recent work showed the usefulness for the Receptor Binding Domain (RBD) and the full Spike

protein to detect SARS-CoV-2 specific antibodies (7) and their correlation with neutralizing

antibodies nAb (8). For these reasons we choose to use a recombinant SARS-CoV-2 Spike

Protein, S1 domain containing the RBD.

With this work we described the validation of a quantitative ELISA, CovIgG-Assay

(https://prsciencetrust.org/the-covigg-assay-kit/), showing a very low background and lack of

cross reactivity with other respiratory and non-respiratory pathogens in more than 132 samples

collected before June 2019. Also the limited correlation with a rapid test and another serology test

availables in the market is described. Finally, we confirm the usefulness of the assay detecting

anti-SARS-CoV-2 antibodies in plasma samples from potential plasma donors. CovIgG-Assay is

a useful tool to characterize, quantify and to study the dynamics of the humoral immune response

to SARS-CoV-2.

Materials and Methods

Study Design

The study population included 49 positive samples from individuals with symptomatic infection and positive SARS-CoV-2. Forty-eight (48) were confirmed by RT-PCR tests EUA authorized and one (1) diagnosed by COVID-19 ELISA IgG Antibody Test – Mount Sinai, also EUA authorized. De-identified serum or plasma specimens were obtained from local clinical laboratories and Blood Banks and no personal identifiers were retained. For specificity, we tested a total of 132 samples taken previously to 2019 from the Virology UPR MSC or the Immunology UPR-MSC serum bank. From these samples, 78 had no previous history of viral, allergic or bacterial infections according to our cross-reactivity panel. Nine (9) were previously diagnosed with Zika, three (3) with Dengue, thirteen (13) with history of respiratory allergies and one (1) with Influenza H1N1. We also included a cross reactivity panel with 28 samples kindly donated by the Centers for Disease Control and Prevention (CDC) Dengue Branch, San Juan, PR. These samples included six (6) positives for Respiratory Syncytial Virus (RSV)-IgM, twelve (12) RT-PCR positive for Influenza A or B, five (5) Zika-IgM positive and five (5) positive por Dengue-IgM. In addition, to complete our cross-reactivity panel, we tested nine (9) samples IgM positive for Mycoplasma and three (3) positives for Chikungunya. This cross-reactive panel was selected according to the most common viral and respiratory infections affecting our population. All samples were stored at -80^{o}C until use.

For comparison with two other serological test (CoronaCheck and Abbott Architect) holding an EUA, we used a set of nine (9) samples assumed to be positive for IgG and IgM and eighteen (18) assumed to be IgG positive for SARS-CoV-2 antibodies. Those samples were also received de-identified from local laboratories.

CovIgG-Assay

CovIgG Assay is an indirect ELISA for quantitative determination of human IgG antibody class,

which was optimized by checkerboard titration. Disposable high bind flat-bottomed polystyrene

96-wells microtiter plates (Costar, Corning MA No. 3361) were coated overnight at 4^{o}C with

2µg/ml of recombinant SARS-CoV-2 S1-RBD protein (GenScript No. Z03483-1) in carbonate-

bicarbonate buffer (Sigma Aldrich No. 08058). Plates were washed 3 times with phosphate

buffered saline (PBS) containing 0.05% Tween-20 (PBST) and blocked for 30 min at 37^{o}C with

250µl/well of 3% non-fat, skim milk in PBST. Samples (serum or plasma) were diluted 1:100 in

PBST; 100µL/well was added in duplicates and incubated at 37^{o}C for 30 min. The excess antibody

was washed off with PBST. Horseradish peroxidase (HRP) labeled-mouse anti-human IgG-Fc

specific (GenScript No. A01854) diluted 1:10,000 in PBST was added (100µl/well) and incubated

for 30 min at 37^{o}C. After another washing step, the substrate solution (Sigma Aldrich No. P4809)

was added (100µl/well) followed by 15 min incubation in dark. The reaction was stopped by the

addition of 50µl/well 10% HCl and the absorbance was measured at 492nm (A$_{492}$) using a

Multiskan FC reader (Thermo Fisher Scientific). In every CovIgG-Assay determination two in-

house controls, a high positive control (HPC) and negative control (NC) were included. HPC and

NC were prepared by diluting an IgG anti-SARS-CoV-2 at a concentration of 30µg/ml and

0.070µg/ml, respectively in PBST containing 10% glycerol. The IgG anti-SARS-CoV-2 was

purified from plasma of a convalescent patient using a 5/5 HiTrap rProtein-A column (GE

Healthcare, USA) (see detailed information about this procedure in **Supplementary method**

No.1).

Antibody class specificity

To confirm that our assay accurately detects antibody IgG class and excludes the potential for

human IgM to cross-react with IgG, five (5) COVID-19 samples (1:100 diluted) were treated with

5mM DTT for 30 min at 37^{o}C prior testing. After treatment, samples were added in duplicate

(100µl/well) followed by the addition of the anti-human IgG-Fc-HRP (GenScript No. A01854)

conjugate (diluted 1:10,000) or the addition of an anti-human IgM-HRP conjugate (Abcam No.

ab97205) diluted 1:8,000 in PBST and the assay progressed as described above.

Estimation of Antibody Titer

To estimate the IgG antibody titer, 26 COVID-19 RT-PCR-confirmed samples were subjected to

serial dilutions from 1:100 to 1:12,800, each dilution was tested in duplicate in the CovIgG-Assay

and each experiment was replicated twice. A standard curve was created in which the mean

individual absorbance (A_{492}) of each sample at 1:100 dilutions was correlated with its

corresponding IgG antibody titer. Antibody titer was defined as the highest serum dilution that

renders A_{492} values greater than the cut point estimated by the ROC analysis.

Comparison with two serological assays approved for emergency use

We tested a set of 9 samples reported as IgM/IgG positives and 18 reported as IgG positives for

SARS-CoV-2 antibodies by CoronaCheck (20/20 BioResponse, 20/20 Genesystems, Inc,

Rockville, MA, USA). The information provided by the manufacturer claims that this assay use

Roche's technology (Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim,

Germany). Same set of 18 samples reported as SARS-CoV-2 IgG positive were also tested by

Abbott Architect SARS-CoV-2 IgG (Abbott Laboratories Diagnostics Division Abbott Park, IL

60064 USA). For comparison, both set of samples (n=27) were tested with our CovIgG-Assay.

Data analysis

Each CovIgG-Assay determination was performed in duplicate and the results expressed as the

mean absorbance at 492 nm (A_{492}) for each determination. The optimal cut point for the assay

was established within a 95% confidence interval (CI) by receiver operating characteristic (ROC)

curve analysis using the EpiTools epidemiological calculator (http://epitools.ausvet.com.au).

Arbitrary guidelines were followed for analyzing the area under curve (AUC) as follows: non-

informative, AUC=0.5; low accurate, 0.5 < AUC < 0.7; moderately accurate, 0.7 < AUC < 1;

perfect, AUC = 1 (9). Intra-plate repeatability was evaluated for CovIgG-Assay by measuring the

coefficient variation (CV) of 60 repeats of a High Positive Control (HPC) and a Negative Control

(NC). For reproducibility evaluation, we completed three independent runs for the CovIgG Assay

using HPC, NC, four (4) negative and four (4) COVID-19 positive sera for each of the three runs.

Correlation between the A_{492} at 1:100 dilutions and the antibody titer as well as between the

results of CovIgG Assay and the RT-PCR test results were evaluated using the Pearson

correlation coefficient (with the 95% CI). To evaluate the agreement between the CovIgG-Assay

and the RT-PCR, CovIgG-Assay and CoronaCheck or Abbott Architect SARS-CoV-2 IgG, inter-

rater agreement (kappa) was calculated according to the method described by Thrusfield (10).

The Kappa values (κ) were considered as follows: poor agreement, κ<0.02); fair agreement,

κ=0.21 to 0.4; moderate agreement, κ=0.41 to 0.6; substantial agreement, κ=0.61 to 0.8; very

good agreement, κ=0.81 to 1.0.

Results

Distribution of absorbance values of sera and ROC analysis

We used the RT-PCR for COVID-19 positive samples, as recommended standard reference

method, to build ROC curves on the basis of the absorbance values (A_{492}) obtained with

specimens from two reference populations: subjects infected with SARS-CoV-2 that were all RT-

PCR positive (assumed infected population) and healthy subjects or subjects that had been

diagnosed with other respiratory or viral infections prior to the COVID-19 pandemic (uninfected

population). The A_{492} values of uninfected population ranged between 0.011 and 0.312 with a

mean ± SD A_{492} value of 0.075 ± 0.052 whereas samples from assumed infected population

showed A_{492} values that ranged between 0.045 (one sample) and 3.21 with a mean A_{492} value of

1.99 ± 0.727. The mean value of the infected population was significantly different from the mean

value of the uninfected population ($p<0.0001$). The distribution of A_{492} values of these two

reference populations was very different. Approximately the 75% of infected population had A_{492}

values between 0.828 and 2.5 (median 2.01), whereas that the 95% of uninfected population had

A_{492} values between 0.011 and 0.176 (median 0.065) (**Figure-1**). Receiving operating

characteristic analysis was used to determine the best cut-points for the CovIgG-Assay. The ROC

optimized cut-point was 0.312. The selection of this cut-point derived from three different

conditions: (a) maximum specificity at which the sensitivity was still 100%, (b) maximum sensitivity

at which the specificity of the assay was also maximized, and (c) maximum value for Youden's J

index (S + Sp-1) and test efficiency (**Table-1**).

Table-1. Specificity (Sp) and sensitivity (Se) of the CovIgG-Assay based on cut-points results
obtained when compared infected and uninfected populations.
202

Target Sp	Cut-point	Specificity	Sp lower 95% CL	Sp Upper 95% CL	Sensitivity	Se Lower 95% CL	Se Upper 95% CL
0.999	0.745	1	0.961	1	0.98	0.893	0.996
0.995	0.745	1	0.961	1	0.98	0.893	0.996
0.990	0.745	1	0.961	1	0.98	0.893	0.996
0.98	**0.312**	**0.989**	**0.943**	**0.998**	**0.98**	**0.893**	**0.996**
0.95	0.202	0.958	0.897	0.984	0.98	0.893	0.996
0.9	0.129	0.905	0.83	0.949	0.98	0.893	0.996
0.8	0.099	0.8	0.709	0.868	0.98	0.893	0.996

The area under curve values (AUC) (accuracy) for the ROC curve was 0.985 (**Figure-2**). Based

on the established cut-point only one seronegative was detected in the infected group whereas

no seropositive was detected in the uninfected group. A sample from the uninfected group,

collected between 1995 and June 2019, had A_{492} values equal to the cut-point and was considered

negative (**Figure-3**).

To verify the cross reactivity of the assay we tested 67 samples know to be positive to common

respiratory and non-respiratory infections (RSV, Flu A and B, Mycoplasma, Zika, dengue,

Chikungunya) or allergies which are very common in the local population. As it is showed in figure

3, all those samples were negative showing no-cross reactivity in CovIgG-Assay. Based on these

results the estimated sensitivity if 98% (95% CI 89.35-99.95%) and the specificity is 100% (95%

CI 97.24-100%). Also the predictive positive value (PPV) and predictive negative value (NPV) for

CovIgG-Assay were estimated to be 100% and 99.25% (95% CI 94.99-99.89%), respectively

(**Table-2**). There was substantial agreement (κ=0.71) between CovIgG-Assay and RT-PCR.

Detailed optical densities (ODs) values of the positive and negative samples, including the cross-

reactivity panel are provided (**Supplementary tables 1 and 2 respectively**)

220
221
222   **Table-2. Agreement between test results of CovIgG-Assay and the PCR-based assay used as**
223   **reference method for COVID-19 diagnosis.**
224

		PCR-based assay		
		Positive	Negative	Total
CovIgG-Assay	**Positive**	48	0	48
	Negative	1	132	133
	Total	49	132	181

Positive Predictive Value (PPV)=100% is calculated as the number of individuals with a positive result by
CovIgG-Assay / the total of true positive individuals x 100. Negative Predictive Value (PNV)= 99.25% is
calculated as the number of individuals reported as negative by CovIgG-Assay / the total of true negative
individuals x 100.

We also assessed the reproducibility of the CovIgG-Assay by calculating the CV of data from 3

different assays and 30 repeats of controls and 6 repeats of selected samples. The intra-assay

and inter-assay reproducibility values were both lower than 15% (**Supplementary Table 3**).

Class antibody specificity of CovIgG-Assay

To confirm that the positivity showed by CovIgG-Assay with the COVID-19 samples was mostly

due to the presence of IgG antibody class and not due to potential cross-reactions with IgM

antibody, five samples treated with DTT were tested in parallel on the CovIgG-Assay using as

secondary antibody anti-human IgG- and anti-human IgM-HRP conjugates and the results

obtained were compared with those obtained for the same samples previous to the DTT

treatment. As expected, the A_{492} values of DTT-treated samples tested with the anti-IgM-HRP

conjugate significantly dropped to values similar to the background. In contrast, the A_{492} values

for the same DTT-treated samples tested with the anti-IgG-HRP conjugate were similar to those

obtained with the untreated samples, confirming that positive results were from IgG antibodies.

(**Table-3**).

245
246   **Table 3. Samples before and after treatment with DTT (Dithiothreitol) showing IgG class**
247   **specifity for CovIgG-Assay. Each result represent OD at 492nm absorbance.**

Numeric ID	IgG			IgG DTT			IgM			IgM DTT		
	OD1	OD2	Average OD	OD1	OD2	Average OD	OD1	OD2	Average OD	OD1	OD2	Average OD
45	2.7594	2.6820	2.7207	1.7638	1.7889	1.7763	1.6503	1.5807	1.6155	0.0244	0.0235	0.0239
121	3.1175	2.9996	3.0585	2.7779	2.6761	2.7270	3.5162	3.5980	3.5571	0.3303	0.3293	0.3298
122	2.7393	2.3745	2.5569	1.9655	1.9516	1.9585	2.6131	2.5388	2.5759	0.1060	0.1019	0.1039
146	2.7958	2.7973	2.7965	2.3467	2.3361	2.3414	1.0818	1.0508	1.0663	0.0898	0.0876	0.0887
147	2.8958	2.7455	2.8206	2.4553	2.3915	2.4234	3.2084	3.2769	3.2426	0.8243	0.8349	0.8296
183	1.4743	1.4506	1.4624	0.8341	0.8457	0.8399	3.3603	3.3997	3.3800	0.0607	0.0519	0.0563

248
249   Samples tested were previously confirmed by PCR and IgG/IgM rapid tests. Sample 183 was previously
250   confirmed by PCR.
251
Correlation between the A_{492} values and the IgG antibody titer

To determine whether the magnitude of the A_{492} values correlate with the antibody titer we

selected 27samples from infected individuals with different A_{492} values, which were titrated at

dilutions from 1:100 to 1:12,800 and each dilution was tested in duplicate in the CovIgG-Assay.

The number of individuals with different antibody titers (defined as the maximal dilution that

renders a positive result) is shown in Table-4. We found a lineal correlation (r^2=0.984) between

the antibody titer (maximal dilution that render A_{492} = 0.312) and the individual A_{492} value at the

working dilution (1:100). Thus, results reported by CovIgG-Assay could be quantitatively reported

by estimating the titer, using the lineal equation (Y= 1.185*X -1.773) derived from the lineal

correlation between antibody titer and the magnitude of absorbance values (**Figure-4**). Titers are

reported from 1:100 to 1:12,800 which were the dilutions tested to establish the OD at A_{492} and

the IgG titers. Positive samples rendering titers below or above of these values are reported as

<1:100 or >1:12,800.

Table 4. Experimental antibody titers of SARS-CoV-2 infected subjects in relation to their absorbance value at 492nm (1:100 dilution).

Antibody Titer*	Number of Individuals	Absorbance range at 492 nm (A_{492})	Mean A_{492} + SD
1:100	0	0	0
1:200	6	0.745-1.03	0.896 + 0.103
1:400	4	1.08-1.55	1.24 + 0.18
1:800	0	0	0
1:1600	7	1.98-2.52	2.2 + 0.203
1:3200	6	2.17-2.57	2.43 + 0.166
1:6400	2	2.37-3.05	2.71 + 0.34
1:12800	2	2.85-3.12	2.985 + 0.135

*Antibody titer is defined as the maximal serum dilution that renders A_{492} greater than the optimized cut-point (A_{492} = 0.312) determined by ROC analysis.

Agreement between CovIgG-Assay and other Rapid tests

To evaluate the performance of CovIgG-Assay with other tests in the market, we analyzed a group of samples that have been previously reported as positive for IgG/IgM (n=9) or only positive for IgG (n=18) by CoronaCheck rapid test. CovIgG-Assay had 100% agreement with the CoronaCheck results for the IgG/IgM positive samples. These samples were all reported as positive by CovIgG-Assay with antibody titers that ranged between 1:100 and 1:3,251 (Supplementary table 4). However, the agreement was fair (38.8%) for samples only reported positive for IgG by CoronaCheck since only 7 from 18 samples were reported as positive. From these sera 3 had antibody titers <1:100 and 4 had antibody titters among 1:219 to 1:254 (**Table-5**). Interestingly, all 18 presumptive IgG positive samples reported by CoronaCheck were found negative by Abbot Architect SARS-CoV-2 IgG method, which might subject that these 18 could be considered as false positive and reveal a better agreement between Abbot Architech SARS-

285 CoV-2 IgG method and our CovIgG-Assay. The correlation between CovIgG-Assay and Abbot

286 Architect method was of 61% (**Table 5**). Detailed OD values for this experiment are provided in

287 **supplementary table 4**.

289 **Table 5. Comparison of two rapid tests against CoVIgG-Assay for the detection of anti SARS CoV-**
290 **2 IgG antibodies**.

Sample	CoronaCheck[1]	CoVIgG-Assay[2]	CoV-2 IgG[3] (Abbott Architect)
	(Rapid COVID-19 IgG)	Estimated titer	Index value
46	+	-	.05
47	+	†	.03
48	+	-	.03
49	+	254	.05
50	+	-	.03
51	+	-	.05
52	+	-	.02
53	+	-	.04
54	+	254	.05
55	+	-	.04
56	+	-	.07
57	+	-	.37
58	+	†	.08
59	+	†	.02
60	+	-	.05
61	+	-	.01
62	+	219	.69
63	+	220	.05

Serum samples obtained during April 2020 were tested for SARS CoV-2 IgG antibodies using two different
assays. All samples tested positive with the Coronacheck test, when tested with the CoVIgG-Assay only 7
were positive for a 38.8% concordance. Same set of samples were tested with Abbott Architect CoV-2
(EUA) assay; all samples were negative for a 61% match when compared with CoVIgG-Assay. Positive
results for CoVIgG-Assay are expressed as estimated titers from the average of duplicates $OD_{492.}$
[1]Lateral flow rapid test intended to qualitatively detect antibodies to SARS-CoV-2. Presence of IgG/IgM
antibodies are indicated by a visible line in the specific region on the device. Positive results may be due
to cross reactivity with other Coronavirus strains (non CoV-2). Information regarding antigen used is not
available. [2]Quantitative ELISA for the detection SARS CoV-2 IgG antibodies using Spike S1-RBD
(GenScript) as capture antigen. Samples with $OD_{492}<0.312$ are considered negative. †IgG titration of
positive samples by CovIgG-Assay range from 1:100 to 1:12,800. Values below or above those limits are
reported as < or > respectively.
[3]Chemiluminescent microparticle immunoassay for the qualitative detection of IgG antibodies against
SARS CoV-2. Assay is designed to detect IgG antibodies to the nucleocapsid protein of SARS CoV-2.
Samples with an index value <1.4 are considered negative.
306

Discussion

Since the circulation of SARS-CoV-2 was spillover outside of China, the global efforts to develop

serological assays have been unprecedently huge. The precise diagnostic of COVID-19 poses

multiples challenges. The proposed method reference is the molecular diagnostic which

determine the presence of an active infection. However the timing of viral replication and the

development of immune response is quite variable (reviewed in (11)) and the presence of IgM

and or IgG at the time of the molecular diagnostic is merely speculative. While the molecular

testing results are a guide, they should not be considered gold standards as it has been the

practice so far. Other factors as clinical presentation and epidemiological aspects need to be

considered at the time of selecting the appropriate samples to validate any assay. Here we

selected 48 samples reported positive by authorized molecular methods and 1 sample reported

as positive by an authorized serological assay which is not considered a rapid test (7). As negative

samples we used a set of 132 sera that were collected in the period of 1995 to June 2019, before

COVID-19 period. The CovIgG-Assay data were subjected to ROC curve analysis. During the last

two decades this type of analysis has become a popular method for evaluating the accuracy of

medical diagnostic systems and has been used not only to evaluate the ability of a test to

discriminate between infected and healthy subjects (12) but also to compare the diagnostic

performance of a number of tests (13). The ROC curve is obtained by plotting the true-positive

rate (sensitivity) as a function of the false-positive rate (100-specificity) that is associated with

each cut-point. The AUC is then used as a measure of the accuracy of the test. If the assay can

distinguish between infected and normal populations the AUC will be equal to 1 and ROC curve

will reach the upper left corner. As our results demonstrate, the AUC value obtained from the

ROC curve analysis conducted on the CovIgG-Assay data was very high, indicating the accuracy

of this test. The only sample that was not detected by CovIgG-Assay could be a false positive in

the RT-PCR. Currently, there are few reports addressing those discordant results. Several

molecular assays have been developed with high specificity and low limit of detections (14-17)

and are considered the reference method for SARS-CoV-2 diagnostic (18). However everyday

there are more reports addressing problems with the RT-PCR accuracy (14, 19-21). Otherwise

that sample may be collected within a window where the immune response was not developed

yet. Nevertheless, our results reinforce the complexity of the diagnostic of COVID-19 and the

need for prospective studies with more samples and better-characterized cohorts to better

understand the dynamics of the immune response to this novel coronavirus.

We also considered fundamental to develop a quantitative test, in addition to suggesting a

qualitative result. This would provide a guide about quantity or the dimension of the immune

response mounted by an individual. Up to today, few works on COVID-19 addressed the relation

between the titer of the IgG and the neutralizing capability of that sample. But all of them coincide

that there is a direct correlation (1, 21-23). While the scientific community develops safe (BSL-2)

and reproducible neutralization assays to determine nAbs against SARS-CoV-2, quantitative

assays like CovIgG-Assay are useful tools for a reliable serological characterization.

When we compared the performance of an available rapid test and a high throughput assay, both

with EUA by the FDA. The results are highly worrisome showing 100% and 0% of positive samples

reported by CoronaCheck and Abbott Architect SARS-CoV-2 IgG respectively while our assay

reported a 38.8% % of the samples as positive. These results reinforce the need for better assays

and for better validations in the context of clinical presentations (24).

These results reinforce the concept recently proposed by the CDC to implement an orthogonal

testing algorithm (i.e., employing two independent tests in sequence when the first test yields a

positive result (shorturl.at/AHUY2). However the best characteristics of CovIgG-Assay are its PPV

of 100% along a PNV of 99.2% which render highly trustable positive results. Theses

characteristics made this assay suitable to be implemented both in a high or low prevalence

settings. A subject confirmed as positive by this assay, with a high degree of confidence can be

reincorporating to a normal life in support of the rest of the community with a high degree of

confidence of having certain level of immunity. On the other hand the high NPV of the assay

increase its value to know which individuals may be still at risk to be infected.

Furthermore, the quantitative aspect of CovIgG-Assay, after a subject being reported as positive,

provides a second step of certainty of possible protection against SARS-CoV-2 infection. However

such correlation studies using a larger set of samples are under way.

Statistical analysis

The sensitivity of the developed ELISAs was calculated as (the number of true positive samples

/ the total number of true positive and false negative samples) x 100. The specificity was

calculated as (the number of true negative samples / the total number of true negative and false

positive samples) x 100. As we do not know the prevalence of the tested population, the PPV

(probability that the disease is present when the test is positive) was calculated as the number of

individuals with a positive result by CovIgG-Assay / the total of true positive individuals x 100. The

NPV was calculated as the number of individuals reported as negative by CovIgG-Assay / the

total of true negative individuals x 100. Agreement was calculated as (the total number of true

positive and negative samples / the total number of samples) X 100 and evaluated with kappa

values. Receiver operating characteristic (ROC) analysis was calculated using GraphPad Prism

software.

Authors Contribution

AME and CAS conceived and supervised the studies. PP and AME performed the experiments.

CAS and AME drafted the manuscript. CAS, AME and PP reviewed the final version of the

manuscript. CAS and AME obtained the funds. CAS prepared and CAS and AME submitted the

Emergency Use Authorization request to the US Federal Drug Administration.

Acknowledgement

Authors want to thank Ilia Toledo, MT and Drs. Consuelo Climent and Gerardo Latoni and Ivelisse

Martin for their contribution and diligent access to the samples from subject presumptive exposed

to SARS-CoV-2 and from plasma donors. Also, thanks to Dr. Jorge L. Muñoz-Jordan for providing

the panel to improve the cross-reactivity testing. Particular acknowledgement deserves all

administrative and supportive staff at the Medical Sciences Campus, University of Puerto Rico,

Laboratorio Clinico Toledo, Laboratorio Clinico Martin, Banco de Sangre Centro Médico and

Banco de Sangre Servicios Mutuos for their availability and commitment during the curfew

imposed by the quarantine period. This work was supported by the COVID Rapid Grant from the

Puerto Rico Science Technology and Research Trust (PRSTRT https://prsciencetrust.org/) to

AME and CAS. The excellent support of Mss. Andreica Maldonado and Grace Rendon from

PRSTRT was instrumental in support of the work described here.

References

1. Amanat F, Stadlbauer D, Strohmeier S, Nguyen THO, Chromikova V, McMahon M, et al. A serological assay to detect SARS-CoV-2 seroconversion in humans. Nat Med. 2020.
2. Al Kahlout RA, Nasrallah GK, Farag EA, Wang L, Lattwein E, Muller MA, et al. Comparative Serological Study for the Prevalence of Anti-MERS Coronavirus Antibodies in High- and Low-Risk Groups in Qatar. J Immunol Res. 2019;2019:1386740.
3. Hsueh PR, Huang LM, Chen PJ, Kao CL, Yang PC. Chronological evolution of IgM, IgA, IgG and neutralisation antibodies after infection with SARS-associated coronavirus. Clin Microbiol Infect. 2004;10(12):1062-6.
4. Ko JH, Müller MA, Seok H, Park GE, Lee JY, Cho SY, et al. Serologic responses of 42 MERS-coronavirus-infected patients according to the disease severity. Diagn Microbiol Infect Dis. 2017;89(2):106-11.
5. Mo HY, Xu J, Ren XL, Zeng GQ, Tan YX, Chen RC, et al. Evaluation by indirect immunofluorescent assay and enzyme linked immunosorbent assay of the dynamic changes of serum antibody responses against severe acute respiratory syndrome coronavirus. Chin Med J (Engl). 2005;118(6):446-50.
6. Hashem AM, Al-Amri SS, Al-Subhi TL, Siddiq LA, Hassan AM, Alawi MM, et al. Development and validation of different indirect ELISAs for MERS-CoV serological testing. J Immunol Methods. 2019;466:41-6.
7. Stadlbauer D, Amanat F, Chromikova V, Jiang K, Strohmeier S, Arunkumar GA, et al. SARS-CoV-2 Seroconversion in Humans: A Detailed Protocol for a Serological Assay, Antigen Production, and Test Setup. Curr Protoc Microbiol. 2020;57(1):e100.
8. Krammer F, Simon V. Serology assays to manage COVID-19. Science. 2020.
9. Swets JA. Measuring the accuracy of diagnostic systems. Science. 1988;240(4857):1285-93.
10. Thrusfield M. Veterinary epidemiology. London, United Kingdom, 2nd ed Blackwell Science. 1995.
11. Vabret N, Britton GJ, Gruber C, Hegde S, Kim J, Kuksin M, et al. Immunology of COVID-19: Current State of the Science. Immunity. 2020.
12. Hajian-Tilaki K. Receiver Operating Characteristic (ROC) Curve Analysis for Medical Diagnostic Test Evaluation. Caspian J Intern Med. 2013;4(2):627-35.
13. Surujballi O, Mallory M. Competitive enzyme-linked immunosorbent assay for detection of Leptospira interrogans serovar pomona antibodies in bovine sera. Clin Diagn Lab Immunol. 2001;8(1):40-3.
14. Arevalo-Rodriguez I, Buitrago-Garcia D, Simancas-Racines D, Zambrano-Achig P, del Campo R, Ciapponi A, et al. FALSE-NEGATIVE RESULTS OF INITIAL RT-PCR ASSAYS FOR COVID-19: A SYSTEMATIC REVIEW. medRxiv. 2020:2020.04.16.20066787.
15. Chan JF, Yip CC, To KK, Tang TH, Wong SC, Leung KH, et al. Improved Molecular Diagnosis of COVID-19 by the Novel, Highly Sensitive and Specific COVID-19-RdRp/Hel Real-Time Reverse Transcription-PCR Assay Validated In Vitro and with Clinical Specimens. J Clin Microbiol. 2020;58(5).
16. Loeffelholz MJ, Tang YW. Laboratory diagnosis of emerging human coronavirus infections - the state of the art. Emerg Microbes Infect. 2020;9(1):747-56.
17. Tang YW, Schmitz JE, Persing DH, Stratton CW. Laboratory Diagnosis of COVID-19: Current Issues and Challenges. J Clin Microbiol. 2020;58(6).
18. CDC. Interim Clinical Guidance for Management of Patients with Confirmed Coronavirus Disease (COVID-19). 2020;Updated June 2.
19. Byambasuren O, Cardona M, Bell K, Clark J, McLaws M-L, Glasziou P. Estimating the extent of asymptomatic COVID-19 and its potential for community transmission: systematic review and meta-analysis. medRxiv. 2020:2020.05.10.20097543.

20. Cohen AN, Kessel B. False positives in reverse transcription PCR testing for SARS-CoV-2.
medRxiv. 2020:2020.04.26.20080911.
21. Jiang S, Hillyer C, Du L. Neutralizing Antibodies against SARS-CoV-2 and Other Human
Coronaviruses. Trends Immunol. 2020;41(5):355-9.
22. Dingens AS, Crawford KH, Adler A, Steele SL, Lacombe K, Eguia R, et al. Seroprevalence
of SARS-CoV-2 among children visiting a hospital during the initial Seattle outbreak.
medRxiv. 2020:2020.05.26.20114124.
23. Okba NMA, Müller MA, Li W, Wang C, GeurtsvanKessel CH, Corman VM, et al. Severe Acute
Respiratory Syndrome Coronavirus 2-Specific Antibody Responses in Coronavirus Disease
2019 Patients. Emerg Infect Dis. 2020;26(7).
24. Woloshin S, Patel N, Kesselheim AS. False Negative Tests for SARS-CoV-2 Infection —
Challenges and Implications. New England Journal of Medicine. 2020.
457
458

FIGURES
460

461
462
463
Figure 1: Distribution of absorbance values obtained with CovIgG-Assay. The CovIgG-
Assay was optimized using as antigen recombinant Spike-S1-RBD from SARS-CoV-2.
Absorbance values were distributed in form of frequency histograms to clearly visualize the
separation between true positives (COVID-19 sera) (upper figure) and true negative populations

(sera from healthy subjects and subjects carrying other respiratory and viral infections) (lower
figure).

470
471

472
473
474   **Figure 2: Receiver-operator characteristic (ROC) curve.** The ROC curve was built for 132
475   sera from healthy subjects or subjects carrying other respiratory or other viral infections and 49
476   COVID-19 confirmed subjects. The area under the ROC curve (accuracy) was 0.985 and 95%
477   for AUC= 0.954-1.
478

480
481
482
Figure 3: Validation of use of Spike S1-RBD ELISA for detection of SARS-CoV-2 IgG antibodies.
Non-CoV-2 infection indicate negative cohorts from samples prior 2019 with no previous history of
selected viral infections or respiratory allergies (n= 78). Black dots indicate samples which tested
positive for Mycoplasma IgM, (n = 9), Influenza A or B (n=13), respiratory allergies (n=13), ZIKV
(n=14), CHIKV (n=3), DENV (n=8). Gray dots indicate SARS CoV-2 infections (n=49). Dotted
horizontal line indicate CoVIgG ELISA cutoff value (OD_{492}= 0.312). S, spike; RBD, receptor-binding
domain; ZIKV, Zika virus; CHIKV, Chikungunya virus; DENV, Dengue virus; SARS-CoV-2, severe
acute respiratory syndrome coronavirus 2. ODs in duplicate for every sample are showed.
491
492

Best-fit values	
Slope	1.185 ± 0.07537
Y-intercept when X=0.0	-1.773 ± 0.2496
X-intercept when Y=0.0	1.497
1/slope	0.8440
95% Confidence Intervals	
Slope	0.9756 to 1.394
Y-intercept when X=0.0	-2.466 to -1.080
X-intercept when Y=0.0	1.101 to 1.779
Goodness of Fit	
R square	0.9841

Equation: Y= 1.185* X- 1.773

494
495
496
497
Figure 4: Correlation between absorbance at 492nm (A_{492}) and antibody titter. A
total of 27 sera from confirmed COVID-19 subjects that resulted positive by CovIgG-
Assay were titrated at dilutions among 1:100 to 1:12,800. A lineal regression analysis
was then done in which the mean A_{492} of sera with similar antibody titer were plotted
with their corresponding A492 values. We found a lineal correlation (r^2=0.984) between
the antibody titer (maximal dilution that render A_{492} = 0.312) and the individual A_{492}
value at the working dilution (1:100). From this analysis the following lineal equation
(Y= 1.185*X -1.773) was obtained, which was further used to estimate the antibody titer
of all sera reported as seropositive by CovIgG-Assay.