1   **Validation and performance of a quantitative IgG assay for the screening of SARS-**

2   **CoV-2 antibodies**

Ana M. Espino[a*#], Petraleigh Pantoja[a*] and Carlos A. Sariol[a*#]

[a]Department of Microbiology and Medical Zoology, University of Puerto Rico-Medical

Sciences Campus, San Juan, PR, USA.

Running Title: UPR-MSC CovIgG-Assay

* Ana M. Espino, Petraleigh Pantoja and Carlos A. Sariol contributed equally to this work.

Author order was determined based on the basis of data analyses

Address correspondence to Ana M Espino, ana.espino1@upr.edu or Carlos A. Sariol

carlos.sariol1@upr.edu

Abstract

The current COVID-19 epidemic imposed an unpreceded challenge to the scientific community in terms of treatment, epidemiology, diagnosis, social interaction, fiscal policies and many other areas. The development of accurate and reliable diagnostic tools (high specificity and sensitivity) is crucial in the current period, the near future and in the long term. These assays should provide guidance to identify immune presumptive protected persons, potential plasma, and/or B cell donors and vaccine development among others. Also, such assays will be contributory in supporting prospective and retrospective studies to identify the prevalence and incidence of COVID-19 and to characterize the dynamics of the immune response. As of today, only thirteen serological assays have received the Emergency Use Authorization (EUA) by the U.S. Federal Drug Administration (FDA). In this work we describe the development and validation of a quantitative IgG enzyme-linked immunoassay (ELISA) using the recombinant SARS-CoV-2 Spike Protein S1 domain, containing the receptor-binding domain (RBD), showing 98% sensitivity, 98.9% specificity and positive and negative predictive values of 100% and 99.2%, respectively. The assay showed to be useful to test for SARS-CoV-2 IgG antibodies in plasma samples from COVID-19-recovered subjects as potential donors for plasmapheresis. This assay is currently under review by the Federal Drug Administration for an Emergency Use Authorization request (Submission Number EUA201115).

Introduction

The current severe acute respiratory syndrome coronavirus 2 (SARS CoV-2) pandemic and the resulting unprecedented outbreak of coronavirus disease 2019 (COVID-19) have shifted the paradigm for viral research, epidemiology and diagnostic. Both molecular and serological methods have been developed at an extraordinary speed. As of April 2, 2020 only four months after the virus was detected for first time in Wuhan region, 28 companies obtained Emergency Use Authorization (EUA) approvals from US Federal Drug Administration (FDA) for their commercial Reverse Transcription-Polymerase Chain Reaction (RT-PCR) diagnostics. Those assays are intended to detect the virus during the acute phase of the infection, providing no information regarding the immunological status of these patients. By the same time, from the more than 25 rapid serological tests available only one had the EUA granted. These rapid tests are relatively simple to perform and interpret and therefore require limited test operator training. The main drawback of these rapid tests is that the specificity and particularly the sensitivity are lower than the standard Enzyme-linked Immunosorbent Assays (ELISA). As of June 1, 2020 FDA had received more than 198 notifications from manufacturers confirming they have validated and intend to distribute their tests in the market. However only 13 of those tests have indeed the EUA from FDA. Moreover, in May 2020, FDA removed 28 SARS-CoV-2 serological tests from the notification list of tests offered during the COVID-10 emergency for not having an EUA request. Choosing an appropriate test to screen for the presence of humoral immune response to SARS-CoV-2 is critical. Such serologic tests are expected to play a key role in the fight against COVID-19 by helping to identify individuals who had developed an adaptive immune response and may be at lower risk of infection. Also, validated serological tests are needed to confirm which subjects, being confirmed positive for COVID-19, truly developed a substantial humoral immune response and may be considered as plasma donors (1). Different antigens have been used to detect antibodies against another novel

coronavirus such as SARS-CoV and MERS-CoV (2-5). From these previous works it can

be concluded that spike-derived (S) antigens are more sensitive, specific and accurate than

nucleocapsid protein-derived (NP) antigens. Also results from assays using S antigens

correlated much better with the neutralizing titers than those using NP antigens (6). A

recent work showed the usefulness for the Receptor Binding Domain (RBD) and the full

Spike protein to detect SARS-CoV-2 specific antibodies (7) and their correlation with

neutralizing antibodies nAb (8). For these reasons we choose to use a recombinant SARS-

CoV-2 Spike Protein, S1 domain containing the RBD.

With this work we described the validation of a quantitative ELISA, CovIgG-Assay

(https://prsciencetrust.org/the-covigg-assay-kit/), showing a very low background and lack

of cross reactivity with other respiratory and non-respiratory pathogens in more than 132

samples collected before June 2019. Also the correlation with three serological tests

available in the market is described. Finally, we confirm the usefulness of the assay

detecting anti-SARS-CoV-2 antibodies in plasma samples from potential plasma donors.

CovIgG-Assay is a useful tool to characterize, quantify and to study the dynamics of the

humoral immune response to SARS-CoV-2.

Materials and Methods

Study Design

The study population included a total of 181 samples. Forty-nine (49) samples were from

individuals with symptomatic infection and positive diagnosis for SARS-CoV-2. Forty-eight

(48) were confirmed by RT-PCR tests EUA authorized and one (1) diagnosed by COVID-19

ELISA IgG Antibody Test – Mount Sinai, also EUA authorized. De-identified serum or

plasma specimens were obtained from local clinical laboratories and Blood Banks and no

personal identifiers were retained. The other 132 de-identified samples had been taken

previously to 2019 and belonged to the Virology or the Immunology UPR-RCM serum bank.

From these samples, 78 had no previous history of viral, allergic or bacterial infections

according to our cross-reactivity panel. Nine (9) were previously diagnosed with Zika, three

(3) with Dengue, thirteen (13) with history of respiratory allergies and one (1) with Influenza

H1N1. We also included a cross reactivity panel with 28 samples kindly donated by the

Centers for Disease Control and Prevention (CDC) Dengue Branch, San Juan, PR. These

samples included six (6) positives for Respiratory Syncytial Virus (RSV)-IgM, twelve (12)

RT-PCR positive for Influenza A or B, five (5) Zika-IgM positive and five (5) positive for

Dengue-IgM. This cross-reactive panel was selected according to the most common viral

and respiratory infections affecting our population. Additionally, we tested nine (9) samples

from individuals that resulted positive for Mycoplasma-IgM and three (3) positives for

Chikungunya, which were collected during COVID-19 pandemic. Although these 12

samples were included in the cross-reactive study they were excluded from the statistical

analysis to establish the cut-point and diagnostic specificity/sensitivity of CovIgG-Assay. All

samples were stored at -80°C until use.

For comparison with two others serological tests (CoronaCheck and Abbott Architect)

holding an EUA, we used a set of nine (9) samples assumed to be positive for IgG and IgM

and eighteen (18) assumed to be IgG positive for SARS-CoV-2 antibodies. Those samples

were also received de-identified from local laboratories.

CovIgG-Assay

CovIgG-Assay is an indirect ELISA for quantitative determination of human IgG antibody

class, which was optimized by checkerboard titration. Disposable high bind flat-bottomed

polystyrene 96-wells microtiter plates (Costar, Corning MA No. 3361) were coated

overnight at 4^{o}C with 2µg/ml of recombinant SARS-CoV-2 S1-RBD protein (GenScript No.

Z03483-1) in carbonate-bicarbonate buffer (Sigma Aldrich No. 08058). Plates were washed

3 times with phosphate buffered saline (PBS) containing 0.05% Tween-20 (PBST) and

blocked for 30 min at 37^{o}C with 250µl/well of 3% non-fat, skim milk in PBST. Samples

(serum or plasma) were diluted 1:100 in PBST; 100µL/well was added in duplicates and

incubated at 37^{o}C for 30 min. The excess antibody was washed off with PBST. Horseradish

peroxidase (HRP) labeled-mouse anti-human IgG-Fc specific (GenScript No. A01854)

diluted 1:10,000 in PBST was added (100µl/well) and incubated for 30 min at 37^{o}C. After

another washing step, the substrate solution (Sigma Aldrich No. P4809) was added

(100µl/well) followed by 15 min incubation in dark. The reaction was stopped by the

addition of 50µl/well 10% HCl and the absorbance was measured at 492nm (A_{492}) using a

Multiskan FC reader (Thermo Fisher Scientific). In every CovIgG-Assay determination two

in-house controls, a high positive control (HPC) and negative control (NC) were included.

HPC and NC were prepared by diluting an IgG anti-SARS-CoV-2 at a concentration of

30µg/ml and 0.070µg/ml, respectively in PBST containing 10% glycerol. The IgG anti-

SARS-CoV-2 was purified from plasma of a convalescent patient using a 5/5 HiTrap

rProtein-A column (GE Healthcare, USA). See detailed information about this procedure in

Supplementary method No.1.

Antibody class specificity

To confirm that our assay accurately detects antibody IgG class and excludes the potential for human IgM to cross-react with IgG, five (5) COVID-19 samples (1:100 diluted) were treated with 5mM DTT for 30 min at 37^{o}C prior testing. After treatment, samples were added in duplicate (100µl/well) followed by the addition of the anti-human IgG-Fc-HRP (GenScript No. A01854) conjugate (diluted 1:10,000) or the addition of an anti-human IgM-HRP conjugate (Abcam No. ab97205) diluted 1:8,000 in PBST and the assay progressed as described above.

Estimation of Antibody Titer

To estimate the IgG antibody titer, 40 COVID-19 samples were subjected to serial dilutions from 1:100 to 1:12,800. Each dilution was tested in duplicate in the CovIgG-Assay and each experiment was replicated twice. A standard curve was created in which the mean individual absorbance (A_{492}) of each sample at 1:100 dilutions was correlated with its corresponding IgG antibody titer. Antibody titer was defined as the highest serum dilution that renders A_{492} values greater than the cut point estimated by the ROC analysis.

Comparison with other serological assays approved for emergency use

We tested a set of 9 samples reported as IgM/IgG positives and 18 reported as IgG positives for SARS-CoV-2 antibodies by CoronaCheck (20/20 BioResponse, 20/20 Genesystems, Inc, Rockville, MA, USA). The information provided by the manufacturer claims that this assay use Roche's technology (Roche Diagnostics GmbH, Sandhofer Strasse 116, D-68305 Mannheim, Germany). Same set of 18 samples reported as SARS-CoV-2 IgG positive were also tested by Abbott Architect SARS-CoV-2 IgG (Abbott Laboratories Diagnostics Division Abbott Park, IL 60064 USA). For comparison, both set of samples (n=27) were tested with our CovIgG-Assay. Moreover, another set of 18 samples

from convalescent COVID-19 subjects, which had been confirmed by PCR were tested by

CovIgG-Assay and Elecsys Anti-SARS-CoV-2 method (Cobas).

Data analysis

Each CovIgG-Assay determination was performed in duplicate and the results expressed

as the mean absorbance at 492 nm (A_{492}) for each determination. The optimal cut point for

the assay was established within a 95% confidence interval (CI) by receiver operating

characteristic (ROC) curve analysis using the EpiTools epidemiological calculator

(http://epitools.ausvet.com.au). Arbitrary guidelines were followed for analyzing the area

under curve (AUC) as follows: non-informative, AUC=0.5; low accurate, $0.5 < AUC < 0.7$;

moderately accurate, $0.7 < AUC < 1$; perfect, $AUC = 1$ (9). Intra-plate repeatability was

evaluated for CovIgG-Assay by measuring the coefficient variation (CV) of 60 repeats of a

High Positive Control (HPC) and a Negative Control (NC). For reproducibility evaluation, we

completed three independent runs in different days for the CovIgG Assay using HPC, NC

(30 replicates), four (4) negative and four (4) COVID-19 positive sera (6 replicates).

Correlation between the A_{492} at 1:100 dilutions and the antibody titer as well as between the

results of CovIgG Assay and the RT-PCR test results were evaluated using the Pearson

correlation coefficient (with the 95% CI). To evaluate the agreement between the CovIgG-

Assay and the RT-PCR, CovIgG-Assay and CoronaCheck, Abbott Architect SARS-CoV-2

IgG or Elecsys, inter-rater agreement (kappa) was calculated according to the method

described by Thrusfield (10). The Kappa values (κ) were considered as follows: poor

agreement, κ<0.02); fair agreement, κ=0.21 to 0.4; moderate agreement, κ=0.41 to 0.6;

substantial agreement, κ=0.61 to 0.8; very good agreement, κ=0.81 to 1.0.

Results

Distribution of absorbance values of sera and ROC analysis

We used the RT-PCR for COVID-19 positive samples, as recommended standard

reference method, to build ROC curves on the basis of the absorbance values (A_{492})

obtained with specimens from two reference populations: subjects infected with SARS-

CoV-2 that were all RT-PCR positive (assumed infected population) and healthy subjects or

subjects that had been diagnosed with other respiratory or viral infections prior to the

COVID-19 pandemic (uninfected population). The A_{492} values of uninfected population

ranged between 0.011 and 0.312 with a mean ± SD A_{492} value of 0.075 ± 0.052 whereas

samples from assumed infected population showed A_{492} values that ranged between 0.045

(one sample) and 3.21 with a mean A_{492} value of 1.99 ± 0.727. The mean value of the

infected population was significantly different from the mean value of the uninfected

population ($p<0.0001$). The distribution of A_{492} values of these two reference populations

was very different. Approximately the 75% of infected population had A_{492} values between

0.828 and 2.5 (median 2.01), whereas that the 95% of uninfected population had A_{492}

values between 0.011 and 0.176 (median 0.065) (**Figure-1**). Receiving operating

characteristic analysis was used to determine the best cut-points for the CovIgG-Assay.

The ROC optimized cut-point was 0.312. The selection of this cut-point derived from three

different conditions: (a) maximum specificity at which the sensitivity was still 100%, (b)

maximum sensitivity at which the specificity of the assay was also maximized, and (c)

maximum value for Youden's J index (S + Sp-1) and test efficiency (**Table-1**).

The area under curve values (AUC) (accuracy) for the ROC curve was 0.985 (**Figure-2**).

Based on the established cut-point only one seronegative was detected in the infected

group whereas no seropositive was detected in the uninfected group. A sample from the

uninfected group, collected between 1995 and June 2019, had A_{492} values equal to the cut-

point and was considered negative. Significant differences ($p<0.0001$) were obtained

between the mean OD values of COVID-19 infection sera (1.99 ± 0.727) compared to those

from uninfected subjects (0.075 ± 0.052).

To verify the cross reactivity of the assay we tested 67 samples know to be positive to

common respiratory and non-respiratory infections (RSV, Flu A and B, Zika, and dengue)

or allergies which are very common in the local population and that had been collected

prior pandemic. As it is showed in figure 3, all those samples were negative showing no-

cross reactivity in CovIgG-Assay. Thus, under these optimized conditions, CovIgG-Assay

reached 98.9% specificity and 98.0% sensitivity with estimated predictive positive value

(PPV) and predictive negative value (PNV) for CovIgG-Assay of 100% and 99.2%,

respectively (**Table-2**). Importantly, all the 12 samples from individuals with Mycoplasma

and Chikungunya that were collected during pandemic also resulted negative in the

CovIgG-Assay (Figure-3), which confirmed the absence of cross-reaction in the CovIgG-

Assay. There was substantial agreement (97.95%, κ=0.657) between CovIgG-Assay and

RT-PCR. Detailed optical densities (ODs) values of the positive and negative samples,

including the cross-reactivity panel are provided (**Supplementary tables 1 and 2**

respectively). We also assessed the reproducibility of the CovIgG-Assay by calculating the

CV of data from 3 different assays and 30 repeats of controls and 6 repeats of selected

negative and positive samples. The intra-assay and inter-assay reproducibility values were

both lower than 10% (**Supplementary Table 3**).

Class antibody specificity of CovIgG-Assay

To confirm that the positivity showed by CovIgG-Assay with the COVID-19 samples was

mostly due to the presence of IgG antibody class and not due to potential cross-reactions

with IgM antibody, five samples treated with DTT were tested in parallel on the CovIgG-

Assay using as secondary antibody anti-human IgG- and anti-human IgM-HRP conjugates

and the results obtained were compared with those obtained for the same samples

previous to the DTT treatment. As expected, the A_{492} values of DTT-treated samples tested

with the anti-IgM-HRP conjugate significantly dropped to values similar to the background.

In contrast, the A_{492} values for the same DTT-treated samples tested with the anti-IgG-HRP

conjugate were similar to those obtained with the untreated samples, confirming that

positive results were from IgG antibodies. (**Table-3**).

Correlation between the A_{492} values and the IgG antibody titer

To determine whether the magnitude of the A_{492} values correlate with the antibody titer we

selected 40 samples from infected individuals with A_{492} values among 0.321 to 3.12, which

were the lowest and the greatest A_{492} obtained from the sample population studied,

respectively. All 40 sera were diluted from 1:100 to 1:12,800 and each dilution was tested

in duplicate in the CovIgG-Assay. The number of individuals with different antibody titers

(defined as the maximal dilution that renders a positive result) is shown in Table-4. We

found a lineal correlation (r^2=0.9946) between the antibody titer (maximal dilution that

render A_{492} > 0.312) and the individual A_{492} value at the working dilution (1:100). Thus,

results reported by CovIgG-Assay could be quantitatively reported by estimating the titer,

using the lineal equation (Y= 1.268*X -2.036) derived from the lineal correlation between

antibody titer and the magnitude of absorbance values (**Figure-4**). Based on this analysis

antibody estimated titers are reported in the range among 1:100 to 1:12,800. Samples with

A_{492} in the range of 0.312 to 0.49 would have antibody titer lower than 1:100. Such a

samples would be considered as weakly positive with undetermined antibody titer. It would

be highly recommendable that another sample from such subjects collected 2-3 weeks

thereafter can be tested. Samples with A_{492} >3.12 are reported with estimated antibody titer

>1:12,800 (**Supplementary table 4**).

Agreement between CovIgG-Assay and other serological tests

To evaluate the performance of CovIgG-Assay with other tests in the market, we analyzed

a group of samples that have been previously reported as positive for IgG/IgM (n=9) or only

positive for IgG (n=18) by CoronaCheck rapid test. CovIgG-Assay had 100% agreement

with the CoronaCheck results for the IgG/IgM positive samples. These samples were all

reported as positive by CovIgG-Assay with antibody titers that ranged between 1:100 and

1:3,251. However, the agreement was fair (38.88%) for samples only reported positive for

IgG by CoronaCheck since 7 from 18 samples were reported as positive by CovIgG-Assay

(**Table-5**). Interestingly, all 18 these presumptive IgG positive samples were found negative

by the Abbot Architec SARS-CoV-2 IgG method, which reveal a better agreement (61.0%)

between our CovIgG-Assay method and Abbot Architec SARS-CoV-2 IgG (**Table-5**) and

might suggest that most of these 18 samples could be false positives.

In another experiment, samples from subjects confirmed by PCR (n=18) were analyzed by

our CovIgG-Assay and Elecsys Anti-SARS-CoV-2 method. CovIgG-Assay reported as

positive all these 18 samples with antibody estimated titers ranging among 1:200 and

>1:12,800 whereas Elecsys Anti-SARS-CoV-2 method reported 17 positive. Thus, very

good agreement (94.4%) between Elecsys and our CovIgG-Assay was observed (**Table-5).**

Discussion

Since the circulation of SARS-CoV-2 was spillover outside of China, the global efforts to

develop serological assays have been unprecedently huge. The precise diagnostic of

COVID-19 poses multiples challenges. The proposed method reference is the molecular

diagnostic, which determines the presence of an active infection. However the timing of

viral replication and the development of immune response is quite variable (reviewed in

(11) and the presence of IgM and or IgG at the time of the molecular diagnostic is merely

speculative. While the molecular testing results are a guide, they should not be considered

gold standards, as it has been the practice so far. Other factors as clinical presentation and

epidemiological aspects need to be considered at the time of selecting the appropriate

samples to validate any assay. Here we selected 48 samples reported positive by

authorized molecular methods and 1 sample reported as positive by an authorized

serological assay which is not considered a rapid test (7). As negative samples we used a

set of 132 sera that were collected in the period of 1995 to June 2019, before COVID-19

period. The CovIgG-Assay data were subjected to ROC curve analysis. During the last two

decades this type of analysis has become a popular method for evaluating the accuracy of

medical diagnostic systems and has been used not only to evaluate the ability of a test to

discriminate between infected and healthy subjects (12) but also to compare the diagnostic

performance of a number of tests (13). The ROC curve is obtained by plotting the true-

positive rate (sensitivity) as a function of the false-positive rate (100-specificity) that is

associated with each cut-point. The AUC is then used as a measure of the accuracy of the

test. If the assay can distinguish between infected and normal populations the AUC will be

equal to 1 and ROC curve will reach the upper left corner. As our results demonstrate, the

AUC value obtained from the ROC curve analysis conducted on the CovIgG-Assay data

was very high, indicating the high accuracy of this test. The only sample that was not

detected by CovIgG-Assay could be a false positive in the RT-PCR. Currently, there are

few reports addressing those discordant results. Several molecular assays have been

developed with high specificity and low limit of detections (14-17) and are considered the

reference method for SARS-CoV-2 diagnostic (18). However everyday there are more

reports addressing problems with the RT-PCR accuracy (14, 19-21). Otherwise that sample

may be collected within a window where the immune response was not developed yet.

Nevertheless, our results reinforce the complexity of the diagnostic of COVID-19 and the

need for prospective studies with more samples and better-characterized cohorts to expand

our understanding of the dynamics of the immune response to this novel coronavirus.

We also considered fundamental to develop a quantitative test, in addition to suggesting a

qualitative result. This would provide a guide about quantity or the dimension of the

immune response mounted by an individual. Up to today, few works on COVID-19

addressed the relation between the titer of the IgG and the neutralizing capability of that

sample. But all of them coincide that there is a direct correlation (1, 21-23). While the

scientific community develops safe (BSL-2) and reproducible neutralization assays to

determine nAbs against SARS-CoV-2, quantitative assays like CovIgG-Assay are useful

tools for a reliable serological characterization.

The notable disagreement observed between CoronaCheck and Abbot Architec SARS-

CoV-2 IgG method for the same set of samples (100% and 0% positive samples,

respectively) resulted surprising and at the same time worrisome since both methods have

FDA EUA. Since these samples were kindly donated by Clinical Laboratories de-identified

we unknown whether the presence of virus was confirmed in any of these subjects.

However, because our assay had significant agreement with the Abbot method (61%) and

very good agreement (94%) with Elecsys method using a set of samples from PCR-

confirmed SARS-CoV-2 infection, we could suggest that most IgG positive samples

reported by CoronaCheck might be false positive.   Furthermore we have recent results

showing that the sample reported as negative by Elecsys and positive by CovIgG-Assay

had neutralizing antibodies (data not shown-manuscript in preparation). Together those

results reinforce the notion that our method is capable to offer accurate and efficient

diagnostic testing for detection of antiviral antibodies in infected individuals.

The finding of contradictory results in the performance of different immunoenzymatic

diagnostic methods strengthen the need for better assays and for better validations in the

context of clinical presentations (24). The best characteristics of CovIgG-Assay are its PPV

of 100% along a PNV of 99.2%, which render highly trustable positive results. A subject

confirmed as positive by this assay, with a high degree of confidence can be

reincorporating to a normal life in support of the rest of the community. On the other hand

the high NPV of the assay increase its value to know which individuals may be still at risk to

be infected. Altogether, the quantification of the IgG, after a subject being reported as

positive, provides a second step of certainty of possible protection against SARS-CoV-2

infection. However such correlation studies using a larger set of samples are under way.

Authors Contribution

AME and CAS conceived and supervised the studies. PP and AME performed the

experiments. CAS and AME drafted the manuscript. CAS, AME and PP reviewed the final

version of the manuscript. CAS and AME obtained the funds. CAS prepared and CAS and

AME submitted the Emergency Use Authorization request to the US Federal Drug

Administration.

Acknowledgement

Authors want to thank Ilia Toledo, MT, Francheska Rivera, MT and Drs. Consuelo Climent,

Gerardo Latoni and Ivelisse Martin for their contribution and diligent access to the samples

from subject presumptive exposed to SARS-CoV-2 and from plasma donors. Also, thanks

to Dr. Jorge L. Muñoz-Jordan for providing the panel to improve the cross-reactivity testing.

Particular acknowledgement deserves all administrative and supportive staff at the Medical

Sciences Campus, University of Puerto Rico, Laboratorio Clinico Toledo, Laboratorio

Clinico Martin, Banco de Sangre Centro Médico and Banco de Sangre Servicios Mutuos for

their availability and commitment during the curfew imposed by the quarantine period. The

Puerto Rico Science, Technology and Research Trust supported research reported in this

work under agreement number 2020-00272 to AME and CAS. The excellent support of Ms.

Andreica Maldonado and Grace Rendon from PRSTRT was instrumental in support of the

work described here.

References

1. Amanat F, Stadlbauer D, Strohmeier S, Nguyen THO, Chromikova V, McMahon M,

Jiang K, Arunkumar GA, Jurczyszak D, Polanco J, Bermudez-Gonzalez M, Kleiner

G, Aydillo T, Miorin L, Fierer DS, Lugo LA, Kojic EM, Stoever J, Liu STH,

Cunningham-Rundles C, Felgner PL, Moran T, García-Sastre A, Caplivski D, Cheng

AC, Kedzierska K, Vapalahti O, Hepojoki JM, Simon V, Krammer F. 2020. A

serological assay to detect SARS-CoV-2 seroconversion in humans. Nat Med

doi:10.1038/s41591-020-0913-5.

2. Al Kahlout RA, Nasrallah GK, Farag EA, Wang L, Lattwein E, Muller MA, El

Zowalaty ME, Al Romaihi HE, Graham BS, Al Thani AA, Yassine HM. 2019.

Comparative Serological Study for the Prevalence of Anti-MERS Coronavirus

Antibodies in High- and Low-Risk Groups in Qatar. J Immunol Res 2019:1386740.

3. Hsueh PR, Huang LM, Chen PJ, Kao CL, Yang PC. 2004. Chronological evolution

of IgM, IgA, IgG and neutralisation antibodies after infection with SARS-associated

coronavirus. Clin Microbiol Infect 10:1062-6.

4. Ko JH, Müller MA, Seok H, Park GE, Lee JY, Cho SY, Ha YE, Baek JY, Kim SH,

Kang JM, Kim YJ, Jo IJ, Chung CR, Hahn MJ, Drosten C, Kang CI, Chung DR,

Song JH, Kang ES, Peck KR. 2017. Serologic responses of 42 MERS-coronavirus-

infected patients according to the disease severity. Diagn Microbiol Infect Dis

89:106-111.

5. Mo HY, Xu J, Ren XL, Zeng GQ, Tan YX, Chen RC, Chan-Yeung M, Zhong NS.

2005. Evaluation by indirect immunofluorescent assay and enzyme linked

immunosorbent assay of the dynamic changes of serum antibody responses

against severe acute respiratory syndrome coronavirus. Chin Med J (Engl) 118:446-

50.

6.   Hashem AM, Al-Amri SS, Al-Subhi TL, Siddiq LA, Hassan AM, Alawi MM, Alhabbab

RY, Hindawi SI, Mohammed OB, Amor NS, Alagaili AN, Mirza AA, Azhar EI. 2019.

Development and validation of different indirect ELISAs for MERS-CoV serological

testing. J Immunol Methods 466:41-46.

7.   Stadlbauer D, Amanat F, Chromikova V, Jiang K, Strohmeier S, Arunkumar GA,

Tan J, Bhavsar D, Capuano C, Kirkpatrick E, Meade P, Brito RN, Teo C, McMahon

M, Simon V, Krammer F. 2020. SARS-CoV-2 Seroconversion in Humans: A

Detailed Protocol for a Serological Assay, Antigen Production, and Test Setup. Curr

Protoc Microbiol 57:e100.

8.   Krammer F, Simon V. 2020. Serology assays to manage COVID-19. Science

doi:10.1126/science.abc1227.

9.   Swets JA. 1988. Measuring the accuracy of diagnostic systems. Science 240:1285-

93.

10.   Thrusfield M. 1995. Veterinary epidemiology. London, United Kingdom, 2nd ed

Blackwell Science.

11.   Vabret N, Britton GJ, Gruber C, Hegde S, Kim J, Kuksin M, Levantovsky R, Malle L,

Moreira A, Park MD, Pia L, Risson E, Saffern M, Salomé B, Esai Selvan M, Spindler

MP, Tan J, van der Heide V, Gregory JK, Alexandropoulos K, Bhardwaj N, Brown

BD, Greenbaum B, Gümüş ZH, Homann D, Horowitz A, Kamphorst AO, Curotto de

Lafaille MA, Mehandru S, Merad M, Samstein RM. 2020. Immunology of COVID-19:

Current State of the Science. Immunity doi:10.1016/j.immuni.2020.05.002.

12.   Hajian-Tilaki K. 2013. Receiver Operating Characteristic (ROC) Curve Analysis for

Medical Diagnostic Test Evaluation. Caspian J Intern Med 4:627-35.

13.   Surujballi O, Mallory M. 2001. Competitive enzyme-linked immunosorbent assay for

detection of Leptospira interrogans serovar pomona antibodies in bovine sera. Clin

Diagn Lab Immunol 8:40-3.

14.   Arevalo-Rodriguez I, Buitrago-Garcia D, Simancas-Racines D, Zambrano-Achig P,

425   del Campo R, Ciapponi A, Sued O, Martinez-Garcia L, Rutjes A, Low N, Perez-

426   Molina JA, Zamora J. 2020. FALSE-NEGATIVE RESULTS OF INITIAL RT-PCR

427   ASSAYS FOR COVID-19: A SYSTEMATIC REVIEW. medRxiv

428   doi:10.1101/2020.04.16.20066787:2020.04.16.20066787.

429   15.   Chan JF, Yip CC, To KK, Tang TH, Wong SC, Leung KH, Fung AY, Ng AC, Zou Z,

430   Tsoi HW, Choi GK, Tam AR, Cheng VC, Chan KH, Tsang OT, Yuen KY. 2020.

431   Improved Molecular Diagnosis of COVID-19 by the Novel, Highly Sensitive and

432   Specific COVID-19-RdRp/Hel Real-Time Reverse Transcription-PCR Assay

433   Validated In Vitro and with Clinical Specimens. J Clin Microbiol 58.

434   16.   Loeffelholz MJ, Tang YW. 2020. Laboratory diagnosis of emerging human

435   coronavirus infections - the state of the art. Emerg Microbes Infect 9:747-756.

436   17.   Tang YW, Schmitz JE, Persing DH, Stratton CW. 2020. Laboratory Diagnosis of

437   COVID-19: Current Issues and Challenges. J Clin Microbiol 58.

438   18.   CDC. 2020. Interim Clinical Guidance for Management of Patients with Confirmed

439   Coronavirus Disease (COVID-19). Updated June 2.

440   19.   Byambasuren O, Cardona M, Bell K, Clark J, McLaws M-L, Glasziou P. 2020.

441   Estimating the extent of asymptomatic COVID-19 and its potential for community

442   transmission: systematic review and meta-analysis. medRxiv

443   doi:10.1101/2020.05.10.20097543:2020.05.10.20097543.

444   20.   Cohen AN, Kessel B. 2020. False positives in reverse transcription PCR testing for

445   SARS-CoV-2. medRxiv doi:10.1101/2020.04.26.20080911:2020.04.26.20080911.

446   21.   Jiang S, Hillyer C, Du L. 2020. Neutralizing Antibodies against SARS-CoV-2 and

447   Other Human Coronaviruses. Trends Immunol 41:355-359.

448   22.   Dingens AS, Crawford KH, Adler A, Steele SL, Lacombe K, Eguia R, Amanat F,

449   Walls AC, Wolf CR, Murphy M, Pettie D, Carter L, Qin X, King NP, Veesler D,

Krammer F, Chu HY, Englund JA, Bloom JD. 2020. Seroprevalence of SARS-CoV-

2 among children visiting a hospital during the initial Seattle outbreak. medRxiv

doi:10.1101/2020.05.26.20114124:2020.05.26.20114124.

23.   Okba NMA, Müller MA, Li W, Wang C, GeurtsvanKessel CH, Corman VM, Lamers

MM, Sikkema RS, de Bruin E, Chandler FD, Yazdanpanah Y, Le Hingrat Q,

Descamps D, Houhou-Fidouh N, Reusken C, Bosch BJ, Drosten C, Koopmans

MPG, Haagmans BL. 2020. Severe Acute Respiratory Syndrome Coronavirus 2-

Specific Antibody Responses in Coronavirus Disease 2019 Patients. Emerg Infect

Dis 26.

24.   Woloshin S, Patel N, Kesselheim AS. 2020. False Negative Tests for SARS-CoV-2

Infection — Challenges and Implications. New England Journal of Medicine

doi:10.1056/NEJMp2015897.

Figure 1: Distribution of absorbance values obtained with CovIgG-Assay. The CovIgG-Assay was optimized using as antigen recombinant Spike-S1-RBD from SARS-CoV-2. Absorbance values were distributed in form of frequency histograms to clearly visualize the separation between true positives (SARS-CoV-2 infected) (upper figure) and true negative population (non-SARS-CoV-2 infection), which include healthy subjects and subjects carrying other respiratory and viral infections collected prior pandemic (lower figure).

Figure 2: Receiver-operator characteristic (ROC) curve. The ROC curve was built for 132 sera from healthy subjects or subjects carrying other respiratory or other viral infections and 49 COVID-19 confirmed subjects. The area under the ROC curve (accuracy) was 0.985 and and the 95% Confident interval (CI) for AUC= 0.954-1.

Figure 3: Validation of use of Spike S1-RBD ELISA for detection of SARS-CoV-2 IgG antibodies. Black dots indicate samples from negative cohorts (non-CoV-2 infection) collected prior 2019 with no previous history of selected viral infections or respiratory allergies (n= 78) and samples that tested positive for Mycoplasma IgM, (n = 9), Influenza A or B (n=13), respiratory allergies (n=13), Zika virus (ZIKV, n=14), Chikungunya virus (CHIKV, n=3), Dengue virus (DENV, n=8) or RSV (n=6). Grey dots indicate samples from patients with confirmed SARS CoV-2 infection (n=49). Dotted horizontal line indicate CovIgG-Assay cut-point value (OD_{492}= 0.312). S1-RBD: Spike subunit-1-Receptor biding domaine. Each dot indicates mean OD of each sample tested in duplicate.

Y= 1.268*X - 2.036

Best-fit values	
Slope	1.268 ± 0.04190
Y-intercept when X=0.0	-2.036 ± 0.1323
X-intercept when Y=0.0	1.606
1/slope	0.7888
95% Confidence Intervals	
Slope	1.160 to 1.375
Y-intercept when X=0.0	-2.376 to -1.696
X-intercept when Y=0.0	1.455 to 1.736
Goodness of Fit	
R square	0.9946

Figure 4: Correlation between absorbance at 492nm (A_{492}) and antibody titter. A total of 40 sera from confirmed COVID-19 subjects that resulted positive by CovIgG-Assay were titrated at dilutions among 1:100 to 1:12,800. A lineal regression analysis was then done in which the mean A_{492} of sera with similar antibody titer were plotted with their corresponding A_{492} values. We found a lineal correlation (r2=0.9946) between the antibody titer (maximal dilution that render $A_{492} \geq 0.312$) and the individual A_{492} value at the working dilution (1:100). From this analysis the following lineal equation (Y= 1.268*X -2.036) was obtained, which was further used to estimate the antibody titer of all sera reported as seropositive by CovIgG-Assay.

Table-1. Specificity (Sp) and sensitivity (Se) of the CovIgG-Assay based on cut-points from the comparison between infected and uninfected populations.

Target Sp	Cut-point	Specificity	Sp Lower 95% CL	Sp Upper 95% CL	Sensitivity	Se Lower 95% CL	Se Upper 95% CL
0.999	0.745	1	0.961	1	0.980	0.893	0.996
0.995	0.745	1	0.961	1	0.980	0.893	0.996
0.990	0.745	1	0.961	1	0.980	0.893	0.996
0.980	**0.312**	**0.989**	**0.943**	**0.998**	**0.980**	**0.893**	**0.996**
0.950	0.202	0.958	0.897	0.984	0.980	0.893	0.996
0.900	0.129	0.905	0.830	0.949	0.980	0.893	0.996
0.800	0.099	0.800	0.709	0.868	0.980	0.893	0.996

Table-2. Agreement between the results of CovIgG-Assay and the PCR-based assay used as reference method for COVID-19 diagnosis.

		PCR-based assay*		
		Positive	**Negative**	**Total**
CovIgG-Assay	**Positive**	48	0	48
	Negative	1	132	133
	Total	49	132	181

Positive Predictive Value (PPV)=100% is calculated as the number of individuals with a positive result by CovIgG-Assay / the total of true positive individuals x 100. Negative Predictive Value (PNV)= 99.2% is calculated as the number of individuals reported as negative by CovIgG-Assay / the total of true negative individuals x 100.

*One sample was reported positive by Mt. Sinai Laboratory COVID-19 ELISA Antibody Test.

Table-3. Samples before and after treatment with DTT (Dithiothreitol) showing IgG class specificity for CovIgG-Assay. Each result represents OD at 492nm absorbance.

ID	IgG			IgG DTT			IgM			IgM DTT		
	OD1	OD2	Average OD	OD1	OD2	Average OD	OD1	OD2	Average OD	OD1	OD2	Average OD
45	2.7594	2.6820	2.7207	1.7638	1.7889	1.7763	1.6503	1.5807	1.6155	0.0244	0.0235	0.0239
121	3.1175	2.9996	3.0585	2.7779	2.6761	2.7270	3.5162	3.5980	3.5571	0.3303	0.3293	0.3298
122	2.7393	2.3747	2.5569	1.9655	1.9516	1.9585	2.6131	2.5388	2.5759	0.1060	0.1019	0.1039
146	2.7958	2.7973	2.7965	2.3467	2.3361	2.3414	1.0818	1.0508	1.0663	0.0898	0.0876	0.887
147	2.8958	2.7455	2.8206	2.4553	2.3915	2.4234	3.2084	3.2769	3.2426	0.8243	0.8349	0.8296
183	1.4743	1.4506	1.4624	0.8341	0.8457	0.8399	3.3603	3.3997	3.3800	0.0607	0.0519	0.0563

Sample 183 was confirmed by PCR and the others were confirmed by PCR and IgG/IgM rapid tests

Table-4. Experimental antibody titers of SARS-CoV-2 infected subjects in relation with their absorbance value at 492nm (1:100 dilution).

Antibody titer*	No. Individuals	Absorbance at 492nm (A_{492}) range	Mean A_{492} ± SD
100	6	0.321-0.538	0.427 ± 0.0983
200	10	0.687-1.111	0.927 ± 0.135
400	8	0.802-1.518	1.224 ± 0.238
800	--	--	--
1600	6	1.980-2.385	2.149 ± 0.169
3200	6	2.175-2.577	2.439 ± 0.165
6400	2	2.372-3.059	2.715 ± 0.343
12800	2	2.851-3.128	2.985 ± 0.138

*Antibody titer is defined as the maximal serum dilution that renders A_{492} greater than the optimized cut-point ($A_{492} \geq 0.312$) determined by ROC analysis.

Table 5. Comparison of three serological tests against CoVIgG-Assay for the detection of anti SARS CoV-2 antibodies

Sample ID	CoronaCheck[1] (Rapid COVID-19 IgG/IgM)	CoronaCheck[1] (Rapid COVID-19 IgG)	CoVIgG-Assay[2] Estimated titer	Abbott Architect SARS-CoV-2 IgG[3] (Index value)	Elecsys[4] Anti SARS CoV-2 Cobas®
32	+		>12800		
33	+		>12800		
34	+		1:649		
35	+		1:11040		
36	+		1:1963		
37	+		>12800		
38	+		1:8241		
39	+		>12800		
40	+		>12800		
46		+	-	.05	
47		+	†	.03	
48		+	-	.03	
49		+	1:254	.05	
50		+	-	.03	
51		+	-	.05	
52		+	-	.02	
53		+	-	.04	
54		+	1:254	.05	
55		+	-	.04	
56		+	-	.07	
57		+	-	.37	
58		+	†	.08	
59		+	†	.02	
60		+	-	.05	
61		+	-	.01	
62		+	1:219	.69	
63		+	1:220	.05	
105			1:3200‡		+
106			1:200‡		+
132			1:3200‡		+
133			1:400‡		+
134			1:200‡		-
154			1:200‡		+
148			1:1600‡		+
149			1:800‡		+
150			1:200‡		+
151			1:1600‡		+
152			1:400‡		+
153			1:6400‡		+
194			>1:12800		+
195			>1:12800		+
197			1:1064		+

220			>1:12800		+
221			>1:12800		+
222			1:3236		+

[‡]Titer determined by endpoint dilution from the average of duplicates OD492.
[1]Lateral flow rapid test intended to qualitatively detect antibodies to SARS-CoV-2. Presence of IgG/IgM antibodies are indicated by a visible line in the specific region on the device (On the manufacturer's website mention holding FDA EUA). Positive results may be due to cross reactivity with other Coronavirus strains (non CoV-2). Information regarding antigen used is not available.
[2]Quantitative ELISA for the detection SARS CoV-2 IgG antibodies using Spike S1-RBD (GenScript) as capture antigen. Samples with OD492<0.312 are considered negative. [†]IgG titration of positive samples by CovIgG-Assay range from 1:100 to 1:12,800.
[3]Chemiluminescent microparticle immunoassay for the qualitative detection of IgG antibodies against SARS CoV-2 (FDA EUA). Assay is designed to detect IgG antibodies to the nucleocapsid protein of SARS CoV-2. Samples with an index value <1.4 are considered negative. [4]Electrochemiluminescence immunoassay intended for qualitative detection of antibodies to SARS-CoV-2 in human serum and plasma using a recombinant protein representing the nucleocapsid antigen of SARS-CoV-2 (FDA EUA).

1 **SUPPLEMENTARY DATA**
2
3 **Supplementary Method-1**
4
5 **IgG Purification for positive and negative controls elaboration**
6
For the assay quality control; positive and negative controls were included for each
assay. These controls were prepared in-house from a convalescent subject with COVID-19.
(A) 20mL of the plasma sample was mixed 1:1 with phosphate buffer saline (PBS) and
loaded onto a 5/5 HiTrap rProtein-A column (GE Healthcare, USA) at a flow rate: of
1.5mL/min. The equilibration buffer used was PBS (pH 7.2), for elution we used 0.1M
Glycine-HCl (pH 2.85) and the neutralization buffer used was 1M Tris pH 8.5. **(B)** A total of
17 fractions were collected and absorbance (OD) at 280nm for each fraction was
measured. Fractions 9 to 14 were pooled in a total volume of 20ml and the pooled sample
had A_{280}=2.179. Pooled sample was desalted by PD-10 column and the desalted
fraction had A_{280}= 2.076). The final concentration for this fraction was 1.48mg/mL (2.076 /14)
x 10) in 28mL, for a total of 41.44 mg. **(C)** A total of 2µg-purified IgG was suspended in
loading SDS-buffer containing 5mM DTT and analyzed for purity by 4-20% sodium dodecyl
sulfate polyacrylamide electrophoresis (SDS-PAGE) and stained with Coomassie-blue. The
purified IgG fraction was used to prepare a high positive control (HPC) and a negative control
(NC) to be used in the CovIgG-Assay. For the HPC preparation IgG was diluted 30µg/mL
and for the NC IgG was diluted 380-fold to get a concentration equivalent to 0.078µgmL, both
controls were prepared in PBST containing 10% glycerol.
24

A

COVID-IgG

Column HiTrap rProtein-A FF 5/5
Loading buffer: PBS pH 7.2
Sample: 20-ml Plasma BB#1 + 20-ml PBS
Elution Buffer: 0.1M Glycine-HCL pH 2.85

B

Fraction	OD 280nm
1	2.438
2	2.563
3	2.438
4	2.341
5	0.353
6	0.135
7	0.149
8	0.13
9	2.126
10	2.251
11	2.184
12	2.184
13	1.16
14	0.132
15	0.087
16	0.078
17	0.036

C

25
26
27

Supplementary Table-1. Positive samples used CovIgG-Assay validation. Each result represent the absorbance measurement at 492nm.

Immunological status/Source	Numeric ID	Confirmatory Method	Sample Date	OD1	OD2	Average OD	Estimated Titer
PCR+/ IgG/IgM Positive Local Laboratory	45	LabCorp	4/12/20	2.034	2.251	2.291	1:2691
Blood Bank 1 PCR+	102	Roche or LabCorp or CDC	4/27/20	2.533	2.591	2.555	1:4487
	103	Roche or LabCorp or CDC	4/27/20	2.584	2.463	2.539	1:4355
	104	Roche or LabCorp or CDC	4/27/20	0.972	0.975	0.973	1:207
	105	Roche or LabCorp or CDC	4/27/20	2.422	2.487	2.447	1:3639
	106	Roche or LabCorp or CDC	4/27/20	0.797	0.876	0.826	1:156
Local Major Hospital PCR+	120	Cepheid	4/27/20	2.022	2.005	2.014	1:1570
Local Major Hospital PCR+ / IgG/IgM Positive	121	Cepheid	4/24/20	2.390	2.355	2.372	1:3148
Local Major Hospital PCR+/ IgG+/IgM-	143	Cepheid	4/28/20	2.885	2.818	2.851	1:7980
PCR+ / IgG/IgM Positive Local Laboratory	122	LabCorp	4/16/20	1.991	1.970	1.981	1:1472
Blood Bank 1 PCR+	132	Roche or LabCorp or CDC	4/30/20	2.153	2.198	2.176	1:2148
	133	Roche or LabCorp or CDC	4/30/20	1.163	0.998	1.081	1:256
	134	Roche or LabCorp or CDC	4/30/20	0.828	0.816	0.822	1:155
	135	Roche or LabCorp or CDC	4/30/20	1.953	2.009	1.981	1:1472
	136	Roche or LabCorp or CDC	4/30/20	0.048	0.043	0.045	N/A
	137	Roche or LabCorp or CDC	4/30/20	0.973	1.100	1.037	1:235
Blood Bank 2 PCR+	155	Quest PCR	5/11/20	1.429	1.314	1.371	1:450
	156	Quest PCR	5/11/20	1.793	1.724	1.759	1:955
	157	Quest PCR	5/11/20	2.464	2.501	2.483	1:3899
	158	Roche	5/11/20	3.080	3.334	3.207	>12800
	159	PR ASEM PCR	5/11/20	2.906	3.206	3.056	1:11,885
	160	Quest PCR	5/11/20	1.916	1.879	1.897	1:1,250
	161	Quest PCR	5/11/20	2.746	3.170	2.958	1:9,817
	162	Roche	5/11/20	2.264	1.981	2.122	1:1936
	163	Roche	5/11/20	1.415	1.460	1.437	1:512
	164	Roche	5/11/20	2.694	2.513	2.604	1:4932
	165	Roche	5/11/20	2.199	2.194	2.196	1:2239
	166	Roche	5/11/20	1.715	1.851	1.783	1:1002
	167	Roche	5/11/20	1.796	1.800	1.798	1:1032

	170	VA Orlando, FL PCR	5/11/20	3.128	3.152	3.140	>12800
	171	Roche	5/11/20	2.468	2.487	2.477	1:3864
	173	PR Dept. of Health PCR	5/11/20	3.185	3.223	3.204	>12800
	176	VA San Juan PCR	5/11/20	2.484	2.541	2.512	1:4130
	177	PR Dept. of Health PCR	5/11/20	2.103	2.296	2.200	1:2254
	178	Quest PCR	5/11/20	1.621	1.521	1.571	1:664
	181	Roche	5/11/20	1.748	1.682	1.715	1:879
	182	Quest PCR	5/11/20	2.857	2.955	2.906	1:8872
	183	COVID-19 Ab Assay - Titer 2880 Mount Sinai	5/11/20	1.758	1.663	1.711	1:871
	184	Quest PCR	5/11/20	1.889	1.616	1.752	1:944
	185	Roche	5/11/20	1.362	1.497	1.430	1:505
PCR+ Local Laboratory	146	LabCorp	5/11/20	2.707	2.756	2.732	1:6324
	147	LabCorp	5/11/20	2.573	2.700	2.636	1:5260
Blood Bank 1 PCR+	148	Roche or LabCorp or CDC	5/11/20	2.480	2.019	2.250	1:2483
	149	Roche or LabCorp or CDC	5/11/20	2.052	2.131	2.092	1:1824
	150	Roche or LabCorp or CDC	5/11/20	0.925	0.848	0.886	1:175
	151	Roche or LabCorp or CDC	5/11/20	2.153	2.220	2.187	1:2193
	152	Roche or LabCorp or CDC	5/11/20	1.316	1.073	1.195	1:319
	153	Roche or LabCorp or CDC	5/11/20	2.365	2.155	2.260	1:2529
	154	Roche or LabCorp or CDC	5/11/20	0.984	0.990	0.987	1:213

Estimated titer was calculated , using the lineal equation (Y= 1.185*X -1.773) derived from the lineal correlation between antibody titer and the magnitude of absorbance values.

32
33
34

Supplementary Table 2. Sera from subjects with non-SARS-CoV-2 infection used for CovIgG-Assay validation. Each result represent the absorbance measurement at 492nm.

Sample Source	Numeric ID	Immunological status	Sample Testing Date	OD1	OD2	Average OD
Virology Serum Bank	VB1	Healthy Subjects	5/23/95	0.028	0.027	0.027
	VB2		12/7/95	0.211	0.194	0.202
	VB3		2/10/97	0.082	0.087	0.085
	VB4		5/25/97	0.054	0.053	0.054
	VB5		2/17/06	0.041	0.054	0.047
	VB6		6/28/16	0.054	0.059	0.056
	VB7		4/10/17	0.092	0.106	0.099
	VB8		8/15/16	0.067	0.102	0.084
	VB9		8/15/16	0.310	0.225	0.268
	VB10		6/27/17	0.075	0.058	0.067
	VB11		1/10/17	0.023	0.000	0.011
	VB12		4/28/17	0.076	0.061	0.068
	VB13		6/26/18	0.330	0.294	0.312
	VB14		6/26/18	0.025	0.025	0.025
	VB15		6/28/18	0.036	0.030	0.033
	VB16		9/8/00	0.040	0.050	0.045
	VB17		5/8/19	0.026	0.033	0.029
	VB18		6/6/19	0.069	0.077	0.073
Immunology Bank	IB95	Healthy Subjects	2012	0.087	0.094	0.090
	IB96		2012	0.042	0.050	0.046
	IB97		2012	0.081	0.076	0.078
	IB100		2012	0.032	0.031	0.032
	IB144		2012	0.100	0.100	0.100
	IB145		2012	0.048	0.036	0.042
	IB146		2012	0.072	0.067	0.069
	IB147		2012	0.035	0.038	0.036
	IB148		2012	0.028	0.031	0.029
	IB149		2012	0.043	0.049	0.046
	IB135		2012	0.044	0.044	0.044
	IB136		2012	0.112	0.095	0.103
	IB137		2012	0.024	0.027	0.025
	IB138		2012	0.023	0.034	0.029
	IB139		2012	0.028	0.027	0.027
	IB140		2012	0.107	0.106	0.106
	IB141		2012	0.033	0.025	0.029

	IB142		2012	0.049	0.040	0.044
	IB143		2012	0.041	0.040	0.040
	IM		2012	0.032	0.029	0.030
	AO		2012	0.096	0.090	0.093
	OF		2012	0.075	0.069	0.072
	VB83		2012	0.064	0.060	0.062
	IB37		2012	0.069	0.078	0.074
	IB58		2012	0.055	0.059	0.057
	IB86		2012	0.016	0.015	0.015
	IB84		2012	0.014	0.016	0.015
Virology Serum Bank	RB	**ZIKV +**	2016	0.084	0.082	0.083
	FM		2016	0.019	0.024	0.021
	VB82		2016	0.083	0.088	0.085
	JN		8/11/16	0.076	0.054	0.065
	JR		8/15/16	0.157	0.157	0.157
	EXP		4/1/17	0.040	0.015	0.028
Immunology Bank	IB133	**Healthy Subjects**	2012	0.029	0.004	0.017
	IB130		2012	0.073	0.061	0.067
	IB131		2012	0.024	0.025	0.024
	IB132		2012	0.253	0.251	0.252
Immunology Bank	IB1	**Resp. Allergies**	2012	0.081	0.084	0.083
	IB2		2012	0.044	0.048	0.046
	IB3		2012	0.048	0.049	0.049
	IB4		2012	0.055	0.123	0.089
	IB5		2012	0.127	0.061	0.094
	IB6		2012	0.050	0.050	0.050
	IB7		2012	0.103	0.107	0.105
	IB8		2012	0.081	0.080	0.081
	IB9		2012	0.090	0.095	0.093
	IB10		2012	0.003	0.045	0.024
	IB11		2012	0.255	0.128	0.192
	IB12		2012	0.065	0.064	0.065
	IB13		2012	0.058	0.317	0.188
Immunology Bank	BVF		2012	0.052	0.059	0.056
	VA		2012	0.058	0.064	0.061
	JJF		2012	0.070	0.075	0.073
	IB53		2012	0.067	0.062	0.065
	IB54		2012	0.058	0.059	0.059
	IB55		2012	0.049	0.050	0.050

	IB56		2012	0.180	0.158	0.169
	IB64		2012	0.095	0.082	0.089
	IB65		2012	0.051	0.062	0.057
	IB66		2012	0.065	0.066	0.066
	IB74	**Healthy Subjects**	2012	0.062	0.063	0.063
	IB75		2012	0.113	0.112	0.113
	IB77		2012	0.063	0.062	0.063
	IB78		2012	0.132	0.146	0.139
	IB79		2012	0.137	0.121	0.129
	IB80		2012	0.108	0.103	0.106
	IB81		2012	0.089	0.093	0.091
	IB83		2012	0.121	0.118	0.120
	IB84		2012	0.106	0.102	0.104
	IB85		2012	0.059	0.055	0.057
	IB86		2012	0.066	0.067	0.067
	IB89		2012	0.064	0.062	0.063
	IB90		2012	0.100	0.139	0.120
	IB22		2012	0.066	0.063	0.065
	IB23		2012	0.068	0.068	0.068
	IB35		2012	0.063	0.061	0.062
	IB69		2012	0.069	0.068	0.069
	IB82		2012	0.059	0.060	0.060
CDC*	CDC233	**DENV+**	4/28/20	0.013	0.013	0.013
	CDC255	**DENV+**	4/28/20	0.010	0.009	0.010
	CDC269	**DENV+**	4/28/20	0.014	0.006	0.010
	CDC713	**DENV+**	4/28/20	0.052	0.050	0.051
	CDC736	**DENV+**	4/28/20	0.082	0.080	0.081
	CDC315	**ZIKV +**	4/28/20	0.058	0.052	0.055
	CDC324	**ZIKV +**	4/28/20	0.017	0.024	0.021
	CDC432	**ZIKV +**	4/28/20	0.011	0.012	0.011
	CDC493	**ZIKV +**	4/28/20	0.075	0.064	0.069
	CDC518	**ZIKV +**	4/28/20	0.055	0.062	0.059
	CDC101	**Influenza A+**	4/28/20	0.011	0.018	0.014
	CDC102	**Influenza A+**	4/28/20	0.023	0.023	0.023
	CDC103	**Influenza A+**	4/28/20	0.042	0.045	0.044
	CDC104	**Influenza A+**	4/28/20	0.015	0.020	0.017
	CDC105	**Influenza B+**	4/28/20	0.064	0.064	0.064
	CDC106	**Influenza B+**	4/28/20	0.257	0.255	0.256
	CDC107	**Influenza B+**	4/28/20	0.077	0.137	0.107

	CDC108	**Influenza B+**	4/28/20	0.035	0.036	0.035
	CDC109	**Influenza B+**	4/28/20	0.014	0.017	0.015
	CDC110	**Influenza B+**	4/28/20	0.080	0.022	0.051
	CDC111	**Influenza B+**	4/28/20	0.093	0.101	0.097
	CDC112	**Influenza B+**	4/28/20	0.058	0.063	0.060
	CDC113	**RSV**	4/28/20	0.058	0.070	0.064
	CDC114	**RSV**	4/28/20	0.036	0.034	0.035
	CDC115	**RSV**	4/28/20	0.050	0.050	0.050
	CDC116	**RSV**	4/28/20	0.022	0.025	0.024
	CDC117	**RSV**	4/28/20	0.098	0.101	0.100
	CDC118	**RSV**	4/28/20	0.036	0.037	0.037
Virology Serum Bank	1		4/26/17	0.061	0.069	0.065
	3	**FluA H1N1+**	6/26/18	0.039	0.041	0.040
	5	**DENV +**	10/11/17	0.092	0.078	0.085
	7	**DENV +**	8/17/14	0.024	0.029	0.026
	9	**DENV +**	6/24/16	0.060	0.062	0.061
	29	**ZIKV +**	6/29/16	0.027	0.022	0.024
	VB83	**ZIKV +**	2016	0.014	0.009	0.012
	VB84	**ZIKV +**	2016	0.012	0.016	0.014
Mycoplasma IgGM+	107	**Mycoplasma IgGM+**	4/28/20	0.0289	0.0309	0.0299
	123		4/28/20	0.0472	0.0456	0.0464
	124		4/28/20	0.0313	0.0331	0.0322
	125		4/28/20	0.1966	0.1743	0.1854
	126		4/28/20	0.0499	0.0554	0.0526
	127		4/28/20	0.0641	0.0804	0.0722
	128		4/28/20	0.0516	0.0536	0.0526
	129		4/28/20	0.0286	0.0402	0.0344
	130		4/28/20	0.0323	0.0315	0.0319
	131		4/28/20	0.0631	0.0738	0.0684
Virology Serum Bank	119	**Chikungunya +**	5/4/20	0.0067	0.0445	0.0256
	20		4/15/20	0.1159	0.0982	0.1070
	118		5/4/20	0.0184	0.0281	0.0232

* All these samples were collected prior to December 2019 in Puerto Rico. The data showed is the date in the panel of those samples was assembled at CDC Dengue Branch in support this study.
35

36 **Supplementary Table-3. Reproducibility of CovIgG-Assay**

Sample	N	Mean A_{492}	Repeatability (Within-Run)		Within-Laboratory [a]	
			SD	% CV	SD	% CV
NC	30	0.022	0.021	N/A[b]	0.0026	N/A[b]
HPC	30	2.476	0.211	8.52	0.219	8.84
NS-1	6	0.049	0.011	N/A[b]	0.014	N/A[b]
NS-2	6	0.043	0.016	N/A[b]	0.032	N/A[b]
NS-3	6	0.042	0.024	N/A[b]	0.035	N/A[b]
NS-4	6	0.062	0.005	N/A[b]	0.006	N/A[b]
PS-1	6	2.085	0.011	0.527	0.075	3.59
PS-2	6	2.37	0.05	2.109	0.012	0.506
PS-3	6	2.235	0.015	0.671	0.15	6.71
PS-4	6	3.17	0.057	1.79	0.28	8.83

38 [a] Includes repeatability (Within-run), between-run and between-day variability
39 [b] Not applicable
40 HPC: High positive control, NS: Negative serum, PC: Positive serum
41 NC: Negative control
42
43

Supplementary Table-4. Samples titrated for establishing lineal correlation between OD (1:100) and antibody titer.

Sample No.	Numeric ID	Test Date	Dilution	OD1	OD2	Odx	Titer
1	34	4/28/20	**100**	**0.4869**	**0.5379**	**0.512**	1:100
			200	0.2833	0.2530	0.268	
			400	0.1394	0.1448	0.142	
			800	0.0960	0.0914	0.094	
			1600	0.0631	0.0628	0.063	
			3200	0.0529	0.0526	0.053	
			6400	0.0481	0.0496	0.049	
			12800	0.0453	0.0448	0.045	
2	47	6/26/20	**100**	**0.3188**	**0.3332**	**0.326**	1:100
			200	0.1033	0.0950	0.099	
			400	0.0290	0.0295	0.029	
			800	0.0128	0.0109	0.012	
			1600				
			3200				
			6400				
			12800				
3	58	6/26/20	**100**	**0.5478**	**0.5276**	**0.538**	1:100
			200	0.2545	0.2686	0.262	
			400	0.1369	0.1376	0.137	
			800	0.0612	0.0705	0.066	
			1600				
			3200				
			6400				
			12800				
4	201	6/26/20	**100**	**0.3051**	**0.3363**	**0.321**	1:100
			200	0.1579	0.1481	0.153	
			400	0.0757	0.0731	0.074	
			800	0.0255	0.0269	0.026	
			1600				
			3200				
			6400				
			12800				
5	216	6/26/20	**100**	**0.3333**	**0.3473**	**0.340**	1:100
			200	0.1503	0.1605	0.155	
			400	0.0795	0.0819	0.081	
			800	0.0455	0.0474	0.046	
			1600				
			3200				
			6400				
			12800				
			100	**0.5076**	**0.5395**	**0.524**	
			200	0.2410	0.2394	0.240	

6	172	6/26/20	400	0.1042	0.1080	0.106	1:100
			800	0.0514	0.0473	0.049	
			1600				
			3200				
			6400				
			12800				
7	104	4/28/20	100	0.9717	0.9752	0.973	1:200
			200	**0.5333**	**0.5404**	**0.537**	
			400	0.2533	0.2778	0.266	
			800	0.1348	0.1434	0.139	
			1600	0.0486	0.0608	0.055	
			3200	0.0306	0.0301	0.030	
			6400	0.0150	0.0176	0.016	
			12800	0.0060	0.0104	0.008	
8	106	4/28/20	100	0.7967	0.8761	0.836	1:200
			200	**0.4181**	**0.4621**	**0.440**	
			400	0.1768	0.2351	0.206	
			800	0.0998	0.1105	0.105	
			1600	0.0460	0.0363	0.041	
			3200	0.0271	0.0292	0.028	
			6400	0.0114	0.0140	0.013	
			12800	0.0071	0.0115	0.009	
9	54	4/28/20	100	0.9500	1.1300	1.040	1:200
			200	**0.6640**	**0.6885**	**0.676**	
			400	0.2636	0.3388	0.301	
			800	0.1524	0.1567	0.155	
			1600	0.0846	0.0782	0.081	
			3200	0.0638	0.0637	0.064	
			6400	0.0504	0.0525	0.051	
			12800	0.0505	0.0449	0.048	
10	134	5/2/20	100	0.9500	1.1300	1.040	1:200
			200	**0.6640**	**0.6885**	**0.676**	
			400	0.2636	0.3388	0.301	
			800	0.1524	0.1567	0.155	
			1600	0.0846	0.0782	0.081	
			3200	0.0638	0.0637	0.064	
			6400	0.0504	0.0525	0.051	
			12800	0.0505	0.0449	0.048	
11	137	5/2/20	100	0.9732	1.0999	1.037	1:200
			200	**0.6146**	**0.5656**	**0.590**	
			400	0.2925	0.2601	0.276	
			800	0.1270	0.1434	0.135	
			1600	0.0675	0.0601	0.064	
			3200	0.0264	0.0289	0.028	
			6400	0.0168	0.0126	0.015	
			12800	0.0076	0.0075	0.008	
			100	0.9245	0.8482	0.886	
			200	0.4733	0.4452	0.459	
			400	0.2409	0.2095	0.225	

12	150	5/12/20	800	0.1094	0.0997	0.105	1:200
			1600	0.0532	0.0460	0.050	
			3200	0.0207	0.0185	0.020	
			6400	0.0085	0.0089	0.009	
			12800	0.0079	-0.0001	0.004	
13	154	5/12/20	100	0.9835	0.9899	0.987	1:200
			200	**0.5533**	**0.5347**	**0.544**	
			400	0.2375	0.2643	0.251	
			800	0.1069	0.1334	0.120	
			1600	0.0507	0.0538	0.052	
			3200	0.0234	0.0252	0.024	
			6400	0.0109	0.0164	0.014	
			12800	0.0071	0.0044	0.006	
14	216	6/24/20	100	0.6982	0.6759	0.687	1:200
			200	**0.3516**	**0.3389**	**0.345**	
			400	0.1579	0.1615	0.160	
			800	0.0650	0.0718	0.068	
			1600	0.0301	0.0303	0.030	
			3200	0.0073	0.0117	0.010	
			6400	-0.0029	0.0011	-0.001	
			12800	0.0039	0.0003	0.002	
15	49	6/26/20	100	1.1191	1.1038	1.111	1:200
			200	**0.5989**	**0.6113**	**0.605**	
			400	0.3058	0.3080	0.307	
			800	0.1876	0.1522	0.170	
			1600				
			3200				
			6400				
			12800				
16	59	6/26/20	100	1.0007	1.0650	1.033	1:200
			200	**0.4657**	**0.5784**	**0.522**	
			400	0.2586	0.2404	0.249	
			800	0.1159	0.1044	0.110	
			1600				
			3200				
			6400				
			12800				
17	36	4/28/20	100	1.2787	1.0590	1.169	1:400
			200	0.7156	0.6355	0.676	
			400	**0.3584**	**0.3298**	**0.344**	
			800	0.1805	0.1539	0.167	
			1600	0.0942	0.0690	0.082	
			3200	0.0631	0.0562	0.060	
			6400	0.0505	0.0469	0.049	
			12800	0.0470	0.0452	0.046	
18	133	5/2/20	100	1.1632	0.9980	1.081	1:400
			200	0.7341	0.7633	0.749	
			400	**0.3873**	**0.4138**	**0.401**	
			800	0.1791	0.2104	0.195	

			1600	0.0881	0.1002	0.094	
			3200	0.0371	0.0465	0.042	
			6400	0.0213	0.0236	0.022	
			12800	0.0119	0.0145	0.013	
19	149	5/15/20	100	2.0521	2.1313	2.092	1:800
			200	1.2680	1.2933	1.281	
			400	0.7667	0.7253	0.746	
			800	**0.3935**	**0.3880**	**0.391**	
			1600	0.2208	0.1876	0.204	
			3200	0.0976	0.0972	0.097	
			6400	0.0451	0.0396	0.042	
			12800	0.0220	0.0144	0.018	
20	152	5/12/20	100	1.3163	1.0732	1.195	1:400
			200	0.7488	0.7716	0.760	
			400	**0.3602**	**0.3523**	**0.356**	
			800	0.1927	0.1857	0.189	
			1600	0.0999	0.0860	0.093	
			3200	0.0404	0.0424	0.041	
			6400	0.0209	0.0195	0.020	
			12800	0.0132	0.0124	0.013	
21	163	6/24/20	100	1.4934	1.5267	1.510	1:400
			200	0.7820	0.8445	0.813	
			400	**0.3944**	**0.3925**	**0.393**	
			800	0.1904	0.1869	0.189	
			1600	0.0931	0.0841	0.089	
			3200	0.0402	0.0415	0.041	
			6400	0.0186	0.0174	0.018	
			12800	0.0114	0.0124	0.012	
22	62	6/26/20	100	1.1285	0.9897	1.059	1:400
			200	0.6192	0.6303	0.625	
			400	**0.3311**	**0.3188**	**0.325**	
			800	0.1542	0.1757	0.165	
			1600				
			3200				
			6400				
			12800				
23	63	6/26/20	100	1.5425	1.4930	1.518	1:400
			200	0.8156	0.8586	0.837	
			400	**0.4495**	**0.4619**	**0.456**	
			800	0.2100	0.2269	0.218	
			1600				
			3200				
			6400				
			12800				
24	42	6/26/20	100	0.8688	0.7354	0.802	1:200
			200	**0.4387**	**0.4302**	**0.434**	
			400	0.2134	0.2007	0.207	
			800	0.0999	0.1000	0.100	
			1600				

			3200				
			6400				
			12800				
25	120	5/2/20	100	2.0222	2.0052	2.014	1:1,600
			200	1.8618	1.6005	1.731	
			400	1.4337	1.5597	1.497	
			800	1.0007	1.0090	1.005	
			1600	**0.5712**	**0.5883**	**0.580**	
			3200	0.2744	0.2653	0.270	
			6400	0.1360	0.1439	0.140	
			12800	0.0781	0.0803	0.079	
26	122	5/2/20	100	1.9911	1.9703	1.981	1:1,600
			200	1.8679	1.7005	1.784	
			400	1.3115	1.2880	1.300	
			800	0.8457	0.9002	0.873	
			1600	**0.4816**	**0.4654**	**0.474**	
			3200	0.2219	0.2303	0.226	
			6400	0.1239	0.1290	0.126	
			12800	0.0627	0.0660	0.064	
27	135	5/2/20	100	1.9532	2.0085	1.981	1:1,600
			200	1.6756	1.6682	1.672	
			400	1.1874	1.1280	1.158	
			800	0.7737	0.7315	0.753	
			1600	**0.3755**	**0.3583**	**0.367**	
			3200	0.1864	0.1863	0.186	
			6400	0.0877	0.0890	0.088	
			12800	0.0511	0.0466	0.049	
28	38	5/6/20	100	2.3136	2.4567	2.385	1:1,600
			200	1.6694	1.7264	1.698	
			400	1.1548	1.1688	1.162	
			800	0.6065	0.7014	0.654	
			1600	**0.3298**	**0.3129**	**0.321**	
			3200	0.1599	0.1609	0.160	
			6400	0.0730	0.0802	0.077	
			12800	0.0398	0.0390	0.039	
29	39	5/7/20	100	2.3385	2.3616	2.350	1:1,600
			200	2.0048	1.9391	1.972	
			400	1.2542	1.3920	1.323	
			800	0.9147	0.9688	0.942	
			1600	**0.4241**	**0.4908**	**0.457**	
			3200	0.2482	0.2410	0.245	
			6400	0.0588	0.1070	0.083	
			12800	0.0560	0.0473	0.052	
30	151	5/12/20	100	2.1525	2.2204	2.186	1:1,600
			200	1.7488	1.7520	1.750	
			400	1.2940	1.2440	1.269	
			800	0.7536	0.7430	0.748	
			1600	**0.3648**	**0.4260**	**0.395**	
			3200	0.2118	0.2133	0.213	

			6400	0.1127	0.1025	0.108	
			12800	0.0622	0.0588	0.061	
31	102	4/28/20	100	2.5331	2.5908	2.562	1:3,200
			200	2.2574	2.3355	2.296	
			400	1.8572	1.8475	1.852	
			800	1.3643	1.2219	1.293	
			1600	0.7555	0.7154	0.735	
			3200	**0.3826**	**0.3698**	**0.376**	
			6400	0.2024	0.1752	0.189	
			12800	0.1045	0.1046	0.105	
32	105	4/28/20	100	2.4223	2.4874	2.455	1:3,200
			200	2.2955	2.3155	2.305	
			400	1.8623	1.8930	1.878	
			800	1.3086	1.2429	1.276	
			1600	0.7716	0.7072	0.739	
			3200	**0.3724**	**0.3818**	**0.377**	
			6400	0.1674	0.2172	0.192	
			12800	0.1049	0.0913	0.098	
33	132	5/2/20	100	2.1529	2.1981	2.176	1:3,200
			200	2.1780	2.3296	2.254	
			400	1.8968	1.4301	1.663	
			800	1.4837	1.5024	1.493	
			1600	0.8612	0.7826	0.822	
			3200	**0.5374**	**0.4550**	**0.496**	
			6400	0.2470	0.2643	0.256	
			12800	0.1507	0.1283	0.140	
34	146	5/15/20	100	2.7072	2.7559	2.732	1:1,600
			200	1.9902	2.0991	2.045	
			400	1.5837	1.4940	1.539	
			800	1.1582	1.0425	1.100	
			1600	**0.6085**	**0.5643**	**0.586**	
			3200	0.3054	0.3015	0.303	
			6400	0.1436	0.1585	0.151	
			12800	0.0849	0.0918	0.088	
35	147	5/12/20	100	2.5725	2.7002	2.636	1:3,200
			200	2.1213	2.4363	2.279	
			400	1.9729	2.1135	2.043	
			800	1.4755	1.5723	1.524	
			1600	1.0330	0.9994	1.016	
			3200	**0.5196**	**0.5754**	**0.547**	
			6400	0.2717	0.2903	0.281	
			12800	0.1414	0.1443	0.143	
36	153	5/12/20	100	2.3651	2.1547	2.260	1:3,200
			200	2.0569	2.1836	2.120	
			400	1.4118	1.4230	1.417	
			800	1.0265	0.8761	0.951	
			1600	0.6102	0.5617	0.586	
			3200	**0.3374**	**0.3209**	**0.329**	
			6400	0.1781	0.1888	0.183	

			12800	0.0959	0.0925	0.094	
37	40	4/28/20	100	3.0662	3.0520	3.059	1:6,400
			200	2.5939	2.5867	2.590	
			400	2.3322	2.0433	2.188	
			800	1.8097	1.7007	1.755	
			1600	0.8596	0.9741	0.917	
			3200	0.6036	0.6416	0.623	
			6400	**0.3190**	**0.3377**	**0.328**	
			12800	0.2060	0.2043	0.205	
38	121	5/2/20	100	2.3896	2.3548	2.372	1:6,400
			200	2.3850	2.3421	2.364	
			400	2.1573	2.1270	2.142	
			800	1.7675	1.9848	1.876	
			1600	1.4231	1.3634	1.393	
			3200	0.8557	0.8547	0.855	
			6400	**0.4810**	**0.4662**	**0.474**	
			12800	0.2406	0.2418	0.241	
39	33	5/6/20	100	3.2099	3.0455	3.128	>1:12,800
			200	3.0308	2.9626	2.997	
			400	2.8015	2.6544	2.728	
			800	2.4501	2.4730	2.462	
			1600	1.6862	1.7521	1.719	
			3200	1.0987	1.1226	1.111	
			6400	0.5937	0.6122	0.603	
			12800	**0.3338**	**0.3550**	**0.344**	
40	143	5/7/20	100	2.8849	2.8177	2.851	>1:12,800
			200	2.8338	2.7436	2.789	
			400	2.4554	2.5898	2.523	
			800	2.3795	2.3766	2.378	
			1600	2.0173	1.9755	1.996	
			3200	1.4074	1.4552	1.431	
			6400	0.9191	0.9061	0.913	
			12800	**0.5609**	**0.5505**	**0.556**	