Development of RNA-based assay for rapid detection of SARS-CoV-2 in clinical samples

[1#]Vinod Kumar, [1#]Suman Mishra, [1#]Rajni Sharma, [2]Jyotsna Agarwal, [3]Ujjala Ghoshal, [4]Tripti Khanna, [1]Lokendra K. Sharma, [1]Santosh Kumar Verma and [1*]Swasti Tiwari

[1]Department of Molecular Medicine & Biotechnology, SGPGIMS, Raibareli Road, Lucknow - 226014, India; [2]Department of Microbiology, Dr. Ram Manohar Lohia Institute of Medical Sciences, Lucknow, 226010; [3]Department of Microbiology, Virology and COVID19 Lab, SGPGIMS, Raibareli Road, Lucknow -226014, India; [4]Indian Council of Medical Research, Ramalingaswami Bhawan, New Delhi, India

Equal authorship

*Corresponding author

Prof. Swasti Tiwari**,**

Department of Molecular Medicine & Biotechnology

4th Floor PMSSY Building

Sanjay Gandhi Post Graduate Institute of Medical Science (SGPGIMS)

Raebareli Rd, Lucknow, Uttar Pradesh 226014, India

Email Address: tiwaris@sgpgi.ac.in

Disclosure: A patent application has been filed (ref#202011018132) on subject matter described in the publication.

Abstract:

The ongoing spread of pandemic coronavirus disease (COVID-19) is caused by Severe Acute

Respiratory Syndrome coronavirus 2 (SARS-CoV-2). In the lack of specific drugs or vaccines for

SARS-CoV-2, it demands rapid diagnosis and management for controlling the outbreak in the

community. Here we report the development of the first rapid-colorimetric assay capable of

detecting SARS-CoV-2 in the human nasopharyngeal RNA sample in less than 30 minutes. We

utilized a nanomaterial-based optical sensing platform to detect RNA-dependent RNA

polymerase (*RdRp*) gene of SARS-CoV-2, where the formation of oligo probe-target hybrid led

to salt-induced aggregation and changes in gold-colloid color from pink to blue in visible range.

Accordingly, we found a change in colloid color from pink to blue in assay containing

nasopharyngeal RNA sample from the subject with clinically diagnosed COVID-19 (n=7). The

colloid retained pink color when the test includes samples from COVID-19 negative subjects

(n=6) or human papillomavirus (HPV) infected women (n=2). The optimized method has

detection limit as little as 0.5 ng of SARS-CoV-2 RNA. Overall, the developed assay rapidly

detects SARS-CoV-2 RNA in clinical samples in a cost-effective manner and would be useful in

pandemic management by facilitating mass screening.

Key words

COVID-19, diagnosis, Ribonucleic acid, Coronavirus disease, colorimetric test

Introduction

Coronavirus disease is rapidly spreading across the world and raising severe global health concerns. In December 2019, China reported the first disease case in its Hubei Province. Based on the phylogenetic analysis, the identified novel coronavirus is named as Severe Acute Respiratory Syndrome Coronavirus 2 (SARS-CoV-2), and the disease spread by SARS-CoV-2 is known as "COVID-19", declared as a pandemic by World Health Organization (WHO).(1) Despite global massive efforts to control the outbreak of COVID-19, this pandemic is still on the rise. To date, lack of approved medicine or vaccine impede escalated the management of the COVID-19 epidemic. In the absence of an effective treatment strategy, developing affordable screening for rapid diagnosis is critically required in the management of COVID-19.(2) According to the WHO, the immediate priority is the development of point-of-care tests for the detection of SARS-CoV-2   at an early stage with improved sensitivity.(3)

Currently, the COVID-19 diagnostic test falls into two categories: antibody and nucleic acid-based detection systems. The Developed immunoassay is rapid but inefficient for the detection of the pathogen at an early stage of infection.(2) Besides, WHO does not currently recommend the use of antigen-based rapid diagnostic tests for patient care, but encourages the related research to improve their performance and potential diagnostic utility.(4)

Among the nucleic acid-based detection systems, the WHO considered a Real-time polymerase chain reaction (RT-PCR) based method as a gold standard for COVID-19 testing.(4) The sample-to-result time of the quantitative RT-PCR (qRT-PCR) was initially >4 hrs; however, constant efforts are underway to improve the turn-around time through automation. Besides, time-consuming process PCR based tests are expensive; they require sophisticated instruments and expertise.(2) Thus, other technologies, such as reverse transcription- loop-mediated isothermal amplification (RT-LAMP) (5) and Genome-editing (6), are being explored. These are

promising technologies; however, the expected turn-around time would still be around 1 hour

and may not be economical for mass screening, especially for the developing countries.

Nanomaterials based sensing platforms hold promise to develop rapid disease diagnostics.

However, their limited use in the clinical setting is due to the need for sophisticated

equipment.(7, 8) A recent attempt to use antisense-nucleotide capped gold nanoparticles for N-

gene based, COVID-19 detection could be a game changer.(9) Whereas, WHO suggested that

N-gene has a relatively weak analytical capability, compared to the *RdRp* gene, to detect

COVID-19 infection.(10) Moreover, the authors have demonstrated COVID-19 detection in a

cellular-model system, and application in human samples has yet to be shown.

In this study, we report for the first time gold nanoparticles (AuNPs) based rapid colorimetric

assay for visual eye detection of COVID-19 RNA in human samples designed to target *RdRp*

specific gene target in very cost-effective manner with wide application of mass screening in

field. We utilized the surface plasmon resonance property of AuNPs to detect unamplified

COVID-19 RNA in human samples. The ability of AuNPs to preferentially adsorb ssRNA/ssDNA

over dsDNA/dsRNA is the crucial concept of this assay. The single and double-stranded

oligonucleotides have different electrostatic properties, which provide stability (via retaining

natural color) and aggregation (which causes color change) of AuNPs in solution,

respectively.(11)

Materials and Methods

Chemicals: Citrate buffer stabilized AuNPs (10 nm diameter) was purchased from Alfa Aesar

(Thermo Fisher Scientific India Private Limited). Phosphate Buffer Saline (PBS pH 7.0) and

NaCl were procured from Sigma Aldrich.

Clinical samples: Nasopharyngeal RNA sample from COVID- negative (n=6) and positive

subjects (n=7) were used. The COVID-19 testing of these subjects were done in the Indian

Council of Medical Research (ICMR, India) approved diagnostic laboratory using Taqman-

based RT-PCR kit (Labgun, lab, Genomics.co. Ltd, Republic of Korea). The study protocol to

use clinical samples was approved by Institutional Human Ethics Committee SGPGIMS,

Lucknow (Ref N. PGI/BE/327/2020). The cervical DNA samples from human papilloma virus

(HPV) infected woman (n=2) was used as non-specific target control. The clinical Samples

(cervical smear) for the HPV DNA test were processed using HPV Test Hybrid Capture® 2

protocol (QIAGEN). Samples with relative light units (RLU) /Cutoff Value ratios > 1.0 were

considered as HPV positive and < 1.0 were considered as HPV negative. Only HPV positive

DNA samples were used during validation experiments.

In Vitro transcription (IVT) of SARS-CoV-2 RNA: The RNA (5ng) from confirmed COVID-19

positive human sample were first reverse transcribed with modified oligo dT primer having a T7

promoter sequence at its 5' end, and the resulting single stranded cDNA was further *In-vitro*

transcribed (IVT) using HiScribe T7 Quick High Yield RNA Synthesis Kit according to the

manufacturer's protocol (New England BioLabs Inc, NEB #E2050). The amplified RNA was than

purified using Monarch RNA Cleanup Kit (New England BioLabs Inc.) and quantified by RNA

HS reagent using Qubit system (Thermo Fisher Scientific).

Colorimetric assay for the detection of SARS-CoV-2 RNA: For colorimetric assay, reaction

was set up in 10 µL of reaction volume in sterile PCR tubes, containing **(a)** hybridization buffer

containing 80mM NaCl), **(b)** 0.5 µM oligo probe for *RdRp* genes, and **(c)** target RNA from

positive patients or IVT RNA. The *RdRp* oligo probe (5'-GTGATATGGTCATGTGTGGCGG-3')

was used to specifically detect the presence of SARS-CoV-2 RNA in the assay. In parallel, to

measure the specificity of the assay, input genomes from different source such nasopharyngeal

RNA from COVID negative subjects and HPV DNA from cervical cancer positive samples were

used as negative controls. Similarly, non-template control (NTC) was also included to measure

the background reactivity. In addition, to confirm the working principle of the assay, a different

RNA template (isolated from pancreas) and pancreas specific *REG-3* (Regenerating islet-

derived protein 3) gene oligo probe (5'- GTGCCTATGGCTCCTATTGCT-3') were used

separately.

The final reaction mixture with above combinations was then denatured at 95°C for 30

seconds, annealed at 60°C for 60 seconds and then cooled to room temperature for 10 minutes.

Subsequently, 10 nM colloidal AuNPs (~10 nm) were added to the assay mixture and allowed to

develop color for 1-2 minute.

***Spectral studies and measurement of sensitivity*:** Absorption spectrum of the assay mixture

was recorded in the range of 300 -700 nm. The peak shift from 520 nm (known as red-shift) and

peak broadening after 520 nm were measured as a characteristic feature of the salt induced

aggregation. Using various combinations of positive and negative controls (as discussed in the

previous section) the specificity of reaction and aggregation were compared. Assay sensitivity

was determined by serially diluting the input SARS-CoV-2 RNA from both the IVT synthesized

and, synthetic SARS-CoV-2 control (nCov19 control kit by Applied Biosystems) ranging from 5-

0.1ng and 1-0.1ng concentrations respectively.

Result and discussion

In this study, we report for the first time the development of a rapid and affordable RNA-based

assay for the visual detection of the SARS-CoV-2 genome in human samples. In this assay, we

use surface plasmon resonance property of gold nanoparticles/colloids (AuNP) and targeted the

RdRp specific gene sequence of SARS-CoV-2. *RdRp* is essential for viral replication and has

higher analytical power than E (envelope protein) and N (nucleocapsid protein) genes of SARS-

CoV-2. (10) our current established assay, salt-induced aggregation and color change of the

gold colloids occurs after *RdRp* oligo probe hybridizes with its specific target RNA of SARS-

CoV-2. With the current escalated demand of cost effective, easy and sensitive diagnostic for

COVID-19, the test was developed using commercially available nCoV19 synthetic DNA and

validated further using clinical samples from COVID-19 subjects (as confirmed using Taqman

based RT-PCR method, Table S1). In our study, we demonstrated a visual change in gold

colloid color from pink to blue when RNA samples from subjects with clinically diagnosed

COVID-19 infection hybridize with *RdRp* oligo probe (n=7). Simultaneously, the color remained

pink in SARS-CoV-2 negative samples due to the absence of hybridization (n=6).

Figure 1 illustrates the sequential schematics of process of the developed test. The color of the

gold colloid solution is dependent on the aggregation property of AuNPs in suspension.(12, 13)

The aggregation-induced color change can be visually monitored (by the naked eye) and

quantitated through absorption spectroscopy.   Generally, in aqueous solution, gold colloids

remain stabilized by the coating of negatively charged citrate ions (14) and have visible

appearance of pink color (Figure 1). In solution, individual particle exhibits a surface plasmon

resonance peak (λmax) at 520 nm (Figure 2a, Green curve). The oligonucleotide probe

preferentially adsorbs on AuNPs and provides additional stability due to the addition of negative

charges.(15) The same phenomenon was observed in our experiments when *RdRp* oligo probe

adsorbed and protected the salt-induced aggregation of colloid in the absence of target RNA

(NTC) (Figure 2a, Grey curve). Except for the reduced intensity of the absorption spectrum,

which was due to dilution of the colloid (Figure 2a, Grey curve), NTC and AuNP assays showed

no change? A pure colloid is pink in color (Inset of Figure 2a left vial), which turns into light pink

when it reacts with hybridization buffer containing oligo probe in equal volume (without target)

(Inset of Figure 2a right vial). Unlike dsDNA, inherent structural flexibility of ssDNA/RNA to

partially uncoil its bases, exposing them to AuNPs, generates the attractive electrostatic forces

causing them to allow over colloids and giving protection against electrostatic interaction

causing salt-induced aggregation.(15)

In this assay, SARS-CoV-2 RNA from human patients or IVT synthesized RNA was added into

hybridization buffer (containing oligo probe), followed by denaturation and annealing at 95°C

(30s) and 60°C (60s), respectively. After cooling at room temperature, the gold colloid was

added into the above reaction mixture. The colloid color changes in visible range from pink to

blue, indicate the formation of hybridized product (Figure 2b-c, and Figure 3). Broadening of the

peak with red-shift (~30 nm) was observed in the spectrum of aggregated colloids than non-

aggregated, confirms the success of developed assay for detection of an unamplified target with

unmodified colloids in a quick and facile way. The principle of binding oligo probe to its specific

target leading to change in color of the solution was independently verified using a different

template RNA (isolated from pancreas tissue) and pancreas specific gene *REG3* oligo probe in

a separate assay. This assay also resulted in a similar change in color and absorption spectra

as optimized earlier for SARS-CoV-2 RNA and *RdRp* oligo probe. It established the working

principle and specificity of the test (Figure S1).

We determined the cross-reactivity using a cervical-DNA sample from women diagnosed with

HPV infection (non-specific target control). No color change of gold colloids was observed with

HPV DNA, indicating no hybridization, and specificity of the developed assay (Figure 3 b, right

vial). Contrary to HPV DNA-negative control and NTC, a positive control sample shows

development of blue color (Figure 3 b, middle vial). Absorption spectrum (of colloids) with HPV

DNA-negative control exhibited characteristics similar to that of NTC, and no red-shift or peak

broadening as found with positive samples (Figure 3a). Cross-reactivity of the developed assay

with other respiratory viruses is warranted. However, we do not anticipate the same as the test

utilizes the detection of *the RdRp* gene of the SARS-CoV-2 virus. The oligo probe sequence

used in our assay is not complementary to any human mRNAs and other members of the SARS

family, as verified by BLAST using the NCBI database. (10)

IVT synthesized SARS-CoV-2 RNA was used to test the sensitivity of the developed assay.

RNA ranging from 0.1 to 5ng resulted in a gradual change in colloid color from light pink to blue

(Figure 4b). The color difference at 0.1 ng, compared to NTC, is barely visible with naked eyes

(Figure 4 b extreme left vial). However, the absorption spectra of colloids show a clear red-shift

with peak broadening up till 0.5 ng target RNA (Figure 4a).

We determined the assay sensitivity using different dilutions (1-0.1ng) of PCR-amplified

synthetic DNA (positive control, with nCoV19 control kit). Similar to IVT synthesized RNA, a

decreasing amount of DNA (from 1 to 0.1 ng) show a gradual change in colloid color from blue

to light pink (left to right, Fig. 4d). A clear red-shift with peak broadening reflects in the

absorption spectra of colloids recorded with control DNA, compared to NTC (Fig. 4c).

Accordingly, positive control showed a clear visual demarcation up to 0.5 ng amount, compared

to NTC (Fig. 4d). A similar recent approach utilizes thiol capped gold nanoparticles to detect N-

gene of the SARS-CoV-2 gene in a cellular system.(9) However, for the detection of SARS-

CoV-2   RNA in human samples, the N-gene has reportedly inferior analytical power than the

RdRp gene.(10) Thus, it would be essential to know the assay's performance, developed by

Moitra et al., with clinical samples.

Conclusions

We have successfully developed an affordable gold nanoparticles-based colorimetric test for the

rapid detection of SARS-CoV-2   RNA in humans. The assay can detect up to 0.5 ng of SARS-

CoV-2 RNA. The turnaround time of our assay is less than 30 minutes. Moreover, the

developed test will be helpful for mass screening, as it does not require sophisticated

equipment. Validation of the developed assay using a large number of clinical samples is

underway to determine the diagnostic accuracy, including sensitivity, specificity, false-positive

and negative rate.

Acknowledgement

The study was supported by overhead funds from the extramural grants to ST from DBT, ICMR

and MHRD. The authors wish to thank the technical staff of the Department of Microbiology

(SGPGIMS, Lucknow and RMLIMS, Lucknow) for RNA extraction and RT-PCR analysis for the

clinical diagnosis of subjects.

Contributions: ST, VK; conceived the idea; ST, VK, SM, RS; performed the experiments, JY,

UG; provided clinical samples and the diagnosis; TK, LS and SKV; provided critical comments on

the manuscript draft; ST, VK and SM; drafted the manuscript.

Competing Interest

The authors declare no conflict of interest.

Supplementary Files

Supplementary information supporting the finding of this study is available in this article as

supporting information.

References-

1.   https://www.who.int/emergencies/diseases/novel-coronavirus-2019?gclid=EAIaIQobChMI-u-

Pz9vW6QIVUyUrCh0kjAlOEAAYASAAEgJ0pvD_BwERemuzzi.

2.   Udugama, B., Kadhiresan, P., Kozlowski, H. N., Malekjahani, A., Osborne, M., Li, V. Y. C., Chen, H.,

Mubareka, S., Gubbay, J. B., and Chan, W. C. W. (2020) Diagnosing COVID-19: The Disease and

Tools for Detection. *ACS Nano* **14**, 3822-3835

3.   Report of the WHO-China Joint Mission on Coronavirus Disease 2019 (COVID-19); WHO. (2017)

4.   Division of Viral Diseases. CDC 2019-Novel Coronavirus (2019-nCoV) Real-Time RT-PCR

Diagnostic Panel; Division of Viral Diseases, U. S. C. f. D. C. a. P. (2020)

5.   Kashir, J., and Yaqinuddin, A. (2020) Loop mediated isothermal amplification (LAMP) assays as a

rapid diagnostic for COVID-19. *Med Hypotheses* **141**, 109786

6.   Broughton, J. P., Deng, X., Yu, G., Fasching, C. L., Servellita, V., Singh, J., Miao, X., Streithorst, J.

A., Granados, A., Sotomayor-Gonzalez, A., Zorn, K., Gopez, A., Hsu, E., Gu, W., Miller, S., Pan, C.-

Y., Guevara, H., Wadford, D. A., Chen, J. S., and Chiu, C. Y. (2020) CRISPR–Cas12-based detection

of SARS-CoV-2. *Nature Biotechnology*

7.   Qiu, G., Gai, Z., Tao, Y., Schmitt, J., Kullak-Ublick, G. A., and Wang, J. (2020) Dual-Functional

Plasmonic Photothermal Biosensors for Highly Accurate Severe Acute Respiratory Syndrome

Coronavirus 2 Detection. *ACS Nano* **14**, 5268-5277

8.   Chen, Z., Zhang, Z., Zhai, X., Li, Y., Lin, L., Zhao, H., Bian, L., Li, P., Yu, L., Wu, Y., and Lin, G. (2020)

Rapid and Sensitive Detection of anti-SARS-CoV-2 IgG, Using Lanthanide-Doped Nanoparticles-

Based Lateral Flow Immunoassay. *Analytical Chemistry* **92**, 7226-7231

9.   Moitra, P., Alafeef, M., Dighe, K., Frieman, M. B., and Pan, D. (2020) Selective Naked-Eye

Detection of SARS-CoV-2 Mediated by N Gene Targeted Antisense Oligonucleotide Capped

Plasmonic Nanoparticles. *ACS nano*, acsnano.0c03822

10.   Corman, V. M., Landt, O., Kaiser, M., Molenkamp, R., Meijer, A., Chu, D. K., Bleicker, T., Brünink,

S., Schneider, J., Schmidt, M. L., Mulders, D. G., Haagmans, B. L., van der Veer, B., van den Brink,

S., Wijsman, L., Goderski, G., Romette, J.-L., Ellis, J., Zambon, M., Peiris, M., Goossens, H.,

Reusken, C., Koopmans, M. P., and Drosten, C. (2020) Detection of 2019 novel coronavirus

(2019-nCoV) by real-time RT-PCR. *Euro Surveill* **25**, 2000045

11.   Jain, P. K., Lee, K. S., El-Sayed, I. H., and El-Sayed, M. A. (2006) Calculated Absorption and

Scattering Properties of Gold Nanoparticles of Different Size, Shape, and Composition:

262   Applications in Biological Imaging and Biomedicine. *The Journal of Physical Chemistry B* **110**,

263   7238-7248

264 12. Lazarides, A. A., and Schatz, G. C. (2000) DNA-Linked Metal Nanosphere Materials: Structural

265   Basis for the Optical Properties. *The Journal of Physical Chemistry B* **104**, 460-467

266 13. Hunter, R. J. F. o. C. S. O. U. P. I. N. Y., 2001. Shaw, D. J. Colloid and Surface Chemistry;

267   Butterworth-Heinemann Ltd.: Oxford. (1991)

268 14. Grabar, K. C., Freeman, R. G., Hommer, M. B., and Natan, M. J. (1995) Preparation and

269   Characterization of Au Colloid Monolayers. *Analytical Chemistry* **67**, 735-743

270 15. Li, H., and Rothberg, L. (2004) Colorimetric detection of DNA sequences based on electrostatic

271   interactions with unmodified gold nanoparticles. *Proceedings of the National Academy of*

272   *Sciences of the United States of America* **101**, 14036-14039

284 **Figure 1.**

Figure 1. Schematic representation of the assay for the visual detection of SARS-CoV-2 RNA

Figure 1. Schematic representation of the assay for the visual detection of SARS-CoV-2 RNA

Schematic illustrates the assay flow to detect *RdRp* (RNA dependent RNA polymerase) gene sequence of SARS-CoV-2 in the nasopharyngeal RNA sample from subject clinically diagnosed with nCOVID infection (**positive control**). Hybridization buffer with *RdRp* oligo probe (forward) was mixed with RNA sample. The reaction mixture was denatured at 95°C for 30 seconds, followed by annealing at 60° C for 60 seconds. After annealing, the tube was kept at room temperature for 10 minutes before colloidal AuNPs were added. A pure colloid is pink in color. It turns blue in the vial containing RNA sample from **positive control** due to salt-induced aggregation upon successful hybridization between oligo probe and target RNA. The color remained pink in the absence of target RNA, or the presence of a non-specific target.

Figure 2.

Figure 2. *Colorimetric assay to detect SARS-CoV-2 RNA.*

(a) Comparative absorption spectra of unmodified AuNPs (Green curve) and oligo probe stabilized AuNPs i.e., NTC (Grey curve). Both spectra are exhibiting characteristics absorption peak at λmax 520 nm, however, the reduced peak intensity in NTC is due to dilution of colloid solution. In NTC, no red-shift in peak position, approve the stabilizing property of single-stranded oligo probe, Optical images of gold colloids (left vial) and NTC (right vial) are shown in the inset. **(b)** Comparative absorption spectra of NTC (Grey curve) and positive control i.e., nasopharyngeal RNA sample from subject clinically diagnosed with nCOVID infection (Brown curve). Broadening of the peak as well as red-shift in peak position confirm the salt-induced aggregation of AuNP due to successful hybrid formation in control. Optical images shown in inset demonstrate the evident change in the color of the solution from pink to blue in the control vial (right) while no change in color of NTC vial (left). **(c)** Representative absorption spectra, and in the inset shows optical images comparing assay performed with NTC (Grey curves, left vial), RNA from clinically diagnosed nCOVID infected subjects (Brown curve, middle vial), and RNA from subjects without nCOVID infection (Yellow curve, right vial). Samples from a total of seven infected and six uninfected individuals were analyzed.

288 **Figure 3.**

Figure 3. Specificity of developed assay to detect SARS-CoV-2 RNA.

Nasopharyngeal-RNA from nCOVID infected subject (positive control), and a cervical DNA sample from Human papillomavirus (HPV, non-specific target control) infected women were tested. **(a)** Comparative absorption spectra for no target control (NTC, Grey curve), positive control (Brown curve), and non-specific target control (HPV, Yellow curve). In the positive control, the broadening of the peak and red-shift in peak position (Brown curve) confirmed the salt-induced aggregation upon successful hybrid formation. **(b)** Optical pictures demonstrate that the colloid color from pink to blue changes only in the vial with the positive control (middle vial), while the color remained pink in the vial with NTC (left vial) or HPV (right vial).

Figure 4.

Figure 4. Sensitivity of developed assay to detect SARS-CoV-2 RNA.

Assay sensitivity was determined using different concentrations (concentration are in ng) of IVT synthesized SARS-CoV-2 RNA **(a-b)**, or nCOVID synthetic DNA **(c-d)**. Absorption spectra **(a and c)** corresponding to aggregated colloids exhibit a clear red-shift in peak with broadening, indicating successful hybridization. Optical pictures **(b &d)** of the assay performed to demonstrate the color change, an extra pair of vials (on the left) to show color change with a higher amount (10ng) of target nucleic acid for the reference

Supporting information

Development of RNA-based visual assay for rapid detection of SARS-CoV-2
in clinical samples

[1#]Vinod Kumar, [1#]Suman Mishra, [1#]Rajni Sharma, [2]Jyotsna Aggarwal, [3]Ujjala Ghoshal, [4]Tripti

Khanna, [1]Lokendra K. Sharma, [1]Santosh Kumar Verma and [1]Swasti Tiwari

[1]Department of Molecular Medicine & Biotechnology, SGPGIMS, Raibareli Road, Lucknow -

226014, India; [2] Department of Microbiology, Dr. Ram Manohar Lohia Institute of Medical

Sciences, Lucknow, 226010; [3]Department of Microbiology, Virology and COVID19 Lab,

SGPGIMS, Raibareli Road, Lucknow -226014, India; [4]Indian Council of Medical Research,

Ramalingaswami Bhawan, New Delhi, India

Figure S1

a)   b)

Figure S1. The assay was performed using an oligo probe specific for regenerating islet-derived (REG 3). The assay was containing RNA extracted from pancreatic tissue (positive control), placental RNA (negative control), or no RNA (NTC). **(a)** Comparative absorption spectra of NTC (Grey curve), positive control (Blue curve), and negative control (red curve). In the positive control, the broadening of the peak and red-shift in peak position (Blue curve) confirm the salt-induced aggregation upon successful hybrid formation. **(b)** Optical pictures demonstrate that the colloid color from pink to blue changes only in the vial with the positive control (middle vial), while the color remained pink in the vial with NTC (left vial) or negative control (non-specific target, right vial).

Table S1

Patient ID	Age in Years (Y)/ Gender	nCOVID19 infection	E gene Ct value	RdRp Ct Value
S1	18 Y / male	Positive	25.45	25.30
S2	18Y / male	Positive	28.37	29.20
S3	11 Y /female	Negative	-	-
R1	1 Y/Female	Positive	22.66	21.20
R2	28 Y/Female	Negative	-	-
R3	35 Y/Male	Negative	-	-
R4	66 Y/Female	Positive	22.81	24.59
R5	18 Y/Male	Positive	23.16	22.66
R6	14 Y/Male	Negative	-	-
R7	70 Y/Female	Negative	-	-
R8	44 Y/Male	Positive	29.64	25.56
R9	35Y /Male	Negative	-	-
R10	9 Y/Female	Positive	28.69	29.70

Table S1. *The table shows the clinical diagnosis of the subjects based on Taqman RT-PCR analysis.* RNA extracted from the nasopharyngeal samples were analyzed in the study.