Highly Sensitive Detection of *Campylobacter spp*. in Chicken Meat using a Silica

Nanoparticle Enhanced Dot Blot DNA Biosensor

Priya Vizzini[a,b], Marisa Manzano[b], Carole Farre[c], Thierry Meylheuc[a], Carole Chaix[c], Nalini Ramarao[a,*], Jasmina Vidic[a,*]

[a]Université Paris-Saclay, INRAE, AgroParisTech, Micalis Institute, 78350, Jouy-en-Josas, France.

[b]Dipartimento di Scienze AgroAlimentari, Ambientali e Animali, Università di Udine, Italy.

[c]Institut des Sciences Analytiques, UMR 5280, CNRS-UCBL, Université de Lyon, Lyon, France.

[*]Corresponding author:

E-mail address: jasmina.vidic@inrae.fr

Phone: + 33134652737

nalini.ramarao@inrae.fr

Abstract

Paper-based DNA biosensors are powerful tools in point-of-care diagnostics since they are affordable, portable, user☐friendly, rapid and robust. However, their sensitivity is not always as high as required to enable DNA quantification. To improve the response of standard dot blots, we have applied a new enhancement strategy that increases the sensitivity of assays based on the use of biotinylated silica-nanoparticles (biotin-Si-NPs). After immobilization of a genomic *Campylobacter* DNA onto a paper membrane, and addition of a biotinylated-DNA detection probe, hybridization was evidenced using streptavidin-conjugated to horseradish peroxidase (HRP) in the presence of luminol and H_2O_2. Replacement of the single biotin by the biotin-Si-NPs boosted on average a 30 fold chemiluminescent read-out of the biosensor. Characterization of biotin-Si-NPs onto a paper with immobilized DNA was done using a scanning electron microscope. A limit of detection of 3 pg/µL of DNA, similar to the available qPCR kits, is achieved, but it is cheaper, easier and avoids inhibition of DNA polymerase by molecules from the food matrices. We demonstrated that the new dot blot coupled to biotin-Si-NPs successfully detected *Campylobacter* from naturally contaminated chicken meat, without needing a PCR step. Hence, such an enhanced dot blot paves the path to the development of a portable and multiplex paper based platform for point-of-care screening of chicken carcasses for *Campylobacter*.

Graphical abstract

Keywords: *Campylobacter,* DNA dot blot, Si-nanoparticles, Food safety, Multiplex bacterial detection.

1. Introduction

Campylobacter is considered the most common bacterial cause of human gastroenteritis in the world. The World Health Organization (WHO) estimates that 550 million people fall ill every year from gastroenteritis, of which 220 million are children (WHO 2020). This zoonosis is transmitted mostly through exposure to under-cooked poultry products (50-80% cases), and in a minor way, to contaminated milk, vegetables and water (ECDC 2018; Hermans et al. 2012; Humphrey et al. 2007). Campylobacteriosis in humans is usually a self-limiting condition involving bloody diarrhea, abdominal cramping, nausea and fever, which can last up to two weeks. In 1 % of cases, campylobacteriosis evolves to the Guillain-Barré syndrome, a severe autoimmune disease that leads to death in 2–12 % of patients, depending on their age (Scallan Walter et al. 2020). The overall economic burden of campylobacteriosis was estimated to–about EUR 3 billion/year in EU (ECDC 2018), and between US$ 1.2 and 4 billion/year for the US (Batz et al. 2014). To prevent the entry of *Campylobacter* contaminated broilers into the market, the European Commission adopted a process hygiene criterion (EU 2017/1495) with the critical limit of 1000 cfu/g of broiler meat or skin and obligatory systematic screening of broiler carcasses for *Campylobacter* spp. (EU No 2017/625).

Identification and quantification of *Campylobacter* spp. rely on official, culture-based methods, and bacterial biochemical/phenotypical characterization (Vizzini et al. 2019). *Campylobacter* is highly infectious, with reported infective doses as low as 500 cells (Black et al. 1988). To ensure the detection of one cell of *Campylobacter* in 25 g of food, an enrichment step in Bolton broth for 24–48 h is needed, before isolation of the colonies by culturing (it takes about 48–72 h) onto selective agar plates incubated in chambers for microaerophilic conditions. *Campylobacter* is usually present in low numbers in food samples when compared to bacteria like coliforms and Enterobacteriaceae making its isolation on

common agar media difficult. The official ISO 10272-1:2006 method for *Campylobacter*

detection may provide false negative results because of the possibility *Campylobacter* death

during handling. Furthermore, *Campylobacter* can enter in a viable but not cultivable (VBNC)

status in food matrices, making its detection based on culturing impossible (Vidic et al. 2017;

Vizzini et al. 2019). Identification of *Campylobacter* by the optical microscope is not easy as

bacteria can change their spiral distinctive shape into spherical or coccoid. Moreover, the

current problem of official methods is the use of a not enough selective medium, and

emergence of bacterial resistance against the antibiotic added to the enrichment broth.

Culture-independent molecular methods, such as PCR, and real-time PCR are used as

alternatives to colony growth (De Boer et al. 2015; Fontanot et al. 2014; Gosselin-Théberge et

al. 2016; Liu et al. 2017; Ricke et al. 2019). However, PCR-based methods may also provide

false-negative responses because of the sensitivity of DNA polymerase to inhibitors present in

food matrices and enrichment broths (Schrader et al. 2012; Vidic et al. 2019). Next generation

sequencing (NGS) which enables sequencing of the entire bacterial genome in a relatively

short time has started to be used as a tool for the identification of infectious bacteria.

However, NGS can hardly be routinely used at farms and slaughterhouse because it requires

special equipment and highly trained personnel for data interpretation (Gosiewski et al. 2017).

Biosensors for detection of *Campylobacter* show marked advantages over traditional

methods in terms of rapidity, facility utilization and cost-effectiveness (Manzano et al. 2015;

Masdor et al. 2016; Morant-Miñana and Elizalde 2015; Vidic et al. 2017; Vidic et al. 2019;

Vizzini et al. 2019; Yang et al. 2013). However, today no commercial biosensor is available

to detect *Campylobacter* in food matrices, mainly due to the difficulty of attaining a high

sensitivity.

Due to significant improvement in biosensor technology over the last two decades,

applications of paper-based sensors in pathogen detection are increasing. Such devices seem

to meet ASSURED criteria (Affordable, Sensitive, Specific, User friendly, Rapid and robust,

Equipment free and Deliverable) recommended by the WHO for point-of-care diagnostics.

Paper as a substrate has the advantage of being inexpensive, lightweight, and easily enables

multiplex analysis (Dincer et al. 2019). Various paper-based methods to detect foodborne

pathogens have been reported including enzymatic-, immuno- and DNA/RNA- tests (Jokerst

et al. 2012; Morales-Narváez et al. 2015; Russell et al. 2017). Tests that target nucleic acids

are among the most accurate and specific although they are usually associated with a PCR

step, including isothermal amplification, to allow for detection of pathogens in low titers

(Trinh et al. 2020; Vidic et al. 2019).

Here, we coupled functionalized silica nanoparticles (Si-NPs) to a paper based DNA

dot blot test to enable sensitive *Campylobacter* nucleic acid detection without a pre-

amplification step. A highly specific DNA probe that recognizes the 16S rRNA gene of the

most prevalent *Campylobacter* spp. causing infections (*C. jejuni*, *C. coli*, *C. lari*, and *C.*

upsaliensis) was used in the test. Si-NPs decorated with biotin (biotin-Si-NPs) were employed

to enhance the strepatavidin–HRP chemiluminescent signal read-out. Our study illustrates the

proof-of-concept that the collective effect of biotin-Si-NPs could become an effective means

for increasing the sensitivity of the cost-effective, specific and easy-to-perform dot blot assay.

2.   Material and Methods

2.1   Materials and reagents

Streptavine-HRP, Proclin, acetonitrile, 1,8-Diazabicyclo[5.4.0]undéc-7-ène (DBU),

Controlled pore glass (120-200 Mesh, CPG-3000 Å) were purchased from Sigma-Aldrich

(Saint Quentin Fallavier, France). Fluorescent rhodamine B silica NPs (Si-NPs, diameter~50

nm) were supplied by Nano-H (Saint Quentin Fallavier, France). DNA phosphoramidite

synthons and all DNA-synthesis reagents were purchased from Glen Research (Sterling,

Virginia, USA). Phosphate buffered saline (PBS) was purchased from Dominique Dutscher

(Brumath, France). Amersham Hybond-N+ nylon and Amercham Hybon-XL nitrocellulose

membaranes were purchased from ThermoFisher (Illkirch, France).

All bacterial media and supplements used were from Oxoid (Milan, Italy), except for

the Violet red Bile glucose (VRBG) agar and Coli ID medium that were purchased from

Biomerieux (Bagno a Ripoli, Italy). Triton, SDS, NaCl, trizma-base, phenol, chloroform,

bacteriological peptone and isoamyl alcohol were purchased from Sigma (Milan, Italy). A 36-

mer oligonucleotide related to the 16S gene encoding for ribosomal *Campylobacter* RNA of

C. jejuni, *C. coli*, *C. lari*, and *C. upsaliensis* (base location: 72-108) was used as a recognition

element (probe named CampyP3). It is interesting to note that the probe matches in three

points for each *Campylobacter* genome at 100%. CampyP3 was labeled at 5' with biotin for

chemiluminescent dot blot assay development. The probe was tested *in silico* by the

OligoAnalyzer3.1 (https://eu.idtdna.com/calc/analyzer), the Amplifix software, Fast PCR 6.1

and Blast (https://blast.ncbi.nlm.nih.gov/Blast.cgi). An ssDNA sequence complementary to

the CampyP3 probe, of the same length as the probe was named CP3 and used as positive

control. Two ssDNA sequences of the same length as the CP3, but not complementary to

CampyP3, were named PR and PE, and used as negative controls to study selectivity of the

sensor through their hybridization with the probe. PR was designed by mismatching positions

of nucleic acids of the CP3 sequence, while PE corresponded to the sequence of *E. coli*

*(*accession number 527445.1) (base location: 338-376 for *E. coli),* which shows some

similarities to *Campylobacter.* These sequences, reported in Table 1S, were provided by

Sigma-Aldrich (Saint Quentin Fallavier, France) as lyophilized powers. All solutions were

prepared using Milli-Q water.

2.2   *Bacterial strains*

Bacteria used in this study are listed in Table 2S. *Campylobacter* strains were grown under

microaerophilic conditions (5% O_2, 10 % CO_2 and 80% N_2, generated with a Sachet Oxoid™

CampyGen™ 2.5 L (Oxoid, Italy) in anaerobic gas jars at 37°C for 48h, on Columbia blood

agar plates. *Campylobacter* isolates, both reference strains and strain isolates from chicken

samples, were subjected to Gram straining and optical microscope observations for cell

morphology and motility (Brucella broth, Thermofisher, Milan, Italy) after oxidase and

catalase tests.

All negative control strains were grown on their specific culture media at optimal

conditions and subjected to the same tests carried out for *Campylobacter* before DNA

extractions.

156   *2.3   Sample collection, plate count enumeration and selective isolation*

Five chickens were purchased from local supermarkets in Italy. 10 g of chicken skin was

transferred into a filter-sterile stomacher bag with 40 mL saline-peptone water (8 g/L NaCl, 1

g/L bacteriological peptone) and homogenized in a Stomacher (PBI, Milan, Italy) for 90 s.

Aliquots of 0.1 mL were spread for the mesophilic aerobic count on Triptone Soya Agar, and

incubated at 30°C for 48 h, while 0.1 mL were spread on Agar Malt tetracycline at 30°C for

48 h to count yeast/molds. Aliquots of 1 mL were used with the pour plate method for

enumeration of Enterobacteriaceae in the VRBG agar (37°C for 24 h), and coliforms and *E.*

coli in Coli ID medium (37°C for 24 h). *Campylobacter* detection was performed according to

the conventional method ISO 10272-1:2006. For selective bacterial isolation, 10 µL of the

Bolton broth were streaked onto Skirrow agar plates, and mCCDA (modified charcoal-

cefoperazone-deoxycholade) plates and incubated at 41.5 °C for 48h, under microaerophilic

conditions. One colony was selected from mCCDA and streaked on two plates of Columbia

blood agar. One plate was incubated at 41.5 °C for 48 h in aerobic condition and one plate at

25 °C for 48 h in microphilic condition. Growth on Skirrow served as a confirmation to

proceed with the identification steps.

2.4   DNA extraction from pure cultures and enrichment broths

DNA was extracted from reference strains and isolates from chicken samples according to

(Manzano et al. 2015). 2 mL of Bolton enrichment broth was centrifuged at 13,000 g at 4 °C

for 10 min, and bacterial pellets were washed three times with PBS 1X, and subjected to

extraction (Manzano et al. 2015). Extracted DNAs were rehydrated using 50 µL of sterile

distilled water and treated with RNase enzyme at 37°C for 1 h. Finally, DNA was quantified

using a Nanodrop™ 2000C (ThermoFisher Scientific, Milan Italy) spectrophotometer. The

concentration of DNA was adjusted to 100 ng/µL using sterile distilled water.

2.5 Si-NP synthesis, labeling and quantification

Si-NPs was functionalized by an innovative solid-phase synthesis technology which

enables functionalization of nanosized particles with DNA fragments, as reported previously

(Bonnet et al. 2018; De Crozals et al. 2012). Briefly, nanoparticle immobilization onto

controlled pore glass (CPG) allows a very high functionalization with a modified

oligonucleotidic based linker. The linker (sequence: dT10-PEG2-dT10) was synthetized using

an applied Biosystems 394 RNA/DNA synthesizer (Applied Biosystems). After grafting

nanoparticles on a CPG support, they were functionalized by automated synthesis using the

phosphoramidite chemistry. First, the linker, that allows a better accessibility of the functions

of interest, was synthesized. Second, the biotin group was incorporated. We controlled the

biotin incorporation to reach a 10% coupling yield. To do it, diluted solutions of biotin

phosphoramidite (10 mM) and tetrazole in acetonitrile (45 mM) were used and the coupling

time was reduced to 10 s. The biotin incorporation was monitored at 498 nm using

dimethoxytrityl quantification by an UV-visible spectrophotometer. Third, biotin-Si-NPs were

released from CPG by incubating the support in 1 mL of 0.1% (m/v) DBU in water-

acetonitrile 1/1 (v/v). The DBU solution was stirred in a thermomixer at 22°C during 1 h

before recovering. A fresh DBU solution was added to the CPG suspension every hour. The

release kinetics was followed by quantifying the NP concentration in each DBU solution with

an UV-visible measurement at 560 nm. Released nanoparticles were washed with milli-Q

water (1x 4 mL then 3 x 2 mL) and concentrated on 30 K Amicon Ultra filter (5000 g, 10

min). Finally, the amount of dT10-PEG2-dT10-10% Biotin strands per NP was estimated

using a Varian Cary 100 Bio UV-visible spectrophotometer (Agilent Technologies) and a

quartz cuvette of 1 cm path length.

The amount of strands per NP was estimated using a Varian Cary 100 Bio UV-visible

spectrophotometer (Agilent Technologies) and a quartz cuvette of 1 cm path length. The

amount of linker grafted to nanoparticles was quantified by measuring absorbance in water

(200 µL) at 260 nm and 560 nm with a microplate reader (Perkin Elmer). The nanoparticle

concentration was estimated as described previously (Bonnet et al. 2018). To fit with this

estimation, we approximated a molar extinction coefficient to 163200 M^{-1} cm^{-1} that considers

the epsilons of the different parts of the sequence corrected by their corresponding coupling

yield.

Functionalized NPs were observed under a transmission electron microscopy (TEM)

using a Philips CM120 instrument operating at an accelerate voltage of 120 kV (Centre

Technologique des microstructures, Lyon). Si-NPs were observed after deposition of 5 µL of

diluted solution on a formvar-carbon coated copper grid followed by evaporation until dry.

2.6 Scanning electron microscopy

Samples were mounted on aluminum stubs (50 mm diameter) with carbon adhesive

discs (Agar Scientific, Oxford Instruments SAS, Gomez-la-Ville, France) and visualized by

field emission gun scanning electron microscopy (SEM FEG) as secondary and backscattered

electrons images (5 kV) under high vacuum conditions with a Hitachi SU5000 instrument

(Milexia, Verrières-le-Buisson, France). Sample preparation and scanning Electron

Microscopy analyses were performed at the Microscopy and Imaging Platform MIMA2

(INRAE, Jouy-en-Josas, France).

2.7 Sample immobilization and detection procedure

Prior to immobilization onto the nylon membrane, extracted DNAs were denatured at

95 °C for 10 min, put immediately on ice and 1 µL spotted on the positively charged nylon

membrane (Amersham Hybon^{TM}-XL, GE Healthcare, France), which was exposed to UV at

254 nm for 10 min to fix DNA. The membrane was then soaked in a pre-wormed

hybridization buffer (0.5 M Na_2HPO_4, 0.5 M NaH_2PO_4, 10 mM EDTA, 1 % SDS, pH 7.5) at

65°C for 30 min under gentle shaking. 4 ng/µL of the denatured biotin labeled CampyP3

probe (100 ng/µL) was added to the hybridization buffer and left overnight at 65°C under

gentile shaking to allow for hybridization.

Subsequently, the membrane was washed twice with 0.1 SDS, 300 mM SSC (Saline

sodium citrate, Meraudex, France) for 5 min and with 75 mM SSC for 15 min. The membrane

was transferred in the blocking buffer (0.1 % Tween, 0.1 % BSA, 0.03% Proclin, PBS, pH

7.4) at room temperature for 15 min to saturate the surface. The membrane was finally

incubated with the blocking solution containing 0.7 µM streptavidin-HRP for 15 min at room

temperature. After washing with 0.1 % SDS, 150 mM SSC at room temperature for 5 min, the

membrane was incubated with 10^6 biotin-Si-NPs/mL, PBS, pH 7.4 under gentile shaking at

room temperature, for 30 min. A 0.7 µM streptavidin-HRP in blocking solution was added

after washing and incubated for 30 min with shaking. The signal was then revealed using the

enhanced chemiluminescent substrate for detection HPR (Thermo Scientific, France). The

membrane was removed from the solution and observed under a ChemiDoc MP imaging

system (Biorad, France). Detection signals were quantified using Image Lab^{TM} software

(Biorad). The normalized value of a spot intensity was calculated by $(PI_0 - PI_n)/PI_0$, where PI_0

and PI_n represent the pixel intensity obtained for detection of 0.1 ng/µL CP3 probe (standard)

and the experimental sample, respectively.

Results and discussion

3.1. Preparation and characterization of biotin-Si-NPs

Biotin-Si-NPs were prepared by an innovative solid-phase synthesis technology that

allowed for a very high functionalization. A hydroxyl group was attached to the Si-NP surface

by silanization to support the oligonucleotide linker synthesis on the NP surfaces, as

previously reported (Bonnet et al. 2018; Farre et al. 2010). In the final step, biotin groups

were grafted to the linker. TEM observations showed that the NP morphology was stable

during synthesis (Fig. 1A). Functionalized biotin-Si-NPs had an average size of 50 ± 3 nm, as

estimated from TEM images (Fig. 1B). The absorbance measurements suggested that about

500 molecules of the linker were attached to each NP, while dimethoxytrityl quantification

indicated the biotin-coupling yield of 10 %, which quantified biotin to about 50 per NP (Fig.

1C).

A quite narrow size distribution of functionalized NPs was confirmed by DLS

measurements. Si-NPs formed monodisperse aqueous solutions of particles with a

hydrodynamic diameter (R_H) of about 90 nm. Conjugation of biotin with the linker moieties

shifted R_H to about 120 nm (Fig. 1D). This increase is probably related to the hydration layer

around biotin units linked to arms bearing negatively charged groups. Biotin-Si-NP solutions

were stable at 4°C for at least two months (Supplementary Material, Figs. 1S and 2S).

3.2.   Concept of biotin-Si-NPs based DNA dot blot

The principle of the DNA dot blot method for *Campylobacter* detection enhanced with

biotin-Si-NPs is described in Fig. 2. Initially, the target DNA was immobilized on a porous

nylon membrane by UV irradiation to enable crosslinking of DNA to the positively charged

surface. Thereafter, the DNA probe CampyP3 labeled with biotin was allowed to hybridize

with the target DNA. Hybridization was detected with a streptavidin-HRP conjugate in

combination with a chemiluminogenic substrate luminol in the presence of the activator H_2O_2.

In enhanced detection, the streptavidin-biotin sandwich enabled attachment of biotin-Si-NPs

to the DNA probe, and consequently amplification of the detection signal compared to a

single biotin.

3.3. *Optimization of key parameters and analytical performances of the DNA dot blot*

sensor

The time and temperature of hybridization of the DNA probe with its target were

optimized as they can markedly influence the sensibility and selectivity of a DNA sensor.

First, hybridization of DNA was tested at room temperature, 44°C, 55°C and 65°C using

complementary and non-complementary short ssDNA targets. A complementary DNA target

sequence CP3 (positive control) was detected at all temperatures tested, but only at 65°C no

non-specific binding was obtained with negative controls, a truncated target sequence (PR)

and *E. coli* sequence (PE) (Fig. 3A). Second, hybridization was studied using 1 ng/µL CP3 at

65°C for different times from 1 h to overnight. The overnight incubation provided the highest

level of hybridization and maximal signal intensity.

Other parameters that may affect hybridization, such as buffer compositions or the

paper support were also tested. The optimization criterion was the best signal-to-background

ratio. The nylon membrane allowed for better immobilization of DNA compared to

nitrocellulose. The effects of formamide (10, 25, 30, 35 and 50 %), DMSO (0 and 10%) and

SDS (0, 1 and 10 %) in the hybridization buffer were tested since these denaturing agents for

double strained DNA may prevent nonspecific hybridization. The optimal hybridization

buffer contained SDS, and no DMSO or formamide. We determined that addition of 1 % SDS

enhanced the intensity of signals obtained and reduced the frequency of unspecific

background staining. The effect of SSC concentrations (0.5x, 1x and 2x) in the washing

buffer was also investigated. A 3-step washing procedure with two buffers (2xSSC, 0.1 %

SDS and 0.5xSSC, 0.1 % SDS) lead to elimination of the background signal while

maintaining high light intensity of specific spots.

Detection of 1 ng/µL CP3 under the optimized conditions using various

concentrations of streptavidin-HRP and biotin-Si-NPs allowed selection of 25 ng/µL of

streptavidin-HRP and 10^6 nanoparticles/mL for the chemiluminometric reaction. It is worth

noting that the chemiluminometric reaction itself does not significantly contribute to

background staining as the emission of light arises from the enzymatic reaction without any

photonic excitation (Laios et al. 2010).

3.4 Calibration curve

Calibration curves were obtained from quantification of chemiluminescent light

intensity for different CP3 concentrations, taken from at least three independent dots per

concentration. The linear range for the CP3 target was from 1 ng/µL to 0.1 ng/µL (with

correlation R^2=0.98684), and from 6 pg/µL to 0.1 ng/µL (with correlation R^2=0.98935), for

classical and biotin-Si-NP enhanced dot blot techniques, respectively (Fig. 3B-D). The limit

of detection (LOD) of 0.08 ng/µL and 0.003 ng/µL for classical and enhanced dot blot,

respectively, were calculated using 3 s/m formula, where 's' is the standard deviation of the

blank solution and 'm' is the slope of the linear calibration graph. Taking into account the

molecular weight of the CP3 of 14395.5, the calculated LOD was 5.5 nM and 0.2 nM, for

classical and enhanced dot blots, respectively. Consequently, biotin-Si-NPs allowed for

almost a 30 fold increase in chemiluminescent signal intensity compared to the classical dot

blot. We estimate LOD of about 600 cells considering that one *Campylobacter* cell contains 2

fg of genomic DNA (Pacholewicz et al. 2013). The low LOD obtained suggests that this

enhanced dot blot method may be suitable for meat sample testing as the *Campylobacter*

infection dose is around 500 cells (2.5 pg DNA).

The proposed biosensor showed comparable or enhanced performance in relation to

recently reported DNA sensors for *Campylobacter* detection where the detection limit was 0.5

ng/µL in (Fontanot et al. 2014), 0.37 ng/µL in (Manzano et al. 2015) and 0.09 nM in (Morant-

Miñana and Elizalde 2015). The biotin-Si-NP enhanced dot blot reproducibility was estimated

to 5 % according to the signal obtained for detection of the same concentration of CP3.

Although our time of analysis is not the lowest reported for *Campylobacter* detection (Table

1), our assay is, as far as we know, the first one for the *Campylobacter* detection that has a

high sensitivity and does not require the specific bacterial morphology or the background

matrix. The colorimetric aptasensor was reported to detect spiral *Campylobacter* cells but not

spherical or coccoid ones (Kim et al. 2018); whereas real-time PCR can be inhibited by

molecules and ions presented in enrichment broths or meat (Alves et al. 2016).

SEM characterization of the biotin-Si-NPs dot blot

SEM was applied to examine different detection steps in order to directly visualize

biotin-Si-NPs on the surface of the paper carrying hybridized DNA. The nylon surface before

DNA immobilization showed a typical membrane structure, including membrane pores (Figs.

4 and 3S). After *C. jejuni* DNA cross-linking to the nylon, the surface was occupied by a large

quantity of ssDNA. Compared to the surface of the bare membrane, the surface of membrane

carrying DNA became irregular showing increased morphological heterogeneity. ssDNA

molecules seemed to rest horizontally positioned over the surface (Fig. 4, right upper panel).

The membrane with hybridized DNA carrying biotin-Si-NPs (final detection step) had thicker

fibril structures than the ones with only immobilized ssDNA molecules (Fig. 4, lower panel).

Biotin-Si-NPs were easily detected on double strain DNA as small beads of about 50 nm in

diameter. It is worth to note, that because the DNA probe matches three points in the

Campylobacter genome, several biotin-Si-NPs may specifically bind to one DNA. Overall,

SEM images confirmed efficient DNA immobilization and hybridization by revealing that the

nylon surface became more complex after each stage of the detection process and that

hybridized DNA were decorated with nanoparticles of the expected size of 50 nm.

3.4. *Specificity and selectivity of detection*

To examine the selectivity and sensitivity of the biotin-Si-NPs dot blot, we tested

various strains of *Campylobacter* spp. for inclusivity, and 17 other bacterial species and 1

yeast strain for exclusivity (Fig. 5). All strains were cultivated as monocultures and genomic

DNA were extracted as explained in the experimental part. Dot blot analyses were performed

using 10 ng/µL of non-amplified extracted DNA. Before immobilization, all genomic DNA

were denatured for 5 min at 95°C to allow double strain opening. CP3 at 0.1 ng/µL was used

as an internal control to normalize signal intensities. The signal ratios for *C. jejuni*, *C. coli*, *C.*

lari, and *C. upsaliensis* over CP3 were higher than 1.0, with the membrane images exhibiting

obvious dots. Moreover, the intensity obtained for *C. jejuni* and *C. coli* were about 4 times

higher than that of the positive control. In contrast, the highest ratio among control DNA and

CP3 was only 0.6 for *S. enterica* (background staining). The obtained results highlight the

specificity of detection. In addition, the CampyP3 detection probe reacted with DNA of *C.*

jejuni and *C. coli* more efficiently than with DNA of other *Campylobacter* spp. tested. Our

results demonstrated that the biotin-Si-NPs enhanced dot blot platform is sensitive enough to

detect the whole DNA extracted from the most frequent *Campylobacter* spp. and could

distinguish between *Campylobacter* and other bacteria. This specificity of the CampyP3

probe suggests that no nonspecific signal will be generated upon testing DNAs of competing

total mesophilic flora in chicken meat.

3.5. *Campylobacter spp. detection in naturally contaminated chicken samples*

To demonstrate the capabilities of our developed biosensor for possible real-world

applications, we have chosen to target the detection of *Campylobacter* in chicken meat

samples. The results of plate count enumeration for background bacteria in five chicken

samples are reported in Table 3S. The ISO 10272-1:2006 indicated that two out of five

samples were contaminated with *Campylobacter* (Table 4S). Indeed, only isolates from

selective media (mCCDA, Skirrow and Columbia agar) of samples C3 and C10 were

confirmed for *Campylobacter* by the temperature growth and motility test.

In dot blot analysis, the CP3 template (0.1 ng/µL) was used as a positive control. The

dot blot enhanced with biotin-Si-NP successfully detected two contaminated samples, C3 and

C10 (Fig. 6). The test achieved a relative specificity of 100 % due to the high specificity of

the CampyP3 detection probe. The proposed biotin-Si-NP enhanced dot blot method, thus,

specifically detected low levels of naturally present *Campylobacter* after enrichment and

could be considered for determination and detection of *Campylobacter* spp. in contaminated

chicken meat. Compared to the official ISO 11272: 2006 method, the new DNA dot blot

biosensor is less laborious, more cost-effective and time-saving. First, the highly selective

DNA detection probe enables selective detection of *Campylobacter* in the presence of meat

background bacteria, and thus no culturing on a selective agar is needed. Second,

enhancement of the detection signal intensity allows DNA detection without PCR pre-

amplification.

Conclusion

This work describes the DNA dot blot method enhanced with biotin-Si-NPs for easy,

fast and reliable detection of *Campylobacter* spp. in contaminated chicken meats. The dot blot

test is widely used in molecular biology and genetics for detecting target DNA/RNA

sequences. This paper-based hybridization can easily analyze multiple samples inexpensively

in a high-throughput fashion but has limitation in a low sensitivity. We demonstrated that

association of highly functionalized biotin-Si-NPs with the chemiluminescent read-out of the

dot blot enhanced the signal about 30 times. Furthermore, biotin-Si-NPs are robust at working

temperature and stable in time. The LOD obtained was 3 pg/µL (0.2 nM). Such a low

concentration of detected DNA is close to values that can be detected by qPCR (Manzano et

al. 2018; Vidic et al. 2019). Hybridization of immobilized genomic DNA with the CampyP3

probe induced positive signals for different *Campylobacter* spp. responsible for human

gastroenteritis, while no background staining was observed with control samples. The

developed system is a promising tool for fast and cheap screening of poultry samples for the

presence of *Campylobacter* since detection is performed on bacterial DNA without a pre-

amplification step. Furthermore, our dot blot can be applied on DNA extracted by different

extraction methods and from various food matrices, as it is not sensitive to DNA polymerase

inhibitors.

The proposed sensitive, miniaturized and multiplex paper-based test is simple to

design and could be used by the food industry and regulatory agencies for the detection of

other pathogens to monitor food quality. We believe that in the future it could be integrated

into a Lab-on-the chip based biosensor that comprises an automatized DNA extraction

protocol and a mobile phone camera (Kalligosfyri et al. 2019; Vidic et al. 2019).

Appendix A. Supplementary material

A list of oligonucleotides and bacterial strains used in this studies, stability of biotin-Si-

nanoparticles, as well as the results obtained by the plate count method on chicken samples

are available in the supplementary material associated with this article.

Acknowledgements

The authors thank the Centre Technologique des Microstructures of the Lyon 1 University

for providing access to its TEM facilities, and the MIMA2 platform Jouy en Josas for access

to electron microscopy equipment (MIMA2, INRAE, 2018. Microscopy and Imaging Facility

for Microbes, Animals and Foods, https://doi.org/10.15454/1.5572348210007727E12). JV

thanks Maria-Vesna Nikolic (University of Belgrade, Serbia) for English editing. PV

acknowledges a doctoral fellowship from the University of Udine, Italy. This research was

supported in part by the European Union's Horizon 2020 research and innovation programme

under the Marie Skłodowska–Curie grant agreement No 872662 (IPANEMA), the European

Union's Horizon 2020 research and by the University Paris-Saclay through the Poc in labs

2019 grant agreement No 00003469 (OSCAR).

5. References

Alves, J., Hirooka, E.Y., de Oliveira, T.C.R.M., 2016. Development of a multiplex real-time PCR assay
with an internal amplification control for the detection of Campylobacter spp. and Salmonella spp. in
chicken meat. LWT-Food Science and Technology 72, 175-181.
Batz, M., Hoffmann, S., Morris Jr, J.G., 2014. Disease-outcome trees, EQ-5D scores, and estimated
annual losses of quality-adjusted life years (QALYs) for 14 foodborne pathogens in the United States.
Foodborne pathogens and disease 11(5), 395-402.
Black, R.E., Levine, M.M., Clements, M.L., Hughes, T.P., Blaser, M.J., 1988. Experimental
Campylobacter jejuni infection in humans. Journal of infectious diseases 157(3), 472-479.
Bonnet, R., Farre, C., Valera, L., Vossier, L., Léon, F., Dagland, T., Pouzet, A., Jaffrézic-Renault, N.,
Fareh, J., Fournier-Wirth, C., 2018. Highly labeled methylene blue-ds DNA silica nanoparticles for
signal enhancement of immunoassays: application to the sensitive detection of bacteria in human
platelet concentrates. Analyst 143(10), 2293-2303.
De Boer, P., Rahaoui, H., Leer, R., Montijn, R., Van der Vossen, J., 2015. Real-time PCR detection of
Campylobacter spp.: a comparison to classic culturing and enrichment. Food microbiology 51, 96-
100.
De Crozals, G., Farre, C., Hantier, G., Léonard, D., Marquette, C.A., Mandon, C.A., Marmuse, L., Louis,
C., Toulmé, J.-J., Billotey, C., 2012. Oligonucleotide solid-phase synthesis on fluorescent nanoparticles
grafted on controlled pore glass. RSC advances 2(31), 11858-11866.
Dincer, C., Bruch, R., Costa-Rama, E., Fernández-Abedul, M.T., Merkoçi, A., Manz, A., Urban, G.A.,
Güder, F., 2019. Disposable sensors in diagnostics, food, and environmental monitoring. Advanced
Materials 31(30), 1806739.
ECDC, E.a., 2018. The European Union summary report on trends and sources of zoonoses, zoonotic
agents and food-borne outbreaks in 2017. EFSa Journal 16(12), e05500.
Farre, C., Lansalot, M., Bazzi, R., Roux, S., Marquette, C.A., Catanante, G., Blum, L.c.J., Charvet, N.,
Louis, C., Chaix, C., 2010. Automated oligonucleotide solid-phase synthesis on nanosized silica
particles using nano-on-micro assembled particle supports. Langmuir 26(7), 4941-4950.
Fontanot, M., Iacumin, L., Cecchini, F., Comi, G., Manzano, M., 2014. Rapid detection and
differentiation of important Campylobacter spp. in poultry samples by dot blot and PCR. Food
microbiology 43, 28-34.
Gosiewski, T., Ludwig-Galezowska, A., Huminska, K., Sroka-Oleksiak, A., Radkowski, P., Salamon, D.,
Wojciechowicz, J., Kus-Slowinska, M., Bulanda, M., Wolkow, P., 2017. Comprehensive detection and
identification of bacterial DNA in the blood of patients with sepsis and healthy volunteers using next-
generation sequencing method-the observation of DNAemia. European Journal of Clinical
Microbiology & Infectious Diseases 36(2), 329-336.
Gosselin-Théberge, M., Taboada, E., Guy, R.A., 2016. Evaluation of real-time PCR assays and standard
curve optimisation for enhanced accuracy in quantification of Campylobacter environmental water
isolates. Journal of microbiological methods 129, 70-77.
Hermans, D., Pasmans, F., Messens, W., Martel, A., Van Immerseel, F., Rasschaert, G., Heyndrickx,
M., Van Deun, K., Haesebrouck, F., 2012. Poultry as a host for the zoonotic pathogen Campylobacter
jejuni. Vector-Borne and Zoonotic Diseases 12(2), 89-98.
Humphrey, T., O'Brien, S., Madsen, M., 2007. Campylobacters as zoonotic pathogens: a food
production perspective. International journal of food microbiology 117(3), 237-257.
Jokerst, J.C., Adkins, J.A., Bisha, B., Mentele, M.M., Goodridge, L.D., Henry, C.S., 2012. Development
of a paper-based analytical device for colorimetric detection of select foodborne pathogens.
Analytical chemistry 84(6), 2900-2907.

Kalligosfyri, P.M., Sevastou, A., Kyriakou, I.K., Tragoulias, S.S., Kalogianni, D.P., Christopoulos, T.K.,
2019. Smartphone-based chemiluminometric hybridization assays and quantitative competitive
polymerase chain reaction. Analytica Chimica Acta 1088, 123-130.
Kim, Y.-J., Kim, H.-S., Chon, J.-W., Kim, D.-H., Hyeon, J.-Y., Seo, K.-H., 2018. New colorimetric
aptasensor for rapid on-site detection of Campylobacter jejuni and Campylobacter coli in chicken
carcass samples. Analytica chimica acta 1029, 78-85.
Laios, E., Ioannou, P.C., Christopoulos, T.K., 2010. Gene Assays Based on Bio (Chemi) luminescence.
Chemiluminescence and Bioluminescence, pp. 334-377. RSC Publishing, Cambridge, UK.
Liu, K.C., Jinneman, K.C., Neal-McKinney, J., Wu, W.-H., Rice, D.H., 2017. Simultaneous identification
of Campylobacter jejuni, Campylobacter coli, and Campylobacter lari with smartcycler-based
multiplex quantitative polymerase chain reaction. Foodborne Pathogens and Disease 14(7), 371-378.
Manzano, M., Cecchini, F., Fontanot, M., Iacumin, L., Comi, G., Melpignano, P., 2015. OLED-based
DNA biochip for Campylobacter spp. detection in poultry meat samples. Biosensors and
bioelectronics 66, 271-276.
Manzano, M., Viezzi, S., Mazerat, S., Marks, R.S., Vidic, J., 2018. Rapid and label-free electrochemical
DNA biosensor for detecting hepatitis A virus. Biosensors and Bioelectronics 100, 89-95.
Masdor, N.A., Altintas, Z., Tothill, I.E., 2016. Sensitive detection of Campylobacter jejuni using
nanoparticles enhanced QCM sensor. Biosensors and Bioelectronics 78, 328-336.
Morales-Narváez, E., Golmohammadi, H., Naghdi, T., Yousefi, H., Kostiv, U., Horak, D., Pourreza, N.,
Merkoçi, A., 2015. Nanopaper as an optical sensing platform. ACS nano 9(7), 7296-7305.
Morant-Miñana, M.C., Elizalde, J., 2015. Microscale electrodes integrated on COP for real sample
Campylobacter spp. detection. Biosensors and Bioelectronics 70, 491-497.
Pacholewicz, E., Swart, A., Lipman, L.J., Wagenaar, J.A., Havelaar, A.H., Duim, B., 2013. Propidium
monoazide does not fully inhibit the detection of dead Campylobacter on broiler chicken carcasses
by qPCR. Journal of microbiological methods 95(1), 32-38.
Ricke, S.C., Feye, K.M., Chaney, W.E., Shi, Z., Pavlidis, H., Yang, Y., 2019. Developments in rapid
detection methods for the detection of foodborne Campylobacter in the United States. Frontiers in
Microbiology 9, 3280.
Russell, S.M., Doménech-Sánchez, A., de la Rica, R., 2017. Augmented reality for real-time detection
and interpretation of colorimetric signals generated by paper-based biosensors. Acs Sensors 2(6),
848-853.
Scallan Walter, E.J., Crim, S.M., Bruce, B.B., Griffin, P.M., 2020. Incidence of Campylobacter-
associated Guillain-Barre Syndrome estimated from health insurance data. Foodborne Pathogens and
Disease 17(1), 23-28.
Schrader, C., Schielke, A., Ellerbroek, L., Johne, R., 2012. PCR inhibitors–occurrence, properties and
removal. Journal of applied microbiology 113(5), 1014-1026.
Trinh, K.T.L., Stabler, R.A., Lee, N.Y., 2020. Fabrication of a foldable all-in-one point-of-care molecular
diagnostic microdevice for the facile identification of multiple pathogens. Sensors and Actuators B:
Chemical, 128057.
Vidic, J., Manzano, M., Chang, C.-M., Jaffrezic-Renault, N., 2017. Advanced biosensors for detection
of pathogens related to livestock and poultry. Veterinary research 48(1), 11.
Vidic, J., Vizzini, P., Manzano, M., Kavanaugh, D., Ramarao, N., Zivkovic, M., Radonic, V., Knezevic, N.,
Giouroudi, I., Gadjanski, I., 2019. Point-of-need DNA testing for detection of foodborne pathogenic
bacteria. Sensors 19(5), 1100.
Vizzini, P., Braidot, M., Vidic, J., Manzano, M., 2019. Electrochemical and optical biosensors for the
detection of campylobacter and listeria: An update look. Micromachines 10(8), 500.
WHO, 2020. World Health Organization. Campylobacter. Available online:
https://www.who.int/news-room/fact-sheets/detail/campylobacter.
Yang, X., Kirsch, J., Simonian, A., 2013. Campylobacter spp. detection in the 21st century: A review of
the recent achievements in biosensor development. Journal of microbiological methods 95(1), 48-56.

Table 1. Analytical parameters of methods for *Campylobacter spp.* detection.

Method	LOD	Analysis time	Reference
Culture-based methods	1 CFU / 25 g	≥ 5 days	(Vizzini et al. 2019)
Real-time PCR	3×10^3 CFU/mL	0.5 to 4 h	(Alves et al. 2016)
Dot blot	0.37 ng/µL	24h	(Manzano et al. 2015)
Colorimetric aptasensor	7.2×10^5 CFU/mL	30 min	(Kim et al. 2018)
Enhanced dot blot	3 pg/µL or 600 CFU	24h	This work

Figure legends:

Figure 1. (A) TEM images of Si-NPs before and after functionalization with linker and biotin. (B) Histogram showing the average particle size 50 ± 3 nm before functionalization. (C) Absorbance spectra of Si-NPs (before functionalization) and biotin-Si-NPs (after functionalization). (D) DLS size plots of native and biotin functionalized Si-NPs in water.

Figure 2. Schematic representation of *Campylobacter* detection based on paper-based DNA hybridization with a complementary biotinylated probe, and a streptavidin-HRP read-out through dot blot (A). The signal was amplified using highly functionalized biotin-Si-NPs instead of a single biotin (B).

Figure 3. (A) Dot blot detection of *Campylobacter* DNA sequence with biotin-Si-NP enhanced read-out. Note that no signal was obtained with a truncated *Campylobacter* sequence (PR) nor with an *E. coli* control sequence (PE). (B) Conventional dot blot detection of biotin labeled CampyP3 probe using dilution of complementary *Campylobacter* sequence CP3 ranging from 1ng/ µL to 78 fg/µL. (C) Enhanced dot blot detection of CP3 (0.1 ng/µL – 78 fg/µL) using CampyP3 probe. (D, E) Corresponding calibration curves were obtained by plotting the chemiluminescent signal intensity of dots as a function of CP3 template concentrations.

Figure 4. SEM images of the basic nylon membrane, nylon membrane functionalized with a *C. jejuni* DNA, and nylon membrane with CampyP3 hybridized with *C. jejuni* DNA and with biotin-Si-NPs. Note that the size of visualized circular beads of about 50 nm (yellow shading) corresponded to the estimated NP size in Fig 1B.

Figure 5. Inclusivity and exclusivity test results of the enhanced dot blot biosensor in pure

bacterial cultures. Signal intensity were normalized using an intern control (CP3 template).

Error bars represent the standard deviation of the mean from triplicates.

Figure 6. Schematic diagram of *Campylobacter* detection from chicken samples using the

enhanced dot blot and the official ISO 10272-1:2006 method. The ISO method involved steps

performed to obtain the results given in Table 4. Dot blot membrane shows representative dot

blot results obtained with naturally infected chicken samples (C3 and C10) and non-infected

chicken samples (C4, C5 and C6). CP3 template sequence (0.1 ng/µL) was used as a positive

control.

〰️	Target DNA	✖️	Streptavidine
〰️	DNA probe	🟠	HRP
🔶	Biotin	🔵	Si-NP

A

CP3

0.8*   0.4   0.2   0.1

PR

0.8   0.4   0.2   0.1

PE

0.8   0.4   0.2   0.1

B   Without NPs

1*   0.5   0.2   0.1

0.05   0.02   0.01   0.005

0.1*   0.05   0.025   0.0125

0.006   0.0031   0.0015   0.00078

C   With NPs

0.1*   0.05   0.025   0.0125

0.006   0.0031   0.0015   0.00078

D

Without NPs

$R^2 = 0.98684$

LOD = 0.08 ng/µl

CL light intensity (A.U.)

Target DNA (ng/µL)

E

With NPs

$R^2 = 0.98935$

LOD = 0.003 ng/µl

CL light intensity (A.U.)

Target DNA (ng/µL)

Nylon membrane

Target DNA

Biotin-Si-NPs-DNA-DNA

10 g of chicken meat

+

90mL Bolton broth

48 h incubation

DNA extraction

Dot blot

Culture on selective agar

48 h incubation

Sub-culture on Blood agar

48 h incubation

Confirmation test