Everyday sound exposure and the association with human heart rate: evidence

from real-world data logging

Authors

Jeppe H. Christensen

Gabrielle H. Saunders

Michael Porsbo

Niels H. Pontoppidan

[1]Eriksholm Research Centre, Oticon A/S, Snekkersten, Denmark

[2]Oticon A/S, Smørum, Denmark

[3]Manchester Centre for Audiology and Deafness, School of Health Sciences, University of

Manchester, Manchester, UK

Abstract

Prolonged acoustic noise negatively affects human cardiovascular health. However, noise is typically defined by acoustic intensity (i.e. pressure levels) alone, which neglect other dimensions of the sound associated with e.g. the ease of listening.

In this study, we used in-market data from hearing aid users to investigate the longitudinal association between multidimensional characteristics of everyday sound logged by the hearing aids and continuous mean heart rate measures logged by the user's own wearable.

We find that increases in ambient sound intensity are significantly related to increases in mean heart rates, corroborating prior laboratory and short-term real-world data. In addition, however, and not previously recognized, we find that increases in the ambient sound quality - that is, the difference between sound signal and noise - are associated with decreases in mean heart rates.

Our findings document a mixed influence of everyday sounds on cardiovascular stress, and that the relationship is more complex than is seen from examination of sound intensity alone. In addition, our findings highlight the importance of including environmental sounds to models predicting human physiology.

Keywords: real-world heart rates; data-logging; everyday sounds; linear mixed-effects models

Introduction

Humans constantly face a multifaceted soundscape. Some content is by choice (e.g. listening to music), while other is not (e.g. traffic noise when commuting), but despite environmental sounds being an integral part of everyday human life, there is very little knowledge about how aspects of the soundscape encountered in everyday life (beyond the intensity component, Flamme et al., 2012) affect the human body. There is, however, an increasing recognition of the harmful effects on human auditory (Kirchner et al., 2012; Kujawa, 2006) and cardiovascular systems from sounds classified as noise, in particular occupational noise, and noise exceeding certain intensity levels but below those known to cause physical damage to hair cells (Burns et al., 2016; Munzel et al., 2014). A recent report from the WHO summarizes the biological mechanism of cardiovascular reactions to noise (M. Nilsson, 2018). In brief, it suggests that auditory noise causes a direct (i.e. through subcortical connections of the brain) or indirect (involving projections from the auditory thalamus to the auditory cortex) stress reaction by disrupting the autonomic nervous system (ANS) balance due to elevated activity in the sympathetic nervous system branch (Belojevic et al., 2008; Björ et al., 2007; M. Nilsson, 2018; Robinson et al., 1966; Tzaneva et al., 2001). This can have acute effects that lead to momentarily elevated heart rate (HR) or blood pressure (Lusk et al., 2004), or cause long-term damage, such as elevating the risk for heart disease due to sustained periods of hypertension (Babisch, 2014).

A recent laboratory study by Shoushtarian et al., (2019) documented a causal link between immediate noise exposure and acute cardiovascular system reactions (Shoushtarian et al., 2019). The study found that noise segments exceeding intensities of 40 dBA significantly increase HRs from baseline within 2 – 5 seconds post exposure. This timing is in line with the dynamic response of sympathetic ANS modulation of HR (Berger et al., 1989), which is considerably slower than the fast acting parasympathetic control (Draghici & Taylor, 2016), suggesting that auditory noise

induces a "fight and flight" stress response in the sympathetic nervous system branch (Munzel et

al., 2014). The increase in HRs with noise was reported for intensities up to 90 dBA. However, the

reported increase in HRs between 65 dBA and 90 dBA was not significant, suggesting a possible

ceiling effect. Interestingly, the study by Shoushtarian et al., (2019) also revealed that low-intensity

noise near the threshold for hearing (15 dBA) had a dampening effect on HRs. That is, the low-

intensity noise resulted in a 2% reduction in mean HR from baseline levels measured in complete

quiet. The authors believe this effect might be due to a fast-acting orienting response caused by the

sound onset at low intensities (Graham & Clifton, 1966). Nevertheless, their finding suggests that

noise can have diverse effects on the ANS depending on its intensity.

Along those lines, Sim C. S., et al., (2015) reported that noise with equal intensity in decibels but of

different type (e.g. traffic noise, background noise, speech noise) had distinct effects on the ANS

measured by heart rate variability (HRV). Specifically, they argued that speech noise, besides

activating a sympathetic stress reaction, also activates the para-sympathetic nervous system leading

to a more stable balance of the ANS. In addition, effects of noise are not always perceivable. In a

study by Evans & Johnson (2000), cleric workers exposed to simulated open-office noise showed

increased levels of adrenaline, indicative of increased sympathetic levels, compared to a control

group, but did not rate perceive levels of stress higher than the control group.

Diverse ANS reactions to sound exposure have also been linked to psychological evaluation.

Umemura M. & Honda K., (1998) found that listening to classical music suppresses sympathetic

nervous activity compared to rock music or noise, and that the degree of suppression (measured by

Mayer Wave related Sinus Arrhythmia and respiratory sinus arrhythmia) positively correlated with

subjective reports of comfort (Umemura & Honda, 1998). On top of this, individual differences in

HR responses to unpleasant noise (high speed dental engine) have been explained by subjects'

familiarity and subjective experience with the acoustic stimuli (Gang & Teft, 1975), indicating a

certain adaptation to typically occurring noise sources in daily life.

Measurements of the association between human ANS balance and noise intensity are typically

performed under controlled conditions (e.g. in laboratories). This neglects the potential impact of

daily life contexts, which, as pointed out above, are expected to play a significant role in ANS

reactions. For example, individual sensitivity towards noise might fluctuate with time-of-day or

location (e.g. home versus at work) making it hard to generalize the causal effects found in

laboratories to real-world situations with complex sound immersion (Di Nisi et al., 1990).

Although, in a recent real-world study, El Aarboui & Chaix found that, over a 7-day period,

increases in sound intensity levels were associated with a concomitant increase in HR and HRV

parameters, but a delayed decrease in the overall HRV (El Aarbaoui & Chaix, 2019), confirming the

role of noise intensity in disrupting ANS stability. In addition, they found that the strength of the

association between sound intensity levels and changes in HR were significantly moderated by

place/mobility contexts, highlighting the relevance of real-world data. However, their study only

considered sound intensity, and as reported above, to understand how everyday sound immersion

objectively affect the human cardiovascular system (in particularly the ANS balance) and general

wellbeing, more dimensions of the sound must be considered (Can et al., 2016; El Aarbaoui &

Chaix, 2019). This notion has long been acknowledged in research of soundscapes (Schafer, 1993)

such as when city planning (M. E. Nilsson & Berglund, 2006) but has not yet been applied in the

study of human stress reactions to everyday sound immersion.

In this paper, we use real-world data to investigate how everyday ANS dynamics (quantified by

changes in mean HR) are associated with the sound environment. First, we describe the

characteristics of the typical sound environments faced by a population of hearing aid users, and

second, we model how characteristics of the sound environment are associated with changes in HRs

throughout the day.

The sound environment is described using four parameters: sound pressure level (SPLs), sound

modulation level (SMLs), signal-to-noise ratio (SNRs), and soundscape class (SS), representing

distinct characteristics of the momentary sound immersion. SPLs represent sound intensity and are

the most commonly used indicator of the sound wave strength. SPL correlates well with human

perception of loudness (Long, 2014). SMLs represent temporal amplitude modulation - that is, the

degree by which the sound wave amplitude oscillates over short-term intervals by e.g. speech

(Elliott & Theunissen, 2009). SNRs represent a spectral dimension of the sound by differentiating

between background noise levels and signal levels. Finally, SSs represent a qualitative dimension of

the sound related to how effortless it is to listen to speech-like sources.

Methods

Participants and Ethics

This study does not require ethical approval according to Danish law. Data were obtained from the

commercially available HearingFitness™ program (Laplante-Lévesque et al., 2017), which is

offered to users of Oticon A/S internet-connected hearing aids. Thus, participants in this study are

in-market hearing aid users. In total, we extracted real-world data from 98 random users from June

through December 2019. Given the real-world nature of the data and the privacy-preserved

extraction, no personal information characterizing the users (e.g. age, gender) are available.

However, the users represent a random sample of typical hearing aid users, thus, we expect that 6 in

10 are male with age centered around 74 years, which is the reported mean age of hearing aid

owners in 2008 (Kochkin, 2010).

Data sources and Apparatus

Users from the sample all wore Oticon Opn™ hearing aids paired with the Oticon ON™ iOS

smartphone app for remote control and logging of sound data from hearing aid microphones. Users

were selected for the study based on their using a smartwatches linked with the Apple HealthKit

app for automatic and continuous logging of 5-min mean HR.

The HearingFitness™ app stores data on a secure server. When signing up for the HearingFitness™

program via the app, users submit written consent agreeing to use of their anonymous data for

research purposes. . In addition, usersconsent to have their GPS location (as a relative measure in

meters between each sample) and health data from Apple Health Kit extracted and stored on the

secure servers. All data collection and storage are conducted in a "privacy by design" manner in

accordance with GDPR.

Variables collected

Besides HR data, sound environment data are logged by the smartphone-connected hearing aids.

Data consist of three continuous (acoustic data) and one discrete (soundscape) variable. The

continuous data represent acoustic characteristics of the momentary soundwave sensed by

calibrated hearing aid microphones at ear level. These variables are: SPLs, SMLs, SNRs all

estimated in a broadband frequency range (0-10kHz) in decibel units (see Figure 1B). SPL is the

level output estimate from a low-pass infinite impulse-response filter with a time constant of 63ms.

SML is then derived as the difference between a top- and bottom tracker (peak- and valley detector)

of the SPL. The bottom tracker is implemented with a slow dynamical attack time of 1s to 5s and a

fast release time of 30ms and the top tracker is implemented with the reverse (see Figure 10-3 in

Kates, 2008). Finally, SNR is calculated as the difference between the bottom tracker and the

immediate SPL. Thus, values of SPL and SNR are dynamically changing on same time scale

whereas the SML changes with a slower time constant of up to 5 seconds. The discrete soundscape

variable classifies the momentary sound environment into four soundscapes by a proprietary

hearing aid algorithm using SPL, SML and SNR values. The soundscapes are: "Quiet"; "Noise";

"Speech"; and "Speech in Noise". These soundscapes have been used in the standard fitting

software for Oticon Hearing Aids since 2004, and similar soundscape classification from hearing

aids have been reported as a potential source for data mining (Mellor et al., 2018). Figure 2 shows

distributions of soundscapes against each of the three continuous acoustic data variables.

Data logging

Sound environment data are logged every 60 seconds together with timestamps when the user's

hearing aids are turned on and connected via Bluetooth to an associated smartphone. In addition, the

5-min running mean HR is logged approximately every 7 minutes (observed mean = 7.6 minutes,

SD = 2.8 minutes) when a wearable with a pulse sensor is connected to the smartphone.

Data extraction and filtration

We extracted data from 98 users from which we obtained 1,144,785 sound environment data logs,

which represents approximately 10,000 hours of bilateral hearing aid use, and 616,893 heart rate

logs, which represents approximately 72,000 hours of data from a wearable. We only considered

data logged between 05:00 and 24:00 to avoid confounds from night-time logs that are likely

collected when neither the hearing aids nor wearables were being worn, Further, we only included

users for whom there were at least 100 HR logs.

We examined data in two steps. First, to investigate general sound exposure, we used all the sound

data regardless of temporal overlap with HR logs. This was done in order to maximize the data

available. Second, for subsequent association analysis, we pre-processed the data to ensure full

overlap between sound and HR logs. The pre-processing consisted of selecting time windows of

five minutes prior to each HR log and computing the average of each sound data variable within

that window. Thus, the fully overlapping data consists of data records with the value of each sound

data variable estimated from the same time window as the running mean HR logs. Note that since

HRs are not uniformly distributed, we excluded data records with extreme low (<5% quantile) and

high (>95% quantile) HR to avoid potential confounds from low incident HRs. This exclusion of

extreme data also ensured normality of residuals from statistical modeling (see comparison in

Supplementary Information) but did not affect the order of regression coefficients (Table 1) or the

statistical significance of the included statistical models. After pre-processing, our data consisted of

24677 data records from 52 users and 917 unique user days. Figure 3A shows the count of data

records separated by time-of-day and weekday, and Figure 3B shows the density distributions of

each variable in the data records.

The effective sampling period of data records was dependent on the proportion of time that a user

simultaneously wore both hearing aid(s), smart phone, and wearable. Furthermore, users wore their

hearing aids for different durations throughout the. Thus, we estimated the *effective* sampling

frequency (i.e. representativeness of samples across time) by computing the average amount of data

records each hour between 5:00 and 24:00 per user across the period covered in the data. The grand

average sampling frequency across all 917 user days and hours was 2.2 data records per hour (SD =

2.5) with a peak at 18:00 (2.3 data records per hour, SD = 1.8).

Adjusting for movement

A subset of the dataset containing GPS information associated with each pulse measure was used to

assess the possible confounding effect of movement activity on the relationship between heart rates

and sound environment data. Movement activity was estimated within the database and extracted

for analysis. Specifically, the distance in meters between two consecutive latitude and longitude

coordinates was computed using the haversine method. That is, for each pair of subsequent latitude

(φ_1, φ_2) and longitude (λ_1, λ_2) coordinates the distance between them in meters, d, is estimated as:

195
$$d = 2r\arcsin\left(\sqrt{\sin^2\left(\frac{\varphi_2-\varphi_2}{2}\right) + \cos(\varphi_1)\cos(\varphi_2)\sin^2\left(\frac{\lambda_2-\lambda_1}{2}\right)}\right),$$

196   with r being equal to the radius of the earth.

Statistical analysis

Due to the unbalanced (unequal samples per user, day, hour etc.) and hierarchical multi-level nature

of the data records, we quantified associations between variables using linear mixed-effects models.

The autocorrelation of residuals was addressed with a first-order autoregressive structure, which

have previously been shown to improve the goodness-of-fit with similar data (El Aarbaoui & Chaix,

2019), while the random-effects structure adjusted for individual differences in both baseline HR

and sensitivity towards each fixed-effect (intercept and slopes), time-of-day (intercepts), and

weekday (intercepts) (Barr et al., 2013). We used R for statistical analysis and visualizations and

used the "nlme" package for implementing linear mixed-effects models. Model diagnostics (see

Supplementary Information) were conducted in accordance with recommendations in the literature

(Harrison et al., 2018), and we ensured that models adhered to assumptions of normality of

residuals and homogeneity of variance among random-effect groups.

Results

Data validation

The most occurring (median) SPL across all users were 54.86 dB (SD = 5.74 dB), which

corresponds to a level just below normal conversational speech (~60 dB). The sound exposure

varied according to the soundscape classified by the hearing aid (see Figure 1 and 2). For example,

as expected, "speech" environments consisted of higher values of raw SNR and SML, while

"speech in noise" and "noise" were the loudest (highest SPL). This suggests that the classification

of soundscape could be reasonably trusted.

Figure 1: Quantile distributions of raw acoustic data.

We also examined the variation in sound data as a function of time of day. In Figure 2A, each

acoustic characteristic is plotted as standard scores, and Figure 2B shows the relative occurrence of

each soundscape. As would be expected, the SPL is lowest late at night and early in the morning

and the proportion of time the sound environment is classified as being 'quiet' is also greatest late at

night and early in the morning. Further, the proportion of time sound is classified being 'speech' is

greatest in the evening, corresponding to the increase in SML in the evening, again, this is to be

expected based on the findings of other studies (Humes et al., 2018) and on real life situations.

Figure 2: Typical daily sound exposure (A) and most often occurring soundscape (B). Computed as median values across all data for

each hour and user and then subsequently taking the median across all users. Shaded area represents the SEM.

Sound environment characteristics fluctuating with time most likely reflects different everyday

contexts (e.g. multi-modal peaks in SPL and increased SML and SNR in the evening). In addition,

Figure 2B suggests that, besides the hourly fluctuations in soundscape occurrence, approximately

65% – 70% of the day is spent with "quiet" or "speech" soundscapes. For comparison, Humes, L. E.

et al., (2018), found that older adults with hearing aids typically spent 60% of their time in "quiet"

or "speech-only" conditions, while around 10% of time is being spent in "pure noise". In summary,

the patterns identified in our sound data document their real-world characteristics and validity.

The grand mean HR was 74.88 bpm (SD = 9.19 bpm). In comparison, real-world HR norms of

66788 individuals spanning all ages above 18 years have been documented to be 79.1 bpm (SD =

14.5), while stratification with age in the range 71 to 80 resulted in a mean HR of 74.2 (SD = 11.1),

indicating that users in the current study are indeed of older age (Avram et al., 2019), suggesting

that users in the current study are likely to be in this age range. This also agrees with the expected

age of the majority of hearing aid users (Kochkin, 2010).

Association between sound and heart rate

Figure 3 shows summary distributions of the data in the data records used for regression analysis).

Figure 33A indicates a close to uniform sampling of observations from 10:00 to 20:00 and across

weekdays except for Friday and Sunday, which exhibits approximately 500 - 1000 less records than

the rest of the week. When averaged across all users, 38% (SD = 20) of data records were

registered as belonging to "quiet", 28% (SD = 17) belonged to "speech", 17% (SD = 13) belonged

to "speech in noise", and 17% (SD = 12) to "noise" environments.

Figure 3: Distribution of combined observations separated by soundscape class and pooled among all users. In A), the total counts

of observations are shown across each hour-of-day (left) and weekday (right) for each soundscape class (colors). Note that one

observation entails one heart rate measure associated with time-averages of the preceding sound pressure level, signal to noise level,

sound modulation level, and soundscape class. In B), density distributions of each of the continuous acoustic variables and heart rate

measures are shown.

Density distributions of the continuous acoustic parameters associated with each data record are

shown in Figure 33B, three first panels. Not surprisingly, the sound intensity (SPL) were highest for

"speech in noise" and "noise" soundscape classes with median SPL L_{eq} equal to 72.3 and 66.2 dBs,

respectively. For comparison, the same value in "quiet" was 49.3 dB. On the other hand, "quiet"

and "noise" soundscapes exhibited lower median signal to noise ratio (6.0 and 3.9) and modulation

levels (15.7 and 11.0) compared to "speech" (SNR: 12.5, SML: 21.6). The last panel of Figure 3B

shows the distribution of HR across all records (excluding the bottom and top 5% most extreme).

The grand average HR was 75.8 bpm (SD = 6.5), and separated by soundscape the mean HR was

74.6 bpm (SD = 6.9) for "quiet", 75.9 bpm (SD = 7.0) for "speech", 78.2 bpm (SD = 8.6) for

"speech in noise", and 77.6 bpm (SD = 6.9) for "noise". A repeated measures ANOVA shows these

to differ significantly across soundscapes ($F(1.92,94.26) = 10.96$, $p = 0.041$) with post hoc

comparisons using Bonferroni correction indicating that the mean heart rate for "quiet" was

significantly lower than for "speech" ($p = 0.013$), "speech in noise" ($p < 0.001$), and "noise" ($p <$

0.001); the mean heart rate for "speech" was significantly lower than for "speech in noise" ($p =$

0.041), and "noise" ($p = 0.019$) while heart rates for "speech in noise" and "noise" did not differ.

Marginal mean HR versus increasing levels of each acoustic characteristic are shown in Figure 4.

The marginal mean HR was computed by first standardizing each user's HR and acoustic data

(centering and scaling) and then computing the pooled average HR within non-overlapping bins for

quantiles of SPL, SNR and SML (see Figure 5). Since high levels of SNR and SML can occur for

both low and high levels of SPL (see Figure 6), we computed the marginal mean HR versus SML

and SNR in two regions of SPL: SPLs below and above the overall median (i.e. 60 dB SPL).

Figure 4: Marginal means of heart rates separated by each acoustic characteristic. Solid lines represent best fitting linear

regression. The marginal means were computed separately for observations within quantiles (width 5%) of the acoustic

characteristics SPL (A), SNR (B-C), and SML (D-E). In addition, prior to computation of the marginal means, SPL, SNR, SML, and

heart rate data were standardized (centered and scaled) for each user individually. In B-E, marginal mean heart rate for increasing

levels of SNR and SML are shown for two regions of SPL: low intensity (i.e. SPL < 60 dB) and high intensity (SPL > 60 dB) sound

pressure levels, with the threshold based on the overall median observed SPL (60 dB).

The marginal means and fitted linear trends in Figure 4 indicate a strong association between

acoustic characteristics and HRs. In addition, Figure 4 B-E reveals that the association between

derived moments of the sound wave (i.e. SNR and SML) and mean HR is conditioned on the

intensity. That is, SNR seems to be more strongly associated with HR in conditions with high SPL

compared to low SPL conditions, whereas SML is either positive or negatively associated with HR

again depended on the intensity (i.e. SPL).

As stated in the introduction, everyday sounds are not only modulated by intensity but also by

dynamic characteristics (i.e. modulated or not), perceptual quality (i.e. noisy or clean signals), and

type (e.g. conversation, music, traffic). These distinct traits can be approximated by SPL, SML,

SNR, and soundscape, respectively. However, the different sound dimensions inherently correlate.

For example, high intensity speech signals will also naturally yield a high sound modulation level,

and if not masked by noise, also a high SNR. Thus, we assessed the mutual relationships between

all sound environment variables to establish a baseline of potential covariance between predictors in

statistical modelling. In Figure 55, 2D density distributions as contours are shown for the mutual

relation between SPL, SML, SNR, and soundscape class. The most correlated parameters are the

SML and SNR (Pearsons' correlation: $r(24675) = 0.72$, $p < 0.001$) since SNR rarely exceeds SML.

On the other hand, a wide range of SNR values are observed below any given SML value due to the

faster temporal dynamics in the SNR estimate (see *Methods*). The next most correlated relationship

was between SPL and SNR (Pearsons' correlation: $r(24675) = 0.37$, $p < 0.001$) while SPL and SML

correlated the least (Pearsons' correlation: $r(24675) = 0.14$, $p < 0.001$).

Figure 5: 2D density distributions as contours of each acoustic variable associated with each heart rate log..

Given the high degree of clustering of the continuous acoustic data with soundscape (see Figure 5),

we first included the two types of data as independent variables in separate models to predict pulse

rates. Thus, the continuous acoustic parameters (SPL, SML, and SN) or the soundscape (contrasted

to "Quiet") were included as fixed effects while controlling for individual baseline offsets, time-of-

day offsets, and weekday offsets with random intercepts. In addition, when estimating the

continuous parameters, individual sensitivity towards each acoustic characteristic were controlled

for by random slopes for each user. Furthermore, we fitted the models to a subset of the

observations which included GPS data. By this, we investigated the influence of controlling for

movement activity estimated from relative GPS coordinates (see next section).

316   *Table 1: Regression coefficients and associated summary statistics for models predicting heart rates (bpm).*

	All obs. (n = 24.677)		W. movement (n = 5.991)	
	β	*95% CI*	β	95% CI
Soundscape model				
S vs Q	+0.94	[+0.68 to +1.19]	+1.33	[+0.81 to +1.84]
SN vs Q	+1.82	[+1.48 to +2.15]	+1.96	[+1.32 to +2.59]
N vs Q	+2.04	[+1.70 to +2.39]	+2.20	[+1.44 to +2.96]
Acoustic data model				
SPL	+1.31	[+1.07 to +1.55]	+1.33	[+0.88 to +1.77]
SML	+0.57	[+0.37 to +0.77]	+0.47	[+0.10 to +0.85]
SNR	-0.85	[-1.04 to -0.66]	-0.89	[-1.21 to -0.57]

S: Speech; Q: Quiet; SN: Speech in Noise; N: Noise

Inclusion of sound data to predict heart rates significantly improved model goodness-of-fit when

compared to intercept-only models with Likelihood-ratio tests (acoustic data: χ^2 (6) = 474, p <

0.001; soundscape data: χ^2 (3) = 174.62, p < 0.001), confirming a significant association between

changes in mean HR and the sound environment. This effect was not caused by motion (i.e. sound

and HR changing in response to latent physical activity). Models adjusted for the momentary

movement activity of users were able to explain additional 32.8% of the HR variance (see

Supplementary Information), but had similar regression coefficient values as the un-adjusted

models with highly overlapping confidence intervals (see Table 1) indicating that the additional

explained variance did not covary with sound data.

The regression coefficients for the soundscape model confirmed the trend present in Figure 3B last

panel; HR increase with the complexity of the soundscape going from quiet, to speech, and then

speech in noise and noise. Thus, despite speech in noise being the loudest soundscape (Figure 1A),

HRs were overall higher in soundscapes classified as pure noise.

The acoustic data model revealed distinct associations of the HRs with the different acoustic

characteristic. Specifically, increasing SPL and SML predicts increasing HRs whereas increasing

SNR predicts decreasing HRs.

Moderating effect of soundscape

The outcomes presented in Table 1 suggest that changes in mean HR throughout the day are

significantly associated with changes in everyday sounds, and that overall mean HRs are lower in

less complex sound environments (e.g. "Quiet" versus "Speech in Noise"). Here, we investigated to

what extend the soundscape, classified by the hearing aids, moderated the strength of the

association between heart rate and the continuous predictors. That is, were users more sensitive

towards acoustic characteristics in certain soundscapes? We would expect this to be true since

differences among soundscapes proxy differences in listening conditions. For instance, less effort is

needed to understand speech in "quiet" and "speech" soundscapes than in a "noise" soundscape.

Prior to analysis we collapsed across "Speech in Noise" and "Noise" since they were highly

overlapping in terms of SPL and SNR (see Figure 5B) and contained the fewest observations

(Figure 3A). Thus, the resulting three soundscapes are "Quiet", "Speech", and "Noisy" listening

conditions, reflecting increasing levels of soundscape complexity.

We fitted an interaction model with each acoustic parameter being allowed to interact with

soundscape while keeping the same random-effects variance structure as before. For this model, we

used all observations and did not include movement as a random-effect term since the additional

degrees-of-freedom would otherwise render the model un-identifiable.

The interaction model produced better predictions than both the soundscape model (dAIC = 270.5,

chi2(11) = 292.54, p < 0.001) and the acoustic data model (dAIC = 12.6, chi2(8) = 28.6, p < 0.001).

Despite the potentially high covariance between soundscape and acoustic data (see Figure 5), the

largest variance inflation factor was estimated to be 1.79. The main effects replicated the outcomes

of fitting the acoustic data and soundscape models separately (see Table 1). That is, higher HR for

more complex soundscape ($\chi^2(2) = 8.03$, $p = 0.018$), more intense and modulated sound

environments (SPL: $\chi^2(1) = 58.8$, $p < 0.001$; SML: $\chi^2(1) = 27.1$, $p < 0.001$), and environments with

more background noise (SNR: $\chi^2(1) = 53.5$, $p < 0.001$). Additionally, all interactions between

soundscape and SPL, SML, and SNR were significant ($\chi^2(2) = 6.7$, $p = 0.035$; $\chi^2(2) = 31.2$, $p <$

0.001; $\chi^2(2) = 10.7$, $p = 0.005$, respectively).

Figure 6 shows each level of the three interactions between acoustic variables and soundscape. It

indicates that interactions were driven by differences in strength of association between

Quiet/Speech and Noisy soundscapes. Thus, observed HRs were more strongly associated with

sound intensities (SPL), modulation levels (SML), and signal quality (SNR) when the soundscapes

were optimal for listening.

Figure 6: Regression lines from LME interaction model. Shaded area represents the standard error around each prediction.

We assessed the significance of each pairwise interaction by contrasting each coefficient (i.e. slope)

with the Quiet soundscape as baseline.

For SPLs (Figure 6A), the slope for "Noisy" soundscape was significantly different (β=-1.38, SE =

0.28, p < 0.0001) whereas the slope for "speech" did not differ from the slope for "quiet" (β=-0.12,

SE = 0.27, p = 0.65). For SNRs (Figure 6B), only the slope for "Noisy" was significantly higher

than for "Quiet" (β=0.97, SE = 0.26, p < 0.001). Finally, for SMLs (Figure 6C), slopes for both for

"Speech" and "Noisy" were significantly lower than for "Quiet" (β=-0.53, SE = 0.18, p = 0.0037;

β=-0.92, SE = 0.23, p < 0.001).

Discussion

Using real-world data from in-market hearing aids we were able to investigate the association

between exposure to sound and short-term (i.e. within 5 minutes) changes in HR. The data reflect

hearing aid use over several weeks and as such represents the sound exposure from everyday life

activities (see Figure 1 and 2), with characteristics of intensity and soundscape occurrences that

agree with previous research (Flamme et al., 2012; Humes et al., 2018; Wu et al., 2018).

We find that characteristics of the sound environment are significantly associated with changes in

heart rate. Specifically, higher SPLs and SMLs are associated with increased HRe, whereas more

favourable (higher) SNRs are associated with lowered HR. In addition, all associations are stronger

in simple listening soundscapes (Quiet and Speech) compared to soundscapes classified as

containing noise, while marginal means reveal that HR moderation by SMLs and SNRs are distinct

depending on the sound intensity (see Figure 4). Specifically, marginal means suggest that the

negative association between SNR and mean HR is most pronounced in louder sound environments

where the SPL exceeds 60 dB SPL, whereas sound modulation seems to be more strongly

associated with increases in HR at intensity levels below 60 dB SPL.

SPLs are directly associated with loudness perception (Long, 2014) and have been found to induce

activation of the sympathetic branch of the human ANS (El Aarbaoui & Chaix, 2019; Munzel et al.,

2014; Shoushtarian et al., 2019). Our study suggests that even everyday SPLs (i.e. levels well

below those typically defined as hazardous to the auditory system) also affect heart rates. Thus, we

speculate that the sympathetic ANS branch is moderated by sound pressure - regardless of level of

intensity.

We also document a positive, albeit smaller, association between HR and SML. SML indicates how

modulated a sound wave is. Highly modulated sound is typically characterized by fast oscillating

SPLs which are indicative of e.g. speech or music. We speculate that the positive association

between SMLs and HRs are, to some extent, caused by task demands (Mackersie & Calderon-

Moultrie, 2016). That is, in highly modulated sound environments, listening and speech demands

are increased, which leads to slightly increased sympathetic ANS activity. This speculation is

further corroborated by the fact that the positive association between SML and HR is stronger for

low-intensity simpler sound environments (i.e. Figure 4D and 6C), which are indicative of task-

related activities, compared to loud and noisy environments.

We also document a negative association between real-world HRs and SNRs (see Table 1 and

Figure 4 B - C). This finding corroborates laboratory research within the hearing sciences, which

shows that difficult listening conditions (e.g. characterized by low SNRs) increases the listening

effort needed for speech understanding, and thus, either elevates the sympathetic stress levels or

decrease the parasympathetic ANS activity seen as increased pupil dilation (Wendt et al., 2017;

Zekveld et al., 2018) or skin conductance (Mackersie & Calderon-Moultrie, 2016) and decreased

HR variability (Seeman & Sims, 2015). This suggests that the ambient SNR, as measured by the

hearing aids, could proxy as a real-life indicator of momentary and contextual listening difficulty,

while associated changes in HR might indicate the level of mobilized listening effort.

The sound data used in this study are measured by commercial hearing aid devices. These devices

are typically optimized for low power consumption, which typically would entail lower resolution

and accuracy of sound estimators. Despite this, summary statistics of our data agree with other

studies using devices specifically developed for research purposes (El Aarbaoui & Chaix, 2019).

This highlights a potential for exploiting commercial devices, and in particularly hearing aids, for

establishing real-world evidence regarding human behavioural and biological reactions to

environmental stimuli. For example, future studies using a similar setup could investigate how

sensitivity towards sound exposure potentially differ between populations due to underlying health

pathologies. This is further corroborated by the fact that previous research has documented the use

of similar hearing aid data logging (Christensen et al., 2019) for policy-making within the hearing

or public health domains (Saunders et al., 2020).

Limitations

The users in our dataset are all wearing hearing aids daily thus, they all have some degree of

hearing loss and therefore do not represent the general population. Further, the mean age in the

group is likely skewed towards older age. Some studies suggest that the older hearing-impaired

people spend more time in quiet sound environments than younger and normal hearing individuals

(Wu & Bentler, 2012). Moreover, their sensitivity towards noisy and loud sound environments are

expected to differ compared to normal hearing individuals. Finally, the data assessed in the current

study is considered real-world in nature, which means that there was no direct control over who

used what devices, when, and for how often. However, by both applying filtration during data

extraction and pre-processing and using mixed-effects modelling, we minimized the impact of

unbalanced observations.

437   Conclusion

This study is the first to assess longitudinal real-world data of human HRs and multidimensional

sound exposure for potential associations. Our results corroborate previously presented laboratory

findings (Shoushtarian et al., 2019) and short-term real-world measures (El Aarbaoui & Chaix,

2019), which both indicate that ambient sound intensity is positively associated with heart rates. In

addition, we document that the real-world ambient signal-to-noise levels are associated with

lowered HRs, suggesting a destressing effect on the human cardiovascular system in conditions

assumed to reduce the auditory perceptual load and listening effort (Howard et al., 2010; Picou et

al., 2016; Wendt et al., 2017; Zekveld et al., 2018). This finding is furthermore supported by a

documented effect of soundscape on the strength of the association between sound characteristics

and ANS reactions. That is, in favorable listening conditions, acoustic characteristics have the

strongest association with changes in HR. In summary, our findings suggest a possible mixed

influence of everyday sounds on cardiovascular stress, and that the relationship is more complex

than is seen from examination of sound intensity alone. Furthermore, our findings highlight the

importance of including exposure to ambient sounds to models predicting human physiology.

Lastly, our study demonstrates the applicability of data logging using commercially available

devices (here, hearing aids) to establish ecologically valid evidence of everyday sound exposure and

human physiological reactions to it.

References

Avram, R., Tison, G. H., Aschbacher, K., Kuhar, P., Vittinghoff, E., Butzner, M., Runge, R., Wu,

N., Pletcher, M. J., Marcus, G. M., & Olgin, J. (2019). Real-world heart rate norms in the

Health eHeart study. *Npj Digital Medicine*, *2*(1), 58. https://doi.org/10.1038/s41746-019-

0134-9

Babisch, W. (2014). Updated exposure-response relationship between road traffic noise and

coronary heart diseases: A meta-analysis. *Noise & Health*, *16*(68), 1–9.

https://doi.org/10.4103/1463-1741.127847

Barr, D. J., Levy, R., Scheepers, C., & Tily, H. J. (2013). Random effects structure for confirmatory

hypothesis testing: Keep it maximal. *Journal of Memory and Language*, *68*(3), 255–278.

https://doi.org/10.1016/j.jml.2012.11.001

Belojevic, G., Jakovljevic, B., Stojanov, V., Paunovic, K., & Ilic, J. (2008). Urban road-traffic noise

and blood pressure and heart rate in preschool children. *Environment International*, *34*(2),

226–231. https://doi.org/10.1016/j.envint.2007.08.003

Berger, R. D., Saul, J. P., & Cohen, R. J. (1989). Transfer function analysis of autonomic

regulation. I. Canine atrial rate response. *American Journal of Physiology-Heart and*

Circulatory Physiology, *256*(1), H142–H152.

https://doi.org/10.1152/ajpheart.1989.256.1.H142

Björ, B., Burström, L., Karlsson, M., Nilsson, T., Näslund, U., & Wiklund, U. (2007). Acute effects

on heart rate variability when exposed to hand transmitted vibration and noise. *International*

Archives of Occupational and Environmental Health, *81*(2), 193–199.

https://doi.org/10.1007/s00420-007-0205-0

Burns, K., Sun, K., Fobil, J., & Neitzel, R. (2016). Heart Rate, Stress, and Occupational Noise

Exposure among Electronic Waste Recycling Workers. *International Journal of*

Environmental Research and Public Health, *13*(1), 140.

https://doi.org/10.3390/ijerph13010140

Can, A., Aumond, P., Michel, S., DE COENSEL, B., Ribeiro, C., Botteldooren, D., &

LAVANDIER, C. (2016). Comparison of noise indicators in an urban context. *Inter-Noise*

2016, 45th International Congress and Exposition of Noise Control Engineering, 9p.

https://hal.archives-ouvertes.fr/hal-01373857

Christensen, J. H., Rossing, R., Pontoppidan, N. H., Anisetti, M., Bamiou, D.-E., Spanoudakis, G.,

Murdin, L., Bibas, A., Kikidis, D., Dimakopoulos, N., Giotis, G., & Economou, A. (2019).

Fully synthetic longitudinal real-world data from hearing aid wearers for public health

policy modeling. *Frontiers in Neuroscience, Auditory Cognitive Neuroscience*, *13*.

https://doi.org/10.3389/fnins.2019.00850

Di Nisi, J., Muzet, A., Ehrhart, J., & Libert, J. P. (1990). Comparison of Cardiovascular Responses

to Noise During Waking and Sleeping in Humans. *Sleep*, *13*(2), 108–120.

https://doi.org/10.1093/sleep/13.2.108

Draghici, A. E., & Taylor, J. A. (2016). The physiological basis and measurement of heart rate

variability in humans. *Journal of Physiological Anthropology*, *35*(1), 22.

https://doi.org/10.1186/s40101-016-0113-7

El Aarbaoui, T., & Chaix, B. (2019). The short-term association between exposure to noise and

heart rate variability in daily locations and mobility contexts. *Journal of Exposure Science &*

Environmental Epidemiology. https://doi.org/10.1038/s41370-019-0158-x

Elliott, T. M., & Theunissen, F. E. (2009). The Modulation Transfer Function for Speech

Intelligibility. *PLoS Computational Biology*, *5*(3), e1000302.

https://doi.org/10.1371/journal.pcbi.1000302

Evans, G. W., & Johnson, D. (2000). Stress and open-office noise. *Journal of Applied Psychology*,

85(5), 779–783. https://doi.org/10.1037/0021-9010.85.5.779

Flamme, G. A., Stephenson, M. R., Deiters, K., Tatro, A., van Gessel, D., Geda, K., Wyllys, K., &

McGregor, K. (2012). Typical noise exposure in daily life. *International Journal of*

Audiology, *51*(sup1), S3–S11. https://doi.org/10.3109/14992027.2011.635316

Gang, M. J., & Teft, L. (1975). Individual Differences in Heart Rate Responses to Affective Sound.

Psychophysiology, *12*(4), 423–426. https://doi.org/10.1111/j.1469-8986.1975.tb00016.x

Graham, F. K., & Clifton, R. K. (1966). Heart-rate change as a component of the orienting

response. *Psychological Bulletin*, *65*(5), 305–320. https://doi.org/10.1037/h0023258

Harrison, X. A., Donaldson, L., Correa-Cano, M. E., Evans, J., Fisher, D. N., Goodwin, C. E. D.,

Robinson, B. S., Hodgson, D. J., & Inger, R. (2018). A brief introduction to mixed effects

modelling and multi-model inference in ecology. *PeerJ*, *6*, e4794.

https://doi.org/10.7717/peerj.4794

Howard, C. S., Munro, K. J., & Plack, C. J. (2010). Listening effort at signal-to-noise ratios that are

typical of the school classroom. *International Journal of Audiology*, *49*(12), 928–932.

https://doi.org/10.3109/14992027.2010.520036

Humes, L. E., Rogers, S. E., Main, A. K., & Kinney, D. L. (2018). The Acoustic Environments in

Which Older Adults Wear Their Hearing Aids: Insights From Datalogging Sound

Environment Classification. *American Journal of Audiology*, 1.

https://doi.org/10.1044/2018_AJA-18-0061

Kates, J. M. (2008). *Digital hearing aids*. Plural Publ.

Kirchner, D. B., Evenson, E., Dobie, R. A., Rabinowitz, P., Crawford, J., Kopke, R., & Hudson, T.

W. (2012). Occupational noise-induced hearing loss: ACOEM task force on occupational

hearing loss. *Journal of Occupational and Environmental Medicine*, *54*(1), 106–108.

http://citeseerx.ist.psu.edu/viewdoc/download?doi=10.1.1.593.6717&rep=rep1&type=pdf

Kochkin, S. (2010). MarkeTrak VIII: The efficacy of hearing aids in achieving compensation equity

in the workplace. *The Hearing Journal*, *63*(10), 19–28.

https://doi.org/10.1097/01.HJ.0000389923.80044.e6

Kujawa, S. G. (2006). Acceleration of Age-Related Hearing Loss by Early Noise Exposure:

Evidence of a Misspent Youth. *Journal of Neuroscience*, *26*(7), 2115–2123.

https://doi.org/10.1523/JNEUROSCI.4985-05.2006

Laplante-Lévesque, A., Pontoppidan, N. H., Mazevski, A., Schum, D., Behrens, T., & Porsbo, M.

(2017). *Data-driven hearing care with HearingFitness^{TM}: Oticon shares openly its*

innovative vision and roadmap (White Paper No. 27896UK; pp. 1–7). Oticon A/S.

Long, M. (2014). Fundamentals of Acoustics. In *Architectural Acoustics* (pp. 39–79). Elsevier.

https://doi.org/10.1016/B978-0-12-398258-2.00002-7

Lusk, S. L., Gillespie, B., Hagerty, B. M., & Ziemba, R. A. (2004). Acute Effects of Noise on

Blood Pressure and Heart Rate. *Archives of Environmental Health: An International*

Journal, *59*(8), 392–399. https://doi.org/10.3200/AEOH.59.8.392-399

Mackersie, C. L., & Calderon-Moultrie, N. (2016). Autonomic Nervous System Reactivity During

Speech Repetition Tasks: Heart Rate Variability and Skin Conductance. *Ear and Hearing*,

37, 118S-125S. https://doi.org/10.1097/AUD.0000000000000305

Mellor, J., Stone, M. A., & Keane, J. (2018). Application of Data Mining to a Large Hearing-Aid

Manufacturer's Dataset to Identify Possible Benefits for Clinicians, Manufacturers, and

Users. *Trends in Hearing*, *22*, 233121651877363.

https://doi.org/10.1177/2331216518773632

Munzel, T., Gori, T., Babisch, W., & Basner, M. (2014). Cardiovascular effects of environmental

noise exposure. *European Heart Journal*, *35*(13), 829–836.

https://doi.org/10.1093/eurheartj/ehu030

Nilsson, M. (2018). *Biological mechanisms related to cardiovascular and metabolic effects by*

environmental noise.

Nilsson, M. E., & Berglund, B. (2006, December). *Soundscape Quality in Suburban Green Areas*

and City Parks [Text].

https://www.ingentaconnect.com/content/dav/aaua/2006/00000092/00000006/art00008

Picou, E. M., Gordon, J., & Ricketts, T. A. (2016). The Effects of Noise and Reverberation on

Listening Effort in Adults With Normal Hearing: *Ear and Hearing*, *37*(1), 1–13.

https://doi.org/10.1097/AUD.0000000000000222

Robinson, B. F., Epstein, S. E., Beiser, G. D., & Braunwald, E. (1966). Control of Heart Rate by the

Autonomic Nervous System: Studies in Man on the Interrelation Between Baroreceptor

Mechanisms and Exercise. *Circulation Research*, *19*(2), 400–411.

https://doi.org/10.1161/01.RES.19.2.400

Saunders, G. H., Christensen, J. H., Gutenberg, J., Pontoppidan, N. H., Smith, A., Spanoudakis, G.,

& Bamiou, D.-E. (2020). Application of Big Data to Support Evidence-Based Public Health

Policy Decision-Making for Hearing: *Ear and Hearing*, 1.

https://doi.org/10.1097/AUD.0000000000000850

Schafer, R. M. (1993). *The soundscape: Our sonic environment and the tuning of the world*. Simon

and Schuster.

Seeman, S., & Sims, R. (2015). Comparison of Psychophysiological and Dual-Task Measures of

Listening Effort. *Journal of Speech, Language, and Hearing Research*, *58*(6), 1781–1792.

https://doi.org/10.1044/2015_JSLHR-H-14-0180

Shoushtarian, M., Weder, S., Innes-Brown, H., & McKay, C. M. (2019). Assessing hearing by

measuring heartbeat: The effect of sound level. *PLOS ONE*, *14*(2), e0212940.

https://doi.org/10.1371/journal.pone.0212940

Tzaneva, L., Danev, S., & Nikolova, R. (2001). Investigation of noise exposure effect on heart rate

variability parameters. *Central European Journal of Public Health*, *9*(3), 130–132.

Umemura, M., & Honda, K. (1998). Influence of music on heart rate variability and comfort—A

consideration through comparison of music and noise. *Journal of Human Ergology*, *27*(1–

2), 30–38.

Wendt, D., Hietkamp, R. K., & Lunner, T. (2017). Impact of noise and noise reduction on

processing effort: A pupillometry study. *The Journal of the Acoustical Society of America*,

141(5), 4040–4040. https://doi.org/10.1121/1.4989330

Wu, Y.-H., & Bentler, R. A. (2012). Do Older Adults Have Social Lifestyles That Place Fewer

Demands on Hearing? *Journal of the American Academy of Audiology*, *23*(9), 697–711.

https://doi.org/10.3766/jaaa.23.9.4

Wu, Y.-H., Stangl, E., Chipara, O., Hasan, S. S., Welhaven, A., & Oleson, J. (2018). Characteristics

of Real-World Signal to Noise Ratios and Speech Listening Situations of Older Adults With

Mild to Moderate Hearing Loss. *Ear and Hearing*, *39*, 293–304.

Zekveld, A. A., Koelewijn, T., & Kramer, S. E. (2018). The Pupil Dilation Response to Auditory

Stimuli: Current State of Knowledge. *Trends in Hearing*, *22*, 233121651877717.

https://doi.org/10.1177/2331216518777174

593