In situ structure determination using single particle cryo-electron microscopy images

Jing Cheng[1,2], Bufan Li[1,2], Long Si[1,2], Xinzheng Zhang[1,2,3,*]

1 National Laboratory of Biomacromolecules, CAS Center for Excellence in Biomacromolecules, Institute of Biophysics, Chinese Academy of Sciences, 100101 Beijing, People's Republic of China.

2 University of Chinese Academy of Sciences, 100049 Beijing, People's Republic of China.

3 Center for Biological Imaging, CAS Center for Excellence in Biomacromolecules, Institute of Biophysics, Chinese Academy of Sciences, 100101 Beijing, China.

* For correspondence: xzzhang@ibp.ac.cn

Cryo-electron microscopy (cryo-EM) tomography is a powerful tool for *in situ* structure determination. However, this method requires the acquisition of tilt series, and its time consuming throughput of acquiring tilt series severely slows determination of *in situ* structures. By treating the electron densities of non-target protein as non-Gaussian distributed noise, we developed a new target function that greatly improves the efficiency of the recognition of the target protein in a single cryo-EM image without acquiring tilt series. Moreover, we developed a sorting function that effectively eliminates the model dependence and improved the resolution during the subsequent structure refinement procedure. By eliminating model bias, our method allows using homolog proteins as models to recognize the target protein. Together, we developed an *in situ* single particle analysis (isSPA) method. Our isSPA method was successfully applied to solve structures of glycoproteins on the surface of a non-icosahedral virus and Rubisco inside the carboxysome. The cryo-EM data from both samples were collected within 24 hours, thus allowing fast and simple structural determination *in situ*.

Keywords: in situ; cryo-EM; single particle; sorting

1. Introduction

Determining the *in situ* structure of working protein machineries in their native context allows for more physiological structural information, together with identifying the interactions with other proteins nearby. One of the best technologies to determine *in situ* structures is cryo-electron tomography (Beck & Baumeister, 2016; Lučić, Leis, & Baumeister, 2008) When combined with sub-tomogram averaging technique (M. Chen et al., 2019; Leigh et al., 2019) that increases the signal noise ratio (SNR) of target protein complexes with multiple copies in the tomogram by aligning and averaging the three dimensional（3D）volume of the protein complex, *in situ* protein structures have previously been determined at sub-nanometer resolution on non-cryo-sectioning sample (Dodonova et al., 2017; Himes & Zhang, 2018; Mattei et al., 2018; Pfeffer et al., 2015; Schur, 2016; F. K. M. Schur et al., 2015; Turoňová, Schur, Wan, & Briggs, 2017; Wan et al., 2017) or nanometer resolution on cryo-sectioning sample (Bäuerlein et al., 2017; Bykov et al., 2017; Freeman Rosenzweig et al., 2017; Q. Guo et al., 2018; Mosalaganti et al., 2018). However, tomography requires the acquisition of a tilt series of the target protein complex. A tilt series typically contains more than 30 images taken at a range of tilt angles typically acquired within a time of 30 min, resulting in a slowdown of data collection throughput. The recent development of a stable sample stage in electron microscopy allows faster data collection by decreasing the waiting time of the stage (Chreifi, Chen, Metskas, Kaplan, & Jensen, 2019). However, the throughput is still dozens of times slower than collection of single particle data.

In a single particle image, when the target protein complex is located in an *in situ* environment, the density of the target protein complex in the image is overlapped by other densities from surrounding proteins or biological molecules. The overlapping densities can be considered as noise that decreases the SNR of the image, especially at the range of low frequencies, where the shot noise is much lower than signals. The low-frequency signals exhibiting high SNR are essential for determining the initial position and orientation of the protein complex which is required by applying a conventional iterative single particle algorithm. A previous study showed that by using a high-resolution model of the target protein complex, the initial position and orientation of this protein complex can be determined from the protein background by incorporating the high-frequency signals of the target protein into the search (Rickgauer, Grigorieff, & Denk, 2017). However, the usage of the high resolution structure of the target protein renders the method less practical, since this structure is usually not available yet. A simple whitening filter was applied to both the reference and the image before calculating the correlation coefficient (Rickgauer et al., 2017), which did not take the overlapping density and SNR oscillation into account. This oscillation stems from the signal oscillation induced by the contrast transfer function (CTF). Furthermore, the shot noise in the image follows a Gaussian distribution, resulting in a smooth background in Fourier space. Therefore, CTF-like weighting has been widely used in score function (F. Guo & Jiang, 2014; Jasenko Zivanov et al., 2018). However, the overlapping densities can be regarded as part of the noise that fails to follow Gaussian distribution. How to optimize the score function when considering the noise distributed in a non-Gaussian manner remains to be investigated.

Our and other's previous studies showed that in cases where the density of target
protein complexes is overlapped by other densities, after extracting information of the
initial center and orientation of the protein complex from single particle results (Zhu et
al., 2018) or from sub tomogram averaging (Song et al., 2019), the structure of a protein
complex can be effectively refined using traditional local refinement procedures
without having to subtract overlapping densities. Both of these methods provide
accurate initial center and orientation parameters, as the low-frequency signal in their
data used for the calculation possesses a high SNR. However, the determination of the
initial center and orientation of a protein complex by incorporating a high-frequency
signal with low SNR introduces false positive solutions, resulting in a reference bias
problem (Shaikh, Hegerl, & Frank, 2003) in the structure refinement procedure, which
therefore limits the resolution of the reconstruction. An effective method to reduce this
problem and improve the resolution is urgently required.

Here, we developed a single particle-like data processing method to determine the
in situ structure of protein complexes by combining an optimized score function to
provide the initial orientation and location of a target protein and a sorting algorithm to
distinguish the correct solution from the false positive solutions, therefore reducing the
model bias problem.

2. Materials and Methods
2.1.Algorithmic details
The weighted CC was expressed as

109
$$cc = \sum_k W(k) \cdot X(k) \cdot M(k),$$
(1)

110 Here, k is the spatial frequency, $W(k)$ is a weighting function, $M(k)$ is a projection of a
111 3D template and $X(k)$ is a raw image. Signal is modulated by CTF in raw image, and
112 considering shot noise $N(k)$ and protein noise which representing overlapping protein
113 densities $PN(k)$, $X(k)$ and $M(k)$ are written as below

114
$$X(k) = CTF(k) \cdot S(k) + N(k) + CTF(k) \cdot PN(k),$$
(2)

115
$$M(k) = S(k) + N'(k),$$
(3)

116 thus we can rewrite cc

117
$$cc = \sum_k W(k) \cdot (CTF(k) \cdot S(k) + N(k) + CTF(k) \cdot PN(k)) \cdot (S(k) + N'(k)).$$
(4)

118 Weighting function is calculated to maximize the SNR of cc

119
$$SNR_{cc} = \frac{\left\{\sum_k W(k) \cdot CTF(k) \cdot S(k) \cdot S(k)\right\}^2}{\left\{\sum_k \begin{array}{l} W(k) \cdot N(k) \cdot S(k) + W(k) \cdot N'(k) \cdot CTF(k) \cdot S(k) + W(k) \cdot N'(k) \cdot N(k) + \\ W(k) \cdot PN(k) \cdot CTF(k) \cdot S(k) + W(k) \cdot PN(k) \cdot CTF(k) \cdot N'(k) \end{array}\right\}^2},$$
(5)

120
$$\frac{d(SNR_{cc})}{dW} = 0.$$
(6)

The numerator of Eq.(5) represents the square of signal of *cc*, which contributes to
particle detection, and the denominator represents the noise in *cc*. We simplify this

calculation by including in two frequency terms (k_1 and k_2), and the corresponding

variables are simplified to *W_1, N_1, N'_1, PN_1, CTF_1,* and *S_1* for *W(k_1), N(k_1), N'(k_1), PN(k_1),*

CTF(k_1) and *S(k_1)*, and *W_2, N_2, N'_2, PN_2, CTF_2,* and *S_2* for *W(k_2), N(k_2), N'(k_2), PN(k_2),*

CTF(k_2) and *S(k_2)*

128
$$SNR_{cc} \rightarrow \left(\frac{W_1 CTF_1 S_1^2 + W_2 CTF_2 S_2^2}{\begin{array}{l} W_1 N_1 S_1 + W_1 N_1' CTF_1 S_1 + W_1 N_1' N_1 + W_1 PN_1 CTF_1 S_1 + W_1 PN_1 CTF_1 N_1' + \\ W_2 N_2 S_2 + W_2 N_2' CTF_2 S_2 + W_2 N_2' N_2 + W_2 PN_2 CTF_2 S_2 + W_2 PN_2 CTF_2 N_2' \end{array}} \right)^2 . \quad (7)$$

129   Ignoring the first order terms, and assuming that noise N' present in 3D template is

130   much smaller than noise N present in 2d raw image, the result is

131
$$W(k) = \frac{CTF(k) \cdot FSC(k)}{N^2(k) + PN^2(k) \cdot CTF^2(k)}. \quad (8)$$

132   Here $FSC(k)$ describes the *FSC* (Fourier Shell Correlation) between the perfect model

133   and the 3D template, which is defined by(Rosenthal & Henderson, 2003) as

134
$$FSC(k) = \frac{SNR(k)}{1 + SNR(k)}. \quad (9)$$

135   In this situation, $SNR(k)$ equals the ratio of $S^2(k)$ to $N'^2(k)$.

136   To remove the impacts caused by structure factor and the B-factor damping in

137   images, raw images and projections were signal-whitened

138
$$M(k) \rightarrow \frac{M(k)}{\sqrt{\left| S(k) \right|_r^2}}, \quad (10)$$

139
$$X(k) \rightarrow \frac{X(k)}{\sqrt{\left| S(k) \right|_r^2}}. \quad (11)$$

140   The term $\left| S(k) \right|_r^2$ is the radially averaged signal intensity, so the weighting function

141   applied to signal-whitened images was

142
$$W(k) = \frac{CTF(k) \cdot FSC(k)}{\dfrac{1}{SSNR(k)} + \dfrac{PN^2(k) \cdot CTF^2(k)}{\left| S(k) \right|_r^2}} \quad (12)$$

143   In the absence of protein noise *PN(k)*, weighting function is calculated as

144
$$W(k) = CTF(k) \cdot FSC(k) \cdot SSNR(k) \quad (13)$$

2.2.Fitting the power spectrum to perform signal whitening

The SSNR used in the weighting function was determined from the power spectrum of the image. First we fitted the shot noise intensity presented in the power spectrum by fitting zero points of CTF with an exponential function. According to this result, we calculated the differences between shot noise intensity and peak values induced by CTF. These values were assigned as signal intensity at the frequencies where the CTF equals 1. These values were then fitted with an exponential function. The two sets of exponential function parameters were used as SSNR parameters in the later processing.

2.3.Image acquisition of test data

Images of Herpes simplex virus type 2 particles (HSV-2) (~120 nm in diameter) and Reovirus particles (~90 nm in diameter) obtained from previous studies were analyzed for testing. HSV-2 data were collected on an FEI Titan Krios microscope operated at 300 kV equipped with a Falcon2 camera, with pixel size of 1.38 Å/pixel and a total dose of ~25 e$^-$/Å^2. Reovirus data was acquired on an FEI Titan Krios microscope operated at 200 kV with a Gatan K2 Summit camera, and the pixel size is 1.04 Å/pixel with a total dose of ~30 e$^-$/Å^2.

2.4.Picking the protein complexes in the testing dataset

Thirty virus particle images from each of the HSV-2 and Reovirus datasets were selected for testing. Before applying a weighting function, 2D images were signal-whitened. Projections of perfect 3D model were whitened first, then the square root of each weighting function was applied to both 2D images and projections. The projections of the initial model were generated by *EMAN* (Ludtke, Baldwin, & Chiu, 1999) with the incremental rotation angle set to 5 degrees. Images and projections were binned by 2 to reduce the CPU hours. For both HSV-1 and Reovirus datasets, any results with the translational error larger than 5 pixels or the rotational error larger than 6 degrees were considered as false positive results.

2.5.Reducing the false positive results by 3D classification

Initially, global detection was conducted with our weighting function on 2000 HSV-2 particles. The top 500 solutions were selected from each virus for further processing. 3D classification by RELION without alignment helped us exclude false particles. 129,119 particles were selected and the proportion of biased image was significantly decreased from 88% to 50%. We performed a local refinement. The resolution was reported at 4.3 Å, which we believe was limited by model biased problem. Another round of 3D classification (three classes) was performed using the parameter from the refinement and skipping alignment analysis. Each of the three classes contained 61,054, 21,756, 44,737 particles (Fig. 3A), respectively. Particles in class 1 and class 3 were selected for local refinement. The proportion of biased image was 44%, and the resolution remained at 4.3 Å, showing that 3D classification no longer helped to reduce model biased problem in the reconstruction.

2.6.Reducing the false positive results by sorting

To further reduce the biased images, we sorted the selected images based on different score.

$$score = FSC \cdot CTF \cdot \cos(\Delta\phi) \qquad (14)$$

Due to the extremely low SNR in high frequencies range, the constant n is much
smaller than *1/SSNR*, thus we ignored the overlapping density and used *FSC* and *CTF*
as weighting factors. In this test dataset, we divided the results into positive and false
positive results and plotted the corresponding curves, respectively. The corresponding
curves as shown in Fig. 3E are two Gaussion-like peaks. In addition, the plot of sorting
results was fitted by two Gaussion peaks using least square method which resemble the
two Gaussion-like peaks representing the plots of positive and false positive results,
respectively (Fig. S2C). It indicates that the fitting of two Gaussion peaks may help to
differentiate false positive results from positive results. In the non-testing dataset of
Rubisco, the plot of sorting was fitted well by two Gaussion peaks.   According to
sorting, we selected 40,000 particles for local refinement using RELION and achieved
a 4.0 Å resolution map. By adding in a further 6,000 viral particles, 180,000 particles
were selected after non-alignment 3D classification and sorting. ~180,000 particles
were selected after non-alignment 3D classification and PR sorting, the gold standard
resolution was 3.7 Å. By combining with CTF refinement in RELION using our
optimized weighting function, resolution was improved to 3.5 Å. The original CTF
refinement procedure implanted in RELION produced a map of 3.6 Å resolution.

2.7. Image acquisition of Bunyavirus

Bunyavirus data were collected on an FEI Titan Krios EM operated at 300 kV
equipped with a Gatan K2 Summit detector. Data collection was performed with Serial
EM using a nominal magnification of 22,500 with a total dose of ~50 e$^-$ resulting in a
pixel size of 1.3 Å/pixel. Each micrograph was recorded as a movie composed of 36
frames. A total of 1,849 such movies were collected and 10,950 viral particles were
extracted from motion-corrected micrographs.

2.8. Details of determining the structure of glycoproteins on non-icosahedral Bunyavirus

According to 2D and 3D classification, 3,140 viral particles displaying icosahedral
symmetry were selected for further processing. The whole virus reconstruction resulted
in a 11.8 Å resolution map due to the flexibility, block-based reconstruction of a block
centered on a pentamer. A pentamer in the center surrounded by 5 hexamers improved
the resolution to 8.7 Å, which was used as model in our isSPA methods. In the search
for pentamers, an area 500 Å in diameter was used as the model but restrained to a
diameter of 350 Å during further processing. Twenty potential solutions were selected
from each virus and 27,000 particles in total were selected according to 3D non-
alignment classification resulting in a 7.7 Å resolution map. In the search for hexamers,
one hexamer from the 5 surrounding a pentamer was segmented as the model, and 200
potential solutions were selected from each virus. After 3D non-alignment
classification, ~87,000 particles were selected and refined to 8.3 Å resolution.

2.9. Image acquisition of Carboxysome

Purified carboxysomes were applied to a glow-discharged Quantifoil R1.2/1.3 300-
mesh copper holey carbon grid (Quantifoil, Micro Tools), blotted under 100% humidity
at 4 ℃ and plunged into liquid ethane using a Mark IV Vitrobot (FEI). Micrographs
were acquired on a Titan Krios G2 microscope (FEI) operated at 300 kV with a K2
Summit direct electron detector (Gatan). A calibrated magnification of 130k times was

used for imaging, yielding a pixel size of 1.04 Å on images. The defocus was set at 1.5
to 2.0 μm. Each micrograph was dose-fractionated to 32 frames under a dose rate of 8
e⁻/pixel/s with a total exposure time of 10.4 s.

2.10. Details of determining the structure of Rubisco inside the carboxysome

From the collected micrographs, ~5,200 carboxysome images were extracted, and
each image was cross-correlated with the 2.7 Å model at a frequencies range from 1/100
$Å^{-1}$ to 1/8 $Å^{-1}$. This procedure yielded a large group of locations and orientations, from
which on average 150 solutions were picked per carboxysome for further processing.
The non-alignment 3D classification was performed using RELION. Approximately
150,000 particles were selected, and a further auto local refinement with angular
sampling of 0.9 degrees and translational sampling of 1.04 Å in RELION reported a 4.3
Å resolution map. To remove the reference bias, the refined particles were sorted with
phase residual using frequencies from 1/8 $Å^{-1}$ to 1/5 $Å^{-1}$. We tested three cutting
thresholds (0.07, 0.08 and 0.09) and selected ~82,000, ~46,000 and ~21,000 particles
to the second-round refinement individually, resulting in 4.0 Å, 3.9 Å and 3.9 Å
resolution. Further CTF refinement subsequently improved the resolutions to 3.9 Å, 3.9
Å and 3.7 Å. According to the fitting of two Gaussian distributions (Fig. 5H), the
portion of true solutions at threshold 0.09 was estimated to ~90%.

3. Results

3.1. Theoretical background

To localize the target protein in a crowded environment, a cross-correlation
coefficient (cc) between a template of the target protein and raw cryo-EM image, had
previously been applied as a picking function (Rickgauer et al., 2017). However, cryo-
EM data are characterized by a frequency-dependent SNR. The template from a cryo-
EM reconstruction also contains frequency-dependent noise. The SNR of the template
can be calculated according to the Fourier Shell Correlation (FSC). The noise in both
the cryo-EM image and the template contribute errors to the cc. Therefore, an
appropriate frequency-dependent weighting function should be applied to the picking
function to minimize the error in cc. In this work, the densities of the target protein in
a cryo-EM image are overlapped with the densities from other surrounding proteins *in*
situ. We treat the densities from surrounding proteins as noise, together with the shot
noises. However, different from the shot noise, the densities from the surrounding
proteins are determined by the structure factor of proteins and modified by the contrast
transfer function.

Our picking function is based on cc with a frequency-dependent weight applied,
which can be expressed in Fourier space as Eq. (1). We calculated the weighting
function $W(k)$ by maximizing the SNR of cc function (see Materials and Methods). The
weighting function is calculated as Eq. (12). Here, we used a cryo-EM map as a 3D
template. The $FSC(k)$ curve associated to the cryo-EM map was used to calculate $N'(k)$,
the *SSNR(k)* denotes the ratio of signal intensity to noise intensity in raw image. The
second term of the denominator is introduced by noise from overlapping protein
densities. Assuming that the structural factors of different proteins are similar, the ratio
of protein noise intensity to projection intensity could be approximated as a constant n,

which describes how proteins are overlapped. Our optimized weighting function is
simplified to

$$W(k) = \frac{CTF(k) \cdot FSC(k)}{\frac{1}{SSNR(k)} + n \cdot CTF^2(k)} \tag{15}$$

Weighting functions from a typical cryo-EM micrograph with different values of n
applied were calculated and plotted in Fig. 1. Here，SSNR was estimated from the
power spectrum of the image (see Materials and Methods) and *FSC* was set to 1. An
increase of n indicates more noise from overlapping protein densities, which decreases
the SNR. However, the SNR at low frequency range decreases faster than that at high
frequency range. Therefore, the weight of score at high frequency range increases along
with an increase of n. Moreover, the shape of the peak of the oscillation of the weighting
function expended in x direction with an increasing n. When n is much larger than
1/SSNR(k) (Fig. 1, n = 999) (e.g. ignoring the shot noise), our weighting function is
similar to a whitening filter (Rickgauer et al., 2017). Under different conditions when
n is zero (e.g. ignoring the overlapping densities), our weighting function is a dot
product of *CTF*(*k*), *SSNR*(*k*) and *FSC*(*k*) (see Materials and Methods). Without
considering the overlapping densities, similar weighting functions have been used in
CisTEM (Grant, Rohou, & Grigorieff, 2018) and for CTF refinement in RELION (J.
Zivanov et al., 2018). In our picking function, n is normally estimated to 3 or 4 in real
cryo-EM data (see below).

3.2.Particle detection
We tested our picking function by finding different protein complexes of different
sizes in cryo-EM images of different icosahedral viruses (HSV, Alphavirus, Reovirus).
Protein complexes are overlapped with densities from other proteins and genome of the
virus, which mimics an *in situ* environment. High resolution single particle analysis
(Yuan et al., 2018) (L. Chen et al., 2018) on these viruses ensured an accurate
determination of rotational and translational parameters of each 2D image. We extracted
the center and the orientation of target protein complexes on the virus and used these
known parameters as positive controls. In our weighting function, n regulates the ratio
of protein density noise to shot noise. To pick these protein complexes, we tested
different picking functions, with n ranging from 0 to 50. The results are displayed as
precise-recall curves (Figs. 2B and C). When n increases from 0 to 3, the ratio of correct
results increases in the datasets of HSV-2 and Reovirus. The ratio becomes steady when
n is between 3 and 6. Thus, we used 4 as default value of n in the picking function. The
ratio decreases when n increases further. The precise-recall curve of the whitening filter
where n approaches infinity shows a much lower ratio of correct results compared with
the curve of n = 4 (Figs. 2B and C).

Our picking function was tested on recognizing protein complexes of different
molecular weights. As shown in Fig. 2D, the ratio of correct results versus all results
decreases quickly as the size of the protein complexes decreases. This finding suggested
that our picking function also generates a significant amount of incorrect particles.
These false positive results are very similar to the model for picking according to the

high score of our picking function, which will induce reference bias in the structure refinement procedure.

3.3. Avoiding reference bias in the refinement

To investigate the ability of our method to determine the protein structure in a crowded environment, we selected a part of the HSV-2 capsid as a target protein. This part is approx. 900 kD in molecular weight and consists of VP5 and VP26 trimeric capsid protein as well as the surrounding triplex (two copies of VP23 and one copy of VP19C) (Yuan et al., 2018).

We set n to 4 and used the frequencies ranging from 1/100 $Å^{-1}$ to 1/8 $Å^{-1}$ for particle detection (see Materials and Methods). Possible locations and orientations of the target protein complex were calculated by our picking function and sorted according to the score of our picking function. After merging the results with similar orientations (within 7 degrees) in neighbor locations (within 10 pixels) into a single result, the first 500 putative target protein complexes with highest score from each virus particle were selected for further data processing, in which around 88% were false positive results according to the criteria we set in the Methods section. We performed 3D classification in RELION skipping alignment using the centers and orientations provided by our picking function. As shown in Fig. 3A, 2 out of 10 classes containing the lowest percentages of false positive result were selected, among which ~50% were false positive results. Further classification failed to improve the ratio of correct result.

Next, we performed auto refinement using only local search, which resulted in a reconstruction at 4.3 Å-resolution. The FSC curve as shown in Fig.3B drops quickly at the frequency of ~ 1/8 $Å^{-1}$ and exhibits a shoulder around the frequency of 1/5 $Å^{-1}$, indicating a reference bias problem in the reconstruction. We also tried to use fewer putative target protein complexes at the top of the sorting list to increase the ratio of correct results. However, this approach also decreased the number of correct results. When we reduced the number of putative target protein complexes in the refinement procedure, the resolution of the reconstruction improved (Fig. S2) presumably due to the increase of the ratio of correct result before decreasing because of the lack of particles.

To further reduce the ratio of false positive results, we calculated the score between reference image and the raw image according to refined parameters using phase residual. For this calculation, we only used the frequencies ranging from 1/8 $Å^{-1}$ to 1/5 $Å^{-1}$. The particles were sorted according to their score. As shown in Fig. 3E, the sorting efficiently separated the correct results from false positive results. When the sorting was based on the score from our picking function using the frequencies ranging from 1/20 $Å^{-1}$ to 1/8 $Å^{-1}$, the power of separation decreased markedly (Fig. 3D). This range of frequency was involved in particle picking, which performed a global search of location and orientation of the target protein on a whole virus. For instance, in a combination of translational (step size of 2.76 Å) and rotational parameters (step size of 5 degrees), 7.5×10^{10} possible locations of the protein complex were searched, from which the top 500 possible locations were selected by the program. Thus, each false positive result was selected from 7.5×10^{10} possible locations. However, further refinement using only local search improved the resolution to ~ 4.3 Å. The local

search strictly limited the possible locations. Thus, in the range of frequencies from 1/8 Å^{-1} to 1/4.3 Å^{-1}, the false positive result exhibits differences from the model. Therefore, excluding the range of frequencies that involves in picking function for sorting exhibits less reference bias. In addition, the refinement only performing local search was based on maximum likelihood score function in RELION. We tested "MaxValueProbDistribution" generated in RELION as score to sort the particles; however, the correct results were barely differentiated from the false positive results (Fig. 3C). The sorting according to parameter of "NrOfSignificantSamples" exhibited similar result to that of "MaxValueProbDistribution". Thus, it is possible that using scores different from the one used in the refinement for sort also helps to reduce the reference bias problem encountered in the refinement.

After sorting by the score calculated using only the frequencies from 1/8 Å^{-1} to 1/5 Å^{-1}, the top 40% particles contained ~93% correct results. Further refinement of this dataset led to a resolution of 4.0 Å (Fig. 3B). We were able to further improve the resolution to 3.7 Å by adding more particles from new micrographs (see Materials and Methods). In addition, we refined the CTF using our optimized weighting function (Equation 5), thus improved the resolution to 3.5 Å.

3.4. Workflow of isSPA method

IsSPA is a single-particle method to reconstruct *in situ* structures from untilted images. The images therefore contain overlapping protein densities, and the workflow includes: (1) estimate SSNR parameters by fitting power spectrum with two exponential functions (one for noise and the other for signal); (2) for particle picking, raw images are filtered by our optimized weighting function with known SSNR parameters, then the filtered images and projections of 3D model are input to program; (3) the output of program presents a large group of locations and orientations, thus we merge the neighbor solutions according to their translations and rotations; (4) we sort solutions from each image by cc value, and the number of solutions are selected depending on the estimated particle number in each image; (5) scripts were used to extract potential particles and perform a local refine and 3D classification on these particles using different cryo-EM software packages such as RELION, JSPR and so on; (7) a sorting algorithm was provided to remove false positive particles; (8) the remaining particles are performed local refinement to achieve the final reconstruction.

3.5. Determining protein structure using a homologous structure as a picking model

Since the expected structure from *in situ* structure determination is usually unknown, we explored the possibility of using homologous structure as a picking model. The differences between homologous structure and expected structure were treated as noise (N'), resulting in a different term named FSC_m (between the cryo-EM map of homologous structure and the expected structure) in the weighting function to replace the FSC.

$$W(k) = \frac{CTF(k) \cdot FSC_m(k)}{\frac{1}{SSNR(k)} + n \cdot CTF^2(k)} \tag{15}$$

To search for a similar protein complex on HSV-2 capsid core, we used a homologous protein complex present in the HSV-1 capsid core as a model. First, we extracted the 3D model of the homologous protein complex from a 4.2 Å map of HSV-1 (Dai & Zhou, 2018). The FSC curve between protein complexes from HSV-1 and HSV-2 showed similarity in the structures with the FSC value decreasing to 0.7 at the frequency of $1/8$ Å^{-1}. Three million potential particles of protein complex on HSV-2 capsid were selected on the basis of the score produced by our picking function. After local 3D classification and further selection by sorting, 60,000 particles were finally selected. As shown in Fig. 4B, the local refinement resulted in a 4.0 Å resolution map. We calculated the FSC curves between this map and the corresponding 3.1 Å map from the single particle result of HSV-2 and between this map and the corresponding 4.2 Å map of HSV-1 (Dai & Zhou, 2018). As shown in Fig.4C and D, our refined structure is closer to the corresponding structure in HSV-2 than that in HSV-1. Assuming that the 3.1 Å map represents a perfect map, the FSC between 3.1 Å map and our map reported a resolution of 4.0 Å using a threshold of 0.5. This result is in agreement with the resolution (4.0 Å) reported by FSC between two half maps using a threshold of 0.143. Together, these results show that the procedure we used in the *in situ* reconstruction avoids reference bias induced by homologous models.

To evaluate the ability of finding protein complexes by homologous models of different similarity between their structures and the structure of the target protein complex, we simulated homologous models by applying different scale factors to the structure of the target protein. As shown in Fig. S3, the similarities between the re-scaled model and the original map are indicated by FSC curves. The precision-recall curves show that the ability of finding a protein complex decreases along with the reduction of the similarity between homologous model and the protein complex (Fig. 4E and F). We downloaded PDB files of different homologous proteins and calculated the potential density maps, and then calculated the FSCs between pairs of homologous maps. As shown in Fig.S4, the similarity between proteins varies greatly. These homologous proteins with high similarities can be used as picking models in our method.

3.6. Determining *in situ* structures at high resolution using isSPA

To evaluate our *in situ* structure determination program, we first processed a dataset of a Bunyavirus by 2D and 3D classification. As shown in Fig. 5B, only approx. 28% of the viral particles exhibited an icosahedral symmetry. Further refinement on the icosahedral viral particles resulted in an 11.8 Å map. To compensate for this limitation in flexibility, one block centered on a pentamer was extracted and refined (Shaikh et al., 2003), which led to a map of 8.7 Å resolution. A centered sub-block (pentamer) and a sub-block adjacent to the pentamer that centered on a hexamer segmented from the 8.7 Å block were used as 3D models to pick the protein complex on the non-icosahedral virus particles previously excluded from data processing. Our isSPA resulted in a 8.3 Å map of hexamer and 7.7 Å map of pentamer (see Materials and Methods). The flexibility of the virus might have limited the resolution of these reconstructions.

In the second sample, carboxysomes were purified from *Halothiobacillus neapolitanus* and the cryo-EM data were collected on this sample. The size of the carboxysomes ranged from 80 nm to 150 nm. We also purified Rubisco from fractured

carboxysomes by adding extra freeze-thaw cycles before the centrifugation step for carboxysome purification. We collected the cryo-EM data of purified Rubisco, and obtained a 2.7 Å map of this complex (around 500 kD). This complex was then used as a model to pick the Rubisco inside the carboxysome in the cryo-EM images. The frequencies between $1/100$ Å^{-1} to $1/8$ Å^{-1} were used for picking. Further data processing (see Materials and Methods) using isSPA resulted in a 3.7 Å map of Rubisco located inside the carboxysome.

4. Discussion

In this work, we showed that the image-processing routines of isSPA method can resolve proteins (larger than 400 kD) at high resolution *in situ* with non-cryo-sectioning data when the thickness of the sample is around 100nm. However, our isSPA method can hardly solve small proteins due to the lack of signal. Moreover, due to the thickness and the quality of the cryo-sectioning data, we believe that there is more limitation in the size of protein complexes in cryo-sectioning data for this method. The major obstacle of this method is to determine the initial parameters of protein complexes. This method will benefit greatly from hardwares such as new generation of direct detectors or phase plate that improves the SNR of images at frequencies ranging from $1/20$ Å^{-1} 1 to $1/8$ Å^{-1} (0.1~0.25 Nyquist at a pixel size of 1 Å).

Acknowledgements

We thank L. Kong for cryo-EM data storage and backup and Z. Y. Lou in Tsinghua University for offering us the cryo-EM data of Bunyavirus. Cryo-EM data collection was carried out at the Center for Biological Imaging at the Institute of Biophysics (IBP), Chinese Academy of Sciences (CAS). We thank X. Huang, B. L. Zhu, G. Ji and other staff members at the Center for Biological Imaging (IBP, CAS) for their support in data collection.

Funding information

The project was funded by the National Key R&D Program of China (2017YFA0504700) and the Key Research Program of Frontier Sciences at the Chinese Academy of Sciences (ZDBS-LY- SM003), X.Z. received scholarships from the 'National Thousand (Young) Talents Program' from the Office of Global Experts Recruitment in China.

References:

Bäuerlein, Felix J B, Saha, Itika, Mishra, Archana, Kalemanov, Maria, Martínez-Sánchez, Antonio, Klein, Rüdiger, . . . Fernández-Busnadiego, Rubén. (2017). In Situ Architecture and Cellular Interactions of PolyQ Inclusions. *Cell, 171*(1), 179-187.e110. doi:10.1016/j.cell.2017.08.009

Beck, Martin, & Baumeister, Wolfgang. (2016). Cryo-Electron Tomography: Can it Reveal the Molecular Sociology of Cells in Atomic Detail? *Trends Cell Biol, 26*(11), 825-837.

doi:10.1016/j.tcb.2016.08.006

Bykov, Yury S, Schaffer, Miroslava, Dodonova, Svetlana O, Albert, Sahradha, Plitzko, Jürgen M,
Baumeister, Wolfgang, . . . Briggs, John Ag. (2017). The structure of the COPI coat determined
within the cell. *Elife, 6*. doi:10.7554/eLife.32493

Chen, Lihong, Wang, Ming, Zhu, Dongjie, Sun, Zhenzhao, Ma, Jun, Wang, Jinglin, . . . Zhang, Xinzheng.
(2018). Implication for alphavirus host-cell entry and assembly indicated by a 3.5A resolution
cryo-EM structure. *Nat Commun, 9*(1), 5326. doi:10.1038/s41467-018-07704-x

Chen, Muyuan, Bell, James M, Shi, Xiaodong, Sun, Stella Y, Wang, Zhao, 0000-0003-4897-9986, Orcid:,
& Ludtke, Steven J. (2019). A complete data processing workflow for cryo-ET and subtomogram
averaging. *Nat Methods, 16*(11), 1161-1168. doi:10.1038/s41592-019-0591-8

Chreifi, Georges, Chen, Songye, Metskas, Lauren Ann, Kaplan, Mohammed, & Jensen, Grant J. (2019).
Rapid tilt-series acquisition for electron cryotomography. *J Struct Biol, 205*(2), 163-169.
doi:10.1016/j.jsb.2018.12.008

Dai, Xinghong, & Zhou, Z. Hong. (2018). Structure of the herpes simplex virus 1 capsid with associated
tegument protein complexes. *Science, 360*(6384). doi:10.1126/science.aao7298

Dodonova, Svetlana O, Aderhold, Patrick, Kopp, Juergen, Ganeva, Iva, Röhling, Simone, Hagen, Wim J
H, . . . Briggs, John A G. (2017). 9Å structure of the COPI coat reveals that the Arf1 GTPase
occupies two contrasting molecular environments. *Elife, 6*. doi:10.7554/eLife.26691

Freeman Rosenzweig, E. S., Xu, B., Kuhn Cuellar, L., Martinez-Sanchez, A., Schaffer, M., Strauss, M., . . .
Jonikas, M. C. (2017). The Eukaryotic CO(2)-Concentrating Organelle Is Liquid-like and Exhibits
Dynamic Reorganization. *Cell, 171*(1), 148-162.e119. doi:10.1016/j.cell.2017.08.008

Grant, Timothy, Rohou, Alexis, & Grigorieff, Nikolaus. (2018). cisTEM, user-friendly software for single-
particle image processing. *Elife, 7*. doi:10.7554/eLife.35383

Guo, Fei, & Jiang, Wen. (2014). Single particle cryo-electron microscopy and 3-D reconstruction of
viruses. In *Electron Microscopy* (pp. 401-443): Springer.

Guo, Qiang, Lehmer, Carina, Martínez-Sánchez, Antonio, Rudack, Till, Beck, Florian, Hartmann,
Hannelore, . . . Fernández-Busnadiego, Rubén. (2018). In Situ Structure of Neuronal C9orf72
Poly-GA Aggregates Reveals Proteasome Recruitment. *Cell, 172*(4), 696-705.e612.
doi:10.1016/j.cell.2017.12.030

Himes, Benjamin A, & Zhang, Peijun. (2018). emClarity: software for high-resolution cryo-electron
tomography and subtomogram averaging. *Nat Methods, 15*(11), 955-961.
doi:10.1038/s41592-018-0167-z

Leigh, Kendra E, Navarro, Paula P, Scaramuzza, Stefano, Chen, Wenbo, Zhang, Yingyi, Castaño-Díez,
Daniel, & Kudryashev, Misha. (2019). Subtomogram averaging from cryo-electron tomograms.
Methods Cell Biol, 152, 217-259. doi:10.1016/bs.mcb.2019.04.003

Lučić, Vladan, Leis, Andrew, & Baumeister, Wolfgang. (2008). Cryo-electron tomography of cells:
connecting structure and function. *Histochemistry and Cell Biology, 130*(2), 185.
doi:10.1007/s00418-008-0459-y

Ludtke, S. J., Baldwin, P. R., & Chiu, W. (1999). EMAN: semiautomated software for high-resolution
single-particle reconstructions. *J Struct Biol, 128*(1), 82-97. doi:10.1006/jsbi.1999.4174

Mattei, Simone, Tan, Aaron, Glass, Bärbel, Müller, Barbara, Kräusslich, Hans-Georg, & Briggs, John A G.
(2018). High-resolution structures of HIV-1 Gag cleavage mutants determine structural switch
for virus maturation. *Proc Natl Acad Sci U S A, 115*(40), E9401-e9410.
doi:10.1073/pnas.1811237115

Mosalaganti, S., Kosinski, J., Albert, S., Schaffer, M., Strenkert, D., Salomé, P. A., . . . Beck, M. (2018). In
situ architecture of the algal nuclear pore complex. *Nat Commun, 9*(1), 2361.
doi:10.1038/s41467-018-04739-y

Pfeffer, Stefan, Burbaum, Laura, Unverdorben, Pia, Pech, Markus, Chen, Yuxiang, Zimmermann,
Richard, . . . Förster, Friedrich. (2015). Structure of the native Sec61 protein-conducting channel.
Nat Commun, 6, 8403. doi:10.1038/ncomms9403

Rickgauer, J Peter, Grigorieff, Nikolaus, & Denk, Winfried. (2017). Single-protein detection in crowded
molecular environments in cryo-EM images. *Elife, 6*. doi:10.7554/eLife.25648

Rosenthal, Peter B, & Henderson, Richard %J Journal of Molecular Biology. (2003). Optimal
Determination of Particle Orientation, Absolute Hand, and Contrast Loss in Single-particle
Electron Cryomicroscopy. *4*(333), 721-745.

Schur. (2016). An atomic model of HIV-1 capsid-SP1 reveals structures regulating assembly and
maturation. *Science, 353*. doi:10.1126/science.aaf9620

Schur, Florian K M, Hagen, Wim J H, Rumlová, Michaela, Ruml, Tomáš, Müller, Barbara, Kräusslich, Hans-
Georg, & Briggs, John A G. (2015). Structure of the immature HIV-1 capsid in intact virus
particles at 8.8 Å resolution. *Nature, 517*(7535), 505-508. doi:10.1038/nature13838

Shaikh, Tanvir R, Hegerl, Reiner, & Frank, Joachim %J Journal of structural biology. (2003). An approach
to examining model dependence in EM reconstructions using cross-validation. *142*(2), 301-310.

Song, Kangkang, Shang, Zhiguo, Fu, Xiaofeng, Lou, Xiaochu, Grigorieff, Nikolaus, & Nicastro, Daniela.
(2019). In situ structure determination at nanometer resolution using TYGRESS. *Nat Methods*.
doi:10.1038/s41592-019-0651-0

Turoňová, Beata, Schur, Florian K. M., Wan, William, & Briggs, John A. G. (2017). Efficient 3D-CTF
correction for cryo-electron tomography using NovaCTF improves subtomogram averaging
resolution to 3.4Å. *Journal of structural biology, 199*(3), 187-195.
doi:10.1016/j.jsb.2017.07.007

Wan, William, Kolesnikova, Larissa, Clarke, Mairi, Koehler, Alexander, Noda, Takeshi, Becker, Stephan, &
Briggs, John A G. (2017). Structure and assembly of the Ebola virus nucleocapsid. *Nature,
551*(7680), 394-397. doi:10.1038/nature24490

Yuan, S, Wang, J, Zhu, D, Wang, N, Gao, Q, Chen, W, . . . Liu, H. (2018). Cryo-EM structure of a herpesvirus
capsid at 3.1 Å. Science 360: eaao7283. In.

Zhu, Dongjie, Wang, Xiangxi, Fang, Qianglin, Van Etten, James L, Rossmann, Michael G, Rao, Zihe, &
Zhang, Xinzheng. (2018). Pushing the resolution limit by correcting the Ewald sphere effect in
single-particle Cryo-EM reconstructions. *Nat Commun, 9*(1), 1552. doi:10.1038/s41467-018-
04051-9

Zivanov, J., Nakane, T., Forsberg, B. O., Kimanius, D., Hagen, W. J., Lindahl, E., & Scheres, S. H. (2018).
New tools for automated high-resolution cryo-EM structure determination in RELION-3. *Elife,
7*. doi:10.7554/eLife.42166

Zivanov, Jasenko, Nakane, Takanori, Forsberg, Björn O, Kimanius, Dari, Hagen, Wim JH, Lindahl, Erik, &
Scheres, Sjors HW %J Elife. (2018). New tools for automated high-resolution cryo-EM structure
determination in RELION-3. *7*, e42166.

Figures and Tables

Figure 1

592   **Weighting functions**.

593   Weighting factors damp and oscillate with spatial frequency (black, red, green and gray) at different n (0, 1, 4, 999).

Figure 2

Efficiency of particle detection.

(A) Left: projection of HSV-2 hexamer model. Right: HSV-2 virus particle imaged at 2.75 μm defocus with 25 electrons per Å[-2] .The location of the projection on the virus particles is indicated by a white square. (B) Precision-recall curves for detections on Reovirus datasets using model in 900 KD molecular weight at n=0 (red),1 (purple),2 (navy),3 (magenta),4 (cyan),5 (orange),6 (gray), 50 (green) and whitening filter (black). (C) Precision-recall curves for detections on HSV-2 datasets with the same processing as in (B). (D) Precision-recall curves (purple, navy, green and orange) for detections (n equals 3) on Alphaviruses using models in different molecular weights (960 kD, 800 kD, 640 kD and 480 kD).

Figure 3

Data processing of HSV-2 hexamer.

(A) Upper panel: 3D classification of raw picked positives in ten classes, the percentage of true detections in each class is shown at the top left corner. The two selected classes are indicated by squares in red. Lower panel: 3D classification of selected particles in three classes, and the percentage of true detections in each class is noted at the top left corner. (B)Upper pannel: 3D reconstruction of the HSV-2 hexamer. Lower pannel: FSC curves denote 3 reconstructions in different conditions using 2000 viral particles before PR sorting (dash) and after PR sorting (black), using 8000 virus particles and after PR sorting (navy). (C) True and false particle distribution with *MaxValueProbDistribution* term in RELION. (D) True and false particle distribution with phase residual using data from 1/30 Å^{-1} to 1/8 Å^{-1}. (E) True and false particle distribution with phase residual using data from 1/8 Å^{-1} to 1/5 Å^{-1}.

655 **Figure 4**

Homologous structure as picking model.

(A) FSC curve of HSV-1 hexamer with HSV-2 hexamer. (B) The resolutions determined by gold standard FSC at threshold 0.143

of HSV-2 hexamer reconstructed from HSV-1 hexamer. (C) PDB of HSV-1 hexamer (yellow) and HSV-2 (green) fitting to the

density of the 4.0 Å cryo-EM map, the disagreements of the two PDBs are pointed out by red arrows. (D) FSC curves show

similarities between our map and HSV-1 hexamer (black), and HSV-2 hexamer (red) respectively. (E) Precision-recall curves of

the ~2 MD protein of HSV-2 at a series of scales corresponding to Fig. S3. (F) Precision-recall curves of the ~1.8 MD protein of

Reovirus at a series of scales corresponding to Fig. S3.

Figure 5

690 A

699 C

707 F

Application.

(A) A micrograph of Bunyavirus particles. (B) Upper panel: 2D classification of viral particles binned by 4. Lower panel: 3D

classification of viral particles selected from 2D classification. (C) Densities in the 7.7 Å map of pentamer. (D) FSC curve shows

the resolution of hexamer map. (E) FSC curve shows the resolution of pentamer map. (F) An image of a carboxysome, Rubiscos

are packaged inside; scale bar represents 10 nm. (G) The 3.7 Å map of Rubisco reconstructed using our isSPA method (right) and

the FSC curve showing the resolution of the 3.7 Å map at gold standard (left). (H) Distribution of particle numbers to scores fitted

by two Gaussian functions.

Supplementary Materials

Figure S1

Data processing of Reovirus.

(A) The 4.0 Å reconstruction of reovirus hexamer by our isSPA methods with previous reconstruction as picking

model. (B) Left: EM map produced by pdb of the crystal structure (1jmu). Right: FSC curve describing the similarity

between the target and homologous model. (C) 3D classification of selected particles from global search with

homologous map as picking model, and particles in class 8 (which is bolded in a red square) were true detections.

Figure S2

749

FSC weight and picking threshold on HSV-2 dataset.

(A) Precision-recall curves of the ~2 MD protein rescaled to ~8 Å at FSC 0.5. We tested on HSV-2 with FSC

weighting (black) and without FSC weighting (red). (B) The relationship of resolution to threshold tested on HSV-2

dataset before PR sorting (black) and after PR sorting (purple). (C) Gaussian fitting of particle number to score

distributions by true solutions (red) and false solutions (green).

Figure S3

A

774 **FSC curves of rescaled maps with the origin maps.**

775 (A) HSV-2 hexamer rescaled at different scales. (B) Reovirus hexamer rescaled at different scales.

776
777
778

779 **Figure S4**

802 **FSC curves between homologous structures.**

803 (A) Glycoproteins of influenza viruses. (B) Ion channels proteins. (C) NDH. (D) Spike proteins of coronaviruses.

Figure S5

Flowchart of isSPA method.

849 **Figure S6**

852 A

861 **FSC with high resolution initial model.**

862 (A)   FSCs of 3D reconstructions by selected particles according to sorting (green), removed bad particles according

863 to sorting (red) and both particles before sorting (navy) with high resolution initial model. (B) the same processing

864 with Rubisco images as in (A).