Full title: Method comparison studies of telomere length measurement using qPCR approaches: a critical appraisal of the literature

Short title: A review of telomere length measurement method comparisons using qPCR

Alyssa R. Lindrose[1]*, Lauren W. Y. McLester-Davis[2], Renee I. Tristano[3], Leila Kataria[4], Shahinaz M. Gadalla[5], Dan T. A. Eisenberg[6], Simon Verhulst[7], Stacy Drury[1,2,8]*

[1] Department of Psychiatry and Behavioral Sciences, School of Medicine, Tulane University, New Orleans, LA, USA

[2] Tulane Brain Institute, Tulane University, New Orleans, LA, USA

[3] School of Public Health and Tropical Medicine, Tulane University, New Orleans, LA, USA

[4] School of Science and Engineering, Tulane University, New Orleans, LA, USA

[5] Division of Cancer Epidemiology and Genetics, National Cancer Institute, National Institutes of Health, Bethesda, MD, USA

[6] Department of Anthropology, Department of Biology, Center for Studies in Demography and Ecology, University of Washington, UA

[7] Groningen Institute for Evolutionary Life Sciences, University of Groningen, Groningen, the Netherlands

[8] Department of Pediatrics, School of Medicine, Tulane University, New Orleans, LA, USA

*Corresponding authors

Email: alindrose@tulane.edu (ARL); sdrury@tulane.edu (SD)

Abstract

Use of telomere length (TL) as a biomarker for various environmental exposures and diseases has increased in recent years. Various methods have been developed to measure telomere length. PCR-based methods remain wide-spread for population-based studies due to the high-throughput capability. While several studies have evaluated TL measurement methods, the results have been variable. We conducted a literature review of TL measurement cross-method comparison studies that included a PCR-based method published between January 1, 2002 and May 25, 2020. A total of 20 articles were found that matched the inclusion criteria. Papers were reviewed for quality of methodologic reporting of sample and DNA quality, PCR assay characteristics, and analytic approaches to determine final TL was low as assessed by two different recommended reporting guidelines. Additionally, few papers reported whether samples were blinded for analysis. Reported correlation between methods (as assessed by Pearson's r) varied, as did reported coefficients of variation or intra-class correlation coefficient. The sample size for nearly all studies was less than 100, raising significant concerns about statistical power. Overall, this review found that the current literature on the relation between TL measurement methods is lacking in validity and scientific rigor. To assist future investigators, we present recommendations for reporting guidelines of PCR-based TL measurement methods and present data to demonstrate the impact of assay variability on statistical power. Future cross-laboratory studies with rigorous methodologic and statistical reporting, adequate sample size, and blinding are essential to determine the true relation and variability of various TL measurement methods and are a key directive to the newly initiated Telomere Research Network funded jointly by then NIA and NIEHS.

Introduction:

Telomeres, the protective nucleic acid and protein cap found at the end of all eukaryotic chromosomes, have captured the attention of scientists, medical and public health professionals, biotechnology companies, and the media over the last two decades. In 1973, Olovnikov proposed his theory of marginotomy, which reasoned that during DNA replication, DNA polymerase would not be able to completely copy the first DNA segment and, to prevent the loss of critical DNA sequences in genes, a noncoding set of DNA nucleotides would be required to act as a buffer protecting the loss of important, gene-encoding, sequences[1]. Subsequently in 1978, Blackburn et al. first reported the actual DNA sequences of telomeres in yeast[2], followed by the first sequencing of the human telomere in 1988[3]. The sequencing of telomeric DNA paved the way for the development of methods that measured the length of telomeres, beginning with the first report of telomere length measurement using the Southern blot method for mammalian chromosomes in 1988 [3]. Since then, thousands of papers assessing telomere length (TL) in human cells have been published across a myriad of different scientific fields (Fig 1). As a result of the broad scientific interest in both the role of TL in disease processes and the influence of environmental factors on TL dynamics, the number of studies evaluating TL in human cells continues to increase, in part facilitated by the regular development of new methods and modifications of existing assays.

Fig 1. Telomere publications 1988-2019. All publications mentioning telomere length of human DNA from 1988 to 2019. Search was completed on May 25, 2020. Data was obtained by searching "telomere," "length," and "human" in PubMed and sorting articles by year.

Currently, over two dozen assays have been developed to measure TL in human studies (Fig 2) [4-33]. These assays are classifiable into four broad categories: hybridization-based, PCR-based, sequence-based, and mixed methods (e.g. hybridization/PCR combination). These assays vary in which

component of TL is measured. While most focus on the measurement of the average TL within the

sample, assays also measure chromosome-specific TL, the complete distribution of TL in a cell

population[33], or the shortest TL. The shortest TL has received considerable attention, given evidence

from in vitro and preclinical models suggesting that the shortest TL is most predictive of cellular

senescence [34, 35]. Several recent reviews have discussed the overall advantages and disadvantages of

each method [26, 36] focusing on cost, scalability, constraints of starting biological samples (e.g. living

cells, amount of DNA, etc) and, to some extent, inter and intra-lab precision as specific challenges facing

the field, including the use of coefficient of variation (CV) compared to intraclass correlation coefficients

(ICC) [26, 37, 38]. While studies of basic telomere biology continue to explore the complex role that

telomeres play in cellular and organismal function, studies testing TL as markers of disease risk or

environmental exposure must balance biological relevance, methodologic precision, and experimental

practicality, similar to other epigenetic markers such as DNA methylation[39].

Fig 2. Novel telomere methods developed from 1988 to 2020. Data were obtained by literature search

and through references of telomere review papers and method comparison papers.

Over the last decade, debates have arisen over the utility and measurement of TL. This debate is

partially fueled by concerns related to the reproducibility and replicability of TL measurements across

studies, methods, and laboratories, particularly as new methods were developed and more laboratories

began performing these assays. In response to this debate, several studies have attempted to compare

TL measured across different assay methods or laboratories. Some of these method comparison studies

examined the direct correlation of TL measurement in the same sample using different assay methods

and/or tested the repeatability of TL with the same method (e.g. the amount of within assay variation)

[40, 41]. Others tested the relative correlation of the TL measured by different assays with an expected

phenotype (e.g. aging, parent-offspring correlation) [42], or examined the relative ability of different

assays of TL to predict a specific disease or health outcome[43, 44]. Each of these approaches requires a

different analytic strategy and study design. To date, the existing evidence remains insufficient to

answer key methodologic questions related to differences in reproducibility and replicability across

measurement assays and how/whether these differences affect the ability of TL to serve as a biological

indicator of exposure or a predictor of disease or health risk [45]. Beyond these concerns, there remains

a lack of consensus as to which, if any, methodology is the "gold standard," as even the classic Southern

blot method is challenged by its inability to capture potentially critical metrics (e.g. full distribution,

shortest telomere length, inclusion of the subtelomeric region). Before applying TL as a biomarker of

environmental exposure and/or a predictor of a disease, measured by any method that is applicable to

human population studies or ecological studies, it is critical to systematically test fundamental issues

related to assay reproducibility and replicability [46].

78   As part of a joint National Institute of Aging and National Institute of Environmental Health

Science initiative that funded a U24 cooperative award and four separate U01 awards, a Telomere

Research Network (TRN) was established in 2019 (trn.tulane.edu). This network is facilitating

interdisciplinary collaborations and dialogues across all scientific fields incorporating telomere research

in relation to environmental exposures, psychosocial stress, and disease susceptibility in a variety of

health-related research studies. The TRN is coordinating cross-method comparison studies with the

long-term goal of developing methodological guidelines and recommendations for telomere research

applicable to population-based studies. As a first step towards the goals of this network, we undertook

a systematic literature review of published studies that directly compared TL measured using at least

one PCR-based method and another approach to determine how these studies might inform the field,

with particular attention to assay precision and accuracy of different measurement assays and what

research gaps remain.

90   As defined by the Committee on Reproducibility and Replicability in Science, precision is the

closeness of agreement between measured quantities obtained by replicate measurements, while

accuracy is the closeness of agreement between a measured quantity and a true value [46].

Reproducibility is defined as precision in measurement under conditions that involve different locations

or different measurement procedures, while repeatability is defined as precision in measurements that

include the same procedures/locations. Beginning from this perspective, this systematic review

evaluated the existing literature related to cross-method comparisons. This review focuses on PCR-

based methodologies due to their increasing use in population-based studies, their central role in the

debate related to assay precision, and the existence of two reporting guidelines - one created through

the TRN, and a second one created by a separate group in a recently published manuscript [47]. The

majority of PCR-based methods are derived from two seminal methodologic papers by Richard Cawthon,

the first describing a monoplex based assay (qPCR) and the second describing a multiplex assay

(MMqPCR) [13, 18]. Our review focuses specifically on the comprehensiveness of methodologic

reporting, correlation between TL measured by different assays, assay repeatability and reproducibility,

and overall scientific design of methodological comparisons. Finally, we suggest areas of needed

scientific examination and provide some guidance related to study design, necessary sample size, and

analytic approach, to address key remaining questions: (1) What is known about the relationship

between TL measured using PCR-based methods and other assays? (2) What is known about the

reproducibility and repeatability of PCR-based methods and how does this relate to other TL

measurement techniques? (3) Given variable methodologic precision, what are the implications for

sample size and power? (4) What are appropriate guidelines to systematically evaluate the precision of

both existing and future TL assays? Addressing these important questions is a requisite step in

advancing our understanding of the ability of TL, measured by any approach, to serve as a sentinel of

psychosocial and environmental exposures and a predictor of future disease.

Methods

Manuscript search

To identify relevant papers that reported on cross method comparisons of any qPCR-based method (qPCR, absolute TL (aTL), and MMqPCR) and another method of TL assessment (PCR-based or otherwise) or the same PCR-based method conducted in separate laboratories, we conducted a critical review beginning with a literature search (Fig 3). The following key terms "telomere," "telomere length," and "human" were searched in PubMed. From these initial results, a second search included the keyword "PCR" to identify the initial titles for screening. Search criteria included papers published since January 1, 2002 (the year the first method to measure TL by qPCR was published) through May 10, 2020. The references of selected papers were also reviewed to identify any additional papers. A list of identified papers was presented to the TRN Steering Committee, who also suggested additional papers. Initial review of papers for inclusion was accomplished through evaluation of both the abstract and methods section, as some manuscripts were not directly focused on methodological comparison and instead only reported the cross-method comparison on a subset of samples. Final inclusion in this review met the following baseline requirements:

1. Article published in a peer-reviewed journal (abstracts and pre-prints not included).
2. Article was not presenting the initial development of a new method or a substantial refinement of an existing methodology. This type of study was excluded due to the intrinsic bias in reporting novel method development.
3. Included a direct comparison of TL using the same biological measured with two distinct TL assays or in two separate laboratories.
4. At least one of the methods used to measure TL was PCR-based.

136 **Fig 3. PRISMA flow diagram of literature search and article selection for inclusion**. Date range searched

137 was between Jan 1 2002 to May 10 2020. Other sources of article identification included suggestions

138 from participants of the TRN and review of references of selected papers.

Quality review:

140   Included papers were evaluated for quality of methodologic reporting using two different

indices of reporting guidelines for PCR-based telomere studies. The first was created through consensus

of the initial participants in the TRN (Supplemental Table 1). The second was derived from

recommendations published by Morinha et al. 2020 (Supplemental Table 2) [47]. We included both

guidelines for two reasons. First, there is not empirical data to distinguish between the two guidelines in

terms of ensuring rigor and reproducibility for the field. Second, as several of the papers reviewed were

authored by participants involved in the creation of the TRN guidelines, the inclusion of both guidelines

provided some degree of impartiality. Both guidelines contain overlap with the MIQE guidelines and

include characterization of the importance of each recommendation [48]. In terms of specific

differences, the Morinha recommendations included several pre-analytic considerations not included in

the TRN guidelines (e.g. volume of sample processed, robotic instrumentation vs manual), while storage

buffer and the percentage of samples tested for DNA integrity were included in the TRN guidelines but

not the Morinha. The Morinha guidelines also required greater detail for qPCR validation such as the

standard curve and calibration samples, as well as a requirement for the melt curve and Ct of the

negative control, which were not included in the TRN guidelines. Both guidelines assess the

comprehensiveness of the information describing the PCR assay itself as well as analytic considerations

for final TL determination. A grading rubric for each set of guidelines was developed to reduce

subjective reviewer interpretation (Supplemental Table 3). A composite assessment for each index was

divided into three subcategories for the TRN guidelines and five subcategories for the Morinha

guidelines. These broadly encompassed sample collection and processing, DNA quality metrics and

storage; PCR assay components and quality control; and data analysis. Two of three reviewers

independently assessed each article for fulfillment of reporting guidelines (ARL, LWYM, and SSD). The

scores for each individual item were compared and discrepancies resolved by the third reviewer.

Additional characteristics assessed included sample blinding prior to analyses, single lab or

multi-lab testing, conversion/transformation of raw TL measurement prior to comparison, and whether

the study design evaluated repeatability and/or reproducibility. Lastly, when available, sample size,

means and standard deviation of TL measurement are included to indicate study power. Although

several studies included means and SD of the entire sample, only a subset reported the means and SD of

the samples utilized in the method comparison analyses. This review only included method comparisons

that involved at least one PCR-based method as currently reporting guidelines are only available for PCR

based methods. As the majority of PCR-based measurements of TL are relative, it was not possible to

address accuracy.

Correlation between methods was assessed by using Pearson's r or r^2 values where provided.

Weighted average correlation coefficients were determined for each type of comparison by converting

reported Pearson's r values (or the square root of reported r^2 values) to Fisher's z values, and weighting

by sample size. A forest plot was generated from the weighted r average, total sample size for that

correlation, and 95% confidence interval (CI) range using DistillerSR Forest Plot Generator from Evidence

Partners (https://www.evidencepartners.com/resources/forest-plot-generator/).

178 ICC calculation:

Given established analytic shortcomings related to the use of the CV as a metric of testing the

repeatability of TL, or the correlation of TL measurement between assays, raw data from cross method

comparison studies was used to calculate ICCs for comparison between methods where available [38,

49, 50]. ICCs for one study were also calculated using a two-way, single measurement, absolute

agreement, random effects model, known as ICC(A,1) and for average measurements ICC(A,k) in

McGraw & Wong's (1996) terminology [51].

To provide guidance for future study design, we present several different power analyses

outlining the relation between sample size, ICC, and ability to detect group differences. For cross

sectional studies, calculations assumed a realistic (true) standard deviation of 650 base pairs (bp). Power

analysis was done using G*Power[52]. N is the combined n of the two groups to be compared and was

assumed to be equally distributed among the two groups. To examine the impact of variation in ICC on

longitudinal TL studies, the statistical power to detect a significant change (paired-t-test) in telomere

length of 25 bp/year for sample sizes of 25, 50 and 100 individuals, and an interval of 8 years between

baseline and follow-up (i.e. on average 200 bp in total), as a function of measurement repeatability (e.g.

reliability) expressed as the ICC (A,1). Population SD of telomere length was assumed to be 650 bp at

both time points and telomere shortening was simulated assuming a Poisson distribution with

mean/variance of 25bp/year, which is within the range of observed shortening in adults, although the

true clinical or phenotypic implications of specific differences in bps has yet to be empirically

determined.

Results

The initial search revealed 5427 articles and the inclusion of "PCR" as a search term limited the

results to 767 articles, whose abstracts and methods were read (Figure 3). An additional 25 articles were

identified through assessment of the references of method validation papers and other included cross-

method validation studies. A review of these 792 abstracts identified 65 articles for assessment of the

full text and supplemental information to determine inclusion in this review. 26 articles were

determined to be novel method validation and excluded. 18 articles were excluded as they were either

reviews or did not include direct method comparisons. One article was excluded due to the

determination that the DNA samples used for cross-method comparison were obtained at different time

points. We also included 6 papers that, while not specifically designed as a cross method comparison,

included sufficient details comparing TL measurement using different assays. This resulted in a total of

20 articles included in this review (Table 1).

Table 1: Characteristics of included papers.

Study	Year	Methods	Multi-lab?	Sample Type	Sample Size	Age Range (years)	TL Mean ± SD
Ehrlenbach*†[53]	2009	qPCR x TRF		blood	56	51-81	T/S = 1.31 (0.95-1.90) (n=669)
Pavesi*[54]	2009	qPCR x TRF		blood, cord blood	28	0-97	NR
Aviv[40]	2011	qPCR x TRF	Yes	blood	50	41-70	NR
Imam*[55]	2012	qPCR x FlowFISH		blood, cord blood, DBS	29	0	T/S = 6.23 ± 1.54 (n=35); 5.64 ± 1.54 (n=51)
Zanet[56]	2013	qPCR x MMqPCR		blood	32	2-59	NR
Gutierrez-Rodrigues[44]	2014	qPCR x FlowFISH x TRF		blood, cord blood	70, 51	0-88	NR
Martin-Ruiz[41]	2015	qPCR x MMqPCR x TRF x STELA	Yes	cell lines, placenta	12	N/A	NR
Eisenberg[42]	2015	MMqPCR x TRF	Yes	blood	190	NR	NR
Jodcyzk*[57]	2015	qPCR x TRF		blood	20	0-35/50	T/S = 1.184 ± 0.371 (n=677); 1.104 ± 0.153 (n=351)
Panero*[58]	2015	aTL x TRF		bone marrow cells	102	30-87	7.2 ± 0.46 kb (n=48); 6.15 ± 0.41 kb (n=54)
Hsieh[59]	2016	qpCR x MMqPCR x TRF x flowFISH		blood, cord blood, DBS, placenta, cell lines	33 - 84	NR	NR
Wand[60]	2016	qPCR x TRF x FlowFISH		blood	11	24-69	NR
Behrens[61]	2017	qPCR x FlowFISH x T/C-FISH x TRF		blood, cord blood	83 - 99	0-81	NR
Khincha†[62]	2017	qPCR x TRF x FlowFISH		blood	35, 53	3-69	T/S = 0.50 (0.21-1.48) (n=35); 0.91 (0.45-1.27) (n=53)
Lee[63]	2017	qPCR x TRF x WGS		cell lines	20	N/A	NR
Tarik[64]	2018	MMqPCR x TRF		blood	94	NR	T/S = 1.02 ± 0.32
Wang[65]	2018	qPCR x FlowFISH		blood	181	19-53	NR
Gadalla*[43]	2018	qPCR x qPCR	Yes	blood	100	NR	NR
Ventura Ferreira[66]	2019	MMqPCR x FlowFISH		blood	105	24-66	NR
Ropio[67]	2020	qPCR x TRF x aTL		cell lines	20	N/A	4.32 ± 0.143 kb

Sample size reflects the n used for method comparisons. If the TL mean and standard deviation (SD)

reflects reported values for the entire sample size, the n is included in parentheses.

* Papers not specifically intended to compare correlation of methods.

† TL median and range were provided in lieu of mean and SD.

N/A: not applicable to this sample type.

NR: not reported

DBS: dried blood spot

Paper Characteristics

The most common methods comparison among the 20 papers evaluated in this review was

monoplex quantitative PCR (qPCR) and the telomere restriction fragment (TRF) method by Southern blot

(n=12). Four studies compared multiplex qPCR (MMqPCR) with TRF. Seven studies compared qPCR with

the flow-FISH method, and two studies compared MMqPCR with flow-FISH. Three studies involved

comparisons of MMqPCR and qPCR, and two studies involved direct comparisons of qPCR methods or

separate laboratories. Other methods examined included STELA (compared with qPCR, MMqPCR, and

TRF in one study) [41], and absolute TL (two studies, one comparing aTL to TRF and one comparing to

aTL to qPCR) [58, 68]. T/C-FISH was also examined in one study as it related to qPCR [61]. Additionally,

one paper compared the correlation of several whole genome sequencing (WGS) platforms to qPCR-

based measurement[63]. Note some studies compared more than two methods [41, 44, 59-63, 65, 68].

Whole blood was the most common sample type used (n=12), but cord blood (n=5), peripheral

blood mononuclear cells (PBMCs) (n=4), and cell lines (n=4) were also utilized as well as a range of other

sample types. Several studies reported on more than one sample type.

The reported sample size for the cross-method comparisons ranged from 12 to 181 and only 7

papers had a total sample size greater than 100. Five studies reported the means and standard

deviations and two reported the median and range of TL measurements for the study. Two studies

provided the raw values of the PCR-based TL measurements.

Overall quality of reporting of PCR assay methodology

Of the 20 studies included in this systematic review, the average completion score across both

reporting guidelines was 54%, with an average of 55% for the TRN guidelines and 52% for the Morinha

guidelines (Table 2). Overall, papers included between 26-75% and 29-78% of the recommended

reporting metrics for the TRN and Morinha guidelines, respectively. Some metrics were consistently

reported in nearly all papers, including the sample type, single copy gene name, and type of PCR method

utilized. However, only 10% of the included papers reported on sample storage, PCR efficiencies, or the

number of samples excluded due to quality concerns with the assay.

Table 2. Results of reporting guidelines analysis for TRN and Morinha Reporting Guidelines.

	Category	Average Score
TRN Reporting Guidelines	Sample	41%
	qPCR assay	77%
	Data analysis	42%
	Total	**55%**
Morinha Reporting Guidelines	Sample	47%
	DNA extraction	34%
	qPCR validation	15%
	qPCR protocol	66%
	Data analysis	70%
	Total	**52%**
	Average Total Score	**54%**

247 Average score reflects the average score across all papers for each reporting category. Total scores

248 reflect the average score across all papers from all reporting items.

DNA

For both the TRN guidelines and Morinha guidelines, reporting of sample type, storage, DNA extraction,

and DNA quality/integrity was poor, with an average of 41% for the TRN Sample/DNA category, 47% for

the Morinha Sample category, and 34% for the DNA category of the Morinha guidelines. Storage

conditions for both the biological samples and extracted DNA were poorly reported, with 20% or less of

studies providing this information (Table 3). Less than half of the studies reported on metrics related to

DNA integrity.

Table 3. Item-specific reporting results of the TRN Reporting Guidelines.

Category	Reporting Item	# Papers Reported
Sample type, storage, DNA extraction and integrity	Sample type	20
	Sample storage temperature	4
	Sample storage time before extraction	2
	Sample storage buffer	6
	DNA extraction method	16
	DNA storage conditions	4
	DNA freeze-thaw cycles	4
	Method of documenting DNA quality/integrity	8
	% of samples tested for DNA quality/integrity	6
qPCR assay	State type of PCR method	20
	PCR machine	16
	Source of master mix & reagents	17
	Final reaction volume	15
	Telomere primer sequences	18
	Telomere primer concentration	16
	Single copy gene name	20
	Single gene primer sequence	18
	Singe gene primer concentration	15
	Full PCR program	14

		PCR efficiency of both primers	2
		Source of control samples	11
		Concentration of DNA standard	9
Data analysis		Mean and standard deviation/median and range of TL	10
		Number of sample replicates	19
		Level of independence of replicates	9
		Analytic method to determine final TL	18
		Method of accounting for variation between replicates	11
		Method of accounting for well position effects	4
		Method of accounting for between-plate effects	12
		% of samples repeated as a result of failed QC	3
		% of samples excluded from analysis	2
		Acceptable range of PCR efficiency for primers	5
		ICCs of sample/study groups	2
		T/S ratio transformed to Z score prior to analysis	1

PCR assay

259   Reporting on PCR assay conditions and quality control varied. While many metrics of the PCR

260   assay required in the TRN guidelines were well-reported, only 14 of 20 studies reported the full cycling

261   conditions. The lowest reporting metric related to PCR efficiencies with only 10% reporting on actual

262   PCR efficiencies. Additionally, just over half (55%) of studies reported the source of their control

263   samples.

Analytic approaches

Several key reporting gaps were noted in relation to assay quality control and analytic

approaches to determining final TL. As with any biologic assay, there is the potential that a specific

sample will fail quality control metrics and need to be repeated. Only three papers reported on the

number of samples repeated and/or the number of samples that failed quality control. While all but

one study reported the numbers of sample replicates, surprisingly, only 9 studies reported on the level

of independence of sample replicates (e.g. replicates run on the same plate or on different

plates/different times) and only half of the studies reported the means and standard deviations (or

median and range) of the T/S ratio.

Cross-laboratory studies

274   Only four studies compared analyses across more than one laboratory (Table 1) [40-43]. Of

275   these studies, three described how samples were blinded before analyses. Further, of these cross-

276   laboratory studies, only two studies included the same assay performed in different laboratories [41,

277   43].

Reproducibility

279   Reproducibility, a critical criterion for biologic assays, refers to the relation between

280   measurements using the same assay in different locations or the comparison of values generated using

281   different measurement procedures. This systematic review attempted to assess the relative

282   reproducibility of PCR based measure of TL in different laboratories as well as the reproducibility

283   precision, e.g. the closeness of two or more measurements, in TL measurement using different methods.

Relative reproducibility

The current literature does not provide sufficient data to address the relative reproducibility,

the relation of measurements using the same assay in different laboratories as, to date, only two studies

have tested this directly. In one study that blinded comparison samples before they were sent to the

external laboratories, the median CV across laboratories for qPCR was 18.3%, while the median CV for

STELA/TRF based TL measurement was 9.2%. However given the dependence of CVs on the y-intercept,

the interpretation of these CVs remains challenging [38]. In the other study, where samples were not

blinded before being assayed, the reported within lab CVs for replicate measurements were 2.5% and

8.6%. As the laboratories involved utilized different PCR primers and slightly different methods, it was

not possible to specifically evaluate relative reproducibility.

Reproducibility precision

Reproducibility precision, e.g. the closeness of two or more measurements using different

techniques, was addressed to some extent in 19 of the 20 studies reviewed. However, only four studies

involved assays performed in different laboratories. The correlation of TL measurement with qPCR-

based assays to other assays was, for the most part, reported as linear regression and correlations

coefficients (Figure 4).

Figure 4. Correlation of methods. Weighted average correlation (Pearson's r) and 95% CI for each

method compared with qPCR, MMqPCR, or aTL. LCL: lower confidence limit. UCL: upper confidence

limit. The point sizes for each comparison are weighted by the number of individual papers with that

comparison.

The widest variability in correlation was found with qPCR and Flow-FISH or TRF. Less variation in

the correlation between measurement was found, in general, with MMQPCR and other measures,

however there are, to date, fewer laboratories and studies using this methodology. No studies have

compared MMQPCR to aTL, or whole genome sequencing (WGS). Only one paper each compared qPCR

and qPCR (in separate labs), aTL and TRF, qPCR and aTL, and qPCR with WGS. To our knowledge no

studies have compared WGS data with aTL, although studies have compared TRF and WGS [23].

In five of the papers in this review, linear regression was used to extrapolate TL into kilobases

(kb) from the T/S ratio using TRF values. One paper converted T/S ratio to kb before analysis of the

correlation between methods [44]. In two cases, T/S ratio were converted to kb prior to TL comparison

utilizing Bland-Altman analysis [61, 66]. In two of the five papers, the conversion of the T/S ratio to bp

was based on analyses extrapolated from different data or measured on a different sample type, raising

substantial concerns on the true measurement with uncertain implications for the r value [44, 66].

Beyond concerns related to the source of the data utilized for conversion from T/S to bp before

comparison across methods, this analytic approach likely to leads to inaccurate reporting [44]. Only one

study utilized the TRN-recommended procedure of transformation to z-score before comparison [41].

Repeatability

Repeatability, the precision in measurements that include the same procedure/locations,

revealed the greatest variation in both lab and assay specific precision and between methods. In these

studies, the number of replicates for a specific DNA sample ranged from 1 replicate (e.g. sample

analyzed twice) to five replicates (each sample analyzed six different times). Additionally, only three

reviewed studies reported the number of samples that were repeated due to within replicate variance,

despite clear acknowledgement in the field that a proportion of all studies will ultimately require

repeated assays of telomere length as a result of between-replicate variance. While of limited utility in

confirming precision, the intra-assay reported CVs for PCR-based methods (qPCR, MMqPCR) in this

review ranged from 2.5 to 12%, and the inter-assay CVs ranged from 3.97 to 15.9%. Inter-assay CVs for

TRF ranged from 1.68 to 6.3%, with intra-assay CV reported in only one paper as 1.20%. Inter-assay CVs

for flow-FISH were reported as 9.3% and 10.8% in two papers, with only one reporting an intra-assay CV

of 9.6%. One paper examining the aTL assay reported its inter-assay CV as 6.7% and intra-assay CV as

2.5%. We emphasize, however, that CVs have several flaws which make comparisons of them across

laboratories and methods particularly problematic [38, 50].

Only two studies utilized ICC analyses to examine the repeatability of replicates, reporting ICCs

of 0.89 and 0.92 for PCR-based measurement [62, 64]. To expand data on the repeatability of PCR-based

and other TL methods, raw data was obtained from authors of a subset of these papers, and ICCs

independently calculated. Calculated/reported ICC for TRF methods ranged from 0.92- 0.99 in the

studies included in this review and are consistent with the ICCs reported in existing studies utilizing the

TRF (0.95 to 0.99). However, it is of note that these ICCs were almost entirely the result of TRF

measurement in one laboratory. The ICCs for qPCR-based methods in reviewed papers ranged from

0.89-0.92, including the two reported in manuscripts and an additional ICC calculated from raw data (ICC

= 0.915, SE = 0.023, 95% confidence interval: [0.860, 0. 946], P<0.001; reported CV 6.5%) [38, 69]. ICCs

for MMqPCR (triplicates on the same plate) from one study were run separately based on year of

analyses. In one set (n=873) run across different PCR plates in initial and duplicate runs, the ICCs were

ICC(A,1) = 0.82 (95% CI 0.79-0.84) and ICC(A,k)=0.90 (95% CI 0.88-0.91). Because these samples were re-

run due to initially high intra-assay CVs, this is likely an under-estimate of the true ICC value. For these

same samples, TRF ICCs were calculated from duplicate gels on a subset (n=159) and the inter-gel

ICC(A,1) = 0.92 (95% CI 0.89-0.95) and ICC(A,k)=0.96 (95% CI 0.94-0.97). However, we note that these

TRF analyses were conducted by a trainee which likely decreased repeatability compared to what is

typical of experienced technicians. Given the significant variation in methodologic and raw data

reporting, and the wide variability in published CVs, it is likely that the majority of existing TL studies not

specifically comparing methods would have significantly lower ICCs.

Determination of effects of ICC variability on sample size and study

power

Our systematic review revealed wide variation in TL measurement repeatability. No biologic

assay is perfect, and laboratories measuring any biologic substrate vary in their own internal quality

control and repeatability. To provide general guidance for investigators, we therefore conducted

analyses to evaluate the impact on power and sample size across a range of ICCs.

In Figure 5A, we present the sample size required to test effect sizes of 150, 200, and 300 bp

with a t-test with a power of 0.9, as a function of measurement error as expressed in the ICC. To

contextualize the differences: 150 bp is the approximate difference found between the sexes, and 300

bp is the approximate difference observed between individuals with and without atherosclerotic

cardiovascular disease [70]. As directly converting bp to T/S ratios is not feasible in this analysis, we

suggest that investigators using PCR-based methods conservatively estimate their potential bp

difference from the existing literature related to their exposure or outcome of interest when considering

study design and sample size. Calculations assumed a realistic (true) standard deviation of 650 bp and

power analysis was done using G*Power [52]. N is the combined n of the two groups to be compared

and was assumed to be equally distributed among the two groups.

Fig 5A. ICC effect on statistical power (cross-sectional studies). The sample size required to test effect

sizes of 150, 200 and 300 bp with a t-test with a power of 0.9, as a function of measurement error as

expressed in the ICC (intraclass correlation coefficient). Calculations assumed a realistic (true) standard

deviation of 650 bp and power analysis was done using G*Power. N is the combined n of the two groups

to be compared and was assumed to be equally distributed among the two groups.

Fig. 5B and 5C present the power of different sample sizes to detect a 200bp difference at

various ICC levels with a presumed mean TL of 650bp in a longitudinal study. For small sample sizes, e.g.

25 subjects, the impact of even moderate variability in ICCs on power is substantial. This calculation was

based on the average loss of 25 base pairs per year in adults. The rate of base pair loss in infants and

children is likely significantly different and, but as of yet is poorly characterized.

Fig 5B, C. ICC effect on statistical power (longitudinal studies). Statistical power to detect a significant

change (paired-t-test) in telomere length of 25 bp/yearfor sample sizes of 25, 50 and 100 individuals,

and with an interval of 4 (B) or 8 years (C) between baseline and follow-up(i.e. respectively on average

100 and 200 bp in total), as a function of measurement reliability expressed as the Intraclass Correlation

Coefficient. Population SD of true telomere length (i.e. without measurement error) was transformed to

be 0.65 kb at both time points and telomere shortening was simulated assuming a Poisson distribution

with mean / variance of 25.

Discussion:

This systematic review found a total of 20 papers documenting comparison between TL

measured using a PCR based methodology and another TL assay. Until recently, there were no

publication reporting guidelines for PCR TL measurement. Our review specifically focused on method

comparison studies with the expectation that reporting of critical assay parameters would be expected

to be higher than in typical studies. Our analyses, using two separately developed reporting guidelines,

found that on average only half of the recommended factors were documented, indicating the need for

increased methodologic reporting. The lowest reporting was related to information about the validation

of PCR-based assays outlined in the Morinha guidelines, with only five papers including any of the

recommended factors. PCR efficiencies, a key reporting requirement in both guidelines and also

considered essential in the MIQE guidelines, was absent from the majority of papers with only five

mentioning the PCR efficiency parameters and only two documenting the actual PCR efficiencies. Given

that all PCR-based methods either assume or specifically calculate the PCR efficiency when determining

the T/S ratio, the absence of this key metric is concerning, particularly if there are significant differences

in PCR efficiencies between the single copy and the telomere primers. A surprising number of studies

failed to comment on pre-analytic factors, specifically sample storage time and conditions, freeze-thaw

cycles, and evaluation of DNA quality and integrity, all potential sources of assay variability for both PCR

and non-PCR based TL assays that may contribute to current debates in the field about the utility of TL

[67, 71, 72]. Lastly, the reporting of the number of samples failing initial quality control, repeated, or

unable to be assayed was low. In laboratories routinely performing TL measurement using any assay, a

certain percentage of samples for each study will require repeating and regularly a small subset may be

unanalyzable for various reasons. While it is possible that these factors were considered and monitored,

the lack of reporting for this metric heightens the need for increased attention to the proposed

reporting recommendations. Moving forward, the widespread dissemination of these qPCR reporting

guidelines to study sections, peer reviewers, and scientists represents an important next step in

enhancing the scientific rigor of the field.

At this time, evaluation of the existing literature fails to provide sufficient evidence of the

relative or precision reproducibility of different TL assays. Our review identified only four studies that

included cross laboratory comparisons and, of these, only two evaluated PCR-based assays performed in

more than one lab. As the number of laboratories performing TL studies using PCR and other methods

continue to increase, the lack of clear data about cross-laboratory reproducibility and the absence of

existing DNA standards or other methods to account for cross-laboratory variation substantially limits

the ability to characterize relative reproducibility. In terms of precision reproducibility across different

methods (e.g. PCR and TRF, or PCR and FISH), the current variability in findings, particularly when

coupled with limited methodologic reporting, highlights the need for additional rigorous and blinded

cross laboratory studies that are adequately powered to accurately determine how TL in a population

measured using different assays truly relates. Although eleven studies evaluated the relationship

between qPCR and TRF, due to the wide range in reported correlations between TL measurements, the

relatively small samples sizes, and the insufficient analytic and assay blinding, there is currently

insufficient data to draw firm conclusions on the general correlation between TL measured with

different assays. The analytic consequences of using CVs to test the relationship between TL measured

using different assays has been discussed previously, as has the issues caused by the use of analytic

strategies such as conversion to base pairs instead of z scores, especially when extrapolating from data

not produced in the same laboratory or with the same samples [38, 49]. In this review, we utilized

existing raw data from 3 included studies to provide preliminary data about precision reproducibility for

PCR and TRF studies. The wide range of ICCs calculated from these few studies, particularly for PCR-

based methods, and the low reporting of ICCs in the papers included in this review highlights the need

to increase attention to this specific reporting requirement. For many of the existing studies, the small

sample size and the lack of reporting of the means and standard deviations of TL prevents objective

determination of whether any of the current studies were adequately powered. Beyond these concerns,

as many studies utilized data, particularly for TRF, from a small number of specialized laboratories, the

applicability of much of the existing data to the wider telomere field is uncertain. For aTL and MMqPCR

based TL measurements, the current paucity of published cross method comparisons limits the ability to

form an opinion of how TL measured with these assays relates to other methods.

Measurement precision is critical. Methodologies that are low cost, practical, and simple to

implement with standard laboratory equipment, especially when they are innovative or high impact, are

often rapidly implemented across laboratories with various levels of expertise in the new methodology.

Invariably this results in diverse protocols, analyses, and methodologic reporting, consequences that are

even more problematic when there is an absence of consensus on best practices[48]. As with many

other biologic assays, the development of reporting guidelines for TL measurement has lagged behind

the broad implementation of the methods themselves [73-77]. The lack of consolidated guidance about

factors, both pre-analytic and within the assay itself, that contribute to measurement error when

combined with the wide popularity of PCR-based TL measurement undoubtedly contributed to

discrepancies in the existing literature and failed study replications. Similar to the MIQE guidelines, the

reporting guidelines presented and tested in this systematic review for PCR-based TL assays are meant

as minimal reporting recommendations focused on enhancing the reliability of results, consistency

between different laboratories performing the same assay, and increased experimental transparency

and accuracy [48]. To assist investigators and reviewers we highlight the overlapping recommendations

with the MIQE guidelines, indicate whether a particular requirement is desirable or essential, and

provide references that support the selection of the particular reporting requirement. Over the course

of the next four years, the TRN expects to develop similar reporting recommendations for other types of

TL assays while conducting adequately powered and scientifically rigorous studies to support these

reporting guidelines, recognizing that individual recommendations have varying levels of initial empirical

support [78].

Despite the strengths of this review, there are several limitations. First, this review only focused

on assays applicable to population-based studies in humans. It does not address issues in other species

or assays that may have clinical utility but for which the requirements for sample types (e.g. fresh tissue

and/or live cells) or the cost/labor/expertise requirements (e.g. TESLA) limit utilization in population

based studies. As part of its scientific charge to establish telomeres as indicators of psychosocial and

environmental exposures and sentinels of disease, the TRN also expects to support both the adaptation

of existing TL assays not currently scalable to population studies as well as the discovery of new

methodologies. A second limitation is that we utilized reporting guidelines for qPCR-based assays only.

To date, specific protocol recommendations and reporting guidelines have not been published for other

TL assays (e.g. TRF, FISH) although detailed methodologic protocols do exist [79]. Finally, we note that

while this article primarily focused on precision and reproducibility, accuracy of measurement is also

critically important. Precisely inaccurate measures will be of no use to the scientific field. In the absence

of a clear gold standard measurement technique, accuracy is difficult to discern. In addition to validating

methods across laboratories, we suggest using measures known to be correlated with TL (e.g. age,

familial relatedness, baseline to follow up TL correlations) to better establish external validity.

Conclusions

After careful examination of the existing literature, it is apparent that rigorous cross laboratory

and methodological studies must be an immediate priority for the TRN and the field. To assist the field

moving forward, we include reporting guidelines for PCR-based TL assays and indicate specific scientific

papers that support these recommendations originally developed through consensus of the initial TRN

members and consultants. These guidelines do not outline a specific PCR methodology and, at this time,

we do not believe there is sufficient data to provide guidance on specific assay approaches or

components. Rather, these guidelines are provided to ensure reviewers and readers can adequately

assess the methodology and consider the implications of these factors for each study's findings. The

consistency in results across reporting guidelines (TRN, Morinha, MIQE) related to the integrity and

quality of both the initial biological sample and the DNA itself support the critical nature of this

reporting metric. In terms of assay reporting, increased attention of investigators and reviewers to

ensuring complete reporting of assay reagents and PCR efficiencies is also expected to enhance the rigor

of the field. TRN investigators are currently testing the impact of different pre-analytic factors, DNA

integrity, and PCR conditions to provide extant evidence of the importance of these parameters in

relation to precision and reproducibility. We recommend that studies be required to report ICCs in lieu

of CVs, as well as either the median or mean and standard deviation of TL. We also provide specific

guidance related to sample size and power that is contingent upon the ICC given the substantial impact

of differences in assay precision on the ability to determine true relationships and with the expectation

that this will be of use for investigators as they develop new research questions. It is important to

balance assay cost, in terms of both time and reagents, with the needed sample size and statistical

power. Moving forward, investigators should carefully consider study design from this perspective,

recognizing that there is currently no "ideal" approach. Telomere research offers significant potential

across a diverse range of scientific fields with potential mechanistic insight into overlapping biological

pathways contributing to many of the leading causes of morbidity and mortality. Ensuring the highest

scientific rigor and precision, through accurate methodological reporting and rigorous testing of the

factors that contribute to assay variability, are requisite steps to ensuring that potential is achieved.

Acknowledgements

The authors would like to acknowledge the TRN Advisory Committee (Daniel Belsky, Eileen Crimmins,

Sonja Entringer, Daniel Nettle, Jerry Shay, Daniel Nettle), U01 grantee P.I.s (Abraham Aviv, Jue Lin, Idan

Shalev, Yun-Ling Zheng), and NIA/NIEHS partners (Max Guo, Michelle Heacock, Lisbeth Nielsen) for their

input in creating the TRN qPCR Reporting Guidelines.

References

1. Olonikov, A., *A theory of marginotomy.* J Theor Biol, 1971. **41**: p. 181-190.
2. Blackburn, E. and J. Gall, *A tandemly repeated sequence at the termini of the extrachromosomal*
ribosomal RNA genes in Tetrahymena. Journal of molecular biology, 1978. **120**(1): p. 33-53.
3. Moyzis, R., et al., *A highly conserved repetitive DNA sequence,(TTAGGG) n, present at the*
telomeres of human chromosomes. Proceedings of the National Academy of Sciences, 1988.
85(18): p. 6622-6626.
4. Therkelsen, A., et al., *Staining of human telomeres with primed in situ labeling (PRINS).*
Cytogenet Cell Genet, 1995. **68**(1-2): p. 115-8.
5. Lansdorp, P., et al., *Heterogeneity in telomere length of human chromosomes.* Hum Mol Genet,
1996. **5**(5): p. 685-91.
6. Bryant, J., et al., *Measurement of telomeric DNA content in human tissues.* Biotechniques, 1997.
23(3): p. 476-8, 480, 482, passim.
7. de Pauw, E., et al., *Assessment of telomere length in hematopoietic interphase cells using in situ*
hybridization and digital fluorescence microscopy. Cytometry, 1998. **32**(3): p. 163-9.
8. Hultdin, M., et al., *Telomere analysis by fluorescence in situ hybridization and flow cytometry.*
Nucleic Acids Res, 1998. **26**(16): p. 3651-6.
9. Narath, R., et al., *Automatic telomere length measurements in interphase nuclei by IQ-FISH.*
Cytometry Part A: the Journal of the International Society for Analytical Cytology, 2005. **68**(2): p.
113-120.
10. Nakamura, Y., et al., *Simple, rapid, quantitative, and sensitive detection of telomere repeats in*
cell lysate by a hybridization protection assay. Clin Chem, 1999. **45**(10): p. 1718-24.
11. Law, H. and Y. Lau, *Validation and development of quantitative flow cytometry-based*
fluorescence in situ hybridization for intercenter comparison of telomere length measurement.
Cytometry: The Journal of the International Society for Analytical Cytology, 2001. **43**(2): p. 150-
153.
12. Rufer, N., et al., *Telomere length dynamics in human lymphocyte subpopulations measured by*
flow cytometry. Nat Biotechnol, 1998. **16**(8): p. 743-7.
13. Cawthon, R., *Telomere measurement by quatitative PCR.* Nucleic Acids Research, 2002. **30**(10):
p. e47- 53.
14. Meeker, A.K., et al., *Telomere length assessment in human archival tissues: combined telomere*
fluorescence in situ hybridization and immunostaining. The American Journal of Pathology, 2002.
160(4): p. 1259-1268.
15. Perner, S., et al., *Quantifying Telomere Lengths of Human Individual Chromosome Arms by*
Centromere-Calibrated Fluorescence in Situ Hybridization and Digital Imaging. The American
Journal of Pathology, 2003. **163**(5): p. 1751-1756.
16. Baird, D., et al., *Extensive allelic variation and ultrashort telomeres in senescent human cells.* Nat
Genet, 2003. **33**(2): p. 203-7.
17. Canela, A., et al., *High-throughput telomere length quantification by FISH and its application to*
human population studies. Proc Natl Acad Sci U S A, 2007. **104**(13): p. 5300-5.
18. Cawthon, R., *Telomere length measurement by a novel monochrome multiplex quantitative PCR*
method. Nucleic Acids Research, 2009. **37**: p. e21.
19. Bendix, L., et al., *The load of short telomeres, estimated by a new method, Universal STELA,*
correlates with number of senescent cells. Aging Cell, 2010. **9**(3): p. 383-97.

20. O'Callaghan, N. and M. Fenech, *A quantitative PCR method for measuring absolute telomere*
length. Biological procedures online, 2011. **13**: p. 3-3.
21. Kimura, M. and A. Aviv, *Measurement of telomere DNA content by dot blot analysis.* Nucleic
Acids Research, 2011. **39**(12): p. e84-e84.
22. Kibriya, M., et al., *Measurement of telomere length: a new assay using QuantiGene chemistry on*
a Luminex platform. 2014, AACR.
23. Ding, Z., et al., *Estimating telomere length from whole genome sequence data.* Nucleic Acids Res,
2014. **42**(9): p. e75.
24. Mender, I. and J. Shay, *Telomere Dysfunction Induced Foci (TIF) Analysis.* Bio-protocol, 2015.
5(22): p. e1656.
25. Nersisyan, L. and A. Arakelyan, *Computel: Computation of Mean Telomere Length from Whole-*
Genome Next-Generation Sequencing Data. PLOS ONE, 2015. **10**(4): p. e0125201.
26. Lai, T., et al., *A method for measuring the distribution of the shortest telomeres in cells and*
tissues. Nature communications, 2017. **8**(1): p. 1-14.
27. McCaffrey, J., et al., *High-throughput single-molecule telomere characterization.* Genome
research, 2017. **27**(11): p. 1904-1915.
28. Jasmine, F., et al., *A novel pooled-sample multiplex luminex assay for high-throughput*
measurement of relative telomere length. American Journal of Human Biology, 2018. **30**(4): p.
e23118.
29. Dahlgren, P., et al., *Development of a New Monochrome Multiplex qPCR Method for Relative*
Telomere Length Measurement in Cancer. Neoplasia, 2018. **20**(5): p. 425-431.
30. Farmery, J.H.R., et al., *Telomerecat: A ploidy-agnostic method for estimating telomere length*
from whole genome sequencing data. Scientific Reports, 2018. **8**(1): p. 1300.
31. Lu, A., et al., *DNA methylation-based estimator of telomere length.* Aging, 2019. **11**(16): p. 5895-
5923.
32. Feuerbach, L., et al., *TelomereHunter–in silico estimation of telomere content and composition*
from cancer genomes. BMC bioinformatics, 2019. **20**(1): p. 1-11.
33. de Pedro, N., et al., *Analytical Validation of Telomere Analysis Technology® for the High-*
Throughput Analysis of Multiple Telomere-Associated Variables. Biological Procedures Online,
2020. **22**(1): p. 2.
34. Hemann, M., et al., *The shortest telomere, not average telomere length, is critical for cell*
viability and chromosome stability. Cell, 2001. **107**(1): p. 67-77.
35. Xu, Z., et al., *The length of the shortest telomere as the major determinant of the onset of*
replicative senescence. Genetics, 2013. **194**(4): p. 847-857.
36. Montpetit, A., et al., *Telomere length: a review of methods for measurement.* Nursing research,
2014. **63**(4): p. 289.
37. Aubert, G., M. Hills, and P. Lansdorp, *Telomere length measurement-caveats and a critical*
assessment of the available technologies and tools. Mutat Res, 2012. **730**(1-2): p. 59-67.
38. Eisenberg, D., *Telomere length measurement validity: the coefficient of variation is invalid and*
cannot be used to compare quantitative polymerase chain reaction and Southern blot telomere
length measurement techniques. International Journal of Epidemiology, 2016. **45**(4): p. 1295-
1298.
39. Ehrlich, M., *DNA hypermethylation in disease: mechanisms and clinical relevance.* Epigenetics,
2019. **14**(12): p. 1141-1163.
40. Aviv, A., et al., *Impartial comparative analysis of measurement of leukocyte telomere*
length/DNA content by Southern blots and qPCR. Nucleic Acids Research, 2011. **39**(20): p. e134.
41. Martin-Ruiz, C., et al., *Reproducibility of telomere length assessment: an international*
collaborative study. International journal of epidemiology, 2015. **44**(5): p. 1673-1683.

42.   Eisenberg, D., C. Kuzawa, and M.G. Hayes, *Improving qPCR telomere length assays: Controlling*
600   *for well position effects increases statistical power.* American Journal of Human Biology, 2015: p.
601   n/a-n/a.
602   43.   Gadalla, S., et al., *No association between donor telomere length and outcomes after allogeneic*
603   *unrelated hematopoietic cell transplant in patients with acute leukemia.* Bone marrow
604   transplantation, 2018. **53**(4): p. 383-391.
605   44.   Gutierrez-Rodrigues, F., et al., *Direct Comparison of Flow-FISH and qPCR as Diagnostic Tests for*
606   *Telomere Length Measurement in Humans.* PLoS ONE, 2014. **9**(11): p. e113747.
607   45.   Bodelon, C., S. Savage, and S. Gadalla, *Telomeres in molecular epidemiology studies.* Prog Mol
608   Biol Transl Sci, 2014. **125**: p. 113-31.
609   46.   Plant, A., R. Hanisch, and National Institute of Standards and Technology, *Reproducibility and*
610   *Replicability in Science, A Metrology Perspective* 2018: A Report to the National Academies of
611   Sciences, Engineering and Medicine Committee on Reproducibility and Replicability in Science.
612   47.   Morinha, F., P. Magalhães, and G. Blanco, *Standard guidelines for the publication of telomere*
613   *qPCR results in evolutionary ecology.* Molecular Ecology Resources, 2020. **20**(3).
614   48.   Bustin, S., et al., *The MIQE Guidelines: Minimum Information for Publication of Quantitative*
615   *Real-Time PCR Experiments.* Clinical Chemistry, 2009. **55**(4): p. 611-622.
616   49.   Verhulst, S., et al., *Response to: Reliability and validity of telomere length measurements.*
617   International journal of epidemiology, 2016. **45**(4): p. 1298-1301.
618   50.   Nettle, D., et al., *Consequences of measurement error in qPCR telomere data: A simulation study.*
619   PLoS One, 2019. **14**(5): p. e0216118.
620   51.   McGraw, K. and S. Wong, *" Forming inferences about some intraclass correlations coefficients":*
621   *Correction.* 1996.
622   52.   Faul, F., et al., *Statistical power analyses using G* Power 3.1: Tests for correlation and regression*
623   *analyses.* Behavior research methods, 2009. **41**(4): p. 1149-1160.
624   53.   Ehrlenbach, S., et al., *Influences on the reduction of relative telomere length over 10 years in the*
625   *population-based Bruneck Study: introduction of a well-controlled high-throughput assay.*
626   International Journal of Epidemiology, 2009. **38**(6): p. 1725-1734.
627   54.   Pavesi, E., et al., *Analysis of telomeres in peripheral blood cells from patients with bone marrow*
628   *failure.* Pediatric Blood & Cancer, 2009. **53**(3): p. 411-416.
629   55.   Imam, T., et al., *Leukocyte Telomere Length in HIV-Infected Pregnant Women Treated With*
630   *Antiretroviral Drugs During Pregnancy and Their Uninfected Infants.* JAIDS Journal of Acquired
631   Immune Deficiency Syndromes, 2012. **60**(5): p. 495-502.
632   56.   Zanet, D.L., et al., *Blood and Dried Blood Spot Telomere Length Measurement by qPCR: Assay*
633   *Considerations.* PLOS ONE, 2013. **8**(2): p. e57787.
634   57.   Jodczyk, S., et al., *Telomere length measurement on the Roche LightCycler 480 Platform.* Genet
635   Test Mol Biomarkers, 2015. **19**(2): p. 63-8.
636   58.   Panero, J., et al., *Absolute qPCR for measuring telomere length in bone marrow samples of*
637   *plasma cell disorders.* Molecular biotechnology, 2015. **57**(2): p. 155-159.
638   59.   Hsieh, A., et al., *Optimization of a relative telomere length assay by monochromatic multiplex*
639   *real-time quantitative PCR on the LightCycler 480: sources of variability and quality control*
640   *considerations.* The Journal of Molecular Diagnostics, 2016. **18**(3): p. 425-437.
641   60.   Wand, T., et al., *Telomere content measurement in human hematopoietic cells: Comparative*
642   *analysis of qPCR and Flow-FISH techniques.* Cytometry A, 2016. **89**(10): p. 914-921.
643   61.   Behrens, Y., et al., *Comparison of different methods for telomere length measurement in whole*
644   *blood and blood cell subsets: recommendations for telomere length measurement in*
645   *hematological diseases.* Genes, Chromosomes and Cancer, 2017. **56**(9): p. 700-708.

62. Khincha, P., et al., *Correlation of leukocyte telomere length measurement methods in patients*
with dyskeratosis congenita and in their unaffected relatives. International journal of molecular
sciences, 2017. **18**(8): p. 1765.
63. Lee, M., et al., *Comparative analysis of whole genome sequencing-based telomere length*
measurement techniques. Methods, 2017. **114**: p. 4-15.
64. Tarik, M., et al., *Validation of quantitative polymerase chain reaction with Southern blot method*
for telomere length analysis. Future Science OA, 2018. **4**(4): p. FSO282.
65. Wang, Y., et al., *Telomere Length Calibration from qPCR Measurement: Limitations of Current*
Method. Cells, 2018. **7**(11): p. 183.
66. Ferreira, M., et al., *Comparison of flow-FISH and MM-qPCR telomere length assessment*
techniques for the screening of telomeropathies. Annals of the New York Academy of Sciences,
2019: p. 1-11.
67. Ropio, J., et al., *Reliable blood cancer cells' telomere length evaluation by qPCR.* Cancer Med,
2020. **9**(9): p. 3153-3162.
68. Ropio, J., et al., *Reliable blood cancer cells' telomere length evaluation by qPCR.* Cancer
Medicine, 2020.
69. Verhulst, S., et al., *Commentary: The reliability of telomere length measurements.* International
Journal of Epidemiology, 2015. **44**(5): p. 1683-1686.
70. Benetos, A., et al., *Response by Benetos et al to letter regarding article,"Short leukocyte*
telomere length precedes clinical expression of atherosclerosis: the blood-and-muscle model".
Circulation research, 2018. **122**(8): p. e73-e74.
71. Dlouha, D., et al., *Comparison of the relative telomere length measured in leukocytes and eleven*
different human tissues. Physiol Res, 2014. **63 Suppl 3**: p. S343-50.
72. Dagnall, C., et al., *Effect of pre-analytic variables on the reproducibility of qPCR relative telomere*
length measurement. PLOS ONE, 2017. **12**(9): p. e0184098.
73. Brazma, A., et al., *Minimum information about a microarray experiment (MIAME)—toward*
standards for microarray data. Nature genetics, 2001. **29**(4): p. 365-371.
74. Taylor, C., et al., *The minimum information about a proteomics experiment (MIAPE).* Nature
biotechnology, 2007. **25**(8): p. 887-893.
75. Field, D., et al., *The minimum information about a genome sequence (MIGS) specification.*
Nature biotechnology, 2008. **26**(5): p. 541-547.
76. Taylor, C., et al., *Promoting coherent minimum reporting guidelines for biological and biomedical*
investigations: the MIBBI project. Nature biotechnology, 2008. **26**(8): p. 889-896.
77. Fiehn, O., et al., *The metabolomics standards initiative (MSI).* Metabolomics, 2007. **3**(3): p. 175-
178.
78. Brazma, A., *Minimum information about a microarray experiment (MIAME)–successes, failures,*
challenges. TheScientificWorldJournal, 2009. **9**.
79. Kimura, M., et al., *Measurement of telomere length by the Southern blot analysis of terminal*
restriction fragment lengths. Nature protocols, 2010. **5**(9): p. 1596-1607.

Supporting information

S1 Table. Telomere Research Network (TRN) Reporting Guidelines. Reporting items of the TRN

Guidelines by category, with assigned importance for each item and comparison with the related

assigned importance of the MIQE guidelines. E: essential information, should be submitted with the

manuscript. D: desired information, should be submitted if available.

S2 Table. Morinha Reporting Guidelines. Reporting items of the guidelines developed by Morinha et al.

by category, with assigned importance for each item and comparison with the related assigned

importance of the MIQE guidelines. A: very important. B: somewhat important. C: somewhat important,

should be submitted if available. E: essential information, should be submitted with the manuscript. D:

desired information, should be submitted if available.

S3 Table. Grading rubric for TRN and Morinha Reporting Guidelines.

S4 Table. Results of reporting guidelines for individual papers. Scores for each category are provided as

a percentage of items in that category sufficiently reported. A total score for each set of reporting

guidelines are calculated from all items as well as an average overall score between the two reporting

guidelines.

Figure 1

Figure 2

PRISMA 2009 Flow Diagram

Identification

Records identified through database searching terms "telomere," "human," "length" (n =5427)

Additional records identified through other sources (n = 25)

Records after inclusion of "PCR" (n = 767)

Screening

Abstracts screened (n = 792)

Abstracts excluded (n = 727)

Eligibility

Full-text articles assessed for eligibility (n = 65)

Full-text articles excluded, reviews/not comparison (n = 18)

Full-text articles excluded, novel method validation (n = 26)

Full-text articles excluded, comparison of different samples (n = 1)

Included

Studies included in qualitative synthesis (n = 20)

Figure 3

Figure 4

Figure 5B 5C

Figure 5A