Microfluidic Antibody Affinity Profiling for In-Solution Characterisation of Alloantibody – HLA Interactions in Human Serum

Matthias M. Schneider,[†,#] Tom Scheidt,[†,#] Ashley J. Priddey,[‡,#] Catherine K. Xu,[†,#] Mengsha Hu,[†,#] Sean R. A. Devenish,[¶] Georg Meisl,[†] Christopher M. Dobson,[†,@] Vasilis Kosmoliaptsis,[*,‡,§,‖,△] and Tuomas P. J. Knowles[*,†,⊥,△]

†Centre for Misfolding Diseases, Department of Chemistry, University of Cambridge, Lensfield Road, Cambridge CB2 1EW, United Kingdom

‡Department of Surgery, University of Cambridge, Addenbrooke's Hospital, Hills Road, Cambridge CB2 0QQ, UK

¶Fluidic Analytics, Unit A, The Paddocks Business Centre, Cherry Hinton Rd, Cambridge CB1 8DH, UK

§NIHR Blood and Transplant Research Unit in Organ Donation and Transplantation, University of Cambridge, Hills Road, Cambridge CB2 0QQ, UK

‖NIHR Cambridge Biomedical Research Centre, Hills Road, Cambridge CB2 0QQ, UK

⊥Cavendish Laboratory, Department of Physics, University of Cambridge, JJ Thomson Ave, Cambridge CB3 0HE, UK

#contributed equally

@Passed away September 2019

△contributed equally as corresponding authors

E-mail: vk256@cam.ac.uk; tpjk2@cam.ac.uk

Abstract

The detection and characterisation of antibodies in human blood is a key for clinical diagnostics and risk assessmentn for autoimmunity, infectious diseases and transplantation. Antibody titre derived from immunoassays is a commonly used measure for antibody response, but this metric does not resolve readily the two fundamental properties of antibodies in solution, namely their affinity and concentration. This difficulty originates from the fact that the fundamental parameters describing the binding interaction, affinity and ligand concentration, are convoluted into the titre measurement; moreover, the difficulty of controlling the surface concentration and activity of the immobilised ligand can make it challenging to distinguish between avidity and affinity. To address these challenges, we developed microfluidic antibody affinity profiling, an assay which allows the simultaneous determination of both affinity and antibody concentration, directly in solution, without surface immobilisation or antibody purification. We demonstrate these measurements in the context of alloantibody characterisation in organ transplantation, using complex patient sera, and quantify the concentration and affinity of alloantibodies against donor Human Leukocyte Antigens (HLA), an extensively used clinical biomarker to access the risk of allograft rejection. These results outline a path towards detection and in depth profiling of antibody response in patient sera.

Introduction

Non-covalent protein-protein interactions underlie many biological and physiological processes, including protein self-assembly,[1–3] protein aggregation,[4–6] antibody-antigen recognition,[7–9] muscle contraction,[10] and cellular communication.[11] One fundamental application of measuring protein-protein interactions is based on immuno-assays for the detection of biomarkers, in body fluids, primarily human serum, related to a various diseases including cancer,[12–15] protein misfolding diseases,[16–21] auto-immune diseases,[22] and graft rejection.[23] In the latter, detection of antibodies against donor Human Leukocyte Antigens (HLA), termed

alloantibodies, in patient serum serve as strong indicator for potential rejection of transplants. The analysis and characterisation of this biomarkers is essential for pre-transplant assessment and post-transplant immune monitoring.[24–27]

Current approaches for the detection and characterisation of antibodies, including alloantibodies, rely mostly on surface immobilisation of one of the binding partners, such as in enzyme-linked immunosorbant assays (ELISA),[28–30] bead-based multiplex assays,[31–34] and surface plasmon resonance (SPR) spectroscopy.[35–37] The requirement for protein immobilisation is associated with a number of challenges, such as non-specific interactions with the surface or suppressed accessibility due to alternations of substrate and ligand structures.[38–40] Furthermore, avidity effects caused by dense substrate occupation are hard to control in surface based measurements.[41] Additionally, the hook/prozone effect, most prevalent in sandwich immunoassays or complement interference, can result in false negative measurements.[42–44] More generally, for measurements on surfaces, the fundamental parameters of antibodies in solution, namely their affinity and concentration, are challenging to resolve due to the difficulty in controlling the concentration of the surface-bound species. For instance, the commonly used EC50 value obtained in surface measurements is only weakly dependent on the affinity of the interaction for strong antibody binding, as illustrated in Fig.1A.

To overcome the limitations of surface-based immunoassays, we demonstrate here an in-solution Microfluidic Antibody Affinity Profiling (MAAP) appporach, allowing quantitative measurement of biophysical parameters governing specific antibody-antigen interactions in unpurified human serum and its complex background with more than 10 million different proteins.[45] Unlike existing microfluidic immunoassays, many of which rely on surface immobilisation,[46,47] the approach described in this paper operates fully in solution. We use a strategy based on microfluidic diffusional sizing, which tracks the spatial and temporal evolution of a fluorescently labelled protein in a microfluidic channel under laminar flow conditions to determine its hydrodynamic radius, R_h (Fig. 1) and hence effective molecu-

lar weight. When an antigen molecule interacts with an antibody in solution, its effective molecular weight increases to that of the antibody-antigen complex and hence its diffusion coefficient decreases.

We focus here on humoral responses against HLA, also known as the human major histocompatibility complex (MHC), a biomolecule of key clinical significance. The extensive polymorphism of the HLA system, evolved to enable immune protection against an ever-

Figure 1: (**A**) Basic Principle of surface based antibody-binding measurements. In a strong binding regime, the antibody concentration can be determined, in a weak binding regime the ratio between antibody concentration and dissociation constant, K_d, can be determined only. (**B**) Basic principle of applying MDS for clinical samples. (1) Patient-derived human serum (2) is incubated with labelled HLA to allow binding. (3) The effective size of the complex is determined by microfluidic diffusional sizing, from which (4) the dissociation constant K_d and the antibody concentration $[Ab]$ are determined. As it is an in-solution approach, both K_d and antibody concentration become accessible. The posterior shows the probability distribution, whereby yellow stands for high, blue for low probability that the according parameters K_d and $[Ab]$ are found at the respective values.

changing environment of human pathogens, is a major barrier in organ and cell transplantation.[48] Exposure to donor HLA through pregnancy, transfusion and/or transplantation leads to development of alloantibodies, principal mediators of acute and chronic allograft loss.[23,49] Detection and characterisation of alloantibodies is essential for evaluation of donor-recipient compatibility and to facilitate post-transplant immune monitoring and provide individual therapeutic intervention.[50–52] In current clinical practice, this is mainly performed using solid phase assays which suffer from the aforementioned disadvantages of surface-based techniques. Such semi-quantitative approaches do not allow the full characterisation of fundamental biophysical properties of the humoral alloresponse such as alloantibody levels (concentration) and the affinity of alloantibody-HLA interaction.[27,53–55] In the following, we show that MAAP is an advanced technique capable of quantifying an analyte in human serum under native solution conditions to yield physiologically relevant results and thus propose this platform as an additional proceudre for immuno-profiling in human serum.

Results and Discussion

Binding Interactions of Covalently Labelled HLA

In order to detect alloantibodies in solution using HLA, a microfluidic platform relying on diffusional sizing (MDS) was used that enables determination of binding parameters utilising the measured hydrodynamic radius, R_h, of a fluorescently labelled protein, as previously described.[56–59] We used a reliable and stable labelling strategy for fluorescence detection, relying on NHS-Chemistry (Fig. S1A) for N-terminal labelling, which yields highly pure HLA with labelling stoichiometry of 0.33 to 1.55, depending on the variant (Fig. S1C-F). This allows to control the stoichiometry of the binding interaction more accurately than traditional strategies utilising biotin-streptavidin-HLA complexes which are highly heterogeneous (Fig. S1B). A fluorophore in the far-red spectral region was chosen ($\lambda_{\text{em, max}} = 650$ nm), since serum autofluorescence is minimised in this spectral region (Fig. S2E).

Rapid and physiologically accurate investigation of antibodies in human samples is crucial for clinical evaluation. Therefore, immunoassays must be able to cope with untreated samples and not be influenced by non-specific binding to surfaces, by the hook/prozone-effect, and by surface-mediated avidity effects.[42–44] To validate the applicability of the assay for quantification of alloantibody-HLA interactions, the binding of the HLA variant A*03:01 to the mouse derived anti-human monoclonal antibody W6/32, which specifically recognises a monomorphic epitope on all HLA class I molecules that includes both the heavy chain and the β_2-microglobulin chain,[60] was investigated first. The hydrodynamic radius, R_h, of pure HLA A*03:01 was determined to be 3.47 $\pm$ 0.13 nm (Fig. 2A), which is consistent with the expected radius for a natively folded protein with a molecular weight of approximately 55 kDa (Fig. 2B). Upon addition of a 200-fold excess of antibody W6/32, a significant increase of the hydrodynamic radius to $R_h = 5.01 \pm 0.13$ nm was observed, indicating an interaction between both species. The hydrodynamic radius of 5.35 nm suggests that two antigens are bound per antibody, which is expected for a bivalent IgG antibody which has two binding sites, and is consistent with the expected size for a natively folded protein complex of 260 kDa (Fig. 2B). Negative control experiments, including alloantibody against fluorescently labelled BSA (Fig. S4A), OUW4F11 (an alloantibody specific to HLA B*08:01) against HLA A*03:01 (Fig 2A), and human IgG binding against HLA A*03:01 (Fig. S4B), did not show an increase in R_h and, therefore, confirmed that the complex formation is based only on specific interactions.

We next explored whether this approach could yield both the dissociation constant of the interaction and the concentration of the antibody. To this effect, an equilibrium binding curve for the interaction between HLA A*03:01 and antibody W6/32 was measured, yielding a $K_d = 0.7$ [0.3, 1.6] nM (95% confidence intervals from Bayesian inference given in square brackets) consistent with a binding ratio of 1 to 2, i.e. a stoichiometry of 2 antigens per antibody (Fig. 2C and Fig. S5). Both cooperative and non-cooperative binding of antibodies have previously been described.[63] Thus, cooperativity was tested by a Hill plot (Fig. S5),[64]

Figure 2: (**A**) Binding experiments of HLA A*03:01 using diffusional sizing. Bars show the average of triplicate measurements with the error bars representing the standard deviation. A significant change in hydrodynamic radius indicates binding of 50 nM HLA A*03:01 to 1.9 μM W6/32. In contrast, no binding can be observed between HLA and OUW4F11 or BSA. (**B**) Correlation of hydrodynamic radii with the number of residues.[61,62] The radii determined for HLA both free and bound to the antibody W6/32 agree well with the assumption of a folded protein. (**C**) Binding curve of 476 pM HLA A*03:01 with varying concentration of antibody W6/32. The blue points give the hydrodynamic radius of each equilibrated sample, averaged over the data of at least three replicates, and the blue curve is the best fit (see Experimental for details). From the fit, the dissociation constant K_d = 0.7 [0.3, 1.6] nM (95% confidence intervals given in square brackets) could be determined, consistent with a ratio of two antigens per antibody.

yielding a Hill parameter $h = 1.01 \pm 0.15$. This shows that the binding of the HLA and the

antibody is non-cooperative, ergo binding events occur independently.

Characterisation of Alloantibody-HLA Interactions in Human Serum

We next determined the applicability of MAAP for the absolute quantification and char-

acterisation of alloantibodies in human serum. We selected two well-characterised, human

monoclonal antibodies, SN23OG6 and OUW4F11, which specifically recognise HLA A*02:01

and HLA B*08:01, respectively,[65] both in human serum of non-transfused, healthy donors,

which did not contain HLA-specific antibodies, and in buffer (PBS). During data processing,

we removed a weak autofluorescence background signal from both the diffused and undiffused

measurement channels. Serum autofluorescence has been reported for human serum in the

spectral region of interest (Fig. S2)[66] and is likely to arise from natural aromatic compounds

including haem complexes, found in haemoglobin or bilirubin, which are stabilised by human serum albumin.[67]

As shown in Fig. 3, the hydrodynamic radii of pure HLA obtained in human serum ($R_h = 3.27 \pm 0.11$ nm for HLA A*02:01 and $R_h = 3.22 \pm 0.19$ nm for HLA B*08:01), were found to be consistent with theoretical values for natively folded 50 kDa proteins, as well as with the radii obtained in buffer ($R_h = 3.22 \pm 0.10$ nm for HLA A*02:01 and $R_h = 3.16 \pm 0.07$ nm for HLA B*08:01), demonstrating the applicability of MDS for measurements in human serum. Importantly, this also shows that measurement in human serum did not affect the antigen size compared to the measurements in PBS, suggesting

Figure 3: (**A**) Binding curve of 5 nM HLA A*02:01 against antibody SN23OG6 in human serum (red) and PBS (blue). The measurements in serum and PBS are in good agreement with each other, yielding K_d values of 6.9 [2.3, 15.6] nM and 4.8 [2.0, 9.3] nM, respectively, with a binding ratio of 2 antigens per antibody. (**B**) Binding curve of 1.2 nM HLA B*08:01 against antibody OUW4F11 in human serum (red) and PBS (blue). Again, the $K_d = 72.1$ [32.9, 163.6] nM in serum and $K_d = 89.1$ [45.7, 173.8] nM in PBS with a binding ratio of 1 to 2 are in good agreement. Summary of the hydrodynamic radii for the fully unbound antigens and of the dissociation constants in human serum in comparison to their values in pure buffer, demonstrating consistent values under both conditions (**C** for SN23OG6 against HLA A*02:01 and (**B**) for OUW4F11 against HLA B*08:01.) .

presence of serum proteins in the sample do not affect the measurements.

Assessment of these two alloantibody-HLA interactions showed that microfluidic affinity measurements were independent of the buffer conditions used (Fig. 3A-B). The interaction between SN23OG6 against HLA A*02:01 yielded a dissociation constant of K_d = 6.9 [2.3, 15.6] nM in human serum and was in good agreement with $K_d = 4.8$ [2.0, 9.3] nM in buffer. Similarly, for the interaction of antibody OUW4F11 against HLA B*08:01, the $K_d =$ 72.1 [32.9, 163.6] nM in human serum which was consistent with $K_d = 89.1$ [45.7, 173.8] nM determined in buffer. The different saturation levels between the two media conditions are within the CI and most likely reflect minor conformational variations between individual antigens. Analysis of the above alloantibody-HLA interactions in PBS using biolayer interferometry showed dissociation constants in a similar order ($K_d = 5.6 \pm 0.03$ nM for SN23OG6 against HLA A*02:01 and $K_d = 314.9 \pm 0.04$ nM for OUW4F11 against HLA B*08:01, Fig. S6).

Taken together, these data show the general feasibility of binding measurements in complex media with diffusional sizing, with hydrodynamic radii, affinities and stoichiometric parameters consistent with theoretical values and measurements under unperturbed (buffer) conditions. The results show that the binding of HLA, even in the background of a complex solution such as human serum, is solely based on specific interactions and not influenced by any other protein species that are abundant in human serum including by soluble HLA and β_2- microglobulin.[68] More generally, these results suggest that MAAP can be used as a platform for molecular level characterisation of protein-protein interactions in complex mixtures such as body fluids.

Quantification of Alloantibody-HLA Interactions in Human Serum

Simultaneous determination of both affinity and alloantibody concentration in patient samples is a key advantage of our method compared to traditional assays. Through varying the concentration of both labelled (i.e. HLA) and unlabelled (i.e. alloantibodies in human serum)

Figure 4: Binding Curves for the interaction of A*02:01 against SN23OG6 which was spiked in human serum at concentrations (**A**) 3 nM, (**B**) 10 nM, (**C**) 30 nM and (**D**) 100 nM in a blinded manner (i.e. the antibody concentrations were unknown to the person performing the analysis). This shows that the calculated K_d corresponds well to the $K_d = 6.9$ [2.3, 15.6] nM obtained for the same interaction previously, and the antibody concentrations determined are in good agreement with the true concentrations, even for (**A**), where the antibody concentration is comparable to the K_d for the interaction. The plots in the middle row show the probability distribution of the dissociation constant, K_d, with the shaded region showing the 95% confidence range obtained from data fitting using the known antibody concentrations in spiked serum. The bottom row plots depict the probability distribution for the concentration of antibody under assumption of 2:1 binding HLA : antibody, with the shaded region showing the 95 % confidence range obtained from direct measurements of antibody concentration by Bayesian inference.

species, it becomes possible to properly constrain the probability distribution of unkown parameters for the interaction (K_d and antibody concentration) as demonstrated by Bayesian inference analysis (see Experimental). This is the key advance of this technology as compared to similar assays previously used to describe affinity measurements. In order to verify the robustness of our method for determination of absolute parameters for reactive antibody species, we spiked human serum from non-sensitised donors with HLA-specific monoclonal antibody in a blinded manner, i.e. the final antibody serum concentrations were not revealed to the person performing the analysis. The interaction investigated was that between

Figure 5: Quantification of reactivity against different HLA variants in a transplant patient serum. (**A**) Antibody binding profile as detected on Luminex Single antigen beads for HLA A*01:01, HLA A*02:01 and HLA A*24:02 (MFI: Mean Fluorescence Intensity). (**B**) Binding Curves for the interaction of alloantibodies in patient serum against HLA A*02:01 and A*24:02, relating the hydrodynamic radius, R_h, to the measured concentration of antibody. The individual data points represent the measured data, the solid line the fit. The error bars report the standard deviation for triplicate measurements. For HLA A*02:01 isoform, an antibody concentration $[Ab]_{spec} = 6.1[2.4, 11.6]$ nM and a dissociation constant $K_d = 0.13$ [0.01, 1.38] nM were determined. For HLA A*24:02 isoform, we determined the $K_d = 1.3[0.2, 9.7]$ nM and $[Ab]_{spec}38.7 = [21.6, 98.5]$ nM. (**C**) Titration Curve for the same patient serum against 1 nM HLA A*01:01, with different serum concentration. As it can be seen, no binding is detected.

alloantibody SN23OG6 and HLA A*02:01 (Fig. 4). For antibody concentrations of 1 nM or above, we were able to determine both the dissociation constant, K_d, and the concentration of specific antibody, $[Ab]_{spec}$, in doped serum simultaneously. The K_d determined in all cases was consistent with previous results for the interaction (as shown above), and the concentration determined was in good agreement with the expected antibody concentration (Fig. 4). This was the case even when the concentration of antibody binding sites was approximately equal to the dissociation constant, K_d (Fig. 4a), demonstrating that our method enables accurate determination of antibody concentrations around the K_d.

Quantification of Alloantibody-HLA Interactions in Patient Sera

We next investigated the ability of our immunoassay to quantify HLA-specific antibodies in the serum of a kidney transplant patient. The patient became sensitised after transplantation with a kidney allograft expressing the HLA A*24:02 alloantigen. Analysis of post-

transplant sera using the Luminex single antigen bead assay (standard of care in the clinical setting) showed a complex profile with reactivity against the priming antigen (A*24:02), cross-reactivity against A*02:01 (A*24:02 and A*02:01 are part of a common serological cross-reactive HLA epitope group), and additional reactivity to A*01:01 (no known serological cross-reactivity to A*24:02 and A*02:01[69]). The mean fluorescence intensity values detected by Luminex were 19314 a.u. for A*24:02, 18653 a.u. for A*02:01 and 7810 a.u. for A*01:01 (Fig. 5A). As shown in Fig. 5B, for the serum interaction with HLA A*02:01, we determined a concentration of alloantibody of $[Ab]_{spec} = 6.1$ [2.4, 11.6] nM and a $K_d = 0.13$ [0.01, 1.38] nM, assuming a binding ratio of 1 to 2. Similarly, for the same patient serum interaction with A*24:02, we detected an antibody concentration of, $[Ab]_{spec} = 38.7$ [21.6, 98.5] nM and an affinity of $K_d = 1.3$ [0.2, 9.7] nM. Thus, we were able to deconvolute the fundamental biophysical properties (affinity and alloantibody concentration) of the humoral response in a complex patient serum and differentiate the reactivity against the priming alloantigen (A*24:02) and a cross-reactive alloantigen (A*02:01) demonstrating higher antibody concentration against the priming HLA. Importantly, we could not demonstrate an interaction between HLA A*01:01 and the patient serum, despite a relatively high MFI value of 7810 a.u. from the Luminex assay. Output from the Luminex assay is avidity driven and is considered semi-quantitative;[40] accordingly, an interaction at the MFI level showed here against A*01:01 would be considered as clinically significant (e.g. potential donors expressing A*01:01 would typically be excluded for a patient with similar levels of reactivity on Luminex). Taken together, the data highlight the potential of MAAP to provide immunologically relevant information not attainable by currently available techniques and to quantify antibody interactions against proteins that share highly similar structures, such as the HLA system.

Conclusion

In this study, we have shown for the first time that, using an in-solution technique, namely Microfluidic Antibody Affinity Profiling (MAAP), it is possible to determine both dissociation constant, K_d, and the absolute concentration of antibody binding sites, which may manifest as the binding stoichiometry in samples of known antibody concentrations, or the total antibody concentration in unknown samples, through a measurement of effective hydrodynamic radii at different antibody and antigen concentrations. By applying this platform to measure immunologically-relevant interactions between specific antibodies and HLA, we were able to determine biophysical parameters describing the binding event fully quantitatively, even in such a complex medium as human serum. The determined dissociation constants range between 10^{-10} M and 10^{-8} M and are consistent with previous work.[70] However, our in-solution approach avoids the commonly reported disadvantages of surface-based assays and does not require serum preparation to reduce non-specific binding thereby enabling determination of fundamental parameters of humoral responses under physiological conditions.

Our results suggest applicability of this method in a wide range of investigations aiming to understand the role of both abundance and dissociation constants implicated in clinically relevant immune responses. For example, further insights into the complex process of graft rejection may be obtained through investigations of the correlation of the concentration and affinity to the occurrence of an immune response. Finally, our results indicate that the platform can be of general use for diagnostics beyond histocompatibility testing, such as immuno-profiling in auto-immunity, in infectious diseases, and to monitor immune responses after vaccination, or the detection of biomarker levels for various diseases in human serum.[15,16]

Experimental

Materials

HLA monomers were obtained through the NIH Tetramer Core Facility in Emory, Atlanta, US, in PBS. PBS and Alexa Fluor 647 was purchased from Thermo Fisher Scientific Inc., Waltham, US. Human IgG ab205198 was from Abcam, Cambridge, UK. All other chemicals were from Sigma Aldrich. All PBS was supplemented with NaN_3 (0.02 % (w/v)).

Plate reader experiments were performed on a Clariostar BMG Labtech, Ortenberg, DE instrument. Size exclusion chromatography was performed on a Superdex 200 increase 10/300 gL column (GE Healthcare, Chicago, US) on an AKTA pure protein purification system (GE Healthcare, Chicago, US). All microfluidic experiments were performed on a Fluidity One-W Serum instrument, Fluidic Analytics, Cambridge, UK. The basic principle of MDS has been described before[71] In brief, labelled protein flows into the diffusion chamber from one side, auxiliary buffer from the other side. Due to the small channel size, laminar flow can be assumed, meaning that the particles can move into the buffer stream by diffusion only, whereby the rate depends on the size of the molecular complex (Fig. 1). At the end of the diffusion chamber, the stream is split again and the fluorescence of both the diffused and the undiffused material is measured. From the ratio between the fluorescence in both chambers, the hydrodynamic radius, R_h, of the protein can be determined.

Determination of Autofluorescence in Human Serum

Human serum from non-sensitised volunteers (not containing HLA-specific antibodies) was supplemented with PBS and the fluorophore Alexa Fluor™ 647 to yield fluorophore concentrations between 10 pM and 1 µM in serum. Similar dilutions of fluorophore in buffer were prepared for comparison. Subsequently, both absorption spectra (Fig. S2 a-b) and the emission spectra upon excitation at two wavelengths $\lambda_{ex,1} = 481$ nm and $\lambda_{ex,2} = 632$ nm (Fig. S2 c-d)were recorded on a plate reader.

Labelling of HLA with Alexa Fluor 647 fluorophore

To HLA (typically 1 nmol, 1 equiv.) in 0.1 M NaHCO$_3$ (pH = 8), Alexa Fluor 647 N-hydroxysuccinimide ester (in DMSO, 3 equiv.) was added. The reaction mixture was incubated for 1 h at ca. 20 °C, protected from light. The sample was purified by size exclusion chromatography with a flow rate of 0.5 mL/min and PBS as eluent buffer, to yield labelled HLA (typically around 0.8 nmol, DOL between 0.33 and 2.25).

Binding Experiments in PBS

For all binding experiments in PBS (supplemented with 0.02 % Tween-20), HLA (typically at a total concentration of 5 nM) was incubated with a varying concentration of specific antibodies at room temperature for 30 minutes and measured by MDS. The same protocol was followed for negative controls with fluorescently labelled BSA.

MAAP measurements in PBS

Labelled HLA, together with a varying concentration of the antibody of interest, were added and diluted in PBS (supplemented with 0.02 % Tween-20). The samples were incubated at room temperature for approximately 30 minutes. Subsequently, the size was determined by MDS.

Binding Experiments in Human Serum

All binding experiments in human serum were carried out using human serum from non-sensitised volunteers (not containng HLA-sepcific antibodies) or from a kidney transplant patient. For all binding experiments in human serum, from HLA (typically at a total concentration of 5 nM) was incubated with a varying concentration of specific antibodies or serum fractions at room temperature for 30 minutes in human serum and measured by MAAP. When fitting the data, the fluorescence intensity of both the diffused and the undiffused

channel was subtracted from the background at the relevant serum concentration.

Bayesian Analysis

The dissociation constant, $K_d = \frac{[Ab][H]}{[AbH]}$, and where necessary the antibody binding site concentration, $[Ab]_0$, were determined through Bayesian inference. The hydrodynamic radii were measured based on the amount of protein that diffuses into the distal channel; in order to relate the measurements, we introduce the parameters ρ_f and ρ_b as the fractions of free and antibody-bound HLA, respectively, that diffuse into the distal channel. The priors used for ρ_f and ρ_b were flat in linear space, while flat priors in logarithmic space were used for K_d and antibody concentration, where applicable. We can therefore express the fraction of HLA that diffuses into the distal channel, f_d, as:

$$f_d = \frac{([AbH]\rho_b + ([H]_0 - [AbH])\rho_f)}{[H]_0} \tag{1}$$

where $[AbH]$ denotes the equilibrium concentration of antibody-HLA complex, and $[H]_0$ the total HLA concentration. Considering mass-balance, we express $[AbH]$ as:

$$[AbH] = \frac{[Ab]_0 + [H]_0 + K_d - \sqrt{([Ab]_0 + [H]_0 + K_d)^2 - 4[Ab]_0[H]_0}}{2} \tag{2}$$

The Bayesian inference analysis method utilises Bayes' theorem, and allows the determination of the probability distribution of unknown parameters, given the observed data, by the following equation

$$P(parameters|data) \propto P(parameters)P(data|parameters) \tag{3}$$

where $P(parameters|data)$ is known as the posterior, $P(parameters)$ as the prior, and $P(data|parameters)$ as the likelihood. The prior probability distribution is an expression of our information about the system before any measurement data can be acquired. For ρ_f and

ρ_b, prior is assumed to be flat in linear space, whereas for the K_d and total concentration of antibody, a prior that is flat in logarithmic space is more appropriate, to reflect the scale invariance of the problem.

Experimental measurement data as described herein are normally distributed about the true value, and the likelihood function is therefore a Gaussian, centred on the theoretical measurement value.

$$P(data|parameters) \propto \exp[-\frac{1}{2\sigma^2}\sum_{i+1}^{N}(y_i - f_d(\rho_f, \rho_b, K_d, [Ab]_tot, \alpha_i, [H]_i))^2] \qquad (4)$$

where f_d is defined in equation 1 with $[AbH]$ defined in equation 2, α_i and $[H]_i$ are the concentrations of serum and antibody, respectively, in the i^{th} measurement and y_i is the pre-processed data point obtained in the i^{th} measurement. In order to define an appropriate standard deviation, σ, for each dataset, the standard deviations of repeats of each measurement are calculated, and the maximum of these values are used as a global standard deviation for that dataset.

In the experiments where the total antibody concentration was unknown, we substitute $[Ab]_0$ for $\alpha[Ab]_{tot}$, where α is the fraction of serum used in the measurement, and $[Ab]_{tot}$ is the total concentration of antibody binding sites in the stock solution. Assuming a binding stoichiometry of 1:2 antibody:HLA, we therefore obtain the antibody concentration, $\frac{[Ab]_{tot}}{2}$. All confidence intervals used were 95% confidence intervals, corresponding to the standard derivation.

Luminex Single Antigen Beads

HLA-specific antibody reactivity in the patient serum was detected using LabScreen single antigen HLA class I detection beads (One Lambda, Canoga Park, CA), as previously described[72]. HLA single antigen bead-defined Ab reactivity was determined using a mean fluorescence intensity (MFI) cut-off threshold of 2000 (MFI cut-off level used clinically in

our center and elsewhere to define a positive alloantibody response to a given HLA).

Bio-Layer Interferometry

Monoclonal antibody affinity of binding to HLA was determined by bio-layer interferometry (BLI) using the Octet RED96 system (ForteBio, Fremont, California). Antibody was immobilised to anit-human IgG Fc kinetic biosensors. To determine the association phase, sensors were dipped into wells containing soluble, recombinant HLA in a 2-fold titration for 300 seconds so an equilibrium was reached. Next, sensors were placed into buffer alone-containnig wells for further 1000 seconds to determine the dissociation phase. Affinity values (K_d) were calculated via steady-state analysis as the ratio of on- and off-rate constants ($\frac{k_o f f}{k_o n}$. All experiments were carried out using standard kinetic buffer (PBS, 0.1% (w/v) bovine serum albumin, 0.02% Tween-20), at a temperature of 30°C and a constant plate shake speed of 1000 rpm.

Acknowledgement

The research presented in this manuscript has received funding from the European Research Council under the European Union's Seventh Framework Programme (FP7/2007-2013) through the ERC grant PhysProt (agreement no. 337969) and under European Union's Horizon 2020 research and innovation programme (ETN grant 674979-NANOTRANS). We gratefully acknowledge financial support from the Engineering and Physical Sciences Research Council (EPSRC) and Frances and Augustus Newman Foundation. We also acknowledge funding from the NIHR Blood and Transplant Research Unit in Organ Donation and Transplantation at the University of Cambridge and from an NIHR Fellowship (PDF-2016-09-065, VK). The views expressed are those of the authors and not necessarily those of the National Health Service, the National Institute for Health Research, the Department of Health, or National Health Service Blood and Transplant. We are grateful to Dr

Arend Mulder and Prof Frans Claas for provision of monoclonal antibodies SN23OG6 and OUW4F11. Ethical Approval for work involving human serum was granted from NRES Committee North-East York, IRAS number 167211.

Author Contribution

T.P.J.K., V.K., C.M.D., T.S., M.M.S. and A.J.P. designed the study. M.M.S., T.S., M.H., A.J.P. performed the experiments. A.J.P., S.R.A.D., V.K. and T.P.J.K. provided material. C.K.X., G.M., M.M.S., T.S., V.K. and T.P.J.K. analysed the data. M.M.S., T.S., C.K.X., G.M., V.K. and T.P.J.K. wrote the paper. All authors discussed the results and commented on the manuscript.

Conflict of Interest

S.R.A.D. is an employee of Fluidic Analytics, which is developing and commercializing microfluidic diffusional sizing instrumentation including the Fluidity One-W Serum used here. V.K. is a consultant with Fluidic Analytics Ltd. T.P.J.K. is a member of the board of directors of Fluidic Analytics Ltd.

Supporting Information Available

The following files are available free of charge online.

- Filename: Supplementary Information

References

(1) King, N. P.; Sheffler, W.; Sawaya, M. R.; Vollmar, B. S.; Sumida, J. P.; André, I.; Gonen, T.; Yeates, T. O.; Baker, D. Computational Design of Self-Assembling Protein Nanomaterials with Atomic Level Accuracy. *Science* **2012**, *336*, 1171–1174.

(2) Edwardson, T. G. W.; Mori, T.; Hilvert, D. Rational Engineering of a Designed Protein Cage for siRNA Delivery. *J. Am. Chem. Soc.* **2018**, *140*, 10439–10442.

(3) Esposito, L.; Sica, F.; Raia, C. A.; Giordano, A.; Rossi, M.; Mazzarella, L.; Zagari, A. Crystal Structure of the Alcohol Dehydrogenase from the Hyperthermophilic Archaeon Sulfolobus solfataricus at 1.85A Resolution. *J. Mol. Biol.* **2002**, *318*, 463–477.

(4) Knowles, T. P. J.; Waudby, C. A.; Devlin, G. L.; Cohen, S. I. A.; Aguzzi, A.; Vendruscolo, M.; Terentjev, E. M.; Welland, M. E.; Dobson, C. M. An Analytical Solution to the Kinetics of Breakable Filament Assembly. *Science* **2009**, *326*, 1533–1537.

(5) Knowles, T. P. J.; Vendruscolo, M.; Dobson, C. M. The amyloid state and its association with protein misfolding diseases. *Nat. Rev. Mol. Cell Biol.* **2014**, *15*, 384.

(6) Hartl, F. U. Protein Misfolding Diseases. *Ann. Rev. Biochem.* **2017**, *86*, 21–26.

(7) Goldberg, R. J. A Theory of Antibody—Antigen Reactions. I. Theory for Reactions of Multivalent Antigen with Bivalent and Univalent Antibody2. *J. Am. Chem. Soc.* **1952**, *74*, 5715–5725.

(8) Goldberg, R. J. A Theory of Antibody—Antigen Reactions. II. Theory for Reactions of Multivalent Antigen with Multivalent Antibody. *J. Am. Chem. Soc.* **1953**, *75*, 3127–3131.

(9) Patrick, G. L. *An Introduction to Medicinal Chemistry*; Oxford University Press: Oxford, 2013.

(10) Huxley, H. E. The Mechanism of Muscular Contraction. *Science* **1969**, *164*, 1356.

(11) Westermarck, J.; Ivaska, J.; Corthals, G. L. Identification of Protein Interactions Involved in Cellular Signaling. *Mol. Cell. Proteom.* **2013**, *12*, 1752.

(12) Suwinski, R.; Giglok, M.; Galwas-Kliber, K.; Idasiak, A.; Jochymek, B.; Deja, R.; Maslyk, B.; Mrochem-Kwarciak, J.; Butkiewicz, D. Blood serum proteins as biomarkers for prediction of survival, locoregional control and distant metastasis rate in radiotherapy and radio-chemotherapy for non-small cell lung cancer. *BMC Cancer* **2019**, *19*, 427.

(13) Srivastava, A.; Creek, D. J. Discovery and Validation of Clinical Biomarkers of Cancer: A Review Combining Metabolomics and Proteomics. *Proteomics* **2019**, *19*, 1700448.

(14) Ivancic, M. M.; Anson, L. W.; Pickhardt, P. J.; Megna, B.; Pooler, B. D.; Clipson, L.; Reichelderfer, M.; Sussman, M. R.; Dove, W. F. Conserved serum protein biomarkers associated with growing early colorectal adenomas. *Proc. Nat. Acad. Sci.* **2019**, *116*, 8471.

(15) Núñez, C. Blood-based protein biomarkers in breast cancer. *Clinica Chimica Acta* **2019**, *490*, 113–127.

(16) Shi, L.; Baird, A. L.; Westwood, S.; Hye, A.; Dobson, R.; Thambisetty, M.; Lovestone, S. A Decade of Blood Biomarkers for Alzheimer's Disease Research: An Evolving Field, Improving Study Designs, and the Challenge of Replication. *J. Alzheimers Dis.* **2018**, *62*, 1181.

(17) Chen-Plotkin, A. S. et al. Finding useful biomarkers for Parkinson's disease. *Sci. Transl. Med.* **2018**, *10*, 454.

(18) Huang, T.; Glass, K.; Zeleznik, O. A.; Kang, J. H.; Ivey, K. L.; Sonawane, A. R.; Birmann, B. M.; Hersh, C. P.; Hu, F. B.; Tworoger, S. S. A Network Analysis of Biomarkers for Type 2 Diabetes. *Diabetes* **2019**, *68*, 281.

(19) Killoran, A.; Biglan, K. Biomarkers for Huntington's disease: A brief overview. *Journal of Rare Diseases Research and Treatment* **2016**, *1*, 46–50.

(20) Sperling, R. A. et al. Toward defining the preclinical stages of Alzheimer's disease: Recommendations from the National Institute on Aging-Alzheimer's Association workgroups on diagnostic guidelines for Alzheimer's disease. *Alzheimers Dement.* **2011**, *7*, 280–292.

(21) Thambisetty, M. et al. Plasma Biomarkers of Brain Atrophy in Alzheimer's Disease. *PLOS ONE* **2011**, *6*, e28527.

(22) Hakki, Y.; Cakmak, M.; Ceydilek, B.; Demir, C.; Aktas, A. Role of interlekin-35 as a biomarker in patients with newly diagnosed Hashimoto's thyroiditis. *Endocr. Regul.* **2016**, *50*, 55–61.

(23) Montgomery, R. A.; Tatapudi, V. S.; Leffell, M. S.; Zachary, A. A. HLA in transplantation. *Nat. Rev. Nephrol.* **2018**, *14*, 558–570.

(24) Loupy, A.; Lefaucheur, C. Antibody-Mediated Rejection of Solid-Organ Allografts. *N. Engl. J. Med.* **2018**, *379*, 1150–1160.

(25) Timofeeva, O. A. Donor-Specific HLA Antibodies as Biomarkers of Transplant Rejection. *Clin. Lab Med.* **2019**, *39*, 45–60.

(26) Terasaki, P. I.; Cai, J. Human leukocyte antigen antibodies and chronic rejection: from association to causation. *Transplantation* **2008**, *86*, 377–83.

(27) Taylor, C. J.; Kosmoliaptsis, V.; Summers, D. M.; Bradley, J. A. Back to the future: application of contemporary technology to long-standing questions about the clinical relevance of human leukocyte antigen-specific alloantibodies in renal transplantation. *Hum. Immunol.* **2009**, *70*, 563–8.

(28) Engvall, E.; Perlmann, P. Enzyme-Linked Immunosorbent Assay, Elisa: III. Quantitation of Specific Antibodies by Enzyme-Labeled Anti-Immunoglobulin in Antigen-Coated Tubes. *J. Immunol.* **1972**, *109*, 129–135.

(29) Ueda, H.; Tsumoto, K.; Kubota, K.; Suzuki, E.; Nagamune, T.; Nishimura, H.; Schueler, P. A.; Winter, G.; Kumagai, I.; Mahoney, W. C. Open sandwich ELISA: A novel immunoassay based on the interchain interaction of antibody variable region. *Nat. Biotechnol.* **1996**, *14*, 1714–1718.

(30) Sakamoto, S.; Putalun, W.; Vimolmangkang, S.; Phoolcharoen, W.; Shoyama, Y.; Tanaka, H.; Morimoto, S. Enzyme-linked immunosorbent assay for the quantitative/qualitative analysis of plant secondary metabolites. *J. Nat. Med.* **2018**, *72*, 32–42.

(31) Carson, R. T.; Vignali, D. A. A. Simultaneous quantitation of 15 cytokines using a multiplexed flow cytometric assay. *J. Immunol. Met.* **1999**, *227*, 41–52.

(32) Fulton, R. J.; McDade, R. L.; Smith, P. L.; Kienker, L. J.; Kettman, J., John R. Advanced multiplexed analysis with the FlowMetrixTM system. *Clin. Chem.* **1997**, *43*, 1749–1756.

(33) Joos, T. O.; Stoll, D.; Templin, M. F. Miniaturised multiplexed immunoassays. *Curr. Opin. Chem. Biol.* **2002**, *6*, 76–80.

(34) Pei, R.; Lee, J. H.; Shih, N. J.; Chen, M.; Terasaki, P. I. Single human leukocyte antigen flow cytometry beads for accurate identification of human leukocyte antigen antibody specificities. *Transplantation* **2003**, *75*, 43–9.

(35) Lyon, L. A.; Musick, M. D.; Natan, M. J. Colloidal Au-Enhanced Surface Plasmon Resonance Immunosensing. *Anal. Chem.* **1998**, *70*, 5177–5183.

(36) Mullett, W. M.; Lai, E. P. C.; Yeung, J. M. Surface Plasmon Resonance-Based Immunoassays. *Methods* **2000**, *22*, 77–91.

(37) Simon, A.; Macdonald, J. In *Wilson and Walker's Techniques of Biochemistry and Molecular Biology*; Hoffmann, A., Clokie, S., Eds.; Cambridge University Press: Cambridge, 2018; pp 500–534.

(38) Tait, B. D. Detection of HLA Antibodies in Organ Transplant Recipients – Triumphs and Challenges of the Solid Phase Bead Assay. *Front. Immunol.* **2016**, *7*.

(39) Reed, E. F.; Rao, P.; Zhang, Z.; Gebel, H.; Bray, R. A.; Guleria, I.; Lunz, J.; Mohanakumar, T.; Nickerson, P.; Tambur, A. R.; Zeevi, A.; Heeger, P. S.; Gjertson, D. Comprehensive Assessment and Standardization of Solid Phase Multiplex-Bead Arrays for the Detection of Antibodies to HLA. *Am. J. Transplant.* **2013**, *13*, 1859–1870.

(40) Gebel, H. M.; Bray, R. A. HLA antibody detection with solid phase assays: great expectations or expectations too great? *Am. J. Transplant.* **2014**, *14*, 1964–75.

(41) Henry, S. M.; Sutlief, E.; Salas-Solano, O.; Valliere-Douglass, J. ELISA reagent coverage evaluation by affinity purification tandem mass spectrometry. *mAbs* **2017**, *9*, 1065–1075.

(42) Butch, A. W. Dilution Protocols for Detection of Hook Effects/Prozone Phenomenon. *Clin. Chem.* **2000**, *46*, 1719–1720.

(43) Kosmoliaptsis, V.; O'Rourke, C.; Bradley, J. A.; Taylor, C. J. Improved Luminex-based human leukocyte antigen-specific antibody screening using dithiothreitol-treated sera. *Hum. Immunol.* **2010**, *71*, 45–9.

(44) Schnaidt, M.; Weinstock, C.; Jurisic, M.; Schmid-Horch, B.; Ender, A.; Wernet, D. HLA antibody specification using single-antigen beads—a technical solution for the prozone effect. *Transplantation* **2011**, *92*, 510–5.

(45) Anderson, N. L.; Anderson, N. G. The Human Plasma Proteome: History, Character, and Diagnostic Prospects. *Mol. Cell. Proteom.* **2002**, *1*, 845–867.

(46) Redman, E. A.; Batz, N. G.; Mellors, J. S.; Ramsey, J. M. Integrated Microfluidic Capillary Electrophoresis-Electrospray Ionization Devices with Online MS Detection for the Separation and Characterization of Intact Monoclonal Antibody Variants. *Anal. Chem.* **2015**, *87*, 2264–2272.

(47) Fan, R.; Vermesh, O.; Srivastava, A.; Yen, B. K. H.; Qin, L.; Ahmad, H.; Kwong, G. A.; Liu, C.-C.; Gould, J.; Hood, L.; Heath, J. R. Integrated barcode chips for rapid, multiplexed analysis of proteins in microliter quantities of blood. *Nat. Biotechnol.* **2008**, *26*, 1373.

(48) Horton, R.; Wilming, L.; Rand, V.; Lovering, R. C.; Bruford, E. A.; Khodiyar, V. K.; Lush, M. J.; Povey, S.; Talbot, C. C.; Wright, M. W.; Wain, H. M.; Trowsdale, J.; Ziegler, A.; Beck, S. Gene map of the extended human MHC. *Nat. Rev. Genet.* **2004**, *5*, 889–899.

(49) Valenzuela, N. M.; Reed, E. F. Antibody-mediated rejection across solid organ transplants: manifestations, mechanisms, and therapies. *J. Clin. Invest.* **2017**, *127*, 2492–2504.

(50) Tait, B. D. et al. Consensus guidelines on the testing and clinical management issues associated with HLA and non-HLA antibodies in transplantation. *Transplantation* **2013**, *95*, 19–47.

(51) Gosset, C.; Viglietti, D.; Rabant, M.; Vérine, J.; Aubert, O.; Glotz, D.; Legendre, C.; Taupin, J. L.; Duong Van-Huyen, J. P.; Loupy, A.; Lefaucheur, C. Circulating donor-specific anti-HLA antibodies are a major factor in premature and accelerated allograft fibrosis. *Kidney. Int.* **2017**, *92*, 729–742.

(52) Gebel, H. M.; Bray, R. A.; Nickerson, P. Pre-transplant assessment of donor-reactive, HLA-specific antibodies in renal transplantation: contraindication vs. risk. *Am. J. Transplant.* **2003**, *3*, 1488–500.

(53) Tambur, A. R. et al. Sensitization in transplantation: Assessment of risk (STAR) 2019 Working Group Meeting Report. *Am. J. Transplant.* **2020**,

(54) Tambur, A. R.; Herrera, N. D.; Haarberg, K. M.; Cusick, M. F.; Gordon, R. A.; Leventhal, J. R.; Friedewald, J. J.; Glotz, D. Assessing Antibody Strength: Comparison of MFI, C1q, and Titer Information. *Am. J. Transplant.* **2015**, *15*, 2421–30.

(55) Konvalinka, A.; Tinckam, K. Utility of HLA Antibody Testing in Kidney Transplantation. *J. Am. Soc. Nephrol.* **2015**, *26*, 1489–502.

(56) Müller, T.; Arosio, P.; Rajah, L.; Cohen, S. I. A.; Yates, V., Emma; Vendruscolo, M.; Dobson, C. M.; Knowles, T. P. J. Particle-Based Monte-Carlo Simulations of Steady-State Mass Transport at Intermediate Péclet Numbers. *Int. J. Nonlin. Sci.* **2016**, *17*, 175.

(57) Arosio, P.; Müller, T.; Rajah, L.; Yates, E. V.; Aprile, F. A.; Zhang, Y.; Cohen, S. I. A.; White, D. A.; Herling, T. W.; De Genst, E. J.; Linse, S.; Vendruscolo, M.; Dobson, C. M.; Knowles, T. P. J. Microfluidic Diffusion Analysis of the Sizes and Interactions of Proteins under Native Solution Conditions. *ACS Nano* **2016**, *10*, 333–341.

(58) Linse, S. et al. Kinetic fingerprint of antibody therapies predicts outcomes of Alzheimer clinical trials. *bioRxiv* **2019**, 815308.

(59) Scheidt, T.; Łapińska, U.; Kumita, J. R.; Whiten, D. R.; Klenerman, D.; Wilson, M. R.; Cohen, S. I. A.; Linse, S.; Vendruscolo, M.; Dobson, C. M.; Knowles, T. P. J.; Arosio, P. Secondary nucleation and elongation occur at different sites on Alzheimers amyloid-β aggregates. *Sci. Adv.* **2019**, *5*, eaau3112.

(60) Kievits, F.; Ivanyi, P. Monomorphic anti-HLA monoclonal antibody (W6/32) recognizes polymorphic H-2 heavy-chain determinants exposed by association with bovine or human but not murine ß2-microglobulin. *Hum. Immunol.* **1987**, *20*, 115–126.

(61) Wilkins, D. K.; Grimshaw, S. B.; Receveur, V.; Dobson, C. M.; Jones, J. A.; Smith, L. J. Hydrodynamic Radii of Native and Denatured Proteins Measured by Pulse Field Gradient NMR Techniques. *Biochemistry* **1999**, *38*, 16424–16431.

(62) Zhang, Y.; Yates, E. V.; Hong, L.; Saar, K. L.; Meisl, G.; Dobson, C. M.; Knowles, T. P. J. On-chip measurements of protein unfolding from direct observations of micron-scale diffusion. *Chem. Sci.* **2018**, *9*, 3503–3507.

(63) Yang, D.; Kroe-Barrett, R.; Singh, S.; Roberts, C. J.; Laue, T. M. IgG cooperativity - Is there allostery? Implications for antibody functions and therapeutic antibody development. *mAbs* **2017**, *9*, 1231–1252.

(64) Hill, A. V. PROCEEDINGS OF THE PHYSIOLOGICAL SOCIETY: January 22, 1910. *J. Physiol. (Lond.)* **1910**, *40*, i–vii.

(65) Mulder, A.; Kardol, M.; Regan, J.; Buelow, R.; Claas, F. Reactivity of twenty-two cytotoxic human monoclonal HLA antibodies towards soluble HLA class I in an enzyme-linked immunosorbent assay (PRA-STAT). *Hum. Immunol.* **1997**, *56*, 106–13.

(66) Wolfbeis, O. S.; Leiner, M. Mapping of the total fluorescence of human blood serum as a new method for its characterization. *Anal. Chem. Acta* **1985**, *167*, 203–215.

(67) Blauer, G.; Wagniere, G. Conformation of bilirubin and biliverdin in their complexes with serum albumin. *J. Am. Chem. Soc.* **1975**, *97*, 1949–1954.

(68) DeVito-Haynes, L. D.; Jankowska-Gan, E.; Meyer, K. C.; Cornwell, R. D.; Zeevi, A.; Griffith, B.; Dauber, J.; Iacono, A.; Burlingham, W. J.; Love, R. B. Soluble donor HLA class I and beta 2m-free heavy chain in serum of lung transplant recipients: steady-state levels and increases in patients with recurrent CMV infection, acute rejection episodes, and poor outcome. *Hum. Immunol.* **2000**, *61*, 1370–82.

(69) Rodey, G. E.; Revels, K.; Fuller, T. C. Epitope specificity of HLA class I alloantibodies: II. Stability of cross-reactive group antibody patterns over extended time periods. *Transplantation* **1997**, *63*, 885–93.

(70) Visentin, J.; Leu, D. L.; Mulder, A.; Jambon, F.; Badier, L.; Lee, J.-H.; Guidicelli, G.; Bouthemy, C.; Ralazamahaleo, M.; Claas, F.; Di Primo, C.; Taupin, J.-L. Measuring anti-HLA antibody active concentration and affinity by surface plasmon resonance: Comparison with the luminex single antigen flow beads and T-cell flow cytometry crossmatch results. *Mol. Immunol.* **2019**, *108*, 34–44.

(71) Yates, E. V.; Müller, T.; Rajah, L.; De Genst, E. J.; Arosio, P.; Linse, S.; Vendruscolo, M.; Dobson, C. M.; Knowles, T. P. J. Latent analysis of unmodified biomolecules and their complexes in solution with attomole detection sensitivity. *Nat. Chem.* **2015**, *7*, 802.

(72) Mallon, D. H.; Kling, C.; Robb, M.; Ellinghaus, E.; Bradley, J. A.; Taylor, C. J.; Kabelitz, D.; Kosmoliaptsis, V. Predicting Humoral Alloimmunity from Differences in Donor and Recipient HLA Surface Electrostatic Potential. *The Journal of Immunology* **2018**, *201*, 3780.

Graphical TOC Entry