Consistent ultra-long DNA sequencing with automated slow pipetting

Trent M. Prall[1]   https://orcid.org/0000-0003-0635-4152

Emma K. Neumann[1]   https://orcid.org/0000-0003-2453-8297

Julie A. Karl[1]   https://orcid.org/0000-0002-4447-4721

Cecilia G. Shortreed[1]   https://orcid.org/0000-0003-0654-5952

David A. Baker[1]   https://orcid.org/0000-0002-6811-8238

Hailey E. Bussan[1]   https://orcid.org/0000-0002-0961-7109

Roger W. Wiseman[1,2]   https://orcid.org/0000-0002-7682-7085

David H. O'Connor[1,2]   https://orcid.org/0000-0003-2139-470X

(corresponding author: doconno@wisc.edu 608-890-0845)

Affiliations

[1]Department of Pathology and Laboratory Medicine, University of Wisconsin, Madison

[2]Wisconsin National Primate Research Center, University of Wisconsin, Madison

15
16

Abstract

Background Oxford Nanopore Technologies' instruments can sequence reads of great length. Long reads improve sequence assemblies by unambiguously spanning repetitive elements of the genome. Sequencing reads of significant length requires the preservation of long DNA template molecules through library preparation by pipetting reagents as slowly as possible in order to minimizing shearing. This process is time consuming and inconsistent at preserving read length as even small changes in volumetric flow rate can result in template shearing.

Results We have designed SNAILS (Slow Nucleic Acid Instrument for Long Sequences), a 3D-printable instrument that automates slow pipetting of reagents used in long read library preparation for Oxford Nanopore sequencing. Across six sequencing libraries, SNAILS preserved more reads exceeding one hundred kilobases in length and increased the average read length of its libraries over manual slow pipetting.

Conclusions SNAILS is a low-cost, easily deployable solution for improving sequencing projects that require reads of significant length. By automating the slow pipetting of library preparation reagents, SNAILS both increases the consistency and throughput of long read Nanopore sequencing.

Keywords

Oxford Nanopore Technologies, GridION, MinION, DNA sequencing, Long read sequencing, Ultra-long, Ultra-high molecular weight DNA, de novo assembly

Background

Oxford Nanopore Technologies' (ONT) MinION and GridION instruments generate reads of unprecedented length [1]. Long reads improve genomic assemblies by spanning repetitive regions and structural variation within the genome. These regions have historically complicated alignments created solely using shorter, second generation sequencing reads [2]. An increased understanding of these genomic features is crucial as evidence linking heritable structural variation with disease phenotypes continues to emerge [3]. ONT sequencing has resolved the nucleotide identity of gaps left by 100 base pair (bp) Illumina reads aligned to GRCh38 better than other long read sequencing technologies [4]. These gaps included coding exons for 76 genes with known disease-associated mutations catalogued in the Human Gene Mutation Database [5]. A pathogenic deletion in TMEM231, a gene known to be involved in Meckel-Gruber syndrome, was uncovered by ONT reads [6]. This mutation had been masked by ambiguous mapping of short reads to a paralogous pseudogene locus. Recently, the first telomere-to-telomere assemblies of the human X and Y chromosomes were created using ONT sequencing [7, 8]. In these assemblies, ultra-long reads in excess of one megabase pair (Mb) were used to bridge the centromeric gaps of previous assemblies. Using similar methods, ONT generated long reads have been utilized to establish more contiguous assemblies for important model organisms [9−12].

Sequencing long reads requires isolating ultra-high molecular weight (UHMW) DNA and preserving those DNA molecules through library preparation [13]. Methods for obtaining and preserving UHMW DNA have been developed for use in the field of optical mapping which benefits from continuous strands of DNA [14]. For example, DNA in solution can be condensed into a compact conformation by the addition of neutral polymers that exert osmotic pressure to nearby double helices [15]. Condensed DNA can also be induced by adding cationic ligands

into solution to promote attractive interactions of DNA segments. DNA is resistant to shearing by

pipetting when in condensed conformation [16]. Unfortunately, elongation of condensed DNA

requires the addition of monovalent salts which could interfere with ONT sequencing pores. In

1984, Schwartz et al. pioneered pulse field gradient gel electrophoresis that allowed for the

separation of chromosomal-sized DNA [17]. Consistent ultra-long DNA isolation was

accomplished by suspending cells in low-melting-point agarose and diffusing lysing reagents

into the resulting suspension. The "nuclei-method" later expanded on this technique by

enriching for cell nuclei prior to embedding in agarose to remove cytoplasmic and other cellular

lysates from the resulting DNA [18]. Second generation short read sequencing can be

performed on DNA extracted using the nuclei method [19]. However, gel extraction protocols

designed for preserving high molecular weight DNA sequences require access to specialized

electroelution chambers or pulsed-field gel electrophoresis systems [20]. UHMW DNA can also

be obtained using phenol-chloroform isolations based on a previously published protocol [21].

This protocol uses caustic chemicals and requires several hours to perform [13]. Alternatively,

Circulomics Inc. has developed a thermoplastic substrate that can bind and release large

quantities of DNA without damaging it [22]. They have created an isolation kit (Nanobind) using

the substrate that produces large quantities of UHMW DNA in approximately sixty minutes.

Recently, Circulomics has reported a 2.44 Mb continuous read sequenced by MinION

representing the longest single read ever sequenced [23]. Regardless of isolation method, the

resulting UHMW DNA must undergo library preparation which introduces more opportunities for

shearing to occur.

Long read ONT libraries are prepared using the Rapid Sequencing Kit (SQK-RAD004).

This protocol involves simultaneous fragmentation and ligation of tags by a transposome

complex followed by the attachment of sequencing adaptors to template molecules before

loading onto flow cells for sequencing. Throughout library preparation, DNA is susceptible to

shearing by pipetting. A single draw using a normal-bore pipette tip shears megabase DNA, so

wide-bore tips are used to reduce the amount of shearing [16]. DNA shearing can also be

mitigated by pipetting at a slow rate [16]. Because the standard RAD004 sequencing protocol

wasn't initially intended for long read sequencing, a modified RAD004 protocol was developed

to preserve template length [24]. In this protocol, UHMW DNA and sequencing reagents are

mixed by pipetting up and down as slowly as possible eighteen total times [25]. N50 denotes

the minimum length for which half of the read lengths in a pool are equal to or longer, and it is

used to estimate the distribution of read lengths within a pool. When implement, the modified

RAD004 protocol increased the average N50 from 10.6 kilobases to 99.7 kilobases over the

standard protocol. Performing the long read RAD004 protocol for a single flow cell typically

requires two hours of manual pipetting to perform [13]. This can be a significant challenge for

technicians to properly execute and is limited in throughput. To address this challenge, we have

designed and tested SNAILS (Slow Nucleic Acid Instrument for Long Sequences), an open-

sourced, 3D-printed robot that automates slow pipetting steps used in long read RAD004 library

preparations. By automating this process, SNAILS eliminates some of the variables inherent to

manual slow pipetting caused by unsteady hands. Additionally, SNAILS dramatically increases

the long read sequencing throughput capabilities of a single technician.

Results

SNAILS - Slow Nucleic Acid Instrument for Long Sequences

We hypothesized that slow pipetting reagents could be simply automated by rotating a

pipette plunger using a DC motor. To accomplish this we created SNAILS, a 3D-printed scaffold

designed around the VistaLab Ovation M micropipette that aligns a twelve-volt direct current

planetary gear motor with a three-axle gear train connected to the pipette's plunger (Fig 1A, B).

An L298N motor driver supplies current to the motor and is controlled through 3.3-volt signals

generated by a raspberry pi 3B+ microcontroller. This allows the direction of current, and by

extension the direction of motion, to be programmatically toggled to aspirate or dispense

volumes (Fig 1C). The three-axle gear train creates a 36:7 ratio between the motor and pipette

plunger that downshifts the motor's twelve rotations per minute to approximately 1.25 rotations

per minute (Fig 1D). This results in a volumetric flow rate of around 6 µl per minute when used

with a P100 micropipette. Printing files, assembly instructions, and source code for SNAILS are

freely available for download from github (https://github.com/dholab/SNAILS).

Comparison of ultra-high molecular weight DNA isolations

Three UHMW DNA isolations were performed using approximately 2×10^7 peripheral

blood mononuclear cells per isolation from a single rhesus macaque. Eighteen sequencing

libraries were prepared from these DNA isolations using ONT SQK-RAD004 library preparation

kits. One of three pipetting methods was used for library preparation: manual aspiration using

an Eppendorf P200 micropipette, manual rotation using an Ovation M P100 micropipette, and

automated rotation of an Ovation M P100 micropipette using SNAILS. Each of the pipetting

methods was used to create two libraries simultaneously from each isolation for a total of six

sequencing runs per pipetting method. It has been observed that MinKNOW arbitrarily

fragments base called reads during sequencing [26]. Therefore, we informatically fused reads

post-sequencing to accurately assess the read length retention between pipetting methods

using whale_watch.py from the BulkVis toolset [26]. The results of each library preparation are

displayed in Table 1. N50 values vary substantially per isolation with means of 23,124 bp;

15,741 bp; and 19,975 bp from isolations one, two, and three, respectively. The length of DNA

within a pool can vary substantially between UHMW isolations. Therefore, DNA length is still

limited by UHMW isolations and automated library preparation serves only to preserve existing long molecules.

To assess the efficiency of SNAILS, the long read retention of pipetting by manual aspiration, manual rotation, and automated rotation performed by SNAILS was compared through various metrics. We observed that SNAILS produced more total reads than either manual pipetting method (Figure 2A). Across six flow cells, SNAILS produced 5,251,774 reads in total, exceeding both manual aspiration which produced 4,216,634 total reads and manual rotation which produced 2,810,324 total reads (Supplementary Table 1). SNAILS produced a mean total reads sequenced per run of 892,641, exceeding the mean of 714,979 reads sequenced by manual aspiration and the mean of 547,176 reads sequenced by manual rotation (Figure 2B). This result is mostly the result of DNA isolation two where SNAILS produced a far greater number of reads than from other isolations. Given this, we cannot definitively say that automated pipetting produces a greater number of reads than manual pipetting.

Across eighteen library preparations, only SNAILS retained ultra-long reads in excess of one Mb: 1,253,204 bp and 1,021,308 bp (Table 2). In contrast, the maximum lengths of reads preserved by manual aspiration and rotation were 798,240 bp and 862,322 bp, respectively. SNAILS produced six reads longer than the maximum read produced by manual rotation and nine reads longer than the longest read produced by manual aspiration. These findings suggest that automated rotation of the pipette plunger via SNAILS is a superior method of read length retention though more comparative library preparations are necessary to confirm the observed trend.

Manual rotation produced a mean read length of 8,583 bp across six libraries. This surpassed SNAILS' mean length of 6,630 bp and manual aspiration's mean length of 5,209 bp (Figure 3A). This hierarchy was mirrored when comparing N50s produced by each method (Figure 3B). Pipetting by manual aspiration resulted in the lowest mean N50 of 16,203 bp across six cells. By comparison, the mean N50 produced by manual and automated rotation of

the plunger were 22,221 bp and 20,381 bp, respectively. Automated rotation produced the

highest N50 for a single run (31,810 bp) despite having a lower mean N50 than manual rotation

(Table 1). When the results of all six flow cells were combined, SNAILS retained more reads of

longer length than either manual method (Figure 3C). In silico modeling has predicted that an

N50 greater than 100 kb will substantially improve the continuity of genomic assemblies [24].

We therefore compared the number of reads greater than or equal to 100 kb retained by each

pipetting method (Figure 3D). SNAILS produced a combined 23,884 reads greater than or equal

to 100 kb in length (Supplementary Table 1). This number exceeded both manual aspiration and

rotation that produced 13,079 and 14,424 reads, respectively. We then compared the mean

number of reads greater than or equal to 100 kb sequenced across libraries to assess whether

SNAILS was more consistently retaining beneficially long reads. SNAILS produced the highest

mean of 3,981. Manual rotation resulted in a mean of 2,968 and aspiration produced a mean of

2,180. The difference in means was significant between automated and aspirated pipetting by

one-way ANOVA (p-value = 0.0428). No significant difference was observed between

automated and manual rotation. Taken together, these findings suggest that SNAILS is

comparable to manual rotation of the pipette plunger in terms of increasing N50 and preserving

length through RAD004 library preparations. Furthermore, SNAILS appears to be superior to the

common practice of slowly aspirating the pipette for preventing DNA shearing.

Discussion

Minimizing shearing of UHMW DNA through library preparation is a necessary

precaution needed to maximize the advantages of ONT long read sequencing. We have

designed and tested SNAILS, a 3D-printed instrument that automates the slow pipetting steps

used during the modified RAD004 library preparation protocol. SNAILS retain more 100 kb

reads and produce larger average N50s compared to libraries prepared by conventional,

manual, slow aspiration pipetting. The amount of manual preparation time saved is arguably the

largest advantage provided by SNAILS. The RAD004 long read protocol typically requires at

least two hours to prepare a single library for only eighteen pipetting steps [13]. Developing the

ability to prepare long read libraries requires substantial training to properly execute. Moreover,

only technicians with steady and dexterous hands are able to effectively pipette at the

consistent, slow rate necessary to preserve length. SNAILS both reduces the amount of hands-

on time required to prepare libraries to a few minutes and the technical burden to executing a

couple of commands. This advantage is further exemplified when considering that multiple

SNAILS can be run in parallel. This allows for a substantial increase in long read throughput

without sacrificing length. Moving forward, we plan to expand on the SNAILS design by creating

a pipette-mounted version. A pipette-mounted chassis could be attached to a poseable or

robotic arm to give SNAILS 360 degrees of motion. This design would eventually eliminate any

manual intervention that remains when using the present SNAILS design, thereby completely

automating long read library preparation. To our knowledge, the current SNAILS design serves

as the first advancement built upon the initial long read RAD004 protocol.

Liquid handling robots manufactured by Tecan, Eppendorf and others are theoretically

capable of automating long read library preparation by designing custom protocols. However,

liquid handling robots are not designed for slow pipetting and require knowledge of the

instrument's programmatic background in order to properly implement custom protocols. By

comparison, SNAILS is designed specifically for slow pipetting and operates with pythonic code

that requires little programmatic expertise to customize the speed, draw volume, and number of

pipetting steps. This specialization sacrifices SNAILS' ability to perform more complex protocols.

However, SNAILS has an extremely low cost of entry when compared to enterprise liquid

handling robots. 3D printing has become an increasingly utilized resource in the development of

biological instrumentation and laboratory consumables [27]. In the past five years, 3D printers

have rapidly decreased in price, fostering an open-sourced, at-home printing community. At the

writing of this manuscript, Creatality 3D's Ender 3 printer can be purchased for under $200 USD

and is equipped with a bed large enough to handle printing pieces for SNAILS. When printed

using 20% infill and support density of 25, the entire structure amounts to roughly 350 grams of

polylactic acid which costs approximately $30 USD. The remaining electronic components can

be inexpensively sourced (Supplementary Table 2). Together, the total cost of a single SNAILS

is approximately $165 USD. This cost of entry, including a printer, is extremely low when

considering the cost of labor required to prepare a single library or to troubleshoot a custom

protocol for an enterprise liquid handling robot. The SNAILS approach to library preparation is

an economically viable solution for consistently obtaining ultra-long reads.

Manual slow pipetting is inherently difficult to perform consistently because volumetric

flow rate isn't easily measurable or reproducible. Minute changes in volumetric flow rate caused

by slight hand movements can result in DNA shearing. SNAILS resolves these inconsistencies

through its voltage-dependent volumetric flow rate. The voltage supplied to SNAILS' motor can

be manipulated using a pulse width modulation signal delivered by a microcontroller to the

L298N driver. The stepwise changes in pipetting rate can be accurately benchmarked to

optimize ultra-long read retention. Any pipetting will inevitably result in some DNA shearing, yet

performing multiple pipetting steps is necessary to ligate sequencing adaptors to DNA

molecules. Theoretically, performing less pipetting steps in library preparation will result in less

DNA shearing. However, this may come at the cost of total reads sequenced. SNAILS can

assist in establishing an optimal protocol to maximize both number of reads sequenced and

average length of reads.

Conclusions

SNAILS is a cost-effective, easily-deployable, robotic pipette which automates long read

RAD004 library preparation for ONT sequencing. SNAILS out performs manual slow pipetting

both in terms of preserving reads exceeding 100 kb in length and acquiring larger N50s within a

sequencing library. Additionally, SNAILS allow untrained technicians to perform effective slow

pipetting while simultaneously increasing throughput of a single technician without sacrificing

quality of the library. SNAILS serves as a platform for immediately increasing the quantity and

quality of long read libraries and provides a basis for future optimization of long read retention.

Methods

CAD, 3D printing

All 3D printed portions of the robot were designed in Solidworks 2019

(https://www.solidworks.com/), a computer-aided designing and computer-aided engineering

program. All 3D printing was conducted on Ultimaker 3, S3, and S5 machines with polylactic

acid. Solidworks part files were converted to STL files and imported to the Ultimaker Cura

program (https://github.com/Ultimaker/Cura). Parts were positioned on the build plate at least

0.5 cm apart with flat faces down. For the towers, the layer height was set to 0.2 mm, infill to

20%, support material unchecked, and the remaining settings left at default. For the base, motor

holder, and bridge parts, the layer height was set at 0.2 mm, infill to 20%, support density to

25%, support material checked, and the remaining settings left at default. Before submitting the

print job to the printer, the Ultimaker was loaded with polylactic acid and the build plate was

covered in glue from the Ultimaker glue sticks to ensure the parts would not shift while printing.

To create the figures, the Solidworks parts were converted to Solidworks drawings. The *e -*

landscape.slddrt size sheet was used as a base, and the sheet format was suppressed. Multiple

angles for each part were included using the Model View tool under View Layout and

dimensions were added using the Smart Dimensions tool under Annotation with the significant

figures, size, and units edited in the Document Properties under Dimensions.

SNAILS assembly

All support material was removed, and dovetail joints were assembled. Due to the margin of

error in printing using the Ultimaker, a Dremel was used to shave down parts and a mallet was

used to squarely mesh dovetail joints as necessary. Raspbian v4.19 was installed using

NOOBS v3.3.1 to a 32 GB micro sd card and booted on a Raspberry Pi model 3B+. Electronic

components were assembled into a solderless circuit as depicted in Figure 1C. Briefly, the

general-purpose input/output pins 23, 24, 25, and ground pin 6 were connected to IN1, IN2,

ENA, and GND terminals of the L298N motor driver board, respectively. The positive and

negative leads of a twelve-volt direct current power supply were connected to the twelve-volt

and GND terminals of the L298N motor driver, respectively. Finally, the two terminals of a

twelve-volt direct current planetary motor were connected to the OUT1 and OUT2 terminals of

the L298N motor driver. Protocols were written in python3 and downloaded to the rpi

(https://github.com/emmakn/SNAILS). All sources for parts used in assembly are displayed in

Supplementary Table 2.

Animal selection and sample collection

Blood was obtained from a rhesus macaque (rh1990) housed at the Wisconsin National Primate

Research Center. Peripheral blood mononuclear cells were isolated from the buffy coat

following density gradient centrifugation with ficoll and frozen in liquid nitrogen. Sampling was

performed following protocols approved by the University of Wisconsin-Madison Institutional

Animal Care and Use Committee and in accordance with the regulations and guidelines outlined

in the Animal Welfare Act, the Guide for the Care and Use of Laboratory Animals, and the

Weatherall report (https://mrc.ukri.org/documents/pdf/the-use-of-non-human-primates-in-

research/).

High Molecular Weight DNA Extraction

DNA was extracted from frozen peripheral blood mononuclear cells using a previously described ultra-long DNA extraction protocol [24] with some modifications. For each extraction, roughly twenty million cells were thawed and then spun at 300 x g for 10 min. The supernatant was removed, and the pelleted cells were resuspended in 200 µl PBS. In a 15 ml conical tube, the resuspended cells were added to 10 ml of TLB (100 mM NaCl, 10 mM Tris-Cl pH 8.0, 25 mM EDTA pH 8.0, 0.5% (w/v) SDS, 20 ug/ml Qiagen RNase A, H_2O) and vortexed at full speed for 5 sec. The samples were then incubated at 37°C for 1 hr. Qiagen Proteinase K (200 ug/ml) was then added and mixed by slow end-over-end inversion three times. Samples were incubated at 50°C for 2 hrs with 3x end-over-end inversion every 30 min. The lysate was slowly pipetted into two phase-lock 15 ml conical tubes in 5 ml increments. The phase lock conical tubes were prepared prior to DNA isolation by adding ~2 ml autoclaved high-vacuum silicone grease into 15 ml conical tubes and spinning max speed for 1 min. To each phase-lock tube of lysate, 2.5 ml buffer-saturated phenol and 2.5 ml chloroform were added before rotational mixing at 20 rpm for 10 min. Tubes were then spun down at 4000 rpm for 10 min in an Eppendorf 5810R centrifuge with a swinging-bucket rotor. After centrifugation, the aqueous phase was poured slowly into a new phase-lock 15 ml conical tube. The addition of phenol and chloroform, rotational mixing, and centrifugation was repeated with the second set of phase-lock tubes. The aqueous phases from both phase-lock tubes were slowly poured into a single 50 ml conical tube before adding 4 ml of 5 M ammonium acetate and 30 ml of ice-cold 100% ethanol. The mixture was incubated at room temperature while the DNA visibly precipitated. Once the DNA rose to the surface of the mixture, a glass capillary hook was used to retrieve the DNA. The hooked DNA was dipped into 70% ethanol, then carefully worked off of the hook into an Eppendorf DNA LoBind 1.5 ml tube. One ml of 70% ethanol was added to the tube before spinning down at 10,000 x g for 1 min and the supernatant was removed. This process was

repeated for a second 70% ethanol wash, and any remaining ethanol was evaporated off during

a 15 min incubation at room temperature. 100 µl of elution buffer (10 mM Tris-Cl pH 8.0, H_2O)

was added before incubating the DNA at 4°C for at least two days. A total of three ultra-long

DNA isolations were performed.

Library Preparation

For each library preparation, pipetting was performed using one of three techniques: manual

aspiration, manual rotation, and automated rotation with SNAILS. All pipetting was performed

with wide-bore pipette tips. For manual aspiration libraries, volumes were pipetted up and down

by hand as slowly as possible. For manual rotation, the pipette plunger was twisted by hand as

slowly as possible to aspirate and dispense. For automated rotation, a SNAILS was

programmed to rotate the pipette plunger at a rate of approximately 1.25 rotations per minute.

An aliquot of 16 µl HMW DNA from each library was loaded into a flow cell on an Oxford

Nanopore Technologies (Oxford, United Kingdom) GridION. We used the ONT SQK-RAD004

Rapid Sequencing Kit with modifications of a previously described library preparation [24]

optimized for this kit. This procedure is also described in Karl et al. (manuscript in publication).

1.5 µl fragmentation mix (FRA) and 3.5 µl of elution buffer were added to the DNA aliquot and

pipetted five times to mix. The samples were then placed on an Applied Biosystems Thermal

Cycler (ThermoFisher Scientific, Waltham, MA, USA) at 30°C for 1 min followed by 80°C for 1

min. One µl of rapid adaptor (RAP) was added to the solution and pipetted five times. The

library was then incubated at room temperature while the flow cells were primed as follows. For

each flow cell, 30 µl of flush tether (FLT) was added to a tube of flush buffer (FLB), and a very

small volume of buffer was removed from the priming port in order to remove the air gap. 800 µl

of the Flush tether + Flush buffer solution were added to the priming port. Following a five

minute incubation, the SpotOn port was opened and 200 µl of the FLT + FLB solution was

slowly added to the priming port so that small volumes of solution rise from the SpotOn port and

then return to the cell. 34 µl of the sequencing buffer (SQB) and 20 µl of water were added to

the sample solution and pipetted three times. Finally, 75 µl of the prepared library was slowly

drawn into a pipette with a wide bore tip. The library was then added dropwise to the SpotON

port.

MinION Sequencing and base calling

Eighteen R9.4 (FLO-MIN106) total MinION flow cells were loaded with ultra-long DNA libraries

and sequenced for 48 hours and default parameters on a GridION instrument according to ONT

guidelines using their MinKNOW software to control each sequencing run. Different versions of

MinKNOW were used throughout the course of the study as updates were released; specific

versions are recorded in the metadata in the fast5 files for each run. Only reads that passed the

preliminary default screening of MinKNOW were used for base calling. Base calling was

performed using the GPU-enabled guppy version 3.2.4 software provided by ONT with the

following settings: [--config dna_r9.4.1_450bps_hac.cfg --num_callers 14 --chunk_size 500 --

gpu_runners_per_device 8 --chunks_per_runner 768]. Currently, Guppy and MinKNOW are

only   available   to   ONT   customers   via   their   community   site

(https://community.nanoporetech.com).

Fusing split reads and read statistics

Whale_watch.py was used to identify and fuse reads that were incorrectly split during

sequencing as previously described [26]. Briefly, basecalled fastq files were merged into a

single file and mapped against the human reference genome GRCh38 [28] using minimap2 [29]

and saved as a .paf output file. Next, whale_merge.py was run on the resulting .paf files along

with their corresponding merged fastq files to output fastq files containing fused reads. Read

lengths and N50s for fused fastq files were calculated using stats.py from the bbtools suite

(https://sourceforge.net/projects/bbmap/). Figures were generated using Prism 8.4.0 for MacOS

(GraphPad Software, San Diego, California USA, www.graphpad.com).

Declarations

Ethical approval and consent to participate

All applicable international, national, and/or institutional guidelines for care and use of animals were followed

Consent for publication

Not applicable.

Availability of data and materials

The datasets used during the current study are available from the corresponding author on reasonable request. The accompanying files and schematics are freely available for use at https://github.com/dholab/SNAILS.

Competing interests

The authors declare that they have no competing interest.

Funding

This work was made possible by financial support through a supplement to contract HHSN272201600007C from the National Institute of Allergy and Infectious Diseases of the NIH and grant T32 GM135119 of the NIH. This research was conducted in part at a facility constructed with support from Research Facilities Improvement Program grants RR15459-01 and RR020141-01. This work was also supported in part by the Office of Research Infrastructure Programs/OD (P51OD011106) awarded to the WNPRC, Madison-Wisconsin.

Authors' contributions

TP designed and supervised the project, performed the analysis, and wrote the manuscript. TP and EN created the SNAILS design. EN created CAD schematics, 3D printed designs, and wrote the manuscript. TP, JK, EN, and CS performed sequencing experiments. DB wrote

387   software code. HB, RW, and DO provided data interpretation. All authors read and approved the

388   final manuscript.

Acknowledgements

The authors would like to thank Phoenix Shepherd for creating the SNAILS acronym.

References

1.   Midha, M. K., M. Wu, and K. P. Chiu. 2019. Long-read sequencing in deciphering human

genetics to a greater depth. *Hum Genet* 138: 1201-1215.

2.   Nagarajan, N., and M. Pop. 2013. Sequence assembly demystified. *Nat Rev Genet* 14:

157-167.

3.   Weischenfeldt, J., O. Symmons, F. Spitz, and J. O. Korbel. 2013. Phenotypic impact of

genomic structural variation: insights from and for human disease. *Nat. Rev. Genet.* 14:

125-138.

4.   Ebbert, M. T. W., T. D. Jensen, K. Jansen-West, J. P. Sens, J. S. Reddy, P. G. Ridge, J.

S. K. Kauwe, V. Belzil, L. Pregent, M. M. Carrasquillo, D. Keene, E. Larson, P. Crane, Y.

W. Asmann, N. Ertekin-Taner, S. G. Younkin, O. A. Ross, R. Rademakers, L. Petrucelli,

and J. D. Fryer. 2019. Systematic analysis of dark and camouflaged genes reveals

disease-relevant genes hiding in plain sight. *Genome Biol* 20: 97.

5.   Stenson, P. D., M. Mort, E. V. Ball, K. Shaw, A. Phillips, and D. N. Cooper. 2014. The

Human Gene Mutation Database: building a comprehensive mutation repository for clinical

and molecular genetics, diagnostic testing and personalized genomic medicine. *Hum*

Genet 133: 1-9.

6.   Watson, C. M., P. Dean, N. Camm, J. Bates, I. M. Carr, C. A. Gardiner, and D. T.

Bonthron. 2020. Long-read nanopore sequencing resolves a TMEM231 gene conversion

event causing Meckel-Gruber syndrome. *Hum. Mutat.* 41: 525-531.

7.   Miga, K. H., S. Koren, A. Rhie, M. R. Vollger, A. Gershman, A. Bzikadze, S. Brooks, E.

Howe, D. Porubsky, G. A. Logsdon, V. A. Schneider, T. Potapova, J. Wood, W. Chow, J.

Armstrong, J. Fredrickson, E. Pak, K. Tigyi, M. Kremitzki, C. Markovic, V. Maduro, A.

Dutra, G. G. Bouffard, A. M. Chang, N. F. Hansen, A. B. Wilfert, F. Thibaud-Nissen, A. D.

Schmitt, J. M. Belton, S. Selvaraj, M. Y. Dennis, D. C. Soto, R. Sahasrabudhe, G. Kaya, J.

Quick, N. J. Loman, N. Holmes, M. Loose, U. Surti, R. A. Risques, T. A. Graves Lindsay,

R. Fulton, I. Hall, B. Paten, K. Howe, W. Timp, A. Young, J. C. Mullikin, P. A. Pevzner, J.

L. Gerton, B. A. Sullivan, E. E. Eichler, and A. M. Phillippy. 2020. Telomere-to-telomere

assembly of a complete human X chromosome. *Nature* 585: 79-84.

8.   Jain, M., H. E. Olsen, D. J. Turner, D. Stoddart, K. V. Bulazel, B. Paten, D. Haussler, H. F.

Willard, M. Akeson, and K. H. Miga. 2018. Linear assembly of a human centromere on the

Y chromosome. *Nat. Biotechnol.* 36: 321-323.

9.   Giordano, F., L. Aigrain, M. A. Quail, P. Coupland, J. K. Bonfield, R. M. Davies, G.

Tischler, D. K. Jackson, T. M. Keane, J. Li, J. X. Yue, G. Liti, R. Durbin, and Z. Ning. 2017.

De novo yeast genome assemblies from MinION, PacBio and MiSeq platforms. *Sci Rep* 7:

3935.

10.   Eccles, D., J. Chandler, M. Camberis, B. Henrissat, S. Koren, G. Le Gros, and J. J.

Ewbank. 2018. De novo assembly of the complex genome of Nippostrongylus brasiliensis

using MinION long reads. *BMC Biol* 16: 6.

11.   Miller, D. E., C. Staber, J. Zeitlinger, and R. S. Hawley. 2018. Highly Contiguous Genome

Assemblies of 15 Drosophila Species Generated Using Nanopore Sequencing. *G3*

(Bethesda) 8: 3131-3141.

12.   Tyson, J. R., N. J. O'Neil, M. Jain, H. E. Olsen, P. Hieter, and T. P. Snutch. 2018. MinION-

based long-read sequencing and assembly extends the Caenorhabditis elegans reference

genome. *Genome Res* 28: 266-274.

13.   Gong, L., C. H. Wong, J. Idol, C. Y. Ngan, and C. L. Wei. 2019. Ultra-long Read

438   Sequencing for Whole Genomic DNA Analysis. *J Vis Exp* 145:

439   14.   Schwartz, D. C., and A. Samad. 1997. Optical mapping approaches to molecular

440   genomics. *Curr. Opin. Biotechnol.* 8: 70-74.

441   15.   Yoshikawa, K. 2001. Controlling the higher-order structure of giant DNA molecules. *Adv.*

442   *Drug Deliv. Rev.* 52: 235-244.

443   16.   Kovacic, R. T., L. Comai, and A. J. Bendich. 1995. Protection of megabase DNA from

444   shearing. *Nucleic Acids Res.* 23: 3999-4000.

445   17.   Schwartz, D. C., and C. R. Cantor. 1984. Separation of yeast chromosome-sized DNAs by

446   pulsed field gradient gel electrophoresis. *Cell* 37: 67-75.

447   18.   Zhang, M., Y. Zhang, C. F. Scheuring, C. C. Wu, J. J. Dong, and H. B. Zhang. 2012.

448   Preparation of megabase-sized DNA from a variety of organisms using the nuclei method

449   for advanced genomics research. *Nat Protoc* 7: 467-478.

450   19.   Łopacińska-Jørgensen, J. M., J. N. Pedersen, M. Bak, M. M. Mehrjouy, K. T. Sørensen, P.

451   F. Østergaard, B. Bilenberg, A. Kristensen, R. J. Taboryski, H. Flyvbjerg, R. Marie, N.

452   Tommerup, and A. Silahtaroglu. 2017. Enrichment of megabase-sized DNA molecules for

453   single-molecule optical mapping and next-generation sequencing. *Sci Rep* 7: 17893.

454   20.   Pollman, M. J., and A. J. Zuccarelli. 1989. Rapid isolation of high-molecular-weight DNA

455   from agarose gels. *Anal. Biochem.* 181: 12-17.

456   21.   Sambrook, J. R. 2001. *Molecular Cloning: A Laboratory Manual.* Cold Spring Harbor

457   Laboratory Press,

458   22.   Zhang, Y., J. M. Burke, K. Gleitsman, S. M. Friedrich, K. J. Liu, and T. H. Wang. 2016. A

459   Simple Thermoplastic Substrate Containing Hierarchical Silica Lamellae for High-

460   Molecular-Weight DNA Extraction. *Adv Mater* 28: 10630-10636.

461   23.   Liu, D. K. 2020. Generating Ultra Long Reads on Oxford Nanopore

462   MinION/GridION/PromethION. *The Association of Biomolecular Resource Facilities*

24.   Jain, M., S. Koren, K. H. Miga, J. Quick, A. C. Rand, T. A. Sasani, J. R. Tyson, A. D.

Beggs, A. T. Dilthey, I. T. Fiddes, S. Malla, H. Marriott, T. Nieto, J. O'Grady, H. E. Olsen,

B. S. Pedersen, A. Rhie, H. Richardson, A. R. Quinlan, T. P. Snutch, L. Tee, B. Paten, A.

M. Phillippy, J. T. Simpson, N. J. Loman, and M. Loose. 2018. Nanopore sequencing and

assembly of a human genome with ultra-long reads. *Nat Biotechnol* 36: 338-345.

25.   Quick, J. Ultra-long read sequencing protocol for RAD004 V.3.

26.   Payne, A., N. Holmes, V. Rakyan, and M. Loose. 2019. BulkVis: a graphical viewer for

Oxford nanopore bulk FAST5 files. *Bioinformatics* 35: 2193-2198.

27.   Duong, L. H., and P. C. Chen. 2019. Simple and low-cost production of hybrid 3D-printed

microfluidic devices. *Biomicrofluidics* 13: 024108.

28.   Schneider, V. A., T. Graves-Lindsay, K. Howe, N. Bouk, H. C. Chen, P. A. Kitts, T. D.

Murphy, K. D. Pruitt, F. Thibaud-Nissen, D. Albracht, R. S. Fulton, M. Kremitzki, V.

Magrini, C. Markovic, S. McGrath, K. M. Steinberg, K. Auger, W. Chow, J. Collins, G.

Harden, T. Hubbard, S. Pelan, J. T. Simpson, G. Threadgold, J. Torrance, J. M. Wood, L.

Clarke, S. Koren, M. Boitano, P. Peluso, H. Li, C. S. Chin, A. M. Phillippy, R. Durbin, R. K.

Wilson, P. Flicek, E. E. Eichler, and D. M. Church. 2017. Evaluation of GRCh38 and de

novo haploid genome assemblies demonstrates the enduring quality of the reference

assembly. *Genome Res.* 27: 849-864.

29.   Li, H. 2018. Minimap2: pairwise alignment for nucleotide sequences. *Bioinformatics* 34:

3094-3100.

Tables

Table 1. Results for eighteen SQK-RAD004 long read sequencing libraries

DNA isolation	Method	Pores available at run start	Number of bases (Gb)	Number of reads	Longest read (bp)	N50 size (bp)	N90 size (bp)
1	Aspiration	1505	4.25	661189	737462	21327	2212
	Aspiration	1528	3.92	644261	645672	19617	2077
	Rotation	1432	2.56	283481	862322	25482	3818

	Method						
	Rotation	1362	4.32	365489	782247	30174	4891
	Automatio	1452	4.97	795794	1021308	13082	2333
	Automatio	1442	4.92	496674	1253204	29063	3933
Mean:				541148	883703	23124	3211
2	Aspiration	883	3.34	673876	682744	12695	1909
	Aspiration	1093	3.16	598310	774556	13510	2030
	Rotation	1046	5.4	617128	508355	17107	3407
	Rotation	1208	7.41	637349	526752	24163	4800
	Automatio	1182	7.21	1209053	508355	12382	2342
	Automatio	1571	10.92	1608258	526752	14589	2654
Mean:				890662	587919	15741	2857
3	Aspiration	1420	5.06	883721	605755	11809	2127
	Aspiration	1629	6.46	755277	798240	18470	3381
	Rotation	1311	8.02	915877	661883	16458	3566
	Rotation	818	4.24	386642	584414	19942	4285
	Automatio	1421	4.97	596368	602104	21358	3539
	Automatio	1410	6.89	545627	717975	31810	5864
Mean:				680585	661729	19975	3794

487 **Table 1.** Results of eighteen SQK-RAD004 long read sequencing libraries. Libraries were

488 prepared in duplicate using one of three pipetting methods. Each pipetting method was applied

489 to three separate UHMW DNA isolations. All sequencing libraries were sequenced using

490 Oxford Nanopore Technologies' GridION instrument.

492 **Table 2.** Ten longest reads per pipetting method

Rank	Aspiration	Rotation	Automation
1	798240	862322	1253204
2	774556	831285	1021308
3	737462	782247	981305
4	716306	775581	929976
5	715053	735123	907341
6	696967	729766	903921
7	694147	719004	862183
8	682744	712077	845167
9	645672	697032	800208
10	637514	673953	769538

494 **Table 2.** Top ten longest reads sequenced per pipetting method. Length is displayed in base

495 pairs and was calculated after programmatic fusion of split reads (see methods). Top ten

496 ranking is taken from the combined six sequencing libraries prepared by each pipetting method.

Figure Captions

Figure 1. Slow Nucleic Acid Instrument for Long Sequences schematics. **A.** Rendering of a full

assembled SNAILS from five individual pieces. Each piece is printed with dovetail joints that are

slid together during assembly. **B.** Dimensions of the fully assembled instrument. **C.** Circuit

diagram for SNAILS motor, L298N, and microcontroller general-purpose input/output pins are

representative of a Raspberry Pi 3B+; however, any microcontroller with configurable pins can

be used. **D.** Rotations per minute ratios for four-axle gear train. M represents the motor's axle.

Figure 2. Total reads sequenced. **A.** Total reads sequenced per DNA isolation. **B.** Comparison

of total reads sequenced per library prepared by each pipetting method. The mean total reads

sequenced is denoted by the middle bar. Standard deviation is denoted by the upper and lower

bounds.

Figure 3. Comparison of pipetting methods for long read retention. **A**. Distribution of log-

transformed read lengths. Each violin plot represents the total reads sequenced across six

library preparations. The Yellow region denotes 500-1000 kb and the orange region denotes 1-

10 Mb **B.** Averages of N50 values per pipetting method. Mean is denoted by center line.

Standard deviation is denoted by upper and lower bounds. **C.** Comparative read length

histogram. Each point represents total base pairs sequenced across 6 libraries. The Y axis

denotes log-transformed total base pairs sequenced. The X axis denotes length in base pairs.

The Yellow region denotes 500-1000 kb and the orange region reads greater than one Mb. **D.**

Number of reads greater than 100 kb sequenced per flow cell. Significant difference is denoted

by the asterisk (p-value = 0.0438).

Supplemental Figure Captions

Supplementary Figure 1. Schematics of SNAILS towers and bridge. **A.** Dimensions of the

right, dovetailed tower. **B.** Dimensions of the left, dovetailed tower. **C.** Dimensions of the

dovetailed bridge which houses gear axles.

Supplementary Figure 2. Schematics of SNAILS base and motor basket. **A.** Dimensions of the

base. The slots on either side of the base accept each towers' dovetail joints. The base is

designed with a protruding slot that fits around the base of an Ovation M micropipette. The back

right corner includes standoffs aligned to L298N boards for attachment. **B.** Dimensions of the

dovetailed motor basket. The basket is fitted to a 12 rpm HD premium planetary gear motor.

Supplementary Figure 3. Schematics of SNAILS gears. **A.** Dimensions of 14 tooth gear cap for

planetary motor. **B.** Dimensions for 16 tooth axle 2 gear. **C.** Dimensions of 24 tooth axle 2 gear.

D. Dimensions of 36 tooth axle 1 gear. **E.** Dimension of 36 tooth axle 3 gear. **F.** Dimensions of

24 tooth axle 3 gear.

Supplementary Figure 4. Schematics of SNAILS axle 1 and pipette cap. **A.** Dimensions for

axle 1. **B.** Dimensions for pipette cap. The cap is fitted for an Ovation M Micropipette.

Supplementary Table 1. Combined number of reads sequenced by length. Numbers are

indicative of the combined reads sequenced from six separate library preparations.

Supplementary Table 2. Key resources. Vendors and product IDs are listen where applicable.

A

Isometric View

B

Top View

14.50cm

14.50cm

S.N.A.I.L.S.
Slow Nucleic Acid
Instrument for
Long Sequences

Side View

5.00cm

3.64cm

4.77cm

22.20cm

4.73cm

14.50cm

Front View

0.80cm

10.50cm

5.10cm

6.00cm

5.83cm

13.70cm

0.70cm

14.50cm

C

Raspberry Pi 3B+

1	3.3v	5V	2
3	GPIO 2	5V	4
5	GPIO 3	GND	6
7	GPIO 4	TXD	8
9	GND	RXD	10
11	GPIO 17	GPIO 18	12
13	GPIO 27	GND	14
15	GPIO 22	GPIO 23	16
17	3.3V	GPIO 24	18
19	GPIO 10	GND	20
21	GPIO 9	GPIO 25	22
23	GPIO 11	GPIO 8	24
25	GND	GPIO 7	26
27	REV	REV	28
29	GPIO 5	GND	30
31	GPIO 6	GPIO 12	32
33	GPIO 13	GND	34
35	GPIO 19	GPIO 16	36
37	GPIO 26	GPIO 20	38
39	GND	GPIO 21	40

M

12V

L298N

12V
GND
5V

ENA
IN1
IN2
IN3
IN4
ENB

OUT2
OUT1

OUT4
OUT3

D

Top View

1:1 rpm
(axle3:axle2)

2:1 rpm
(axle2:axle1)

18:7 rpm
(motor:axle3)

M

A

Reads sequenced by isolation

B

Reads sequenced by method