Active focus stabilization for fluorescence microscopy by tracking actively generated infrared spots.

Birthe van den Berg[1,2]   Robin Van den Eynde[2]   Marcel Müller[3]
Wim Vandenberg[2]   Peter Dedecker[2]

[1]Declarative Languages and Artificial Intelligence, Department of Computer Science, KU Leuven, Belgium
[2]Biochemistry, Molecular and Structural Biology, Department of Chemistry, KU Leuven, Belgium
[3]Biomolecular Photonics, Faculty of Physics, Bielefeld University, Germany

birthe.vandenberg@kuleuven.be

September 22, 2020

1   Introduction

Keeping a sample in focus remains a challenging, yet important, task in optical microscopy. Even a minimal discrepancy in Z-position can cause severe disturbances for time-lapse, scanning or TIRF-based experiments. Focus drift can arise from a range of sources, such as vibrations, mechanical instabilities, samples that are not flat or thermal fluctuations of the microscope's components.

Various techniques and systems have been proposed to address or circumvent these problems. They can be divided into two major categories: active and passive methods. Active methods accept that there is drift and compensate it using a control loop. Image-based methods [1, 2] rely on complex software that analyzes images and compares them with in-focus images captured beforehand. Sensor-based methods use additional sensing circuits to detect drift. Passive methods use optical components [3, 4] to monitor the sample-objective distance and include a range of mechanic practicalities, from using a weight to push down the sample to avoiding a perturbing airflow over the system.

Optical z-drift correction systems use an infrared light source that is reflected on the cover glass-sample interface and captured by a detector. These detectors range from quadrant detectors [5] to cameras (e.g. Leica's Adaptive Focus Control) and line sensors (e.g. Nikon's Perfect Focus System). In these systems, a spatial disturbance between the sample and objective influences the position of the reflected light, which is then mitigated by moving the sample or objective towards the desired focus position. These systems are rapidly becoming standard equipment on new motorized microscopes, though may be more challenging to engineer in existing systems.

In this paper we present a z-drift correction system that was implemented both on a popular commercial microscope body (Olympus IX-71) and on a bespoke modular microscope system, though which should also be compatible with systems from other manufacturers.

2   Concept

Figure 1 shows the concept of our automated out-of-focus system, which is based on the in-situ real-time focus detection method of Binh et al. [6]. The main idea is to illuminate the sample with two light beams that are reflected from the glass surface carrying the sample, and to relate the relative motion of these beams to the

sample position. We obtain these two beams by passing collimated infrared laser light through a double-hole mask. The beams are sent into the microscope body, onto the sample, where they will reflect on the coverslip, or any other glass-other refractive index surface. As sample and glass surface are close enough together and move more or less in unison, there's no drawback to this. The reflected beams are captured by an industrial CMOS camera.

The horizontal distance between the detected spots is proportional to the z-position of the cover glass. The control software then adjusts the stage position to regain focus by restoring the distance between both spots. Our software determines the positions of the reflected spots by simply calculating the 'center of mass' of each spot (after initial filtering to avoid artifacts, see below).

Figure 1: Concept.

Moreover, our system is designed to eliminate any residual light originating from reflections in the beam path.

In regular correction techniques that employ a single infrared beam that is detected by a quadrant detector, the shape of the spot is of the utmost importance. Any distortions (also not focus-drift related) of this beam profile will directly result in a focusing error as the center-of-mass algorithm is largely influenced by imperfections at the border of the spot. Our technique, with the double-hole mask, will result in a smaller misconception of the z-postion, as long as the disturbance influences both beams in a similar way.

3   Description of the System

We provide a full description of the setup components and alignment procedure. This setup is developed with the eye on being compact and straight-forward to construct and use. We also provide software for automatically detecting and correcting z-drift, including a user-friendly GUI.

3.1   Setup

Figure 2 gives a schematic overview of the hardware that is used to construct the autofocus system.

The light source is a 3.0 mW, 830 nm infrared laser with an elliptical beam profile, mounted in a pitch/yaw adaptor for tilting the laser on two axes. This wavelength is chosen to avoid the color channels used for fluorescence detection. The double-hole mask is positioned in a rotation mount to tilt the laser in the long axis of the elliptical beam. These holes are 600 μm apart and 500 μm wide. Figure 3 shows this mask and its dimensions.

Two mirrors M1 and M2 can flexibly position the two beams, so that they can be guided towards the objective-sample interface. Component ND1 is a neutral density filter that reduces the power of the laser to less than 1 mW (safety category 2), while still retaining a clear signal on the detector. Lenses L1 and L2 both have a focal

Figure 2: Schematic overview of the system.

length of 50.0 mm and are aligned in a 4f configuration. These lenses extend the beam path in order to be able to modify the incident angle and position of the beams with mirrors M1 and M2.

A 50/50 beam splitter passes the light coming from the laser into the direction of the microscope. The two beams will be reflected on the glass-other refractive coverslip that holds the specimen. Upon reflection, the light is send back towards the 50/50 beam splitter. This light is then split again, this time with the aim to detect the two beams on a camera.

As the infrared laser provides a reasonably coherent source, pattering artifacts can arise in this process due to interference. However, as these occur at rather high spatial frequency, they can easily be filtered in post-processing. Component ND2 is a laser line filter in order to compensate for imperfections in the beam splitter (e.g. low-intensity secondary laser lines) and D1 and to filter out ambient light, that may come from the fluorescence path.

An extended version of the system is shown in the red dashed frame in Figure 2. This extension is required if extra reflections occur on the camera that might disturb or overwrite the spots to be measured. These reflections may originates from several sources: imperfections in optical components (beam splitter, dichroics), infrared light reflected off of lenses shared with the fluorescence path.

Mirror M3 is used to center the beams on the iris, to maximally filter out light noise. Lenses L3 and L4 both

Figure 3: Cross-section of the double-hole mask with technical dimensions (μm).

have a focal length of 50.0 mm and are in a 4f configuration. The spatial filter is placed in the (back)focal plane of the two lenses and is the primary component for eliminating reflections. Lens L5 focusses the two parallel infrared laser beams on the IDS camera window. The distance between this imaging lens and the sensor chip of the IDS camera influences the shape of the spots and the distance between them.

3.2 Software: GUI

Figure 4a shows the graphical user interface (GUI) of the software, and Figure 4b shows the actual spots on the camera window.

(a) Graphical User Interface (GUI).

(b) IDS camera view.

Figure 4: Software for adapting the z-position of the stage.

The `Focus tracking` window offers two functionalities: (de)activating the camera, and (de)activating the focus tracking algorithm. The current distance between the two spots on the camera is defined as `Current position` (in pixels). The ideal z-position of the stage (i.e. when in focus) is defined as `Focus position`. When focus tracking is enabled with the `Track focus` button, the software aims at keeping the distance between the detected spots constant. It adapts to reach this state by moving the stage or objective up and down when the

sample is drifting.

The `Settings` window allows the user to set `Exposure time`, `Min.deviation` and `Step size`. Depending on the intensity of the reflecting laser beams, the exposure time can be set as desired with the `Exposure time` button. The exposure time also determines the speed at which the autofocus system can detect and correct disturbances in the relative position of the spots.

The `Min.deviation` button refers to the least difference (in pixels) there should be between the current position and the focus position in order for the stage or objective to move.

The `Step size` button represents the relation of the z-position of the stage and the distance between the spots on the IDS camera. If the distance between the lobes changes with 1 pixel, the stage will move $(\text{focus pos} - \text{aim}) * \text{step size}$ μm. This step size changes depending on the magnification of the objective and lenses that are used in the system. The higher the magnification, the smaller the working distance and focus depth.

Since the user can set these parameters as desired, it is easy to adapt the software to work for different samples, objectives or microscopes.

The `Stages` window has the absolute positions of the stage in the x-, y- and z-direction. It also provide buttons to move the stage up and down. The `Calibrate` button sets the absolute x-, y- and z-value to zero. This button should be used as a calibration step before finding the correct focus position. It allows the user to set a neutral position, where the objective is relatively far away from the camera but the spots are clearly visible. When the system is not able to find its focus back within a reasonable time, it returns to this position to start correcting the stage.

The autofocus system is compatible with different hardware components. Currently, we support IDS cameras for spot detection but others can be easily added. Moreover, we can control anything that is used through Micromanager but also different kinds of stages, as long as they are motorized.

4 Results

We validated our z-drift correction system by visualising the mitigation of temperature-dependent drift on a fluorescence microscope. As a sample we used sparcely disperced 200 nm Tetraspeck beads. The microscope itself consists of: an Olympus IX71 inverted microscope body, Hamamatsu ORCA-Flash4.0 LT+ Digital CMOS camera, Spectra X light engine (Lumencor), a dichroic turret wheel mounting a Chroma ZT561RDC dichroic filter with a Chroma HQ572lp emission filter. This microscope supports a wide range of objective lenses. We use an UMPlanFI $50\times$ objective, with a numerical aperture of 0,80 and a working distance of 0,66 mm. We have 542 nm illumination light at a power of 50%; the exposure time is 0,1s.

We will visualize z-drift as a result of a change in temperature. For this purpose, we use a time lapse experiment, imaging every second, while the temperature is varied between 32°C and 45°C using the Okolab H301-T-UNIT-BL-PLUS stage top incubator.

When the autofocus sytem is activated, we first have to determine the focus position, find good spots and set the step size correctly for the microscope and the objective. For this, we take a z-stack of images (Figure 6) to relate the number of pixels with the deviation in z-position of the stage.

We use a step size of 0,1 μm/pixel. The minimal deviation is set to 2 pixels.

As Figure 5 shows, the sample clearly moves out of focus when the temperature is changing. With our system, the focus position is recovered by adapting the stage. During the experiment, also xy-drift occuers, for which our system does not compensate.

Similar results were achieved with a UPlanSApo $10\times$ objective with a numerical aperture of 0,40 and a working distance of 3,1mm. Moreover, a variant of the same tracking system has also been implemented on the SIM.

5 Conclusion

We have described and built an automated out-of-focus detection and correction system that is inexpensive, easy to build, and compatible with any microscope that has an actuated stage or objective. This system keeps a sam-

Figure 5: Time-lapse experiment with autofocus deactivated.

Figure 6: Z-stack of the autofocus system.

ple in focus during time-lapse experiments where z-drift occurs. Extending the system with different hardware components is quite easy as new stages or cameras just need to be added to the API.

For more sensitive systems, it would be easy to synchronize both the laser and the camera so that there is only infrared light when really needed. Also, the rest of the measurement protocol, e.g. so that infrared light does not leak into the main image acquisition. The software could be extended to reacquire different focus positions so that, when they have been stored once, the microsope can be back to them. Moreover, we could improve the software to support scanning.

Acknowledgements

Robin Van den Eynde thanks the FWO for a doctoral fellowship.

References

[1] Siavash Yazdanfar, Kevin B. Kenny, Krenar Tasimi, Alex D. Corwin, Elizabeth L. Dixon, and Robert J. Filkins. Simple and robust image-based autofocusing for digital microscopy. *Opt. Express*, 16(12):8670–

Figure 7: Time-lapse experiment with autofocus activated.

8677, Jun 2008.

[2] Sihong Li, Xiaodong Cui, and Wei Huang. High resolution autofocus for spatial temporal biomedical research. *The Review of scientific instruments*, 84:114302, 11 2013.

[3] Yuvalal Liron, Yael Paran, N Zatorsky, Benjamin Geiger, and Z Kam. Laser autofocusing system for high-resolution cell biological imaging. *Journal of microscopy*, 221:145–51, 03 2006.

[4] Alexandros Pertsinidis, Yunxiang Zhang, and Steven Chu. Subnanometre single-molecule localization, registration and distance measurements. *Nature*, 466:647–51, 07 2010.

[5] M. Bathe-Peters, P. Annibale, and M. J. Lohse. All-optical microscope autofocus based on an electrically tunable lens and a totally internally reflected ir laser. *Opt. Express*, 26(3):2359–2368, Feb.

[6] Binh Cao, Phuong Hoang, Sanghoon Ahn, Jeng-O Kim, Heeshin Kang, and Jiwhan Noh. In-situ real-time focus detection during laser processing using double-hole masks and advanced image sensor software. *Sensors*, 17, 07 2017.

A List of Components

This section gives an overview of the necessary components to build this system, together with their reference number, their price and a description of the component.

The system as described in Section 3.1 needs the following components:

#	Item	Part number	Unit price (€)	Description
1x	Infrared laser	CPS830	103,28	Collimated Laser Diode Module, 830 nm, 3.0 mW, Elliptical Beam, ⌀11 mm Housing
1x	Tilting mount	KAD11F	63,44	Threaded Kinematic Pitch/Yaw Adapter for ⌀11 mm Cylindrical Components
1x	Rotation mount	LRM1	84,86	Rotation Mount for ⌀1" (25.4 mm) Optics, External SM1 Threads
1x	Double-hole mask	*Custom*	/	Space between holes: 600 mm, size of holes 500 mm
3x	Mirror M1-M2-M3	BB1-E03	70,31	⌀1" Broadband Dielectric Mirror, 750 - 1100 nm
2x	Lens L1-L2	AC254-050-B	85,83	f = 50.0 mm, ⌀1" Achromatic Doublet, ARC: 650 - 1050 nm
1x	Beam splitter	BT610/M	291,16	Beam Trap, 400 nm - 2.5 µm, 30 W Max Avg. Power, Pulsed and CW, M4 Tap
1x	Lens L3	LA1509-B	30,50	N-BK7 Plano-Convex Lens, ⌀1", f = 100 mm, AR Coating: 650 - 1050 nm
1x	IDS camera	UI-3260CP Rev. 2	364,00	Sony's 2.35 MP sensor IMX 249 (1936 x 1216 px)
1x	Filter ND1	NE10A-B	72,06	⌀25 mm AR-Coated Absorptive Neutral Density Filter, 650-1050 nm, SM1-Threaded Mount, OD:1.0
1x	Filter ND2	FL830-10	92,46	⌀1" Laser Line Filter, CWL = 830 ±2 nm, FWHM = 10 ±2 nm

The components needed for this system cost €1484,35 in total, with as most expensive components the beam splitter and the IDS camera. If extra reflections are eliminated with a spatial filter, extra components are needed, in addition to the previously listed components.

#	Item	Part number	Unit price (€)	Description
1x	Mirror M4	BB1-E03	70,31	⌀1" Broadband Dielectric Mirror, 750 - 1100 nm
2x	Lens L4-L5	AC254-050-B	85,83	f = 50.0 mm, ⌀1" Achromatic Doublet, ARC: 650 - 1050 nm
1x	Iris diaphragm	SM1D12SZ	71,57	SM1 Lever-Actuated Zero Aperture Iris Diaphragm, ⌀11.9 mm Max Aperture

This extra list has a price of €313,54. The total price of the system is €1.797,89.

B Alignment Procedure

We have the freedom to align the system as desired. It is an easy alignment procedure, that is described in the following guide. Activating the crosshair may help centering the beams on the camera window.

1. Align the infrared laser: Make the elliptical infrared light beam as bright as possible using the tilting mount, center it using a detector card and a spatial filter.

2. Align the rotation mount and mask together so that the two spots are centered and as bright as possible (Figure **??**). Use a detector card and a spatial filter.

3. Put lenses L1 and L2 in 4F configuration.

4. Use the two mirrors M1 and M2 to center the two spots. Use a detector card and a spatial filter.

5. Place the IDS camera with L3 behind the beam splitter in an unfixed position. Use the mirrors M1 and M2 to make the spots as bright and as circular as possible on the IDS camera. Make sure that these spots move horizontally and in different directions.

6. Place the IDS camera and L3 in a fixed position behind M3. Use M3 to center the two spots on the camera window.

7. (E) Put lenses L4 and L5 in 4F configuration. Place the iris diaphragm (spatial filter) at the focus point of the two lenses. This focus point should be in the middle between L4 and L5.

8. (E) Place the IDS camera and L3 in a fixed position behind M4. Use M3 and M4 to center the two spots on the camera window.

9. Vary the distance between lens L3 and the IDS camera to obtain the desired spacing between the two spots. This also has an impact on the form of the spots.

10. (E) Use the spatial filter to eliminate noise and undesired spots.

11. Place the IDS camera in the correct angle in order to let the spots move horizontally.

Steps 7, 8 and 10 are only required in case there are extra spots to be eliminated.