1 **A polymer gel index-matched to water enables diverse applications in fluorescence**
2 **microscopy**
3

4 Xiaofei Han[1,2*,^], Yijun Su[1,3*,^], Hamilton White[4,5*,^], Kate M. O'Neill[6, 7], Nicole Y. Morgan[8], Ryan
5 Christensen[1], Deepika Potarazu[1], Harshad D. Vishwasrao[3], Stephen Xu[1], Yilun Sun[9], Shar-yin
6 Huang[9], Mark W. Moyle[10], Qionghai Dai[2], Yves Pommier[9], Edward Giniger[6], Dirk R. Albrecht[4,11,]
7 [&], Roland Probst[12, &], Hari Shroff[1, 3, 13, &]
8

1. Laboratory of High Resolution Optical Imaging, National Institute of Biomedical Imaging
and Bioengineering, National Institutes of Health, Bethesda, Maryland, USA
2. Department of Automation, Tsinghua University, 100084, Beijing, China
3. Advanced Imaging and Microscopy Resource, National Institutes of Health, Bethesda,
Maryland, USA
4. Department of Biomedical Engineering, Worcester Polytechnic Institute, 100 Institute
Road, Worcester, MA, 01609, USA
5. Department of Neurobiology, University of Massachusetts Medical School, Worcester,
MA 01655, USA
6. National Institute of Neurological Disorders and Stroke, NIH, Bethesda, MD, 20892, USA.
7. Institute for Physical Science and Technology, University of Maryland College Park,
College Park, Maryland, USA
8. National Institute of Biomedical Imaging and Bioengineering, National Institutes of
Health, Bethesda, Maryland, USA
9. Developmental Therapeutics Branch and Laboratory of Molecular Pharmacology, Center
for Cancer Research, National Institutes of Health, Bethesda, Maryland, USA
10. Department of Neuroscience and Department of Cell Biology, Yale University School of
Medicine, New Haven, CT 06536, USA.
11. Department of Biology and Biotechnology, Worcester Polytechnic Institute, 100 Institute
Road, Worcester, MA, 01609, USA.
12. ACUITYnano, Innovation in Biomedical Imaging, North Bethesda, Maryland, USA
13. Marine Biological Laboratory Fellows Program, Woods Hole, MA, USA
31

32 *Equal contribution
33 & Equal contribution
34 ^ Correspondence to hxf16@mails.tsinghua.edu.cn, suy4nih@gmail.com, hawhite@wpi.edu
35

36 **Abstract**
37

We demonstrate diffraction-limited and super-resolution imaging through thick layers (tens-
hundreds of microns) of BIO-133, a biocompatible, UV-curable, commercially available polymer
with a refractive index (RI) matched to water. We show that cells can be directly grown on BIO-
133 substrates without the need for surface passivation and use this capability to perform
extended time-lapse volumetric imaging of cellular dynamics 1) at isotropic resolution using
dual-view light-sheet microscopy, and 2) at super-resolution using instant structured

illumination microscopy. BIO-133 also enables immobilization of 1) *Drosophila* tissue, allowing
us to track membrane puncta in pioneer neurons, and 2) *Caenorhabditis elegans*, which allows
us to image and inspect fine neural structure and to track pan-neuronal calcium activity over
hundreds of volumes. Finally, BIO-133 is compatible with other microfluidic materials, enabling
optical and chemical perturbation of immobilized samples, as we demonstrate by performing
drug and optogenetic stimulation on cells and *C. elegans*.
50
51 **Introduction**
52
Fluorescence microscopy spurs biological discovery, especially if imaging is performed at
high spatiotemporal resolution and under physiologically relevant conditions. Coupling
fluorescence microscopy with strategies for immobilizing or confining samples enables further
applications, particularly when studying organisms that move rapidly. For example, the
transparency and genetic accessibility[1] of the nematode *C. elegans* has made it an ideal system
for studying the growth, morphology and function of individual cells in the context of the whole
organism[2,3]; yet imaging the living animal without motion blur usually requires immobilization
with chemical[4,5], steric[6,7], or microfluidic[8-11] means.
Microfluidic systems provide efficient immobilization and handling[12-18] for studying
cellular morphology[19,20] and dynamics[21], neuronal function[22-24], behavior[25-27], and lifespan[28,29].
Hydrogels (either independently[7,30] or in conjunction with microfluidics[31]) have also been
demonstrated as highly useful materials with tunable mechanical[32], diffusive, and optical
properties[33] that are well-suited for long-term imaging applications[7,34,35].
Unfortunately, relatively few attempts have been made to index-match immobilization
devices[36-39]. The high refractive index (RI, n) of materials commonly used in microfluidics, such
as polydimethoxylsilane (PDMS), causes significant optical aberrations[40,41] due to the RI
mismatch that occurs at the interface between the polymer ($n_{PDMS} \sim 1.41$) and an aqueous
sample ($n_{water} = 1.33$). These aberrations severely degrade image focus, resolution, and signal,
compromising the performance of immobilization devices by reducing the information content
of the resulting data. Hydrogels offer a lower RI (n ranging from $1.34 - 1.41$) depending on
thickness and polymerization conditions[33]. Although the RI of these materials is better matched
to living samples, even a small mismatch in RI causes a noticeable deterioration in image
quality[42]. Image degradation is particularly obvious when using water-dipping lenses designed
for imaging living samples, such as those employed in high-resolution light-sheet fluorescence
microscopy (LSFM)[43-45].
Here we demonstrate a broadly applicable refractive-index-matched specimen
mounting method that introduces negligible aberration when imaging living samples with high-
resolution light-sheet microscopy and super-resolution microscopy. We show its utility in
combination with microfluidics, enabling applications in high-resolution, volumetric imaging of
cells, *Drosophila* tissue, and *C. elegans* adults and larvae. Our method takes advantage of the
commercially available UV curable optical polymer BIO-133 (MY Polymers Ltd.) that has a
refractive index matched to water (n = 1.333), is non-fluorescent, and is non-toxic. We show
that 1) BIO-133 provides a gas permeable, inert, and biocompatible scaffold on which to grow
and image tissue culture cells, 2) enables rapid tissue or animal encapsulation, and 3) is
compatible with other microfluidic mounting schemes and optical or chemical perturbations.

Results

BIO-133 does not introduce additional optical aberrations

We assessed the optical properties of BIO-133 by using dual-view light-sheet microscopy
(diSPIM[43,46]) to image 100 nm yellow-green beads placed under polymer layers of progressively
increasing thickness (**Fig. 1, Methods**). In the most common diSPIM implementation, two
identical 0.8 numerical aperture (NA) water-dipping objectives mounted above a planar
substrate (usually a glass coverslip) alternately illuminate the sample with a light sheet and
detect the resulting fluorescence. Since both illumination and detection planes are angled at
~45 degrees with respect to the glass coverslip, imaging through a polymer gel with surface
parallel to the coverslip will introduce significant lateral ('x' direction, **Fig. 1a**) and axial ('z'
direction, **Fig. 1a**) aberrations if the gel's refractive index differs from that of water.
As predicted, we observed this effect when imaging beads embedded under polymers
with different RIs (**Fig. 1b, c**). Under no polymer, images of beads approximate the system point
spread function, with measured lateral and axial full width at half maximum (FWHM) 395.9 +/-
7.7 nm and 1527.9 +/- 119.5 nm (*N* = 70 beads), respectively. Imaging beads under PDMS
caused severe aberrations (**Fig. 1c**), more than doubling the lateral FWHM under a 25 μm layer
(816.8 +/-24.9 nm) with progressive deterioration under thicker polymer layers (**Fig. 1b**). We
also observed aberrations (**Fig. 1b, c**) under poly(ethylene-glycol) diacrylate (PEG-DA)[7] and
fluorinated ethylene polymer (FEP)[47], albeit to lesser extent as the refractive indices of these
polymers are closer to water. By contrast, beads imaged under BIO-133 showed negligible
visual aberrations or measurable degradation in image quality (**Fig. 1b, c, Supplementary Table
1**), even under a 150 μm thick film, the largest thickness we tested (lateral FWHM 416.5 +/- 8.5
nm). We attribute this result to the refractive index of BIO-133, which we measured to be
1.333. We also verified, under 488 nm, 561 nm, and 637 nm illumination, that BIO-133
introduces negligible autofluorescence (**Supplementary Fig. 1**).

High- and super-resolution imaging of cells through a layer of BIO-133

BIO-133 has permeability to oxygen about 2-3 times greater than PDMS and has water
repellent properties (Personal Communication, Ehud Shchori, My Polymers, Ltd.). These
material properties are advantageous for maintaining physiologically relevant cell culture
conditions. Thus, we investigated whether BIO-133 could provide an inert and biocompatible
scaffold for single cell imaging (**Supplementary Table 2a**). After a curing and leaching treatment
(**Methods**), U2OS cells seeded on a 50 μm layer of BIO-133 adhered and displayed similar
morphology and growth rate to cells grown on glass coverslips (**Fig. 2a, Supplementary Fig. 2**).
Similar results were obtained using HCT-116 (human colon carcinoma) cells that express
endogenous topoisomerase I-GFP, and we also observed similar expression and localization of
tagged proteins compared to cells cultured on glass coverslips (**Supplementary Figs. 2, 3**).
To demonstrate that transfected cells seeded directly on BIO-133 could be imaged at
high spatiotemporal resolution, we created BIO-133 substrates on PDMS supports
(**Supplementary Fig. 4**) and imaged cells expressing mEmerald-Tomm20, a fluorescent marker

of the outer mitochondrial membrane, through a 50 μm thick BIO-133 layer using diSPIM (**Fig.**
2b, c). The jointly registered and deconvolved data acquired from two views displayed isotropic
spatial resolution (**Fig. 2c, d**), allowing us to clearly visualize individual mitochondria and their
dynamics (**Supplementary Video 1**), including mitochondrial fusion and fission (**Fig. 2d**). We
also used BIO-133 in conjunction with diSPIM to construct a simple gravity-driven flow
cytometry setup, obtaining clear images of DAPI stained nuclei as they flowed through the
chamber (**Supplementary Fig. 5, 6, Supplementary Video 2**).
Next, we sought to image subcellular targets at spatial resolution beyond the diffraction
limit, so we turned to instant structured illumination microscopy (iSIM)[48], which enables high-
speed super-resolution imaging. We again seeded U2OS cells expressing mEmerald-Tomm20 on
a 50 μm BIO-133 film, this time imaging them with iSIM using a water dipping lens in an
inverted geometry (**Fig. 2e**). Again, BIO-133 provided aberration-free imaging, enabling us to
visualize the internal mitochondrial space absent Tomm20 (**Fig. 2f, g, Supplementary Video 3**).
We also visualized LAMP-1-GFP-stained lysosome dynamics in wild type HCT-116 cells grown on
another BIO-133 film of 50 μm thickness (**Supplementary Fig. 7, Supplementary Video 4**). As a
third example, we grew multiple layers of HCT-116 cells on BIO-133, and immunostained the
cells for lamin A/C, Tomm20, and actin (**Fig. 2h**), obtaining clear images of these structures
through the volume of the sample (**Fig. 2i, Supplementary Video 5**). We conclude that BIO-133
is compatible with multicolor, super-resolution imaging in live and fixed targets.
151
152 *BIO-133 enables subcellular imaging, segmentation, and tracking within immobilized living*
153 *tissue*
154
In addition to monitoring the dynamics of organelles within single cells, we also
immobilized and imaged multicellular structures in flies and worms at high spatiotemporal
resolution, using diSPIM in conjunction with BIO-133 for sample immobilization (**Fig. 3,**
Supplementary Table 2b). The developing *Drosophila* wing has long been a model for axon
growth and neuronal pathfinding and differentiation[49,50]. More recently, spinning disk confocal
microscopy was used to dissect the role of cytoskeletal organization and dynamics in shaping
the morphogenesis and growth of the TSM1 pioneer sensory neuron axon in explanted early-
pupal wing imaginal discs[51,52]. In those experiments, phototoxicity and photobleaching limited
imaging duration to ~30 volumes, with image volumes acquired every 3 minutes. Wings were
sandwiched between two glass coverslips to immobilize the preparation and prevent it from
moving during imaging, but this scheme introduces unacceptable aberrations if imaging with
the less phototoxic diSPIM. Instead, we immobilized wings with a thin layer of BIO-133 (**Fig. 3a,**
Supplementary Fig. 8), which enabled sustained volumetric imaging with diSPIM. We acquired
360 single-view volumes (5 s inter-volume interval, spanning 30 minutes) of tdTomato-CD4
expressed in TSM1 and the neighboring L3 neuron, marking neuronal membranes (**Fig. 3b**). In
addition to observing slower remodeling of the TSM1 growth cone (**Fig. 3c, top**) our imaging
rate also enabled us to capture rapid movement of membrane-labeled puncta that appeared to
traffic along the L3 axon shaft (**Fig. 3c, bottom, Supplementary Video 6**). Puncta were also
evident in comparative spinning disk confocal datasets (**Supplementary Fig. 9**).
In another example, we used soft lithography techniques[53] to cast BIO-133 into
microfluidic devices suitable for trapping *C. elegans* (**Supplementary Fig. 10**). Introducing adult

worms into the channels via suction (**Fig. 3a**), we imaged fine structures (**Fig. 3d**) and functional
activity (**Fig. 3e-j**) in living animals at isotropic resolution. Animals were sufficiently immobile
that we could serially acquire and fuse the two diSPIM views[54] to obtain reconstructions free of
motion blur. In strains expressing GFP sparsely targeted to a few neurons, we resolved axons
and dendrites (likely from amphid neurons, **Fig. 3d**) within anesthetized *C. elegans*. When
imaging the genetically encoded calcium indicator GCaMP6s[55] and mCherry targeted pan-
neuronally[56] in immobilized adult animals without anesthetic, our volume imaging rate of 1.25
Hz (simultaneous acquisition of red and green channels) enabled us to segment and track 126
nuclei in the animal head (**Fig. 3e,f Supplementary Video 7**), permitting inspection of
spontaneously active nuclei (**Fig. 3f, g**) over our 450 volume (6 minute) experiment. Intriguingly,
we observed a pair of nuclei (#79 and #15, **Fig. 3e, g**) that exhibited in-phase, rhythmic activity
with slow (45-80 s) period (**Supplementary Video 7**), as well as nuclei showing out-of-phase
activity with respect to this pair (#27). In another experiment, we simply embedded *C. elegans*
larvae expressing the same pan-nuclear GCaMP6s marker in a cured disk of BIO-133 (**Fig. 3a**),
recording volumes from one side to obtain volumes at 4 Hz, for 250 volumes. Despite the
poorer axial resolution of single-sided imaging, and the smaller size of the larval nuclei, we were
able to again segment and track 110 nuclei in the head of the animal, identifying calcium
transients in spontaneously active nuclei with a time resolution of 0.25 s (**Fig. 3h-j,**
Supplementary Video 8).
The droplet-based design also enabled easy recovery of animals post-imaging. 26 / 28
animals were recovered even ~12 hours after embedding, confirming our suspicion that cured
BIO-133 is inert, gas permeable, water repellant, and does not obviously affect animal viability
(the remaining two animals died within the BIO-133 capsule due to internal hatching of
embryos within the animals). The water repellency of BIO-133 likely contributes to retaining the
animal's intrinsic hydration and thus viability during encapsulation. The ease at which *C.*
elegans can be immobilized and imaged at high spatiotemporal resolution suggests useful
synergy with multicolor strategies that permit unambiguous neural identification[23].
203
204 *BIO-133 is compatible with chemical and optogenetic perturbations*
205
The ability to specifically perturb and subsequently follow biological processes by
observing morphological or functional changes is valuable in dissecting biological processes.
We conducted several studies to show that BIO-133-mounted samples are compatible with
such perturbations (**Fig. 4**). First, we conducted a simple drug assay by modifying our BIO-
133/PDMS cellular scaffolds (**Fig. 2b, Supplementary Fig. 11**) so that U2OS cells could be
exposed (**Fig. 4a**) to carbonyl cyanide m-chlorophenyl hydrazine (CCCP), an inhibitor of
oxidative phosphorylation. Because we could clearly observe cells through the BIO-133 layer
using diSPIM, we observed that, compared to control cells in a neighboring well (**Fig. 4a-c**),
within minutes of exposure the treated cells showed mitochondrial fragmentation, eventually
exhibiting major disruption to the mitochondrial network (**Fig. 4b, c, Supplementary Video 9**).
Chemically stimulating animals directly embedded in BIO-133 is difficult, since BIO-133 is
not permeable to aqueous solutions. One solution is to introduce chemicals via microfluidic
channels (such as those shown in **Fig. 3d-g**). Alternatively, we explored using PEG-DA for
immobilization and aqueous permeability, above a microfluidic layer for stimulus introduction

and control, and beneath a BIO-133 layer to enclose the fluidic path. Using a thin PEG-DA disk
allows easy transfer of different embedded organisms on the same imaging setup, and
repeated imaging of the same animals over many hours if desired. We constructed a hybrid
multi-material device composed of a PDMS microfluidic base bonded to a BIO-133 upper
membrane that sealed in a small PEG-DA disk containing tens of embedded nematodes[7]
(**Supplementary Fig. 12, Fig. 4a**). Chemicals applied via flow channels diffuse into the PEG-DA
disk, evoking neural responses that can be imaged through the BIO-133 viewing layer with
widefield microscopy or diSPIM (**Fig. 4d-g, Methods**). We embedded 15 animals expressing
GCaMP2.2b in AWA chemosensory neurons[57] in a PEG-DA disk and applied 1.1 µM diacetyl
pulses, which directly activate these neurons via the ODR-10 chemoreceptor[57,58]. Using
widefield microscopy, we recorded robust calcium transients from all animals and observed a
characteristic sensory adaptation to repeated stimulation (**Fig. 4e**). The initiation of sensory
neural responses varied slightly due to diffusion of diacetyl stimulus through the PEG-DA disk to
animals embedded in different planes (**Supplementary Fig. 13**). Next, we examined individual
neuron responses by using the same apparatus with diSPIM (**Fig. 4f,g Supplementary Video
10**). Our imaging provided sufficient spatial resolution to distinguish subcellular responses,
observing faster on and off dynamics of fluorescent transients in the dendrites than in the
soma[59].

Other stimulation modalities (such as light, temperature, and mechanical vibration) are
directly transmissible through BIO-133, and these can be applied directly to cells and organisms
embedded in the polymer in a simpler preparation. Optogenetic neural activation is a
particularly advantageous tool, allowing remote light-induced activation or suppression of
neurons. We embedded nematodes expressing the red light activated cation channel
Chrimson[60] and GCaMP2.2b[7] directly in BIO-133 disks and monitored calcium readout with
diSPIM during repeated red light stimulation pulses (**Fig. 4a,h,i, Supplementary Video 11**). We
observed increases in fluorescence after each stimulus (**Fig. 4h**), and again could clearly localize
such transients to subcellular areas including soma, dendrite, and axon (**Fig. 4i**).

Discussion

BIO-133 is commercially available, rapidly curing, gas permeable, inert, water repellent,
and biocompatible. It is not autofluorescent under visible illumination and does not introduce
additional aberration when imaging with water-dipping or water-immersion objective lenses
designed for aqueous specimens. These characteristics make it well-suited to microfluidic
experiments under physiological conditions, particularly with the many LSFM systems that use
such lenses. We suspect that capillary mounting[47], often used for mounting zebrafish in LSFM,
could be improved if BIO-133 were used instead of the FEP material commonly used in this
application. Our finding that cells can be directly grown on BIO-133 without additional surface
treatment may prove useful in non-standard LSFM geometries that previously employed RI-
matched materials with passivated surfaces[45], or in ultra-high-throughput light-sheet
imaging[61,62].
We also found that BIO-133 does not noticeably degrade imaging performed in more
traditional inverted microscope geometries, including in super-resolution imaging. This
capability suggests that using BIO-133 could improve imaging in studies of cell morphology,

mechanics, migration, and motility, e.g., when using micropillars[20,63] or in traction force
microscopy[64].
We bonded BIO-133 to glass and PDMS with silicone-based adhesive tape, or reversibly
to glass via van der Waals forces. The elastic modulus of cured BIO-133 (5 MPa) is similar to that
of PDMS (3.7 MPa). Thus, similar to PDMS[65], BIO-133 conforms to minor imperfections in glass
and bonds to it by weak van der Waal forces, creating a reversible bond and a watertight
seal. We suspect further tuning of adhesive, optical and mechanical properties of this intriguing
polymer is possible but will depend on knowing the chemical formula, which is currently
proprietary.
273
274 **Author Contributions**
275
Conceived project: R.P., H.S. Designed experiments: X.H., Y.S., H.W., K.M.O., H.D.V., E.G., D.A.,
R.P., H.S. Provided technical advice and resources for microfluidics: N.M., R.P., H.W., D.A.
Provided biological advice: R.C., Y.S., S.H., Y.P. Created new reagents: S.H., Y.P, M.W.M.
Performed experiments: X.H., Y.S., H.W., K.M.O., D.P., H.D.V., R.P. Tracked nuclei in GCaMP
imaging experiments: X.H., Y.S., S.X. Wrote paper with input from all authors: X.H., Y.S., H.W.,
D.A., R.P., H.S. Supervised research: Q.D., Y.P., E.G., D.A., R.P., H.S. Directed research: H.S.
282

283 **Acknowledgements**
284
This research was funded in part by the National Institute of Biomedical Imaging and
Bioengineering, the National Institute of Neurological Disorders and Stroke, and the Center for
Cancer Research of the National Cancer Institute within the National Institutes of Health (Z01
BC 006161), and the National Science Foundation (CBET 1605679). We thank George Patterson
for the use of his cell culture facilities, Leighton Duncan and Daniel Colón-Ramos for kindly
providing strains and for conducting initial pilot experiments and their careful read, Evan Ardiel
and Eviatar Yemini for providing helpful feedback on pan-nuclear GCaMP recordings, Ron Zohar
and Ehud Shchori for providing useful information on BIO-133, and Hank Eden for providing
helpful feedback on the manuscript. K.M.O. and E.G. were supported by NINDS Z01-NS003013
to E.G. K.M.O. was also jointly supported by AFOSR grant number FA9550-16-1-0052 to W.
Losert at UMD College Park. M.W.M was supported by NIH grant F32-NS098616.

Disclaimer: The NIH, its officers, and staff do not recommend or endorse any company,
product, or service.
299
300 **Methods**
301
302 **Sample preparation**
303
U2OS, wild type (WT) HCT-116, and HCT-116 TOP1-GFP cell culture
U2OS (ATCC, HTB-96), WT HCT-116 (ATCC, CCL-247), and HCT-116 TOP1-GFP (see below) cells
were cultured in DMEM (Lonza, 12-604F) media with 10% Fetal Bovine Serum at 37°C and 5%
CO_2.

To tag the genomic topoisomerase I (TOP1) in WT HCT-116 cells, sequence
CCTCACTTGCCCTCGTGCCT targeting a CRISPR site 77nt after the stop codon of TOP1 was cloned
into pX330. Homology arms (of ~1 kb) upstream and downstream of the target site were cloned
to flank a blasticidin‐resistance gene, where the upstream homology arm was modified to
replace the stop codon with a GFP domain connected to the protein-coding region of the last
exon of TOP1 via a short poly-lysine linker. Both constructs were co-transfected into WT-
HCT116 cell, followed by selection with 5 μg/mL of blasticidin 48 hours post transfection. GFP-
positive cells were further selected by FACS.

Transfection of cells
Cells were cultured to 50% confluency and transfected using xTreme Gene HP DNA Transfection
Reagent (Sigma, 6366236001). The transfection mixture contained 100 μL 1X PBS, 2 μL
Transfection Reagent, and 200-1000 ng plasmid DNA. Cells were imaged 24-48 hours after
transfection.

C. elegans samples
Nematode strains were grown on NGM plates seeded with OP50 bacteria. *C. elegans* imaged as
young adults were synchronized by picking L4 stage worms 24 hours prior to the experiment
and transferring them to seeded plates, and *C. elegans* imaged as larvae were directly picked
from plates. Strain DCR6268 (*olaEx3632[pttx-3b::SL2::Pleckstrin homology domain::GFP::unc-54
3'UTR + pelt-7::mCh::NLS::unc-54 3'UTR]*) was used for imaging axons and dendrites (**Fig. 3d**).
olaEx3632 was made by injecting plasmid DACR2285 (pttx-3b::SL2::Pleckstrin homology
domain::GFP::unc-54 3'UTR) at 25 ng/uL and DACR2436 (pelt-7::mCh::NLS::unc-54 3'UTR) at 10
ng/μL. Strain AML32[56] (*wtfIs5 [rab-3p::NLS::GCaMP6s + rab-3p::NLS::tagRFP]*) was used for pan-
nuclear neuronal calcium imaging (**Fig. 3e-j**); strain NZ1091(*kyIs587 [gpa-6p::GCaMP2.2b; unc-
122p::dsRed]; kyIs5662 [odr-7p::Chrimson:SL2:mCherry; elt-2p::mCherry][7]*) was used for
chemical (**Fig. 4d-g**) and optogenetic stimulation (**Fig. 4h, i**) . For optogenetic stimulation, L4
stage animals were transferred to agar plates seeded with 62.5 μM all trans-retinal (ATR, Sigma
Aldrich, R2500) for over 12 hours.

Drosophila samples
Drosophila stocks were obtained from the Bloomington *Drosophila* Stock Center: *neur-GAL4*
(BL6393) and *UAS-CD4-td-Tomato* (BL35837). White prepupae were selected and aged for 13-
14 h at 18°C followed by 1 h at 25°C (equivalent to 7.5–8 h at 25°C). These aged pupae were
dissected in fresh culture media (CM; Schneider's Drosophila media + 10% fetal bovine serum,
both from Life Technologies), and wing discs were isolated from the aged pupae and stored in
fresh CM prior to mounting with BIO-133.

Characterizing the optical properties of polymers

Characterization of aberrations by visualizing fluorescent beads under different polymer layers

100 nm diameter yellow-green fluorescent beads (Invitrogen, F8803, 1:10000 dilution in water)
were coated on #1.5 coverslips (24 mm x 50 mm, VWR, 48393-241) coated with 0.1% w/v poly-
L-lysine (Sigma, P8920-100ML). We placed spacers (Precision Brand, 44910) of variable
thickness on one of the coverslips and deposited droplets of BIO-133, 10% PEG-DA (ESIBIO,
GS705) or UV-curable PDMS (Shin-Etsu Chemical, KER-4690) on the beads. The droplet was
then covered with another coverslip coated with beads and compressed with an iron ring. BIO-
133 and PEG-DA hydrogels were crosslinked at 312 nm (Spectroline ENB-280C) for 2 minutes.
PDMS was crosslinked at 312 nm for 5 minutes and post-cured at room temperature for one
day. Once cured, we separated the two cover glasses and kept the one with polymer/hydrogel
on it. For imaging through FEP films of different thickness (CS Hyde, 23-1FEP-24, 25 μm; 23-
2FEP-24, 50 μm; 23-3FEP-24, 75 μm; 23-5FEP-24, 125 μm), we immersed beads in 1μL water
and covered the sample with the FEP film. All bead images were acquired with a symmetric
0.8/0.8 NA diSPIM[43].
364
365   *Measurement of refractive index of polymers*
366   A refractometer (American Optical) was used to measure the refractive index of pure water,
367   BIO-133 film (My Polymer, BIO-133, 25 μm), PDMS film (Shin-Etsu Chemical, KER-4690, 25 μm),
368   FEP film (CS Hyde, 23-1FEP-24, 25 μm), and 10% PEG-DA hydrogel (100 μm). The refractive
369   index for each material was measured 3 times and the average value reported in **Fig. 1**.
370
Measurement of BIO-133 autofluorescence
A 50 μm thick BIO-133 film was deposited on a glass bottom dish (MatTek, P35G-1.5-14-C).
Images were acquired both on the BIO-133 area and an area without BIO-133, using an instant
structured illumination microscope (iSIM[48]) with 40 ms exposure time and 45 mW 488 nm
excitation, 70 mW 561 nm excitation, or 90 mW 639 nm excitation (measured with a power
meter immediately prior to the objective). Care was taken to ensure the illumination was
focused within the BIO-133 film. The two images were subtracted to measure the
autofluorescence of BIO-133 relative to glass (**Supplementary Fig. 1**).
379
380   **Cell growth and imaging using BIO-133 substrates**
381
Fabrication of BIO-133 cell culture wells for diSPIM experiments
BIO-133-sided PDMS substrates with 2.5 mm diameter wells (**Supplementary Fig. 4**) were used
for live cell imaging experiments (**Fig. 2b-d**). To make the BIO-133 bottom, a BIO-133 droplet
was positioned on a #1.5 glass coverslip (24 mm x 50 mm, VWR 48393-241) between two 50
μm plastic spacers (Precision Brand, 44910), covered with another coverslip , compressed with
a glass slide (Ted Pella, 260386) and cured with a UV lamp (365nm, Spectroline ENB-280C) for
15 minutes. After curing, the BIO-133 film was peeled off and exposed to UV light for another 2
hours in 70% ethanol. To make the PDMS well, 15 mL PDMS (Dow Inc. Sylgard 184) was poured
into a 10 cm plastic dish (Kord-Valmark, 2910) and cured for 2 hours at 80 °C to obtain a 2 mm
thick PDMS slab. We punched 2.5 mm diameter matching holes on the PDMS slab and a piece
of double-sided tape (Adhesives Research, ARCare 90880) using a 2.5 mm diameter circular
punch (Acuderm Inc., P2550). The PDMS slab and the tape were then cut into smaller pieces (~5

mm on a side) with a razor blade (Sparco, 01485). BIO-133 membranes, PDMS chunks and
double-sided tape were further disinfected in 70% ethanol for 2 hours. After disinfection, the
BIO-133 membrane was adhered to PDMS using the adhesive tape, so that the matching holes
became wells for cell culture. After seeding and growing cells in wells, the assembly was flipped
over for diSPIM imaging.
399
Quantification of cell growth
Cured and leached 50 μm thick BIO-133 films were deposited on glass bottom dishes (MatTek,
P35G-1.5-14-C). Similar aliquots of U2OS (or HCT 116 TOP1-GFP) cells were seeded onto BIO-
133 films or on another glass bottom dish without BIO-133 (MatTek, P35G-1.5-14-C). Dishes
seeded with cells were maintained between imaging experiments in an incubator at 37°C, 5%
CO_2. On each dish, a small area was selected and imaged using a widefield microscope equipped
with a 10x/0.25 NA objective lens each day, for three days. Cells numbers were estimated from
images with the Cell Counter ImageJ plugin (https://imagej.nih.gov/ij/plugins/cell-
counter.html). Each experiment was repeated three times. Raw images were divided by
Gaussian-blurred versions of themselves (sigma = 5 pixels) to flat-field images prior to display
(**Fig. 2a, Supplementary Fig. 2**).
411
Live cell imaging through BIO-133 with diSPIM
U2OS cells were cultured and transfected with 100-200 ng of mEmerald-Tomm20 plasmid
(Addgene, 54281) directly on the BIO-133 bottomed well plate. The well plate was inverted and
immersed in live cell imaging solution (Invitrogen, A14291DJ). Cells were imaged with a
symmetric 0.8/0.8 NA diSPIM, through the BIO-133 layer. 50 volumes were acquired with 3 s
intervals between dual-view volumes. Dual-view data were jointly deconvolved with ImageJ
plugin DiSPIM Fusion[54], and were drift- (with ImageJ plugin Correct 3D Drift
(https://imagej.net/Correct_3D_Drift) and bleach- corrected (with ImageJ function Bleach
Correction (https://imagej.net/Bleach_Correction, exponential fitting method) prior to display.
421
Super-resolution imaging through BIO-133 with iSIM
U2OS or WT HCT-116 cells were cultured and transfected with mEmerald-Tomm20 or LAMP1-
EGFP (Taraska Lab, NHLBI) on a 50 μm thick BIO-133 film. The BIO-133 film was cured on a glass
bottom dish (MatTek, P35G-1.5-14-C). A 60X, NA = 1.2 water objective (Olympus, PSF grade),
correction collar adjusted to 0.17, was used to image the cells through the glass and BIO-133
film using our home built iSIM system[48] and 488 nm excitation. Volumes were acquired every 3
s for U2OS cells expressing mEmerald-Tomm20 and every 7 s for WT HCT-116 cells expressing
LAMP1-EGFP. We used an exposure time of 80 ms, and a z-step of 0.25 μm for U2OS cells and
0.5 μm for WT HCT-116 cells. Live HCT-116 TOP1-GFP cultured on a 50 μm thick BIO-133 film or
on a glass bottom dish were imaged to acquire volumes with a step size of 0.5 μm. Raw images
were deconvolved with the Richardson-Lucy algorithm for 20 iterations, destriped in Fourier
space to remove striping artifacts[66], and bleach corrected
(https://imagej.net/Bleach_Correction). A median filter with kernel size 0.5 pixel was applied to
denoise mEmerald-Tomm20 and GFP-LAMP1 images prior to display.
436
Immunolabeling and imaging of multilayered WT HCT-116 cells on BIO-133

WT HCT-116 cells were cultured on a 50 μm thick BIO-133 film on a glass bottom dish until a
thick layer was visible by eye. Cells were fixed with 4% paraformaldehyde (Electron Microscopy
Sciences) in 1X PBS for 30 minutes at room temperature (RT). Cells were rinsed 3 times in 1X
PBS and permeabilized with 0.1% Triton X-100/PBS (Sigma, 93443) for 15 min at RT.
Permeabilized cells were rinsed 3 times with 1X PBS and incubated in 1X PBS with primary
antibody Rabbit-α-Tomm20 (Abcam, ab186735) and Mouse-α-LaminA/C (Abcam, ab244577) at
a concentration of 1:100 for 1 hour at RT. After primary antibody staining, cells were washed in
1X PBS for 5 min, three times. Cells were stained in 1X PBS with secondary antibody Donkey-α-
Rabbit Alexa Fluor 488 (Jackson Immuno Research, 711-547-003), Donkey-α-Mouse JF549
(Novusbio, NBP1-75119JF549) and Alexa Fluor 647 Phalloidin (Thermofisher, A22287) at a
concentration of 1:100 for 1 hour at RT. Cells were washed in 0.1% Triton X-100/PBS for 5 min,
three times. In each spectral channel, 46 slices were acquired on iSIM with an exposure time of
100 ms and a z-step of 0.5 μm. Raw images were deconvolved with the Richardson-Lucy
algorithm for 20 iterations and destriped in Fourier space to remove striping artifacts[66]. The 633
nm channel (Alexa Fluor 647 Phalloidin) was bleach corrected
(https://imagej.net/Bleach_Correction) across the z stack to compensate for decreased signal
further into the stack.
455
Flow cytometry preparation
Sample handling channels, 1 mm wide and 70 μm high, were formed by pouring 20 mL PDMS
(Dow Corning, Sylgard 184) on a positive mold made of packaging tape (Duck Brand) cut to the
desired dimensions with a craft cutter (Silhouette Cameo) and stuck in a 10 cm Petri dish. A thin
PDMS membrane (~0.5 mm) was air plasma (Harrick Plasma, PDC-32G (115V)) bonded to the
channel surface. Holes at the endpoints of the channel were created by punching the PDMS
membrane with a 1 mm diameter circular punch (Acuderm Inc., P150) after plasma treatment.
A 400 μm wide (142 μm height) imaging channel was cut directly from double-sided silicon-
based adhesive tape (Adhesives Research, ARCare 90880) with a craft cutter and stuck to the
PDMS device, thus creating a connection between the two sample handling channels in the
lower layer. A thin BIO-133 membrane (50 μm) was placed on top of the tape to seal the
channel. Two thick PDMS pieces with holes (6 mm and 2 mm diameter) were cut and air plasma
bonded to the device to provide fluidic access. When imaging, cells were added to the 6 mm
diameter reservoir, and output tubing (Dow Corning, 508-004) was connected to the 2 mm
hole. Flow speed was adjusted by changing the height of the output tubing. See also
Supplementary Fig. 5.
472
Flow cytometry imaging of fixed, DAPI-stained U2OS cells
U2OS cells were fixed in 4% Paraformaldehyde/PBS and subsequently stained with DAPI in 0.1%
Triton X-100/PBS (Sigma, 93443). The flow device was mounted on a 10 cm petri dish for
imaging with diSPIM. Fixed cells were added to the input port, producing steady flow through
the channel after several minutes. 1000 frames of the same image plane were acquired with
diSPIM at 50 frames per second under 'fixed sheet mode'.
479
480   **Live animal/tissue imaging through BIO-133 with diSPIM**
481

Live imaging of BIO-133 embedded Drosophila wings

A 13 μm thick BIO-133 film was created and cut into two rectangular pieces (4.5 mm x 10 mm) and a square piece (10 mm x 10 mm). The rectangular pieces were deposited on a 10 cm petri dish to form a 1 mm wide open-top channel. Early pupal fly wings were deposited into the channel (convex side up) with 20-40 μL culture media, the square BIO-133 piece placed on top to close the channel and additional culture media was carefully added to the dish (**Fig. 3a, Supplementary Fig. 8)**. Single-view diSPIM imaging was then performed. 360 volumes were acquired with 5 s inter-volume spacing, over 30 minutes. Volumes were deconvolved with MATLAB and bleach corrected with ImageJ (https://imagej.net/Bleach_Correction, exponential fitting method).

Fabrication of microfluidics for C. elegans immobilization

Standard soft lithography techniques[67] were used to fabricate an SU-8 (Kayaku Advanced Materials, formerly Microchem Corp.) master mold for sets of four microfluidic funnels for worm confinement as described[53]. To fabricate devices in BIO-133 (MY Polymers Ltd.), we placed two spacers (100 μm, Precision Brand 44910) beside the pattern, poured polymer onto the mold, covered the mold with a glass slide and cured the polymer under a UV lamp (365nm, Spectroline ENB-280C) for 2 minutes. After curing, we peeled the BIO-133 off the mold, punched inlet and outlet holes with a 1 mm diameter circular punch (Acuderm Inc., P150), and sealed the device to a #1.5 cover glass (24 mm x 50 mm, VWR 48393-241) with double-sided silicone-based adhesive tape (Adhesives Research, ARCare 90880). We cut out an aperture from a 10 cm petri dish and used UV-curable optical cement (Norland Products Inc., Norland Optical Adhesive NOA 68) to secure the coverslip carrying the microfluidic device over the aperture in the petri dish. Inlet and outlet tubing (Dow Corning, 508-004) was clamped to the assembly using a pair of hollow magnets (K&J Magnetics, R211-N52) placed above and below the coverslip, as described[68]. Optical cement was again used to secure tubing to the magnets. See also **Supplementary Fig. 10**.

Live imaging of C. elegans through BIO-133 chambers

To load worms into the immobilization device, we added a drop of M9 buffer containing worms to the inlet and created vacuum at the outlet using a syringe. Within several minutes (for a 4-channel chip), worms were observed to align in the channels. The petri dish was then filled with water and worms were imaged with symmetric 0.8/0.8 NA diSPIM. For structural imaging, we added 0.25 mM levamisole to the buffer to stop residual worm motion. For calcium imaging, the outlet was connected to a peristaltic pump (Dolomite Microfluidics, 3200243) which provided negative pressure to immobilize worms without using anesthetics. We simultaneously imaged nuclei structure (TagRFP) and the nuclear-localized calcium response (GCaMP) with 488 nm and 561 nm excitation (Coherent) and image splitting devices on the detection side (Hamamatsu W-VIEW GEMINI), using a previously described fiber-coupled diSPIM system[46]. Dual-view stacks were acquired every 0.8 s over 500 time points. Dual-color, dual-view images were deconvolved and registered with ImageJ plugin DiSPIM Fusion[54].

Droplet-based immobilization of nematodes prior to imaging, and recovery after imaging

C. elegans were directly transferred from agar plates into a drop (~10 μL) of BIO-133 or 10%
PEG-DA (ESIBIO GS705) on a #1.5 cover glass (24 mm x 50 mm, VWR, 48393-241). The droplet
was positioned between two 100 μm spacers (Precision Brand, 44910), and was compressed by
a glass slide followed by 2-minute polymerization under a UV lamp (365nm, Spectroline ENB-
280C). After polymerization, worms were immobilized in the resulting gel disk. The gel disk was
then placed in a 10 cm petri dish or a standard chamber for diSPIM imaging. Single-view stacks
were acquired every 0.25 s for 250 time points. After imaging, worms could be released and
checked for viability by gently breaking the droplet with forceps. In some experiments, we
immersed the disks in M9 buffer for up to 12 hours, finding that live worms could also be
recovered after this period.
535
Tracking nuclei, calcium imaging analysis
TagRFP volumes were imported into Imaris and neurons tracked with Imaris for Tracking
(https://imaris.oxinst.com/products/imaris-for-tracking) to obtain the center of each neuron at
every timepoint. A custom MATLAB script was used to extract the calcium signal. For every
neuron, the average intensity of TagRFP channel I_{561} and the intensity of GCaMP channel I_{488}
were computed by averaging pixels within a 2 μm (adult) or 1.5 μm (larval) diameter sphere
placed around each center position. I_{561} and I_{488} were calculated from dual-view deconvolved
images (Fig. 3 f, g) or single-view raw data (Fig. 3 i, j). The ratio $R=I_{488}/I_{561}$ was used to minimize
non-GCaMP fluctuations. Neuronal activity for the datasets in **Fig. 3** was reported as $dR/R = (R -
R_0)/R_0$, where R_0 is the baseline for an individual neuron defined as its lower 20th percentile
intensity value.
547
548 **Chemical and optical perturbations in BIO-133 based imaging devices**
549
Fabrication of chemical delivery devices for cells
A modified PDMS well plate design (**Supplementary Fig. 11**) was applied to deliver chemical
perturbations to cells (**Fig. 4a-c**). A 400 μm thick BIO-133 film was created using the method
described above. 10 mL PDMS was cured in a 10 cm dish, PDMS tubing (Dow Corning, 508-004)
was placed on the cured PDMS layer at 8 mm intervals, and another 15 mL PDMS was added to
obtain a ~4 mm thick PDMS slab with channels contained inside. Holes crossing the channels
were punched at 8 mm intervals using a 5 mm diameter circular punch (Acuderm Inc., P550). A
piece of double-sided silicone-based adhesive tape (Adhesives Research, ARCare 90880) was
also punched at 8 mm intervals using a 2.5 mm diameter circular punch (Acuderm Inc., P2550).
The PDMS slab and the tape were cut into two-well pieces. BIO-133 film, PDMS chunks and
double-sided tape were disinfected in 75% ethanol. After disinfection, the BIO-133 membrane
was adhered to PDMS via the double-sided adhesive tape, so that the matching holes became
wells for cell culture. After growing cells, tubing (Scientific Commodities Inc., BB31695-PE/2)
was inserted into both sides of the channel for introducing chemical flow and another piece of
double-sided tape without holes was used to seal the wells. The assembly was flipped over for
diSPIM imaging through the BIO-133 membrane.
566
Mitochondrial imaging in the presence of CCCP

U2OS cells were cultured in two 5 mm diameter BIO-133 bottomed wells with ports and
transfected with 300-400 ng of mEmerald-Tomm20. Before imaging, the wells were filled with
live cell imaging solution (Invitrogen, A14291DJ), flipped over and attached to a 10 cm petri dish
with double-sided silicone-based adhesive tape (Adhesives Research, ARCare 90880). For the
well containing control cells, tubing was left disconnected from a source. For the well
containing cells that experienced chemical perturbation, input tubing was connected to a
syringe containing 0.05 mM carbonyl cyanide m-chlorophenyl hydrazine (CCCP, Sigma, C2759).
The syringe is higher than the output tube so that drug flow was induced by gravity. We used a
valve (McMaster-Carr, 7033T21) placed between the input tube and syringe to control the flow.
The valve was closed prior to imaging. For each well, two cells were chosen for imaging. A
multi-position acquisition was set in the Micro-manager[69] diSPIM plugin[70] to sequentially image
the four cells. Volume acquisition time was 3 s, and 90 volumes were acquired for each cell with
60 s intervals between volumes. 10 minutes after the imaging started, the valve was opened
and drug flow was induced in ~60 s. Dual-view images were deconvolved with ImageJ plugin
DiSPIM Fusion[54], drift corrected (ImageJ plugin Correct 3D Drift,
https://imagej.net/Correct_3D_Drift) and bleach corrected (ImageJ function Bleach Correction,
https://imagej.net/Bleach_Correction, exponential fitting method).
585
Encapsulation of C. elegans into PEG hydrogels
C. elegans were encapsulated into PEG hydrogel disks as described in our prior work[7]. PEG
hydrogel precursor solutions were prepared by combining 20% w/v poly(ethylene glycol)
diacrylate (PEG-DA, 3350 MW, 94.45% acrylation, ESI BIO) with 0.10% w/v Irgacure 2959
photoinitiator (2-hydroxy-4'-(2-hydroxyethoxy)-2-methylpropiophenone, I2959, BASF) in 1x S-
basal buffer (100 mM NaCl, 50 mM KPO_4 buffer, pH 6.0). Clean 24 mm x 50 mm glass coverslips
(VWR) were rendered permanently hydrophobic by exposure to vapors of (tridecafluoro-
1,1,2,2-tetrahydrooctyl) trichlorosilane (Gelest). For covalent attachment of the PEG hydrogel
to glass, coverslips (Thermo Scientific) were silanized by coating with 3-(trimethoxysilyl)propyl
methacrylate (TMSPMA, Sigma-Aldrich). Both methods of surface modification were applied to
1" x 3" glass slides (VWR). A small volume (1.75 µL) of PEG hydrogel solution with photoinitiator
was pipetted onto a hydrophobic glass slide flanked by two PDMS spacers whose thickness
matched the desired hydrogel thickness of 150 µm. Animals were transferred into the hydrogel
solution by worm pick. A coverslip, TMSPMA treated for making mounted PEG-DA gels, or
untreated for making freestanding gels, was placed over the hydrogel droplet and supported by
the PDMS spacers. The glass slide/coverslip sandwich was then placed over a UV light source
(312 nm, International Biotechnologies, Inc, model UVH, 12W) and illuminated for two minutes
until gelation. Hydrogel disks were immediately transferred to wet agar dishes to keep
embedded animals hydrated.
605
Fabrication of microfluidic devices for chemical stimulation of C. elegans
Microfluidic chambers were prepared using poly(dimethyl siloxane) (PDMS; Sylgard 184, Dow
Corning) in a ratio of 1:10 and poured to a depth of 5 mm on a silicon master positive mold of
the microchannels used previously[71]. Once cut free of the master, devices were punched so
that two balanced-length inlets and the outlet had 1.5 mm holes going through the thickness of
the material, and 1 mm holes punched from the side to allow flexible tubing to be inserted from

the sides. The smooth PDMS device surface opposite the microchannels was irreversibly
bonded to a glass slide using oxygen plasma (Harrick PDC-32G, 18W, 45 seconds). A thin PDMS
membrane (150 μm) was cut with a 3.5 mm diameter dermal punch and then oxygen plasma
bonded to the microfluidic channel surface with the hole in the membrane exposing the
micropost array. The hole in the thin PDMS membrane formed a "well" that hydrogel disks
could be gently placed in with forceps. A thin BIO-133 membrane (75-80 μm) was prepared by
gelation of BIO-133 liquid polymer between two glass slides rendered permanently
hydrophobic as described above. Two layers of clear cellophane tape (height of ~80 μm)
formed the standoffs that determined final membrane height. After degassing the microfluidic
device in a vacuum chamber for approximately 45 minutes, the device was removed from the
desiccator, connected to tubing, and flushed with S. basal buffer before use to remove any air
bubbles. Hydrogel disks could be interchanged between the well formed by the PDMS above
the micropost array easily using forceps and then the system sealed for microfluidic flow using
the thin BIO-133 membrane described above. See also **Supplementary Fig. 12**.
626
627   *Preparation of Chemosensory Stimulus*
628   For both wide-field and diSPIM assays, diacetyl (2,3-butanedione, Sigma) was diluted to 1.1 μM
629   in 1x S. basal (10^{-7} dilution). 1 μL of 1 mg/mL fluorescein solution was added to 40 mL of
630   diacetyl solution to visualize stimulus delivery.
631
Wide-field Imaging with Chemical Stimulation
For wide-field, single plane imaging of multiple *C. elegans* at once[57], the microfluidic chamber,
valves, tubing and reservoirs were prepared as above and placed on a Zeiss AxioObserver
epifluorescence microscope with a 5x, 0.25 NA objective, EGFP filter set, and Hamamatsu Orca-
Flash 4 sCMOS camera. Micromanager scripts[72] automatically synchronized capture of ten 30-s
trials recording at 10 fps with 10 ms excitation pulses and 10 s chemical stimulation.
NeuroTracker software[71] analyzed the wide-field neural imaging data, from which background-
corrected fluorescence changes were calculated in MATLAB as $\Delta F/F_0$, where F_0 is baseline
neural fluorescence during the four seconds prior to stimulation. Data for multiple individual
animals were also presented as a population mean to show the relative decrease in average
calcium response after multiple stimulation periods.
643
DiSPIM Imaging with Optical and Chemical Stimulation
Stimulus control for optical illumination or chemical pulses was integrated with diSPIM
volumetric imaging using a custom Micromanager script controlling an Arduino Uno and
enabling independent digital switching of 6 TTL channels at the beginning of specified image
acquisition timepoints. One TTL channel controlled the intensity of a red LED (617 nm, 3W,
Mightex) connected to the bottom port of the diSPIM and illuminated the sample through a
Nikon 4x, 0.1 NA lower objective. A second TTL channel controlled a 12V fluidic valve system for
chemical stimulation (ValveLink 8.2, Automate). Pinch valves allowed flow of either buffer or
chemical stimulus lines into the microfluidic channel network, flowing to a common outlet.
653
For optogenetic stimulation experiments, animals were embedded in BIO-133 disks bonded to a
cover glass placed in the diSPIM sample chamber. To embed animals, they were first transiently

immobilized by being picked onto seeded (OP50 *E. coli*) plates with 1 mM tetramisole, and
allowed to rest for 1.5 hours. Subsequently, worms were picked into a droplet of BIO-133
polymer liquid and gelled in the same manner as the PEG-DA hydrogel disks above, using a
TMSPMA silanized coverslip for covalent bonding.
660
For chemical stimulation experiments, animals were embedded in PEG hydrogel disks. Animals
can be maintained in these disks for many hours if they are kept hydrated[7]. Just prior to an
experiment, an animal-embedded disk was inserted into the sample cavity of the diSPIM
microfluidic chamber. A 75-80 μm thick BIO-133 membrane was sealed to the microfluidic
device surface, closing the fluidic channel with the PEG disk and animals contained within. The
hydrogel disk was inserted into a droplet of S. basal buffer present in the well to avoid the
introduction of bubbles that would disrupt microfluidic flow. To assure continuous microfluidic
flow through the chamber without leaking, we balanced inlet and outlet flows by adjusting the
reservoir heights. Specifically, inlet reservoir heights were held slightly above the stage (Δh_{in}),
and the outlet reservoir level was placed further below the stage ($\Delta h_{out} > \Delta h_{in}$) to ensure a slight
negative pressure in the chamber. Microfluidic stimulus switching was achieved using a dual
pinch valve (NResearch Inc., 161P091), that alternately allows either a buffer or stimulus line to
flow through the microfluidic chamber to the single outflow line.
674
A typical diSPIM acquisition captured one volume per second (10 ms exposure, minimum slice
time setting, 55 slices per volume, 166.4 x 166.4 x 82.5 μm total volume space) for 10 minutes,
with 20-s duration stimulation every minute. Z-Projection time series videos were produced in
ImageJ from cropped versions of the total number of images, then analyzed for GCaMP neural
fluorescence using rectangular boxes for integrated fluorescence density, with a nearby region
void of signal used for background subtraction.
681
682

684
685
Fig. 1, **Diffraction-limited imaging is possible when imaging through BIO-133, unlike other**
polymers. a) Imaging geometry. A light sheet is used to illuminate the 100 nm yellow-green
bead sample, which is embedded under progressively thicker polymer. Illumination and
detection occur through 0.8 NA water-dipping objectives. **b)** Full width at half maximum
(FWHM) in the 'x' direction under different thicknesses of polymer. Means and standard
deviations are shown. **c)** Exemplary lateral (top row) and axial (bottom row) images of beads
imaged through 75 μm of polymer, demonstrating that BIO-133 provides diffraction-limited
performance whilst the other polymers do not. Single images, rather than maximum intensity
projections, are shown. The refractive index of each polymer as measured with a refractometer
is also indicated (average value from 3 independent trials). Scale bar: 1 μm. See also
Supplementary Table 1.
697

Fig. 2, BIO-133 provides an inert and biocompatible scaffold on which to grow and image
cells. **a)** U2OS growth on BIO-133 is similar to growth on glass coverslips. *Top*: quantifying cell
growth on 50 μm thick BIO-133 layer vs. glass coverslip. Means and standard deviations from 3
fields of view (10x magnification, ~800 μm x 800 μm field of view) are shown over 3 days.
Bottom: example fields of view from day 1 and day 3, cells on BIO-133 layer. Scale bar: 200 μm.
See also **Supplementary Fig. 2, 3**. **b)** Schematic of diSPIM imaging geometry. 50 μm film with
adherent cells is inverted and imaged in the diSPIM setup. See also **Supplementary Fig. 4**. **c)**
Example maximum intensity projections of deconvolved images of U2OS cells expressing
mEmerald-Tomm20 in lateral (top) and axial (bottom) views. Scale bar: 10 μm. **d)**. Higher
magnification views of the red and yellow rectangles in **c)**, highlighting examples of
mitochondrial fusion (top, red arrowhead) and fission (bottom, yellow arrowhead). 50 volumes
were taken with a 3 s inter-volume interval. See also **Supplementary Video 1**. Note that primed
coordinates refer to the plane of the BIO-133 layer (x', y) and the direction normal to the BIO-
133 layer (z'). Scale bar: 2 μm. **e)** iSIM imaging geometry. Cells were cultured on 50 μm BIO-133
film, and the film placed in a glass-bottom dish and immersed in cell culture medium. Imaging
was performed with a 60x, NA 1.2 water-immersion lens. **f)** Example deconvolved iSIM
maximum intensity projection showing live U2OS cells expressing mEmerald-Tomm20. Scale
bar: 10 μm. **g)** Higher magnification view of orange rectangular region in **f)**. Orange arrowhead
marks the same mitochondrion. 25 volumes were acquired with a 3 s inter-volume interval. See
also **Supplementary Videos, 3, 4**. Scale bar: 2 μm. A 0.5 pixel median filter was used to denoise
images in **f, g)** prior to display. **h)** Multiple layers of HCT-116 cells were grown on 50 μm BIO-
133 layer and immunostained against Tomm20 (green), lamin A/C (red), and actin (blue). See
also **Supplementary Video 5**. Scale bar: 10 μm. **i)** Maximum intensity projection over indicated
axial range (measured from the bottom of the cell layer) for cyan rectangular region in **h)**. Scale
bar: 5 μm.

Fig. 3 Live imaging of BIO-133 encapsulated fly wings and *C. elegans*. a) Experimental
schematic. Top: A thin layer of BIO-133 membrane covers excised fly wings, immobilizing them
so that axon dynamics can be recorded at high resolution over an extended period. Middle: A
simple microfluidic device is used to trap adult worms for structural and functional imaging of
the nervous system. Bottom: Worms can also be encapsulated in a gelled droplet of BIO-133.
See also **Supplementary Figs. 8, 10. b)** Deconvolved, single-view maximum intensity projection
of a fly wing with TdTomato-labelled CD4 showing the axons of two neurons (upper: TSM1;
lower: L3) shortly before fasciculation in the developing *Drosophila melanogaster* wing disc.
Scale bar: 10 μm. 360 volumes were taken with 5 s inter-volume intervals. (30 min in total, see
also **Supplementary Video 6**). **c)** Magnified regions of TSM1 and L3 axons, corresponding to
yellow and blue rectangles in **b)**, highlighting morphological changes and apparent motion of
CD4 puncta. Scale bars: 4 μm. **d)** Isotropic, high-resolution imaging of GFP-labeled axons and
dendrites in anesthetized adult *C. elegans*, as shown by orthogonal, jointly deconvolved diSPIM
maximum intensity projections. Cell bodies are circled and axons entering the nerve ring region

are overlaid with dotted lines. Scale bar: 10 μm. **e)** Calcium imaging of adult worm (red channel:
TagRFP, green channel: GCaMP6s; both labels targeted to nuclei), 2 views imaged at 1.25 Hz
volumetric rate. Joint deconvolution diSPIM results are shown; red and green channels were
simultaneously collected and colors are overlaid in display. Scalebar: 10 μm. See also
Supplementary Video 7. f) dR/R traces for all 126 tracked nuclei. **g)** dR/R traces for selected
individual neurons. Note correspondence with numbered neurons and marked neurons in **e). h)**
Calcium imaging of larval worm with higher temporal resolution (4 Hz volumetric rate), single-
view results are shown. GCaMP channel and associated segmented dR/R signal are indicated for
3 successive time points. Scale bar: 5 μm. See also **Supplementary Video 8. i)** dR/R traces for all
110 nuclei segmented and tracked in **h). j)** traces for selected individual neurons. Note
correspondence with numbered neurons and marked neurons in **h)**.
751
752

Fig. 4, BIO-133 is compatible with chemical and optical perturbations. a) Experimental
schematic for perturbations. Cells grown on BIO-133 were placed on PDMS wells and were
either perturbed by flowing 0.05 mM CCCP or left as controls (top higher magnification view).
Alternatively, worms were embedded in PEG-DA bonded to a PDMS flow chip and imaged
through a layer of BIO-133 (middle higher magnification view) to examine response to chemical
stimulation; or embedded in BIO-133, repetitively stimulated with red light from lower, 4x
objective and imaged using upper diSPIM objectives (bottom higher magnification view). See
also **Supplementary Figs. 11-13. b)** Example cells with (left column) and without (right column)
CCCP treatment at early (top, 0 min) and late (bottom, 70 min) time points. CCCP was added at
10 minutes. Maximum intensity projections of deconvolved diSPIM data are shown. Scale bar:
10 μm. See also **Supplementary Video 9. c)** Higher magnification views of yellow (left column)
and blue (right column) regions in **b)**. Yellow arrowhead shows CCCP-induced morphological

change of mitochondrion. Scale bar: 5 μm. **d)** Example images of worms expressing GCaMP
immobilized in PEG-DA disk with (bottom) and without (top) 1.1 μM diacetyl. Fluorescein added
to stimulus highlights the rapid addition/removal of chemical. Scale bar: 500 μm. Right
schematics show layered structure of assembly, including direction of flow and diffusion
(arrowheads) into PEG-DA layer. Line plots show intensity of fluorescein over time in channel
(red) and PEG-DA (blue) layers. **e)** Top: dF/F heatmaps derived from widefield microscopy
measurements from 15 animals (rows) in response to 10 repeated stimulus pulses (once per
minute). Bottom: responses averaged over all animals show neural adaptation. **f)** Single-view
diSPIM images recorded from a single animal, showing subcellular response in AWA neuron to
1.1 μM diacetyl compared to control (buffer flow) conditions. Contrast has been adjusted to
better highlight the response from different cell regions. Scale bar: 10 μm. See also
Supplementary Video 10. **g)** Graphs show average intensity from boxed regions in **f)**
highlighting fluorescence intensity changes in soma and dendrite. **h)** Worms expressing
Chrimson and GCaMP are repetitively stimulated with red light and imaged using upper diSPIM
objectives. Maximum intensity projection of GCaMP fluorescence from single-view diSPIM
recordings are shown before (left) and after (right) optogenetic stimulation. Scale bar: 10 μm.
See also **Supplementary Video 11**. **i)** dF/F traces for dendrite, axon, and soma, corresponding to
boxed regions in **h)**.
784

References

Chalfie, M., Tu, Y., Euskirchen, G., Ward, W. W. & Prasher, D. C. Green fluorescent protein as a marker for gene expression. *Science* **263**, 802-805 (1994).

Ou, G., Stuurman, N., D'Ambrosio, M. & Vale, R. D. Polarized Myosin Produces Unequal-Size Daughters During Asymmetric Cell Division. *Science* **330**, 677-680 (2010).

Kato, S. *et al.* Global brain dynamics embed the motor command sequence of Caenorhabditis elegans. *Cell* **163**, 656-669 (2015).

Chokshi, T. V., Ben-Yakar, A. & Chronis, N. CO_2 and compressive immobilization of C. elegans on-chip. *Lab Chip* **9**, 151-157 (2009).

Snow, J. J. *et al.* Two anterograde intraflagellar transport motors cooperate to build sensory cilia on *C. elegans* neurons. *Nature Cell Biology* **6**, 1109-1113 (2004).

Kim, E., Sun, L., Gabel, C. V. & Fang-Yen, C. Long-Term Imaging of Caenorhabditis elegans Using Nanoparticle-Mediated Immobilization. *PLoS One*, e53419 (2013).

Burnett, K., Edsinger, E. & Albrecht, D. R. Rapid and gentle hydrogel encapsulation of living organisms enables long-term microscopy over multiple hours. *Communications biology* **1**, 73 (2018).

Rohde, C. B., Zeng, F., Gonzalez-Rubio, R., Angel, M. & Yanik, M. F. Microfluidic system for on-chip high-throughput whole-animal sorting and screening at subcellular resolution. *Proc Natl Acad Sci U S A* **104**, 13891-13895 (2007).

Cornaglia, M., Lehnert, T. & Gijs, M. A. M. Microfluidic systems for high-throughput and high-content screening using the nematode Caenorhabditis elegans. *Lab Chip* **17**, 3736-3759 (2017).

Berger, S. *et al.* Long-term *C. elegans* immobilization enables high resolution developmental studies in vivo. *Lab Chip* **18**, 1359-1368 (2018).

Mondal, S. *et al.* Large-scale microfluidics providing high-resolution and high-throughput screening of Caenorhabditis elegans poly-glutamine aggregation model. *Nat Commun.* **7**, 13023 (2016).

Levario, T. J., Zhan, M., Lim, B., Shvartsman, S. Y. & Lu, H. Microfluidic trap array for massively parallel imaging of Drosophila embryos. *Nat Protoc.* **8**, 721-736 (2013).

Yanik, M. F., Rohde, C. B. & Pardo-Martin, C. Technologies for micromanipulating, imaging, and phenotyping small invertebrates and vertebrates. *Annu Rev Biomed Eng* **13**, 185-217 (2011).

Shorr, A. Z., Sönmez, U. M., Minden, J. S. & LeDuc, P. R. High-throughput mechanotransduction in Drosophila embryos with mesofluidics *Lab Chip* **19**, 1141-1152 (2019).

Stavrakis, S., Holzner, G., Choo, J. & DeMello, A. High-throughput microfluidic imaging flow cytometry. *Current Opinion in Biotechnology* **55**, 36-43 (2018).

Probst, R., Cummins, Z., Ropp, C., Waks, E. & Shapiro, B. Flow Control of Small Objects On-Chip: Manipulating Live Cells, Quantum Dots, and Nano-Wires. *IEEE Control Systems Magazine* **32**, 26-53 (2012).

Riba, J., Schoendube, J., Zimmermann, S., Koltay, P. & Zengerle, R. Single-cell dispensing and 'real-time' cell classification using convolutional neural networks for higher efficiency in single-cell cloning. *Scientific Reports* **10**, 1193 (2020).

18   Anagnostidis, V. *et al.* Deep learning guided image-based droplet sorting for on-demand
selection and analysis of single cells and 3D cell cultures *Lab Chip* **20**, 889-900 (2020).
19   Lam, J. *et al.* Adaptation of a Simple Microfluidic Platform for High-Dimensional
Quantitative Morphological Analysis of Human Mesenchymal Stromal Cells on
Polystyrene-Based Substrates. *SLAS Technol.* **22**, 646-661 (2017).
20   Doolin, M. T. & Stroka, K. M. Integration of Mesenchymal Stem Cells into a Novel
Micropillar Confinement Assay. *Tissue Engineering Part C: Methods* **25**, 662-676 (2019).
21   Albrecht, D. R. *et al.* Microfluidics-integrated time-lapse imaging for analysis of cellular
dynamics. *Integrative Biology* **2**, 278-287 (2010).
22   Reilly, D. K., Lawler, D. E., Albrecht, D. R. & Srinivasan, J. Using an Adapted Microfluidic
Olfactory Chip for the Imaging of Neuronal Activity in Response to Pheromones in Male
C. Elegans Head Neurons. *JoVE* **127**, e56026 (2017).
23   Yemini, E. *et al.* NeuroPAL: A Neuronal Polychromatic Atlas of Landmarks for Whole-
Brain Imaging in *C. elegans*. *bioRxiv*, https://doi.org/10.1101/676312 (2019).
24   Cáceres, I. d.-C., Valmas, N., Hilliard, M. A. & Lu, H. Laterally Orienting *C. elegans* Using
Geometry at Microscale for High-Throughput Visual Screens in Neurodegeneration and
Neuronal Development Studies. *PLoS One* **7**, e35037 (2012).
25   Albrecht, D. R. & Bargmann, C. I. High-content behavioral analysis of Caenorhabditis
elegans in precise spatiotemporal chemical environments. *Nat Methods* **8**, 599-605
(2011).
26   Belfer, S. J. *et al.* Caenorhabditis-in-Drop Array for Monitoring *C. elegans* Quiescent
Behavior. *Sleep* **36**, 689-698 (2013).
27   Scholz, M., Lynch, D. J., Lee, K. S., Levine, E. & Biron, D. A scalable method for
automatically measuring pharyngeal pumping in *C. elegans*. *Journal of Neuroscience
Methods* **274**, 172-178 (2016).
28   Atakan, H. B., Cornaglia, M., Mouchiroud, L., Auwerx, J. & Gijs, M. A. M. Automated
high-content phenotyping from the first larval stage till the onset of adulthood of the
nematode Caenorhabditis elegans. *Lab Chip* **19**, 120-135 (2018).
29   Hulme, S. E. *et al.* Lifespan-on-a-chip: microfluidic chambers for performing lifelong
observations of C. elegans. *Lab Chip* **10**, 589-597 (2010).
30   Hwang, H., Krajniak, J., Matsunaga, Y., Benian, G. M. & Lu, H. On-demand optical
immobilization of *Caenorhabditis elegans* for high-resolution imaging and
microinjection. *Lab Chip* **14**, 3498-3501 (2014).
31   Krajniak, J. & Lu, H. Long-term high-resolution imaging and culture of *C. elegans* in chip-
gel hybrid microfluidic device for developmental studies. *Lab Chip* **10**, 1862-1868 (2010).
32   S., N., Stowers, R., Lou, J., Xia, Y. & Chaudhuri, O. Varying PEG density to control stress
relaxation in alginate-PEG hydrogels for 3D cell culture studies *Biomaterials* **200**, 15-24
(2019).
33   Zhang, Z. F., Ma, X., Wang, H. & Ye, F. Influence of polymerization conditions on the
refractive index of poly(ethylene glycol) diacrylate (PEGDA) hydrogels. *Applied Physics A*
124, 283 (2018).
34   Wevers, N. R. *et al.* High-throughput compound evaluation on 3D networks of neurons
and glia in a microfluidic platform. *Scientific Reports* **6**, 38856 (2016).

35 Pittman, W. E., Sinha, D. B., Zhang, W. B., Kinser, H. E. & Pincus, Z. A simple culture
system for long-term imaging of individual *C. elegans*. *Lab Chip* **17**, 3909-3920 (2017).
36 Kim, D. N. H., Kim, K. T., Kim, C., Teitell, M. A. & Zangle, T. A. Soft lithography fabrication
of index-matched microfluidic devices for reducing artifacts in fluorescence and
quantitative phase imaging. *Microfluidics and Nanofluidics* **22** (2017).
37 Polanco, E. R., Western, N. & Zangle, T. A. Fabrication of Refractive-index-matched
Devices for Biomedical Microfluidics. *J Vis Exp* **139**, 58296 (2018).
38 Levario, T. J., Insley, P., Hwang, H., Shaham, S. & Lu, H. in *18th International Conference*
on Miniaturized Systems for Chemistry and Life Sciences, MicroTAS 2014   727-729 (San
Antonio, TX, 2014).
39 Xu, T. *et al.* Modified inverted selective plane illumination microscopy for
submicrometer imaging resolution in polydimethylsiloxane soft lithography devices. *Lab*
Chip **Advance Online Publication**, doi:https://doi.org/10.1039/D0LC00598C (2020).
40 Chang, T.-Y., Pardo-Martin, C., Allalou, A., Wählby, C. & Yanik, M. F. Fully automated
cellular-resolution vertebrate screening platform with parallel animal processing. *Lab*
Chip **12**, 711-716 (2012).
41 Tonin, M., Descharmes, N. & Houdré, R. Hybrid PDMS/glass microfluidics for high
resolution imaging and application to sub-wavelength particle trapping. *Lab Chip* **16**,
465-470 (2016).
42 Turaga, D. & Holy, T. E. Miniaturization adn defocus correction for objective-coupled
planar illumination microscopy. *Optics Letters* **33**, 2302-2304 (2008).
43 Wu, Y. *et al.* Spatially isotropic four-dimensional imaging with dual-view plane
illumination microscopy. *Nat Biotechnol.* **31**, 1032-1038 (2013).
44 Chen, B. C. *et al.* Lattice light-sheet microscopy: imaging molecules to embryos at high
spatiotemporal resolution. *Science* **346**, 1257998 (2014).
45 Hedde, P. N., Malacrida, L., Ahrar, S., Siryaporn, A. & Gratton, E. sideSPIM - selective
plane illumination based on a conventional inverted microscope. *Biomedical Optics*
Express **8**, 3918-3937 (2017).
46 Kumar, A. *et al.* Dual-view plane illumination microscopy for rapid and spatially isotropic
imaging. *Nature Protocols* **9**, 2555-2573 (2014).
47 Kaufmann, A., Mickoleit, M., Weber, M. & Huisken, J. Multilayer mounting enables long-
term imaging of zebrafish development in a light sheet microscope. *Development* **139**,
3242-3247 (2012).
48 York, A. G. *et al.* Instant super-resolution imaging in live cells and embryos via analog
image processing. *Nat Methods* **10**, 1122-1126 (2013).
49 Murray, M. A., Schubiger, M. & Palka, J. Neuron Differentiation and Axon Growth in the
Developing Wing of *Drosophila melanogaster*. *Developmental Biology* **104**, 259-273
(1984).
50 Jan, Y. N., Ghysen, A., Christoph, I., Barbel, S. & Jan, L. Y. Formation of Neuronal
Pathways in the Imaginal Discs of *Drosophila melanogaster*. *The Journal of Neuroscience*
5, 2453-2464 (1985).
51 Clarke, A. *et al.* Dynamic Morphogenesis of a Pioneer Axon in *Drosophila* and Its
Regulation by Abl Tyrosine Kinase. *Mol Biol Cell* **31**, 452-465 (2020).

52   Clarke, A. *et al.* Abl Signaling Directs Growth of a Pioneer Axon in *Drosophila* by Shaping
the Intrinsic Fluctuations of Actin. *Mol Biol Cell* **31**, 466-477 (2020).

53   Hulme, S. E., Shevkoplyas, S. S., Apfeld, J., Fontana, W. & Whitesides, G. A
Microfabricated Array of Clamps for Immobilizing and Imaging *C. Elegans*. *Lab Chip* **7**,
1515-1523 (2007).

54   Guo, M. *et al.* Rapid image deconvolution and multiview fusion for optical microscopy.
Nature Biotechnol., in press, doi:https://doi.org/10.1038/s41587-020-0560-x (2020).

55   Chen, T.-W. *et al.* Ultra-sensitive fluorescent proteins for imaging neuronal activity.
Nature **499**, 295-300 (2013).

56   Nguyen, J. P., Linder, A. N., Plummer, G. S., Shaevitz, J. W. & Leifer, A. M. Automatically
tracking neurons in a moving and deforming brain. *PLoS Comput Biol.* **13**, e1005517
(2017).

57   Larsch, J., Ventimiglia, D., Bargmann, C. I. & Albrecht, D. R. High-throughput imaging of
neuronal activity in Caenorhabditis elegans. *PNAS* **110**, E4266-4273 (2013).

58   Sengupta, P., Chou, J. H. & Bargmann, C. I. odr-10 Encodes a Seven Transmembrane
Domain Olfactory Receptor Required for Responses to the Odorant Diacetyl. *Cell* **84**,
899-909 (1996).

59   Shidara, H., Hotta, K. & Oka, K. Compartmentalized cGMP Responses of Olfactory
Sensory Neurons in *Caenorhabditis elegans*. *Journal of Neuroscience* **37**, 3753-3763
(2017).

60   Klapoetke, N. C. *et al.* Independent optical excitation of distinct neural populations.
Nature Methods **11**, 338-346 (2014).

61   Martin, C. *et al.* Line excitation array detection fluorescence microscopy at 0.8 million
frames per second *Nature Communications* **9**, 4499 (2018).

62   Han, Y. *et al.* Cameraless high-throughput three-dimensional imaging flow cytometry.
Optica **6**, 1297-1304 (2019).

63   Ravasio, A., Vaishnavi, S., Ladoux, B. & Viasnoff, V. High-resolution imaging of cellular
processes across textured surfaces using an index-matched elastomer. *Acta Biomater.*
14, 53-60 (2015).

64   Xiao, F., Wen, X., Tan, X. H. M. & Chiou, P.-Y. Plasmonic micropillars for precision cell
force measurements across a large field-of-view *Appl. Phys. Lett.* **112**, 033701 (2018).

65   Duffy, D. C., McDonald, J. C., Schueller, O. J. & Whitesides, G. M. Rapid Prototyping of
Microfluidic Systems in Poly(dimethylsiloxane). *Anal Chem* **70**, 4974-4984 (1998).

66   Guo, M. *et al.* Single-shot super-resolution total internal reflection fluorescence
microscopy. *Nature Methods* **15**, 425-428 (2018).

67   Xia, Y. & Whitesides, G. M. Soft Lithography. *Angew. Chem. Int. Ed. Engl.* **37**, 550-575
(1998).

68   Atencia, J. *et al.* Magnetic connectors for microfluidic applications *Lab Chip* **10**, 246-249
(2010).

69   Edelstein, A. D. *et al.* Advanced methods of microscope control using μManager
software. *Journal of Biological Methods* **1**, e11 (2014).

70   Ardiel, E. L. *et al.* Visualizing Calcium Flux in Freely Moving Nematode Embryos. *Biophys.*
J. **112**, 1975-1983 (2017).

71   Lagoy, R. C. & Albrecht, D. R. Microfluidic Devices for Behavioral Analysis, Microscopy,
and Neuronal Imaging in Caenorhabditis elegans. *Methods Mol Biol.* **1327**, 159-179
(2015).
72   Lagoy, R. C. & Albrecht, D. R. Automated fluid delivery from multiwell plates to
microfluidic devices for high-throughput experiments and microscopy. *Sci. Reports* **8**,
6217 (2018).
963