1
2
3
4

# Automated Segmentation of Fractured Distal Radii by 3D Geodesic Active Contouring of *in vivo* HR-pQCT Images

10 Authors: Nicholas Ohs(1), Duncan C. Tourolle(1), Penny R. Atkins(1), Caitlyn J.

11 Collins(1), Bryant Schroeder(1), Michael Blauth(2), Patrik Christen (1,3), Ralph

12 Müller(1)

13 *Affiliations:*

    1. *Institute for Biomechanics, ETH Zurich, Zurich, Switzerland*
    2. *Department of Orthopedics and Trauma Surgery, Medical University of Innsbruck, Innsbruck, Austria*
    3. *Institute for Information Systems, FHNW University of Applied Sciences and Arts Northwestern Switzerland, Olten, Switzerland*

*Corresponding Author:*

Ralph Müller
ETH Zurich, Institute for Biomechanics
Leopold-Ruzicka-Weg 4
CH-8093 Switzerland
ram@ethz.ch

# Abstract

Radius fractures are among the most common fracture types; however, there is limited consensus on the standard of care. A better understanding of the fracture healing process could help to shape future treatment protocols and thus improve functional outcomes of patients. High-resolution peripheral quantitative computed tomography (HR-pQCT) allows monitoring and evaluation of the radius on the micro-structural level, which is crucial to our understanding of fracture healing. However, current radius fracture studies using HR-pQCT are limited by the lack of automated contouring routines, hence only including small number of patients due to the prohibitively time-consuming task of manually contouring HR-pQCT images.

In the present study, a new method to automatically contour images of distal radius fractures based on 3D morphological geodesic active contours (3D-GAC) is presented. Contours of 60 HR-pQCT images of fractured and conservatively treated radii spanning the healing process up to one year post-fracture are compared to the current gold standard, hand-drawn 2D contours, to assess the accuracy of the algorithm. Furthermore, robustness was established by applying the algorithm to HR-pQCT images of intact radii of 73 patients and comparing the resulting morphometric indices to the gold standard patient evaluation. Reproducibility was evaluated using repeat scans of intact radii of 19 patients.

The new 3D-GAC approach offers accurate contours within inter-operator variability for images of fractured distal radii. The generated contours for images of intact radii yielded morphometric indices within *in vivo* reproducibility limits compared to the current gold standard. Additionally, the 3D-GAC approach shows an improved robustness against failure when dealing with cortical interruptions, fracture fragments, etc. Using the 3D-GAC approach assures consistent results, while reducing the need for time-consuming semi-automatic contouring.

52      Keywords: HR-pQCT, fracture, active contours, automatic, radius

# 1 Introduction

In our ageing population, bone fractures are increasingly common and have become a major socioeconomic burden [1]. Specifically, fractures of the distal radius are among the most common fracture types and indicative of reduced bone quality [2]. However, there is currently limited consensus on the optimal treatment protocol [3]. The conservative treatment of many radius fractures provides the opportunity to study the fracture healing process in humans. A better understanding of this process could help to shape future treatment protocols of distal radius fractures, ultimately resulting in better functional outcomes for patients.

With the introduction of high-resolution peripheral quantitative computed tomography (HR-pQCT), longitudinal changes in the microstructure of the radius can be monitored non-invasively [5–9]. Use of this technology has revealed critical changes in bone microstructure as a result of aging as well as pharmaceutical and surgical interventions. For example, an increase in cortical porosity, which is highly associated with bone frailty, was revealed in a longitudinal study of patients who had undergone kidney transplantation [9]. Preliminary evidence suggests that HR-pQCT might also be a viable technique to monitor bone fractures *in vivo* [10]. The microstructure in such studies is of fundamental importance, as techniques such as micro-finite-element analysis can be utilized to estimate the capability of a bone scanned with HR-pQCT to withstand mechanical load, a key parameter to assess successful fracture healing.

Image-based clinical research, such as that using HR-pQCT, requires accurate, reproducible and scalable tools for image processing. For this reason, automatic contouring approaches have been developed to reduce the need for hand-drawn contours [11–14] in HR-pQCT studies. The method developed by Buie et al. [11], specifically, is integrated into the software of the manufacturer of HR-pQCT devices (XtremeCT I and II, Scanco Medical AG, Switzerland), making it the de-facto standard. All of these automatic contouring approaches segment the target bone (i.e. radius) by first finding the outer contour of the bone with the assumption that the largest cortical interruptions imaged

with HR-pQCT are small (on the sub-mm length scale) and that the cortex forms a well-defined, high-

contrast edge in the HR-pQCT image. However, these assumptions break down when processing

images of fractured radii where the cortex is fragmented, and the presence of cartilaginous callus

may blur the boundary between bone and soft tissue. Furthermore, radius fractures often occur in

the ultra-distal region, contributing to the complexity of the image segmentation, as the cortex is

both thinner and less mineralized. To our knowledge, approaches to generate automatic outer

contours for HR-pQCT images of distal radius fractures have yet to be proposed.

Active contouring [15] is a promising technique that has already been shown to successfully segment

HR-pQCT images of healthy distal radii [16,17]. The advantage of the active contouring algorithm is

the inherent ability to pass over larger gaps in the bone.  This is possible because the requirements

for a contour to be both close to the object edges and smooth are balanced. Recently, three-

dimensional morphological geodesic active contours (3D-GAC) have been implemented, allowing for

efficient calculation of active contours on 3D images [18,19]. Hence, 3D-GAC appears to be a

promising approach to create outer contours of HR-pQCT images of distal radius fractures.

The goal of present study was to investigate the use of fully automatic 3D-GAC to generate outer

contours on HR-pQCT images of both intact and fractured radii. We validated the accuracy of these

contours against hand-drawn contours using HR-pQCT images acquired throughout the first year of

the healing process from 10 patients with fractured distal radii. We further assessed the robustness

of the algorithm by comparing computed morphometric indices from intact distal radii HR-pQCT

images of 73 subjects using contours generated from (i) the 3D-GAC and (ii) the scanner

manufacturer default approach. Lastly, the reproducibility of the 3D-GAC based algorithm was

assessed based on contours of 19 subjects imaged six times using HR-pQCT.

## 2  Materials and Methods

### 2.1  Data

HR-pQCT images (XtremeCT II, 60.7 μm voxels, 68 kV, 1470 μA, 43 ms integration time) were

obtained from the database of a previous fracture study conducted at Innsbruck Medical University.

Patients provided informed consent and participated in a study approved by the ethics committee of

the Medical University of Innsbruck (UN 0374344/4.31). For each patient, six scans of the fractured

(up to 504 slices per scan distributed over three stacks) and contralateral (up to 168 slices per scan)

radius were taken over the course of one year (1, 3, 5, 13, 26, and 52 weeks post-fracture).

### 2.1.1  Fractured Radii

Images of radius fractures for 10 out of the 75 patients that completed the study were selected from

the database based on the highest available visual grading scores (VGS) [20] and well aligned stacks

resulting in 60 HR-pQCT fractured radius images. For each image, the most distal slice before the

appearance of the lunate fossa was identified, and all evaluations on these images were performed

on slices proximal to that slice.

### 2.1.2  Intact Radii

Seventy-three patients were selected which had at least one HR-pQCT image of the contralateral,

intact radius with a VGS of three or better resulting in 438 images. Of these 73 patients, 19 had a VGS

of three or better across all time points and were used in the reproducibility study.

### 2.2  Image Pre-Processing

HR-pQCT images of each patient were registered using rigid image registration [21] to allow direct

comparison of generated contours across all time-points per patient.

### 2.3  Contours: Current Gold Standards

For ten randomly selected slices per fractured radius HR-pQCT image, hand-drawn 2D contours were

generated by three researchers experienced in the processing of medical images (DCT, PRA, CJC),

respectively, using the software of the scanner manufacturer. For this procedure, the 2D slice is

visualised in the scanner manufacturer's software, and using a computer mouse, the contour is

drawn onto the image. The software provides zooming functionality and local edge detection to

assist the operator in defining the outer contour. These hand-drawn 2D contours were used as the

gold standard to validate the results of the 3D-GAC algorithm for fractured radii and to quantify the

inherent inter-observer variability in hand-drawn contours of fractured radii.

For intact radii, the default manufacturer pipeline based on a 3D algorithm by Buie et al. [11], written

in the Image Processing Language (IPL) of the scanner manufacturer (Scanco Medical AG,

Switzerland), was used as the gold standard. All contours generated using the manufacturer's

software were exported as binary 3D images to h5 files [22], a hierarchical data format, for further

processing using Python [23].

## 2.4  Novel Contouring Algorithm

The proposed algorithm uses 3D morphological geodesic active contours (3D-GAC) [18] implemented

in the Python library Scikit-Image [24] using morphological operators [19]. The guiding principle for

geodesic active contours is to minimize an energy term consisting of both internal and image energy.

The internal energy penalizes non-smooth contours, while the image energy penalizes contours away

from the voxels of interest. The image energy landscape, in this approach, is generated using an

inverse Gaussian gradient, where the Gaussian blur removes local minima and the gradient is an

operation to detect object edges.

The proposed contouring algorithm can be separated into four distinct steps. First, the image is

segmented into sections that only contain one bone (here, radius or ulna). In a second step, an initial

guess of the radius contour is generated. Third, the image is pre-processed and converted to an

energy landscape. Finally, the 3D-GAC algorithm is applied in an iterative loop to isolate the surface

of the radius.

### 2.4.1 Segment Radius and Ulna

3D-GAC are drawn to all intensity-based edges in an image and do not differentiate between objects of interest and other objects. Therefore, it is necessary to remove extraneous objects. Here, the ulna must be removed, as it is in close proximity to the radius. The watershed algorithm implemented in Scikit-Image [24] is used to separate these two bones, based on seed voxels of the different objects.

#### 2.4.1.1 Seed Voxels of Radius and ulna

Since the radius and ulna are typically only a few voxels apart in ultra-distal scans, segmenting using a global threshold has the risk of identifying both bones as a single object. Therefore, the radius and ulna are identified in a proximal slice (here, 35-39 slices from proximal to avoid end effects from registration) using a threshold of 250 mg HA / cm$^3$ and component labelling implemented in the Python library NumPy [25]. The largest component identified is the radius and the second largest one the ulna. Finally, a seeds image is generated by copying the labels of the radius and ulna in the proximal slice into the entire proximal half of a zeroed image. The distal part was left empty to avoid incorrect assignments of seeds in the distal region, where radius and ulna are not well separated.

#### 2.4.1.2 Watershed Algorithm

The energy landscape for the watershed algorithm is generated from the HR-pQCT image by setting all negative values to zero and inverting the sign of all remaining elements. The algorithm is applied to the entire image such that every voxel – background or bone – is assigned to exactly either the radius or ulna.

### 2.4.2 Initial Guess

To generate the initial guess contour for the 3D-GAC algorithm, the image is first normalized to the range of zero to one. Afterwards, the image is equalized in the two orthogonal longitudinal planes independently using a contrast limited adaptive histogram equalization implemented in Scikit-Image [24] and the resulting two images are averaged. A threshold of 0.5 is applied and the ulna portion of the image identified during pre-processing is removed from the thresholded image. Finally, the greatest connected component is extracted via component labelling, the structure is closed using a

174    closing distance of 50 voxels, the remaining interior holes are filled using an additional component-

175    labelling step and the structure is dilated using five iterations to obtain the initial guess.

### 176    2.4.3   Pre-Processing of Image

To prepare the image from the scanner to be used as an input to the 3D-GAC algorithm, the image is

first cropped to the bounds of the initial guess contour to reduce the image size, which reduces

computational cost. The image is then normalized to the range of zero to one and equalized as is

done during the creation of the initial guess. The catchment basins from the pre-processing

watershed algorithm are used to set those voxels not belonging to the radius section to zero. Finally,

the voxels outside the initial guess are set to the mean value of the initial guess surface voxels, which

are those voxels that make up the perimeter of the initial guess.

### 184    2.4.4   3D Geodesic Active Contours (3D-GAC)

185    The 3D-GAC algorithm takes an initial guess, an energy landscape and outputs an optimized contour.

186    The energy landscape needs to be generated from the input image. Since the algorithm is easily

187    trapped in local minima, an iterative application of 3D-GAC to energy landscapes containing

188    increasing levels of details is part of the proposed approach.

#### 189    *2.4.4.1   Density to Energy Landscape Conversion*

190    The pre-processed image is used to create energy landscapes. For this, the image is first padded in

191    the longitudinal direction by 20 voxels at each end using the NumPy [25] edge mirroring setting. The

192    padding is followed by the application of two iterations of a Gaussian filter. The sigma of the

193    Gaussian filter is a parameter that is decreased step-wise during the iterative application of 3D-GAC.

194    Finally, the padding voxels are removed.

#### 195    *2.4.4.2   Iterative Application of 3D-GAC*

To avoid finding an unwanted local minima (that is, a contour that does not match the surface of the

radius), Gaussian blurring is used to smooth the energy landscape. However, finer details of the

radius can get lost when a Gaussian blurring with a large sigma is used. Therefore, three iterations of

3D-GAC are run using decreasing sigmas (14.0, 3.0, 1.5 voxels) for the Gaussian blurring to allow for

an iterative approximation of the true radius contour. Additionally, for the first iteration, image

resolutions are halved after blurring to speed up computations. Each 3D-GAC application is then run

using the default parameters of SciPy [26] with the number of iterations set to five. As a final step to

the entire algorithm, a component labelling followed by a selection of everything not connected to

the background is performed to remove holes in the contour that can appear at the distal image

boundary.

## 2.5 Morphometric Indices

The Scanco system provides the standard patient evaluation script for XtremeCT II devices, which is

the gold standard approach of generating cortical and trabecular masks [13] and computing bone

morphometric indices. Total volumetric bone mineral density (Tt.vBMD), trabecular volumetric bone

mineral density (Tb.vBMD), cortical volumetric bone mineral density (Ct.vBMD), trabecular area

(Tb.Ar), cortical area (Ct.Ar), trabecular bone volume fraction (Tb.BV/TV), trabecular number (Tb.N),

trabecular thickness (Tb.Th), trabecular separation (Tb.Sp), cortical perimeter (Ct.Pm), and cortical

thickness (Ct.Th) were computed for this study. The default setup uses outer contours generated by

the approach of Buie et al. as input [11]. However, the standard patient evaluation pipeline also

accepts alternative outer contours as input. Outer contours generated from the proposed novel 3D-

GAC method were used as input to the standard patient evaluations and the resultant

morphometrics were compared to those of the manufacturer's default pipeline.

## 2.6 Study Design

219 To assess the quality of the generated contours, we compared them to the respective gold standard

220 for two different datasets (fractured and intact distal radii).

### 2.6.1 Fractured Radii: Comparison of the 3D-GAC Approach to Hand-Drawn Contours

The 3D-GAC method was run for all fracture images. Contours were visually categorised as acceptable, mistakenly including parts of the ulna, having obvious missing parts, or having both of these issues. Only images categorised as acceptable were used for following quantitative analysis steps. Agreement between hand-drawn contour operators was assessed by computing the Dice score between their respective 2D contours for each image. The Dice score is a typical measure to compare automatically generated contours against reference contours yielding values in the range from zero to one with zero indicating no overlap while one indicates a perfect match. The Dice score is close to percent agreement when comparing contours that are similar to each other. Additionally, the distance transform of the area between the edges of 2D contours between operators was computed. Accuracy of the 3D-GAC approach was determined by taking the median Dice score between the automatic and the three hand-drawn 2D contours for each of the slices per image for which hand-drawn 2D contours exists. Furthermore, all three hand-drawn 2D contours were combined to yield a smallest and largest contour. The voxels between these two contours describes an area of uncertainty for the hand-drawn contours. Voxels, which were in the smallest hand-drawn 2D contour but not in the automatic one and voxels, which were in the 3D-GAC contour but not in the largest hand-drawn 2D contour were extracted and a distance transform was applied as a measure for deviations of the 3D-GAC contours from the hand-drawn 2D ones. Bland-Altman plots between the 3D-GAC and the closest hand-drawn 2D contour in terms of area from all three operators for each slice were generated to identify systematic differences between hand-drawn contours and the 3D-GAC outer contours. The default manufacturer pipeline was also run on all fracture images and categorisation as for the 3D-GAC contours was performed.

### 2.6.2 Intact Radii: Comparison of the 3D-GAC and Manufacturer's Approach

The 3D-GAC approach and the default manufacturer pipeline were run on all intact radii. Contours of both approaches were visually assessed and categorized as was done for the contours of the

fractured radii. For each patient, one image for which both approaches were categorized as being

acceptable was randomly chosen for further analysis. Accuracy of the 3D-GAC method was assessed

by computing the Dice score between each contour of the 3D-GAC method and the respective

contour of the manufacturer's method. Additionally, the distance transform of the volume between

the surfaces of both approaches was computed as another measure of how much the two methods

deviated from each other. Bland-Altman plots were generated to assess systematic differences

between the two methods. Reproducibility was determined by computing the Dice score of each

follow-up image with the baseline image for each patient. Morphometrics were computed to assess

how strongly the computation of standard morphometric indices is influenced by the choice of outer

contour.

## 2.7 Bland-Altman Plots & Statistics

Bland-Altman plots were created to assess systematic deviations for morphometric indices. To allow

visual comparison of the magnitudes of deviations between different parameters, all Bland-Altman

plots had their y-axis scaled by 13% of the maximum x-value throughout this paper, which was the

maximum value not leading to clipping of data. Robust linear models were fitted for Bland-Altman

plots with Huber's T for M estimation using the statsmodels package in Python [27]. This package

also provides two-sided Student's t-tests for the linear models with zero slope and intercept of each

Bland-Altman plot being the null-hypotheses, respectively.

Statistical significance for the comparison of morphometrics was performed using a paired Student's

t-test. Significance level was set to $P<0.05$.

# 3 Results

## 3.1 Fractured Radii: Comparison of the 3D-GAC Approach to Hand-Drawn

Contours

From the 60 images, 20 were successfully contoured by both 3D-GAC and the default manufacturer

pipeline. The latter approach failed 40 times, while the 3D-GAC approach failed for four images

(Figure 1a). The default manufacturer pipeline failed due to missing parts of the radius as a result of

fracture gaps in the cortex (N=33), including parts of the ulna (N=2) or both (N=5). The 3D-GAC

approach failed due to missing parts of the radius (N=2), including parts of the ulna (N=1), or

including parts of the background (N=1). Furthermore, the default manufacturer pipeline only

showed a high success rate for images taken one year post fracture (Figure 1b). It should be noted,

that the default manufacturer pipeline has not been designed to handle images of fractured radii.

Visually, contours matched the given bone structures well within the inter-operator variability (Figure

2). Agreement was found between the 3D-GAC approach and hand-drawn contours with a mean Dice

score of 0.992 ± 0.004. This is within the range of inter-operator agreement: (operator 1 vs operator

2) 0.993 ± 0.005, (operator 1 vs operator 3) 0.993 ± 0.006, and (operator 2 vs operator 3) 0.994 ±

0.005. Evaluating the six appointments individually, Dice scores were high (>0.985), though some

increase in the inter-quartile range of the Dice scores was observed between three and thirteen

weeks post fracture (Figure 1c). This coincided with a decrease in contouring success (Figure 1b). For

voxels of the 3D-GAC deviating from the area of uncertainty, 82% of these contour voxels were only

one voxel away from at least one hand-drawn contour. This was within the inter-observer variability,

as we observed (operator 1 vs operator 2) 57%, (operator 1 vs operator 3) 60%, and (operator 2 vs

operator 3) 58% of deviations being less than one voxel away between operators (Figure 1d).

Comparing the contour areas between the 3D-GAC method and the three operators, agreement was

high with errors being: -1.14 ± 0.76%, -0.40 ± 0.62%, and -0.22 ± 0.73% for each operator,

respectively. An underestimation of the total area was observed for the 3D-GAC contours (Figure 3).

This underestimation is less than e.g. comparing operator 1 to operator 3: -0.011*[±0.004] +

0.090[±0.134], i.e.: slope (-0.003 vs 0.011) and intercept (-0.076 vs 0.090 (both not significant))

(Figure 3).

## 3.2 Intact Radii: Comparison of the 3D-GAC and Manufacturer's Approach

From the 438 images of intact radii, 341 were successfully contoured by both approaches. The

default manufacturer pipeline failed for 85 images, the 3D-GAC approach failed 3 times, and both

approaches failed 9 times (Figure 4a). The default manufacturer pipeline failed due to the inclusion

of parts of the ulna (N=42) into the radius contour, omitting large parts of the radius (N=37) (Figure

5), having parts of the background in the contour (N=5), or including parts of the ulna while omitting

large parts of the radius (N=10). The 3D-GAC approach failed due inclusion of background (N=8),

large parts of the radius missing (N=1), missing parts of the radius and inclusion of parts of the ulna

(N=2), and all mentioned failure modes combined (N=1). The VGS for the failed contours of the

manufacturer's approach ranged from one to five (N=7 for VGS 1, N=14 for VGS 2, N=16 for VGS 3,

N=30 for VGS 4, N=27 for VGS 5). For the thirteen contours that only failed for the 3D-GAC approach

(N=1 for VGS 2, N=5 for VGS 4, N=6 for VGS 5).

Of the 341 images, 72 images (one image per patient) with sufficient image quality (VGS of three or

better) and acceptable contours from both methods were used for further analysis. Overall

agreement between both approaches was high with a mean Dice score of 0.996 ± 0.001, with typical

Dice score values for contours with missing parts of the radius being around 0.85. Reproducibility is

on the same level as the default manufacturer pipeline, with agreement with the baseline image for

both methods and for all appointments (median Dice score > 0.994 for all time-points) (Figure 4b).

For those voxels that were either part of the 3D-GAC contour and not part of the contour from the

default manufacturer pipeline or vice versa, 91% are at most one voxel away from the

manufacturer's contour (Figure 4c). Bland-Atman plots reveal agreement in morphometric indices

with a slight thickness and density based bias in cortical parameters and no bias in trabecular

parameters (Figure 6). Mean deviations were below 1% for all parameters except for Ct.Th (1.341%)

(Table 1).

# 4 Discussion

## 4.1 Fractured Radii: Comparison of the 3D-GAC Approach to Hand-Drawn Contours

56 out of 60 images were contoured without major issues by the 3D-GAC approach. We observed clear inter-operator variability that is a known issue of hand-contouring bone [28]. On average, the generated contours by the 3D-GAC algorithm agree well with operators. The slightly larger inter-quartile range especially for appointments 2 and 3 can be explained by the formation of a callus in this time period and the resulting low mineralization of background voxels visible in the HR-pQCT images. This makes it harder for both operators and the algorithm to decide on the correct contour. Bland-Altman plots show no systematic deviations between operators and the 3D-GAC approach that is larger than between operators themselves (Figure 3). Hence, the 3D-GAC contouring approach provides accurate contours for HR-pQCT images of fractured distal radii compared to the gold standard.

Interestingly, the default manufacturer pipeline only reliably contoured HR-pQCT images taken one year post fracture (Figure 1b). Given the fact that this pipeline was not designed to work with images of impaired cortices, this might indicate structural weaknesses in the cortex even 6 months post fracture. This observation is in agreement with recent findings suggesting that fracture repair is a long term process which might take as long as two years [29].

## 4.2 Intact Radii: Comparison of the 3D-GAC and Manufacturer's Approach

Comparing the 3D-GAC approach to the manufacturer default pipeline, the failure rate of the latter approach was much higher (94 vs 12 failed contours). Since this study used HR-pQCT images of fracture patients, the used cohort may inherently have poorer bone quality than standard populations, which could be one reason for the high number of failed contours for the default manufacturer approach. One of the major reasons for failure was the inclusion of the ulna into the outer contour of the radius, which is a known issue of this approach that is implementation

dependent [11]. The second reason for failure is having obvious missing parts, which appears to

happen if the cortex has very thin or low-density structures (Figure 5). The fallback option of the

default manufacturer pipeline is to create hand-drawn contours, which can be done using a built-in

tool of the manufacturer's software. Another proposed solution for this is to increase of the number

of dilation steps of the algorithm [11], however this can lead to unwanted smoothing of the contour

[11]. Defining per-image numbers of dilation steps on the other hand would reduce the degree of

automation drastically. The 3D-GAC approach only failed for one image with sufficient image quality

to be typically included in HR-pQCT studies [9,28,30], showing and overall improvement in

robustness compared to the current gold standard.

Accuracy of the generated contours was high with the mean Dice score being 0.996 ± 0.001. Using

these generated masks, trabecular morphometric indices agreed well with those generated using the

gold standard method with deviations well below the precision error of these indices [31].

Interestingly, for the cortical parameters, accuracy was slightly worse with Ct.Th showing deviations

of 1-2%; however, these differences are still less than 1 voxel in thickness and thus at the accuracy

limit of the device (Figure 6). Furthermore, differences in cortical thickness were at the

reproducibility limit of the standard patient evaluation (1.3-3.9% dependent on subject age) [13] and

were still below differences observed between patient groups from the literature, e.g. healthy and

osteopenic women (12.8%) [32], indicating that the 3D-GAC approach is a viable option for clinical

studies of non-fractured radii as well.

### 363  4.3  Limitations

This study is not without limitations. The repeat scans of the contralateral site were spaced over a

year, thus may include variations in bone due to natural remodelling. Additional scans for

reproducibility were not possible due to radiation concerns. However, it would be expected that

these natural variations would only decrease the reproducibility found. The same intervals between

scans was used for the images of the fractured radii. Due to the large structural changes happening

during fracture repair, the images of the fracture site could not regarded as repeat scans. Therefore,

no direct assessment of reproducibility was possible for images of fractured radii. However, the high

accuracy observed for 60 images does indicates also high levels of reproducibility. Another limitation

is the low number of hand-drawn contour slices (270). While time was a limiting factor (operators

reporting to take roughly 1-2 minutes per slice) for creating more hand-drawn contours, we included

images of the intact contralateral site to cover a large variety of bone morphologies, which we could

benchmark against existing automatic approaches for intact radii. Another limitation is that the 3D-

GAC contours were generated only proximal to the lunate fossa. However, the morphometric

analysis of the standard patient evaluation has only been developed for that region, making this

region the most relevant for clinical studies. Furthermore, no hand-corrected contours were

generated for intact radii to estimate the bias introduced by the automated protocol. However,

recent studies have already done this for the gold standard [33] and, given the good agreement of

the 3D-GAC approach with the gold standard, the requirement for minor manual corrections should

be on a similar level.

## 5   Conclusion

384 Our proposed 3D-GAC algorithm provides a unified pipeline for generating outer contours from HR-

385 pQCT images of distal radii, including both fractured and intact bones. Importantly, our method

386 automatically contours radii both accurately and reproducibly, while showing robustness when

387 dealing with a wide variety of cortex structures. This will facilitate future clinical studies using large

388 patient cohorts to support the development of improved treatment protocols of radius fractures.

# 6  References

[1]  O. Johnell, J.A. Kanis, An estimate of the worldwide prevalence and disability associated with

osteoporotic fractures, Osteoporos. Int. 17 (2006) 1726–1733.

https://doi.org/10.1007/s00198-006-0172-4.

[2]  C.M. Court-Brown, B. Caesar, Epidemiology of adult fractures: A review, Injury. 37 (2006) 691–

697. https://doi.org/10.1016/j.injury.2006.04.130.

[3]  C.Y. Ng, M.M. McQueen, What are the radiological predictors of functional outcome following

fractures of the distal radius?, J. Bone Jt. Surg. - Ser. B. 93 B (2011) 145–150.

https://doi.org/10.1302/0301-620X.93B2.25631.

[4]  F.E. Turrentine, H. Wang, V.B. Simpson, R.S. Jones, Surgical Risk Factors, Morbidity, and

Mortality in Elderly Patients, J. Am. Coll. Surg. 203 (2006) 865–877.

https://doi.org/10.1016/j.jamcollsurg.2006.08.026.

[5]  A.J. Burghardt, G.J. Kazakia, M. Sode, A.E. De Papp, T.M. Link, S. Majumdar, A longitudinal HR-

pQCT study of alendronate treatment in postmenopausal women with low bone density:

Relations among density, cortical and trabecular microarchitecture, biomechanics, and bone

turnover, J. Bone Miner. Res. 25 (2010) 2558–2571. https://doi.org/10.1002/jbmr.157.

[6]  V. V. Shanbhogue, R.K. Støving, K.H. Frederiksen, S. Hanson, K. Brixen, J. Gram, N.R.

Jørgensen, S. Hansen, Bone structural changes after gastric bypass surgery evaluated by HR-

pQCT: A two-year longitudinal study, Eur. J. Endocrinol. 176 (2017) 685–693.

https://doi.org/10.1530/EJE-17-0014.

[7]  L.A. Burt, D.A. Hanley, S.K. Boyd, Cross-sectional Versus Longitudinal Change in a Prospective

HR-pQCT Study, J. Bone Miner. Res. 32 (2017) 1505–1513. https://doi.org/10.1002/jbmr.3129.

[8]  R. Ellouz, R. Chapurlat, B. van Rietbergen, P. Christen, J.B. Pialat, S. Boutroy, Challenges in

longitudinal measurements with HR-pQCT: Evaluation of a 3D registration method to improve

bone microarchitecture and strength measurement reproducibility, Bone. 63 (2014) 147–157.

https://doi.org/10.1016/j.bone.2014.03.001.

[9]    K.K. Nishiyama, Y. Pauchard, L.E. Nikkel, S. Iyer, C. Zhang, D.J. McMahon, D. Cohen, S.K. Boyd,

E. Shane, T.L. Nickolas, Longitudinal HR-pQCT and image registration detects endocortical

bone loss in kidney transplantation patients, J. Bone Miner. Res. 30 (2015) 456–463.

https://doi.org/10.1002/jbmr.2358.

[10]   J.J.A. De Jong, P.C. Willems, J.J. Arts, S.G.P. Bours, P.R.G. Brink, T.A.C.M. van Geel, M. Poeze,

P.P. Geusens, B. van Rietbergen, J.P.W. van den Bergh, Assessment of the healing process in

distal radius fractures by high resolution peripheral quantitative computed tomography,

Bone. 64 (2014) 65–74. https://doi.org/10.1016/j.bone.2014.03.043.

[11]   H.R. Buie, G.M. Campbell, R.J. Klinck, J.A. MacNeil, S.K. Boyd, Automatic segmentation of

cortical and trabecular compartments based on a dual threshold technique for in vivo micro-

CT bone analysis, Bone. 41 (2007) 505–515. https://doi.org/10.1016/j.bone.2007.07.007.

[12]   A.J. Burghardt, G.J. Kazakia, S. Majumdar, A local adaptive threshold strategy for high

resolution peripheral quantitative computed tomography of trabecular bone, Ann. Biomed.

Eng. 35 (2007) 1678–1686. https://doi.org/10.1007/s10439-007-9344-4.

[13]   A.J. Burghardt, H.R. Buie, A. Laib, S. Majumdar, S.K. Boyd, Reproducibility of direct

quantitative measures of cortical bone microarchitecture of the distal radius and tibia by HR-

pQCT, Bone. 47 (2010) 519–528. https://doi.org/10.1016/j.bone.2010.05.034.

[14]   R. Zebaze, A. Ghasem-Zadeh, A. Mbala, E. Seeman, A new method of segmentation of

compact-appearing, Transitional and trabecular compartments and quantification of cortical

porosity from high resolution peripheral quantitative computed tomographic images, Bone.

54 (2013) 8–20. https://doi.org/10.1016/j.bone.2013.01.007.

[15]   M. Kass, A. Witkin, D. Terzopoulos, Snakes: Active contour models, Int. J. Comput. Vis. 1

(1988) 321–331. https://doi.org/10.1007/BF00133570.

[16]   M. Hafri, H. Toumi, S. Boutroy, R.D. Chapurlat, E. Lespessailles, R. Jennane, Fuzzy energy

based active contours model for HR-PQCT cortical bone segmentation, in: 2016 IEEE Int. Conf.

Image Process., IEEE, 2016: pp. 4334–4338. https://doi.org/10.1109/ICIP.2016.7533178.

[17]   M. Hafri, R. Jennane, E. Lespessailles, H. Toumi, Dual active contours model for HR-pQCT

cortical bone segmentation, in: 2016 23rd Int. Conf. Pattern Recognit., IEEE, 2016: pp. 2270–

2275. https://doi.org/10.1109/ICPR.2016.7899974.

[18]   V. Caselles, R. Kimmel, G. Sapiro, Geodesic Active Contours, Int. J. Comput. Vis. 22 (1997) 61–

79. https://doi.org/10.1023/A:1007979827043.

[19]   P. Marquez-Neila, L. Baumela, L. Alvarez, A morphological approach to curvature-based

evolution of curves and surfaces, IEEE Trans. Pattern Anal. Mach. Intell. 36 (2014) 2–17.

https://doi.org/10.1109/TPAMI.2013.106.

[20]   J.B. Pialat, A.J. Burghardt, M. Sode, T.M. Link, S. Majumdar, Visual grading of motion induced

image degradation in high resolution peripheral computed tomography: Impact of image

quality on measures of bone density and micro-architecture, Bone. 50 (2012) 111–118.

https://doi.org/10.1016/j.bone.2011.10.003.

[21]   P. Thévenaz, U.E. Ruttimann, M. Unser, A pyramid approach to subpixel registration based on

intensity, IEEE Trans. Image Process. 7 (1998) 27–41. https://doi.org/10.1109/83.650848.

[22]   The HDF Group, Hierarchical Data Format, version 5, (2020).

http://www.hdfgroup.org/HDF5/.

[23]   Python Software Foundation, Python Language Reference, Version 3.7, (2020).

https://www.python.org.

[24]   S. Van Der Walt, J.L. Schönberger, J. Nunez-Iglesias, F. Boulogne, J.D. Warner, N. Yager, E.

Gouillart, T. Yu, Scikit-image: Image processing in python, PeerJ. 2014 (2014) 1–18.

https://doi.org/10.7717/peerj.453.

[25]    S. Van Der Walt, S.C. Colbert, G. Varoquaux, The NumPy array: A structure for efficient

numerical computation, Comput. Sci. Eng. 13 (2011) 22–30.

https://doi.org/10.1109/MCSE.2011.37.

[26]    P. Virtanen, R. Gommers, T.E. Oliphant, M. Haberland, T. Reddy, D. Cournapeau, E. Burovski,

P. Peterson, W. Weckesser, J. Bright, S.J. van der Walt, M. Brett, J. Wilson, K.J. Millman, N.

Mayorov, A.R.J. Nelson, E. Jones, R. Kern, E. Larson, C. Carey, İ. Polat, Y. Feng, E.W. Moore, J.

VanderPlas, D. Laxalde, J. Perktold, R. Cimrman, I. Henriksen, E.A. Quintero, C.R. Harris, A.M.

Archibald, A.H. Ribeiro, F. Pedregosa, P. van Mulbregt, S. 1. 0 Contributors, SciPy 1.0--

Fundamental Algorithms for Scientific Computing in Python, (2019) 1–22.

http://arxiv.org/abs/1907.10121.

[27]    S. Seabold, J. Perktold, statsmodels: Econometric and statistical modeling with python, in: 9th

Python Sci. Conf., 2010.

[28]    E.A.C. de Waard, C. Sarodnik, A. Pennings, J.J.A. de Jong, H.H.C.M. Savelberg, T.A. van Geel,

C.J. van der Kallen, C.D.A. Stehouwer, M.T. Schram, N. Schaper, P.C. Dagnelie, P.P.M.M.

Geusens, A. Koster, B. van Rietbergen, J.P.W. van den Bergh, Reliability of HR-pQCT Derived

Cortical Bone Structural Parameters When Using Uncorrected Instead of Corrected

Automatically Generated Endocortical Contours in a Cross-Sectional Study: The Maastricht

Study, Calcif. Tissue Int. 103 (2018) 252–265. https://doi.org/10.1007/s00223-018-0416-2.

[29]    J.J. de Jong, F.L. Heyer, J.J.C. Arts, M. Poeze, A.P. Keszei, P.C. Willems, B. van Rietbergen, P.P.

Geusens, J.P. van den Bergh, Fracture Repair in the Distal Radius in Postmenopausal Women:

A Follow-Up 2 Years Postfracture Using HRpQCT, J. Bone Miner. Res. 31 (2016) 1114–1122.

https://doi.org/10.1002/jbmr.2766.

[30]    S.L. Manske, E.M. Davison, L.A. Burt, D.A. Raymond, S.K. Boyd, The Estimation of Second-

Generation HR-pQCT From First-Generation HR-pQCT Using In Vivo Cross-Calibration, J. Bone

Miner. Res. 32 (2017) 1514–1524. https://doi.org/10.1002/jbmr.3128.

[31]    T.L. Mueller, M. Stauber, T. Kohler, F. Eckstein, R. Müller, G.H. van Lenthe, Non-invasive bone

competence analysis by high-resolution pQCT: An in vitro reproducibility study on structural

and mechanical properties at the human radius, Bone. 44 (2009) 364–371.

https://doi.org/10.1016/j.bone.2008.10.045.

[32]    K.K. Nishiyama, H.M. Macdonald, H.R. Buie, D.A. Hanley, S.K. Boyd, Postmenopausal women

with osteopenia have higher cortical porosity and thinner cortices at the distal radius and tibia

than women with normal aBMD: An in vivo HR-pQCT study, J. Bone Miner. Res. 25 (2010)

882–890. https://doi.org/10.1359/jbmr.091020.

[33]    D.E. Whittier, A.N. Mudryk, I.D. Vandergaag, L.A. Burt, S.K. Boyd, Optimizing HR-pQCT

workflow: a comparison of bias and precision error for quantitative bone analysis,

Osteoporos. Int. 31 (2020) 567–576. https://doi.org/10.1007/s00198-019-05214-0.

# 7 Tables

502 Table 1: Deviations of the bone morphometrics derived using 3D-GAC based contours compared to

503 the manufacturer's default pipeline.

| Morphometric Index | Deviation from Gold Standard |
|---|---|
| Tt.vBMD | 0.190 ± 3.668% |
| Tb.vBMD | 0.662 ± 3.226% |
| Ct.vBMD | -0.131 ± 2.437% |
| Tb.Ar | 0.510 ± 4.920% |
| Ct.Ar | -0.710* ± 2.530% |
| Tb.BV/TV | 0.540 ± 2.830% |
| Tb.N | -0.093 ± 1.497% |
| Tb.Th | 0.147 ± 1.017% |
| Tb.Sp | 0.181 ± 1.625% |
| Ct.Pm | -0.333 ± 2.402% |
| Ct.Th | -1.341* ± 4.911% |
| Morphometric Indices: total volumetric bone mineral density (Tt.vBMD), trabecular volumetric bone mineral density (Tb.vBMD), cortical volumetric bone mineral density (Ct.vBMD), trabecular area (Tb.Ar), cortical area (Ct.Ar), trabecular bone volume fraction (Tb.BV/TV), trabecular number (Tb.N), trabecular thickness (Tb.Th), trabecular separation (Tb.Sp), cortical perimeter (Ct.Pm), cortical thickness (Ct.Th).<br>*indicates significant deviation (P<0.05) | |

## 8  Figures

### 8.1  Figure 1

Figure 1: Accuracy of the 3D-GAC approach vs hand-drawn contours for scans of fractured radii. a)
Histogram of successful and failed contours for the 3D-GAC and the default manufacturer pipeline.
The final bar shows the images selected for further analysis. Note that this pipeline was not designed
to handle scans of fractured radii. b) Histogram of successful contours for the two approaches per
patient time-appointment. Dashed line indicates 100% success. c) Dice scores computed for all
images between the hand-drawn slices of all operators (op) respectively. The corresponding slices of
the 3D-GAC contours were also compared against the hand-drawn slices, respectively, and the
average median Dice score is shown. d) Histogram of the distance of all voxels different between the
3D-GAC and the area of uncertainty (AOU) for the hand-drawn contours of each operator,

518   respectively. The AOU is defined as the area between the contour obtained by only accepting voxels

519   present in all hand-drawn contours and the contour obtained by accepting voxels present in at least

520   one hand-drawn contour.

## 8.2   Figure 2

523   Figure 2: Visual comparison of hand-drawn contours (by three different operators (op)) and 3D

524   geodesic active contour (3D-GAC) generated contours for a representative challenging slice from one

525   week post-fracture. Visual differences between the 3D-GAC and hand-drawn contours are similar to

526   those between hand-drawn contours of different operators. Scale bar is 2 mm.

## 8.3 Figure 3

529 Figure 3: Bland-Altman plots of the mean contour area per image. a) Comparing operator 1 (op. 1) to

530 operator 2 (op. 2). b) Comparing operator 1 to operator 3 (op. 3). c) Comparing operator 2 to

531 operator 3. d) Comparing the automatic 3D-GAC contours with the median (computed per slice) of all

532 three operators. Area computed with the 3D-GAC approach agrees within the observable inter-

533 operator variability with the area computed from hand-draw contours.

 ## 8.4   Figure 4

Figure 4: Accuracy and reproducibility of the 3D-GAC vs the default manufacturer pipeline (DMP). a)

Histogram of successful and failed contours for the 3D-GAC and the DMP with indicated image

quality. The final bar shows the images selected for the reproducibility analysis. b) Dice scores of

contours of all time points relative to the initial scan for both approaches show the reproducibility of

each method. c) Histogram of the distance of all voxels different between the 3D-GAC and the DMP

contour from the respective DMP contour.

## 8.5 Figure 5

Figure 5: Examples of failure modes when contouring distal radii. Contours from the 3D-GAC and the

default manufacturer pipeline are compared for four challenging contouring cases: A thin cortex

leading to missing parts of the radius, the ulna (partly being visible on the right) being close to the

radius (bone on the left) and hence being included in the contour, the combination of the previous

two, and a low visual grading score. Only the 3D-GAC approach manages to generate visually correct

contours in most cases. Scale bar is 2 mm.

## 8.6 Figure 6

Figure 6: Bland-Altman plots of morphometric indices characterize potential differences in 3D-GAC

and contours from the default manufacturer pipeline. Robust linear model fits are shown in red.

Regression parameters are presented above each plot with slope and intercept parameters

statistically significantly different from zero being indicated by (*, P<0.05). Indices are total

volumetric bone mineral density (Tt.vBMD), trabecular volumetric bone mineral density (Tb.vBMD),

cortical volumetric bone mineral density (Ct.vBMD), trabecular area (Tb.Ar), cortical area (Ct.Ar),

trabecular bone volume fraction (Tb.BV/TV), trabecular number (Tb.N), trabecular thickness (Tb.Th),

trabecular separation (Tb.Sp), cortical perimeter (Ct.Pm), cortical thickness (Ct.Th).