Label-free composition determination for biomolecular condensates with an arbitrarily large number of components

Patrick M. McCall[1,2,3], Kyoohyun Kim[4,5], Martine Ruer-Gruß[1], Jan Peychl[1],
Jochen Guck[4,5,6], Anthony A. Hyman[1,6,#], Jan Brugués[1,2,3,6,#]

[1] Max Planck Institute of Molecular Cell Biology and Genetics, Pfotenhauerstraße 108, 01307 Dresden, Germany
[2] Max Planck Institute for the Physics of Complex Systems, Nöthnitzerstraße 38, 01187 Dresden, Germany
[3] Center for Systems Biology Dresden, Pfotenhauerstraße 108, 01307 Dresden, Germany
[4] Biotechnology Center, Technische Universität Dresden, Tatzberg 47/49, 01307 Dresden, Germany
[5] Max Planck Institute for the Science of Light, Staudstraße 2, 91058 Erlangen, Germany
[6] Cluster of Excellence Physics of Life, Technische Universität Dresden, 01062 Dresden, Germany

[#] Address correspondence to: hyman@mpi-cbg.de; brugues@mpi-cbg.de

ABSTRACT (150 words)

Biomolecular condensates are membrane-less organelles made of multiple components, often including several distinct proteins and nucleic acids. However, current tools to measure condensate composition are limited and cannot capture this complexity quantitatively, as they either require fluorescent labels, which we show can perturb composition, or can distinguish only 1-2 components. Here, we describe a label-free method based on quantitative phase microscopy to measure the composition of condensates with an arbitrarily large number of components. We first validate the method empirically in binary mixtures, revealing sequence-encoded density variation and complex aging dynamics for condensates composed of full-length proteins. In simplified multi-component protein/RNA condensates, we uncover a regime of constant condensate density and a large range of protein:RNA stoichiometry when varying average composition. The unexpected decoupling of density and composition highlights the need to determine molecular stoichiometry in multi-component condensates. We foresee this approach enabling the study of compositional regulation of condensate properties and function.

INTRODUCTION

Many compartments in living cells exist as condensed phases of biopolymers, termed biomolecular condensates, which are demixed from the surrounding cytoplasm or nucleoplasm [1,2], and are implicated in a wide range of cellular processes [3]. Phase separation of a simple binary mixture of a polymer in solvent results in a dilute phase coexisting with a polymer-rich condensed phase (**Fig. 1a**). Although demixing of a single full-length protein in a binary mixture is often sufficient to reconstitute simplified condensates [4–7], condensates *in vivo* contain dozens of components [8–10]. Indeed, the functional identity of a particular condensate

Fig. 1: Phase diagrams of binary and ternary mixtures

a Binary phase diagram

A — 1-phase
B — 2-phase
C — 2-phase
D — 1-phase

Temperature

c_{Dil} c_{Cond}

Concentration

b Ternary phase diagram

A — 1-phase
B — 2-phase
C — 2-phase
D — 1-phase

[Polymer r]

r_B^{II}
r_C^{II}
r_B^{I}
r_C^{I}

p_B^{I} p_C^{I} p_B^{II} p_C^{II}

[Polymer p]

Fig 1: Phase diagrams of binary and ternary mixtures
a, Phase diagram in a prototypical binary mixture with sketches of samples prepared in the 1-
phase (A,D) and 2-phase regimes (B,C). Compositions of coexisting phases, c_{Dil} and c_{Cond}, lie
on the binodal curve (solid black) that separates the 1-phase (mixed) and 2-phase (demixed)
regimes. A tie-line (dashed black) connects compositions of coexisting phases to the
corresponding average composition (open dashed black circle). Varying average solute
concentration in the 2-phase regime (gray) changes the relative volumes of coexisting phases
but not their composition. **b**, Phase diagram in a prototypical ternary mixture with sketches of
samples prepared at different average compositions. Unlike the binary case, ternary mixtures
prepared at different points in the two-phase regime may lie on different tie-lines, yielding
compositionally distinct pairs of coexisting phases.
51

inside a cell is determined by its composition. Unlike binary systems, such multi-component

condensates possess a continuum of compositions [11], each connected to a coexisting dilute

phase via a tie-line in a phase diagram (**Fig. 1b**). Changes in component abundance can thus

shift the system to a new tie-line, altering condensate composition and physical properties [12,13].

Despite its central role in physically defining condensates and specifying their properties, the

composition of multi-component condensates *in vivo*, and component stoichiometries in

reconstituted systems *in vitro*, are largely unknown.

In multi-component systems, condensate composition is typically estimated by fusing

each molecular species to a spectrally-distinct fluorescent tag [12,14,15]. Although powerful,

fluorescent tags contribute to interactions between species and with solvent, potentially shifting

the thermodynamic balance that specifies phase composition [16]. Interactions with the

condensed phase can also drive strong deviations in fluorophore characteristics relative to

behavior calibrated in the dilute phase [17,18,15], confounding quantification. Label-free

techniques avoid these issues entirely, but existing approaches require harsh treatments and are

effectively limited to binary systems for native-like molecules. For example, traditional bulk

approaches like ultra-violet (UV) absorption [20,21] and thermogravimetric analysis [22] require

condensate dissolution or destruction and impose sample requirements that are often

inaccessible with the modest yields obtained by recombinant expression and purification of

endogenous cellular condensate components (**Supplementary Note 1**). Though confocal

Raman spectroscopy enables measurements of intact condensates [18,19], the experimental

requirements of high laser exposure or nanoparticles for surface-enhancements may alter

condensate dynamics, and a calibration sufficient to resolve multi-component composition has

yet to be demonstrated.

Reflecting these limitations, recent UV absorption measurements of the composition of

condensates reconstituted from recombinant proteins looked exclusively at binary systems with

intrinsically disordered protein regions (IDRs) rather than the full-length proteins [23–25]. This is

in part because those fragments could be purified with sufficient yield from bacteria following

denaturation. That is not an option in many cases, however, as bacteria lack the machinery

needed to add post-translational modifications (PTMs) or assist folding of certain protein

domains, and denaturants may irreversibly alter protein conformational ensembles. Thus, it

remains unclear how the additional features of native-like proteins, including PTMs, native-

like conformational ensembles, and the additional domains present in full-length proteins, will

change the picture now emerging for condensates formed from IDRs alone. As functional roles

for PTMs and structured protein domains accumulate alongside molecular parts lists [8–10], there

is a pressing need for methods to reveal condensate composition in more faithful reconstitutions

that include multiple native-like components.

Here, we present a label-free method to precisely measure the composition of micron-

sized condensates reconstituted from an arbitrary number of components. Overall, this method

requires 1000-fold less material than bulk label-free alternatives, which enables dynamic and

temperature-dependent measurements of condensates formed from full-length native-like

components. We demonstrate that quantitative phase imaging (QPI) can be used to extract

refractive index differences between demixed phases, Δn, which we convert to condensed-

phase concentration for binary systems. By combining these Δn with tie-line measurements,

we then show that the concentrations of an arbitrary number of individual species can be

resolved quantitatively in multi-component condensates. We demonstrate this explicitly for a

model ternary system of full-length FUS protein and RNA, and reveal unexpected features in

the composition of these multi-component condensates. Additionally, we find that protein

concentrations vary from 87 to more than 470 mg/ml in condensates depending on sequence

and use of labels, and uncover a dramatic increase in density with age that may underly

previous reports of mechanical hardening [26]. By resolving the chemical composition of multi-

component condensed phases *in situ* and in unprecedented detail, we anticipate that this label-

free method will enable mechanistic studies of complex composition regulation of

biomolecular condensate properties and function with multiple native components.

RESULTS

Optical concept

To measure the compositional difference between micron-sized droplets and the coexisting

dilute phase, we employ quantitative phase imaging (QPI) [27,28]. Physically, QPI measures the

optical phase shift accumulated along a wavefront as it traverses spatial inhomogeneities in

refractive index within a sample, such as high-refractive index droplets immersed in a lower-

index medium **(Fig. 2a)**. Within the first-order Born approximation [29], the optical phase shift

$\Delta\varphi$ measured by QPI at pixel (x, y) is proportional to the product of refractive index difference,

Δn, and droplet shape

$$\Delta\varphi(x, y) = \frac{2\pi}{\lambda} \Delta n H(x, y), \qquad \text{(Eq. 1)}$$

116   where λ is the imaging wavelength and $H(x, y)$ is the projected height of the droplet along the

117   imaging axis. Thus, the refractive index difference between coexisting thermodynamic phases

118   can be obtained from the droplet's shape and optical phase shift. The refractive index of an

Fig. 2: QPI measures the refractive index of micron-size droplets *in situ*

a Phase Delay
Sample

$$\Delta\varphi(x) = \frac{2\pi}{\lambda}\Delta n H(x)$$

$\Delta\varphi$

x position (μm)

b Spherical Cap

(x_c, y_c, z_{eq})

R z_{eq}

x position (μm)

c Height

x position (μm)

d QPI Image

10 μm

e Fit

f Residuals

$\Delta\varphi$ (rad)

1.4

g

Prob.

$\Delta n \times 10^{-3}$ (-)

- ⊕ Sample 1; N = 140
- ⊕ Sample 2; N = 336
- ⊕ Sample 3; N = 162

Radius, R (μm)

h

Δn_{QPI} (-)

- ⊕ Sphere
- ⊡ Droplet

$\Delta n_{Refractometer}$ (-)

i

n (-)

dn/dc (mL/g)

BSA   Dex.   PEG

- ⊡ BSA
- △ Dex.
- ⊡ PEG

$$n = n_0 + \sum_i \frac{dn}{dc_i} c_i$$

[Polymer] (mg/mL)

Fig 2: QPI measures the refractive index of micron-sized droplets *in situ*
a, Schematic of optical wavefronts (red) distorted by droplets on a flat surface (top) and the
cumulative optical phase shift (bottom). **b**, Schematic defining the geometry of a spherical cap.
c, Projected height profile *H(x)* for the geometry in **b**. Shading in **b**,**c** denotes separate terms in
the analytic expression for *H(x)*, (see Methods). **d**, Phase image of dextran-rich droplet on
passivated glass. **e**, Fit of droplet in **d** to spherical cap. **f**, Residuals from fit. Scalebar in **e**-**f** is
10 µm. **g**, Refractive index difference Δn vs. droplet radius *R* extracted from fits to individual
dextran-rich droplets obtained from 3 different PEG/Dextran mixtures (colors). Error bars are
95 % confidence intervals. Probability histograms at right. **h**, Refractive index differences
measured with QPI vs. bulk refractometry for PEG/Dextran mixtures (red) or silica spheres in
different glycerol-water mixtures (black). Error bars are standard deviation. Solid line is *y = x*.
i, Refractive index is a linear function of component concentration for three model biopolymers
(BSA, Dextran, PEG). Inset: Slopes (*dn/dc*) for each polymer with 95 % confidence intervals.
133

134   aqueous protein solution is, in turn, a linear function of concentration over a wide range [30,31].

135   In this regime (see **Supplementary Note 2**), the condensed-phase protein concentration, c_{Cond},

136   in a binary system is given by

137
$$c_{Cond} = \frac{\Delta n}{dn/dc} + c_{Dil}, \qquad\qquad (Eq.\ 2)$$

where c_{Dil} is the dilute phase concentration and *dn/dc* is the slope of the concentration-

dependence of refractive index, which can be estimated from amino acid sequence [32]. This

suggests QPI is suitable to measure the concentration difference between a condensate and its

coexisting dilute phase in a binary system when condensate shape is known.

Measuring droplet shape on flat surfaces

To characterize the shape of condensates typical of *in vitro* reconstitution experiments, we

model them physically as sessile fluid droplets on a flat substrate (**Fig. 2a**). In this context,

droplet shape is determined at equilibrium by interfacial tension opposing gravitational settling

of the denser fluid [33]. For droplets smaller than the capillary length $l_c = (\gamma/\Delta\rho g)^{1/2}$, the

interfacial tension γ is sufficiently strong to suppress expansion of the interfacial area driven

by gravitational effects and the resulting shape is described to an excellent approximation as a

spherical cap [33] (**Fig. 2b**). This approximation is valid for most reconstituted condensates,

which are typically smaller than our capillary length estimate of $\gtrsim 30\ \mu m$ (**Extended Data Fig.**

1). In this limit, four parameters suffice to fully characterize droplet shape: droplet radius R,

the position of the droplet center (x_0, y_0), and the height of the equatorial plane above the

substrate, Z_{eq} (**Fig. 2b, Methods**). Importantly, the corresponding projected height function

$H(x, y | R, x_0, y_0, Z_{eq})$ has a closed analytic form that we use to determine the refractive index

and shape parameters of droplets from QPI images by fitting (**Fig. 2c**, **Methods**).

Experimental validation and conversion from Δn to concentration

To validate the approach experimentally, we use two different reference systems. The first is a

well-characterized PEG/Dextran aqueous two-phase system [34] for which we can readily

measure Δn independently by bulk refractometry (**Methods**). QPI images of Dextran-rich

droplets on passivated coverglass and surrounded by a coexisting PEG-rich phase (**Fig. 2d**) are

well-modeled as spherical caps (**Fig. 2e**), as evidenced by small and spatially unstructured

residuals in the droplet interior following fitting (**Fig. 2f**). For the best fits, Δn is independent

of size for three different PEG/Dextran compositions (**Fig. 2g, left),** and approximately

symmetrically distributed (**Fig. 2g, right**), indicative of equilibrated phases and uncertainty

dominated by statistics rather than systematics, respectively. Crucially, the Δn values extracted

from fL-droplets in QPI images are in excellent agreement with those measured independently

from 100-µL volumes of each phase using a digital refractometer (**Fig. 2h**). To validate

application of the first-order Born approximation for larger Δn, we used silica microspheres

suspended in glycerol-water mixtures as a second reference system, where Δn is set by the

glycerol/water ratio. As with the dextran droplets, we recovered the expected shape without

bias and find excellent agreement between Δn measurements extracted from QPI images and

those expected on the basis of digital refractometry measurements, now over a much larger

range (**Fig. 2h, Extended Data Fig. 2**). Taken together, these data demonstrate that the present

analysis pipeline extracts accurate geometric and optical measurements of homogeneous

sessile droplets from QPI image data for Δn of at least 0.085.

178   Conversion from Δn to compositional differences requires knowledge of the refractive

index increments, *dn/dc*, for each partitioning component. Using bulk refractometry, we

determined *dn/dc* from the concentration-dependence of solution refractive index for several

representative (bio)polymers (**Fig. 2i, Extended Data Fig. 3**). In each case, we found excellent

linearity over the entire range probed. Importantly, the measured dn/dc value for BSA is

consistent with estimates from amino acid sequence [32]. This validates the use of sequence-

based dn/dc estimates in the following, particularly given the impracticality of direct

measurement for many recombinant proteins (**Supplementary Note 1**). With *dn/dc* estimated

from protein sequence, Δn measured by QPI, and c_{Dil} measured by standard analytical methods

or neglected (**Supplementary Note 3**), Eq. 2 enables calculation of condensed-phase protein

concentrations in binary systems.

Condensates of native-like proteins

To demonstrate the suitability of our method for condensates reconstituted with native full-

length proteins, we first investigated PGL-3, which is a major component of P granules in *C.*

elegans [8]. PGL-3 forms condensates *in vitro* under low salt [6]. Using QPI, (**Fig. 3a**), we found

the concentration in individual PGL3 condensates is symmetrically distributed about a mean

of 87.0 ± 0.1 mg/ml (s.e.m., N = 269) (**Fig. 3c**), approximately 1000-fold higher than that in

the coexisting dilute phase. The standard deviation of the measured population is only 1.7

mg/ml (**Fig. 3c**), yielding a low coefficient of variation (1.9 %) which reflects the high precision

of the QPI method as well as the low droplet-to-droplet variation expected near phase

equilibrium. Given the impracticality of bulk label-free measurements for this untagged full-

Fig. 3: Label-free composition measurement of multi-domain protein condensates

Fig 3: Label-free composition measurement of multi-domain protein condensates
Quantitative phase, **a**, and refractive index, **b**, images of untagged full-length PGL3
condensates acquired by QPI and ODT, respectively. Scalebar is 10 µm.   **c**, PGL3
concentration measured from individual condensates by QPI (N = 269) or ODT (N = 355). **d**,
Confocal fluorescence images of SNAP-TAF15(RBD) condensates doped with either 23 %
mEGFP-TAF15(RBD) or 12 % AlexaFluor546-SNAP-TAF15(RBD). Scale bar is 5 µm.
Intensity profiles along the dashed yellow lines are shown at right. Gray, cyan, and lavender
lines denote the average detector background, dilute-phase intensity, and condensed-phase
intensity. **e,** Comparison of SNAP-TAF15(RBD) concentrations in condensates (left) and
partition coefficients (right) measured by QPI (N = 119) or confocal fluorescence intensity
ratios of mEGFP (N = 107) or AF546 (N = 104). White circles denote medians, thick black
bars are the interquartile range, and whiskers extend 1.5x beyond the interquartile range.
Protein concentration in the dilute phase is taken to be 1.97 µM. **f**, Temperature-dependence of
protein concentration in mEGFP-TAF15(RBD) condensates (i.e. condensed-branch of binodal)
measured by QPI. Error bars are s.e.m.
216

length protein (**Supplementary Note 1**), we employed Optical Diffraction Tomography (ODT)

for comparison (**Fig. 3b,c**). ODT is a related approach recently applied to stress granules [10,35].

We find ODT to provide a comparable accuracy to QPI, though with reduced precision

(**Fig. 3c**).

We next formed condensates using constructs derived from full-length FUS and the

RNA-binding domain of TAF15, TAF15(RBD), both reported previously in [7], and measured

their composition with QPI. Interestingly, we find these condensates to be much denser than

those of PGL3 (**Table 1**), with the 34% polymer volume fraction in TAF15(RBD) condensates

comparable to that in protein crystals [36]. Taken together, these data not only reveal that protein

sequence can tune condensate composition over at least a 5-fold range, but also demonstrate

that the QPI method enables precise label-free measurements on condensates of full-length

proteins. This removes the primary practical barrier preventing study of native full-length

proteins that are available only in limited quantities but which are most physiologically

relevant.

Influence of fluorescent labels

Conventional approaches to study condensates of full-length proteins typically require

fluorescent labels, introducing two complications in phase-separating systems. First, large

GFP- or SNAP-tags may alter the same phase behavior they are used to measure by shifting

the balance of polymer-solvent interactions that drive demixing. Second, fluorophore photo-

physics may vary between the starkly different chemical environments presented by the two

phases (**Table 1**). We leveraged our label-free method to assess both of these potential effects.

In the case of PGL3, we found that fusion to an mEGFP-tag increases the protein mass

concentration in the condensed phase by 14 % (**Table 1**). As the tag increases the construct's

molecular weight by more than 14 %, this actually corresponds to a decrease in the molar

protein concentration, consistent with the tag imparting a modest solubilizing effect.

To test for environmentally-sensitive fluorescence, we used scanning confocal

microscopy to measure fluorescence in SNAP-TAF15(RBD) condensates doped with either

mEGFP- or AlexaFluor546-SNAP-tagged (AF546) constructs (**Fig. 3d**, **Methods**). The

partition coefficient $P \equiv c_{Cond}/c_{Dil}$ of ~350 obtained from mEGFP fluorescence suggests a

condensed-phase concentration of only 43 mg/ml, underestimating the 477 ± 14 mg/ml value

we measure with QPI by over 10-fold (**Fig. 3e**). This indicates that the relationship between

fluorescence intensity and concentration differs between phases, and we suspect that enhanced

quenching from the high protein concentration in the condensed phase is largely responsible

Table 1. Compositions of biomolecular condensates in vitro.

Construct	Conditions (T in °C, [salt], pH)	c_{cond} mg/ml[a]	Protein Volume Fraction[b]	Partition Coefficient	Method
PGL3	(25.0, 75 mM, 7.4)	87.0 ± 1.7	0.0649 ± 0.0013	n.d.	QPI
PGL3	(21.5, 87 mM, 7.4)	99.2 ± 5.9	0.0740 ± 0.0044	n.d.	ODT
PGL3-mEGFP	(21.5, 87 mM, 7.4)	113.2 ± 8.6	0.0844 ± 0.0064	n.d.	ODT
SNAP-TAF15(RBD)	(37.0, 100 mM, 7.4)	477.1 ± 13.7	0.3400 ± 0.0098	3850[c]	QPI
FUS-mEGFP	(21.0, 150 mM, 7.4)	337.3 ± 8.2	0.2395 ± 0.0058	860[d]	QPI

[a] Uncertainty represents standard deviation from a population of at least 100 individual condensates

[b] Fraction of the condensed phase volume occupied by protein, $\phi \equiv c_{Cond}\bar{v}$; $\bar{v} \approx 0.75$ mL/g for PGL3 constructs and $\bar{v} \approx 0.71$ mL/g for TAF15 and FUS constructs

[c] $c_{Dil} = 1.97 \pm 0.09$ µM, $M_w = 62.92$ kDa; [7]

[d] $c_{Dil} = 4.87 \pm 0.48$ µM, $M_w = 80.38$ kDa; [7]

for the decreased quantum yield we infer there [37]. Surprisingly, assessing partitioning by

fluorescence of the more solvent-accessible AlexaFluor546-labeled construct underestimates

the concentration by 3.6-fold (**Fig. 3e**). The differential sensitivity we see with different

fluorophores suggests that brightness may vary for each dye/condensate pair and be

challenging to correct for *a priori*. Taken together, these data demonstrate that fluorescent

labels compromise condensate composition measurements in two distinct ways, sometimes

dramatically, underscoring the importance of label-free approaches like QPI.

Temperature-dependent phase behavior

Biomolecular condensates are intrinsically temperature-dependent as thermodynamic phases,

making temperature an important control parameter subject to evolutionary selection [38]. To test

whether we could detect temperature-induced composition variation with QPI, we analyzed

phase images of TAF15(RBD) condensates acquired at temperatures between 5 and 50 ºC set

with a custom temperature stage (**Methods**, [39]). After accounting for the temperature-

dependence of optical constants in Eq. 1 (**Extended Data Fig. 4**), we find that the condensed-

phase protein concentration decreases significantly from ~ 460 to ~ 390 mg/ml with increasing

temperature (**Fig. 3f**). This is indicative of an upper-critical solution temperature (compare to

Fig. 1a), as has been reported for several other [23,40], though not all [38], RNA-binding proteins.

Finally, the dry objective lenses used for QPI enable fast temperature equilibration by avoiding

direct coupling of their thermal mass to the temperature stage.

Complex aging dynamics in binary systems

Motivated by recent work demonstrating that the mechanical properties of many protein

condensates undergo an aging process [41–43,26], we hypothesized that there may be a

corresponding change in composition as condensates age. To this end, we used QPI to measure

the composition of individual PGL3 condensates over 20 hours (**Fig. 4**). During this time

period, we observed droplets to noticeably shrink (**Fig. 4a, top, Supplementary Video 1**), as

previously shown [26]. While the shrinkage would be apparent by simple brightfield imaging,

QPI indicated that the optical phase shift also increased with time, despite the reduction in

droplet size (**Fig. 4a, bottom**). By fitting the QPI data as before, we were able to precisely

measure the composition (**Fig. 4b**) and volume (**Fig. 4c**) of individual condensates over this

timeframe, revealing surprisingly coordinated dynamics. From these concentration and volume

data, we calculated the number of proteins in the condensate (**Fig. 4d**) and found that the nearly

2-fold concentration increase was approximately balanced by a volume decrease, such that the

Fig. 4: QPI reveals complex aging dynamics in binary systems

a Untagged PGL3

Fig 4: QPI reveals complex aging dynamics in binary systems
a, Timelapse of untagged PGL3 condensate (top, QPI amplitude; bottom, QPI phase). Scale
bar 10 µm. Concentration **b**, volume **c**, and number of protein molecules in the condensate **d**
for the example in **a**. Time is relative to induction of phase separation. **e**, Time-dependence of
concentration is identical for N = 25 differently sized condensates. **f**, Normalized volume varies
continuously with initial condensate size R_0. Initial shrinkage rate decreases with increasing
R_0. **g**, Fraction of molecules in 25 individual condensates over time. Thick black lines in **f**,**g**
show volume and molecule count dynamics if protein mass were conserved inside condensates.
301

total number of protein molecules in the condensate decreased by only 15 %. These

observations indicate that the condensate necessarily expelled a significant amount of solvent

while aging.

To quantify whether this near cancellation was serendipitous for this particular

condensate, we analyzed the dynamics of 24 additional condensates with a range of initial sizes

over the same period (**Fig. 4e-g**). Strikingly, we found that the kinetics and extent of

concentration increase were identical for all condensates, independent of size (**Fig. 4e**). In

contrast, the kinetics and extent of the volume decrease both showed systematic size

dependencies, with smaller condensates losing volume faster and to greater extent than larger

condensates (**Fig. 4f**). As a result, the fraction of molecules retained shows a marked

dependence on condensate size, with larger condensates retaining more molecules (**Fig. 4g**).

We speculate that the size dependence in the volume kinetics may stem from Ostwald ripening

operating in parallel with an additional as yet unknown process driving the contraction and

water expulsion. Taken together, these data demonstrate the suitability of QPI to monitor the

composition of many individual droplets in parallel, providing insight into the complex

interplay between the physical processes driving ripening and aging.

Measuring binodals and tie-lines with an arbitrary number of components

As condensates in cells are typically enriched in several distinct biomolecular species, we next

asked whether QPI could specify the composition of reconstituted condensates with multiple

components. In the linear approximation (**Extended Data Fig. 5**), the refractive index

difference between two phases with N solutes is

324                    $$\Delta n \approx \sum_{i=1}^{N} \frac{dn}{dc_i} \Delta c_i,$$ (Eq. 3)

where Δc_i is the concentration difference of the i^{th} component between the two phases. While

knowledge of Δn constrains Δc to a single value in the binary systems studied above, the

challenge for systems with multiple solutes is that Δn constrains the Δc_i only to an (N-1)-

dimensional manifold populated by compositions of equal refractive index. For a ternary

system with $N = 2$ solutes and 1 solvent, for instance, this manifold can be visualized as a line

in the (c_1, c_2)-plane (**Fig. 5a**). As all compositions along this isorefractive line are compatible

with the measured Δn, additional relationships between the Δc_i are required to uniquely specify

condensate composition.

Here, we take advantage of the fact that a tie-line connects the total system composition

averaged over both phases, $(\bar{c}_1, \bar{c}_2, \ldots, \bar{c}_N) \equiv \bar{c}$, to the compositions of the coexisting phases c^I

Fig. 5: Composition determination of multi-component biomolecular condensates

Fig 5: Composition determination of multi-component biomolecular condensates
a, Top: Schematic of multi-component measurement approach illustrated for a model ternary
system. Condensate composition (p^{II}, r^{II}) is determined by the intersection of the tie-line with
the line of constant refractive index (isorefractive line). Bottom: Schematic of coexisting multi-
component phases. **b**, Workflow to obtain all parameters in the system of linear equations. i)
Refractive index increments are determined once for each partitioning component. ii) For each
tested composition $(\bar{p}, \bar{r})$ in the two-phase regime, Δn is measured with QPI and the dilute
phase composition (p^{I}, r^{I}) is inferred from decomposition of UV-VIS absorbance spectra. The
points $(\bar{p}, \bar{r})$ and (p^{I}, r^{I}) suffice to define the tie-line, whose slope m_{TL} is the final required
parameter to solve the linear system, iii). **c**, QPI images of FUS/RNA condensates for a range
of RNA concentrations $\bar{r}$. Scalebar is 3 µm. **d**, Measured Δn distributions for the samples in **c**.
e, Experimentally-determined ternary phase diagram. Inset shows the same data on loglog axes.
Error bars denote standard deviation for dilute-phase and average concentrations, and error
propagation via Jacobian for condensed-phase concentrations. **f**, Polymer volume fractions in
the condensed phase as a function of $\bar{r}$ with $\bar{p} = 2$ mg/ml. Error bars correspond to the
uncertainty in condensed-phase concentration rescaled by polymer partial specific volumes.
352

and c^{II} located on the (N-1)-dimensional dilute and condensed binodal manifolds,

respectively. For ternary systems, the binodals can be visualized as bounded curves in the

(c_1, c_2)-plane (**Fig. 5a**). Mass conservation guarantees that tie-lines are straight (**Fig. 5a**,

Supplementary Note 4). Crucially, this provides N-1 linearly independent constraints, which

can be seen by noting that projections of the tie-line in each of the N-1 $(c_1, c_{i \neq 1})$-planes must

all be straight. In principle, tie-line constraints of the form $c_i = m_{1i}c_1 + b_{1i}$ where $m_{1i} =$

$(\bar{c}_i - c_i^I)/(\bar{c}_1 - c_1^I)$ and $b_{1i} = \bar{c}_i - m_{1i}\bar{c}_1$ for $i \neq 1$ can be obtained from knowledge of overall

sample conditions, $\bar{c}$, and composition measurements of the abundant dilute-phase c_i^I using

traditional approaches from analytic chemistry (**Supplementary Notes 3,5**). Combined with

Eq. 3, these form a linear system of N equations of the form $Mc^{II} = x$, where the matrix M and

vector x contain measured quantities (see **Fig. 5b**, **Supplementary Note 5** for explicit

expressions for ternary and (N+1)-component systems, respectively). By solving this system

of equations, the composition of the multi-component condensate is written in terms of known

optical quantities and concentrations as

$$c_i^{II} = \Delta n M_{i1}^{-1} + c_i^I, \tag{Eq. 4}$$

368   where M_{i1}^{-1} are elements of the system's matrix inverse (**Supplementary Note 5**). Thus, by

369   leveraging knowledge of tie-lines to provide the missing the constraints, this method is, in

370   principle, capable of resolving the composition of biomolecular condensates with an arbitrary

371   number of components.

Phase diagram for ternary system with full-length FUS protein and RNA

To validate this approach, we used poly(A) RNA and full length FUS-mEGFP, which localizes

to RNA-rich stress granules in eukaryotes [10], as solutes in a model ternary system (N = 2).

Following the workflow in **Fig. 5b**, we prepared systems in the two-phase region for a range

of total RNA concentrations. To obtain the dilute-phase composition, we decomposed UV

absorbance spectra of the dilute phase into protein and RNA contributions (**Fig. 5b, Methods**).

Combined with the system average compositions, the decomposed spectral data produce a set

of physically compatible tie-lines (**Extended Data Fig. 6**). To obtain Δn, we analyzed QPI

data (**Fig. 5c**), finding that Δn decreases and saturates with increasing total RNA (**Fig. 5d**).

Using these data as inputs in Eq. 4 (**Fig. 5b**), we calculated the corresponding points on the

condensed binodal branch and plotted these together with the dilute binodal and tie-lines as a

full phase diagram (**Fig. 5e**). Reassuringly, this phase diagram captures the re-entrant behavior

with increasing RNA inferred previously in related systems without actually measuring the

phase boundary [44,45]. In addition to connecting variation in system average composition directly

to its consequences on the condensate, which was seldom accessible previously [21], this phase

diagram exposes surprising features in the condensed binodal branch, including a kink and

linear section (**Fig. 5e**). The ability of this method to resolve the molecular composition of

multi-component condensates reveals that rising RNA concentrations are compensated by

decreasing protein concentrations such that a constant polymer volume fraction is maintained

near cytoplasmic levels (**Fig. 5f**), which would have been challenging to infer using

conventional techniques.

DISCUSSION

Our results show that the present QPI-based method enables quantitative composition

measurements of biomolecular condensates reconstituted with full-length proteins in binary

and multi-component mixtures. Our finding that a common fluorescent tag can significantly

alter condensate composition (**Table 1**) highlights the need for label-free measurement

approaches like QPI. As a microscopic method, QPI offers several advantages over traditional

label-free strategies based on bulk measurements. The first and most dramatic of these is the

~1000-fold reduction in sample requirements, which removes the primary barrier to

measurements on condensates reconstituted from native full-length proteins (**Fig. 3a,b**).

Second, whereas time-resolved measurements of composition in response to changing solution

conditions, temperature, or intrinsic sample dynamics are challenging with bulk approaches,

QPI's compatibility with open samples and dry objectives makes it easy to record sample

dynamics with second-scale time resolution (**Fig. 4**) and in response to environmental changes

(**Fig. 3f**). Third, measuring composition of micron-sized condensates *in situ* provides access to

information like size-dependent composition (**Fig. 2g**), homogeneity (**Fig. 2f**), and dynamics

(**Fig. 4e-g**) that are not available with bulk techniques. Finally, we note that the high precision

demonstrated here with QPI not only enables the method to resolve subtle compositional

differences (**Table 1**) and unexpected binodal features (**Fig. 5f**) with high confidence, but could

also serve to constrain competing thermodynamic models of condensate formation [22,7,46,24,47,25].

An important benefit of the present method is its ability to disentangle the composition

of multi-component biomolecular condensates. All endogenous condensates are expected to

contain multiple components, with some of the best-characterized known to house dozens [8–10].

Knowledge of component identity and quantity will be essential for design and interpretation

of increasingly faithful reconstitution experiments as well as studies of chemical reactions in

synthetic condensates that account for reacting species. Here, we validated our methodology

by measuring the full phase diagram for a ternary mixture of full-length FUS protein and RNA,

including the dilute- and condensed-binodal branches as well as tie-lines (**Fig 5f**). We note that

three pieces of information, tie-lines along with both binodals, are strictly required to physically

relate the compositions of coexisting phases to the average composition specified in

experiments, and that no other methods suitable for low-yield proteins provide all three.

Further, we proved mathematically that the proposed methodology can, in principle, resolve

the composition of condensates containing an arbitrary number of distinct molecular

components (**Supplementary Note 5**). Whether partitioning of a particular species can be

resolved is determined primarily by the sensitivity and precision of the dilute-phase detection

method employed (**Extended Data Fig. 7**). We emphasize that our methodology is agnostic to

the choice of dilute-phase detection method. This flexibility allows experimenters to select or

combine established analytic approaches best-suited for the molecules used. In particular,

mass-spectrometry can resolve complex mixtures with 100s or 1000s of components and

distinguish between protein some PTM-states [48].

In future applications of this method, there are three classes of practical considerations

that must be kept in mind. The first concerns condensate size. Though QPI measures refractive

index robustly for droplets over the size-range most commonly encountered in reconstitution

experiments, from a few to a few 10s of microns, systematic errors are incurred due to

scattering and gravitational settling for sufficiently small and large droplets, respectively. We

anticipate that additional development of fitting routines to account for these physical effects

can further extend the size range for reliable measurements [49,50]. The second consideration

involves treating the refractive index of a mixture as a linear sum of contributions from its

components (see **Eq. 3**, **Supplementary Note 2**). This is a very good approximation for protein

solutions over a wide range [30], though higher-order terms could potentially contribute for some

molecules [51]. In the cases we checked, we found the linear sum to describe the refractive index

of multi-component mixtures accurately to within a few % (**Extended Data Fig. 5**). We also

note that while contributions from PTMs are currently neglected when estimating protein

refractive index increments [32], which are the coefficients in the linear sum, this could be readily

addressed in the future with molar refractivity measurements of modified amino acids in

solution [52]. This could improve accuracy for condensates containing highly-modified proteins.

The third consideration is in regard to measurements in cells. Though our QPI-based method

is best-suited for reconstituted condensates, QPI and related techniques like ODT can provide

some information on cellular condensates whose refractive index differs sufficiently from the

surrounding material [53,31,35]. For condensates with sufficient contrast, reconstruction of

binodals and tie-lines would require knowledge of molecular abundances in individual cells as

well as the coexisting phase. Absent this, phase-based imaging data can still provide a

quantitative measure of the average macromolecular mass difference between phases *in vivo*,

which may be particularly informative in the context of perturbations.

As composition ultimately influences all other condensate properties and associated

cellular functions, we anticipate a central role for this method in addressing many pressing

biological questions. Measuring condensate composition as the abundance of individual

components are systematically varied will reveal the thermodynamic contributions of these

molecules to the phase, potentially clarifying the biological function of individual components.

We anticipate that correlating composition with other physical properties like viscoelasticity,

interfacial tension, and dielectric constant will likewise provide insight into condensate

function. By providing a ground-truth with which to calibrate fluorophore behavior, we expect

that this method will enable the use of dyes to both follow reactions localized to condensates

as well as quantitatively probe the chemical environment within condensates. The latter will

likely be essential to understand and potentially tune the partitioning of therapeutic drugs into

condensates in treatment of diseases like cancer [54].

MATERIALS AND METHODS

Sample preparation. Recombinant protein constructs used in this work were purified and

stored as described previously [6,7]. To induce phase separation, we mixed protein in high-salt

storage buffer (300 mM KCl for PGL3 constructs, 500 mM KCl for TAF15 and FUS

constructs) with storage buffer lacking monovalent salt, "Dilution Buffer", to reach the desired

final salt concentration. Generally, an aliquot of Dilution Buffer was supplemented to 1 mM

with fresh DTT prior to each day's experiments. After induction of phase separation, dilute

phase was obtained by centrifugation at RCF = 20,800 x g for 30 min in a tabletop centrifuge

(Centrifuge 5417 R, Eppendorf) pre-equilibrated at the desired temperature. For control

measurements, 10-µm silica microspheres were purchased from Whitehouse Scientific

(Waverton, UK), and glycerol-water-mixtures were prepared by weight to the desired refractive

index. Bead-containing dispersions were prepared by gently dipping a 10-µL pipette tip into a

stock of dry beads, transferring the pipette tip to a 40 µL volume of glycerol-water mixture,

and pipette mixing to disperse. Aqueous two-phase systems with PEG-35k (Sigma) and

Dextran T500 (Pharmocosmos) were prepared as described previously [34]. BSA was purchased

from Sigma and used without further purification.

Quantitative phase imaging and analysis. QPI measurements were performed using a

coherence-controlled digital holographic microscope (Q-Phase, Telight (formerly TESCAN),

Brno, CZ) based on the set-up in [28]. Most data were acquired on a Generation-1 instrument

with a tungsten-halogen bulb as lightsource, though some data were acquired on a Generation-

2 instrument with a 660-nm LED as lightsource. In each case, the holography lightsource was

filtered by a 10-nm bandwidth notch filter centered at 650 nm. All measurements were

performed with 40x dry objectives (0.9 NA, Nikon) except those for SNAP-TAF15(RBD)

reported in **Fig. 3f**, for which 20x dry objectives were used. In all cases, the condenser aperture

was set to an NA of 0.30. Immediately following phase separation, ~ 5 µL of sample was

loaded into a temperature-controlled flowcell, sealed with two-component silicone glue

Twinsil (Picodent, Wipperfürth, DE), and allowed to settle under gravity for ~ 10 minutes prior

to data collection. Flowcells were constructed with a 30x24x0.17 mm^3 PEGylated coverslip

and a 75x25x1 mm^3 sapphire slide as bottom and top surfaces, respectively, using parafilm

strips as spacers. Proportional-integral-derivative (PID)-controlled Peltier elements affixed to

the sapphire slide enabled regulation of flowcell temperature, as previously described [39]. The

sapphire, coverslip, and spacers were adhered by heating the assembled flowcell to 50 °C for

2-5 min, then returning to the desired temperature for the first measurement, typically to 20 °C.

For each sample, hologram z-stacks ($dz = 0.2$ µm, first plane typically near the coverglass

surface) were acquired for several fields of view. SophiQ software (Telight, Brno, CZ) was

used to construct amplitude and compensated phase images from the raw holograms. Pixels in

the phase images are 0.157 µm per side for the 40x, and pixel intensities are in units of radians.

To aid interpretation by persons with red/green color perception deficiencies, phase images are

displayed using the Ametrine colormap [55].

All phase images were analyzed in MATLAB using custom code. For each z-plane,

compensated phase images were first segmented to identify individual droplets. To determine

the background phase value, φ_0, the image's pixel intensity histogram was fit to a Gaussian,

and the Gaussian center taken as φ_0. Pixel intensities $\varphi \geq n_{sig}\sigma_\varphi$ are considered above

threshold, where σ_φ is the standard deviation extracted from the Gaussian fit. Typically,

$n_{sig} = 5$. A binary mask was generated with this threshold and individual objects were identified

using the MATLAB function bwconncomp.m. For each object, a region of interest slightly

larger than the object's bounding box was fit twice to phase functions of the form given by Eq.

(1) in the main text. First, we fit using the projected height of a sphere,

$H_{sphere}(x, y|R, x_c, y_c) = \Theta(R^2 - (x - x_c)^2 - (y - y_c)^2)\sqrt{R^2 - (x - x_c)^2 - (y - y_c)^2},$

in order to obtain estimates for the parameters $\Delta n, R, x_c, y_c$, where $\Theta(x)$ is the Heaviside

function. These estimates were then used to initialize a fit to a regularized version of Eq. (1),

$$\Delta\varphi_{Reg}(x, y) = \frac{2\pi}{\lambda}\Delta n H_{cap}(x, y|R, x_c, y_c, Z_{eq}) + \varphi_0 + A(Z_{eq}, R),$$

using the projected height of a spherical cap,

$H_{cap}(x, y|R, x_c, y_c, Z_{eq}) = \sqrt{R^2 - (x - x_c)^2 - (y - y_c)^2}\left(1 + \Theta(Z_{eq}^2 + (x - x_c)^2 + (y - $

$y_c)^2 - R^2)\right)\Theta(R^2 - (x - x_c)^2 - (y - y_c)^2) + Z_{eq}\Theta(R^2 - Z_{eq}^2 - (x - x_c)^2 - (y - y_c)^2).$

The regularization is given by $A(Z_{eq}, R) = A_0(Z_{eq} - R)^2\Theta(Z_{eq} - R)$ with $A_0 = 10^5$ and φ_0

fixed at the value obtained from the pixel intensity histogram fit. After all z-planes were

processed,   the   objects   were   tracked   through z using   track.m

(https://site.physics.georgetown.edu/matlab/index.html).   For   each   tracked   object,   the

representative fit parameters are taken as those from the fit with the highest Adj. R^2 value,

which are typically in the plane acquired nearest to the equatorial plane of a given droplet. The

particle list was then automatically filtered for fit quality (typically retaining only Adj. R^2-

values > 0.95) and overlap with dead-pixels on the detector. Droplets with irregular wetting or

that are not isolated in z (i.e. are situated beneath other droplets in solution) are removed

manually following inspection of the raw data.

Optical Diffraction Tomography. ODT measurements were performed using a custom-built

microscope employing a 532-nm laser, as described previously [56]. Tomogram reconstruction

and image analysis was performed as described previously [57,10].

Confocal Fluorescence Microscopy and Analysis. Confocal imaging was performed on an

inverted Zeiss LSM880 point-scanning confocal microscope with a 40x water-immersion

objective (1.2 NA, C-Apochromat, Zeiss) at room temperature. mEGFP was excited with a

488-nm argon laser and emission detected with a 32-channel GaAsP photomultiplier tube

(PMT) array set to accept photon wavelengths between 499 and 569 nm. AlexaFluor546 was

excited with a 561-nm diode-pumped solid-state laser and emission detected between 570 and

624 nm with the same spectral PMT array. For both fluorophores, the confocal pinhole

diameter was set to 39.4 µm, corresponding to 0.87 and 0.96 Airy Units for mEGFP and

AlexaFluor546, respectively. For each field of view, scanning was performed with a lateral

pixel size of 0.415 µm and z-stacks acquired with a spacing of 0.482 µm.

All confocal fluorescence images were analyzed in MATLAB using custom code.

Partition coefficients of fluorescently-labeled species into condensates are estimated on the

basis of the fluorescence intensity along a line-scan through the droplet center. The analysis

pipeline begins with determining the location of each condensate and an appropriate line-scan

orientation angle. To determine lateral positions of condensates in each field of view, a z-plane

was selected slightly above the coverglass such that even small droplets appeared bright.

Following convolution with a 2D Gaussian ($\sigma_x = \sigma_y = 0.5$ pixels) to suppress shot noise, a

threshold of $I_{thresh} = \max(I(x,y))/2$ was applied to obtain a binary mask. The lateral

positions and approximate sizes of objects were determined from the mask with

bwconncomp.m. Only the largest ~120 objects for each condition were analyzed further. For

each object, partition coefficients were calculated using the z-plane for which the mean

intensity in a 5-pixel-radius disk concentric with the object was greatest. Line-scans were 51

pixels long, concentric with the object, and averaged over a width of 3 pixels. Suitable line-

scan orientations were determined in a semi-automated manner by superimposing reference

lines rotated through 15° increments on each object and manually selecting an orientation that

best avoided neighboring objects. Objects for which no suitable line-scan orientation could be

found were discarded.

Line-scans for each droplet were automatically subdivided into three domains,

corresponding to pixels in the dilute phase, the condensed phase, or the exclusion zone. The

positions of the left and right dilute/condensed interfaces are estimated as those at which the

intensity profile reaches its half-maximal value above detector background I_{Bkgd} (see below).

To reduce artefacts stemming from the finite point-spread function of the microscope, pixels

within the greater of 1 pixel or $l_{EZ} = 1.22\lambda/(2NA)$ on either size of the half-maximum were

excluded from the analysis. Remaining profile pixels outside the droplet were averaged to give

I_{Dil}, while profile pixels inside are averaged to give I_{Cond}. The partition coefficient for each

object was calculated according to

575
$$P = \frac{I_{Cond} - I_{Bkgd}}{I_{Dil} - I_{Bkgd}},$$

where I_{Bkgd} is the average of all pixels in a background image acquired immediately following

the fluorescence z-stack. Background images were acquired with the light source blocked to

measure the contribution of detector noise to the signal.

Bulk Refractometry. Data in Fig. 2i were acquired at λ = 653.3 nm and 21 °C with a DSR-L

multi-wavelength refractometer (Schmidt + Haensch, Berlin) using a 200-µL sample volume.

All other bulk refractive index measurements were acquired at λ = 589.3 nm with a J457

refractometer (Rudolph Research Analytical, Hackettstown, NJ). The refractive index of

glycerol/water mixtures was adjusted from $\lambda = 589.3$ to 650 nm using empirical dispersion

relations for distilled water, $n_{Water}(\lambda)$ [58], and glycerol, $n_{Glycerol}(\lambda)$ [59] according to

$$n_{Mix}(\lambda) = w_{Glycerol}n_{Glycerol}(\lambda) + (1 - w_{Glycerol})n_{Water}(\lambda).$$

The glycerol weight fraction in the mixture was calculated from the refractive index

measurement of the mixture at 589.3 nm as

$$w_{Glycerol} = (n_{Mix}(\lambda) - n_{Water}(\lambda))/(n_{Glycerol}(\lambda) - n_{Water}(\lambda))\Big|_{\lambda=589.3}.$$

Bead porosity models. The two models used to account for the porosity p of the silica

microspheres (Fig. 2h, Extended Data Fig. 2) are a weighted linear sum, $\Delta n = pn_{Silica} +$

$(1 - p)n_{Water}$ (simple model), and a more detailed model,

$$\Delta n = \left(\frac{1 + 2(1-p)f(n_{Silica}) + 2pf(n_{Mix})}{1 - (1-p)f(n_{Silica}) + pf(n_{Mix})}\right)^{1/2} - n_{Mix}$$

based on the Lorentz-Lorenz relation [60], wherein

$$p = \frac{f(n_{Silica}) - f(n_{Microsphere})}{f(n_{Silica}) - f(n_{GlycerolWaterMixture})}$$

with $f(n) \equiv (n^2 - 1)/(n^2 + 2)$.

Calculation of dn/dc, $\bar{v}$ and polymer volume fraction. The refractive index increment and

partial specific volume were estimated for each protein construct using the calculator tool

within SEDFIT [32] and the protein sequences listed in the SI. The partial specific volume of

0.5773 mL/g for poly(A) RNA was estimated using consensus volumes per base from [61] and

assuming a typical chain length of 500 bases. The partial specific volumes of 0.8321 and 0.6374

mL/g for PEG-35k and Dextran-500k, respectively, were taken from [34]. Polymer volume

fraction (Fig. 5f) for polymer i is given by $\phi_i = c_{Cond}\bar{v}$.

Calculations for aging systems. Volumes of individual aging condensates (Fig. 4c,f) were

calculated as $V = \frac{4}{3}\pi R^3$. The number of molecules in each condensate (Fig. 4d,g) were

calculated as $N(t) = c(t)V(t)$. Given $c(t)$, the relative volume change expected if $N(t) =$

$N(0)$ is given by $V(t)/V(0) = c(0)/c(t)$ (Fig. 4f, black line).

UV-Vis spectroscopy. Absorption spectra of dilute-phase and reference samples for the

FUS/RNA system (Fig. 5) were collected on an NP-80 spectrophotometer (IMPLEN,

München). All spectra were acquired at room temperature over $\lambda \in [200\text{ nm}, 900\text{ nm}]$. For

each raw spectra $\tilde{S}(\lambda)$, a linear fit on $\lambda \in [550\text{ nm}, 750\text{ nm}]$ was used to determine a baseline

correction, $S_{BL}(\lambda)$. Corrected spectra are given by $S(\lambda) = \tilde{S}(\lambda) - S_{BL}(\lambda)$. At least 3 replicate

spectra were acquired for each condition and averaged following baseline correction to give

the final representative spectra. The uncertainty in the spectra at each wavelength was

estimated as the standard deviation of the corrected replicates. Dilute-phase spectra were

demixed (Fig. 5b) into a weighted sum of three contributions

616
$$S_{Dil}(\lambda) = p^I S_p(\lambda) + r^I S_r(\lambda) + p^I r^I a_{int} S_{int}(\lambda),$$

where (p^I, r^I) are the protein and RNA concentrations in the dilute phase. S_p and S_r are

reference spectra for protein and RNA, respectively. The final term captures the effect of

protein-RNA interactions on the absorbance of a mixture, which could physically stem from

binding-induced changes in extinction coefficients. Reference spectra for the interaction was

calculated from spectra of a protein/RNA mixture of known composition (p, r) in the 1-phase

regime according to $S_{int} = S(p, r) - pS_p - rS_r$. The parameter a_{int} captures the

approximately linear increase of S_{int} with p. The same value of a_{int} was used to demix all

dilute-phase spectra. The dilute-phase concentrations p^I, r^I were therefore the only free

parameters for the demixing.

ACKNOWLEDGEMENTS

We thank E. Filippidi, T. Franzmann, B. Diederich, T. Harmon, L. Jawerth, A.W. Fritsche, and J.M. Iglesias-Artola, as well as members of the Brugués, Hyman, Alberti, Jülicher, and Weber groups for helpful discussions. We thank T. Slabý, M. Šicner, and V. Procházka of Telight for support with the Q-Phase microscope and E. Bittrich and R. Schulze of the Leibniz Institute for Polymer Research for assistance with multi-wavelength refractometry measurements. We thank A. Poznyakovskiy for creating several constructs for this project, J. Wang for purification of TAF15 protein constructs, R. Lemaitre for purification of full-length FUS protein, D. Kuster for the gift of poly(A) RNA, and A.W. Fritsche for the gift of silica beads. We thank A. Schwager and S. Ernst for preparation of passivated coverglass, L. Jawerth for assistance with aging PGL3 samples, and A.W. Fritsche for assistance with the temperature stage. We also thank the MPI-CBG Core Facilities, particularly the Light Microscopy Facility and the Protein Expression and Purification Core, for their support. PMM was supported by an ELBE Postdoctoral Fellowship from the CSBD. PMM and JB were also supported by Volkswagen 'Life' grant number 96827.

AUTHOR CONTRIBUTIONS

PMM, JP, AAH, and JB conceived the project. PMM conceived and developed the multi-component measurement approach and derived the analytical solution. PMM designed and performed QPI, refractometry, spectroscopy, and fluorescence microscopy measurements, developed the analysis pipelines for these data types, and performed the formal analysis. KK and PMM designed and performed ODT measurements. KK analyzed ODT data. MR-G purified multiple constructs and articulated essential distinctions related to the handling of IDRs and multi-domain proteins. JB, AAH, and JG supervised the work. PMM and JB wrote the paper with input from all authors.

COMPETING INTERESTS

AAH is a founder of Dewpoint Therapeutics and a member of the board as well as a shareholder in Caraway Therapeutics. All other authors have no competing interests.

DATA AVAILABILITY

Data generated and analyzed supporting the findings of this Article will be made available by the Authors upon reasonable request.

CODE AVAILABILITY

Custom code generated supporting the findings of this manuscript will be made available by the Authors upon reasonable request.

SUPPLEMENTARY INFORMATION

- Note 1: Requirements of bulk measurements
- Note 2: Linear approximation to refractive index of mixtures
- Note 3: Measurement of c_{Dil}
- Note 4: Requirement that tie-lines are straight
- Note 5: Derivation of condensate composition for multi-component systems
- Supplementary References
- Table S1: Protein constructs used
- List S1: List of protein sequences used

675   **EXTENDED DATA FIGURES AND CAPTIONS**
676

Extended Data Fig. 1: Estimate of capillary length for biomolecular condensates
a, Capillary length as a function of the condensed-phase polymer mass concentration. The
capillary length increases with increasing interfacial tension, so we plot traces for different
values of interfacial tension. From bottom to top, interfacial tensions are 1, 2, 5, and 10 µN/m.
This spans the range of tensions reported for PGL-3 condensates by [62]. For a given droplet
density and interfacial tension, the spherical cap approximation for the droplet shape is valid
so long as the droplet size is less than the capillary length. In the case of PGL-3, our density
measurements of ~ 90 mg/ml indicate a capillary length of 67 µm at the lowest interfacial
tension, which is much larger than the 1-8 µm radii of the droplets. Although the capillary
length is reduced to ~ 30 µm for the higher ~ 400-500 mg/ml densities we measure for
TAF15(RBD) condensates, this length is still much larger than the 0.67-3.2 µm radii of those
condensates analyzed in this work.
689

Extended Data Fig. 1: Estimate of capillary length for biomolecular condensates

a

Extended Data Fig. 2: Silica bead measurements and porosity
a, (left) Refractive index difference between silica beads and the surrounding glycerol/water
mixture as a function of bead size as measured by QPI for a range of glycerol/water ratios.
Glycerol weight fraction of the medium is 0.6641 (yellow, $N = 66$), 0.808 (orange, $N = 83$),
0.8688 (purple, $N = 102$), and 0.8751 (blue, $N = 74$). Datapoints represent individual beads,
error bars are 95% confidence intervals returned from fits, and are typically smaller than the
datapoint. Though the nominal bead diameter is 10 µm, the samples show significant
polydispersity. Importantly, there is no strong systematic variation of the extracted Δn with
bead size. (right) Measured refractive index distribution is approximately symmetric for each
condition. **b**, Population mean ($\pm$ standard deviation) of Δn measured by QPI at $\lambda = 650$ nm
for silica beads as a function of the refractive index of the glycerol/water mixture, n_{medium}. Note
that n_{medium} was measured at $\lambda = 589$ nm on a digital refractometer and is plotted following
adjustment to $\lambda = 650$ nm using published dispersion relationships for glycerol and distilled
water and an approximate mixing rule, as described in Methods. Solid red line is a linear fit to
the data with the 4 largest x-values, with slope -0.6455 $\pm$ 0.01806. The x-intercept of the fit
line, at which the refractive index of the silica bead is indistinguishable from that of the
medium, is 1.461 $\pm$ 0.057. Importantly, this is comparable to published values of fused silica
under similar conditions (e.g. n = 1.4565 at 650 nm, 20 °C, [63]), independently of whether we
adjust for dispersion. However, the slope of the fit line (-0.6455 $\pm$ 0.0181) differs significantly
from the value of -1 expected for beads of pure fused silica (dashed black line). This suggests
that the beads may be porous, as has been reported previously [60]. **c**, The refractive index
difference expected for silica beads with porosity p as a function of the refractive index of the
surrounding glycerol/water mixture, assuming that the bead pores are filled with the
glycerol/water mixture. Here p is the fraction of the bead volume occupied by pores rather than
silica. Different colors represent different values of porosity from 0 to 0.4. For each p, Δn is
calculated using two different refractive index mixing rules, either a weighted linear
combination (approx., dashed line) or the Lorentz-Lorenz relation (full model, solid line,
Methods). Both models predict that fluid-filled pores in the silica beads would give a Δn that
decreases (nearly) linearly with slope (1-p) as n_{medium}. increases. The two models are almost
indistinguishable except at low n_{medium} and high p, where the simple approximation noticeably
overestimates Δn. **d**, Same data as in **b** above, but now overlaid with model predictions for
silica beads with $p = 0.3545$. All measurements at 25 °C.
723

Extended Data Fig. 2: Silica bead measurements and porosity

a

b

c

d

Extended Data Fig. 3: Measurement of dn/dc for poly(A) RNA at 589 nm
a, Refractive index of poly(A) RNA in water as a function of RNA concentration.
Measurements were performed with a digital refractometer at 589 nm at 20 °C. Error bars are
the larger of the standard deviation of $N = 5$ measurements or the instrument resolution 0.00001
(if all repeat measurements were identical). The refractive index increment is given by a linear
fit as 0.1655 ± 0.0045 ml/g.
731

Extended Data Fig. 3: Measurement of dn/dc for polyA RNA at λ = 589 nm

a

Extended Data Fig. 4: Temperature-dependent optical constants
a, Refractive index of buffer as measured with a digital refractometer at 589 nm at the
temperatures used in Fig. 3f. Error bars are the larger of the standard deviation of $N = 5$
measurements or the instrument resolution 0.00001 (if all repeat measurements were identical).
The refractive index of the solution decreases with increasing temperature primarily due to the
reduced electron density accompanying thermal expansion. Note that the reduction in refractive
index due to this effect, $n_{buffer}(6\ °C) - n_{buffer}(50\ °C) = 0.0051$, is much smaller than the variation
of Δn we measure for TAF15(RBD) condensates over the same range, $\Delta n(6\ °C) - \Delta n(50\ °C) =$
0.0144. **b**, Refractive index increment for SNAP-TAF15(RBD) estimated using SEDFIT
software [32] as a function of temperature. These values were used in Fig. 3f to convert the Δn
measured by QPI to the c_{cond} values reported at each temperature.
744

Extended Data Fig. 4: Temperature-dependent optical constants

a

b

Extended Data Fig. 5: Validation of linear sum approximation for refractive index of
mixtures
a, Refractive index difference between either homogeneous PEG/Dextran mixtures prepared
in the one-phase regime and water (blue and red) or between coexisting phases in the two-
phase regime (yellow and purple). Average system composition is specified below each bar as
weight fractions of PEG-35k and Dextran-500k. Blue bars show the refractive index measured
experimentally on a digital refractometer at 589 nm, while red bars show the refractive index
predicted for the same compositions using a linear sum approximation for the mixing rule (see
Methods). Yellow bar shows the difference between the refractive indices of the coexisting
phases, each measured individually at 589 nm on a digital refractometer. Purple bar shows the
population mean refractive index difference between the same phases measured via QPI for N
= 205 droplets. Error bars for digital refractometry measurements are the larger of the standard
deviation of N = 5 measurements or the instrument resolution 0.00001 (if all repeat
measurements were identical). Error bars for the predictions are estimated at 0.00001. Error
bar for the QPI measurement is the standard deviation of the measured Δn distribution. All
measurements were performed at 24 °C. **b**, Relative error of the paired measurements in **a**. The
relative error incurred by application of the linear sum mixing rule is typically much less than
5 %. For comparison, natural variation in optical properties with wavelength (dispersion)
results in an 8% change in Δn between 589 nm and 650 nm.
765

Extended Data Fig. 5: Validation of linear sum approximation for refractive index of mixtures

Extended Data Fig. 6: Examples of physically permissible tie-lines
a, Examples of ternary phase diagrams with varying tie-line orientations. While tie-lines can
in principle diverge (left), be parallel (center-left), or converge (center-right), it is unphysical
and thus forbidden for them to cross within the multi-phase coexistence region (right). **b**,
Ternary phase diagram for FUS/poly(A) RNA on linear scales zoomed in on the dilute
binodal. **c**, Slopes of the tie-lines shown in **b**. Relative to the classification scheme introduced
in **a**, this system displays diverging tie-lines.
774

Extended Data Fig. 6: Examples of physically permissible tie-lines

a

Diverging tie-lines: **allowed**   Parallel tie-lines: **allowed**   Converging tie-lines: **allowed**   Crossing tie-lines: **forbidden**

b

Legend:
- ● Dilute Binodal (Set 1)
- ○ Avg Composition (Set 1)
- ■ Dilute Binodal (Set 2)
- □ Avg Composition (Set 2)
- — Tie-lines
- - - Approximate binodal line

x-axis: [FUS-GFP] (μM)
y-axis: [poly(A) RNA] (mg/mL)

c

x-axis: [poly(A) RNA]$_{avg}$ (mg/ml)
y-axis: Tieline slope (mg/mL/μM)

Extended Data Fig. 7: Multi-component detection limits
a, The minimum partition coefficient P_{min} for which the concentration of a species in the dilute
phase c^I is measurably different from its total average concentration in the system $\bar{c}$ depends
on the relative precision $\delta c/c$ of the dilute-phase detection strategy. For a given $\delta c/c$, P_{min}
decreases as the volume fraction of the condensed phase increases. In this diagram, curves
show the minimum partition coefficient for which partitioning of an arbitrary species would be
detectable in a system as a function of condensed-phase volume fraction for $\delta c/c = 1\%$ (solid
black), 10% (blue dashed), and 100 % (red dotted). The shading denotes that all partition
coefficients above the corresponding lower-limit are detectable.
785

Extended Data Fig. 7: Multi-component detection limits

a

787   **Supplementary Video 1**
788   Quantitative phase imaging of aging timecourse for reconstituted PGL3 condensates.
789

REFERENCES

1. Hyman, A. A., Weber, C. A. & Julicher, F. Liquid-Liquid Phase Separation in Biology. *Annu.*

Rev. Cell Dev. Biol. **30**, 39–58 (2014).

2. Banani, S. F., Lee, H. O., Hyman, A. A. & Rosen, M. K. Biomolecular condensates:

organizers of cellular biochemistry. *Nat. Rev. Mol. Cell Biol.* **18**, 285–298 (2017).

3. Alberti, S., Gladfelter, A. & Mittag, T. Considerations and Challenges in Studying Liquid-

Liquid Phase Separation and Biomolecular Condensates. *Cell* **176**, 419–434 (2019).

4. Elbaum-Garfinkle, S. *et al.* The disordered P granule protein LAF-1 drives phase

separation into droplets with tunable viscosity and dynamics. *Proc. Natl. Acad. Sci.* **112**,

7189–7194 (2015).

5. Nott, T. J. *et al.* Phase Transition of a Disordered Nuage Protein Generates

Environmentally Responsive Membraneless Organelles. *Mol. Cell* **57**, 936–947 (2015).

6. Saha, S. *et al.* Polar Positioning of Phase-Separated Liquid Compartments in Cells

Regulated by an mRNA Competition Mechanism. *Cell* **166**, 1572-1584.e16 (2016).

7. Wang, J. *et al.* A Molecular Grammar Governing the Driving Forces for Phase Separation

of Prion-like RNA Binding Proteins. *Cell* **174**, 688-699.e16 (2018).

8. Updike, D. & Strome, S. P Granule Assembly and Function in Caenorhabditis elegans

Germ Cells. *J. Androl.* **31**, 53–60 (2010).

9. Xing, W., Muhlrad, D., Parker, R. & Rosen, M. K. A quantitative inventory of yeast P body

proteins reveals principles of composition and specificity. *eLife* **9**, e56525 (2020).

10. Guillén-Boixet, J. *et al.* RNA-Induced Conformational Switching and Clustering of G3BP

Drive Stress Granule Assembly by Condensation. *Cell* **181**, 346-361.e17 (2020).

11. Riback, J. A. *et al.* Composition-dependent thermodynamics of intracellular phase

separation. *Nature* **581**, 209–214 (2020).

12. Kaur, T. *et al.* Sequence-encoded and composition-dependent protein-RNA interactions

control multiphasic condensate morphologies. *Nat. Commun.* **12**, 872 (2021).

13. Alshareedah, I., Thurston, G. M. & Banerjee, P. R. Quantifying viscosity and surface

tension of multicomponent protein-nucleic acid condensates. *Biophys. J.* **120**, 1161–

1169 (2021).

14. Banani, S. F. *et al.* Compositional Control of Phase-Separated Cellular Bodies. *Cell* **166**,

651–663 (2016).

15. Choi, S., Meyer, M. O., Bevilacqua, P. C. & Keating, C. D. Phase-specific RNA

accumulation and duplex thermodynamics in multiphase coacervate models for

membraneless organelles. *Nat. Chem.* 1–8 (2022) doi:10.1038/s41557-022-00980-7.

16. Morin, J. A. *et al.* Sequence-dependent surface condensation of a pioneer transcription

factor on DNA. *Nat. Phys.* 1–6 (2022) doi:10.1038/s41567-021-01462-2.

17. Lakowicz, J. R. Solvent and Environmental Effects. in *Principles of Fluorescence*

Spectroscopy 205–235 (Springer, Boston, MA, 2006). doi:10.1007/978-0-387-46312-4_6.

18. Küffner, A. M. *et al.* Acceleration of an Enzymatic Reaction in Liquid Phase Separated

Compartments Based on Intrinsically Disordered Protein Domains. *ChemSystemsChem* **2**,

e2000001 (2020).

19. Avni, A., Joshi, A., Walimbe, A., Pattanashetty, S. G. & Mukhopadhyay, S. Single-droplet

surface-enhanced Raman scattering decodes the molecular determinants of liquid-liquid

phase separation. *Nat. Commun.* **13**, 4378 (2022).

20. Broide, M. L., Berland, C. R., Pande, J., Ogun, O. O. & Benedek, G. B. Binary-liquid phase

separation of lens protein solutions. *Proc. Natl. Acad. Sci.* **88**, 5660–5664 (1991).

21. Zhang, F. *et al.* Charge-controlled metastable liquid–liquid phase separation in protein

solutions as a universal pathway towards crystallization. *Soft Matter* **8**, 1313–1316

(2012).

22. Li, L. *et al.* Phase Behavior and Salt Partitioning in Polyelectrolyte Complex Coacervates.

Macromolecules (2018) doi:10.1021/acs.macromol.8b00238.

23. Brady, J. P. *et al.* Structural and hydrodynamic properties of an intrinsically disordered

region of a germ cell-specific protein on phase separation. *Proc. Natl. Acad. Sci.* **114**,

E8194–E8203 (2017).

24. Martin, E. W. *et al.* Valence and patterning of aromatic residues determine the phase

behavior of prion-like domains. *Science* **367**, 694–699 (2020).

25. Bremer, A. *et al. Deciphering how naturally occurring sequence features impact the*

phase behaviors of disordered prion-like domains. 2021.01.01.425046

https://www.biorxiv.org/content/10.1101/2021.01.01.425046v1 (2021)

doi:10.1101/2021.01.01.425046.

26. Jawerth, L. *et al.* Protein condensates as aging Maxwell fluids. *Science* **370**, 1317–1323

(2020).

27. Park, Y., Depeursinge, C. & Popescu, G. Quantitative phase imaging in biomedicine. *Nat.*

Photonics **12**, 578–589 (2018).

28. Slabý, T. *et al.* Off-axis setup taking full advantage of incoherent illumination in

coherence-controlled holographic microscope. *Opt. Express* **21**, 14747–14762 (2013).

29. Born, M. & Wolf, E. *Principles of Optics*. (Cambridge University Press, 2019).

30. Barer, R. & Tkaczyk, S. Refractive Index of Concentrated Protein Solutions. *Nature* **173**,

821–822 (1954).

31. Biswas, A., Kim, K., Cojoc, G., Guck, J. & Reber, S. The Xenopus spindle is as dense as the

surrounding cytoplasm. *Dev. Cell* **56**, 967-975.e5 (2021).

32. Zhao, H., Brown, P. H. & Schuck, P. On the Distribution of Protein Refractive Index

Increments. *Biophys. J.* **100**, 2309–2317 (2011).

33. Park, J., Park, J., Lim, H. & Kim, H.-Y. Shape of a large drop on a rough hydrophobic

surface. *Phys. Fluids* **25**, 022102 (2013).

34. Atefi, E., Fyffe, D., Kaylan, K. B. & Tavana, H. Characterization of Aqueous Two-Phase

Systems from Volume and Density Measurements. *J. Chem. Eng. Data* **61**, 1531–1539

(2016).

35. Schlüßler, R. *et al.* Correlative all-optical quantification of mass density and mechanics of

subcellular compartments with fluorescence specificity. *eLife* **11**, e68490 (2022).

36. Matthews, B. W. Solvent content of protein crystals. *J. Mol. Biol.* **33**, 491–497 (1968).

37. Lakowicz, J. R. Quenching of Fluorescence. in *Principles of Fluorescence Spectroscopy*

277–330 (Springer, Boston, MA, 2006). doi:10.1007/978-0-387-46312-4_8.

38. Iserman, C. *et al.* Condensation of Ded1p Promotes a Translational Switch from

Housekeeping to Stress Protein Production. *Cell* **181**, 818-831.e19 (2020).

39. Mittasch, M. *et al.* Non-invasive perturbations of intracellular flow reveal physical

principles of cell organization. *Nat. Cell Biol.* **20**, 344–351 (2018).

40. Fritsch, A. W. *et al.* Local thermodynamics govern formation and dissolution of

Caenorhabditis elegans P granule condensates. *Proc. Natl. Acad. Sci.* **118**, e2102772118

(2021).

41. Lin, Y., Protter, D. S. W., Rosen, M. K. & Parker, R. Formation and Maturation of Phase-

Separated Liquid Droplets by RNA-Binding Proteins. *Mol. Cell* **60**, 208–219 (2015).

42. Strom, A. R. *et al.* Phase separation drives heterochromatin domain formation. *Nature*

advance online publication, (2017).

43. Shin, Y. & Brangwynne, C. P. Liquid phase condensation in cell physiology and disease.

Science **357**, eaaf4382 (2017).

44. Banerjee, P. R., Milin, A. N., Moosa, M. M., Onuchic, P. L. & Deniz, A. A. Reentrant Phase

Transition Drives Dynamic Substructure Formation in Ribonucleoprotein Droplets.

Angew. Chem. Int. Ed. **56**, 11354–11359 (2017).

45. Maharana, S. *et al.* RNA buffers the phase separation behavior of prion-like RNA binding

proteins. *Science* eaar7366 (2018) doi:10.1126/science.aar7366.

46. Choi, J.-M., Holehouse, A. S. & Pappu, R. V. Physical Principles Underlying the Complex

Biology of Intracellular Phase Transitions. *Annu. Rev. Biophys.* **49**, 107–133 (2020).

47. Lin, Y.-H., Brady, J. P., Chan, H. S. & Ghosh, K. A unified analytical theory of

heteropolymers for sequence-specific phase behaviors of polyelectrolytes and

polyampholytes. *J. Chem. Phys.* **152**, 045102 (2020).

48. Valverde, J. M. *et al. Single-embryo phosphoproteomics reveals the importance of*

intrinsic disorder in cell cycle dynamics. 2021.08.29.458076

https://www.biorxiv.org/content/10.1101/2021.08.29.458076v1 (2021)

doi:10.1101/2021.08.29.458076.

49. Müller, P., Schürmann, M., Girardo, S., Cojoc, G. & Guck, J. Accurate evaluation of size

and refractive index for spherical objects in quantitative phase imaging. *Opt. Express* **26**,

10729–10743 (2018).

50. Müller, P., Cojoc, G. & Guck, J. DryMass: handling and analyzing quantitative phase

microscopy images of spherical, cell-sized objects. *BMC Bioinformatics* **21**, 226 (2020).

51. Barer, R. & Joseph, S. Refractometry of Living Cells: Part I. Basic Principles. *J. Cell Sci.* **s3-**
95, 399–423 (1954).

52. McMeekin, T. L., Groves, M. L. & Hipp, N. J. Refractive Indices of Amino Acids, Proteins,
and Related Substances. in *Amino Acids and Serum Proteins* vol. 44 54–66 (American
Chemical Society, 1964).

53. Handwerger, K. E., Cordero, J. A. & Gall, J. G. Cajal Bodies, Nucleoli, and Speckles in the
Xenopus Oocyte Nucleus Have a Low-Density, Sponge-like Structure. *Mol. Biol. Cell* **16**,
202–211 (2005).

54. Klein, I. A. *et al.* Partitioning of cancer therapeutics in nuclear condensates. *Science* **368**,
1386–1392 (2020).

55. Geissbuehler, M. & Lasser, T. How to display data by color schemes compatible with red-
green color perception deficiencies. *Opt. Express* **21**, 9862–9874 (2013).

56. Abuhattum, S. *et al.* Intracellular Mass Density Increase Is Accompanying but Not
Sufficient for Stiffening and Growth Arrest of Yeast Cells. *Front. Phys.* **6**, (2018).

57. Kim, K. *et al.* High-resolution three-dimensional imaging of red blood cells parasitized by
Plasmodium falciparum and in situ hemozoin crystals using optical diffraction
tomography. *J. Biomed. Opt.* **19**, 011005 (2013).

58. Daimon, M. & Masumura, A. Measurement of the refractive index of distilled water
from the near-infrared region to the ultraviolet region. *Appl. Opt.* **46**, 3811–3820 (2007).

59. Rheims, J., Köser, J. & Wriedt, T. Refractive-index measurements in the near-IR using an
Abbe refractometer. *Meas. Sci. Technol.* **8**, 601–605 (1997).

60. Cheong, F. C., Xiao, K., Pine, D. J. & Grier, D. G. Holographic characterization of individual
colloidal spheres' porosities. *Soft Matter* **7**, 6816–6819 (2011).

61. Voss, N. R. & Gerstein, M. Calculation of standard atomic volumes for RNA and

comparison with proteins: RNA is packed more tightly. *J. Mol. Biol.* **346**, 477–492 (2005).

62. Jawerth, L. M. *et al.* Salt-Dependent Rheology and Surface Tension of Protein

Condensates Using Optical Traps. *Phys. Rev. Lett.* **121**, 258101 (2018).

63. Malitson, I. H. Interspecimen Comparison of the Refractive Index of Fused Silica*,†. *JOSA*

55, 1205–1209 (1965).