Label-free cell viability assay using phase imaging with computational specificity

Chenfei Hu[1, 2, †], Shenghua He[3, †], Young Jae Lee[2, 4], Yuchen He[1, 2], Edward M. Kong[2], Hua Li[5,6]

*Mark A. Anastasio[5, *], and Gabriel Popescu[1,2,5 *]*

[1]Department of Electrical and Computer Engineering, [2]Beckman Institute for Advanced Science and Technology, [4]Neuroscience Program, [5]Department of Bioengineering, University of Illinois at Urbana-Champaign, Urbana, Illinois, 61801, USA

[3]Department of Computer Science & Engineering, Washington University in St. Louis, St. Louis, Missouri, 63130, USA

[6]Carle Cancer Center, Carle Foundation Hospital, Urbana, IL, 61801, USA

[†]These authors contributed equally to this work

[*]Corresponding author: Gabriel Popescu (gpopescu@illinois.edu) and Mark Anastasio (maa@illinois.edu)

Abstract

Existing approaches to evaluate cell viability involve cell staining with chemical reagents. However, this step of exogenous staining makes these methods undesirable for rapid, nondestructive and long term investigation. Here, we present instantaneous viability assessment of unlabeled cells using phase imaging with computation specificity (PICS). This new concept utilizes deep learning techniques to compute viability markers associated with the specimen measured by quantitative phase imaging. Demonstrated on HeLa cells culture, the proposed method reports approximately 95% accuracy in identifying injured and dead cells. Further comparison of cell morphology with labeled HeLa cells suggests that potential adverse effect on

cell dynamics introduced by the viability reagents can be avoided using the label-free investigation

method, which would be valuable for a broad range of biomedical applications.

Introduction

Rapid and accurate estimation of viability of biological cells is important for assessing the impact of drug, physical or chemical stimulants, and other potential factors on cell dynamics. The existing methods to evaluate cell viability commonly require mixing a population of cells with reagents to convert a substrate to a colored or fluorescent product [1]. For instance, using membrane integrity as an indicator, the live and dead cells can be separated by trypan blue exclusion assay, where only nonviable cells are stained and appear as a distinctive blue color under a microscope [2, 3]. Starting in 1970s, fluorescence imaging has developed as a more accurate, faster, and reliable method to determine cell viability [4-7]. Similar to the principle of trypan blue test, this method identifies nonviable cells by using fluorescent reagents only taken up by cells that lost their membrane permeability barrier. Unfortunately, the step of exogenous staining generally requires some incubation time for optimal staining intensity, making these methods difficult for quick evaluation. More importantly, the toxicity introduced by stains eventually kills the cells and, thus, prevents the long term investigation.

Quantitative phase imaging (QPI) is a label-free modality that has gained significant interest due to its broad range of potential biomedical applications [8, 9]. QPI measures the optical phase delay across the specimen as an intrinsic contrast mechanism, and thus, allows visualizing transparent specimen (i.e., cells and thin tissue slices) with nanoscale sensitivity, which makes this modality particularly useful for nondestructive investigations of cell dynamics (i.e. growth, proliferation, and mass transport) in both 2D and 3D [10-14]. In addition, the optical phase delay is linearly related to the non-aqueous content in cells (referred to as dry mass), which directly yields biophysical properties of the sample of interest [15-17]. More recently, with the concomitant advances in deep learning, we have witnessed exciting new avenues for QPI. Underlying the

assumption that a QPI map encodes structure and biophysical information, it is possible to apply deep learning techniques to extract subcellular structures, perform signal reconstruction and artifact correction, convert QPI data into virtually stained or fluorescent images, diagnose and classify various specimens [18-26].

In this article, we demonstrate that rapid viability assay can be conducted in a label-free manner using spatial light interference microscopy (SLIM), a highly sensitive QPI method, and deep learning. We apply the concept of our newly-developed phase imaging with computational specificity (PICS) to digitally stain for the live and dead markers. Demonstrated on cultured HeLa cells, we predict the viability of individual HeLa cell measured with QPI as either live, injured or dead by using a joint EfficientNet [27] and transfer learning [28] strategy. Using the standard fluorescent viability imaging as ground truth, the prediction accuracy of 73.4%, 97.0%, and 94.3% in identifying live, injured and dead cells, can be achieved by the trained network. Furthermore, by tracking the cell morphology over time, unstained HeLa cells show significantly higher viability compared to the cells stained with viability reagents. These findings suggest that PICS method enables rapid, nondestructive, and unbiased cell viability assessment, potentially, of great benefit to the general biomedical community.

Results

The procedure of image acquisition is summarized in Fig. 1. We employed spatial light interference microscopy (SLIM) [29, 30] to measure quantitative phase map of cells *in vitro*. The system is built by attaching a SLIM module (CellVista SLIM Pro, Phi Optics, Inc.) to the output port of an existing phase-contrast microscope (Fig. 1a). By modulating the optical phase delay

between the incident and the scattered field, a quantitative phase map is retrieved from four intensity images via phase-shifting interferometry [31]. SLIM employs a broadband LED as illumination source and common-path imaging architecture, which yields sub-nanometer sensitivity to optical pathlength change and high temporal stability [31, 32]. By switching to epi-illumination, the optical path of SLIM is also used to record the fluorescent signals over the same field of view (FOV). A detailed description about the microscope configuration and operations can be found in *Methods*. Live HeLa cell cultures were prepared in a 6-well plate. Cell preparation protocols can be found in the *Methods* section. Before imaging started, two drops of cell-viability-assay reagents (ReadyProbes Cells Viability Imaging Kit; Thermofisher) were added into the growth media, and the cells were then incubated for approximately 15 minutes to achieve optimal staining intensity. The viability-assay kit contains two fluorescently labeled reagents: NucBlue (the "live" reagent) combines with the nuclei of all cells and can be imaged with a DAPI fluorescent filter set, and NucGreen (the "dead" reagent) stains the nuclei of cells with compromised membrane integrity, which is imaged with a FITC filter set. In this assay, live cells produce blue fluorescent signal; dead cells emit both green and blue fluorescence; and the injured cells yield blue and weak green light.

After the staining step, we transferred the 6-well plate to the microscope stage, and imaged the HeLa cells with SLIM and epi-fluorescence microscopy. In order to generate a heterogeneous cell distribution that shifts from predominantly alive to mostly dead cells, we performed the imaging under room environmental conditions, such that the low-temperature and imbalanced pH level in the media would adversely impact the metabolism and eventually lead to cell death. We took one measurement every 30 or 60 minutes, and the entire imaging process lasted for approximately 10 hours. This experiment was repeated four times to capture the variability among

different batches. Figure 1b shows the SLIM images of a representative FOV measured at 1 hour, 3, and 8.5 hours, respectively, and the corresponding fluorescent measurements are shown in Figs. 1c-d. The SLIM images reveal that injured cells suffer from blebbing of the plasma membrane and swelling (shown as higher phase values), while the dead cells are typically characterized by membrane disruption and severe nuclear fragmentation. Our observations agree with the results in literature [33]. On the other hand, these morphological alterations are correlated with the changes in fluorescence signals, where the intensity of NucGreen ("dead" fluorescent channel) continuously increases, as cells transit to injured and dead states. By computing the ratio between NucGreen and NucBlue signals within the cell nuclei, semantic segmentation maps were generated to label individual cell as either live, injured, or dead, as shown in Fig. 1e. The *Methods* section and Fig. S1 describe in detail the semantic map generation.

Deep neural network architecture, training, and testing

Using the SLIM data as input and fluorescence-based semantic maps as ground truth, we trained a deep neural network to perform semantic segmentation. The deep neural network was employed to assign labels for "live", "injured", "dead", and background pixels from the SLIM images. We employed a U-Net based on EfficientNet (E-U-Net) [27], with its architecture shown in Fig.2a. Compared to conventional U-Nets, the E-U-Net uses EfficientNet [27] as a powerful network of relatively lower complexity as the encoding part. This architecture allows for learning an efficient and accurate end-to-end segmentation model, while avoiding training a very complex network. The network was trained using a transfer learning strategy [28] that performs well with only few labeled ground truth images. At first, the EfficientNet of E-U-Net (the encoding part) was pre-trained for image classification on a publicly available dataset ImageNet [34], containing

millions of natural images from 1000 categories. The entire E-U-Net was then further trained for a semantic segmentation task by using 899 labeled SLIM images as the training set and 199 labeled images as the validation set. The network training was performed by updating the weights of parameters in the E-U-Net using an Adam optimizer [35] to minimize a focal loss [36] computed in the training set. More details about the EfficientNet module and loss function can be found in the *Supplemental Information*. The network was trained for 100 epochs. At the end of each training epoch, the focal loss of the network was evaluated, and the weights that yielded the lowest focal loss on the validation set were selected for the E-U-Net model. The curves related to training and validation focal losses vs. epochs are shown in Fig. 2d. The *Methods* section and Figs. 2a-c present more details about the E-U-Net architecture and network training.

To demonstrate the performance of phase imaging with computational specificity (PICS) as a label-free live/ dead assay, we applied the trained network to 200 SLIM images not used in the training process. Figure 3a shows the three representative testing phase maps, with their corresponding semantic maps in Fig. 3b and PICS in Fig. 3c. This direct comparison indicates that PICS successfully identifies the region of cell nucleus with high confidence. As reported in previous publications, the conventional deep learning evaluation metrics focus on assessing pixel-wise segmentation accuracy, which overlooks some biologically relevant instances [37]. Thus, we adopted an object-based accuracy metric, which relies on comparing the dominant semantic label between the predicted cell nuclei and the ground truth at the cell nucleus scale. The confusion matrix and the corresponding evaluation (e.g. precision, recall and F1-score) are shown in Table. 1. A comparison with standard pixel-wise evaluation is included in Table. S1 in *Supplemental Information*. The entries of the confusion matrix are normalized with respect to the number of cells in each category. Our PICS strategy reports an F1 score of 73.4%, 97.0%, and 94.3% in segmenting

live, injured, and dead HeLa cells, respectively. We speculate that the relatively lower performance on identifying live cells are due to subtle morphological differences between live cells and cells just starting to lose membrane permeability. In Fig. S1 in the *Supplemental Information*, the histogram of fluorescent ratio between NucGreen and NucBlue shows a continuous distribution pattern, which suggests the potential ambiguity in predicting the cell status. However, for routine assays, one can use PICS to label only the injured and dead cells with high accuracy, while labeling the remaining cells as "live".

PICS on label-free cells *in vitro*

Performing viability assay on unlabeled cells essentially circumvents the cell injury effect caused by exogenous staining and produces an unbiased evaluation. To demonstrate this advantage, a fresh HeLa cell culture was prepared in a 6-well plate, transferred to the microscope stage, and maintained under room conditions. Half of the wells were mixed with viability assay reagents, where the viability was determined by both PICS and fluorescence imaging. The remaining wells did not contain reagents, such that the viability of these cells were only evaluated by PICS. The procedure of cell preparation, staining, and microscope settings were consistent with the previous experiments. We took measurements every 30 minutes, and the entire experiment lasted for 12 hours.

Figures 4a, c show SLIM images of HeLa cells with and without fluorescent reagents at t = 0, 2.5, and 12 hours, respectively, whereas the resulting PICS predictions are shown in Figs. 4b, d. Supplemental Video 1 shows a time-lapse SLIM measurement, PICS prediction, and standard live-dead assay based on fluorescent measurements. Supplemental Video 2 shows HeLa cells without reagents. As expected, the PICS method depicts the transition from live to injured and

dead state. In addition, the visual comparison from Figs. 4a-d suggests that HeLa cells with viability stains in the media appear smaller in size, and more rapidly entering the injured state, as compared to their label-free counterparts. Using TrackMate [38], an imageJ plugin, we were able to extract the trajectory of individual cells and track the cell morphology over time. As a result, the cell nucleus, area, and dry mass at each moment in time can be obtained by integrating the pixel value over the segmented area in the PICS prediction and SLIM image, respectively. We successfully tracked 57 labeled and 34 unlabeled HeLa cells. Figures 4e-f show the area and dry mass change (mean ± standard error), where the values are normalized with respect to the one at t=0. Running two sample t-tests, we found a significant difference in cell nuclear areas between the labelled and unlabeled cells, during the interval t=2 and t=7 hours ($p < 0.05$). Similarly, cell dry mass showed significant differences between the two groups during the time interval t=2 and t=5 hours ($p < 0.05$). As expected, these differences vanish as the cells in both groups eventually die.

Summary and Discussion

We demonstrated PICS as a method for high-throughput, label-free, unbiased viability assessment of adherent cells. This approach utilizes quantitative phase imaging to record high-resolution morphological structure of unstained cells, combined with deep learning techniques to extract intrinsic viability markers. Tested on HeLa cells, the proposed method achieves 98.7% accuracy in distinguish background and cell nucleus, 97% accuracy in identifying injured cells and 94% accuracy in labeling dead cells. By integrating the trained network on NVIDIA graphic processing units, the proposed label-free method enables real-time acquisition and viability prediction (see Supplemental Video 3 for a demonstration). One SLIM measurement and deep

learning prediction takes ~100 ms, which is approximately 8 times faster than the acquisition time required for fluorescence imaging with the same camera. Of course, the cell staining process itself takes time, 15 minutes in our case. In addition, using optical pathlength as an intrinsic contrast marker, PICS rules out the adverse effect on cell function caused by the exogenous staining and allows for an unbiased assessment of cell viability over long periods of time (*e.g.,* many days). As shown in Figs. 4e-f, label-free HeLa cells showed significantly higher viability than their labelled counterparts.

The translation of this approach to other cell types is straight forward: the ground truth data is generated automatically from the SLIM and fluorescence colocalized channels, while the training process takes between approximately 6 hours on a NVIDIA Tesla V100 GPU of 32 GB VRAM. Thus, the label-free assay for a new cell type can be completed in less than a day, while the resulting PICS inference takes 48 ms and is implemented in the acquisition software, in real-time.

Label-free imaging methods are valuable to study biological samples without destructive fixation or staining. The capabilities of PICS can be furthermore enriched by combining with other imaging modalities. For example, by employing infrared spectroscopy, the bond-selective transient phase imaging measures molecular information associated with lipid droplet and nucleic acids [39]. In addition, harmonic optical tomography can be integrated into an existing QPI system to report both non-centrosymmetric structures [40]. These additional chemical signals would potential enhance the effective learning and produce more biophysical information. We anticipate that the PICS method will benefit high-throughput cell screening for a variety of applications, ranging from basic research to therapeutic development and protein production in cell reactors [8].

Methods

HeLa cell preparation

HeLa cervical cancer cells (ATCC CCL-2) were purchased from ATCC and kept frozen in liquid nitrogen. Prior to the experiments, we thawed and cultured the cells into T75 flask in Dulbecco's Modified Eagle Medium (DMEM with low glucose) containing 10% fetal bovine serum (FBS) and incubated in 37°C with 5% CO2. As the cells reach 70% confluence, the flask was washed thoroughly with phosphate-buffered saline (PBS) and trypsinized with 3 mL of 0.25% (w/v) Trypsin EDTA for three minutes. When the cell starts to detach, the cells were suspended in 5 mL DMEM and passaged onto a glass bottom 6 well plate to grow. HeLa cells were then imaged after two days.

SLIM imaging

The SLIM optical setup in shown in Fig. 1a. Detailed design and calibration procedure have been described previously [29]. In brief, the microscope is built upon an inverted phase contrast microscope (Axiovert Z1; Zeiss) using a SLIM module (CellVista SLIM Pro; Phi Optics) attached to the output port. Inside the module, a spatial light modulator (Meadowlark Optics) is placed at the system pupil plane via a Fourier transform lens to constantly modulate the phase delay between the scattered and incident light. By recording four intensity images with phase shifts of 0, $\pi/2$, π, and $3\pi/2$, a quantitative phase map, φ, can be computed by combining the 4 acquired frames in real-time.

For both SLIM and fluorescence imaging, cultured HeLa cells were measured by a 40×/ 0.75 NA objective (EC Plan-Neofluar; Zeiss), and the images were recorded by a CMOS camera (ORCA-Flash 4.0; Hamamatsu) with a pixel size of 6.5 μm. In this experiment, acquisition time of each SLIM and fluorescent measurement is 50 ms and 400 ms, respectively. For deep learning training and predicting, the recorded SLIM images were downsampled by a factor of 2. This step saves computational cost and does not sacrifice information content.

Semantic segmentation map generation

Semantic segmentation maps were generated automatically in MATLAB using a customized script. For each image used for training, validation, and testing dataset, an adaptive thresholding was applied to the live and dead fluorescent images, such that the segmented cell nuclei were obtained by computing the union of the binarized fluorescent images. Next, the scattered labeling artifacts mainly due to fluorescent signals in the background were removed. Based on the segmentation masks, we calculated the ratio between the dead and live fluorescence signal. A histogram of the average ratio within the cell nucleus is plotted in Fig. S1 in the *Supplemental Information*, where three distinctive peaks were observed corresponding to the live, injured and dead state. Finally, semantic labels were assigned to cell nuclei by comparing to the thresholds set at values producing lowest histogram counts between two classes.

E-U-Net architecture and network training

The E-U-Net is a U-Net-like fully convolutional neural network that performs an efficient end-to-end mapping from SLIM images to the corresponding probability maps, from which the desired segmentation maps are determined by use of a softmax decision rule. The network architecture of E-U-Net is shown in Fig. 2a. Similar to a standard U-Net [41], it includes an encoding path to extract multi-scale features from input images and a decoding path to integrate these features for inferring the corresponding probability maps. The decoding path includes five decoding stages that correspond to five scales of feature maps from down to top. The sizes of filters in the five decoding stages were set to 3×3, and the numbers of filers in the five stages were set to 256, 128, 64, 32, and 16, respectively. Different from conventional U-Nets, the E-U-Net uses a more efficient network architecture, EfficientNet [27], for feature extraction in the encoding path. Here, EfficientNets refers to a family of deep convolutional neural networks that possess a powerful capacity of feature extraction but require much fewer network parameters compared to other state-of-the-

art network architectures, such VGG-Net, ResNet, Mask R-CNN, etc. This can significantly benefit the segmentation model learning in this study in two aspects. First, its strong capacity of feature extraction allows to extract informative viability-related features from SLIM images for viability inference, which is very challenging due to the low nuclei contrast and indistinguishable nuclei structures between live and injured cells in the images. Second, it does not require the training of large-size network to extract representative features for semantic label inference, which can be also very challenging especially with limited training images. The EfficientNet family includes eight network architectures, EfficientNet-B0 to EfficientNetB7, with an increasing network complexity. EfficientNet-B3 was selected in this study considering it yields the most accurate segmentation performance on the validation set among all the eight EfficientNets. See Supplemental information and Fig.2b-2c for more details about the EfficientNet-B3.

The E-U-Net was trained with randomly cropped patches of 512×512 pixels drawn from the training set by minimizing a focal loss [36] with an Adam optimizer [35]. In regard to Adam optimizer, the exponential decay rates for 1st and 2nd moment estimates were set to 0.9 and 0.999, respectively; a small constant ε for numerical stability was set to 10^{-7}. See *Supplemental information* for more details related to the focal loss. The batch size was set to 14. The learning rate was initially set to 5×10^{-4}. At the end of each epoch, the focal loss of the being-trained E-U-Net was computed on the whole validation set. When the validation loss did not decrease for 10 training epochs, the learning rate was multiplied by a factor of 0.8. This validation loss-aware learning rate decaying strategy benefits for mitigating the overfitting issue that commonly occurs in deep neural network training. Furthermore, data augmentation techniques, such as random cropping, flipping, shifting, and random noise and brightness adding etc., were employed to augment training samples on-the-fly for further reducing the overfitting risk. The E-U-Net was trained for 100 epochs. The parameter weights that yield the lowest validation loss were selected, and subsequently used for model testing and further model investigation. The similar trends related to the training and validation curves can be observed in the Fig. 2d, which indicates that the network was reasonably trained without obvious over-fitting.

The E-U-Net was implemented using the Python programming language with libraries including Python 3.6 and Tensorflow 1.14. The model training, validation and testing were performed on a NVIDIA Tesla V100 GPU of 32 GB VRAM.

Data availability

The images and generated that support the findings of this study are available from the corresponding author upon request.

Code availability

The MATLAB script for semantic segmentation maps and object-based evaluation are available from the corresponding author upon request. The trained network model along with sample testing images are available for download at https://github.com/shenghh2020/label-free-viability-assay.

Acknowledgement

This work is sponsored in part by the National Science Foundation (0939511, 1353368), and the National Institute of Health (R01CA238191, R01GM129709). C. H is supported by National Institute of Health (T32EB019944).

Author contribution

C. H. designed the experiment, collected data, generated semantic segmentation maps, performed deep learning testing, and conducted data analysis. S. H. led the deep learning training and validation. Y. L.

prepared HeLa cell culture. Y. H. performed software integration. E. M. K prepared the script for PICS evaluation. C. H. and S. H wrote the manuscript with input from all authors. M. A. A. supervised the deep learning. G. P. supervised the project.

Competing financial interest

G.P. has financial interest in Phi Optics, a company developing quantitative phase imaging technology for materials and life science applications, which, however, did not sponsor the research.

List of supplemental materials.

1. Supplemental Information

2. Supplemental Movie 1:

   Animation of HeLa cells with viability reagents measured by SLIM (left), corresponding PICS prediction (middle), and standard viability assay (right)

3. Supplemental Movie 2:

   Animation of HeLa cells without staining reagents measured by SLIM (left), and corresponding PICS prediction (right)

4. Supplemental Movie 3:

   Real-time SLIM measurement and PICS viability prediction

References:

1. Riss, T.L., et al., *Cell viability assays*, in *Assay Guidance Manual [Internet]*. 2016, Eli Lilly & Company and the National Center for Advancing Translational Sciences.

2. Strober, W., *Trypan blue exclusion test of cell viability.* Curr Protoc Immunol, 2001. **Appendix 3**: p. Appendix 3B.

3. Phillips, H.J., *Dye exclusion tests for cell viability*, in *Tissue culture*. 1973, Elsevier. p. 406-408.

4. Altman, S.A., L. Randers, and G. Rao, *Comparison of trypan blue dye exclusion and fluorometric assays for mammalian cell viability determinations.* Biotechnol Prog, 1993. **9**(6): p. 671-4.

5. Breeuwer, P. and T. Abee, *Assessment of viability of microorganisms employing fluorescence techniques.* Int J Food Microbiol, 2000. **55**(1-3): p. 193-200.

6. Haugland, R.P., I.C. MacCoubrey, and P.L. Moore, *Dual-fluorescence cell viability assay using ethidium homodimer and calcein AM*. 1994, Google Patents.

7. Millard, P.J., et al., *Fluorescent viability assay using cyclic-substituted unsymmetrical cyanine dyes*. 1996, Google Patents.

8. Park, Y., C. Depeursinge, and G. Popescu, *Quantitative phase imaging in biomedicine.* Nature Photonics, 2018. **12**(10): p. 578-589.

9. Popescu, G., *Quantitative phase imaging of cells and tissues*. McGraw-Hill biophotonics. 2011, New York: McGraw-Hill. xx, 362 p.

10. Hu, C., et al., *Optical excitation and detection of neuronal activity.* J Biophotonics, 2018: p. e201800269.

11. Kandel, M.E., et al., *Epi-illumination gradient light interference microscopy for imaging opaque structures.* Nature Communications, 2019. **10**(1): p. 4691.

12. Merola, F., et al., *Tomographic flow cytometry by digital holography.* Light-Science & Applications, 2017. **6**.

13. Lee, K., et al., *Quantitative phase imaging techniques for the study of cell pathophysiology: from principles to applications.* Sensors (Basel), 2013. **13**(4): p. 4170-91.

14. Kim, K., et al., *Real-time visualization of 3-D dynamic microscopic objects using optical diffraction tomography.* Optics Express, 2013. **21**(26): p. 32269-32278.

15. Eldridge, W.J., et al., *Optical Phase Measurements of Disorder Strength Link Microstructure to Cell Stiffness.* Biophys J, 2017. **112**(4): p. 692-702.

16. Popescu, G., et al., *Optical measurement of cell membrane tension.* Phys Rev Lett, 2006. **97**(21): p. 218101.

17. Hu, C., et al., *Imaging collagen properties in the uterosacral ligaments of women with pelvic organ prolapse using spatial light interference microscopy (SLIM).* Frontiers in Physics, 2019. **7**: p. 72.

18. Jo, Y., et al., *Quantitative Phase Imaging and Artificial Intelligence: A Review.* Ieee Journal of Selected Topics in Quantum Electronics, 2019. **25**(1).

19. Rivenson, Y., Y. Wu, and A. Ozcan, *Deep learning in holography and coherent imaging.* Light Sci Appl, 2019. **8**: p. 85.

20. Rivenson, Y., et al., *PhaseStain: the digital staining of label-free quantitative phase microscopy images using deep learning.* Light-Science & Applications, 2019. **8**.

21. Nygate, Y.N., et al., *Holographic virtual staining of individual biological cells.* Proceedings of the National Academy of Sciences of the United States of America, 2020. **117**(17): p. 9223-9231.

22. Kandel, M.E., et al., *Reproductive outcomes predicted by phase imaging with computational specificity of spermatozoon ultrastructure.* Proc Natl Acad Sci U S A, 2020. **117**(31): p. 18302-18309.

23.   Bostan, E., et al., *Deep phase decoder: self-calibrating phase microscopy with an untrained deep neural network.* Optica, 2020. **7**(6): p. 559-562.

24.   Liu, R.R., et al., *Deep spectral learning for label-free optical imaging oximetry with uncertainty quantification.* Light-Science & Applications, 2019. **8**.

25.   Kandel, M.E., et al., *PICS: Phase Imaging with Computational Specificity.* arXiv preprint arXiv:2002.08361, 2020.

26.   Rivenson, Y., et al., *Phase recovery and holographic image reconstruction using deep learning in neural networks.* Light-Science & Applications, 2018. **7**.

27.   Tan, M. and Q.V. Le, *Efficientnet: Rethinking model scaling for convolutional neural networks.* arXiv preprint arXiv:1905.11946, 2019.

28.   Pan, S.J. and Q. Yang, *A survey on transfer learning. IEEE Transactions on knowledge and data engineering.* 22 (10): 1345, 2010. **1359**.

29.   Wang, Z., et al., *Spatial light interference microscopy (SLIM).* Optics Express, 2011. **19**(2): p. 1016-1026.

30.   Kim, T., et al., *White-light diffraction tomography of unlabeled live cells.* Nature Photonics, 2014. **8**(3): p. 256-263.

31.   Hu, C. and G. Popescu, *Quantitative Phase Imaging: Principles and Applications*, in *Label-Free Super-Resolution Microscopy*. 2019, Springer. p. 1-24.

32.   Hu, C. and G. Popescu, *Quantitative Phase Imaging (QPI) in Neuroscience.* IEEE Journal of Selected Topics in Quantum Electronics, 2019. **25**(1): p. 1-9.

33.   Miller, M.A. and J.F. Zachary, *Mechanisms and morphology of cellular injury, adaptation, and death.* Pathologic Basis of Veterinary Disease, 2017: p. 2.

34.   Deng, J., et al., *ImageNet: A Large-Scale Hierarchical Image Database.* Cvpr: 2009 Ieee Conference on Computer Vision and Pattern Recognition, Vols 1-4, 2009: p. 248-255.

35.   Kingma, D.P. and J. Ba, *Adam: A method for stochastic optimization.* arXiv preprint arXiv:1412.6980, 2014.

36.   Lin, T.Y., et al., *Focal Loss for Dense Object Detection.* IEEE Trans Pattern Anal Mach Intell, 2020. **42**(2): p. 318-327.

37.   Caicedo, J.C., et al., *Evaluation of Deep Learning Strategies for Nucleus Segmentation in Fluorescence Images.* Cytometry A, 2019. **95**(9): p. 952-965.

38.   Tinevez, J.Y., et al., *TrackMate: An open and extensible platform for single-particle tracking.* Methods, 2017. **115**: p. 80-90.

39.   Zhang, D., et al., *Bond-selective transient phase imaging via sensing of the infrared photothermal effect.* Light Sci Appl, 2019. **8**: p. 116.

40.   Hu, C.F., et al., *Harmonic optical tomography of nonlinear structures.* Nature Photonics, 2020. **14**(9): p. 564-+.

41.   Ronneberger, O., P. Fischer, and T. Brox, *U-Net: Convolutional networks for biomedical image segmentation. arXiv 2015.* arXiv preprint arXiv:1505.04597, 2019.

Figure. 1 **Schematic of the imaging system and representative results**. **a.** CellVista SLIM Pro microscope (Phi Optics, Inc.) consists of an existing phase contrast microscope and an external module attached to the output port. By switching between transmission and reflection excitation, both SLIM and co-localized fluorescence images can be recorded via the same optical path. Before time-lapse imaging started, fluorescence viability reagents were mixed with HeLa cell culture. **b.** Representative SLIM measurements of HeLa cell at 1, 3, and 8.5 hours. **c.** Corresponding DAPI fluorescent signals of the live viability reagent. **d.** Corresponding fluorescent signals of the dead viability reagents measured by a FITC filter. **e.** Viability states of the individual cells. Scale bars: 50 microns

Figure 2 **Principle of E-U-Net training. a.** The E-U-Net. architecture includes an EfficientNet as the encoding path and five stages of decoding. The E-U-Net includes a Down+Conv+BN+ReLU block and 7 other blocks. The Down-Conv-BN-ReLU block represents a chain of down-sampling layer, convolutional layer, batch normalization layer, and ReLU layer. Similarly, the Conv+BN+ReLU is a chain of convolutional layer, batch normalization layer, and ReLU layer. **b** The network architecture of EfficientNet-B3. Different blocks are marked in different colors. They correspond to the layer blocks of EfficientNet in **a**. **c.** The major layers inside the MBConvX module. X =1 and X = 6 indicate the ReLU and ReLU6 are used in the module, respectively. The skip connection between the input and output of the module is not used in the first MBConvX module in each layer block. **d.** Training and validation focal losses vs epochs.

Figure 3. **Results of E-U-Net on testing dataset**. **a.** representative SLIM measurements of HeLa cells not used during training. **b.** The ground truth of viability of frames corresponding to **a**. **c.** The PICS prediction shows high level accuracy in segmenting the nuclear regions and inferring viability states. Scale bars: 50 microns

Figure 4. **Viability of HeLa cells with and without reagent stains. a.** SLIM images of cells recorded at 0, 2.5 and 12 hours after staining. **b.** The PICS prediction associated with the frames in **a**. **c.** SLIM images of unstained HeLa cells measured at same time points as **a**. **d.** The corresponding PICS prediction associated with the frames in **c**. **e.** Relative cell nuclear area change of tracked cells. The shaded region represents the standard error. **f.** Relative cell nucleus dry mass change. The shaded region represents of the standard error.

Table. 1. **Evaluation of the E-U-Net performance.** An object-based accuracy metric is used to estimate the deep learning prediction by comparing the dominant semantic label of cell nuclei with the ground truth. The entries of the confusion are normalized with respect to number of cells in each class.

Actual / Prediction	Live (n= 146)	Injured (n = 1828)	Dead (n = 246)
Live	82.2%	3.3%	0%
Injured	17.1%	95.5%	2.4%
Dead	0.7%	1.2%	97.6%
Precision	66.3%	98.6%	91.3%
Recall	82.2%	95.5%	97.6%
F1 Score	73.4%	97.0%	94.3%