Single gene targeted nanopore sequencing enables simultaneous identification and

antimicrobial resistance detection of sexually transmitted infections

Running title: Identification and antimicrobial resistance detection of sexually transmitted

infections

Liqing Zhou[1], Andrea Lopez Rodas[2], Luz Marina Llangarí[2], Natalia Romero Sandoval[2,3],

Philip Cooper[1,2], Syed Tariq Sadiq[1*]

[1] Applied Diagnostic Research and Evaluation Unit

Institute for Infection and Immunity

St George's, University of London

Cranmer Terrace

London SW17 0RE

UK

[2] School of Medicine,

Faculty of Health and Life Sciences

Universidad Internacional del Ecuador

Quito, Ecuador

[3] Grups de Recerca d'Amèrica i Àfrica Llatines, GRAAL, Barcelona, Spain

* Corresponding author

Tel: +44 2087255740

Fax: +44 2087250137

E-mail: ssadiq@sgul.ac.uk

Abstract

Objectives

To develop a simple DNA sequencing test for simultaneous identification and antimicrobial

resistance (AMR) detection of multiple sexually transmitted infections (STIs).

Methods

Real-time PCR (qPCR) was initially performed to identify *Neisseria gonorrhoeae* (NG),

Chlamydia trachomatis (CT), *Mycoplasma genitalium* (MG) and *Trichomonas vaginalis* (TV)

infections among a total of 200 vulvo-vaginal swab samples from female sex workers in

Ecuador. qPCR positive samples plus qPCR negative controls for these STIs were subjected

to single gene targeted PCR MinION-nanopore sequencing using the smartphone operated

MinIT.

Results

Among 200 vulvo-vaginal swab samples 43 were qPCR positive for at least one of the STIs.

Single gene targeted nanopore sequencing generally yielded higher pathogen specific read

counts in qPCR positive samples than qPCR negative controls. Of the 26 CT, NG or MG

infections identified by qPCR, 25 were clearly distinguishable from qPCR negative controls

by read count. Discrimination of TV qPCR positives from qPCR negative controls was

poorer as many had low pathogen loads (qPCR cycle threshold >35) which produced few

specific reads. Real-time AMR profiling revealed that 3/3 NG samples identified had *gyrA*

mutations associated with fluoroquinolone resistance, while none of the MG samples

possessed *23S rRNA* gene mutations contributing to macrolide resistance.

Conclusions

Single gene targeted nanopore sequencing for diagnosing and simultaneously identifying key

antimicrobial resistance markers for four common genital STIs shows promise. Further work

to optimise accuracy, reduce costs and improve speed may allow sustainable approaches for

managing STIs and emerging AMR in resource poor and laboratory limited settings.

Introduction

Sexually transmitted infections (STIs) remain a major public health problem worldwide, with an estimated 357 million new cases of *Chlamydia trachomatis* (CT), *Neisseria gonorrhoeae* (NG), *Treponema pallidum* (TP) and *Trichomonas vaginalis* (TV) per year [1], all which commonly cause a genital discharge syndrome in men and women. High rates of the STI, *Mycoplasma genitalium* (MG), have also been reported worldwide [2] and is associated with genital discharge syndrome in men and reproductive sequelae in women [3]. All these STIs are generally curable with existing, effective single-dose antibiotic regimens but, if left undiagnosed and/or untreated, can result in serious long-term reproductive health sequelae, particularly for women. Antimicrobial resistance (AMR) among some STIs to multiple classes of antibiotics has spread rapidly in recent years. For NG, loss of extended spectrum cephalosporins as first line empirical treatment is a major concern [4] as circulating multi- and extensive-drug resistance clones have been detected internationally [5]. For MG, macrolide resistance is now widely but not universally reported with increasing rates of resistance to fluoroquinolones also detected [6]. These developments have made treatment, and particularly empirical treatment challenging.

New World Health Organization (WHO) guidelines reinforce the need to treat these STIs with the right antibiotic, at the right dose, and the right time to reduce spread and improve sexual and reproductive health. For both NG and MG, accurate and rapid diagnostics which also predict antibiotic susceptibility are likely to be needed to achieve this. Laboratory based nucleic-acid amplification tests (NAAT) for detection of NG, CT, TV and MG are the current gold standard for detection, providing high sensitivity and specificity. NAAT are widely used in high-income countries but are often unavailable in resource-poor settings. Culture-based antimicrobial susceptibility testing (AST) for NG remains gold standard for phenotypically predicting AMR and commonly takes usually two to five days to obtain a result in UK

microbiology laboratories (personal communication, Sadiq). This in practical terms may be

too late to initiate targeted antibiotic therapy, particularly for hard to reach vulnerable

populations. MG is difficult to cultivate and requires cell culture, not usually feasible in most

clinical settings.

Advances in understanding AMR in NG, MG and TV have allowed for development of

NAAT- based AMR detection. For example, absence of *gyrA* mutations in NG accurately

predicts fluoroquinolone susceptibility [7,8], presence of *23S rRNA* mutations in MG is

associated with failure of treatment with azithromycin [9-11], and *ntr6* mutations in TV may

have diagnostic value for metronidazole resistance [12]. Use of NAAT-based AMR tests,

however, may have limitations due to continually changing mutations and novel mechanisms

of resistance evolving under ongoing treatment selection pressures. Whole genome

sequencing (WGS) using high throughput sequencing platforms may address this challenge to

some degree and give added value in identifying phylogenetic relationships in identified

infections [13,14]. However, for diagnostic purposes WGS itself can be constrained by

sample preparation, cost etc., making it unsuitable for near patient applications.

Oxford Nanopore Technologies' (ONT) portable MinION DNA sequencer, together with its

recent MinIT hand-held processor, may offer advantages as an accurate diagnostic in

resource-limited settings. Herein we report early phase study among female sex workers

(FSWs) in Ecuador, of using the MinION, controlled by a smartphone-operated MinIT to

detect the four common STIs and profile AMR to ciprofloxacin in NG, azithromycin in MG

and metronidazole in TV utilising a single gene targeted approach.

Materials and Methods

Ethics statement

Written and verbal informed consent was obtained from FSW participants all of whom were

18 years or older and none of whom received any compensation for their participation. The

study was conducted according to the Declaration of Helsinki and approved by the Ethical

Committee of Universidad Internacional del Ecuador (02-02-17).

Clinical sample collection and DNA preparation

Vulvo-vaginal swab samples were collected by clinicians using Xpert® CT/NG Patient-

Collected Vaginal Swab Specimen Collection Kit (Cepheid). DNA from swab samples was

prepared using PureLink Genomic DNA Mini Kit (Thermo Fisher Scientific) according to the

manufacturer's instructions and quantified using a NanoDrop Spectrometer.

STI identification of swab samples by qPCR

Swab samples were initially screened by qPCR to identify NG, TV, MG and CT infections,

with gene targets and primers shown in Table 1.   qPCR was performed using Applied

Biosystems 7500 Fast Real-Time PCR System in a volume of 10 μl containing 5 μl of

TaqMan™ Fast Universal PCR Master Mix, 1 µl of 10x Exogenous Internal Positive Control

(IPC) Mix, 0.2 µl of 50x IPC DNA, 250 nM each of the primers, 100 nM each of the probes

and 50 ng of template DNA. Cycling parameters were 95˚C for 10 min, followed by 40

cycles of 95°C for 15 sec, 60°C for 1 min.

Table 1 Gene targets, primers and probes used in this study

a) Initial real-time PCR (qPCR) for STI testing on samples

Trichomonas vaginalis: *T. vaginalis* specific repeat DNA fragment [15]
    TV_F: AAAGATGGGTGTTTTAAGCTAGATAAGG
    TV_R: TCTGTGCCGTCTTCAAGTATGC
    TV_P: [6FAM]AGTTCATGTCCTCTCCAAGCGTAAGT[BHQ1]
 	
Mycoplasma genitalium: MgPa gene [16]
    MG_F: GAGAAATACCTTGATGGTCAGCAA

MG_R: GTTAATATCATATAAAGCTCTACCGTTGTTATC
MG_P: [6FAM]ACTTTGCAATCAGAAGGT[MGBNFQ]

Chlamydia trachomatis: The cryptic plasmid [17]
   CT_F: CATGAAAACTCGTTCCGAAATAGAA
   CT_R: TCAGAGCTTTACCTAACAACGCATA
   CT_P: [6FAM]TCGCATGCAAGATATCGA[MGBNFQ]

Neisseria gonorrhoeae: opa [18]
   NG_F: TTGAAACACCGCCCGGAA
   NG_R: TTTCGGCTCCTTATTCGGTTTGA
   NG_P: [6FAM]CCGATATAATCCGCCCTTCAACATCAG[BHQ1]

b) Targeted PCR (tPCR) for nanopore sequencing*

Trichomonas vaginalis: Nitroreductase family protein (*ntr6*)
   TV_NTR6_BC_F:
   5'TTTCTGTTGGTGCTGATATTGCCTTCATTGAATTTATTCGTTCAAAATT
   TV_NTR6_BC_R:
   5'ACTTGCCTGTCGCTCTATCTTCTTATTCAATGTATGTAACCTTTCTAA

Mycoplasma genitalium: 23S ribosomal RNA
   Mg 23S_BC_1992F:
   5' TTTCTGTTGGTGCTGATATTGCCCATCTCTTGACTGTCTCGG
   Mg 23S_BC_2679R:
   5' ACTTGCCTGTCGCTCTATCTTCTCCTCTCGTACTAGAAGCAAAG

Chlamydia trachomatis: Major outer membrane protein 1 (*omp1*)
   CT_OMP1_BC_F:
   5' TTTCTGTTGGTGCTGATATTGCTTTGCCGCTTTGAGTTCTGCT
   CT_OMP1_BC_R:
   5' ACTTGCCTGTCGCTCTATCTTCCAATACCGCAAGATTTTCTAGATTTC

Neisseria gonorrhoeae: *gyrA*
   NG_gyrA_BC_F:
    5' TTTCTGTTGGTGCTGATATTGCATCCGCCACGACCACAAATT
   NG_gyrA_BC_R:
    5' ACTTGCCTGTCGCTCTATCTTCATATTGGACAGTGCGACGGC

*developed in this study

Targeted PCR (tPCR) amplification and barcoding

Both qPCR positive and negative samples for NG, MG, TV and CT plus an NG positive

control were amplified by tPCR, targeting genes *gyrA* (NG), *23S rRNA* (MG), *ntr6* (TV) and

omp1 (CT) with the primers listed in Table 1. tPCR reactions for each target were performed

in a 50 µl volume consisting of 25 µl LongAmp® Taq 2X Master Mix (NEB), 200 nM each

of tPCR primers and 100 ng DNA template as follows: 95 °C for 3 min followed by 35 cycles

of 95 °C for 30 sec, 58 °C for 30 sec, 65 °C for 1 min 30 sec, and one cycle of 65 °C for 5

min. tPCR amplicons were purified and subjected to a second PCR for barcoding using PCR

Barcoding Expansion 1-96 (EXP-PBC096, ONT) according to the manufacturer's

instructions.

Nanopore sequencing and data analysis

A DNA sequencing library was constructed from a pool of barcoded tPCR amplicons of

different clinical samples using Ligation Sequencing Kit (SQK-LSK108) and sequenced

using FLO-MIN106 R9 Version Flow Cell MK I Spot-ON on portable DNA sequencer

MinION MK I which was controlled by a smartphone operated MinIT (ONT) according to

the manufacturer's instructions. Sequencing data were uploaded onto Metrichor Epi2ME

(ONT) and analyzed using the workflow Fastq Antimicrobial Resistance r3.3.2 which

included three components: QC and Barcoding [rev. 3.10.4], WIMP [rev. 3.4.0] and ARMA

CARD [rev. 1.1.6]. QC and Barcoding component contains quality score (qscore) and

barcode filter, cutting off reads with a qscore below the threshold (min qscore:7) which will

not be uploaded for analysis, and demultiplexing barcodes if the sequenced samples are

barcoded. WIMP (what's In My Pot) allows for species identification by classifying read

sequences against the standard Centrifuge database (including RefSeq complete genomes for

bacteria). ARMA (antimicrobial resistance mapping application) performs antimicrobial

resistance identification by aligning input reads with minmap2 against all reference

sequences available in the CARD database (CARD version 1.1.3). TV, an anaerobic,

flagellated protozoan parasite, was not included in RefSeq for bacteria and therefore a TV G3

nitroreductase family protein (TVAG_354010, ntr6) reference was uploaded using Fasta

Reference Upload r3.2.2 and TV identification was performed using Fastq Custom

Alignment r3.2.2 against the *ntr6* reference sequence. Read counts were log-transformed and

compared between qPCR positive and negative samples by t-test.

BLAST search

DNA sequences of NG, MG and TV reads were extracted from read FASTQ files and

BLAST-aligned against the reference sequences: *Neisseria gonorrhoeae* FA 1090 *gyrA*

reference (NC_002946.2:c621189-618439), *Mycoplasma genitalium* strain G-37 *23S rRNA*

gene (NR_077054.1) and *Trichomonas vaginalis* G3 nitroreductase family protein reference

(TVAG_354010) respectively. Part of the aligned sequences flanking AMR-associated

mutations were used to construct a consensus sequence using MultAlin [19]. AMR-associated

mutations were confirmed by manually comparing the consensus sequence against the

reference.

Results

STI identification of swab samples by qPCR

A total of 200 vulvo-vaginal swab samples from FSWs were initially screened by qPCR, and

43 were found positive, including 37 single infections (11 MG, 19 TV and 7 CT) and six co-

infections (two CT/TV and one each of CT/NG, NG/TV, CT/NG/TV and CT/TV/MG), as

shown in Table S1 (Supplementary data).

STI identification and AMR detection by single gene tPCR nanopore sequencing

All 43 qPCR positives plus 12 negatives (3 for each pathogen) and an NG positive control

were subjected to single gene tPCR nanopore sequencing in one barcoded sequencing library.

A sequencing run of 10 hours produced 1,499,872 reads of which 1,127,703 (76%) reads

passed QC, with total yield of 1.2 gigabases, average qscore 8.34 (Figure S1, Supplementary

data) and average sequence length 816 bases.   To identify MG, NG and CT, all reads

(1,127,703) which passed QC were analyzed by WIMP, resulting in 810,683 (72%) classified

reads of which 43,491 (5.4%) were non-barcoded. For TV identification, Fastq Custom

Alignment workflow produced 148,458/1,127,703 (13%) read alignments with TV *ntr6*

sequence of which 66,500 (5.9%) were non-barcoded.

Targeted amplicon sequencing from clinical samples, as expected, still non-specifically

produced some reads from other microbial and/or host DNA in the sample. However, for STI

pathogens, qPCR positive samples generally yielded higher specific read counts than qPCR

negative controls, all of the latter of which had fewer than 15 reads (Figure 1). Mean $\log_{10}$

read counts were higher in qPCR positive NG, MG and CT samples compared to negative by

t-test (Table S2, Supplementary data) but not significantly for TV as 14/24 TV read counts in

TV qPCR positive samples were very low (<20 reads). NG, MG and CT read counts clearly

separated qPCR positive and negative samples except for one MG qPCR positive sample that

had an absolute read count of 1 (Figure 1). The samples having <20 reads generally had a low

load of pathogen, as evidenced by qPCR cycle threshold (Ct) being >35 (Table S1,

Supplementary data).

Figure 1. Pathogen read counts in the qPCR positive and negative STI samples analyzed

by single gene targeted PCR nanopore sequencing

The ARMA workflow assigned 86,430/810,683 (11%) classified reads to antibiotic resistance

genes in the CARD database, >99.99% of which were identified to be fluoroquinolone

resistant *gyrA* gene in NG, using protein variant model. This analysis indicated that 3/3 NG

clinical strains identified were fluoroquinolone-resistant, while none of the MG strains were

resistant to macrolide antibiotics. The NG positive control, phenotypically susceptible to

fluoroquinolones, and genotypically wild-type by BLAST (see below), was mis-identified to

be fluoroquinolone-resistant. TV antimicrobial resistance was not profiled due to the absence

of TV *ntr6* entries in the databases.

AMR confirmation by manual BLAST search

A manual BLAST search was performed to ascertain the accuracy of the pathogen AMR

identified by the ARMA workflow. The BLAST confirmed, as shown in Table 2, that all (3/3)

of NG had fluoroquinolone-resistant mutations, 2/10 (20%) of TV had mutations related to

metronidazole resistance, and none of MG had macrolide resistance-associated mutations. It

also showed that the gyrA of NG positive control did not possess fluoroquinolone-resistant

mutations.

Table 2. BLAST confirmation of antimicrobial resistance associated mutations identified by single gene targeted PCR nanopore sequencing*

Bacterial gene	Sample	Sequence of antimicrobial resistance region
Neisseria gonorrhoeae: *gyrA*	Wild type	-**T**CCGCAGTTTACG**A**C-
	All NG samples	-**TT**CGCAGTTTACG**CC**-
	Positive control	-**T**CCGCAGTTTACG**A**C-
Mycoplasma genitalium: 23S rRNA gene	Wild type	-CGGGACGG**AA**AGACC-
	All MG samples	-CGGGACGG**AA**AGACC-
Trichomonas vaginalis: *ntr6*	Wild type	-AATGCAA**A**AGCAGAC-
	2 TV samples	-AATGCAA**T**AGCAGAC-
	8 TV samples	-AATGCAA**A**AGCAGAC-
*One or more changes of the bold bases in the wild type cause antimicrobial resistance of strains		

Discussion

We demonstrated the potential to test four common genital STIs and simultaneously detect AMR in three of them among vulnerable FSWs in Quito, Ecuador using the portable MinION DNA sequencer, controlled by smartphone, through a single gene targeted approach.

Single gene approaches for detection of infection, such as 16S rRNA sequencing have been increasingly described [20,21], particularly with the advent of long read sequencing such as PacBio and ONT [22], which has greater likelihood of taxonomically identifying organisms at species level. We intentionally evaluated the accuracy of sequencing of a single AMR gene, as opposed to a two-stage gene approach [23], to both diagnose infection and predict AMR simultaneously in order to evaluate potential for use in resource poor field settings where simplicity and cost become increasingly important factors to consider.

Our approach appeared to have some advantages over NAAT. We demonstrated initially that genes responsible for AMR have potential to be as useful as 16S rRNA genes for diagnosis and thus may serve as a 'two-in-one' target. We chose for both pathogen and AMR detection the NG *gyrA*, MG *23S rRNA* and TV *ntr6*, all gene mutations in which have been shown to be predictive of AMR [7-12,24,25], and CT *omp-1* for CT detection [26]. As sequencing allows for identification of evolving mutations in the same genes without changing the test, our approach will therefore have value as it is adapted to other gene targets undergoing continuous evolution under selection pressure such as penA in which changes may give rise to penicillin and extended spectrum cephalosporin resistance [27]. The approach can also allow for sequencing multiple genes simultaneously alongside vigilance over any changing nature of phenotypic to genotypic association for antibiotics where AMR involves multiple mechanisms.

Clinical STI samples contain a wide range of pathogen loads being drawn frequently from

largely asymptomatic patients as in the case of the FSW cohort in this study. We found that

except for TV, diagnosis by the single gene targeted nanopore sequencing appeared to be

hardly limited by pathogen loads. In fact, our single gene approach correctly identified STI

infections in the samples which had qPCR Ct<35 (higher pathogen loads). In the group with

lower pathogen loads (Ct>35), tPCR amplification appeared not to enhance the absolute read

counts of pathogen targets, possibly due to low tPCR efficiency caused by the presence of a

large quantity of background DNA and non-optimal tPCR conditions. Further optimization

of tPCR conditions may maximize the diagnostic sensitivity for this group of samples.

Regarding the Fastq Antimicrobial Resistance workflow used in this study, the component

WIMP correctly identified each pathogen of interest, but we found the component ARMA

CARD appeared to mis-profile the AMR of the NG control strain. This might be because the

ARMA CARD used individual reads for matching AMR associated mutations, instead of a

consensus sequence, supported by the result obtained with our manual BLAST search using

the consensus sequence. Any further development of ARMA could help improve accuracy of

AMR prediction.

This study had some limitations. Firstly, there was a small sample size, which impacted

particularly on measures of specificity. Secondly no further confirmatory sequencing was

performed to confirm the mutations present, largely due to the local regulations which

prevented transporting material out of Ecuador. Thus, this study should be regarded as a

proof of concept study. Future work is required on larger sample sets with composite

reference standards in a formal diagnostic evaluation. Optimisation of amplification and

target multiplexing will also be important.

Nanopore sequencing with MinION requires substantially lower infrastructure and startup

cost compared to other high throughput sequencing platforms. Although arguably

consumables are not optimally priced for resource poor settings, the promise of combining

automated library preparations and use of disposable flow-cells has potential to use nanopore

sequencing for field diagnostics. We recently demonstrated that even in developed settings

implementation of rapid ciprofloxacin NAAT resistance tests for NG still requires net

investment [28]. More recently ONT has developed a tablet version, MinION™ Mk1C which

enables full sequencing and analysis to be performed in the lab and field without need for

internet connection. This development could reduce cost and improve speed further to use

nanopore sequencing as a diagnostic tool. However, this will be formally evaluated in a large-

scale clinical study.

In conclusion, this study demonstrated that single gene targeted nanopore sequencing for

diagnosing and simultaneously identifying key antimicrobial resistance markers for four

common STIs shows promise. Further work to optimise accuracy, reduce costs and improve

speed may allow sustainable approaches for managing STIs and emerging AMR in resource

poor and laboratory limited settings.

Contributions

STS and LZ conceived the study. NRS, PC and STS conceived the wider study for FSWs and

developed protocols for recruitment of patients with LML and ALR. ALR, LZ and LML

performed qPCR. LZ performed tPCR and nanopore sequencing. LZ and STS performed

analysis. LZ and STS wrote the paper with contributions from all authors.

Acknowledgements

The authors would like to thank the FSWs associations for supporting the development of this

study.

Funding

This work was supported by University Internacional del Ecuador (grant number UIDE-

EDM04-2016-2017), the Wellcome Trust Institutional Strategic Support Fund (grant number

204809/Z/16/Z) awarded to St. George's, University of London. The sponsors had no role in

the study design, data collection, and analysis, or in the decision to prepare these data for

publication.

Transparency declarations

LZ and STS are inventors on patents in the field of STI diagnostics.

References

1. Rowley J, Vander Hoorn S, Korenromp E *et al*. Chlamydia, gonorrhoea,

trichomoniasis and syphilis: global prevalence and incidence estimates, 2016. Bull

World Health Organ. 2019;97(8):548-562P. doi:10.2471/BLT.18.228486

2. Golden MR, Workowski KA, Bolan G. Developing a Public Health Response to

Mycoplasma genitalium. J Infect Dis. 2017;216(suppl_2):S420–6.

3. Soni S, Horner P, Rayment M, Pinto-Sander N, Naous N, Parkhouse A, Bancroft D,

Patterson C, Fifer H. British Association for Sexual Health and HIV national

guideline for the management of infection with Mycoplasma genitalium (2018). Int J

STD AIDS. 2019 Sep;30(10):938-950. doi: 10.1177/0956462419825948. Epub 2019

Jul 7. Erratum in: Int J STD AIDS. 2019 Aug 12;:956462419870463. PMID:

31280688.

4. Wi T, Lahra MM, Ndowa F *et al.* Antimicrobial resistance in Neisseria gonorrhoeae:

Global surveillance and a call for international collaborative action. PLoS Med.

2017;14(7):e1002344. Published 2017 Jul 7. doi:10.1371/journal.pmed.1002344

5. Ndowa F, Lusti-Narasimhan M, Unemo M. The serious threat of multidrug-resistant

and untreatable gonorrhoea: the pressing need for global action to control the spread

of antimicrobial resistance, and mitigate the impact on sexual and reproductive health.

Sex Transm Infect. 2012;88(5):317-318. doi:10.1136/sextrans-2012-050674

6. Muller EE, Mahlangu MP, Lewis DA *et al*. Macrolide and fluoroquinolone resistance-

associated mutations in Mycoplasma genitalium in Johannesburg, South Africa, 2007-

2014. BMC Infect Dis. 2019;19(1):148. Published 2019 Feb 13. doi:10.1186/s12879-

019-3797-6

7. Donà V, Low N, Golparian D *et al*. Recent advances in the development and use of

molecular tests to predict antimicrobial resistance in Neisseria gonorrhoeae. Expert

Rev Mol Diagn. 2017;17(9):845-859. doi:10.1080/14737159.2017.1360137

8. Pond MJ, Hall CL, Miari VF *et al*. Accurate detection of Neisseria gonorrhoeae

ciprofloxacin susceptibility directly from genital and extragenital clinical samples:

towards genotype-guided antimicrobial therapy. J Antimicrob Chemother.

2016;71(4):897-902. doi:10.1093/jac/dkv432

9. Jensen JS, Bradshaw CS, Tabrizi SN *et al*. Azithromycin treatment failure in

Mycoplasma genitalium-positive patients with nongonococcal urethritis is associated

with induced macrolide resistance. Clin Infect Dis. 2008;47(12):1546-1553.

doi:10.1086/593188

10. Yew HS, Anderson T, Coughlan E *et al*. Induced macrolide resistance in Mycoplasma

genitalium isolates from patients with recurrent nongonococcal urethritis. J Clin

Microbiol. 2011;49(4):1695-1696. doi:10.1128/JCM.02475-10

11. Ito S, Shimada Y, Yamaguchi Y *et al*. Selection of Mycoplasma genitalium strains

harbouring macrolide resistance-associated 23S rRNA mutations by treatment with a

single 1 g dose of azithromycin. Sex Transm Infect. 2011;87(5):412-414.

doi:10.1136/sextrans-2011-050035

12. Paulish-Miller TE, Augostini P, Schuyler JA *et al*. Trichomonas vaginalis

metronidazole resistance is associated with single nucleotide polymorphisms in the

nitroreductase genes ntr4Tv and ntr6Tv. Antimicrobial Agents and Chemotherapy.

2014 May;58(5):2938-2943. DOI: 10.1128/aac.02370-13.

13. Köser CU, Ellington MJ, Peacock SJ. Whole-genome sequencing to control

antimicrobial resistance. Trends Genet. 2014 Sep;30(9):401-7. doi:

10.1016/j.tig.2014.07.003. Epub 2014 Aug 3.

14. Schürch AC, van Schaik W. Challenges and opportunities for whole-genome

sequencing-based surveillance of antibiotic resistance. Ann N Y Acad Sci.

2017;1388(1):108-120. doi:10.1111/nyas.13310

15. Pillay A, Radebe F, Fehler G *et al*. Comparison of a TaqMan-based real-time

polymerase chain reaction with conventional tests for the detection of Trichomonas

vaginalis. Sex Transm Infect. 2007;83(2):126-129. doi:10.1136/sti.2006.022376

16. Jensen JS, Björnelius E, Dohn B *et al*. Use of TaqMan 5' nuclease real-time PCR for

quantitative detection of Mycoplasma genitalium DNA in males with and without

urethritis who were attendees at a sexually transmitted disease clinic. J Clin Microbiol.

2004;42(2):683-692. doi:10.1128/jcm.42.2.683-692.2004

17. Jaton K, Bille J, Greub G. A novel real-time PCR to detect Chlamydia trachomatis in

first-void urine or genital swabs. J Med Microbiol. 2006;55(Pt 12):1667-1674.

doi:10.1099/jmm.0.46675-0

18. Tabrizi SN, Chen S, Tapsall J *et al*. Evaluation of opa-based real-time PCR for

detection of Neisseria gonorrhoeae. Sex Transm Dis. 2005;32(3):199-202.

doi:10.1097/01.olq.0000154495.24519.bf

19. Corpet F. Multiple sequence alignment with hierarchical clustering. Nucleic Acids Res. 1988;16(22):10881-10890. doi:10.1093/nar/16.22.10881

20. Janda JM, Abbott SL. 16S rRNA gene sequencing for bacterial identification in the diagnostic laboratory: pluses, perils, and pitfalls. J Clin Microbiol. 2007;45(9):2761-2764. doi:10.1128/JCM.01228-07

21. Woo PC, Lau SK, Teng JL *et al*. Then and now: use of 16S rDNA gene sequencing for bacterial identification and discovery of novel bacteria in clinical microbiology laboratories. Clin Microbiol Infect. 2008;14(10):908-934. doi:10.1111/j.1469-0691.2008.02070.x

22. Johnson JS, Spakowicz DJ, Hong BY *et al*. Evaluation of 16S rRNA gene sequencing for species and strain-level microbiome analysis. Nat Commun. 2019;10(1):5029. Published 2019 Nov 6. doi:10.1038/s41467-019-13036-1

23. Graham RM, Doyle CJ, Jennison AV. Epidemiological typing of Neisseria gonorrhoeae and detection of markers associated with antimicrobial resistance directly from urine samples using next generation sequencing. Sex Transm Infect. 2017;93(1):65-67. doi:10.1136/sextrans-2015-052422

24. Hemarajata P, Yang S, Soge OO *et al*. Performance and Verification of a Real-Time PCR Assay Targeting the gyrA Gene for Prediction of Ciprofloxacin Resistance in Neisseria gonorrhoeae. J Clin Microbiol. 2016;54(3):805-808. doi:10.1128/JCM.03032-15.

25. de Korne-Elenbaas J, Pol A, Vet J *et al*. Simultaneous Detection of Neisseria gonorrhoeae and Fluoroquinolone Resistance Mutations to Enable Rapid Prescription of Oral Antibiotics. Sex Transm Dis. 2020;47(4):238-242. doi:10.1097/OLQ.0000000000001141

26. Molano M, Meijer CJ, Morré SA *et al*. Combination of PCR targeting the VD2 of

omp1 and reverse line blot analysis for typing of urogenital Chlamydia trachomatis

serovars in cervical scrape specimens. J Clin Microbiol. 2004;42(7):2935-2939.

doi:10.1128/JCM.42.7.2935-2939.2004

27. Yahara K, Ma KC, Mortimer TD *et al*. Emergence and evolution of antimicrobial

resistance genes and mutations in Neisseria gonorrhoeae. bioRxiv 2020.10.26.354993;

doi: https://doi.org/10.1101/2020.10.26.354993

28. Harding-Esch EM, Huntington SE , Harvey MJ *et al*. Antimicrobial resistance point-

of-care testing for gonorrhoea treatment regimens: cost-effectiveness and impact on

ceftriaxone use of five hypothetical strategies compared with standard care in England

sexual   health   clinics.   Euro   Surveill.   2020;25(43):pii=1900402.

https://doi.org/10.2807/1560-7917.ES.2020.25.43.1900402

401