**Single gene targeted nanopore sequencing enables simultaneous identification and**

**antimicrobial resistance detection of sexually transmitted infections**

Liqing Zhou[1], Andrea Lopez Rodas[2], Luz Marina Llangarí[2], Natalia Romero Sandoval[2,3],

Philip Cooper[1,2], Syed Tariq Sadiq[1*]

[1] Applied Diagnostic Research and Evaluation Unit

Institute for Infection and Immunity

St George's, University of London

Cranmer Terrace

London SW17 0RE

UK

[2] School of Medicine,

Faculty of Health and Life Sciences

Universidad Internacional del Ecuador

Quito, Ecuador

[3] Grups de Recerca d'Amèrica i Àfrica Llatines, GRAAL, Barcelona, Spain

* Corresponding author

ssadiq@sgul.ac.uk (STS)

**Abstract**

**Objectives**

To develop a simple DNA sequencing test for simultaneous identification and antimicrobial

resistance (AMR) detection of multiple sexually transmitted infections (STIs).

**Methods**

A total of 200 vulvo-vaginal swab samples from 200 female sex workers in Ecuador were

initially tested by real-time PCR for *Neisseria gonorrhoeae* (NG), *Chlamydia trachomatis* (CT),

*Mycoplasma genitalium* (MG) and *Trichomonas vaginalis* (TV). Samples positive for these

STIs and controls were subjected to single gene targeted PCR MinION-nanopore sequencing

by using the smartphone operated MinIT.

**Results**

Among 200 vulvo-vaginal swab samples 43 were positive for at least one of the STIs by real-

time PCR. Single gene targeted nanopore sequencing generally yielded higher pathogen

specific reads in positive samples than negative controls. Of the 26 CT, NG or MG infections

identified by real-time PCR, 25 were clearly distinguishable from the negative controls using

sequence read count. Discrimination of TV positives from controls was poorer as many had

low pathogen loads (cycle threshold > 35) with associated low sequence read counts. AMR

analysis workflow indicated that 11% of the classified reads aligned with antibiotic resistance

genes, all of which identified fluoroquinolone resistance in NG.

**Conclusions**

Single gene targeted nanopore sequencing for diagnosing and simultaneously identifying key

antimicrobial resistance markers for four common STIs, associated with genital tract discharge

in women shows promise. Further work to optimise accuracy, reduce costs and improve speed

may allow sustainable approaches for managing STIs and emerging AMR in resource poor and

laboratory limited settings.

## Introduction

Sexually transmitted infections (STIs) remain a major public health problem worldwide, with an estimated 357 million new cases of *Chlamydia trachomatis* (CT), *Neisseria gonorrhoeae* (NG), *Treponema pallidum* (TP) and *Trichomonas vaginalis* (TV) per year.[1] High rates of the STI, *Mycoplasma genitalium* (MG), have also been reported worldwide.[2] CT, NG, MG and TV commonly cause a genital discharge syndrome and are generally curable with existing, effective single-dose antibiotic regimens but, if left undiagnosed and/or untreated, can result in serious long-term reproductive health sequelae, particularly for women. Antimicrobial resistance (AMR) among some STIs to multiple classes of antibiotics has spread rapidly in recent years. For NG, loss of extended spectrum cephalosporins as first line empirical treatment is a major concern[3] as circulating multi- and extensive-drug resistance clones have been detected internationally.[4] For MG, macrolide resistance is now widely but not universally reported with increasing rates of resistance to fluoroquinolones also detected.[5] These developments have made treatment, and particularly empirical treatment challenging.

New World Health Organization (WHO) guidelines reinforce the need to treat these STIs with the right antibiotic, at the right dose, and the right time to reduce spread and improve sexual and reproductive health. For both NG and MG, accurate and rapid diagnostics which also predict antibiotic susceptibility are likely to be needed to achieve this. Laboratory based nucleic-acid amplification tests (NAAT) for detection of NG, CT, TV and MG are the current gold standard for detection, providing high sensitivity and specificity. NAAT are widely used in high-income countries but are often unavailable in resource-poor settings. Where testing is available, it is still frequently expensive and geographically inaccessible.[6] Culture-based antimicrobial susceptibility testing (AST) for NG remains gold standard for phenotypically predicting AMR and takes usually two to five days to obtain a result. This in practical terms may be too late to initiate targeted antibiotic therapy, particularly for hard to reach vulnerable populations. MG is difficult to cultivate and requires cell culture, not usually feasible in most clinical settings.

Advances in understanding AMR in NG, MG and TV have allowed for development of NAAT-based AMR detection. For example, absence of *gyrA* mutations in NG accurately predicts fluoroquinolone susceptibility,[7,8] presence of *23S rRNA* mutations in MG is associated with failure of treatment with azithromycin,[9-11] and *ntr6* mutations in TV may have diagnostic value for metronidazole resistance.[12] Use of NAAT-based AMR tests, however, may have limitations due to continually changing mutations and novel mechanisms of resistance evolving under ongoing treatment selection pressures. Whole genome sequencing (WGS) using high

throughput sequencing platforms may address this challenge to some degree and give added
value in identifying phylogenetic relationships in identified infections.[13,14] However, for
diagnostic purposes WGS itself can be constrained by need to first culture from clinical
samples or to use nucleic acid capture techniques, increasing turnaround time and cost, and
making them additionally unsuitable for near patient applications.
Oxford Nanopore Technologies' (ONT) portable MinION DNA sequencer, together with its
recent MinIT hand-held processor, may offer advantages as an accurate diagnostic in resource-
limited settings. Herein we report early phase evaluation among female sex workers (FSWs)
in Ecuador where the STIs are prevalent,[15] of the MinION, controlled by a smartphone-
operated MinIT to detect the four common STIs and profile AMR to ciprofloxacin in NG,
azithromycin in MG and metronidazole in TV utilising a single gene targeted PCR approach,
followed by long read sequencing in clinical samples.

**Materials and Methods**

*Clinical sample collection and DNA preparation*

A cross-sectional study of STIs was conducted among FSWs in a primary health centre in Quito,
Ecuador, during the last quarter of 2017. Written and verbal informed consent was obtained
from all participants all of whom were 18 years or older and none of whom received any
compensation for their participation. Participants were interviewed by a local study investigator
using a questionnaire designed for the study. The study was conducted according to the
Declaration of Helsinki and approved by the Ethical Committee of Universidad Internacional
del Ecuador (02-02-17).
Vulvo-vaginal swab samples were collected by clinicians using Xpert® CT/NG Patient-
Collected Vaginal Swab Specimen Collection Kit (Cepheid). DNA from swab samples was
prepared using PureLink Genomic DNA Mini Kit (Thermo Fisher Scientific) according to the
manufacturer's instructions and quantified using a NanoDrop Spectrometer. All nanopore
sequencing and experiments were done in Ecuador.

*STI identification of swab samples by real-time PCR*

DNA prepared from the swab samples was initially tested by real-time PCR for NG, TV, MG
and CT infections, with gene targets and primers for real-time PCR detection shown in Table
1a. Real-time PCR was performed using Applied Biosystems 7500 Fast Real-Time PCR
System in a volume of 10 µl containing 5 µl of TaqMan™ Fast Universal PCR Master Mix, 1
µl of 10x Exogenous Internal Positive Control (IPC) Mix, 0.2 µl of 50x IPC DNA, 250 nM

each of the primers, 100 nM each of the probes and 50 ng of template DNA. Cycling parameters

were 95˚C for 10 min, followed by 40 cycles of 95°C for 15 sec, 60°C for 1 min.

**Table 1 Gene targets, primers and probes used in this study**

| a) Initial real-time PCR for STI testing on samples |
| --- |
| *Trichomonas vaginalis*: *T. vaginalis* specific repeat DNA fragment[16] <br>      TV_F: AAAGATGGGTGTTTTAAGCTAGATAAGG <br>      TV_R: TCTGTGCCGTCTTCAAGTATGC <br>      TV_P: [6FAM]AGTTCATGTCCTCTCCAAGCGTAAGT[BHQ1] <br><br> *Mycoplasma genitalium*: MgPa gene[17] <br>      MG_F: GAGAAATACCTTGATGGTCAGCAA <br>      MG_R: GTTAATATCATATAAAGCTCTACCGTTGTTATC <br>      MG_P: [6FAM]ACTTTGCAATCAGAAGGT[MGBNFQ] <br><br> *Chlamydia trachomatis*: The cryptic plasmid[18] <br>      CT_F: CATGAAAACTCGTTCCGAAATAGAA <br>      CT_R: TCAGAGCTTTACCTAACAACGCATA <br>      CT_P: [6FAM]TCGCATGCAAGATATCGA[MGBNFQ] <br><br> *Neisseria gonorrhoeae*: *opa*[19] <br>      NG_F: TTGAAACACCGCCCGGAA <br>      NG_R: TTTCGGCTCCTTATTCGGTTTGA <br>      NG_P: [6FAM]CCGATATAATCCGCCCTTCAACATCAG[BHQ1] |
| **b) Targeted PCR for nanopore sequencing\*** |
| *Trichomonas vaginalis*: Nitroreductase family protein (*ntr6*) <br>      TV_NTR6_BC_F: <br>      5'TTTCTGTTGGTGCTGATATTGCCTTCATTGAATTTATTCGTTCAAAATT <br>      TV_NTR6_BC_R: <br>      5'ACTTGCCTGTCGCTCTATCTTCTTATTCAATGTATGTAACCTTTCTAA <br><br> *Mycoplasma genitalium*: 23S ribosomal RNA <br>      Mg 23S_BC_1992F: <br>      5' TTTCTGTTGGTGCTGATATTGCCCATCTCTTGACTGTCTCGG <br>      Mg 23S_BC_2679R: <br>      5' ACTTGCCTGTCGCTCTATCTTCTCCTCTCGTACTAGAAGCAAAG <br><br> *Chlamydia trachomatis*: Major outer membrane protein 1 (*omp1*) <br>      CT_OMP1_BC_F: <br>      5' TTTCTGTTGGTGCTGATATTGCTTTGCCGCTTTGAGTTCTGCT <br>      CT_OMP1_BC_R: <br>      5' ACTTGCCTGTCGCTCTATCTTCCAATACCGCAAGATTTTCTAGATTTC <br><br> *Neisseria gonorrhoeae*: *gyrA* <br>      NG_gyrA_BC_F: <br>      5' TTTCTGTTGGTGCTGATATTGCATCCGCCACGACCACAAATT |

| NG_gyrA_BC_R: 5' ACTTGCCTGTCGCTCTATCTTCATATTGGACAGTGCGACGGC |
|---|
| *developed in this study |

### Targeted PCR amplification and barcoding

DNA of NG, MG, TV and CT positive samples was amplified by PCR, targeting genes *gyrA*

(NG), *23S rRNA* (MG), *ntr6* (TV) and *omp1* (CT) with the primers listed in Table 1b. PCR

reactions for each target were performed in a 50 µl volume consisting of 25 µl LongAmp® Taq

2X Master Mix (NEB), 200 nM each of PCR primers and 100 ng DNA template as follows:

95 °C for 3 min followed by 35 cycles of 95 °C for 30 sec, 58 °C for 30 sec, 65 °C for 1 min

30 sec, and one cycle of 65 °C for 5 min.  PCR amplicons were purified and subjected to a

second PCR for barcoding using PCR Barcoding Expansion 1-96 - EXP-PBC096 (Oxford

Nanopore Technologies) according to the manufacturer's instructions.

### Nanopore sequencing and data analysis

A DNA sequencing library was constructed from a pool of  barcoded PCR amplicons of

different clinical samples using Ligation Sequencing Kit - SQK-LSK108 and sequenced using

FLO-MIN106 R9 Version Flow Cell MK I Spot-ON on portable DNA sequencer MinION MK

I which was controlled by a smartphone operated MinIT (Oxford Nanopore Technologies)

according to the manufacturer's instructions. Sequencing data were uploaded onto Metrichor

Epi2ME (Oxford Nanopore Technologies) and analysed using the workflow Fastq

Antimicrobial Resistance r3.3.2 which included three components: QC and Barcoding [rev.

3.10.4], WIMP [rev. 3.4.0] and ARMA CARD [rev. 1.1.6]. QC and Barcoding component

contains quality score (qscore) and barcode filter, cutting off reads with a qscore below the

threshold (min qscore: 7) which will not be uploaded for analysis, and demultiplexing barcodes

if the sequenced samples are barcoded. A qscore is calculated as Phred scores, ie -10log(Pe),

where Pe is the probability of a particular base call. The mean qscore for a read directly

corresponds to the alignment accuracy. WIMP (what's In My Pot) allows for species

identification by classifying read sequences against the standard Centrifuge database (including

RefSeq complete genomes for bacteria). ARMA (antimicrobial resistance mapping application)

performs antimicrobial resistance identification by aligning input reads with minmap2 against

all reference sequences available in the CARD database (CARD version 1.1.3). TV, an

anaerobic, flagellated protozoan parasite, was not included in RefSeq for bacteria and therefore

a TV G3 nitroreductase family protein (TVAG_354010, ntr6) reference was uploaded using

Fasta Reference Upload r3.2.2 and TV identification was performed using Fastq Custom
Alignment r3.2.2 against the *ntr6* reference sequence. Read counts were log-transformed and
compared between positive and negative samples by t-test.

***BLAST search***

NG, MG and TV read DNA sequences were extracted from read FASTQ files and BLAST-
aligned against the reference sequences: *Neisseria gonorrhoeae* FA 1090 *gyrA* reference
(NC_002946.2:c621189-618439), *Mycoplasma genitalium* strain G-37 *23S rRNA* gene
(NR_077054.1) and *Trichomonas vaginalis* G3 nitroreductase family protein reference
(TVAG_354010) respectively. Part of the aligned read sequences flanking AMR gene
mutations were aligned to construct a consensus sequence using MultAlin.[20] AMR gene
mutations were confirmed by manually comparing the consensus sequence against the
reference.

**Results**

***STI identification of swab samples by real-time PCR***

A total of 200 vulvo-vaginal swab samples from 200 FSWs were initially tested by real-time
PCR, and 43 were found positive, including 37 single infections (11 MG, 19 TV and 7 CT) and
six co-infections (two CT/TV and one each of CT/NG, NG/TV, CT/NG/TV and CT/TV/MG).
These positive samples together with their Ct values are shown in Table S1 (Supplementary
data).

***STI identification and AMR detection by single gene targeted PCR nanopore sequencing***
All 43 PCR positives plus 12 negatives (3 for each pathogen) and an NG positive control were
amplified by PCR. The amplicons from each sample were barcoded and combined in one
sequencing library. A sequencing run of 10 hours produced 1,499,872 reads of which 1,127,703
(76%) reads passed QC (min qscore: 7), with total yield of 1.2 gigabases, average qscore 8.34
(Figure S1, Supplementary data) and average sequence length 816 bases.  To identify MG,
NG and CT, the 1,127,703 reads were analyzed by WIMP, resulting in 810,683 (72%)
classified reads of which 43,491 (5.4%) were non-barcoded. For TV identification, Fastq
Custom Alignment workflow was used, producing 148,458/1,127,703 (13%) read alignments
with TV *ntr6* sequence of which 66,500 (5.9%) were non-barcoded.
Targeted amplicon sequencing from clinical samples, as expected,  still non-specifically
produced some sequence reads from other microbial and/or host DNA in the sample. However,

for all target pathogens, PCR positive samples generally yielded higher pathogen specific read

counts than PCR negative controls, all of the latter of which had fewer than 15 reads (Figure

1). Mean $\log_{10}$ read counts were higher in positive NG, MG and CT samples compared to

negative by t-test (Table S2, Supplementary data) but not significantly for TV as 14/24 TV

read counts in TV positive samples were very low (<20 reads).   For NG, MG and CT read

counts clearly separated PCR positive and negative samples except for one MG positive sample

that had an absolute read count of 1 (see Figure 1). The samples having <20 reads generally

had a low load of pathogen, as evidenced by their Ct values being >35 (Table S1,

Supplementary data).

**Figure 1. Pathogen read counts in the clinical swab samples analyzed by single gene**

**targeted PCR nanopore sequencing**

All the classified reads were analyzed by the ARMA workflow and 86,430/810,683 (11%)

classified reads aligned with antibiotic resistance genes in the CARD database, > 99.99% of

which were identified to be fluoroquinolone resistant *gyrA* gene in NG, using protein variant

model. This analysis indicated that 3/3 NG clinical strains identified were resistant to

fluoroquinolones, while none of the MG strains were resistant to macrolide antibiotics.

Surprisingly, the NG positive control, which was phenotypically susceptible to

fluoroquinolones, and genotypically wild-type by BLAST (see below) was also identified to

be fluoroquinolone resistant by ARMA. TV antimicrobial resistance was not profiled due to

the absence of TV *ntr6* entries in the databases.

**Confirmation of antimicrobial resistance by manual BLAST search**

A manual BLAST search was performed to ascertain the accuracy of the pathogen AMR
identified through the sequence analysis of the ARMA workflow. The manual BLAST search
confirmed, as shown in Table 2, that all (3/3) of NG had fluoroquinolone-resistant mutations,
2/10 (20%) of TV had mutations related to metronidazole resistance, and none of MG had
macrolide resistance-associated mutations. However, the manual BLAST search confirmed
that the NG positive control was incorrectly identified to be fluoroquinolone resistant by
ARMA workflow.

**Table 2. BLAST confirmation of antimicrobial resistance associated mutations identified**
**by single gene targeted PCR nanopore sequencing***

| Gene | Sample | Sequence of antimicrobial resistance region |
|---|---|---|
| NG *gyrA* | Wild type | -T**C**CGCAGTTTACG**A**C- |
| | All NG samples | -T**T**CGCAGTTTACG**C**C- |
| | Positive control | -T**C**CGCAGTTTACG**A**C- |
| MG 23S rRNA gene | Wild type | -CGGGACGG**AA**AGACC- |
| | All MG samples | -CGGGACGG**AA**AGACC- |
| TV *ntr6* | Wild type | -AATGCAA**A**AGCAGAC- |
| | 2 TV samples | -AATGCAA**T**AGCAGAC- |
| | 8 TV samples | -AATGCAA**A**AGCAGAC- |
| *One or more changes of the bold bases in the wild type cause antimicrobial resistance of strains | | |

**Discussion**
We demonstrated the potential to test four common STIs responsible for genital discharge
syndrome and simultaneously detect AMR in three of them among vulnerable FSWs in Quito,
Ecuador using the portable MinION DNA sequencer, controlled by smartphone, through a
single gene targeted sequencing approach.
Single gene approaches for detection of infection, such as 16S rRNA sequencing have been
increasingly described,[21,22] particularly with the advent of long read sequencing such as PacBio
and ONT[23], which has greater likelihood of taxonomically identifying organisms at species
level. We intentionally evaluated the accuracy of sequencing of a single AMR gene, as opposed
to a two-stage gene approach,[24] to both diagnose infection and predict AMR simultaneously in
order to evaluate potential for use in resource poor field settings where simplicity and cost
become increasingly important factors to consider.
Our approach appeared to have some advantages over NAAT. We demonstrated initially that
genes responsible for AMR have potential to be as useful as 16S rRNA genes for diagnosis and
thus may serve as a 'two-in-one' target. The approach is most advantageous for antimicrobials

for which the number of essential AMR genes are limited. We chose for both pathogen and

AMR detection the NG *gyrA*, MG *23S rRNA* and TV *ntr6*, all gene mutations in which have

been shown to be predictive of AMR,[7-12,25,26] and CT *omp-1* for CT detection.[27,28] Although for

*gyrA* in NG and to some extent *23S rRNA* in MG, mutations that confer resistance to

ciprofloxacin and azithromycin respectively appear to be constrained to limited possibilities,

for other genes not assessed here, such as *penA* in which changes may give rise to penicillin

and extended spectrum cephalosporin resistance, there appears to be continuous evolution

under selection pressure.[29] As sequencing allows for identification of evolving mutations in the

same genes without changing the test, our approach will therefore have value as it is adapted

to other gene targets, although for antibiotics where AMR involves multiple mechanisms,

multiple genes  would need to be sequenced alongside vigilance over any changing nature of

phenotypic to genotypic association. Although previously the possibility of identifying NG and

AMR markers on enriched NG positive urine samples has been demonstrated,[24] we did not

perform WGS directly from clinical samples because of the costly and complex process of

enrichment, host DNA depletion[30] and the need for sufficient pathogen load. Approaches that

directly sequence from clinical samples involving capture probe technology, immunomagnetic

separation and optimal extraction may in future allow for sequencing without amplification

routinely.[31-33]

Many clinical STI samples in women contain a wide range of pathogen loads being drawn

frequently from largely asymptomatic patients as in the case of the FSW cohort in this study.

We found that except for TV, diagnosis accuracy by the single gene targeted nanopore

sequencing appeared to be high and hardly limited by pathogen loads. Compared to the real-

time PCR assay, the single gene approach correctly identified STI infections in samples with

cycle thresholds (Ct) <35, that is with higher pathogen loads. In the group with lower pathogen

loads (CT>35), PCR amplification did not enhance the absolute read counts of pathogen targets,

possibly due to low PCR efficiency caused by the presence of a large quantity of background

DNA and non-optimal PCR conditions.  Further optimization of PCR conditions will be

required to maximize the diagnostic sensitivity for this group of samples.

Regarding the Fastq Antimicrobial Resistance workflow used in this study, the component

WIMP correctly identified each pathogen of interest, but we found the component ARMA

CARD appeared to mis-profile the AMR of the NG control strain.  This might be because the

ARMA CARD used individual reads for matching AMR associated mutations, instead of a

consensus sequence, supported by the result obtained with our manual BLAST search using

the consensus sequence. Any further development of ARMA could help improve accuracy of
AMR prediction.

Our study although small suggested fluoroquinolone resistance in circulating NG strains may
be high in Ecuador, consistent with neighboring Peru, where 82.3% of NG strains were
fluoroquinolone resistant.[34] However, we did not find macrolide resistance associated
mutations in *23S rRNA* gene of MG strains, even though macrolide-resistance of MG was
prevalent in the U.S.[35] and common worldwide.[36] Interestingly, we also previously described
lack of macrolide resistant MG in the Solomon islands[37] despite a high prevalence of infection
in the islands. Among the TV strains detected, 2/10 had the *A238T* SNP which results in a
premature stop codon in the *ntr6* gene that is associated with metronidazole resistance.[12] The
presence of this mutation has potential to be useful in indicating where metronidazole may fail,
but other mutations in a number of other genes are also likely to be important in confidently
identifying susceptibility.[38] Adding other gene targets might therefore be useful to guide
clinicians as to the reasons for treatment failure.

This study had some limitations. Firstly, there was a small sample size, impacting particularly
on measures of specificity. Although future development is currently using larger sample sizes,
species-specific sequencing gives inherent high specificity compared to real-time PCR.
Secondly no further confirmatory sequencing or NAAT were performed to confirm diagnoses
or mutations present, largely due to the local regulations which prevented transporting material
out of Ecuador. Thus, this study should not be regarded as a formal diagnostic evaluation.
Future work is required on larger sample sets with composite reference standards. Optimisation
of amplification and target multiplexing will also be important.

Nanopore sequencing with MinION requires substantially lower infrastructure and startup cost
compared to other high throughput sequencing platforms. Although arguably consumables are
not optimally priced for resource poor settings, the promise of combining automated library
preparations and use of disposable flow-cells has potential to use nanopore sequencing for field
diagnostics. We recently demonstrated that even in developed settings implementation of rapid
ciprofloxacin NAAT resistance tests for NG still requires net investment.[39] Small-scale local
automated approaches have further potential to increase turn-around times to within time
frames that may suite point-of care applications. More recently ONT has developed a tablet
version, MinION™ Mk1C which enables full sequencing and analysis to be performed in the
lab and field without need for internet connection. This development could reduce cost and
improve speed further to use nanopore sequencing as a diagnostic tool.

In conclusion, this study demonstrated that single gene targeted nanopore sequencing for
diagnosing and simultaneously identifying key antimicrobial resistance markers for four
common STIs, associated with genital tract discharge in women shows promise. Further work
to optimise accuracy, reduce costs and improve speed may allow sustainable approaches for
managing STIs and emerging AMR in resource poor and laboratory limited settings.
305

306 **Contributions**
307 STS and LZ conceived the study. NRS, PC and STS conceived the wider study for FSWs and
308 developed protocols for recruitment of patients with LML and ALR. ALR, LZ and LML
309 performed PCR. LZ performed nanopore sequencing. LZ and STS performed analysis. LZ and
310 STS wrote the paper with contributions from all authors.
311

312 **Acknowledgements**
313 The authors would like to thank the FSWs associations for supporting the development of this
314 study.
315

**Funding**
This study was supported by University Internacional del Ecuador [Grant Number UIDE-
EDM04-2016-2017] and the Wellcome Trust Institutional Strategic Support Fund
[204809/Z/16/Z] awarded to St. George's, University of London. The sponsors had no role in
the study design, data collection, and analysis, or in the decision to prepare these data for
publication.
322

323 **Transparency declarations**
324 LZ and STS are inventors on patents in the field of STI diagnostics.
325

**References**
1. Rowley J, Vander Hoorn S, Korenromp E, et al. Chlamydia, gonorrhoea,
trichomoniasis and syphilis: global prevalence and incidence estimates, 2016. Bull
World Health Organ. 2019;97(8):548-562P. doi:10.2471/BLT.18.228486
2. Golden MR, Workowski KA, Bolan G. Developing a Public Health Response to
Mycoplasma genitalium. J Infect Dis. 2017;216(suppl_2):S420–6.

3. Wi T, Lahra MM, Ndowa F, et al. Antimicrobial resistance in Neisseria gonorrhoeae: Global surveillance and a call for international collaborative action. PLoS Med. 2017;14(7):e1002344. Published 2017 Jul 7. doi:10.1371/journal.pmed.1002344

4. Ndowa F, Lusti-Narasimhan M, Unemo M. The serious threat of multidrug-resistant and untreatable gonorrhoea: the pressing need for global action to control the spread of antimicrobial resistance, and mitigate the impact on sexual and reproductive health. Sex Transm Infect. 2012;88(5):317-318. doi:10.1136/sextrans-2012-050674

5. Muller EE, Mahlangu MP, Lewis DA, Kularatne RS. Macrolide and fluoroquinolone resistance-associated mutations in Mycoplasma genitalium in Johannesburg, South Africa, 2007-2014. BMC Infect Dis. 2019;19(1):148. Published 2019 Feb 13. doi:10.1186/s12879-019-3797-6

6. Sexually transmitted infections (STIs). World Health Organization; 2019. Available at:https://www.who.int/news-room/fact-sheets/detail/sexually-transmitted-infections-(stis)

7. Donà V, Low N, Golparian D, Unemo M. Recent advances in the development and use of molecular tests to predict antimicrobial resistance in Neisseria gonorrhoeae. Expert Rev Mol Diagn. 2017;17(9):845-859. doi:10.1080/14737159.2017.1360137

8. Pond MJ, Hall CL, Miari VF, et al. Accurate detection of Neisseria gonorrhoeae ciprofloxacin susceptibility directly from genital and extragenital clinical samples: towards genotype-guided antimicrobial therapy. J Antimicrob Chemother. 2016;71(4):897-902. doi:10.1093/jac/dkv432

9. Jensen JS, Bradshaw CS, Tabrizi SN, et al. Azithromycin treatment failure in Mycoplasma genitalium-positive patients with nongonococcal urethritis is associated with induced macrolide resistance. Clin Infect Dis. 2008;47(12):1546-1553. doi:10.1086/593188

10. Yew HS, Anderson T, Coughlan E, Werno A. Induced macrolide resistance in Mycoplasma genitalium isolates from patients with recurrent nongonococcal urethritis. J Clin Microbiol. 2011;49(4):1695-1696. doi:10.1128/JCM.02475-10

11. Ito S, Shimada Y, Yamaguchi Y, et al. Selection of Mycoplasma genitalium strains harbouring macrolide resistance-associated 23S rRNA mutations by treatment with a single 1 g dose of azithromycin. Sex Transm Infect. 2011;87(5):412-414. doi:10.1136/sextrans-2011-050035

12. Paulish-Miller TE, Augostini P, Schuyler JA, et al. Trichomonas vaginalis metronidazole resistance is associated with single nucleotide polymorphisms in the

366       nitroreductase genes ntr4Tv and ntr6Tv. Antimicrobial Agents and Chemotherapy.
367       2014 May;58(5):2938-2943. DOI: 10.1128/aac.02370-13.

13. Köser CU, Ellington MJ, Peacock SJ. Whole-genome sequencing to control antimicrobial resistance. Trends Genet. 2014 Sep;30(9):401-7. doi: 10.1016/j.tig.2014.07.003. Epub 2014 Aug 3.

14. Schürch AC, van Schaik W. Challenges and opportunities for whole-genome sequencing-based surveillance of antibiotic resistance. Ann N Y Acad Sci. 2017;1388(1):108-120. doi:10.1111/nyas.13310

15. Llangarí-Arizo LM, Sadiq ST, Márquez C, et al. Biological and behavioural factors associated with risky sexual practices among female sex workers: a cross sectional study in a large Andean city. BMJ Open (Submitted)

16. Pillay A, Radebe F, Fehler G, et al. Comparison of a TaqMan-based real-time polymerase chain reaction with conventional tests for the detection of Trichomonas vaginalis. Sex Transm Infect. 2007;83(2):126-129. doi:10.1136/sti.2006.022376

17. Jensen JS, Björnelius E, Dohn B, Lidbrink P. Use of TaqMan 5' nuclease real-time PCR for quantitative detection of Mycoplasma genitalium DNA in males with and without urethritis who were attendees at a sexually transmitted disease clinic. J Clin Microbiol. 2004;42(2):683-692. doi:10.1128/jcm.42.2.683-692.2004

18. Jaton K, Bille J, Greub G. A novel real-time PCR to detect Chlamydia trachomatis in first-void urine or genital swabs. J Med Microbiol. 2006;55(Pt 12):1667-1674. doi:10.1099/jmm.0.46675-0

19. Tabrizi SN, Chen S, Tapsall J, Garland SM. Evaluation of opa-based real-time PCR for detection of Neisseria gonorrhoeae. Sex Transm Dis. 2005;32(3):199-202. doi:10.1097/01.olq.0000154495.24519.bf

20. Corpet F. Multiple sequence alignment with hierarchical clustering. Nucleic Acids Res. 1988;16(22):10881-10890. doi:10.1093/nar/16.22.10881

21. Janda JM, Abbott SL. 16S rRNA gene sequencing for bacterial identification in the diagnostic laboratory: pluses, perils, and pitfalls. J Clin Microbiol. 2007;45(9):2761-2764. doi:10.1128/JCM.01228-07

22. Woo PC, Lau SK, Teng JL, et al. Then and now: use of 16S rDNA gene sequencing for bacterial identification and discovery of novel bacteria in clinical microbiology laboratories. Clin Microbiol Infect. 2008;14(10):908-934. doi:10.1111/j.1469-0691.2008.02070.x

23. Johnson JS, Spakowicz DJ, Hong BY, et al. Evaluation of 16S rRNA gene sequencing
for species and strain-level microbiome analysis. Nat Commun. 2019;10(1):5029.
Published 2019 Nov 6. doi:10.1038/s41467-019-13036-1

24. Graham RM, Doyle CJ, Jennison AV. Epidemiological typing of Neisseria gonorrhoeae
and detection of markers associated with antimicrobial resistance directly from urine
samples using next generation sequencing. Sex Transm Infect. 2017;93(1):65-67.
doi:10.1136/sextrans-2015-052422

25. Hemarajata P, Yang S, Soge OO, et al. Performance and Verification of a Real-Time
PCR Assay Targeting the gyrA Gene for Prediction of Ciprofloxacin Resistance in
Neisseria　　gonorrhoeae.　　J　　Clin　　Microbiol.　　2016;54(3):805-808.
doi:10.1128/JCM.03032-15.

26. de Korne-Elenbaas J, Pol A, Vet J, et al. Simultaneous Detection of Neisseria
gonorrhoeae and Fluoroquinolone Resistance Mutations to Enable Rapid Prescription
of　　Oral　　Antibiotics.　　Sex　　Transm　　Dis.　　2020;47(4):238-242.
doi:10.1097/OLQ.0000000000001141

27. Molano M, Meijer CJ, Morré SA, et al. Combination of PCR targeting the VD2 of omp1
and reverse line blot analysis for typing of urogenital Chlamydia trachomatis serovars
in　　cervical　　scrape　　specimens.　　J　　Clin　　Microbiol.　　2004;42(7):2935-2939.
doi:10.1128/JCM.42.7.2935-2939.2004

28. Schaeffer A, Henrich B. Rapid detection of Chlamydia trachomatis and typing of the
Lymphogranuloma venereum associated L-Serovars by TaqMan PCR. BMC Infect Dis.
2008;8:56. Published 2008 Apr 30. doi:10.1186/1471-2334-8-56

29. Yahara K, Ma KC, Mortimer TD, et al.  Emergence and evolution of antimicrobial
resistance genes and mutations in Neisseria gonorrhoeae. bioRxiv 2020.10.26.354993;
doi: https://doi.org/10.1101/2020.10.26.354993

30. Bachmann NL, Rockett RJ, Timms VJ, Sintchenko V. Advances in Clinical Sample
Preparation for Identification and Characterization of Bacterial Pathogens Using
Metagenomics.　　Front　　Public　　Health.　　2018;6:363.　　Published　　2018　　Dec　　12.
doi:10.3389/fpubh.2018.00363

31. Seth-Smith HM, Harris SR, Skilton RJ, et al. Whole-genome sequences of Chlamydia
trachomatis　　directly　　from　　clinical　　samples　　without　　culture.　　Genome　　Res.
2013;23(5):855-866. doi:10.1101/gr.150037.112

32. Christiansen MT, Brown AC, Kundu S, et al. Whole-genome enrichment and sequencing of Chlamydia trachomatis directly from clinical samples. BMC Infect Dis. 2014;14:591. Published 2014 Nov 12. doi:10.1186/s12879-014-0591-3

33. Street TL, Barker L, Sanderson ND, et al. Optimizing DNA Extraction Methods for Nanopore Sequencing of Neisseria gonorrhoeae Directly from Urine Samples. J Clin Microbiol. 2020;58(3):e01822-19. Published 2020 Feb 24. doi:10.1128/JCM.01822-19

34. Jorge-Berrocal A, Mayta-Barrios M, Fiestas-Solórzano V. Resistencia antimicrobiana de Neisseria gonorrhoeae en Perú [Antimicrobial resistance of Neisseria gonorrhoeae in Peru]. Rev Peru Med Exp Salud Publica. 2018;35(1):155-156. doi:10.17843/rpmesp.2018.351.3552

35. Getman D, Jiang A, O'Donnell M, Cohen S. Mycoplasma genitalium Prevalence, Coinfection, and Macrolide Antibiotic Resistance Frequency in a Multicenter Clinical Study Cohort in the United States. J Clin Microbiol. 2016;54(9):2278-2283. doi:10.1128/JCM.01053-16

36. Lau A, Bradshaw CS, Lewis D, et al. The Efficacy of Azithromycin for the Treatment of Genital Mycoplasma genitalium: A Systematic Review and Meta-analysis. Clin Infect Dis. 2015;61(9):1389-1399. doi:10.1093/cid/civ644

37. Harrison MA, Harding-Esch EM, Marks M, et al. Impact of mass drug administration of azithromycin for trachoma elimination on prevalence and azithromycin resistance of genital Mycoplasma genitalium infection. Sex Transm Infect. 2019 Nov;95(7):522-528. doi: 10.1136/sextrans-2018-053938. Epub 2019 Apr 13. PMID: 30981999; PMCID: PMC6860407.

38. Bradic M, Warring SD, Tooley GE, et al. Genetic Indicators of Drug Resistance in the Highly Repetitive Genome of Trichomonas vaginalis. Genome Biol Evol. 2017;9(6):1658-1672. doi:10.1093/gbe/evx110

39. Harding-Esch EM, Huntington SE , Harvey MJ, et al. Antimicrobial resistance point-of-care testing for gonorrhoea treatment regimens: cost-effectiveness and impact on ceftriaxone use of five hypothetical strategies compared with standard care in England sexual health clinics. Euro Surveill. 2020;25(43):pii=1900402. https://doi.org/10.2807/1560-7917.ES.2020.25.43.1900402