EPIDURAL ELECTRICAL STIMULATION OF THE CERVICAL DORSAL ROOTS RESTORES VOLUNTARY UPPER LIMB CONTROL IN PARALYZED MONKEYS

B. Barra[1,2,*], S. Conti[1,*], M.G. Perich[3], K. Zhuang[1], G. Schiavone[4], F. Fallegger[4], K. Galan[5,7], N. D. James[5],Q. Barraud[5,7], M. Delacombaz[1,7], M. Kaeser[1], E. M. Rouiller[1], T. Milekovic[3,7], S. Lacour[4], J. Bloch[6,7], G. Courtine[5,6,7] and M. Capogrosso[1,2,8]

[1] Platform of Translational Neuroscience, Department of Neuroscience and Movement Sciences, Faculty of Sciences and Medicine, University of Fribourg, Fribourg, Switzerland
[2] Rehab and Neural Engineering Labs, University of Pittsburgh, Pittsburgh, USA.
[3] Department of Fundamental Neuroscience, Faculty of Medicine, University of Geneva, Geneva, Switzerland
[4] Bertarelli Foundation Chair in Neuroprosthetic Technology, Laboratory for Soft Bioelectronic Interfaces, Institute of Microengineering, Institute of Bioengineering, Centre for Neuroprosthetics, École Polytechnique Fédérale de Lausanne, Geneva, Switzerland.
[5] Center for Neuroprosthetics and Brain Mind Institute, School of Life Sciences, Ecole Polytechnique Fédérale de Lausanne, Lausanne, Switzerland.
[6] Department of Neurosurgery, CHUV, Lausanne, Switzerland;
[7] Defitech Center for Interventional Neurotherapies (NeuroRestore), University Hospital Lausanne (CHUV), University of Lausanne (UNIL) and Ecole Polytechnique Fédérale de Lausanne, Lausanne, Switzerland.
[8] Department of Neurological Surgery, University of Pittsburgh, Pittsburgh, USA

*these authors contributed equally to this work

Correspondence to: Marco Capogrosso - mcapo@pitt.edu

SUMMARY
Recovering arm control is a top priority for people with paralysis. Unfortunately, the complexity of the neural mechanisms underlying arm control practically limited the effectiveness of neurotechnology approaches. Here, we exploited the neural function of surviving spinal circuits to restore voluntary arm and hand control in three monkeys with spinal cord injury using spinal cord stimulation. Our neural interface leverages the functional organization of the dorsal roots to convey artificial excitation via electrical stimulation to relevant spinal segments at appropriate movement phases. Stimulation bursts targeting specific spinal segments produced sustained arm movements enabling monkeys with arm paralysis to perform an unconstrained reach-and-grasp task. Stimulation specifically improved strength, task performances and movement quality. Electrophysiology suggested that residual descending inputs were necessary to produce coordinated movements. The efficacy and reliability of our approach hold realistic promises of clinical translation.

INTRODUCTION

More than 5 million people in the US currently live with some form of motor paralysis[1]. Stroke and spinal cord injury (SCI) are the main causes with hundreds of thousands of new cases per year[2]. Impairments of the hand and arm are particularly problematic, representing a major unmet need for both SCI and stroke patient populations[3,4]. Indeed, even mild deficits in hand function lead to significant degradation of quality of life. Unfortunately, recovery of hand and arm motor function is still an unsolved clinical challenge.

Generated in the cerebral cortex, upper limb motor commands are relayed to subcortical and spinal circuits that activate motoneurons and regulate sensory inputs to produce skilled motor actions[5–8]. Spinal cord injury (SCI), or stroke, damage these communication pathways generating impairments in sensory regulation and motor functions that lead to motor paralysis.
Historically, neurotechnologies were conceived around the idea of restoring movements in paralyzed subjects via a technological bypass. Such solution would use signals from cortical areas as inputs and artificially compensate for lack of motoneuron activation by producing desired muscle activity below the lesion[9]. For example, functional electrical stimulation (FES) was used to activate arm muscles in response to intracortical neural activity from the motor cortex[10,11]. This pioneering concept allowed paralyzed monkeys and humans to perform voluntary grasping tasks[10–13]. However, translation of these concepts into daily clinical practice is hindered by two distinct limitations. First, the artificial motoneuron recruitment order generated by FES induces muscle fatigue[14] which is particularly problematic for arm movements. Indeed, fatigue prevents the generation of sustained forces and consequently FES fails to enable sustained three-dimensional arm movements that are required for daily activities. Second, since FES bypasses surviving circuits in the spinal cord, complex stimulation protocols[15] and sophisticated decoding algorithms[10,13] are required to orchestrate the activation of multiple muscles and produce functional movements. As a result, these systems require an articulated combination of hardware and software. Unfortunately, this complexity does not cope well with dynamic clinical environments that need robust and practical solutions for a rapid set up and large-scale use.

In contrast, epidural electrical stimulation (EES) of the lumbar spinal cord exploits surviving spinal circuits and supra-spinal connections after injury to produce movements[16]. Similar to intraspinal stimulation[17–19], EES engages motoneurons via direct recruitment of large sensory afferents[20,21] leading to widespread excitatory post-synaptic potentials in the spinal cord. More importantly, since motoneurons are recruited via natural synaptic inputs, EES generates a natural recruitment order[22,23] that is resistant to artificial fatigue. This enables the production of forces that can sustain the whole-body weight[24]. Moreover, engagement of motoneurons from pre-synaptic pathways allows residual descending inputs and spinal circuits to control motoneurons excitability and produce voluntary movement after complete motor paralysis[25,26].

Building on animal models[27–29], recent clinical studies have shown that continuous stimulation delivered through epidural implants on the dorsal aspect of the lumbosacral spinal cord increased muscle strength, voluntary muscle activation and single joint movements in people with complete leg paralysis[26,30,31]. More strikingly, when coupled with targeted physical rehabilitation protocols, continuous EES restored weight bearing locomotion in subjects with severe SCI[32,33]. These outstanding clinical results prompted experimental studies aiming at verifying whether EES could be used to promote also upper limb movements after SCI[34]. Unfortunately, while clinical studies showed some success in improving hand grip force with both epidural and non-invasive approaches[35,36], continuous EES did not produce results of similar outstanding efficacy as those

observed for the lower limbs[32,33]. In fact, clinical outcomes were similar to those obtained with
surface FES[37].
Reasons for this discrepancy may stem from the complexity of upper limb motor control and
biomechanics compared to locomotion. Indeed, in contrast to pattern-driven[38,39] and repetitive
locomotor movements, upper limb movements are composed by a non-repetitive and task-
dependent combination of movement modules which are highly dependent from sophisticated
cortico-spinal control[7,40–44] and accurate sensory feedback[42,45–47]. Because of this intrinsic
complexity, non-specific neuromodulation could limit the efficacy of EES by exciting all spinal
segments simultaneously, irrespectively of movement phase. More importantly, unspecific and
continuous stimulation of the sensory afferents through EES disrupts natural sensory inputs[23]
thus hindering spinal regulation of movements which is critical in dexterous upper limb control[45–
47].
We and others have shown that it is possible to direct electrical stimulation of the spinal cord to
target restricted segments during appropriate times[17,48,49]. These spatio-temporal stimulation
protocols enabled voluntary locomotion in monkeys with SCI as early as day 6 post injury without
any physical training[50] and within 2 weeks post implantation in humans with complete leg
paralysis[51]. This approach exploits the somato-topography of the spinal sensory system to
selectively engage restricted spinal regions[21,49]. Unfortunately, non-invasive technologies and
clinically approved electrodes are unfit for this scope[52,53] because of their limits in selectivity.
Therefore, we hypothesized that a neural interface, specifically designed to target the cervical
dorsal roots, could enable the administration of spatio-temporal stimulation patterns to the cervical
spinal cord. We tested this hypothesis in three monkeys with a unilateral cervical SCI. We
designed a personalized epidural interface to target primary afferents within the cervical dorsal
roots. We hypothesized that the electrical stimulation of the roots with bursts linked to movement
attempts would enable voluntary motor control and improve functional deficits of the arm and hand
that emerge after SCI. Specifically we tested for improvements in muscle strength, dexterity and
ability to execute three-dimensional functional tasks in full independence. Finally, we verified that
the mechanisms enabling the voluntary recruitment of motoneurons in the cervical spinal cord
were similar to those occurring during EES of the lumbosacral circuits.
115
116 **<u>Results</u>**
117
Natural arm movements
Clinically effective systems should enable truly functional arm movements rather than simplified
tasks such as single-joint movements. A functional arm movement entails a coordinated activation
of arm muscles to achieve a desired movement while supporting the arm weight at all times. Most
of daily activities require arm extension (reach) and flexion (pull), combined with a hand-grasp
without a constrained timing or structure. Consequently, we developed a robotic platform allowing
the quantification of reach, grasp and pull movements[54] that would feel natural and unconstrained
to monkeys both in trajectory and timings (**Figure 1A**). We trained three adult Macaca fascicularis
monkeys to reach for, grasp, and pull an instrumented object placed on the end effector of our
robotic arm (**Figure 1B**). Movement trajectories were not constrained neither kinematically nor in
time. Monkeys waited for the go signal, reached for the object and pulled to receive a food or juice
reward when the object crossed a pre-defined displacement threshold[54]. Monkeys intuitively and
rapidly[29,30] learned this task by developing their own individual kinematic strategies (**Extended**
Data Figure 1) and personal movement speeds. We then designed a battery of electrophysiology
and kinematic measurements to evaluate functional outcomes on task performances, muscle

Figure 1. Experimental framework. **(A)** Schematic of the behavioral experimental platform. While the animals were performing a robotic reach, grasp and pull task, we measured 3D forces applied to the robot joints, full-limb kinematics, electromyographic (EMG) activity from eight muscles of the arm and hand, and intra-cortical signals from sensorimotor areas. **(B)** Schematic illustration of the task. Monkeys were trained to reach for, grasp, and pull a target object placed at the end effector of a robotic arm. We considered a movement complete when a target spatial threshold was crossed during pull. **(C)** Motoneurons pool distribution of arm and hand muscles in the cervical spinal cord in relation to vertebrae and spinal segments (adapted from Jenny and Inukai, 1983). Deltoid (DEL), Biceps Brachii (BIC), Flexor Carpi Radialis (FCR), Triceps Brachii (TRI), Extensor Digitorium Communis (EDC), Extensor Carpi Radialis (ECR), Flexor Digitorium Profundis (FDP), Abductor Pollicis Brevis (ABP). **(D)** Schematic representation of spinal implant positioning and X-ray scans of the epidural implant in the three monkeys (Mk-Sa, Mk-Br and Mk-Yg). **(E)** Anatomical reconstruction of the cervical spinal cord lesion (black area) for the 3 monkeys, shown on a transversal section (the percentage indicates the portion of the total spinal cord area that was injured on this transversal plane). On the right, representative image of longitudinal section of the spinal cord of Mk-Br around the lesion site stained with NeuN (neuronal cell bodies) and Iba1 (microglia). Copyright Jemère Ruby.

activation, muscle strength and movement dexterity. Specifically, we quantified full-limb 3D
kinematics (Vicon Motion Systems, Oxford, UK), pulling forces, and electromyographic (EMG)
signals from intramuscular leads in eight arm muscles (**Figure 1A, Extended Data Figure 1**).
Before SCI, we observed clear bursts of EMG activity from all hand and arm muscles during the
three movement phases: reach, grasp, and pull in all monkeys. Finally, to document the
involvement of cortical neurons during movement enabled by EES and to extract signals that
could also be used to link stimulation bursts to movement phase onset, we implanted multi-
microelectrode arrays (Blackrock Microsystems, Salt Lake City, USA) in the arm/hand region of
the right sensorimotor (M1, S1) and ventral premotor (PMv) cortex. We validated these recordings
by verifying that neural activity was consistently modulated with kinematics pre-injury and with the
three movement phases as largely expected[54] (**Figure 1, Extended Data Figure 1**). In summary,

we analyzed natural arm movements in monkeys and concluded that in order for stimulation
protocols to be effective, it was important to support reach, grasp and pull independently with
specific parameters for each animal.
147
148
Personalized spinal interface
To design an optimal interface, we studied the anatomy of the monkey cervical spinal cord. We
extrapolated available anatomical information from literature and found that, similar to humans,
motoneurons innervating arm muscles in the monkeys are segmentally organized[55] (**Figure 1C**).
We previously showed that stimulation of a single cervical dorsal root will recruit motoneurons
that receive direct afferent inputs from that root[53]. Exploiting this property allows to obtain a
segmental recruitment order of motoneurons that can be targeted to promote specific movement
phases[49,51,56]. Therefore, we designed a spinal interface that could target each root independently.
We achieved this by placing contacts on the lateral aspect of the cord to target the entry zone of
each individual root[53]. Since each monkey displayed a unique anatomy, we tailored the design of
our interface to each specific subject. For this, we measured white matter diameter and vertebral
canal features from computed tomography (CT) and magnetic resonance imaging (MRI). We then
spaced the electrodes rostro-caudally and medio-laterally to match the transversal and
longitudinal dimensions of the cord of each animal (**Extended Data Figure 2A, 2B**). This allowed
us to simplify the neural interface architecture by minimizing the number of contacts while
maintaining high muscle recruitment specificity[57]. We then designed a surgical strategy to position
the epidural interface between the C6 and T1 dorsal roots (**Figure 1D**). We performed
laminectomies between the T1 and T2 vertebrae and the C5 and C6 vertebrae, then pulled the
neural interface through the intermediate epidural space with the help of a custom soft inserter[57].

Figure 2. Muscle recruitment of spinal stimulation. (A) Examples of muscle recruitment obtained by stimulating (1 Hz) at C5, C6/C7, and T1 spinal segments (Mk-Yg). **(B)** Average muscle activations elicited from C6/C7 and T1 contacts in n=3 monkeys (grey bullets: for each animal, average recruitment across all stimulation currents. Big bullets: mean of average recruitments across animals).

Figure 3. EES produces functional joint movements in anesthetized animals. (A) Stick diagram schematic of elbow extension and wrist flexion movements elicited by pulse-trains of stimulation in anesthetized conditions in Mk-Yg. **(B)** Modulation of maximal joint angles achieved by pulse-trains of stimulation at different frequencies, in anesthetized conditions in Mk-Yg. Stimulation was delivered at C7 (blue) and T1 (yellow)**.** Statistics performed with Wilcoxon Ranksum test and Bonferroni correction. Asterisks: *p<0.05, **p<0.01, ***p<0.001 **(C)** Triceps (blue), biceps (yellow), and flexor digitorium superficialis (yellow with black border) activity elicited by pulse-trains of stimulation at different frequencies, in anesthetized conditions in Mk-Yg. Bullets represent mean values and bars are standard deviation. Statistics performed with Wilcoxon Ranksum test and Bonferroni correction. Asterisks: *p<0.05, **p<0.01, ***p<0.001 **(D)** Stick diagram schematic of elbow extension and wrist flexion movements elicited by pulse-trains of stimulation in anesthetized conditions in Mk-Br. Statistics performed with Wilcoxon Ranksum test and Bonferroni correction. Asterisks: *p<0.05, **p<0.01, ***p<0.001 **(E)** Modulation of maximal joint angles achieved by pulse-trains of stimulation at different frequencies, in anesthetized conditions in Mk-Br. Stimulation was delivered at C7 (blue) and T1 (yellow) **(F)** Triceps (blue), biceps (yellow), and flexor digitorium superficialis (yellow with black border) activity elicited by pulse-trains of stimulation at different frequencies, in anesthetized conditions in Mk-Br. Bullets represent mean values and bars are standard deviation. Statistics performed with Wilcoxon Ranksum test and Bonferroni correction. Asterisks: *p<0.05, **p<0.01, ***p<0.001

We verified that the position of the array remained stable for the entire duration of the study (up
to 3 weeks) through repeated X-ray imaging (**Figure 1D**, **Extended Data Figure 2C**). During the
same surgery, we performed a unilateral spinal cord injury at the C5/C6 segments (**Figure 1E**)

aiming at transecting the cortico-spinal tract that is located on the lateral aspect of the white matter
in monkeys. This type of lesion is amply described in literature and induces unilateral arm and
hand paralysis[58,59] while preserving important bodily functions such as bladder control.
Postmortem immunohistochemistry analysis of the spinal cords showed that the spinal interface
did not damage the cervical cord in any of the three monkeys but did reveal that Mk-Br received
an unplanned compression injury at the insertion site (T3 spinal segment). Given the caudal
position of this contusion it is likely for it to have occurred during implantation (**Extended Data**
Figure 2D). Since the T3 segment is below the innervation of the arm motoneurons, this lesion
did not affect the phenotype of arm and hand motor deficits which did not differ from the other
monkeys (see Methods).
In summary, we designed a spinal interface to selectively recruit the cervical dorsal roots. We
tailored the interface to the specific anatomy of each monkey and designed a surgical strategy to
perform a consistent and stable implantation.
184
Cervical EES produces functional joint movements and grasp in anaesthetized moneys
We next assessed the selectivity of the epidural interface. In propofol anaesthetized monkeys, we
delivered asymmetric, charge-balanced biphasic pulses of EES at low repetition rate (1Hz) at
various current amplitudes from each contact. Minimum and maximum amplitude values were
selected as the first subthreshold and first saturation current value respectively. As predicted[53],
different stimulation contacts generated muscle recruitment patterns that mirrored the segmental
organization of cervical motoneurons (**Figure 2A**, **Extended Data Figure 3**). Specifically,
contacts located at C8/T1 level (caudal) elicited spinal reflexes mostly in the hand and forearm
muscles, contacts located at C7 level elicited triceps and contacts located at C5/C6 recruited
biceps and deltoids (rostral). Those results were consistent in all animals (**Figure 2B, Extended**
Data Figure 3). To ensure that this segmental selectivity translated into separate functional arm
and hand movements, we delivered supra-threshold stimulation at various frequencies (20-120
Hz) from each contact in two animals (Mk-Br and Mk-Yg). Indeed, since recruitment of
motoneuron is pre-synaptic, EES may not be able to produce sustained muscle activation
because of frequency dependent suppression[60]. This effect is an observed substantial
suppression of muscle evoked potentials during repetitive stimulation of the afferents. Instead, we
observed large and sustained arm movements during EES bursts. Muscle selectivity was
preserved during long stimulation trains (**Figure 3C, F**) and different contacts elicited distinct
functional joint movements (**Figure 3A, B, D, E, Video 1**) such as shoulder abduction, elbow
extension and whole hand grasp. When looking at the energy of the EMGs, we found a monotonic
relationship between muscle activation and stimulation frequency in most of the upper arm
muscles (**Figure 3C, F**). However, not all muscles showed such clear frequency dependent
responses (**Extended Data Figure 4A**). Moreover, peak-to-peak responses (**Extended Data**
Figure 4B) were generally decreased during a burst at high frequency but were not abolished
and tended to vary during the burst and while the movement was produced. We used these
observations to optimize stimulation parameters to be used in a behavioral reach and grasp task
(see Methods and **Extended Data Figure 5**). In summary, we found that single contacts of our
spinal interface elicited segmental recruitment of arm flexors, extensors and hand flexors. Bursts
of stimulation from these contacts produced sustained joint movements that were graded by
stimulation frequency (**Extended Data Figure 6**).
215

Figure 4. EES improves task performance. (A) Snapshots of Mk-Yg performing the task before SCI, after SCI without EES, and after SCI with EES. A full successful trial is composed of a reach, a grasp, and a pull. After SCI, Mk-Yg could only perform reaching movements without EES, while when EES was delivered the full task could be performed. **(B)** Task performance rate over all available sessions, computed as the percentage of successful movements across all attempted movements. Performance rate are shown for reach (blue), grasp (yellow to blue gradient) and pull (yellow movements). Data are shown as mean (bullets) and standard deviation (bars). Statistics was performed with Bootstrap. Asterisks: *p<0.05, **p<0.01, ***p<0.001.**(C)** Bar plots report the rate of successful movements after SCI without and with stimulation, for all the days in which animals performed the task. Rates were computes as number of successful trials per units of time. Data are presented as mean ± STD and normalized on the mean value in stimulation condition. Statistics was performed with Bootstrap, Asterisks: *p<0.05, **p<0.01, ***p<0.001. (P-values. Mk-Br: reach, n.s; grasp, p =0.01; pull, p =0.08. Mk-Yg: reach, grasp and pull n.s) **(D)** Bar plots report the rate of successful movements after SCI without and with stimulation, for the best session of Mk-Br and Mk-Yg. Data are presented as mean ± STD and normalized on the mean value in stimulation condition. Statistics was performed with Bootstrap, Asterisks: *p<0.05, **p<0.01, ***p<0.001. (P-values. Mk-Br: reach, p = 0.04; grasp, p = 0.002; pull, p =0.08. Mk-Yg: reach, grasp and pull n.s)

217
Cervical EES substantially improves arm and hand motor function after spinal cord injury
We next tested whether our stimulation protocol could improve functional outcomes of upper limb movements after SCI. Specifically, we tested the efficacy of EES to improve muscle activation, pulling forces, functional task performance, and kinematic quality of three-dimensional movements after SCI when stimulation was on against stimulation off as a control. In all monkeys, the lesion led to substantial motor deficits of the left arm and hand.

While each monkey retained the ability to activate proximal shoulder and biceps muscles, elbow extension and hand functions were severely compromised. Severity of the impairment and extent of spontaneous recovery (**Extended Data Figure 7**) varied across monkeys because of the variability in lesion size (**Figure 1E**). Generally, animals showed severe paralysis immediately after lesion, and then gradually regained some movement capabilities (**Extended Data Figure 7**). Due to the initial impairment, immediately after the lesion, monkeys were not able to perform the behavioral task. Consequently, during the first week, we simplified the task by presenting an object close to the monkeys and triggering stimulation bursts manually to encourage the animal to perform the task. After the first week, all monkeys spontaneously attempted to perform the task, making it possible to link the delivery of movement-specific stimulation bursts to real-time detection of movement onset using intra-cortical signals. Whenever the monkeys strived for a reach, grasp or pull movement, we delivered bursts of stimulation promoting reach or grasp/pull respectively (movement specific EES). Outcomes were computed for each animal independently and compared between EES on and EES off. In terms of functional task performances, without stimulation, the monkeys were rarely capable of completing any part of the task (defined as reach, grasp and pull). Instead, with the support of EES, both the percentage of success and the rate of success improved with rates that depended on the level of function of the animals over time. For example, reach was recovered immediately with larger improvements at the beginning, when deficits were larger in all three animals. Instead, improvements in grasps emerged only later when the animals spontaneously recovered some movement capacity (**Figure 4, Video 2,3,4**). More specifically Mk-Br improved grasp and pull only after 2 weeks with stimulation while Mk-Yg was never able to grasp and pull except during stimulation which we could test only until day 7 when the grasp contact E6 failed (see Methods).Instead, when we used our interface to deliver continuous EES that was not related to movement onsets, only non-significant and modest improvements were observed in Mk-Br while Mk-Yg did not show ability to grasp and pull during continuous EES (**Extended Data Figure 8A**). Moreover, we analyzed trials in which stimulation bursts were not triggered at movement onset, for example when pull stimulation was erroneously triggered during reach. In these trials the reach movement was abruptly interrupted, and the animal did not complete the task (**Extended Data Figure 8B, Video 5**).

During phase dependent stimulation, EES enhanced muscles activity and forces (**Figure 5A,B**) compared to no stimulation. In terms of movement quality, EES bursts triggered at movement onset significantly improved the overall quality of arm movements (**Figure 5B**). Indeed, principal component analysis (PCA) of three-dimensional kinematic parameters (i.e., timing, force, arm trajectories, joint angles) revealed that during EES, movement kinematics were significantly closer to pre-lesion kinematics than the few successful movements performed without stimulation (distance from pre-lesion performances in the multi-parametric kinematic space, **Figure 5B**). Notably, animals sustained the weight of the arm and lifted their elbow more, performed wider movements, and generated stronger forces (**Figure 5B**), getting closer to normal kinematic trajectory patterns without any long-term training.

In summary, we showed that EES bursts triggered at movement phase onsets, improved muscle strength, task performance and quality of arm movements. This allowed monkeys to perform reach, grasp and pull movements that were otherwise not able to perform without EES.

Figure 5. EES improves muscle strength and movement quality. (A) Bar plots of signal energy of biceps, triceps and FDS EMG profiles during movement with no stimulation (black) and stimulation (yellow). Data are shown for different sessions (one for each week) in Mk-Br and Mk-Yg. Mk-Sa performed only one session. All individual data points are represented by bullets. Statistical analysis with Wilcoxon Ranksum test. **(B)** PC analysis of kinematic features for Mk-Yg (top) and Mk-Br (bottom). From left to right: (1) first and second PC space. Each bullet represents one trial. Trials performed after injury (black) are consistently separated from the trials performed in intact conditions, highlighting a change in the quality of resulting kinematics. Trials performed with the support of stimulation (blue for reach and yellow for pull) are located closer to the intact trials in the PC space, denoting an improvement in kinematic features. (2) euclidean distance in the feature space of trials without stimulation (black) and with stimulation (blue for Mk-Yg, yellow for Mk-Br) from the centroid of the trials in intact condition; (3) example violin plots of movement quality features in the three conditions: intact, after SCI, and after SCI with stimulation. Statistics with Wilcoxon Ranksum test. Asterisks: *p<0.05, **p<0.01, ***p<0.001.

267
268
269
270

Figure 6. EES must be synchronized with motor intention. (A) Schematic of the interactions between EES and residual neural structures during anesthetized stimulation. During anesthesia, cortical control has no interaction, therefore EES interacts solely with sensory feedback spinal circuits. **(B)** Quantification of EMG activity during EES in two conditions: unconstrained arm (no load, black); arm constrained by load applied at the hand (load, gray). White and grey bullets: individual data points for no load and load conditions. Black and yellow bullets: mean values for no load and load conditions. Black and yellow lines: interpolation of mean values for no load and load conditions. On the bottom, example of EMG traces obtained during stimulation in the no-load (black) and load (gray) conditions. Stimulation artifacts have been removed. Data from Mk-Br **(C)** Schematic of interactions between EES and residual neural structures during the performance of the behavioral task. EES interacts with descending cortical drive sent through residual pathways after SCI, as well as with sensory spinal circuits. **(D)** Schematic illustrating the kinematic outcome of the interaction between EES and residual cortical inputs. The same EES pulse train (top) applied to Mk-Br can result in different motor outputs: no movement output when the cortex is silent (yellow, top), movement is produced when the cortex is active (blue, bottom). **(E)** Distribution of average firing rates across all M1 channels during stimulation trains that evoked no movement (yellow) and movement (blue). **(F)** Left: State space view of M1 activity for all time points during rest (gray), successful stimulation (blue) and unsuccessful stimulation (yellow). The brain states during unsuccessful stimulation (yellow) overlapped with the rest states, while the successful stimulation (blue) did not. Right: we computed a relative Mahalanobis distance between the two stimulation conditions and the cluster of neural states at rest. For both monkeys, neural states during stimulation periods with no movement were close to rest.

Sensory inputs can decrease EES-induced motor output

We then investigated the role of spinal circuits and sensory inputs in the production of the movements that we observed. Indeed, since activation of motoneurons was pre-synaptic, spinal reflexes and sensory inputs can influence EES evoked spinal reflexes in the legs[22,61]. In order to exclude influences of residual supraspinal voluntary inputs, we conducted experiments under propofol anesthesia (**Figure 6A**) with Mk-Br. We then delivered bursts of EES from the contact eliciting elbow flexion at varying stimulation frequencies in two distinct conditions (**Figure 6B**): in isometric and unconstrained conditions. In the isometric condition, we constrained the wrist, elbow and shoulder of the animal and measured force production at the wrist joint. Under unconstrained conditions we left the arm free to move under the effect of stimulation. This setup only differs from the sensory feedback generated at the load when pull forces are produced by EES. We found that EES induced EMG activity during unconstrained movement that was significantly different from the EMG activity induced during isometric movements (**Figure 6B**). In particular, overall EMGs and peak-to-peak amplitudes of elicited spinal reflexes were significantly lower when the arm was attached to a load (isometric) compared to when it was free to move. Albeit present at all frequencies, this difference was particularly important within the 40 to 60Hz range, thus overlapping with the functional frequency ranged that we selected for our study.

These results show that force loads at the hand decreased EMG activity induced by EES as compared to no load applied at the hand. Under anesthesia, only changes on spinal circuit excitability induced by sensory inputs can explain the observed changes on EES evoked muscle activity.

Some residual cortical input is necessary for cervical EES to be effective

The influence of spinal sensory inputs showed that EES output may be decreased because of spinal sensory inputs when loads are applied at the hand. This would decrease the efficacy of EES which is supposed to enhance force production. Therefore, to explain the results we obtained in behaving monkeys (**Figure 6**) we investigated the contribution of residual cortical inputs in the production of forces and movements during EES. Specifically, since cortical inputs actively modulate spinal circuits, they should be able to both enhance and suppress EES output by modulating spinal circuit excitability[30]. Since we showed that monkeys could use EES to amplify their movement and forces (**Figure 6D**) we focused on demonstrating that cortical inputs could also suppress unwanted EES-generated movements. We hypothesized that if monkeys did not want to move, EES would not produce the large joint movements that we observed when the monkeys were anesthetized. Therefore, we identified trials in which our decoder detected a false-positive reach movement (**Figure 6C**). In this situation our system would deliver a burst of stimulation even if the animal was not attempting to execute the task. We then compared intracortical activity from the primary motor cortex (M1) of Mk-Br and Mk-Yg during these false-positive trials to the signals recorded during correctly detected trials. We identified trials where EES was present and the monkey moved, and trials when EES was present but the monkey did not move (**Figure 6D**). We verified that the same neural units were present in both conditions and found that the overall firing rates of all units in motor cortex was significantly higher when EES produced movement (**Figure 6E**) than when it did not. This suggested that movement happened only if the motor cortex was active, despite EES was delivered at amplitudes that generated large joint movements when the same monkey was anesthetized. To further validate this hypothesis we applied dimensionality reduction using Principal Component Analysis to the firing rates in each electrode and reduced the M1 population activity to low-dimensional states[62]. In this low-dimensional space each point represents the global neural state of the motor cortex at a given time point (**Figure 6F**). We compared the neural states present when EES was associated movements and those when EES was not associated movement with the neural states associated

to rest, e.g. when the monkeys were resting before the go signals between trial repetitions. When
looking at the spatial distribution of neural states, trials in which EES was not associated to
movement seemed to overlap with states of rest. We then computed the distance between each
neural state to the subspace representing neural states at rest and found that the neural states
associated to movements during EES were significantly further away from neural states at rest
than neural states associated to EES and no movement. In summary, we found that the motor
cortex activity was similar to the activity at rest whenever we delivered EES but the monkey did
not move (**Figure 6F**). Instead, the monkey moved when the motor cortex was significantly active.
This implies that the residual cortical inputs via direct and indirect pathway can either suppress
or enable movement during EES.
332
333

Discussion
We showed that EES of cervical spinal cord immediately enhanced muscle activation and strength,
task performances and movement quality during a natural-like reach and grasp task in monkeys
with unilateral cervical SCI compared to no stimulation controls in three monkeys. Importantly,
our technique allowed monkeys to support the weight of their arm during reach, grasp and pull
movements. These results are important in light of clinical translation of our technology. Stronger
forces and better arm weight bearing can empower patients with the capacity to perform a larger
spectrum of movements than they would normally be capable of doing without the need of support.
This may provide for more independence in daily living as well as better outcomes of physical
therapy.
344

Exploiting subject-specific anatomy to simplify technology
We obtained our results with relatively simple stimulation protocols that engaged up to three
monopolar contacts (one for reach, one for grasp and one for pull). The combination of simple
bursts through these contacts enabled whole arm multi-joint movements. We believe that the
design of our interface was key to achieve this result. The dorsal roots are a robust anatomical
target that we could easily identify through standard imaging to personalize surgical planning and
interface design. A similar surgical planning approach can be imagined in humans where MRIs
and CT can guide surgical planning[51,63].
Our results were enabled by the relative mapping between each dorsal root and the rostro-caudal
distribution of motoneurons in the cervical spinal cord, which is similar in monkeys and
humans[53,55,64]. The anatomical separation of roots in the cervical enlargement allowed us to recruit
each root independently which generated distinct joint movements to a degree that was not
observed in applications of EES for the lower limbs[49]. Stimulation of the C6 root elicited distinct
arm flexion, C7 stimulation produced arm extension and C8/T1 stimulation produced hand grasp.
However, similarly to other spinal cord stimulation studies we could not identify contacts that
selectively produced finger extension[18,65,66]. This is likely caused by the overlap of extensor motor-
pools in the forearm[55,64] but possibly also because flexors may be biomechanically stronger and
dominate hand kinematics in the case of co-contraction at rest. Despite these limitations in
specificity, we were able to restore a whole three-dimensional arm movement by solely detecting
movement onset signals to trigger pre-determined stimulation bursts through two or three contacts.
Unlike FES, this is possible because EES activates cervical motoneurons via pre-synaptic inputs
thus allowing modulation of elicited muscle responses that can compensate for reduced
specificity[30,49].
368

Supporting arm movement phases independently

Contrary to previous pilot applications of epidural and transcutaneous spinal cord stimulation of
the cervical spinal cord[35,36], we utilized a soft epidural interface that allowed selective and
independent support of each movement phase rather than providing continuous stimulation to the
whole spinal cord. This approach is not possible with transcutaneous technologies[67] or current
design of human leads[53] and would require new interfaces designed for the cervical cord.
Selective spatiotemporal stimulation was shown to be more effective in animal models and
humans than continuous stimulation in the sense that it was able to immediately produce
coordinated locomotion compared to continuous stimulation that instead required long training
periods[28,48,49,49,56]. In the case of the upper limb we believe that this approach was critical. Indeed,
while continuous stimulation did provide some level of facilitation, it failed to entirely promote
grasp and pull in one of the monkeys. Perhaps the intrinsically unstructured nature of arm and
hand control makes a continuous stimulation approach less effective than it is in locomotion that
instead has an intrinsic repetitive structure[38]. For example, stimulation parameters that promote
grasp, may impair reach if they are delivered continuously throughout movement. Indeed, when
a pull stimulation was triggered at mid-reach it generated the interruption of the reach movement.
Perhaps a different interface design or lower stimulation amplitudes could be used to optimize
continuous stimulation protocols, but it would be at the expense of power of elicited movements
potentially preventing the weight bearing component necessary for three-dimensional movements.
In summary, the complex articulation of arm and hand movements may exacerbate the difference
in efficacy between continuous and phase-specific stimulation protocols that was already
observed for EES in locomotion, possibly explaining the difference in effect size that was obtained
so far for application in the upper limb.
392
The role of sensory feedback and residual cortical inputs in cervical EES
We showed that sensory feedback when the hand was constrained to a force load reduced the
EMG power produced by EES compared to free movements. This is likely caused by afferent
inhibitory feedback coming from Ib afferents[68]. Unfortunately, lower muscle power while resisting
a force load would decrease the clinical usability of this technology. We believe that this
phenomenon is particularly relevant for the upper limb. Indeed, also during EES of the
lumbosacral cord, the EES motor output is influenced by sensory inputs[22,61], however sensory
inputs are instrumental for locomotion and heavily contribute to the generation of the repetitive
movement patterns that are required to walk[16,22,23,38,69]. Therefore, in the case of locomotion these
inputs amplified and sustained EES-induced activity[16,22,23,28]. Instead arm and hand movements
are produced by an unstructured sequence of primitive movements[41] and reflexes[45] in parallel
with a sophisticated gating of sensory inputs through mechanisms such as pre-synaptic
inhibition[8,70]. Therefore, residual cortical inputs become instrumental to obtain arm and hand
movement with EES as shown by our analysis of intra-cortical signals during the production of
movement combined with the observation that functional grasp was achieved only when the
animals had recovered some level of function. Indeed, our lesions were non-complete and while
most of the cortico-spinal tract was transected, multiple residual descending pathways were
spared. These indirect inputs could have been used by the animals to mediate the inputs required
to integrate EES and sensory inputs to produce voluntary movements. In summary, we believe
that even during phase-specific EES residual cortical inputs play a critical role in enabling arm
movement for cervical EES.
414
415   **Clinical significance and challenges**
416   The most important challenge for clinical translation of EES to humans concerns the role of
417   residual inputs. Our data show that some level of residual inputs and of function is required to
418   enable movement, first because in awake animals EES did not initiate movements, and second

because it lacks the selectivity to achieve selective finger activity. However, previous studies showed that even completely paralyzed subjects retain residual but functionally silent descending inputs[25,32,51]. Therefore, while overall efficacy may depend on injury severity, even severely injured patients may obtain benefits from cervical EES. After a period of physical training combined with EES[71] these subjects may be able to use EES to achieve simple but functional grasp. Alternatively, more selective technologies targeting hand muscles such as FES could be combined with EES to obtain powerful yet selective movements.

The adaptation of EMG output to stimulation frequency that we observed in consequence of pre-synaptic activation of motoneurons may lead to a reduction in efficacy during long-term clinical use. Additionally, stimulation of afferent fibers may cause uncontrolled reflexes which may affect function. While we did not observe these phenomena in our data, this may be due to the relatively small size of the lesion compared to severe contusion in humans. However, data in humans with SCI suggest that stimulation protocols can be adapted to be functional even in subjects with chronic severe thoracic lesions[32,51], therefore we expect that this will be the case also for cervical lesions. At any rate both risks can be reduced by accurate stimulation tuning and real-time adaptation of stimulation patterns[22,24,72].

Concerning complexity of our system, in our study we detected movement onsets from intracortical activity which may be seen as a limitation for a realistic implementation of our protocol in clinical settings. However, given the simplicity of our protocol which is essentially constituted by alternation of pre-defined bursts, brain recordings may not be required in clinics. Indeed, most patients suffer from a severe but incomplete paralysis[51,73], which spares some residual muscle activity in few muscles. While this residual activity is not sufficient to produce functional movements, it can be reliably detected and used to trigger stimulation bursts with standard clinical technologies[49,51]. In summary, we believe that by exploiting the functionality of residual spinal circuits and supra-spinal inputs, cervical EES constitutes a simple yet robust approach to the restoration of arm motor control with significant translational potential.

Acknowledgements

The authors would like to thank Jacques Maillard and Laurent Bossy for the care provided to the animals, Dr Eric Schmidlin and Dr Simon Borgognon for their help with anaesthesia and surgery preparations, Dr Marion Badi for her help and advice during experiment preparations and experimental procedures, Dr. Andrina Zbinden for her contribution to the health survey of the monkeys, André Gaillard and Andrea Francovich for their help with the implementation of the hardware and the students of the University of Fribourg Amélie Jeanneret, Alen Jelusic, Laora Marie Jacquemet and Samia Borra for their help in processing data.

Funding

The authors would like to acknowledge the financial support from the Wyss Center grant (WCP 008) to MC, GC and TM, an industrial grant from GTX medicals to GC and MC; the Bertarelli Foundation (Catalyst Fund Grant to MC and TM and funds to SL) a Swiss National Science Foundation Ambizione Fellowship (No. 167912 to MC), a Swiss National Science Foundation Doc-Mobilit Grant to BB, The European Union's Horizon 2020 research and innovation program under the Marie Skłodowska-Curie grant agreement no. 665667 (GS) the Swiss National foundation grant BSCGI0_157800 (SL), a Whitaker International Scholars Program fellowship to MGP, and an internal pilot grant of the University of Fribourg to MC.

Author Contributions

MC, BB and SC conceived the study; BB, MGP, and TM designed and implemented the hardware
and software tools; SC designed the behavioral task and training strategy; GS and SL designed
and manufactured the implantable interface; BB, SC, MGP and MC conducted the experiments;
BB, SC, MGP and KZ performed the data analysis; SC, MD and MK trained the animals; SC, KG,
NJ and QB processed the histological data; JB, GC and MC designed surgical implantation
strategies and stimulation strategies. GC and JB, performed surgical implantations and lesions.
EMR and MC implemented and supervised procedures on monkeys; MC, BB, SC and MGP wrote
the manuscript; all authors edited the manuscript; SL, TM, JB, GC and MC secured funding for
the study; MC supervised the study.

Competing Interests
G.C., J.B., S.L., M.C., B.B. and K.Z. hold various patents in relation to the present work. G.C.,
S.L. and J.B. are founders and shareholders of GTX medical, a company developing an EES-
based therapy to restore movement after spinal cord injury.

Data and materials availability
All software and data will be available upon reasonable request to the corresponding author.

486
487

A Monkey's specific movement performance

B Kinematic strategy

C Decoding performances

Extended Data Figure 1. (A) Portfolio of signals recorded during intact movement for each animal. These signals have been recorded during the experimental session prior to the lesion. Motor cortex recordings show firing rate profiles for the 64 microelectrodes. Each row shows the firing rate of a specific electrode. Electrodes are displayed from top to bottom by order of first activation in a reference trial. Arbitrary units in motor cortex recording indicate normalized firing rate for each electrode (see methods). In kinematic and EMG plots, black lines correspond to the mean profile across all trials, shaded area shows the SEM across all trials. Kinematic scales are expressed in mm. For Mk-Yg, arbitrary units on kinematic plots represent displacement units derived by the count of video pixels. EMG scales are expressed in mV. **(B)** Kinematic strategies implemented by each monkey. Stick diagrams representations of the arm kinematic during reach (blue) and pull (yellow). The black line highlights the elbow trajectory. Pie charts represent the percentage of success and failure in task performance before lesion. **(C)** Offline decoding performance for Mk-Br and Mk-Yg before lesion. Histograms show timing accuracy of reach (blue) and pull (yellow) event decoding. The height of bars (y coordinate) illustrates the amount of events decoded with a specific timing accuracy (x coordinate). Pie charts (inset) show the percentage of correctly identified (true positive) reaches (blue) and pulls (yellow), across all decoded events. The black portion of the pie chart highlights the percentage of false positive decoded events.

Extended Data Figure 2. (A) Anatomical landmarks used to tailor the epidural interface to each monkey's anatomy (Length of dorsal aspect of spinal canal L_{cs}, length of C5-T1 spinal segment L_{C5-T1}, electrode width W_{el}, electrode length L_{el}). Three-dimensional reconstructions of vertebras are obtained by CT-reconstruction (Osirix, Pixmeo, Switzerland). **(B)** Personalized design of the epidural implant for each animal. All measures are in millimeters. Yellow traces at the bottom of the electrode identify connectors. **(C)** Position stability of the epidural array over time, illustrated through X-rays imaging taken during 3 consecutive weeks after the implantation, images from Mk-Yg **(D)** Compression injury at the insertion level of the array (T2-T3 segment) in Mk-Br, discovered post-mortem, stained with NeuN (neuronal cell bodies) and Iba1 (microglia).

Recruitment curves

Mk-Br

Mk-Sa

DEL TRI FCR FDS
BIC EDC ECR ABP

Extended Data Figure 3. Muscle recruitment obtained by stimulating (1 Hz) at C5, C6/C7, and T1 spinal segments for Mk-Br and Mk-Sa. Mk-Sa only had three muscles implanted: biceps, triceps, and flexor digitorium superficialis.

A Graded muscle activation during train pulses

B Muscle responses are modulated at higher frequencies

Extended Data Figure 4. (A) Energy of EMG signals of triceps (Mk-Br and Mk-Yg), Flexor Digitorium Superficialis (Mk-Yg) and abductor pollicis (Mk-Br) muscles, following pulse-train stimulation at different frequencies (on the x-axis). Black bullets represent mean values. **(B)** Evolution over time of the peak-to-peak value of stimulation evoked responses during a stimulation burst. Each plot shows the evolution for a specific muscle following pulse-train stimulation at 50 and 100Hz. Triceps is shown for Mk-Br and Mk-Yg, Flexor Digitorium Superficialis for Mk-Yg and abductor pollicis for Mk-Br. Each data point is represented as a bullet and lines represent mean values over time.

Extended Data Figure 5. Design of stimulation protocol. (A) Combined representation of movement smoothness, elbow and finger flexion, and pulling force during anesthetized stimulation. Shades of gray highlight three frequency ranges that produce: (1) smooth trajectory, but little movement and low force (20Hz), (2) smooth trajectory, extended movement and medium force (40 and 50Hz), (3) abrupt and very extended movement and low force (80 and 100Hz). Kinematics and force reported here were measured in different experiments, kinematics was unconstrained, force data were acquired in isometric conditions (see Methods). The range 40-50 Hz was selected as the best optimization of sufficient movement, smoothness and force production. **(B)** Schematic representation of arm and hand kinematics during stimulation delivered from the selection of three contacts to produce elbow extension (blue), hand and wrist flexion (yellow and black), and elbow flexion (yellow). **(C)** Example of comparison between EMG activity during intact movement (left) and movement elicited by chaining stimulation from the three selected contacts (right). **(D)** Scheme illustrating how stimulation is triggered from movement-related intra-cortical signals. On the right, online performances of movement attempt decoder in two animals with SCI. Pie charts represent percentage of predicted (blue) and unpredicted (black) reach events by our decoder.

Extended Data Figure 6. (A) Stick diagram schematic of movements elicited by pulse-trains of stimulation in anesthetized conditions. Mk-Br: on the left, arm kinematic obtained by delivering stimulation at different frequencies from contact number 5, on the bottom-left, arm kinematics obtained by repetitive delivery of a burst at 50 Hz; on the bottom right, superimposition of stick diagrams obtained with stimulation at 20 Hz and at higher frequencies (50 or 100 Hz) from different contacts. For Mk-Yg: arm kinematic obtained by delivering stimulation at different frequencies from contact number 2 and superimposition of stick diagrams obtained with stimulation at 20 Hz and at higher frequencies (50 or 100 Hz) from different contacts. **(B)** On the left, finger flexion produced by stimulation at different frequencies from the grasp contact in Mk-Br. Black bullets represent the mean value across different pulse-trains. On the right, wrist flexion obtained by stimulation at different frequencies from the grasp contact in Mk-Yg.

Extended Data Figure 7. (A) Evolution (in weeks) of rates at which Mk-Br performed reach movements after SCI (black), compared to the performances before injury (gray). **(B)** Evolution (in weeks) of rates at which Mk-Br performed grasp movements after SCI (black), compared to the performances before injury (gray). **(C)** Evolution (in weeks) of rates at which Mk-Br performed pull movements after SCI (black), compared to the performances before injury (gray). **(D)** Evolution (in days) of pull force after SCI without stimulation for Mk-Br. Values are plotted as the mean ± SEM. Statistical analysis was carried out with Wilcoxon Ranksum test. **(E)** Evolution (in weeks) of rates at which Mk-Yg performed reach movements after SCI (black), compared to the performances before injury (gray). **(F)** Evolution (in weeks) of rates at which Mk-Yg performed grasp movements after SCI (black), compared to the performances before injury (gray). **(G)** Evolution (in weeks) of rates at which Mk-Yg performed pull movements after SCI (black), compared to the performances before injury (gray). **(H)** Evolution (in days) of pull force after SCI without stimulation for Mk-Yg. Values are plotted as the mean ± SEM. Statistical analysis was carried out with Wilcoxon Ranksum test.

Extended Data Figure 8. (A) Bar plots report the rate of successful movements after SCI, without stimulation (black), with continuous stimulation (gray) and with phase-dependent stimulation (blue or yellow) for Mk-Br and Mk-Yg. Data are presented as mean ± STD and normalized on the mean value in stimulation condition. Statistics was performed with Bootstrap. **(B)** Left: wrist frontal displacement in trials in which pull stimulation was erroneously triggered during reach (gray and yellow), compared to trials in which pull stimulation was not delivered (black). Yellow bullets highlight the instant at which stimulation was delivered: yellow lines highlight the trajectories during and after stimulation. Middle: barplot of the length of the reach movement when pull stimulation was erroneously delivered and when pull stimulation was not delivered. Data are presented as mean ± STD. Right: stick diagram of arm kinematics during reach without (black) and with (yellow) erroneous pull stimulation.

Materials and Methods

Animals involved in the study

All procedures were carried out in accordance to the Guide for Care and Use of Laboratory Animals[74] and the principle of the 3Rs. Protocols were approved by local veterinary authorities of the Canton of Fribourg (veterinary authorization No 2017_04_FR and 2017_04E_FR), including the ethical assessment by the local (cantonal) Survey Committee on Animal Experimentation and final acceptance by the Federal Veterinary Office (BVET, Bern, Switzerland). Three adult female *Macaca Fascicularis* monkeys were involved in the study (Mk-Sa 9 years old, 4.0 kg, Mk-Br 3 years old, 3.4 kg, Mk-Yg 3 years old, 4.0 kg). Animals were not food deprived, could freely access water at any time and were housed in collective rooms designed in accordance to the Swiss guidelines (detention in groups of 2-5 animals in a room of at least 45 m³). Rooms were enriched with toys, food puzzles, tree branches and devices to climb and hide, as well as access to an outdoor space of 10-12 m³ (see www.unifr.ch/spccr/about/housing). Detailed information on which animals were involved in specific experimental procedures are reported in **Supplementary Table 1.**

Surgical procedures

For each animal, we performed three surgical procedures, (1) intracortical electrodes implantation, (2) intramuscular electrodes implantation, and (3) epidural implant insertion and spinal cord injury. Mk-Sa deviated from this protocol. Mk-Sa was first implanted with the epidural interface before injury, however an infection occurred and resulted in the explanation of the lead to treat the infection. After recovery, the animal was re-implanted, and lesion performed following the same protocol of Mk-Br and Mk-Yg. All the surgical procedures were performed under full anaesthesia induced with midazolam (0.1 mg/kg, i.m.), methadone (0.2 mg/kg, i.m.), and ketamine (10 mg/kg, i.m.) and maintained under continuous intravenous infusion of propofol (5 ml/kg/h) and fentanyl (0.2-1.7 ml/kg/h) using standard aseptic techniques. A certified neurosurgeon (Dr. Jocelyne Bloch, CHUV, Lausanne, Switzerland) performed all the surgical procedures.

During the first surgical procedure, we implanted multi-microelectrode arrays in the primary motor cortex (M1-42 channels), ventral premotor cortex (PMv-32 channels) and primary somatosensory cortex (S1-42 channels) for a total of 128 channels for Mk-Br and Mk-Yg (Blackrock Microsystems, 400 *µ*m pitch and electrodes tip lengths 1.5 mm 1.5 mm and 1mm for M1, PMv and S1 respectively). Instead, Mk-Sa was implanted with 2 microelectrode arrays of 64 channels each and pitch of 1.5 and 1 mm in M1 and PMd respectively. Functional motor areas of the arm were identified through anatomical landmarks and intra-surgical micro-stimulation. In order to access the brain areas of interest we performed a 20 mm diameter craniotomy and we incised the dura. The arrays implantation was achieved using a pneumatic compressor system (Impactor System, Blackrock Microsystems). A pedestal (*Pedestal A*) was then fixated to a compliant titanium mesh (Medtronic Ti-Mesh) modelled to fit the skull shape and implanted in a previous surgery a few weeks earlier[54].

During the second surgical procedure we implanted intramuscular electrodes (Teflon-coated stainless-steel wires, Cooner Wire, cat. no. AS631). Mk-Yg received electrodes in the following arm and hand muscles: Deltoid (DEL), Biceps Brachii (BIC), Triceps Brachii (TRI), Extensor Digitorium Communis (EDC), Flexor Carpi Radialis (FCR), Extensor Carpi Radialis (ECR), Flexor Digitorium Superficialis (FDS). Mk-Br received an additional electrode in the Abductor Pollicis

Brevis (ABP). Due to practical constraints, Mk-Sa received electrodes only in Biceps Brachii (BIC),
Triceps Brachii (TRI) and Flexor Digitorium Superficialis (FDS). In all animals, wires were then
connected to an additional pedestal (Pedestal B), fixated to the titanium mesh.

During the third surgical procedure, monkeys were subjected to a lesion at the cervical level
(C5/C6) of the spinal cord. The surgeon used a micro-blade to cut approximately one third of the
dorsolateral aspect of the spinal cord, in order to interrupt the main component of the corticospinal
tract unilaterally. All monkeys retained autonomic functions, as well as limited arm flexion and
shoulder adduction capabilities. We monitored the animals for the first hours after surgery and
several times daily during the following days. Monitoring scales (score sheets) were used to
assess post-operative pain and general health condition during 1-2 weeks. Antibiotics were given
immediately after the surgery and then once per day for 10 subsequent days, anti-inflammatory
drugs were given once per day for 5 days (Rymadyl 4mg/kg, s.c.; Dexamethasone 0.3mg/kg, s.c.),
and analgesic was given twice per day for 5 days (Temgesic 0.01mg/kg, i.m.). Within the same
procedure, each monkey received a tailored epidural implant. The implant was inserted in the
epidural space of the cervical spinal cord, according to methods described in Schiavone 2020[57]
and Capogrosso 2018[49]. The implant was inserted below the T1 vertebra and pulled until it
covered spinal segments from C6 to T1. We performed intra-operative electrophysiology in order
to assess and refine the implant positioning so that electrodes are aligned to the animal-specific
anatomical features. In particular, we verified that single pulses of stimulation delivered from the
most rostral and most caudal electrodes elicited contractions in the BIC and FDS muscles
respectively. We re-routed the wires subcutaneously in order to connect them to the *Pedestal B.*
All surgical and post-operative care procedures were developed in details in previous reports[49,50].

Data acquisition

For Mk-Sa and Mk-Br, we acquired three-dimensional spatial coordinates of arm and hand joints
using a 14-camera motion tracking system (Figure 1, Vicon Motion Systems, Oxford, UK) that
tracked the Cartesian position of 6 infrared reflective markers (6 to 9 mm in diameter each, Vicon
Motion Systems, Oxford, UK) at a 100 Hz framerate. All markers were placed on the left arm, one
below the shoulder, three on the elbow (proximal, medial and distal position), and two on the left
and right side of the wrist. For each subject, a model of the marker placement was calibrated in
Vicon's Nexus software at the beginning of each experimental session. For Mk-Yg spatial
coordinates of arm and hand joints were recorded using two cameras placed parallel to the sagittal
and transversal plane of the animal (Vicon Motion Systems, Oxford, UK). The 3D coordinates of
the arm and hand joints were extracted using DeepLabCut[75]. Due to the reduced informative
content extracted from the camera parallel to the transverse plane, we then only used 2D
coordinates on the animals' sagittal plane. The training set needed for automatic data labeling
was created by manually labeling a subset of recorded videos. An investigator was blinded to the
experimental condition and was instructed to mark four anatomical landmarks that mirrored the
position of markers in Mk-Sa and Mk-Br (shoulder, medial elbow, left and right wrist). Neural
signals were acquired with a Neural Signal Processor (Blackrock Microsystems, USA) using the
Cereplex-E headstage with a sampling frequency of 30 kHz. Electromyographic signals were
acquired with a Behavioral Neurophysiology chronic recording system (RZ2 BioAmp Processor,
Tucker-Davis Technologies, USA) at a sampling frequency of 12207 Hz.

Electrophysiology in sedated monkeys

Monkeys were sedated with a continuous intravenous infusion of propofol (5 ml/kg/h) that
minimizes effects on spinal cord stimulation[76]. We delivered single pulses of cathodic, charge
balanced, asymmetric square pulses (0.3 ms, 1 Hz) from each electrode contact while recording
compound potentials from all implanted arm and hand muscles. Electromyographic signals were
acquired with a Behavioral Neurophysiology chronic recording system (RZ2 BioAmp Processor,
Tucker-Davis Technologies, USA) at a sampling frequency of 12207 Hz. We then delivered 10
repetitions of pulse trains from each contact, at several frequencies ranging from 20 to 120 Hz.
We recorded compound potentials from all implanted arm and hand muscles and arm kinematics
through two high resolution cameras (Sony FDR-X3000 Action Cam 4K). Through this procedure
we identified three contacts that primarily elicited (1) arm flexors, (2) arm extensors and (3) hand
flexors. In a reduced set of trials, we also recorded the force produced by arm flexion through a
10 N range force sensor (Dual-Range Force Sensor, DFS-BTA, Vernier, Beaverton, Oregon,
USA). To record the pulling force produced during isometric arm flexion, the hand was fixated to
the sensor hook through a string, and the sensor and the elbow were kept in place by two
experimenters, in order to optimally capture the strength produced by muscle contraction.

Behavioral experimental recordings

All animals were trained to perform a three-dimensional robotic reach, grasp and pull task,
previously described in detail in (Barra 2019[54]) and briefly recalled here for simplicity.
All animals were instructed to wait for a start signal by resting the left hand on a metallic bar.
When the "go-cue" was given, monkeys had to reach for and grasp a small spherical object
attached to the robot end effector and located in the three-dimensional space. The object was
placed approximately 180 mm above the animal seating height, 150 mm far from the
shoulder/head coronal plane and 30 mm left of the animal's left arm. Once animals got a hold on
the object, they had to pull it towards their own body until trespassing a virtual spatial threshold.
The accomplishment of such virtual threshold was automatically detected by the robot control
through online monitoring of the end effector position. Once attained the threshold, monkeys had
to let go on the object and go back to the metallic bar. Fruits and vegetables were used to reward
successful movements. Animals were trained daily (5 days per week) and every session ended
as soon as the animals showed any sign of fatigue or impatience.

For Mk-Sa, data presented in this paper were collected several weeks pre lesion and 1week post
lesion, unfortunately a severe infection of the spinal array and EMGs that recurred after day 7
lead to the premature euthanasia of the monkey before the study could be completed, in
agreement with the endpoints in our veterinary authorization. For Mk-Br and Mk-Yg data
presented in this paper were collected several weeks pre lesion and until 3 weeks post lesion. At
the end of week 3 post lesion, Mk-Br had 2 episodes of self-mutilation on the foot ipsi-lateral to
the lesion. In consequence we euthanized the animal before the end of the protocol according to
the endpoints in our veterinary authorization. As described in the results section, we found post-
mortem that Mk-Br had a medial spinal cord contusion at the T3 level. While this lesion did not
affect motor control of the legs or the arms, it may have generated neuropathic pain. Mk-Yg could
perform the entire protocol without any adverse event, however after day 7, the caudal contact of
the spinal interface (E8) identified to promote grasp failed, thus preventing us to perform
experiments with optimal stimulation configuration and impacting the efficacy of grasp movements.

Optimization of EES parameters

To optimized stimulation parameters we exploited the frequency/kinematic relationship that we
observed during single contact stimulation (**Figure 3B,E**). We then analyzed single joint
movements at different frequencies and contacts and weighted joint excursion angles against
movement smoothness[77], we found that stimulation frequencies of 50-60 Hz (**Extended Data**
Figure 5) produced smooth[77] and full-range movements and maximal forces. Instead, movements
elicited at frequencies lower than 40 Hz were too weak to complete a full joint movement while
frequencies higher than 60 Hz produced either abrupt movements or incomplete movements
(**Extended Data Figure 5A**). Next, we identified among all the tested contacts, those that could
consistently elicit arm extension (reach), hand flexion (grasp) and arm flexion (pull) (**Extended**
Data Figure 5B). We chose these contacts and 50-60Hz to sustain full arm and hand movement
and tested their effect in anesthetized animals by sequentially executing bursts on each of these
three contacts. We verified that the sequence triggered whole arm and hand movements that
mimicked smooth[77] and natural multi-joints movements (**Extended Data Figure 5C, Video 1**).
Specifically, extension, grasping and pulling movements produced clear EMG bursts as well as
robust and smooth kinematics. These stimulation protocols could be triggered by an operator at
the beginning of each reach movement or automatically from intra-cortical signals in real-time.
Therefore, we verified that movement onset could be detected from intra-cortical signals even
after SCI (**Extended Data Figure 5**).
655

Stimulation during three-dimensional reach and pull task in injured monkeys

All monkeys were recorded after injury as soon as they could independently move in their housing,
feed themselves autonomously and did not show signs of discomfort. This corresponded to 3, 5
and 6 days after injury respectively for Mk-Yg, Mk-Br and Mk-Sa. After injury, the animals were
reluctant to perform the task which required intense manual activity by the trainers to encourage
them with the use of special positive rewards. Moreover, in consequence of the arm and hand
impairments animals were quickly exhausted. As a result, the output of consistent behavior/day
was low, and we were able to collect robust data in about 1day/week per animal after SCI. Each
session was organized as follows. First, we executed two blocks without stimulation, each of the
duration of approximately 2 minutes. During those blocks we visually evaluated the impairment
level of the animal and the performance of the brain decoder. Second, we used the brain decoder
to trigger specific stimulation patterns. Contacts used to elicit those functions were defined
through the experiments described in the previous paragraph and combined together to create
stimulation protocols that allowed the animal to perform a full reach, grasp and pull movement.

Identification and classification of arm movements for kinematic analysis

We defined the movement performed by the animals as composed of three different phases:
reach, grasp and pull. The identification of the reach phase was done by marking the moment in
which the left hand left the metallic bar to when the hand closed around the object secured to the
robot hand effector (the grasp event). The grasp phase was considered to be a window of 100
ms around the moment in which hand closed around the object. The pull phase started from the
grasp event and finished when the animal accomplished the task by pulling the object across the
virtual spatial threshold and placed the hand back on the resting bar. Events related to the 3
phases of the movement (movement onset: reaching, grasp onset: grasping and release of the
object, and pulling) were identified manually by inspecting video recordings from Vicon Motion

Systems (Oxford, UK). The same method was applied to mark successful and complete
performance of reach, grasp and pull movements as events. A successful reach was defined as
a complete extension of the arm that brought the hand at the position of the target (even when
grasp could not be performed). A successful grasp was defined as a successful closure of the
hand around the target. A successful pull was defined as the accomplishment of a complete
flexion movement that brought the target past the virtual spatial threshold. Events were then
extracted from Vicon and used to perform analysis on the kinematic of the movements and to
train the brain decoder by automatic routines (Matlab 2019b). All the analysis was conducted as
blinded experiments.

Decoding motor states from intracortical signals

We designed a neural decoder that detected reaching and grasping events using intracortical
spiking activity. In order to detect spikes, we set a threshold on each channel of -4 times the root-
mean-square voltage recorded during a brief period while the monkey was at rest. We estimated
firing rates in each of the motor cortical array channels by summing the multiunit spikes with a
150 ms history every 0.5 ms. We used these multiunit firing rate estimates to compute a twenty-
dimensional neural manifold capturing the majority of population variance[62]. We projected the
spiking activity onto this manifold to calibrate a multiclass regularized linear discriminant analysis
decoder[50] that predicted the labeled timing of reach and grasp events. The decoder used 500 ms
of past neural activity and output the probability of observing the reach and grasp events. During
calibration, we defined a probability threshold for each event ranging from 0.8 to 0.99 to optimize
predictions of the timing of each event using cross-validation. Since the monkeys could not
complete the task after SCI, we were unable to consistently acquire labeled training data. We
therefore calibrated a decoding algorithm using reaches from a recording session of a healthy
monkey. We then manually labeled attempted reaches after SCI by manual inspection of video
recordings. Using canonical correlation analysis, we aligned the neural dynamics[78] preceding
reaches on the healthy sessions to the observed neural dynamics preceding attempted reaches
after SCI. These aligned dynamics were used to control the decoder trained on the healthy
reaches.

We implemented a custom C++ software application running a control suite that used the
decoding algorithm to trigger EES stimulation in real-time. The application received neural data
over UDP and made predictions using the decoding algorithm at 15 ms intervals. When the output
probabilities crossed the defined threshold, the application triggered preprogrammed patterns of
EES.

Analysis of muscle recruitment curves

Electromyographic activity was bandpass filtered between 30 and 800 Hz with an offline 3[rd] order
Butterworth filter and stimulus artifact were removed. For each animal, stimulation contact, muscle
and stimulation amplitude, we extracted compound potentials from 50ms-long segments of
electromyographic activity following a stimulation pulse. We then computed the peak-to-peak
amplitude of compound potentials. Since we gave four pulses of stimulation for each selected
current amplitude, we averaged across values corresponding to the same stimulation amplitude
and represented as the mean recruitment value of each muscle as a function of the injected
current. For each muscle, recruitment values have been subsequently normalized by the
maximum value obtained for that specific muscle, provided that we obtained response saturation
(and therefore maximal contraction) in at least one occasion during the session. In addition, we

computed a selectivity index for each muscle[79].

In order to obtain a comprehensive measure of muscle recruitment for each contact that would
allow to compare across animals, we computed, for each animal, each muscle and each contact,
an Average Recruitment Index (ARI) as the average of the recruitment values across all
stimulation amplitudes used from a specific stimulation site.

To compute muscle recruitment during the delivery of pulse train stimulation, we computed the
energy of the EMG signal during the duration of stimulation. We then applied the same
normalization procedure described above for single pulse recruitment.

Analysis of muscle activity during EES

Electromyographic activity was bandpass filtered between 30 and 800 Hz with an offline 3rd order
Butterworth filter and stimulus artifact were removed. In all animals we computed the energy EMG
signals, for each implanted muscle. Energy of EMG signals during stimulation was computed on
each segment in which stimulation was delivered after the animal started a movement attempt,
with the formula here below:

$EN_{EMG} = \frac{1}{N} \sum_{i}^{N} \|EMG_i\|^2 \, dt$

Where EMG_i is the value of EMG activity at sample i, N is the number of samples in the signal
and dt is the sampling resolution.

Energy of EMG signals without stimulation was computed on each segment in which stimulation
was not delivered and the animal started a movement attempt. A movement attempt was defined
as an increased EMG activity of the Biceps and Deltoid muscles.

Analysis of task performance

We computed task performance with two different measures. First, we computed the success rate
as the percentage of successful movement across all movement attempts. Successful
movements were identified by a blind experimenter as movements performed skillfully and until
the end (see above, *Identification and classification of arm movements for kinematic analysis*).
Movement attempts were identified as all movements executed in response to a *go* cue and
included successful movements too. Second, we computed the task performance frequency as
the rate of successful movements per unit of time. In order to do this, we subdivided sessions in
time bins of 2 seconds and we marked the presence or absence of successful trials, both with
and without stimulation. We then used bootstrap to analyze significance of those results. We
normalized all the results by the mean success rate during stimulation.

Analysis of kinematics performance

We performed Principal Component Analysis on a large set of kinematic features. We computed
the features on data segments during the reach phase and the pull phase (see movement
identification explained above, section *Identification and classification of arm movements for
kinematic analysis*). All kinematic signals were previously low pass filtered at 6 Hz. Segments

were not interpolated nor resampled. Before performing PCA analysis, features were centered to
have mean 0 and scaled to have standard deviation of 1 (Matlab 2019). The computed features
for Mk-Br included: minimum value, maximum value and total excursion of joint angles (shoulder
flexion, elbow flexion, and wrist pronation); maximum, minimum and average angular velocity (for
the shoulder flexion, elbow flexion and wrist pronation); minimum, maximum and average position
along the sagittal, frontal and vertical axis of each arm joint (shoulder, elbow, wrist); maximum
minimum and average wrist velocity along the sagittal, frontal and vertical axis; movement
smoothness[77]; trajectory length during and time required to complete movements. All the listed
features have been computed identically during the reach phase and the pull phase separately
and treated as different features. In addition, computed maximal applied three-dimensional pulling
force and the average position along the sagittal, frontal and vertical axis of each arm joint
(shoulder, elbow, wrist) during grasp.

Since for Mk-Yg we only extracted 2D kinematics on the sagittal plane, the kinematic features for
Mk-Yg included: minimum value, maximum value and total excursion of joint angles (shoulder
flexion and elbow flexion); maximum and average angular velocity (for the shoulder flexion and
elbow flexion); minimum, maximum and average position along the sagittal and vertical axis of
each arm joint (shoulder, elbow, wrist); maximum and average wrist velocity along the sagittal
and vertical axis; movement smoothness[77]; trajectory length during and time required to complete
movements. All the listed features have been computed during the reach phase.

Processing of cortical signals

We identified spiking events on each channel when the band-pass filtered signal (250 Hz–5kHz)
exceeded 3.0–3.5 times its root-mean-square value calculated over a period of 5s. We removed
artifacts by deleting all the spikes that synchronously in at least 30 channels. We computed the
firing rate of each channel as the number of spikes detected over non-overlapping bins of 10ms.
Whenever we showed average firing rate activity, we sorted channels in order of activation in one
reference trial, and subsequently applied the same ordering method to all other trials. Finally, we
normalized the activity of each channel by its maximum firing rate.

Comparison of motor cortical activity during EES evoking movement and no movement

To study how motor cortical activity interacted with EES, we analyzed the neural recordings from
Mk-Br and Mk-Yg. We identified periods where EES pulse trains produced no discernible
movements by setting a threshold on hand velocity. We compared multi-unit neural firing rates on
each channel in this period to neural firing rates in the previously identified trials where EES
enabled reaching and grasping. First, we counted the number of spikes within the window of
stimulation and divided by the duration of stimulation. We then averaged across stimulus
repetitions of the movement and no movement conditions and pooled across recording sites in
motor cortex.

We next computed instantaneous estimates of multi-unit firing rates on each channel by counting
the number of spikes in non-overlapping 20 ms bins and convolving with a gaussian kernel of 50
ms width. We applied Principal Component Analysis (PCA) to compute 10-dimensional neural
manifolds spanning this multi-unit population activity[62]. We projected the neural activity onto these
manifold axes during the periods where EES evoked either movement or no movement. We then
identified periods where the monkey was at rest with no EES, as well as periods where the
monkey attempted movements of the arm with no EES. To compare the similarity of neural activity

between these conditions, we computed the Mahalananobis distance between activity at rest and
the three other periods: EES with movement, EES with no movement, and attempted movements
with no EES.

Histology

Monkeys were deeply anesthetized (lethal dose of pentobarbital, 60mg/kg, injected i.v.) and
transcardially perfused with saline (about 200 ml), followed by 3 liters of 4% paraformaldehyde
(PFA). Dissected spinal cord were post-fixed in 4% PFA overnight, and then immersed in 30%
sucrose solution for 2 weeks. 50μm transverse or horizontal sections were cut using a cryostat
and kept in 0.1M PBS azide (0.03%) at 4°C. Primary antibodies were: rabbit anti-Iba1 (1:1000,
Wako) and guinea pig anti-NeuN (1:300, Millipore). Fluorescence secondary antibodies were
conjugated to: Alexa fluor 647 and Alexa fluor 555 (Life technologies). Sections were coverslipped
using Mowiol. Immunofluorescence was imaged digitally using a slide scanner (Olympus VS-120).
Lesions were reconstructed using image analysis software (Neurolucida) to trace the lesion over
serial sections (200 μm apart).

Statistical procedures

All data are reported as mean values ± standard error of the mean (s.e.m.) or mean values ±
standard deviation (std). The choice is highlighted directly in the figures or in the relative caption.
Significance was analyzed using the non-parametric Wilcoxon rank-sum test. In the comparisons
shown in Figure 3 we subsequently applied the Bonferroni correction. In only one case (Figure
4A, 4B, 4C), significance was analyzed using bootstrap. The level of significance was set at
*p<0.05, **p<0.01, ***p<0.001.

References

1. ICCP. International Campaign for Cures of Spinal Cord Injury Paralysis.

   http://www.campaignforcure.org.

2. National Center for Chronic Disease Prevention and Health Promotion , Division for Heart

   Disease and Stroke. Stroke Facts. https://www.cdc.gov/stroke/facts.htm (2020).

3. Anderson, K. D. Targeting recovery: priorities of the spinal cord-injured population. *Journal*

   of neurotrauma **21**, 1371–1383 (2004).

4. Moreland, J. D. *et al.* Needs assessment of individuals with stroke after discharge from

   hospital stratified by acute Functional Independence Measure score. *Disability and*

   rehabilitation **31**, 2185–2195 (2009).

5. Lemon, R. N. Descending pathways in motor control. *Annual review of neuroscience* **31**,

   195–218 (2008).

6. Griffin, D. M., Hoffman, D. S. & Strick, P. L. Corticomotoneuronal cells are 'functionally

   tuned'. *Science* **350**, 667–70 (2015).

7. Griffin, D. M. & Strick, P. L. The motor cortex uses active suppression to sculpt movement.

   Sci Adv **6**, eabb8395 (2020).

8. Seki, K., Perlmutter, S. I. & Fetz, E. E. Sensory input to primate spinal cord is

   presynaptically inhibited during voluntary movement. *Nat Neurosci* **6**, 1309–16 (2003).

9. Lebedev, M. A. & Nicolelis, M. A. Brain-machine interfaces: From basic science to

   neuroprostheses and neurorehabilitation. *Physiological reviews* (2017).

10. Ethier, C., Oby, E. R., Bauman, M. J. & Miller, L. E. Restoration of grasp following paralysis

    through brain-controlled stimulation of muscles. *Nature* **485**, 368–371 (2012).

11. Moritz, C. T., Perlmutter, S. I. & Fetz, E. E. Direct control of paralysed muscles by cortical

    neurons. *Nature* **456**, 639–642 (2008).

12. Bouton, C. E. *et al.* Restoring cortical control of functional movement in a human with

quadriplegia. *Nature* (2016) doi:10.1038/nature17435.

13. Ajiboye, A. B. *et al.* Restoration of reaching and grasping movements through brain-

controlled muscle stimulation in a person with tetraplegia: a proof-of-concept demonstration.

The Lancet (2017) doi:10.1016/S0140-6736(17)30601-3.

14. Giat, Y., Mizrahi, J. & Levy, M. A musculotendon model of the fatigue profiles of paralyzed

quadriceps muscle under FES. *IEEE transactions on biomedical engineering* **40**, 664–674

(1993).

15. Popovic, M. R., Popovic, D. B. & Keller, T. Neuroprostheses for grasping. *Neurological*

research **24**, 443–452 (2002).

16. Edgerton, V. R. *et al.* Training locomotor networks. *Brain Res Rev* **57**, 241–254 (2008).

17. Holinski, B. J. *et al.* Intraspinal microstimulation produces over-ground walking in

anesthetized cats. *J Neural Eng* **13**, 056016 (2016).

18. Zimmermann, J. B., Seki, K. & Jackson, A. Reanimating the arm and hand with intraspinal

microstimulation. *Journal of neural engineering* **8**, 054001 (2011).

19. Gaunt, R. A., Prochazka, A., Mushahwar, V. K., Guevremont, L. & Ellaway, P. H. Intraspinal

microstimulation excites multisegmental sensory afferents at lower stimulus levels than local

α-motoneuron responses. *Journal of neurophysiology* **96**, 2995–3005 (2006).

20. Rattay, F., Minassian, K. & Dimitrijevic, M. R. Epidural electrical stimulation of posterior

structures of the human lumbosacral cord: 2. quantitative analysis by computer modeling.

Spinal cord **38**, 473–489 (2000).

21. Capogrosso, M. *et al.* A computational model for epidural electrical stimulation of spinal

sensorimotor circuits. *The Journal of neuroscience : the official journal of the Society for*

Neuroscience **33**, 19326–40 (2013).

22. Moraud, E. M. *et al.* Mechanisms Underlying the Neuromodulation of Spinal Circuits for

Correcting Gait and Balance Deficits after Spinal Cord Injury. *Neuron* **89**, 814–28 (2016).

23. Formento, E. *et al.* Electrical spinal cord stimulation must preserve proprioception to enable

locomotion in humans with spinal cord injury. *Nature neuroscience* **21**, 1728 (2018).

24. Wenger, N. *et al.* Closed-loop neuromodulation of spinal sensorimotor circuits controls

refined locomotion after complete spinal cord injury. *Science translational medicine* **6**,

255ra133-255ra133 (2014).

25. Harkema, S. *et al.* Effect of epidural stimulation of the lumbosacral spinal cord on voluntary

movement, standing, and assisted stepping after motor complete paraplegia: a case study.

The Lancet **377**, 1938–1947 (2011).

26. Angeli, C. A., Edgerton, V. R., Gerasimenko, Y. P. & Harkema, S. J. Altering spinal cord

excitability enables voluntary movements after chronic complete paralysis in humans. *Brain*

137, 1394–1409 (2014).

27. Ichiyama, R. M., Gerasimenko, Y. P., Zhong, H., Roy, R. R. & Edgerton, V. R. Hindlimb

stepping movements in complete spinal rats induced by epidural spinal cord stimulation.

Neurosci Lett **383**, 339–44 (2005).

28. Courtine, G. *et al.* Transformation of nonfunctional spinal circuits into functional states after

the loss of brain input. *Nat Neurosci* **12**, 1333–1342 (2009).

29. van den Brand, R. *et al.* Restoring Voluntary Control of Locomotion after Paralyzing Spinal

Cord Injury. *Science* **336**, 1182–1185 (2012).

30. Harkema, S. *et al.* Effect of epidural stimulation of the lumbosacral spinal cord on voluntary

movement, standing, and assisted stepping after motor complete paraplegia: a case study.

Lancet **377**, 1938–47 (2011).

31. Grahn, P. J. *et al.* Enabling Task-Specific Volitional Motor Functions via Spinal Cord

Neuromodulation in a Human With Paraplegia. *Mayo Clin Proc* **92**, 544–554 (2017).

32. Angeli, C. A. *et al.* Recovery of Over-Ground Walking after Chronic Motor Complete Spinal

Cord Injury. *N Engl J Med* **379**, 1244–1250 (2018).

33. Gill, M. L. *et al.* Neuromodulation of lumbosacral spinal networks enables independent

stepping after complete paraplegia. *Nat Med* (2018) doi:10.1038/s41591-018-0175-7.

34. Alam, M. *et al.* Evaluation of optimal electrode configurations for epidural spinal cord

stimulation in cervical spinal cord injured rats. *Journal of neuroscience methods* **247**, 50–57

(2015).

35. Lu, D. C. *et al.* Engaging Cervical Spinal Cord Networks to Reenable Volitional Control of

Hand Function in Tetraplegic Patients. *Neurorehabil Neural Repair* **30**, 951–962 (2016).

36. Inanici, F., Brighton, L. N., Samejima, S., Hofstetter, C. P. & Moritz, C. T. Transcutaneous

spinal cord stimulation restores hand and arm function after spinal cord injury. *IEEE Trans*

Neural Syst Rehabil Eng Online ahead of print-Online ahead of print (2021)

doi:10.1109/TNSRE.2021.3049133.

37. Kapadia, N., Zivanovic, V. & Popovic, M. Restoring voluntary grasping function in individuals

with incomplete chronic spinal cord injury: pilot study. *Topics in spinal cord injury*

rehabilitation **19**, 279–287 (2013).

38. Grillner, S. The motor infrastructure: from ion channels to neuronal networks. *Nat Rev*

Neurosci **4**, 573–586 (2003).

39. McCrea, D. A. & Rybak, I. A. Organization of mammalian locomotor rhythm and pattern

generation. *Brain Res Rev* **57**, 134–46 (2008).

40. Bizzi, E., Giszter, S. F., Loeb, E., Mussa-Ivaldi, F. A. & Saltiel, P. Modular organization of

motor behavior in the frog's spinal cord. *Trends in neurosciences* **18**, 442–446 (1995).

41. Giszter, S. F. Motor primitives--new data and future questions. *Curr Opin Neurobiol* **33**,

156–65 (2015).

42. Alstermark, B. & Isa, T. Circuits for skilled reaching and grasping. *Annu Rev Neurosci* **35**,

559–78 (2012).

43. Lemon, R. N. & Griffiths, J. Comparing the function of the corticospinal system in different

species: organizational differences for motor specialization? *Muscle & Nerve: Official*

Journal of the American Association of Electrodiagnostic Medicine **32**, 261–279 (2005).

44. Kinoshita, M. *et al.* Genetic dissection of the circuit for hand dexterity in primates. *Nature*

487, 235–238 (2012).

45. Weiler, J., Gribble, P. L. & Pruszynski, J. A. Spinal stretch reflexes support efficient hand

control. *Nat. Neurosci* 1–11 (2019).

46. Omrani, M., Murnaghan, C. D., Pruszynski, J. A. & Scott, S. H. Distributed task-specific

processing of somatosensory feedback for voluntary motor control. *Elife* **5**, e13141 (2016).

47. Sauerbrei, B. A. *et al.* Cortical pattern generation during dexterous movement is input-

driven. *Nature* **577**, 386–391 (2020).

48. Wenger, N. M. M., E; Gandar, J; Musienko, P; Capogrosso, M; Baud, L; Legoff, C; Barraud,

Q; Pavlova, N; Domonici N; Minev, I; Asboth, L; Hirsch, A; Duis, S; Kreider, J; Mortera, A;

Haverbeck, O; Kraus, S; Schmitz, F; DiGiovanna, J; van den Brand, R; Bloch, J; Detemple,

P; Lacour, S; Bezard, E; Micera, S; Courtine G. Spatiotemporal neuromodulation therapies

engaging muscle synergies to improve motor control after spinal cord injury. *Nature*

Medicine (2015).

49. Capogrosso, M. *et al.* Configuration of electrical spinal cord stimulation through real-time

processing of gait kinematics. *Nat Protoc* (2018) doi:10.1038/s41596-018-0030-9.

50. Capogrosso, M. *et al.* A brain–spine interface alleviating gait deficits after spinal cord injury

in primates. *Nature* **539**, 284–288 (2016).

51. Wagner, F. B. *et al.* Targeted neurotechnology restores walking in humans with spinal cord

injury. *Nature* **563**, 65 (2018).

52. Ladenbauer, J., Minassian, K., Hofstoetter, U. S., Dimitrijevic, M. R. & Rattay, F. Stimulation

of the Human Lumbar Spinal Cord With Implanted and Surface Electrodes: A Computer

Simulation Study. *Neural Systems and Rehabilitation Engineering, IEEE Transactions on*

18, 637–645 (2010).

53. Greiner, N. *et al.* Recruitment of Upper-Limb Motoneurons with Epidural Electrical

Stimulation of the Primate Cervical Spinal Cord. *Nat Commun* (2021).

54. Barra, B. *et al.* A Versatile Robotic Platform for the Design of Natural, Three-Dimensional

Reaching and Grasping Tasks in Monkeys. *Journal of Neural Engineering* (2019)

doi:10.1088/1741-2552/ab4c77.

55. Jenny, A. B. & Inukai, J. Principles of motor organization of the monkey cervical spinal cord.

J Neurosci **3**, 567–75 (1983).

56. Capogrosso, M. *et al.* A brain-spine interface alleviating gait deficits after spinal cord injury

in primates. *Nature* **539**, 284–288 (2016).

57. Schiavone, G. *et al.* Soft, Implantable Bioelectronic Interfaces for Translational Research.

Advanced Materials **32**, 1906512 (2020).

58. Chao, Z. C., Sawada, M., Isa, T. & Nishimura, Y. Dynamic Reorganization of Motor

Networks During Recovery from Partial Spinal Cord Injury in Monkeys. *Cereb Cortex* (2018)

doi:10.1093/cercor/bhy172.

59. Freund, P. *et al.* Nogo-A–specific antibody treatment enhances sprouting and functional

recovery after cervical lesion in adult primates. *Nature medicine* **12**, 790–792 (2006).

60. Sharpe, A. N. & Jackson, A. Upper-limb muscle responses to epidural, subdural and

intraspinal stimulation of the cervical spinal cord. *J Neural Eng* **11**, 016005 (2014).

61. Gerasimenko, Y. P. *et al.* Spinal cord reflexes induced by epidural spinal cord stimulation in

normal awake rats. *Journal of Neuroscience Methods* **157**, 253–263 (10).

974 62. Gallego, J. A., Perich, M. G., Miller, L. E. & Solla, S. A. Neural Manifolds for the Control of

975   Movement. *Neuron* **94**, 978–984 (2017).

976 63. Lempka, S. F. *et al.* Patient-specific analysis of neural activation during spinal cord

977   stimulation for pain. *Neuromodulation: Technology at the Neural Interface* **23**, 572–581

978   (2020).

979 64. Schirmer, C. M. *et al.* Heuristic map of myotomal innervation in humans using direct

980   intraoperative nerve root stimulation. *J Neurosurg Spine* **15**, 64–70 (2011).

981 65. Moritz, C. T., Lucas, T. H., Perlmutter, S. I. & Fetz, E. E. Forelimb movements and muscle

982   responses evoked by microstimulation of cervical spinal cord in sedated monkeys. *Journal*

983   *of neurophysiology* **97**, 110–120 (2007).

984 66. Kato, K., Nishihara, Y. & Nishimura, Y. Stimulus outputs induced by subdural electrodes on

985   the cervical spinal cord in monkeys. *Journal of Neural Engineering* **17**, 016044 (2020).

986 67. de Freitas, R. M. *et al.* Selectivity and excitability of upper-limb muscle activation during

987   cervical transcutaneous spinal cord stimulation in humans. *Journal of Applied Physiology*

988   (2021).

989 68. Kirsch, R. & Rymer, W. Neural compensation for muscular fatigue: evidence for significant

990   force regulation in man. *Journal of neurophysiology* **57**, 1893–1910 (1987).

991 69. Song, S. & Geyer, H. A neural circuitry that emphasizes spinal feedback generates diverse

992   behaviours of human locomotion. *The Journal of physiology* (2015) doi:10.1113/JP270228.

993 70. Confais, J., Kim, G., Tomatsu, S., Takei, T. & Seki, K. Nerve-specific input modulation to

994   spinal neurons during a motor task in the monkey. *Journal of Neuroscience* **37**, 2612–2626

995   (2017).

996 71. Seáñez, I. & Capogrosso, M. Motor improvements enabled by spinal cord stimulation

997   combined with physical training after spinal cord injury: review of experimental evidence in

998   animals and humans. *Bioelectronic Medicine* **7**, 1–13 (2021).

72. Granat, M., Heller, B., Nicol, D., Baxendale, R. & Andrews, B. Improving limb flexion in FES

gait using the flexion withdrawal response for the spinal cord injured person. *Journal of*

biomedical engineering **15**, 51–56 (1993).

73. Ting, J. *et al.* A wearable neural interface for detecting and decoding attempted hand

movements in a person with tetraplegia. in 1930–1933 (IEEE, 2019).

74. National Research Council (US) Institute for Laboratory Animal Research. *Guide for the*

Care and Use of Laboratory Animals. (National Academies Press (US), 1996).

75. Mathis, A. *et al.* DeepLabCut: markerless pose estimation of user-defined body parts with

deep learning. *Nature neuroscience* **21**, 1281–1289 (2018).

76. Toossi, A. *et al.* Effect of anesthesia on motor responses evoked by spinal neural

prostheses during intraoperative procedures. *Journal of neural engineering* **16**, 036003

(2019).

77. Teulings, H.-L., Contreras-Vidal, J. L., Stelmach, G. E. & Adler, C. H. Parkinsonism

Reduces Coordination of Fingers, Wrist, and Arm in Fine Motor Control. *Experimental*

Neurology **146**, 159–170 (1997).

78. Gallego, J. A., Perich, M. G., Chowdhury, R. H., Solla, S. A. & Miller, L. E. Long-term

stability of cortical population dynamics underlying consistent behavior. *Nature*

neuroscience **23**, 260–270 (2020).

79. Raspopovic, S., Capogrosso, M. & Micera, S. A Computational Model for the Stimulation of

Rat Sciatic Nerve Using a Transverse Intrafascicular Multichannel Electrode. *IEEE*

Transactions on Neural Systems and Rehabilitation Engineering **19**, 333–344 (2011).

1020