MinION™, a portable long-read sequencer, enables rapid vaginal microbiota

analysis in a clinical setting

Shinnosuke Komiya[1,2], Yoshiyuki Matsuo[3], So Nakagawa[4], Yoshiharu Morimoto[2], Kirill

Kryukov[5], Hidetaka Okada[1], and Kiichi Hirota[3]

1 Department of Obstetrics and Gynecology, Kansai Medical University Graduate

School of Medicine, Osaka, Japan

2 HORAC Grand Front Osaka Clinic, Osaka, Japan

3 Department of Human Stress Response Science, Institute of Biomedical Science,

Kansai Medical University, Osaka, Japan

4 Department of Molecular Life Science, Tokai University School of Medicine,

Kanagawa, Japan

5 Department of Genomics and Evolutionary Biology, National Institute of Genetics,

Shizuoka, Japan

Correspondence:

Yoshiyuki Matsuo

E-mail: ysmatsuo-kyt@umin.ac.jp

Keywords:

16S rRNA; bacterial vaginosis; long-read sequencer; MinION™; nanopore

Abbreviations:

BV, bacterial vaginosis; CST, community state type; PCoA, principal coordinate

analysis; PCR, polymerase chain reaction; rRNA, ribosomal RNA.

Running heading:

MinION™ enables rapid vaginal microbiota analysis.

Abstract

It has been suggested that the local microbiota in the reproductive organs is relevant to

women's health and may also affect pregnancy outcomes. Analysis of partial 16S

ribosomal RNA (rRNA) gene sequences generated by short-read sequencers has been

used to identify vaginal and endometrial microbiota, but it requires a long time to obtain

the results, making it unsuitable for the rapid analysis of small samples in a clinical

context. We demonstrated a simple workflow using the nanopore sequencer MinION™

that allows high-resolution and rapid differentiation of vaginal microbiota. Vaginal

samples collected from 18 participants were subjected to DNA extraction and full-

length 16S rRNA gene sequencing with MinION™. The principal coordinate analysis

showed no differences in the bacterial compositions regardless of the sample collection

method. The vaginal microbiota results could be reported within 2 days of specimen

receipt. Although bacterial vaginosis (BV) was not diagnosed by the Nugent score in

any cases, groups with both healthy and BV-like vaginal microbiota were clearly

characterized by MinION™ sequencing. We conclude that full-length 16S rRNA gene

sequencing analysis with MinION™ provides a rapid means for identifying vaginal

bacteria with higher resolution. Species-level profiling of human vaginal microbiota by

MinION™ sequencing can allow the analysis of associations with conditions such as

genital infections, endometritis, and threatened miscarriage.

Introduction

Bacterial vaginosis (BV) is one of the most common gynecological disorders among women of reproductive age [1]. It is caused by disruption of the vaginal microbiota due to various factors such as psychological stress, poor physical condition, menstrual cycle, and pregnancy. It has relatively mild subjective symptoms, such as abnormal vaginal discharge (odor, color, amount of discharge) and itching, but there are cases of unrecognized infection and recurrent infections without significant symptoms [1]. BV has been clinically reported to cause adverse reproductive outcomes, including sexually transmitted infections, genitourinary viral infections such as HIV-1/2, HPV, and HSV-2 [2], and preterm delivery [3].

Currently, Amsel's criteria based on clinical outcomes [4] and the Nugent score using Gram stain findings [5] are used as diagnostic criteria for BV. These methods are easy to use, but since they are based on the subjective opinion of the physician, the diagnosis largely depends on the skill of the examiner [6]. Bacterial culture of vaginal discharge is also a standard method to investigate the vaginal microbiota, but some bacteria are difficult to culture [6]. Moreover, in routine gynecological practice, vaginal discharge is not sampled unless patients have symptoms suggestive of vaginosis. Therefore, accurately assessing the vaginal microbiota has been difficult with these regular clinical methods. Due to the lack of a reliable methodology, the association of BV with other vaginal disorders has not been fully elucidated. Until recently, a more detailed analysis of vaginal microbiota in asymptomatic women was considered excessive and might not be clinically relevant. Given that BV has been implicated in women's health outcomes in various ways, it is crucial to develop a rapid and accurate means to diagnose BV with higher sensitivity. The increased risk of various infections, infertility, miscarriage and preterm birth due to BV has been shown to be impacted by the species that make up the vaginal microbiota [1-3]. In this context, accurate representation of the vaginal microbiota will allow for more appropriate medical treatment. If abnormal vaginal

microbiota is identified with persistent cervical lesions, the dysbiosis should be treated

even in the absence of symptoms. Detection of asymptomatic BV will potentially be

beneficial for the diagnosis and prevention of gynecological diseases.

16S ribosomal RNA (rRNA) gene sequencing analysis using next-generation sequencers

can be used to identify bacteria without culture in a manner not dependent on subjective

assessment by the examiner. Previous studies have shown that the vaginal microbiota of

healthy women is classified into community state type (CST) I to V [7]. CST is defined

by the dominant bacteria as follows: CST I: *Lactobacillus crispatus*, CST II:

Lactobacillus gasseri, CST III: *Lactobacillus iners*, CST IV: lack of *Lactobacillus* spp.,

and CST V: *Lactobacillus jensenni*. These narrow definitions mean that undiagnosed

cases of vaginal dysbiosis may be present. This new classification of the vaginal

microbiota has been clarified by performing sequencing analysis on cases with no

clinical symptoms. Because a deeper and more accurate understanding of the vaginal

microbiota may provide new clinical insights, we should demonstrate that 16S rRNA

analysis is quick and easy to perform in a broad range of contexts in clinical practice.

MinION™ detects the signal of a DNA nucleotide that passes through a nanopore

arranged on a flow cell [8]. This long-read sequencing technology can be used to

identify bacterial microbiota with the whole region of the 16S rRNA. MinION™ is

smaller, lighter, and easier to deploy than conventional massively parallel sequencing

platforms, and it also offers the advantage of high-resolution microbiota analysis with

long-read sequencing [9-11]. We constructed an in-hospital vaginal microbiota-

analyzing workflow using MinION™. This is the first report to demonstrate that

MinION™ can provide rapid patient feedback regarding vaginal microbiota analysis.

Materials and methods

Ethics statement

The study adhered to the Declaration of Helsinki throughout the protocol, and written

informed consent was obtained from all participants. This study is a cross-sectional

observational trial approved by the institutional review board of the medical corporation

Sankeikai (Approval Number: 2019-38). After the approval, the study protocol was

submitted to the UMIN-CTR clinical trial registry site

(https://www.umin.ac.jp/ctr/index.htm) (trial registration number: UMIN000038032),

which met the criteria of the International Committee of Medical Journal Editors

(ICMJE).

Recruitment of participants

From April 2019 to May 2020, participants who met all of the following enrollment

criteria were recruited: 1) primary infertility patients attending HORAC Grand Front

Osaka Clinic in Osaka, Japan; 2) premenopausal Japanese women who were not

pregnant; 3) no clinical symptoms strongly suggestive of BV; 4) scheduled hormone

replacement embryo transfer cycle with frozen-thawed blastocyst; and 5) those who

provided written consent. The exclusion criteria were as follows: 1) known history of

HIV-1/2, hepatitis B/C, syphilis, or genital chlamydia infection; 2) known history of

diabetes; 3) known history of cervical/vaginal surgery; 4) known history of intrauterine

device use; and 5) known history of antibiotic, steroid, or vaginal suppository use within

2 weeks. In addition to collecting vaginal samples, details of age, body mass index, and

basal hormone levels (anti-Mullerian hormone, luteinizing hormone, and follicle-

stimulating hormone on days 2–4 of the menstrual cycle), medical history, and infertility

treatment history were collected from the medical records. All vaginal samples were

collected on the day of embryo transfer.

Nugent score

Vaginal smears were analyzed by two pathologists affiliated with an external laboratory,

and the Nugent score was calculated based on microscopic findings: the numbers of

Lactobacillus (scored as 0 to 4), *Gardnerella* (scored as 0 to 4), and *Mobiluncus* (scored

as 0 to 2). A total score of 0 to 3 was diagnosed as representative of healthy vaginal

microbiota, 4 to 6 as an intermediate group, and 7 or more as BV [5]. Inappropriate

samples with defective specimen collection were diagnosed as indeterminate. If BV was

diagnosed based on the Nugent score, our policy was to treat it using antibiotics.

Vaginal sample collection method

Vaginal lavage samples were collected from 18 participants; from four participants,

additional swab samples were collected with written consent. When both lavage and

swab samples were collected, the swab samples were collected first. For the lavage

method, the inside of the vagina was washed with 10 ml of sterile saline, and at least 2

ml was collected using a sterile syringe. Collected lavage samples were stored at –30 °C

until DNA extraction. For the swab method, OMNIgene vaginal kit (OMR-130; DNA

Genotek Inc., Ottawa, Canada) was used to infiltrate vaginal mucus from the

uterovaginal region, and the swab tip was stored in the preservative solution of the

sampling kit. Swab samples were stored at room temperature until DNA extraction in

accordance with the manufacturer's instructions.

Sequencing sample preparation

DNA was extracted from 22 human vaginal samples using the QIAamp UCP Pathogen

Mini Kit (QIAGEN, Venlo, Netherlands) following the manufacturer's instructions.

Briefly, samples were subjected to mechanical cell lysis by bead-beating [12], and DNA

was isolated using silica membrane-based spin columns. After extraction, the DNA

content was measured using a NanoDrop® 1000 Spectrophotometer (Thermo Fisher

Scientific, MA, USA), and the concentrations of extracted DNA from the swab and

lavage samples were 6.1–32.3 ng/µl and 53.1–1092.4 ng/µl, respectively.

MinION™ sequencing

As described in our previous report [12], a slightly modified version of four-primer

polymerase chain reaction (PCR) with rapid adapter attachment was performed

(https://dx.doi.org/10.17504/protocols.io.bs9knh4w). For amplification of the V1–9

region of the 16S rRNA gene, a forward primer (S-D-Bact-0008-c-S-20) with the

anchor sequence 5'-

TTTCTGTTGGTGCTGATATTGCAGRGTTYGATYMTGGCTCAG-3' and a reverse

primer with the anchor sequence 5'-

ACTTGCCTGTCGCTCTATCTTCCGGYTACCTTGTTACGACTT-3' were used as

inner primers. For amplification of the V3–4 region, 341F with the anchor sequence 5'-

TTTCTGTTGGTGCTGATATTGCCCTACGGGNGGCWGCAG-3' and 806R with the

anchor sequence 5'-

ACTTGCCTGTCGCTCTATCTTCGGACTACHVGGGTWTCTAAT-3' were used as

inner primers. PCR amplification of the 16S rRNA gene was performed using the

KAPA2G™ Robust HotStart ReadyMix PCR Kit (Kapa Biosystems, MA, USA) with an

inner primer pair (50 nM for each) and the outer primer mixture (10 nM) of PCR

Barcoding Kit (SQK-PBK004; Oxford Nanopore Technologies, Oxford, UK) in a total

volume of 25 μl. Each sample was assigned an individual barcode, which was used to

identify the PCR product. Amplification was performed under the following PCR

conditions: initial denaturation at 95 °C for 3 min; five cycles of 15 s at 95 °C, 15 s at

55 °C, and 30 s at 72 °C; and 30 cycles of 15 s at 95 °C, 15 s at 62 °C, and 30 s at

72 °C; followed by a final extension at 72 °C for 1 min. The amplification products

were confirmed using 1% agarose gel electrophoresis (1 × TAE buffer) to be

approximately 1600 base pairs (V1–9 region) and 400 base pairs (V3–4 region), which

are the expected amplification sizes. After confirmation, the amplified DNA was

purified using AMPure® XP (Beckman Coulter, CA, USA). A total of 100 ng (10 μl) of

purified DNA was incubated with 1 μl of Rapid Adapter at room temperature for 5 min.

The concentrations of swab samples and lavage samples were 8.3–34 ng/µl and 3.58–61

ng/µl, respectively, as determined using a Quantus™ Fluorometer (Promega, WI, USA).

The prepared DNA library (total of 11 µl) was mixed with 34 µl of Sequencing Buffer,

25.5 µl of Loading Beads, and 4.5 µl of water. The final adjusted sample was loaded

into the flow cell R.9.4.1 (FLO-MIN106; Oxford Nanopore Technologies Ltd.) and

attached to the MinION™ Mk1B, which was connected to a personal computer. In this

study, four to six samples were loaded at one time and each sequencing session lasted

about 90 min. DNA sequencing data were acquired in FAST5 format using MINKNOW

software ver. 1.11.5 (Oxford Nanopore Technologies Ltd.).

Bioinformatic analysis workflow

We performed the bioinformatic analysis using a pipeline of multiple programs built in

accordance with our previous reports [9,11]. An overview is given below.

1) Computer: Apple iMac 27-inch, Late 2015 (OS, macOS 10.14.6; CPU, 3.3 GHz

Intel Core i5-6600; memory, 16 GB)

2) GUPPY software ver. 3.1.5 (Oxford Nanopore Technologies Ltd.): base calling,

conversion of FAST5 files to FASTQ files with quality control based on quality

score > 7.

3) SeqKit software ver. 0.10.0 [13]: extraction of a read length of 1,300 to 1,950 bp,

based on the 16S rRNA length registered in the SILVA rRNA database ver. 132

(https://www.arb-silva.de/) [14].

4) TANTAN program ver. 18 [15]: removing simple repeat sequences.

5) Minimap2 program ver. 2.14 [16]: eliminating human genome information (Human

Genome Assembly GRCh38,

https://www.ncbi.nlm.nih.gov/assembly/GCF_000001405.26/) and matching each

genome read to the 5,850 representative bacterial genome sequences (Table S1)

stored in the GenomeSync database (http://genomesync.org) [11,12].

6) In-house Perl scripts (Genome Search Toolkit, http://kirill-

kryukov.com/study/tools/gstk/): selecting the species with the highest Minimap2

score and determining the taxa based on the NCBI taxonomy database.

7) Krona software version 2.7 [17]: visualizing the frequency of detected bacterial

species in a given sample.

Statistical analysis

Principal coordinate analysis was performed and the results were visualized with R

version 4.0.2.

Results

Characteristics of participants

Our study consisted of 18 Japanese, ranging from 30 to 43 years of age, with a median

age of 36.5. All participants had primary infertility with normal menstrual cycles, along

with no underlying medical conditions (Table 1).

Identification of vaginal bacteria

The composition of the vaginal microbiota was investigated by 16S rRNA gene

amplicon sequencing on the MinION™ platform (Figure 1). The V1–9 MinION™

sequencing for almost 90 min yielded 15,836–119,745 reads at first and then filtered

10,688–103,212 reads satisfying the conditions of 1,300–1,950 bp and QC > 7 (details

in Table 2). The reads were mapped against the GenomeSync reference database for

taxonomic assignment. The results of Nugent scoring and microbiota analysis of the

filtered 3,000 reads using MinION™ are shown in Table S2. Based on the Nugent score

criteria, four out of 18 cases were in the intermediate group, but none of them were

diagnosed as BV. The MinION™ sequencing data were also analyzed by an alternative

bioinformatics tool. The FASTQ 16S workflow produced almost similar taxonomic

profiles (Table S3). This result confirmed the validity of our method for taxonomic

classification and the bacterial compositions were comparable regardless of the program

and database used.

Comparison of the results of swab and lavage samples

For subjects 1–4, lavage and swab samples were collected and bacterial DNA extraction

and 16S rRNA analysis were performed to determine the effect of the sample collection

method on the results of the microbiota analysis. After filtering, 3,000 reads were

randomly extracted in eight samples (1-4_lavage and 1-4_swab), the results of which

are shown in Figure 2, with almost the same identification with the two sampling

methods. We performed principal coordinate analysis (PCoA) to assess the equivalence

of the lavage and swab sampling methods, the results of which are shown in Figure 3.

Regardless of the sample collection method, the results of the analysis of the same

subjects involved almost the same coordinates.

Comparison of the results of sequence read number

To assess the effect of read count on the results of the MinION™ sequencing, three

samples that showed characteristic bacterial microbiota [5_lavage: *Lactobacillus iners*

dominant (>99%), 6_lavage: *Lactobacillus crispatus* dominant (>99%), 8_lavage: BV

like] were selected and their taxonomic profiles were compared after filtering for 3,000

reads (random sampling) and 10,000 reads (random sampling) (Figure 4). In all

specimens, the results of vaginal microbiota analysis in MinION™ were similar for

different numbers of reads.

Comparison of the results of V1–9 and V3–4 16S rRNA gene sequencing

For the taxonomic classification of bacteria, we compared the resolution of long-read

(V1–9) and short-read (V3–4) 16S amplicon sequencing for six samples (sample

numbers 5, 6, 9, 10, 11, and 14). Of these, for sample numbers 5, 9, and 10,

Lactobacillus iners was confirmed to be present at a rate of more than 99% (Group I),

while for sample numbers 6, 11, and 14, *Lactobacillus crispatus* was confirmed to be

present at more than 99% (Group C) in the V1–9 16S rRNA analysis. The V3–4 region

was amplified by four-primer PCR from the bacterial DNA of six samples and

sequenced with MinION™. From each sequencing result, 3,000 reads were randomly

extracted and the results of the analysis of the V1–9 region were compared with those of

the V3–4 region (Figure 5). In Group I, *Lactobacillus iners* was classified as being

present at a rate of more than 98% in the V3–4 S16 rRNA analysis, which was

comparable to the results obtained by the V1-9 sequencing (Figure 5A). In Group C, the

V3-4 sequence alignment resulted in ambiguous identification of *Lactobacillus* species

(Figure 5B). All samples showed *Lactobacillus crispatus*, *Lactobacillus acidophilus*,

Lactobacillus helveticus, *Lactobacillus amylovorus*, *Lactobacillus kefiranofaciens*,

Lactobacillus hamsteri, *Lactobacillus gallinarum*, *Lactobacillus kalixensi*s, and

Lactobacillus psittaci in the V3–4 S16 rRNA analysis as potentially existing species,

which were found in similar proportions across all samples in Group C.

Discussion

The vaginal microbiota is characterized by particularly low diversity compared with the

human gut harboring a complex microbial community. [18]. Conventional methods,

such as Amsel's criteria, the Nugent score, and bacterial culture, are not suitable for an

accurate bacterial identification. Next-generation sequencing has revolutionized the

profiling of the bacterial microbiota. In short, the combination of metagenomic

sequencing and bioinformatic technology has made it possible to more accurately assess

the bacterial microbiota without the influence of uncertainties such as examiner

subjectivity or the capturability of each bacterium [6]. The MinION™ platform allows a

single flow cell to be used repeatedly by assigning individual barcodes. MinION™

allows data generation by immediately reading the amplicon sequences that pass

through the nanopores placed on the substrate of the flow cell. The real-time data

generation allows the sequencing to end when the data reach the target amount.

In particular, analysis of the vaginal microbiota is a field in which MinION™'s

capabilities can be maximized because of the low complexity of this microbiota under

healthy conditions. However, to date, no studies have examined clinical vaginal

metagenomic samples with MinION™ technology. In the present study, the number of

bacterial species detected with MinION™ ranged from 1 to 22 (mean value 3.3),

confirming the low diversity of the vaginal microflora (Table S2). Although no cases

were diagnosed as BV by the Nugent scoring method, the 16S rRNA gene analysis

using MinION™ revealed the presence of BV-causing bacteria such as *Streptococcus*,

Atopobium, *Prevotella*, and *Gardnerella* [7] in the group categorized as intermediate

based on Nugent's criteria. For the most part, the 16S rRNA gene analysis appeared to

correlate with the Nugent scoring to diagnose BV. However, there was an inconsistency

between the two methods in one case (14_lavage), where it was categorized as

intermediate, even though only *Lactobacillus* was identified by the MinION™

sequencing. While Nugent scoring is a subjective evaluation and depends on the

acumen of the examiner, the molecular method targeting the 16S rRNA gene could

serve as a practical and reliable measure for the diagnosis of BV. In vaginal microbiota

analysis, it is important to be able to profile *Lactobacillus* spp. and the causative

organisms of BV at the species level. The Nugent score is calculated based on the

number of bacterial morphotypes and therefore has limitations for discriminating

bacterial taxa at the species level. Meanwhile, the sequence-based assessment of the

vaginal microbiota can allow us to obtain an accurate representation of the bacterial

composition. It has been reported that species-specific characteristics of *Lactobacillus*

spp. are critical for the stability and maintenance of human vaginal microbiota [19,20].

In particular, *Lactobacillus iners*-dominated community has been more associated with

vaginal dysbiosis, suggesting its relevance in the pathophysiology of gynecological

diseases.

Previous studies on short-read sequencing have reported that the sample collection

method had no effect on the results [21]. We collected swab and lavage samples from

the same cases and determined the species composition of vaginal communities. (Figure

2). When a single bacterial taxon comprised 98%–99% of the total bacteria (sample

numbers 1 and 4), there was no change in the composition ratio depending on the

sample collection method. When the vaginal microbiota was composed of more than

two bacteria (sample numbers 2 and 3), the bacterial species comprising the microbiota

were found to be the same, although there were slight differences in the proportions

present. Thus, we concluded that the effect of the sample collection method was

negligible. In contrast to the characterization of highly complex bacterial communities

in the human gut [12], the analysis of vaginal microbiota of lower complexity would

require a much smaller number of reads. A calculation using the median filtered reads

predicted that 3,000 reads would be accumulated in a sequence of approximately 8 min,

indicating that the time required for vaginal microbiota analysis using MinION™ could

be further reduced.

The conventional parallel-type short-read sequencer cannot yield reads covering the full

length of the 16S rRNA gene, and the partial 16S rRNA gene sequencing has been

insufficient to identify bacterial taxa at the species level. Benchmarking with the

conventional sequencing method (Illumina MiSeq™ technology) demonstrated that the

full-length 16S gene sequencing by MinION™ gives a better resolution for bacterial

identification [12]. To clarify the usefulness of long-read sequencing in vaginal

microbiota testing, we picked up *Lactobacillus iners*-dominant (> 99% in V1–9

analysis) samples (Group I: sample numbers 5, 9, and 10) and *Lactobacillus crispatus*-

dominant (>99% in V1–9 analysis) samples (Group C: sample numbers 6, 11, and 14)

and compared the results of 16S rRNA analyses of the V1–9 and V3–4 regions (Figure

5). The bioinformatic pipeline used a method where each amplicon sequence was

directly compared with the database, scored for similarity (Minimap2 score) according

to a proprietary algorithm, and then applied to the bacterial species with the highest

score. Hence, if there were multiple species with the same score, we treated the reads as

potentially existing in all of the possible species. The reads assigned to "Others" in this

study could be (1) bacteria with low abundance, or (2) bacteria that could not be

assigned to a single species due to competition between multiple species. In Group I

(Figure 5A), *Lactobacillus iners* was well recognized regardless of the choice of the

regions to be sequenced, but the numbers of reads assigned to "Others" increased in the

V3–4 analysis. In Group C (Figure 5B), V3–4 analysis showed that *Lactobacillus*

crispatus accounted for about 50% of the reads and *Lactobacillus acidophilus*,

Lactobacillus helveticus, *Lactobacillus amylovorus*, *Lactobacillus kefiranofaciens*,

Lactobacillus hamsteri, and *Lactobacillus gallinarum* were classified in similar

proportions in all three specimens, which had not been found in V1–9 analysis. Due to

the lower discriminatory power of the V3-4 region, the considerable number of V3-4

reads could not be assigned to one species and evenly allocated to multiple taxonomic

bins represented as potentially existing species. The numbers of reads assigned to

"Others" were greater in the V3-4 data set as compared to the V1-9 data set. This

suggested that assigning bacteria using the V1–9 region is more informative for

determining the appropriate bacterial species than using only the V3–4 region. Similar

to our previous report [12], the long-read sequencing with MinION™ showed high

identification accuracy compared with the short-read sequencing under the same

bioinformatic pipeline.

In terms of the usefulness of the MinION™ sequencer in gynecological practice, it is

interesting that we were able to show the detailed condition of vaginal dysbiosis in a

short period of time, which was not noted by the Nugent score. It has been suggested

that vaginal dysbiosis may have a connection with the incidence of sexually transmitted

infections [22] and genital viral infections [23,24] or cervical lesions [25]. The full-

length 16S rRNA analysis using MinION™ could be a promising option for bacterial

identification in a reasonable time frame for diagnostic purposes.

Having used MinION™ in our vaginal microbiota analysis workflow, we have seen four

key benefits. 1) Convenience: The unit is extremely light, small, and portable, and

sequencing can occur via a USB connection to an in-clinic PC. 2) Speed: The average

time between specimen collection and result disclosure is 2 days. 3) Economy: Analysis

of a small number of samples can be performed without increasing the unit cost of the

test, by using individual barcodes. 4) Functionality: Although MinION™ has been

reported to have a slightly higher error rate, the technical issues are being resolved and

even the intestinal microbiota can now be identified [12]. In particular, the analysis of

vaginal microbiota characterized by low diversity, as in the present study, allowed for

comparable species-level identification [26]. Although this study focused on the vaginal

microbiota, a similar workflow could be applied to many clinical areas in the future, and

the benefits of MinION™ could make it easier for clinicians to successfully perform

bacterial metagenome analysis. In addition, future large-scale microbiota studies could

lead to new clinical findings.

Conclusions

In conclusion, our validation shows that the MinION™ long-read sequencer provides a

low-cost, rapid workflow for identifying vaginal microbiota with higher resolution in a

clinical setting. Detecting vaginal microbiota at the species level has the potential to

identify risks to women's health, as well as facilitating large-scale clinical studies in any

medical field.

Data accessibility

Sequence data from this article have been deposited in the DDBJ DRA database

(www.ddbj.nig.ac.jp/dra/index-e.html) under accession numbers DRR244979–

DRR245006.

Author contributions

Conceptualization, SK, YMa, YMo, HO, and KH; methodology, SK, YMa, SN, KK, and

KH; software, SN and KK; validation, SK, YMa, and KH; investigation, SK and YMa;

data curation, SK and YMa; writing—original draft preparation, SK; writing—review

and editing, YMa, SN, YMo, KK, HO, and KH; visualization, SK and YMa;

supervision, YMa and KH; funding acquisition, YMa, KK, and KH. All authors have

read and agreed to the published version of the manuscript.

Funding

This research was funded by Japan Society for the Promotion of Science KAKENHI

Grant Numbers JP19K09339 (to YMa), JP17H07123 (to KK), JP20K06612 (to KK),

and the Branding Program as a World-leading Research University on Intractable

Immune and Allergic Diseases supported by the Ministry of Education, Culture, Sports,

Science and Technology of Japan.

Acknowledgments

We are grateful to Tadashi Imanishi (Tokai University School of Medicine) for his

helpful comments. We would like to thank Edanz (https://en-author-

services.edanzgroup.com/ac) for editing the English text of a draft of this manuscript.

Conflicts of interest

The authors declare no conflict of interest.

Supplementary information

Supplementary Table S1: Representative bacterial genomes stored in the GenomeSync database.

Supplementary Table S2: Details of identified species with V1-9 16S rRNA analysis using MinION™ by minimap2 best hit score (>0.5%).

Supplementary Table S3: Datils of identified taxa with V1-9 16S rRNA analysis using MinION by FASTQ 16S workflow.

References

Thomason JL, Gelbart SM, Scaglione NJ (1991) Bacterial vaginosis: current review with indications for asymptomatic therapy. *Am J Obstet Gynecol* **165**: 1210-1217, doi: 10.1016/s0002-9378(12)90729-2.

Nardis C, Mosca L, Mastromarino P (2013) Vaginal microbiota and viral sexually transmitted diseases. *Ann Ig* **25**: 443-456, doi: 10.7416/ai.2013.1946.

Larsson PG, Bergstrom M, Forsum U, Jacobsson B, Strand A, Wolner-Hanssen P (2005) Bacterial vaginosis. Transmission, role in genital tract infection and pregnancy outcome: an enigma. *APMIS* **113**: 233-245, doi: 10.1111/j.1600-0463.2005.apm_01.x.

Amsel R, Totten PA, Spiegel CA, Chen KC, Eschenbach D, Holmes KK (1983) Nonspecific vaginitis. Diagnostic criteria and microbial and epidemiologic associations. *Am J Med* **74**: 14-22, doi: 10.1016/0002-9343(83)91112-9.

Nugent RP, Krohn MA, Hillier SL (1991) Reliability of diagnosing bacterial vaginosis is improved by a standardized method of gram stain interpretation. *J Clin Microbiol* **29**: 297-301, doi: 10.1128/JCM.29.2.297-301.1991.

van den Munckhof EHA, van Sitter RL, Boers KE, Lamont RF, te Witt R, Cessie S, Knetsch CW, van Doorn LJ, Quint WGV, Molijn A, van Hall MAL (2019) Comparison of Amsel criteria, Nugent score, culture and two CE-IVD marked quantitative real-time PCRs with microbiota analysis for the diagnosis of bacterial vaginosis. *Eur J Clin Microbiol Infect Dis* **38**: 959-966, doi: 10.1007/s10096-019-03538-7.

Ravel J, Gajer P, Abdo Z, Schneider GM, Koenig SS, McCulle SL, Karlebach S, Gorle R, Russell J, Tacket CO, Brotman RM, Davis CC, Ault K, Peralta L, Forney LJ (2011) Vaginal microbiome of reproductive-age women. *PNAS* **108 Suppl 1**: 4680-4687, doi: 10.1073/pnas.1002611107.

Mikheyev AS, Tin MM (2014) A first look at the Oxford Nanopore MinION

464 sequencer. *Mol Ecol Resour* **14**: 1097-1102, doi: 10.1111/1755-0998.12324.

465 9 Mitsuhashi S, Kryukov K, Nakagawa S, Takeuchi JS, Shiraishi Y, Asano K,

466  Imanishi T (2017) A portable system for rapid bacterial composition analysis using a

467  nanopore-based sequencer and laptop computer. *Sci Rep* **7**: 5657, doi:

468  10.1038/s41598-017-05772-5.

469 10 Kai S, Matsuo Y, Nakagawa S, Kryukov K, Matsukawa S, Tanaka H, Iwai T,

470  Imanishi T, Hirota K (2019) Rapid bacterial identification by direct PCR

471  amplificatiaon of 16S rRNA genes using the MinION™ nanopore sequencer. *FEBS*

472  *Open Bio* **9**: 548-557, doi: 10.1002/2211-5463.12590.

473 11 Nakagawa S, Inoue S, Kryukov K, Yamagishi J, Ohno A, Hayashida K, Nakazwe R,

474  Kalumbi M, Mwenya D, Asami N, Sugimoto C, Mutengo MM, Imanishi T (2019)

475  Rapid sequencing-based diagnosis of infectious bacterial species from meningitis

476  patients in Zambia. *Clin Transl Immunology* **8**, e01087, doi: 10.1002/cti2.1087.

477 12 Matsuo Y, Komiya S, Yasumizu Y, Yasuoka Y, Mizushima K, Takagi T, Krykov K,

478  Fukuda A, Morimoto Y, Naito Y, Okada H, Bono H, Nakagawa S, Hirota K (2021)

479  Full-length 16S rRNA gene amplicon analysis of human gut microbiota using

480  MinION nanopore sequencing confers species-level resolution. *BMC Microbiol* **21**:

481  35, possidoi: 10.1186/s12866-021-02094-5.

482 13 Shen W, Le S, Li Y, Hu F (2016) SeqKit: A Cross-Platform and Ultrafast Toolkit for

483  FASTA/Q File Manipulation. *PLoS One* **11**: e0163962, doi:

484  10.1371/journal.pone.0163962.

485 14 Quast C, Pruesse E, Yilmaz P, Gerken J, Schweer T, Yarza P, Peplies J, Glöckner FO

486  (2013) The SILVA ribosomal RNA gene database project: improved data processing

487  and web-based tools. *Nucl Acids Res* **41**: D590-D596, doi: 10.1093/nar/gks1219.

488 15 Frith MC (2011) A new repeat-masking method enables specific detection of

489  homologous sequences. *Nucleic Acids Res* **39**: e23, doi: 10.1093/nar/gkq1212.

490 16 Li H (2018) Minimap2: pairwise alignment for nucleotide sequences.

Bioinformatics **34**: 3094-3100, doi: 10.1093/bioinformatics/bty191.

17 Ondov BD, Bergman NH, Phillippy AM (2011) Interactive metagenomic

visualization in a Web browser. *BMC Bioinformatics* **12**: 385, doi: 10.1186/1471-

2105-12-385.

18 Human Microbiome Project Consortium (2012) Structure, function and diversity of

the healthy human microbiome. *Nature* **486**: 207-214, doi: 10.1038/nature11234.

19 Zheng N, Guo R, Yao Y, Jin M, Cheng Y, Ling Z (2019) Lactobacillus iners Is

Associated with Vaginal Dysbiosis in Healthy Pregnant Women: A Preliminary

Study. *Biomed Res Int* **article ID 6070734**, doi: 10.1155/2019/6079734

20 Petrova MI, Reid G, Vaneechoutte M, Lebeer S (2017) Lactobacillus iners: Friend

or Foe? *Trends Microbiol* **25**: 182-191, doi: 10.1016/j.tim.2016.11.007

21 Virtanen S, Kalliala I, Nieminen P, Salonen A (2017) Comparative analysis of

vaginal microbiota sampling using 16S rRNA gene analysis. *PLoS One* **12**:

e0181477, doi: 10.1371/journal.pone.0181477.

22 Shipitsyna E, Khusnutdinova T, Budilovskaya O, Krysanova A, Shalepo K,

Savicheva A, Unemo M (2020) Bacterial vaginosis-associated vaginal microbiota is

an age-independent risk factor for Chlamydia trachomatis, Mycoplasma genitalium

and Trichomonas vaginalis infections in low-risk women, St. Petersburg, Russia.

Eur J Clin Microbiol Infect Dis **39**: 1221-1230, doi: 10.1007/s10096-020-03831-w.

23 Borgogna JC, Shardell MD, Santori EK, Nelson TM, Rath JM, Glover ED, Ravel J,

Gravitt PE, Yeoman CJ, Brotman RM (2020) The vaginal metabolome and

microbiota of cervical HPV-positive and HPV-negative women: a cross-sectional

analysis. *BJOG* **127**: 182-192, doi: 10.1111/1471-0528.15981.

24 Hoang T, Toler E, DeLong K, Mafunda NA, Bloom SM, Zierden HC, Moench TR,

Coleman JS, Hanes J, Kwon DS, Lai SK, Cone RA, Ensign LM (2020) The

cervicovaginal mucus barrier to HIV-1 is diminished in bacterial vaginosis. *PLoS*

Pathog **16**: e1008236, doi: 10.1371/journal.ppat.1008236.

25 Dahoud W, Michael CW, Gokozan H, Nakanishi AK, Harbhajanka A (2019)

Association of Bacterial Vaginosis and Human Papilloma Virus Infection with

Cervical Squamous Intraepithelial Lesions. *Am J Clin Pathol* **152**: 185-189, doi:

10.1093/ajcp/aqz021.

26 Kilianski A, Haas JL, Corriveau EJ, Liem AT, Willis KL, Kadavy DR, Rosenzweig

CN, Minot SS (2015) Bacterial and viral identification and differentiation by

amplicon sequencing on the MinION nanopore sequencer. *Gigascience* **4**: 12, doi:

10.1186/s13742-015-0051-z.

Table 1. Participants' background.

Number	Age (y.o.)	AMH (ng/mL)	LH (mIU/mL)	FSH (mIU/mL)	BMI (kg/m^2)
1	36	0.47	7.2	6.8	18.8
2	36	1.15	3.7	5.0	25.6
3	42	0.23	8.3	11.3	21.0
4	30	9.16	6.8	7.2	26.0
5	43	1.67	2.8	5.5	20.8
6	33	3.11	3.2	5.5	25.0
7	35	5.89	5.1	7.9	20.9
8	35	4.98	2.6	4.9	18.8
9	43	2.26	1.8	6.0	19.4
10	39	0.59	6.5	6.3	22.4
11	37	2.34	8.9	6.0	18.6
12	34	4.52	5.2	5.8	20.3
13	35	2.62	8.8	8.2	23.0
14	37	3.16	2.0	3.8	22.1
15	40	3.69	8.2	7.6	20.4
16	38	2.27	6.4	8.4	18.1
17	35	3.66	6.9	7.8	20.4
18	38	0.44	3.5	6.0	19.1

528 AMH: anti-Mullerian hormone, LH: luteinizing hormone, FSH: follicle-stimulating

529 hormone, BMI: body mass index.

531   **Table 2.** Statistics of MinION™ V1–9 sequencing data.

Sample	Pass reads				Filtered reads	
	No. of reads	Min (bp)	Avg (bp)	Max (bp)	No. of reads	Avg (bp)
1_lavage_V1-9	115,957	128	1,434.3	4,377	91,449 (78.9%)	1,633.5
1_swab_V1-9	63,547	155	1,078.0	4,193	31,462 (49.5%)	1,619.3
2_lavage_V1-9	33,721	158	1,507.2	4,486	28,718 (85.2%)	1,607.0
2_swab_V1-9	53,584	145	1,203.6	3,746	33,294 (62.1%)	1,613.6
3_lavage_V1-9	28,071	154	1,396.3	4,875	20,958 (74.7%)	1,613.8
3_swab_V1-9	23,053	162	1,521.0	3,990	20,073 (87.1%)	1,614.5
4_lavage_V1-9	30,603	155	1,502.9	4,977	25,756 (84.2%)	1,639.5
4_swab_V1-9	37,597	161	1,103.2	3,383	19,588 (52.1%)	1,644.9
5_lavage_V1-9	15,836	147	1,346.1	3,393	10,688 (67.5%)	1,632.8
6_lavage_V1-9	38,020	144	1,517.2	3,787	32,255 (84.8%)	1,642.4
7_lavage_V1-9	61,432	158	1,218.8	5,035	37,018 (60.3%)	1,631.4
8_lavage_V1-9	52,071	111	1,257.5	3,344	33,906 (65.1%)	1,603.2
9_lavage_V1-9	119,745	125	1,530.5	4,664	103,212 (86.2%)	1,649.4
10_lavage_V1-9	102,333	150	1,496.5	4,980	84,028 (82.1%)	1,650.2
11_lavage_V1-9	61,756	121	1,541.3	3,519	54,116 (87.6%)	1,647.8
12_lavage_V1-9	101,755	120	1,433.9	4,759	77,756 (76.4%)	1,635.8
13_lavage_V1-9	71,041	134	1,437.5	4,828	56,601 (79.7%)	1,599.2
14_lavage_V1-9	42,629	159	1,367.7	4,981	31,447 (73.8%)	1,629.0
15_lavage_V1-9	52,908	172	1,452.9	3,333	42,910 (81.1%)	1,630.4
16_lavage_V1-9	52,230	153	1,480.3	4,345	43,549 (83.4%)	1,633.0
17_lavage_V1-9	38,219	154	1,478.3	3,453	31,757 (83.1%)	1,625.3
18_lavage_V1-9	28,886	185	1,464.7	3,328	23,816 (82.4%)	1,631.8

Min: minimum read length, Avg: average read length, Max: maximum read length.

Figure legends

Figure 1. Workflow of 16S rRNA amplicon sequencing with the MinION™ platform

and bioinformatic analysis. After collecting the vaginal sample, it should be stored in a

freezer at −30°C until DNA extraction begins. Some swab collection kits should be

stored at room temperature, in accordance with the manufacturer's instructions.

Sequencing libraries are generated by a four-primer PCR-based strategy; in the initial

stages of PCR, the 16S rRNA gene is amplified with an inner primer pair. The PCR

product is amplified with the outer primers and targeted to introduce identical barcode

and tag sequences at both ends, allowing for the attachment of adapter molecules in a

one-step reaction. The library is then loaded into a MinION™ connected to a personal

computer. With our experimental method, sequencing runtime of 10 min is sufficient for

3,000 reads to be obtained. A final report of the microbiota can be presented within 2

days of initiating the analysis.

Figure 2. Taxonomic profiles comparing the different sampling method (lavage and

swab) results of V1–9 16S rRNA MinION™ sequencing with 3,000 randomly sampled

reads after filtration. In the analytical algorithm, we assigned the bacterial name that

showed the highest Minimap2 score for each read; bacteria with an assignment of less

than 1% were included in "Others."

Figure 3. PCoA of vaginal microbiota by different sampling techniques. The

coordinates of swab samples and lavage samples were nearly equivalent and did not

show any effect of the vaginal specimen collection method in the 3,000 filtered reads of

V1–9 MinION™ sequencing.

Figure 4. Taxonomic profiles comparing the results obtained for 3,000 and 10,000

filtered reads of V1–9 MinION™ sequencing. In the analytical algorithm, we assigned

the bacterial name that showed the highest Minimap2 score for each read; bacteria with

an assignment of less than 1% were included in "Others."

Figure 5. Taxonomic profiles comparing the results of V3–4 and V1–9 16S rRNA

sequences using MinION™ in 3,000 filtered reads. (A) In three cases, *Lactobacillus*

iners was shown to constitute more than 99% of the bacteria by MinION™ analysis

with the V1–9 region (Group I). (B) In three cases, *Lactobacillus crispatus* was shown

to constitute more than 99% of the bacteria by MinION™ analysis with the V1–9 region

(Group C).

Figure 1

Duration

Sample collection
          and -30°C freezing until DNA extraction   -

DNA extraction using QIAamp UCP Pathogen Mini Kit   90 min.

V1-9 16S rRNA gene PCR with 4-primer method
PCR: KAPA2G™ Robust HotStart ReadyMix PCR Kit   60 min.
Barcoding: PCR Barcording Kit SQK-PBK004

MinION™ sequencing
                                                           10 min.
Live basecalling with MinKNOW software during the run

Read mapping and annotation using the GSTK software
Extraction of non-human reads   approximately
                                                       1 hour per sample
Sequence search against the GenomeSync database

Output as a report visualized with Krona chart sofrware

          The whole process will be finished in two days.

Figure 2

Lactobacillus iners
Lactobacillus jensenii
Others

Streptococcus agalactiae
Streptococcus dysagalactiae
Others

Lactobacillus gasseri
Atopobium vaginae
Staphylococcus epidermidis
Streptococcus anginosus
Others

Lactobacillus crispatus
Others

Figure 3

Figure 4

■	*Lactobacillus iners*
■	*Others*

■	*Streptococcus agalactiae*
■	*Others*

■	*Streptococcus anginosus*
■	*Streptococcus mitis*
■	*Prevotella oris*
■	*Anaerococcus vaginalis*
■	*Prevotella fusca*
■	*Prevotella copri*

■ *Peptoniphilus harei*	■ *Streptococcus pneumoniae*	■ *Peptoniphilus coxii*
■ *Peptoniphilus timonensis*	■ *Ezakiella massiliensis*	■ *Prevotella histicola*
■ *Prevotella intermedia*	■ *Prevotella scopos*	■ *Others*

Figure 5

A

5_lavage_V1-9 5_lavage_V3-4

9_lavage_V1-9 9_lavage_V3-4

10_lavage_V1-9 10_lavage_V3-4

■ *Lactobacillus iners*
■ *Others*

B

6_lavage_V1-9 6_lavage_V3-4

11_lavage_V1-9 11_lavage_V3-4

14_lavage_V1-9 14_lavage_V3-4

■ *Lactobacillus crispatus*
■ *Lactobacillus acidophilus*
■ *Lactobacillus helveticus*
■ *Lactobacillus amylovorus*
■ *Lactobacillus kefiranofaciens*
■ *Lactobacillus hamsteri*
■ *Lactobacillus gallinarum*
■ *Lactobacillus kalixensis*
■ *Lactobacillus psittaci*
■ *Others*