MinION™, a portable long-read sequencer, enables rapid vaginal microbiota

analysis in a clinical setting

Running title: MinION™ enables rapid vaginal microbiota analysis

Shinnosuke Komiya[1,2], Yoshiyuki Matsuo[3,*], So Nakagawa[4], Yoshiharu Morimoto[2],

Kirill Kryukov[5], Tadashi Imanishi[4], Hidetaka Okada[1], and Kiichi Hirota[3]

[1]Department of Obstetrics and Gynecology, Kansai Medical University Graduate

School of Medicine, Osaka, Japan

[2]HORAC Grand Front Osaka Clinic, Osaka, Japan

[3]Department of Human Stress Response Science, Institute of Biomedical Science,

Kansai Medical University, Osaka, Japan

[4]Department of Molecular Life Science, Tokai University School of Medicine,

Kanagawa, Japan

[5]Department of Genomics and Evolutionary Biology, National Institute of Genetics,

Shizuoka, Japan

*Correspondence:

Yoshiyuki Matsuo

ysmatsuo-kyt@umin.ac.jp

Department of Human Stress Response Science, Institute of Biomedical Science,

Kansai Medical University, 2-5-1 Shin-machi, Hirakata, Osaka 573-1010, Japan

Abstract

It has been suggested that the local microbiota in the reproductive organs is relevant to women's health and may also affect pregnancy outcomes. Analysis of partial 16S ribosomal RNA (rRNA) gene sequences generated by short-read sequencers has been used to identify vaginal and endometrial microbiota, but it requires a long time to obtain the results, making it unsuitable for the rapid analysis of small samples in a clinical context. We demonstrated a simple workflow using the nanopore sequencer MinION™ that allows high-resolution and rapid differentiation of vaginal microbiota. Vaginal samples collected from 18 participants were subjected to DNA extraction and full-length 16S rRNA gene sequencing with MinION™. The principal coordinate analysis showed no differences in the bacterial compositions regardless of the sample collection method. The vaginal microbiota results could be reported within 2 days of specimen receipt. Although bacterial vaginosis (BV) was not diagnosed by the Nugent score in any cases, groups with both healthy and BV-like vaginal microbiota were clearly characterized by MinION™ sequencing. We conclude that full-length 16S rRNA gene sequencing analysis with MinION™ provides a rapid means for identifying vaginal bacteria with higher resolution. Species-level profiling of human vaginal microbiota by MinION™ sequencing can allow the analysis of associations with conditions such as genital infections, endometritis, and threatened miscarriage.

Keywords

16S rRNA, bacterial vaginosis, next-generation sequencer, long-read sequencer, MinION™

Introduction

Bacterial vaginosis (BV) is one of the most common gynecological disorders among women of reproductive age. It is caused by disruption of the vaginal microbiota, and has relatively mild subjective symptoms, such as abnormal vaginal discharge (Thomason, et al., 1991). BV has been clinically reported to cause adverse reproductive outcomes, including sexually transmitted infections, genitourinary viral infections such as HIV-1/2, HPV, and HSV-2 (Nardis, et al., 2013), and preterm delivery (Larsson, et al., 2005). Currently, Amsel's criteria based on clinical outcomes (Amsel, et al., 1983) and the Nugent score using Gram stain findings (Nugent, et al., 1991) are used as diagnostic criteria for BV. These methods are easy to use, but since they are based on the subjective opinion of the physician, the diagnosis largely depends on the skill of the examiner (van den Munckhof, et al., 2019). Bacterial culture of vaginal discharge is also a standard method to investigate the vaginal microbiota, but some bacteria are difficult to culture (van den Munckhof, et al., 2019). Moreover, in routine gynecological practice, vaginal discharge is not sampled unless patients have symptoms suggestive of vaginosis. Therefore, accurately assessing the vaginal microbiota has been difficult with these regular clinical methods.

16S ribosomal RNA (rRNA) gene sequencing analysis using next-generation sequencers can be used to identify bacteria without culture in a manner not dependent on subjective assessment by the examiner. Previous studies have shown that the vaginal microbiota of healthy women is classified into community state type (CST) I to V (Ravel, et al., 2011). CST is defined by the dominant bacteria as follows: CST I: *Lactobacillus crispatus*, CST II: *Lactobacillus gasseri*, CST Ⅲ: *Lactobacillus iners*, CST Ⅳ: lack of *Lactobacillus* spp., and CST V: *Lactobacillus jensenni*. These narrow definitions mean that undiagnosed cases of vaginal dysbiosis may be present. This new classification of the vaginal microbiota has been clarified by performing sequencing analysis on cases with no clinical symptoms. Because a deeper and more accurate

understanding of the vaginal microbiota may provide new clinical insights, we should

demonstrate that 16S rRNA analysis is quick and easy to perform in a broad range of

contexts in clinical practice.

MinION™ (Oxford Nanopore Technologies Ltd., Oxford, UK) detects the signal of a

DNA nucleotide that passes through a nanopore arranged on a flow cell (Mikheyev and

Tin, 2014). This long-read sequencing technology can be used to identify bacterial

microbiota with the whole region of the 16S rRNA. MinION™ is smaller, lighter, and

easier to deploy than conventional short-read sequencers, and it also offers the

advantage of high-resolution microbiota analysis with long-read sequencing

(Mitsuhashi, et al., 2017) (Kai, et al., 2019) (Nakagawa, et al., 2019). We constructed an

in-hospital vaginal microbiota-analyzing workflow using MinION™. This is the first

report to demonstrate that MinION™ can provide rapid patient feedback regarding

vaginal microbiota analysis.

Materials and Methods

Ethics statement

This study is a cross-sectional observational trial approved by the institutional review

board of the medical corporation Sankeikai (Approval Number: 2019-38). After the

approval, the study protocol was submitted to the UMIN-CTR clinical trial registry site

(https://www.umin.ac.jp/ctr/index.htm) (trial registration number: UMIN000038032),

which met the criteria of the International Committee of Medical Journal Editors

(ICMJE).

Recruitment of participants

From April 2019 to May 2020, participants who met all of the following enrollment

criteria were recruited: 1) primary infertility patients attending HORAC Grand Front

Osaka Clinic in Osaka, Japan; 2) premenopausal Japanese women who were not

pregnant; 3) no clinical symptoms strongly suggestive of BV; 4) scheduled hormone

replacement embryo transfer cycle with frozen-thawed blastocyst; and 5) those who

provided written consent. The exclusion criteria were as follows: 1) known history of

HIV-1/2, hepatitis B/C, syphilis, or genital chlamydia infection; 2) known history of

diabetes; 3) known history of cervical/vaginal surgery; 4) known history of intrauterine

device use; and 5) known history of antibiotic, steroid, or vaginal suppository use within

2 weeks. In addition to collecting vaginal samples, details of age, body mass index, and

basal hormone levels (anti-Mullerian hormone, luteinizing hormone, and follicle-

stimulating hormone on days 2–4 of the menstrual cycle), medical history, and

infertility treatment history were collected from the medical records. All vaginal

samples were collected on the day of embryo transfer.

Nugent score

Vaginal smears were analyzed by pathologists affiliated with an external laboratory, and

the Nugent score was calculated based on microscopic findings: the numbers of

Lactobacillus (scored as 0 to 4), *Gardnerella* (scored as 0 to 4), and *Mobiluncus* (scored

as 0 to 2). A total score of 0 to 3 was diagnosed as representative of healthy vaginal

microbiota, 4 to 6 as an intermediate group, and 7 or more as BV (Nugent, et al., 1991).

Inappropriate samples with defective specimen collection were diagnosed as

indeterminate. If BV was diagnosed based on the Nugent score, our policy was to treat

it using antibiotics.

Vaginal sample collection method

Vaginal lavage samples were collected from 18 participants; from four participants,

additional swab samples were collected with written consent. When both lavage and

swab samples were collected, the swab samples were collected first. For the lavage

method, the inside of the vagina was washed with 10 ml of sterile saline, and at least 2

ml was collected using a sterile syringe. Collected lavage samples were stored at −30 °C

until DNA extraction. For the swab method, vaginal mucus from the uterovaginal

region was collected using a packaged OMNIgene vaginal kit (OMR-130; DNA

Genotek Inc., Ottawa, Canada). Swab samples were stored at room temperature until

DNA extraction, in accordance with the manufacturer's instructions.

Sequencing sample preparation

DNA was extracted from 22 human vaginal samples using the QIAamp UCP Pathogen

Mini Kit (QIAGEN, Venlo, Netherlands) by the bead beating and spin protocol, in

accordance with the manufacturer's instructions. After extraction, the DNA content was

measured using a NanoDrop® 1000 Spectrophotometer (Thermo Fisher Scientific, MA,

USA), and the concentrations of extracted DNA from the swab and lavage samples were

6.1–32.3 ng/µl and 53.1–1092.4 ng/µl, respectively.

MinION™ sequencing

As described in our previous report (Matsuo, et al., 2020), a slightly modified version of

four-primer polymerase chain reaction (PCR) with rapid adapter attachment was

performed. For amplification of the V1-9 region of the 16S rRNA gene, a forward

primer (S-D-Bact-0008-c-S-20) with the anchor sequence 5′-

TTTCTGTTGGTGCTGATATTGCAGRGTTYGATYMTGGCTCAG-3′ and a reverse

primer with the anchor sequence 5′-

ACTTGCCTGTCGCTCTATCTTCCGGYTACCTTGTTACGACTT-3′ were used as

inner primers. For amplification of the V3–4 region, 341F with the anchor sequence 5′-

TTTCTGTTGGTGCTGATATTGCCCTACGGGNGGCWGCAG-3′ and 806R with the

anchor sequence 5′-

ACTTGCCTGTCGCTCTATCTTCGGACTACHVGGGTWTCTAAT-3′ were used as

inner primers. PCR amplification of the 16S rRNA gene was performed using the

KAPA2G™ Robust HotStart ReadyMix PCR Kit (Kapa Biosystems, MA, USA) with

an inner primer pair (50 nM for each) and the outer primer mixture (10 nM) of PCR

Barcoding Kit (SQK-PBK004; Oxford Nanopore Technologies, Oxford, UK) in a total

volume of 25 μl. Each sample was assigned an individual barcode, which was used to

identify the PCR product. Amplification was performed under the following PCR

conditions: initial denaturation at 95 °C for 3 min; five cycles of 15 s at 95 °C, 15 s at

55 °C, and 30 s at 72 °C; and 30 cycles of 15 s at 95 °C, 15 s at 62 °C, and 30 s at

72 °C; followed by a final extension at 72 °C for 1 min. The amplification products

were confirmed using 1% agarose gel electrophoresis (1 × TAE buffer) to be

approximately 1600 base pairs (V1–9 region) and 400 base pairs (V3–4 region), which

are the expected amplification sizes. After confirmation, the amplified DNA was

purified using AMPure® XP (Beckman Coulter, CA, USA). A total of 100 ng (10 μl) of

purified DNA was incubated with 1 μl of Rapid Adapter (Oxford Nanopore

Technologies Ltd.) at room temperature for 5 min. The concentrations of swab samples

and lavage samples were 8.3–34 ng/µl and 3.58–61 ng/µl, respectively, as determined

using a Quantus™ Fluorometer (Promega, WI, USA). Based on the manufacturer's

instructions, the prepared DNA library (total of 11 µl) was mixed with 34 µl of

Sequencing Buffer, 25.5 µl of Loading Beads, and 4.5 µl of water. The final adjusted

sample was loaded into the flow cell R.9.4.1 (FLO-MIN106; Oxford Nanopore

Technologies Ltd.) and attached to the MinION™ Mk1B, which was connected to a

personal computer. In this study, four to six samples were loaded at one time and each

sequencing session lasted about 90 min. DNA sequencing data were acquired using

MINKNOW software ver. 1.11.5 (Oxford Nanopore Technologies Ltd.), in accordance

with the manufacturer's instructions.

Bioinformatic Analysis Workflow

We performed the bioinformatic analysis using a pipeline of multiple programs built in

accordance with our previous reports (Mitsuhashi, et al., 2017) (Nakagawa, et al., 2019).

An overview is given below.

1) Computer: Apple iMac 27-inch, Late 2015 (OS, macOS 10.14.6; CPU, 3.3 GHz Intel

Core i5-6600; memory, 16 GB)

2) GUPPY software ver. 3.1.5 (Oxford Nanopore Technologies Ltd.): base calling,

conversion of FAST5 files to FASTQ files with quality control based on quality score >

7.

3) SeqKit software ver. 0.10.0 (Shen, et al., 2016): extraction of a read length of 1,300

to 1,950 bp, based on the 16S rRNA length registered in the SILVA rRNA database ver.

132 (https://www.arb-silva.de/).

4) TANTAN program ver. 18 (Frith, 2011): removing simple repeat sequences.

5) Minimap2 program ver. 2.14 (Li, 2018): eliminating human genome information

(Human Genome Assembly GRCh38) and matching each genome read to the 5,850

representative bacterial genome sequences (Table S1) stored in the GenomeSync

database (http://genomesync.org).

6) In-house Perl scripts: selecting the species with the highest Minimap2 score and

determining the taxa based on the NCBI taxonomy database.

7) Krona software version 2.7 (Ondov, et al., 2011): visualizing the frequency of

detected bacterial species in a given sample.

Statistical analysis

Principal coordinate analysis was performed and the results were visualized with R

version 4.0.2.

204
205

Results

Characteristics of participants

Our study consisted of 18 Japanese, ranging from 30 to 43 years of age, with a median

age of 36.5. All participants had primary infertility with normal menstrual cycles, along

with no underlying medical conditions (Table I).

Identification of vaginal bacteria

The composition of the vaginal microbiota was investigated by 16S rRNA gene

amplicon sequencing on the MinION™ platform (Fig. 1). The V1–9 MinION™

sequencing for almost 90 min yielded 15,836–119,745 reads at first and then filtered

10,688–103,212 reads satisfying the conditions of 1,300–1,950 bp and QC > 7 (details

in Table II). The results of microbiota analysis of the filtered 3,000 reads using

MinION™ are shown in Supplementary Table S2.

Comparison of the results of swab and lavage samples

For subjects 1–4, lavage and swab samples were collected and bacterial DNA extraction

and 16S rRNA analysis were performed to determine the effect of the sample collection

method on the results of the microbiota analysis. After filtering, 3,000 reads were

randomly extracted in eight samples (1-4_lavage and 1-4_swab), the results of which

are shown in Fig. 2, with almost the same identification with the two sampling methods.

We performed principal coordinate analysis (PCoA) to assess the equivalence of the

lavage and swab sampling methods, the results of which are shown in Fig. 3. Regardless

of the sample collection method, the results of the analysis of the same subjects

involved almost the same coordinates.

Comparison of the results of sequence read number

To assess the effect of read count on the results of the MinION™ sequencing, three

samples that showed characteristic bacterial microbiota [5_lavage: *Lactobacillus iners*

dominant (>99%), 6_lavage: *Lactobacillus crispatus* dominant (>99%), 8_lavage: BV

like] were selected and their taxonomic profiles were compared after filtering for 3,000

reads (random sampling) and 10,000 reads (random sampling) (Fig. 4). In all

specimens, the results of vaginal microbiota analysis in MinION™ were similar for

different numbers of reads.

Comparison of the results of V1–9 and V3–4 16S rRNA gene sequencing

For the taxonomic classification of bacteria, we compared the resolution of long-read

(V1–9) and short-read (V3–4) 16S amplicon sequencing for six samples (sample

numbers 5, 6, 9, 10, 11, and 14). Of these, for sample numbers 5, 9, and 10,

Lactobacillus iners was confirmed to be present at a rate of more than 99% (Group I),

while for sample numbers 6, 11, and 14, *Lactobacillus crispatus* was confirmed to be

present at more than 99% (Group C) in the V1–9 16S rRNA analysis.

The V3–4 region was amplified by four-primer PCR from the bacterial DNA of six

samples and sequenced with MinION™. From each sequencing result, 3,000 reads were

randomly extracted and the results of the analysis of the V1–9 region were compared

with those of the V3–4 region (Fig. 5). In Group I, *Lactobacillus iners* was classified as

being present at a rate of more than 98% in the V3–4 S16 rRNA analysis, which was

comparable to the results obtained by the V1-9 sequencing (Fig. 5a). In Group C, the

V3-4 sequence alignment resulted in ambiguous identification of *Lactobacillus* species

(Fig. 5b). All samples showed *Lactobacillus crispatus*, *Lactobacillus acidophilus*,

Lactobacillus helveticus, *Lactobacillus amylovorus*, *Lactobacillus kefiranofaciens*,

Lactobacillus hamsteri, *Lactobacillus gallinarum*, and *Lactobacillus kalixensis* in the

V3–4 S16 rRNA analysis as potentially existing species, which were found in similar

proportions across all samples in Group C.

Discussion

The vaginal microbiota is characterized by particularly low diversity compared with that

in other parts of the body (Human Microbiome Project, 2012). Conventional methods,

such as Amsel's criteria, the Nugent score, and bacterial culture, have enabled only

rough assessment of the vaginal microbiota. Next-generation sequencing (NGS) has

revolutionized the profiling of the bacterial microbiota. In short, the combination of

metagenomic sequencing and bioinformatic technology has made it possible to more

accurately assess the bacterial microbiota without the influence of uncertainties such as

examiner subjectivity or the capturability of each bacterium (van den Munckhof, et al.,

2019). The MinION™ platform allows a single flow cell to be used repeatedly by

assigning individual barcodes. MinION™ allows data generation by immediately

reading the amplicon sequences that pass through the nanopores placed on the substrate

of the flow cell. The real-time data generation allows the sequencing to end when the

data reach the target amount. In particular, analysis of the vaginal microbiota is a field

in which MinION™'s capabilities can be maximized because of the low complexity of

this microbiota under healthy conditions. However, to date, no studies have examined

clinical vaginal metagenomic samples with MinION™ technology. In vaginal

microbiota analysis, it is important to be able to profile *Lactobacillus* spp. and the

causative organisms of BV at the species level. Interestingly, no cases were diagnosed

with BV by the Nugent score in this study, but the results of vaginal microbiota analysis

using MinION™ to determine the profile by Minimap2 best-hit score identified not only

healthy vaginal microbiota but also cases with BV-like microbiota (Supplementary

Table S2).

In this study, 90-min MinION™ sequencing yielded 15,836 to 119,745 reads (Table II).

To estimate the shortest sequencing time for accurately assessing constituents of the

vaginal microbiota, the results of the analyses of 3,000, 10,000, and all reads were

compared in the typical cases of healthy vaginal microbiota and BV. The results were

not affected by increasing the number of reads to 3,000 or more (Fig. 4), so we

concluded that 3,000 reads were sufficient for vaginal microbiota analysis. A

calculation using the median filtered reads predicted that 3,000 reads would be

accumulated in a sequence of approximately 8 min, indicating that the time required for

vaginal microbiota analysis using MinION™ could be further reduced.

Previous studies on short-read sequencing have reported that the sample collection

method had no effect on the results (Virtanen, et al., 2017). Therefore, we collected

swab and lavage samples from the same cases to determine whether the effect of the

sample collection method was negligible in the case of long-read sequencing with our

bioinformatic pipeline in 3,000 randomly sampled reads (Fig. 2). When a single

bacterial taxon comprised 98%–99% of the total bacteria (sample numbers 1 and 4),

there was no change in the composition ratio depending on the sample collection

method. When the vaginal microbiota was composed of more than two bacteria (sample

numbers 2 and 3), the bacterial species comprising the microbiota were found to be the

same, although there were slight differences in the proportions present. PCoA was

performed to visualize the influence of the sample collection method (Fig. 3).

Regardless of the sampling method, each sample was located in approximately the same

coordinates and we thus concluded that the sample collection method had no effect in

MinION™ sequencing. As this is the first report of vaginal microbiota analysis using

MinION™, we mainly selected lavage samples to ensure the recovery of more

organisms, but we were able to show that swab samples were sufficient.

To clarify the usefulness of long-read sequencing, one of the properties of MinION™,

we picked up *Lactobacillus iners*-dominant (> 99% in V1–9 analysis) samples (Group I:

sample numbers 5, 9, and 10) and *Lactobacillus crispatus*-dominant (>99% in V1–9

analysis) samples (Group C: sample numbers 6, 11, and 14) and compared the results of

16S rRNA analyses of the V1–9 and V3–4 regions (Fig. 5). The bioinformatic pipeline

used a method where each amplicon sequence was directly compared with the database,

scored for similarity (Minimap2 score) according to a proprietary algorithm, and then

applied to the bacterial species with the highest score. Hence, if there were multiple

species with the same score, we treated the reads as potentially existing in all of the

possible species. The reads assigned to "Others" in this study could be (1) bacteria with

low abundance, or (2) bacteria that could not be assigned to a single species due to

competition between multiple species. In Group I (Fig. 5a), *Lactobacillus iners* was

well recognized in the V3–4 16S rRNA analysis, but the numbers of reads assigned to

"Others" increased in the V3–4 analysis with 41.9, 44.8, and 57.4 reads compared with

6.0, 5.0, and 6.0 reads in the V1–9 analysis, respectively. In Group C (Fig. 5b), V3–4

analysis showed that *Lactobacillus crispatus* accounted for about 50% of the reads and

Lactobacillus acidophilus, *Lactobacillus helveticus*, *Lactobacillus amylovorus*,

Lactobacillus kefiranofaciens, *Lactobacillus hamsteri*, and *Lactobacillus gallinarum*

were classified in similar proportions in all three specimens, which had not been found

in V1–9 analysis. The numbers of reads assigned to "Others" were 15.5, 25.5, and 36.0

reads in groups V1–9 and 130.2, 110.5, and 176.6 reads in groups V3–4, respectively.

This suggested that assigning bacteria using the V1–9 region is more informative for

determining the appropriate bacterial species than using only the V3–4 region. Similar

to our previous report (Matsuo, et al., 2020), the long-read sequencing with MinION™

showed high identification accuracy compared with the short-read sequencing under the

same bioinformatic pipeline.

In terms of the usefulness of the MinION™ sequencer in gynecological practice, it is

interesting that we were able to show the detailed condition of vaginal dysbiosis in a

short period of time, which was not noted by the Nugent score. It has been suggested

that vaginal dysbiosis may have a connection with the incidence of sexually transmitted

infections (Shipitsyna, et al., 2020) and genital viral infections (Borgogna, et al., 2020)

(Hoang, et al., 2020) or cervical lesions (Dahoud, et al., 2019). It has also been shown

that metagenomic analysis using NGS applications to assess vaginal microbiota can

provide more accurate results than conventional methods such as the Nugent score,

Amsel's criteria, and bacterial cultures (van den Munckhof, et al., 2019). Despite the

potential value of vaginal microbiota analysis in protecting women's health, the

challenges of all currently used sequencing equipment, including their price and running

costs, and the 2–3 weeks required to obtain results in outsourced testing (in Japan), have

prevented widespread use of vaginal microbiota analysis.

Having used MinION™ in our vaginal microbiota analysis workflow, we have seen

four key benefits. 1) Convenience: The unit is extremely light, small, and portable, and

sequencing can occur via a USB connection to an in-clinic PC. 2) Speed: The average

time between specimen collection and result disclosure is 2 days. 3) Economy: Analysis

of a small number of samples can be performed without increasing the unit cost of the

test, by using individual barcodes. 4) Functionality: Although MinION™ has been

reported to have a slightly higher error rate, the technical issues are being resolved and

even the intestinal microbiota can now be identified (Matsuo, et al., 2020). In particular,

the analysis of vaginal microbiota characterized by low diversity, as in the present

study, allowed for comparable species-level identification (Kilianski, et al., 2015).

Although this study focused on the vaginal microbiota, a similar workflow could be

applied to many clinical areas in the future, and the benefits of MinION™ could make it

easier for clinicians to successfully perform bacterial metagenome analysis. In addition,

future large-scale microbiota studies could lead to new clinical findings.

In conclusion, our validation shows that the MinION™ long-read sequencer provides a

low-cost, rapid workflow for identifying vaginal microbiota with higher resolution in a

clinical setting. Detecting vaginal microbiota at the species level has the potential to

identify risks to women's health, as well as facilitating large-scale clinical studies in any

medical field.

Authors' roles

S.K., Y.Ma., Y.Mo., H.O. and K.H. designed and supervised the study. S.K., Y.Mo. and

H.O. contributed to sample collection. S.N., K.K. and T.I. built the bioinformatics

pipeline. S.K., Y.Ma. and K.H. contributed to the experiments. S.K. analyzed the data.

S.K. wrote the manuscript. Y.Ma., S.N., K.K., H.O. and K.H. contributed to editing the

manuscript. All authors read and approved the final manuscript.

Acknowledgments

We would like to thank Edanz (https://en-author-services.edanzgroup.com/ac) for

editing the English text of a draft of this manuscript.

Funding

This work was supported by Japan Society for the Promotion of Science KAKENHI

Grant Numbers JP19K09339 (to Y.Ma.), JP17H07123 (to K.K.), JP20K06612 (to K.K.),

and the Branding Program as a World-leading Research University on Intractable

Immune and Allergic Diseases supported by the Ministry of Education, Culture, Sports,

Science and Technology of Japan.

Conflict of interest

No authors have any conflicts of interest to declare.

Abbreviations

BV: Bacterial vaginosis, CST: community state type, NGS: Next-generation

sequencing, PCoA: principal coordinate analysis, PCR: polymerase chain reaction,

rRNA: ribosomal RNA

Data availability

Sequence data from this article have been deposited in the DDBJ DRA database

(www.ddbj.nig.ac.jp/dra/index-e.html) under accession numbers DRR244979–

DRR245006.

Additional files

Supplementary Table S1. Representative bacterial genomes stored in the GenomeSync

database.

Supplementary Table S2. Details of identified species by V1–9 16S rRNA analysis

using MinION™ by Minimap2 best-hit score (>1%).

References

Amsel R, Totten PA, Spiegel CA, Chen KC, Eschenbach D, Holmes KK. Nonspecific vaginitis. Diagnostic criteria and microbial and epidemiologic associations. *Am J Med* 1983; **74**: 14-22.

Borgogna JC, Shardell MD, Santori EK, Nelson TM, Rath JM, Glover ED, Ravel J, Gravitt PE, Yeoman CJ, Brotman RM. The vaginal metabolome and microbiota of cervical HPV-positive and HPV-negative women: a cross-sectional analysis. *BJOG* 2020; **127**: 182-192.

Dahoud W, Michael CW, Gokozan H, Nakanishi AK, Harbhajanka A. Association of Bacterial Vaginosis and Human Papilloma Virus Infection With Cervical Squamous Intraepithelial Lesions. *Am J Clin Pathol* 2019; **152**: 185-189.

Frith MC. A new repeat-masking method enables specific detection of homologous sequences. *Nucleic Acids Res* 2011; **39**: e23.

Hoang T, Toler E, DeLong K, Mafunda NA, Bloom SM, Zierden HC, Moench TR, Coleman JS, Hanes J, Kwon DS *et al.* The cervicovaginal mucus barrier to HIV-1 is diminished in bacterial vaginosis. *PLoS Pathog* 2020; **16**: e1008236.

Human Microbiome Project C. Structure, function and diversity of the healthy human microbiome. *Nature* 2012; **486**: 207-214.

Kai S, Matsuo Y, Nakagawa S, Kryukov K, Matsukawa S, Tanaka H, Iwai T, Imanishi T, Hirota K. Rapid bacterial identification by direct PCR amplificatiaon of 16S rRNA genes using the MinION™ nanopore sequencer. *FEBS Open Bio* 2019; **9**: 548-557.

Kilianski A, Haas JL, Corriveau EJ, Liem AT, Willis KL, Kadavy DR, Rosenzweig CN, Minot SS. Bacterial and viral identification and differentiation by amplicon sequencing on the MinION nanopore sequencer. *Gigascience* 2015; **4**: 12.

Larsson PG, Bergstrom M, Forsum U, Jacobsson B, Strand A, Wolner-Hanssen P. Bacterial vaginosis. Transmission, role in genital tract infection and pregnancy outcome: an enigma. *APMIS* 2005; **113**: 233-245.

Li H. Minimap2: pairwise alignment for nucleotide sequences. *Bioinformatics* 2018; **34**:

3094-3100.

Matsuo Y, Komiya S, Yasumizu Y, Yasuoka Y, Mizushima K, Takagi T, Krykov K,

Fukuda A, Morimoto Y, Naito Y, Okada H, Bono H, Nakagawa S, Hirota K. Full-length

16S rRNA gene amplicon analysis of human gut microbiota using MinION nanopore

sequencing confers species-level resolution. *bioRxiv* 2020

https://doi.org/10.1101/2020.05.06.078147

Mikheyev AS, Tin MM. A first look at the Oxford Nanopore MinION sequencer. *Mol*

Ecol Resour 2014; **14**: 1097-1102.

Mitsuhashi S, Kryukov K, Nakagawa S, Takeuchi JS, Shiraishi Y, Asano K, Imanishi T.

A portable system for rapid bacterial composition analysis using a nanopore-based

sequencer and laptop computer. *Sci Rep* 2017; **7**: 5657.

Nakagawa S, Inoue S, Kryukov K, Yamagishi J, Ohno A, Hayashida K, Nakazwe R,

Kalumbi M, Mwenya D, Asami N *et al.* Rapid sequencing-based diagnosis of infectious

bacterial species from meningitis patients in Zambia. *Clin Transl Immunology* 2019; **8**:

e01087.

Nardis C, Mosca L, Mastromarino P. Vaginal microbiota and viral sexually transmitted

diseases. *Ann Ig* 2013; **25**: 443-456.

Nugent RP, Krohn MA, Hillier SL. Reliability of diagnosing bacterial vaginosis is

improved by a standardized method of gram stain interpretation. *J Clin Microbiol* 1991;

29: 297-301.

Ondov BD, Bergman NH, Phillippy AM. Interactive metagenomic visualization in a

Web browser. *BMC Bioinformatics* 2011; **12**: 385.

Ravel J, Gajer P, Abdo Z, Schneider GM, Koenig SS, McCulle SL, Karlebach S, Gorle

R, Russell J, Tacket CO *et al.* Vaginal microbiome of reproductive-age women. *Proc*

Natl Acad Sci U S A 2011; **108 Suppl 1**: 4680-4687.

Shen W, Le S, Li Y, Hu F. SeqKit: A Cross-Platform and Ultrafast Toolkit for FASTA/Q

File Manipulation. *PLoS One* 2016; **11**: e0163962.

Shipitsyna E, Khusnutdinova T, Budilovskaya O, Krysanova A, Shalepo K, Savicheva

A, Unemo M. Bacterial vaginosis-associated vaginal microbiota is an age-independent

risk factor for Chlamydia trachomatis, Mycoplasma genitalium and Trichomonas

vaginalis infections in low-risk women, St. Petersburg, Russia. *Eur J Clin Microbiol*

Infect Dis 2020; **39**: 1221-1230.

Thomason JL, Gelbart SM, Scaglione NJ. Bacterial vaginosis: current review with

indications for asymptomatic therapy. *Am J Obstet Gynecol* 1991; **165**: 1210-1217.

van den Munckhof EHA, van Sitter RL, Boers KE, Lamont RF, Te Witt R, le Cessie S,

Knetsch CW, van Doorn LJ, Quint WGV, Molijn A *et al.* Comparison of Amsel criteria,

Nugent score, culture and two CE-IVD marked quantitative real-time PCRs with

microbiota analysis for the diagnosis of bacterial vaginosis. *Eur J Clin Microbiol Infect*

Dis 2019; **38**: 959-966.

Virtanen S, Kalliala I, Nieminen P, Salonen A. Comparative analysis of vaginal

microbiota sampling using 16S rRNA gene analysis. *PLoS One* 2017; **12**: e0181477.

Tables

Table I Participants' background

Sample	Age (y.o.)	AMH (ng/mL)	LH (mIU/mL)	FSH (mIU/mL)	BMI (kg/m²)
1	36	0.47	7.2	6.8	18.8
2	36	1.15	3.7	5.0	25.6
3	42	0.23	8.3	11.3	21.0
4	30	9.16	6.8	7.2	26.0
5	43	1.67	2.8	5.5	20.8
6	33	3.11	3.2	5.5	25.0
7	35	5.89	5.1	7.9	20.9
8	35	4.98	2.6	4.9	18.8
9	43	2.26	1.8	6.0	19.4
10	39	0.59	6.5	6.3	22.4
11	37	2.34	8.9	6.0	18.6
12	34	4.52	5.2	5.8	20.3
13	35	2.62	8.8	8.2	23.0
14	37	3.16	2.0	3.8	22.1
15	40	3.69	8.2	7.6	20.4
16	38	2.27	6.4	8.4	18.1
17	35	3.66	6.9	7.8	20.4
18	38	0.44	3.5	6.0	19.1

AMH: anti-Mullerian hormone, LH: luteinizing hormone, FSH: follicle-stimulating

hormone, BMI: body mass index

475 **Table II** Statistics of MinION™ V1–9 sequencing data

Sample	Pass reads				Filtered reads	
	No. of reads	Min (bp)	Avg (bp)	Max (bp)	No. of reads	Avg (bp)
1_lavage_V1-9	115,957	128	1,434.3	4,377	91,449 (78.9%)	1,633.5
1_swab_V1-9	63,547	155	1,078.0	4,193	31,462 (49.5%)	1,619.3
2_lavage_V1-9	33,721	158	1,507.2	4,486	28,718 (85.2%)	1,607.0
2_swab_V1-9	53,584	145	1,203.6	3,746	33,294 (62.1%)	1,613.6
3_lavage_V1-9	28,071	154	1,396.3	4,875	20,958 (74.7%)	1,613.8
3_swab_V1-9	23,053	162	1,521.0	3,990	20,073 (87.1%)	1,614.5
4_lavage_V1-9	30,603	155	1,502.9	4,977	25,756 (84.2%)	1,639.5
4_swab_V1-9	37,597	161	1,103.2	3,383	19,588 (52.1%)	1,644.9
5_lavage_V1-9	15,836	147	1,346.1	3,393	10,688 (67.5%)	1,632.8
6_lavage_V1-9	38,020	144	1,517.2	3,787	32,255 (84.8%)	1,642.4
7_lavage_V1-9	61,432	158	1,218.8	5,035	37,018 (60.3%)	1,631.4
8_lavage_V1-9	52,071	111	1,257.5	3,344	33,906 (65.1%)	1,603.2
9_lavage_V1-9	119,745	125	1,530.5	4,664	103,212 (86.2%)	1,649.4
10_lavage_V1-9	102,333	150	1,496.5	4,980	84,028 (82.1%)	1,650.2
11_lavage_V1-9	61,756	121	1,541.3	3,519	54,116 (87.6%)	1,647.8
12_lavage_V1-9	101,755	120	1,433.9	4,759	77,756 (76.4%)	1,635.8
13_lavage_V1-9	71,041	134	1,437.5	4,828	56,601 (79.7%)	1,599.2
14_lavage_V1-9	42,629	159	1,367.7	4,981	31,447 (73.8%)	1,629.0
15_lavage_V1-9	52,908	172	1,452.9	3,333	42,910 (81.1%)	1,630.4
16_lavage_V1-9	52,230	153	1,480.3	4,345	43,549 (83.4%)	1,633.0
17_lavage_V1-9	38,219	154	1,478.3	3,453	31,757 (83.1%)	1,625.3
18_lavage_V1-9	28,886	185	1,464.7	3,328	23,816 (82.4%)	1,631.8

Min: minimum read length, Avg: average read length, Max: maximum read length

Figure legends

Fig. 1 Workflow of 16S rRNA amplicon sequencing with the MinION™ platform and

bioinformatic analysis. After collecting the vaginal sample, it should be stored in a freezer

at −30°C until DNA extraction begins. Some swab collection kits should be stored at

room temperature, in accordance with the manufacturer's instructions. Sequencing

libraries are generated by a four-primer PCR-based strategy; in the initial stages of PCR,

the 16S rRNA gene is amplified with an inner primer pair. The PCR product is amplified

with the outer primers and targeted to introduce identical barcode and tag sequences at

both ends, allowing for the attachment of adapter molecules in a one-step reaction. The

library is then loaded into a MinION™ connected to a personal computer. With our

experimental method, sequencing runtime of 10 min is sufficient for 3,000 reads to be

obtained. A final report of the microbiota can be presented within 2 days of initiating the

analysis.

Fig. 2 Taxonomic profiles comparing the different sampling method (lavage and swab)

results of V1–9 16S rRNA MinION™ sequencing with 3,000 randomly sampled reads

after filtration. In the analytical algorithm, we assigned the bacterial name that showed

the highest Minimap2 score for each read; bacteria with an assignment of less than 1%

were included in "Others."

Fig. 3 PCoA of vaginal microbiota by different sampling techniques. The coordinates of

swab samples and lavage samples were nearly equivalent and did not show any effect of

the vaginal specimen collection method in the 3,000 filtered reads of V1–9 MinION™

sequencing.

Fig. 4 Taxonomic profiles comparing the results obtained for 3,000 and 10,000 filtered

reads of V1–9 MinION™ sequencing. In the analytical algorithm, we assigned the

bacterial name that showed the highest Minimap2 score for each read; bacteria with an

assignment of less than 1% were included in "Others."

Fig. 5 Taxonomic profiles comparing the results of V3–4 and V1–9 16S rRNA sequences

using MinION™ in 3,000 filtered reads. **(a)** In three cases, *Lactobacillus iners* was shown

to constitute more than 99% of the bacteria by MinION™ analysis with the V1–9 region

(Group I). **(b)** In three cases, *Lactobacillus crispatus* was shown to constitute more than

99% of the bacteria by MinION™ analysis with the V1–9 region (Group C).

Duration

Sample collection
        and -30℃ freezing until DNA extraction   -

↓

DNA extraction using QIAamp UCP Pathogen Mini Kit   90 min.

↓

V1-9 16S rRNA gene PCR with 4-primer method

PCR: KAPA2G™ Robust HotStart ReadyMix PCR Kit   60 min.

Barcoding: PCR Barcording Kit SQK-PBK004

↓

MinION™ sequencing   10 min.

Live basecalling with MinKNOW software during the run

↓

Read mapping and annotation using the GSTK software

Extraction of non-human reads   approximately
                                                 1 hour per sample
Sequence search against the GenomeSync database

↓

Output as a report visualized with Krona chart sofrware

                The whole process will be finished in two days.

Figure 1

Figure 2

Figure 3

Figure 4

Fig. 5a

Fig. 5b

Lactobacillus iners
Othres

Lactobacillus crispatus
Lactobacillus acidophilus
Lactobacillus helveticus
Lactobacillus amylovorus
Lactobacillus kefiranofaciens
Lactobacillus hamsteri
Lactobacillus gallinarum
Lactobacillus kalixensis
Lactobacillus psittaci
Othres