Large-scale calcium imaging with a head-mounted axial scanning 3D fluorescence microscope

Yuichiro Hayashi[1]*, Ko Kobayakawa[1] and Reiko Kobayakawa[1]

[1] Institute of Biomedical Sciences, Kansai Medical University, 2-5-1 Shin-machi,

Hirakata, Osaka 573-1010, Japan

* Correspondence

Correspondence: hayasyui@hirakata.kmu.ac.jp

Summary

Miniaturized fluorescence microscopes are becoming more important for deciphering the neural codes underlying various brain functions. With gradient index (GRIN) lenses, these devices enable recording neuronal activity in deep brain structures. However, to minimize any damage to brain tissues and local circuits, the diameter of the GRIN lens should be 0.5–1 mm, resulting in a small field of view. Volumetric imaging capability might increase the number of neurons imaged through the lenses considering the three-dimensional (3D) structure of neural circuits in the brain. To observe 3D calcium dynamics, we developed a miniaturized microscope with an electrically tunable lens and a novel CNMF-based neural signal extraction algorithm for wide-field 3D imaging data. By combining the hardware and software, approximately 1000 neurons were imaged from the cortices of freely behaving mice. Compared with the state-of-the-art 2D imaging technique, the proposed 3D method imaged 1.7–2.6 times more cells with a higher separation of cellular signals.

Introduction

Calcium imaging enables to record the activity of hundreds of neurons at cellular resolution simultaneously. In particular, the combined use of miniature head-mounted microscopes and genetically encoded calcium indicators allows tracking the neuronal activity for days to months in freely moving small animals [1,2]. For deep brain imaging, gradient index (GRIN) lenses have been used for deep brain imaging to relay images to microscopes [1-4]. However, the implantation of GRIN lenses damages brain tissue. Lenses with a smaller diameter (0.5–1 mm) and a small field of view might be used for deep brain imaging to minimize the damage. Considering the three-dimensional (3D) structure of neural circuits in the brain, 3D imaging represents a possible strategy to maximize the number of recorded cells through a limited field of view.

Light-field microscopy (LFM) has performed 3D fluorescence calcium imaging [5,6]. Typically, LFM maps a 3D image onto an image sensor using a microlens array. The 3D image is then computationally reconstructed from the 2D image captured by the image sensor [7]. A miniature implementation of LFM, MiniLFM, imaged fluorescence calcium signals within a volume of $700 \times 600 \times 360$ µm³ at near single-cell resolution in the hippocampal CA1 area of freely moving mice [8]. In LFM, spatial and depth data are mapped onto a single image sensor. Therefore, the LFM frame rate can be as high as 2D

imaging, but the spatial resolution is usually low. To overcome this limitation,

Miniscope3D employs an optimized microlens array and a new reconstruction algorithm

[9]. The system achieves ~2× higher spatial resolution than that of MiniLFM. However,

its calcium imaging performance *in vivo* calcium has not been evaluated.

An alternative strategy for 3D imaging is to add a focusing element to the

microscope and sequentially capture 2D images at different depths. The electrowetting

lens (EWTL) is a small, fast, and lightweight variable focus lens. EWTL-based 3D

imaging systems combined with a one- or two-photon laser scanning microscope have

been used in small animals [10,11]. In these systems, the head-mounted optical unit

contains only the objective lens. The EWTL connects to the laser scanning device with

an optical fiber bundle. Therefore, the head-mounted unit is lightweight. However, the

optical fiber bundle limits the spatial resolution of the system. Moreover, these systems

rely on point scanning, which restricts the scan rate of 3D volumes.

Owing to the high and fluctuating background, extracting neuronal signals

from wide-field imaging data is another challenging issue in 2D and 3D imaging.

Principal component analysis/independent analysis (PCA/ICA) was the first attempt to

extract neuronal signals from wide-field imaging data automatically [12] and has been

applied to 3D imaging [6]. However, it is a linear demixing method and therefore often

fails to decorrelate signals from spatially overlapping cells [13].

Sparse matrix factorization [14] and non-negative matrix factorization (NMF) [15] overcome this problem. These methods can deal with overlapping cells by introducing nonlinearity. Constrained NMF (CNMF) improved the demixing performance by introducing temporal dynamics of the calcium activity [13]. However, the original CNMF implementation was designed for images with low-background fluorescence, such as two-photon and light-sheet microscopy. The algorithm performance is often poor on wide-field imaging data [16]. Although several CNMF-based methods have been proposed for handling wide-field imaging data [17,18], they are unsuitable for 3D imaging.

Here, we developed an EWTL-based wide-field 3D fluorescence microscope and a novel CNMF-based neuronal signal extraction algorithm termed CNMFw3 to record the calcium activity from freely behaving small animals at cellular resolution. Combining the hardware and software, we imaged ~1000 neurons from the cortex of freely behaving mice with a higher separation of cellular signals than that obtained with conventional 2D imaging.

Results

Microscope design

We designed a miniature microscope to capture 3D images from freely behaving small animals. Because of its accessibility and low cost, the design was based on the UCLA miniscope v3 (miniscope.org). An EWTL was inserted in the optical path between the objective lens and the dichroic mirror for axial scanning (Figure 1A). The microscope body was 3D-printed to allow rapid prototyping, and the parts were reusable [19]. The weight of the microscope was 3.7 g, allowing the mice to run freely (Figure 1B). We used several types of objective lenses for different applications. Indeed, the aspherical singlet (Asph; Figure 1C, *Left*) is suitable for relay optics because of its short length and high numerical aperture. A long focal length lens was used for large fields of view (Ach; Figure 1C, *Center*). We used tandem achromatic doublets (TAch; Figure 1C, *Right*) to image the tissue through a cranial window because of their small tip diameter (2 mm), which allows access to the small cranial window on the tissue surface (Figure 1G). The axial scan range with these lenses was 300–600 μm (Figure 1D). Figure 1E and Supplementaly Table 1 show the optical performances. The microscope captured volumetric images at 5 Hz, and each volume image consisted of 12 slices (Figure 1F). The last image in the Z-scan sequence was blurry because of the limited scanning speed of the EWTL (Figure 1G). Therefore, 11 of 12 slices were used to reconstruct the 3D image.

CNMF-based neuronal activity extraction algorithm for wide-field 3D imaging data

To extract images from the wide-field 3D microscope, we developed a new algorithm

based on CNMF. Our new method, termed CNMFw3 (CNMF for wide-field 3D data),

first estimated the background of the raw image stream with rank-1 matrix

factorization (Figure 2A,B). The background-subtracted images were then spatially

high-pass-filtered to remove the remaining background (Figure 2C). Next, we looked for

initial cell candidates employing the greedy initialization method used in CNMF-E [18]

with modifications to fit 3D data. The method seeks the local maxima of the product of

local correlation image and peak-to-noise ratio image (Figure 2D, E). Finally, the initial

spatial and temporal components were refined with CNMF [13] (Figure 2F, G).

CNMFw3 accurately recovers neuronal activity in simulated data

We compared the neuronal activity extraction performance of CNMFw3 with that of

other algorithms for 3D imaging data, namely PCA/ICA and CNMF (plain CNMF) and

CNMF-E, the most common ones widely used algorithm for 2D wide-field imaging data.

We synthesized 4 image streams with various signal-to-background ratios (SBR, defined

as the amplitude ratio between the signal and the background) ranging from 0.1 to 0.8.

Each image stream (100 × 100 × 10 pixels, 1000 frames) contained 100 synthesized

neurons and background fluctuations extracted from real imaging data (Figure 3A).

Poisson noise was added to each pixel to simulate photon shot noise. The number of

well-extracted neurons (temporal correlation to the ground truth > 0.9) was higher

using CNMFw3 than that obtained with the other methods (Figure 3B,C). PCA/ICA and

plain CNMF showed poor performances at the low-SBR range (SBR = 0.4). The

performance of CNMF-E was excellent even at SBR = 0.4. However, the number of

well-extracted neurons was approximately half of that acquired with CNMFw3 because

CNMF-E only handles one Z-slice of the entire volume. Notably, unlike other methods,

CNMFw3 still extracted cells at SBR = 0.2 (Figure 3B), indicating that the method can

extract more neurons than other methods in real data.

Application of CNMFw3 to *in vivo* imaging in freely behaving mice

We next applied the hardware and software to *in vivo* microendoscopic imaging. We

imaged neurons in the mouse prefrontal cortex (PFC) through 1.0 mm diameter GRIN

lens (Figure 4A, B). Volumetric fluorescence images of GCaMP6f-expressing neurons in

the PFC were processed with CNMFw3 (Figure 4C, D). CNMFw3 extracted 973

putative neurons. Figure 4E shows fluorescence traces extracted from the 3D movie.

3D imaging is more beneficial for deep brain imaging because the smaller

diameter GRIN lens minimizes tissue damage when imaging deeper brain structures.

Therefore, we used the 3D microscope to image the mouse central amygdala (CeA),

located approximately 4 mm below the brain surface, through a 0.5-mm diameter GRIN

lens (Figure 5A, B). In total, 130 neurons situated at different depths were visible, as

shown in Figure 5C and D, and their fluorescence traces were extracted by CNMFw3

(Figure 5E).

Advantages of 3D over 2D imaging

To evaluate the merit of the proposed 3D imaging technique in real data analysis, we

compared it with conventional 2D imaging. We used CNMF-E, a popular algorithm for

wide-field 2D imaging data [18], to extract neural activity from 2D images. The number of

cells extracted from 2D imaging was defined as the largest number of cells extracted

from single Z-slices in the 3D volume (Figure 6A). Significantly more cells were detected

from the 3D image than from the 2D image in the PFC and CeA (Figure 6B,C; $t = 19.54$,

$p = 0.0026$, n = 3 mice for the PFC, $t = 3.278$, $p = 0.031$, n = 5 mice for the CeA;

one-sample two-tailed t-test).

In wide-field 2D imaging, neurons in, above, and below the focal plane are

recorded. This is problematic for extracting single-cell activity from the image because

signals from multiple neurons tend to overlap. Although recent methods for neural

activity extraction partially allow demixing such overlapping signals, reducing the

overlap will improve the accuracy of the neural activity extraction [13,18]. To test whether

our 3D imaging method obtained a reduced cell-cell overlap, we compared the minimum

cell-cell distance and overlap of the spatial footprint of extracted cells. As shown in

Figure 6D and E, the minimum distance between neighboring cells was significantly

larger with 3D imaging than that in 2D (U = 489100, p < 0.0001, n = 2083 cells for 3D

imaging of the PFC, n = 1227 cells for 2D imaging of the PFC; U = 37030, p < 0.0001, n =

515 cells for 3D imaging of the CeA, n = 231 cells for 2D imaging of the CeA;

Mann–Whitney U test). We then measured the overlap of the spatial footprint. The

spatial footprint was defined as the 50th percentile of the peak value of each spatial

filter extracted by the algorithms. The overlap rate for each cell was defined as the

proportion of shared pixels with other cells in its spatial footprint. We compared the

spatial footprint sizes. In the recording from the PFC and CeA, the sizes of the spatial

footprints were larger in 3D imaging than those in 2D (Figure 6F,G; U = 208000, p <

0.0001, n = 2083 cells for 3D imaging of the PFC, n = 1227 cells for 2D imaging of the

PFC; U = 7117, p < 0.0001, n = 515 cells for 3D imaging of the CeA, n = 231 cells for 2D

imaging of the CeA; Mann–Whitney U test). However, the overlap rate was significantly

smaller in 3D imaging than that in 2D (Figure 6H,I; $U = 347100$, $p < 0.0001$, n = 2083

cells for 3D imaging of the PFC, n = 1227 cells for 2D imaging of the PFC; $U = 29570$, p

< 0.0001, n = 515 cells for 3D imaging of the CeA, n = 231 cells for 2D imaging of the

CeA; Mann–Whitney U test). The longer cell-cell distance, larger spatial footprint size,

and smaller overlap in 3D imaging might contribute to the accurate estimation of

neural activity.

Discussion

In this study, we developed a simple, low cost 3D fluorescence microscope to use

in freely moving small animals. The microscope was made from off-the-shelf optical

components, CMOS image sensor electronics designed by an open-source hardware

project (UCLA miniscope), and a custom 3D-printed microscope body. These parts were

easily available at a low cost. Moreover, the 3D-printed parts were obtained using a

consumer-grade 3D printer (Form2; Formlabs). For these reasons, the design was easy

to test and modify. The weight of the microscope might affect the animal behavior and

must be kept as low as possible. The microscope designed for this study weighed 3.7 g,

which was acceptable for its use in mice. A further weight reduction is required to

record images from smaller animals such as zebra finches [19]. Using the new, lightweight

version of the UCLA miniscope (miniscope v4; miniscope.org/index.php/Miniscope_v4)

in place of the miniscope v3 might provide a suitable solution to reduce the weight.

Moreover, the miniscope v4 includes an EWTL in its optical path and can capture 3D

images by adding electronics for fast focus-scanning.

Not only the methods for image acquisition but also the neural activity

extraction algorithms affect the quality and quantity of neuronal activity traces.

However, existing algorithms for 3D imaging have not shown satisfactory performances

for wide-field 3D imaging data (Figure 3B–D). Based on the CNMF algorithm [13], we

developed a novel algorithm, CNMFw3, to extract individual cell activity from 3D

images (Figure 2). The algorithm showed superior performances for simulated and real

wide-field 3D imaging data (Figure 4,5). CNMFw3 uses a single, global background

model and spatial high-pass filtering for rejecting background fluorescence (Figure 2B,

C). CNMFw3 was significantly more performant for the simulated data than the

existing methods (Figure 3B–D). However, CNMF-E uses a multiple local background

model to achieve higher accuracy for wide-field 2D imaging data compared to other

methods using the single background model [18]. Therefore, adapting the local

background model employed by the CNMF-E algorithm to use wide-field 3D imaging

data will be an important research field for the future.

Using the new hardware and software, we could image 1.7–2.6 times more cells

than with standard 2D imaging techniques (Figure 6C). Moreover, neurons extracted

from 3D images showed longer cell-cell distances (Figure 6D, E), larger spatial

footprints (Figure 6F,G), and smaller overlap of the spatial footprints (Figure 6H,I) than

those from 2D images. Because the overlap of spatial cell footprints is a large obstacle to

the neural activity extraction from wide-field imaging data [13,18], these features lead to a

higher quality of the neural activity extraction.

Overall, by combining the newly designed hardware and software for 3D

imaging, greater quality and amount of neuronal activity was recorded from freely

behaving mice than with conventional methods.

Methods

Head-mounted 3D microscope

The miniature microscope design was based on the UCLA miniscope v3 (miniscope.org).

Excitation light emitted from a blue LED (LXML-PB01-0030; Lumileds, Aachen,

Germany). The light passed through an excitation filter (ET470/40x; Chroma

Technology, Bellows Falls, VT) and was reflected by a dichroic mirror (T495lpxr;

Chroma Technology) onto the tissue through an EWTL (Varioptic A-25H0; Corning, Corning, NY), an objective lens (aspherical singlet; #83-605, Edmund Optics, Burlington, NJ, achromatic doublet; #49-270, Edmund optics, or tandem configuration of achromatic doublets; #84-126 and #49-271, Edmund Optics), and a GRIN lens (0.5-mm diameter, 7.9-mm length; CLHS050GFT009; Gofoton, Tsukuba, Japan, or 1.0-mm diameter, 3.7-mm length CLH-100-WD002-SSI-GF3; Gofoton). The fluorescent emissions collected by the lenses were passed through the dichroic mirror and an emission filter (ET525/50m; Chroma Technology). The fluorescence image was focused by a tube lens (#49-277; Edmund Optics) and captured by a CMOS image sensor (MT9V032; ON Semiconductor, Phoenix, AZ). The microscope body was made using a 3D printer (From2, resin type FGPBLK03; Formlabs, Somerville, MA). The design files are available at https://github.com/yuichirohayashi/Zscope. The driving voltage for the EWTL (square wave, 2 kHz) was generated using a data acquisition board (PCIe-6259, National Instruments, Dallas, TX) controlled by custom software written in LABVIEW 7.1 (National Instruments) and amplified with a linear amplifier (As-904-150B, NF Corporation, Yokohama, Japan). The EWTL driver was synchronized to the image acquisition of the CMOS image sensor (Supplementaly Figure 1).

Point spread function measurements

Three-dimensional stacks of 1-µm fluorescent beads were captured with a spacing of 10

µm. The point speed function was estimated as a Gaussian function (Figure 1E).

Animals

All experiments complied with the relevant guidelines and regulations. Animal care and

use followed protocols approved by the Animal Research Committee of Kansai Medical

University. All experiments complied with the ARRIVE guidelines

(https://arriveguidelines.org/arrive-guidelines/experimental-animals). Ten adult male

C57BL/6N mice were used. The mice were purchased from Japan SLC, Inc. (Shizuoka,

Japan), housed under a standard 12-h light/dark cycle, and allowed *ad libitum* access to

food and water. At the time of the viral injection surgery, the mice were more than 10

weeks old and had body weights of 25.3 ± 0.4 g (mean $\pm$ SEM).

Surgery

The animals were anesthetized using isoflurane. The skull was exposed, and small

craniotomies (<0.5 mm) were made over the target brain regions.

AAV10-syn-GCaMP6f-WPRE (University of Pennsylvania vector core) was diluted to $5 \times$

10^{12} particles/mL in phosphate-buffered saline, and 200 nL were injected. The

stereotaxic coordinates of the virus injection were as follows: hippocampus, AP +2.3 mm,

ML 1.5 mm, and DV −1.2 mm; PFC, AP +2.0 mm, ML 0.5 mm, and DV −1.0 mm; CeA,

AP −1.1 mm, ML 2.5 mm, and DV −4.5 mm. One week after the viral injection, the

second surgical procedure to implant the optics proceeded. A cranial window assembly

composed of a glass coverslip (0.12 mm thick) and a stainless-steel cannula (2.76-mm

outer diameter, 2.40-mm inner diameter, and 1.5-mm height) for hippocampus imaging

was placed over the dorsal CA1 [20]. For PPC and CeA imaging, we implanted the GRIN

lens to target these regions and glued it to the skull with dental cement (Shofu, Kyoto,

Japan).

Four weeks after implantation of the optical parts, GCaMP-expressing neurons

were imaged using the microscope with a tandem achromatic lens (for CA1), achromatic

doublet lens (for the PFC), or aspherical singlet lens (for the CeA and PFC). The

aluminum baseplate of the microscope was cemented in a position at which neurons

could were visible.

Imaging sessions

The excitation light intensity was approximately 0.1–0.5 mW/mm^2. Images were

captured at 60 Hz for 5 min. Image data were first stored in the computer's memory before being transferred to the hard drive once the image acquisition had been completed to prevent missing frames.

Image processing

ImageJ 1.52 (National Institutes of Health, MA) and MATLAB 2019a (Mathworks, Natick, MA) were used for all analyses. The lateral displacements of the brain were corrected by image translation using TurboReg [21]. The registered image stream was processed with CNMFw3 or other neural activity extraction algorithms. The extracted spatial filters were then visually inspected, and those showing non-neuronal morphology were removed.

Simulation experiments

The synthetic data used for Figure 3 were generated as follows: 100 neurons (ellipsoid of length: 4, width: 4, height: 2 pixels) were randomly placed in a volume of $100 \times 100 \times 10$ pixels, maintaining a minimum center distance of 4 pixels in the XY axis and 1 pixel in the Z-axis. Calcium transients were generated with a Poisson process (mean transient rate 0.1 Hz, 1000-time steps at a 5 Hz frame rate) and convolved with an exponentially

decaying kernel ($\tau = 0.5$ s). Background fluctuations were extracted from real imaging

data by moving the average filter with a window size of 10 pixels. The background was

linearly mixed with the neural activity image. The volumetric data were convolved with

a Gaussian kernel (SD = 2 pixels) to approximate the PSF of the microscope and

scattering by the tissue. Poisson noise with a peak value of 200 was added to each pixel

to simulate the photon shot noise.

Code availability

The MATLAB implementation of CNMFw3 is available at

https://github.com/yuichirohayashi/CNMFw3.

Acknowledgements

This work is supported by JSPS KAKENHI (Grant No. 17K19436, 20K07715, 20H03547) to Y.H.,

JSPS KAKENHI 20H04849 to R.K.

Author Contributions

Conceptualization, Y.H.; Methodology, Y.H.; Investigation, Y.H.; Writing –Original Draft,

Y.H.; Writing –Review & Editing, Y.H., K.K. and R.K.; Funding Acquisition, Y.H. and

R.K.; Resources, Y.H., K.K. and R.K.; Supervision, Y.H.

Declaration of Interests

The authors declare no conflict of interest.

References

1.   Ziv, Y. *et al.* Long-term dynamics of CA1 hippocampal place codes. *Nat Neurosci* **16**,

        264–266 (2013).

2.   Ghosh, K. K. *et al.* Miniaturized integration of a fluorescence microscope. *Nat.*

        Methods **8**, 871–8 (2011).

3.   Jung, J. C., Mehta, A. D., Aksay, E., Stepnoski, R. & Schnitzer, M. J. In vivo

        mammalian brain imaging using one- and two-photon fluorescence microendoscopy. *J*

        Neurophysiol **92**, 3121–3133 (2004).

4.   Levene, M. J., Dombeck, D. A., Kasischke, K. A., Molloy, R. P. & Webb, W. W. In vivo

        multiphoton microscopy of deep brain tissue. *J. Neurophysiol.* **91**, 1908–1912 (2004).

5.   Grosenick, L. M. *et al.* Identification Of Cellular-Activity Dynamics Across Large

        Tissue Volumes In The Mammalian Brain. *bioRxiv* (2017). doi:10.1101/132688

6.   Prevedel, R. *et al.* Simultaneous whole- animal 3D imaging of neuronal activity using

light-field microscopy. *Nat. Methods* **11**, 727–730 (2014).

7. Levoy, M., Ng, R., Adams, A., Footer, M. & Horowitz, M. Light field microscopy. *ACM Trans. Graph.* **25**, 924–934 (2006).

8. Skocek, O. *et al.* High-speed volumetric imaging of neuronal activity in freely moving rodents. *Nat. Methods* **15**, 429–432 (2018).

9. Yanny, K. *et al.* Miniscope3D: optimized single-shot miniature 3D fluorescence microscopy. *Light Sci. Appl.* **9**, (2020).

10. Ozbay, B. N. *et al.* Miniaturized fiber-coupled confocal fluorescence microscope with an electrowetting variable focus lens using no moving parts. *Opt. Lett.* **40**, 2553–2556 (2015).

11. Ozbay, B. N. *et al.* Three dimensional two-photon brain imaging in freely moving mice using a miniature fiber coupled microscope with active axial-scanning. *Sci. Rep.* **8**, 1–14 (2018).

12. Mukamel, E. A., Nimmerjahn, A. & Schnitzer, M. J. Automated analysis of cellular signals from large-scale calcium imaging data. *Neuron* **63**, 747–760 (2009).

13. Pnevmatikakis, E. A. A. *et al.* Simultaneous Denoising, Deconvolution, and Demixing of Calcium Imaging Data. *Neuron* **89**, 299 (2016).

14. Andilla, F. D. & Hamprecht, F. Sparse Space-Time Deconvolution for Calcium Image

Analysis. *undefined* (2014).

15.   Maruyama, R. *et al.* Detecting cells using non-negative matrix factorization on

        calcium imaging data. *Neural Networks* **55**, 11–19 (2014).

16.   Barbera, G. *et al.* Spatially Compact Neural Clusters in the Dorsal Striatum Encode

        Locomotion Relevant Information. *Neuron* **92**, 202–213 (2016).

17.   Lu, J. *et al.* MIN1PIPE□: A Miniscope 1-Photon-Based Calcium Resource

        MIN1PIPE□: A Miniscope 1-Photon-Based Calcium Imaging Signal Extraction

        Pipeline. 3673–3684 (2018). doi:10.1016/j.celrep.2018.05.062

18.   Zhou, P. *et al.* Efficient and accurate extraction of in vivo calcium signals from

        microendoscopic video data. *Elife* **7**, 1–37 (2018).

19.   Liberti, W. A., Perkins, L. N., Leman, D. P. & Gardner, T. J. An open source, wireless

        capable miniature microscope system. *J. Neural Eng.* **14**, 045001 (2017).

20.   Hayashi, Y. NMDA receptor-dependent dynamics of hippocampal place cell

        ensembles. *J. Neurosci.* **39**, 5173–5182 (2019).

21.   Thévenaz, P., Ruttimann, U. E. & Unser, M. A pyramid approach to subpixel

        registration based on intensity. *IEEE Trans. Image Process.* **7**, 27–41 (1998).

Figure legends

Figure 1

3D fluorescence microscope

(A) Section diagram of a 3D microscope. (B) Photo of a mouse with the 3D microscope.

(C) Representation of the three types of objective lenses. *Left,* aspherical singlet (Asph).

Center, achromatic doublet (Ach). *Right,* two achromatic doubles in tandem (TAch). (D)

Relationship between the driving voltage and the depth of focus. (E) Optical resolution

of the microscope equipped with the aspherical singlet lens. Images and profiles of the

1-μm fluorescent beads. Red lines indicate Gaussian fit. (F) Image acquisition sequence.

The solid line indicates the driving signal for the ETWL. The dashed line indicates the

actual depth of focus. (G) Image examples for each Z-slice. The images were taken from

the mouse hippocampal CA1 area using the tandem achromatic doublet lens.

Figure 2

CNMFw3 algorithm

(A) Raw 3D image. Images in panels A–E are maximum intensity projections of z- and

t-stacks. (B) Background estimation by rank-1 factorization. (C) Spatial high-pass

filtering. (D) *Left,* map of temporal correlation between the filtered temporal traces of

each pixel and those of neighboring pixels. *Right, the* peak-to-noise ratio of the filtered

traces of each pixel. **(E)** Local maxima of the product of local correlation image and

peak-to-noise ratio image. Red asterisks indicate the local maxima as the center of

initial cell candidates. **(F)** Contours of all neurons were detected by CNMFw3 on three

orthogonal planes. The contours of 80% of the peak values of each neuron are displayed.

(G) Fluorescence traces of initial cell candidates (*Left*) and cell candidates refined with

CNMF (*Right*).

Figure 3

Performances of CNMFw3

(A) Representative images of simulated data for high (0.8) and low (0.2) SBR. Images

are the maximum projection of an image volume on three orthogonal planes. **(B)**

Number of recovered cells from the simulated data using four methods. The number of

cells with a high correlation between their fluorescence trace and the ground truth

(>0.9) are plotted. The data represent mean ± SEM (n = 10). **(C)** Identification accuracy

of the four methods. Correlation coefficients of the recovered trace to the ground truth

are shown. **(D)** Recovered traces for intermediate SBR (0.4) dataset (blue) and the

corresponding ground truth traces (green) are shown. The 1st, 30th, and 60th from the

highest correlation to the ground truth are displayed.

Figure 4

Three-dimensional fluorescence calcium imaging of PFC neurons in a freely behaving

mouse

(A) Scheme of the anatomical location of the optical path. A GRIN lens (1.0-mm

diameter, 3.7-mm length) was inserted in the PFC. **(B)** Maximum projection of an image

volume on three orthogonal planes. **(C)** Contours of all neurons detected by CNMFw3.

The contours of 80% of the peak values of each neuron are displayed. **(D)** Isometric

perspective of data shown in (C). **(E)** Fluorescence traces extracted by CNMFw3.

Figure 5

3D calcium imaging in the CeA

(A) Scheme of the anatomical location of the optical path. A GRIN lens (0.5-mm

diameter, 7.4-mm length) was inserted in the CeA. **(B)** Maximum projection of an image

volume on three orthogonal planes. **(C)** Contours of all neurons detected by CNMFw3.

The contours of 80% of the peak values of each neuron are displayed. **(D)** The isometric

perspective of data shown in (C). **(E)** Fluorescence traces extracted by CNMFw3.

Figure 6

Comparison between 3D and 2D imaging

(**A**) Experimental design. Recorded 3D images were processed with CNMFw3 (*Left*), and each Z-slices in the 3D image were processed with CNMF-E (*Right*). The number of cells extracted from the 2D image is defined as the largest number of cells extracted from a single Z-slice (shown in red) in the 3D volume. (**B**) Comparison of the number of cells extracted from the 3D and 2D images. (**C**) The ratio between the cell numbers extracted from the 3D image and that from the 2D image. The data represent the individual (solid circles) and mean (bars) values. (**D**) Cumulative distribution of the minimum cell-cell distances in 2D and 3D. The cell-cell distance was defined as the distance between the center of mass of each spatial footprint. (**E**) Box-whisker plot indicating the median, quartiles, and range of minimum cell-cell distances. (**F**) Cumulative distribution of the size of the spatial footprints in 2D and 3D. (**G**) Box-whisker plot indicating the median, quartiles, and range of the size of spatial footprints. (**H**) Cumulative distribution of the overlap rate of spatial footprints in 2D and 3D. (**I**) Box-whisker plot indicating the median, quartiles, and range of the overlap rate of spatial footprints. Asterisks indicate significant differences. $*p < 0.05$, $***p < 0.0001$.

Figure 1

A
Image Sensor
Tube Lens
Emission Filter
Dichroic Mirror
Electrowetting Lens
Base Plate
Objective Lens
Excitation Filter
Drum Lens
LED

B

C
Asph
Ach
TAch
1 mm

D
Depth (μm)
Driving Voltage (V)
◆ Asph
○ Ach
△ TAch

E
Lateral
10 μm
10 μm
Axial
100 μm
10 μm

F
Driving Voltage
Depth
Time (s)
1 2 3 4 5 6 7 8 9 10 11 12

G
1 2 3 4 5 6
7 8 9 10 11 12
100 μm

Figure 2

A Raw data

Background estimation using rank-1 factorization

B Estimated Background   Background-subtracted data

+

Spatial HPF

C High-pass filtered data

Compute Local corr. & PNR

D Local Correlation   PNR

Initialization

E [Local corr.] × [PNR]

F

G Initial   Refined with CNMF

Cell number

0   Time (s)   300

0   Time (s)   300

Figure 3

A

SBR = 0.8   SBR = 0.2

B

Number of recovered neurons

CNMFw3
plainCNMF
PCAICA
CNMF-E(2D)

SBR

C

Temporal correlation to the ground truth

CNMFw3   plainCNMF   PCAICA   CNMF-E (2D)

SBR=0.8
0.4
0.2
0.1

Cumurative fraction

D

CNMFw3   plainCNMF

1st Corr=0.994   1st Corr=0.930

30th Corr=0.987   30th Corr=0.688

60th Corr=0.974   60th Corr=0.163

Extracted
Ground Truth

Figure 4

A

B

Y (µm)

0   X (µm)   1080

Z (µm)

0   250

C

D

E

Cell ID

1 min

Figure 5

Figure 6