Intrinsic variability of current MESF calibrators impacts the assignment of absolute fluorescence values to nanoparticles and extracellular vesicles by flow cytometry

Estefanía Lozano-Andrés[1, #], Tina Van Den Broeck[2, #], Lili Wang[3], Majid Mehrpouyan[4], Ye Tian[5], Xiaomei Yan[5], Marca H.M. Wauben*[1, #], Ger J.A. Arkesteijn*[1]

1 Department of Biomolecular Health Sciences, Faculty of Veterinary Medicine, Utrecht University, Utrecht, The Netherlands

2 BD Biosciences, Erembodegem, Belgium

3 Biosystems and Biomaterials Division, National Institutes of Standards and Technology (NIST), Gaithersburg, MD 20899

4 BD Biosciences, San Jose, CA 95131

5 Department of Chemical Biology, MOE Key Laboratory of Spectrochemical Analysis & Instrumentation, Key Laboratory for Chemical Biology of Fujian Province, College of Chemistry and Chemical Engineering, Xiamen University, Xiamen 361005, People's Republic of China

TRAIN-EV Marie Skłodowska-Curie Action-Innovative Training Network, train-ev.eu

*Both authors contributed equally

Corresponding author:

Estefanía Lozano-Andrés

Department of Biomolecular Health Sciences

Faculty of Veterinary Sciences

Utrecht University

Yalelaan 2, 3584 CM

Utrecht, The Netherlands

T: +31 30 2535451/2535833

E: e.lozanoandres@uu.nl

Abstract

Flow cytometry is a promising technique to characterize nanoparticles (NPs) and extracellular

vesicles (EVs). However, the majority of reported experiments use arbitrary units to indicate

fluorescence intensity. This hampers comparison of results from different laboratories and different

platforms. We investigated the advised use of calibrated molecules of equivalent soluble

fluorophores (MESF)-beads for assignment of absolute fluorescence to NPs and EVs. Firstly, we

evaluated the use of two different FITC-MESF bead sets as calibrators on three different flow

cytometry platforms (BD Influx™, CytoFLEX LX™ and SORP BD FACSCelesta™). Secondly,

NPs and biological 4T1 mammary carcinoma-EVs were analyzed using the BD Influx to calibrate

their fluorescence signals by using four different sets of FITC and PE-MESF beads. Fluorescence

calibration using bright calibrators designed for cellular flow cytometry makes inter-platform

comparison possible. However, the intrinsic uncertainty in the assignment of the currently available

MESF beads, which have a far brighter fluorescence range than NPs and EVs, hampers a reliable

assignment of absolute MESF numbers based on extrapolation into the dim fluorescence range.

The findings that we present in this study demonstrate that for robust comparison of calibrated

values to submicron sized particles the same calibration materials should be used. Furthermore, our

results illustrate the need for dim calibration materials specifically designed for calibration of NPs

and EVs to enable a reliable assignment of absolute fluorescence values in the lower fluorescent

ranges and to perform inter-platform comparison.

Key words: fluorescence, calibration, standardization, MESF, extracellular vesicles,

nanoparticles, flow cytometry

Introduction

The assignment of a specific value is inextricably bound to a variation around this value. The

calibration methods used in flow cytometry are no exception to this rule. A well-known calibration

method in flow cytometry is the use of beads to which a specific fluorescence value is assigned

and expressed in molecules of equivalent soluble fluorophores (MESF) or equivalent number of

reference fluorophore (ERF). These are units established and accepted by the National Institute of

Standards and Technology (NIST). Calibration beads are commercially available and suited for a

large number of fluorescence channels. Flow cytometry is most commonly used for the purpose of

cell analysis by determining the level of fluorescence intensity for the marker of interest on labeled

cells. The fluorescence intensity on the calibrator beads matches the intensity that can be expected

on labeled cells. Therefore the calibrated cellular fluorescence variation that has to be taken into

account closely compares to the intrinsic variation on the calibrator, allowing for data interpolation.

Besides cellular measurements, flow cytometry has attracted the attention of many researchers

working in the field of nanoparticles (NPs) and cell-derived extracellular vesicles (EVs) as a

valuable tool for high-throughput single particle-based analysis [1, 2]. It allows for quantitative

and qualitative multiparametric characterization of subsets of EVs based on the detection of light

scatter signals and fluorescence. Although flow cytometry is one of the methods of choice for EV

analysis, it is not free of challenges and pitfalls, mainly because EV signals are lying close to the

instrument's detection limit. Important requirements need to be met and data need to be reported

in a proper way to improve rigor and reproducibility of flow cytometric EV analysis [3]. Despite

the advancements made in the past years, the vast majority of EV flow cytometric data is still

published in arbitrary units and results are expressed as percentages of subpopulations or median

fluorescence intensity (MFI) relative to the experimental controls. This approach is cumbersome

for the analysis of dim and small particles, whereby positively stained events cannot be fully

discriminated from negative counterparts and background signals. Furthermore, although the

number of flow cytometers that are suitable for NP and EV analysis is increasing, these platforms

are very diverse in terms of fluidics, optics and electronics. Calibrated measurements will not only

improve inter-platform comparison but also contributes to determine the sensitivity of a flow

cytometer, which is especially relevant when a fluorescent threshold is applied. Altogether, the use

of standardized units of fluorescence (MESF or ERF) will strongly contribute to rigor and

reproducibility of EV flow cytometric analysis.. The three-society ISEV-ISAC-ISTH EV flow

cytometry working group has recommended in the published MIFlowCyt-EV framework for

standardized reporting of EV flow cytometry experiments, the use of MESF beads for fluorescence

calibration [3]. Fluorescence calibration is successfully applied for cell flow cytometry in which

arbitrary units are translated to standardized units of fluorescence [4-7]. Key in proper calibration

is the precision and accuracy of the available calibrators. Importantly, a NIST/ISAC

standardization study reported in 2012 observed differences in the assigned standardized number

to calibrators provided by different manufacturers [8]. Therefore we decided in our study to focus

our evaluation on different calibrator sets provided by the same manufacturer to exclude any

additional variations due to inter-manufacturer discrepancy. However, inherent to the method used,

an intrinsic uncertainty of the assignment of absolute numbers of MESF or ERF to calibrator beads

will always exist. This uncertainty will translate in a variation that remains within acceptable range

as long as the sample values are within the range of this calibrator. However, the currently available

calibrator materials are designed for the fluorescence calibration of labeled cells, being much

brighter in fluorescence when compared to the fluorescence of stained EVs and other NPs. We here

analyzed the usability of the available calibrator materials for assigning absolute fluorescence

values for EVs and NPs.

We investigated how the variation within the number of MESF or ERF of the calibrator impacts

slope and position of the regression line and consequently influenced the assignment of absolute

units of fluorescence to NPs and EVs that are far out of the range of these calibrator intensity levels.

To this end, we evaluated the performance of different custom-made calibrator beads sets provided

by the same manufacturer and their inter-platform variation on three different flow cytometers

capable of single NP and EV analysis. Finally, we studied how the use of different bead sets would

affect the calibration of fluorescence signals from NPs and fluorescently stained and antibody

labeled EVs on the same platform.

Material and Methods

Calibration beads

For calibration of the fluorescence axis in the fluorescein isothiocyanate (FITC) channel we used

two different sets of FITC MESF beads (custom-made, 6 μm lot MM2307 #131-10; #131-8; #130-

6; #130-5; #130-3 and 2 μm lot MM2307#156; #159.1; #159.2; #122.3, BD Biosciences, San Jose,

CA). For calibration of the fluorescence axis in the PE channel, we used two different sets of PE

MESF beads (6 μm, commercial QuantiBrite, Catalog No. 340495 lot 62981 and 2 μm, custom

made, lot MM2327#153.1; #153.2; #153.4; #153.5; #153.6, BD Biosciences, San Jose, CA).

The 6 μm FITC MESF beads were prepared by reacting various concentration of FITC with PMMA

Beads (Bangs Labs) in borate buffer at pH 9.2. The 2 μm FITC beads were prepared by reacting

various concentrations of FITC-BSA (with a FITC/BSA molar ratio of 2) with 2 μm carboxylic

beads (Bangs Labs) using EDC/NHS chemistry. The 2 μm PE beads were made as described above

except various concentrations of PE were used with 2 μm carboxylic beads in EDC/NHS chemistry.

These beads were analyzed on a BD LSRFortessa™ (BD Biosciences) and their MESF values were

assigned by cross-calibration using commercially available MESF beads (Flow Cytometry

Standards Corp.). The FITC ERF values were assigned to both 2 μm and 6 μm beads using a

specific lot of FITC-FC Bead (BD Biosciences) as a calibrator with known ERF value, which has

been assigned by NIST. This provided us with two distinct calibrator bead sets that were produced

through the same manufacturing process and assigned using the same instruments and the same

internal NIST traceable calibrator to exclude internal processing variations.

In addition, we measured commercially available Quantum™ FITC-5 MESF (7 μm, Catalog No.

555, lot 14609, Bangs Laboratories) and AccuCheck ERF Reference Particles Kit (3 μm, Catalog

No. A55950, lot #081220207, #081220203, #081220208, Thermo Fisher) which were prepared

according to the manufacturer's instructions.

All calibration bead sets were measured with gain or voltage settings as would be used for the

analysis of small particles (i.e. EVs). In addition, not all beads could be measured on every

instrument. For fair cross-platform comparison of the slopes of the regression lines, only the same

bead populations that were measured on all instruments were included for linear regression

analysis.

Flow cytometer platforms

In this study three flow cytometers were used. A jet in air-based BD Influx (BD Biosciences, San Jose, CA), a BC CytoFLEX LX (Beckman Coulter, Brea, CA) with a cuvette-based system and a cuvette-based SORP BD FACSCelesta™ (BD Biosciences, San Jose, CA) equipped with a prototype small particle side scatter module.

The BD Influx flow cytometer was modified and optimized for detection of submicron-sized particles [9]. In brief, FITC was excited with a 488 nm laser (Sapphire, Coherent 200 mW) and fluorescence was collected through a 530/40 bandpass filter. PE was excited with a 562 nm laser (Jive, Cobolt 150 mW) and fluorescence was collected through a 585/42 bandpass filter. Optical configuration of the forward scatter detector was adapted by mounting a smaller pinhole and an enlarged obscuration bar in order to reduce optical background. This reduced wide-angle FSC (rwFSC) allowed detection of sub-micron particles above the background based on forward scatter [9, 10]. Upon acquisition, all scatter and fluorescence parameters were set to a logarithmic scale. To minimize day to day variations, the BD Influx was standardized at the beginning of each experiment by running 100 and 200 nm yellow-green (505/515) FluoSphere beads (Invitrogen, F8803 and F8848). The instrument was aligned until predefined MFI and scatter intensities where reached with the smallest possible coefficient of variation (CV) for rwFSC, SSC and fluorescence. After optimal alignment, PMT settings required no or minimal day to day adjustment and ensured that each measurement was comparable. MESF beads and NPs were measured with a FSC threshold set at 1.0 while for biological EVs a fluorescence threshold was set at 0.67 by allowing an event rate of 10-20 events/second while acquiring a clean PBS control sample.

When performing quantitative and qualitative analysis of synthetic NPs and biological EVs, preparations were diluted in PBS as indicated. Upon loading on the Influx, the sample was boosted into the flow cytometer until events appeared, after which the system was allowed to stabilize for

30 seconds. Measurements were performed either by a fixed 30 second time or by setting a gate

around the spike-in beads and allowing to record a defined number of events in the gate (80 000

events) using BD FACS Sortware 1.01.654 (BD Biosciences).

The CytoFLEX LX was used without any tailor-made modifications in the configuration. Before

measurements, the manufacturer recommended startup and QC procedure were run first. All scatter

and fluorescence parameters were set to a logarithmic scale. FITC was measured with a 50 mW

488 nm laser and fluorescence was measured through a 525/40 band pass filter at gain 1.0. FITC

MESF beads were recorded with an FSC threshold at 1000. Measurements were performed using

CytExpert 2.1 (Beckman Coulter).

The SORP BD FACSCelesta™ was equipped with a prototype small particle SSC module for

improved scatter detection. Before measurement, the recommended CS&T performance check was

run to monitor performance on a daily basis and to optimize laser delay. All scatter and fluorescence

parameters were set to a logarithmic scale. 100 nm yellow-green (505/515) FluoSphere beads

(Invitrogen, F8803) were acquired and used to set optimal fluorescence (FITC detector) PMT-V

values. FITC was measured with a 100 mW 488 nm laser through a 530/30 band pass filter. FITC-

MESF beads were recorded with an SSC threshold at 200. Measurements were performed using

BD FACSDiva™ Software v8.0.3 (BD Biosciences).

Further descriptions of each instrument and methods are provided in Data S1 (MIFlowCyt

checklist) and Data S2 (MiFlowCyt-EV framework).

Preparation of FITC-doped silica nanoparticles

Synthetic silica nanoparticles (SiNPs) of 550 nm diameter with six different FITC fluorescence

intensities were produced by using a modified method of literature reports [11-13]. Briefly, the

amine reactive FITC molecules were covalently linked to the silane coupling agent, (3-

aminopropyl)-triethoxylsilane (APTES) in anhydrous ethanol. Monodisperse silica seeds of ~90

nm prepared by using amino acid as the base catalyst [11, 12] were suspended in a solvent mixture

containing ethanol, water and ammonia. Then tetraethyl orthosilicate (TEOS) and different

volumes of APTES−FITC solutions were added for growing FITC-doped SiNPs by a modified

Stöber method [13, 14]. Upon washing three times with anhydrous ethanol, the FITC-doped SiNPs

were reacted with TEOS in the solvent mixture to allow growth of a silica layer. The synthesized

SiNPs were washed three times with anhydrous ethanol and stocked in anhydrous ethanol. The

diameters of SiNPs were measured by transmission electron microscopy.

Isolation and fluorescent staining of extracellular vesicles for flow cytometric analysis

EV-containing samples were obtained from 4T1 mouse mammary carcinoma cell culture

supernatants (ATCC, Manassas, VA) as previously described [10, 15, 16]. EVs were stained with

5-(and-6)-Carboxyfluorescein diacetate succinimidyl ester (CFDA-SE, hereinafter referred as

CFSE) (Thermo Fisher, Catalog No. C1157) and separated as described previously [9]. Briefly, 2

µl of the isolated 4T1 EVs (corresponding to a concentration of 1.44 E12 particles/mL as

determined by nanoparticle tracking analysis) were mixed with 18 µl PBS/0.1% aggregate-depleted

(ad)BSA. For antibody labeling, samples were first resuspended in 15.5 µL PBS/0.1% adBSA and

incubated with 0.5 µg of rat anti-mouse CD9-PE (Clone: KMC8, IgG2a, κ, lot 7268877, BD

Biosciences) or matched isotype antibodies (Rat IgG2a, κ, PE-conjugated, lot 8096525, BD

Biosciences) for 1h at RT while protected from light. EVs were then stained with 40 µM CFSE in

a final volume of 40 µl. The sealed tube was incubated for 2h at 37°C while protected from light.

Next, staining was stopped by adding 260 µl PBS/0.1% adBSA. After fluorescent staining, EVs

were separated from protein aggregates and free reagents by bottom-up density gradient

centrifugation in sucrose for 17.30 h at 192,000 g and 4°C using a SW40 rotor (k-factor 144.5;

Beckman Coulter, Fullerton, California, USA). Twelve fractions of 1 mL were then collected from

the top of the gradient and respective densities were determined by refractometry using an Atago

Illuminator (Japan). For analysis by flow cytometry, EV samples corresponding to a 1.14 g/mL

density were diluted 1:20 in PBS prior measurement. MIFlowCyt-EV guidelines [3] were followed

whenever applicable (Data S2).

Concentration determination by using spike-in beads

EV concentration was normalized using a spiked-in external standard containing 200 nm orange

(540/560) fluorescent beads (Invitrogen, F8809). The concentration of the beads was determined

by Flow NanoAnalyzer N30 (NanoFCM, Xiamen, China) and stocked at 5.7E10 particles/mL.

Beads were diluted $1:10^4$ in PBS and added to the EV samples, mixed and measured on the flow

cytometer. Bead count was used to calculate the EV concentration for BD Influx measurements.

Data analysis

For fluorescence calibration each bead peak population was gated using FlowJo Version 10.5.0

and MFI were obtained for further least square linear regression analysis. Data was handled in

Microsoft Excel and figures were prepared using GraphPad Prism version 8.0 (GraphPad Software

Inc). The software FCMPASS Version v2.17 was used to generate files with calibrated axis units

in   the   histograms   and   dot   plots   shown   [17]   (Software   is   available   on

http://go.cancer.gov/a/y4ZeFtA).

Data availability

All EV data of our experiments have been submitted to the EV-TRACK knowledgebase (EV-

TRACK ID: EV210047) [18]. All flow cytometric data files have been deposited at the Flow

Repository (FR-FCM-Z3FJ).

Results

The inclusion of different numbers of bead populations impacts linear regression analysis of

custom-made and commercial MESF-bead sets

We noticed that available calibrator sets differ in the number of fluorescent bead populations,

typically ranging from 3 to 5 populations. Therefore, we first assessed whether the inclusion of

more or less populations impacts the consistency of the slope generated by linear regression

analysis. To this end, we used custom-made standard FITC MESF beads (6 μm) and commercially

available beads from Bangs Labs, both containing five fluorescent bead populations with relatively

similar FITC MESF values (ranging respectively from 7094 to 715225 and 8452 to 1203595).

On the BD Influx, we could detect all 5 populations from brightest to dimmest for both FITC MESF

bead sets when using the EV analysis settings. Next, provided MESF values by the manufacturers

and measured MFI were transformed into log-log and three graphs with three equations were

obtained for each FITC MESF bead set (Figure 1). The graphs represent linear regressions by

including either: (i) all five bead populations, (ii) 4 bead populations (from brightest to second

dimmest) or (iii) 3 bead populations (from brightest to third dimmest). We found that the number

of populations that are used for the subsequent linear regression analysis has an impact on the slope

that belongs to that particular equation (Figure 1). This was observed for both custom-made and

commercially available beads and showed that when fluorescence calibration is performed ideally

the same bead populations should be used for linear regression analysis.

Assessment of different MESF bead sets for the calibration of fluorescent signals across platforms

Two FITC MESF bead sets of 6 μm and 2 μm, containing either five or four bead fluorescent populations were measured using three different flow cytometer platforms, namely a BD Influx, a BC CytoFLEX and a SORP BD FACSCelesta™. For fair comparison between the three instruments, we included 4 bead fluorescent populations from both FITC MESF bead sets, as these were consistently measured on all three platforms. Respective singlet gated populations are displayed as overlays in histograms showing the FITC fluorescence (Figure 2a). The 4 populations in the histograms of both bead sets are indicated from dim to bright fluorescence as p1, p2, p3 and p4 (Figure 2b). The bigger sized 6 μm FITC MESF beads contained overall brighter fluorescent intensities, whose assigned values for the four populations used in the cross-platform comparison range from 25,910 to 715,225 FITC MESF. The smaller sized 2 μm FITC MESF beads covered a dimmer part of the fluorescence intensity range with assigned values ranging from 3,634 to 103,706 FITC MESF. As expected, when MFI values are expressed in arbitrary units, data cannot be directly compared between instruments (Figure 2b). However, when the FITC MFI of every bead population was next correlated to the FITC MESF values provided for each population using least square linear regression, we obtained comparable FITC MESF units (Figure 2c). Importantly, the two bead sets revealed a different slope, consistent within each platform, indicating that this variation was introduced by an inherent attribute of the beads. To evaluate the precision and accuracy of FITC MESF assignments we used specific bead populations from one set, referred to as 'unknown' in Figure 2d-h, to calculate the FITC MESF value using the regression line from the other bead set. The 'unknown' samples used were: (i) p1 and p4 of the 2 μm bead set for which the FITC MESF values were calculated using the regression line of the 6 μm bead set and (ii) p1 and p2 of the 6 μm bead set for which the FITC MESF values were calculated using the regression line

of the 2 μm bead set (Figure 2c). Using this approach, the extrapolated FITC MESF values of the

'unknown' p4 from the 2 μm beads and the 'unknown' p2 from the 6 μm beads could successfully

be compared between platforms, proving the precision of the fluorescence calibration and

facilitating the interpretation of data acquired on the three different platforms (Figure 2d-e and

Figure 2h).

To quantify accuracy, we decided to set a range of 20% difference as acceptable variation of the

assigned values (10% above or 10% below actual values, as indicated by the green band in Figure

2d-e). We found less than 20% difference from the expected value for all calculated FITC MESF

values of the 'unknown' signal from the p4 of the 2 μm beads on the 6 μm bead-set regression line

and the p2 of the 6 μm beads on the 2 μm bead-set regression line. This observation was

independent of the platform and bead set used. Thus, further demonstrating not only precision but

also the accuracy of fluorescence calibration (Figure 2d-e and Figure 2h).

However, when the 'unknown' p1 bead signals of the 2 and 6 μm beads were calculated using

respectively the 6 or 2 μm bead-set regression lines (Figure 2f-g) , we found that although

calculated FITC MESF values were again comparable between platforms and hence calibration

proved to be precise, the accuracy was uncertain. Comparison of the calculated MESF values of

the 'unknown' beads to the expected MESF values in these cases revealed more than a 20%

difference, leading to either an underestimation or an overestimation of the absolute number of

FITC MESF (Figure 2f-h).

This can be explained by the fact that slight differences in the slope of the calibration lines of the

different bead sets become more prominent when extrapolation needs to be extended into the dim

area beyond the fluorescence intensities of the calibration beads themselves (Figure 2f). It is

unlikely that differences in slopes are caused by instrument non-linearity since it is a recurring

pattern on all three platforms (Figure 2c) and linear plotted values do not show non-linearity

(Figure S1a-c). Instrument linearity was further confirmed on the BD Influx following the approach

described by Bagwell et al   [19] (Figure S2). Additionally, variability between separate

measurements could be excluded (Figure S3). Furthermore, by using additional commercially

available calibration beads, slope variability was confirmed to be inherent to the use of calibrator

beads in general (Figure 1) further indicating that this is not a phenomenon restricted to the custom-

made beads only. Analysis of a second commercial set of ERF Reference Particles showed that

only two bead populations could be detected using our EV settings. Although for this specific bead

set additional information about the uncertainty of the fluorescent value of the beads was provided,

the robustness of the linear regression analysis based on only two data points is questionable

(Figure S4). Finally, a similar phenomenon as described for the FITC MESF calibrator sets was

observed when 6 μm and 2 μm calibrator bead sets for PE MESF were used, albeit in this case the

dimmer 2 μm beads displayed a steeper slope than the brighter 6 μm beads (Figure S1d-f).

Altogether, our observations are pointing towards an uncertainty in the MESF assignment of both

the FITC and the PE MESF beads rather than towards instrument or other external factors.

MESF assignments of FITC fluorescence intensities from synthetic silica nanoparticles

depend on the MESF-bead calibrator set used

So far, we have demonstrated that MESF assignment of unknow particles potentially has a higher

chance of being inaccurate when the signal of the unknown particle is significantly higher/lower

than the fluorescence intensities of the used MESF calibrators themselves. We next investigated

how calibration with the FITC MESF bead sets impacts absolute fluorescent assignment when

nanoparticles with even dimmer fluorescence were measured. For this purpose, we focused on one

platform (BD Influx) to measure and analyze 550 nm silica NPs, containing 6 populations with

different FITC fluorescence intensities. Singlets were gated (Figure S5) and an overlay histogram

showing 6 different FITC fluorescence intensities was generated (Figure 3a, left). We next performed fluorescence calibration and generated two additional histograms showing calibrated FITC MESF units based on either the 6 µm (Figure 3a, middle) or the 2 µm FITC MESF calibrator bead set (Figure 3a, right). The obtained MFI and CV values for each population, as well as the calculated FITC MESF values based on the two different calibrator sets are shown in Figure 2b. Our data show that when using the 6 µm calibration beads, the calculated FITC MESF values were always lower as compared to the use of the 2 µm bead set regression line (Figure 3b). In addition, the multi-intensity peak analysis of the silica NPs in one measurement revealed that the difference in the absolute FITC MESF values of these NPs obtained by the two calibrator bead sets increased in the dimmer range of the fluorescence, with a 27.3% variation for the brightest fluorescent peak (p6) and 76.5% variation for the dimmest population of these NPs (Figure 3b). To further visualize this, we used the calculated FITC MESF values (Figure 3b) to generate an overlay bar graph showing the differences in calculated FITC MESF from the six FITC fluorescent populations within the 550 nm NPs (Figure 3c). Our results demonstrate that the variation between using the 6 µm (dark green) or 2 µm (light green) FITC MESF bead set progressively increased from brightest (p6) to dimmest (p1) NP populations (Figure 3c), as expected. This phenomenon can be ascribed to the differences in the slopes of the regression lines of the two bead sets and the extrapolation into the dim area. To confirm that this phenomenon was not limited to the use of FITC MESF beads we additionally analysed 6 µm and 2 µm PE MESF calibrator bead sets on the BD Influx (Figure S1d-f). By measuring a sample containing 200 nm broad spectrum fluorescent polystyrene NPs we could determine its fluorescent intensity in terms of ERF (Figure S6a). The variation in the PE ERF value obtained by the two calibrator sets was in this case 41.3% (Figure S6b), thereby demonstrating that the phenomenon observed with FITC MESF beads also occurs with PE MESF calibrator beads.

Different MESF-bead sets assign different absolute numbers of ERF and MESF to fluorescently CFSE stained and CD9 labeled extracellular vesicles

Next, we investigated the impact of this phenomenon on the assignment of absolute FITC ERF

units and PE MESF units to an EV sample measured on the same platform by using the four

different calibration bead sets, i.e. 6 and 2 μm FITC MESF and the 6 and 2 μm PE MESF beads.

Since EVs scatter less light than beads and NPs, we used a fluorescence threshold to measure the

EVs [9]. To this end, we used CFSE as a luminal dye to fluorescently stain 4T1-derived EVs. We

analyzed the expression of CD9, a tetraspanin enriched on the surface of 4T1-derived EVs, by

using a tetraspanin CD9-PE antibody. As controls, we measured unstained EVs and CFSE stained

EVs with a matching isotype-PE control (Figure 4a). Since 200 nm spike-in beads were added to

the samples for EV concentration measurement, we defined an EV-gate and a spike-in bead gate

in these samples and relevant additional controls (Figure S7a). Our data revealed a 76.6% variation

in the MFI expressed in absolute CFSE ERF units on CFSE stained EV and a 156.9% variation in

the MFI expressed in absolute PE MESF units on CD9-PE labeled EVs when the two different

calibrator bead sets were used on the same platform (Figure 4a-b). The fluorescent threshold value

of 0.67 used on the BD Influx corresponds to an equivalent of 150 FITC MESF based on the 6 μm

beads or 300 FITC MESF based on the 2 μm beads (Figure 4a) and indicates the level of detection

using this fluorescence threshold. Therefore, especially when a fluorescence threshold is used to

measure EVs, accurate calibrators are required to be able to report the fluorescence threshold in

calibrated units. Lastly, in the overlay histograms it is shown how the fluorescence intensities of

the calibrators relate to the dimmer signals generated by CFSE stained and CD9-PE labeled EVs

(Figure 4c-d).

Discussion

The calibration of fluorescence measurements obtained by flow cytometry in terms of MESF or

ERF units is important for robust comparison of data between different instruments and

laboratories. The use of flow cytometry in the expanding EV and NP research field is evolving

very rapidly, where the definition of what and how much can be detected is crucial. Besides the

inter-comparability of data, the use of calibrated values generates awareness about the range of

fluorescence that can be expected for single EVs [3, 20].

In line with previous findings we here show that linear regression curves derived from bright

calibration beads can be used to obtain MESF/ERF values for dim NPs and EVs, allowing inter-

comparability of data when the same calibrator is being used [20]. Furthermore, calibration also

provides information on the detection sensitivity of the instrument and makes it possible to report

the fluorescence threshold in calibrated units. However, we show that the robustness and

reproducibility of a calculated MESF or ERF value is varying upon a series of parameters,

including (1) the number of bead subpopulations that are used from the same calibrator set and (2)

the use of different calibrator beads sets, even when they are from the same manufacturer.

Since earlier studies already pointed out towards variabilities in MESF or ERF assignments

between different manufacturer's [8], we designed our study to explore the usability of bright

calibration bead sets from the same manufacturer for assignment of fluorescence values to dim NPs

and EVs. We found that the calculation of MESF or ERF values for dim submicron sized particles

is substantially affected by the intrinsic variation within the assignment of MESF/ERF numbers to

the bright calibrator beads and that this intrinsic variation also differs within calibrator sets from

the same manufacturer. Consequently, the slope of the regression line generated with a calibrator

set varies with the number of data points used for linear regression within the calibrator bead set

and between different calibrator bead sets. Hence, there is an uncertainty for a particular point to

be assigned a number of MESF/ERF based on the spread of the points [21]. To measure the impact

of this intrinsic uncertainty on the assignment of absolute fluorescence values to dim NPs and EVs

we used two distinct calibrator bead sets assigned by using the same instruments and the same

internal NIST traceable calibrator. Hence following the same manufacturing process. The results

of the two bead sets demonstrate that the two separately obtained regression lines result in different

MESF/ERF values to particles with fluorescent intensities well below the intensities of the

calibrator beads themselves. Due to the further separation of the two regression lines at the lower

end, increasing uncertainties exist in the MESF/ERF assignment for these dim particles. The

currently available commercial MESF beads are designed to calibrate brighter fluorescence

intensities, in the range of fluorescently labeled cells. Furthermore, available calibrators do not

provide uncertainty values around the assignment of the calibrator beads themselves. It is important

to notice that EVs are orders of magnitude smaller than cells and their fluorescent intensities based

on generic staining and/or on antibody labeling are far dimmer than the calibrators. To assign

numbers of fluorophores to these weak signals, we need to rely on extrapolation of the regression

line. In this proof-of-principle study, we show that the intrinsic uncertainty of the number of

assigned fluorescent molecules per calibrator bead introduces a higher variation in the assigned

MESF/ERF numbers when values need to be extrapolated for dim particles.

It is clear from our results with the unknown synthetic NPs and EVs that the outcome of the

MESF/ERF values depends on the number of data points included for linear regression analysis

and the calibrator set used, even if beads are provided by the same manufacturer. Since calibration

is an essential part in the interpretation and inter-comparability of EV data, we therefore advise

that (1) for benchmarking studies, the same number of data points and the same calibrator sets

should be used and (2) that detailed description of calibration materials is included in the

manuscript. To that extend, we propose that not only calibrated fluorescence values need to be

reported but also the calibration bead set used, including all MFI (a.u.) and MESF/ERF values of

the beads themselves. Clearly, a calibrator set with MESF/ERF values closer to the range of

fluorescence intensities of the sample of interest and with a low uncertainty of assignment is highly

preferable. Importantly, novel state-of-the-art flow cytometers that are designed to measure small

particles rely obligately on sub-micron sized beads to perform MESF/ERF calibration. These state-

of-the-art platforms cannot measure the 'standard' 6 μm MESF beads [22]. Therefore, there is an

urgent need for small and dim calibration beads that are especially designed for NPs and EVs to be

widely used in cross-platform and inter-laboratory comparisons.

In summary, our results confirm that calibration of fluorescent signals allows for the comparison

between platforms and provides information on the detection sensitivity of the instrument, as

reported in previous studies [23, 24] but also urge for rigor and reproducibility when fluorescence

calibration is being employed for EV and NP measurements.

In general, for more robust assignments of fluorescence values to NPs and EVs, there is a need for

multi-institutional collaborations (between research labs, companies and metrology institutions,

such as NIST) to produce and validate calibration materials that have low and well-characterized

uncertainty of assigned fluorescent values, ideally allowing for data interpolation and with a size

range that is compatible for all flow cytometer platforms used for analysis of NPs and EVs.

Funding Statement

This research is supported by the European Union's Horizon 2020 research and innovation

programme under the Marie Skłodowska-Curie grant agreement No 722148 and by the National

Natural Science Foundation of China (21934004 and 21627811). E. L. A. is supported by the

European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-

Curie grant agreement No 722148.

Author Contributions

E.L.A. designed and performed experiments, analyzed data and wrote the manuscript. T.B.

performed experiments and gave conceptual advice. L.W. gave technical and conceptual advice.

M.M., Y.T. and X.Y. prepared materials and gave technical advice. G.J.A.A. and M.H.M.W.

supervised the research, designed (performed) experiments and wrote the manuscript. G.J.A.A. and

M.H.M.W. contributed equally as senior author. All authors critically reviewed and edited the

manuscript.

Acknowledgments

The authors would like to thank Prof. An Hendrix (Laboratory of Experimental Cancer Research,

Ghent University, Belgium) for the possibility to prepare and analyze M4T1 derived EV in her lab,

Dr. Joshua A. Welsh (National Cancer Institute, Bethesda, MD) for helpful discussion and Ludo

Monheim (BD Biosciences, Erembodegem, Belgium) for helpful technical support. FITC-MESF

and 2 μm PE-MESF beads were kindly provided by BD Biosciences (prepared by Dr. Majid

Mehrpouyan) as part of the European Union's Horizon 2020 research and innovation programme

under the Marie Skłodowska-Curie grant agreement No 722148.

Declaration of interest disclosure

Tina Van Den Broeck and Majid Mehrpouyan are both employees of BD Biosciences, a business

unit of Becton, Dickinson and Company. During the course of this study, the Wauben research

group, Utrecht University, Faculty of Veterinary Medicine, Department of Biomolecular Health

Sciences and BD Biosciences collaborated as a co-joined partner in the European Union's Horizon

2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement No

722148. Xiaomei Yan declares competing financial interests as a cofounder of NanoFCM Inc., a

company committed to commercializing the nano-flow cytometry (nFCM) technology.

Supporting Information

Additional supporting information can be found online in the corresponding section at the end of

the article.

Data S1. Author Checklist: MIFlowCyt-Compliant Items.

Data S2. MIFlowCyt-EV framework.

Figure S1. Analysis of different MESF bead sets measured on the BD Influx.

Figure S2. Analysis of instrument linearity on the BD Influx.

Figure S3. Evaluation of measurement variability of 2 μm FITC MESF beads measured on the BD

Influx.

Figure S4. Evaluation of commercially available ERF Reference Particles measured on the BD

Influx.

Figure S5. Gating strategy for synthetic silica NPs.

Figure S6. 200 nm fluorescent polystyrene NPs.

Figure S7. Gating strategy for external spiked-in beads and biological EV samples.

ORCID

Estefanía Lozano-Andrés https://orcid.org/0000-0002-7305-0776

Lili Wang https://orcid.org/0000-0003-2456-898X

Xiaomei Yan http://orcid.org/0000-0002-7482-6863

Marca H.M. Wauben https://orcid.org/0000-0003-0360-0311

Ger. J.A. Arkesteijn https://orcid.org/0000-0001-7739-582X

References

1.   Nolan, J.P., *Flow Cytometry of Extracellular Vesicles: Potential, Pitfalls, and Prospects.* Curr Protoc Cytom,
2015. **73**: p. 13 14 1-16.
2.   Lian, H., et al., *Flow Cytometric Analysis of Nanoscale Biological Particles and Organelles.* Annu Rev Anal
Chem (Palo Alto Calif), 2019. **12**(1): p. 389-409.
3.   Welsh, J.A., et al., *MIFlowCyt-EV: a framework for standardized reporting of extracellular vesicle flow
cytometry experiments.* J Extracell Vesicles, 2020. **9**(1): p. 1713526.
4.   Hoffman, R.A., *Standardization, calibration, and control in flow cytometry.* Curr Protoc Cytom, 2005. **Chapter
1**: p. Unit 1 3.
5.   Wang, L. and R.A. Hoffman, *Standardization, Calibration, and Control in Flow Cytometry.* Curr Protoc Cytom,
2017. **79**: p. 1 3 1-1 3 27.
6.   Gaigalas, A.K., et al., *Quantitating Fluorescence Intensity From Fluorophore: Assignment of MESF Values.* J
Res Natl Inst Stand Technol, 2005. **110**(2): p. 101-14.
7.   Wang, L., et al., *Quantitating Fluorescence Intensity From Fluorophores: Practical Use of MESF Values.* J Res
Natl Inst Stand Technol, 2002. **107**(4): p. 339-53.
8.   Hoffman, R.A., et al., *NIST/ISAC standardization study: variability in assignment of intensity values to
fluorescence standard beads and in cross calibration of standard beads to hard dyed beads.* Cytometry A,
2012. **81**(9): p. 785-96.
9.   van der Vlist, E.J., et al., *Fluorescent labeling of nano-sized vesicles released by cells and subsequent
quantitative and qualitative analysis by high-resolution flow cytometry.* Nat Protoc, 2012. **7**(7): p. 1311-26.
10.   Arkesteijn, G.J.A., et al., *Improved Flow Cytometric Light Scatter Detection of Submicron-Sized Particles by
Reduction of Optical Background Signals.* Cytometry A, 2020. **97**(6): p. 610-619.
11.   Yokoi, T., et al., *Periodic arrangement of silica nanospheres assisted by amino acids.* J Am Chem Soc, 2006.
128(42): p. 13664-5.
12.   Yokoi, T., et al., *Mechanism of Formation of Uniform-Sized Silica Nanospheres Catalyzed by Basic Amino Acids.*
Chemistry of Materials, 2009. **21**(15): p. 3719-3729.
13.   Stöber, W., A. Fink, and E. Bohn, *Controlled growth of monodisperse silica spheres in the micron size range.*
Journal of Colloid and Interface Science, 1968. **26**(1): p. 62-69.
14.   Giesche, H., *Synthesis of monodispersed silica powders II. Controlled growth reaction and continuous
production process.* Journal of the European Ceramic Society, 1994. **14**(3): p. 205-214.
15.   Vergauwen, G., et al., *Confounding factors of ultrafiltration and protein analysis in extracellular vesicle
research.* Sci Rep, 2017. **7**(1): p. 2704.
16.   Geeurickx, E., et al., *The generation and use of recombinant extracellular vesicles as biological reference
material.* Nat Commun, 2019. **10**(1): p. 3288.
17.   Welsh, J.A., et al., *FCMPASS Software Aids Extracellular Vesicle Light Scatter Standardization.* Cytometry A,
2019.
18.   Consortium, E.-T., et al., *EV-TRACK: transparent reporting and centralizing knowledge in extracellular vesicle
research.* Nat Methods, 2017. **14**(3): p. 228-232.
19.   Bagwell, C.B., et al., *A simple and rapid method for determining the linearity of a flow cytometer amplification
system.* Cytometry, 1989. **10**(6): p. 689-94.
20.   Welsh, J.A. and J.C. Jones, *Small Particle Fluorescence and Light Scatter Calibration Using FCMPASS Software.*
Curr Protoc Cytom, 2020. **94**(1): p. e79.

21.   Draper, N.R. and H. Smith, *Applied regression analysis*. 1981, New York: Wiley.
22.   Tian, Y., et al., *Protein Profiling and Sizing of Extracellular Vesicles from Colorectal Cancer Patients via Flow*
546   *Cytometry.* ACS Nano, 2018. **12**(1): p. 671-680.
547   23.   Welsh, J.A., J.C. Jones, and V.A. Tang, *Fluorescence and Light Scatter Calibration Allow Comparisons of Small*
548   *Particle Data in Standard Units across Different Flow Cytometry Platforms and Detector Settings.* Cytometry
549   A, 2020. **97**(6): p. 592-601.
550   24.   Nolan, J.P. and S.A. Stoner, *A trigger channel threshold artifact in nanoparticle analysis.* Cytometry A, 2013.
551   **83**(3): p. 301-5.

Main figure legends

Figure 1. Comparison of the inclusion of different data points into linear regression analysis

by using custom-made and commercial FITC MESF bead sets. (a) Least square linear

regression analysis of custom-made 6 μm FITC MESF beads and **(b)** 7 μm Quantum^{TM} FITC-5

MESF beads measured on the BD Influx. Provided FITC MESF for both sets (indicated in the

figure) and measured FITC MFI values were transformed to log and plotted in a log-log fashion.

Graphs including either five, four or three fluorescent populations (from left to right).

Figure 2. Evaluation of two different FITC MESF bead sets for the calibration of fluorescent

intensities across three flow cytometer platforms. (a) Histogram overlays (axis in arbitrary units)

of FITC fluorescent intensity peaks derived from the 6 μm (upper row) or the 2 μm (lower row)

FITC MESF beads **(b)** Table showing the median fluorescence intensity (MFI) statistic derived from each of the fluorescent intensity peaks from dimmer to brighter being expressed in arbitrary units as well as the assigned MESF values (right column). **(c)** Least square linear regression analysis of 6 μm (black circles) and 2 μm (grey triangles) FITC MESF beads. Provided FITC MESF and measured FITC MFI values were transformed to log and plotted in a log-log fashion for the three platforms. **(d-g)** Bar graphs showing the expected absolute FITC MESF and the calculated FITC MESF values for different unknowns as indicated in the figure. The absolute number of FITC MESF was calculated for the indicated fluorescent intensity peak population and compared to the expected value on the three flow cytometers. For comparison, we included a 20% variation from the expected FITC MESF of the population as a reference point (green). **(h)** Table indicating the expected and calculated FITC MESF values for each sample used in the analysis.

Figure 3. MESF bead-based calibration of fluorescence signals from synthetic silica NPs. (a) Histogram overlay showing FITC fluorescence in arbitrary units (a.u.) (left), FITC MESF calibrated axis based on the 6 μm (middle) and 2 μm FITC MESF beads (right) from the six differently FITC-labeled 550 nm NP gated populations. **(b)** Table showing the MFI and CV as well as the calculated FITC MESF values for each of the unknown populations with the percentage of variation between the two calculated reference values. **(c)** Bar graph showing the calibrated FITC MESF values for each of the six differently FITC-labeled populations based on the 6 μm (dark green, generally lower numbers) and 2 μm (light green, generally higher numbers). Bars from 6 μm bead values and bars from 2 μm bead values are presented as overlays.

Figure 4. MESF bead-based calibration of fluorescent signals from biological EV samples. (a) Analysis of EV samples by using a fluorescence threshold. Unstained EVs control (left), CFSE and

isotype-PE stained EVs (middle) and CFSE and CD9-PE stained EVs (right) dot plots showing

CFSE fluorescence Vs PE fluorescence in arbitrary units (a.u.) (upper row) or CFSE ERF Vs PE

MESF calibrated axis based on either the 6 μm (middle row) or 2 μm calibration beads (lower

row). The dashed line in each dot plot indicates the fluorescence threshold value used. Number of

events within the EV gate and MFI values for either CFSE or PE fluorescence (a.u.) are indicated

in the top row. CFSE ERF and PE MESF values are indicated based on the 6 μm (middle row) or

2 μm beads (lower row). **(b)** Table showing the ERF or MESF values obtained after calibration for

the CFSE and CD9-PE stained EVs and the percentage of variation between the use of the 6 μm or

2 μm bead sets. **(c)** Histogram overlays displaying fluorescence in arbitrary units from CFSE

stained EVs (blue) next to the 6 μm or 2 μm FITC-MESF bead set (green). **(d)** Histogram overlays

displaying fluorescence in arbitrary units from CD9-PE labeled EVs (purple) next to the 6 μm or

2 μm PE-MESF bead set (red).

Figure 1

625   **Figure 2**

Figure 3

| Unknown | MFI | CV | Calculated FITC MESF values | | % variation |
| | | | Used as calibrator | | |
			6 μm FITC beads	2 μm FITC beads	
p1	6.35	25.53	337	595	76.5
p2	17.20	17.83	1022	1654	61.8
p3	39.60	10.89	2590	3894	50.3
p4	104.90	6.86	7671	10581	37.9
p5	167.00	5.12	12881	17050	32.4
p6	258.90	5.10	20995	26734	27.3

634   **Figure 4**