# Considerations for MESF-bead based assignment of absolute fluorescence values to nanoparticles and extracellular vesicles by flow cytometry

**Estefanía Lozano-Andrés[1, #], Tina Van Den Broeck[2, #], Lili Wang[3], Majid Mehrpouyan[4], Ye Tian[5], Xiaomei Yan[5], Marca H.M. Wauben*[1, #], Ger J.A. Arkesteijn*[1]**

1 Department of Biomolecular Health Sciences, Faculty of Veterinary Medicine, Utrecht University, Utrecht, The Netherlands

2 BD Biosciences, Erembodegem, Belgium

3 Biosystems and Biomaterials Division, National Institutes of Standards and Technology (NIST), Gaithersburg, MD 20899, USA

4 BD Biosciences, San Jose, CA 95131, USA

5 Department of Chemical Biology, MOE Key Laboratory of Spectrochemical Analysis & Instrumentation, Key Laboratory for Chemical Biology of Fujian Province, College of Chemistry and Chemical Engineering, Xiamen University, Xiamen 361005, People's Republic of China

# TRAIN-EV Marie Skłodowska-Curie Action-Innovative Training Network, train-ev.eu

*Both authors contributed equally

**Corresponding author:**

Dr. G.J.A. Arkesteijn

Department of Biomolecular Health Sciences

Faculty of Veterinary Sciences

Utrecht University

Yalelaan 2, 3584 CM

Utrecht, The Netherlands

T: +31 30 2535451/2535833

E: G.J.A.Arkesteijn@uu.nl

**Abstract**

Flow cytometry is a promising technique to characterize nanoparticles (NPs) and extracellular

vesicles (EVs). However, the majority of reported experiments, use arbitrary units to indicate

fluorescence intensity. This hampers comparison of results from different laboratories and different

platforms. We investigated the advised use of calibrated molecules of equivalent soluble

fluorophores (MESF)-beads for assignment of absolute fluorescence to NPs and EVs. Firstly, we

evaluated the use of two different FITC MESF bead sets as calibrators on three different flow

cytometry platforms (BD Influx™, CytoFLEX LX™ and SORP BD FACSCelesta™). Secondly,

NPs and biological 4T1 mammary carcinoma-EVs were analyzed using the BD Influx and their

fluorescence signals calibrated by using different sets of FITC and PE MESF beads. Although

fluorescence calibration, using bright calibrators designed for cellular flow cytometry, makes inter-

platform comparison possible for fluorescently labeled cells and brightly labeled particles, but the

uncertainty of the currently available calibrators, which are far out of the fluorescence range of the

sub-micron particles, hampers a reliable assignment of absolute MESF numbers based on

extrapolation into the dim fluorescence range. Our results illustrate the need for calibration

materials specifically designed for NPs and EVs to enable a reliable assignment of absolute

fluorescence values in the lower fluorescent ranges.

50
**Key words:** flow cytometry, fluorescence calibration, standardization, MESF, extracellular

vesicles, nanoparticles

**Introduction**

Flow cytometry has attracted the attention of many researchers working in the field of cell-derived

extracellular vesicles (EVs) as a valuable tool for high-throughput single particle-based analysis

[1]. It allows for quantitative multiparametric characterization of subsets of EVs within the

heterogeneous EV-population based on the detection of light scatter signals and fluorescence.

Although flow cytometry is one of the methods of choice for EV analysis, several requirements

need to be met and data need to be reported in a proper way to improve rigor and reproducibility

of flow cytometric EV analysis [2]. In spite of the advancements made in the past years, the vast

majority of EV flow cytometric data is still published in arbitrary units. While expressing

percentages of subpopulations or median fluorescence intensity (MFI) relative to the experimental

controls is a common practice for cell measurements by flow cytometry, this approach is less

desirable for the reliable analysis of the much smaller and dimmer EVs whereby positively-stained

events cannot be discriminated from negative counterparts as a clearly separate population.

Especially for these small and dim particles, the variability between the measured outcome caused

by the diversity in fluidics, optics and electronics between different platforms inevitably will be

reflected in the results reported as arbitrary units of fluorescence. Calibration could make clear

whether a specific instrument can fully resolve or partly detect a certain EV or NP population.

Altogether, the use of standardized units for EV flow cytometric analysis will strongly contribute

to rigor and reproducibility and makes inter-platform and inter-laboratory analysis possible.

Therefore, the three-society EV flow cytometry working group, i.e. ISEV-ISAC-ISTH EV flow

workgroup, has recommended in the recently published MIFlowCyt-EV framework for

standardized reporting of EV flow cytometry experiments, the use of fluorescence calibration [2].

Fluorescence calibration is successfully applied for cell flow cytometry in which arbitrary units are

translated to standardized units of fluorescence that are established by National Institute of

Standards and Technology (NIST), such as molecules of equivalent soluble fluorophores (MESF)

or equivalent number of reference fluorophore (ERF) [3-6]. The currently available calibrator

materials are designed for the calibration of fluorescence of labeled cells and hence are much

brighter in fluorescence when compared to the fluorescence from stained EVs and other NPs. We

here investigated whether regression analysis could be used to assign absolute units of fluorescence

to stained EVs that are far out of the range of these calibrator intensity levels. We evaluated the

performance of the different calibrator beads sets and their inter-platform variation on three

different flow cytometers capable of single EV analysis. Finally, we studied how the use of

different bead sets would affect the calibration of fluorescence signals from synthetic NPs and

biological fluorescently stained and antibody labeled EVs.

**Material and Methods**

**MESF calibration beads**

For calibration of the fluorescence axis in the fluorescein isothiocyanate (FITC) channel we used

two different sets of FITC MESF beads (custom-made, 6 µm lot MM2307 #131-10; #131-8; #130-

6; #130-5 and 2 µm  lot MM2307#156; #159.1; #159.2; #122.3,  BD Biosciences, San Jose, CA).

For calibration of the fluorescence axis in the PE channel, we used two different sets of PE MESF

beads (6 μm, commercial QuantiBrite, Catalog. No. 340495 lot 62981 and 2 μm, custom made, lot

MM2327#153.1; #153.2; #153.4; #153.5; #153.6,  BD Biosciences, San Jose, CA).

The 6 μm FITC MESF beads were prepared by reacting various concentration of FITC with PMMA

Beads (Bangs Labs) in borate buffer at pH 9.2.  The 2 μm FITC beads were prepared by reacting

various concentrations of FITC-BSA (with a FITC/BSA molar ratio of 2) with 2 μm carboxylic

beads (Bangs Labs) using EDC/NHS chemistry. The 2 μm PE beads were made as described above

except various concentrations of PE were used with 2 μm carboxylic beads in EDC/NHS chemistry.

The beads were analyzed on LSR Fortessa™ (BD Biosciences) and their MESF values were

assigned using commercially available MESF beads (Flow Cytometry Standards Corp.) as

calibrators. The FITC ERF values were assigned to both 2 μm and 6 μm beads using a specific lot

of FITC-FC Bead (BD Biosciences) as a calibrator with known ERF value, which has been

assigned by NIST.

**Flow cytometer platforms**

In this study three flow cytometers were used. A jet in air-based BD Influx (BD Biosciences, San

Jose, CA), a BC CytoFLEX LX (Beckman Coulter, Brea, CA) with a cuvette-based system and a

cuvette-based SORP BD FACSCelesta™ (BD Biosciences, San Jose, CA) equipped with a small

particle side scatter (SSC) module.

The BD Influx flow cytometer was modified and optimized for detection of submicron-sized

particles [7]. In brief, FITC was excited with a 488 nm laser (Sapphire, Coherent 200 mW) and

fluorescence was collected through a 530/40 bandpass filter. PE was excited with a 562 nm laser

(Jive, Cobolt 150 mW) and fluorescence was collected through a 585/42 bandpass filter. Optical

configuration of the forward scatter detector was adapted by mounting a smaller pinhole and an

enlarged obscuration bar in order to reduce optical background. This reduced wide-angle FSC

(rwFSC) allowed detection of sub-micron particles above the background based on forward scatter

[7, 8].

Upon acquisition, all scatter and fluorescence parameters were set to a logarithmic scale. To

minimize day to day variations, the BD Influx was standardized at the beginning of each

experiment by running 100 and 200 nm yellow-green (505/515) FluoSphere beads (Invitrogen,

F8803 and F8848). The instrument was aligned until predefined MFI and scatter intensities where

reached with the smallest possible coefficient of variation (CV) for rw-FSC, SSC and fluorescence.

After optimal alignment, PMT settings required no or minimal day to day adjustment and ensured

that each measurement was comparable. MESF beads and NPs were measured with a FSC

threshold set at 1.0 while for biological EVs a fluorescence threshold was set at 0.67 by allowing

an event rate of 10-20 events/second while acquiring a clean PBS sample.

When performing quantitative and qualitative analysis of synthetic NP and biological EV samples,

preparations were diluted in PBS as indicated. Upon loading on the Influx, the sample was boosted

into the flow cytometer until events appeared, after which the system was allowed to stabilize for

30 seconds. Measurements were performed by a fixed 30 seconds time or by setting a gate around

the spike-in beads and allowing to record a defined number of events in the gate (80 000 events)

using BD FACS Sortware 1.01.654 (BD Biosciences).

The CytoFLEX LX was used without any tailor-made modifications in the configuration. Before

measurements, the manufacturer recommended startup and QC procedure were run first. All scatter

and fluorescence parameters were set to a logarithmic scale. FITC was measured with a 50 mW

488 nm laser and fluorescence was measured through a 525/40 band pass filter at gain 1.0. FITC

MESF beads were recorded with a FSC threshold at 1000. Measurements were performed using

CytExpert 2.1 (Beckman Coulter).

The SORP BD FACSCelesta™ was equipped with a small particle SSC module for improved

scatter detection. Before measurement, the standard setup and QC procedures were performed and

the recommended CS&T performance check was run to monitor performance on a daily basis and

to optimize laser delay. All scatter and fluorescence parameters were set to a logarithmic scale. 100

nm yellow-green (505/515) FluoSphere beads (Invitrogen, F8803) were acquired and used to set

optimal small particle-SSC and fluorescence (FITC detector) PMT-V values. FITC was measured

with a 100 mW 488 nm laser through a 530/30 band pass filter. FITC MESF beads were recorded

with a SSC threshold at 200. Measurements were performed using BD FACSDiva™ Software

v8.0.3 (BD Biosciences).

Further descriptions of each instrument and methods are provided in Data S1 (MIFlowCyt

checklist) and Data S2 (MiFlowCyt-EV framework).

**Preparation of FITC-doped silica nanoparticles**

Synthetic silica nanoparticles (SiNPs) of 550 nm diameter with six different FITC fluorescence

intensities were produced by using a modified method of literature reports [9-12]. Briefly, the

amine reactive FITC molecules were covalently linked to the silane coupling agent, (3-

aminopropyl)-triethoxylsilane (APTES) in anhydrous ethanol. Monodisperse silica seeds of ~90

nm prepared by using amino acid as the base catalyst [9, 10] were suspended in a solvent mixture

containing ethanol, water and ammonia. Then tetraethyl orthosilicate (TEOS) and different

volumes of APTES−FITC solutions were added for growing FITC-doped SiNPs by a modified

Stöber method [11, 12]. Upon washing three times with anhydrous ethanol, the FITC-doped SiNPs

were reacted with TEOS in the solvent mixture to allow growth of a silica layer. The synthesized

SiNPs were washed three times with anhydrous ethanol and stocked in anhydrous ethanol. The

diameters of SiNPs were measured by transmission electron microscopy.

**Isolation and fluorescent staining of biological Extracellular Vesicles for flow cytometric analysis**

EV-containing samples were obtained from 4T1 mouse mammary carcinoma cell culture supernatants (ATCC, Manassas, VA) as previously described [8, 13, 14]. EVs were stained with 5-(and-6)-Carboxyfluorescein diacetate succinimidyl ester (CFDA-SE, hereinafter referred as CFSE) (ThermoFisher, catalog number C1157) and separated as described previously [7]. Briefly, 2 µl of the isolated 4T1 EVs (corresponding to a concentration of 1.44 E12 particles/mL as determined by nanoparticle tracking analysis) were mixed with 18 µl PBS/0.1% aggregate-depleted (ad)BSA. For antibody labeling, samples were first resuspended in 15.5 µL PBS/0.1% adBSA and incubated with 0.5 µg of rat anti-mouse CD9-PE (Clone: KMC8, IgG2a, κ, lot 7268877, BD Biosciences) or matched isotype antibodies (Rat IgG2a, κ, PE-conjugated, lot 8096525, BD Biosciences) for 1h at RT while protected from light. EVs were then stained with 40 µM CFSE in a final volume of 40 µl. The sealed tube was incubated for 2h at 37°C while protected from light. Next, staining was stopped by adding 260 µl PBS/0.1% adBSA. After fluorescent staining, EVs were separated from protein aggregates and free reagents by bottom-up density gradient centrifugation in sucrose for 17.30 h at 192,000 g and 4°C using a SW40 rotor (k-factor 144.5; Beckman Coulter, Fullerton, California, USA). Twelve fractions of 1 mL were then collected from the top of the gradient and respective densities were determined by refractometry using an Atago Illuminator (Japan). For analysis by flow cytometry, EV samples corresponding to a 1.14 g/mL density were diluted 1:20 in PBS prior measurement. MIFlowCyt-EV guidelines [2] were followed whenever applicable (Data S2).

**Concentration determination by using spike-in beads**

EV concentration was normalized using a spiked-in external standard containing 200 nm orange

(540/560) fluorescent beads (Invitrogen, F8809). The concentration of the beads was determined

by Flow NanoAnalyzer N30 (NanoFCM, Xiamen, China) and stocked at 5.7E10 particles/mL.

Beads were diluted $1:10^4$ in PBS and added to the EV samples, mixed and measured on the flow

cytometer. Bead count was used to calculate the EV concentration.

**Data analysis**

For fluorescence calibration each bead peak population was gated using FlowJo Version 10.5.0

and MFI were obtained for further least square linear regression analysis. Data was handled in

Microsoft Excel and figures were prepared using GraphPad Prism version 8.0 (GraphPad Software

Inc). The software FCMPASS Version v2.17 was used to generate files with calibrated axis units

in  the  histograms  and  dot  plots  shown  [15]  (Software  is  available  on

http://go.cancer.gov/a/y4ZeFtA).

**Data availability**

All EV data of our experiments have been submitted to the EV-TRACK knowledgebase (EV-

TRACK ID: EV210047)[16]. All flow cytometric data files have been deposited at the Flow

Repository (FR-FCM-Z3FJ).

**Results**

**Cross-platform assessment of MESF-bead based calibration of fluorescent signals using**

**different bead sets**

Two differently-sized FITC MESF bead sets of 6 μm and 2 μm, each containing 4 bead populations

per set and respectively representing an overall brighter (6 μm) and an overall dimmer (2 μm)

beads set, were measured and their use for fluorescence calibration of FITC fluorescent signals was

evaluated using three different flow cytometer platforms, namely a BD Influx, a BC CytoFLEX

and a SORP BD FACSCelesta. Every bead population measured on each of these three platforms

was gated and respective singlet gated populations are displayed as overlays in histograms (Figure

1a) showing their FITC fluorescence. The 4 peaks in the histograms of both the 6 μm and 2 μm

beads are indicated from dim to bright fluorescent as p1, p2, p3 and p4 (Figure 1b). The bigger

sized 6 μm FITC MESF beads contained overall brighter fluorescent intensities, whose assigned

values range from 25,910 to 715,225 FITC MESF.  The smaller sized 2 μm FITC MESF beads

covered a dimmer part of the fluorescence intensity range with assigned values ranging from 3,634

to 103,706 FITC MESF. As expected, Figure 1b shows that MFI values alone expressed in arbitrary

units do not allow the comparison of measurements across the three instruments.

Therefore the FITC MFI of every bead population was next correlated to the FITC MESF values

provided for each bead population using least square linear regression. We found a different slope

in the two lines generated by either the 6 μm or the 2 μm MESF beads, consistent within each

platform and indicating that this is an inherent quality of these specific bead sets (Figure 1c). To

evaluate the precision and accuracy of FITC MESF assignments we used specific bead populations

from one set, referred to as 'unknown' in Figure 1, to calculate the FITC MESF by extrapolation

of the regression line based on the other bead set. The 'unknown' samples used where: 1) p1 and

p4 of the 2 μm bead set for which the FITC MESF values were calculated using the regression line

of the 6 μm bead set and 2) p1 and p2 of the 6 μm bead set for which the FITC MESF values were

calculated using the regression line of the 2 μm bead set (Figure 1c). Using this approach, the

extrapolated FITC MESF values of the 'unknown' p4 from the 2 μm beads and the 'unknown' p2

from the 6 μm beads could successfully be compared between platforms, proving the precision of

the fluorescence calibration and facilitating the interpretation of data acquired on the three different

platforms (Figure 1d-e and Figure 1h).

To quantify accuracy we decided to set a range of 20% difference as acceptable (10% above or

10% below actual values, as indicated by the green band in Figure 1d-e) for proper assignment to

MESF values. We found less than 20% difference from the expected value for all calculated FITC

MESF values of the 'unknown' signal from the p4 of the 2 μm beads on the 6 μm bead-set

regression line and the p2 of the 6 μm beads on the 2 μm bead-set regression line. This observation

was consistent for all platforms and independent of the bead set that was used. Thus, further

demonstrating not only precision but also the accuracy of fluorescence calibration (Figure 1d-e and

Figure 1h).

However, when extrapolating the 'unknown' bead signal from p1 of the 2 μm beads on the 6 μm

bead-set regression line and from p1 of the 6 μm beads on the 2 μm bead-set regression line (Figure

1f-g) we found that although calculated FITC MESF values were again comparable between

platforms and hence calibration proved to be precise, the accuracy was uncertain. Comparison of

the calculated MESF values of the 'unknown' beads to the expected MESF values in these cases

revealed more than a 20% difference, leading to either an underestimation or an overestimation of

the absolute number of FITC MESF (Figure 1f-h).

This can be explained by the fact that slight differences in the slope of the calibration lines of the

different bead sets become more prominent when the extrapolation is extended into the dimmer or

brighter range of the fluorescence intensities of the calibration beads, so out of the range of the

calibrators. Importantly, the same trend was observed in all three platforms evaluated. It is unlikely

that differences in slopes are caused by instrument non-linearity since it is a recurrent pattern on

all three platforms (Figure 1) and linear plotted values do not show non-linearity either (Figure

S1a-c). Also, variability between separate measurements could be excluded (Figure S2). Finally,

the same phenomenon occurred when 6 µm and 2 µm calibrator bead sets for PE MESF were used. In this case the opposite situation occurred; the dimmer 2 µm beads displayed a steeper slope than the brighter 6 µm beads (Figure S1d-f). Hence, our observations are pointing towards an uncertainty in the assignment of the both FITC and PE beads.

**Figure 1. Evaluation of two different FITC MESF bead sets for the calibration of fluorescent intensities across three flow cytometer platforms. (a)** Histogram overlays (axis in arbitrary units) of FITC fluorescent intensity peaks derived from the 6 μm or the 2 μm FITC MESF beads **(b)** Table showing the median fluorescence intensity (MFI) statistic derived from each of the fluorescent intensity peaks from dimmer to brighter being expressed in arbitrary units. **(c)** Least square linear regression analysis of 6 μm (black circles) and 2 μm (grey triangles) FITC MESF beads. Provided FITC MESF and measured FITC MFI values were transformed to log and plotted in a log-log fashion for the three platforms. **(d-g)** Bar graphs showing the expected absolute FITC MESF and the calculated FITC MESF values for different unknowns as indicated in the figure. The absolute number of FITC MESF was calculated for the indicated fluorescent intensity peak population and compared to the expected value on the three flow cytometers. For comparison, we included a 20% variation from the expected FITC MESF of the population as a reference point (green). **(h)** Table indicating the expected and calculated FITC MESF values for each sample used in the analysis.

## MESF assignments of FITC fluorescence intensities from synthetic silica nanoparticles depend on the MESF-bead calibrator set used

So far we have demonstrated that accuracy of assignment is negatively influenced when signal strength of the unknown particle is further away from the average fluorescence intensities of the calibrator beads. We next investigated how calibration with the FITC MESF bead-sets impact absolute fluorescent assignment when nanoparticles with even dimmer fluorescence were measured. For this purpose we selected one platform (BD Influx) to measure and analyze 550 nm silica NPs, containing 6 populations with different FITC fluorescence intensities. Singlets were gated (Figure S3) and an overlay histogram showing their 6 different FITC fluorescence intensities was generated (Figure 2a, left). We next performed fluorescence calibration and generated two additional histograms showing calibrated FITC MESF units based on either the 6 μm (Figure 2a, middle) or the 2 μm FITC MESF calibrator bead set (Figure 2a, right). We obtained the MFI and

CV values for each population and further calculated their FITC MESF values based on the two

different FITC MESF calibrator bead sets (Figure 2b). Our data show that when using brighter 6

µm calibration beads, the calculated FITC MESF values were always lower as compared to the use

of the 2 µm bead-set regression line (Figure 2b). In addition, the multi-intensity peak analysis of

the silica NPs in one measurement revealed that the difference in the absolute FITC MESF values

of the synthetic silica NPs obtained by the two calibrator bead sets increased in the dimmer range

of the fluorescence, with a 27.3% variation for the highest fluorescent peak (p6) and 76.5%

variation for the dimmest population of these NPs (Figure 2b). To further visualize this, we used

the calculated FITC MESF values (Figure 2b) to generate an overlay bar graph showing the

differences in calculated FITC MESF from the six FITC fluorescent populations within the 550 nm

NPs (Figure 2c). Our results demonstrate that the variation between using the 6 µm (dark green) or

2 µm (light green) FITC MESF bead set progressively increased from brightest (p6) to dimmest

(p1) NP populations (Figure 2c). As indicated before, this can be ascribed to the differences in the

slopes of the regression lines of the two different bead sets and the extrapolation into the dim area.

To confirm that this phenomenon was not limited to the use of FITC MESF beads we additionally

analysed 6 µm and 2 µm PE MESF calibrator bead sets on the BD Influx (Figure S1d-f). By

measuring a sample containing 200 nm broad spectrum fluorescent polystyrene NPs we could

determine its fluorescent intensity in terms of ERF (Figure S4a). The variation in the absolute PE

ERF value obtained by the two different calibrator sets used to assign the absolute calibrated ERF

value was in this case 41.3% (Figure S4b).

**b**

|  | | | Calculated FITC MESF values | | |
|  | | | Used as calibrator | | |
| Unknown | MFI | CV | 6 µm FITC beads | 2 µm FITC beads | % variation |
|---|---|---|---|---|---|
| p1 | 6.35 | 25.53 | 337 | 595 | 76.5 |
| p2 | 17.20 | 17.83 | 1022 | 1654 | 61.8 |
| p3 | 39.60 | 10.89 | 2590 | 3894 | 50.3 |
| p4 | 104.90 | 6.86 | 7671 | 10581 | 37.9 |
| p5 | 167.00 | 5.12 | 12881 | 17050 | 32.4 |
| p6 | 258.90 | 5.10 | 20995 | 26734 | 27.3 |

**Figure 2. MESF bead-based calibration of fluorescence signals from synthetic silica NPs. (a)**

Histogram overlay showing FITC fluorescence in arbitrary units (a.u.) from each population next to FITC

MESF calibrated axis based on the 6 µm or 2 µm FITC MESF beads from the six differently FITC-labeled

550 nm NP gated populations. **(b)** Table showing the MFI and CV of each population and the calculated

FITC MESF values for each of the unknown populations with the percentage of variation between the two

calculated reference values. **(c)** Bar graph showing the calibrated FITC MESF values for each of the six

differently FITC-labeled populations based on the 6 µm (dark green, generally lower numbers) and 2 µm

(light green, generally higher numbers). Bars from 6 µm bead values and bars from 2 µm bead values are

presented as overlays.

**Different MESF-bead sets assign different absolute numbers of ERF and MESF to**

**fluorescently CFSE stained and CD9 labeled extracellular vesicles**

Next, we investigated the impact of this phenomenon on the assignment of absolute FITC-ERF

units and PE MESF units to a biological EV sample measured on the same platform by using the

four different calibration bead sets, i.e. 6 and 2 µm FITC MESF and the 6 and 2 µm PE MESF

beads. Since EVs scatter less light than beads and NPs, we used a fluorescence threshold to measure

the biological EVs [7]. To this end, we used CFSE as a luminal dye to fluorescently stain EVs and

a tetraspanin CD9 antibody, since this tetraspanin transmembrane protein is known to be enriched

in the surface of EVs and is considered as EV-marker. Next to the CFSE stained and CD9-PE

labeled EVs, we measured an unstained control containing EVs but without CFSE or antibody and

a matching isotype control (Figure 3a). Since 200 nm spike-in beads were added to the samples for

EV concentration measurement we defined an EV-gate and a spike-in bead gate in these samples

and relevant additional controls (Figure S5a). Our data revealed a 76.6% variation in the MFI

expressed in absolute CFSE ERF units on CFSE stained EV and a 156.9% variation in the MFI

expressed in absolute PE MESF units on CD9-PE labeled EVs when the two different calibrator

bead sets were used on the same platform (Figure 3a-b). The fluorescent threshold value of 0.67

used on the BD Influx corresponds to an equivalent of 150 FITC MESF based on the 6 µm beads

or 300 FITC MESF based on the 2 µm beads (Figure 3a) and indicates the level of detection using

this fluorescence threshold. Therefore, especially when a fluorescence threshold is used to measure

EVs, accurate calibrators are required to be able to report the fluorescence threshold in calibrated

units.

Lastly, in the overlay histograms it is shown how the fluorescence intensities of the calibrators

relate to the dimmer signals generated by CFSE stained and CD9-PE labeled EVs (Figure 3c-d).

**Figure 3. MESF bead-based calibration of fluorescence signals from biological EV samples. (a)**

Analysis of three biological EV samples by using a fluorescence threshold. Unstained EVs control, CFSE

+ isotype-PE EVs next to CFSE + CD9-PE stained EVs (from left to right plots) with dot plots showing

CFSE fluorescence vs PE fluorescence in arbitrary units (a.u.) (upper row) or CFSE ERF vs PE MESF

calibrated axis based on either the 6 µm (middle row) or 2 µm beads (lower row). The discontinuous line in

each dot plot indicates the fluorescence threshold value used, which is 0.67 and on this BD Influx

corresponds to an equivalent of 150 FITC MESF based on the 6 µm beads or 300 FITC MESF based on the

2 µm beads. Number of events for each sample within the EV gate and MFI values for either CFSE or PE

fluorescence (a.u.) are indicated in the top row. Once calibrated dot plots were generated the median

fluorescence intensity in either CFSE ERF or PE MESF standardized units of fluorescence were obtained

based on the 6 µm (middle row) or 2 µm beads (lower row) and are indicated for each dot plot. **(b)** Table

showing the ERF or MESF values obtained after calibration for the CFSE + CD9-PE EVs and the percentage

of variation between the use of the 6 µm or 2 µm bead sets. **(c)** Histogram overlays displaying fluorescence

in arbitrary units from CFSE stained EVs (blue) next to the 6 µm or 2 µm FITC MESF bead set (green). **(d)**

Histogram overlays displaying fluorescence in arbitrary units from CD9-PE labeled EVs (purple) next to

the 6 µm or 2 µm PE MESF bead set (red).

**Discussion**

The calibration of fluorescence measurements obtained by flow cytometry in terms of MESF or

ERF is essential for proper and robust comparison of data between different instruments, different

labs and different time points. EV flow cytometry is pursued as a tool in the rapidly expanding EV

research field and the definition of what and how much can be detected is only starting to evolve.

Besides the inter-comparability of EV data, the use of calibrated values also allows to create

consensus about the range of fluorescence that can be expected by EV flow cytometry in case of

generic staining and epitope density, thereby facilitating the further identification of EV in flow

cytometry data. Therefore, it is desirable to express flow cytometry data from EVs in calibrated

units as well [2]. Our results clearly illustrate why it is important to perform fluorescence

calibration, as previously reported in other studies [17, 18], since MFI values in arbitrary units

cannot be directly compared when samples are measured on different flow cytometers.

Our findings show that extrapolating bright reference beads to dim signals is possible and calibrated

values are precise irrespective the instrument used, allowing standard comparisons, which is in line

with earlier findings [18]. However, the accuracy of assigning absolute MESF or ERF values relies

for a great deal on the exactness by which the calibrator beads are assigned. The effect of small

variations in assignment may cause uncertainty in the slopes of the regression lines. The uncertainty

in the slope affects the uncertainty for a particular point based on the spread of the points. Since

the points at the ends (low intensity end) are farthest away, the uncertainty increases with the

distance to the other points on average [19]. On the other hand, difference in slope between two

calibrator sets per definition denotes an interception point somewhere along the line. Comparability

may fully depend on the position of the interception point. In our study the interception point is

close to the range where the beads reside and goes hand in hand with precision and accuracy for

values in this range. However away from this point, larger uncertainties occur due to the further

separation of the two regression lines. The currently available commercial MESF beads are

designed to calibrate fluorescence intensity in the range of regular cell flow cytometry. However,

EVs are orders of magnitude smaller than cells and their fluorescence based on generic staining or

based on antibody labeling and epitope density is far dimmer than the available calibrators. To

assign numbers of fluorophores to these weak signals, we rely on extrapolation of the regression

line into the dim area. In this proof-of-principle study, we show that the intrinsic uncertainty of the

number of assigned fluorescent molecules per calibrator bead induces a higher deviation from the

expected value of MESF when the regression line is followed into the area of the dimmer particles.

It may be clear from our results with the unknown synthetic NPs and the biological EVs that the

outcome of the results definitively depends on the calibrator used. Since calibration is an essential

part in the identification and inter-comparability of EV data, we therefore advise that proper

reporting of how fluorescent calibration is being done and adequate description of materials needs

to be included to avoid confusing interpretation of calibrated data. Hence, we propose that not only

absolute fluorescence values need to be reported but also the calibration bead set and lot number

used, including all MFI and MESF values. Furthermore, for future benchmarking and comparison

studies, ideally the same calibrator sets should be used. Clearly, a calibrator that has a low

uncertainty of assignment and has MESF values closer to the range of fluorescence intensity of the

unknown particle of interest is most preferable and determination of absolute MESF/ERF values

to NPs and EVs needs to be further investigated and validated before robust conclusions can be

made. Moreover, novel state-of-the-art flow cytometers that are designed to measure small particles

rely obligately on sub-micron beads to perform MESF calibration. These state-of-the-art platforms

cannot measure the 'standard' 6 µm or 2 µm MESF beads [20]. Therefore, there is an urgent need

for small and dim calibration beads that are especially designed for NPs and EVs to be widely used

in cross-platform and inter-laboratory comparisons.

In summary, our results indicate that calibration of fluorescent signals is crucial for comparison

between platforms, and that it should be utilized and properly reported. Nevertheless, accurate

assignment of MESF/ERF to particles in the sub-micron size range could be hampered by the

inherent uncertainty of the calibrators. Therefore, statements on the absolute number of

fluorescence units, either MESF or ERF, present on NPs and EVs are related to the calibrator set

used. To avoid inconsistencies, so far calibration for inter-comparison may preferably be performed

using identical calibrator bead lots. For more robust assignments of fluorescence units to NPs and

EVs there is a need for calibration materials that have low uncertainty of assigned fluorescent

values and have a size range that is compatible for all flow cytometer platforms used for analysis

of sub-micron sized particles, including both NPs and EVs.

**Funding Statement**

This research is supported by the European Union's Horizon 2020 research and innovation
programme under the Marie Skłodowska-Curie grant agreement No 722148 and by the National
Natural Science Foundation of China (21934004 and 21627811). E.L.A. is supported by the
European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-
Curie grant agreement No 722148.

**Author Contributions**

E.L.A. designed and performed experiments, analyzed data and wrote the manuscript. T.B.
performed experiments and gave conceptual advice. L.W. gave technical and conceptual advice.
M.M., Y.T. and X.Y. prepared materials and gave technical advice. G.J.A.A. and M.H.M.W.
supervised the research, designed (performed) experiments and wrote the manuscript. G.J.A.A. and
M.H.M.W. contributed equally as senior author. All authors critically reviewed and edited the
manuscript.

**Acknowledgments**

The authors would like to thank Prof. An Hendrix (Laboratory of Experimental Cancer Research,
Ghent University, Belgium) for the possibility to prepare and analyze M4T1 derived EV in her lab,
Dr. Joshua A. Welsh (National Cancer Institute, Bethesda, MD) for helpful discussion and Ludo
Monheim (BD Biosciences, Erembodegem, Belgium) for helpful technical support. FITC MESF
and 2 μm PE MESF beads were kindly provided by BD Biosciences (prepared by Dr. Majid
Mehrpouyan) as part of the European Union's Horizon 2020 research and innovation programme
under the Marie Skłodowska-Curie grant agreement No 722148.

**Declaration of interest disclosure**

Tina Van Den Broeck and Majid Mehrpouyan are both employees of BD Biosciences, a business

unit of Becton, Dickinson and Company. During the course of this study, the Wauben research

group, Utrecht University, Faculty of Veterinary Medicine, Department of Biomolecular Health

Sciences and BD Biosciences collaborated as a co-joined partner in the European Union's Horizon

2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement No

722148. Xiaomei Yan declares competing financial interests as a cofounder of NanoFCM Inc., a

company committed to commercializing the nano-flow cytometry (nFCM) technology.

**Supporting Information**

Additional supporting information can be found online in the corresponding section at the end of

the article.

**Data S1.** Author Checklist: MIFlowCyt-Compliant Items.

**Data S2.** MIFlowCyt-EV framework.

**Figure S1.** Analysis of different MESF bead sets measured on the BD Influx.

**Figure S2.** Evaluation of measurement variability of 2 μm FITC MESF beads measured on the BD

Influx.

**Figure S3.** Gating strategy for synthetic silica NPs.

**Figure S4.** 200 nm fluorescent polystyrene NPs.

**Figure S5.** Gating strategy for external spiked-in beads and biological EV samples.

**ORCID**

Estefanía Lozano-Andrés https://orcid.org/0000-0002-7305-0776

Lili Wang https://orcid.org/0000-0003-2456-898X

Xiaomei Yan http://orcid.org/0000-0002-7482-6863

Marca H.M. Wauben https://orcid.org/0000-0003-0360-0311

Ger. J.A. Arkesteijn https://orcid.org/0000-0001-7739-582X

## References

1.    Nolan, J.P., *Flow Cytometry of Extracellular Vesicles: Potential, Pitfalls, and Prospects.* Curr Protoc Cytom,
2015. **73**: p. 13 14 1-16.
2.    Welsh, J.A., et al., *MIFlowCyt-EV: a framework for standardized reporting of extracellular vesicle flow
cytometry experiments.* J Extracell Vesicles, 2020. **9**(1): p. 1713526.
3.    Hoffman, R.A., *Standardization, calibration, and control in flow cytometry.* Curr Protoc Cytom, 2005. **Chapter
1**: p. Unit 1 3.
4.    Wang, L. and R.A. Hoffman, *Standardization, Calibration, and Control in Flow Cytometry.* Curr Protoc Cytom,
2017. **79**: p. 1 3 1-1 3 27.
5.    Gaigalas, A.K., et al., *Quantitating Fluorescence Intensity From Fluorophore: Assignment of MESF Values.* J
Res Natl Inst Stand Technol, 2005. **110**(2): p. 101-14.
6.    Wang, L., et al., *Quantitating Fluorescence Intensity From Fluorophores: Practical Use of MESF Values.* J Res
Natl Inst Stand Technol, 2002. **107**(4): p. 339-53.
7.    van der Vlist, E.J., et al., *Fluorescent labeling of nano-sized vesicles released by cells and subsequent
quantitative and qualitative analysis by high-resolution flow cytometry.* Nat Protoc, 2012. **7**(7): p. 1311-26.
8.    Arkesteijn, G.J.A., et al., *Improved Flow Cytometric Light Scatter Detection of Submicron-Sized Particles by
Reduction of Optical Background Signals.* Cytometry A, 2020. **97**(6): p. 610-619.
9.    Yokoi, T., et al., *Periodic arrangement of silica nanospheres assisted by amino acids.* J Am Chem Soc, 2006.
**128**(42): p. 13664-5.
10.   Yokoi, T., et al., *Mechanism of Formation of Uniform-Sized Silica Nanospheres Catalyzed by Basic Amino Acids.*
Chemistry of Materials, 2009. **21**(15): p. 3719-3729.
11.   Stöber, W., A. Fink, and E. Bohn, *Controlled growth of monodisperse silica spheres in the micron size range.*
Journal of Colloid and Interface Science, 1968. **26**(1): p. 62-69.
12.   Giesche, H., *Synthesis of monodispersed silica powders II. Controlled growth reaction and continuous
production process.* Journal of the European Ceramic Society, 1994. **14**(3): p. 205-214.
13.   Vergauwen, G., et al., *Confounding factors of ultrafiltration and protein analysis in extracellular vesicle
research.* Sci Rep, 2017. **7**(1): p. 2704.
14.   Geeurickx, E., et al., *The generation and use of recombinant extracellular vesicles as biological reference
material.* Nat Commun, 2019. **10**(1): p. 3288.
15.   Welsh, J.A., et al., *FCMPASS Software Aids Extracellular Vesicle Light Scatter Standardization.* Cytometry A,
2019.
16.   Consortium, E.-T., et al., *EV-TRACK: transparent reporting and centralizing knowledge in extracellular vesicle
research.* Nat Methods, 2017. **14**(3): p. 228-232.
17.   Nolan, J.P. and S.A. Stoner, *A trigger channel threshold artifact in nanoparticle analysis.* Cytometry A, 2013.
**83**(3): p. 301-5.
18.   Welsh, J.A., J.C. Jones, and V.A. Tang, *Fluorescence and Light Scatter Calibration Allow Comparisons of Small
Particle Data in Standard Units across Different Flow Cytometry Platforms and Detector Settings.* Cytometry
A, 2020. **97**(6): p. 592-601.
19.   Draper, N.R. and H. Smith, *Applied regression analysis*. 1981, New York: Wiley.
20.   Tian, Y., et al., *Protein Profiling and Sizing of Extracellular Vesicles from Colorectal Cancer Patients via Flow
Cytometry.* ACS Nano, 2018. **12**(1): p. 671-680.