1

MEYE: Web-app for translational and real-time pupillometry

Raffaele Mazziotti[1,2,$,*], Fabio Carrara[5,$], Aurelia Viglione[2,4], Leonardo Lupori[2,4], Luca Lo Verde[2], Alessandro Benedetto[2], Giulia Ricci[1], Giulia Sagona[2,3,], Giuseppe Amato[5], Tommaso Pizzorusso[1,2,4]

1. Department of Neuroscience, Psychology, Drug Research and Child Health NEUROFARBA University of Florence, Area San Salvi – Pad. 26, 50135 Florence, Italy
2. Institute of Neuroscience, National Research Council, Via Moruzzi, 1 56124 Pisa, Italy
3. Department of Developmental Neuroscience, IRCCS Stella Maris Foundation, 56128 Pisa, Italy
4. BIO@SNS lab, Scuola Normale Superiore via G. Moruzzi, 1 56124 Pisa, Italy
5. ISTI – Istituto di Scienza e Tecnologia dell'Informazione, Via G. Moruzzi, 1 – 56124 – Pisa (PI) – Italy

* Corresponding author: Raffaele Mazziotti, Istituto Neuroscienze CNR, Via G. Moruzzi,1 56125 Pisa ITALY, tel +390503153167, Fax +390503153220 e-mail: raffaele.mazziotti@in.cnr.it

$ These authors contributed equally to this work

Abstract

Pupil dynamics alterations have been found in patients affected by a variety of neuropsychiatric conditions, including autism. Studies in mouse models have used pupillometry for phenotypic assessment and as a proxy for arousal. Both in mice and humans, pupillometry is non-invasive and allows for longitudinal experiments supporting temporal specificity, however its measure requires dedicated setups. Here, we introduce a Convolutional Neural Network that performs on-line pupillometry in both mice and humans in a web app format. This solution dramatically simplifies the usage of the tool for non-specialist and non-technical operators. Because a modern web browser is the only software requirement, this choice is of great interest given its easy deployment and set-up time reduction. The tested model performances indicate that the tool is sensitive enough to detect both spontaneous and evoked pupillary changes, and its output is comparable with state-of-the-art commercial devices.

Keywords

pupillometry, Convolutional Neural Network, Pupil Diameter, Arousal, U-net, Web App; Oddball, Eyelink, MEYE

Introduction

Pupillometry, the measurement of pupil size fluctuations over time, provides useful insights in clinical settings and basic research activity. Light level is the primary determinant of pupil size, even though non-light-driven pupil fluctuations, widely assumed as an indicator of arousal through Locus Coeruleus (LC) activity, can be used to index brain state across species [1–3]. Higher cognitive and emotional processes are also able to evoke tonic or phasic pupillary changes, such as attention [4], memory load [5], novelty [6–8], pain [9–11], and more general cortical sensory processing [2,12] in humans and in animal models.

A growing body of work shows how pupillometry can be used as a possible biomarker for numerous neurological and psychiatric conditions in early development and adult subjects [13–

27]. Spontaneous and voluntary modulation of pupil fluctuations have also been used to

facilitate Human-Computer Interaction in normal subjects [28–30] and patients with severe

motor disabilities. For example, pupil dynamics is used to assess communication capability in

Locked-in Syndrome, a crucial factor for the determination of a minimally conscious state

[31,32]. Pupillometry is also becoming a valuable tool for child neurology, to facilitate risk

assessment in infants. For example, the Pupil Light Reflex (PLR) during infancy seems to

predict the later diagnosis and severity of Autism Spectrum Disorders (ASD) [27]. Intriguingly,

pupil alterations are also present in several ASD mouse models [26].

Pupillometry has several advantages as compared with other physiological methods: it

is non-invasive and can be performed by non-specialized personnel on non-collaborative and

preverbal subjects (like infants), allowing the design of longitudinal experiments to permit

temporal specificity. More importantly, it can be conducted similarly across different species

from mice to humans, guaranteeing maximal translatability of the protocols and results

[13,16,26]. Given these assumptions, it is vital to introduce a simple, versatile tool used in a

range of settings, from the laboratory to the clinical or even domestic environment. Available

open source methods require complicated steps for the installation and configuration of custom

software not suitable for non-technical operators. Moreover, these tools were tested exclusively

in one species (mice [33], humans [34]), and none of them were applied in cognitive

experiments that usually involve small pupil changes associated with high variability.

In this work, we have developed a deep learning tool called MEYE, using convolutional

neural networks (CNNs) to detect and measure real-time changes in pupil size both in humans

and mice in different experimental conditions. Furthermore, MEYE web app, performs pupil area

quantification and blink detection, all within a single network. By embedding artificial intelligence

algorithms in a web browser to process real-time webcam streams or videos of the eye, MEYE

can be used by non-technical operators, opening the possibility to perform pupillometry widely,

cost-effectively, and in a high-throughput manner. This architecture is resistant to different

illumination conditions, allowing the design of basic neuroscience experiments in various

experimental settings, such as behavior coupled with electrophysiology or imaging like 2-photon

microscopy. To describe the performances of MEYE web app in different settings, we tested the

app in both mice and humans. In mice we recorded both running speed and pupil size during

auditory stimulation. In humans we tested MEYE capabilities to detect the PLR. Furthermore,

we performed a visual oddball paradigm [35–37], comparing pupil size and eye position

measurements obtained from MEYE with one of the most used commercial eye-tracker

systems: the EyeLink 1000. Finally we released a dataset of more than 11897 eye images that

can be used to train other artificial intelligence tools.

Methods

Datasets

For this study, we collected a dataset (Fig. 1 A) composed of 11897 grayscale images of humans (4285) and mouse (7612) eyes. The pictures' majority depicts mouse eyes during head-fixation sessions (HF: 5061) in a dark environment using infrared (IR, 850 nm) light sources. In this environment, the pupil is darker than the rest of the image. We also collected mouse eyes (2P: 2551) during 2-photon Ca2+ imaging. In this particular condition, the pupil is inverted in color and tends to be brighter than the iris. Finally, we acquired images of human eyes in IR light (H: 4285) during virtual reality experiments (wearing a headset for virtual reality), using an endoscopic camera (www.misumi.com.tw/). The dataset contains 1596 eye blinks, 841 images in the mouse, and 755 photos in the human datasets. Five human raters segmented the pupil in all pictures (one per image), using custom labeling scripts implemented in Matlab or Python, by manual placement of an ellipse or polygon over the pupil area. Raters flagged blinks using the same code.

Fig.1: Dataset, CNN architecture, and performances. **A:** examples of images taken from the dataset. The

first image depicts a head-fixed mouse with dark pupils, the second one is a head-fixed mouse with a

bright pupil, during 2-photon microscope sessions. The last image is a human eye taken during

experiments wearing virtual reality goggles. **B:** 64 examples of data augmentation fed to CNN. The

images are randomly rotated, cropped, flipped (horizontally or vertically), and changed in

brightness/contrast/sharpness. **C:** CNN architecture with an encoder-decoder "hourglass" shape. The

encoder part comprises a sequence of convolutional layers. Starting from the last encoder output, the

decoder part iteratively upsamples and fuses feature maps with corresponding encoder's maps, to

produce the output pixel map. The pixel probability map and eye/blink probabilities are computed by

applying the sigmoid activation to the network outputs element-wise.

CNN Architecture

The adopted CNN (Fig. 1 C) takes a grayscale 128x128 image as input and produces three

outputs: a) a 128x128 probability map of each pixel belonging to the pupil, b) the probability the

image contains an eye, and c) the probability the image depicts a blinking eye. We adopted a U-
Net variant [38], a widely used CNN in image segmentation tasks. The model has an encoder-
decoder "hourglass" architecture; the encoder part comprises a sequence of convolutional
layers with ReLU activation and 2x2 max pooling operation, each halving the spatial resolution
of feature maps at every layer; this produces a sequence of feature maps of diminishing spatial
dimensions that provides both spatially local information and global context for the subsequent
steps. Starting from the last encoder output, the decoder part iteratively upsamples and fuses
feature maps with corresponding encoder maps, using convolutional layers, to produce the
output pixel map. All convolutional layers have 16 3x3 kernels and pad their input to obtain a
same-shaped output. Upsampling and downsampling operations have factor 2. Eye and blink
probabilities are predicted by an additional branch that applies global average pooling and a
two-output fully-connected layer to the bottleneck feature map. The pixel probability map and
eye/blink probabilities are computed by applying the sigmoid activation to the network outputs
element-wise.
142

Augmentation, Training, and Validation

We randomly split the dataset into training, validation, and test subsets following a 70/20/10%
split. We perform strong data augmentation during the training phase by applying random
rotation, random cropping, random horizontal and vertical flipping, and random
brightness/contrast/sharpness changes; images are resized to 128x128 before feeding them to
the CNN (Fig. 1 B).
For validation and test images, we take a 128x128 crop centered on the pupil. We compute the
binary cross-entropy for all outputs (pixels and eye/blink logits) and take the sum as the loss
function to minimize. The network is trained with the AdaBelief optimizer [39] for 750 epochs
with a learning rate of 0.001. The best performing snapshot on the validation set is selected and
evaluated on the test set.
154

MEYE: Web-browser Tool

We built a web app for pupillometry on recorded or live-captured videos harnessing our model
as the core component. The trained model has been converted to a web-friendly format using
TensorFlow.js, thus enabling predictions on the user machine using a web browser.

This choice greatly facilitates the deployment and reduces set-up time, as a modern web

browser is the only minimum requirement. Once loaded, an internet connection is not

mandatory, as no data leaves the user's browser, and all the processing is performed on the

user's machine. This implies that performance greatly depends on the user's hardware; if

available, hardware (GPU) acceleration is exploited automatically by *TensorFlow.js*. In our tests,

a modern laptop shipping an Intel(R) Core(TM) i7-9750H 2.60GHz CPU and an Intel(R) UHD

Graphics 630 GPU can process up to 28 frames per second.

The web app also offers additional features that facilitate the recording process, such as:

● Processing of pre-recorded videos or real-time video streams captured via webcam;

● ROI placement via user-friendly web UI (drag&drop) and automatic repositioning

following tracked pupil center;

● Embedded tunable post-processing (map thresholding and refinement via mathematical

morphology);

● Support for registering trigger events;

● Live plotting of pupil area and blink probability;

● Data export in CSV format including: pupil area, blink probability, eye position and trigger

channels.

Behavioral Experiments on Mice

Animal Handling: Mice were housed in a controlled environment at 22 C with a standard 12-h

light-dark cycle. During the light phase, a constant illumination below 40 lux from fluorescent

lamps was maintained. Food (standard diet, 4RF25 GLP Certificate, Mucedola) and water were

available ad libitum and changed weekly. Open-top cages (36.5×20.7×14 cm; 26.7×20.7×14 cm

for up to 5 adult mice or 42.5×26.6×15.5 cm for up to 8 adult mice) with wooden dust-free

bedding were used. All the experiments were carried out following the directives of the

European Community Council and approved by the Italian Ministry of Health (1225/2020-PR).

All necessary efforts were made to minimize both stress and the number of animals used. The

subjects used in this work were three female C57BL/6J mice at 3 months of age.

Surgery: The mouse was deeply anesthetized using isoflurane (3% induction, 1.5%

maintenance). Then it was mounted on a stereotaxic frame through the use of ear bars.

Prilocaine was used as a local anesthetic for the acoustic meatus. The eyes were treated with a

dexamethasone-based ophthalmic ointment (Tobradex, Alcon Novartis) to prevent cataract

formation and keep the cornea moist. Body temperature was maintained at 37 degrees using a

heating pad monitored by a rectal probe. Respiration rate and response to toe pinch were

checked periodically to maintain an optimal level of anesthesia. Subcutaneous injection of

Lidocaine (2%) was performed prior to scalp removal. Skull surface was carefully cleaned and

dried, and a thin layer of cyanoacrylate is poured over the exposed skull to attach a custom

made head post that is composed of a 3D printed base equipped with a glued set screw (12 mm

long, M4 thread, Thorlabs: SS4MS12). The implant is secured to the skull using cyanoacrylate

and UV curing dental cement (Fill Dent, Bludental). At the end of the surgical procedure, the

mice recovered in a heated cage. After 1 hour, mice were returned to their home cage.

Paracetamol was used in the water as antalgic therapy for three days. We wait seven days

before performing head-fixed pupillometry to provide sufficient time for the animal to recover.

Head Fixation: In the awake mouse head-fixation experiments, we employed a modified version

of the apparatus proposed by Silasi et al. [40], equipped with a 3D printed circular treadmill

(diameter: 18cm). Components listed in Table1. A locking ball socket mount (TRB1/M) is

secured to an aluminum breadboard (MB2020/M) using two optical posts (TR150/M-P5) and a

right angle clamp (RA90/M-P5). The circular treadmill is blocked between the base plate pillar

rod and the optical post through a ball-bearing element (BU4041, BESYZY) to allow the disk's

spinning with low effort. To couple the head-fixing thread on the mouse to the locking ball, an

ER025 post was modified by re-tapping one end of it with M4 threads to fit the ball and socket

mount. Velocity is detected using an optical mouse under the circular treadmill. Pupillometry is

performed using a USB camera (oCam-5CRO-U, Withrobot) equipped with a 25 mm M12 lens

connected to a Jetson *AGX Xavier Developer Kit* (NVIDIA) running a custom Python3 script

(30fps). The Jetson hardware is connected with an Arduino UNO through GPIO digital

connection. The Arduino UNO manages the auditory stimuli through a speaker (W3-1364SA 3",

Tang Band).

Part Number	Description	Qty	Price(euro)
TRB1/M	Locking ball and socket mount M4	1	55.83
TR150/M-P5	Optical post M4 - M6 150mm 5pack	1	29.97
RA90/M-P5	Right-Angle Clamp	1	45.7
MB2020/M	Aluminium breadboard	1	72.3
RS075P/M	Pedestal Pillar Post	1	21.63
SS4MS12	Setscrews 12mm long M4	1	5.61

AP4M3M	Adaptor M4-M3	5	1.91
ER025	Cage assembly rod	5	4.73
SS6MS12	Setscrews 12mm long M6	1	5.55
CF038C-P5	Clamping Fork	1	46.49
TOTAL			289.72

Table1. Head-fixation apparatus components

Behavioral Procedure: Mice were handled for 5 minutes each day during the week preceding

the experiments; then, they were introduced gradually to head-fixation for an increasing amount

of time for five days. During Day 1 and 2, we performed two sessions of 10 minutes of head-

fixation, one in the morning and one in the afternoon. On Day 3, we performed one session of

twenty minutes, Day 4 thirty minutes, and Day 5 thirty-five minutes. Each recording started with

5 minutes of habituation. We exposed the animal to the auditory stimuli during the last day.

During each head-fixation session, a curved monitor (24 inches Samsung, CF390) is placed in

front of the animal showing a uniform gray with a mean luminance of 8.5 cd/m2. The frequency

of tone 1 is 3000Hz, tone 2 is 4000Hz, both at 70dB, 10 seconds duration, and 120 seconds of

interstimulus.

Data Analysis: Data has been analyzed using Python 3 and Jupyter notebooks. Correlation has

been performed using *pingouin.corr* (spearman method)*.* Permutation tests were carried out

permuting single subject samples for 3000 times using the function *scipy.random.permutation*

to calculate mean and standard deviation of the chance level. Then each sample of the ERT

was compared with the corresponding null hypothesis distribution using *scipy.stats.norm.cdf.*

All the obtained p-values were corrected for multiple comparisons using the

Benjamini/Hochberg FDR correction of *pingouin.multcomp* Python function. T-tests were

performed using the *pingouin.ttest* Python function. Eyes Movements comparison is carried out

normalizing (in the range between -1 and 1) data from both setups, upsamplig MEYE data from

15 to 1000 fps using linear interpolation and then calculating the Mean Absolute Error (MAE),

performed using the Python function *sklearn.metrics.mean_absolute_error*.

Behavioral Experiments on Humans

PLR: Pupillometry has been performed using a MacBook Pro (Retina, 13-inch, Early 2015, Intel

Core i5 Dual-core 2.7GHz, 8GB of RAM, Intel Iris Graphics 6100 1536 MB) running MEYE

application on Firefox (84.0). The tool is able to compute online pupil size quantification, plotting

the instantaneous pupil area and saving the results on file. Furthermore the tool accepts four

independent manual push button triggers (keys T or Y on the keyboard). This feature allowed us

to annotate stimulation events. A USB IR webcam (Walfront5k3psmv97x, Walfront) equipped

with a Varifocal 6-22mm M12 objective (149129, Sodial) was used to acquire images of the eye.

The camera is equipped with 6 IR LEDs to illuminate the eye uniformly, optimising contrast

between the iris and the pupil. Photic stimulation is delivered using an Arduino Due (Arduino)

microcontroller connected via USB to the notebook and programmed to emulate a keyboard.

The Arduino emulates a keyboard (using the *keyboard.h* library) to send event triggers to MEYE

in the form of keystroke events. The microcontroller drives a stripe of four LEDs (WS2813,

WorldSemi) using the *FastLED.h* library, flashing bright white light for 500 ms with an

interstimulus of 5 seconds (Fig. 3 A). The Subject sat in front of a monitor screen (24 inches

Samsung, CF390) at a distance of 60 cm, with the head stabilized by a chin rest and instructed

to maintain fixation on a small dot presented in the center of the screen for the whole duration of

the recording (57 seconds). A total of 10 flash stimuli have been presented through the strip of

LED mounted above the screen.

Oddball Paradigm Corecordings: To compare the performances shown by the CNN system with

that of a state of the art commercial software, we coregistered pupillometry using MEYE and an

EyeLink 1000 system while 9 (3 males, 6 females, average age 28.78 years) participants

executed an oddball paradigm. The experiment was conducted in a quiet, dark room. The

participant sat in front of a monitor screen (88x50 cm) at a distance of 100 cm, with their head

stabilized by a chin rest. Viewing was binocular. Stimuli were generated with the PsychoPhysics

Toolbox routines [41,42] for MATLAB (MATLAB r2010a, The MathWorks) and presented on a

gamma-calibrated PROPixx DLP LED projector (VPixx Technologies Inc., Saint-Bruno-de-

Montarville, Canada) with a resolution of 1920x1080 pixels, and a refresh rate of 120 Hz. Pupil

diameter was monitored at 1kHz with an EyeLink 1000 system (SR Research) with an infrared

camera mounted below the screen and recording from the right eye. The participant was

instructed to maintain fixation on a small dot (0.5 deg) presented in the center of the screen for

the whole duration of the recording (300 seconds). In this study, the visual stimuli consisted in
the appearance of a high probability stimulus (80% of times) defined as "Standard" and a lower
probability stimulus (20% of times) defined as "Target". The Standard stimulus consisted in a
100% contrast-modulated annular grating (mean luminance 25 cd/m^2), horizontally orientated,
spatial frequency of 0.5 cpd, with an inner and outer diameter of 1.5 and 5 deg, respectively.
The edges of the annulus were smoothed by convolving the stimulus with a gaussian mask
(sigma = 0.5 deg). The Target stimulus has the same parameters of the Standard stimulus
except the orientation that was 45 deg (see Fig. 4 A). The presentation duration of each trial,
either the Standard (0 deg) or Target (45 deg) trial, was 200 ms with the intertrial interval
between two consecutive trials being 2800 ms. The phase of both the *Target* and the *Standard*
stimuli was randomized across trials.The participants were instructed to press a button for a
Target stimulus and not to respond for a Standard stimulus.
293
294 *Eye Movements Corecordings:* For eye tracking recording we employed both the MEYE tool
295 and EyeLink 1000 as described above. In the smooth pursuit condition a small dot (0.5 deg),
296 moved on the screen horizontally, changing direction each 20 degrees of the visual field with a
297 constant velocity of 8 deg/sec. In the Saccades condition every 2.5 s the small dot changes
298 abruptly position horizontally with a span of 20 degrees.
299
300
301

302 Data Availability

303 The code and web app are freely available on Github: github.com/fabiocarrara/meye
304 MEYE is available at: www.pupillometry.it
305 The dataset is available on: https://doi.org/10.5281/zenodo.4488164
306

Results

Pupillometry in Head-Fixed Mice

Our primary goal in this work is to examine if our CNN-based pupillometry can detect pupil

fluctuations in real-time. We designed a behavioral experiment to characterize spontaneous and

evoked pupillary changes. To confirm that pupil size changes are coupled with arousal

transitions, we simultaneously recorded animal running speed during head fixation movements

on the circular treadmill (Fig. 2 A). The experimental protocol included an initial period of

habituation lasting 5 minutes followed by auditory stimulation using two tones (Tone1: 3 kHz,

Tone2: 4kHz, using 120 seconds interstimulus, Fig. 2 B). We first set out to evaluate if CNN can

detect event-related transients (ERT) due to auditory stimulation. For all the acoustic events, we

averaged pupil size and running velocity during a 80 sec temporal window centered on acoustic

stimulation (Fig. 2 C). We detected a significant pupillary dilation after the onset of the auditory

stimulus, together with a similar peak in locomotion (Pupil: P-value < 0.01, Permutation Test,

Velocity: P-value < 0.01, Permutation Test). This event-related transient is a proxy of the

arousal change due to acoustic detection and can be considered a manifestation of cognitive

and emotional processing of the stimulus [2]. By analyzing the overall traces, we found that the

pupil diameter and the trace related to animal running, were characterized by a significant

correlation (r: 0.79, P-value < 0.001, Pearson Correlation), in agreement with the hypothesis

that fluctuations of the arousal level mainly characterize spontaneous pupillary events.

Moreover, we calculated pupil size during habituation in different arousal states, finding that

during locomotion, pupil size is significantly larger than in stationary periods (P-value < 0.01,

Paired T-Test, Fig. 2 D). These results indicate that CNN pupillometry can detect spontaneous

and elicited pupillary changes and can be used to monitor the mouse's arousal state during

head fixation experiments.

Fig.2: Pupillometry in head-fixed mice. **A:** Setup for head-fixed pupillometry in the awake mouse. The mouse is head-fixed to a custom made metal arm equipped with a 3D printed circular treadmill to monitor running behavior. In the meantime, pupillometry is performed using CNN. **B:** The average fluctuation of pupillometry and velocity in all experimental mice. Dashed pink and yellow areas represent the onset and duration of auditory stimuli. Evoked peaks in both pupil size (blue line) and velocity (green line) are clearly noticeable during auditory stimulation. **C:** Average event-related transients for both pupil size and velocity. Gray areas represent stimulus onset and duration. Red areas represent statistically significant data points with respect to random permutation testing. **D:** Sensibility of the system to detect spontaneous arousal fluctuations. Average pupil size is significantly affected by the behavioral states of the animal. During running epochs (Moving) the pupil is significantly more dilated than during the resting state (Stationary).

Web-Browser Application to Perform Pupillometry Experiments

To test the implementation of the CNN in a web-browser (MEYE, Fig. 3 B), we designed a
simple experiment aimed to measure PLR evoked by brief flashes of light on the human eye.
The experiment included 10 flash events with an interstimulus of 5 seconds (dashed vertical
lines in Fig. 3 C). The results showed a clear light-induced modulation of pupil size in
correspondence with each flashes onset. Aligning and averaging all the traces along with the
events, PLR can be quantified in both the raw (44.53%±0.67% change from baseline) and z-
scored (14.59±2.05 st.dev. from baseline) trace (Fig. 3 D-E). To detect if it is possible to
measure cognitively driven pupil signals using the MEYE tool reliably, we performed
pupillometry while participants executed an oddball task, a commonly used paradigm for
cognitive and attentional measurement. This task is based on the principle by which pupil
dilation is stronger in response to rare stimuli and can be used as a physiological marker for the
detection of deviant stimuli [37]. This experiment has been carried out recording the same eye
using both the MEYE tool and an EyeLink 1000 system. According to Google Scholar, the
Eyelink system is one of the most utilized eye trackers in psychology, psychophysics, and
neuroscience, with more than 17K scientific publications mentioning this tool. During the oddball
experiment, the subject was instructed to maintain fixation on a small dot presented in the
center of the screen, pushing a button only when the *Target* stimulus appears on the screen and
not responding to the *Standard* stimulus (Fig. 4 A). Averaging and comparing the responses to
Standard and *Target* gratings results in a significant stronger pupil dilation for the *Target*
stimulus than the *Standard* stimulus, that is detected by both the recording systems (MEYE: P-
value < 0.001, T-Test Paired, EyeLink: P-value < 0.001, T-Test Paired, Fig. 4 B-C). No
differences have been found for the responses evoked by the *Target* stimulus between the
MEYE tool and the EyeLink system (P-value:0.327, T-test Paired, Fig. 4 B-inset). Moreover, the
single-subject pupillary evoked amplitudes show a significant positive correlation between the
two techniques (P-value:0.01, r:0.88, Pearson Correlation) with more than 75% of the variability
explained by the linear model. Pupil-size is known to covary with eye position in video-based
measurements [43], producing foreshortening of the pupillary image because the camera is
fixed but the eye rotates. To overcome this issue, there are several possible solutions: the most
simple requires to maintain constant fixation throughout each trial, but, if this requirement
cannot be satisfied (such as in sentence reading), the position of the pupil at each sample is
required to correct and mitigate the estimation error. Thus, we decided to quantify the

agreement between positional outputs provided by MEYE and Eyelink for horizontal eye

movements. We designed two tasks: in the first task, a dot smoothly traveled horizontally on the

screen from left to right and vice versa at a velocity of 8 deg/sec and spanning 20 degrees,

producing slow and smooth pursuit eye movements. In the other experiment, a dot jumped

every 5 seconds from a position to the other (spanning 20 degrees), producing large, fast, and

abrupt saccades. Results (Fig. 4 D) show that smooth pursuit movements generate a sinusoidal

change of position with a good agreement between both systems (MAE: 0.04). The second

task, inducing saccades, produces a slightly larger error (MAE: 0.073). This error is mainly due

to the much lower sampling rate of MEYE (MEYE:15 fps; Eyelink: 1000 fps). This means that

even if MEYE provides the exact positional information for each sample, it has a lower

performance in adequately describing fast eye movements, such as saccades. Thus, MEYE

provides the data required for post-hoc correction of pupil measures although it should be used

with caution for measuring saccades. This factor should be taken into account when designing

experiments using MEYE. Finally, MEYE can also be used as an offline tool to analyze

pupillometry videos in various file formats, depending on the video codec installed in the web-

browser. We successfully analyzed videos in different experimental conditions, including head-

fixed mice running on a treadmill, during 2-photon calcium imaging, and in humans (Fig. 4 A-C).

Fig.3: Web-browser Pupillometry Experiment. **A:** Experimental setup for running the PLR stimulation and in the meantime perform pupillometric recordings. The PC is connected to the internet running an instance of MEYE tool in the web browser. A USB camera, equipped with an IR light source, is focused on the eye of the subject. The photic stimulus is delivered using a LED array driven by an Arduino Due. The Arduino is connected to the PC emulating a keyboard and sending keystrokes stimulus triggers to the MEYE tool. **B:** A picture of MEYE graphical user interface. The subject during the recording is visualized as a streaming video. A ROI is used to locate the eye and a preview of the estimation of the pupil is superimposed to the image of the subject. The GUI allows to set different parameters of post-processing (map thresholding and refinement via mathematical morphology). C: Raw trace of the experiment (blue). Dashed lines locate the onset of flash stimuli. The green rectangles locate the onset and duration of blinks. The samples corresponding to blinks are removed and linearly interpolated (in red). **D:** Average event related transient to flash stimulation in raw values. After the onset of the stimulus (dashed line) a strong constriction of the pupil is observed (44.53%). **E:** Z-score of the average event related transient seen in D. The average nadir amplitude is 14.59 standard deviations from baseline.

413
414
415

Fig.4: Cognitively driven pupillary changes. **A:** Visual Oddball procedure. The participant is instructed to fixate a small red dot placed at the center of the screen and to push a button only when the Target visual stimulus appears. **B:** Average Pupil waveforms**.** Average pupil response to Standard and Target stimulus for both MEYE (blue, left) and EyeLink (red, right). In the inset is represented the comparison between the evoked response to the Target stimulus in both setups. **C:** Average pupil response. Difference between the Standard and Target stimuli recording using MEYE (uppermost) and Eyelink (middle). The lowermost graph represents the correlation between MEYE and Eyelink data. **D:** Eye movements data. comparison between the MEYE tool(blue) and Eyelink system (red) during smooth pursuit task (upper) and saccades (lower).

426
427
428

429
430

Fig.5: Offline Movies Analysis. **A:** Awake Head-fixed mouse running on a treadmill, recorded for 40

seconds. The grey area represents a blink, the trace of the blink is removed and linearly interpolated (Red

line). **B:** Awake mouse during 2-photon calcium imaging. Here is clearly visible a brighter pupil with

respect to A. Blinking epochs are removed and linearly interpolated. **C:** Pupillometry performed on a

human subject, with a higher blinking rate with respect to mice. In all figures the insets images represent

the ROIs.

440 Discussion

In this work, we demonstrated that MEYE is a sensitive tool that can be used to study pupil

dynamics in both humans and mice. Furthermore, by providing eye position MEYE allows post-

hoc control of possible effects of eye movements on pupil measures[43]. MEYE can detect both

spontaneous and evoked pupil changes in a variety of conditions: mice with black pupils in

normal illumination conditions, and mice with bright pupils resulting from laser infrared

illumination. This flexibility allows the use of MEYE in combination with 2-photon, wide field

imaging and electrophysiological techniques widely adopted in the awake or anaesthetized

mice. Furthermore, MEYE can be employed to design standalone experiments using cost-

effective hardware with performance comparable with that of state-of-the-art commercial

software. In this experiment, we used a USB webcam with a varifocal objective that allows focal

adjustment concentrated on the eye. The cost of the imaging equipment is less than 50 euros

(see Table 2), and requires no knowledge of coding to set up. The flashing stimulus apparatus

requires a basic understanding of Arduino boards and can be assembled at a price lower than

50 euros. The overall cost of the apparatus is less than 100 euros. Our code can be used in two

different ways, to satisfy many needs. One way relies on the standalone web-browser tool, that

allows running MEYE on almost any device, from scientific workstations to notebooks or even

smartphones. The other way utilizes a dedicated Python script running the CNN locally on a

workstation. This latter case is suited for experiments with specific requirements, like high and

stable framerate or online processing of pupil size in which on-the-fly pupil computer-interaction

is required.

Valid open source and commercial alternatives exist, most of them are dedicated to gazing

tracking and/or pupillometry. Commercial options are costly (tobii.com, sr-research.com,

neuroptics.com), whereas open-source code instead requires programming knowledge and

most of them are explicitly dedicated to one species [33,34]. One of these papers [33] assessed

pupil dilation in mice through DeepLabCut [44], a technique for 3D markerless pose estimation

based on transfer learning. This approach, albeit powerful, is conceptually different, since it is

trained on user-defined key-points instead of using the entire pupil to perform semantic

segmentation. The former technique is more suited to track and locate arbitrary objects on an

image, the latter technique is focused on a more precise quantification of even small changes of

the object area, since pixel-wise segmentation masks are refined iteratively using local and

global context. The possible contribution of the web app technology resides in its portability: no

software needs to be manually installed and configuration is minimal. Only a clear IR image of

the subject's eye is required. The performances of the tool are dependent on the host computer

but it runs at >10 fps in most of the machines tested. This advantage is particularly useful for

settings with limited resources and space or for educational purposes. Web browser embedded

pupillometry will also be crucial for human scientific research, clinical and preventive medicine.

It would also be a promising tool in the recently growing field of telemedicine given its minimal

setup that can run on an average notebook or even on a smartphone, it allows possible large-

scale recruitment of subjects directly in their own homes. This greatly facilitates infants,

psychiatric, and motor-impaired patients' compliance, particularly for longitudinal research
designs. We also released an open-source database of eyes composed of more than 11.000
images in various settings: head-fixed mice (black pupil), head-fixed two-photon imaging mice
(white pupil), and human eyes. This dataset will grow over time to introduce new species and
new use cases to increase, update, and strengthen MEYE performances. The possible
scenarios can be further expanded in the future, due to the dynamic nature of CNN. It can be
updated from the source, providing instantaneous updates on each computer running an
instance of the program. Our hope is to create a community that refines and consolidates
pupillometric performances, to produce a tool that can be applied in different environments.
489
490

Part Number	Description	Qty	Price(euro)	store	Manufacturer
Walfront5k3psmv97x	USB webcam	1	33.48	Amazon	Walfront
149129	Varifocal M12 Lens	1	12.03	Amazon	Sodial
A000062	Microcontroller Arduino Due	1	35	Arduino Store	Arduino
1312	4 NeoPixel RGB LEDs	1	6.53	Adafruit	Adafruit
Total Amount			87.04		

491
492 Table2: Hardware equipment for PLR.

Authors' contributions

494
495 RM, FC and TP designed the research; RM, FC and GA trained and designed the AI tools; AV,
496 RM, LL, GR and LLV labeled manually the dataset, RM, AV, LL, GS, LLV and AB performed the
497 experiments, RM and FC performed data analysis.
498
499

Acknowledgments

501 We gratefully acknowledge NVIDIA Corporation's support with the Jetson AGX Xavier

502 Developer Kit's donation for this research. Authors would also like to thank Dr. Viviana Marchi,

503 Dr. Grazia Rutigliano and Dr. Carlo Campagnoli for the critical reading of the manuscript.

506 Funding

This work was partially supported by H2020 projects AI4EU under GA 825619 and AI4Media

under GA 951911. Funding from the Italian Ministry for university and research MIUR-PRIN

2017HMH8FA; AIRETT Associazione Italiana per la sindrome di Rett Project "Validation of

pupillometry as a biomarker for Rett syndrome and related disorders: longitudinal assessment

and relationship with disease"; Orphan Disease Center University of Pennsylvania grant MDBR-

19-103-CDKL5; and Associazione "CDKL5 - Insieme verso la cura".

515 Competing interests

516 The authors declare that they have no competing interests.

517 References

1.   Reimer J, McGinley MJ, Liu Y, Rodenkirch C, Wang Q, McCormick DA, et al. Pupil

fluctuations track rapid changes in adrenergic and cholinergic activity in cortex. Nat

Commun. 2016;7: 13289.

2.   Lee CR, Margolis DJ. Pupil Dynamics Reflect Behavioral Choice and Learning in a

Go/NoGo Tactile Decision-Making Task in Mice. Frontiers in Behavioral Neuroscience.

2016. doi:10.3389/fnbeh.2016.00200

3.   McGinley MJ, Vinck M, Reimer J, Batista-Brito R, Zagha E, Cadwell CR, et al. Waking

State: Rapid Variations Modulate Neural and Behavioral Responses. Neuron. 2015;87:

1143–1161.

4.   Binda P, Pereverzeva M, Murray SO. Attention to bright surfaces enhances the pupillary
light reflex. J Neurosci. 2013;33: 2199–2204.

5.   Wierda SM, van Rijn H, Taatgen NA, Martens S. Pupil dilation deconvolution reveals the
dynamics of attention at high temporal resolution. Proc Natl Acad Sci U S A. 2012;109:
8456–8460.

6.   Angulo-Chavira AQ, García O, Arias-Trejo N. Pupil response and attention skills in Down
syndrome. Res Dev Disabil. 2017;70: 40–49.

7.   Krebs RM, Park HRP, Bombeke K, Boehler CN. Modulation of locus coeruleus activity by
novel oddball stimuli. Brain Imaging Behav. 2018;12: 577–584.

8.   Montes-Lourido P, Kar M, Kumbam I, Sadagopan S. Pupillometry as a reliable metric of
auditory detection and discrimination across diverse stimulus paradigms in animal models.
Cold Spring Harbor Laboratory. 2020. p. 2020.11.16.385286.
doi:10.1101/2020.11.16.385286

9.   Charier D, Vogler M-C, Zantour D, Pichot V, Martins-Baltar A, Courbon M, et al. Assessing
pain in the postoperative period: Analgesia Nociception IndexTM versus pupillometry.
British Journal of Anaesthesia. 2019. pp. e322–e327. doi:10.1016/j.bja.2018.09.031

10. Azevedo-Santos IF, DeSantana JM. Pain measurement techniques: spotlight on
mechanically ventilated patients. J Pain Res. 2018;11: 2969–2980.

11. Connelly MA, Brown JT, Kearns GL, Anderson RA, St Peter SD, Neville KA. Pupillometry: a
non-invasive technique for pain assessment in paediatric patients. Arch Dis Child. 2014;99:
1125–1131.

12. Binda P, Pereverzeva M, Murray SO. Pupil constrictions to photographs of the sun. J Vis.
2013;13. doi:10.1167/13.6.8

13. Aleman TS, Jacobson SG, Chico JD, Scott ML, Cheung AY, Windsor EAM, et al.
Impairment of the Transient Pupillary Light Reflex inRpe65−/−Mice and Humans with Leber
Congenital Amaurosis. Investigative Opthalmology & Visual Science. 2004. p. 1259.
doi:10.1167/iovs.03-1230

14. Winston M, Zhou A, Rand CM, Dunne EC, Warner JJ, Volpe LJ, et al. Pupillometry

measures of autonomic nervous system regulation with advancing age in a healthy

pediatric cohort. Clinical Autonomic Research. 2020. pp. 43–51. doi:10.1007/s10286-019-

00639-3

15. Blaser E, Eglington L, Carter AS, Kaldy Z. Pupillometry reveals a mechanism for the Autism

Spectrum Disorder (ASD) advantage in visual tasks. Sci Rep. 2014;4: 4301.

16. Rorick-Kehn LM, Witcher JW, Lowe SL, Gonzales CR, Weller MA, Bell RL, et al.

Determining pharmacological selectivity of the kappa opioid receptor antagonist LY2456302

using pupillometry as a translational biomarker in rat and human. Int J

Neuropsychopharmacol. 2014;18. doi:10.1093/ijnp/pyu036

17. Oh AJ, Amore G, Sultan W, Asanad S, Park JC, Romagnoli M, et al. Correction:

Pupillometry evaluation of melanopsin retinal ganglion cell function and sleep-wake activity

in pre-symptomatic Alzheimer's disease. PLOS ONE. 2020. p. e0230061.

doi:10.1371/journal.pone.0230061

18. Oh AJ, Amore G, Sultan W, Asanad S, Park JC, Romagnoli M, et al. Pupillometry

evaluation of melanopsin retinal ganglion cell function and sleep-wake activity in pre-

symptomatic Alzheimer's disease. PLOS ONE. 2019. p. e0226197.

doi:10.1371/journal.pone.0226197

19. Chougule PS, Najjar RP, Finkelstein MT, Kandiah N, Milea D. Light-Induced Pupillary

Responses in Alzheimer's Disease. Front Neurol. 2019;10: 360.

20. Burley DT, van Goozen SHM. Pupil Response to Affective Stimuli: a Biomarker of Early

Conduct Problems in Young Children. J Abnorm Child Psychol. 2020;48: 693–701.

21. Gajardo AIJ, Madariaga S, Maldonado PE. Autonomic nervous system assessment by

pupillary response as a potential biomarker for cardiovascular risk: A pilot study. Journal of

Clinical Neuroscience. 2019. pp. 41–46. doi:10.1016/j.jocn.2018.11.015

22. Iadanza E, Goretti F, Sorelli M, Melillo P, Pecchia L, Simonelli F, et al. Automatic Detection

of Genetic Diseases in Pediatric Age Using Pupillometry. IEEE Access. 2020. pp. 34949–

34961. doi:10.1109/access.2020.2973747

23. Frost S, Robinson L, Rowe CC, Ames D, Masters CL, Taddei K, et al. Evaluation of

Cholinergic Deficiency in Preclinical Alzheimer's Disease Using Pupillometry. Journal of

Ophthalmology. 2017. pp. 1–8. doi:10.1155/2017/7935406

24. Obinata H, Yokobori S, Shibata Y, Takiguchi T, Nakae R, Igarashi Y, et al. Early automated
infrared pupillometry is superior to auditory brainstem response in predicting neurological
outcome after cardiac arrest. Resuscitation. 2020;154: 77–84.

25. El Ahmadieh TY, Bedros N, Stutzman SE, Nyancho D, Venkatachalam AM, MacAllister M,
et al. Automated Pupillometry as a Triage and Assessment Tool in Patients with Traumatic
Brain Injury. World Neurosurg. 2021;145: e163–e169.

26. Artoni P, Piffer A, Vinci V, LeBlanc J, Nelson CA, Hensch TK, et al. Deep learning of
spontaneous arousal fluctuations detects early cholinergic defects across
neurodevelopmental mouse models and patients. Proc Natl Acad Sci U S A. 2020;117:
23298–23303.

27. Nyström P, Gliga T, Nilsson Jobs E, Gredebäck G, Charman T, Johnson MH, et al.
Enhanced pupillary light reflex in infancy is associated with autism diagnosis in
toddlerhood. Nat Commun. 2018;9: 1678.

28. Ponzio F, Villalobos AEL, Mesin L, de'Sperati C, Roatta S. A human-computer interface
based on the "voluntary" pupil accommodative response. International Journal of Human-
Computer Studies. 2019. pp. 53–63. doi:10.1016/j.ijhcs.2019.02.002

29. Mathôt S, Melmi J-B, van der Linden L, Van der Stigchel S. The mind-writing pupil: A
human-computer interface based on decoding of covert attention through pupillometry.
doi:10.7287/peerj.preprints.1361

30. Beggiato M, Hartwich F, Krems J. Using Smartbands, Pupillometry and Body Motion to
Detect Discomfort in Automated Driving. Front Hum Neurosci. 2018;12: 338.

31. Stoll J, Chatelle C, Carter O, Koch C, Laureys S, Einhäuser W. Pupil responses allow
communication in locked-in syndrome patients. Curr Biol. 2013;23: R647–8.

32. Olivia C, Josef S, Camille C, Christof K, Steven L, Wolfgang E. The use of pupil dilation to
communicate with locked-in syndrome patients. Frontiers in Human Neuroscience. 2013.
doi:10.3389/conf.fnhum.2013.212.00126

33. Privitera M, Ferrari KD, von Ziegler LM, Sturman O, Duss SN, Floriou-Servou A, et al. A

complete pupillometry toolbox for real-time monitoring of locus coeruleus activity in rodents.
Nat Protoc. 2020;15: 2301–2320.

34. Yiu Y-H, Aboulatta M, Raiser T, Ophey L, Flanagin VL, Zu Eulenburg P, et al. DeepVOG:
Open-source pupil segmentation and gaze estimation in neuroscience using deep learning.
J Neurosci Methods. 2019;324: 108307.

35. Aggius-Vella E, Gori M, Animali S, Campus C, Binda P. Non-spatial skills differ in the front
and rear peri-personal space. Neuropsychologia. 2020;147: 107619.

36. LoTemplio S, Silcox J, Federmeier KD, Payne BR. Inter- and intra-individual coupling
between pupillary, electrophysiological, and behavioral responses in a visual oddball task.
Psychophysiology. 2020 [cited 24 Feb 2021]. doi:10.1111/psyp.13758

37. Liao H-I, Yoneya M, Kidani S, Kashino M, Furukawa S. Human Pupillary Dilation Response
to Deviant Auditory Stimuli: Effects of Stimulus Properties and Voluntary Attention. Front
Neurosci. 2016;10: 43.

38. Ronneberger O, Fischer P, Brox T. U-Net: Convolutional Networks for Biomedical Image
Segmentation. Medical Image Computing and Computer-Assisted Intervention – MICCAI
2015. Springer, Cham; 2015. pp. 234–241.

39. Zhuang J, Tang T, Ding Y, Tatikonda S, Dvornek N, Papademetris X, et al. AdaBelief
Optimizer: Adapting Stepsizes by the Belief in Observed Gradients. 2020. Available:
http://arxiv.org/abs/2010.07468

40. Silasi G, Xiao D, Vanni MP, Chen ACN, Murphy TH. Intact skull chronic windows for
mesoscopic wide-field imaging in awake mice. J Neurosci Methods. 2016;267: 141–149.

41. Brainard DH. The Psychophysics Toolbox. Spat Vis. 1997;10: 433–436.

42. Pelli DG. The VideoToolbox software for visual psychophysics: transforming numbers into
movies. Spat Vis. 1997;10: 437–442.

43. Hayes TR, Petrov AA. Mapping and correcting the influence of gaze position on pupil size
measurements. Behav Res Methods. 2016;48: 510–527.

44. Mathis A, Mamidanna P, Cury KM, Abe T, Murthy VN, Mathis MW, et al. DeepLabCut:
markerless pose estimation of user-defined body parts with deep learning. Nat Neurosci.

2018;21: 1281–1289.

642