MicroMator: Open and Flexible Software for Reactive Microscopy

Zachary R Fox[1,2,3,*], Steven Fletcher[1,2,*], Achille Fraisse[2,1,§], Chetan Aditya[1,2,§], Sebastián Sosa-

Carrillo[1,2], Sébastien Gilles[4], François Bertaux[2,1], Jakob Ruess[1,2], Gregory Batt[1,2]

[1] Inria Paris, 2 rue Simone Iff, 75012 Paris, France

[2] Institut Pasteur, 28 rue du Docteur Roux, 75015 Paris, France

[3] Center for Nonlinear Studies, Los Alamos National Laboratory, Los Alamos, NM, USA

[4] Inria Saclay – Ile-de-France, 1 rue H. d'Estienne d'Orves, 91120 Palaiseau, France

[*] These authors contributed equally to this work

[§] These authors contributed equally to this work

11   Abstract

12   *Microscopy image analysis has recently made enormous progress both in terms of accuracy and speed*

13   *thanks to machine learning methods. This greatly facilitates the online adaptation of microscopy*

14   *experimental plans using real-time information of the observed systems and their environments. Here*

15   *we report MicroMator, an open and flexible software for defining and driving reactive microscopy*

16   *experiments, and present applications to single-cell control and single-cell recombination.*

18   Introduction

Software for microscopy automation are essential to support reproducible high-throughput

microscopy experiments[1]. Samples can now be routinely imaged using complex spatial and temporal

patterns. Yet, in the overwhelming majority of cases, executions of experiments are still cast in stone

at the beginning, with little to no possibility for human or computer-driven interventions during the

experiments. This is all the more surprising given that image analysis has recently made a giant leap

in terms of accuracy and rapidity thanks to deep learning methods. Software empowering microscopy

with real-time adaptation capabilities is needed to exploit the full potential of automated

microscopes.

Several dedicated software solutions have been developed for applications requiring real-time

analysis, most notably efficient scanning of large and complex samples (eg Refs[2–5]) and real-time

control of cellular processes (eg Refs[6–12]). However only a handful generic solutions have been

proposed so far[13–16]. AutomatedFCS is a Fiji plugin that is part of the Microscopy Pipeline Constructor.

Its potential has been demonstrated but the tool applies only to Zeiss confocal microscopes[15]. The

three other tools are all extensions of μManager, a widely-used software[17,18] controlling a large range

of microscopy hardware. However, no in-depth case studies demonstrating their practical applicability

–and showing possible limitations– have been reported so far.

35   MicroMator Software

To fill this gap, we developed MicroMator, a modular open-source Python software to support

reactive microscopy, and demonstrate its potential via two challenging case studies. In MicroMator,

events play a fundamental role (Fig 1A). They consist of Triggers and Effects. They can be defined by

the user in a flexible manner. Examples of triggers include "at the 10th frame", "if more than 100 cells

are in the field of view", "if the fluorescence of the 3rd newborn cell exceeds 100 a.u." and "if message

!update position=10 frame=last is received from Discord". Examples of effects include changing a

microscope configuration, sending light in the field of view with a given pattern, actuating a

microfluidic pump, or starting an optimization routine. Microscopy experiments are defined by a main
image acquisition loop, that serves as a backbone for the experiment, and by event creation functions
(Fig 1B). Naturally, the main acquisition loop itself can be modified by event effects in the course of
the experiment. For controlling hardware, MicroMator primarily uses the powerful Python API of
μManager MMCorePy, but can also use other dedicated Python or web-based APIs provided by
vendors, as done for our CellAsic ONIX microfluidic platform. Various types of analysis can be
performed using dedicated software modules, such as on-line image analysis or real-time control and
optimization. Communication modules can be used to interface MicroMator with digital distribution
platforms such as Discord to track experiment progress and potential issues. Lastly, MicroMator
leverages Python's multiprocessing module to perform computations concurrently and possesses an
extensive and customizable logging system, gathering logs of all modules in a unique file and fostering
reproducible research (see SI Text).

We also provide SegMator, a software which uses U-Net for bright-field yeast cell segmentation
(provided by DeLTA[19]), and tracking using TrackPy[21]. U-Net is a convolutional neural network with a
structure that excels at image segmentation[20]. U-Net can analyze dense fields of cells in a few seconds
and with good accuracy (see SI Text and Fig S1 and S2).

To showcase the full potential of reactive experiments performed with MicroMator, we designed
experiments in which cellular processes are controlled in real-time. Single-cell stimulations are
computed on line based on the cell state and/or position, demonstrating that reactive loops can be
implemented at the level of individual cells.

Model predictive control of gene expression at the single-cell level in yeast

For our first application, we use the EL222 optogenetic system and the mScarletI fluorescent reporter
to engineer light-responsive yeast cells (Fig 2A). Our goal is to compare different model predictive
control (MPC) strategies for controlling the expression levels of a protein in a cell population. Using
real-time imaging, segmentation and cell tracking, different cells can be stimulated differently in the
field of view using a digital micromirror device (DMD). The cellular response was then characterized
for different light stimulation profiles in the same experiment (Fig 2B). Only the most central part of
the cell receives significant light stimulation. This erosion of the stimulation region helps improving
the precision of single-cell light stimulations in dense cell regions because of illumination bleed-
through of DMD systems (see SI Text and Fig S3). We then developed and calibrated an "average cell"
(deterministic) and a "single cell" (stochastic) model of light-driven gene expression (SI Text and Fig
S4, and S5).

In open loop control, the average cell model is used to precompute a temporal pattern of light
stimulation so that cells follow a target behavior. This light pattern is then applied to all cells in the
field of view (Fig 2C). In closed loop population-based control, the average cell model and the average
of the measured fluorescence of cells are used by classical state estimators and model predictive
controllers to compute in real-time the appropriate light stimulation to drive the mean fluorescence
to its target (Fig 2D). Finally, in closed loop single-cell control, a stochastic model of gene expression
and single-cell fluorescence measurements are used by advanced state estimators and controllers to
compute in real-time the appropriate light stimulations to drive the fluorescence of each and every
cell in the field of view to its target (Fig 2E). This control problem is quite challenging and needs to be
solved for hundreds of cells in parallel. Advanced methods for numerical simulation and state
estimation were essential (see SI Text and Fig S6).

Defining control performance as the time averaged deviation to target, we found that the single-cell
control method leads to a modest reduction of error of the population averaged fluorescence but to
a drastic improvement of the average error of the single-cell fluorescence (Fig 2F).

Patterns of recombined yeast cells

For our second application, we constructed a light-driven artificial recombination system in yeast and
employed different light stimulation strategies to obtain various structures of recombined cells.

We again used the EL222 optogenetic induction system but this time to drive the expression of the
Cre recombinase. The Cre recombinase induces the expression of a fluorescent reporter, mCerulean,
via an amplification step using the ATAF1 transcription factor (Fig 3A & 3B).

Firstly, we tried to create islets of recombined cells. To this end, we dynamically searched for cells that
were far from any previously-targeted cell, and targeted these cells for recombination. To maximize
the chances that the chosen cells do recombine, we tracked each chosen cell and targeted it
repeatedly with light stimulations. Our strategy was effective in creating isolated micro-colonies of
recombined cells (Fig 3C). Analysis of the lineage trees of targeted cells and non-targeted cells
confirmed that recombined cells have a slow growth phenotype.

Secondly, we applied a ring-like recombination signal. More specifically, every cell that was in the
designated zone at any moment throughout the experiment has been targeted for recombination. As
a result, we did obtain a ring-like pattern of recombined cells (Fig 3D). Experimental and biological
limitations can been revealed by the analysis of the tails of the distributions of the recombination
readout (*i.e.*, mCerulean fluorescence) within the cell populations (Fig 3D). For example, we found
that some cells have been erroneously targeted for recombination because of tracking issues, and that
only a few cells have not shown the recombined phenotype at the end of the experiment despite
having being effectively targeted for recombination (Fig 3D and S8).

Significance

We presented MicroMator together with two challenging applications demonstrating how this
software can help using automated microscopy platforms to their full potential. Each application goes
beyond the state of the art. Indeed, we showed the first application of single-cell model predictive
control using a full stochastic model of gene expression and the first demonstration of cell
recombination driven at the single-cell level, enabling single-cell resolution patterns.

Yet, we foresee that reactiveness in microscopy will primarily be used to enhance and automate
classical experiments. Examples of simple use cases abound: triggering autofocus only when needed,
dynamically adjusting the imaging condition to the signal strength, identifying novel regions of
interest, or following the course of experiments via easily accessible online services (e.g. warning
messages sent on Discord), to provide but a few examples. Thanks to its modular nature and to its use
of a simple but powerful event system, MicroMator capacities can be conveniently expanded to drive
novel hardware or perform a wide range of analyses. MicroMator is a relatively simple software
extension that significantly empowers laboratory equipment that is present in most quantitative
biology laboratory worldwide. MicroMator is a relatively simple software extension that significantly
enhances heavily-used research equipment found in most quantitative biology laboratories.

Figure 1. MicroMator overview. A. Modular software architecture. MicroMator consists of an extensible set of modules that
control various hardware and software aspects of microscopy experiments and of a central core that handles user-defined
events. It is written in the high-level programming language Python. **B.** Event-based reactive microscopy workflow. Imaging
can be followed by online analysis of the samples. This typically involves segmentation, tracking, quantification of cell
properties, and possibly advanced additional computations. Effects may then be triggered based on the result of the analysis.
These may include the physical actuation of the hardware or the initiation of communications or of additional computations.

Figure 2. Control gene expression at the single cell level in yeast. A. The red fluorescent protein mScarletI is placed under the control of the light-responsive transcription factor EL222. **B.** To efficiently characterize cell responses to light stimulations, cells in the field of view are partitioned in 3 groups, each group being stimulated with a different temporal profile. Bright-field images are segmented and cells are tracked. Then, based on their groups, cells are stimulated during the appropriate time with eroded masks. The temporal evolution of the mean mScarletI fluorescence of the cells in the three groups is shown with envelopes indicating one standard deviation. **C.** Open-loop control experiment in which a model of the response of the cell population is used to precompute a light stimulation profile that drives the cell population to the target behavior. The application of the light profile leads to significant deviations from the target of the individual cell trajectories. **D.** Closed-loop control experiment in which the same model is used jointly with real-time observations of the population state to decide which light profile to apply to all cells, using a receding horizon strategy. **E.** A stochastic model of individual cell response is used jointly with single-cell observations to decide which light profile to apply to each cell. **F.** The different strategies have similar performances to drive the mean fluorescence to its target, but the single-cell feedback strategy is significantly better to drive individual cells to their target profiles.

Figure 3. Patterns of recombined yeast cells. A. Upon light exposure, the Cre recombinase is expressed and triggers recombination, leading to the expression of ATAF1 and then of mCerulean-Far1M. Stars indicate nuclear localization of the protein. **B.** Targeted cells are stimulated for 1 second every 6 minutes until the end of the experiment. Fluorescence levels emitted by targeted cells can be recorded. At the end of the experiment, all cells are imaged and a recombined or non-recombined phenotype is attributed. **C.** Cells are dynamically selected such that no target cells are close to each other. Cell lineages of targeted and non-targeted cells can be manually reconstructed and statistics can be extracted. **D.** A ring-like region in the field of view is selected at the beginning of the experiment and all cells entering the designated region at some time point are targeted for recombination. The distributions of the fluorescence levels of the targeted and non-targeted cells can be computed at the end of the experiment. The vast majority of cells present the expected phenotype and outliers can be further analyzed.

Methods

Software and data availability

MicroMator is an open-source software. It contains a core part and an extensible list of modules. The
MicroMator core manages the user-specified events and also the metadata and logging system. The current list
of modules includes a Microscope Controller module, an Image Analysis module, a Model Predictive Control
module, and a Discord Bot module. The Microscope Controller module is an interface with the Python wrapper
for MicroManager MMCorePy. The Image Analysis module uses deep learning methods to segment yeast cells
from bright-field images. It also uses an efficient algorithm for cell tracking. This module is also available as a
standalone tool called SegMator. The Model Predictive Control module implements state estimation and model
predictive control routines for deterministic and stochastic systems, at either the population or single-cell level.
The Discord Bot module uses a web app running on the microscope's computer and connected to the Discord
communication system.

MicroMator, SegMator, the events for the experiments of Fig 2E and S3, and data analysis code for the
experiments of Fig 2E, 3D, and S3 can be found online: https://gitlab.inria.fr/InBio/Public/micromator. Data for
all figures is currently available upon request and will soon be uploaded on a public repository.

Genetic constructions and yeast strains

All plasmids and strains were constructed using the *Yeast Tool Kit*, a modular cloning framework for yeast
synthetic biology[22], the common laboratory strain BY4741 (Euroscarf), and the EL222 optogenetic system[23]. The
light responsive strain (IB44) harbors a constitutively expressed EL222 light-responsive transcription factor (NLS-
VP16AD-EL222) and an EL222-responsive promoter (5xBS-CYC180pr) driving the expression of the mScarletI
protein. The IB44 strain genotype is MATa his3Δ1 leu2Δ0::5xBS-CYC180pr-mScarletI-Leu2 met15Δ0 ura3Δ::NLS-
VP16AD-EL222-URA3. The recombining strain (IB237) harbors a constitutively expressed EL222 light-responsive
transcription factor (NLS-VP16AD-EL222) floxed between two LoxP sites that upon recombination expresses the
ATAF1 transcription factor. This factor expresses (pATAF1_4x) in turn the mCerulean fluorescent protein fused
to a constitutively active Far1 protein (FAR1M_mCerulean). Lastly, the strain also harbors the Cre recombinase
placed under the control of an EL222-responsive promoter (5BS-Gal1pr). The IB237 strain genotype is MATa
his3Δ1::pATAF1_4x-FAR1M_mCerulean-tDIT1-HIS3   leu2Δ::5BS-Gal1pr-CRE-tENO1-LEU2   met15Δ0   ura3Δ::
pTDH3-LoxP-NLS-VP16AD-EL222-tENO1-LoxP-ATAF1-tTDH1-URA3. Lastly, we also used the IB84 strain as a
constitutive 3-color strain to characterize DMD precision. The genotype of this strain is MATa his3Δ1
leu2Δ0::pTDH3-mCerulean-tTDH1-pTDH3-NeonGreen-tTDH1-pTDH3-mScarlet-tTDH1-LEU2   met15Δ0   ura3Δ::
NLS-VP16AD-EL222-URA3.

Culture preparation

195 Cells were grown at 30°C in synthetic medium (SD) consisting of 2% glucose, low fluorescence yeast nitrogen
196 base (Formedium CYN6510), and complete supplement mixture of amino acids and nucleotides (Formedium
197 DCS0019). For each experiment, cells were grown overnight in SC media at 30°C, then diluted 50 times and
198 grown for 4 to 5 hours before being loaded in microfluidic plates.

Microscopy setup, microfluidics and imaging

Images were taken under a Leica DMi8 inverted microscope (Leica Microsystems) with a ×63 oil-immersion
objective (HC PL APO), an LTM200 V3 scanning stage, and an sCMOS camera Zyla 4.2 (ANDOR). Bright-field
images were acquired using a 12V LED light source from Leica Microsystems. Fluorescence images were acquired
using a pE-4000 light source from CoolLED and the following filter cubes: EX:436/20nm DM:455nm
EM:480/40nm (CFP), EX:500/20nm DM:515nm EM:535/30nm (YFP), and EX:546/10nm DM:560nm
EM:585/40nm (RHOD) from Leica Microsystems. Light stimulation was performed using the pE-4000 light source
and the CFP filter. Spatially-resolved illuminations were obtained thanks to a digital mirror device (DMD)
reflecting the light of a pE-4000 light source. We used a MOSAIC3 DMD from ANDOR. The device is used both

for targeted fluorescence imaging and for optogenetic stimulations. A CellASIC ONIX2 system (Merck) was used
together with the Y04C CellASIC microfluidic plates to grow yeast cells in monolayers. Media flow was
maintained by a 7.5 kPa pressure gradient. The media was the same as for pre-culture. The temperature was
maintained at 30 °C by an opaque environmental box and a temperature controller 2000-2, both from PECON.
The microscope was operated using MicroMator.

Model predictive control of gene expression

To compare single-cell and population control strategies, we developed stochastic and deterministic models of
gene expression. Both have been calibrated with respect to the dataset presented in Fig 2B and Fig S4. For
population control, we used the deterministic model, assumed Gaussian measurement noise and used a Kalman
filter for state estimation. Each model assumes a deterministic delay between the time the light signal is applied
and the time protein production is effective. For MPC, fluorescence measurements were taken every 6 minutes
and we considered receding time horizons of 24 minutes. The controller explores the set of light stimulation
profiles in which a 1000ms light stimulation is either applied or not for each measurement time interval, and
selects the profile minimizing mean square deviations. For tracking purposes, brightfield measurements were
taken every 3 minutes. For single-cell control, we used the stochastic model and simulated the cell behavior
using a finite state projection approximation. For each and every cell, state estimation is performed using a
Bayesian approach which conditions the probability distribution for each cell on the most recent measurement,
and light stimulation profiles are selected using the approach outlined above and the expected absolute
deviation as selection criterion. More information is provided in SI Text.

Acknowledgments

228 This work was supported by ANR grants CyberCircuits (ANR-18-CE91-0002), MEMIP (ANR-16-CE33-0018), and
229 Cogex (ANR-16-CE12-0025), by the H2020 Fet-Open COSY-BIO grant (grant agreement no. 766840) and by the
230 Inria IPL grant COSY. We acknowledge the support of the U.S. Department of Energy through the LANL/LDRD
231 Program and the Center for Non Linear Studies for this work.

Author contributions

S.F. developed the MicroMator software and performed experiments. Z.F. wrote the image analysis and real-
time control modules, and analyzed the data. A.F. performed experiments and helped develop the MicroMator
software and analyze the data. C.A. developed strains and helped perform experiments. S.S.-C. developed
strains. S.G. helped with software development. F.B. helped with software development and platform
integration. J.R. helped with controller development. Z.F., F.B., J.R., and G.B. supervised the project. Z.F. and
G.B. wrote the manuscript with input from all authors.

Declaration of interests

240 The authors declare no competing financial interests.

References

1. Eisenstein, M. Smart solutions for automated imaging. *Nat Methods* **17**, 1075–1079 (2020).
2. Conrad, C. *et al.* Micropilot: automation of fluorescence microscopy–based imaging for systems biology. *Nat Methods*
8, 246–249 (2011).
3. Carro, A., Perez-Martinez, M., Soriano, J., Pisano, D. G. & Megias, D. iMSRC: converting a standard automated
microscope into an intelligent screening platform. *Sci Rep* **5**, 10502 (2015).
4. Pinkard, H., Stuurman, N., Corbin, K., Vale, R. & Krummel, M. F. Micro-Magellan: open-source, sample-adaptive,
acquisition software for optical microscopy. *Nat Methods* **13**, 807–809 (2016).
5. Li, T. *et al.* MAARS: a novel high-content acquisition software for the analysis of mitotic defects in fission yeast. *MBoC*
28, 1601–1611 (2017).
6. Toettcher, J. E., Gong, D., Lim, W. A. & Weiner, O. D. Light-based feedback for controlling intracellular signaling
dynamics. *Nat Methods* **8**, 837–839 (2011).
7. Uhlendorf, J. *et al.* Long-term model predictive control of gene expression at the population and single-cell levels. *PNAS*
109, 14271–14276 (2012).

8.   Lugagne, J.-B. *et al.* Balancing a genetic toggle switch by real-time feedback control and periodic forcing. *Nat Commun*
8, 1671 (2017).
9.   Chait, R., Ruess, J., Bergmiller, T., Tkačik, G. & Guet, C. C. Shaping bacterial population behavior through computer-
interfaced control of individual cells. *Nat Commun* **8**, 1535 (2017).
10.   Rullan, M., Benzinger, D., Schmidt, G. W., Milias-Argeitis, A. & Khammash, M. An Optogenetic Platform for Real-Time,
Single-Cell Interrogation of Stochastic Transcriptional Regulation. *Molecular Cell* **70**, 745-756.e6 (2018).
11.   Perrino, G., Wilson, C., Santorelli, M. & di Bernardo, D. Quantitative Characterization of α-Synuclein Aggregation in
Living Cells through Automated Microfluidics Feedback Control. *Cell Reports* **27**, 916-927.e5 (2019).
12.   Perkins, M. L., Benzinger, D., Arcak, M. & Khammash, M. Cell-in-the-loop pattern formation with optogenetically
emulated cell-to-cell signaling. *Nat Commun* **11**, 1355 (2020).
13.   Henriksson, J. *et al.* Endrov: an integrated platform for image analysis. *Nat Methods* **10**, 454–456 (2013).
14.   Binder, J. M. *et al.* Qudi: A modular python suite for experiment control and data processing. *SoftwareX* **6**, 85–90
(2017).
15.   Politi, A. Z. *et al.* Quantitative mapping of fluorescently tagged cellular proteins using FCS-calibrated four-dimensional
imaging. *Nat Protoc* **13**, 1445–1464 (2018).
16.   Pinkard, H. *et al.* Pycro-Manager: open-source software for customized and reproducible microscope control. *Nat*
Methods (2021) doi:10.1038/s41592-021-01087-6.
17.   Edelstein, A., Amodaj, N., Hoover, K., Vale, R. & Stuurman, N. Computer Control of Microscopes Using μManager.
Current Protocols in Molecular Biology **92**, 14.20.1-14.20.17 (2010).
18.   Edelstein, A. D. *et al.* Advanced methods of microscope control using μManager software. *J Biol Methods* **1**, 10 (2014).
19.   Lugagne, J.-B., Lin, H. & Dunlop, M. J. DeLTA: Automated cell segmentation, tracking, and lineage reconstruction using
deep learning. *PLoS Comput Biol* **16**, e1007673 (2020).
20.   Falk, T. *et al.* U-Net: deep learning for cell counting, detection, and morphometry. *Nat Methods* **16**, 67–70 (2019).
21.   Allan, D. *et al. soft-matter/trackpy: Trackpy v0.4.2*. (Zenodo, 2019). doi:10.5281/ZENODO.3492186.
22.   Lee, M. E., DeLoache, W. C., Cervantes, B. & Dueber, J. E. A Highly Characterized Yeast Toolkit for Modular, Multipart
Assembly. *ACS Synth Biol* 12 (2015).
23.   Benzinger, D. & Khammash, M. Pulsatile inputs achieve tunable attenuation of gene expression variability and graded
multi-gene regulation. *Nat Commun* 10 (2018).
284