OCRbayes: A Bayesian hierarchical modeling framework for Seahorse extracellular flux
oxygen consumption rate data analysis

Xiang Zhang[1,2], Taolin Yuan[2], Jaap Keijer[2], Vincent C. J. de Boer[2,*]

1. Theoretical Biology and Bioinformatics, Utrecht University, Padualaan 8, 3584 CH
Utrecht, The Netherlands
2. Human and Animal Physiology, Wageningen University, De Elst 1, 6708 WD
Wageningen, The Netherlands

Abstract

Background: Mitochondrial dysfunction is involved in many complex diseases. Efficient
and accurate evaluation of mitochondrial functionality is crucial for understanding
pathology as well as facilitating novel therapeutic developments. As a popular platform,
Seahorse extracellular flux (XF) analyzer is widely used for measuring mitochondrial
oxygen consumption rate (OCR) in living cells. A hidden feature of Seahorse XF OCR
data is that it has a complex data structure, caused by nesting and crossing between
measurement cycles, wells and plates. Surprisingly, statistical analysis of Seahorse XF
data has not received sufficient attention, and current methods completely ignore the
complex data structure, impairing the robustness of statistical inference.

Results: To rigorously incorporate the complex structure into data analysis, here we
developed a Bayesian hierarchical modeling framework, OCRbayes, and demonstrated its
applicability based on analysis of published data sets.

Conclusions: We showed that OCRbayes can analyze Seahorse XF OCR experimental
data derived from either single or multiple plates. Moreover, OCRbayes has potential to
be used for diagnosing patients with mitochondrial diseases.

Keywords: Oxygen consumption rate; mitochondria; Bayesian; hierarchical modeling;
seahorse extracellular flux

Introduction

27 Mitochondria are double-membrane organelles that are central hubs in regulating energy
28 generation and partitioning. Patients with genetic defects in mitochondrial function are
29 often affected by severe and progressive disease in early life [1]. Furthermore,
30 mitochondrial disorders have also been found to be involved in cardiovascular diseases
31 [2], type II diabetes [3], neurodegenerative disease [4] and cancer [5]. Thus, restoring

mitochondrial function is emerging to be a therapeutic target for both common diseases [6] as well as genetic mitochondrial diseases [7].

Mitochondria produce energy in oxidative phosphorylation (OXPHOS) primarily by transferring electrons along the electron transport chain (ETC) on the inner membrane. Along the ETC, there are four complexes (complex I, II, III and IV), together building up a proton gradient that is ultimately used by ATP synthase to generate ATP. Since electrons are primarily accepted by O_2 to produce H_2O, the OXPHOS activity can be assessed by measuring oxygen consumption rate (OCR). As a reliable and efficient platform, Seahorse XF analyzer provides a multiwell plate based respirometry assay that is widely used to quantify OCR in living cells [8]. Typically, the Seahorse XF analyzer measures the OCR of cells in a 96-well plate under different ETC/OXPHOS perturbation scenarios, used for assessing mitochondrial functionality such as maximal respiration, leak respiration and ATP-linked respiration [9].

A typical Seahorse assay includes three measurement cycles for each phase. Every measurement cycle starts by lowering the cartridge and creating a temporary semi-closed $\sim 2\ \mu L$ chamber [10]. During a measurement cycle, fluorescent oxygen sensors capture oxygen concentration changes in the chamber and outputs OCR. The Seahorse XF analyzer measures OCR in tens to hundreds of thousands of cells per well and typically requires 4–5 replicate wells per experimental group, allowing the analysis of multiple experimental groups in one plate [11]. Since measurement cycles are nested within phases, phases are crossed with wells, and wells are nested in experimental groups, a very complex structure is inherently embedded in the Seahorse XF OCR data. Surprisingly, current Seahorse XF data analyses most often ignore this complex structure, and by default, data sets are often chopped into subgroups followed by performing ANOVA-like statistical tests. Although there are advanced tools developed for Seahorse data analysis [12,13], none of them comprehensively take the complexity of the data structure into account. As a result, OCR variation between measurement cycles, replicate wells and replicate plates is overlooked, impairing the robustness of the interpretation of Seahorse XF OCR outcomes, and eventually the development of mitochondrial targeted therapies as well as our understanding of mitochondrial biology.

A natural way to incorporate the complex structure into data analysis is to use hierarchical modeling, which has been extensively developed for analyzing gene expression data [14,15]. Here we developed a Bayesian hierarchical modeling framework, OCRbayes, for the Seahorse XF OCR data analysis. Compared to the currently most advanced Seahorse data analysis tool, OCR-stats which is a frequentist approach, OCRbayes is based on a fully Bayesian approach due to its flexibility and convenience for constructing the hierarchical models. To demonstrate the applicability of our approach as well as its potential implication for mitochondrial disease diagnostics, we applied OCRbayes to a publicly available OCR data set [13], which contains over 200

Seahorse experiments performed on human fibroblasts derived from patients with
mitochondrial diseases and controls.

Method

OCRbayes: A Bayesian hierarchical modeling framework for
Seahorse OCR data analysis

In order to incorporate the complex data structure into the analysis of Seahorse XF OCR
measurements, we developed a Bayesian hierarchical modeling framework, OCRbayes.
In this study, we focused on experimental data containing two groups, such as patient and
control group. From our perspective, the Seahorse OCR data include three levels,
including 1) measurement cycle, 2) well and 3) plate.

During a measurement cycle, the Seahorse XF analyzer uses fluorescent oxygen sensors
to track OCR. For every interval, multiple measurement cycles are performed in order to
accurately measure the OCR. A typical Seahorse assay contains four intervals. The first
interval refers to initial phase, and the second, third, fourth interval refer to phase after
injecting oligomycin (blocking proton translocation through ATP synthase), FCCP
(allowing protons to move into the mitochondrial matrix independent of the ATP
synthase) and antimycin/rotenone (inhibiting complex I and III, shutting down
mitochondrial respiration).
The variation between the OCR values within an interval is the between measurement
cycle variation. Since OCR values must be positive, we used a lognormal distribution
with the true OCR value at the log scale ($\log OCR_{true}$) and the between measurement
cycle standard deviation (also at the log scale, $\log OCR_{sd}$) to model each observed OCR
value (OCR_{obs}).

$$OCR_{obs}[i] \sim Lognormal(\log OCR_{true}[P[i],W[i],I[i]], \log OCR_{se}[P[i],I[i]]) (1)$$

OCR_{obs} is a vector of length $N_{plate} \times N_{well} \times N_{interval} \times N_{measurement}$, where N_{plate}, N_{well},
$N_{interval}$ and $N_{measurement}$ are number of plates, wells, intervals and measurement cycles,
respectively. The $\log OCR_{true}$ refers to N_{plate} layers of matrices. Each layer is a matrix
with N_{well} rows and $N_{interval}$ columns. $\log OCR_{sd}$ is a matrix of N_{plate} rows and $N_{interval}$
columns.

In a typical Seahorse XF assay, one cell line undergoing the same experimental treatment
is seeded in multiple wells. However, the OCRs in these replicate wells will not be the
same, and the difference is called between well variation. One obvious reason causing the
between well variation is that the number of cells in these replicate wells are not
identical. To adjust for the effect of cell number difference, we modeled the true OCR

value as a function of the cell number, and from there estimated $\text{OCR}_{\text{per 1k cells}}$, which
represents OCR value per 1000 cells that received the same treatment (or from the same
group) and injection. In addition to cell number difference, technical, procedural or
instrumental noise can also contribute to between well variation. We captured this well-
to-well variation after accounting for cell number difference with the residual parameter
σ_{well}.

$$\text{logOCR}_{\text{true}}[P[i],W[i],I[i]]$$
$$\sim \text{Normal}(\text{OCR}_{\text{per 1k cells}}[P[i],G[i],I[i]] \times \text{N}_{\text{cells}}[W[i]], \sigma_{\text{well}}[P[i],G[i],I[i]])(2)$$

$\text{OCR}_{\text{per 1k cells}}$ is a three-dimensional matrix with N_{plate} layers, and each layer has N_{group}
rows and $\text{N}_{\text{Interval}}$ columns. N_{cells} is a vector of length N_{well}, and every entry represents
the cell number in that well.

Apart from the technical replicates in one plate, the biological insight for a (patient) cell
line or a specific condition is generally validated by repeating the Seahorse assay on
different days. As a result, OCR data are distributed on more than one plate. Due to batch
effects such as plating, culturing or environmental differences between time and
laboratories, OCR measurements will differ between plates. To take into account the
between plate variation, we used another lognormal distribution with the logarithm
transformed OCR value per 1000 cells, $(\mu_{\text{OCR}_{\text{per 1k cells}}})$ and the between plate standard
deviation (σ_{plate}) to model the OCR value per 1000 cells $\text{OCR}_{\text{per 1k cells}}$.

$$\text{OCR}_{\text{per 1k cells}}[P[i],G[i],I[i]]$$
$$\sim \text{Lognormal}(\log(\mu_{\text{OCR}_{\text{per 1k cells}}}[G[i], I[i]]), \sigma_{\text{plate}}[G[i], I[i]])(3)$$

Bayesian inference

OCRbayes focuses on calculating posterior distributions for OCR per 1000 cells in
various experimental conditions $(\text{P}(\mu_{\text{OCR}_{\text{per 1k cells}}}|\text{OCR}_{\text{obs}}))$. The posterior distributions
were combinations of prior distributions $(\text{P}(\text{logOCR}_{\text{se}}), \text{P}(\sigma_{\text{well}}), \text{P}(\sigma_{\text{plate}})$ and
$\text{P}(\mu_{\text{OCR}_{\text{per 1k cells}}}))$, and the likelihood function
$(\text{P}(\text{OCR}_{\text{obs}}|\text{N}_{\text{cell}}, \mu_{\text{OCR}_{\text{per 1k cells}}}, \text{logOCR}_{\text{sd}}, \sigma_{\text{well}}, \sigma_{\text{plate}}))$

In this study, since our case studies focused on human fibroblast cells, we used
informative prior distributions for $\text{logOCR}_{\text{sd}}, \sigma_{\text{well}}, \sigma_{\text{plate}}$ and $\mu_{\text{OCR}_{\text{per 1k cells}}}$. Since values of
all these parameters must be positive, we used four lognormal distributions. We applied
maximum likelihood estimation to calculate the lognormal distribution parameters based
on the Seahorse XF OCR data in OCR-stats [13] by running the fitdistr function built in
the MASS r package. In this study, our prior distributions are

$$\begin{aligned}
\text{logOCR}_{\text{sd}} &\sim \text{Lognormal}(-3.23, 0.79) \\
\sigma_{\text{well}} &\sim \text{Lognormal}(-1.62, 0.53) \\
\sigma_{\text{plate}} &\sim \text{Lognormal}(-1.18, 0.05) \\
\mu_{\text{OCR}_{\text{per 1k cells}}} &\sim \text{Lognormal}(0.3, 0.79)
\end{aligned}$$

133   The Bayesian multi-level model of OCRbayes was implemented in Stan (version 2.19.3)
134   [16]. We fitted the model by running Hamiltonian Markov Chain Monte Carlo. We ran
135   four Markov chains with 2000 iterations in each chain. The code can be found at
136   https://github.com/XiangZhangSC/seahorse. We also provided the R code of OCRbayes
137   in the supplementary file.

Calculation of bioenergetic measures

139   Based on OCR (per 1000 cells), we calculated various bioenergetic measures, such as
140   basal respiration, proton leak, ATP-linked respiration, spare respiratory capacity and
141   maximal respiration. These bioenergetic measures are defined as below.

$$\text{Basal respiration} = \text{OCR}_{\text{per 1k cells, initial}} - \text{OCR}_{\text{per 1k cells, antimycin/rotenone}} \quad (4)$$

$$\text{ATP-linked respiration} = \text{OCR}_{\text{per 1k cells, initial}} - \text{OCR}_{\text{per 1k cells, oligomycin}} \quad (5)$$

$$\text{Proton leak} = \text{OCR}_{\text{per 1k cells, oligomycin}} - \text{OCR}_{\text{per 1k cells, antimycin/rotenone}} \quad (6)$$

$$\text{Spare respiratory capacity} = \text{OCR}_{\text{per 1k cells, FCCP}} - \text{OCR}_{\text{per 1k cells, initial}} \quad (7)$$

$$\text{Maximal respiration} = \text{OCR}_{\text{per 1k cells, FCCP}} - \text{OCR}_{\text{per 1k cells, antimycin/rotenone}} \quad (8)$$

Human fibroblast OCR data

To benchmark OCRbayes as well as illustrate how OCRbayes can be used for analyzing
OCR in patients with mitochondrial diseases, we used the OCR data set provided in
Yepez et al. [13]. This data set contains Seahorse OCR measurements from 203 human
fibroblast cell lines that have been assayed in 126 plates. Normal human dermal
fibroblast (NHDF) reference cell lines were used as controls in all plates. The other 202
cell lines were derived from patients suspected to suffer from rare mitochondrial diseases.
Among these 202 cell lines, 26 fibroblast cell lines were measured in multiple plates. We
used the 176 patient fibroblast cell lines that were assayed in a single plate as well as the
control cell line in the same plates for estimating prior distributions. The other 26
fibroblast cell lines were used for benchmarking OCRbayes. Among the 202 cell lines
that were analyzed, Yepez et al. [13] labeled 6 patient cell lines as positive controls that
have shown statistically significant reduction in maximum respiration. Meanwhile, Yepez
et al. also labeled another two patient cell lines as negative controls, since these cell lines
did not show changes in OCR in earlier experiments [13].

We processed the original data by removing wells in which single or more OCR
measurements were missing. After filtering, we used 176 patient cell lines together with
the NHDF control cell line on 78 plates for estimating the prior distributions for between
measurement cycle variation ($\log OCR_{sd}$) and between well variation (σ_{well}). To estimate
the prior distribution for between plate variation (σ_{plate}) and mean OCR per 1000 cells
($\mu_{OCR_{per\ 1k\ cells}}$), we used the OCR values from the NHDF cell lines that were plated in all
78 plates.

Statistical analysis

For benchmarking OCRbayes, we compared the patient cell lines to the control cell line
(NHDF). Since the maximal respiration was reported as the primary mitochondrial
dysfunction outcome in the Yepez et al. study [13], we reported here the mean log2 fold
change of maximal respiration together with the False Discovery Rate (FDR). FDR was
calculated based on the posterior error probability ($1 - P(\log_2(\frac{patient}{control}|OCR_{obs}) < 0)$),
where $P(\log_2(\frac{patient}{control}|OCR_{obs})$ represented the posterior distribution of fold change after
fitting the Bayesian model to the experimental OCR data. If the FDR was below 0.05, we
considered that the difference between patient and control cell line was statistically
significant.

Results

OCRbayes: from OCR to respiration metric difference

To demonstrate the applicability, we applied OCRbayes to analyze OCR data derived
from two patient cell lines with known mutations in either *BOLA3* or *PET100* gene.
These two genes encode proteins that are essential for biogenesis or assembly of
mitochondrial complexes [17–19]. We compared the patient cell line to the control cell
lines within their respective plates. These two cell lines were both assayed on two plates
on two different days. For both cell lines, we observed a clear decrease in maximal
respiration compared to the control cell lines in both plates (Figure 1A and 1E).
Meanwhile, we noticed that the range of OCR values in the two plates used for profiling
BOLA3 patient cell line differed considerably (Figure 1A). In particular, maximal OCR
value in the first plate was around 200 pmol/min, whereas in the second plate the
maximal OCR value was around 100 pmol/min. In contrast, the range of OCR values in
the two plates used for profiling *PET100* were similar to each other. In addition, it was
obvious that there was considerable variation between the replicate wells in both plates
for both cell lines.

By applying OCRbayes, we combined the two plates for each cell line, calculated the
posterior distributions for OCR per 1000 cells during the four intervals of a typical
Seahorse assay, including initial phase without injection (Int1), oligomycin phase (Int2),
FCCP phase (Int3) and antimycin/rotenone phase (Int4) (Figure 1B and 1F). Next, the
calculated posterior OCR per 1000 cells for the four intervals were transformed into the
various respiration metrics, including ATP-linked respiration, basal respiration, maximal
respiration, proton leak and spare respiratory capacity (Figure 1C and 1G). In the last
step, we compared the respiration metrics in patient cell line to the control cell line, and
calculated the posterior distribution for log2(fold change) (Figure 1D and 1H). We
observed that *BOLA3* patient cell line showed reduction in basal respiration (posterior
mean log2(fold change) -0.974, 95% credible interval [-1.93, -0.0970]) and maximal
respiration (-1.06 [-1.87,-0.228]), compared to the control cell line (Figure 1D). On the
other hand, we observed no difference in ATP-linked respiration (-1.08 [-2.38, 0.117]),
proton leak (-0.630 [-3.87, 2.32]) and spare respiratory capacity (-1.26 [-4.41,1.62])
between *BOLA3* patient cell line and the control cell line (Figure 1D). Meanwhile we
observed lower ATP-basal respiration (-1.75 [-3.18, -0.546]), basal respiration (-1.58 [-
2.52, -0.667]) and maximal respiration (-1.28 [-2.10, -0.436]) in the *PET100* patient cell
line than the control cell line (Figure 1H). We observed no difference in proton leak (-
0.906 [-4.14,2.54]) and spare respiratory capacity (-0.826 [-3.29, 2.34]) in the *PET100*
patient cell line compared to the control cell line (Figure 1H).

OCRbayes accounts for various technical variations during Seahorse XF OCR data analysis

OCR measurements generated by the Seahorse XF analyzer were affected by various
technical variations, including between measurement cycle variation, between well
variation and between plate variation. To visualize the technical variations, we used the
cell line with a genetic mutation in *NSUN3* gene (patient cell line 76065). This cell line
was measured on five different plates, allowing us to visualize all three technical
variations.

OCRbayes calculated the between measurement cycle variation for each plate during
each interval. For each cell line in each plate, OCRbayes calculated the between well
variation during each interval. Regarding the between plate variation, OCRbayes
estimated it for each cell line during each interval. All the variation values were on the
log scale. We observed that for both patient and control cell line in the initial phase (Int1)
and FCCP phase (Int3), the between plate variation was larger than the between well
variation, which itself was larger than the between measurement cycle variations (Figure
2). However, in the oligomycin phase (Int2) and antimycin/rotenone phase (Int4), the
between plate variation was not always larger than the other two technical variations
(Figure 2). In particular, we observed that the between well variation was larger than the

between measurement cycle and the between plate variation in three plates during the
antimycin/rotenone phase (Int4) in the patient cell line.

Benchmark OCRbayes

To demonstrate that OCRbayes works properly, we applied it to analyze the published
Seahorse XF OCR data set containing 26 patient cell lines as well as a control cell line
reported by Yepez et al. [13]. This data set contains 6 cell lines that were labeled as
positive controls and 2 cell lines that were labeled as negative controls (as explained in
the material and methods section).

Based on our analysis, we found 6 patient cell lines that had lower maximal respiration
compared to the control cell line with False Discovery Rate (FDR) below 0.05 (Figure
3A and B). Among the 6 patient cell lines that showed statistically significant reduction
in maximal respiration compared to the control cell line, the patient cell line 73387
(mutation in *PET100* gene) showed the largest effect whereas the patient cell line 76065
(mutation in *NUSN3* gene) showed the smallest effect (Figure 3A). Five of these 6 patient
cell lines were labeled as positive controls in the original study [13]. Meanwhile, the two
negative controls (patient cell lines 73901 and 91410) had FDR above 0.05 in our
analysis.

Our analysis successfully recalled both negative controls and five out of six positive
controls. Interestingly, our analysis showed that there was no significant reduction of
maximal respiration in the cell line with a genetic mutation in *SFXN4* gene (patient
61818) that was previously labeled as positive control [13]. This patient cell line was
measured on three independent plates on three different days. The range of OCR values
of these three plates differed, indicating considerable between plate variation (Figure
4A). The OCR profiles of the patient and control cell line were overlapping in the first
plate (Figure 4A). In contrast, the OCR profiles derived from the second and third plate
showed OCR decreasing in the patient cell line compared to the control cell line,
especially during the FCCP phase (Figure 4A). Meanwhile, we observed considerable
variation between the replicate wells as well as measurement cycles in these Seahorse
assays (Figure 4A).
In addition, the between measurement cycle variation in the third plate seemed to be
larger than the other two plates (Figure 4A). We analyzed these three repeated
experiments separately as well as combined them together. Our separate analysis showed
that none of the three experiments showed significant reduction in maximal respiration in
the *SFXN4* patient cell line (Figure 4B). The posterior mean log2(fold change) and the
corresponding 95% credible intervals derived from the first, second and third Seahorse
assay were 0.122 [-0.834, 1.04], -0.431 [-1.40, 0.499] and -0.0977 [-1.04, 0.827],
respectively. When we used OCRbayes to analyze all the three plates together, we found
that the posterior mean log2(fold change) of maximal respiration was -0.200 [-0.877,

0.500] compared to the control cell line (Figure 4C). However, this tendency of reduction
in maximal respiration in *SFXN4* patient cell line was not statistically significant.

268　OCRbayes can be used to evaluate the probability that a
269　patient fibroblast cell line has an abnormality in mitochondrial
270　respiration.

A feature that makes OCRbayes unique from other methods is that OCRbayes can
evaluate what is the probability that a patient has abnormality in his or her fibroblast
mitochondrial respiration based on a single Seahorse assay or multiple Seahorse assays.
To demonstrate this feature, we used two patient cell lines that showed significant
reduction in maximal respiration in our analysis. One patient cell line was the *PET100*
gene mutation fibroblast (patient cell line 73387) and the other patient cell line was
NSUN3 gene mutation fibroblast (patient cell line 76065). Among the 6 patient cell lines
that showed statistically significant reduction in maximal respiration compared to the
control cell line, the *PET100* and *NSUN3* cell line showed the largest and smallest effect,
respectively (Figure 3B).

The *PET100* patient cell line was measured on two plates. Based on the first assay, our
analysis showed that the posterior probability of this *PET100* mutation carrier having
lower maximal respiration than the control was 97.8%. By repeating the experiment once
more, the posterior probability increased to 99.8% (Figure 5A). In contrast, the cell line
with mutation in *NSUN3* gene (patient cell line 76065) was measured on five different
plates. Our analysis showed that based on the first assay, the posterior probability of this
NSUN3 mutation carrier having lower maximal respiration than the control was 41.3%.
Repeating the assay once increased the posterior probability from 41.3% to 73.4%. When
the assay was repeated for the third, fourth and fifth time, the corresponding posterior
probabilities increased to 78.7%, 87.7% and 90.0%, respectively (Figure 5B). In
summary, we demonstrate that OCRbayes can be used to evaluate the probability that a
patient fibroblast cell line has an abnormality in mitochondrial respiration based on a
single Seahorse assay or multiple Seahorse assays.

294　Discussion

Although Seahorse XF analyzer is widely used in bioenergetic profiling, its data analysis
has not received sufficient attention. A hidden feature of Seahorse XF OCR data is that it
contains a complex data structure. The complex data structure is due to the fact that
measurement cycles are nested within injections, injections are crossed with wells, and
wells are nested in plates. As far as we know, currently there is no data analysis protocol
that takes into account this complex data structure, impairing the robustness of Seahorse

XF OCR data analysis outcomes. This is because when one ignores the data structure, one also ignores the variations between measurement cycles, between wells and between plates. In order to make the Seahorse data analysis more robust, in this study we developed a Bayesian hierarchical modeling approach, OCRbayes, which accounts for all these technical variations during the data analysis.

Seahorse XF OCR measurements are noisy

An OCR value is determined not only by mitochondrial activity, but also by technical noise including 1) between measurement cycle variation, 2) between well variation and 3) between plate variation.

Every phase typically contains three measurement cycles, resulting in three OCR values. Since every measurement cycle starts with a "mix and wait" step to ensure the same baseline of cell values, cell physiology should not substantially change within a phase. Existing tools such as OCR-stats [13] and SHORE [12] use different strategies to select a single data point to represent an injection phase. When different data points were chosen, one can get different outcomes, making the current Seahorse XF data analysis less robust. To avoid this ambiguity, our approach did not select any particular data point, instead modeled all three OCR data. By doing so, we incorporated the uncertainty about "which data point should I choose?" into the data analysis and focused on the average behavior.

The between well variation refers to a common observation that OCR values differ among the replicate wells. One important factor leading to the variation between replicate wells is that cell numbers are not identical in these wells. The wells with more cells would have higher OCR than the wells containing fewer cells. Fortunately, cell number in each well can be quantified experimentally and used for the data analysis [20,21]. In addition to cell number difference, initial conditions, treatment concentration, or fluorophore sleeve calibration can also contribute to variation between wells. OCRbayes also takes into account the between well variation caused by these unobserved factors.

Between plate variation takes place when the same Seahorse experiment is repeated on different days and on more than one plate. Due to differences in temperature, seeding time, growth time, growth medium, sensor cartridge as well as treatment efficiency, the OCR outcomes will differ between plates [13,22]. Often the between plate variation is assumed to be the dominant technical variation involved in Seahorse OCR data. Based on our analysis, we showed that this assumption may be appropriate for OCR measurements derived from the initial and FCCP phase, but may not work for the OCR values derived from oligomycin and antimycin/rotenone phase. OCR values in the oligomycin and antimycin/rotenone phase were very small and possibly close to the detection limit of the Seahorse XF in a well. Thus, it is more challenging to accurately measure the OCR in these phases.

Comparison with other statistical tool for Seahorse XF OCR data analysis

In this work, we compared the maximal respiration in 26 patient cell lines to the control cell line individually as what was done in OCR-stats [13]. Overall our analysis recalled successfully all negative controls and five out of six positive controls. Patient cell line 61818 was labeled as a positive control since this patient was found having a mutation in *SFXN4* gene. A recent study based on erythroleukemic cell line showed that *SFXN4* knockout resulted in significant decrease in all parameters of respiration, including baseline respiration, respiratory ATP synthesis, maximal respiration, and spare respiratory capacity [23]. However, our analysis showed that the maximal respiration of this patient was not significantly different from the control cell line. Our further analysis showed that the patient cell line only showed a tendency of having lower maximal respiration than control in one of the three repeated experiments.

We also noticed that other outcomes were also not identical as what was presented in the OCR-Stats publication [13]. Strikingly, in our analysis *PET100* mutation fibroblast (patient cell line 73387) showed a significant decrease in maximal respiration compared to the control cell line. However, in the OCR-stats [13], the difference in maximal respiration was not statistically significant in this patient. This patient was diagnosed carrying a homozygous loss of function mutation in the *PET100* gene, which encodes a mitochondrial complex IV biogenesis factor [13,17,18]. A homozygous truncating variant (c.142C>T, p.(Gln48*)) in the *PET100* gene was found to lead to a complete loss of enzyme activity, and caused deficiency in complex IV [24]. Therefore, our observation of significant reduction in maximal respiration in *PET100* mutation fibroblast based on the OCRbayes is in line with the observed loss of enzymatic activity in patients carrying the genetic mutations in *PET100*. This difference in analysis outcome highlighted the advantage of OCRbayes which used hierarchical models to incorporate the complex data structure into Seahorse OCR data analysis, helping separate the technical variations from the OCR measurements and identify difference in the biological OCR.

OCRbayes is potentially used for screening patients with mitochondrial diseases

Since an increase in proton leak or a decrease in basal or maximal respiration are indicators of mitochondrial dysfunction [25], Seahorse XF is potentially to be used for screening mitochondrial disease patients. Besides providing solely statistical significance information as other methods do, OCRbayes allows us to calculate posterior probability that the maximal respiration (or any other respiration metrics) was abnormal in a patient even based on a single Seahorse assay. This feature is also helpful for deciding whether we need to run a single or multiple Seahorse assays for a patient.

For example, the patient cell line with mutation in *PET100* gene showed the largest
decrease in maximal respiration in our analysis. The posterior probability for this patient
having abnormal mitochondrial respiration after observing a single Seahorse assay was
already about 98%. It is reasonable to run just a single assay in this case. In contrast, the
cell line with mutation in *NSUN3* gene had the smallest effect size among all the patient
cell lines that showed significant reduction in maximal respiration. The posterior
probability of this patient having impaired maximum respiration after running a single
Seahorse assay was modest (about 40%). Repeating the experiment once more increased
the posterior probability from 40% to 70%. Thus in this case, it is beneficial to run
multiple Seahorse assays. We think that the posterior probability is a useful metric to help
scientists to decide whether to perform extra Seahorse assays on the patient cell lines.
The ability of posterior probability calculation by OCRbayes not only allows to make
better conclusions about the significance of the effect of the experimental perturbation, it
can also prevent from repeating unnecessary, often expensive, experiments. Furthermore,
when the availability of patient material or the number of target cells (specific isolated
immune cell subsets) is limited, OCRbayes is valuable to exploit the limited data and
facilitate proper validation the experiments.

Strengths and limitations

OCRbayes has several advantages. The first advantage of our approach is that it
incorporates various technical variations including between measurement cycle variation,
between well variation as well as between plate variation during the estimation of
bioenergetic measures. All current methods need to choose a single data point from the
three measurement cycles to represent the OCR during a particular phase. This procedure
ignores the uncertainty and makes analysis less robust because different choice of data
points may lead to different results. The second advantage is that OCRbayes can calculate
posterior probability for difference in various bioenergetic measures based on Seahorse
OCR data consisting of a single plate or multiple plates. This is a useful feature for
screening samples derived from patients with mitochondrial diseases. A third advantage
of OCRbayes is that it can be used for guiding improvement of experimental protocols
for running Seahorse assays, because the hierarchical model can explicitly quantify the
changes in technical variations resulted from different protocols.

One limitation of this study is that OCRbayes is extensively tested with experimental data
derived from human fibroblasts cell lines. This is because our model development was
restricted to the limited publicly available Seahorse OCR data sets. To generalize our
statistical method to other cell lines than human skin fibroblasts, the prior distributions of
those cell lines should be estimated from previously collected Seahorse OCR data. This is
because each cell line or cell system has its own growth conditions and metabolic
characteristics, making the prior distributions based on the human fibroblasts data not

suitable for other cell lines. Furthermore, the prior distributions for plate-to-plate or well-
to-well variation are likely dependent on factors such as laboratory and machine.
Although many Seahorse studies have been published, few of them provided the raw
OCR measurements together with the cell number quantification information.
Generalization of our method would benefit from the open availability of raw Seahorse
OCR data. This would also facilitate making OCRbayes applicable for other cell lines
than human skin fibroblasts.

Declarations

Ethics approval and consent to participate

Not applicable

Consent for publication

Availability of data and materials

The datasets analysed during the current study are available in the Github repository,
https://github.com/XiangZhangSC/seahorse.

Competing interests

The authors declare that they have no competing interests

Funding

Not applicable

Author's contributions

XZ developed the OCRbayes and wrote the manuscript. TY and JK contributed to the
conceptualization of the project and edited the manuscript. VB supervised the project and
edited the manuscript. All authors read and approved the final manuscript.

Acknowledgements

Not applicable

Reference

1. Koene S, Wortmann SB, Vries MC de, Jonckheere AI, Morava E, Groot IJM de, et al.
Developing outcome measures for pediatric mitochondrial disorders: which complaints

and limitations are most burdensome to patients and their parents? Mitochondrion.
2013;13: 15–24. doi:10.1016/j.mito.2012.11.002

2. Ballinger SW. Mitochondrial dysfunction in cardiovascular disease. Free Radic Biol
Med. 2005;38: 1278–95. doi:10.1016/j.freeradbiomed.2005.02.014

3. Szendroedi J, Phielix E, Roden M. The role of mitochondria in insulin resistance and
type 2 diabetes mellitus. Nat Rev Endocrinol. 2011;8: 92–103.
doi:10.1038/nrendo.2011.138

4. Johri A, Beal MF. Mitochondrial dysfunction in neurodegenerative diseases. J
Pharmacol Exp Ther. 2012;342: 619–30. doi:10.1124/jpet.112.192138

5. Erez A, DeBerardinis RJ. Metabolic dysregulation in monogenic disorders and cancer -
finding method in madness. Nat Rev Cancer. 2015;15: 440–8. doi:10.1038/nrc3949

6. Murphy MP, Hartley RC. Mitochondria as a therapeutic target for common
pathologies. Nat Rev Drug Discov. 2018;17: 865–886. doi:10.1038/nrd.2018.174

7. Weissig V. Drug Development for the Therapy of Mitochondrial Diseases. Trends Mol
Med. 2020;26: 40–57. doi:10.1016/j.molmed.2019.09.002

8. Wu M, Neilson A, Swift AL, Moran R, Tamagnine J, Parslow D, et al. Multiparameter
metabolic analysis reveals a close link between attenuated mitochondrial bioenergetic
function and enhanced glycolysis dependency in human tumor cells. Am J Physiol Cell
Physiol. 2007;292: C125–36. doi:10.1152/ajpcell.00247.2006

9. Chacko BK, Kramer PA, Ravi S, Benavides GA, Mitchell T, Dranka BP, et al. The
Bioenergetic Health Index: a new concept in mitochondrial translational research. Clin
Sci (Lond). 2014;127: 367–73. doi:10.1042/CS20140101

10. Gerencser AA, Neilson A, Choi SW, Edman U, Yadava N, Oh RJ, et al. Quantitative
microplate-based respirometry with correction for oxygen diffusion. Anal Chem.
2009;81: 6868–78. doi:10.1021/ac900881z

11. Artyomov MN, Van den Bossche J. Immunometabolism in the Single-Cell Era. Cell
Metab. 2020;0. doi:10.1016/j.cmet.2020.09.013

12. Nicholas D, Proctor EA, Raval FM, Ip BC, Habib C, Ritou E, et al. Advances in the
quantification of mitochondrial function in primary human immune cells through
extracellular flux analysis. PLoS One. 2017;12: e0170975.
doi:10.1371/journal.pone.0170975

13. Yépez VA, Kremer LS, Iuso A, Gusic M, Kopajtich R, Koňaříková E, et al. OCR-
Stats: Robust estimation and statistical testing of mitochondrial respiration activities

using Seahorse XF Analyzer. PLoS One. 2018;13: e0199938.
doi:10.1371/journal.pone.0199938

14. Robinson MD, McCarthy DJ, Smyth GK. edgeR: a Bioconductor package for
differential expression analysis of digital gene expression data. Bioinformatics. 2010;26:
139–40. doi:10.1093/bioinformatics/btp616

15. Love MI, Huber W, Anders S. Moderated estimation of fold change and dispersion
for RNA-seq data with DESeq2. Genome Biol. 2014;15: 550. doi:10.1186/s13059-014-
0550-8

16. Carpenter B, Gelman A, Hoffman MD, Lee D, Goodrich B, Betancourt M, et al. Stan
: A Probabilistic Programming Language. J Stat Softw. 2017;76.
doi:10.18637/jss.v076.i01

17. Oláhová M, Hardy SA, Hall J, Yarham JW, Haack TB, Wilson WC, et al. LRPPRC
mutations cause early-onset multisystem mitochondrial disease outside of the French-
Canadian population. Brain. 2015;138: 3503–19. doi:10.1093/brain/awv291

18. Lim SC, Smith KR, Stroud DA, Compton AG, Tucker EJ, Dasvarma A, et al. A
founder mutation in PET100 causes isolated complex IV deficiency in Lebanese
individuals with Leigh syndrome. Am J Hum Genet. 2014;94: 209–22.
doi:10.1016/j.ajhg.2013.12.015

19. Haack TB, Rolinski B, Haberberger B, Zimmermann F, Schum J, Strecker V, et al.
Homozygous missense mutation in BOLA3 causes multiple mitochondrial dysfunctions
syndrome in two siblings. J Inherit Metab Dis. 2013;36: 55–62. doi:10.1007/s10545-012-
9489-7

20. Little AC, Kovalenko I, Goo LE, Hong HS, Kerk SA, Yates JA, et al. High-content
fluorescence imaging with the metabolic flux assay reveals insights into mitochondrial
properties and functions. Commun Biol. 2020;3: 1–10. doi:10.1038/s42003-020-0988-z

21. Janssen JJE, Lagerwaard B, Bunschoten A, Savelkoul HFJ, Neerven RJJ van, Keijer
J, et al. Novel standardized method for extracellular flux analysis of oxidative and
glycolytic metabolism in peripheral blood mononuclear cells. Sci Rep. 2021;11: 1662.
doi:10.1038/s41598-021-81217-4

22. Koopman M, Michels H, Dancy BM, Kamble R, Mouchiroud L, Auwerx J, et al. A
screening-based platform for the assessment of cellular respiration in Caenorhabditis
elegans. Nat Protoc. 2016;11: 1798–816. doi:10.1038/nprot.2016.106

23. Paul BT, Tesfay L, Winkler CR, Torti FM, Torti SV. Sideroflexin 4 affects Fe-S
cluster biogenesis, iron metabolism, mitochondrial respiration and heme biosynthetic
enzymes. Sci Rep. 2019;9: 19634. doi:10.1038/s41598-019-55907-z

24. Oláhová M, Haack TB, Alston CL, Houghton JA, He L, Morris AA, et al. A
truncating PET100 variant causing fatal infantile lactic acidosis and isolated cytochrome
c oxidase deficiency. Eur J Hum Genet. 2015;23: 935–9. doi:10.1038/ejhg.2014.214

25. Brand MD, Nicholls DG. Assessing mitochondrial dysfunction in cells. Biochem J.
2011;435: 297–312. doi:10.1042/BJ20110162

Figures

Figure 1 (A) and (E) are OCR profiles for patient cell lines that have a genetic mutation in either BOLA3 (A) or PET100 (E), compared to control cell lines in two plates. (B) and (F) are posterior distributions for OCR per 1000 cells in BOLA3 (B) and PET100 (F) mutated patient cell lines during initial phase (Int1), oligomycin phase (Int2), FCCP phase (Int3) and antimycin/rotenone phase (Int4). (C) and (G) are posterior distributions for respiration metrics in BOLA3 (C) and PET100 (G) mutated patient cell lines and control cell lines. (D) and (H) are posterior distributions of log2 fold change in the respiration metrics between patient and control cell line

Figure 2 Technical variations in Seahorse XF OCR data of experiments based on the
patient cell line with mutation in NSUN3 gene and the control cell line. Int1, Int2, Int3
and Int4 represent initial, oligomycin, FCCP and antimycin/rotenone phase. All the
variations were on log scale. The dots are the posterior mean and the line segment
represent the 95% credible interval.

Figure 3 Maximal respiration change patient vs. control on multiple plates. (A) average log2 fold change (y-axis) of maximal respiration of all cell lines repeated across plates (x-axis) and their respective controls, sorted by the False Discovery Rate (FDR). Red and blue dots represent positive and negative controls, respectively. (B) similar to (A), but depicting FDR. Red dashed line represents FDR = 0.05.

Figure 4 (A) Raw OCR values for the cell line with genetic mutation in SFXN4 gene (patient cell line 61818) in three repeated plates. (B) Posterior distributions of maximal

539   *respiration of the cell line with genetic mutation in SFXN4 gene (patient cell line 61818)*
540   *and the control cell line in the same plate. (C) Posterior distributions of maximal*
541   *respiration of the cell line after combing all three repeated experiments.*

Figure 5 (A) Posterior probability of decreasing maximal respiration in patient cell line
with mutation in PET100 gene compared to the control cell line in the scenario with 1 or
2 plates. (B) Posterior probability of decreasing maximal respiration in patient cell line
with mutation in NUSN3 gene compared to the control cell line in the scenario with 1, 2,
3, 4 or 5 plates