**High Gamma and Beta Temporal Interference Stimulation in the Human Motor Cortex**

**Improves Motor Functions**

*Ru Ma, Xinzhao Xia, Wei Zhang, Zhuo Lu, Qianying Wu, Jiangtian Cui, Hongwen Song, Chuan Fan, Xueli Chen, Junjie Wei, Gongjun Ji, Kai Wang, Xiaoxiao Wang\*, Bensheng Qiu\*, Xiaochu Zhang\**

R. Ma, W. Zhang, Q. Wu, Dr. H. Song, X. Chen, Prof. X. Zhang

Department of Radiology

the First Affiliated Hospital of USTC

Hefei National Laboratory for Physical Sciences at the Microscale and Division of Life

Science and Medicine

University of Science and Technology of China

Hefei, Anhui 230027, China

E-mail: zxcustc@ustc.edu.cn

X. Xia, Z. Lu, J. Cui, C. Fan, Dr. X. Wang, Prof. B. Qiu, Prof. X. Zhang

Centers for Biomedical Engineering

School of Information Science and Technology

University of Science and Technology of China

Hefei, Anhui 230027, China.

E-mail: Wang506@ustc.edu.cn; bqiu@ustc.edu.cn

J. Wei, Dr. G. Ji, Prof. K. Wang

Department of Neurology

the First Affiliated Hospital of Anhui Medical University

Hefei, Anhui 230022, China

Q. Wu

Division of Humanities and Social Sciences

California Institute of Technology

1200 E California Blvd Pasadena, CA 91125, USA

J. Cui

School of Optometry and Vision Science

Cardiff University

Cardiff, CF24 4HQ, United Kingdom

Prof. X. Zhang

Department of Psychology

School of Humanities & Social Science

University of Science and Technology of China

Hefei, Anhui 230026, China

Prof. X. Zhang

Academy of Psychology and Behavior

Tianjin Normal University

Tianjin, 300387, China

Keywords: temporal interference stimulation, noninvasive brain stimulation, brain oscillation,

motor function, motor cortex excitability

**Abstract**

Temporal interference (TI) stimulation is a new technique of noninvasive brain stimulation.

Envelope-modulated waveforms with two high-frequency carriers can activate neurons in target

brain regions without stimulating the overlying cortex, which has been validated in mouse

brains. However, whether TI stimulation can work on the human brain has not been elucidate.

In this study, this issue is investigated in the human primary motor cortex. Participants attended

three sessions of TI stimulation during a random reaction time task (RRTT) or a serial reaction

time task (SRTT). Motor cortex excitability was measured before and after TI stimulation. The

results indicate that TI stimulation with different envelope frequencies influenced different

motor functions. In the RRTT experiment, only 70 Hz TI stimulation had a promoting effect on

the reaction time performance and excitability of the motor cortex compared to sham

stimulation. Meanwhile, compared with the sham condition, only 20 Hz TI stimulation

significantly facilitated motor learning in the SRTT experiment, which was significantly

positively correlated with the increase in motor evoked potential. These results indicate that the

envelope-modulated waveform of TI stimulation has a significant promoting effect on human

motor functions, experimentally suggesting the effectiveness of TI stimulation in humans for

the first time.

## 1. Introduction

The human motor system can quickly react to external stimuli through delicate control of skeletal muscles by the neural activity of the motor cortex.[1] Repetitively performing the same actions in sequential manners allows humans to acquire new motor skills.[2] During these functions, high gamma and beta brain oscillations of the motor cortex play important roles. Previous studies have found that high gamma brain oscillations are transiently increased during movement and they have a promoting effect on movement initiation.[3-5] Meanwhile, beta activities in the motor cortex are considered an important component of motor learning.[6-8] Modulating these oscillations may be useful to improve motor skills and has shown promise for treating motor-related diseases, such as Parkinson's disease.

Electrical stimulation is the most direct way to regulate electrically oscillating neural activities. Two kinds of electrical stimulation techniques have been extensively used. The first is deep brain stimulation (DBS), which has been proven to be an effective treatment for treating Parkinson's disease.[9, 10] The delivery of DBS requires implants to be placed via invasive surgery, thus presenting the potential for surgical complications.[11] Another method is transcranial electrical stimulation (tES),[12, 13] which can modulate brain activities in noninvasive ways.[14-16] TES applied with alternating current, i.e., transcranial alternating current stimulation (tACS), has been used to facilitate oscillation activity within specific frequency ranges.[17, 18] Many studies have shown that tACS can modulate motor-related oscillation brain activities, which could result in changes in cortical excitability and motor function improvement.[19-22] However, currents of tES applied over the scalp were found to be significantly attenuated when traveling through the skin, subcutaneous soft tissue and skull.[23]

Participants reported side effects during tES stimulation, such as phosphene perception, skin

sensations and even skin burns under the stimulation electrode.[24-26] Furthermore, stimulating

relatively deep targets in the brain by tES might require stronger electric stimulation in the

overlying cortex,[27] which might breach the safety guidelines.[28]

To overcome the limitations of these two brain electrical stimulation techniques, temporal

interference (TI) stimulation (**Figure 1**) has been recently proposed,[29] which has caused

considerable excitement in the research community.[30-35] This new technique can be applied by

delivering two electric fields at frequencies that are too high ($\geq$1 kHz) to elicit neural firing

(Figure 1b-c). The frequency difference between these two electric fields is within the range of

brain oscillations (e.g., 20 Hz, 70 Hz, etc.), which could result in a prominent envelope

modulated electric field (Figure 1d) in a targeted brain region. TI stimulation has been proven

to be effective in driving the firing patterns of hippocampal neurons without recruiting neurons

in the overlying brain cortex and evoking different motor behaviors in mice.[29]

Based on the concept of TI stimulation, several modeling and computation studies have been

performed to explore the feasibility of TI stimulation in the human brain but have revealed

different results. Several studies reported that TI stimulation generated field strength in deep

brain regions similar to tACS with smaller electric field intensity in superficial areas.[27, 36-39] A

computational modeling study explored the impact of TI stimulation on neuron levels but found

that not all types of neurons responded to TI stimulation.[40] Another simulation study reported

that TI stimulation might have effects other than activating neurons in the target area, such as

conduction block in off-target areas.[41]

However, no data about the actual effect of TI stimulation on human brains have been

reported thus far. There are many structural differences between the brains of humans and mice,

causing the field generated by TI stimulation in the human brain to not reach the same intensity

in the mouse brain.[42] The stimulation waveform of TI stimulation is an envelope-modulated

waveform produced by the superposition of two sine waves (Figure 1d), which is much more

complex than conventional tES. Whether TI stimulation has a comparable effect with

conventional tES on the human brain is unknown. More importantly, whether TI stimulation is

safe and tolerable to people is also an urgent issue to explore.

**Figure 1.** Schematic diagram of TI stimulation. a) Stimulation of a specific brain region by TI

stimulation. b) One high-frequency (2000 Hz) electric current involved in TI stimulation. c)

Another high-frequency (2020 Hz) electric current involved in TI stimulation. d) Envelope

modulated current waveform of TI stimulation, which is generated by the superposition of the

two current waves shown in (b) and (c).

In this study, we implemented TI stimulation targeting the left primary motor cortex (M1) of

healthy participants to validate the effectiveness of TI stimulation on the human brain. M1 has

a fundamental function in controlling voluntary movements, and considerable plasticity of M1

representations is related to motor skill learning.[43] Participants attended three experimental

sessions with different stimulation conditions (**Figure 2**a). Considering the prior investigations

of oscillations related to M1, we designed two stimulation conditions with envelope frequencies

of 20 Hz (beta) and 70 Hz (high gamma). A sham condition was used as a control.

To explore the influence of TI stimulation on different levels of motor functions, two motor

tasks were employed, including a random reaction time task (RRTT) and a serial reaction time

task (SRTT) (Figure 2b-d). RRTT is a single reaction time task, and the order of the reactions

is totally randomized. Thus, the mean reaction time (RT) of all blocks was calculated as the

behavior measure. SRTT contains repeatedly recurring response sequences, which can be

learned implicitly by participants.[44] The RT reduction between S blocks and R blocks was

defined as a motor learning measure. There are two motor learning processes in the SRTT: R1

to S4 (the first implicit learning, FIL) and R2 to S6 (the second implicit learning, SIL).

Due to the distinct functions of high gamma and beta oscillations in the human motor cortex,

we hypothesized a promotion of reaction speed induced by 70 Hz TI stimulation in RRTT and

a more significant effect of 20 Hz TI stimulation on motor learning in SRTT. We also measured

motor cortex excitability before and after TI stimulation using an input-output (IO) curve or

motor evoked potential (MEP) elicited by single pulse transcranial magnetic stimulation

(TMS),[45, 46] which was hypothesized to be facilitated by TI stimulation based on previous

findings.[19, 21, 47, 48]

**Figure 2.** Experimental design and motor tasks. a) The experimental procedures were the same

for the RRTT and SRTT experiments except for the detailed motor tasks and the MEP

procedures. Participants attended three experimental sessions with different stimulation

conditions (20 Hz/70 Hz/sham). First, the individual M1 location was identified by single pulse

TMS, and baseline motor cortex excitability was measured. Formal experimental tasks (RRTT

or SRTT) started 10 minutes after the beginning of TI stimulation. After 30 minutes of

stimulation, motor cortex excitability was measured again. b) Motor task implemented in our

experiments. Participants were instructed to press one of four buttons (V, B, N, M) on the

keyboard as fast as possible, spatially corresponding to the position of the light rectangles shown on the screen. c) RRTT. The locations corresponding to the light rectangles were pseudorandomly distributed in all 8 blocks (R blocks) in RRTT. d) SRTT. In the first block and the sixth block, the locations corresponding to the light rectangles were pseudorandomly distributed (R blocks). In the remaining blocks, the locations of the light rectangles were repeated in a 12-item sequential manner ten times in each block (S blocks). There were three 12-item (bnmvnbmnvbvm, nvnmbvmnbmvb, mvbmnbnvmbvn) sequences with comparable difficulty for each experimental session in a counterbalanced way.

## 2. Results

### 2.1. TI stimulation at 70 Hz promoted the reaction time and M1 excitability

In the RRTT experiment, the stimulation condition of 70 Hz showed the lowest mean RT, which was significantly different from the sham condition ($t = -2.953$, $p_{corrected} = 0.019$, Cohen's $d = 0.716$) (**Figure 3**a). There was no significant difference in the comparison between the 20 Hz and sham groups ($t = -1.199$, $p_{corrected} = 0.498$).

For the slope of the IO curve, we found no significant results either in the main effects of condition or the main effects of testing time or the interaction of the comparison between 70 Hz or 20 Hz with sham (all ps > 0.05). Paired t-tests revealed a significant increase in the IO slope after TI stimulation at 70 Hz (70 Hz: $t = 2.395$, $p_{corrected} = 0.040$, Cohen's $d = 0.523$, one-tailed) but not at 20 Hz or in the sham condition (20 Hz: $t = -1.075$, $p_{corrected} = 0.443$, one-tailed; sham: $t = 1.597$, $p_{corrected} = 0.189$, one-tailed) (Figure 3b).

**Figure 3.** Results of the RRTT experiment. (a) The mean RT of the 70 Hz condition was significantly smaller than that of the sham condition. (b) Significant increases in the IO slope after TI stimulation were found only in the 70 Hz condition. Error bars represent SEM; *significant at $p_{corrected} < 0.05$.

**2.2. TI stimulation at 20 Hz improved implicit motor learning and MEP amplitude**

In the SRTT experiment, TI stimulation at 20 Hz showed the highest RT reduction in FIL, which was significantly different from the sham condition, while another comparison did not show significance (20 Hz vs sham: t = 2.577, $p_{corrected}$ = 0.041, Cohen's d = 0.625; 70 Hz vs sham: t = 0.197, $p_{corrected}$ = 1) (**Figure 4**a). No significant differences were found in SIL between the stimulation conditions and sham conditions (20 Hz vs sham: t = 0.5116, $p_{corrected}$ = 1; 70 Hz vs sham: t = 1.5716, $p_{corrected}$ = 0.269).

For MEP amplitude, when comparing the 20 Hz condition and sham condition, repeated measures analysis of variance (ANOVA) revealed a significant main effect of testing time (F = 4.230, p = 0.050, $\eta^2$ = 0.145), while the main effect of condition (F = 1.463, p = 0.238) and the interaction (F = 0.345, p = 0.563) were not significant. There was also a significant main effect of testing time (F = 6.523, p = 0.017, $\eta^2$ = 0.207) in the comparison between 70 Hz and sham, and no significant result was found in the main effect of condition (F = 2.028, p = 0.167) or in

the interaction (F = 0.942, p = 0.341). MEP amplitudes increased after 20 Hz TI stimulation

compared with MEP measured before stimulation at a marginally significant level (t = 2.137,

$p_{corrected}$ = 0.062, Cohen's d = 0.397, one-tailed) (Figure 4b). The increase in MEP amplitudes

in the 70 Hz and sham conditions was not significant (70 Hz: t = 1.570, $p_{corrected}$ = 0.192; sham:

t = 1.254, $p_{corrected}$ = 0.330).

After controlling for age, education level and handedness score, the significant reduction in

RT during FIL in the 20 Hz condition was positively correlated with the MEP increase (r =

0.580, $p_{corrected}$ = 0.027) (Figure 4c), while RT reductions in the other two conditions showed no

significant correlations with the MEP increase (70 Hz: r = 0.073, $p_{corrected}$ = 1; sham: r = -0.360,

$p_{corrected}$ = 0.426).

**Figure 4.** Results of the SRTT experiment. (a) Implicit motor learning during FIL in SRTT. A

significant RT reduction was only obtained in the 20 Hz condition. (b) Marginally significant

increases in MEP amplitude after TI stimulation at 20 Hz. (c) RT reduction of FIL and the MEP

increase in the 20 Hz condition was significantly positively correlated in the 20 Hz condition.

Error bars represent SEM; + marginally significant at 0.05< $p_{corrected}$ <0.1, *significant at

$p_{corrected}$ <0.05.

**2.3 TI stimulation caused minor side effects on participants**

Side effects occurring during TI stimulation were minor and tolerable according to the participants' descriptions and our observations. In RRTT, 60 out of 63 sessions were considered sessions with no discomforts (95.238%). Among the 3 sessions with discomforts, there were 2 sham sessions and 1 stimulation session. In SRTT, 84 out of 87 sessions were considered sessions with no discomforts (96.552%). All sessions with discomforts were stimulation sessions. More details are given in Table 1 and Table 2. Notably, all discomforts during the sham sessions occurred in the middle of the session or at the end of the session, which could imply that sham stimulation did not directly cause the sensations. Our subsequent investigations of the participants also reported no other side effects after the experiments.

**Table1.** Discomforts in RRTT

| Sensations | Stimulation sessions | Sham sessions | Percentage |
|---|---|---|---|
| None | 41 | 19 | 95.238% |
| Fatigue | 0 | 1 | 1.587% |
| Vertigo | 0 | 1 | 1.587% |
| Headache | 1 | 0 | 1.587% |

**Table2.** Discomforts in SRTT

| Sensations | Stimulation sessions | Sham sessions | Percentage |
|---|---|---|---|
| None | 55 | 29 | 96.552% |
| Fatigue | 3 | 0 | 3.448% |

## 3. Discussion

In this study, we applied TI stimulation to healthy human participants to explore the modulatory effects of TI stimulation on motor functions. We investigated the changes in motor performance resulting from TI stimulation applied over M1 in two experiments involving different motor tasks. TI stimulation with an envelope frequency of 70 Hz promoted the reaction time performance of the motor task compared with the sham condition in the RRTT experiment. TI stimulation with an envelope frequency of 20 Hz applied over M1 enhanced the first implicit motor learning performance compared with sham stimulation, and the performance was positively correlated with the MEP increase in the SRTT experiment.

### 3.1 TI stimulation is effective in the human motor cortex

Our study, for the first time, suggests that the idea of TI stimulation is plausible, not only in computational models and experiments on mice[27, 29, 36-39] but also in actual experiments performed on healthy human participants. Since the idea of TI stimulation has been raised, the only in vivo investigations have been performed on mouse brains.[29] The human brain is much larger, and the layers around the brain in humans are thicker, which causes up to 100 times weaker electric fields in the human brain than in the mouse brain at the same stimulation intensity.[42] The stimulation waveform of TI stimulation is an envelope-modulated waveform produced by the superposition of two sine waves (Figure 1d), which has not been previously tested in humans. Conventional tACS mostly use normal sine waves to modulate the brain oscillations.[12, 13, 17, 18] Envelope-tACS using only envelope waveforms of speech without carrier waves have been used to improve speech perception and processing,[49-51] but the effects

are still controversial.[52-54] Amplitude-modulated tACS (AM-tACS) was proposed as a

promising way to allow for effective magnetoencephalography (MEG) or

electroencephalography (EEG) signal reconstruction during electrical stimulation. However,

similar to TI stimulation, studies on AM-tACS have also mostly focused on simulations, and

no systematic experimental test to validate the effectiveness of AM-tACS in humans has been

performed. Whether envelope modulated waveforms have comparable effects to conventional

tACS is unknown.

To solve these problems, we applied TI stimulation to the human M1 area and found some

significant effects. In the RRTT experiment, 70 Hz TI stimulation promoted RT performance

and motor cortical excitability. An increase in motor cortex excitability is often accompanied

by better performance in motor functions;[55-58] for example, an increase in motor acceleration

may be associated with an increase in MEP amplitude.[59] In the SRTT experiment, 20 Hz TI

stimulation increased the first implicit motor learning and MEP amplitudes. We found

significant main effects of testing time in the analysis of MEP amplitude, which might indicate

a training effect of the motor learning task.[60, 61] But only the correlation between FIL and MEP

increase in 20 Hz condition was significant, indicating the increase of motor cortex excitability

related with TI stimulation only occurred in 20 Hz TI stimulation. This can be supported by a

meta-analysis, which shows that exogenously applied electric fields in beta frequency range

can increase motor cortex excitability.[62] Based on the significant correlations we observed in

the SRTT experiment, we speculated that TI stimulation promoted implicit motor learning by

influencing the excitability of M1.

Anyway, to be honest, the effect of TI stimulation shown in this study is not as phenomenal

as that in the mice study. Future studies could explore the mechanisms of TI stimulation at the

level of brain regions and networks by corresponding neuroimaging techniques, e.g., functional

magnetic resonance imaging (fMRI), and try to build a more effective TI stimulation system

for human with better understanding to it.

**3.2 High gamma and beta oscillations may represent different motor functions in M1**

Distinct effects of 20 Hz TI stimulation and 70 Hz TI stimulation may indicate different

functions of these two motor cortical brain oscillations. High gamma and beta are considered

vital neural rhythms corresponding to the activation of M1. High gamma oscillations transiently

increase during movement and have a promoting effect on movement initiation.[3-5] TACS (70

Hz) has been reported to increase motor velocity and motor acceleration during stimulation in

visually guided motor tasks.[20, 63] Meanwhile, beta oscillations in the motor cortex, specifically

post-movement beta rebound, are considered an important component in motor learning.[6-8]

TACS at 20 Hz has been reported to improve the performance of implicit learning of SRTT in

previous studies.[22, 64] These findings could imply that beta and high-gamma neural rhythms

predominate in different motor functions in M1. Our results duplicate the functional separation

between brain oscillations at 20 Hz and 70 Hz. In the RRTT experiment, the reaction time

performance was only promoted by 70 Hz TI stimulation. In the SRTT experiment, the motor

learning process was only significantly promoted by 20 Hz TI stimulation. The functional

separation between 70 Hz (2000 Hz & 2070 Hz) and 20 Hz (2000 Hz & 2020 Hz) TI stimulation

also supports the hypothesis that electric fields of high-frequency carriers (2000 Hz) may have

little contribution to the results because of the intrinsic feature of the neural membrane that

filters electrical signals in a low-pass manner.[65, 66] Additional studies could explore the effects

of carrier frequency and envelope frequency more deeply.

**3.3 The application potential of TI stimulation as a noninvasive brain stimulation**

**technique**

We assessed the side effects of TI stimulation by subjective reporting of the participants. In

most sessions (>95%), participants reported no side effects. No sensations related to the skin,

such as tingling, itching and burning, were reported, and no burns or reddening of the skin were

observed by the experimenters. Only fatigue, vertigo and headache were reported in several

sessions, including two sham sessions. The side effects reported by participants in this study

were far less than those reported for conventional tES,[24-26] which indicates that TI stimulation

may have advantages over conventional tES in safety, user-friendliness and blinding

performance.

Our study indicates that TI stimulation can be used as a new technique to modulate neural

activities in a noninvasive way. We speculate that TI stimulation could be a feasible tool for

exploring distinct roles of different brain oscillations in various cognition tasks, especially those

neural activities originating from deep brain regions. We preliminarily explored the effects of

TI stimulation on human M1. Stimulation effects on other deeper brain regions with more

sophisticated functions rely on a better understanding of the working mechanisms and prospects

of TI stimulation in humans, which needs to be explored in additional research utilizing

combinations of neuron models, finite element modeling simulations and experiments.[36] It has

been speculated that regions that are deep but not too small when considered as a fraction of

the total tissue volume (e.g., those in stroke, obsessive-compulsive disorder, epilepsy, depression, and spinal cord injury) may be attractive initial indications.[35] Future studies could explore the effect of TI stimulation in deep brain regions and promote the applications of TI stimulation in clinical practice.

**3.4 Limitations and Future Directions**

This study has two limitations. First, we chose M1 as our target brain region rather than brain regions in deeper areas. TI stimulation is expected to be an alternative method of DBS applied in a noninvasive manner. However, previous studies have suggested that in order to achieve an effect comparable to DBS, the stimulation protocol might need to be modified to improve the stimulation strength and focality.[36] A possible method could be multielectrode stimulation combined with optimization of the electrode montage, which has been explored recently.[39] Second, our study only employed TMS and motor tasks. We speculate that other experimental techniques, i.e., EEG or fMRI, could be involved to capture more details about the mechanism of TI stimulation.

**4. Conclusion**

Our study reveals the promoting effect of TI stimulation on human motor functions and motor cortex excitability. In the SRTT experiment, the promoting effect of motor learning performance had a significant correlation with an increase in MEP, indicating that the possible mechanism of TI stimulation in our experiments might be increasing the excitability of M1. TI stimulation with different envelope frequencies showed separate promoting effects on different

motor tasks, which implied that TI stimulation may work through a low-frequency envelope. Future investigations of TI stimulation in humans could explore TI stimulation in deeper brain regions under the guidance of modeling works. In summary, TI stimulation could be a promising new technique for noninvasive brain stimulation in humans with clinical application potential.

**5. Experimental Section**

*Participants*: We recruited 27 right-handed healthy adult volunteers in the first RRTT experiment, and 6 participants were removed from the analysis because of technical issues (a decrease in current due to poor contact and current crosstalk due to the flow of conductive paste) during experiments. Data from the remaining 21 participants were included in the analysis (6 females, mean age ± SD: 22.429 ± 2.249 years, mean education level ± SD: 15.762 ± 2.166 years, mean handedness score ± SD: 86.667 ± 17.127). Another 33 right-handed healthy adults volunteered to participate in the SRTT experiment, but 1 participant was removed due to the sliding of electrodes, 1 participant was rejected because he switched his performing hand during SRTT, and 2 participants' data were removed because of technical issues (current crosstalk due to the flow of conductive paste). Therefore, 29 participants remained to be analyzed in the SRTT experiment (15 females, mean age ± SD: 22.103 ± 2.024 years, mean education level ± SD: 15.966 ± 1.991 years, mean handedness score ± SD: 77.672 ± 23.792).

All participants reported no history of craniotomy or injury to the head, no personal or family history of neurological or psychiatric disease, no metal implants or implanted electronic devices, no skin sensitivity and no use of medicine during the experiment. For safety reasons, any

participant who was pregnant or could be pregnant was rejected. All participants were right-handed as assessed using the Edinburgh handedness inventory[67] and had normal or corrected-to-normal vision. Informed consent was obtained prior to any involvement in the study. This study was approved by the Human Ethics Committee of the University of Science and Technology of China (IRB Number: 2020KY161).

*Experimental design*: The experimental procedures were the same for both the RRTT and SRTT experiments except for the detailed motor tasks and MEP procedures. At the beginning of the procedures, individual M1 location was identified by single pulse TMS, and baseline motor cortex excitability was measured. Stimulating electrodes were placed on the participant's head around the location of M1. Before stimulation, the participants were asked to perform a practice task with 24 random button presses. Formal experimental tasks (RRTT or SRTT) started 10 minutes after the beginning of TI stimulation. After the 30-minute stimulation, motor cortex excitability was measured again to detect the change in excitability of M1 so that the possible involvement of M1 during the motor task could be assessed (Figure 2a). Participants visited the laboratory three times, at least three days apart, to avoid any influence of the carry-over effects of stimulation.

*TMS and MEP*: Single pulse TMS was delivered manually using a 70-mm air-cooled figure-eight coil and a Magstim Rapid$^2$ stimulator (The Magstim Company Ltd., Whitland, United Kingdom) with the navigation system of Brainsight (Brainsight, Quebec, Canada). The coil was placed tangentially over the scalp of the left hemisphere to search for the hotspot of the right first dorsal interosseus (FDI), which was marked on a medical elastic bandage on the participants' heads after the search process. The handle of the coil was pointing posterolaterally

45° from the midline.[68, 69] An electromyogram (EMG) of the right FDI was recorded by a pair of Ag-AgCl electrodes in a belly tendon montage using the EMG module of the navigation system. The resting motor threshold (RMT) was defined as the lowest stimulus intensity that could elicit an MEP in the resting muscle with an amplitude of 50 µV (peak-to-peak) or greater in at least 5 out of 10 recordings.[46, 70]

In the RRTT experiment, we applied 15 pulses over the FDI hotspot with an interval of 7 seconds at stimulation intensities of 120%, 100%, 130%, 110% and 140% of RMT to obtain details of the variability of the MEPs before and after TI stimulation.[71] We measured 30 MEPs at a stimulation intensity of 120% of the RMT in the SRTT experiment. Only 120% RMT was used because this intensity corresponds to the linear increase range of the input-output curve and is sensitive to the change in M1 excitability. [46]

*Motor tasks*: The motor tasks were both modified from an SRTT task, which was previously involved in tACS experiments.[22, 64] Participants were instructed to press one of four buttons (V, B, N, M) on the keyboard as fast as possible, according to the position of the light rectangles shown on the screen (Figure 2b). The stimulus remained on the screen until the correct response was made. After 500 ms, a new stimulus was displayed on the screen. Eight blocks were included, with 120 trials in each block. The locations corresponding to the light rectangles were pseudorandomly distributed in all 8 blocks (R) in RRTT (Figure 2c). The only difference between SRTT and RRTT was that the locations corresponding to the light rectangles were not randomized in some SRTT blocks (Figure 2d). In the first block and the sixth block, the locations corresponding to the light rectangles were pseudorandomly distributed (R). In the remaining blocks, the locations of the light rectangles were repeated in a 12-item sequential

manner ten times in each block (S). There were three 12-item (bnmvnbmnvbvm,

nvnmbvmnbmvb, mvbmnbnvmbvn) sequences with comparable difficulty for each

experimental session in a counterbalanced way. All of the information about the order of the

locations was unknown to the participant, which allowed them to acquire the sequence in an

implicit manner. The task presentation and the recording of the reaction times were conducted

using E-Prime 2.0 (Psychology Software Tools, Sharpsburg, USA).

*Temporal interference stimulation*: We used five Ag-AgCl electrodes with a radius of 1 cm

(Pistim electrode, Neuroelectrics, Barcelona, Spain), four of which were stimulating electrodes

and one was the ground electrode located on the mastoid behind the participant's left ear to

avoid current accumulations due to safety considerations. The delivery of the current was

provided by a customized battery-driven stimulator with strict safety standards (Figure S1,

Supporting Information 2.1). The current output of each pair of electrodes (a single channel)

had the same frequency with a phase difference of 180° to avoid current crosstalk between

channels.[29] The stimulation intensity was peak-to-peak 2 mA in a single channel. The

stimulation electrodes were located 30 mm away from the FDI hot-spot along the axis of the

Fpz-Oz and T3-T4 in the EEG 10-20 system (Figure S2, Supporting Information 2.2). The

frequency difference between the two channels was the envelope frequency. Three experimental

conditions were included in our design: 20 Hz (2000 Hz & 2020 Hz), 70 Hz (2000 Hz & 2070

Hz) and sham. For the sham condition, electric stimulation (20 Hz or 70 Hz TI stimulation)

only lasted for approximately 60 seconds (30 s ramp up and 30 s ramp down) at the beginning

of this procedure. Stimulation started 10 minutes before the motor task and lasted for 30 minutes.

*Safety aspects*: After the TI stimulation, we asked the participants to complete a subjective

questionnaire,[24, 25] which asked the participants to rate their sensations, including itching,

headache, burning, warmth/heat, tingling, metallic/iron taste, fatigue, vertigo, nausea and

phosphene, during the stimulation and to what extent they thought these sensations were

relevant to the stimulation so that the potential side effects could be measured. Participants were

asked to rate the extent of these sensations that they experienced, from 0 to 4, representing none,

mild, moderate, considerable and strong, respectively. Similarly, the relevance to the

stimulation was also rated from 0 to 4, representing none, remote, possible, probable and

definite, respectively. The influence of those sensations, if any, on the motor task was also

evaluated by the participants. The following criteria were used: 1) Only sensations with a score

larger than 1 were taken into consideration, and 2) Any sensations that had no relevance at all

to the stimulation based on the participants' descriptions were not considered discomforts

elicited by the stimulations.

*Data analysis*: All analyses were performed on MATLAB 2018a (MathWorks, Natick, USA).

The mean RT of the correct trials of each block in RRTT or SRTT was calculated. Accuracy

was not considered a primary measure when assessing the effect of TI stimulation on motor

tasks because of the ceiling effect (Figure S3). Because the calculation of behavior measures

needed to integrate the RT of different blocks, any session containing RT of any block beyond

2SD from all participants' mean RT of that block in the same stimulation condition was removed.

The mean RT of all blocks was considered the behavior measure in the RRTT experiment.

Motor learning performance was measured as the RT reduction between S blocks and R blocks

in the SRTT experiment. There are two implicit motor learning processes in SRTT. The first is

the reduction of reaction time from the first R block to the first four S blocks (Equation (1)):

$$451 \qquad \text{FIL} = RT_{R1} - \frac{RT_{S1} + RT_{S2} + RT_{S3} + RT_{S4}}{4} \qquad (1)$$

452 The second is the reduction of reaction time from the second R block to the sixth S block

453 (Equation (2)):

$$454 \qquad \text{SIL} = RT_{R2} - \frac{RT_{S5} + RT_{S6}}{2} \qquad (2)$$

IO curves linearly fitted using the amplitude of MEPs elicited by 100%, 110%, 120%, 130%

and 140% RMT were involved in each stimulation condition in the RRTT experiment, and the

slope of the IO curve was extracted. Mean MEP amplitudes before and after TI stimulation in

each condition were calculated in the SRTT experiment. The difference between the slope of

the IO curve in the RRTT experiment and the amplitude of the MEP in the SRTT experiment

before and after TI stimulation were calculated, which reflected the variability of the excitability

of M1.

Differences in the behavior measures between the stimulation conditions and the control

conditions (20 Hz vs sham, 70 Hz vs sham) were assessed by two-tailed paired t-tests. Two 2

(condition: 20 Hz vs sham/70 Hz vs sham) × 2 (testing time: before TI stimulation vs after TI

stimulation) repeated measures ANOVA was performed on the slopes of the IO curve and MEP

amplitudes. We set age, education level and handedness score as covariables to control their

potential influence to the motor cortex excitability.[72-74] Since there were significant promoting

effects found in the behavior measures, slopes of the IO curve and MEP amplitude before and

after TI stimulation were compared by one-tailed paired t-tests with the hypothesis that MEPs

would also be facilitated by TI stimulation. Correlations between the behavior measures and

increases in the IO slopes or MEP amplitudes in each condition were tested by two-tailed partial

correlations, with age, education level and handedness score controlled as covariables.

Bonferroni correction was used to correct for multiple comparisons.

**Acknowledgements**

R.M. and X.X. contributed equally to this work. We would like to thank Prof. Bettina Pollok

for her kindly help with programming of the experimental tasks. And we thank Dr. Wei Lu

for his help with circuit testing. This work was supported by grants from The National Key

Basic Research Program (2018YFC0831101), The National Natural Science Foundation of

China (31771221, 71942003, 61773360, 31800927, 31900766 and 71874170), Major Project

of Philosophy and Social Science Research, Ministry of Education of China (19JZD010),

CAS-VPST Silk Road Science Fund 2021 (GLHZ202128), Collaborative Innovation Program

of Hefei Science Center, CAS (2020HSC-CIP001). A portion of the numerical calculations in

this study were performed with the supercomputing system at the Supercomputing Centre of

USTC.

**Conflict of Interest**

The authors declare no conflict of interest.

References

[1]  G. Rizzolatti, G. Luppino, *Neuron* **2001**, *31*, 889.

[2]  D. M. Wolpert, J. Diedrichsen, J. R. Flanagan, *Nature Reviews Neuroscience* **2011**, *12*, 739.

[3]  S. D. Muthukumaraswamy, *Journal of Neurophysiology* **2010**, *104*, 2873.

[4]  W. Gaetz, C. Liu, H. Zhu, L. Bloy, T. P. L. Roberts, *NeuroImage* **2013**, *74*, 245.

[5]  D. Cheyne, S. Bells, P. Ferrari, W. Gaetz, A. C. Bostan, *NeuroImage* **2008**, *42*, 332.

[6]  S. Espenhahn, B. C. M. van Wijk, H. E. Rossiter, A. O. de Berker, N. D. Redman, J.

Rondina, J. Diedrichsen, N. S. Ward, *NeuroImage* **2019**, *195*, 340.

[7]  S. Espenhahn, H. E. Rossiter, B. C. M. van Wijk, N. Redman, J. M. Rondina, J. Diedrichsen,

N. S. Ward, *Brain communications* **2020**, *2*, fcaa161.

[8]  S. Haar, A. A. Faisal, *Frontiers in Human Neuroscience* **2020**, *14*.

[9]  A. L. Benabid, *Current Opinion in Neurobiology* **2003**, *13*, 696.

[10] G. Tinkhauser, A. Pogosyan, S. Little, M. Beudel, D. M. Herz, H. Tan, P. Brown, *Brain*

**2017**, *140*, 1053.

[11] A. Ramirez-Zamora, J. J. Giordano, A. Gunduz, P. Brown, J. C. Sanchez, K. D. Foote, L.

Almeida, P. A. Starr, H. M. Bronte-Stewart, W. Hu, C. McIntyre, W. Goodman, D. Kumsa, W.

M. Grill, H. C. Walker, M. D. Johnson, J. L. Vitek, D. Greene, D. S. Rizzuto, D. Song, T. W.

Berger, R. E. Hampson, S. A. Deadwyler, L. R. Hochberg, N. D. Schiff, P. Stypulkowski, G.

Worrell, V. Tiruvadi, H. S. Mayberg, J. Jimenez-Shahed, P. Nanda, S. A. Sheth, R. E. Gross, S.

F. Lempka, L. Li, W. Deeb, M. S. Okun, *Frontiers in Neuroscience* **2018**, *11*, 734.

[12] W. Paulus, *Neuropsychological Rehabilitation* **2011**, *21*, 602.

[13] S. Bestmann, V. Walsh, *Current biology : CB* **2017**, *27*, R1258.

[14] L. Z. Yang, W. Zhang, B. Shi, Z. Y. Yang, Z. D. Wei, F. Gu, J. Zhang, G. B. Cui, Y. Liu, Y.

F. Zhou, X. C. Zhang, H. Y. Rao, *Plos One* **2014**, *9*.

[15] L. Z. Yang, B. Shi, H. Li, W. Zhang, Y. Liu, H. Z. Wang, Y. F. Zhou, Y. Wang, W. W. Lv, X.

B. Ji, J. Hudak, Y. F. Zhou, A. J. Fallgatter, X. C. Zhang, *Social Cognitive and Affective*

*Neuroscience* **2017**, *12*, 1296.

[16] L. Z. Yang, W. Zhang, W. J. Wang, Z. Y. Yang, H. Z. Wang, Z. D. Deng, C. F. Li, B. S. Qiu, D. R. Zhang, R. C. Kadosh, H. Li, X. C. Zhang, *Advanced Science* **2020**, *7*.

[17] C. S. Herrmann, S. Rach, T. Neuling, D. Strueber, *Frontiers in Human Neuroscience* **2013**, *7*.

[18] A. Vossen, J. Gross, G. Thut, *Brain Stimulation* **2015**, *8*, 499.

[19] M. Feurra, G. Bianco, E. Santarnecchi, M. Del Testa, A. Rossi, S. Rossi, *Journal of Neuroscience* **2011**, *31*, 12165.

[20] R. A. Joundi, N. Jenkinson, J. S. Brittain, T. Z. Aziz, P. Brown, *Current biology : CB* **2012**, *22*, 403.

[21] M. Feurra, P. Pasqualetti, G. Bianco, E. Santarnecchi, A. Rossi, S. Rossi, *Journal of Neuroscience* **2013**, *33*, 17483.

[22] B. Pollok, A.-C. Boysen, V. Krause, *Behavioural Brain Research* **2015**, *293*, 234.

[23] M. Vöröslakos, Y. Takeuchi, K. Brinyiczki, T. Zombori, A. Oliva, A. Fernández-Ruiz, G. Kozák, Z. T. Kincses, B. Iványi, G. Buzsáki, A. Berényi, *Nature Communications* **2018**, *9*, 483.

[24] A. R. Brunoni, J. Amadera, B. Berbel, M. S. Volz, B. G. Rizzerio, F. Fregni, *The international journal of neuropsychopharmacology* **2011**, *14*, 1133.

[25] A. Fertonani, C. Ferrari, C. Miniussi, *Clinical neurophysiology : official journal of the International Federation of Clinical Neurophysiology* **2015**, *126*, 2181.

[26] H. Matsumoto, Y. Ugawa, *Clinical Neurophysiology Practice* **2017**, *2*, 19.

[27] S. Lee, C. Lee, J. Park, C.-H. Im, *Scientific Reports* **2020**, *10*, 11730.

[28] A. Antal, I. Alekseichuk, M. Bikson, J. Brockmöller, A. R. Brunoni, R. Chen, L. G. Cohen, G. Dowthwaite, J. Ellrich, A. Flöel, F. Fregni, M. S. George, R. Hamilton, J. Haueisen, C. S.

Herrmann, F. C. Hummel, J. P. Lefaucheur, D. Liebetanz, C. K. Loo, C. D. McCaig, C. Miniussi, P. C. Miranda, V. Moliadze, M. A. Nitsche, R. Nowak, F. Padberg, A. Pascual-Leone, W. Poppendieck, A. Priori, S. Rossi, P. M. Rossini, J. Rothwell, M. A. Rueger, G. Ruffini, K. Schellhorn, H. R. Siebner, Y. Ugawa, A. Wexler, U. Ziemann, M. Hallett, W. Paulus, *Clinical Neurophysiology* **2017**, *128*, 1774.

[29] N. Grossman, D. Bono, N. Dedic, S. B. Kodandaramaiah, A. Rudenko, H.-J. Suk, A. M. Cassara, E. Neufeld, N. Kuster, L.-H. Tsai, A. Pascual-Leone, E. S. Boyden, *Cell* **2017**, *169*, 1029.

[30] J. Dmochowski, M. Bikson, *Cell* **2017**, *169*, 977.

[31] A. Opitz, W. J. Tyler, *Nature Biomedical Engineering* **2017**, *1*, 632.

[32] A. M. Lozano, *New England Journal of Medicine* **2017**, *377*, 1096.

[33] C. H. Halpern, K. J. Miller, H. Wu, P. A. Tass, *Neurosurgery* **2018**, *82*, E19.

[34] N. Grossman, *Science* **2018**, *361*, 461.

[35] N. Grossman, M. S. Okun, E. S. Boyden, *JAMA Neurology* **2018**, *75*, 1307.

[36] S. Rampersad, B. Roig-Solvas, M. Yarossi, P. P. Kulkarni, E. Santarnecchi, A. D. Dorval, D. H. Brooks, *NeuroImage* **2019**, *202*, 116124.

[37] K. Fariba, A. Ahmadreza, A. Rassoul, N. Abolghasem Zeidaabadi, *Physics in Medicine & Biology* **2019**.

[38] Y. Huang, L. C. Parra, *Brain Stimulation: Basic, Translational, and Clinical Research in Neuromodulation*.

[39] Y. Huang, A. Datta, L. C. Parra, *Journal of Neural Engineering* **2020**, *17*, 036023.

[40] J. Cao, P. Grover, *IEEE Transactions on Biomedical Engineering* **2020**, *67*, 726.

[41] E. Mirzakhalili, B. Barra, M. Capogrosso, S. F. Lempka, *Cell Systems*.

[42] I. Alekseichuk, K. Mantell, S. Shirinpour, A. Opitz, *Neuroimage* **2019**, *194*, 136.

[43] J. N. Sanes, J. P. Donoghue, *Annual Review of Neuroscience* **2000**, *23*, 393.

[44] E. M. Robertson, *The Journal of Neuroscience* **2007**, *27*, 10073.

[45] R. Chen, *Muscle & Nerve* **2000**, *23*, S26.

[46] P. M. Rossini, D. Burke, R. Chen, L. G. Cohen, Z. Daskalakis, R. Di Iorio, V. Di Lazzaro, F. Ferreri, P. B. Fitzgerald, M. S. George, M. Hallett, J. P. Lefaucheur, B. Langguth, H. Matsumoto, C. Miniussi, M. A. Nitsche, A. Pascual-Leone, W. Paulus, S. Rossi, J. C. Rothwell, H. R. Siebner, Y. Ugawa, V. Walsh, U. Ziemann, *Clinical Neurophysiology* **2015**, *126*, 1071.

[47] V. Moliadze, A. Antal, W. Paulus, *Journal of Physiology-London* **2010**, *588*, 4891.

[48] A. Guerra, F. Asci, V. D'Onofrio, V. Sveva, M. Bologna, G. Fabbrini, A. Berardelli, A. Suppa, *Journal of Neuroscience* **2020**, *40*, 4788.

[49] L. Riecke, E. Formisano, B. Sorger, D. Başkent, E. Gaudrain, *Current Biology* **2018**, *28*, 161.

[50] A. Wilsch, T. Neuling, J. Obleser, C. S. Herrmann, *NeuroImage* **2018**, *172*, 766.

[51] S. Kadir, C. Kaza, H. Weissbart, T. Reichenbach, *IEEE transactions on neural systems and rehabilitation engineering : a publication of the IEEE Engineering in Medicine and Biology Society* **2020**, *28*, 23.

[52] B. Asamoah, A. Khatoun, M. Mc Laughlin, *Brain Stimulation* **2019**, *12*, 1001.

[53] J. Erkens, M. Schulte, M. Vormann, C. S. Herrmann, *Neuroscience Insights* **2020**, *15*, 2633105520936623.

[54] A. Kosem, H. R. Bosker, O. Jensen, P. Hagoort, L. Riecke, *Journal of Cognitive*

*Neuroscience* **2020**, *32*, 1428.

[55] M. K. Sohn, B. O. Kim, H. T. Song, *Annals of rehabilitation medicine* **2012**, *36*, 1.

[56] M. Christova, D. Rafolt, E. Gallasch, *Behavioural Brain Research* **2015**, *287*, 27.

[57] V. Lopez-Alonso, B. Cheeran, M. Fernandez-del-Olmo, *Brain Stimulation* **2015**, *8*, 1209.

[58] T. Abe, S. Miyaguchi, N. Otsuru, H. Onishi, *Neuroscience Letters* **2019**, *705*, 138.

[59] W. Muellbacher, U. Ziemann, B. Boroojerdi, L. Cohen, M. Hallett, *Experimental brain research* **2001**, *136*, 431.

[60] J. Classen, J. Liepert, S. P. Wise, M. Hallett, L. G. Cohen, *J Neurophysiol* **1998**, *79*, 1117.

[61] S. Fattinger, T. T. de Beukelaar, K. L. Ruddy, C. Volk, N. C. Heyse, J. A. Herbst, R. H. R. Hahnloser, N. Wenderoth, R. Huber, *Nature Communications* **2017**, *8*, 15405.

[62] M. Wischnewski, D. J. L. G. Schutter, M. A. Nitsche, *Brain Stimulation* **2019**, *12*, 1381.

[63] M. Moisa, R. Polania, M. Grueschow, C. C. Ruff, *The Journal of Neuroscience* **2016**, *36*, 12053.

[64] V. Krause, A. Meier, L. Dinkelbach, B. Pollok, *Front Behav Neurosci* **2016**, *10*, 4.

[65] B. Hutcheon, Y. Yarom, *Trends in Neurosciences* **2000**, *23*, 216.

[66] Z. Esmaeilpour, M. Jackson, G. Kronberg, T. Zhang, R. Esteller, B. Hershey, M. Bikson, *eNeuro* **2020**, *7*, ENEURO.0368.

[67] R. C. Oldfield, *Neuropsychologia* **1971**, *9*, 97.

[68] K. Sakai, Y. Ugawa, Y. Terao, R. Hanajima, T. Furubayashi, I. Kanazawa, *Experimental Brain Research* **1997**, *113*, 24.

[69] A. Opitz, W. Legon, A. Rowlands, W. K. Bickel, W. Paulus, W. J. Tyler, *NeuroImage* **2013**, *81*, 253.

[70] B. Boroojerdi, F. Battaglia, W. Muellbacher, L. G. Cohen, *Clinical Neurophysiology* **2001**,

*112*, 931.

[71] J. A. Kleim, E. D. Kleim, S. C. Cramer, *Nature protocols* **2007**, *2*, 1675.

[72] A. Eisen, S. Siejka, M. Schulzer, D. Calne, *Electroencephalography and Clinical*

*Neurophysiology/Evoked Potentials Section* **1991**, *81*, 209.

[73] L. De Gennaro, R. Cristiani, M. Bertini, G. Curcio, M. Ferrara, F. Fratello, V. Romei, P. M.

Rossini, *Clinical Neurophysiology* **2004**, *115*, 1305.

[74] M. S. Thomas, in *BJEP Monograph Series II, Number 8-Educational Neuroscience*, Vol.

143, British Psychological Society, 2012, 143.

614