An Artificial Neural-Network Approach for Motor Hotspot Identification Based on Electroencephalography: A Proof-of-Concept Study

Ga-Young Choi[1], Chang-Hee Han[2], Hyung-Tak Lee[1], Nam-Jong Paik[3], Won-Seok Kim[3,*],

Han-Jeong Hwang[1,*]

E-mail: cgy326@naver.com, zeros8706@naver.com, falling_slow@naver.com, njpaik@snu.ac.kr, wondol77@gmail.com, hwanghj@korea.ac.kr

[1]Department of Electronics and Information Engineering, Korea University, Sejong 30019, Republic of Korea

[2]Machine Learning Group, Berlin Institute of Technology (TU Berlin), Berlin 10623, Germany

[3]Department of Rehabilitation Medicine, Seoul National University College of Medicine, Seoul National University Bundang Hospital, Seongnam-si 13620, Republic of Korea

*Corresponding Authors

Corresponding Author Information:

Name: Won-Seok Kim

Address: Department of Rehabilitation Medicine, Seoul National University College of Medicine, Seoul National University Bundang Hospital, Seongnam-si 13620, Republic of Korea

E-mail: wondol77@gmail.com

Name: Han-Jeong Hwang

Address: Department of Electronics and Information Engineering, Korea University, Sejong 30019, Republic of Korea

E-mail: hwanghj@korea.ac.kr

Abstract

Background: To apply transcranial electrical stimulation (tES) to the motor cortex, motor hotspots are generally identified using motor evoked potentials by transcranial magnetic stimulation (TMS). The objective of this study is to validate the feasibility of a novel electroencephalography (EEG)-based motor-hotspot-identification approach using a machine learning technique as a potential alternative to TMS.

Methods: EEG data were measured using 63 channels from thirty subjects as they performed a simple finger tapping task. Power spectral densities of the EEG data were extracted from six frequency bands (delta, theta, alpha, beta, gamma, and full) and were independently used to train and test an artificial neural network for motor hotspot identification. The 3D coordinate information of individual motor hotspots identified by TMS were quantitatively compared with those estimated by our EEG-based motor-hotspot-identification approach to assess its feasibility.

Results: The minimum mean error distance between the motor hotspot locations identified by TMS and our proposed motor-hotspot-identification approach was 0.22 ± 0.03 cm, demonstrating the proof-of-concept of our proposed EEG-based approach. A mean error distance of 1.32 ± 0.15 cm was measured when using only nine channels attached to the middle of the motor cortex, showing the possibility of practically using the proposed motor-hotspot-identification approach based on a relatively small number of EEG channels.

Conclusion: We demonstrated the feasibility of our novel EEG-based motor-hotspot-identification method. It is expected that our approach can be used as an alternative to TMS for motor hotspot identification. In particular, its usability would significantly increase when using a recently developed portable tES device integrated with an EEG device.

Keywords –Motor hotspot; Electroencephalography; Transcranial electrical stimulation; Machine learning; Artificial neural-network

Introduction

Motor impairment is a frequent symptom occurring after neurological disorders, such as stroke and Parkinson's disease [1–3]. Although motor functions are not completely restored after neurological disorders, continuous rehabilitation is necessary to prevent muscle loss and the retrogression of intact motor functions. Conventional motor rehabilitation interventions involve specific movements related to the affected limbs, which are enforced by a therapist or an assistive rehabilitation device.

Recently, transcranial electrical stimulation (tES) capable of modulating cortical excitability using a weak electrical current has been introduced for motor rehabilitation, and its positive effects have been proven in many interventional studies even though the mechanisms have not yet been fully understood [4–10]. For example, one study showed that anodal transcranial direct current stimulation (tDCS) on the ipsilesional primary motor cortex could improve overall motor functions of the upper limbs in stroke patients, and its effects persisted for at least 3 months post-intervention [11]. Another study also demonstrated the positive effects of transcranial alternating current stimulation (tACS) on motor performance improvements in Parkinson's disease [12].

To maximize the positive effects of tES on motor rehabilitation, it is important to find an optimal tES target location. The lesional primary motor area (M1) and its contralateral motor area have been traditionally used as stimulation target locations, to which anodal and cathodal tES electrodes are applied, respectively [13–15]. A cathodal tES electrode is sometimes applied to the contralateral supraorbital area of the lesional hemisphere instead of the contralateral motor area, called the M1-supraorbital prominence [15–17]. Identifying M1 can be performed by either the international 10-20 coordinate system used for electroencephalography (EEG) measurements or motor evoked potentials (MEPs), induced by transcranial magnetic

stimulation (TMS). The former method uses C3 on the left hemisphere or C4 on the right hemisphere as the M1 location, depending on the lesional side [13, 18]. However, as the international 10-20 coordinate system does not consider inter-subject variability in cortical anatomy, the M1 location (C3 or C4) found by the international 10-20 coordinate system might not be an appropriate tES target location for motor rehabilitation. In particular, patients with neurological disorders, who are the main recipients of tES treatment, have shown significant changes in cortical morphologies owing to neural reorganization and plasticity after the occurrence of neurological diseases, such as stroke and cerebral palsy [19], requiring better methods to find individualized tES target locations more precisely.

As an alternative to the international 10-20 coordinate system, many studies have used TMS-induced MEP to find an individualized tES target location, called motor hotspot, for motor rehabilitation. This has been done because applying tES to the motor hotspot found by TMS-induced MEP could provide focal and accurate neuromodulatory effects on the motor network [20–22]. TMS is a useful tool to find motor hotspots, but requires a relatively bulky device and a somewhat cumbersome procedure accompanying the empirical judgment of a technician. Moreover, another device that measures MEP is required to find an individual motor hotspot using TMS.

In this study, we propose a novel alternative to TMS for motor hotspot identification based on EEGs measured during a simple finger-tapping motor task. A machine learning technique based on an artificial neural network (ANN) was applied to the measured EEG to localize individualized motor hotspots. The motor hotspot positions estimated by our proposed EEG-based machine-learning approach were then compared to those found by the traditional TMS-induced MEP to verify the feasibility of our approach. Preliminary results have been shown in [23].

Methods

Subjects

Thirty healthy subjects (10 females and 20 males; 25 ± 1.39 years; all right-handed) participated in this study, and they had no history of psychiatric diseases that might affect research results. All subjects received the information about the details of the experimental procedure and signed an informed consent for participation in the study. Appropriate monetary compensation for their participation was provided after the experiment. This study was approved by the Institutional Review Board (IRB) of Kumoh National Institute of Technology (No. 6250) and was conducted in accordance with the principles of the declaration of Helsinki.

Motor Hotspot Identification by TMS-Induced MEP

Before measuring the EEGs, the motor hotspots of both hands were identified for each subject using the MEPs of the first dorsal interosseous (FDI) muscle. To this end, Ag-AgCl disposable electrodes were attached to the FDI muscles of both hands to measure the MEPs (actiCHamp, Brain Products GmbH, Gilching, Germany). We applied single-pulse TMS to the contralateral motor cortex during the resting state (REMED., Daejeon, Korea), and determined the motor hotspot location that showed the maximum MEP with individual minimum stimulation intensity [24–26]. The motor hotspot locations were marked in a 3D coordinate system based on the vertex (Cz in the international 10-20 system) using a 3D digitizer (Polhemus Inc., Colchester, Vermont, USA), and they were used as the ground truth to compare with those estimated by our EEG-based motor-hotspot-identification approach.

EEG Recording

After identifying the individual motor hotspots using TMS for both hands, task-related EEG

data were measured using 63 EEG electrodes attached to the scalp based on the international

10-20 system (Figure 1). The ground and reference electrodes were attached to Fpz and FCz,

respectively. The positions of the EEG electrodes were marked in the 3D coordinate system as

for the motor hotspot locations identified by TMS-induced MEPs. The EEG data were sampled

at 1,000 Hz using a multi-channel EEG acquisition system (actiCHamp, Brain Products GmbH,

Gilching, Germany) while the subjects were performing a simple finger-tapping motor task.

Each subject performed the task 30 times, which consisted of pressing a button using their

index fingers whenever a red circle was presented in the center of a monitor (Figure 2); two

EEG measurement sessions were independently conducted using the left and right index fingers,

respectively. The subjects were given sufficient rest whenever they wanted during the

experiment to avoid fatigue. Moreover, they were instructed to remain relaxed during the

experiment without any movement to prevent unwanted physiological artifacts. Because we

carried out preliminary experiments with the first two subjects to validate our experimental

paradigm before the main experiment, the right hand was only employed for these two subjects.

In addition, we excluded the EEG data of one subject for both hands, and those of the left hand

for other three subjects due to high contamination of EEG data caused by physiological artifacts.

Thus, 29 and 25 EEG datasets were used for the right and left hands, respectively, for data

analysis.

EEG-Based Motor Hotspot Identification

EEG signal preprocessing was performed using the EEGLAB toolbox based on MATLAB

2017b (MathWorks, Natick, MA, USA). The raw EEG data were first down-sampled into 200

Hz to which we sequentially applied common average reference and bandpass filtering between

0.5 and 50.5 Hz (zero-phase 3rd-order Butterworth filter). Independent component analysis

(ICA) was then applied to the filtered EEG data to remove physiological artifacts. As

mentioned above, from visual inspection we excluded the EEG data significantly contaminated,

even after ICA application.

After preprocessing, we epoched EEG data between -0.5 and 0.5 s based on the key press

point for each trial to extract motor-related EEG features. Power spectral densities (PSDs) of

each EEG epoch were estimated in six frequency bands (delta: 1–4 Hz, theta: 4–8 Hz, alpha:

8–13 Hz, beta: 13–30 Hz, gamma: 30–50 Hz, and full: 1–50 Hz) using the fast Fourier

transform (FFT). An ANN was used to estimate the motor hotspot location using PSD features

where a subject-wise 10-fold cross-validation was performed with early stopping to prevent

overfitting. The PSD features of each channel and the 3D location coordinates of the motor

hotspot identified by TMS-induced MEP (ground truth) were used as the input for the ANN,

and the 3D location coordinates estimated using the PSD features were used as the output. The

numbers of input nodes, hidden nodes, and output nodes were 66 (63 channels + 3D motor

hotspot coordinate), 40 (empirically selected), and 3 (3D coordinate estimated by ANN),

respectively. We calculated the Euclidean distance between the 3D motor hotspot coordinates

identified by TMS and EEG to quantitatively estimate the error distance of our EEG-based

motor-hotspot-identification approach. The data analysis was independently performed for

each of the six EEG frequency bands.

To investigate the impact of the number of channels on the error distance, we repeatedly

performed the mentioned analysis by reducing the number of channels, in particular focusing

on the channels attached to the motor area (Figure 1). For this investigation, only gamma-band

PSD features (30–50 Hz) were used because they showed the best mean error distance when

using all 63 channels.

Results

Figure 3 shows the grand-average movement-related cortical potential (MRCP) measured during finger tapping between -0.5 to 0.5 s based on the task onset, which was obtained by averaging the MRCPs of four channels located on each hemisphere of the motor cortex to confirm the reliability of our EEG data (baseline period: -1 to -0.5 s). MRCP was clearly observed on the motor cortex during finger tapping for both hands, and, in particular, stronger MRCP was observed on the contralateral motor area [27–28].

Figure 4(a) shows the mean error distances of the motor hotspot locations estimated by our EEG-based machine learning approach for each hand with respect to the frequency band when using all 63 channels. The PSD features of the relatively high frequency bands (beta, gamma, and full bands) resulted in significantly lower mean error distances than those of the low frequency bands (delta, theta, and alpha) for both hands (RM-ANOVA with Bonferroni corrected p-value < 0.05: delta = theta = alpha > beta = gamma = full). No significant difference was observed between the left and right hand for all the frequency bands in terms of the error distance (paired t-test $p > 0.05$), except for the full band (right hand > left hand). Figure 4(b) shows a representative example of a single subject displaying the 3D locations of the motor hotspot identified by TMS (blue rectangle) and those of our EEG-based approach for both hands with respect to the frequency band. The motor hotspots estimated using the relatively high frequency bands were found closer to the ground truth motor hotspot identified by TMS, as compared to those estimated using the low frequency bands.

Figure 5(a) presents the mean error distances of the motor hotspot locations estimated by our EEG-based approach for each hand with respect to the number of channels. In general, the mean error distance monotonically and significantly increased as the number of channels

decreased (RM-ANOVA with Bonferroni corrected *p*-value < 0.05: Ch_Set5 > Ch_Set4 >

Ch_Set3 = Ch_Set2 > Ch_Set1 for the right hand and Ch_Set5 > Ch_Set4 = Ch_Set3 > Ch_Set2

> Ch_Set1 for the left hand). No significant difference was observed between the left and right

hand for all channel sets in terms of the error distance (paired t-test *p* > 0.05), except Ch_Set3

(left hand > right hand). Figure 5(b) shows the 3D locations of the motor hotspots identified by

TMS and our EEG-based approach with respect to the number of channels, showing a similar

trend to the result shown in Figure 5(a).

Discussion

In this study, we proposed a novel EEG-based motor-hotspot-identification method using an

ANN as a potential alternative to TMS to determine a tES target location for motor

rehabilitation. A minimum mean error distance of 0.22 ± 0.03 cm was attained when using the

gamma band PSD information extracted using all 63 channels, demonstrating the proof-of-

concept of the proposed EEG-based motor-hotspot-identification method. An EEG device is

required for the use of the proposed motor-hotspot-identification approach based on a machine

learning technique. In recent years, a portable tES device integrated with an EEG device was

introduced (e.g., NeuroElectronics Starstim), which could facilitate the use of our motor-

hotspot-identification approach without using TMS.

To check the practical feasibility of the proposed motor-hotspot-identification method, the

effect of the number of channels was investigated using the gamma band PSD features. As

expected, the lowest mean error was obtained using all channels, and the mean error distance

increased as the number of channels decreased (Figure 5). We obtained a mean error distance

of less than 1 cm with a number of channels above 17 (Ch_Set4) broadly attached on the motor

cortex. In addition, a mean error distance of approximately 1.32 cm was found when using only

nine channels attached on the midline of the motor cortex. It was documented that reliable FDI

MEP was observed with an area of approximately 12.9 cm^2 (3.6 cm $\times$ 3.6 cm), meaning that

a reliable MEP could be evoked with a distance of up to approximately 1.8 cm based on the

center of the motor hotspot area [29]. Therefore, although the mean error distance of our

proposed motor-hotspot-identification method statistically increased as the number of channels

decreased, it is expected that our EEG-based machine-learning approach could be utilized by

employing only the motor cortex channels for motor hotspot identification, thereby improving

its practicality.

The performance of our proposed motor-hotspot-identification method was better when

using higher frequency EEG features (delta: 1.20 ± 0.09 cm; theta: 1.17 ± 0.10 cm, alpha: 1.05

± 0.09 cm, beta: 0.33 ± 0.03 cm, gamma: 0.22 ± 0.03 cm, full: 0.30 ± 0.03 cm). In particular,

beta and gamma PSDs showed significantly lower mean error distances than those of delta,

theta, and alpha PSDs. Much evidence has been accumulated indicating that a hand motor task

significantly changes EEG frequency information in relatively higher frequency bands, i.e.,

alpha, beta, and gamma bands [30–34], and thus the higher performance obtained using higher

frequency EEG features can be explained from a neurophysiological point of view. However,

alpha PSDs did not show better performance as compared to delta and theta PSDs even though

alpha band is also closely associated with motor tasks, which should be further investigated in

future studies to optimize EEG frequency bands for more accurate motor hotspot identification

based on the proposed EEG-based machine-learning approach.

The difference between the mean error distances for the left and right hands were not

statistically significant for most comparison cases; two cases showed significant difference

(full band and Ch_Set3). This might indicate that our proposed approach is not sensitive to

handedness for motor hotspot identification. However, as all subjects recruited in this study

were right-handed, additional experiments are required with left-handed subjects to carefully

address the mentioned hypothesis.

The application of tES is not only limited to motor rehabilitation but can also be applied to

various psychiatric disorders, such as depression, schizophrenia, and attention deficit

hyperactivity disorder, to improve not only their cognitive functions, but also neurologically

relieve their symptoms [35–39]. In order to apply tES to psychiatric disorders, a target location

should be first determined, similarly to motor hotspots for motor rehabilitation. For cognitive

rehabilitation, in general, the anodal electrode is attached to F3 according to the international

10-20 system to stimulate the dorsolateral prefrontal cortex (DLPFC), which is known to be

associated with various cognitive functions, and the cathodal electrode is attached to F4 or the

supraorbital area of the contralateral hemisphere [39–41]. However, as the location of motor

hotspots varies from an individual to another, it could be assumed that the tES target location

for cognitive rehabilitation is also slightly different between individuals. Therefore, our

proposed motor-hotspot-identification approach might be also used for precisely finding a tES

target location to maximize the positive effect of cognitive rehabilitation.

On the other hand, the proposed motor-hotspot-identification method used the motor hotspot

coordinates identified by TMS as an input to construct the ANN model. Considering the

practical use of our proposed method without TMS, the motor hotspot information obtained

using TMS should be excluded in the process of finding a motor hotspot using a machine

learning technique, and thus we will continuously advance our proposed algorithm in such a

way that the motor-hotspot-identification algorithm ultimately does not require the motor

hotspot information obtained using TMS for a new subject. Despite the mentioned limitation,

it is thought that the results shown in this study could prove the proof-of-concept of the

proposed EEG-based motor-hotspot-identification method.

Conclusion

In this study, we proposed a novel EEG-based motor-hotspot-identification approach as an alternative to TMS and demonstrated its feasibility via EEG experiments. We also confirmed the possibility of using our proposed method to the development of a practical EEG-based motor-hotspot-identification system with a small number of channels attached only on the motor cortex. Because the brain activity patterns of patients with motor impairment would be different from those of healthy subjects, our proposed motor-hotspot-identification method should be further verified with patients to carefully demonstrate its clinical feasibility.

Abbreviations

tES: transcranial electrical stimulation; NIBS: non-invasive brain stimulation; tDCS: transcranial direct current stimulation; tACS: transcranial alternating current stimulation; tRNS: transcranial random noise stimulation; TMS: transcranial magnetic stimulation; ADHD: attention deficit hyperactivity disorder; DLPFC: dorsolateral prefrontal cortex; MEP: motor evoked potential; EEG: electroencephalography; IRB: institutional review board; FDI: first dorsal interosseous; EMG: electromyography; ICA: independent component analysis; PSD: power spectral density; FFT: fast Fourier transform; ANN: artificial neural network; MRCP: movement-related cortical potential.

Acknowledgements

Not applicable.

Authors' contributions

GYC, NJP, WSK, and HJH participated in designing the experimental paradigm, interpreted

295 the data, and drafted the manuscript. GYC and HTL conducted data collection, and GYC and

296 CHH analyzed the data.

Funding

298 This work was supported by Ministry of Trade Industry & Energy (MOTIE, Korea), Ministry

299 of Science & ICT (MSIT, Korea), and Ministry of Health & Welfare (MOHW, Korea) under

300 Technology Development Program for AI-Bio-Robot-Medicine Convergence (20001650), and

301 by Basic Science Research Program through the National Research Foundation of Korea (NRF)

302 funded by the Ministry of Education (No. 2019R1I1A3A01060732).

Availability of data and material

304 All data measured and analyzed during this study will be published in a public repository after

305 official publication.

Competing interests

GYC, WSK, and HJH have a patent-pending entitled "APPARATUS AND METHOD FOR

DETECTING MOTOR HOTSPOT POSITION USING BRAINWAVE, AND

TRANSCRANIAL ELECTRICAL STIMULATION APPARATUS USING THE SAME",

number 10-2020-0080383, and GYC, NJP, WSK, and HJH have a patent-pending entitled

"APPARATUS AND METHOD FOR DETECTING MOTOR HOTSPOT POSITION USING

BRAINWAVE, AND TRANSCRANIAL ELECTRICAL STIMULATION APPARATUS

USING THE SAME", number PCT/KR/2021/002118, which are relevant to this work.

Ethics approval and consent to participate

317 This study was approved by the Institutional Review Board (IRB) of Kumoh National Institute

318 of Technology (No. 6250) and was conducted in accordance with the principles of the

319 declaration of Helsinki.

Consent for publication

Not applicable. No subject is identifiable by the data presented in the figures and tables.

Reference

1. Hummel FC, Cohen LG: **Non-invasive brain stimulation: a new strategy to improve neurorehabilitation after stroke?** *Lancet Neurol* 2006, **5:** 708-712.

2. Lang AE, Lozano AM: **Parkinson's disease.** *N Engl J Med* 1998, **339:** 1130-1143.

3. Rowland LP, Shneider NA: **Amyotrophic lateral sclerosis.** *N Engl J Med* 2001, **344:** 1688-1700.

4. Jeffery DT, Norton JA, Roy FD, Gorassini MA: **Effects of transcranial direct current stimulation on the excitability of the leg motor cortex.** *Exp Brain Res* 2007, **182:** 281-287.

5. Dmochowski JP, Datta A, Huang Y, Richardson JD, Bikson M, Fridriksson J, Parra LC: **Targeted transcranial direct current stimulation for rehabilitation after stroke.** *Neuroimage* 2013, **75:** 12-19.

6. Claflin ES, Krishnan C, Khot SP: **Emerging treatments for motor rehabilitation after stroke.** *Neurohospitalist* 2015, **5:** 77-88.

7. Klomjai W, Lackmy-Vallée A, Roche N, Pradat-Diehl P, Marchand-Pauvert V, Katz R: **Repetitive transcranial magnetic stimulation and transcranial direct current stimulation in motor rehabilitation after stroke: an update.** *Ann Phys Rehab Med* 2015, **58:** 220-224.

8. Koh CL, Lin JH, Jeng JS, Huang SL, Hsieh CL: **Effects of Transcranial direct current stimulation with sensory modulation on stroke motor rehabilitation: a randomized controlled trial.** *Arch Phys Med Rehab* 2017, **98:** 2477-2484.

9. Nitsche MA, Paulus W: **Excitability changes induced in the human motor cortex by weak transcranial direct current stimulation.** *J Physiol* 2000, **527:** 633-639.

10. Antal A, Boros K, Poreisz C, Chaieb L, Terney D, Paulus W: **Comparatively weak after-effects of transcranial alternating current stimulation (tACS) on cortical excitability in humans.** *Brain Stimul* 2008, **1:** 97-105.

11. Allman C, Amadi U, Winkler AM, Wilkins L, Filippini N, Kischka U, Johansen-Berg H**:**
Ipsilesional anodal tDCS enhances the functional benefits of rehabilitation in patients
after stroke. *Sci Transl Med* 2016, **8:** 330re1-330re1.

12. Del Felice A, Castiglia L, Formaggio E, Cattelan M, Scarpa B, Manganotti P, Masiero S:
Personalized transcranial alternating current stimulation (tACS) and physical
therapy to treat motor and cognitive symptoms in Parkinson's disease: A randomized
cross-over trial. *NeuroImage Clin* 2019, **22:** 101768.

13. Bolognini N, Vallar G, Casati C, Latif LA, El-Nazer R, Williams J, Fregni F:
Neurophysiological and behavioral effects of tDCS combined with constraint-induced
movement therapy in poststroke patients. *Neurorehab Neural Re* 2011, **25:** 819-829.

14. Waters-Metenier S, Husain M, Wiestler T, Diedrichsen J: **Bihemispheric transcranial**
direct current stimulation enhances effector-independent representations of motor
synergy and sequence learning. *J Neurosci* 2014, **34:** 1037-1050.

15. Ferreira IS, Costa BT, Ramos CL, Lucena P, Thibaut A, Fregni F: **Searching for the**
optimal tDCS target for motor rehabilitation. *J NeuroEng Rehab* 2019, **16:** 90.

16. Boggio PS, Castro LO, Savagim EA, Braite R, Cruz VC, Rocha RR, Fregni F:
Enhancement of non-dominant hand motor function by anodal transcranial direct
current stimulation. *Neuroscience* 2006, **404:** 232-236.

17. Wade S, Hammond G: **Anodal transcranial direct current stimulation over premotor**
cortex facilitates observational learning of a motor sequence. *Eur J Neuroscience* 2015,
41: 1597-1602.

18. Ferrucci R, Vergari M, Cogiamanian F, Bocci T, Ciocca M, Tomasini E, de Riz M, Scarpini
E, Priori A: **Transcranial direct current stimulation (tDCS) for fatigue in multiple**
sclerosis. *NeuroRehabilitation* 2014, **34:** 121-127.

19. Volpe JJ: **Brain injury in the premature infant–from pathogenesis to prevention.** *Brain*
Dev 1997, **19:** 519-534.

20. Yousry TA, Schmid UD, Alkadhi H, Schmidt D, Peraud A, Buettner A, Winkler P:
Localization of the motor hand area to a knob on the precentral gyrus. A new
landmark. *Brain* 1997, **120:** 141-157.

21. Boroojerdi B, Foltys H, Krings T, Spetzger U, Thron A, Töpper R: **Localization of the**
motor hand area using transcranial magnetic stimulation and functional magnetic
resonance imaging. *Clin Neurophysiol* 1999, **110:** 699-704.

22. Weise K, Numssen O, Thielscher A, Hartwigsen G, Knösche TR: **A novel approach to**

localize cortical TMS effects. *NeuroImage* 2020, 209:116486.

23. Choi GY, Han CH, Lim H, Lim, Ku J, Kim WS, Hwang HJ: **Electroencephalography-based Motor Hotspot Detection.** *In Procs of the 13th International Joint Conference on Biomedical Engineering Systems and Technologies* 2020, **4:** 195-198.

24. Szubski C, Burtscher M, Loscher WN: **The effects of short-term hypoxia on motor cortex excitability and neuromuscular activation.** *J Appl Physiol* 2006, **101:** 1673-1677.

25. Wiethoff S, Hamada M, Rothwell JC: **Variability in response to transcranial direct current stimulation of the motor cortex.** *Brain Stimul* 2014, **7:** 468-475.

26. Hannah R, Iacovou A, Rothwell JC: **Direction of TDCS current flow in human sensorimotor cortex influences behavioural learning.** *Brain Stimul* 2019, **12:** 684-692.

27. Shibasaki H, Hallett M: **What is the Bereitschaftspotential?** *Clin Neurophysiol* 2006, **117:** 2341-2356. https://doi.org/10.1016/j.clinph.2006.04.025.

28. Jochumsen M, Niazi IK, Mrachacz-Kersting N, Farina D, Dremstrup K: **Detection and classification of movement-related cortical potentials associated with task force and speed.** *J Neural Eng* 2013, **10:** 056015. doi: 10.1088/1741-2560/10/5/056015.

29. Malcolm MP, Triggs WJ, Light KE, Shechtman O, Khandekar G, Rothi LG: **Reliability of motor cortex transcranial magnetic stimulation in four muscle representations.** *Clin Neurophysiol* 2006, **117:** 1037-1046. https://doi.org/10.1016/j.clinph.2006.02.005.

30. Hwang HJ, Kim KH, Jung YJ, Kim DW, Lee YH, Im CH: **An EEG-based real-time cortical functional connectivity imaging system.** *Med Biol Eng Comput* 2011, **49:** 985-995.

31. Lee NG, Kang SK, Lee DR, Hwang HJ, Jung JH, You JSH, Im CH, Kim DA, Jung A, Lee PT, Ki S, Kim KS: **Feasibility and test-retest reliability of an electroencephalography-based brain mapping system in children with cerebral palsy: a preliminary investigation.** *Arch Phys Med Rehab* 2012, **93:** 882-888.

32. Murta T, Chaudhary UJ, Tierney TM, Dias A, Leite M, Carmichael DW, Carmichael P, Lemieux L: **Phase–amplitude coupling and the BOLD signal: a simultaneous intracranial EEG (icEEG)-fMRI study in humans performing a finger-tapping task.** *Neuroimage* 2017, **146:** 438-451.

33. Stavrinou ML, Moraru L, Cimponeriu L, Della Penna S, Bezerianos A: **Evaluation of cortical connectivity during real and imagined rhythmic finger tapping.** *Brain Topogr* 2007, **19:** 137-145.

34. Ginter JrJ, Blinowska KJ, Kamiński M, Durka PJ, Pfurtscheller G, Neuper C: **Propagation**

of EEG activity in the beta and gamma band during movement imagery in humans. *Method Inform Medicine* 2005, **44:** 106-113.

35. Dallmer-Zerbe I, Popp F, Lam AP, Philipsen A, Herrmann CS: **Transcranial Alternating Current Stimulation (tACS) as a Tool to Modulate P300 Amplitude in Attention Deficit Hyperactivity Disorder (ADHD): Preliminary Findings.** *Brain Topogr* 2020, **33:** 191-207.

36. Breitling C, Zaehle T, Dannhauer M, Bonath B, Tegelbeckers J, Flechtner HH, Krauel K: **Improving interference control in ADHD patients with transcranial direct current stimulation (tDCS).** *Front Cell Neurosci* 2016, **10:** 72.

37. Brunelin J, Mondino M, Gassab L, Haesebaert F, Gaha L, Suaud-Chagny MF, Poulet E, et al: **Examining transcranial direct-current stimulation (tDCS) as a treatment for hallucinations in schizophrenia.** *Am J Psychiatry* 2012, **169:** 719-724.

38. George MS, Nahas Z, Molloy M, Speer AM, Oliver NC, Li XB, Ballenger JC, et al: **A controlled trial of daily left prefrontal cortex TMS for treating depression.** *Biol Psychiatry* 2000, **48:** 962-970.

39. Nitsche MA, Boggio PS, Fregni F, Pascual-Leone A: **Treatment of depression with transcranial direct current stimulation (tDCS): a review.** *Exp Neurol* 2009, **219:** 14-19.

40. Fregni F, Boggio PS, Nitsche MA, Rigonatti SP, Pascual-Leone A: **Cognitive effects of repeated sessions of transcranial direct current stimulation in patients with depression.** *Depress Anxiety* 2006, **23:** 482-484.

41. Padberg F, Kumpf U, Mansmann U, Palm U, Plewnia C, Langguth B, Keeser D, et al: **Prefrontal transcranial direct current stimulation (tDCS) as treatment for major depression: study design and methodology of a multicenter triple blind randomized placebo controlled trial (DepressionDC).** *Eur Arch Psy Clin N* 2017, **267:** 751-766.

Figure Captions

Figure 1. Electrode sites used for EEG data measurement. Five different channel sets were used for data analysis in order to study the impact of the number of channels on the error distance of motor hotspot location. Channels denoted by different colors in Ch_Set4 represent those selected on the contralateral motor cortex for the data analysis of each hand. The numbers in parentheses indicate the numbers of channels for the corresponding channel set.

454   Figure 2. Experimental paradigm used in this study. Each subject presses a space bar whenever

455   the red circle is presented in the middle of a screen. The red circle is maintained until the subject

456   presses the space bar. After the task period, the fixation ('+') mark is presented to indicate a

457   rest period.

459
460 Figure 3. Grand averaged movement-related cortical potential (MRCP) patterns of all subjects

461 on both hemispheres. Contralateral MRCPs are clearly observed.

Figure 4. (a) Mean error distances between the motor hotspot locations identified by TMS-induced MEP and our EEG-based machine learning approach with respect to the frequency band (RM-ANOVA with Bonferroni corrected p-value < 0.05: delta = theta = alpha > beta = gamma = full for both hands). The numbers below the bar graphs represent the mean error distances of the left and right hands. No significant difference was observed between the left and right hands for all the frequency bands in terms of the error distance (paired t-test $p > 0.05$), except the full band (right hand > left hand). (b) A representative example showing the 3D locations of the motor hotspots identified by TMS-induced MEP (blue rectangle) and our EEG-based approach with respect to the frequency band.

474
Figure 5. (a) Mean error distances between the motor hotspot locations identified by TMS-
induced MEP and our EEG-based machine learning approach with respect to the number of
channels (RM-ANOVA with Bonferroni corrected p-value < 0.05: Ch_Set5 > Ch_Set4 >
Ch_Set3 = Ch_Set2 > Ch_Set1 for the right hand and Ch_Set5 > Ch_Set4 = Ch_Set3 > Ch_Set2
> Ch_Set1 for the left hand). The numbers below the bar graphs represent the mean error
distances of the left and right hands. No significant difference was observed between the left
and right hand for all channel sets in terms of the error distance (paired t-test $p > 0.05$), except
Ch_Set3 (left hand > right hand). The abbreviation, *n.s.,* means no significant difference. (b) A
representative example showing the 3D locations of the motor hotspots identified by TMS (blue
rectangle) and the EEG-based approach with respect to the number of channels.