Fluorescence signatures of SARS CoV-2 spike S1 proteins and an human ACE-2: excitation-emission maps and fluorescence lifetimes

Jonas Grzesiak[a]**, Lea Fellner**[a]**, Karin Grünewald**[a]**, Christoph Kölbl**[a]**, Arne Walter**[a]**, Reinhold Horlacher**[b]**, Frank Duschek**[a]

[a]German Aerospace Center (DLR), Institute of Technical Physics, Im Langen Grund 1, 74239 Hardthausen, Germany
[b]Trenzyme GmbH, Byk-Gulden-Str. 2, 78467 Konstanz, Germany

Abstract. Fast and reliable detection of infectious virus loads of the SARS CoV-2 virus is still an important issue even after more than one year of the pandemic's outbreak. The spike protein's S1 subunit (including its receptor-binding domain) and human angiotensin-converting enzyme 2 (hACE2) are known as key players in the reproduction mechanism of the SARS CoV-2 virus. Because of its high sensitivity and simple application, fluorescence spectroscopy is promising to meet the sensitivity requirements for a virus detection in a clean environment. In such highly complex protein systems, a comprehensive knowledge of fluorescence data is presumed in order to evaluate the specificity of the spectra with respect to a possible detection. This includes full featured information on the fluorescence process, i. e. wavelength and time-dependent data. In this work, fluorescence spectral excitation-emission maps of the involved proteins are presented, namely the S1 part of the spike protein and its receptor-binding domain as well as the hACE2 enzyme. In addition, measurements of fluorescence lifetimes of the proteins are presented and analyzed by a bi-exponential kinetic approach.

Keywords: excitation-emission maps, fluorescence decay times, fluorescence, SARS CoV-2 , angiotensin-converting enzyme 2 .

*Jonas Grzesiak, tp-apw@dlr.de

1 Introduction

The global spread of the SARS CoV-2 virus, which started in early 2020 [1] and which was declared pandemic by the WHO in March 2020 [2], has promoted the intensive research on human coronaviruses [3], which are a global public health thread [4]. A major tool in monitoring and controlling the spread of the viruses is a fast, accurate and sensitive detection of the virus or the infection[5]. In the case of SARS CoV-2, a significant number of research articles on sampling techniques, nano biosensor technologies and antigen testing have been published and are reviewed in Ref. [6], [7, 8] and [9], respectively. An overview on photonic approaches for detecting virus loads and infections by optical means is given in Ref. [10]. The authors summarize that spectroscopic techniques, such as FTIR, Raman and fluorescence, still face challenges from the huge

diversity of viruses in the human organism. Among those methods, laser induced fluorescence (LIF) has been widely used tool to investigate proteins [11, 12, 13, 14]. Elastic light scattering and fluorescence have been used to observe virus particles and virus like particles [15, 16, 17]. The LIF technology takes advantage of increased signal intensities and has been applied to detect and classify viruses with excitation in the near UV [18, 19]. As pointed out in Ref. [20] , additional, orthogonal information on fluorescence characteristics can be retrieved not only from two-dimensional, spectral signatures (excitation-emission maps) but also from fluorescence decay times. Such unique features are of high importance for reliable sample classification[21].

In this work, high resolution fluorescence data of the proteins involved in the SARS CoV-2 binding process, namely the S1 part of the spike protein of SARS CoV-2 , the receptor-binding domain thereof and human angiotensin-converting enzyme 2 (hACE2) are reported as well as results on their fluorescence decay times. Spectral data and fluorescence decay times of the investigated proteins are compared and discussed in terms of uniqueness as classification features.

2 Experimental setup

The proteins used for this study are SARS-CoV-2 S1 RBD (S1 RBD domain with His-Tag, M=27.5 kDa), SARS-CoV-2 S1 (S1 domain with His-tag, 77.1 kDa) and hACE2 (ECD domain, tag free processes M=80 kDa). The proteins were produced by trenzyme GmbH using a transient production system in HEK293 suspension cells followed by purification either via the encoded His-Tag by IMAC (RBD and S1 protein) or by ion exchange chromatography (hACE2). After purification, the buffer of the purified proteins was exchanged to DPBS and the proteins were analyzed. The samples were dissolved in phosphate-buffered saline (Dulbecco's phosphate buffered saline (DPBS), pH=7.1-7.5, Sigma-Aldrich) to obtain comparable, molar concentrations of the pro-

teins. The final concentrations of the solutions were $22.6 \mu g/ml$ in case of the hACE2, $7.8 \mu g/ml$ in the RBD case and $22.2 \mu g/ml$ for the full length S1 protein.

The samples were investigated in UV-transparent cuvettes (fused silica glass, Hellma 117F), with a light travel path of 10 mm.

A FS5 fluorescence spectrometer from Edinburgh Instruments was used to obtain full fluorescence signatures of the proteins. The spectrometer uses a $90°$ setup and a variable wavelength excitation. Additionally it allows for time correlated single photon measurements of the samples. The fluorescence maps contain data for excitation wavelengths ranging from 220 nm – 300 nm and emission from 230 nm – 500 nm. Time dependent fluorescence signals have been recorded at an excitation at 260 nm, with 0.9 ns and the respective optimal emission wavelengths.

3 Results and Discussion

3.1 LIF spectral signatures of the hACE2 enzyme and the S1 protein

As an overview, the emission maps of the SARS CoV-2 S1 spike protein, the receptor-binding domain (RBD) thereof and the hACE2 are shown in Fig. 1. The diagonal peaks, at $+3430 cm^{-1}$ from the excitation wavelengths λ_{ex} are assigned to be the Raman signals of H_2O. The narrow peaks in the lower right corner of the graphs are assigned to be the second order diffraction maximum. Each map consists of two major maxima: the first one located at an excitation wavelength $\lambda_{ex} = 220$ nm with maximum emission at $\lambda_{em} = 325$ nm for S1 and hACE2 and with $\lambda_{em} = 320$ nm for the receptor-binding domain. The second major maximum appears at an excitation wavelength of $\lambda_{ex} = 280$ nm showing a maximum emission at $\lambda_{em} = 325$ nm for all three proteins.

Fig 1 Excitation-emission fluorescence map for hACE2 (left), the receptor-binding domain (center) and SARS CoV-2 S1 (right). Data are available on biorxiv.org as supplementary material.

Fig. 2 displays the normalized fluorescence response of the three samples in detail at excitation wavelengths $\lambda_{ex} = 220$ nm and $\lambda_{ex} = 280$ nm. For clearness the emission spectra obtained at $\lambda_{ex} = 280$ nm have been adjusted by the background signal from a pure PBS sample. The signals at $\lambda_{ex} = 220$ nm clearly show the Raman signal of the solvent as well as the second order diffraction maximum. The intensities of the respective signals are indicated by the Raman signal intensities and are more explicit in Fig. 3. The fluorescence peaks are located between 290 nm and 410 nm, which indicates that the excitation light mainly probes tryptophane and tyrosine sites inside the proteins. At both respective excitation wavelengths, the three proteins show significantly different fluorescence responses. This is manifested in differently shaped peaks, which in turn give full width half maximum values of the peaks from 55 nm for hACE2 and the receptor-binding do-

main to 60 nm for the S1 protein at an excitation wavelength of $\lambda_{ex} = 280$ nm. The same pattern can be seen for $\lambda_{ex} = 220$ nm, with slightly broader FWHM values: 58 nm for hACE2 and RBD and 63 nm for the S1 protein. These slightly broader spectra may emerge from a excited hot (i.e. vibrationally excited) state with broad emission spectrum due to the lower excitation wavelength.

Fig 2 Detailed, normalized fluorescence signals of the three proteins at $\lambda_{ex} = 220nm$ (left) and $\lambda_{ex} = 280nm$ (right).

Besides the pure protein signals, also the bound complex of the virus with its receptor is of interest for a possible virus detection. To obtain a protein-enzyme complex solution we combined both the S1 full length protein solution and the RBD protein solution with the hACE2 solution: 1 ml of each solution with the respective concentrations noted above to ensure equimolarity of both substances in the combined solution. Straight after combination, the mixture was stirred to ensure a fully mixed solution. From the receptor binding affinity of about 1-40 nM [22] and the concentration of the S1 protein and the RBD protein analytes of about $3.0 \cdot 10^{-7} \frac{Mol}{l}$ one obtains a half life time of the binding process in the range of $\tau \approx 3 - 35s$. Thus, the full binding of all hACE2 receptors to the virus proteins in our solution is ensured within 5 min, when the spectra were taken. Fig. 3 shows the fluorescence results on the bound complexes at an excitation wavelength of 280 nm. The signals have been adjusted by the background signal from a pure PBS sample and for

simplicity the normalized signals are shown. For comparison the single contributions are shown too. We compared the signal of the bound complexes with the sum of the signals of the single contributions (black dashed lines in the spectra). The difference spectra between the summed signals and the bound complexes are shown in Fig. 4. Here, the significant change of about -1% of the signal at 312 nm is due to differences in the H_2O Raman signal. At 350 - 400 nm however, a slight change in the slopes of the peak is found, indicating a possible conformation change under the binding process, especially in the S1 case.

Fig 3 Fluorescence spectra of bound protein complexes and their single contributions: a) Receptor-binding domain of the S1 protein and b) the full S1 protein.

3.2 Fluorescence decay times

Additionally, the fluorescence decay of the base solutions of the hACE2, the RBD protein and the S1 protein as well as of their bound complex solutions were measured at an excitation of 260 nm as a time correlated single photon counting signal, shown in Fig. 5. The decay signals were evaluated in the well established double exponential decay model [23, 24]. Here, the fits were performed as global fits, where all signals are taken into account at once and the fit for the bound complexes is fitted as a sum of the single contributions. Results of the fits are summarized in

Fig 4 Differences between the added single contributions and the bound complex for the RBD and the S1 case.

Fig 5 Time correlated single photon counting signals for the different solutions and their respective double exponential decay fit. a) for the RBD case, b) for the S1 case.

Table 1, the error noted is the error due to the fit routine. For both global fits the coefficient of determination is $R^2 = 0.99$. The fit analysis for the decays yields that the two time constants in the exponential decays, τ_1 and τ_2, are only slightly different for all proteins and bound protein-complexes. The recorded signals consist of a slow and a fast decay with $\tau_1 \approx (2.0 \pm 0.1)$ ns and $\tau_2 \approx (8.0 \pm 1.0)$ ns for all proteins.

	τ_1 (ns)	τ_2 (ns)
hACE2	2.0 ± 0.1	8.0 ± 1.4
RBD	1.9 ± 0.1	9.0 ± 1.9
S1	2.1 ± 0.1	6.9 ± 1.5

Table 1 Results of the double exponential fit to the different protein time correlated single photon signals at an excitation wavelength of 260 nm.

4 Summary

We have presented fluorescence data for the S1 spike protein of the SARS CoV-2, its receptor-binding domain and for its most important receptor in human organisms, hACE2. The spectra show distinguishable features such as slight, but clear shifts in the shapes of the LIF signatures for all proteins, e.g. manifested in different FWHMs of the respective fluorescence peaks. Additionally, the signals vary with the excitation wavelength and show for example a clear blueshift of about 5 nm for the receptor-binding domain at $\lambda_{ex} = 220$ nm in contrast to $\lambda_{ex} = 280$ nm. The fluorescence decay times have been determined to be $\tau_1 \approx (2.0 \pm 0.1)$ns and $\tau_2 \approx (8.0 \pm 1.0)$ns. For further insight, the limitations are the signal to noise ratios and the maximum possible integration times in order to avoid photolytic decomposition of the proteins in the UV. The data are publicly available for external R&D as described in section 4.

Disclosures

The authors declare no financial or ethical conflicts of interest.

Data and Materials Availability

Measurement data will be available on biorxiv.org.

References

[1] P. Zhou, X.-L. Yang, X.-G. Wang, *et al.*, "A pneumonia outbreak associated with a new coronavirus of probable bat origin," *Nature* **579**, 270–273 (2020).

[2] D. Cucinotta and M. Vanelli, "WHO Declares COVID-19 a Pandemic," *Acta Bio Medica Atenei Parmensis* **91**, 157–160 (2020).

[3] M. Haghani and M. C. J. Bliemer, "Covid-19 pandemic and the unprecedented mobilisation of scholarly efforts prompted by a health crisis: Scientometric comparisons across SARS, MERS and 2019-nCov literature," *Scientometrics* **125**, 2695–2726 (2020).

[4] E. de Wit, N. van Doremalen, D. Falzarano, *et al.*, "SARS and MERS: recent insights into emerging coronaviruses," *Nature Reviews Microbiology* **14**, 523–534 (2016).

[5] G. Guglielmi, "Rapid coronavirus tests: a guide for the perplexed," *Nature* **590**, 202–205 (2021).

[6] A. Mohammadi, E. Esmaeilzadeh, Y. Li, *et al.*, "SARS-CoV-2 detection in different respiratory sites: A systematic review and meta-analysis," *EBioMedicine* **59**, 102903 (2020).

[7] A. D. Teklemariam, M. Samaddar, M. G. Alharbi, *et al.*, "Biosensor and molecular-based methods for the detection of human coronaviruses: A review," *Molecular and Cellular Probes* **54**, 101662 (2020).

[8] B. Alhalaili, I. N. Popescu, O. Kamoun, *et al.*, "Nanobiosensors for the detection of novel coronavirus 2019-nCoV and other pandemic/epidemic respiratory viruses: A review, sars," *Sensors* **20**, 6591 (2020).

[9] T. Toptan, L. Eckermann, A. E. Pfeiffer, *et al.*, "Evaluation of a SARS-CoV-2 rapid antigen test: Potential to help reduce community spread?," *Journal of Clinical Virology* **135**, 104713 (2021).

[10] J. Lukose, S. Chidangil, and S. D. George, "Optical technologies for the detection of viruses like COVID-19: Progress and prospects," *Biosensors and Bioelectronics* **178**, 113004 (2021).

[11] J. R. Lakowicz, Ed., *Principles of Fluorescence Spectroscopy*, Springer US (2006).

[12] L. R. Dartnell, T. A. Roberts, G. Moore, *et al.*, "Fluorescence characterization of clinically-important bacteria," *PLoS ONE* **8**, e75270 (2013).

[13] M. S. Ammor, "Recent advances in the use of intrinsic fluorescence for bacterial identification and characterization," *J Fluoresc* **17**, 455–459 (2007).

[14] S. Cabredo, A. Parra, and J. Anzano, "Bacteria spectra obtained by laser induced fluorescence," *Journal of Fluorescence* **17**, 171–180 (2007).

[15] A. Alimova, A. Katz, R. Podder, *et al.*, "Virus particles monitored by fluorescence spectroscopy: A potential detection assay for macromolecular assembly," *Photochemistry and Photobiology* **80**, 41–46 (2004).

[16] M. Fang, W. Diao, B. Dong, *et al.*, "Detection of the assembly and disassembly of PCV2b virus-like particles using fluorescence spectroscopy analysis," *Intervirology* **58**(5), 318–323 (2015).

[17] M. Rüdt, P. Vormittag, N. Hillebrandt, *et al.*, "Process monitoring of virus-like particle re-

assembly by diafiltration with UV/vis spectroscopy and light scattering," *Biotechnology and Bioengineering* **116**, 1366–1379 (2019).

[18] B. S, G. JL, B. M, *et al.*, "Non-contact, real-time laser-induced fluorescence detection and monitoring of microbial contaminants on solid surfaces before, during and after decontamination," *Journal of Biosensors & Bioelectronics* **09**(02) (2018).

[19] V. Gabbarini, R. Rossi, J.-F. Ciparisse, *et al.*, "Laser-induced fluorescence (LIF) as a smart method for fast environmental virological analyses: validation on picornaviruses," *Scientific Reports* **9**, 1–7 (2019).

[20] F. Gebert, M. Kraus, L. Fellner, *et al.*, "Novel standoff detection system for the classification of chemical and biological hazardous substances combining temporal and spectral laser-induced fluorescence techniques⋆," *The European Physical Journal Plus* **133**, 1–5 (2018).

[21] M. Carestia, R. Pizzoferrato, M. Gelfusa, *et al.*, "Development of a rapid method for the automatic classification of biological agents' fluorescence spectral signatures," *Optical Engineering* **54**, 114105 (2015).

[22] Y. M. Bar-On, A. Flamholz, R. Phillips, *et al.*, "SARS-CoV-2 (COVID-19) by the numbers," *eLife* **9** (2020).

[23] K. Willis, A. Szabo, J. Drew, *et al.*, "Resolution of heterogeneous fluorescence into component decay-associated excitation spectra. application to subtilisins," *Biophysical Journal* **57**, 183–189 (1990).

[24] A. G. Szabo and D. M. Rayner, "Fluorescence decay of tryptophan conformers in aqueous solution," *Journal of the American Chemical Society* **102**, 554–563 (1980).

List of Figures

List of Tables