High-throughput, multiplex microfluidic test strip for the determination of antibiotic susceptibility in uropathogenic *E. coli* with smartphone detection

Sarah H. Needs[1,*], Zara Rafaque[2], Wajiha Imtiaz[3], Partha Ray[4], Simon Andrews[3], Alexander D. Edwards[1,5,*]

[1]School of Pharmacy, University of Reading, RG6 6DX, UK

[2]Department of Microbiology, Faculty of Health Sciences, Hazara University, Mansehra, Pakistan

[3]School of Biological Sciences, University of Reading, RG6 6DX, UK

[4]School of Agriculture, Policy and Development, University of Reading RG6 6DX, UK

[5]CFT Ltd, Daux Road, Billingshurst, RH14 9SJ, UK

*Corresponding author contact details:

Sarah H. Needs: *E-mail address:* s.h.needs@reading.ac.uk

Alexander D. Edwards: *E-mail address*: a.d.edwards@reading.ac.uk

Keywords

Antibiotic susceptibility testing, minimum inhibitory concentration, broth microdilution, antibiotic resistance, microfluidics, smartphone

Abstract

Antibiotic resistance in urinary tract infections is a major global challenge and improved cost-effective and high throughput antibiotic susceptibility tests (AST) are urgently needed to inform correct antibiotic selection. We evaluated a high throughput microfluidic test strip for AST and minimum inhibitory concentration (MIC) determination in 20 urinary pathogenic *E.*

coli (UPEC) isolates using six commonly prescribed or therapeutically beneficial antibiotics. The microfluidic MIC performs broth microdilution in 1 microliter volume capillaries, 100 X smaller than standard broth microdilution. Each test strip contains 10 parallel capillaries which are dipped into a single well of a 96 well plate, significantly increasing throughput over a microtitre plate. When tested with clinical UPEC isolates at standardised inoculum density, these devices gave 100% essential agreement (+/- 1 doubling dilution of antibiotic) to the gold standard microplate broth microdilution method described by CLSI. Although for some antibiotic/isolate combinations an earlier endpoint readout reduced accuracy, MIC test strips read at a 6h endpoint still gave 69 – 100 % essential agreement depending on the antibiotic. Growth could be detected significantly earlier than 6h, but with a trade-off between speed vs accuracy. These high-throughput, multiplexed test strips could be used to increase throughput and give faster results than microplates while retaining the core broth microdilution methodology of gold standard techniques for AST and MIC determination.

1 Introduction

Although improved diagnostic technology is widely acknowledged to be essential to combat antimicrobial resistance (AMR) (Trevas et al., 2020), antibiotic susceptibility testing (AST) of clinical samples such as urine remains challenging. Although urinary tract infections (UTI) are often self-limiting, a high prevalence places a significant burden on health systems, and when infections escalate hospital treatment can be required. Current primary care and hospital treatment protocols rely on empirical antibiotic administration at first presentation (Grigoryan et al., 2014, O'Grady et al., 2019), which exacerbates the clinical and economic burden because of a high level of broad and rapidly changing AMR profile necessitating constant updates to guidelines. When urine samples are taken and tested, adjustment of antibiotic is often required when results are returned days later. Resistance ranges up to ~50% of uropathogenic E. coli (UPEC) isolates in OECD countries and as high as >80% outside OECD, depending on antibiotic (Bryce et al., 2016) which limit the use of many common antibiotics, and reduce effectiveness of empirical prescribing. In antibiotics with

resistance rates > 20% empiric treatment is unsuitable (O'Grady et al., 2019). This leads to

consequences for the patient (infection may not be controlled if pathogen is resistant), the

clinician (limited treatment choices) and the public (emergence and spread of AMR and

subsequent drug-resistant infections). Even if susceptibility testing is requested, urine culture

plus AST takes 2-3 days, during which time infections worsen if the empirical antibiotic is

ineffective, or if treatment is delayed prior to culture results.

Several emerging strain-characterisation technologies to determine AST have been

developed to the point of large scale uptake into central microbiology laboratories (Pulido et

al., 2013). Many labs are now using mass spectroscopy for identification of clinically

important microbes by peptide mass fingerprinting to replace traditional analytical

microbiology techniques (Lavigne et al., 2013, Florio et al., 2020). Nucleic acid tests (NAT)

such as PCR can detect specific AMR genes, allowing rapid identification of known

resistance markers, but do not always reliably predict phenotypic antibiotic susceptibility and

can only detect target sequences – as these targets change, tests will need regular updating

informed by phenotypic-genotypic surveillance programs (Needs et al., 2020).

Phenotypic AST, including the current gold standard methods (Andrews, 2001, Andrews,

2005, Jorgensen and Turnidge, 2015) of broth microdilution (BMD) performed in a microtitre

plate (MTP) and disc diffusion assays performed on agar plates, demonstrate antibiotic

suitability by direct phenotypic identification of which antibiotics achieve bactericidal or

bacteriostatic effects (i.e. susceptible), vs which isolates grow in the presence of antibiotic

(i.e. resistant). Phenotypic AST currently requires a pure bacterial culture isolated by

overnight plating, laborious handling and requiring centralised laboratory equipment.

There are several automated phenotypic AST machines currently available. The Vitek[®] 2

(Biomerieux) reports bacterial identification and MIC results within 5-8 h, following bacterial

colony isolation after overnight agar plating. In contrast, the Accelerate Pheno[R] system is

designed to be used directly with samples (Kaprou et al., 2021). These methods are able to

reduce the time to result by hours and potentially days (Kinn et al., 2019) but still require the

use of centralised labs for testing.

Microfluidic technologies offer several advantages, especially in point-of-care applications,

including portability and rapid result times. Recently many examples of microfluidic analytical

microbiology devices illustrate the potential for AST miniaturisation (Avesar et al., 2017, Azizi

et al., 2018, Wu and Dekker, 2016, Kim et al., 2015, Kang et al., 2019). A "millifluidic" system

made from simple FEP tubing illustrated that analysis of AST can be miniaturised using

microdroplets allowing parallel analysis of large populations of individual cells (Baraban et

al., 2011). Whilst this microdroplet system offers improved resolution in quantitative AST

(Jiang et al., 2016) pure cultures from overnight plating and complex instrumentation are still

required. Microfluidic devices have also been developed for quantitative susceptibility

testing. Multi-chamber devices can distribute bacterial samples into chambers loaded with

antibiotics (Cira et al., 2012, Avesar et al., 2017, Azizi et al., 2018, He et al., 2020,

Matsumoto et al., 2016), or cells can be observed in microfluidic systems that generate

antibiotic gradients (Kim et al., 2019, Li et al., 2014). Microfluidic culture can be combined

with digital microscopy and image analysis to allow direct observation of the effects of

antibiotics on cell behaviour or growth (Choi et al., 2013, Wistrand-Yuen et al., 2020,

Baltekin et al., 2017, Matsumoto et al., 2016). However, microscopic observation of bacterial

growth inhibition by antibiotics does not directly map to traditional BMD or disc diffusion

results, and extensive equivalency testing is required. Electrochemical detection has been

combined with the conventional colorimetric metabolic dye resazurin to detect bacterial

growth in microfluidic devices (Besant et al., 2015). Most microfluidic microbiology studies

have focused either on exploring microdevice engineering science, or on addressing

fundamental biology research questions. One notable exception is a microplate-based

microfluidic AST device that has been developed into a fully automated product for testing

positive blood cultures to rapidly select antibiotics for the treatment of sepsis (Choi et al.,

2014, Choi et al., 2013). However, this laboratory device still requires pre-incubation of blood

sample prior to AST, to achieve a threshold level of organisms required for testing.

Incubation time and inoculum size are critical parameters in AST and deviation from

standardised methods can lead to errors in categorical agreement compared to the gold

standard. It is therefore important to understand how both of these may affect any new AST

devices or methods, however there are limited data in the literature exploring how these

parameters affect microfluidic analytical microbiology.

Shorter incubation times of 6- 8 h have been explored for disc diffusion assays

(Mancini et al., 2020, Hombach et al., 2017, Fröding et al., 2017, Chandrasekaran et al.,

2018). While categorical agreement differed depending on antibiotic/bacteria tested, overall

agreement was generally >80% and it was concluded this approach was useful as a

preliminary result to base urgent clinical decisions on (Fröding et al., 2017). Disc diffusion

assays of 11123 Enterobacteroles/antibiotic combinations incubated for 6 h found sharper,

more defined areas of inhibition with categorical agreement of 89.8% and major errors and

very major errors below 4%; after 8 h the categorical agreement rose to 98.5%. The most

common errors came from strains categorised as intermediate (Mancini et al., 2020); for

such strains where the inhibitory concentration for antimicrobial is close to the breakpoint for

scoring resistance vs susceptibility, there is always greater uncertainty in AST, even with

gold standard methods. In a clinical setting it is important to identify resistant strains early,

and false susceptibility is more of a risk than false resistance as the former can lead to

treatment failure, so for rapid test methods to be clinically useful must be optimised to

minimise false susceptibility.

The inoculum cell density is likewise an important factor when measuring MIC, and

an increased starting inoculum can increase the observed MIC for a number of reasons. This

well-known phenomenon is termed the inoculum effect (IE), and varying the starting bacterial

cell density can influence inhibition of growth by many antimicrobials; hence guidelines for

broth microdilution define acceptable ranges for inoculum density. Even within the CLSI

recommended range of inoculum density, categorical errors in antibiotic sensitivity have

been observed. An 8-fold difference was found in MIC between 2-8X10^5 CFU/mL for

carbapenem-resistant *Enterobacteriaceae* with meropenem (Smith and Kirby, 2018).

Estimating and adjusting cell density therefore remains an essential but laborious step for

many functional microbiology methods.

Healthcare budgets remain a major constraint, with high prevalence of UTI resulting in very

large volume of patients presenting with UTI symptoms in primary care. AMR is more

prevalent in parts of the world with most limited healthcare budgets (Pokharel et al., 2019).

Very low device costs are therefore needed and expensive capital equipment should be

avoided, and highly skilled technical sample processing or complex interpretation of results

are likewise undesirable. There remains an urgent need for novel AST methods that can

reduce workload, provide high quality data and directly test patient urine to without the need

for isolation or overnight plating. Ideally new methods would replicate conventional

phenotypic assays as closely as possible, to permit ease of interpretation by clinicians

accustomed to current gold standard methods. For example, miniaturised BMD utilising

growth detection that is similar to current automated lab BMD equipment would accelerate

adoption of new technology by permitting simple clinical interpretation of similar results

measured and presented in similar ways to current gold standard AST, such as MIC values

and scoring resistance/susceptibility at the same breakpoint concentrations. As breakpoint

concentrations are constantly monitored, ensuring new methods can make direct use of

current guidelines is vital to avoid continuous equivalency testing.

Here, we fully evaluate a high throughput system for analytical microbiology that uses a

fluoropolymer microcapillary film (MCF) that contains a parallel array of ten microcapillaries.

The MCF is manufactured by melt-extrusion and multiplex test strips can be produced in

large batches (Barbosa et al., 2015, Castanheira et al., 2015, Barbosa et al., 2014, Edwards

et al., 2011), allowing us to make and test hundreds of very low-cost test devices,

representing thousands of individual microcapillary tests. These devices incorporate a

hydrophilic polymer coating to modify the internal walls of the capillaries allowing sample

uptake by capillary action. Using this method, samples can be loaded easily by dipping the

test strips into the well of a 96-well plate with the sample drawn up into the 10 capillaries by

capillary action. Previously, we showed that these "Lab-on-a-stick" devices can perform a

range of classical microbiology tests to identify and count viable cells (Reis et al., 2016,

Needs et al., 2019, Dönmez et al., 2020, Needs et al., 2021). Here we present the first full

demonstration of MIC measurement and antibiotic susceptibility testing for a range of

antibiotics in clinically relevant UPEC isolates. We explore the impact of inoculum cell

density on microfluidic broth microdilution and assess the kinetics of antibiotic susceptibility

testing in a convenient miniaturised format, to understand better how to use microfluidics to

speed up analytical microbiology.

168 2 Material and Methods

169 2.1 Bacterial Isolates

Reference strain *E. coli* 25922 was purchased from ATCC and was used as a quality control.

E. coli 13352 was purchased from NCTC. The uropathogenic *E. coli* (UPEC) strains were

collected at a tertiary care hospital of Pakistan from community acquired UTI patients

(Rafaque et al., 2018, Rafaque et al., 2019) under a study that was approved by the Ethical

Review Board (ERB) of the Pakistan Institute of Medical Sciences. UPEC were identified

using standard microbiological and biochemical tests as described before (Ali et al., 2019).

176 2.2 Preparation of antibiotic loaded microcapillary test strips

MCF of 1 to 5 m lengths were given an internal hydrophilic coating by incubation with a 5

mg/mL solution of polyvinyl alcohol (PVOH) in water (MW 146,000-186,000, >99%

hydrolysed, Sigma-Aldrich, UK) at room temperature for 2h (Pivetal et al., 2017, Reis et al.,

2016). Coated strips were washed with 5 ml of PBS with 0.5 % Tween 20 (Sigma-Aldrich,

UK) to remove residual PVOH, and dried on a vacuum manifold for 20 minutes per metre

using a SLS Lab Basics Mini Vacuum Pump with PTFE Coated Diaphragm (Scientific

Laboratory Supplies, UK).

Ciprofloxacin, cefoxitin, trimethoprim and amoxicillin were purchased from Sigma Aldrich. Nitrofurantoin and Cephalexin were from purchased from Fischer Scientific. Serial dilutions of antibiotics in sterilised milliQ water were injected using a 30-gauge needle into individual capillaries in up to 1.5 m long strips of MCF. The MCF strip was cut to 17 mm individual test strips and frozen overnight at – 80 °C. Test strips were freeze-dried for >4 h on an Edwards Modulyo freeze drier. Test strips were vacuum packed and stored at -20 °C until use.

2.3 Microplate Broth Microdilution

Bacterial inocula were prepared according to CLSI standard for growth preparation. Briefly, 3-5 individual colonies were inoculated into Mueller-Hinton broth and grown for 4-6 h until visibly turbid. The suspension was diluted to 0.5 McFarland equivalent which corresponds to approximately $1X10^8$ CFU/mL. The culture was diluted 1:150 in Mueller-Hinton to give $1X10^6$ CFU/mL suspension and finally 50μL was added to 50μL of antibiotic solutions in a microplate to give a 1:2 dilution to give a final suspension of bacteria at $5X10^5$ CFU/mL in two-fold dilutions of antibiotic. For BMD including resazurin, dye was added to a final concentration of 0.25 mg/mL. Plates were incubated at 37 °C overnight. MIC measurements were measured in duplicate (Fig 1a). The MIC was recorded as the lowest concentration of antibiotic that did not show resazurin conversion or turbidity. For tests that varied within duplicate measurements, the highest MIC was recorded. A growth control (no antibiotic) and a sterile control (MH broth only) was included for all isolates.

2.4 Microcapillary Broth Microdilution

Individual 17 mm test strips were clipped into 3D printed holders, able to hold up to 12 test strips. For MIC tests in MCF strips, bacteria inoculum was prepared as per the CLSI standard for growth preparation as before at a final concentration of $5X10^5$ CFU/mL and mixed in a 96 well plate with a final concentration of 0.25 mg/mL resazurin in a total volume of 200 µL. The test strips were dipped into each well. Once all capillaries were filled (approximately 3 seconds) end covers filled with silicone grease were placed on each end to

stop sample evaporation. Samples were incubated overnight at 37 °C and color change

monitored using an in-house automated raspberry pi camera imaging system (Needs et al.,

2019). MIC measurements were taken in duplicate with two wells used for each

measurement (Fig 1a). The MIC was recorded as the lowest concentration of antibiotic that

did not show resazurin conversion or turbidity. For tests that varied in duplicate

measurements, the highest MIC was recorded. A growth control (no antibiotic) capillary was

included for all test strips.

For microcapillary BMD experiments using a higher inoculum density, CFU/mL was

calculated using a spot plating protocol. Briefly, 10 μL of serially diluted culture were taken

from the microtitre wells and plated on both LB agar and the Gram-negative selective media,

MacConkey agar to determine cell count and ensure no contamination with other organisms.

Plates were incubated overnight at 37 °C.

2.5 Data Analysis

Two recordings of the test strips were made. A single endpoint was recorded by colorimetric

measurement after overnight incubation at 37 °C using a smartphone camera (iPhone 6s) or

digital camera (Canon Powershot S120). The strips were placed on an even white light

illumination screen. Isolates were scored suspectable or resistant based on EUCAST v11.0

breakpoint values in Table 1.

For time-lapse imaging, resazurin conversion was recorded every 15 minutes using the

POLIR robot (Needs et al., 2019), with 3280 x 2464 resolution images taken with a

Raspberry Pi v2 camera. MatLab scripts were used to analyse time-lapse image series of

bacterial growth in MCF, and the code can be accessed here:

https://gitlab.com/sneeds/code-repository. Briefly, color images were split into red, blue and

green (RGB) channels and the red channel analysed for absorbance using:

$$A = log_{10}\left(\frac{I_o}{I}\right)$$

Time to resazurin conversion was calculated at which half the starting absorbance in the red

channel was reached.

3 Results

3.1 Broth microdilution in microcapillary test strips accurately determines MIC for

uropathogenic *E. coli*

The novel microcapillary BMD method conducts AST in dip-and-test microcapillary film

(MCF) strips, with each test strip containing 10 different conditions (9 antibiotic

concentrations plus no antibiotic control) in 1 µL samples, thereby reducing the test volume

over 100-fold from microplates. To understand if miniaturisation affects BMD performance,

we validated this microcapillary BMD AST, by comparing to a standard microplate BMD

following CLSI guidelines. Growth was quantified using the colorimetric growth indicator,

resazurin, and although full growth kinetics were recorded, we initially focussed on an

overnight endpoint readout. Microplate BMD and microcapillary BMD were performed in

parallel, testing a panel of 6 antibiotics commonly used in the treatment of UTIs, against an

E. coli quality control (QC) strain, an *E. coli* reference strain, and 20 UPEC isolates.

E. coli cultured in Mueller-Hinton broth were diluted to a standard inoculum cell density of

$5X10^5$ CFU/mL in Mueller-Hinton broth with a final resazurin concentration of 0.25 mg/mL in

both microplates and MCF test strips in parallel, and overnight colorimetric endpoint images

taken to score growth/no growth, and the same colour changes were observed in both

microplate and microcapillaries (Fig 2). The capillaries in MCF test strips have a 20-fold

shorter light pathlength than a microtitre plate, therefore the concentration of resazurin had

to be high enough to be clearly visible in the capillaries. Endpoint colour change was scored

after overnight growth from digital images, with pink/white turbid capillaries indicating

bacterial growth whereas samples with a susceptible bacterial population remained blue (Fig

1b and 2).

BMD in MCF and microplates was performed in duplicate and the highest MIC value for

each isolate/antibiotic combination was recorded in Table S1. A total of 2640 data points

were collected from the combination of 6 antibiotics and 22 samples (20 UPEC isolates plus

1 QC strain and 1 reference strain, totalling 132 antibiotic/isolate combinations). The *E. coli*

25922 strain was used as the QC strain. All MIC measured by microplate BMD and

microcapillary BMD were within the acceptable MIC range according to EUCAST QC Table v

11.0 for this QC strain. Out of all combinations of *E. coli* strains tested against the 6

antibiotics, there were 56 combinations where MIC was measurable and for these there was

100% (56/56) essential agreement between microcapillary and microplate, with MIC values

for each method within ±1 doubling dilution of antibiotic. No MIC value was attributed to 74

microplate BMD tests and 71 microcapillary BMD test strips. Of these, no MIC value was

attributed for 3 microplate BMD conditions in which growth was only detected in the control

without antibiotic (i.e. MIC was below the lowest concentration; scored as susceptible), and

for 71 microplate BMD and 71 microcapillary BMD test strips where all antibiotic containing

capillaries showed growth thus MIC was greater than the highest included concentration

(scored as resistant; Table 2). Categorical agreement was achieved when isolates were

scored as susceptible or resistant at the EUCAST breakpoint values shown in Table 1 by

both methods. There was a 96% (127/132 isolate/antibiotic combinations) categorical

agreement across all antibiotics tested and 100% categorical agreement for all isolates

tested against trimethoprim, nitrofurantoin, cephalexin, and amoxicillin (Table 2). The

microcapillary BMD had 2.3% (3/132) very major errors where 3 isolates were categorised

as susceptible in contrast to microplate BMD that categorised them as resistant. All three

major errors occurred with cefoxitin for which the MIC determined by microplate BMD was

within ±1 $\log_2$ fold change of the breakpoint, indicating the breakpoint value is close to the

threshold for inhibition. In this situation the variability in resistance/susceptibility scoring is

known to increase significantly, leading to higher uncertainty in AST results even with gold

standard methods. Microcapillary BMD showed a further 1.5% (2/132) major errors in which

isolates tested for cefoxitin and ciprofloxacin were identified as resistant but categorised as

susceptible by microplate BMD.

Overall, we conclude that microcapillary BMD performs very well and that the miniaturisation

of broth microdilution from microplate wells down to microcapillaries has little if any impact

on MIC determination and scoring for antibiotic resistance.

3.2 Time-resolved growth analysis in microcapillary test strips allows earlier endpoint

MIC giving accurate antibiotic susceptibility within 6h

Previously we established that even at low cell densities, *E. coli* growth can be detected

within 7 h and at the inoculum recommended for AST within approximately 3-5 h (Needs et

al., 2021) (Fig S1). MIC were measured and susceptibility/resistance scored for

microcapillary BMD after 6 h incubation (Fig S2-7), and results compared to overnight

microplate BMD results (Table 3). In only two cases, MIC was not determinable due to no

growth detected at 6h in the no antibiotic control, for *E. coli* 13352 tested against

trimethoprim and cefoxitin. No MIC value was attributed for 9 microcapillary nitrofurantoin

BMD test strips, in which growth was only detected in the growth control without antibiotic

(i.e. MIC below lowest concentration; scored as susceptible), or for a further 72

microcapillary BMD test strips where all antibiotic containing capillaries showed growth and

thus the MIC was above the highest concentration (scored as resistant; Table 3).

For those conditions where MIC was within the measurable range, overall essential

agreement in MIC level across the antibiotics tested dropped to 78% (38/49 isolate/antibiotic

combinations). However, overall categorical agreement remained high at 91% (120/132

isolate/antibiotic combinations). Very major errors were identified in 7.7% (10/130) of cases.

The accuracy was reduced in a number of ways. Where growth was not clearly detected at

the 6h timepoint, because the antibiotic slows growth, false susceptibility could be scored,

but at an overnight timepoint even the slower growth still leads to a positive score. On the

other hand, slower growing isolates growth may not be strongly detected at 6h even in

control without antibiotic; for these strips susceptibility and MIC cannot be scored.

Kinetics of growth was clearly influenced more by some but not other antibiotics, depending

on susceptibility/resistant status. Cefalexin and ciprofloxacin show very similar MIC results at

6 and 16 h varying by only one doubling dilution of antibiotic (Table S3, Fig 3a-b). Isolates

categorised as resistant show indistinguishable growth kinetics in the presence of antibiotics

(Figure 3c). Nitrofurantoin, in which all isolates were categorised as susceptible, showed the

most distinct delay in growth kinetics in the presence of antibiotic (Fig 3d). Shorter incubation

times may be less suitable for high resolution quantitation of MIC, and may be more

accurate for some antibiotics than others, however overall accuracy of microcapillary BMD

for scoring resistance vs sensitivity remained high even with a 6h endpoint readout. As with

overnight endpoint, the growth could readily be recorded using a smartphone camera or if

necessary, scored by eye. Overall, this indicates that early endpoints can be reliable for

determining resistance and may aid in decision making for resistant isolates.

3.3 Effect of inoculum density on microfluidic MIC determination

Detecting antibiotic susceptibility and measuring MIC is known to be significantly

affected by the initial inoculum density, and in the development of new methodologies it is

vital to identify how sensitive results are to variables such as inoculum cell density that are

not easy to control. Here, we focussed on understanding how starting cell density affects

microcapillary BMD performance, and seeing if IE reduces accuracy for both MIC

quantitation and AST scoring, as it does for conventional microplate methods. Microcapillary

BMD test strips were dipped into serial 10-fold dilutions of each bacterial isolate, starting with

10^7 CFU/mL (Fig 4), 100X higher than the recommended density. Inoculum densities 100X

above the CLSI recommended resulted in growth at antibiotic concentrations significantly

above the MIC, giving false resistance scores for *E. coli* 25922 and UPEC 2165 for

ciprofloxacin and cefalexin and in UPEC 2165 for nitrofurantoin (Fig 4b and Table 4). This

confirms that as for microplate BMD, the starting inoculum should be standardised to avoid

false resistance from high starting cell density.

Inoculum densities below the CLSI recommended density gave MIC measurements within

the $\pm$ 1 log$_2$ fold acceptable range, suggesting the microcapillary BMD is relatively robust as

long as inoculum density remains low enough to avoid false resistance or raised MIC levels.

This indicates that the microcapillary BMD test strips are compatible with lower bacterial

inoculum, suggesting this method may be robust enough to be used in environments where

precise inoculum density may be harder to achieve (e.g. less well-resourced laboratories).

3.4 Detecting delay in growth from kinetic analysis allows fast and accurate

categorical determination at higher cell densities

Rapid AST microdevices have been developed that use time-lapse digital microscopy

imaging to determine delayed growth in the presence of antibiotics compared to a growth

control (Avesar et al., 2017, Kang et al., 2019, Busche et al., 2019, Sun et al., 2019, Osaid et

al., 2021). Higher inoculum densities convert resazurin faster. The use of miniaturised BMD

tests combined with kinetic data collection offers further opportunity to gain more detailed

insight into bacterial response to antimicrobial agents, for example through detection of

differential growth kinetics indicating susceptibility. MIC was calculated based on differential

growth kinetics for different inoculum densities (Table 5, Fig S8-9). MIC was calculated by

scoring susceptibility when growth was delayed, but resistant when at a particular antibiotic

concentration, growth was detected within 1 h of the no antibiotic control. It was evident that

isolates that were categorised as susceptible, and with MIC well below the breakpoint

concentration, were consistently scored as susceptible by this method, even at inoculum

densities 100X higher than the recommended range. Nitrofurantoin showed significantly

delayed growth compared to the growth control at all inoculum densities (Fig 4c). False

resistance at higher inocula was still observed for some isolate/antibiotic combinations but

with an MIC still within one doubling dilution of the breakpoint antibiotic concentration. For

example, *E. coli* 25922 has an MIC for ciprofloxacin of 0.015 µg/mL. Based on growth

kinetics, inocula as high as 10^7 CFU/mL would be scored as susceptible. The UPEC isolate

2165 with an MIC of 0.25 µg/mL – the same value as the breakpoint – exhibited false

resistance at inoculum densities greater than 10^5 CFU/mL. It is clear therefore that the

fastest way to score resistance or susceptibility, or to estimate MIC, is through detecting

delayed growth with a high inoculum density that allows early growth detection. The

accuracy of this method varies depending on the antibiotic, and as with even gold standard

methods, if the MIC of isolates is close to breakpoint concentrations, AST is less reliable.

4. Discussion

Overall, by measuring growth with large panels of antibiotic combined with a set of *E. coli*

strains with a range of sensitivities, we observed that microfluidic broth microdilution

performs very similarly to microplate broth microdilution. While agreement between the

methods was very close (Table 2), the conditions where different results were obtained

represented samples with marginal MIC where variability in AST is known to be highest. We

observed the IE, that is well characterised for microplate measurements, also affects

microfluidic broth microdilution (Fig 4), with implications for direct testing methods where

bacterial inoculum density may be unknown or difficult to control. As might be expected,

faster results can be obtained with earlier growth readouts, yet some trade-off is apparent

between speed and accuracy. Nevertheless, even with 6h endpoint (Table 3) or when

differential growth is used to score susceptibility (Table 5), accuracy remained excellent for

some antibiotics and overall was 91%.

Antibiotics were chosen based on their use in UTI treatment but also to cover a range of

antibiotic classes. UTIs are the most common outpatient infection (Medina and Castillo-Pino,

2019). In the UK, the 2018 NICE guidelines recommend a first-choice treatment for

uncomplicated UTI in women over 16 years old of Nitrofurantoin or Trimethoprim, despite

resistance rates to trimethoprim around 30% (Watts et al., 2020). During pregnancy, the

need for safe and effective drugs is increased, while nitrofurantoin remains first choice,

amoxicillin is only recommended if the susceptibility is known due to resistance rates above

50% in some demographics (Watts et al., 2020). Cefoxitin, while not currently used for

clinical prescription, has been cited as an effective *in vitro* against UTI pathogens including

extended-spectrum beta-lactamase producing Enterobacteriaceae (ESBL-E) (Mambie et al.,

2016, Senard et al., 2018, Guet-Revillet et al., 2014). Since antibiotic loading of the strips is

by freeze-drying, it is expected that the test can be adapted to most antibiotics for different

applications.

Bacterial growth was monitored using the metabolic indicator resazurin. Resazurin

has been used in many different plate assays studying bacterial and cellular growth,

mitochondrial function, and MIC of antimicrobial agents. However, with many cell viability

test methods the resazurin is added towards the end of the experiment and incubated for

several hours to measure metabolic activity after a primary incubation period with

antimicrobial agent but no dye, to reduce the risk of the dye interacting with the antimicrobial

(Kim and Jang, 2018, Teh et al., 2017, Rakhmawatie et al., 2019, Elshikh et al., 2016,

Sarker et al., 2007). In contrast, for the novel microcapillary BMD method, the resazurin is

present from the start. Microplate BMD with the addition of resazurin at the start was

therefore compared to microplate BMD without dye to evaluate if the presence of resazurin

dye affected antimicrobial susceptibility and MIC results. There was 100% essential

agreement when microplate BMD was compared with and without resazurin dye for the MIC

values of nitrofurantoin and cephalexin, and 100% essential agreement between

microcapillary BMD (that includes resazurin) vs microplate BMD conducted without resazurin

(Table S2). This confirms that the inclusion of resazurin does not interfere with MIC

determination for *E. coli* when mixed with the sample at 0.25 mg/mL and incubated

overnight. Resazurin can be used with fluorescence detection (Needs et al., 2021) with the

potential for earlier growth detection compared to colorimetric detection, however, in

previous studies, some isolates showed a decay in fluorescence after the initial increase

(Needs et al., 2019). This has been observed previously in assays in which fluorescent

resorufin is further reduced to the colorless, non -fluorescence hydroresorufin. This reduces

the reliability of an endpoint measurement as loss of fluorescent signal may be mistakenly

interpreted as negative growth i.e. scored as false susceptible (Uzarski et al., 2017). For this

reason, we measured growth colorimetrically. Although small, the individual capillaries are

large enough to be scored by the naked eye if digital camera is not available to record result.

A smartphone camera is ideal for digitally recording these test results, and wide range of

different quality digital cameras have been shown to be equally effective for recording

microcapillary bioassay results (Needs et al., 2021, Jégouic et al., 2021).

The ability to detect growth of very low viable cell numbers – ideally down to a single colony

forming unit – may be critical for some analytical applications. We previously identified the

limit of detection of a 1 μL microcapillary test strip to be in the order of 5×10^3 CFU/mL; at this

concentration 99% of capillaries are predicted to contain at least 1 bacteria every test,

allowing MIC measurement (Needs et al., 2021). The initial inoculum density correlates to

the speed of detecting resazurin conversion, as early growth at low density does not convert

enough dye to influence the capillary color, thus lower densities take longer before growth

can be detected (Needs et al., 2021). Thus, although reduced inoculum density still gave

categorical agreement with microplate BMD, early timepoint endpoint readings at 6h are not

possible as growth is not always detected without antibiotic at this timepoint (Fig S1).

Whilst overnight plating to obtain individual colony isolates, plus cell density adjustment prior

to testing is still required, these results confirm that microcapillary BMD has equivalent

performance to standard microplate BMD. The MCF test strips significantly increase

throughput compared to microplates alone, whilst remaining compatible with conventional

microtitre plate format found in most laboratories for sample preparation. While one standard

microplate BMD can test 6 isolates in duplicate with 6 concentrations of a single antibiotic,

the microcapillary BMD can test 96 isolates, with a single isolate per well. Each

microcapillary BMD test strip provides 9 antibiotic concentrations plus control, and each well

can be dipped with multiple test strips for replicates. Assuming picking and adjusting isolate

concentration, and plating samples out, is rate limiting, microcapillary BMD is therefore 16

times higher throughput than microplates.

4 Conclusions

We present a high throughput method for determining antibiotic resistance or susceptibility

and for measuring MIC, using a simple to use microfluidic dipstick device, and validate this

method with uropathogenic *E. coli* and antibiotics commonly used to treat UTI. The test is

fully compatible with existing lab equipment such as multichannel pipettes and 96-well

format microplates for sample preparation and does not require additional equipment such

as pumps or controllers for microfluidic liquid handling. The colorimetric detection of bacterial

growth allows easy readout and simple analysis, reducing the need for equipment

investment – most digital cameras including smartphone cameras are capable of recording

color change, or growth can be scored by eye. The microcapillary BMD can provide rapid

categorical susceptibility/resistance scoring at time points within 6 h based on differential

growth. Kinetic analysis can provide robust S/R categorisation above the recommended

inoculum density, however, isolates with MIC within two-fold dilution of the breakpoint may

show false resistance.

Tables

Table 1. EUCAST v 11.0 antibiotic MIC breakpoints used for categorical analysis

Antibiotic	EUCAST MIC breakpoints (mg/L) ≤S/>R
Cefoxitin	≤8
Ciprofloxacin	≤0.25
Trimethoprim	≤4
Nitrofurantoin	≤64
Cephalexin	≤16
Amoxicillin	≤8

Table 2. AST agreement between microplate BMD and microcapillary BMD.

Antibiotic	Microplate BMD		Microcapillary BMD		% Essential Agreement	% Categorical Agreement
	Susceptible	Resistant	Susceptible	Resistant		
Trimethoprim	18% (4/22)	82% (18/22)	18% (4/22)	82% (18/22)		**100%** (22/22)
Nitrofurantoin	100% (22/22)	0% (0/22)	100% (22/22)	0% (0/22)		**100%** (22/22)
Cephalexin	14% (3/22)	86% (19/22)	14% (3/22)	86% (19/22)	**100%** (55/55)	**100%** (22/22)
Amoxicillin	5% (1/22)	95% (21/22)	5% (1/22)	95% (21/22)		**100%** (22/22)
Ciprofloxacin	18% (4/22)	82% (18/22)	17% (3/22)	86% (19/22)		**95%** (21/22)
Cefoxitin	27% (6/22)	73% (16/22)	36% (8/22)	64% (14/22)		**82% (**18/22)

Table 3. Agreement between microcapillary test strips read at 6h compared to overnight

Antibiotic	BMD (16 – 18 h)		Microcapillary BMD (6h)		% Essential Agreement	% Categorical Agreement
	Susceptible	Resistant	Susceptible	Resistant		
Trimethoprim[1]	18% (4/22)	82% (18/22)	14% (3/21)	86% (18/21)	**100%** (3/3)	**100%** (21/21)
Nitrofurantoin	100% (22/22)	0% (0/22)	100% (22/22)	0% (0/22)	**69%** (9/13)	**100%** (22/22)
Cephalexin	14% (3/22)	86% (19/22)	14% (3/22)	86% (19/22)	**83%** (5/6)	**100%** (22/22)
Amoxicillin	5% (1/22)	95% (21/22)	5% (1/22)	95% (21/22)	**100%** (1/1)	**100%** (22/22)
Ciprofloxacin	18% (4/22)	82% (18/22)	27% (6/22)	73% (16/22)	**100%** (6/6)	**91%** (20/22)
Cefoxitin[2]	27% (6/22)	73% (16/22)	62% (13/21)	38% (8/21)	**71%** (15/21)	**62%** (13/21)

---

[1] *E. coli* 13352 unable to determine MIC at 6h in microcapillary BMD.
[2] *E. coli* 13352 unable to determine MIC at 6h in microcapillary BMD.

Table 4. Determination of endpoint MIC using varying inoculum density.

	UPEC 2165					
	Nitrofurantoin		Cefalexin		Ciprofloxacin	
CFU/mL	Endpoint Microplate BMD ($5X10^5$CFU/mL)	Microcapillary BMD	Endpoint Microplate BMD ($5X10^5$CFU/mL)	Microcapillary BMD	Endpoint Microplate BMD ($5X10^5$CFU/mL)	Microcapillary BMD
$4X10^7$		>512 (R)		> 64 (R)		>1 (R)
$4X10^6$		8 (S)		16 (S)		1 (R)
$4X10^5$	4 (S)	4 (S)	8 (S)	16 (S)	0.25 (S)	0.25 (S)
$4X10^4$		≤2 (S)		16 (S)		0.25 (S)
$4X10^3$		≤2 (S)		16 (S)		0.25 (S)
	E. coli 25922					
$2X10^7$		64 (S)		>64 (R)		1 (R)
$2X10^6$		4 (S)		>64 (R)		0.06 (S)
$2X10^5$	4 (S)	4 (S)	16 (S)	32 (R)	0.015 (S)	0.015 (S)
$2X10^4$		4 (S)		16 (S)		0.03 (S)
$2X10^3$		4 (S)		16 (S)		0.015 (S)

Table 5. MIC determination based on delayed growth in the presence of antibiotics

		UPEC 2165					
		Nitrofurantoin		Cefalexin		Ciprofloxacin	
CFU/mL	Growth Control - Time to resazurin conversion (h)	Endpoint Microplate BMD ($5X10^5$CFU/mL)	Kinetic MIC	Endpoint Microplate BMD ($5X10^5$CFU/mL)	Kinetic MIC	Endpoint Microplate BMD ($5X10^5$CFU/mL)	Kinetic MIC
$4X10^7$	1		8 (S)		> 64 (R)		>1 (R)
$4X10^6$	2.8		≤2 (S)		16 (S)		1 (R)
$4X10^5$	4.4	4 (S)	≤2 (S)	8 (S)	16 (S)	0.25 (S)	0.25 (S)
$4X10^4$	5.8		≤2 (S)		16 (S)		0.25 (S)
$4X10^3$	7.7		≤2 (S)		8 (S)		0.125 (S)
		E. coli 25922					
$2X10^7$	1.5		≤2 (S)		>64 (R)		0.25 (S)
$2X10^6$	3.1		≤2 (S)		>64 (R)		0.015 (S)
$2X10^5$	4.9	4 (S)	≤2 (S)	16 (S)	8 (S)	0.015 (S)	0.015 (S)
$2X10^4$	5.6		≤2 (S)		8 (S)		0.007 (S)
$2X10^3$	7.7		≤2 (S)		4 (S)		0.007 (S)

Figures

504 **Figure 1.**

507 **Figure 2.**

510 **Figure 3.**

Figure 4.

Figure Legends

Figure 1. Microcapillary test use. (a) An array of up to 12 test strips were clipped into a 'ladder' holder with a 9mm pitch. Each test strip contains 10 capillaries each loaded to release doubling dilutions of antibiotic. Test strip arrays were dipped directly into 96-well microtitre plate wells allowing 1 microlitre of the sample to be drawn up by capillary action into all 10 microcapillaries. Plastic end covers filled with silicone grease were slid over the ends of the test strips, sealing the capillary ends to prevent evaporation. Image created using servier medical art **(b)** Bacteria growth and antibiotic inhibition is determined by resazurin color change. Blue to pink/white indicates growth and if antibiotic is present, indicates that concentration is below the minimum inhibitory concentration (MIC). At or above MIC, the capillary remains blue. **(c)** Each ladder of microcapillary BMD test strips provides up to 16X higher throughput than microtitre plates. Image illustrates the number of plates vs microcapillaries required to test 22 isolates in duplicate for a single antibiotic at 9 concentrations plus no antibiotic.

Figure 2. Agreement between microwell broth microdilution and microcapillary film MIC. Representative images showing MIC determination of two reference strains, *E. coli* 25922 and *E. coli* 13352 and urinary pathogen *E. coli* isolate 2165 MIC in microplate and microcapillary film. Error bar on microplates indicate 10 mm. Error bar on microcapillary film indicates 2 mm. Similar agreement was seen with all isolates tested in duplicate.

Figure 3. Microcapillary BMD with a fixed inoculum density growth kinetics in the presence of antibiotics. (a) UPEC 2165 isolate gives the same MIC result after 6h incubation compared to overnight. **(b)** UPEC 2158 recorded MIC is within 1 doubling dilution of antibiotic for ciprofloxacin. **(c)** Growth of UPEC 2164 is identical at all antibiotic concentrations of amoxicillin. **(d)** Nitrofurantoin shows the greatest amount of delayed growth. UPEC 2161 only observes a positive growth control at 6h, indicated an MIC ≤2 mg/L compared to an endpoint MIC of 8 mg/L.

Figure 4. Endpoint MIC measurement of isolates using range of inoculum density. (a) Schematic showing antibiotic cocnetration is highest in capilary 10. Capillary 1 contains no antibtioc and acts as a growth control. Representative images of capillary test strips at various incoculum density ranging from $10^3 - 10^7$ CFU/mL. **(b)** MIC observed at different inoculum density. Data indicates average of duplicate capillaries for two independent experiments. Vertical error bars indicate ± SEM and horizontal error bars indicate ± SD for inoclum density determiend by overnight agar plating. **(c)** Growth kinetics of UPEC 2165 against nitrofurantoin. Dotted line indicates detection time for the growth control, solid line indicates 1 h cuttoff from growth control. Data points indicate average of duplicate capillaries.

Author Contributions

SHN: Conceptualisation, Data curation, Formal analysis, Investigation, Methodology, Project administration, Software, Visualization, Writing – original draft, Writing – review & editing. WI: Resources, Writing – review & editing. ZR: Resources, Writing – review & editing .SA: Resources, Writing – review & editing. PR: Writing – review & editing. ADE: Conceptualisation, Funding acquisition, Methodology, Project administration, Supervision, Writing – original draft, Writing – review & editing

Acknowledgements

Figure 1a was adapted from elements from Servier medical art under CC BY 3.0 licence.

Declaration of Competing Interests

ADE is one of the inventors of patent application protecting aspects of the novel microfluidic devices tested in this study and is a director and shareholder in Capillary Film Technology Ltd, a company holding a commercial license to this patent application: WO2016012778 "Capillary assay device with internal hydrophilic coating" AD Edwards, NM Reis.

Funding

572 This research was supported by the EPSRC (EP/R022410/1) and Higher Education

573 Commission, Pakistan (Pin Number: 2BM2-093) and Commonwealth Scholarship

574 Commission (CSC Ref# PKCN-2017-215).

Ethical Approval

576 The collection of urinary pathogenic *E. coli* from a tertiary care hospital of Pakistan from

577 community acquired UTI patients and was approved by Ethical Review Board (ERB) of

578 Pakistan Institute of Medical Sciences.

References

ALI, I., RAFAQUE, Z., AHMED, I., TARIQ, F., GRAHAM, S. E., SALZMAN, E., FOXMAN, B. & DASTI, J. I.
2019. Phylogeny, sequence-typing and virulence profile of uropathogenic Escherichia coli
(UPEC) strains from Pakistan. *BMC Infectious Diseases,* 19**,** 620.
ANDREWS, J. M. 2001. BSAC standardized disc susceptibility testing method. *J Antimicrob*
Chemother, 48 Suppl 1**,** 43-57.
ANDREWS, J. M. 2005. BSAC standardized disc susceptibility testing method (version 4). *J Antimicrob*
Chemother, 56**,** 60-76.
AVESAR, J., ROSENFELD, D., TRUMAN-ROSENTSVIT, M., BEN-ARYE, T., GEFFEN, Y., BERCOVICI, M. &
LEVENBERG, S. 2017. Rapid phenotypic antimicrobial susceptibility testing using nanoliter
arrays. *Proceedings of the National Academy of Sciences,* 114**,** E5787.
AZIZI, M., ZAFERANI, M., DOGAN, B., ZHANG, S., SIMPSON, K. W. & ABBASPOURRAD, A. 2018.
Nanoliter-Sized Microchamber/Microarray Microfluidic Platform for Antibiotic Susceptibility
Testing. *Anal Chem,* 90**,** 14137-14144.
BALTEKIN, Ö., BOUCHARIN, A., TANO, E., ANDERSSON, D. I. & ELF, J. 2017. Antibiotic susceptibility
testing in less than 30 min using direct single-cell imaging. *Proceedings of the National*
Academy of Sciences, 114**,** 9170.
BARABAN, L., BERTHOLLE, F., SALVERDA, M. L., BREMOND, N., PANIZZA, P., BAUDRY, J., DE VISSER, J.
A. & BIBETTE, J. 2011. Millifluidic droplet analyser for microbiology. *Lab Chip,* 11**,** 4057-62.
BARBOSA, A. I., CASTANHEIRA, A. P., EDWARDS, A. D. & REIS, N. M. 2014. A lab-in-a-briefcase for
rapid prostate specific antigen (PSA) screening from whole blood. *Lab on a Chip,* 14**,** 2918-
2928.
BARBOSA, A. I., GEHLOT, P., SIDAPRA, K., EDWARDS, A. D. & REIS, N. M. 2015. Portable smartphone
quantitation of prostate specific antigen (PSA) in a fluoropolymer microfluidic device.
Biosens Bioelectron, 70**,** 5-14.
BESANT, J. D., SARGENT, E. H. & KELLEY, S. O. 2015. Rapid electrochemical phenotypic profiling of
antibiotic-resistant bacteria. *Lab on a Chip,* 15**,** 2799-2807.
BRYCE, A., HAY, A. D., LANE, I. F., THORNTON, H. V., WOOTTON, M. & COSTELLOE, C. 2016. Global
prevalence of antibiotic resistance in paediatric urinary tract infections caused by
Escherichia coli and association with routine use of antibiotics in
primary care: systematic review and meta-analysis. *BMJ,* 352**,** i939.
BUSCHE, J. F., MOLLER, S., STEHR, M. & DIETZEL, A. 2019. Cross-Flow Filtration of Escherichia coli at a
Nanofluidic Gap for Fast Immobilization and Antibiotic Susceptibility Testing. *Micromachines*
(Basel), 10.

CASTANHEIRA, A. P., BARBOSA, A. I., EDWARDS, A. D. & REIS, N. M. 2015. Multiplexed femtomolar
quantitation of human cytokines in a fluoropolymer microcapillary film. *Analyst,* 140**,** 5609-
5618.
CHANDRASEKARAN, S., ABBOTT, A., CAMPEAU, S., ZIMMER, B. L., WEINSTEIN, M., THRUPP, L.,
HEJNA, J., WALKER, L., AMMANN, T., KIRN, T., PATEL, R. & HUMPHRIES, R. M. 2018. Direct-
from-Blood-Culture Disk Diffusion To Determine Antimicrobial Susceptibility of Gram-
Negative Bacteria: Preliminary Report from the Clinical and Laboratory Standards Institute
Methods Development and Standardization Working Group. *Journal of Clinical Microbiology,*
56**,** e01678-17.
CHOI, J., JUNG, Y.-G., KIM, J., KIM, S., JUNG, Y., NA, H. & KWON, S. 2013. Rapid antibiotic
susceptibility testing by tracking single cell growth in a microfluidic agarose channel system.
Lab on a Chip, 13**,** 280-287.
CHOI, J., YOO, J., LEE, M., KIM, E.-G., LEE, J. S., LEE, S., JOO, S., SONG, S. H., KIM, E.-C., LEE, J. C., KIM,
H. C., JUNG, Y.-G. & KWON, S. 2014. A rapid antimicrobial susceptibility test based on single-
cell morphological analysis. *Science Translational Medicine,* 6**,** 267ra174.
CIRA, N. J., HO, J. Y., DUECK, M. E. & WEIBEL, D. B. 2012. A self-loading microfluidic device for
determining the minimum inhibitory concentration of antibiotics. *Lab on a Chip,* 12**,** 1052-
1059.
DÖNMEZ, S. İ., NEEDS, S. H., OSBORN, H. M. I. & EDWARDS, A. D. 2020. Label-free smartphone
quantitation of bacteria by darkfield imaging of light scattering in fluoropolymer micro
capillary film allows portable detection of bacteriophage lysis. *Sensors and Actuators B:*
Chemical, 323**,** 128645.
EDWARDS, A. D., REIS, N. M., SLATER, N. K. & MACKLEY, M. R. 2011. A simple device for multiplex
ELISA made from melt-extruded plastic microcapillary film. *Lab Chip,* 11**,** 4267-73.
ELSHIKH, M., AHMED, S., FUNSTON, S., DUNLOP, P., MCGAW, M., MARCHANT, R. & BANAT, I. M.
2016. Resazurin-based 96-well plate microdilution method for the determination of
minimum inhibitory concentration of biosurfactants. *Biotechnology letters,* 38**,** 1015-1019.
FLORIO, W., BALDESCHI, L., RIZZATO, C., TAVANTI, A., GHELARDI, E. & LUPETTI, A. 2020. Detection of
Antibiotic-Resistance by MALDI-TOF Mass Spectrometry: An Expanding Area. *Frontiers in*
Cellular and Infection Microbiology, 10**,** 616.
FRÖDING, I., VONDRACEK, M. & GISKE, C. G. 2017. Rapid EUCAST disc diffusion testing of MDR
Escherichia coli and Klebsiella pneumoniae: inhibition zones for extended-spectrum
cephalosporins can be reliably read after 6 h of incubation. *Journal of Antimicrobial*
Chemotherapy, 72**,** 1094-1102.
GRIGORYAN, L., TRAUTNER, B. W. & GUPTA, K. 2014. Diagnosis and management of urinary tract
infections in the outpatient setting: a review. *Jama,* 312**,** 1677-84.
GUET-REVILLET, H., EMIRIAN, A., GROH, M., NEBBAD-LECHANI, B., WEISS, E., JOIN-LAMBERT, O.,
BILLE, E., JULLIEN, V. & ZAHAR, J. R. 2014. Pharmacological Study of Cefoxitin as an
Alternative Antibiotic Therapy to Carbapenems in Treatment of Urinary Tract Infections Due
to Extended-Spectrum-β-Lactamase-Producing <span class="named-content genus-
species" id="named-content-1">Escherichia coli.
Antimicrobial Agents and Chemotherapy, 58**,** 4899.
HE, P. J. W., KATIS, I. N., KUMAR, A. J. U., BRYANT, C. A., KEEVIL, C. W., SOMANI, B. K., MAHOBIA, N.,
EASON, R. W. & SONES, C. L. 2020. Laser-patterned paper-based sensors for rapid point-of-
care detection and antibiotic-resistance testing of bacterial infections. *Biosensors and*
Bioelectronics, 152**,** 112008.
HOMBACH, M., JETTER, M., KELLER, P. M., BLÖCHLIGER, N., KOLESNIK-GOLDMANN, N. & BÖTTGER,
E. C. 2017. Rapid detection of ESBL, carbapenemases, MRSA and other important resistance
phenotypes within 6–8 h by automated disc diffusion antibiotic susceptibility testing. *Journal*
of Antimicrobial Chemotherapy, 72**,** 3063-3069.

JÉGOUIC, S., JONES, I. & EDWARDS, A. 2021. Affordable mobile microfluidic diagnostics: minimum
requirements for smartphones and digital imaging for colorimetric and fluorometric anti-
dengue and anti-SARS-CoV-2 antibody detection [version 1; peer review: awaiting peer
review]. *Wellcome Open Research,* 6.
JIANG, L., BOITARD, L., BROYER, P., CHAREIRE, A. C., BOURNE-BRANCHU, P., MAHÉ, P., TOURNOUD,
M., FRANCESCHI, C., ZAMBARDI, G., BAUDRY, J. & BIBETTE, J. 2016. Digital antimicrobial
susceptibility testing using the MilliDrop technology. *European Journal of Clinical*
Microbiology & Infectious Diseases, 35**,** 415-422.
JORGENSEN, J. H. & TURNIDGE, J. D. 2015. Susceptibility Test Methods: Dilution and Disk Diffusion
Methods. *Manual of Clinical Microbiology***,** 1253-1273.
KANG, W., SARKAR, S., LIN, Z. S., MCKENNEY, S. & KONRY, T. 2019. Ultrafast Parallelized Microfluidic
Platform for Antimicrobial Susceptibility Testing of Gram Positive and Negative Bacteria.
Anal Chem, 91**,** 6242-6249.
KAPROU, G. D., BERGŠPICA, I., ALEXA, E. A., ALVAREZ-ORDÓÑEZ, A. & PRIETO, M. 2021. Rapid
Methods for Antimicrobial Resistance Diagnostics. *Antibiotics (Basel),* 10.
KIM, H. J. & JANG, S. 2018. Optimization of a resazurin-based microplate assay for large-scale
compound screenings against Klebsiella pneumoniae. *3 Biotech,* 8**,** 3.
KIM, S., MASUM, F., KIM, J.-K., CHUNG, H. J. & JEON, J. S. 2019. On-chip phenotypic investigation of
combinatory antibiotic effects by generating orthogonal concentration gradients. *Lab on a*
Chip, 19**,** 959-973.
KIM, S. C., CESTELLOS-BLANCO, S., INOUE, K. & ZARE, R. N. 2015. Miniaturized Antimicrobial
Susceptibility Test by Combining Concentration Gradient Generation and Rapid Cell
Culturing. *Antibiotics (Basel, Switzerland),* 4**,** 455-466.
KINN, P. M., PERCIVAL, K. M., FORD, B. A. & INCE, D. 2019. 1997. Real-World Impact of Accelerate
Pheno Implementation with Antimicrobial Stewardship Intervention. *Open Forum Infectious*
Diseases, 6**,** S669-S670.
LAVIGNE, J. P., ESPINAL, P., DUNYACH-REMY, C., MESSAD, N., PANTEL, A. & SOTTO, A. 2013. Mass
spectrometry: a revolution in clinical microbiology? *Clin Chem Lab Med,* 51**,** 257-70.
LI, B., QIU, Y., GLIDLE, A., MCILVENNA, D., LUO, Q., COOPER, J., SHI, H.-C. & YIN, H. 2014. Gradient
Microfluidics Enables Rapid Bacterial Growth Inhibition Testing. *Analytical Chemistry,* 86**,**
3131-3137.
MAMBIE, A., VUOTTO, F., POITRENAUD, D., WEYRICH, P., CANNESSON, O., DESSEIN, R., FAURE, K.,
GUERY, B. & GALPÉRINE, T. 2016. Cefoxitin: An alternative to carbapenems in urinary tract
infections due to extended-spectrum beta-lactamase-producing Enterobacteriaceae. *Med*
Mal Infect, 46**,** 215-9.
MANCINI, S., RÖTHLIN, K., BODENDOERFER, E., HERREN, S., KOLESNIK-GOLDMANN, N., COURVALIN,
P., ZBINDEN, R. & BÖTTGER, E. C. 2020. Tentative breakpoints and areas of technical
uncertainty for early reading automated disc diffusion for Enterobacterales. *Journal of*
Antimicrobial Chemotherapy, 75**,** 1495-1505.
MATSUMOTO, Y., SAKAKIHARA, S., GRUSHNIKOV, A., KIKUCHI, K., NOJI, H., YAMAGUCHI, A., IINO, R.,
YAGI, Y. & NISHINO, K. 2016. A Microfluidic Channel Method for Rapid Drug-Susceptibility
Testing of Pseudomonas aeruginosa. *PLOS ONE,* 11**,** e0148797.
MEDINA, M. & CASTILLO-PINO, E. 2019. An introduction to the epidemiology and burden of urinary
tract infections. *Therapeutic advances in urology,* 11**,** 1756287219832172-
1756287219832172.
NEEDS, S. H., DIEP, T. T., BULL, S. P., LINDLEY-DECAIRE, A., RAY, P. & EDWARDS, A. D. 2019. Exploiting
open source 3D printer architecture for laboratory robotics to automate high-throughput
time-lapse imaging for analytical microbiology. *PLOS ONE,* 14**,** e0224878.
NEEDS, S. H., DONMEZ, S. I., BULL, S. P., MCQUAID, C., OSBORN, H. M. I. & EDWARDS, A. D. 2020.
Challenges in Microfluidic and Point-of-Care Phenotypic Antimicrobial Resistance Tests.
Frontiers in Mechanical Engineering, 6**,** 73.

NEEDS, S. H., OSBORN, H. M. I. & EDWARDS, A. D. 2021. Counting bacteria in microfluidic devices:
Smartphone compatible 'dip-and-test' viable cell quantitation using resazurin amplified
detection in microliter capillary arrays. *Journal of Microbiological Methods*, 106199.
O'GRADY, M. C., BARRY, L., CORCORAN, G. D., HOOTON, C., SLEATOR, R. D. & LUCEY, B. 2019.
Empirical treatment of urinary tract infections: how rational are our guidelines? *Journal of*
Antimicrobial Chemotherapy, 74, 214-217.
OSAID, M., CHEN, Y.-S., WANG, C.-H., SINHA, A., LEE, W.-B., GOPINATHAN, P., WU, H.-B. & LEE, G.-B.
2021. A multiplexed nanoliter array-based microfluidic platform for quick, automatic
antimicrobial susceptibility testing. *Lab on a Chip*.
PIVETAL, J., PEREIRA, F., BARBOSA, A. I., CASTANHEIRA, A. P., REIS, N. M. & EDWARDS, A. D. 2017.
Covalent immobilisation of antibodies in Teflon-FEP microfluidic devices for sensitive
quantification of clinically relevant protein biomarkers. *Analyst*.
POKHAREL, S., RAUT, S. & ADHIKARI, B. 2019. Tackling antimicrobial resistance in low-income and
middle-income countries. *BMJ Global Health,* 4, e002104.
PULIDO, M. R., GARCÍA-QUINTANILLA, M., MARTÍN-PEÑA, R., CISNEROS, J. M. & MCCONNELL, M. J.
2013. Progress on the development of rapid methods for antimicrobial susceptibility testing.
Journal of Antimicrobial Chemotherapy, 68, 2710-2717.
RAFAQUE, Z., DASTI, J. I. & ANDREWS, S. C. 2018. Draft genome sequence of a multidrug-resistant
CTX-M-15 β-lactamase-producing uropathogenic Escherichia coli isolate (ST131-O25b-H30)
from Pakistan exhibiting high potential virulence. *Journal of Global Antimicrobial Resistance,*
15, 164-165.
RAFAQUE, Z., DASTI, J. I. & ANDREWS, S. C. 2019. Draft genome sequence of a uropathogenic
Escherichia coli isolate (ST38 O1:H15) from Pakistan, an emerging multidrug-resistant
sequence type with a high virulence profile. *New Microbes and New Infections,* 27, 1-2.
RAKHMAWATIE, M. D., WIBAWA, T., LISDIYANTI, P., PRATIWI, W. R. & MUSTOFA 2019. Evaluation of
crystal violet decolorization assay and resazurin microplate assay for antimycobacterial
screening. *Heliyon,* 5, e02263.
REIS, N. M., PIVETAL, J., LOO-ZAZUETA, A. L., BARROS, J. M. S. & EDWARDS, A. D. 2016. Lab on a
stick: multi-analyte cellular assays in a microfluidic dipstick. *Lab on a Chip,* 16, 2891-2899.
SARKER, S. D., NAHAR, L. & KUMARASAMY, Y. 2007. Microtitre plate-based antibacterial assay
incorporating resazurin as an indicator of cell growth, and its application in the in vitro
antibacterial screening of phytochemicals. *Methods (San Diego, Calif.),* 42, 321-324.
SENARD, O., BOUCHAND, F., DECONINCK, L., MATT, M., FELLOUS, L., ROTTMAN, M., PERRONNE, C.,
DINH, A. & DAVIDO, B. 2018. Efficacy of cefoxitin for the treatment of urinary tract infection
due to extended-spectrum-beta-lactamase-producing Escherichia coli and Klebsiella
pneumoniae isolates. *Therapeutic advances in infectious disease,* 6, 2049936118811053-
2049936118811053.
SMITH, K. P. & KIRBY, J. E. 2018. The Inoculum Effect in the Era of Multidrug Resistance: Minor
Differences in Inoculum Have Dramatic Effect on MIC Determination. *Antimicrobial Agents*
and Chemotherapy, 62, e00433-18.
SUN, H., CHAN, C. W., WANG, Y., YAO, X., MU, X., LU, X., ZHOU, J., CAI, Z. & REN, K. 2019. Reliable
and reusable whole polypropylene plastic microfluidic devices for a rapid, low-cost
antimicrobial susceptibility test. *Lab Chip,* 19, 2915-2924.
TEH, C. H., NAZNI, W. A., NURULHUSNA, A. H., NORAZAH, A. & LEE, H. L. 2017. Determination of
antibacterial activity and minimum inhibitory concentration of larval extract of fly via
resazurin-based turbidometric assay. *BMC Microbiology,* 17, 36.
TREVAS, D., CALIENDO, A. M., HANSON, K., LEVY, J., GINOCCHIO, C. C. & INFECTIOUS DISEASES
SOCIETY OF, A. 2020. Diagnostic Tests Can Stem the Threat of Antimicrobial Resistance:
Infectious Disease Professionals Can Help. *Clinical Infectious Diseases*.

UZARSKI, J. S., DIVITO, M. D., WERTHEIM, J. A. & MILLER, W. M. 2017. Essential design
considerations for the resazurin reduction assay to noninvasively quantify cell expansion
within perfused extracellular matrix scaffolds. *Biomaterials,* 129**,** 163-175.
WATTS, V., BROWN, B., AHMED, M., CHARLETT, A., CHEW-GRAHAM, C., CLEARY, P., DECRAENE, V.,
DODGSON, K., GEORGE, R., HOPKINS, S., ESMAIL, A. & WELFARE, W. 2020. Routine
laboratory surveillance of antimicrobial resistance in community-acquired urinary tract
infections adequately informs prescribing policy in England. *JAC-Antimicrobial Resistance,* 2.
WISTRAND-YUEN, P., MALMBERG, C., FATSIS-KAVALOPOULOS, N., LÜBKE, M., TÄNGDÉN, T. &
KREUGER, J. 2020. A Multiplex Fluidic Chip for Rapid Phenotypic Antibiotic Susceptibility
Testing. *mBio,* 11**,** e03109-19.
WU, F. & DEKKER, C. 2016. Nanofabricated structures and microfluidic devices for bacteria: from
techniques to biology. *Chem Soc Rev,* 45**,** 268-80.