Colocalized, Bidirectional Optogenetic Modulations in Freely Behaving Animals with a Wireless Dual-Color Optoelectronic Probe

Lizhu Li,[1, #] Lihui Lu,[2, 3, #] Yuqi Ren,[2, 3] Guo Tang,[1] Yu Zhao,[1] Xue Cai,[1] Zhao Shi,[1] He Ding,[4] Changbo Liu,[5] Dali Cheng,[1] Yang Xie,[1] Huachun Wang,[1] Xin Fu,[6] Lan Yin,[6] Minmin Luo,[2, 3, 7, *] Xing Sheng [1, 7, *]

Affiliations

[1]Department of Electronic Engineering, Beijing National Research Center for Information Science and Technology, Center for Flexible Electronics Technology, Tsinghua University, Beijing, China

[2]Chinese Institute for Brain Research, Beijing, China

[3]National Institute of Biological Sciences and School of Life Sciences, Tsinghua University, Beijing, China

[4]Beijing Engineering Research Center of Mixed Reality and Advanced Display, School of Optics and Photonics, Beijing Institute of Technology, Beijing, China

[5]School of Materials Science and Engineering and Hangzhou Innovation Institute, Beihang University, Beijing, China

[6]School of Materials Science and Engineering, Tsinghua University, Beijing, China

[7]IDG/McGovern Institute for Brain Research, Tsinghua University, Beijing, China

# L. Li and L. Lu contributed equally to this work.

*Correspondence to: luominmin@cibr.ac.cn; xingsheng@tsinghua.edu.cn

ABSTRACT

The precise control of neural activities at both cellular and circuit levels reveals significant impacts on the fundamental neuroscience explorations and medical applications. Optogenetic methods provide efficient cell-specific modulations, but state-of-the-art technologies lack the ability of simultaneous neural activation and inhibition in a same brain region of freely moving animals. Here we report bidirectional neuronal activity manipulation accomplished by a wireless, dual-color optogenetic probe in synergy with the co-expression of two spectrally distinct opsins (ChrimsonR and stGtACR2) in a rodent model. Based on vertically assembled, thin-film microscale light-emitting diodes (micro-LEDs) on flexible substrates, the dual-color probe shows colocalized red and blue emissions and allows chronic *in vivo* operations with desirable biocompatibilities. In addition, we discover that neurons co-expressing the two opsins can be deterministically evoked or silenced under red or blue irradiations. Implanted in behaving mice, the wirelessly controlled dual-color probe interferes with dopaminergic neurons in the ventral tegmental area (VTA), increasing or decreasing dopamine levels with colocalized red and blue stimulations. Such bidirectional regulations further generate rewarding and aversive behaviors of freely moving animals in a place preference test. The technologies established here will create numerous opportunities and profound implications for the brain research.

INTRODUCTION

Understanding the brain functions and treating neurological disorders rely on the continuous development of advanced technologies to interrogate with the complex nervous systems[1, 2]. Over the past decades, optogenetic methods have been emerging as a powerful toolset for effective and precise neural manipulation, owing to their capability of specific cell targeting[3, 4]. Until now, a range of genetically encoded, light-sensitive proteins (opsins) with distinct spectral responses have been engineered to regulate flows of different cations and anions[5–16], triggering various biological responses, from cells and *C. elegans* to rodents and non-human primates[17–20]. Accompanying the development of opsins, advanced technologies for light delivery into the brain have also been progressing rapidly. Apart from various fiber-based emitters interfacing with external light sources[21–23], recently developed implantable probes based on thin-film, microscale optoelectronic devices have offered a viable solution to versatile neural modulations in untethered animals, when incorporated with wirelessly operated systems[24–26]. In addition to these achievements, bidirectional neural modulations, for example, by activating or suppressing the same or different populations of cells in behaving animals with light at different wavelengths, are still highly demanded. Such capabilities would enable more precise control of neural activities to advance both neuroscience explorations and disease medications[27]. Certain efforts have been attempted, including explorations based on experiments by expressing excitatory and inhibitory opsins in separate animals[28, 29] or distinct cells in different neural regions[17, 30]. There are few reports on dual-color optogenetic activation and inhibition by co-expressing different opsins in the same cells[31–33], but the efficacy and utility of such bidirectional modulations could be limited by the potential spectral overlay of the two opsins and the lack of miniaturized light delivery systems. Therefore, we envision that wirelessly operated light sources with polychromatic

emissions, which collaborate with co-expressed opsins, can be implemented for untethered

use in freely moving animals and enable previously inaccessible applications.

This study presents synergic optoelectronic and biological strategies to overcome

challenges of existing techniques, and realize colocalized, bidirectional neural manipulations

in untethered behaving animals. With heterogeneously integrated thin-film, microscale light-

emitting diodes (micro-LEDs), a wirelessly operated, implantable dual-color probe allows

independently controlled red and blue emissions in the same region. We also discover that the

co-expression of two channelrhodopsins (ChrimsonR and stGtACR2) in the same group of

neurons enables efficient light-evoked cell depolarization and hyperpolarization with

controlled cation and anion flows under red and blue illuminations, respectively. With the

wireless, dual-color micro-LED probes implanted into the ventral tegmental area (VTA) of

freely behaving mice, preference and aversion tests clearly demonstrate the bidirectional

optogenetic excitation and inhibition capability. Furthermore, *in vivo* behavioral results

associated with the red and blue stimulations in the VTA are validated by increased or

decreased dopamine levels detected in the nucleus accumbens (NAc) in real time.

RESULTS

Wirelessly operated, dual-color micro-LED probes and circuit systems

We design and implement a wireless, dual-color micro-LED probe that generates red

and blue emissions in the same brain region for bidirectional optogenetic modulations *in vivo*,

as schematically illustrated in Figure 1a. Specifically, with the co-expression of spectrally

distinct ion channels, ChrimsonR[7] and stGtACR2[9], in the same population of cells, red and

blue light regulate the inflow of cations and anions, causing cell depolarization and

hyperpolarization, respectively (Figure 1b). Figure 1c displays an exploded view of the dual-

color probe, with design and fabrication details provided in the supplemental information and

Figure S1. From bottom to top, the probe structure comprises a copper (Cu) coated polyimide (PI) thin-film substrate, an indium gallium phosphide (InGaP) red LED, a silicon oxide (SiO_2) / titanium oxide (TiO_2) based dielectric filter, an indium gallium nitride (InGaN) blue LED. All the devices (red LEDs, blue LEDs and filters) are formed in thin-film, microscale formats (lateral dimension ~ 125 μm × 180 μm, thickness ~ 5–7 μm) through epitaxial liftoff, and vertically assembled via transfer printing[34–36] (Figure 1d). The intrinsic optical transparency of active layers in the InGaN blue micro-LED allows emissions from the red micro-LED passing through with minimal losses. Additionally, the designed filter selectively reflects blue light and transmits red light, enhancing the emissive efficiencies of both micro-LEDs.

Flexible PI substrates with vertically stacked micro-LEDs are laser milled and encapsulated with a waterproof coating, forming a needle shaped probe with a dimension (thickness ~120 μm and width ~320 μm) similar to a conventional silica fiber (diameter ~220 μm) used in optogenetic stimulations. Separate metallization and electrical insulation ensure that red and blue micro-LEDs can be independently lighted up and display spectrally varied illuminations (red, blue or combined) in the same location (Figure 1e and Movie S1). Figure 1f plots measured electroluminescence (EL) spectra. The red and blue micro-LEDs exhibit emission peaks at 630 nm and 480 nm, which are in good accordance with the excitation spectra of ChrimsonR and stGtACR2 (data extracted from references[7, 8]), respectively. Moreover, the spectral separation between the two opsins, as well as their different operational modes (pulsed red light activation for ChrimsonR, and continuous blue light inhibition for stGtACR2), guarantee precise optogenetic activation and inhibition with minimal optical interferences. More optoelectronic properties of the dual-color micro-LED probe, including current-voltage characteristics, quantum efficiencies, power irradiance and angular-dependent emissive profiles are presented in Figure S2. Under injection currents of

10 mA, the red and the blue micro-LEDs can reach irradiances up to 50 mW/mm^2 and 200

mW/mm^2 on the device surface, which are sufficient for optogenetic modulations.

To illustrate the optical performance of the devices, we insert a dual-color probe into

a brain phantom (Figure 1g, top). For comparison, established models calculate optical

profiles based on the optical properties (absorption and scattering) of brain tissues (Figure 1g,

bottom). Both experimental and simulation results reveal that vertically stacked dual-color

micro-LEDs show colocalized red and blue emission profiles, ensuring effective activation

and inhibition in the same brain region. Power irradiances for blue and red emissions

decrease to 10% of the original power at propagating distances of ~400 μm and ~800 μm,

respectively. Such a unique design of vertical assembly holds prominent advantages

compared to a probe with laterally arranged red and blue micro-LEDs displaying highly

misaligned emissions, which could severely limit its capability for bidirectional modulations.

Thermal behaviors of the micro-LED probe are also systematically analyzed, with

detailed results presented in the supplemental material and Figure S3. Probes are inserted into

the brain phantom and the living mouse brain, with temperature rises measured by an infrared

camera and a thermocouple, respectively. Both experiments obtain similar results and are in

agreement with the thermal models established based on finite-element analysis. Under a

pulsed operation (20-Hz frequency, 10-ms pulse duration, current 0–10 mA), the tissue

temperature rises caused by the red micro-LED can be restricted within 1 °C. The operation

of the blue micro-LED induces a higher thermal effect, because practical neural inhibition for

stGtACR2 requires continuous current injection; nonetheless, the tissue heating can be

controlled within 2 °C by limiting the current within 5 mA. Such operation modes for red and

blue micro-LEDs will be employed throughout *in vivo* experiments, to mitigate undesired

neural responses and possible tissue damage due to overheating.

We also design a customized, miniaturized flexible circuit module wirelessly operate

the dual-color micro-LED probe. Controlled by radio frequency (RF) communication at 2.4

GHz, the battery-powered circuit (weight ~1.9 g) can independently address the red and the

blue micro-LEDs, by adjusting their pulse frequencies, durations and injection currents for

versatile neural modulations *in vivo* (Figure 1h, and details in the supplemental material and

Figures S4–S7). Designed programming commends also allow the independent operation of

multiple micro-LED probes in synchronized or non-synchronized modes (Movie S2). After

subcranially implanting the micro-LED probe into a behaving mouse, the light-weighted

circuit mounted onto the mouse head provides the capability for optogenetic control in an

untethered manner (Figure 1i).

Bidirectional optogenetic manipulations of neurons *in vitro*, with dual-color stimulations

We first evaluate the feasibility of dual-color, bidirectional optogenetic modulations for

neurons co-expressing ChrimsonR and stGtACR2 in brain slices with electrophysiological

tests.   Mixed   vectors   of   AAV-CAG-DIO-ChrimsonR-mCherry   and   AAV-EF1a-DIO-

stGtACR2-EGFP are injected into the ventral tegmental area (VTA) of *DAT-Cre* mice

(Figure 2a). The deployment of the soma-targeted variant stGtACR2 exhibits performance

superior to previously reported anion channel GtACR2, by enhancing the membrane

trafficking and reducing the axonal excitation[8, 9]. Figure 2b shows the fluorescence images of

stained cells 3 weeks after viral injection in the VTA, including 4',6-diamidino-2-

phenylindole (DAPI, blue), stGtACR2 (green) and ChrimsonR (red). The merged graph and

zoomed in views (Figure 2c) clearly reveal the desirable co-expression of the two opsins in

target cells. Statistical results show that ~30% of cells are labeled with stGtACR2 or

ChrimsonR, and > 83% of cells co-express both opsins (Figure 2d).

Whole-cell patch recordings capture intracellular signals of these cells under optical

stimulations (Figure 2e). Other than using our designed micro-LED probe, here we employ a

collimated light spot with coupled red (633 nm) and blue (480 nm) laser sources for *in vitro*

experiments, to achieve uniform illuminations for better power calibrations. Figures 2f and 2g

validate that pulsed red stimulations (20-Hz frequency, 10-ms pulse duration) induce action

potentials (APs or spikes) by depolarizing these cells, and spike activation probability

increases at elevated red irradiances. In these same cells, continuous blue stimulations

suppress activities by hyperpolarization (Figures 2h–2j). Voltage clamp results illustrate that

blue light evokes photocurrent reversion when the membrane potential becomes more

positive (Figure 2h), and probability for spike inhibition also increases at elevated blue

irradiances (Figures 2i and 2j). Threshold irradiances for complete spike activation and

inhibition are around 10 mW/mm^2 and 1 mW/mm^2 for red and blue illuminations,

respectively. The results are consistent with previous reports on cells expressing only

ChrimsonR or stGtACR2[7, 9].

Notably, we do not only demonstrate that the co-expression of both opsins does not

affect the efficacy of cell activity excitation or suppression, but also accomplish bidirectional

activation and inhibition in the same cell by instantaneously altered red and blue irradiations

(Figure 2k). In accordance with activation spectra of two opsins (Figure 1f), red light does

not interact with stGtACR2, which exhibits nearly zero responses at wavelengths longer than

600 nm. On the other hand, ChrimsonR should have responded to blue light, which could

raise concern about possible optical crosstalk. Nevertheless, spike activations caused by blue

light do not occur in cells co-expressing ChrimsonR and stGtACR2, since cell suppressions

by turning on stGtACR2 channels arise and saturate at a much lower blue irradiance. This

fact can be confirmed by patching the cells under combined red and blue illuminations

(Figures 2l and 2m). Example traces in Figure 2l show that red-light-evoked spikes are

suppressed under simultaneous blue illuminations even with very low irradiances. More

traces are reported in Figure S8, and Figure 2m summarizes the spike probability (measured

and averaged among $n = 5$ cells) with combined red and blue illuminations. Introducing blue

irradiances orders of magnitude lower than red irradiances can effectively suppress cell

spikes and cause activity inhibition.

193 Bidirectional optogenetic modulation of preference and aversion behaviors *in vivo*, with

194 wireless dual-color micro-LED probes

Activities of the dopaminergic neurons in the VTA we manage to manipulate in Figure 2

significantly correlate with the reward and aversion based behaviors[37, 38, 39]. In Figure 3, we

further exploit the capability of bidirectional optogenetic manipulations by implanting dual-

color micro-LED probes into behaving animals. After the injection of AAV-CAG-DIO-

ChrimsonR-mCherry and AAV-EF1a-DIO-stGtACR2-EGFP vectors, the dual-color micro-

LED probe is implanted into the VTA of *DAT-cre* mice (Figure 3a). Figure 3b illustrates the

locations of the micro-LEDs in the VTA, with a slice showing the co-expression of

ChrimsonR and stGtACR2, combined with the luminescence of the red and the blue micro-

LEDs under corresponding optical excitations. To access the devices' biocompatibility *in*

vivo, histological examinations of brain slices show that inflammatory reactions of the micro-

LED probe are similar to a conventional silica fiber (diameter ~220 μm) 1 day or 3 weeks

post implantation (Figure S9), owing to its thin-film geometry, mechanical flexibility and

bio-friendly encapsulation. Furthermore, the parylene-based, waterproof encapsulation

maintains the micro-LEDs' chronic and stable operation for more than 200 days after

implantation (Figure S10).

In vivo experiments are performed within a real-time place preference / avoidance

(RTPP) paradigm. After 14 days of virus and micro-LED probe injection, we place mice in a

two-compartment apparatus and monitor their behaviors with a camera during optogenetic

modulations (Figure 3c and Figure S11). Figure 3d provides the test protocol, including a

baseline pretest (10 mins), red LED stimulation (pulsed wave, 30 mins), a rest phase (30

mins), and blue LED stimulation (continuous wave, 30 mins). Representative heat maps in

Figure 3e show the change of real-time behaviors during optogenetic stimulations, for mice

expressing different combinations of opsins. Summarized results in Figure 3f clearly show

that animals co-expressing ChimsonR and stGtACR2 present strong preference and

avoidance behaviors following red and blue stimulations, respectively.

By contrast, the control groups without photosensitive opsins (expressing only EGFP

and mCherry) show little difference of behaviors under optical stimulations. More results are

provided in Figures S12 and S13, comparing these animals with other group expressing only

ChimsonR or stGtACR2. When introducing only ChimsonR, animals exhibit place

preference under both red and blue stimulations. For the group with only stGtACR2, blue

light effectively evokes place aversion while red light generates little influence on animal

behaviors. Figures 3g and 3h present the dynamic behavior change at different time courses.

These results are consistent with the spectral characteristics of the opsins as well as the *in*

vitro neural responses at the cellular level (Figure 2), elucidating the utility of bidirectional

modulations with wireless, dual-color micro-LED probes. In conclusion, the implantable

wireless dual-color LED probe combining with two spectrally distinct channerhodopins can

realize bi-directional control the place preference or aversion in freely behaving mice.

Bidirectional optogenetic regulation of dopamine release *in vivo*, with wireless dual-

color micro-LED probes

The effective bidirectional behavior manipulations demonstrated in Figure 3 can be ascribed

to the light induced activation and inhibition of dopaminergic neurons in the VTA. We

further investigate the effect of optogenetic stimulations on the dopamine release using *in*

vivo experiments described in Figure 4a. Similarly, combined opsins are expressed and

followed by implanting the dual-color micro-LED probe in the VTA of *DAT-cre* mice. While

stimulations are performed in the VTA, dopamine signals are simultaneously monitored in

one of its main projection areas, the nucleus accumbens (NAc)[40, 41, 42]. A genetically encoded

indicator GRAB$_{DA2m}$[43] is expressed in the NAc, where the fluorescence signals are recorded

with a fiber photometer[28, 29, 39, 44, 45, 46]. The combination of micro-LED based optogenetic

stimulations and fiber based photometric recordings realizes all-optical interrogation of

neural circuits. During the surgery, we carefully select the insertion direction of the micro-

LED probe, in order to obtain desirable stimulation efficacy in the VTA and minimize the

optical crosstalk that could disturb photometric recording in the NAc (Figure S14).

Fluorescence images in Figure 4b show the expressions of GRAB$_{DA2m}$ in the NAc as well as

ChrimsonR and stGtACR2 in the VTA.

Figure 4c plots temporally resolved dopamine release in respond to stimulations

caused by red and blue micro-LEDs in a same mouse. More representative curves are

presented in Figure S15, with results summarized in Figure 4d. Illuminating the VTA by red

or blue micro-LEDs (with injection currents from 0 to 10 mA) clearly induces increased or

decreased dopamine signals in the NAc, which correlates with neuron responses at the

cellular level (Figure 2). Experiments operated among animal groups expressing different

opsins are also performed, with results compared in Figure 4e. In contrast to the group

expressing both opsins (ChrimsonR and stGtACR2), mice with only ChrimsonR show

increased dopamine release under both red and blue illuminations, and dopamine suppression

is observed under blue illumination for the mice with only stGtACR2. The control group

(with EGFP and mCherry) exhibits no response under optical stimulations. These results

evidently explain *in vivo* preference and avoidance behaviors presented in Figure 3,

showcasing the capability for precise control of neural activities based on bidirectional, dual-

color modulations.

We further investigate neural activities with coincident red and blue LED radiations.

Figure 4f plots example traces of dopamine transients, in response to the red and blue micro-

LEDs under pulse and continuous operations (6-s duration) with varied currents. Statistical

data for accumulative dopamine concentration variations (unit: area under the curve during

stimulation, or AUC) are summarized in Figure 4g and Figure S16. Decreasing power of the

red LED and increasing the power of the blue LED cause the decline of dopamine release.

Additionally, the blue LED current required to suppress the dopamine release is much lower

than that of the red LED required for dopamine level rises, which further confirms with the

cellular activities reported in Figure 2.

Figure 4h (and more examples in Figure S17) presents dopamine dynamics in

response to alternated or combined red and blue LED stimulations, for mice expressing

ChrimsonR and stGtACR2. The red illumination triggers the elevated dopamine release,

while the use of blue light can effectively and promptly suppress the signal. Similar

illumination patterns are implemented on animals expressing different opsins (Figure S18)

and results are summarized and compared in Figures 4i–4k. Evidently, red or blue

illuminations bidiretionally regulate dopamine signals in the experiment group (with both

ChrimsonR and stGtACR2). On the other hand, animals expressing only ChrimsonR can be

excited by both red and blue light, and those with only stGtACR2 can only be inhibited by

blue light and show no reaction under red light. Finally, the control group (with EGFP and

mCherry) does not respond to any optical stimulation patterns. Taken together, these results

clearly demonstrate that wireless dual-color LED combined with optogenetics enable

efficient, bidirectional control of neuronal activity in the VTA and DA release of the terminal

in the NAc.

DISCUSSION

Optogenetic modulations offer a viable solution for cell-type-specific targeting, but still

demand biological and engineering tools for versatile functionalities. In this work, we report

that bidirectional optogenetic activation and inhibition in behaving animals can be achieved

by co-expressing spectrally distinct excitatory and inhibitory channelrhodopsins, cooperating

with a wireless, dual-color micro-LED probe for colocalized red and blue stimulations. With

collective results at cellular, circuit and behavioral levels, successful proof-of-concept

demonstrations in the VTA-NAc circuitry in freely moving mice clearly establish the utility

of such a synergistic methodology.

The unique characteristics of thin-film, microscale optoelectronic devices and their

integration strategies implicate prominent advantages over conventional fibers or single color

LED sources. In terms of technology developments, more directions can be explored and

some preliminary efforts are presented below: 1) Multiple red and blue micro-LEDs can be

assembled on the implantable probe via the stacking approach, realizing multi-channel,

bidirectional modulations imposed on multiple brain regions (Movie S3). It is noted that the

efficacy of preference behavior regulations demonstrated in Figure 3 would also be enhanced

by implementing dual channel stimulations bilaterally in the VTA[47, 48, 49, 50]. 2) The use of

miniaturized wireless control circuits allows independently addressing multiple, untethered

animal objects, making the interrogations of complex social behaviors more feasible (Movie

S4)[51, 52, 53, 54]. 3) Micro-LEDs with other colors can also be included and target on specific

opsins more efficiently. For example, a tri-color probe can be achieved by similarly stacking

red, green and blue micro-LEDs in a vertical structure (Movie S5). Independently controlled

micro-LEDs can interact with different opsins that express in the same or different

populations of cells. 4) Micro-LED based optogenetics can also be combined with previously

reported electrical, optical and chemical sensors as well as microchannel devices for drug

administrations[55, 56, 57, 58], achieving closed-loop neural activity stimulations and detections.

Taken together, we envision a multi-spectral, multi-channel, multi-objective and multi-

functional optical neural interface, realizing wireless, real-time, bi-directional, close-loop and

simultaneous interrogations of neural signals in living animals.

On the other hand, the efficacy of optogenetic modulations can be further optimized

by advanced biological engineering. A recent effort fuses excitatory and inhibitory

channelrhodopsins (ChrimsonR and GtACR2) in a single, trafficking-optimized tandem

protein (BiPOLES), improving the efficiency of opsin co-expression[59]. In addition to the

bidirectional modulation of a same cell population, the dual-color (or even tri-color) micro-

LED probe can also selectively activate genetically distinct and spatially intermingled cells

expressing spectrally resolved opsins (for example, blue for ChR2 or GtACR2, green for

C1V1 or GtACR1 and red for ChrimsonR or halorhodopsin)[3, 7, 8, 22] and provide prolonged

activation of neuronal activity by regulating step-function opsins (for example, SSFO and

SOUL)[11, 13, 60], which is also of critical importance for understanding the circuit structures

and functions. Possible applications would include bidirectional modulations of neural

activities by leveraging electrophysiological or neurotransmitter signals for medications of

seizures or Parkinson's disease[27, 61, 62]. Besides central nervous systems, peripheral neural

circuits can also be targeted for medical treatments[63, 64]. In summary, the results presented

here provide a viable means for fundamental neuroscience studies and advanced

neuroengineering applications.

REFERENCES

1. Won SM, Song E, Reeder JT, Rogers JA. Emerging modalities and implantable technologies for neuromodulation. *Cell* 2020, **181**(1)**:** 115-135.

2. Vázquez-Guardado A, Yang Y, Bandodkar AJ, Rogers JA. Recent advances in neurotechnologies with broad potential for neuroscience research. *Nat Neurosci* 2020, **23**(12)**:** 1522-1536.

3. Boyden ES, Zhang F, Bamberg E, Nagel G, Deisseroth K. Millisecond-timescale, genetically targeted optical control of neural activity. *Nat Neurosci* 2005, **8**(9)**:** 1263-1268.

4. Zhang F, Wang L-P, Brauner M, Liewald JF, Kay K, Watzke N*, et al.* Multimodal fast optical interrogation of neural circuitry. *Nature* 2007, **446**(7136)**:** 633-639.

5. Yizhar O, Fenno LE, Davidson TJ, Mogri M, Deisseroth K. Optogenetics in neural systems. *Neuron* 2011, **71**(1)**:** 9-34.

6. Chow BY, Han X, Dobry AS, Qian X, Chuong AS, Li M*, et al.* High-performance genetically targetable optical neural silencing by light-driven proton pumps. *Nature* 2010, **463**(7277)**:** 98-102.

7. Klapoetke NC, Murata Y, Kim SS, Pulver SR, Birdsey-Benson A, Cho YK*, et al.* Independent optical excitation of distinct neural populations. *Nat Methods* 2014, **11**(3)**:** 338-346.

8. Govorunova EG, Sineshchekov OA, Janz R, Liu X, Spudich JL. Natural light-gated anion channels: A family of microbial rhodopsins for advanced optogenetics. *Science* 2015, **349**(6248)**:** 647.

9. Mahn M, Gibor L, Patil P, Cohen-Kashi Malina K, Oring S, Printz Y*, et al.* High-efficiency optogenetic silencing with soma-targeted anion-conducting channelrhodopsins. *Nat Commun* 2018, **9**(1)**:** 4125.

10. Zhang F, Prigge M, Beyriere F, Tsunoda SP, Mattis J, Yizhar O, *et al.* Red-shifted

optogenetic excitation: a tool for fast neural control derived from Volvox carteri. *Nat*

Neurosci 2008, **11**(6)**:** 631-633.

11. Berndt A, Yizhar O, Gunaydin LA, Hegemann P, Deisseroth K. Bi-stable neural state

switches. *Nat Neurosci* 2009, **12**(2)**:** 229-234.

12. Hochbaum DR, Zhao Y, Farhi SL, Klapoetke N, Werley CA, Kapoor V, *et al.* All-

optical electrophysiology in mammalian neurons using engineered microbial

rhodopsins. *Nat methods* 2014, **11**(8)**:** 825-833.

13. Gong X, Mendoza-Halliday D, Ting JT, Kaiser T, Sun X, Bastos AM, *et al.* An ultra-

sensitive step-function opsin for minimally invasive optogenetic stimulation in mice

and macaques. *Neuron* 2020, **107**(1)**:** 38-51.

14. Brown J, Behnam R, Coddington L, Tervo DGR, Martin K, Proskurin M, *et al.*

Expanding the optogenetics toolkit by topological inversion of rhodopsins. *Cell* 2018,

175(4)**:** 1131-1140.

15. Chuong AS, Miri ML, Busskamp V, Matthews GAC, Acker LC, Sørensen AT, *et al.*

Noninvasive optical inhibition with a red-shifted microbial rhodopsin. *Nat Neurosci*

2014, **17**(8)**:** 1123-1129.

16. Chuong AS, Miri ML, Busskamp V, Matthews GAC, Acker LC, Sørensen AT, *et al.*

Noninvasive optical inhibition with a red-shifted microbial rhodopsin. *Nature*

neuroscience 2014, **17**(8)**:** 1123-1129.

17. Schild LC, Glauser DA. Dual Color Neural Activation and Behavior Control with

Chrimson and CoChR in Caenorhabditis elegans. *Genet* 2015, **200**(4)**:** 1029-1034.

18. Maimon BE, Sparks K, Srinivasan S, Zorzos AN, Herr HM. Spectrally distinct

channelrhodopsins for two-colour optogenetic peripheral nerve stimulation. *Nat*

Biomed Eng 2018, **2**(7)**:** 485-496.

19.   Lu Y, Truccolo W, Wagner FB, Vargas-Irwin CE, Ozden I, Zimmermann JB*, et al.*

Optogenetically induced spatiotemporal gamma oscillations and neuronal spiking

activity in primate motor cortex. *J Neurosci* 2015, **113**(10)**:** 3574-3587.

20.   Galvan A, Caiola MJ, Albaugh DL. Advances in optogenetic and chemogenetic

methods to study brain circuits in non-human primates. *J Neural Transm* 2018,

125(3)**:** 547-563.

21.   Park S, Loke G, Fink Y, Anikeeva P. Flexible fiber-based optoelectronics for neural

interfaces. *Chem Rev* 2019, **48**(6)**:** 1826-1852.

22.   Yizhar O, Fenno LE, Prigge M, Schneider F, Davidson TJ, O'Shea DJ*, et al.*

Neocortical excitation/inhibition balance in information processing and social

dysfunction. *Nature* 2011, **477**(7363)**:** 171-178.

23.   Kampasi K, English DF, Seymour J, Stark E, McKenzie S, Voroslakos M*, et al.* Dual

color optogenetic control of neural populations using low-noise, multishank

optoelectrodes. *Microsyst Nanoeng* 2018, **4:** 1-16.

24.   Kim T-i, McCall JG, Jung YH, Huang X, Siuda ER, Li Y*, et al.* Injectable, Cellular-

Scale Optoelectronics with Applications for Wireless Optogenetics. *Science* 2013,

340(6129)**:** 211.

25.   Montgomery KL, Yeh AJ, Ho JS, Tsao V, Mohan Iyer S, Grosenick L*, et al.*

Wirelessly powered, fully internal optogenetics for brain, spinal and peripheral

circuits in mice. *Nat Methods* 2015, **12**(10)**:** 969-974.

26.   Shin G, Gomez AM, Al-Hasani R, Jeong YR, Kim J, Xie Z*, et al.* Flexible Near-Field

Wireless Optoelectronics as Subdermal Implants for Broad Applications in

Optogenetics. *Neuron* 2017, **93**(3)**:** 509-521.e503.

27.   Wykes RC, Kullmann DM, Pavlov I, Magloire V. Optogenetic approaches to treat

epilepsy. *J Neurosci Methods* 2016, **260:** 215-220.

28. Li Y, Zeng J, Zhang J, Yue C, Zhong W, Liu Z, *et al.* Hypothalamic circuits for

predation and evasion. *Neuron* 2018, **97**(4)**:** 911-924.

29. Lu L, Ren Y, Yu T, Liu Z, Wang S, Tan L, *et al.* Control of locomotor speed, arousal,

and hippocampal theta rhythms by the nucleus incertus. *Nat Commun* 2020, **11**(1)**:** 1-

16.

30. Roseberry TK, Lee AM, Lalive AL, Wilbrecht L, Bonci A, Kreitzer AC. Cell-Type-

Specific Control of Brainstem Locomotor Circuits by Basal Ganglia. *Cell* 2016,

164(3)**:** 526-537.

31. Gradinaru V, Zhang F, Ramakrishnan C, Mattis J, Prakash R, Diester I, *et al.*

Molecular and cellular approaches for diversifying and extending optogenetics. *Cell*

2010, **141**(1)**:** 154-165.

32. Stuber GD, Sparta DR, Stamatakis AM, van Leeuwen WA, Hardjoprajitno JE, Cho S,

et al. Excitatory transmission from the amygdala to nucleus accumbens facilitates

reward seeking. *Nature* 2011, **475**(7356)**:** 377-380.

33. Noked O, Levi A, Someck S, Amber-Vitos O, Stark E. Bidirectional optogenetic

control of inhibitory neurons in freely-moving mice. *IEEE Transactions on*

Biomedical Engineering 2020.

34. Li L, Liu C, Su Y, Bai J, Wu J, Han Y, *et al.* Heterogeneous Integration of Microscale

GaN Light-Emitting Diodes and Their Electrical, Optical, and Thermal

Characteristics on Flexible Substrates. *Adv Mater Technol* 2018, **3**(1)**:** 1700239.

35. Liu C, Zhang Q, Wang D, Zhao G, Cai X, Li L, *et al.* High Performance,

Biocompatible Dielectric Thin-Film Optical Filters Integrated with Flexible

Substrates and Microscale Optoelectronic Devices. *Adv Opt Mater* 2018, **6**(15)**:**

1800146.

36. Ding H, Lu L, Shi Z, Wang D, Li L, Li X*, et al.* Microscale optoelectronic infrared-

to-visible upconversion devices and their use as injectable light sources. *Proc Natl*

Acad Sci U S A 2018, **115**(26)**:** 6632-6637.

37. Lammel S, Lim BK, Malenka RC. Reward and aversion in a heterogeneous midbrain

dopamine system. *Neuropharmacology* 2014, **76:** 351-359.

38. Morales M, Margolis EB. Ventral tegmental area: cellular heterogeneity, connectivity

and behaviour. *Nat Rev Neurosci* 2017, **18**(2)**:** 73.

39. Cohen JY, Haesler S, Vong L, Lowell BB, Uchida N. Neuron-type-specific signals

for reward and punishment in the ventral tegmental area. *Nature* 2012, **482**(7383)**:** 85-

88.

40. Al-Hasani R, McCall JG, Shin G, Gomez AM, Schmitz GP, Bernardi JM*, et al.*

Distinct Subpopulations of Nucleus Accumbens Dynorphin Neurons Drive Aversion

and Reward. *Neuron* 2015, **87**(5)**:** 1063-1077.

41. Sun F, Zeng J, Jing M, Zhou J, Feng J, Owen SF*, et al.* A Genetically Encoded

Fluorescent Sensor Enables Rapid and Specific Detection of Dopamine in Flies, Fish,

and Mice. *Cell* 2018, **174**(2)**:** 481-496 e419.

42. Yuan L, Dou YN, Sun YG. Topography of Reward and Aversion Encoding in the

Mesolimbic Dopaminergic System. *J Neurosci* 2019, **39**(33)**:** 6472-6481.

43. Sun F, Zhou J, Dai B, Qian T, Zeng J, Li X*, et al.* Next-generation GRAB sensors for

monitoring dopaminergic activity in vivo. *Nat Methods* 2020, **17**(11)**:** 1156-1166.

44. Gunaydin Lisa A, Grosenick L, Finkelstein Joel C, Kauvar Isaac V, Fenno Lief E,

Adhikari A*, et al.* Natural Neural Projection Dynamics Underlying Social Behavior.

Cell 2014, **157**(7)**:** 1535-1551.

45. Lin R, Liang J, Wang R, Yan T, Zhou Y, Liu Y*, et al.* The Raphe Dopamine System

Controls the Expression of Incentive Memory. *Neuron* 2020, **106**(3)**:** 498-514.e498.

46. Cui G, Jun SB, Jin X, Pham MD, Vogel SS, Lovinger DM, *et al.* Concurrent

activation of striatal direct and indirect pathways during action initiation. *Nature*

2013, **494**(7436)**:** 238-242.

47. Yuan L, Liang T-Y, Deng J, Sun Y-G. Dynamics and functional role of dopaminergic

neurons in the ventral tegmental area during itch processing. *J Neurosci Res* 2018,

38(46)**:** 9856-9869.

48. Li Y, Li C-Y, Xi W, Jin S, Wu Z-H, Jiang P, *et al.* Rostral and caudal ventral

tegmental area GABAergic inputs to different dorsal raphe neurons participate in

opioid dependence. *Neuron* 2019, **101**(4)**:** 748-761.

49. Taylor NE, Van Dort CJ, Kenny JD, Pei J, Guidera JA, Vlasov KY, *et al.* Optogenetic

activation of dopamine neurons in the ventral tegmental area induces reanimation

from general anesthesia. *Proc Natl Acad Sci USA* 2016, **113**(45)**:** 12826-12831.

50. Chang CY, Gardner M, Di Tillio MG, Schoenbaum G. Optogenetic blockade of

dopamine transients prevents learning induced by changes in reward features. *Curr*

Biol 2017, **27**(22)**:** 3480-3486.

51. de Chaumont F, Ey E, Torquet N, Lagache T, Dallongeville S, Imbert A, *et al.* Real-

time analysis of the behaviour of groups of mice via a depth-sensing camera and

machine learning. *Nat Biomed Eng* 2019, **3**(11)**:** 930-942.

52. Kingsbury L, Huang S, Raam T, Letizia SY, Wei D, Hu RK, *et al.* Cortical

representations of conspecific sex shape social behavior. *Neuron* 2020, **107**(5)**:** 941-

953.

53. Kingsbury L, Huang S, Wang J, Gu K, Golshani P, Wu YE, *et al.* Correlated neural

activity and encoding of behavior across brains of socially interacting animals. *Cell*

2019, **178**(2)**:** 429-446.

54.   Hong W, Kim D-W, Anderson DJ. Antagonistic control of social versus repetitive

482   self-grooming behaviors by separable amygdala neuronal subsets. *Cell* 2014, **158**(6)**:**

483   1348-1361.

484   55.   Chiang C-H, Won SM, Orsborn AL, Yu KJ, Trumpis M, Bent B*, et al.* Development

485   of a neural interface for high-definition, long-term recording in rodents and nonhuman

486   primates. *Sci Transl Med* 2020, **12**(538)**:** eaay4682.

487   56.   Qazi R, Gomez AM, Castro DC, Zou Z, Sim JY, Xiong Y*, et al.* Wireless optofluidic

488   brain probes for chronic neuropharmacology and photostimulation. *Nat Biomed Eng*

489   2019, **3**(8)**:** 655-669.

490   57.   Lu L, Gutruf P, Xia L, Bhatti DL, Wang X, Vazquez-Guardado A*, et al.* Wireless

491   optoelectronic photometers for monitoring neuronal dynamics in the deep brain. *Proc*

492   *Natl Acad Sci USA* 2018, **115**(7)**:** E1374-E1383.

493   58.   Liu C, Zhao Y, Cai X, Xie Y, Wang T, Cheng D*, et al.* A wireless, implantable

494   optoelectrochemical probe for optogenetic stimulation and dopamine detection.

495   *Microsyst Nanoeng* 2020, **6**(1)**:** 1-12.

496   59.   Vierock J, Rodriguez-Rozada S, Pieper F, Dieter A, Bergs A, Zeitzschel N*, et al.*

497   BiPOLES: a tool for bidirectional dual-color optogenetic control of neurons. *BioRxiv*

498   2020. https://doi.org/10.1101/2020.07.15.204347

499   60.   Selimbeyoglu A, Kim CK, Inoue M, Lee SY, Hong ASO, Kauvar I*, et al.* Modulation

500   of prefrontal cortex excitation/inhibition balance rescues social behavior in

501   CNTNAP2-deficient mice. *SCI TRANSL MED* 2017, **9**(401)**:** eaah6733.

502   61.   Chen S, Weitemier AZ, Zeng X, He L, Wang X, Tao Y*, et al.* Near-infrared deep

503   brain stimulation via upconversion nanoparticle–mediated optogenetics. *Science*

504   2018, **359**(6376)**:** 679.

505   62.   Kravitz AV, Freeze BS, Parker PRL, Kay K, Thwin MT, Deisseroth K*, et al.* Regulation of parkinsonian motor behaviours by optogenetic control of basal ganglia circuitry. *Nature* 2010, **466**(7306)**:** 622-626.

508   63.   Keppeler D, Schwaerzle M, Harczos T, Jablonski L, Dieter A, Wolf B*, et al.* Multichannel optogenetic stimulation of the auditory pathway using microfabricated LED cochlear implants in rodents. *Sci Transl Med* 2020, **12**(553)**:** eabb8086.

511   64.   Michoud F, Seehus C, Schönle P, Brun N, Taub D, Zhang Z*, et al.* Epineural optogenetic activation of nociceptors initiates and amplifies inflammation. *Nat Biotechnol* 2021, **39**(2)**:** 179-185.

Acknowledgements

This work is supported by National Natural Science Foundation of China (NSFC) (61874064), Beijing Municipal Natural Science Foundation (4202032), Beijing Innovation Center for Future Chips at Tsinghua University, Center for Flexible Electronics Technology at Tsinghua University, Beijing National Research Center for Information Science and Technology (BNR2019ZS01005). We thank Prof. Y. Li (Peking University) for support on $GRAB_{DA2m}$, Y. Zhou and Y. Ju (Chinese Institute for Brain Research, Beijing) for support on stGtACR2.

Author contributions

X. S. and M. L. developed the concepts. L. Li, G. T., Y. Z., X. C., Z. S., H. D., C. L., Y. X., H. W., L. Y. and X. S. performed material, device and circuit design, fabrication and characterization. L. Li, X. C., D. C, and X. S. performed the simulations. L. Li, L. Lu, Y. R., G. T., X. F., L. Y., M. L., and X. S. designed and performed biological experiments. L. Li, L. Lu, M. L. and X. S. wrote the paper in consultation with other authors.

Competing interests

The authors declare no competing interests.

Data and materials availability

All data needed to evaluate the conclusions in the paper are present in the paper and/or the Supplementary information.

Figure 1

a

b

Na⁺   K⁺   Cl⁻   Cl⁻   Cl⁻

c

blue LED

filter

red LED

polyimide

d

blue LED

filter

red LED

5 µm

e

LEDs off   blue LED on

200 µm

red LED on   both LEDs on

f

red LED   ChrimsonR
blue LED   stGtACR2

Intensity (a.u.)

Wavelength (nm)

g

Experiment

Simulation

200 µm

blue on   red on   both on   lateral LEDs

h

5 mm

i

1 cm

Figure 1. Wirelessly operated, dual-color microscale LED (micro-LED) probes for bidirectional, colocalized optogenetic control of neural activities. a, Schematic view of the probe inserted into the target brain area to perform dual-color neural activation and inhibition. **b**, Schematic, cellular scale depiction of the modulation function: red light regulates the cation channel ChrimsonR for depolarization, while blue light regulates the anion channel stGtACR2 for hyperpolarization. **c**, Exploded view of the dual-color LED probe made from vertically stacked blue and red thin-film micro-LEDs assembled on a flexible polyimide (PI) substrate, with a thin-film filter interface for spectrally selective reflection and transmission. **d**, Cross-sectional scanning electron microscopic (SEM) image of the probe structure, with false colors highlighting (from bottom to top) the indium gallium arsenide (InGaP) red micro-LED, the multilayer silicon dioxide / titanium dioxide (SiO_2/TiO_2) filter, and the indium gallium nitride (InGaN) blue micro-LED. Layers of SU-8 photoresists serve as the bonding materials. **e**, Optical images of a micro-LED probe (with a thickness of ~120 μm and a width of ~320 μm), showing independent control of blue and red emissions under current injections. **f**, Emission spectra of the blue and red micro-LEDs, in comparison with the action spectra of ion channels stGtACR2 and ChrimsonR (data extracted from references[7, 8]). **g**, Experimental photographs and simulated results illustrating blue and red light propagations for a micro-LED probe embedded into a brain phantom. Results for a probe with laterally assemble blue and red LEDs are also presented, showing misplaced blue and red emissions. Iso-intensity lines show 10% of the maximum power. **h**, Images of a micro-LED probe integrated with a wireless circuit for independent control of dual-color emissions. **i**, Photos of a behaving mouse after intracranial implantation of a probe, with (bottom) and without (top) the wireless circuit module.

Figure 2

Figure 2. Bidirectional optogenetic activation and inhibition of neurons with red and blue light stimulation. a, Schematic strategy for co-expressing ChrimsonR and stGtACR2 in the ventral tegmental area (VTA) of *DAT-cre* mice. **b**, Immunostained fluorescence images of the VTA region expressing DAPI (blue), stGtACR2 (green), and ChrimsonR (red) and a merged image, 3 weeks after viral injection. **c**, Zoomed view showing cells co-expressing stGtACR2 (green) and ChrimsonR (red). **d**, Percentages of stGtACR2-, ChrimsonR- and co-expressing neurons among DAPI cells in the VTA (mean ± s.e.m., 3 slices per animal, $n = 3$ mice). **e**, Illustration of the experiment setup with combined blue (480 nm) and red (633 nm) laser beams (spot diameter ~1 mm) for optogenetic stimulating brain slices during patch-clamp recordings *in vitro*. **f**, Example traces showing neural activation by pulsed red light at different irradiances (20 Hz, 10-ms pulse, 20-pulse train). **g**, Activated spike probability as a function of red irradiances (mean ± s.e.m., $n = 4$ cells). **h**, Example photocurrents of neurons at membrane potentials varied from −105 mV to −5 mV (bottom to top), under continuous blue light illumination (duration 5 s, irradiance 8 mW/mm^2). **i**, Example traces of neural inhibition by continuous blue light at different irradiances (duration 0.5 s). **j**, Inhibited spike probability as a function of blue irradiances (mean ± s.e.m., $n = 5$ cells). **k**, Traces presenting bidirectional spike activation and inhibition with alternating red and blue illuminations (Patching current +90 pA; Red: duration 3 s, 20 Hz, 10-ms pulse, 2.2 mW/mm^2; Blue: duration 3 s, 8.5 mW/mm^2). **l**, Example traces of neural activities with simultaneous red and blue illuminations at different irradiances (duration 6 s). **m**, Summarized plot of measured spike probability (normalized mean values) as a function of red and blue irradiances ($n = 5$ cells).

Figure 3

Figure 3. Bidirectional, *in vivo* optogenetic regulation of animals' preference or aversion behaviors with wirelessly operated, dual-color microscale LED (micro-LED) probes. a, Diagram illustrating the combined viral injection (AAV2/9-CAG-DIO-ChrimsonR-mCherry and AAV2/9-EF1a-DIO-stGtACR2-EGFP) followed by implanting the micro-LED probe into the VTA of *DAT-cre* mice. **b**, Micrographs of the VTA showing co-expression of stGtACR2 and ChrimsonR, with a dual-color micro-LED probe embedded into the tissue. **c**, Illustration of the real-time place preference test with a two-compartment arena, where a mouse receives signals for dual-color control and its locomotion is recorded by a camera. **d**, Patterns used for optogenetic modulations, including a 10-min pretest, a 30-min red LED stimulation (20 Hz, 10-ms pulse, current 7 mA), a 30-min rest, and a 30-min blue LED stimulation (continuous, current 5 mA). **e**, Representative heat maps showing real-time preference and aversion behavior following red or blue stimulation for mice co-expressing stGtACR2 + ChrimsonR (experiment, or Exp) or EGFP + mCherry (control, or Ctrl). **f**, Summary of preference indices (the ratio of the time that mice spend in the left chamber to the whole recorded time) under red and blue stimulations for Exp and Ctrl groups ($n = 4$ mice each). Student's t test, ** $P < 0.01$, *** $P < 0.001$, n.s. $P > 0.1$. **g**, Preference indices measured at different times under red and blue stimulations for Exp and Ctrl groups ($n = 4$ mice each). **h**, Preference indices measured at different times under red and blue stimulations for groups expressing only ChrimsonR or stGtACR2 ($n = 3$ mice each).

Figure 4

Figure 4. Bidirectional, *in vivo* optogenetic control of dopamine (DA) release simultaneously recorded in the nucleus accumbens (NAc) during dual-color stimulation in the VTA. a, Diagram illustrating the combined viral injection (AAV2/9-CAG-DIO-ChrimsonR-mCherry and AAV2/9-EF1a-DIO-stGtACR2-EGFP) followed by implanting the micro-LED probe into the VTA, and the viral injection (AAV2-hSyn-GRAB$_{DA2m}$) followed by optical fiber implantation into the NAc of *DAT-cre* mice. **b**, Micrographs of (top) the NAc showing expression of DA2m and the fiber track, (bottom) the VTA showing co-expression of stGtACR2 and ChrimsonR (merged color) and the micro-LED probe track, 2 weeks after viral injection and probe implantation. **c**, Representative traces of DA signal increase or decrease in response to stimulations by the red LED (7 mA, 20 Hz, 20-ms pulse, 6 s) or the blue LED (2.4 mA, continuous, 12 s). The solid lines and shaded areas indicate the mean and s.e.m., respectively (*n* = 4 mice). **d**, Accumulative DA signals in response to red and blue stimulations under different LED currents. AUC: area under the curve during stimulation. **e**, Summary of measured DA signals (AUC) under red or blue stimulations (Red LED: 20 Hz, 20-ms pulse, 10 mA, 6 s; Blue LED, continuous, 3 mA, 12 s) for different animal groups expressing: ChrimsonR + stGtACR2 (*n* = 4 mice), ChrimsonR + EGFP (*n* = 3 mice), stGtACR2 + mCherry (*n* = 3 mice), EGFP + mCherry (*n* = 3 mice). Student's *t* test. **f**, Example traces of DA signals under combined red and blue stimulations with varied LED currents. **g**, Summarized plot of measured DA signals (AUC) (mean value) as a function of red and blue LED currents for 6 s (*n* = 3 mice). **h**, Plots of optogenetically evoked DA transients for mice co-expressing stGtACR2 + ChrimsonR, in response to different stimulation patterns (Red: red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s; Blue: blue LED on, continuous, 3 mA, 6–12 s; Green: red LED on 0–12 s and blue LED on 6–12 s). The solid lines and shaded areas indicate the mean and s.e.m., respectively. (*n* = 4 mice). **i–k**, Summary of measured DA signals (AUC) for mice with and without stGtACR2 and ChrimsonR expressions (*n* = 3 or 4 mice each), in response to different stimulation patterns. **i**, red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s. **j**, blue LED on, continuous, 3 mA, 6–12 s. **k**, red LED on 0–12 s and blue LED on 6–12 s.

Figure S1

Figure S1. Schematic illustration of the process flow for fabricating the dual-color micro-LED probes, with vertically stacked red LED, filter and blue LED printed on PI substrates.

Figure S2

Figure S2. Measured optoelectronic properties of the red and the blue micro-LEDs in a probe structure. (a) Current–voltage curves. (b) External quantum efficiencies (EQEs) as a function of current. (c) Irradiance (power density on the LED surface) versus current. (d) Angular dependent emission profiles (in arbitrary unit).

Figure S3

Figure S3. Thermal behaviors of a dual-color micro-LED probe. (a, b) Images and infrared thermographs showing a probe embedded into a brain phantom (~0.5 mm below the phantom surface). (a) Red LED on (injection current 7 mA, pulse frequency 20 Hz, pulse duration 10 ms). (b) Blue LED on (injection current 2 mA, continuous operation). (c–j) Measured and simulated temperature rise in phantom or brain when the red LED or the blue LED is on. (c) Temperature rise in phantom, measured by an infrared thermal camera when the red LED is on (current 0–10 mA, frequency 10–50 Hz, pulse duration 10 ms). (d) Temperature rise in the brain of living mice, measured with a thin wire thermocouple attached to the probe. (e) Simulated steady-state temperature distribution in the brain tissue surrounding the probe, when the red LED is on (current 7 mA, frequency 20 Hz, pulse duration 10 ms). (f) Simulated maximum temperature rise of the brain tissue when the red LED is in different operation conditions. (g–j) Corresponding results when the blue LED is on (continuous mode). current = 2 mA in (i).

Figure S4

Figure S4. Schematic diagram showing the operational principle for the wireless circuit system controlling the dual-color micro-LED probe.

Figure S5

Figure S5. Detailed circuit design diagram of the wireless circuit system.

Figure S6

Figure S6. Layout of the designed printed circuit board (top: front view; bottom: back view).

Figure S7

	components
L1, L2	High frequency chip inductor +/- 5%, 6.8 nH
L3	High frequency chip inductor +/- 5%, 5.6 nH
C1, C2	NP0 +/- 2%, 15 pF
C3, C4	X7R +/- 10%, 2.2 nF
C5	NP0 +/- 0.1pF, 1.5 pF
C6	NP0 +/- 0.1pF, 1.0 pF
C7,C9,C11	X7R +/- 10%, 100 nF
C8,C10	X7R +/- 10%, 33 nF
C12,C14	X7R +/- 10%, 1 μF
Cout1,Cout2	2.2 μF
C13	NP0 50V, 10 pF
D0, D1	Indicating red LED, 610 nm, SMD 0402, 1 mm * 0.5 mm * 0.4 mm
R1	22 kΩ, 1%
nRF24LE1	nRF24LE1-F16Q24, QFN24 4 mm * 4 mm package, MCU
Component_1 Component_2	ZLED7012 ZI1R, LED current driver
X1	Crystal oscillator, SMD-3225, 16 MHz, CL=9 pF, +/- 60 ppm 3.2 mm * 2.5 mm package
E1	Antenna, H2U34WGTQW0100, AA055, 3.2 mm * 1.6 mm * 0.5 mm package
P1, P2	Pin Header, 6 pin, 1.25 mm, for programming
P3	FPC connector, 4 pin, 0.5 mm, for LED probe
P4	PCB header Connector, male, DIP Straight, 2 pin, 2.54 mm, for battery
Battery	Rechargeable lithium ion battery, 3.7 V, 45 mAh

Figure S7. List of components used in the wireless circuit system.

Figure S8

Figure S8. Recorded traces of a representative neuron with combined red and blue illumination at different irradiances. Red LED: 20 Hz, 10-ms pulse; Blue LED: continuous operation.

Figure S9

Figure S9. Immunoreactivity results showing the biocompatibility of the micro-LED probe, in comparison with those obtained with a silica based optical fiber. (a) Representative confocal fluorescence images of horizontal brain slices, showing immunohistochemical staining of astrocytes (GFAP) and activated microglia (Iba1) for both the LED probe and the fiber, after 1 day and 21 days implantation. Green: GFAP; Red: Iba1; Blue: DAPI. (b) Percentages of GFAP and Iba1 cell populations among DAPI cells collected at a distance of 200 μm from the edge of implantation. The LED probe and the fiber show similar results of inflammatory glial responses occurring after implantation. ($n = 3$ mice for each group).

Figure S10

Figure S10. Number of functional micro-LEDs measured after chronic implantation in behaving mice.

Figure S11

Figure S11. (a) Schematic diagram and (b) top-view photograph of the real-time place preference test with a two-compartment arena.

Figure S12

Figure S12. Representative heat maps showing real-time preference and aversion behavior following red or blue stimulation for mice expressing stGtACR2 + ChrimsonR (experiment, or Exp), EGFP + mCherry (control, or Ctrl), ChrimsonR + EGFP and stGtACR2 + mCherry. The data for ChrimsonR + stGtACR2 and EGFP + mCherry groups are identical to those presented in Fig. 3e.

Figure S13

Figure S13. Summary of preference indices (the ratio of the time that mice spend in the left chamber to the whole recorded time) under red and blue stimulations for different animal groups expressing ChrimsonR + stGtACR2 (n = 4 mice), EGFP + mCherry (n = 4 mice), ChrimsonR + EGFP (n = 3 mice), stGtACR2 + mCherry (n = 3 mice). Student's t test, ** P < 0.01, *** P < 0.001, n.s. P > 0.1. The data for ChrimsonR + stGtACR2 and EGFP + mCherry groups are identical to those presented in Fig. 3f.

Figure S14

Fiber in NAc

LED Probe in VTA

Figure S14. Schematic illustration (horizontal view) for the implantation of the fiber for DA recording in the NAc, and the dual-color LED probe for optogenetic stimulations in the VTA. The specific insertion angle is chosen for the LED probe, to minimize the optical crosstalk that disturbs the fiber photometric recording in the NAc.

Figure S15

Figure S15. DA signals recorded as the fluorescence of GRAB$_{DA2m}$ in the NAc, in response to optogenetic stimulations applied in the VTA by red or blue micro-LEDs. (a, b) Representative traces of DA signals in response to stimulations by (a) the red LED (20 Hz, 20 ms, 6 s) and (b) the blue LED (continuous, 12 s) at various currents. The solid lines and shaded areas indicate the mean and s.e.m., respectively. (n = 4 mice). AUC: area under the curve during stimulation. The data for red LED at 7 mA and blue LED at 2.4 mA are identical to those presented in Figure 4c.

Figure S16

Figure S16. Detailed results of measured DA signals (AUC) with the combination of red and blue irradiance at different LED currents simultaneously for 6 s ($n = 3$ mice).

Figure S17

Figure S17. Example traces of optogenetically evoked DA transients for mice co-expressing stGtACR2 + ChrimsonR, in response to different stimulation patterns. (a) Red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s. (b) Blue LED on, continuous, 3 mA, 0–12 s. (c) Blue LED on, continuous, 3 mA, 6–12 s. (d) Red LED on 0–12 s and Blue LED on 6–12 s. (e) Red LED on 0–6 s and Blue LED on 0–12 s. The solid lines and shaded areas indicate the mean and s.e.m., respectively.

Figure S18

Figure S18. Plots of optogenetically evoked DA transients in response to different stimulation patterns (Red: red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s; Blue: blue LED on, continuous, 3 mA, 6–12 s; Green: red LED on 0–12 s and blue LED on 6–12 s), for different animal groups expressing (a) ChrimsonR + stGtACR2 (n = 4 mice), (b) ChrimsonR + EGFP (n = 3 mice), (c) stGtACR2 + mCherry (n = 3 mice), (d) EGFP + mCherry (n = 3 mice). The solid lines and shaded areas indicate the mean and s.e.m., respectively. The data in Figure S18a are identical to those presented in Figure 4h.

Movie S1

Movie S1. Video for dual-color micro-LED probe, displaying alternating blue and red emissions.

Movie S2

Movie S2. Video for multiple dual-color LED probes assembled with wirelessly operated circuit modules, showing capabilities for independent light emission control.

Movie S3

Movie S3. Video for multi-channel dual-color probes with 6 pairs of stacked red-blue micro-LEDs, displaying alternating blue and red emissions.

Movie S4

Movie S4. Video for four freely moving mice implanted with micro-LED probes and head-mounted circuits.

Movie S5

Movie S5. Video for a tri-color (red-green-blue) micro-LED needle structure, displaying different colors under current injection.

Supplementary Information

Device Fabrication

Fabrication of red and blue micro-LEDs

Details about the thin-film micro-LED structures and fabrication processes can be found in our previous work.[1,2] Via metal-organic chemical vapor deposition (MOCVD), red and blue LEDs are originally grown on gallium arsenide (GaAs) and sapphire substrates, respectively. The red LED active device contains an indium gallium phosphide (InGaP) based multiple quantum well structure (thickness ~5.6 μm) grown on GaAs with an $Al_{0.95}Ga_{0.05}As$ sacrificial interlayer. The blue LED active device contains an indium gallium nitride (InGaN) based multiple quantum well structure (thickness ~7 μm) grown on sapphire. Photolithographic process, wet / dry etching and metallic electrode deposition define the LED geometry (size ~ 180 μm × 125 μm). Freestanding, thin-film red LEDs are formed by selectively removing the $Al_{0.95}Ga_{0.05}As$ sacrificial layer in a hydrofluoric acid (HF) based solution. Freestanding, thin-film blue LEDs are formed by laser lift off (with a KrF excimer laser at 248 nm). Poly(dimethylsiloxane) (PDMS) based stamps are employed to pick up released thin-film micro-LEDs and transferred them onto other foreign substrates.

Fabrication of thin-film filters

Details about the thin-film filter fabrication can be found in our previous work.[3] The wavelength selective thin-film filters is based on a multilayered titanium dioxide

(TiO_2) and silicon dioxide (SiO_2) structure (thickness ~ 6.6 μm) comprising (from top

to bottom): 19 nm TiO_2 / 15 periods (89 nm SiO_2 + 52 nm TiO_2) / 63 nm SiO_2 / 66 nm

TiO_2 / 22 periods of (73 nm SiO_2 + 42.5 nm TiO_2) / 154 nm SiO_2. The structure is

deposited on GaAs by ion beam-assisted sputter deposition. The filter shape (size ~

180 μm × 125 μm) is defined by laser milling (Nd:YVO_4 laser, 1064 nm). After

removing the GaAs substrate in solutions ($NH_4OH:H_2O_2:H_2O$ = 1:1:2), freestanding

thin-film filters can be picked up and transfer printed onto other foreign substrates by

PDMS stamps.

Fabrication of dual-color micro-LED probes

The process flow to fabricate the dual-color micro-LED probe is described in Figure

S1. The micro-LED probe is based on a flexible substrate made of composite

polyimide (PI) and copper (Cu) films (from top to bottom: Cu/PI/Cu = 18/25/18 μm,

from DuPont). From bottom to top, a red LED, an optical filter and a blue LED are

sequentially transfer printed on the film by PDMS, in a vertical stack. Between each

device layer, SU-8 based epoxy (thickness ~ 2–5 μm) serves an optically transparent

and electrically insulating bonding layer. Each individual LED is metallized with

sputtered metal layers (Cr/Au/Cu/Au = 10/100/500/100 nm). Laser milling (ultraviolet

laser 365 nm) defines the needle shape (width ~ 320 μm, length ~ 5 mm). A bilayer of

PDMS (~20 μm) and parylene (~15 μm) is used for waterproof encapsulation. For

comparison in Fig. 1g, laterally arranged red-blue micro-LED probes are also

fabricated with similar procedures.

Device Characterization and Modeling

Structural and optoelectronic characterizations

Scanning electron microscopic (SEM) images are taken with a ZEISS Auriga SEM / FIB Crossbeam System. Optical images are captured by a microscope MC-D800U(C). The LED emission spectra are measured with a spectrometer (HR2000+, Ocean Optics). Current–voltage characteristics are recorded by a Keithley 2400 source meter. LED irradiances and external quantum efficiencies are measured using a spectroradiometer with an integrating sphere (LabSphere Inc.). The angular dependent emission profiles are measured by a standard Si photodetector (DET36A, Thorlabs) with the device mounted on a goniometer. For *in vitro* test, micro-LED probes are implanted into the brain phantom. The brain phantom is made of agarose (0.5% w/v), hemoglobin (0.2% w/v), and intralipid (1% w/v) mixed in phosphate buffer solutions (98.3% w/v). After stirring, the mixed liquid is heated to boiling and then naturally cooled to room temperature, forming the gels.

Optical modeling

A ray tracing method based on Monte Carlo simulations is used to simulate the light propagation in the brain tissue. In the model, the tissue has absorption coefficients of 0.2 /mm and 0.08 /mm, and scattering coefficients of 47 /mm and 35 /mm, for blue (475 nm) and red (630 nm), respectively. The tissue has an anisotropy factor 0.85 and a refractive index of 1.36.[4]

Thermal characterizations

For *in vitro* experiments, the micro-LED probes are implanted into the brain phantom

(~0.5 mm underneath the surface, at a depth of ~5 mm), and various currents are

injected into the LEDs. The temperature distributions are mapped with an infrared

camera (FOTRIC 228). For *in vivo* experiments, the micro-LED probe is bonded with

an ultra fine flexible micro thermocouple (Physitemp, W-TW-44P/50) and inserted

into the anesthetized mouse brain. Steady-state temperature rises are recorded from

the thermocouple with a thermometer (Physitemp, BAT-12).

Thermal modeling

3D steady-state heat transfer models are established by finite element analysis

(Comsol Multiphysics). The parameters of the brain use in the model include: density

1.1 g/cm^3, thermal conductivity 0.5 W/m/K, heat capacity 3650 J/kg/K, and the

natural heat convection to air. The micro-LED probe serves as the heat source, with

the input thermal power estimated by $P = V \times I \times (1 - EQE)$, where V, I and EQE are

the measured voltage, current, and corresponding external quantum efficiencies for

LEDs.

Wireless Circuit Design

A customized circuit module is designed to wirelessly operate the implantable

micro-LED probe by independently controlling the red and the blue micro-LEDs for

bi-directional modulation. The schematic diagram is depicted in Figure S4, with

detailed circuit diagram, layouts and components presented in Figures S5, S6 and S7,

respectively. The core components include: a micro-controller (nRF24LE1, Nordic

Semiconductor) operated at 2.4 GHz for data communication, and two LED drivers

(ZLED7012, Renesas Electronics Corp.) with programmable current levels,

frequencies, pulse widths for providing a constant current to micro-LEDs.[5] Two red

LEDs are used on the circuit board outside the brain for signal indication, without

affecting the mouse behaviors.[6] Control signals are wirelessly transmitted from the

transmitting module connected to a laptop computer via an antenna operated at a radio

frequency of 2.4 GHz, with a communication distance up to 50 m. The wireless circuit

has a footprint of ~2.2 cm × 1.3 cm and a weight of 1.9 g (including a 0.9-g

rechargeable lithium-ion battery with a capacity of 35 mAh). Used in the experimental

conditions in this paper, the battery lifetime can reach up to 2 hours. The circuit can

be connected with the implanted micro-LED probes via a flexible printed circuit

(FPC) connector before animal experiment.

Biological Studies

Animals

Animal care is in accordance with the institutional guidelines of National Institute of

Biological Sciences in Beijing (NIBS), with protocols proved by Institutional Animal

Care and Use Committee (IACUC). All animals are socially housed in a 12 h/12 h

(lights on at 8 am) light/dark cycle, with food and water *ad alibitum*. Wildtype

C57BL6/N mice are purchased from VitalRiver (Beijing, China). *DAT-Cre* mice

(strain name B6.SJLSlc6a3tm1.1(cre) Bkmn; JAX Strain 006660) are at least six

weeks old at the time of surgery.

Virus production

The AAV2/9-CAG-DIO-ChrimsonR-tdTomato (Plasmid #130909) vector[7] is from

Addgene. The AAV2/9-EF1a-DIO-stGtACR2-EGFP-Kv2.1 plasmid is synthesized

and constructed according to the original reports.[8, 9] The AAV-hSyn-GRAB$_{DA2m}$

plasmid is provided by Prof. Yulong Li (Peking University).[10] AAV vectors are

packaged into serotype2/9 vectors, which consist of AAV2 ITR genomes pseudotyped

with AAV9 serotype capsid proteins. AAV vectors are replicated in HEK293 cells

with the triple plasmid transfection system and purified by chloroform,[11] resulting in

AAV vector titers of about 5×10^{12} particles/mL. Virus suspension is aliquoted and

stored at −80 °C. AAV titers are measured using real-time qPCR. Adeno-associated

viral particles are produced at the Vector Core Facility at Chinese Institute for Brain

Research, Beijing (CIBR).

Stereotaxic Surgery

Adult *DAT-Cre* mice are anesthetized with an intraperitoneal injection of Avertin (250

mg/kg) before surgery, and then placed in a mouse stereotaxic instrument. After

disinfection with 0.3% hydrogen peroxide, a small incision of the scalp is created to

expose the skull. Then, 0.3% hydrogen peroxide is applied to clean the skull. A small

craniotomy is made and a calibrated pulled-glass pipette (Sutter Instrument) is

lowered to the VTA (coordinates 3.64 mm posterior to the bregma, 0.5 mm lateral to

midline, and 4.6 mm ventral to the skull surface). Virus is delivered through a small

durotomy by a glass micropipette using a microsyringe pump (Nanoliter 2000 injector

with the Micro4 controller, WPI).

Four different combinations of virus are used for *in vivo* experiments:

1) ChrimsonR + stGtACR2 (AAV2/9-CAG-DIO-ChrimsonR-tdTomato and

AAV2/9-EF1a-DIO-stGtACR2-EGFP-Kv2.1, 1:1);

2) mCherry + EGFP (AAV2/9-EF1a-DIO-mCherry and AAV2/9-EF1a-DIO-EGFP,

1:1);

3) ChrimsonR + EGFP (AAV2/9-CAG-DIO-ChrimsonR-tdTomato and

AAV2/9-EF1a-DIO-EGFP, 1:1)

4) stGtACR2 + mCherry (AAV2/9-EF1a-DIO-stGtACR2-EGFP-Kv2.1 and

AAV2/9-EF1a-DIO-mCherry, 1:1)

A bolus of 0.4 μL of mixed virus is injected into the VTA at 46 nL/min. The pipette is

held in place for 10 min after the injection and then slowly withdrawn. For real-time

place preference (RTPP) behavioral test, the dual-color LED probe implantation is

carried out after virus injection and diffusion. The probe was slowly implanted at the

side of the target brain areas with an angle of 45° to midline (coordinates 3.64 mm

posterior to the bregma, 0.8 mm lateral to midline, and 4.85 mm ventral to the skull

surface). The probe is fixed to the skull with Cyanoacrylate adhesive (TONSAN

1454) and dental cement. For dopamine detection with fiber photometric recording,

the AAV-hSyn-GRAB$_{DA2m}$ (0.2 μL) is injected into the NAc (coordinates: 1.18 mm

front to the bregma, 0.75 mm lateral to midline, and 3.75 mm ventral to the skull

surface). Optical fiber implantation is carried out after virus injection and diffusion. A

piece of optical fiber (FT200UMT, 200 μm O.D., 0.39 NA, Thorlabs) is fit into an

LC-sized ceramic fiber ferrule (Shanghai Fiblaser, China) and implanted over the

target brain areas with the tip 0.1 mm above the virus injection sites in the NAc. The

ceramic ferrule is fixed to the skull with Cyanoacrylate adhesive (TONSAN 1454)

and dental cement. After the virus injection and device implantation, mice are housed

individually to prevent potential damage to the probe. All subsequent experiments are

performed at least 2 weeks after injection to allow sufficient time for transgene

expression and animal recovery.

Immunohistochemistry

The micro-LED probes are implanted into the VTA region and fixed to the skull with

Cyanoacrylate adhesive (TONSAN 1454) and dental cement. After 1 day or 3 weeks

following the surgery, animals are transcardially perfused with 4% PFA in 0.1 M PBS.

Brains are postfixed in 4% PFA and 40-μm-thick vibratome sections are prepared.

After 5-min PBS rinses for 3 times, the sections are blocked in PBST (PBS + 0.3%

Triton X-100) with 3% bovine serum albumin for 1 hour, and incubated in rabbit

anti-Iba1 (1:1000, Wako, Cat.#019-19741) and rabbit anti-GFAP antibody (1:2000,

Sigma, Cat#AB5541) dissolved in PBST at 4 °C for 24 hours. After 5-min washes in

PBS for 5 times, the sections are incubated with Alexa Cy3-conjugated goat

anti-rabbit antibody (1:500, Jackson ImmunoResearch) and Alexa 488 goat

anti-chicken antibody (1:500, Thermo Fisher Scientific) dissolved in PBST for 2

hours at room temperature followed by 5-min washes in PBS for 3 times. At last, the

samples are coverslipped in 50% glyceroland slides and photographed with an

automated slide scanner (VS120 Virtual Slide, Olympus) or a confocal scanning

microscope (DigitalEclipse A1, Nikon). The numbers of DAPI, stGtACR2 and

ChrimsonR stained cells or the numbers of Iba1+ microglia and GFAP+ astrocytes are

analyzed with Fiji software (https://fiji.sc/) and their percentages in cell population

indicated by DAPI are calculated.

In vitro optogenetics evaluation

Electrophysiological properties of the neurons in brain slices are measured with

whole-cell patch-clamp recording technique following a previously described

procedure.[12] After two weeks of virus expression (ChrimsonR + stGtACR2), mice are

deeply anesthetized with pentobarbital (100 mg/kg i.p.) and intracardially perfused

with 5 mL ice-cold oxygenated modified artificial cerebrospinal fluid (ACSF) at a rate

of 2 mL/min. After perfusion with modified ACSF containing (in mM): 225 sucrose,

119 NaCl, 2.5 KCl, 0.1 $CaCl_2$, 4.9 $MgCl_2$, 1.0 NaH_2PO_4, 26.2 $NaHCO_3$, 1.25 glucose,

3 kynurenic acid, and 1 Na L-ascorbate (Sigma-Aldrich), brains are quickly removed

and placed in ice-cold oxygenated ACSF containing (in mM): 110 choline chloride,

2.5 KCl, 0.5 $CaCl_2$, 7 $MgCl_2$, 1.3 NaH_2PO_4, 25 $NaHCO_3$, 20 glucose, 1.3 Na

ascorbate, and 0.6 Na pyruvate. Sagittal sections (200 μm thick) are prepared using a

Leica VT1200S vibratome. Brain slices are incubated for 1 hour at 34 °C with ACSF

saturated with 95% O_2/5% CO_2 and containing (in mM) 125 NaCl, 2.5 KCl, 2 $CaCl_2$,

1.3 $MgCl_2$, 1.3 NaH_2PO_4, 25 $NaHCO_3$, 10 glucose, 1.3 Na ascorbate, and 0.6 Na

pyruvate. The internal solution within whole-cell recording pipettes (3–6 MΩ)

contains (in mM): 130 K-gluconate, 10 HEPES, 0.6 EGTA, 5 KCl, 3 Na_2ATP, 0.3

Na_3GTP, 4 $MgCl_2$, and 10 Na_2-phosphocreatine (pH 7.2–7.4). Voltage-clamp

recordings are performed using a MultiClamp 700B amplifier (Molecular Devices).

Traces are low-pass filtered at 2.6 kHz and digitized at 10 kHz (DigiData 1440,

Molecular Devices). The data are acquired and analyzed using Clampfit 10.0 software

(Molecular Devices). Red and blue illuminations are provided by a fiber coupled to

two laser sources (wavelengths 633 nm and 473 nm).

Real-time place preference test

After 2 weeks of recovery and AAV expression in *DAT-cre* mice, real-time place

preference / avoidance (RTPP) tests are performed in a custom-made

two-compartment apparatus. The RTPPT apparatus comprises two white plastic

chambers (size: 25 × 25 × 30, L × W × H in cm), separated by a half-open opaque

plastic walls in the middle (8 × 30, L × H cm). A RTPP test consists of 3 phases:

10-min baseline (pretest), 30-min red LED stimulation, 30-min rest (no stimulation),

and 30-min blue LED simulation. During the pretest phase, individual mice are

allowed to freely explore the entire apparatus for 10 min. Animals that exhibit a

strong initial preference or avoidance for one of the side chambers (preference index >

60% or < 40%) are excluded from further experiments. During the red LED

stimulation phase, animals that entered the left chamber receive red light pulses (20

Hz, 20-ms pulse, 7 mA). During the blue light stimulation phase, animals that entered

the left chamber receive constant blue light (continuous, 5 mA). The locomotion of

animals is assessed from the video recording data using a custom-developed

MATLAB program.

 Fiber photometry for dopamine (DA) detection

Details for the fiber photometric system are described previously.[13] The excitation

source is a 488 nm semiconductor laser (Coherent, Inc. OBIS 488 LS, tunable power

up to 60 mW). A dichroic mirror with a 452–490 nm reflection band and a 505–800

nm transmission band (Thorlabs, Inc. MD498) is employed for wavelength selection.

A multimode fiber (Thorlabs, Inc., 200 μm in diameter and 0.39 in numerical

aperture) coupled to an objective lens (Olympus Inc,10×, NA 0.3) is used for optical

transmission. The fluorescence signals are collected with a photomultiplier tube

(PMT) (Hamamatsu, Inc. R3896) after filtering by a GFP bandpass filter (Thorlabs,

Inc. MF525-39). An amplifier (C7319, Hamamatsu) is used to convert the current

output from the PMT to voltage signals, which are further filtered through a lowpass

filter (40 Hz cut-off; Brownlee 440, USA). The fluorescence signals are digitalized at

100 Hz and recorded by a Power 1401 digitizer and a Spike2 software (CED, UK).

Photometric data are exported and analyzed with MATLAB. The fluorescence signals

($\Delta F/F_0$) are processed by averaging the baseline signal F_0 over a 4.5-s long control

time window, and presented as heatmaps or per-event plots. AUC (area under the

curve) of DA signal is used for statistical analysis. AUC is the integral of a curve that

describes the variation of DA signals during light stimulation.

References

1.	Ding H, Lu L, Shi Z, Wang D, Li L, Li X, *et al.* Microscale optoelectronic infrared-to-visible upconversion devices and their use as injectable light sources. *Proc Natl Acad Sci U S A* 2018, **115**(26): 6632-6637.

2.	Li L, Liu C, Su Y, Bai J, Wu J, Han Y, *et al.* Heterogeneous Integration of Microscale GaN Light-Emitting Diodes and Their Electrical, Optical, and Thermal Characteristics on Flexible Substrates. *Adv Mater Technol* 2018, **3**(1): 1700239.

3.	Liu C, Zhang Q, Wang D, Zhao G, Cai X, Li L, *et al.* High Performance, Biocompatible Dielectric Thin-Film Optical Filters Integrated with Flexible Substrates and Microscale Optoelectronic Devices. *Adv Opt Mater* 2018, **6**(15): 1800146.

4.	Gebhart SC, Lin WC, Mahadevan-Jansen A. In vitro determination of normal and neoplastic human brain tissue optical properties using inverse adding-doubling. *Phys Med Biol* 2006, **51**(8): 2011-2027.

5.	Zhao Y, Liu C, Liu Z, Luo W, Li L, Cai X, *et al.* Wirelessly Operated, Implantable Optoelectronic Probes for Optogenetics in Freely Moving Animals. *IEEE Trans Electron Devices* 2019, **66**(1): 785-792.

6.	Peirson SN, Brown LA, Pothecary CA, Benson LA, Fisk AS. Light and the laboratory mouse. *J Neurosci Methods* 2018, **300:** 26-36.

7.	Klapoetke NC, Murata Y, Kim SS, Pulver SR, Birdsey-Benson A, Cho YK, *et al.* Independent optical excitation of distinct neural populations. *Nat Methods* 2014, **11**(3): 338-346.

8.	Govorunova EG, Sineshchekov OA, Janz R, Liu X, Spudich JL. Natural light-gated anion channels: A family of microbial rhodopsins for advanced optogenetics. *Science* 2015, **349**(6248): 647.

9.	Mahn M, Gibor L, Patil P, Cohen-Kashi Malina K, Oring S, Printz Y, *et al.* High-efficiency optogenetic silencing with soma-targeted anion-conducting channelrhodopsins. *Nat Commun* 2018, **9**(1): 4125.

10.	Sun F, Zhou J, Dai B, Qian T, Zeng J, Li X, *et al.* Next-generation GRAB sensors for monitoring dopaminergic activity in vivo. *Nat Methods* 2020, **17**(11): 1156-1166.

11.	Wu X, Dong X, Wu Z, Cao H, Niu D, Qu J, *et al.* A novel method for purification of recombinant adenoassociated virus vectors on a large scale. *Chinese Sci Bull* 2001, **46**(6): 485-488.

12.	Lu L, Ren Y, Yu T, Liu Z, Wang S, Tan L, *et al.* Control of locomotor speed, arousal, and hippocampal theta rhythms by the nucleus incertus. *Nat Commun* 2020, **11**(1): 1-16.

13.	Lu L, Wang R, Luo M. An optical brain-to-brain interface supports rapid information transmission for precise locomotion control. *Sci China Life Sci* 2020, **63**(6): 875-885.